The Foster method: Rapid and non-invasive detection of clinically significant

American Foulbrood disease levels using eDNA sampling and a dual-target

qPCR assay, with its potential for other hive pathogens.

John F. Mackay[1]*, Rebecca E. Hewett[1] , Noa T. Smith[1] Tammy L. Waters[1], John S.

Scandrett[2]

[1] dnature diagnostics & research Ltd, 60 Carnarvon St., Gisborne, New Zealand

[2] Scandrett Rural Ltd, Invercargill, New Zealand

*Email: john@dnature.co.nz

Abstract

Clinical signs of American Foulbrood (AFB) can be difficult to diagnose and thus disease is often missed and leads to further spread. Diagnosis is centred on the beekeeper's skill in recognising clinical symptoms – a highly subjective and time-consuming activity. Previous testing methods have relied on sampling that necessitates dismantling the hive and/or requires multiple visits to retrieve passive samples. The Foster method is a novel environmental DNA sampling method using colony entrance swabs together with a dual-target qPCR for *Paenibacillus larvae*: the causative bacteria of AFB disease. The quantification data generated can be used to detect hives with clinically significant infections, even before visual symptoms are apparent. Such a sampling method will be applicable to other bee pathogens and incursion pests.

Keywords: honeybee, American foulbrood, entrance, qPCR, diagnosis, swab, quantification, non-invasive.

Importance

Discovery of American foulbrood disease in a honeybee colony typically means the destruction of the bees and hive by burning, in New Zealand and many other countries. This discovery is typically by visual examination of capped brood by the beekeeper - a subjective skill that means the disease is being missed or not recognised. It is a time-consuming and exacting process to inspect hives for AFB. Here we present a novel rapid sampling method that does not require opening/ dismantling the hive, in conjunction with a dual-target quantitative PCR assay for the bacteria responsible, *Paenibacillus larvae*. Using the resulting quantitative data,

hives presenting visual clinical symptoms or likely to soon become clinical, can be

determined and the hives dealt with appropriately before further spread occurs. This

study provides the basis for a novel way of sampling for honeybee pathogens and

pests.

Introduction

American Foulbrood (AFB) is one of the most destructive diseases in honeybees

(*Apis mellifera*) and is caused by spores of the Gram-positive bacteria *Paenibacillus*

larvae infecting bees during their larval stage. These spores are extremely hardy;

being resistant to heat, caustic, and other chemical treatments with the ability to

remain infectious for over 30 years (Genersch, 2010). Only a few spores are

required to cause an infection in a new larva and yet a single infected larvae can

generate many millions of spores - leading to rapid spread and decline in a bee

colony. Once the colony has weakened or died then it may be 'robbed' by bees from

neighbouring hives where contaminated honey is brought back to a neighbouring

hive, which leads to further infections. Infection is also spread unknowingly by the

beekeeper (Fries and Camazine, 2001), through the introduction of hive frames

carrying an undetected infection into a new colony of bees or the feeding of

contaminated honey to colonies.

In New Zealand, hive numbers have more than doubled since 2013

(www.afb.org.nz). Similarly, the number of beekeepers has also increased –

primarily due to hobby and conservation efforts (several hives per beekeeper) but

also in the commercial sector (hundreds or thousands of hives per operation). Much

of the growth in hive and commercial beekeeping has been due to the increased

demand for high-value New Zealand manuka honey. The increase in hive numbers

(and densities) combined with newer beekeepers, who are less likely to be skilled in

identifying the clinical signs of AFB, has led to an increase in the percentage of hives

reported by beekeepers as infected from 0.21% in 2015 to 0.46% in 2021

(www.afb.org.nz).

The diagnosis of AFB is based primarily on visual symptoms – typically unusual,

capped brood cells on a beehive frame that warrant investigation through a 'roping'

field test. In a hive with three to seven frames of brood or more, this means typically

thousands of brood cells to visually examine per hive. Other diseases mediated by

varroa can affect brood cells meaning that examining suspicious cells for clinical

AFB signs is laborious and prone to disease being missed. Clinical signs can vary

based on the AFB genotype present, with ERIC II strains exhibiting reduced clinical

symptoms (Genersch, 2010). In addition, hives must be dismantled, and the colony

disrupted to inspect for AFB.

Confirmation of AFB in beehive materials has been traditionally performed by

culture, using selective media, and induced germination of spores. However, this is a

variable process (Forsgren, 2008) and may take up to a week for results. PCR and

qPCR (quantitative PCR) have been demonstrated as effective tools for the

detection of *P. larvae*, with qPCR being the preferred diagnostic tool in recent years

due its faster time to results and lower chance for amplicon contamination. However,

most applications have not used the quantitative data generated; rather using the

tool for rapid confirmation of culture isolates (Kňazovická et al, 2011). The

quantification aspect is important as low levels of AFB are often not clinically relevant

(Bassi et al., 2018 , Pernal and Melathopoulos, 2006) and such hives may never

develop clinical symptoms. Much of this low level may be due to the hygienic

behaviour of bees to reduce overt infections (Spivak and Reuter, 2001).

Most molecular publications relating to AFB to date have relied on either

conventional PCR or SYBR Green-based qPCR and the sole use of the 16S gene

due to its high sensitivity as a multicopy target (Dobbelaere *et al.*, 2001; Piccini *et al.*,

2002). However this sensitivity can be at the expense of specificity, as evidenced by

the initial design (Martinez, 2010) and subsequent re-evaluation/re-design of qPCR

primers to improve specificity (Rossi *et al.*, 2018). Sensitivity for bacterial targets can

also be compromised by deletions of target sequences or sequence polymorphisms

which affect primer or probe binding (Dahlberg *et al.*, 2018; Johansen *et al.*, 2019;

Xiu *et al.*, 2021) and many microbiological qPCR tests now routinely utilise two

independent targets to enhance both sensitivity and specificity, such as many of the

commercial qPCR tests for chlamydia and the virus behind COVID-19: SARS-CoV-2.

Hydrolysis probe ("TaqMan®") qPCR assays permit quantification of each target by

selected fluorescent wavelengths (Kušar *et al.*, 2021) while also allowing the use of

an internal control in an additional detection channel for confirmation of suitable

quality DNA present in the reaction. These internal controls may improve the

sensitivity of previous studies whereby the presence of inhibitors possibly prevented

the amplification of *P. larvae* targets. (Forsgren and Laugen, 2013)

We used a commercially available multiplex qPCR assay for *P. larvae* (Hall et al.,

2021) and developed a new and rapid environmental DNA (eDNA) sampling method

for the detection of hives having - or at risk of developing - clinical AFB infections.

Previously *P. larvae* has been detected in debris from a hive baseboard (Bassi et al.,

2018) but this study required installing collection sheets, necessitating dismantling

the hive twice as well as repeat visits. The ability to detect relevant levels of *P. larvae*

from a single sample taken from the entrance of a hive would permit far more rapid

sampling of hives and allow infected hives to be dealt with more quickly (Lyall et al.,

2019). In this study, we evaluated and compared the levels of *P. larvae* among bees,

the entire baseboard, and the entrance region of the baseboard.

Methods

Quarantining hives in an apiary upon discovery of a clinical AFB hive is a common

international practice, in order to minimise spread of the disease. From five such

quarantine apiaries around New Zealand (East Coast, Wellington, Wairarapa and

South Canterbury regions) multiple hives were sampled two-monthly from late 2018

to early 2020. In addition, samples from visually clinical hives and neighbouring hives

were collected by local apiary inspectors for testing during this time. Samples

included nurse bees, swabs from the hive entrance and swabs from over the entire

baseboard. Subsequent samples comprised of comparisons between bees and

solely the hive entrance. Bees were collected from the brood frames using 50 mL

sterile containers and frozen to euthanise the bees. Sterile foam-tipped swabs

(Puritan) were moistened in sterile water before swabbing either the whole

baseboard or just through and across the hive entrance for three to five seconds.

Instructions were provided to participating beekeepers to rotate the swabs during

collection to ensure all swab surfaces could collect hive material. Swab heads were

then snapped off into a 2 mL microcentrifuge tube, capped and sent to the laboratory

with the bee samples, at ambient temperature by overnight courier. Once in the

laboratory, bee samples were frozen at -20°C and swabs were stored at ambient

temperature until DNA extraction was performed.

Genomic DNA extraction from bees and associated bacteria was performed from

each sample containing 10 bees. DNA was extracted using the Bee Pathogen

DNA/RNA Extraction Kit (dnature diagnostics & research Ltd, Gisborne, New

Zealand); a column-based kit in conjunction with beadbeating. 10 bees were used for

the practical purpose of ease of sample handling and throughput. In short, 3 mL DXL

lysis buffer supplied in the kit was added to 10 bees in a 5 mL tube, containing a

mixture of 0.5 mm (~2.4 g) and 2.3 mm zirconia silica (~1 g) beadbeating beads

(BioSpec Products, Bartlesville, USA). The bees were homogenised in a Mini

Beadbeater 16 instrument (BioSpec Products, Bartlesville, USA) for 3 minutes.

Tubes were then incubated at 65°C for 10 minutes with manual inversion every 3

minutes. 1 mL of the homogenate was transferred to a 2 mL microcentrifuge tube

and centrifuged at 15,000 x g for 5 minutes. 500 µL of supernatant was transferred to

a new microcentrifuge tube containing 450 µL AD buffer and mixed well. The solution

was then applied to a nucleic acid binding column from the kit (in two passes) and

centrifuged through before successive washing steps and elution of nucleic acids

from the column in 100 µL elution buffer.

A similar process was used for swabs: they were placed in a screwcap 2 mL tubes

with 0.5mm zirconia silica beads (~0.8g) and 2.3 mm zirconia silica beads (0.3g).

800 µL DXL buffer was added to each tube and the swab material homogenised in

the Mini Beadbeater 16 for 3 minutes. The tubes were then incubated at 65°C with

shaking for 10 minutes, before centrifuging and processing 500 µL of the resulting

supernatant as for the bees. For pooled applications, individual swabs were added to

500 µL DXL buffer in a separate 2 mL microcentrifuge tube and vortexed for 10

seconds to dislodge cells and bacteria off the swab. 60 µL of each swab eluate (up

to twelve swabs) was added to a 2 mL beadbeating tube as for single swabs and

processed as for single swabs from the beadbeating step.

Assessment of DNA quality and presence of potential inhibitors was performed by

measurement (spectroscopy) on a DS-11 FX (DeNovix, Wilmington, USA) and

OD260/280 measurements recorded. Dilutions of extracted DNA were also tested by

qPCR to measure the expected Cq increases with increasing dilution and thus

monitor any qPCR inhibition.

The AFBduo qPCR test (dnature diagnostics & research Ltd, Gisborne, NZ) is a

commercially available multiplex hydrolysis probe assay that uses two targets for *P.*

larvae: a single copy gene for quantification purposes (ftsZ, a prokaryotic homologue

to tubulin) and the multicopy 16S gene. The ftsZ hydrolysis probe is detected in the

FAM channel and the 16S target in the HEX channel. The test also includes a

simultaneous internal control hydrolysis probe reaction (ROX channel) for an animal

mitochondrial marker, to confirm validity of results (in the cases where *P. larvae* was

not detected) by way of successful internal control amplification.

Each AFBduo qPCR reaction was performed in a 10 μL volume and consisted of 5.5

μL PCR grade water, 0.5 μL of the 20X Oligo Mix and 2 μL of the 5X Mastermix (all

reagents supplied in the AFBduo real-time PCR kit, dnature). 2 μL genomic DNA,

positive control DNA (DNA extracted from larvae with clinical disease) or sterile

water (for no template controls) was added per reaction. The amplifications were

performed on the Mic qPCR Cycler (BioMolecular Systems, Australia) or the Eco

qPCR instrument (Illumina, San Diego, USA) and the cycling conditions comprised:

denaturation at 95°C for 2 minutes followed by 40 cycles of 95°C for 5 seconds and

60°C for 15 seconds for both instruments. Data was acquired on the Green (ftsZ),

Yellow (16S) and Orange (Internal control) channels of the Mic cycler or FAM, HEX

and ROX channels on the Eco instrument. Cq's (Cycle quantities) were automatically

generated by the Mic software, using the dynamic analysis setting (based on second

derivative maximum method).

A synthetic DNA standard (gBlock, IDT, Singapore) was utilised that incorporated

both amplicon sequences (available on request). The synthesised 153 bp standard

was resuspended in TE buffer, pH8, quantified in nanograms per µL on a DS-11 FX

microvolume spectrophotometer (DeNovix, Wilmington, USA) and its copy number

calculated using the equation:

Copies = ng * Avogadro's constant copies/mole
189   length of standard x 660 g/mole x 1×10^9 ng/g
190
191
Quantification standards were generated from 10^6 copies per µL down to 2 copies

per µL (Svec et al., 2015). For all dilutions, the diluent was a pooled DNA sample

extracted from bees free of *P. larvae*. Dilutions of extracted DNA from samples

positive for *P. larvae* were diluted in a pool of negative bee DNA.

The standards and their resulting Cq's (supplementary S1) were also used to convert

the Cq's resulting from bee and swab samples to spore levels using the amplification

efficiency of the single copy ftsZ gene, the y-intercept, and the formula:

Spores per swab or 10 bees = $10^{(Cq\ ftzZ-39/-3.48)}$ x 50

Statistical analysis was performed after imputing data points by regression analysis

of the other paired data, for two clinical hives (due to spore numbers being skewed

lower by the new hiveware baseboards at time of sampling) resulting in a total of 26

data points with paired clinical data. A binary logistic regression model was

performed in Minitab (Minitab®, State College, PA, USA) to identify the relationship

between spore numbers at the entrance and clinical disease likelihood, and analysis

of variance calculated to confirm the correlation between increased entrance spores

and the clinical signs of AFB. A *p*-value of <0.05 was considered significant.

Results

Testing of the AFBduo assay on synthetic DNA standards showed high analytical
sensitivity, with both target reactions in the assay able to reliably detect down to 4
copies per 10 µL reaction. The reaction targets amplified with high efficiency in the
multiplex (ftsZ = 94% and efficiency for 16S = 96%) as well as excellent linearity
over the 7 orders of magnitude (r = 0.998). On testing *P. larvae*-positive bee, honey,
and hive material, the 16S target typically amplified 3-5 cycles earlier (i.e. lower
Cq's) than the ftsZ target as expected, due to the multiple copies of the 16S
sequence per AFB bacterium compared to the single copy ftsZ gene.

The DNA extraction method from bees generated high purity DNA as shown by
measurements on the microvolume spectrophotometer with OD260/280
measurements of between 1.8 and 1.9. Extracted DNA samples serially diluted in
elution buffer showed the expected increase in Cq's with decreasing concentration,
indicating no inhibitors were present in the samples following DNA extraction. DNA
yields from swabs were too low to be measured on the spectrophotometer, however
internal control Cq's for serial dilutions as for the bees, indicated the DNA was
similarly of suitable quality.

In this study, we evaluated and compared the levels of *P. larvae* among bees, the
entire baseboard, and the entrance region of the baseboard. In the first cohort of
quarantine beehives from multiple regions in NZ (beehives from quarantine apiaries
i.e. an apiary that had contained a hive exhibiting clinical signs), 45 hives were
sampled up to 6 times during the 2018-2019/20 seasons. The samples taken were
nurse bees, swabs from the entire baseboard and entrance swabs. Among these
hives, 8 were found to have clinical signs of AFB ('roping' larval remains) during

sampling while another 6 were presumed to have AFB due to their condition and the

high prevalence in the apiary. Examples of the progression of *P. larvae* levels in

three of the colonies is shown in Table 1, where the resulting Cq data has been

converted to spores per 10 bee or swab sample processed.

In a second cohort of hives from one beekeeper, 23 hives were tested in a

quarantine apiary using bee and hive entrance swab samples. Four of these hives

showed clinical signs of AFB. Levels of AFB in nurse bee samples were compared

to the levels indicated by the entrance swabs. Once again, most hives showed no or

low levels of *P. larvae*, while those with clinical AFB demonstrated much higher *P.*

larvae spore levels in both the bee and entrance swab samples as shown in Table 2.

A third cohort of hives came from a beekeeper with a high incidence of AFB that was

introduced through the purchase of contaminated equipment into his operation. Here

we performed initial trials on pooling swab eluates into single extractions to reduce

time and costs. Swab eluates were pooled into groups of eight and later twelve, the

DNA extracted from the pooled sample and tested for the presence of P. larvae

DNA. The remnant samples from pools that tested positive for both markers, were

then tested individually as described. The results (Table 3) show that that even in the

pools of twelve hive samples, single hives with higher levels of AFB could be clearly

detected within a pool, prompting testing of those hives contributing to the positive

pool to identify potentially clinically affected hive(s).

Statistical analyses using a binary logistic regression model (Minitab) and analysis

of variance (Supplementary figure 4) indicated the strong relationship between the

increasing spore levels and clinical signs ($p = 0.00$). Indeed, in this work a spore

quantity of 565,000 or more, revealed a 95% probability of clinical signs being

present in the hive.

Discussion

The AFBduo assay provided high confidence in the results due to the integrated

internal control and detection of two markers for AFB. As expected, the 16S marker

had higher analytical sensitivity (amplifying 3-5 cycles earlier) due to the multi-copy

nature of this marker (Dingman, 2017). However, the variability in Cq difference

between the two markers among samples indicated the 16S copy number possibly

varied among AFB isolates and thus could not be relied on for quantification (Dahllof

et al., 2000). This variability in copy number for diagnostic targets has been noted in

other diagnostic qPCR assays for AFB (Rossi et al., 2018; Dainat et al., 2018).

Therefore the use of the single copy ftsZ gene was used for quantification purposes

using a standard curve prepared with a quantified synthetic template diluted in AFB-

negative bee DNA.

Three sample types were compared in the first cohort – bees from the brood nest

(nurse bees), swabs from the whole baseboard surface and swabs through the

entrance of the hive only. The ability to sample the hive through the entrance

removed the need to dismantle and disrupt the hive as with the first two sample

types (bees and whole baseboard swabbing). The levels of both *P. larvae* and

internal control DNA seen on the swabs taken from the whole baseboard or just the

entrance area of the baseboard were similar (Table 4) possibly due to the maximum

amount able to be collected on a given swab. This showed dismantling the hive to

swab the whole baseboard was not required, in order to detect the higher levels of *P.*

larvae in clinically infected colonies. While the spore levels varied among clinically

affected hives, they were still easily differentiated from the very low or undetectable

spore levels in non-infected hives. While spore levels never increased during the

monitoring time for many of the hives or became undetectable in some low spore

cases, some of the hives showed increasing levels at each testing point before

clinical signs were observed in the respective hive and the hive destroyed (Table 1).

Spore levels were the highest in the bees from the brood nest area of the hive and

while other studies have employed up to 100 bees per sample for culturing, the use

of 10 bees permitted easier and more convenient handling for DNA extraction.

Despite the lower spore amounts on swabs, the levels from the entrance swabs

could clearly be differentiated from levels seen in hives not exhibiting any visual

signs of AFB (Table 2). While most hives without clinical signs of AFB had low or

undetectable levels of *P. larvae* DNA detected, one hive without visual signs of AFB

(Hive 11) showed levels similar to two other hives with clinical AFB signs. Indeed, at

the next inspection by the beekeeper (approximately two weeks after sampling), hive

11 showed clinical signs of AFB and was destroyed thus demonstrating the ability of

qPCR to find infected hives that may have pre-clinical infections or have symptoms

missed by visual inspection.

However these results also demonstrated the importance of knowing the history of

the hives sampled. In two cases (hives 11 and 12 in Table 2), the bees showed high

levels of *P. larvae* spores while the hiveware showed low or undetectable levels.

Subsequent investigation showed that the hiveware was new (< 2 weeks old at time

of sampling) for both these hives and that the colonies were newly established from

an unknowingly infected parent colony. Therefore, while the bees carried high levels

of *P. larvae* there had not been enough time for spores to be deposited around the

hive entrance surfaces. Further research will be required to assess this accumulation

rate of spores on the hive as well as further refine the clinical spore threshold levels

as more hives are tested over time.

Given the ability to detect clinically relevant levels of AFB disease through this

simple hive entrance swabbing method, preliminary work with these entrance swab

samples indicate that the same DNA sample can be used to test for other bee

pathogens. *Nosema ceranae* has been detected through the use hive debris and

qPCR (Copley *et al.*, 2012) and showed good correlation of levels between bees and

debris. Since the development of this current work, a report has used hive debris

collected over the period of a week to estimate loading of European foulbrood (a

disease not present in New Zealand currently). As with the results described here,

the bees offered a higher sensitivity of pathogen detection, yet the hive debris results

still provided predictive information as compared to non-infected hives (Biová *et al.*,

2021). Another application could for biosecurity surveillance for unwanted hive pests

such as tracheal mites (*Acarapis woodi*), tropilaelap mites or small hive beetle

(*Aethina tumida*). The use of the sampling technique here combined with an

appropriate qPCR assay (e.g. Li et al., 2018; Ward *et al.*, 2007) could provide rapid

detection of any incursion to countries such as New Zealand (where tracheal and

tropilaelap mites as well as small hive beetle are not found) by providing faster

sampling of hives than dismantling hives to extract hive debris for testing (Ward *et*

al., 2007). The quantification provides an important risk element of the hive as

demonstrated - as to whether visual signs are likely present at the time of sampling

or likely to demonstrate these clinical signs in the near future. *P. larvae* has been

detected from winter debris previously using conventional PCR (Ryba et al., 2009)

and thus the quantification described in this current work, was not possible. All the

methods published to date have required either dismantling the hive to sample from

the baseboard or repeated visits to the hive to insert and remove sampling sheets. In

contrast the rapid sampling method described here - that we have named the Foster

method - is a fast, one-time sampling and quantification method amenable to

composite testing and one that provides clinically significant information as to the

AFB status of the hive.

References

Bassi S, Carpana E, Bergomi P, Galletti G. 2018. Detection and quantification of
Paenibacillus larvae spores in samples of bees, honey, and hive debris as a tool for
American foulbrood risk assessment. Bulletin of Insectology 71 (2): 235-241.

Biová J, Charrière J-D, Dostálková S, Škrabišová M, Petřivalský M, Bzdil J, Danihlík,
J. 2021. *Melissococcus plutonius* Can Be Effectively and Economically Detected
Using Hive Debris and Conventional PCR. Insects 12, no. 2: 150

Copley T, Giovenazzo P, Jabaji S. 2012. Detection of *Nosema apis* and *N. ceranae*
in honeybee bottom scraps and frass in naturally infected hives. Apidologie 43, 753–
760.

Dahlberg J, Hadad R, Elfving K, Larsson I, Isaksson J, Magnuson A, Fredlund H,
Unemo M, Herrmann B. 2018. Ten years transmission of the new variant of
Chlamydia trachomatis in Sweden: Prevalence of infections and associated
complications. Sex. Transm. Infect., 94, 100–104.

Dahllöf I, Baillie H, Kjelleberg S. rpoB-based microbial community analysis avoids
limitations inherent in 16S rRNA gene intraspecies heterogeneity. Appl Environ
Microbiol. 2000 Aug;66(8):3376-80.

Dainat B, Grossar D, Ecoffey B, Haldemann C. 2018. Triplex real-time PCR method
for the qualitative detection of European and American foulbrood in honeybee. J.
Microbio. Methods. 146:61-63.

Dahllöf I, Baillie H, Kjelleberg S. rpoB-based microbial community analysis avoids
limitations inherent in 16S rRNA gene intraspecies heterogeneity. Appl Environ
Microbiol. 2000;66(8):3376-3380.

Dingman DW. Four Complete Paenibacillus larvae Genome Sequences. Genome
Announc. 2017 Jun 1;5(22):e00407-17. doi: 10.1128/genomeA.00407-17

Dobbelaere W, de Graaf DC, Peeters JE, Jacobs FJ. 2001. Development of a fast
and reliable diagnostic method for American foulbrood disease (Paenibacillus larvae
subsp. larvae) using a 16S rRNA gene based PCR. Apidologie 32, 363–370.

Forsgren E, Stevanovic J, Fries I. 2008. Variability in germination and in temperature

and storage resistance among Paenibacillus larvae genotypes, Vet. Microbiol. 129,

342–349.

Forsgren E, Laugen A. 2014. Prognostic value of using bee and hive debris samples

for the detection of American foulbrood disease in honeybee colonies. Apidologie 45,

10–20.

Fries I, Carmazine S. 2001. Implications of horizontal and vertical pathogen

transmission for honeybee epidemiology. Apidologie 32:199–214

Genersch E. 2010. American foulbrood in honeybees and its causative agent,

Paenibacillus larvae. J Invertebr Pathol 103:S10–S19

Govan V, Allsopp M, Davidson S. 1999. A PCR detection method for rapid

identification of *Paenibacillus larvae*. Appl. Environ. Microbiol., 65, 2243–2245.

Hall R, Pragert H, Bernard J. Phiri B, Fan Q, Li X, Parnell A, Stanislawek W,

McDonald C, Ha H, McDonald W, Taylor M. (2021). Apicultural practice and disease

prevalence in Apismellifera, New Zealand: a longitudinal study, Journal of Apicultural

Research 60:5, 644-658.

Johansen T, Kløvstad H, Rykkvin R, Herrfurth-Erichsen E, Sorthe J, Njølstad G.

2019. The 'Finnish new variant of Chlamydia trachomatis' escaping detection in the

Aptima Combo 2 assay is widespread across Norway, June to August 2019. Euro

Surveill 24: 1900592.

Kňazovická V, Miluchová M, Gábor M, Kačániová M, Melich M, Kročko M,

Trakovická A. 2011. Using Real-time PCR for Identification of *Paenibacillus larvae*.

Animal Science and Biotechnologies, 44 (1).

Kušar D, Papić B, Zajc U, Zdovc, I, Golob, M, Žvokelj, L, Knific, T, Avberšek, J,

Ocepek, M, Pislak Ocepek, M. Novel TaqMan PCR Assay for the Quantification of

Paenibacillus larvae Spores in Bee-Related Samples. Insects 2021, 12, 1034.

Li D, Waite D, Fan Q, George S, Semeraro L, Blacket M. 2018. Molecular detection

of small hive beetle Aethina tumida Murray (Coleoptera: Nitidulidae): DNA barcoding

and development of a real-time PCR assay. Sci Rep 8, 9623.

Lyall D, Hansbro P, Horvat J, Stanwell P. 2020. Quantitative Nondestructive Assessment of Paenibacillus larvae in Apis mellifera Hives. In: Ahram T., Karwowski W., Pickl S., Taiar R. (eds) Human Systems Engineering and Design II. IHSED 2019. Advances in Intelligent Systems and Computing, vol 1026. Springer, Cham. https://doi.org/10.1007/978-3-030-27928-8_88.

Martínez J, Simon V, Gonzalez B, Conget P. 2010. A real-time PCR-based strategy for the detection of Paenibacillus larvae vegetative cells and spores to improve the diagnosis and the screening of American foulbrood. Lett. Appl. Microbiol., 50, 603–610.

Pernal S, Melathopoulos A. 2006. Monitoring for American Foulbrood spores from honey and bee samples in Canada. Apiacta, 41, 99–109.

Piccini C; D'Alessandro B; Antúnez K; Zunino P. 2002. Detection of *Paenibacillus larvae* subspecies larvae spores in naturally infected bee larvae and artificially contaminated honey by PCR. World J. Microbiol. Biotechnol., 18, 761–765.

Rossi F, Amadoro C, Ruberto A, Ricchiuti L. 2018. Evaluation of quantitative PCR (qPCR) Paenibacillus larvae targeted assays and definition of optimal conditions for its detection/quantification in honey and hive debris. Insects, 9, 165.

Spivak M, Reuter G. 2001. Resistance to American foulbrood disease by honeybee colonies Apis mellifera bred for hygienic behavior. Apidologie 32:555–565.

Svec D, Tichopad A, Novosadova V, Pfaffl M, Kubista M. How good is a PCR efficiency estimate: Recommendations for precise and robust qPCR efficiency assessments. Biomol Detect Quantif. 2015 Mar 11;3:9-16.

Ward L, Brown M, Neumann P, Wilkins S, Pettis J, Boonham N. 2007. A DNA method for screening hive debris for the presence of small hive beetle (Aethina tumida). Apidologie 38, 272–280.

Xiu L, Li Y, Zhang C, Li Y, Zeng Y, Wang F, Peng J. 2021. A molecular screening assay to identify Chlamydia trachomatis and distinguish new variants of C. trachomatis from wild-type. Microb Biotechnol. Mar;14(2):668-676.

419   Ryba S, Titera D, Haklova M, Stopka P. 2009. A PCR method of detecting American
420   foulbrood (Paenibacillus larvae) in winter beehive wax debris. Vet. Microbiol., 139,
421   193–196.

423   Hive FL5

Date	14 Sept 2018	4 Dec 2018	16 Jan 2019	Clinical status
Bees (spores/10 bees)	-	24,448	779,266	Clinical signs seen 16 Jan after sampling and hive destroyed
Entrance (spores/swab)	ND	10,870	678,363	

425   Hive STF4

Date	22 Sept 2018	Clinical status
Bees (spores/10 bees)	2,591,937	Clinical signs seen early October 2018 and hive destroyed
Entrance (spores/swab)	4,281,312	

427   Hive RL02

Date	31 Oct 2018	27 Dec 2018	11 Mar 2019	23 Aug 2019	Clinical status
Bees (spores/10 bees)	ND	ND	ND	ND	No AFB seen
Entrance (spores/swab)	ND	ND	ND	ND	

429   ND   Not Detected

430   -   Test not performed

432   **Table 1:** AFB spore levels seen in bee samples and entrance swabs from three

433   quarantine hives examples followed over time. Total number of hives followed over

434   time was thirty.

	Clinical signs of AFB?	Bees Spores/10 bees	Entrance Swab Spores/swab
MLHive 1	**Yes**	**37,097,589**	**3,138,994**
MLHive 2	**Yes**	**514,061**	**24,008**
MLHive 3	No	ND	ND
MLHive 4	No	387	1,107
MLHive 5	No	539	377
MLHive 6	No	ND	ND
MLHive 7	No	80	ND
MLHive 8	No	210	ND
MLHive 9	No	205	ND
MLHive 10	**Yes**	**55,866,177**	**962,621**
MLHive 11	No**	9,518,751	ND*
MLHive 12	**Yes**	**12,561,091**	**890***
MLHive 13	No	<10	<10
MLHive 14	No	ND	<10
MLHive 15	No	2049	1,318
MLHive 16	No	ND	861
MLHive 17	No	ND	326
MLHive 18	No	ND	1,092
MLHive 19	No	97	207
MLHive 20	No	ND	11,235
MLHive 21	No	ND	326
MLHive 22	No	ND	201
MLHive 23	No	943,739	2,365

Table 2: Testing an apiary comparing bee and entrance swab samples. The spore

levels marked with * indicate low level hive entrance spore on colonies recently

placed into new hiveware (hives 11 and 12 were started from splits of hive 10). While

at the time of sampling hive 11 showed no AFB symptoms (**), it developed clinical

symptoms 2 weeks later and was destroyed.

ND Not detected

	Pooled result (12) ftsZ spore levels	Individual Swab Result ftsZ spore levels (per swab)	Clinical Signs of AFB (following testing)
Hive SB216		Not detected	No
Hive SB217		Not detected	No
Hive SB218		Not detected	No
Hive SB219		**165,103**	**Yes**
Hive SB220		Not detected	No
Hive SB221	**15,630**	Not detected	No
Hive SB222		Not detected	No
Hive SB250		Not detected	No
Hive SB251		Not detected	No
Hive SB252		Not detected	No
Hive SB253		Not detected	No
Hive SB254		Not detected	No
Hive SB255		-	No
Hive SB256		-	No
Hive SB257		-	No
Hive SB258		-	No
Hive SB259		-	No
Hive SB260	Not Detected	-	No
Hive SB261		-	No
Hive SB262		-	No
Hive SB263		-	No
Hive SB264		-	No
Hive SB265		-	No

Table 3: composite pooling of swabs to rapidly find clinically relevant levels in hives.

In this instance the first pool showed raised levels and the remnant swab eluates

were tested individually, whereupon elevated levels were found in hive SB219.

Subsequent hive inspections showed this hive to have clinical symptoms and it was

destroyed.

– test not performed

Hive	Location	Estimated Spore level (ftsZ target)	Hive status
ML Shearer	Baseboard	52 million	Clinical signs of AFB
	Entrance	25 million	
ML Ellis	Baseboard	3.6 million	Clinical signs of AFB
	Entrance	3.4 million	
FL 10	Baseboard	61	No AFB seen
	Entrance	335	

452   **Table 4:** Comparison of baseboard and entrance swabs showing the differentiation

453   of spore levels among two clinical hives and a hive without clinical symptoms.

Conflict of interest statement

JFM, REH, TLW and NTS are or were previously employees of dnature diagnostics
& research Ltd., that previously developed the commercial AFBduo qPCR kit as well
as the commercial bee DNA extraction kit method used in this study.

Acknowledgements and memoriam

This work was partly funded by Eastland Community Trust (now Trust Tairawhiti)
(JFM, REH., NTS.,TLW) and the Sustainable Food and Fibre Futures from Ministry
of Primary Industries, Alpine Honey and NZ Honey Trust (JSS). We thank all
beekeepers who provided samples. We also thank Richard Hall (Ministry for Primary
Industries) for early discussions, Dr Graeme Wood for statistical assistance and also
Prof. Phil Lester and Dr Bart Janssen for comments on the manuscript. The method
is named for New Zealand beekeeper Barry Foster - a long-time AFB educator,
inspector, and former New Zealand AFB management agency board member. This
paper is dedicated to the memories of Russell Poulter, a long-standing collaborator
who passed away suddenly in January 2021 and beekeeper/AFB inspector John
Maynard who passed away suddenly in February 2021.

477   Supplementary information

Standard (copies/µL)	Cq ftsZ (FAM)	Cq 16S (HEX)	Cq IC (ROX)
1,000,000	17.06	15.6	31.14
100,000	20.73	19.20	30.15
10,000	24.33	22.77	29.42
1,000	27.84	26.22	29.15
100	31.29	29.55	29.15
10	34.94	33.17	29.36
5	36.06	34.07	29.03
2	37.08	36.38	29.39

480   S1 Sensitivity of AFBduo qPCR assay

481   Synthetic standard was diluted in a *P. larvae*-free hive swab DNA matrix from 10^6

482   copies down to 2 copies per µL. Dilutions were amplified in duplicate and the mean

483   Cq values presented.

	Clinical signs of AFB?	Entrance Swab		
		ftsZ FAM Cq	16S HEX Cq	Internal Control
MLHive 1	Yes	22.3	18.4	27.3
MLHive 2	Yes	29.7	25.4	29.3
MLHive 3	No	ND	35.0	28.5
MLHive 4	No	34.3	29.1	28.0
MLHive 5	No	35.9	32.5	28.1
MLHive 6	No	ND	35.8	29.1
MLHive 7	No	ND	ND	27.5
MLHive 8	No	ND	33.9	27.4
MLHive 9	No	ND	34.1	28.7
MLHive 10	Yes	24.1	20.1	28.3
MLHive 11	No	ND	ND	29.2
MLHive 12	Yes	34.6	31.3	27.4
MLHive 13	No	ND	34.1	28.6
MLHive 14	No	36.8	33.2	27.4
MLHive 15	No	26.5	22.5	26.5
MLHive 16	No	34.7	36.5	28.2
MLHive 17	No	36.2	33.8	26.9
MLHive 18	No	34.3	31.0	26.0
MLHive 19	No	36.9	33.4	26.7
MLHive 20	No	30.8	26.9	28.3
MLHive 21	No	36.2	36.3	28.2
MLHive 22	No	36.9	32.3	28.4
MLHive 23	No	33.2	28.8	28.9

492   S2: Cq data for hive series ML – *entrance swab* (spore levels in table 2)

	Clinical signs of AFB?	Bees		
		ftsZ FAM Cq	16S HEX Cq	Internal Control
MLHive 1	Yes	18.5	15.1	25.2
MLHive 2	Yes	25.0	21.0	24.2
MLHive 3	No	ND	34.8	21.9
MLHive 4	No	35.9	34.5	23.2
MLHive 5	No	35.4	31.2	23.7
MLHive 6	No	ND	ND	23.3
MLHive 7	No	30.8	27.0	23.0
MLHive 8	No	36.8	33.2	23.3
MLHive 9	No	36.9	33.3	22.6
MLHive 10	Yes	17.9	14.6	21.2
MLHive 11	No	20.6	16.2	22.0
MLHive 12	Yes	20.2	16.3	23.4
MLHive 13	No	38.2	34.3	21.2
MLHive 14	No	36.7	ND	21.7
MLHive 15	No	25.8	21.5	22.3
MLHive 16	No	ND	35.3	25.0
MLHive 17	No	ND	ND	23.7
MLHive 18	No	ND	ND	20.9
MLHive 19	No	38.0	33.7	22.0
MLHive 20	No	ND	ND	23.4
MLHive 21	No	ND	36.2	22.9
MLHive 22	No	ND	ND	24.2
MLHive 23	No	24.1	20.1	22.3

497   S3: Cq data for hive series ML – *bees* (spore levels in table 2)

502   Analysis of variance

Source	DF	Adj Dev	Adj Mean	Chi-Square	P-Value
Regression	1	15.42	15.4192	15.42	0.000
Entrance Spores/swab	1	15.42	15.4192	15.42	0.000
Error	24	10.04	0.4182		
Total	25	25.46			

504   Supplementary figure 4

505   Binary logistic regression model showing the relationship between increasing

506   entrance swab levels and the presence of clinical disease, and the calculated

507   analysis of variance with associated p-value (0.000).