Nanopore ion sources deliver single amino acid and peptide ions directly into high vacuum

Nicholas Drachman[1], Mathilde LePoitevin[1], Hannah Szapary[1], Benjamin Wiener[1], William Maulbetsch[1] and Derek Stein[1,2]

[1]Physics Department, Brown University, 182 Hope street, Providence, 02906, RI, USA.

[2]School of Engineering, Brown University, 184 Hope street, Providence, 02906, RI, USA.

Abstract

We report a nanopore ion source for mass spectrometry that delivers ions directly into high vacuum from aqueous solutions. The ion source comprises a pulled quartz pipette with a sub-100 nm opening. Ions escape an electrified meniscus by ion evaporation and travel along collisionless trajectories to the ion detector. We measure mass spectra of 16 different amino acid ions, post-translationally modified variants of glutathione, and the peptide angiotensin II, showing that these analytes can be emitted as desolvated ions. The emitted current is composed of ions rather than charged droplets, and more than 90% of the current can be recovered in a distant collector. By circumventing the sample loss mechanisms inherent to conventional electrospray ionization (ESI), where charged droplets are sprayed into a background gas that scatters ions and degrades their transmission, the nanopore ion source could enable more sensitive proteomic analyses.

Keywords: proteomics, nanopore, mass spectrometry, ion source, single molecule

1 Introduction

Mass spectrometry (MS) is the workhorse of proteomics research thanks to its ability to distinguish amino acids and small peptides by their mass to charge ratio [1]. Its utility also significantly derives from the availability of soft ionization techniques for transferring peptide ions into the gas phase intact [2]. In particular, electropray ionization (ESI) transfers analyte into a mass spectrometer via a plume of charged droplets that emerge from a liquid cone-jet at the end of a voltage-biased emitter, as illustrated in Fig. 1a [3, 4]. The droplets pass through a background gas that induces a series of evaporation and Coulomb explosion cycles that ultimately release analyte ions into the gas phase [5]. However, the background gas needed to liberate ions from droplets is also a source of significant sample loss that limits the sensitivity of MS.

The background gas and the plume of charged droplets it creates widely disperse ions, the majority of which collide with the transfer capillary, which bridges the ambient-pressure ion source and the first pumping stage of the mass spectrometer, and other hardware components upstream of the detector [2]. The ion transmission efficiency of an ion source, which refers to the fraction of emitted ions that enters the mass analyzer, is a key determinant of the sensitivity of mass spectrometry [6]. Early ESI sources had emitter tips with diameters of hundreds of micrometers, and only one ion in $\sim 10^4$ reached the mass analyzer [7]. Nano-electrospray ionization (nano-ESI) increased the ion transmission efficiency to $\sim 1\%$ in typical measurements (occasionally reaching as high as 12%) by using emitters with micrometer-scale tips that reduced the flow rates to the range of several nL/min [8–10]. Marginean *et al.* demonstrated that by optimizing the background gas pressure, the sheath gas flow rate, and the ion funnel voltages of a subambient

pressure nano-ESI source, they could achieve an ion utilization efficiency (the fraction of analyte originally in solution reaching the detector) as high as 50% for a particular small peptide [11]. However, co-optimizing the efficiency of multiple analytes is fundamentally challenging because ESI involves processes that physically separate different ion species within the plume [12]. State-of-the-art MS instrumentation still requires thousands to millions of copies of proteins for their identification [13, 14]. This sensitivity falls short of what is desired for single-cell proteomics and single-molecule analyses [15–18]. Achieving single-molecule sensitivity requires an ion source that circumvents the loss mechanisms intrinsic to spraying charged droplets into a background gas.

We have developed a nanopore ion source that emits amino acid and small peptide ions directly into high vacuum from aqueous solution at its tip (Fig. 1b). The heart of the ion source is a pulled quartz pipette with a tip whose inner diameter (ID) is smaller than 100 nm [19, 20]. We reason that the smallness of the tip will influence ion emission in three important ways: First, the surface tension of water can maintain a stable liquid-vacuum interface that supports many atmospheres of pressure when stretched across a nanoscale opening [21]. Second, the fluid flow rate, which scales as the inverse cube of the tip diameter, is too low for a stable electrospray cone-jet to form [22, 23], and this may prevent charged droplets from being emitted altogether. Third, electric fields will concentrate at a sharp, conductive tip like an electrolyte-filled nanopipette, and may reach ∼1 V/nm at the meniscus, the approximate field needed to draw out ions at high rates by the process of ion evaporation [24, 25].

Here we characterize the emission of amino acid and small peptide ions from aqueous solutions directly into high vacuum using a nanopore ion source. We

Fig. 1 Comparison of conventional electrospray ion source with nanopore ion source. a, Schematic of conventional electrospray ionization highlighting the background gas that stimulates evaporation of solvent from droplets, as well as the transfer capillary where significant ion loss occurs. **b**, Schematic of a nanopore ion source showing the liquid-filled nanopipette tip, the extractor electrode, and the extraction voltage V_E applied between them. Inset shows an SEM image of the tip of a pulled quartz nanopipette with a tip ID of 30 nm. **c**, Schematic of the mass spectrometer used in this study. Ion optics comprising an extractor electrode and an einzel lens extract ions from the liquid meniscus at the ion source and focus them through a quadrupole mass filter and an electrostatic ion bender. The transmitted ions strike a channel electron multiplier detector which is sensitive to single ions. **d**, Mass spectrum of 100 mM arginine in aqueous solution obtained with a 41 nm ID nanopore ion source in our quadrupole mass spectrometer.

obtain mass spectra using a custom quadrupole mass spectrometer in which the nanopore ion source operates at pressures below 10^{-6} torr (Fig. 1c) [26]. Separately, we measure the transmission efficiency between the ion source and a downstream Faraday cup for both aqueous salt solutions and aqueous amino acid solutions and find it to exceed that of ESI by two orders of magnitude. We further use a magnetic sector to separate charged droplets from small ions and thereby demonstrate that the nanopore source can be made to emit only ions. We also measured the pH-dependence of amino acid ion emission rates to probe the emission mechanism.

2 Results

2.1 Emitting amino acid ions from a nanopore ion source

We characterize the emission of amino acids from aqueous solution in a custom quadrupole mass spectrometer shown in Fig. 1c [26]. In a typical experiment, ion emission from a nanopore ion source is initiated by applying an extraction voltage, V_E, in the range +190 V to +380 V between the tip and the extraction electrode. The emission current, I_E, is typically in the range of 2-20 pA. Significantly higher currents, ranging from 100s of pA to a few nAs are often measured when working with aqueous salt solutions. The onset of current is abrupt and accompanied by the measurement of ions striking the instrument's detector. We collect easily interpretable mass spectra within seconds to minutes at these low emission currents. We observed that an individual nanopore ion source can be kept in operation for several weeks without clogging.

¹⁰¹ Fig. 1d shows a typical mass spectrum obtained from a 100 mM solution
¹⁰² of arginine in water. This spectrum is obtained in positive ion mode using a
¹⁰³ nanopore ion source with inner tip diameter of 41 nm. Five peaks are clearly
¹⁰⁴ visible. The peak at 175 m/z corresponds to the singly protonated arginine ion
¹⁰⁵ (labelled as Arg^+, shorthand for the protonated form $[Arg + H^+]$). The higher
¹⁰⁶ m/z peaks are all separated by 18 m/z, the shift induced by an additional
¹⁰⁷ water molecule. Thus, the other peaks correspond to solvated arginine clusters
¹⁰⁸ (labelled as $Arg^+(H_2O)_n$, where the integer n denotes the number of attached
¹⁰⁹ water molecules).

¹¹⁰ Figure 2a illustrates how the tip diameter can influence the mass spec-
¹¹¹ trum obtained from an aqueous amino acid solution, again using 100 mM
¹¹² arginine. The spectra shown are obtained using tips with inner diameters of
¹¹³ 20 nm, 125 nm, and 300 nm. The largest tip produces a broad spectrum of
¹¹⁴ peaks that includes the bare arginine ion, eight incrementally hydrated argi-
¹¹⁵ nine ion clusters, and a peak at 349 m/z that corresponds to the arginine
¹¹⁶ dimer ion $((Arg)_2^+)$. The intermediate-sized tip produces a narrower spectrum
¹¹⁷ that includes the bare arginine ion, six incrementally hydrated arginine ion
¹¹⁸ clusters, and a relatively diminished arginine dimer ion peak. The smallest tip
¹¹⁹ primarily produces the bare arginine ion, but attenuated peaks correspond-
¹²⁰ ing to the singly and doubly hydrated arginine ion clusters are also visible in
¹²¹ the spectrum. Smaller tips tend to produce relatively stronger signals and less
¹²² noisy spectra than larger tips, as can be seen by comparing the baselines of
¹²³ the three spectra in Fig. 2a. We observe some variance in the distribution of
¹²⁴ solvation states between nanopore ion sources with similar tip sizes, and even
¹²⁵ in repeated measurements with the same tip. This is illustrated, for example,
¹²⁶ by the different distributions of arginine solvation states in Fig 2a (20 nm I.D.)
¹²⁷ and in Fig 5b (24 nm I.D.). However, only tips with inner diameters smaller

 than about 65 nm produced spectra where most of the amino acid ions were
 in the desolvated state.

     Figure 2b shows mass spectra obtained from 16 different aqueous amino
 acid solutions, all at 100 mM concentration with the exception of tryptophan
 which was 50 mM. Four different tips with inner diameters of 20, 25, 57 and
 58 nm were used for these measurements. The most prominent amino acid
 peak in every spectrum shown in Fig. 2b corresponds to a singly charged and
 unsolvated ion. The spectra for glycine, alanine, proline, valine, cysteine, glu-
 tamine, and phenylalanine show no additional peaks which could correspond
 to solvated amino acid ions. The spectra for serine, threonine, asparagine,
 lysine, methionine, histidine, arginine, and tryptophan show a secondary peak
 18 m/z to the right of the unsolvated peak, corresponding to the singly
 hydrated amino acid ion. Leucine shows a third and fourth peak corresponding
 to higher solvation states. The spectra for leucine, methionine, and histidine
 show additional small peaks corresponding to bare, singly charged amino acid
 dimers. A number of spectra also show peaks below 200 m/z corresponding
 to singly protonated water clusters $(H_3O^+(H_2O)_n)$; the same peaks appear in
 control measurements of aqueous solutions with no amino acid present (see
 Supplemental Information Fig. S1). Four proteinogenic amino acids are absent
 from Fig. 2b: we do not attempt to measure aspartic acid and glutamic acid
 in positive ion mode because of their low isoelectric points; we also left out
 isoleucine because it is indistinguishable from leucine based on m/z; and tyro-
 sine gave poor emission characteristics, likely related to its low solubility in
 water. We worked with high (50 mM - 100 mM) analyte concentrations to
 obtain interpretable spectra quickly. Reliable identification of amino acids in
 these measurements required the release of only attomoles of analyte from
 the ion source, as shown in the Supplemental Information §3. The minimum

Fig. 2 Amino acid and peptide spectra. a, Mass spectrum of 100mM arginine solution in H_2O using nanopore ion sources with 3 different inner tip diameters (20 nm, 125 nm, and 300 nm). **b**, Gallery of 16 amino acid mass spectra, ordered from top left to bottom right by mass. Peaks labelled w_n correspond to water clusters ($H_3O^+(H_2O)_n$), peaks labelled 0 correspond to the desolvated amino acid ion, 1 corresponds to the singly solvated amino acid ion, etc, and peaks labelled D correspond to the desolvated amino acid dimer ion. All experiments were carried out using nanopore ion sources with 20-60 nm inner tip diameters (see Supplementary Information Table 1). Each spectrum comprises thousands of individual detection events. Supplementary Fig. S2 shows how the mass spectrum develops with increasing number of events. **c**, Overlaid mass spectra of glutathione (GSH, blue) and two of its PTM variants, s-nitrosoglutathione (GSNO, red) and s-acetylglutathione (SAG, green).

concentration of analyte at which identification is possible is an alternative measure of sensitivity that we did not explore in this work.

2.2 Measuring post-translationally modified peptides

Figure 2c shows mass spectra of glutathione and two chemically modified variants, s-nitrosoglutathione, and s-acetylglutathione. Glutathione is a tripeptide (amino acid sequence: GSH) found in high concentrations in most cells [27], and the variants we studied result from common post-translational modifications [28, 29]. Ion sources with 20 nm ID tips generate the peptide ions from 100 mM aqueous solutions with a pH between 3.1 and 3.9, adjusted by the addition of acetic acid. The glutathione spectrum shows a single peak at 308

m/z, which corresponds to the singly protonated, unsolvated glutathione ion. The spectra of s-acetylglutathione and s-nitrosoglutathione show dominant peaks at 350 m/z and 337 m/z, respectively, corresponding to the singly protonated, unsolvated peptide ions; each spectrum also shows two progressively smaller peaks 18 and 36 m/z to the right of the dominant peak, corresponding to singly and doubly solvated peptide ions, respectively.

2.3 Measuring a multiply charged peptide

Figure 3 shows mass spectra obtained from an experiment where an aqueous glutathione solution was replaced with an angiotensin II solution (amino acid sequence: DRVYIHPF) inside a 29 nm ID tip. The mass spectrum of the 100 mM glutathione solution (pH 3.0) initially inside the tip shows a large peak at 308 m/z that we assign to protonated glutathione, along with additional smaller peaks at 326, 344, and 615 m/z that we assign, respectively, to singly protonated, singly hydrated glutathione, singly protonated, doubly hydrated glutathione, and singly protonated glutathione dimer. We gradually introduced a 1 mM angiotensin II solution into the tip using a previously described tube-in-tube fluid handling system [26]. Additional peaks appeared in the mass spectrum at 524 m/z and 350 m/z, which we assign to doubly and triply protonated angiotensin II, respectively. The emission current throughout the measurement was 10 ± 3 pA with an extraction voltage of 400 V.

2.4 Ion transmission efficiency

We measure the efficiency with which ions pass from the nanopore source to a distant detector in a high vacuum environment as illustrated in Fig. 4a. Ions

Fig. 3 Mass spectra of 100 mM glutathione (dashed blue line) and a mixture of glutathione and angiotensin II obtained about 45 minutes after introducing a 1 mM angiotensin II solution into the tip of the ion source (solid red line). Peaks assigned to glutathione (GSH^+), singly hydrated glutathione (GSH^+)+H_2O)), glutathione dimer (($GSH)_2^+$), and the doubly and triply protonated states angiotensin II ($DRVYIHPF^{n+}$, $n = 2, 3$) are indicated.

emitted by the source are focused into the 2.4 cm diameter opening of a Faraday cup located $\sim$ 50 cm away. The ion transmission efficiency is the ratio of the current transmitted to the Faraday cup, I_T, to the current emitted from the nanopore, I_E. Figure 4b shows I_E, I_T, and the ion transmission efficiency measured over several minutes using a 39 nm ID tip filled with a 100 mM aqueous solution of sodium iodide. The mean ion transmission efficiency in this measurement is 93.4 % $\pm$ 1.7 %. Figure 4c shows the results of a similar measurement made with a 49 nm ID tip filled with 100 mM arginine in aqueous solution at pH 8.95. The mean ion transmission efficiency is 65.3 % $\pm$ 8.1 %. In both measurements, small changes in I_E are mirrored by I_T. We had more difficulty optimizing the focusing voltages in the amino acid measurement because of the very low emission currents, which were at the single pA level.

2.5 Separating ions and charged droplets

We investigate the possibility that the nanopore ion source emits charged droplets in addition to ions by adding a magnetic sector to the flight path as

Fig. 4 Ion transmission efficiency and ion/droplet current fractions. a, Setup used to measure the ion transmission efficiency of a nanopore ion source. **b**, I_E and I_T (above) and the resulting transmission efficiency (below) of a 39 nm ID nanopore ion source containing a 100 mM aqueous solution of NaI. **c**, I_E and I_T (above) and the resulting transmission efficiency (below) of a 49 nm ID nanopore ion source containing a 100 mM aqueous solution of arginine at pH 8.95. **d**, Setup used to separately measure I_{Ion} and I_{Drop}. **e**, I_{Ion} and I_{Drop} (above) and the resulting ion fraction (below) emitted by a 39 nm ID nanopore ion source containing a 100 mM aqueous solution of NaI. **f**, I_{Ion} and I_{Drop} (above) and the resulting ion current fraction (below) emitted by a 41 nm ID nanopore ion source containing a 100 mM aqueous solution of Arginine at pH 8.95.

²⁰³ shown in Fig. 4d. The 6 cm diameter, 0.54 T magnetic sector deflects charged

²⁰⁴ species based on their mass-to-charge ratio. Droplets with an $m/z > 33,000$

²⁰⁵ deflect by less than 3.1° and enter the Faraday cup where their current, I_{Drop},

²⁰⁶ is measured. We do not expect the droplet size to change appreciably between

²⁰⁷ emission and detection because nanoscale aqueous droplets freeze rapidly in

²⁰⁸ vacuum [30, 31]. Meanwhile, ions with m/z in the range of about 70 to 325

²⁰⁹ deflect onto a separate Faraday plate where the ion current, I_{Ion}, is measured

²¹⁰ (see Supplementary Information §5). Figure 4e shows I_{Ion}, I_{Drop}, and the ion

²¹¹ fraction, defined as ($\frac{I_{\mathrm{Ion}}}{I_{\mathrm{Ion}}+I_{\mathrm{Drop}}}$), for a measurement performed using a 28 nm

²¹² ID tip filled with a 100 mM aqueous solution of NaI. I_{Ion} varies between about

²¹³ 60 pA and 80 pA while no I_{Drop} is observed. Figure 4f shows the results of a

²¹⁴ similar measurement made using a 41 nm ID tip filled with a 100 mM aqueous

solution of arginine at pH 8.75. I_{Ion} varies between about 1 pA and 2 pA while there is no detectable I_{Drop}. In both measurements, the nanopore source was only observed to emit ions.

2.6 pH dependence of amino acid emission

To further investigate the emission mechanism, we measured the pH-dependence of the emission rates of histidine and arginine. For each amino acid, we prepared solutions at two different pH values but with a fixed concentration of Na^+ to serve as a reference ion. We made all measurements with a single, 24 nm ID tip. Figure 5a shows the mass spectra of 100 mM aqueous histidine solutions with pH values of 5.49 and pH 8.52; the concentration of Na^+ is 24 mM. At pH 5.49, peaks corresponding to the bare histidine ion and several solvated states appear between approximately 150 m/z and 275 m/z; the histidine peaks are more intense than the solvated sodium peaks which appear between approximately 50 m/z and 150 m/z. At pH 8.97, the histidine peaks are still visible, but their intensity is significantly diminished relative to the sodium peaks. Figure 5b shows the mass spectra of 100 mM arginine solutions with pH values of 9.10 and 10.78; the solutions also contain 50 mM of Na^+. Similarly to histidine, bare and solvated arginine peaks, which appear in the range between 170 m/z and 350 /mz, have a higher intensity than the sodium peaks at the lower pH of 9.10. The relative intensity of the arginine peaks decreases dramatically at the higher pH of 10.78. The results shown in Fig. 5 are representative of repeated measurements. To quantify the pH-dependence of the amino acid emission rates, we normalized them by the sodium emission rates. The histidine emission rate increased by an average factor of 8.14 ± 1.65

Fig. 5 Amino acid ion emission pH a Normalized mass spectra of 100 mM Histidine and 24 mM NaI/NaOH in aqueous solution at pH 5.49 (darker solid curve) and 8.52 (lighter dashed curve) respectively. **b** Normalized mass spectra of 100 mM Arginine and 50mM NaI at pH 9.10 (darker curve) and 10.78 (lighter curve) respectively. The vertical dotted line shows the m/z value used in the analysis to separate arginine signal from sodium signal. In both panels, the vertical dashed line shows the m/z value used in the analysis to separate the amino acid signal from sodium signal.

239 as the pH decreased from 8.52 to 5.49. The arginine emission rate increased

240 by an average factor of 22.6 ± 10.5 as the pH decreased from 10.78 to 9.10.

2.7 Calculating the probability of ions scattering

Calculations indicate that most ions trace collisionless trajectories from the

ion source to the detector. Figure 6a shows the probability that an amino acid

ion dressed with a hydration shell will collide with a gas molecule, based on

the kinetic theory of gases. We assume that ions pass through a distribution of

evaporating water molecules [32] and a homogeneous low-pressure background

of N_2. Figure 6a sketches the physical situation and plots the number density

of gas molecules and the cumulative collision probability as functions of the distance from the meniscus. The cumulative probability of an ion colliding with a gas molecule over the entire 50 cm trajectory from the source to the detector is just 2.1%, meaning the vast majority of ions do not experience any collisions. Most of the collisions occur within 200 nm of the liquid meniscus due to the high density of evaporated water molecules there. A detailed description of these calculations is found in §4 of the Supplementary Information.

3 Discussion

There are two distinct mechanisms by which the nanopore source could generate the ions that we measure. The first is conventional electrospray ionization, a multi-step mechanism in which an ion source first produces charged droplets which go on to evaporate ions into the vapor phase [2]. The second is a single-step mechanism in which ions evaporate directly from the electrified meniscus of the ion source [33]. We highlight the important physical distinctions between these two mechanisms before presenting an analysis that rules out the former.

ESI begins when a voltage applied to the liquid sample generates electrical forces on the meniscus that are strong enough to overcome the surface tension and draw out charged droplets (Fig. 1a). The droplets typically measure hundreds of nanometers to tens of micrometers in radius and contain fully solvated analyte molecules [34]. A background gas is used to evaporate solvent from the droplets [34]. As a droplet shrinks, its charge density increases, and the resulting internal Coulomb forces can once again overcome surface tension when the electric field strength at the surface reaches the Rayleigh limit, $E_R = 2\sqrt{\frac{\gamma}{\epsilon_0 r}}$, where γ is the surface tension, ϵ_0 is the permittivity of free space,

and r is the droplet radius [3, 35]. The unstable droplet undergoes a Coulomb explosion and expels smaller charged droplets [3, 5]. The cycle repeats, producing ever smaller droplets requiring ever stronger electric fields to reach the Rayleigh limit. When the field strength at the surface reaches about 10^9 V/m, it becomes possible for ions to quickly escape the droplet by a different process: Ion evaporation is a thermal process in which ions overcome an energy barrier associated with their solvation energy, ΔG_0^S, to enter the vapor phase [24]. Coulomb explosions cease and ion evaporation becomes the main charge release mechanism for small molecules when droplets reach around 10 nm in radius [24, 36, 37].

Alternatively, the nanopore source might deliver ions and ion clusters into the vapor phase directly from the meniscus at the tip by an ion evaporation process. This alternative mechanism circumvents the emission of droplets and their interactions with a background gas that are integral to ESI. Figure 6b illustrates the direct evaporation of a solvated ion cluster from an electrified meniscus and a schematic of the energy landscape [24, 38]. An ion inside the nanopore source is fully solvated by water molecules, which reduce the free energy by an amount ΔG_0^S, called the solvation energy. The free energy increases as the ion protrudes from the meniscus, distorting it against surface tension. At this point, the ion is also attracted by its own image charge as it begins to break free of the solvent. Far from the meniscus, Coulomb forces drive the ion (or cluster) away from the tip and its free energy decreases. Modelling the height of the energy barrier as $\Delta G_0^S - G(E)$, where $G(E)$ is the field-dependent reduction in the barrier height [3, 24, 38], the current density of ions escaping the meniscus is given by

$$j = \sigma \frac{k_B T}{h} \exp\left(-\frac{\Delta G_0^S - G(E)}{k_B T}\right),\tag{1}$$

where σ is the surface charge density, $k_B T$ is the thermal energy, and h is the Planck constant. The ion evaporation current density only becomes appreciable when the electric field at the meniscus approaches $\sim 10^9$ V/m, at which point $\Delta G_0^S - G(E) \sim k_B T$ [39]. The ion evaporation mechanism applies equally well to bare ions and small ion clusters, as each is a charged species that can escape the liquid over a solvation energy barrier [24, 40, 41]. On the other hand, charged nano-droplets comprising hundreds of solvent molecules or more are too large to evaporate.

Multiple lines of evidence lead us to conclude that the nanopore source produces ions directly by ion evaporation rather than by the conventional ESI mechanism in which ions emerge from charged droplets. First, the chain of Coulomb explosions that characterizes ESI cannot occur under the high vacuum conditions of our experiments. Droplets quickly freeze before shedding a significant fraction of their mass when there is no background gas to provide the thermal energy that normally sustains solvent evaporation [30]. A previous study showed that nanoscale charged aqueous droplets survive under high vacuum for more than a second without undergoing Coulomb explosions [42].

Secondly, we searched for droplets and measured none among the charged species delivered by the source (Fig. 4d-f). Our measurements, which are sensitive to species with $m/z > 33,000$, exclude droplets with a radius larger than 14 nm and a charge as high as the Rayleigh limit. For comparison, nano-ESI begins with a spray of droplets that are in the detectable range of tens to hundreds of nanometers in diameter [2, 43]. If even smaller charged droplets were produced [44, 45], they would evaporate ions readily, lose a significant

Fig. 6 Ion emission mechanism. a Plot of the cumulative probability that an emitted amino acid with a small hydration shell (radius = 7 Å) will collide with an evaporated water molecule or background gas molecule as a function of distance from the meniscus, r. The green curve shows the calculated maximum water vapor density as a function of the distance from the meniscus. **b** Potential energy diagram showing the contributions to the energy barrier for ion evaporation: the solvation energy, ΔG_0, image charge attraction, U_{image}, and electrostatic repulsion, U_E. **c** Plot of the approximate limiting fields for ion evaporation and droplet generation (Rayleigh limit). The region where ion evaporation should dominate is highlighted in green.

fraction of their charge, and move into the detectable range of m/z; when the

electric field at the surface falls to 0.75 V/m, the approximate threshold for

ion evaporation, droplets as small as 7 nm in radius are detectable.

₃₂₄   Thirdly, the nanopore source is sufficiently small that as the electric field
₃₂₅ at the liquid meniscus increases, it reaches the threshold for ion evaporation
₃₂₆ well before the Rayleigh limit. Figure 6c, which compares the critical electric
₃₂₇ field for ion evaporation with the Rayleigh limit, illustrates why ions evaporate
₃₂₈ from the meniscus while the meniscus remains stable. If we consider a nanopore
₃₂₉ source with a radius of 15 nm, we expect ions to evaporate from the meniscus
₃₃₀ when the electric field reaches about 0.75 V/nm, whereas droplets cannot
₃₃₁ emerge unless the electric field grows about twice as strong. On the other hand,
₃₃₂ a conventional ESI source with a radius of 1 μm produces charged droplets
₃₃₃ when the electric field at the meniscus reaches about 0.2 V/nm, which is well
₃₃₄ below the threshold for ion evaporation. The crossover from a conventional
₃₃₅ droplet-emitting source to a nanopore source that can evaporate ions directly
₃₃₆ is predicted to occur at a radius of about 60 nm.

₃₃₇   Finally, the pH-dependence of the emission rates of arginine and histi-
₃₃₈ dine we measured (fig. 5) is consistent with a simple kinetic model of ion
₃₃₉ evaporation from the meniscus. We consider a two-step process in which an
₃₄₀ analyte molecule M can undergo reversible protonation/deprotonation with
₃₄₁ rates $k_{\mathrm{on}}[\mathrm{H^+}]$ & k_{off}, and non-reversible evaporation from the protonated state
₃₄₂ with rate k_{evap},

$$H^+ + M \underset{k_{\mathrm{off}}}{\overset{k_{\mathrm{on}}}{\rightleftharpoons}} MH^+ \xrightarrow{k_{\mathrm{evap}}} MH^+_{\mathrm{vapor}}. \tag{2}$$

₃₄₃ We assume that evaporation is the rate-limiting step and that $k_{\mathrm{evap}} \ll k_{\mathrm{off}}$,
₃₄₄ justified by the picoampere-level currents measured in this work. Holding the
₃₄₅ concentration $[\mathrm{MH^+}]$ constant, we find that the quasi-steady-state emission
₃₄₆ rate, $v \equiv [MH^+]k_{\mathrm{evap}}$, scales as

Fig. 7 Amino acid ionization versus pH Summary of repeated measurements of the relative emission rates of histidine (circles) and arginine (triangles) compared to theoretical predictions. The emission rate at the lower pH, v_{low}, is calculated using Eq. (3), and the emission rate at the higher pH is calculated using v_{low}/F, where F is the ratio of measured signal between low and high pH. The error bars represent standard error of the mean.

$$v \propto \frac{10^{-\mathrm{pH}}}{10^{-\mathrm{pKa}} + 10^{-\mathrm{pH}}}. \tag{3}$$

Figure 7 plots the relative emission rates predicted by this model for both histidine and arginine and compares them to the measured values. For histidine, two protonation reactions occur within the experimental pH range (with $\mathrm{pKa}_1 = 6.04$ and $\mathrm{pKa}_2 = 9.09$); a single dissociation constant corresponding to the isoelectric point (pKa=7.59) models the combined effect [46]. For arginine, only the protonation of the amino group (pKa= 9.00) is relevant [46, 47]. The kinetic model predicts that the emission rates should increase by factors of 9.2 and 27.3 over the pH ranges investigated for histidine and arginine, respectively, in agreement with our measured values of 8.14 ± 1.65 and 22.5 ± 10.5. By contrast, ESI emits arginine and histidine at rates that depend much more weakly on pH (by a factor less than three over the same pH range), and this observation was attributed to the acidification of droplets that accompanies the ESI process [48, 49].

The nanopore source appears to operate in a pure ion mode where no droplets are emitted. Pure ion emission was previously observed from highly conductive liquids including liquid metals [50], ionic liquids [33, 51], and concentrated electrolyte solutions in formamide [39, 40], but not aqueous solutions. Previous attempts to evaporate ions directly into vacuum from a fixed water meniscus encountered significant challenges, as the evaporation of water led to electrical arcing and freezing [52]. The nanopore source overcomes those challenges because the evaporation rate from the nanoscale meniscus is minuscule (less than 1 pL of water evaporating per minute); this low gas load also eases the pumping requirements of the instrument [26]. Previously, even the smallest ion emitters were used in combination with a background gas and the emission mechanism was presumed to be conventional ESI [53].

It is striking that we measure mostly desolvated amino acid and small peptide ions in many experiments (fig. 2b and c, fig. 3). This property simplifies the interpretation of mass spectra without requiring a drying gas. Were the ions emitted in an already desolvated state, or clustered with one or more water molecules which were shed on the way to the detector? Given that just 2% of emitted ions collide with a gas molecule even once (see Fig. 6a), post-emission collisions can be dismissed as a primary desolvation mechanism. Furthermore, the tip size appears to influence the solvation state, as only sub-100 nm nanopores produced bare ions (Fig. 2a). These observations suggest that it is the local environment at the source that controls the solvation state rather than processes occurring in flight. The conditions that give rise to bare ion emission are not entirely clear. The electric fields that develop near the smallest tips are strong enough to reorient and weaken hydration bonds, which may play a role [54, 55]. Joule heating may also play a role by raising the local temperature and further weakening hydration bonds [56, 57]. In addition to

single amino acid ions, we observed desolvated amino acid dimer ions in some spectra (e.g., Fig. 2b, His). This is likely due to the aggregation of amino acids in solution, which is known to occur at concentrations of 1 mM and above [58]. We also observed glutathione and its post-translationally modified variants as bare, singly charged ions. Subsequent work should investigate how the charge and hydration states depend on the peptide length, composition, and other properties.

The present study makes three significant contributions: First, our measurements and analyses of the nanopore source provide strong evidence for the direct ion evaporation mechanism. Second, the nanopore source achieves near unit ion transmission efficiency. This holds the promise of significant improvements in the sensitivity of MS measurements, which are universally desirable and could be achieved by combining a nanopore source with a sensitive mass filter (e.g., a trap-based mass filter). Third, the nanopore source is compatible with biologically relevant aqueous samples.

The findings reported here open new avenues for research. We envision sequencing a single protein molecule by fragmenting it and delivering its constituent amino acids into a mass spectrometer in sequential order [18, 26, 59]. The nanopore ion source provides the crucial means of transmitting virtually every amino acid to the detector, and because they avoid collisions in flight, the timing of detection events can be used to accurately reconstruct the order in which they left the source. However, significant challenges persist. A means of reliably fragmenting polypeptides into free monomers is required, as is a means of preserving the order of monomers prior to emission, perhaps using force gradients [59] or a narrow nanotube emitter [60] to prevent amino acids from diffusing past one another. These findings could also impact various fields

outside of single-molecule sequencing or proteomics. Electrospray propulsion, a method for generating thrust through ion emission, is currently constrained by a limited selection of non-volatile ionic liquids that can be used as propellants [61, 62], but the nanopore ion source highlights the possibility of using volatile liquids. Atom probe tomography, a technique for generating compositional maps of materials using field-emitted ions, is entering a new frontier with studies of liquid samples [63–65], which may be simplified by using a nanopore ion source. Finally, the nanopore source may shed new experimental light on the ionization of water, its orientation, or its evaporation at highly electrified surfaces [66, 67].

Methods

Preparing nanopore ion sources

We pulled nanopipettes from 7.5 cm long quartz capillaries with 0.7 mm inner diameter and 1 mm outer diameter (QF100-70-7.5, Sutter Instruments) using a laser puller (P-2000, Sutter Instruments) according to the following single-line recipe: heat = 650, velocity = 45, delay = 175, pull = 190. To prepare a nanopore for imaging, we deposited approximately 5 nm of carbon to the nanopipettes using thermal evaporation (CED 030, BalTec) with double-stranded carbon thread (Leica Microsystems). We imaged the nanopipettes by scanning electron microscopy (LEO 1530 VP, Zeiss inc.) to measure the inner and outer tip diameters. Nanopipettes were plasma cleaned in air for two minutes (Plasma Preen II-973, Plasmatic Systems Inc.) to make the surfaces hydrophilic prior to filling with sample solution.

Amino acid solutions

Amino acid solutions comprised the amino acid of interest (Sigma-aldrich) dissolved in deionized water (18 MΩcm, Millipore) at a concentration of 100 mM, with the exception of tryptophan, whose concentration concentration was 50 mM. We reduced the pH below the amino acid's isoelectric point by adding 0.1-0.5% v/v glacial acetic acid (Sigma-Aldrich). Peptide solutions comprised the peptide of interest (glutathione, s-acetylglutathione, or s-nitrosoglutathione) dissolved in deionized water at a concentration of 100 mM. We purchased glutathione and s-acetylglutathione in powder form (Sigma-aldrich), and synthesized S-nitrosoglutathione from glutathione according to a protocol from T.W. Hart [68]. We measured the pH and conductivities of each solution using a pH meter (Ultrabasic Benchtop, Denver Instruments) and conductivity meter (Sension+ EC71 GLP, Hach), respectively.

The nanopore mass spectrometer

A custom-built quadrupole mass spectrometer, described previously in Bush et al [26], measured all the amino acid and peptide mass spectra presented here. The instrument, illustrated schematically in Fig. 1c, comprises a custom einzel lens, a quadrupole mass filter (QC MAX-500, Extrel inc.), an ion bender (Extrel inc.), and a channel electron multiplier detector with a conversion dynode (DeTech 413) sensitive to single ions. The quadrupole is driven by a 440 kHz RF oscillator and scans across a mass range of 50 m/z to 1000 m/z in about 0.6 s. The quadrupole's Δm and Δres parameters were periodically tuned to optimize the peak shape while acquiring spectra of NaI in formamide.

The pole bias was set to ensure the ions had a kinetic energy of about 5-10 eV as they traversed the mass filter. The base pressure of the instrument is 10^{-8} torr. A tube-in-tube system (previously described in Bush *et al* [26]) enables us to deliver fresh sample solution to the nanopipette tip while flushing away old solution without breaking the vacuum. A 150 μm i.d., 360 μm o.d. PEEK tube (IDEX Health and Science) carries fresh sample solution into the tip, displacing and pushing out old solution through a wider outer PEEK tube (0.04" i.d., 1/16" o.d., IDEX Health and Science). We prepare Ag/AgCl electrodes by placing a length of 0.5 mm diameter silver wire (Alfa-Aesar) in bleach for 15-20 minutes. The Ag/AgCl electrode is threaded far into the outer PEEK tubing, enabling us to control the voltage of the solution near the tip of the ion source.

Measurements of amino acid and glutathione mass spectra

We prefilled the nanopipette with the amino acid or peptide solution using a Microfil flexible needle (World Precision Instruments) before mounting the nanopipette inserting them into the mass spectrometer. We continuously refreshed the solution at the tip by pumping solution through the tube-in-tube system at rate of 0.4 mL/hour using a syringe pump (NE-300, New Era Pump Systems). Note that this is not the flow rate of fluid emitted from the tip, as the tube-in-tube system carries the vast majority of the sample solution out of the tip through the rear, as described in Bush *et al* [26]. A voltage of +100 V or +200 V was applied to the electrode in contact with the solution inside the nanopore ion source using a high voltage source measure unit (2657A, Keithley Instruments) while a negative voltage applied to the extraction electrode was

 slowly strengthened using a high voltage power supply (Bertan 230, Spellman

 HV) until ionization was observed. The onset of emission typically occurred

 when V_E was between 200 and 350 V.

Ion transmission efficiency measurements

We measured ion transmission efficiencies in a custom vacuum chamber containing a nanopore ion source and a Faraday cup (Fig. 4a). A high voltage source measure unit (Keithley 2657a or Keithley 2410) measured the current emitted from the nanopore, I_E, while also applying a voltage, V_T, to it. A current-preamplifier (SR570, Stanford Research Systems) converted I_T at the Faraday cup into a voltage that was recorded by a DAQ (NI PCIe-6251, National instruments). An 8-channel high-voltage power supply (DT8033, CAEN) supplied the focusing voltages to the ion optics. A custom Labview (National Instruments) program controlled V_T and recorded I_E and I_T. The currents emitted from aqueous salt solutions (Fig. 4b), which ranged from 100 pA to 1 nA, were measured with a Keithley 2410. To reliably measure the low, picoampere-level currents emitted from aqueous amino acid solutions (Fig. 4c), we used a more sensitive source measure unit (Keithley 2657a) and a shielded, tri-axial cable that more effectively suppressed leakage currents. To further suppress noise, the source measure unit averaged three sequential measurements of I_E with an NPLC setting of 5, then an 8^{th}-order, low-pass Bessel filter with a cutoff frequency of 0.1 Hz filtered the data to obtain the results plotted in Fig. 4c. I_T was sampled at a rate of 50 kHz, and 10k samples were averaged to give each data point in Fig. 4c. The setup records I_E and I_T data points sequentially with an overall sampling rate of ~ 0.6 Hz. A small leakage current measured at the source measure unit with $V_E = 0$ V and

$V_T = 500$ V was subtracted from measurements of I_E. The leakage current was about 5 pA when using the Keihtley 2657a (arginine measurements) and about 100 pA when using the Keithley 2410 (NaI measurements) for the NaI efficiency measurement. An instrumental offset of about 200 fA was subtracted from measurements of I_T.

Magnetic sector measurements

We constructed a simple magnetic sector mass spectrometer by adding a magnetic sector and a Faraday plate downstream of the ion source in the vacuum chamber described above, as illustrated in Fig. 4d. The magnetic sector comprises a neodymium magnet and a yoke made of low-carbon magnetic iron (ASTM A848) which concentrate magnetic fields of 0.54 ± 0.02 T in a flat circular region 6 cm in diameter and 1 cm in height. The Faraday plate is a stainless steel disk, 5 inches in diameter, mounted about 25 cm from the center of the magnetic sector, 45° from the undeflected axis of the ions. Two separate current pre-amplifiers (Stanford Research Systems SR570) measured the currents striking the Faraday plate and cup. The currents were measured and processed in the same manner described above for the ion transmission efficiency measurements.

Data availability statement

The data and analysis code used in this study are available at the Harvard Dataverse as Drachman, Nicholas; LePoitevin, Mathilde; Szapary, Hannah; Wiener, Benjamin; Maulbetsch, William; Stein, Derek, 2024, "Replication Data for: Nanopore ion sources deliver single amino acid and peptide ions

References

[1] Aebersold, R. & Mann, M. Mass-spectrometric exploration of proteome structure and function. *Nature* **537** (7620), 347–355 (2016) .

[2] Cole, R. B. *Electrospray and MALDI mass spectrometry: fundamentals, instrumentation, practicalities, and biological applications* (John Wiley & Sons, 2011).

[3] Fenn, J. B., Mann, M., Meng, C. K., Wong, S. F. & Whitehouse, C. M. Electrospray ionization for mass spectrometry of large biomolecules. *Science* **246** (4926), 64–71 (1989) .

[4] Taylor, G. I. Electrically driven jets. *Proceedings of the Royal Society of London. A. Mathematical and Physical Sciences* **313** (1515), 453–475 (1969) .

[5] Kebarle, P. & Verkerk, U. H. Electrospray: from ions in solution to ions in the gas phase, what we know now. *Mass spectrometry reviews* **28** (6), 898–917 (2009) .

[6] Page, J. S., Kelly, R. T., Tang, K. & Smith, R. D. Ionization and transmission efficiency in an electrospray ionization—mass spectrometry interface. *Journal of the American Society for Mass Spectrometry* **18** (9), 1582–1590 (2007) .

[7] Smith, R. D., Loo, J. A., Edmonds, C. G., Barinaga, C. J. & Udseth, H. R. New developments in biochemical mass spectrometry: electrospray ionization. *Analytical chemistry* **62** (9), 882–899 (1990) .

[8] Wilm, M. & Mann, M. Analytical properties of the nanoelectrospray ion source. *Analytical chemistry* **68** (1), 1–8 (1996) .

[9] El-Faramawy, A., Siu, K. M. & Thomson, B. A. Efficiency of nano-electrospray ionization. *Journal of the American Society for Mass Spectrometry* **16** (10), 1702–1707 (2005) .

[10] Hawkrdige, A. M. in *Quantitative proteomics* (eds Eyers, C. E. & Gaskell, S. J.) *Practical considerations and current limitations in quantitative mass spectrometry based proteomics* Ch. 1, 3–25 (Royal Society of Chemistry, 2014).

[11] Marginean, I., Page, J. S., Tolmachev, A. V., Tang, K. & Smith, R. D. Achieving 50% ionization efficiency in subambient pressure ionization with nanoelectrospray. *Analytical chemistry* **82** (22), 9344–9349 (2010) .

[12] Tang, K. & Smith, R. D. Physical/chemical separations in the break-up of highly charged droplets from electrosprays. *Journal of the American Society for Mass Spectrometry* **12** (3), 343–347 (2001) .

[13] Li, S. *et al.* An integrated platform for isolation, processing, and mass spectrometry-based proteomic profiling of rare cells in whole blood. *Molecular & Cellular Proteomics* **14** (6), 1672–1683 (2015) .

[14] Callahan, N., Tullman, J., Kelman, Z. & Marino, J. Strategies for development of a next-generation protein sequencing platform. *Trends in Biochemical Sciences* **45** (1), 76–89 (2020) .

[15] Budnik, B., Levy, E., Harmange, G. & Slavov, N. Scope-ms: mass spectrometry of single mammalian cells quantifies proteome heterogeneity during cell differentiation. *Genome biology* **19** (1), 1–12 (2018) .

[16] Kelly, R. T. Single-cell proteomics: progress and prospects. *Molecular & Cellular Proteomics* **19** (11), 1739–1748 (2020) .

[17] Restrepo-Pérez, L., Joo, C. & Dekker, C. Paving the way to single-molecule protein sequencing. *Nature Nanotechnology* **13** (9), 786–796 (2018) .

[18] Alfaro, J. A. *et al.* The emerging landscape of single-molecule protein sequencing technologies. *Nature methods* **18** (6), 604–617 (2021) .

[19] Steinbock, L. J., Otto, O., Chimerel, C., Gornall, J. & Keyser, U. F. Detecting dna folding with nanocapillaries. *Nano letters* **10** (7), 2493–2497 (2010) .

[20] Jordan, J. S., Xia, Z. & Williams, E. R. Tips on making tiny tips: Secrets to submicron nanoelectrospray emitters. *Journal of the American Society for Mass Spectrometry* **33** (3), 607–611 (2022) .

[21] Liu, H. & Cao, G. Effectiveness of the young-laplace equation at nanoscale. *Scientific reports* **6** (1), 1–10 (2016) .

[22] Maulbetsch, W. *Nanopore Mass Spectrometry*. Ph.D. thesis, Brown University (2018). URL https://repository.library.brown.edu/studio/item/bdr:792908/.

[23] Higuera, F. Qualitative analysis of the minimum flow rate of a cone-jet of a very polar liquid. *Journal of Fluid Mechanics* **816**, 428–441 (2017) .

[24] Iribarne, J. V. & Thomson, B. A. On the evaporation of small ions from charged droplets. *The Journal of Chemical Physics* **64** (6), 2287–2294 (1976) .

[25] Higuera, F. Model of the meniscus of an ionic-liquid ion source. *Physical Review E* **77** (2), 026308 (2008) .

[26] Bush, J. *et al.* The nanopore mass spectrometer. *Review of Scientific Instruments* **88** (11), 113307 (2017) .

[27] Meister, A. & Anderson, M. E. Glutathione. *Annual review of biochemistry* **52** (1), 711–760 (1983) .

[28] Broniowska, K. A., Diers, A. R. & Hogg, N. S-nitrosoglutathione. *Biochimica et Biophysica Acta (BBA)-General Subjects* **1830** (5), 3173–3181 (2013) .

[29] Mann, M. & Jensen, O. N. Proteomic analysis of post-translational modifications. *Nature biotechnology* **21** (3), 255–261 (2003) .

[30] Caleman, C. & van der Spoel, D. Temperature and structural changes of water clusters in vacuum due to evaporation. *The Journal of chemical physics* **125** (15), 154508 (2006) .

[31] Sellberg, J. A. *et al.* Ultrafast x-ray probing of water structure below the homogeneous ice nucleation temperature. *Nature* **510** (7505), 381–384 (2014) .

[32] Smith, J. D., Cappa, C. D., Drisdell, W. S., Cohen, R. C. & Saykally, R. J. Raman thermometry measurements of free evaporation from liquid water

droplets. *Journal of the American Chemical Society* **128** (39), 12892–12898 (2006) .

[33] Romero-Sanz, I., Bocanegra, R., Fernandez De La Mora, J. & Gamero-Castano, M. Source of heavy molecular ions based on taylor cones of ionic liquids operating in the pure ion evaporation regime. *Journal of Applied Physics* **94** (5), 3599–3605 (2003) .

[34] Kebarle, P. & Tang, L. From ions in solution to ions in the gas phase-the mechanism of electrospray mass spectrometry. *Analytical chemistry* **65** (22), 972A–986A (1993) .

[35] Rayleigh, L. Xx. on the equilibrium of liquid conducting masses charged with electricity. *The London, Edinburgh, and Dublin Philosophical Magazine and Journal of Science* **14** (87), 184–186 (1882) .

[36] Loscertales, I. & Fernández De La Mora, J. Experiments on the kinetics of field evaporation of small ions from droplets. *The Journal of chemical physics* **103** (12), 5041–5060 (1995) .

[37] Konermann, L., Ahadi, E., Rodriguez, A. D. & Vahidi, S. Unraveling the mechanism of electrospray ionization (2012).

[38] Magnani, M. & Gamero-Castaño, M. Energy barrier for ion field emission from a dielectric liquid sphere. *Physical Review E* **105** (5), 054802 (2022) .

[39] Gamero-Castano, M. & Fernandez De La Mora, J. Direct measurement of ion evaporation kinetics from electrified liquid surfaces. *The Journal of Chemical Physics* **113** (2), 815–832 (2000) .

[40] Luedtke, W. *et al.* Nanojets, electrospray, and ion field evaporation: Molecular dynamics simulations and laboratory experiments. *The Journal of Physical Chemistry A* **112** (40), 9628–9649 (2008) .

[41] Kwan, V. & Consta, S. Conical shape fluctuations determine the rate of ion evaporation and the emitted cluster size distribution from multicharged droplets. *The Journal of Physical Chemistry A* (2022) .

[42] Harper, C. C., Brauer, D. D., Francis, M. B. & Williams, E. R. Direct observation of ion emission from charged aqueous nanodrops: effects on gaseous macromolecular charging. *Chemical science* **12** (14), 5185–5195 (2021) .

[43] Hollerbach, A. *et al.* Sizing sub-diffraction limit electrosprayed droplets by structured illumination microscopy. *Analyst* **143** (1), 232–240 (2018) .

[44] Li, H., Allen, N., Li, M. & Li, A. Conducting and characterizing femto flow electrospray ionization. *Analyst* **147** (6), 1071–1075 (2022) .

[45] Baez Bolivar, E. G. *et al.* Submicron emitters enable reliable quantification of weak protein–glycan interactions by esi-ms. *Analytical Chemistry* **93** (9), 4231–4239 (2021) .

[46] Lide, D. R. *CRC handbook of chemistry and physics* Vol. 85 (CRC press, 2004).

[47] Fitch, C. A., Platzer, G., Okon, M., Garcia-Moreno E, B. & McIntosh, L. P. Arginine: Its pka value revisited. *Protein science* **24** (5), 752–761 (2015) .

[48] Mansoori, B. A., Volmer, D. A. & Boyd, R. K. 'wrong-way-round'electrospray ionization of amino acids. *Rapid Communications in Mass Spectrometry* **11** (10), 1120–1130 (1997) .

[49] Ojakivi, M., Liigand, J. & Kruve, A. Modifying the acidity of charged droplets. *ChemistrySelect* **3** (1), 335–338 (2018) .

[50] Prewett, P., Mair, G. & Thompson, S. Some comments on the mechanism of emission from liquid metal ion sources. *Journal of Physics D: Applied Physics* **15** (7), 1339 (1982) .

[51] Lozano, P. C. Energy properties of an emi-im ionic liquid ion source. *Journal of Physics D: Applied Physics* **39** (1), 126 (2005) .

[52] Cook, K. D. Electrohydrodynamic mass spectrometry. *Mass Spectrometry Reviews* **5** (4), 467–519 (1986) .

[53] Yuill, E. M., Sa, N., Ray, S. J., Hieftje, G. M. & Baker, L. A. Electrospray ionization from nanopipette emitters with tip diameters of less than 100 nm. *Analytical chemistry* **85** (18), 8498–8502 (2013) .

[54] Vegiri, A. & Schevkunov, S. V. A molecular dynamics study of structural transitions in small water clusters in the presence of an external electric field. *The Journal of chemical physics* **115** (9), 4175–4185 (2001) .

[55] He, Z., Cui, H., Hao, S., Wang, L. & Zhou, J. Electric-field effects on ionic hydration: a molecular dynamics study. *The Journal of Physical Chemistry B* **122** (22), 5991–5998 (2018) .

[56] Gallud, X. & Lozano, P. C. The emission properties, structure and stability of ionic liquid menisci undergoing electrically assisted ion evaporation.

690   *Journal of Fluid Mechanics* **933**, A43 (2022) .

[57]   Zhan, D., Rose, J. & Fenn, J. B.   Solvation studies of electrospray
        ions—method and early results. *Journal of the American Society for Mass
        Spectrometry* **9** (12), 1241–1247 (1998) .

[58]   Toyama, N., Kohno, J.-y., Mafuné, F. & Kondow, T. Solvation structure of
        arginine in aqueous solution studied by liquid beam technique. *Chemical
        physics letters* **419** (4-6), 369–373 (2006) .

[59]   Maulbetsch, W., Wiener, B., Poole, W., Bush, J. & Stein, D. Preserving
        the sequence of a biopolymer's monomers as they enter an electrospray
        mass spectrometer. *Physical Review Applied* **6** (5), 054006 (2016) .

[60]   Secchi, E. *et al.*   Massive radius-dependent flow slippage in carbon
        nanotubes. *Nature* **537** (7619), 210–213 (2016) .

[61]   Lozano, P. & Martinez-Sanchez, M. Ionic liquid ion sources: characteriza-
        tion of externally wetted emitters. *Journal of colloid and interface science*
        282 (2), 415–421 (2005) .

[62]   Miller, C. E. *Characterization of ion cluster fragmentation in ionic liquid
        ion sources.* Ph.D. thesis, Massachusetts Institute of Technology (2019).

[63]   Gault, B. *et al.*   Atom probe tomography.   *Nature Reviews Methods
        Primers* **1** (1), 51 (2021) .

[64]   Schwarz, T. *et al.* Field evaporation and atom probe tomography of pure
        water tips. *Scientific Reports* **10** (1), 20271 (2020) .

[65]   Qiu, S. *et al.* Direct imaging of liquid–nanoparticle interfaces with atom
        probe tomography.   *The Journal of Physical Chemistry C* **124** (35),

19389–19395 (2020) .

[66] Stuve, E. M. Ionization of water in interfacial electric fields: An electrochemical view. *Chemical Physics Letters* **519**, 1–17 (2012) .

[67] Segreto, N. *et al.* Understanding the underlying field evaporation mechanism of pure water tips in high electrical fields. *The Journal of Physical Chemistry A* **126** (33), 5663–5671 (2022) .

[68] Hart, T. W. Some observations concerning the s-nitroso and s-phenylsulphonyl derivatives of l-cysteine and glutathione. *Tetrahedron Letters* **26** (16), 2013–2016 (1985) .

Acknowledgements

We thank Oxford Nanopore Technologies LTD. for supporting this work. Aspects of this work were also supported by the Institute for Molecular and Nanoscale Innovation at Brown University. N.D. was partially supported by the Brown IMSD program (NIH R25GM083270) and Graduate Fellowships for STEM Diversity (GFSD). H.S. was partially supported by the Brown UTRA program. B.W. was partially supported by NSF 1409577.

Competing Interests Statement

D.S. declares a competing interest in Oxford Nanopore Technologies, Ltd.

Author Contributions

Conceptualization & supervision: D.S.; Research design: D.S., M.L., W.M.; Mass spectrometry measurements: M.L., N.D., and H.S.; Ion source characterization: N.D., W.M., and B.W.; Manuscript writing and editing: N.D. and D.S.; Figure preparation: N.D.; Funding acquisition: D.S.; All authors contributed to the discussions.