Absolute quantitation of serum antibody reactivity using the Richards growth model for

antigen microspot titration

Krisztián Papp, Ágnes Kovács, Anita Orosz, Zoltán Hérincs, Judit Randek, Károly Liliom,

Tamás Pfeil, József Prechl

Author information

These authors contributed equally: Krisztián Papp, Ágnes Kovács, Anita Orosz

R&D Laboratory, Diagnosticum Zrt, Budapest, Hungary

Krisztián Papp, Zoltán Hérincs, József Prechl

Department of Applied Analysis and Computational Mathematics, Eötvös Loránd University,

Budapest, Hungary

Ágnes Kovács, Tamás Pfeil

Department of Immunology, Eötvös Loránd University, Budapest, Hungary

Anita Orosz

BME, Budapest University of Technology and Economics, Budapest, Hungary

Judit Randek

Department of Biophysics and Radiation Biology, Semmelweis University, Budapest,

Hungary

Károly Liliom

Corresponding author: József Prechl, jprechl@diagnosticum.hu

1047 Attila út 126. Budapest, Hungary; T: +36-1-8810161

Current addresses

Anita Orosz : Department of Physiology, Semmelweis University, Budapest, Hungary;

Judit Randek : F. Hoffmann - La Roche Ltd., Basel, Switzerland

Abstract

In spite of its pivotal role in the characterization of humoral immunity, there is no accepted

method for the absolute quantitation of antigen specific serum antibodies.

We devised a novel method to quantify polyclonal antibody reactivity, which exploits protein

microspot assays and employs a novel analytical approach. Microarrays with a density series

of disease specific antigen were treated with different serum dilutions and developed for IgG

and IgA binding. By fitting binding data of both dilution series to a product of two

generalized logistic functions, we obtained estimates of antibody reactivity of two

immunoglobulin classes simultaneously. These estimates are the antigen concentrations

required for reaching the inflection point of thermodynamic activity coefficient of antibodies

and the limiting activity coefficient of antigen.

By providing universal chemical units, this method may improve standardization of

serological testing, the quality control of antibodies and the quantitative mapping of

antibody-antigen interaction space.

Keywords:

polyclonal, antibody, quantitative, serology, immunoglobulin, antigen, immune response,

chemical potential, activity coefficient, microarray, Richards curve

Introduction

Antibodies are glycoproteins belonging to the immunoglobulin protein superfamily, produced by B cells of the adaptive immune system. Antibodies in an organism are structurally highly heterogeneous owing to the gene recombination-mediated diversity of the adaptive immune response and the clonal nature of B-cell expansions[1,2]. A single given B cell and its progeny of plasmablasts and plasma cells produce antibody molecules with a given target specificity and binding affinity, determined by the fit of the rearranged heavy and light chain variable regions. Millions of B cells carrying different rearranged heavy and light chain variable regions generate a network of diverse antibodies present in an individual [3]. A humoral immune response against a particular antigen involves the evolution of antibody producing plasmablast and plasma cell clones with various isotypes and affinities, these latter properties also being characteristic for the polyclonal response besides target specificity. The affinity of an antibody molecule towards its target antigen is an important characteristic both from the biochemical and immunological point of view. Biochemically, it reflects the strength of interaction between the two binding partners [4]. Immunologically, it is closely correlated with effector functions such as neutralization efficiency and complement activation [5]. The isotype of the antibody molecule is the other key determinant of effector function. Understanding these properties of serum antibodies generated against particular antigenic targets is the goal of serological measurements, which aim at the laboratory diagnosis of infectious, allergic and autoimmune diseases [6–8]. Evidently, medical laboratory diagnostics requires the standardization of serological measurements.

Standardization of the measurement of antibody reactivity has always been a difficult issue [9] and in spite of much effort it has not been fully resolved [10]. Among the analytical issues is the fact that the quantitation of a heterogeneous population of antibodies would require a heterogenous reference material. Currently this is achieved by the utilization of affinity purified polyclonal serum pools or pools of monoclonal antibodies. These reagents, along with the standardization and harmonization of measurement protocols led to current best laboratory practices, which provide results in international units. However, quantitative comparison of different assays with different units is not meaningful, therefore systems level quantitative representation of antibody reactivity is currently impossible.

Determination of antibody affinity has always been considered of key importance for the characterization of humoral immunity. From the immunochemical and biophysical point of view, the techniques employed range from equilibrium dialysis [11], precipitation assays using radiolabeling [11], use of denaturants to disrupt binding forces [12], setting up competition assays

[13] and varying Ag density in enzyme-linked immunosorbent-assays (ELISA) [14]. A non-

competitive ELISA [15] for the measurement of monoclonal antibody affinity that relies on the

determination of Ab binding at different antigen coating densities shares some conceptual

basis with our assay, however it requires the use of known concentrations of monoclonal

antibodies. Alternatively, equilibrium binding constants can be obtained from kinetics

measurements by surface plasmon resonance analysis [16], biolayer interferometry [17],

microscale thermophoresis [18] and similar technologies. While these measurements are often

regarded as the golden standard of affinity determination, the instrumentation and data-

intensive nature of measurement may not be optimal for routine diagnostic use. Additionally,

label-free methods are not easily adaptable for the selective measurement of antibody classes,

a must for immunodiagnostics.

Serum antibody titration is a simple and widely used way of estimating antibody reactivity,

wherein the sample is serially diluted and the highest dilution that gives a signal reliably

discernible from the background, or the dilution resulting in half signal, are regarded as the

titer for endpoint and midpoint titration, respectively. Such titers may provide fast and cost-

effective semi-quantitative results but are not suitable for real quantitation of serum

antibodies because we are changing an unknown variable when we decrease antibody

concentration by serum dilution.

The unique properties of microspot immunoassays were described by Ekins [19], who

propagated their use for capture immunoassays [20]. The general idea is that owing to the

negligible amount of molecules present in the microspot probe compared to the amount of

analyte in the tested solution, these assays are mass independent [20]: interactions are governed

by concentration and affinity only. When affinity is constant and homogenous, as in a capture

immunoassay utilizing monoclonal antibody, the density of analyte bound to the surface is

determined solely by its concentration. Here we propose that mass independence of

microspot immunoassays in fact requires an analytical approach that is different from

conventional calculations based on the law of mass action [15,21], which applies to reactions in

solutions. We suggest the use of concepts and terminology of physical chemistry, which are

readily applicable to reactions at a solid-liquid interface [22,23].

In this paper we present a microspot-based approach that allows the estimation of apparent

chemical potential of distinct antibody isotypes in human serum without any prior

purification steps. The measurement relies on the simultaneous titration of antigen surface

density and serum dilution and on an improved mathematical model of fitting. Most

importantly, by changing a known variable of the system, antigen concentration, the

technology allows the absolute quantification of antigen specific reactivity of polyclonal

serum antibodies, allowing quantitative comparison of different samples, isotypes or

antigens.

Results

Experimental setup and properties of the measurement system

The experimental system we use is characteristic of protein microarray technology with some

key differences to traditional indirect ELISA. The solid surface used as antigen adsorbent has

higher binding capacity than conventional polystyrene. This results in a dose dependent

adsorptive binding over a wider concentration range. Therefore, the concentration of antigen

in the solution used for bio-printing (spotting of antigen) is expected to show linear

relationship with the surface density of the antigen on the microarray over a correspondingly

wider range. Antibodies that bind to the spotted antigens are detected by fluorescently

labelled secondary antibodies. Owing to the properties of fluorescent measurement, bound

antibodies can be detected over a wide range as well. To allow for the analysis and

visualization of signals spreading over several orders of magnitude, the logarithm of both

antigen concentrations and fluorescent signal intensities is used.

To establish a distribution of immune complex concentrations as a function of antigen density

we employed two different, complementary strategies (Fig. 1). First, increasing

concentrations of antigen [Ag] were printed on a solid surface as microspots. Owing to the

negligible amount of antigen in a microspot compared to the amount of antibody in the

reaction solution, these measurement conditions are called mass-independent or ambient

analyte immunoassay [19,20]. Antigen microspots therefore probe antibody reactivity in the

sample without significantly altering composition of the sample. Second, dilutions of the

serum samples of interest were prepared to vary the concentration of antibody-antigen

complexes [AbAg] in the measurement range and to examine the relationship between

antibody concentration [Ab] and [AbAg]. Even though the absolute free antibody

concentration [Ab] in the tested serum is not known, the relative concentrations of the

dilution series can be used to follow [Ab] effects by mathematical curve fitting. Dilutions of

serum samples were applied to the series of antigen density microspots and incubated to bring

the system in equilibrium.

Figure 1. Simultaneous titration of antigen density and serum antibody

The resulting measurement system examines the effect of changes in relative [Ab] and [Ag]

on the extent of Ab binding. The obtained experimental data are fitted with a growth function

to derive values of the parameters that characterize antibody reactivity in the sample (Fig. 1).

Curve fitting

The formation of AbAg complexes and the extent of antibody binding under equilibrium in

our assay can be interpreted as the indicator of Ab thermodynamic activity in the microspots.

The equilibrium concentration of AbAg complexes is a logistic function of ln[Ag], and the

logarithm of the fluorescent intensity is a linear function of ln[AbAg], therefore the growth of

fluorescent intensity can be described by the generalized logistic function [24,25] or Richards

function R(x) of x=ln[Ag], an extended form of the function most frequently used for

immunoassays. We use the following parametrization

$$R(x) = A * (1 + (d-1)e^{-k*(x-x_i)})^{\frac{1}{1-d}} \tag{1}$$

where A is Ab capacity, [Ab]; k is rate of exponential growth; d determines asymmetry, the

relative value of R at fastest growth; x_i is the point of fastest growth.

The capacity of the system is determined by the availability of antibody, [Ab]. While the

exact value is unknown, relative values corresponding to steps in dilution series can be used

in the fitting procedure. In our assay exponential growth of [AbAg] is allowed by the

provision of exponentially increasing [Ag] in a series of microspots. We assume that antigen

titration results in a binding curve following a generalized logistic curve with the rate of

exponential growth k=1, which corresponds to the exponential increase of antigen

concentration [Ag]=e$^{1*ln[Ag]}$. Shape parameter d of the generalized logistic function allows for

asymmetry in the binding curve. In our assay d is an index, which characterizes the Ab

composition of serum. Location of the fastest growth x_i with respect to ln[Ag] provides a

general measure of antibody affinity.

Affinities and concentrations of serum antibodies are distributed over a very wide range,

equilibrium dissociation constants from 10^{-5} to 10^{-11} and molar concentrations from pM to

nM, respectively. In order to better assess interactions over this range and to weigh curve

fitting against signal intensity-dependent variation, we use logarithmically transformed

signals of binding. Thus, fitting the Richards curve to our measurements therefore requires

the logarithmic form of the above equation

$$\ln R(x) = \ln(A) + \frac{1}{1-d} * \ln(1 + (d-1) * e^{-k*(x-xi)}) \tag{2}$$

By using the logarithm we transform thermodynamic activity to chemical potential of the Ab,

as determined by its standard molar Gibbs free energy, concentration and activity coefficient

$$\mu_{Ab} = \mu° + RTln(a_{Ab}) = \mu° + RTln(c_{Ab} * \gamma_{Ab}) = \mu° + RTln(c_{Ab}) + RTln(\gamma_{Ab}) \qquad (3$$

where μ is chemical potential; $\mu°$ is standard chemical potential or molar Gibbs free energy;

a is relative thermodynamic activity; γ is activity coefficient; and c is molar concentration.

Since the standard term serves as a reference point, we can still obtain relative potentials after

its removal. Then we can interpret ln(A) of the lnR(x) function as the mole fraction term of

chemical potential and the remaining part of the function describes the activity coefficient as

a function of [Ag] (for details see **Supplementary text 1**).

To demonstrate the dependence of ln(A) from the serum dilution we extended the formula of

lnR(x) to obtain the combined fitting of serum dilutions and antigen dilutions:

$$\ln R_2(x) = \ln(C) - m * \ln(B + z) + \frac{1}{1-d} * \ln\left(1 + (d-1) * e^{-(x-x_i)}\right) \qquad (4$$

where z is serum dilution, and m, B, C are auxiliary parameters of the logarithm of the other

generalized logistic function that replaced ln(A) and has a parametrization different from

formula (2.

Reliable curve fitting depends on the number and location of x=ln[Ag] measurement points

on the curve and requires the value of x_i to be in the measurement range.

Characterization of anti-deamidated gliadin peptide serum antibodies

As an exemplary antigen for the proof of our measurement principle we choose a deamidated

peptide sequence known to be the target of antibodies in celiac disease. Such an antigenic

peptide is a well-defined molecular target of both IgA and IgG antibodies in celiac patients

and is used in laboratory diagnostics. We used peptide concentrations corresponding to the

range of K_D values expected to occur in serum and peptide dilutions that well extended

beyond that range. Serum samples were diluted to span about two orders of magnitude and

correspond to dilutions conventionally used in serological diagnostics. Binding data were

fitted using the lnR$_2$(x) function of equation (4 introduced above and generated curves were

overlain on the binding data (Fig. 2)

Figure 2. Fitting in two dimensions

Using the values of parameters from the fitted binding curves of experimental data, we can generate binding curves normalized to Ab concentrations, which are curves of thermodynamic activity coefficients γ_{Ab}, as illustrated in Figure 3.

Figure 3. Comparative characterization of serum antibody binding

These curves are independent of antibody concentrations and fluorescent intensity and are therefore comparable for different classes of Ab in a given sample, for different samples and any combination thereof. Parameter x_i and γ_{Ab} are sufficient to quantitatively characterize the distribution of Ab thermodynamic activity in the tested serum sample as long as the Richards curve models binding events.

To confirm that conventional titration cannot provide exact results we calculated titers from the binding data obtained from the microspot experiments. The classical approach of serum titration was heavily dependent on the antigen density, as observed by others [13,26,27]. We calculated end-point titers and mid-point titers of the same measurements by logistic fitting of binding curves to distinct epitope densities, and obtained different titers for different antigen densities (Fig. 4). This antigen density dependence is not only avoided but is instead exploited in our approach of fitting curves to antigen density.

Figure 4. Conventional titration of serum

We also confirmed the applicability of our method to other antigens using a microarray printed with dilution series of citrullinated peptides known to be targets of autoantibodies of rheumatoid arthritis patients (**Supplementary Fig. S1**).

Measurement of reference monoclonal antibody properties

Current serological assays use standardized antibody preparations as reference material for quality control and calibration [10]. These preparations are monoclonal antibodies alone or in combinations of monoclonal antibodies, or pooled raw or isolated serum antibodies from positive samples. Because of the individual variance of affinity distribution of serum antibodies these approaches have the major drawback of assuming identical affinity distribution of reference and test samples. Titration by antigen density avoids this pitfall since it does not a priori assume but it rather identifies distribution of samples and reference standards. The incorporation of a reference antibody is also suitable to control antigen density

variation. A critical point of antigen density titration is the precise deposition of Ag on the

microarray surface, so that the real density of antigen corresponds to the nominal antigen

concentration used for printing. The precision can be estimated by measuring and

characterizing a reference antibody. We successfully used the monoclonal antibody 14D15

specific for an epitope in our deamidated gliadin peptide sequence to characterize their

interaction (Fig. 5).

Figure 5. Characterization of a monoclonal Ab

Discussion

The concentrations of serum antibodies and of targets of these antibodies, along with the

range of the strength of these interactions span several orders of magnitude. To obtain a

systems level structured map of these interactions we need technologies that provide

quantitative results of molecular interaction measurements over this wide range, from low

picomolar to micromolar concentrations. Here we present results, which together with earlier

observations [28,29] provide evidence and theoretical support for protein microspot-based

fluorescent detection as a method of choice.

The difficulty of rendering the measurement of polyclonal antibody responses quantitative

lies partly in the reproducible generation of rigorously characterized antigen, partly in the

mathematical description of heterogeneity of polyclonal antibody response. The generation of

synthetic epitopes for serological assays, along with functional characterization using

reference affinity reagents may solve the first issue [10,30]. Our approach may provide a solution

for the second one: establishing standardized assays that generate quantitative data with SI

units of measurement. The key for generating a quantitative description of polyclonal serum

antibody reactivity is the ability to dissect antibody concentration and affinity, the two

parameters that determine antigen saturation. This is basically achieved by titrating both

concentration and affinity: in addition to the conventional serum dilution series, a series of

antigen dilutions is used concomitantly. By diluting serum antibody, a series of antibody

concentrations is examined, while by additionally using microspots of antigen dilution series,

affinity distribution is assessed simultaneously. Microspots allow a mass-independent

interrogation of interactions on the chip surface, whereby the measurement becomes

independent of the relative total masses of Ag and Ab in the reaction volume. Complex

formation is determined only by concentration and affinity, which two measures can be

deconvolved by two-dimensional titration. By gradually increasing the Ag density of the

surface, an Ab with a given affinity will engage in lower affinity interactions as well. This results in the measurement of a range of affinities from the highest (detected on low density Ag spots) to more degenerate lower affinity interactions.

There have been several attempts to characterize the affinity of antibodies associated with autoimmunity [31–34], infectious disease [35,36], vaccination [37] and allergy [14,38–41]. The strength of binding of serum antibodies is determined by the distribution of clones with different affinities specific for the tested epitope. While the importance of antibody affinity in conferring pathogenicity or protectivity is widely acknowledged, affinity determination is not generally considered as part of diagnostic serology. The possible reason is that affinity measurement is technologically challenging and is not standardized, as outlined above. Affinity dependence of traditional immunoassays (e.g., RIA, ELISA, Farr assay, hemagglutination, complement-mediated haemolysis, precipitation) [42] practically excludes low affinity interactions from measurement and may skew affinity determinations when such assays are used. Such a progressive decrease of sensitivity to low affinity antibodies should lead to estimations of antibody affinity distribution with a higher mean affinity. The method we propose here presumes the existence of low affinity antibodies and measures them by using Ag at very high densities as well.

Microspot immunoassays were introduced for the measurement of analyte concentration, taking advantage of the very low amount of capture antibody required by the technology. If in the complete volume of measurement, the concentration of antibody on the solid support is less than 1% of the K_D of interaction, then the concentration of the analyte does not change significantly ($<1\%$) during the measurement. The signal from a detection antibody is therefore correlated only with analyte concentration when other variables are kept constant, which conditions are characteristic of ambient analyte immunoassay [19]. The analyte in our assay is serum antibody and the K_D of interactions are heterogeneous. Nevertheless, the amount of antigen and therefore its concentration in the total measurement volume is still negligible. Our method takes further advantage of microspots by varying the surface density of antigen for antibody capture. Since the relative concentrations of antibodies with different affinities are not changed by serum dilution, the affinity profile of antibodies that make contact with antigen on the solid support does not change either. The activity coefficient of this population of antibodies will depend solely on the density of antigen [Ag] relative to the average affinity of interactions. By varying [Ag] we obtain a distribution curve of chemical potentials suitable for estimating values of the parameters of fitted function.

The interpretation of Richards curve parameters provides insight into several aspects of
serum antibody reactivity. Parameter A is related to the molar concentration of functional
antibody binding sites, called paratopes, belonging to the measured immunoglobulin class.
Because the measurement of distribution of affinity reaches into the low affinity range, the
estimated concentrations are approximations of total serum antibody concentrations. We
expect this value to be dependent and correlated with serum antibody concentrations and
therefore to have less diagnostic value. Thus, it is presumably not necessary to adjust for
antibody valency and convert binding site concentrations into isotype concentrations.
Antibody responses that exploit the increase of antibody concentrations without major
increase in affinity, like the thymus-independent response, are therefore expected to appear as
a change in this parameter. A technical advantage of rendering the binding parameters
independent of total serum antibody concentrations is that we also exclude absolute
fluorescence intensity values from the analysis, thus removing a factor of interlaboratory
variance.

The central point or inflection point x_i is related to the apparent chemical potential of serum
antibodies. We propose to use the short name "lnK_D" for the natural logarithm of molar
concentration of antigen required to reach maximal relative growth of antibody activity
coefficient γ_{Ab} (inflection point of sigmoid curve in log-lin scale). While the unit of chemical
potential is that of energy (Joule/mole) here we would retain the unitless number derived
from antigen concentration titration.

The Richards curve uses parameter d to introduce asymmetry on the two sides of the
inflection point. At d=2 the curve is symmetric and it is a logistic curve. At d→1 the curve
approaches Gompertz growth and becomes asymmetric as the inflection point shifts from
y_i=A/2 to y_i=A/e.

The slope of the lnR_2(x) function s=k/(d-1) at minus infinity, in the case k=1, meanwhile
increases from s=1 to ∞. This slope characterizes interactions as antigen is diluted out to
infinity. Infinite dilution is a special thermodynamic state when antigen molecules are in
contact with antibody molecules only [43]. In this state binding is determined only by
interactions between Ag and Ab, without the interference of homotypic interactions. This
ideal state is characterized by the limiting activity coefficient γ_i^∞. In our parametrization d-1
changes between 0 and 1 and gives the probability of Ag binding to and remaining bound to
Ab until measurement. Thus, d-1 is the limiting activity coefficient of Ag, γ_{Ag}^∞; it
characterizes the composition of Ab mixture with 0 representing no interacting Abs and 1

representing 100% presence of binding Abs. We expect that parameter d can be used to

characterize disease activity when antibody diversity and affinity is related to disease

pathogenesis, with lower d implying immunological activity, and higher d indicating

approach of equilibrium concentrations and stability of the immune response.

A potential disadvantage of the proposed method is that it is technologically challenging,

compared to established automated assays. A critical point is the generation of antigen spots

with real densities as close to nominal densities as possible. This requires the introduction of

extra quality control steps into the production. The fitting procedure requires a minimal

number of data points around the estimated inflection point, therefore weakly reactive

samples cannot be quantitated. These disadvantages position our assay as a fine analysis

method, suitable not for the screening and identification of presence of specific antibodies but

rather for the detailed characterization of pre-screened sera.

A key conceptual novelty in our assay is that instead of assessing the concentration of a

heterogeneous antibody population we assess distribution of apparent chemical potential.

Current immunoassays focus on establishing conditions that are ideal for estimation of the

concentration of antibodies against a particular target antigen. For solid phase assays this

involves the optimization of coating antigen density, where actually an antigen density

suitable for identifying the shift of average K_D in a diagnostically relevant region of affinity is

identified. By using microspots with antigen densities spanning the whole range of relevant

concentrations, a single measurement can provide a distribution of apparent Ab chemical

potentials.

The procedure we describe here assumes that equilibrium is reached when microspots are

incubated in serum. In order to confirm that measurement conditions are appropriate and to

allow reproducibility and interassay comparisons, antibody standards can be introduced into

the assay. Pooled positive and borderline serum samples could be used in our assay just as in

any other immunoassay for specific antibody measurement. Additionally, monoclonal

antibody preparations are also suitable as internal reference, since calculations of sample Ab

affinity would not be based on the binding of reference antibody, but these would rather serve

as controls for validation and perhaps for adjusting calculations.

The introduction of such quantitative assays could not only improve diagnostic accuracy of

immunoassays but would also pave the way to our quantitative understanding of the adaptive

immune system. The generalized quantitative model of antibody homeostasis [3] provides a

conceptual framework for the experimental mapping of the "antibodyome" [44–46], a goal that

could be achieved by the strategic mapping of antigen structural space. The integration of

quantitative binding information into a network and the mapping of the interaction network [45]

to the sequence space generated by next generation sequencing [47] is expected to bring about

unprecedented systemic understanding of adaptive immunity in particular and protein

evolution in general.

Materials and methods

Microarray production and measurements

Experiments were carried out on hydrogel coated glass slides (Nexterion Slide H) by using a

BioOdyssey Calligrapher MiniArrayer. Different peptide dilutions, as indicated in the text,

were spotted in quintuplicates of ½ serial dilution in 7 steps starting from 200 μM. Slides

were dried for 1h at 37°C then soaked in 0.1 M Tris buffer (pH=8) for 1h at 37°C in order to

block reactive residues on the surface. Once prepared, slides were kept in sealed non-

transparent bags at 4 °C. The fusion peptide GLN1-2F was based on the deamidated gliadin

sequence motifs LQPFPQPELPYPQPQ and PLQPEQPFP.

Previously characterized, frozen serum samples collected for studies with contract numbers:

24933-6/2018/EKU and 24973-1/2012/EKU (658/PI/2012.) were used for the experiments.

Those studies were aimed at developing serological assays for gluten sensitivity and

autoimmune diseases, respectively, and all procedures were in accordance with the ethical

standards of the responsible committee on human experimentation and with the Helsinki

declaration of 1975, as revised in 2008.

Sample handling and signal detection

The basic protocol was the following: dried arrays were rehydrated in 110 μl PBS (3×5 min)

before using, then sub-arrays were incubated in 70 μl diluted sample at 37 °C for 1 h. Sample

dilution was carried out in PBS-BSA-Tween (PBS, 2% BSA, 0.05% Tween 20). Serum

treated slides were washed in 0.05% Tween-PBS, then incubated at room temperature for 30

minutes with fluorescently labelled antibodies that were diluted in the blocking buffer (0.05%

Tween 20, 2% BSA, PBS). Detecting antibodies were the following: Fc γ fragment specific

DyLight 649-conjugated AffinityPure F(ab')2 fragment goat anti-human IgG; DyLight 549-

conjugated AffinityPure F(ab')2 fragment goat anti human-IgA. Chips were washed again

and following drying, slides were scanned using laser microarray scanner.

Measurements with 14D5 monoclonal antibody

The previously described basic protocol was altered for the assays with monoclonal antibody

against deamidated gliadin (14D5). First, slides were rehydrated in PBS (3×5 min) then

blocked in blocking buffer (0.05% Tween 20, 2% BSA, PBS) for 30 minutes at 37°C. A

dilution series was prepared from the monoclonal antibody starting from 200 μg/ml by ½ in

10 steps. Detection was carried out by goat anti-mouse IgG2a-Alexa488.

Analysis of the microarray data

Images of the slides were analyzed with GenePix Pro 6.0 software after visual inspection.

Spots were recognized by the program, then gpr files containing the spot coordinates for

individual spots were created. Then R (version 3.5.2), a statistical programming environment,

was used to re-analyse 16-bit tiff images by using these coordinates in order to align the

spots. Relative fluorescence intensity (RFI) medians were calculated for each spot using 80%

of the diameter of the circle shapes that were previously adjusted on the features by the

software. In the next step, means of the 5 parallel spots were taken and values of the least

concentration antigen spots (considered as background) were subtracted of all RFI values,

thus providing the final signal intensities.

Fitting of binding curves

In the indirect assay, we take into consideration that a bound antibody inhibits nearby free

antigens from forming complexes with other antibodies [48,49]. This inhibition makes the

concentration of immune complexes a logistic function of the logarithm of the total antigen

concentration:

$[AbAg] = [Ab] * [Ag]/(K_D + [Ab] + D * [Ab] * [Ag])$ (6

where D is positive constant.

It is the nature of our measurement system that the logarithm fluorescent intensity is a linear

function of the logarithm concentration of AbAg complexes. By using these, we obtain that

the logarithm of fluorescent intensity is a Richards function of the logarithm of total antigen

concentration

$R(x) = A * (1 + (d - 1) * e^{-k(x-xi)})^{\frac{1}{1-d}}$ (7

with k = 1, where A is the total antibody concentration [Ab] (limit of function R(x) at

infinity),

x_i is the inflection point, d is the asymmetry parameter.

The upper limit of the fluorescent intensity A depends on the serum dilution. Formula (6) also

implies that the concentration of immune complexes is a logistic function of the logarithm of

the total antibody concentration ln[Ab] with growth rate 1, i.e. a logistic function of −ln(z)

with growth rate 1, where z is the serum dilution factor. Due to the linear dependence of the

logarithm fluorescent intensity on the logarithm printed Ag concentration, the upper limit of

the logarithm fluorescent intensity is a Richards function of the negative logarithm serum

dilution z with growth rate 1.

Therefore, the fluorescent signal intensity against the logarithm total antigen concentration x,

and the logarithm serum dilution z, is the product of two generalized logistic functions of

form

$R_2(x) = C * (B + z)^{-m} * (1 + (d - 1) * e^{-(x - xi)})^{\frac{1}{1-d}}$ (8

The logarithm transformation converts the proportional variance pattern to a constant

variance pattern and thus the conversion makes the transformed data more suitable for fitting

the model. The above multiplicative relationship now changes to an additive one in the form

of

$lnR_2(x) = \ln(C) - m * \ln(B + z) + \frac{1}{1-d} * \ln(1 + (d - 1) * e^{-(x - x_i)})$ (9

This generalized logistic model on the log-log scale was fit to the data, and parameter

estimates and 95% confidence intervals were obtained for each serum and both isotypes. All

analyses were carried out using the R software (version 3.6.2).

Nonlinear least squares estimates for the model parameters were calculated using the Gauss-

Newton algorithm of the nls function from the statistical software package R (version 3.5.1).

The 95% confidence intervals generated for the model parameters were based on the profile

likelihood technique.

Bibliography

1.   Imkeller, K. & Wardemann, H. Assessing human B cell repertoire diversity and
convergence. *Immunol. Rev.* **284**, 51–66 (2018).

2.   Briney, B., Inderbitzin, A., Joyce, C. & Burton, D. R. Commonality despite exceptional
diversity in the baseline human antibody repertoire. *Nature* **566**, 393–397 (2019).

3.   Prechl, J. A generalized quantitative antibody homeostasis model: antigen saturation,
natural antibodies and a quantitative antibody network. *Clin. Transl. Immunology* **6**,
e131 (2017).

4.   Marillet, S., Lefranc, M.-P., Boudinot, P. & Cazals, F. Novel Structural Parameters of
Ig-Ag Complexes Yield a Quantitative Description of Interaction Specificity and
Binding Affinity. *Front. Immunol.* **8**, 34 (2017).

5.   Prechl, J. A generalized quantitative antibody homeostasis model: maintenance of
global antibody equilibrium by effector functions. *Clin. Transl. Immunology* **6**, e161
(2017).

6.   Bacarese-Hamilton, T., Bistoni, F. & Crisanti, A. Protein microarrays: from
serodiagnosis to whole proteome scale analysis of the immune response against
pathogenic microorganisms. *BioTechniques* **Suppl**, 24–29 (2002).

7.   Meyer, P. W. A., Ally, M. M. T. M. & Anderson, R. Reliable and cost-effective
serodiagnosis of rheumatoid arthritis. *Rheumatol. Int.* **36**, 751–758 (2016).

8.   Chokshi, N. Y. & Sicherer, S. H. Interpreting IgE sensitization tests in food allergy.
Expert Rev. Clin. Immunol. **12**, 389–403 (2016).

9.   Bozsóky, S. The problem of standardization in rheumatoid arthritis serology. *Arthritis*
& Rheumatism **6**, 641–649 (1963).

10.   Meroni, P. L. *et al.* Standardization of autoantibody testing: a paradigm for serology in
498   rheumatic diseases. *Nat. Rev. Rheumatol.* **10**, 35–43 (2014).

499   11.   Stanley, C., Lew, A. M. & Steward, M. W. The measurement of antibody affinity: a
500   comparison of five techniques utilizing a panel of monoclonal anti-DNP antibodies and
501   the effect of high affinity antibody on the measurement of low affinity antibody. *J.*
502   *Immunol. Methods* **64**, 119–132 (1983).

503   12.   Macdonald, R. A., Hosking, C. S. & Jones, C. L. The measurement of relative
504   antibody affinity by ELISA using thiocyanate elution. *J. Immunol. Methods* **106**, 191–
505   194 (1988).

506   13.   Bruderer, U., Deusinger, M., Schürch, U. & Lang, A. B. Analyses of affinity
507   distributions within polyclonal populations of antigen-specific antibodies. Evaluation of
508   their accuracy in population detection using monoclonal antibodies. *J. Immunol.*
509   *Methods* **151**, 157–164 (1992).

510   14.   Nieto, A., Gaya, A., Jansa, M., Moreno, C. & Vives, J. Direct measurement of
511   antibody affinity distribution by hapten-inhibition enzyme immunoassay. *Mol. Immunol.*
512   **21**, 537–543 (1984).

513   15.   Beatty, J. D., Beatty, B. G. & Vlahos, W. G. Measurement of monoclonal antibody
514   affinity by non-competitive enzyme immunoassay. *J. Immunol. Methods* **100**, 173–179
515   (1987).

516   16.   Sem, D. S., McNeeley, P. A. & Linnik, M. D. Antibody affinities and relative titers in
517   polyclonal populations: surface plasmon resonance analysis of anti-DNA antibodies.
518   *Arch. Biochem. Biophys.* **372**, 62–68 (1999).

519   17.   Abdiche, Y., Malashock, D., Pinkerton, A. & Pons, J. Determining kinetics and
520   affinities of protein interactions using a parallel real-time label-free biosensor, the
521   Octet. *Anal. Biochem.* **377**, 209–217 (2008).

522   18.   Jerabek-Willemsen, M., Wienken, C. J., Braun, D., Baaske, P. & Duhr, S. Molecular

523   interaction studies using microscale thermophoresis. *Assay Drug Dev. Technol.* **9**,

524   342–353 (2011).

525   19.   Ekins, R., Chu, F. & Biggart, E. Multispot, multianalyte, immunoassay. *Ann Biol Clin*

526   *(Paris)* **48**, 655–666 (1990).

527   20.   Saviranta, P. *et al.* Evaluating sandwich immunoassays in microarray format in terms

528   of the ambient analyte regime. *Clin. Chem.* **50**, 1907–1920 (2004).

529   21.   Larsson, A. Assessment of polyclonal antibody binding of ligand by Sips' equation or

530   by the exact polyclonal equation. Comparison of models. *Mol. Immunol.* **25**, 1239–

531   1249 (1988).

532   22.   Kallay, N., Preočanin, T., Kovačević, D., Lützenkirchen, J. & Villalobos, M.

533   Thermodynamics of the reactions at solid/liquid interfaces. *Croat. Chem. Acta* 1–10

534   (2011) doi:10.5562/cca1864.

535   23.   Moradi, O. Thermodynamics of Interfaces. in *Thermodynamics - Interaction Studies -*

536   *Solids, Liquids and Gases* (ed. Moreno Pirajn, J. C.) (InTech, 2011).

537   doi:10.5772/20083.

538   24.   Gottschalk, P. G. & Dunn, J. R. The five-parameter logistic: a characterization and

539   comparison with the four-parameter logistic. *Anal. Biochem.* **343**, 54–65 (2005).

540   25.   Tsoularis, A. & Wallace, J. Analysis of logistic growth models. *Math. Biosci.* **179**, 21–

541   55 (2002).

542   26.   Hadzhieva, M. *et al.* Impact of antigen density on the binding mechanism of igg

543   antibodies. *Sci. Rep.* **7**, 3767 (2017).

544   27.   Lew, A. M. The effect of epitope density and antibody affinity on the ELISA as

545   analysed by monoclonal antibodies. *J. Immunol. Methods* **72**, 171–176 (1984).

28.   Kusnezow, W. *et al.* Optimal design of microarray immunoassays to compensate for
kinetic limitations: theory and experiment. *Mol. Cell. Proteomics* **5**, 1681–1696 (2006).

29.   Engelmann, B. W. *et al.* The development and application of a quantitative peptide
microarray based approach to protein interaction domain specificity space. *Mol. Cell.*
Proteomics **13**, 3647–3662 (2014).

30.   Rizwan, M. *et al.* Serology in the Digital Age: Using Long Synthetic Peptides Created
from Nucleic Acid Sequences as Antigens in Microarrays. *Microarrays (Basel)* **5**,
(2016).

31.   Achenbach, P. *et al.* Combined testing of antibody titer and affinity improves insulin
autoantibody measurement: Diabetes Antibody Standardization Program. *Clin.*
Immunol. **122**, 85–90 (2007).

32.   Szarka, E. *et al.* Affinity Purification and Comparative Biosensor Analysis of Citrulline-
Peptide-Specific Antibodies in Rheumatoid Arthritis. *Int. J. Mol. Sci.* **19**, (2018).

33.   Real-Fernández, F. *et al.* Surface Plasmon Resonance Method to Evaluate Anti-
citrullinated Protein/Peptide Antibody Affinity to Citrullinated Peptides. *Methods Mol.*
Biol. **1348**, 267–274 (2015).

34.   Sakakibara, S. *et al.* Clonal evolution and antigen recognition of anti-nuclear
antibodies in acute systemic lupus erythematosus. *Sci. Rep.* **7**, 16428 (2017).

35.   Poulsen, T. R., Jensen, A., Haurum, J. S. & Andersen, P. S. Limits for antibody
affinity maturation and repertoire diversification in hypervaccinated humans. *J.*
Immunol. **187**, 4229–4235 (2011).

36.   Reddy, S. B. *et al.* Differences in affinity of monoclonal and naturally acquired
polyclonal antibodies against Plasmodium falciparum merozoite antigens. *BMC*
Microbiol. **15**, 133 (2015).

37.   Pratesi, F. *et al.* BNT162b2 mRNA SARS-CoV-2 Vaccine Elicits High Avidity and

571   Neutralizing Antibodies in Healthcare Workers. *Vaccines (Basel)* **9**, (2021).

572   38.   Xiong, H., Dolpady, J., Wabl, M., Curotto de Lafaille, M. A. & Lafaille, J. J. Sequential

573   class switching is required for the generation of high affinity IgE antibodies. *J. Exp.*

574   *Med.* **209**, 353–364 (2012).

575   39.   Mita, H., Yasueda, H. & Akiyama, K. Affinity of IgE antibody to antigen influences

576   allergen-induced histamine release. *Clin. Exp. Allergy* **30**, 1583–1589 (2000).

577   40.   Wang, J. *et al.* Correlation of IgE/IgG4 milk epitopes and affinity of milk-specific IgE

578   antibodies with different phenotypes of clinical milk allergy. *J. Allergy Clin. Immunol.*

579   **125**, 695–702, 702.e1 (2010).

580   41.   Pierson, L., Allauzen, S., Blumenthal, M. & Rosenberg, A. An automated method for

581   determination of antibody affinity distribution functions with nanogram quantities. *J.*

582   *Immunol. Methods* **211**, 97–109 (1998).

583   42.   Péterfy, F., Kuusela, P. & Mäkelä, O. Affinity requirements for antibody assays

584   mapped by monoclonal antibodies. *J. Immunol.* **130**, 1809–1813 (1983).

585   43.   Alessi, P., Fermeglia, M. & Kikic, I. Significance of dilute regions. *Fluid Phase Equilib*

586   **70**, 239–250 (1991).

587   44.   Dimitrov, D. S. Therapeutic antibodies, vaccines and antibodyomes. *MAbs* **2**, 347–

588   356 (2010).

589   45.   Prechl, J. Network organization of antibody interactions in sequence and structure

590   space: the RADARS model. *Antibodies (Basel)* **9**, (2020).

591   46.   Prechl, J. Why current quantitative serology is not quantitative and how systems

592   immunology could provide solutions. *Biologia Futura* (2021) doi:10.1007/s42977-020-

593   00061-1.

594   47.   Prechl, J. Quantitative network organization of interactions emerging from the

595   evolution of sequence and structure space of antibodies: the RADARS model. *BioRxiv*

596   (2018) doi:10.1101/438804.

597   48.   De Michele, C., De Los Rios, P., Foffi, G. & Piazza, F. Simulation and Theory of

598   Antibody Binding to Crowded Antigen-Covered Surfaces. *PLoS Comput. Biol.* **12**,

599   e1004752 (2016).

600   49.   Edwards, D. A. Steric hindrance effects on surface reactions: applications to BIAcore.

601   *J. Math. Biol.* **55**, 517–539 (2007).

Acknowledgments

We thank participants of the GAPAID consortium, especially László Czirják, for providing serum samples for the study and Paolo Rovero for providing VCP2 peptide.

Author contributions

Á.K. analyzed data, developed model, wrote fitting program

J.P. conceptualized of strategy, designed experiments, wrote manuscript, created figures, supervised study

J.R. carried out measurements

K.L. designed experiments

K.P. designed experiments, analyzed data, contributed to figures

O.A. carried out measurement, analyzed data, wrote manuscript

T.P. conceptualized and developed model

Z.H. produced microarrays, carried out measurements

Conflict of interest

K.P., Z.H. and J.P. are employed by Diagnosticum Zrt, a company that provides custom microarray services. A provisional patent application related to this work has been filed by Diagnosticum zrt.

Funding

JP was partially supported by VEKOP-2.1.1-15-2016-00098 grant, part of Széchenyi 2020 project, to Diagnosticum. This work of TP was completed in the ELTE Institutional Excellence Program (TKP2020-IKA-05) financed by the Hungarian Ministry of Human Capacities.

631
632

Figure legends

Figure 1. Simultaneous titration of antigen density and serum antibody

Steps of the technology, starting with microarray fabrication and measurement, through

image analysis to curve fitting and visualization of results are shown. Serum dilution is

indicated by red drops, antigen density differences are represented by shades of blue circles.

Fitting data with generalized logistic curves yields two parameters, x_i and d, which

characterize chemical potential distribution of serum antibodies.

Figure 2. Fitting in two dimensions

Examples of parallel measurements of IgG and IgA binding to diagnostic celiac disease

peptide epitope are shown, with serum dilutions in three-fold (A) and two-fold (B) steps. Dot

symbols stand for measurement data, lines are fitted curves.

Figure 3. Comparative characterization of serum antibody binding

Using the values of the parameters obtained by curve fitting, calculated thermodynamic

activity coefficient distributions are comparable not only for serum samples (A) but also for

distinct antibody classes (B). Results shown in Figure 1 were used the generation of

normalized comparisons. Bar charts (C) show mean d and lnK_D values, and 95% confidence

intervals as whiskers.

Figure 4. Conventional titration of serum.

Mid-point and end-point titers of IgA and IgG are calculated by logistic fitting of binding to

distinct antigen densities and are displayed as a function of antigen density.

Figure 5. Characterization of a monoclonal Ab for reference.

Binding data of monoclonal Ab 14D15 was fitted using the same algorithm as for serum

antibodies. Binding data and fitted curves (A), calculated distribution of activity coefficient

(B) and estimates of parameters lnK_D and d (C) are shown.

Symbols and abbreviations

a	relative thermodynamic activity
A	antibody capacity [FI]
Ab	antibody
[Ab]	antibody concentration [mol/L]
AbAg	antibody-antigen complex
[AbAg]	antibody-antigen complex concentration [mol/L]
Ag	antigen
[Ag]	antigen concentration [mol/L]
B	auxiliary parameter
BSA	Bovine Serum Albumin
c	molar concentration [mol/L]
C	auxiliary parameter
d	shape parameter that determines asymmetry [-]
D	positive constant [1/M]
ELISA	Enzyme-Linked Immunosorbent Assay
γ	thermodynamic activity coefficient [-]
IgA	immunoglobulin A
IgG	immunoglobulin G
k	rate of exponential growth
K_D	equilibrium dissociation constant

m	auxiliary parameter
μ	chemical potential [J/mol]
$\mu\square$	standard chemical potential [J/mol]
mAb	monoclonal antibody
PBS	Phosphate-Buffered Saline
R	ideal gas constant [~8.314 J/K*mol]
R(x)	Richards function
RFI	Relative Fluorescence Intensity
RIA	Radioimmunoassay
s	slope of lnR(x) at minus infinity
T	temperature [K]
x_i	value of x at inflection point
y_i	value of y at inflection point
z	serum dilution factor [1/dilution]

Figure 1. Simultaneous titration of antigen density and serum antibody

Steps of the technology, starting with microarray fabrication and measurement, through image analysis to curve fitting and visualization of results are shown. Serum dilution is indicated by red drops, antigen density differences are represented by shades of blue circles. Fitting data with generalized logistic curves yields two parameters, x_i and d, which characterize chemical potential distribution of serum antibodies.

Figure 2. Fitting in two dimensions

Examples of parallel measurements of IgG and IgA binding to diagnostic celiac disease peptide epitope are shown, with serum dilutions in three-fold (A) and two-fold (B) steps. Dot symbols stand for measurement data, lines are fitted curves, numbers are sample identification numbers.

Figure 3. Comparative characterization of serum antibody binding

Using the values of the parameters obtained by curve fitting, calculated thermodynamic activity coefficient distributions are comparable not only for serum samples (A) but also for distinct antibody classes (B). Results from data shown in Figure 2 were used for the generation of normalized comparisons. Bar charts (C) show mean d and x_i values, and 95% confidence intervals as whiskers.

Figure 4. Conventional titration of serum.

Mid-point and end-point titers of IgA and IgG are calculated by logistic fitting of binding to distinct antigen densities and are displayed as a function of antigen density.

Figure 5. Characterization of a monoclonal Ab for reference.

Binding data of monoclonal Ab 14D15 was fitted using the same algorithm as for serum antibodies. Binding data and fitted curves (A), calculated distribution of activity coefficient (B) and estimates of parameters x_i and d (C) are shown.