1   **Absolute quantitation of serum antibody reactivity using the Richards growth model for**
2   **antigen microspot titration**
3

4   Krisztián Papp, Ágnes Kovács, Anita Orosz, Zoltán Hérincs, Judit Randek, Károly Liliom,
5   Tamás Pfeil, József Prechl
6
7   Author information
8   These authors contributed equally: Krisztián Papp, Ágnes Kovács, Anita Orosz
9
R&D Laboratory, Diagnosticum Zrt, Budapest, Hungary
Krisztián Papp, Zoltán Hérincs, József Prechl
Department of Applied Analysis and Computational Mathematics, Eötvös Loránd University,
Budapest, Hungary
Ágnes Kovács, Tamás Pfeil
Department of Immunology, Eötvös Loránd University, Budapest, Hungary
Anita Orosz
BME, Budapest University of Technology and Economics, Budapest, Hungary
Judit Randek
Department of Biophysics and Radiation Biology, Semmelweis University, Budapest, Hungary
Károly Liliom
21
22 Corresponding author: József Prechl, jprechl@diagnosticum.hu
23 1047 Attila út 126. Budapest, Hungary; T: +36-1-8810161
24
25 Current addresses
26 Anita Orosz : Department of Physiology, Semmelweis University, Budapest, Hungary;
27 Judit Randek : F. Hoffmann - La Roche Ltd., Basel, Switzerland
28

Abstract

In spite of its pivotal role in the characterization of humoral immunity, there is no accepted

method for the absolute quantitation of antigen specific serum antibodies.

We devised a novel method to quantify polyclonal antibody reactivity, which exploits protein

microspot assays and employs a novel analytical approach. Microarrays with a density series

of disease specific antigen were treated with different serum dilutions and developed for IgG

and IgA binding. By fitting binding data of both dilution series to a product of two generalized

logistic functions, we obtained estimates of antibody reactivity of two immunoglobulin classes

simultaneously. These estimates are the antigen concentrations required for reaching the

inflection point of thermodynamic activity coefficient of antibodies and the limiting activity

coefficient of antigen.

By providing universal chemical units, this method may improve standardization of serological

testing, the quality control of antibodies and the quantitative mapping of antibody-antigen

interaction space.

Keywords:

polyclonal, antibody, quantitative, serology, immunoglobulin, antigen, immune response,

chemical potential, activity coefficient, microarray, Richards curve

Introduction

Antibodies are glycoproteins belonging to the immunoglobulin protein superfamily, produced by B cells of the adaptive immune system. Antibodies in an organism are structurally highly heterogeneous owing to the gene recombination-mediated diversity of the adaptive immune response and the clonal nature of B-cell expansions[1,2]. A single given B cell and its progeny of plasmablasts and plasma cells produce antibody molecules with a given target specificity and binding affinity, determined by the fit of the rearranged heavy and light chain variable regions. Millions of B cells carrying different rearranged heavy and light chain variable regions generate a network of diverse antibodies present in an individual [3]. A humoral immune response against a particular antigen involves the evolution of antibody producing plasmablast and plasma cell clones with various isotypes and affinities, these latter properties also being characteristic for the polyclonal response besides target specificity. The affinity of an antibody molecule towards its target antigen is an important characteristic both from the biochemical and immunological point of view. Biochemically, it reflects the strength of interaction between the two binding partners [4]. Immunologically, it is closely correlated with effector functions such as neutralization efficiency and complement activation [5]. The isotype of the antibody molecule is the other key determinant of effector function. Understanding these properties of serum antibodies generated against particular antigenic targets is the goal of serological measurements, which aim at the laboratory diagnosis of infectious, allergic and autoimmune diseases [6–8]. Evidently, medical laboratory diagnostics requires the standardization of serological measurements.

Standardization of the measurement of antibody reactivity has always been a difficult issue [9] and in spite of much effort it has not been fully resolved [10]. Among the analytical issues is the fact that the quantitation of a heterogeneous population of antibodies would require a heterogenous reference material. Currently this is achieved by the utilization of affinity purified polyclonal serum pools or pools of monoclonal antibodies. These reagents, along with the standardization and harmonization of measurement protocols led to current best laboratory practices, which provide results in international units. However, quantitative comparison of different assays with different units is not meaningful, therefore systems level quantitative representation of antibody reactivity is currently impossible.

Determination of antibody affinity has always been considered of key importance for the characterization of humoral immunity. From the immunochemical and biophysical point of view, the techniques employed range from equilibrium dialysis [11], precipitation assays using radiolabeling [11], use of denaturants to disrupt binding forces [12], setting up competition assays

[13] and varying Ag density in enzyme-linked immunosorbent-assays (ELISA) [14]. A non-
competitive ELISA [15] for the measurement of monoclonal antibody affinity that relies on the
determination of Ab binding at different antigen coating densities shares some conceptual basis
with our assay, however it requires the use of known concentrations of monoclonal antibodies.
Alternatively, equilibrium binding constants can be obtained from kinetics measurements by
surface plasmon resonance analysis [16], biolayer interferometry [17], microscale thermophoresis
[18] and similar technologies. While these measurements are often regarded as the golden
standard of affinity determination, the instrumentation and data-intensive nature of
measurement may not be optimal for routine diagnostic use. Additionally, label-free methods
are not easily adaptable for the selective measurement of antibody classes, a must for
immunodiagnostics.

Serum antibody titration is a simple and widely used way of estimating antibody reactivity,
wherein the sample is serially diluted and the highest dilution that gives a signal reliably
discernible from the background, or the dilution resulting in half signal, are regarded as the
titer for endpoint and midpoint titration, respectively. Such titers may provide fast and cost-
effective semi-quantitative results but are not suitable for real quantitation of serum antibodies
because we are changing an unknown variable when we decrease antibody concentration by
serum dilution.

The unique properties of microspot immunoassays were described by Ekins [19], who propagated
their use for capture immunoassays [20]. The general idea is that owing to the negligible amount
of molecules present in the microspot probe compared to the amount of analyte in the tested
solution, these assays are mass independent [20]: interactions are governed by concentration and
affinity only. When affinity is constant and homogenous, as in a capture immunoassay utilizing
monoclonal antibody, the density of analyte bound to the surface is determined solely by its
concentration. Here we propose that mass independence of microspot immunoassays in fact
requires an analytical approach that is different from conventional calculations based on the
law of mass action [15,21], which applies to reactions in solutions. We suggest the use of concepts
and terminology of physical chemistry, which are readily applicable to reactions at a solid-
liquid interface [22,23].

In this paper we present a microspot-based approach that allows the estimation of apparent
chemical potential of distinct antibody isotypes in human serum without any prior purification
steps. The measurement relies on the simultaneous titration of antigen surface density and
serum dilution and on an improved mathematical model of fitting. Most importantly, by
changing a known variable of the system, antigen concentration, the technology allows the

absolute quantification of antigen specific reactivity of polyclonal serum antibodies, allowing

quantitative comparison of different samples, isotypes or antigens.

Results

Experimental setup and properties of the measurement system

The experimental system we use is characteristic of protein microarray technology with some

key differences to traditional indirect ELISA. The solid surface used as antigen adsorbent has

higher binding capacity than conventional polystyrene. This results in a dose dependent

adsorptive binding over a wider concentration range. Therefore, the concentration of antigen

in the solution used for bio-printing (spotting of antigen) is expected to show linear relationship

with the surface density of the antigen on the microarray over a correspondingly wider range.

Antibodies that bind to the spotted antigens are detected by fluorescently labelled secondary

antibodies. Owing to the properties of fluorescent measurement, bound antibodies can be

detected over a wide range as well. To allow for the analysis and visualization of signals

spreading over several orders of magnitude, the logarithm of both antigen concentrations and

fluorescent signal intensities is used.

To establish a distribution of immune complex concentrations as a function of antigen density

we employed two different, complementary strategies (Fig. 1). First, increasing concentrations

of antigen [Ag] were printed on a solid surface as microspots. Owing to the negligible amount

of antigen in a microspot compared to the amount of antibody in the reaction solution, these

measurement conditions are called mass-independent or ambient analyte immunoassay [19,20].

Antigen microspots therefore probe antibody reactivity in the sample without significantly

altering composition of the sample. Second, dilutions of the serum samples of interest were

prepared to vary the concentration of antibody-antigen complexes [AbAg] in the measurement

range and to examine the relationship between antibody concentration [Ab] and [AbAg]. Even

though the absolute free antibody concentration [Ab] in the tested serum is not known, the

relative concentrations of the dilution series can be used to follow [Ab] effects by mathematical

curve fitting. Dilutions of serum samples were applied to the series of antigen density

microspots and incubated to bring the system in equilibrium.

Figure 1. Simultaneous titration of antigen density and serum antibody

The resulting measurement system examines the effect of changes in relative [Ab] and [Ag] on

the extent of Ab binding. The obtained experimental data are fitted with a growth function to

derive values of the parameters that characterize antibody reactivity in the sample (Fig. 1).

Curve fitting

The formation of AbAg complexes and the extent of antibody binding under equilibrium in our

assay can be interpreted as the indicator of Ab thermodynamic activity in the microspots. The

equilibrium concentration of AbAg complexes is a logistic function of ln[Ag], and the

logarithm of the fluorescent intensity is a linear function of ln[AbAg], therefore the growth of

fluorescent intensity can be described by the generalized logistic function [24,25] or Richards

function R(x) of x=ln[Ag], an extended form of the function most frequently used for

immunoassays. We use the following parametrization

$$R(x) = A * (1 + (d - 1)e^{-k*(x-x_i)})^{\frac{1}{1-d}}$$ (1

where A is Ab capacity, [Ab]; k is rate of exponential growth; d determines asymmetry, the

relative value of R at fastest growth; x_i is the point of fastest growth or inflection point.

The capacity of the system is determined by the availability of antibody, [Ab]. While the exact

value is unknown, relative values corresponding to steps in dilution series can be used in the

fitting procedure. In our assay exponential growth of [AbAg] is allowed by the provision of

exponentially increasing [Ag] in a series of microspots. We assume that antigen titration results

in a binding curve following a generalized logistic curve with the rate of exponential growth

k=1, which corresponds to $[Ag]=e^{1*\ln[Ag]}$. Shape parameter d of the generalized logistic function

allows for asymmetry in the binding curve. In our assay d is an index, which characterizes the

Ab composition of serum. Location of the fastest growth x_i with respect to ln[Ag] provides a

general measure of antibody affinity.

Affinities and concentrations of serum antibodies are distributed over a very wide range,

equilibrium dissociation constants from 10^{-5} to 10^{-11} and molar concentrations from pM to nM,

respectively. In order to better assess interactions over this range and to weigh curve fitting

against signal intensity-dependent variation, we use logarithmically transformed signals of

binding. Thus, fitting the Richards curve to our measurements therefore requires the

logarithmic form of the above equation

$$\ln R(x) = \ln(A) + \frac{1}{1-d} * \ln(1 + (d - 1) * e^{-k*(x-x_i)})$$ (2

By using the logarithm we transform thermodynamic activity to chemical potential of the Ab,

as determined by its standard molar Gibbs free energy, concentration and activity coefficient

$$\mu_{Ab} = \mu° + RTln(a_{Ab}) = \mu° + RTln(c_{Ab} * \gamma_{Ab}) = \mu° + RTln(c_{Ab}) + RTln(\gamma_{Ab}) \qquad (3$$

where μ is chemical potential; μ° is standard chemical potential or molar Gibbs free energy;

a is relative thermodynamic activity; γ is activity coefficient; and c is molar concentration.

Since the standard term serves as a reference point, we can still obtain relative potentials after

its removal. Then we can interpret ln(A) of the lnR(x) function as the mole fraction term of

chemical potential and the remaining part of the function describes the activity coefficient as a

function of [Ag] (for details see **Supplementary text 1**).

To demonstrate the dependence of ln(A) from the serum dilution we extended the formula of

lnR(x) to obtain the combined fitting of serum dilutions and antigen dilutions:

$$lnR_2(x, z) = ln(C) - m * ln(B + z) + \frac{1}{1-d} * ln\left(1 + (d-1) * e^{-(x-x_i)}\right) \qquad (4$$

where z is serum dilution, and m, B, C are auxiliary parameters of the logarithm of another

generalized logistic function(Tjørve and Tjørve 2010) of -ln(z) that replaced ln(A) and has a

parametrization different from formula (2.

Reliable curve fitting depends on the number and location of x=ln[Ag] measurement points on

the curve and requires the value of x_i to be in the measurement range.

Characterization of anti-deamidated gliadin peptide serum antibodies

As an exemplary antigen for the proof of our measurement principle we choose a deamidated

peptide sequence known to be the target of antibodies in celiac disease. Such an antigenic

peptide is a well-defined molecular target of both IgA and IgG antibodies in celiac patients and

is used in laboratory diagnostics. We used peptide concentrations corresponding to the range

of K_D values expected to occur in serum and peptide dilutions that well extended beyond that

range. Serum samples were diluted to span about two orders of magnitude and correspond to

dilutions conventionally used in serological diagnostics. Binding data were fitted using the

$lnR_2(x)$ function of equation (4 introduced above and generated curves were overlain on the

binding data (Fig. 2)

Figure 2. Fitting in two dimensions

Using the values of parameters from the fitted binding curves of experimental data, we can

generate binding curves normalized to Ab concentrations, which are curves of thermodynamic

activity coefficients γ_{Ab}, as illustrated in Figure 3.

Figure 3. Comparative characterization of serum antibody binding

These curves are independent of antibody concentrations and fluorescent intensity and are

therefore comparable for different classes of Ab in a given sample, for different samples and

any combination thereof. Parameter x_i and γ_{Ab} are sufficient to quantitatively characterize the

distribution of Ab thermodynamic activity in the tested serum sample as long as the Richards

curve models binding events.

To confirm that conventional titration cannot provide exact results we calculated titers from

the binding data obtained from the microspot experiments. The classical approach of serum

titration was heavily dependent on the antigen density, as observed by others [13,26,27]. We

calculated end-point titers and mid-point titers of the same measurements by logistic fitting of

binding curves to distinct antigen densities, and obtained different titers for different antigen

densities (Fig. 4). This antigen density dependence is not only avoided but is instead exploited

in our approach of fitting curves to antigen density.

Figure 4. Conventional titration of serum

We also confirmed the applicability of our method to other antigens using a microarray printed

with dilution series of citrullinated peptides known to be targets of autoantibodies of

rheumatoid arthritis patients (**Supplementary Fig. S1**).

<u>*Measurement of reference monoclonal antibody properties*</u>

Current serological assays use standardized antibody preparations as reference material for

quality control and calibration [10]. These preparations are monoclonal antibodies alone or in

combinations of monoclonal antibodies, or pooled raw or isolated serum antibodies from

positive samples. Because of the individual variance of affinity distribution of serum antibodies

these approaches have the major drawback of assuming identical affinity distribution of

reference and test samples. Titration by antigen density avoids this pitfall since it does not a

priori assume but it rather identifies distribution of samples and reference standards. The

incorporation of a reference antibody is also suitable to control antigen density variation. A

critical point of antigen density titration is the precise deposition of Ag on the microarray

surface, so that the real density of antigen corresponds to the nominal antigen concentration

used for printing. The precision can be estimated by measuring and characterizing a reference

antibody. We successfully used the monoclonal antibody 14D15 specific for an epitope in our

deamidated gliadin peptide sequence to characterize their interaction (Fig. 5).

Figure 5. Characterization of a monoclonal Ab

Discussion

The concentrations of serum antibodies and of targets of these antibodies, along with the range

of the strength of these interactions span several orders of magnitude. To obtain a systems level

structured map of these interactions we need technologies that provide quantitative results of

molecular interaction measurements over this wide range, from low picomolar to micromolar

concentrations. Here we present results, which together with earlier observations [28,29] provide

evidence and theoretical support for protein microspot-based fluorescent detection as a method

of choice.

The difficulty of rendering the measurement of polyclonal antibody responses quantitative lies

partly in the reproducible generation of rigorously characterized antigen, partly in the

mathematical description of heterogeneity of polyclonal antibody response. The generation of

synthetic epitopes for serological assays, along with functional characterization using reference

affinity reagents may solve the first issue [10,30]. Our approach may provide a solution for the

second one: establishing standardized assays that generate quantitative data with SI units of

measurement. The key for generating a quantitative description of polyclonal serum antibody

reactivity is the ability to dissect antibody concentration and affinity, the two parameters that

determine antigen saturation. This is basically achieved by titrating both concentration and

affinity: in addition to the conventional serum dilution series, a series of antigen dilutions is

used concomitantly. By diluting serum antibody, a series of antibody concentrations is

examined, while by additionally using microspots of antigen dilution series, affinity

distribution is assessed simultaneously. Microspots allow a mass-independent interrogation of

interactions on the chip surface, whereby the measurement becomes independent of the relative

total masses of Ag and Ab in the reaction volume. Complex formation is determined only by

concentration and affinity, which two measures can be deconvolved by two-dimensional

titration. By gradually increasing the Ag density of the surface, an Ab with a given affinity will

engage in lower affinity interactions as well. This results in the measurement of a range of

affinities from the highest (detected on low density Ag spots) to more degenerate lower affinity

interactions.

There have been several attempts to characterize the affinity of antibodies associated with

autoimmunity [31–34], infectious disease [35,36], vaccination [37] and allergy [14,38–41]. The strength of

binding of serum antibodies is determined by the distribution of clones with different affinities

specific for the tested epitope. While the importance of antibody affinity in conferring

pathogenicity or protectivity is widely acknowledged, affinity determination is not generally

considered as part of diagnostic serology. The possible reason is that affinity measurement is

technologically challenging and is not standardized, as outlined above. Affinity dependence of

traditional immunoassays (e.g., RIA, ELISA, Farr assay, hemagglutination, complement-

mediated haemolysis, precipitation) [42] practically excludes low affinity interactions from

measurement and may skew affinity determinations when such assays are used. Such a

progressive decrease of sensitivity to low affinity antibodies should lead to estimations of

antibody affinity distribution with a higher mean affinity. The method we propose here

presumes the existence of low affinity antibodies and measures them by using Ag at very high

densities as well.

Microspot immunoassays were introduced for the measurement of analyte concentration,

taking advantage of the very low amount of capture antibody required by the technology. If in

the complete volume of measurement, the concentration of antibody on the solid support is less

than 1% of the K_D of interaction, then the concentration of the analyte does not change

significantly (<1%) during the measurement. The signal from a detection antibody is therefore

correlated only with analyte concentration when other variables are kept constant, which

conditions are characteristic of ambient analyte immunoassay [19]. The analyte in our assay is

serum antibody and the K_D of interactions are heterogeneous. Nevertheless, the amount of

antigen and therefore its concentration in the total measurement volume is still negligible. Our

method takes further advantage of microspots by varying the surface density of antigen for

antibody capture. Since the relative concentrations of antibodies with different affinities are

not changed by serum dilution, the affinity profile of antibodies that make contact with antigen

on the solid support does not change either. The activity coefficient of this population of

antibodies will depend solely on the density of antigen [Ag] relative to the average affinity of

interactions. By varying [Ag] we obtain a distribution curve of chemical potentials suitable for

estimating values of the parameters of fitted function.

The interpretation of Richards curve parameters provides insight into several aspects of serum

antibody reactivity. Parameter A is related to the molar concentration of functional antibody

binding sites, called paratopes, belonging to the measured immunoglobulin class. Because the

measurement of distribution of affinity reaches into the low affinity range, the estimated

concentrations are approximations of total serum antibody concentrations. We expect this value

to be dependent and correlated with serum antibody concentrations and therefore to have less

diagnostic value. Thus, it is presumably not necessary to adjust for antibody valency and

convert binding site concentrations into isotype concentrations. Antibody responses that exploit

the increase of antibody concentrations without major increase in affinity, like the thymus-

independent response, are therefore expected to appear as a change in this parameter. A

technical advantage of rendering the binding parameters independent of total serum antibody

concentrations is that we also exclude absolute fluorescence intensity values from the analysis,

thus removing a factor of interlaboratory variance.

The central point or inflection point x_i is related to the apparent chemical potential of serum

antibodies. We propose to use the short name "$\ln K_D$" for the natural logarithm of molar

concentration of antigen required to reach maximal relative growth of antibody activity

coefficient γ_{Ab} (inflection point of sigmoid curve in log-lin scale). While the unit of chemical

potential is that of energy (Joule/mole) here we would retain the unitless number derived from

antigen concentration titration.

The Richards curve uses parameter d to introduce asymmetry on the two sides of the inflection

point. At d=2 the curve is symmetric and it is a logistic curve. At d→1 the curve approaches

the Gompertz growth curve and becomes asymmetric as the inflection point shifts from $y_i = A/2$

to $y_i = A/e$.

The slope of the $\ln R_2(x)$ function $s = k/(d-1)$ at minus infinity, in the case k=1, meanwhile

increases from s=1 to ∞. This slope characterizes interactions as antigen is diluted out to

infinity. Infinite dilution is a special thermodynamic state when antigen molecules are in

contact with antibody molecules only [43]. In this state binding is determined only by interactions

between Ag and Ab, without the interference of homotypic interactions. This ideal state is

characterized by the limiting activity coefficient γ_i^∞. Using our parametrization d-1 changed

between 0 and 1. We propose that d-1 is the limiting activity coefficient of Ag, γ_{Ag}^∞. It

characterizes the composition of Ab mixture with 0 representing no interacting Abs and 1

representing 100% presence of binding Abs at infinite Ag dilution. We expect that parameter

d can be used to characterize disease activity when antibody diversity and affinity is related to

disease pathogenesis, with lower d implying immunological activity, and higher d indicating

approach of equilibrium concentrations and stability of the immune response.

A potential disadvantage of the proposed method is that it is technologically challenging,
compared to established automated assays. A critical point is the generation of antigen spots
with real densities as close to nominal densities as possible. This requires the introduction of
extra quality control steps into the production. The fitting procedure requires a minimal number
of data points around the estimated inflection point, therefore weakly reactive samples cannot
be quantitated. These disadvantages position our assay as a fine analysis method, suitable not
for the screening and identification of presence of specific antibodies but rather for the detailed
characterization of pre-screened sera.

A key conceptual novelty in our assay is that instead of assessing the concentration of a
heterogeneous antibody population we assess distribution of apparent chemical potential.
Current immunoassays focus on establishing conditions that are ideal for estimation of the
concentration of antibodies against a particular target antigen. For solid phase assays this
involves the optimization of coating antigen density, where actually an antigen density suitable
for identifying the shift of average K_D in a diagnostically relevant region of affinity is
identified. By using microspots with antigen densities spanning the whole range of relevant
concentrations, a single measurement can provide a distribution of apparent Ab chemical
potentials.

The procedure we describe here assumes that equilibrium is reached when microspots are
incubated in serum. In order to confirm that measurement conditions are appropriate and to
allow reproducibility and interassay comparisons, antibody standards can be introduced into
the assay. Pooled positive and borderline serum samples could be used in our assay just as in
any other immunoassay for specific antibody measurement. Additionally, monoclonal antibody
preparations are also suitable as internal reference, since calculations of sample Ab affinity
would not be based on the binding of reference antibody, but these would rather serve as
controls for validation and perhaps for adjusting calculations.

The introduction of such quantitative assays could not only improve diagnostic accuracy of
immunoassays but would also pave the way to our quantitative understanding of the adaptive
immune system. The generalized quantitative model of antibody homeostasis [3] provides a
conceptual framework for the experimental mapping of the "antibodyome" [44–46], a goal that
could be achieved by the strategic mapping of antigen structural space. The integration of
quantitative binding information into a network and the mapping of the interaction network [45]
to the sequence space generated by next generation sequencing [47] is expected to bring about
unprecedented systemic understanding of adaptive immunity in particular and protein
evolution in general.

Materials and methods

Microarray production and measurements

Experiments were carried out on hydrogel coated glass slides (Nexterion Slide H) by using a BioOdyssey Calligrapher MiniArrayer. Different peptide dilutions, as indicated in the text, were spotted in quintuplicates of ½ serial dilution in 7 steps starting from 200 μM. Slides were dried for 1h at 37°C then soaked in 0.1 M Tris buffer (pH=8) for 1h at 37°C in order to block reactive residues on the surface. Once prepared, slides were kept in sealed non-transparent bags at 4 °C. The fusion peptide GLN1-2F was based on the deamidated gliadin sequence motifs LQPFPQPELPYPQPQ and PLQPEQPFP.

Previously characterized, frozen serum samples collected for studies with contract numbers: 24933-6/2018/EKU and 24973-1/2012/EKU (658/PI/2012.) were used for the experiments. Those studies were aimed at developing serological assays for gluten sensitivity and autoimmune diseases, respectively, and all procedures were in accordance with the ethical standards of the responsible committee on human experimentation and with the Helsinki declaration of 1975, as revised in 2008.

Sample handling and signal detection

The basic protocol was the following: dried arrays were rehydrated in 110 µl PBS (3×5 min) before using, then sub-arrays were incubated in 70 µl diluted sample at 37 °C for 1 h. Sample dilution was carried out in PBS-BSA-Tween (PBS, 2% BSA, 0.05% Tween 20). Serum treated slides were washed in 0.05% Tween-PBS, then incubated at room temperature for 30 minutes with fluorescently labelled antibodies that were diluted in the blocking buffer (0.05% Tween 20, 2% BSA, PBS). Detecting antibodies were the following: Fc γ fragment specific DyLight 649-conjugated AffinityPure F(ab')2 fragment goat anti-human IgG; DyLight 549-conjugated AffinityPure F(ab')2 fragment goat anti human-IgA. Chips were washed again and following drying, slides were scanned using laser microarray scanner.

Measurements with 14D5 monoclonal antibody

The previously described basic protocol was altered for the assays with monoclonal antibody against deamidated gliadin (14D5). First, slides were rehydrated in PBS (3×5 min) then blocked in blocking buffer (0.05% Tween 20, 2% BSA, PBS) for 30 minutes at 37°C. A dilution series was prepared from the monoclonal antibody starting from 200 μg/ml by ½ in 10 steps. Detection was carried out by goat anti-mouse IgG2a-Alexa488.

Analysis of the microarray data

Images of the slides were analyzed with GenePix Pro 6.0 software after visual inspection. Spots were recognized by the program, then gpr files containing the spot coordinates for individual spots were created. Then R (version 3.5.2), a statistical programming environment, was used to re-analyse 16-bit tiff images by using these coordinates in order to align the spots. Relative fluorescence intensity (RFI) medians were calculated for each spot using 80% of the diameter of the circle shapes that were previously adjusted on the features by the software. In the next step, means of the 5 parallel spots were taken and values of the least concentration antigen spots (considered as background) were subtracted of all RFI values, thus providing the final signal intensities.

Fitting of binding curves

In the indirect assay, we take into consideration that a bound antibody inhibits nearby free antigens from forming complexes with other antibodies [48,49]. This inhibition makes the concentration of immune complexes a logistic function of the logarithm of the total antigen concentration:

$$[AbAg] = [Ab] * [Ag]/(K_D + [Ab] + D * [Ab] * [Ag]) \qquad (6$$

where D is positive constant.

It is the nature of our measurement system that the logarithm fluorescent intensity is a linear function of the logarithm concentration of AbAg complexes. By using these, we obtain that the logarithm of fluorescent intensity is a Richards function of the logarithm of total antigen concentration

$$R(x) = A * (1 + (d - 1) * e^{-k(x-x_i)})^{\frac{1}{1-d}} \qquad (7$$

with k = 1, where A is the total antibody concentration [Ab] (limit of function R(x) at infinity), x_i is the inflection point, d is the asymmetry parameter.

The upper limit of the fluorescent intensity A depends on the serum dilution. Formula (6) also implies that the concentration of immune complexes is a logistic function of the logarithm of the total antibody concentration ln[Ab] with growth rate 1, i.e. a logistic function of $-\ln(z)$ with growth rate 1, where z is the serum dilution factor. Due to the linear dependence of the logarithm fluorescent intensity on the logarithm printed Ag concentration, the upper limit of the logarithm fluorescent intensity is a Richards function of the negative logarithm serum dilution z with growth rate 1.

Therefore, the fluorescent signal intensity against the logarithm total antigen concentration x,
and the negative logarithm of serum dilution, -ln(z), is the product of two generalized logistic
functions of form

$$R_2(x,z) = C * (B + z)^{-m} * (1 + (d - 1) * e^{-(x-x_i)})^{\frac{1}{1-d}} \tag{8}$$

The logarithm transformation converts the proportional variance pattern to a constant variance
pattern and thus the conversion makes the transformed data more suitable for fitting the model.
The above multiplicative relationship now changes to an additive one in the form of

$$lnR_2(x,z) = \ln(C) - m * \ln(B + z) + \frac{1}{1-d} * \ln(1 + (d - 1) * e^{-(x-x_i)}) \tag{9}$$

This generalized logistic model on the log-log scale was fit to the data, and parameter estimates
and 95% confidence intervals were obtained for each serum and both isotypes. All analyses
were carried out using the R software (version 3.5.2).

Nonlinear least squares estimates for the model parameters were calculated using the Gauss-
Newton algorithm of the nls function from the statistical software package R (version 3.5.2).
The 95% confidence intervals generated for the model parameters were based on the profile
likelihood technique.

Bibliography

1.   Imkeller, K. & Wardemann, H. Assessing human B cell repertoire diversity and

467   convergence. *Immunol. Rev.* **284**, 51–66 (2018).

468   2.   Briney, B., Inderbitzin, A., Joyce, C. & Burton, D. R. Commonality despite exceptional

469   diversity in the baseline human antibody repertoire. *Nature* **566**, 393–397 (2019).

470   3.   Prechl, J. A generalized quantitative antibody homeostasis model: antigen saturation,

471   natural antibodies and a quantitative antibody network. *Clin. Transl. Immunology* **6**,

472   e131 (2017).

473   4.   Marillet, S., Lefranc, M.-P., Boudinot, P. & Cazals, F. Novel Structural Parameters of

474   Ig-Ag Complexes Yield a Quantitative Description of Interaction Specificity and

475   Binding Affinity. *Front. Immunol.* **8**, 34 (2017).

476   5.   Prechl, J. A generalized quantitative antibody homeostasis model: maintenance of

477   global antibody equilibrium by effector functions. *Clin. Transl. Immunology* **6**, e161

478   (2017).

479   6.   Bacarese-Hamilton, T., Bistoni, F. & Crisanti, A. Protein microarrays: from

480   serodiagnosis to whole proteome scale analysis of the immune response against

481   pathogenic microorganisms. *BioTechniques* **Suppl**, 24–29 (2002).

482   7.   Meyer, P. W. A., Ally, M. M. T. M. & Anderson, R. Reliable and cost-effective

483   serodiagnosis of rheumatoid arthritis. *Rheumatol. Int.* **36**, 751–758 (2016).

484   8.   Chokshi, N. Y. & Sicherer, S. H. Interpreting IgE sensitization tests in food allergy.

485   *Expert Rev. Clin. Immunol.* **12**, 389–403 (2016).

486   9.   Bozsóky, S. The problem of standardization in rheumatoid arthritis serology. *Arthritis

487   & Rheumatism* **6**, 641–649 (1963).

488   10.   Meroni, P. L. *et al.* Standardization of autoantibody testing: a paradigm for serology in
489   rheumatic diseases. *Nat. Rev. Rheumatol.* **10**, 35–43 (2014).

490   11.   Stanley, C., Lew, A. M. & Steward, M. W. The measurement of antibody affinity: a
491   comparison of five techniques utilizing a panel of monoclonal anti-DNP antibodies and
492   the effect of high affinity antibody on the measurement of low affinity antibody. *J.*
493   *Immunol. Methods* **64**, 119–132 (1983).

494   12.   Macdonald, R. A., Hosking, C. S. & Jones, C. L. The measurement of relative
495   antibody affinity by ELISA using thiocyanate elution. *J. Immunol. Methods* **106**, 191–
496   194 (1988).

497   13.   Bruderer, U., Deusinger, M., Schürch, U. & Lang, A. B. Analyses of affinity
498   distributions within polyclonal populations of antigen-specific antibodies. Evaluation of
499   their accuracy in population detection using monoclonal antibodies. *J. Immunol.*
500   *Methods* **151**, 157–164 (1992).

501   14.   Nieto, A., Gaya, A., Jansa, M., Moreno, C. & Vives, J. Direct measurement of
502   antibody affinity distribution by hapten-inhibition enzyme immunoassay. *Mol. Immunol.*
503   **21**, 537–543 (1984).

504   15.   Beatty, J. D., Beatty, B. G. & Vlahos, W. G. Measurement of monoclonal antibody
505   affinity by non-competitive enzyme immunoassay. *J. Immunol. Methods* **100**, 173–179
506   (1987).

507   16.   Sem, D. S., McNeeley, P. A. & Linnik, M. D. Antibody affinities and relative titers in
508   polyclonal populations: surface plasmon resonance analysis of anti-DNA antibodies.
509   *Arch. Biochem. Biophys.* **372**, 62–68 (1999).

510   17.   Abdiche, Y., Malashock, D., Pinkerton, A. & Pons, J. Determining kinetics and
511   affinities of protein interactions using a parallel real-time label-free biosensor, the
512   Octet. *Anal. Biochem.* **377**, 209–217 (2008).

513   18.   Jerabek-Willemsen, M., Wienken, C. J., Braun, D., Baaske, P. & Duhr, S. Molecular

514   interaction studies using microscale thermophoresis. *Assay Drug Dev. Technol.* **9**,

515   342–353 (2011).

516   19.   Ekins, R., Chu, F. & Biggart, E. Multispot, multianalyte, immunoassay. *Ann Biol Clin*

517   *(Paris)* **48**, 655–666 (1990).

518   20.   Saviranta, P. *et al.* Evaluating sandwich immunoassays in microarray format in terms

519   of the ambient analyte regime. *Clin. Chem.* **50**, 1907–1920 (2004).

520   21.   Larsson, A. Assessment of polyclonal antibody binding of ligand by Sips' equation or

521   by the exact polyclonal equation. Comparison of models. *Mol. Immunol.* **25**, 1239–

522   1249 (1988).

523   22.   Kallay, N., Preočanin, T., Kovačević, D., Lützenkirchen, J. & Villalobos, M.

524   Thermodynamics of the reactions at solid/liquid interfaces. *Croat. Chem. Acta* 1–10

525   (2011) doi:10.5562/cca1864.

526   23.   Moradi, O. Thermodynamics of Interfaces. in *Thermodynamics - Interaction Studies -*

527   *Solids, Liquids and Gases* (ed. Moreno Pirajn, J. C.) (InTech, 2011).

528   doi:10.5772/20083.

529   24.   Gottschalk, P. G. & Dunn, J. R. The five-parameter logistic: a characterization and

530   comparison with the four-parameter logistic. *Anal. Biochem.* **343**, 54–65 (2005).

531   25.   Tsoularis, A. & Wallace, J. Analysis of logistic growth models. *Math. Biosci.* **179**, 21–

532   55 (2002).

533   26.   Hadzhieva, M. *et al.* Impact of antigen density on the binding mechanism of igg

534   antibodies. *Sci. Rep.* **7**, 3767 (2017).

535   27.   Lew, A. M. The effect of epitope density and antibody affinity on the ELISA as

536   analysed by monoclonal antibodies. *J. Immunol. Methods* **72**, 171–176 (1984).

537   28.   Kusnezow, W. *et al.* Optimal design of microarray immunoassays to compensate for
538   kinetic limitations: theory and experiment. *Mol. Cell. Proteomics* **5**, 1681–1696 (2006).

539   29.   Engelmann, B. W. *et al.* The development and application of a quantitative peptide
540   microarray based approach to protein interaction domain specificity space. *Mol. Cell.*
541   *Proteomics* **13**, 3647–3662 (2014).

542   30.   Rizwan, M. *et al.* Serology in the Digital Age: Using Long Synthetic Peptides Created
543   from Nucleic Acid Sequences as Antigens in Microarrays. *Microarrays (Basel)* **5**,
544   (2016).

545   31.   Achenbach, P. *et al.* Combined testing of antibody titer and affinity improves insulin
546   autoantibody measurement: Diabetes Antibody Standardization Program. *Clin.*
547   *Immunol.* **122**, 85–90 (2007).

548   32.   Szarka, E. *et al.* Affinity Purification and Comparative Biosensor Analysis of Citrulline-
549   Peptide-Specific Antibodies in Rheumatoid Arthritis. *Int. J. Mol. Sci.* **19**, (2018).

550   33.   Real-Fernández, F. *et al.* Surface Plasmon Resonance Method to Evaluate Anti-
551   citrullinated Protein/Peptide Antibody Affinity to Citrullinated Peptides. *Methods Mol.*
552   *Biol.* **1348**, 267–274 (2015).

553   34.   Sakakibara, S. *et al.* Clonal evolution and antigen recognition of anti-nuclear
554   antibodies in acute systemic lupus erythematosus. *Sci. Rep.* **7**, 16428 (2017).

555   35.   Poulsen, T. R., Jensen, A., Haurum, J. S. & Andersen, P. S. Limits for antibody
556   affinity maturation and repertoire diversification in hypervaccinated humans. *J.*
557   *Immunol.* **187**, 4229–4235 (2011).

558   36.   Reddy, S. B. *et al.* Differences in affinity of monoclonal and naturally acquired
559   polyclonal antibodies against Plasmodium falciparum merozoite antigens. *BMC*
560   *Microbiol.* **15**, 133 (2015).

37.   Pratesi, F. *et al.* BNT162b2 mRNA SARS-CoV-2 Vaccine Elicits High Avidity and
Neutralizing Antibodies in Healthcare Workers. *Vaccines (Basel)* **9**, (2021).

38.   Xiong, H., Dolpady, J., Wabl, M., Curotto de Lafaille, M. A. & Lafaille, J. J. Sequential
class switching is required for the generation of high affinity IgE antibodies. *J. Exp.*
Med. **209**, 353–364 (2012).

39.   Mita, H., Yasueda, H. & Akiyama, K. Affinity of IgE antibody to antigen influences
allergen-induced histamine release. *Clin. Exp. Allergy* **30**, 1583–1589 (2000).

40.   Wang, J. *et al.* Correlation of IgE/IgG4 milk epitopes and affinity of milk-specific IgE
antibodies with different phenotypes of clinical milk allergy. *J. Allergy Clin. Immunol.*
125, 695–702, 702.e1 (2010).

41.   Pierson, L., Allauzen, S., Blumenthal, M. & Rosenberg, A. An automated method for
determination of antibody affinity distribution functions with nanogram quantities. *J.*
Immunol. Methods **211**, 97–109 (1998).

42.   Péterfy, F., Kuusela, P. & Mäkelä, O. Affinity requirements for antibody assays
mapped by monoclonal antibodies. *J. Immunol.* **130**, 1809–1813 (1983).

43.   Alessi, P., Fermeglia, M. & Kikic, I. Significance of dilute regions. *Fluid Phase Equilib*
70, 239–250 (1991).

44.   Dimitrov, D. S. Therapeutic antibodies, vaccines and antibodyomes. *MAbs* **2**, 347–
356 (2010).

45.   Prechl, J. Network organization of antibody interactions in sequence and structure
space: the RADARS model. *Antibodies (Basel)* **9**, (2020).

46.   Prechl, J. Why current quantitative serology is not quantitative and how systems
immunology could provide solutions. *Biologia Futura* (2021) doi:10.1007/s42977-020-
00061-1.

47.   Prechl, J. Quantitative network organization of interactions emerging from the

evolution of sequence and structure space of antibodies: the RADARS model. *BioRxiv*

(2018) doi:10.1101/438804.

48.   De Michele, C., De Los Rios, P., Foffi, G. & Piazza, F. Simulation and Theory of

Antibody Binding to Crowded Antigen-Covered Surfaces. *PLoS Comput. Biol.* **12**,

e1004752 (2016).

49.   Edwards, D. A. Steric hindrance effects on surface reactions: applications to BIAcore.

J. Math. Biol. **55**, 517–539 (2007).

Acknowledgments

We thank participants of the GAPAID consortium, especially László Czirják, for providing serum samples for the study and Paolo Rovero for providing VCP2 peptide.

Author contributions

Á.K. analyzed data, developed model, wrote fitting program

J.P. conceptualized of strategy, designed experiments, wrote manuscript, created figures, supervised study

J.R. carried out measurements

K.L. designed experiments

K.P. designed experiments, analyzed data, contributed to figures

O.A. carried out measurement, analyzed data, wrote manuscript

T.P. conceptualized and developed model

Z.H. produced microarrays, carried out measurements

Conflict of interest

K.P., Z.H. and J.P. are employed by Diagnosticum Zrt, a company that provides custom microarray services. A provisional patent application related to this work has been filed by Diagnosticum zrt.

Funding

JP was partially supported by VEKOP-2.1.1-15-2016-00098 grant, part of Széchenyi 2020 project, to Diagnosticum. This work of TP was completed in the ELTE Institutional Excellence Program (TKP2020-IKA-05) financed by the Hungarian Ministry of Human Capacities.

Figure legends

Figure 1. Simultaneous titration of antigen density and serum antibody

Steps of the technology, starting with microarray fabrication and measurement, through image

analysis to curve fitting and visualization of results are shown. Serum dilution is indicated by

red drops, antigen density differences are represented by shades of blue circles. Fitting data

with generalized logistic curves yields two parameters, x_i and d, which characterize chemical

potential distribution of serum antibodies.

Figure 2. Fitting in two dimensions

Examples of parallel measurements of IgG and IgA binding to diagnostic celiac disease peptide

epitope are shown, with serum dilutions in three-fold (A) and two-fold (B) steps. Dot symbols

stand for measurement data, lines are fitted curves.

Figure 3. Comparative characterization of serum antibody binding

Using the values of the parameters obtained by curve fitting, calculated thermodynamic activity

coefficient distributions are comparable not only for serum samples (A) but also for distinct

antibody classes (B). Results shown in Figure 1 were used for the generation of normalized

comparisons. Bar charts (C) show mean d and x_i values, and 95% confidence intervals as

whiskers.

Figure 4. Conventional titration of serum.

Mid-point and end-point titers of IgA and IgG are calculated by logistic fitting of binding to

distinct antigen densities and are displayed as a function of antigen density.

Figure 5. Characterization of a monoclonal Ab for reference.

Binding data of monoclonal Ab 14D15 was fitted using the same algorithm as for serum

antibodies. Binding data and fitted curves (A), calculated distribution of activity coefficient (B)

and estimates of parameters x_i and d (C) are shown.

Symbols and abbreviations

a   relative thermodynamic activity

A   antibody capacity [FI]

Ab   antibody

[Ab]   antibody concentration [mol/L]

AbAg   antibody-antigen complex

[AbAg]   antibody-antigen complex concentration [mol/L]

Ag   antigen

[Ag]   antigen concentration [mol/L]

B   auxiliary parameter

BSA   Bovine Serum Albumin

c   molar concentration [mol/L]

C   auxiliary parameter

d   shape parameter that determines asymmetry

D   positive constant [L/mol]

ELISA   Enzyme-Linked Immunosorbent Assay

γ thermodynamic activity coefficient

IgA   immunoglobulin A

IgG   immunoglobulin G

k   rate of exponential growth

K_D equilibrium dissociation constant

m	auxiliary parameter
μ	chemical potential [J/mol]
μ^o	standard chemical potential [J/mol]
mAb	monoclonal antibody
PBS	Phosphate-Buffered Saline
R	ideal gas constant [~8.314 J/K*mol]
R(x)	Richards function
RFI	Relative Fluorescence Intensity
RIA	Radioimmunoassay
s	slope of lnR(x) at minus infinity
T	temperature [K]
x_i	value of x at inflection point
y_i	value of y at inflection point
z	serum dilution factor [1/dilution]

Figure 1. Simultaneous titration of antigen density and serum antibody

Steps of the technology, starting with microarray fabrication and measurement, through image

analysis to curve fitting and visualization of results are shown. Serum dilution is indicated by

red drops, antigen density differences are represented by shades of blue circles. Fitting data

with generalized logistic curves yields two parameters, x_i and d, which characterize chemical

potential distribution of serum antibodies.

Figure 2. Fitting in two dimensions

Examples of parallel measurements of IgG and IgA binding to diagnostic celiac disease peptide epitope are shown, with serum dilutions in three-fold (A) and two-fold (B) steps. Dot symbols stand for measurement data, lines are fitted curves, numbers are sample identification numbers.

Figure 3. Comparative characterization of serum antibody binding

Using the values of the parameters obtained by curve fitting, calculated thermodynamic activity
coefficient distributions are comparable not only for serum samples (A) but also for distinct
antibody classes (B). Results from data shown in Figure 2 were used for the generation of
normalized comparisons. Bar charts (C) show mean d and x_i values, and 95% confidence
intervals as whiskers.

683
684

686 **Figure 4**. Conventional titration of serum.

687 Mid-point and end-point titers of IgA and IgG are calculated by logistic fitting of binding to

688 distinct antigen densities and are displayed as a function of antigen density.

Figure 5. Characterization of a monoclonal Ab for reference.

Binding data of monoclonal Ab 14D15 was fitted using the same algorithm as for serum antibodies. Binding data and fitted curves (A), calculated distribution of activity coefficient (B) and estimates of parameters x_i and d (C) are shown.

698
699