Towards prevention of re-entrant arrhythmias: Injectable hydrogel electrodes enable direct capture of previously inaccessible cardiac tissues

Authors: Gabriel J. Rodriguez-Rivera,[1] Allison Post,[2] Mathews John,[2] Skylar Buchan,[2] Megan Wancura,[4] Malgorzata Chwatko,[5†] Christina Waldron,[5] Nikhith Kalkunte,[5] Janet Zoldan,[5] Mathieu Arseneault,[5‡] Mehdi Razavi,[2,3*] Elizabeth Cosgriff-Hernandez[5*]

Affiliations:

[1]McKetta Department of Chemical Engineering, The University of Texas at Austin; Austin, TX 787212, USA.

[2]Electrophysiology Clinical Research and Innovations, Texas Heart Institute; Houston, TX 77030, USA.

[3]Division of Cardiology, Department of Medicine, Baylor College of Medicine; Houston, TX 77030, USA.

[4]Department of Chemistry, The University of Texas at Austin; Austin, TX 78712, USA.

[5]Department of Biomedical Engineering, The University of Texas at Austin; Austin, TX 78712, USA.

† Present location: Department of Chemical and Materials Engineering; University of Kentucky, Lexington, KY 40506, USA.

‡Present location: OmegaChem Inc.; Quebec G6W 7V6, Canada.

*Corresponding authors:

Elizabeth Cosgriff-Hernandez: cosgriff.hernandez@utexas.edu

Mehdi Razavi: mehdirazavi1@gmail.com

One Sentence Summary: Injectable hydrogel electrodes achieve pacing that mimics physiologic conduction by capturing midmyocardial tissue.

Abstract: Re-entrant arrhythmias—the leading cause of sudden cardiac death—are caused by diseased and delayed myocardial conduction. Access to the coronary veins that cross the "culprit" scarred regions where re-entry originates provides improved pacing to prevent ventricular arrhythmias and circumvent the need for painful defibrillation, risky cardiac ablation, or toxic and often ineffective antiarrhythmic medications. To date, this goal has not been achieved due to the lack of pacing electrodes which are small or focal enough to navigate these tributaries. We have developed an injectable conductive hydrogel that can fill the epicardial coronary veins and their mid-myocardial tributaries. When connected to a standard pacing lead, these injected hydrogels can be converted into flexible electrodes that directly pace the previously inaccessible mid-myocardial tissue. In our two-component system, hydrogel precursor solutions can be injected through a dual lumen catheter in a minimally invasive deployment strategy to provide direct access to the diseased regions with relative precision and ease. Mixing of the two solutions upon injection into the vein activates redox-initiated crosslinking of the gel for rapid *in situ* cure without an external stimulus. An *ex vivo* porcine model was used to identify the requisite viscosity and cure rate for gel retention and homogeneity. Ionic species added to the hydrogel precursor solutions conferred conductivity above target myocardium values that was retained after implantation.

Successful *in vivo* deployment demonstrated that the hydrogel electrode filled the anterior interventricular vein with extension into the septal (mid-myocardial) venous tributaries, far deeper than current technologies allow. In addition to successful capture and pacing of the porcine heart, analysis of surface ECG tracings revealed a novel pacing paradigm not observed in traditional single-point pacing: capture of extensive swaths of the native conduction system. This is the first report of an injectable electrode used to successfully pace the mid-myocardium and mimic physiologic conduction. As such, this injectable hydrogel electrode can be deployed to any region affected by prior myocardial infarction and consequent scar tissue to provide a reliable pacing modality that most closely resembles native conduction.

NATIVE

HYDROGEL ELECTRODE

METAL ELECTRODE

INTRODUCTION

Ventricular arrhythmias (VA) remain the leading cause of sudden cardiac death, and current pacing modalities are insufficient to prevent or treat lethal arrhythmias (*1*). Post-infarction scarring or otherwise diseased myocardium leads to diseased and slowed conduction, and this spatial and temporal heterogeneity leads to wave break, re-entry, and chaotic cardiac activity and fibrillation (*2-5*). Much like chemotherapy, current approaches to prevent re-entrant circuits rely on antiarrhythmic drugs with high toxicity profiles or destruction of more cardiac tissue in the diseased regions via the "curative" process of ablation. Antiarrhythmic drugs further slow conduction velocity in an effort to prevent sustenance of re-entrant wavefronts (*6*). This, at the same time, can make them pro-arrhythmic due to the role of decreased conduction velocity in sustenance of re-entry (*7*). Although widely adopted, ablative strategies have a high failure rate with recurrent arrhythmias in 18 - 40% of cases (*8*). None of these technologies attempt to correct re-entry's underlying pathophysiology of delayed conduction. For many patients, the only available option is an implantable cardiac defibrillator that uses a high energy shock to extinguish re-entrant circuits. These high-energy shocks far exceed pain thresholds and have a well-documented negative impact on patient quality of life, including post-traumatic stress disorder and depression (*9*). A preventative treatment regimen addressing its underlying pathophysiology is needed to prevent re-entry.

In this work, we developed an injectable hydrogel electrode that can fill coronary veins and tributaries and convert these veins into flexible electrodes that span the heretofore inaccessible mid-myocardium (Figure 1). Transforming these tributaries into flexible electrodes enables simultaneous pacing from multiple sites along the electrode rather than from a single point

stimulus. We hypothesize that the resultant planar wave propagation from the length of the electrode will stimulate wide areas of ventricular tissue that would have been otherwise subject to delayed activation, thus normalizing the delayed conduction which serves as the lynchpin for re-entry. To this end, we developed a unique hydrogel system that has the requisite conductivity, biostability, and rapid *in situ* cure to be delivered via a transvenous catheter. First, a new polyethylene glycol-based macromer with improved hydrolytic stability and durability was designed and synthesized. Redox chemistry was used to provide rapid *in situ* cure without an external stimulus and an *ex vivo* porcine model was used to identify the requisite viscosity and cure rate for gel retention and homogeneity. Ionic species added to the hydrogel precursor solutions conferred conductivity above target myocardium values and the stability of the gel conductivity was assessed *in vitro* and *in vivo*. A porcine model was then used to confirm intravital delivery and successful pacing via the hydrogel electrode. Collectively, this work demonstrates for the first the ability to confer direct electrical stimulation of the midmyocardium using an injectable hydrogel electrode as a new pacing modality.

A Injectable hydrogel electrode **B Connection to pacemaker lead** **C Wavefront propagation pulse**

Figure 1. Schematic of injectable hydrogel system to transform coronary veins into flexible electrodes that capture inaccessible cardiac tissue. (**A**) Hydrogel precursors solutions are delivered using a dual lumen catheter to fill coronary veins and tributaries that span the myocardium near the scarred tissue. Upon mixing of the two hydrogel solutions in the coronary veins, redox-initiated crosslinking results in rapid cure of the ionic hydrogel. (**B**) Subsequent connection to a pacemaker lead to the ionic hydrogel electrode provides increased tissue contact across the myocardium. (**C**) Wavefront activation of the myocardium along the hydrogel electrode reduces the energy required for defibrillation.

RESULTS

Injectable hydrogels with rapid *in situ* cure

There are no current hydrogels that can provide the multifaceted needs of a cardiac electrode with endovascular delivery to the midmyocardium: injectability, conductivity, and biostability. Poly(ethylene glycol) (PEG) hydrogels are a popular choice based on their established biocompatibility and highly tunable, soft tissue-like properties; however, traditional acrylate-derivatized PEG hydrogels are susceptible to slow degradation *in vivo* due to hydrolysis. To generate a PEG-based hydrogel suitable for long-term implantable applications, we synthesized a new hydrogel chemistry that combines biostability with flexibility and durability. Poly(ether urethane diacrylamide) (PEUDAm) contains urethane and amide groups that are resistant to hydrolysis at physiological conditions. The detailed synthetic route is described in **Figure S1** and successful conversion (>95%) was confirmed with H^1 NMR (**Figure S2-S4**). The resulting PEUDAm macromer displayed a number average molecular weight of 19,960 ± 40 g/mol and a narrow polydispersity index of 1.08 ± 0.01. A small degree of polymer coupling was indicated by

a secondary peak at Mn = 49,340 ± 220 g/mol (**Figure S5**). In addition to the biostable PEG macromer, we synthesized N-acryloyl glycinamide (NAGA), a small molecule crosslinker with bidentate hydrogen bonding (*10*). Successful acrylation and purity of NAGA were confirmed by H[1] NMR analysis (**Figure S6**). The resulting hydrogel (20kDa, 20wt% PEUDAm+ 1% NAGA) properties are summarized in **Table 1**. The hydrogel stiffness, 59.1 ± 7.5 kPa, was within the range of myocardial tissue, reported as 20 to 200 kPa by Ramadan et al. for ventricles (*11*) or 20 to 500 kPa reported from mathematical models. The hydrogel ultimate elongation of 111 ± 28% was above the average myocardium strains reported from imaging analysis ranging from 10-55% (*12*) and above the axial strain of cardiac veins, which is up to 30% strain for the larger veins (*13*). In addition to tissue-matching mechanical properties, accelerated hydrolytic degradation studies indicated that the new hydrogel was biostable, **Figure S7-S8**.

The proposed design also requires rapid *in situ* cure of the hydrogel without external stimuli (e.g UV). Our lab has previously shown tunable gelation using the redox pair APS and IG (*14*). In this study, we demonstrated faster gelation kinetics than previously described by increasing APS concentration at constant APS: IG of 1:2 molar ratio. A double-barrel syringe with mixing head was used to deliver the precursor solution as an analog for future delivery from a dual lumen catheter, **Figure 2A**. The effect of compositional variables on network formation was investigated by monitoring changes in storage and loss modulus with rheology. The crossover point was achieved too quickly for data quantification for most samples, **Figure S9**. Complete network formation was determined as the third point, after which there was a less than 1.0% change in modulus. Iterative testing was used to identify a target cure rate of < 2 minutes which corresponded to an APS concentration of 1.0 mM. The inclusion of 1 w/w% NAGA increased the cure rate and gel fraction (**Figure S10, Table S1**). We hypothesized that this was due to the increase in

acrylamide groups and mobility of the smaller molecules. In the presence of blood, full network formation was delayed from 1.7 to 5 minutes (**Figure 2B**).

Figure 2. In situ cure kinetics and cytocompatibility of redox hydrogel. (**A**) Redox initiation reaction of new PEUDAm macromer + NAGA delivered using double barrel syringe with a mixing head. Resulting hydrogels display bidentate hydrogen bonding at netpoints for improved durability. (**B**) Effect of redox concentration (APS:IG ratio 1:2) and blood on hydrogel cure rate (time to complete network formation). (**C**) Representative images of HUVECs and hiPSC-cardiomyocytes exposed to ionic hydrogels shortly after cure. Scale bar is 100 mm. Individual data points, means, and standard deviation are presented (n = 3). Significant differences between groups are marked as follows: * (p < 0.05).

Table 1. Physical properties of the hydrogel electrode: PEUDAm 20kDa, 20% + 1% NAGA, 0.75 mM APS/1.5 mM IG

Physical Property	Average value
Gel fraction	$95.2 \pm 1.4\%$
Mass swelling ratio	17.4 ± 0.1 g/g
Young's modulus	59.1 ± 7.5 kPa
Ultimate tensile strength	50.6 ± 13.3 kPa
Ultimate elongation	$111 \pm 28\%$

As a final determination of the suitability of this new system as an injectable hydrogel, the cytocompatibility of the *in situ* cure mechanism was examined prior to *in vivo* deployment. In the target application, the hydrogels will be cured in direct contact with the endothelial cells on the cardiac veins with potential diffusion to the myocardium. To assess *in vitro* cytocompatibility, the potential cytotoxic effects of hydrogel leachables were evaluated on human umbilical vein endothelial cells (HUVECs) and human-induced pluripotent stem cells – cardiomyocytes (hiPSC-MCs). **Figure 2C** shows that cell viability 24 hours after indirect exposure to a freshly prepared hydrogels (20% 20kDa PEUDAm, 1 w/w% NAGA, 1mM APS/2mM IG) was tested on two different cell lines. hydrogel was above 90% and comparable to positive controls. No adverse effect from indirect contact to the hydrogels immediately after gelation was observed indicating overall cytocompatibility of the redox-initiated hydrogel.

Precursor solution viscosity and cure rate determine hydrogel retention in an *ex vivo* model

An *ex vivo* porcine heart model was used to test hydrogel retention and homogeneity after injection. First, an *in vitro* method developed to visualize polymer mixing during delivery and gelation (SI **M1-M2**) indicated that poor mixing strongly correlated to low gel fractions. This was attributed to insufficient initiation due to a lack of radical formation that is dependent on the reducing agent and initiator reaction. A mixing head at the end of the double-barrel syringe ensured adequate mixing upon delivery and high gel fractions. Once suitable mixing was confirmed, the effect of precursor solution viscosity and cure rate on hydrogel retention in the cardiac vein was studied. Hydrogel compositions with lower viscosity precursor solutions (~42 mPa-s) and slower cure rates (~5.8 min) did not form contiguous gels. Increasing the viscosity (~83 mPa-s) by increasing the precursor macromer concentration and increasing the cure rate (~2.2 min) with higher initiator concentrations resulted in contiguous gel formation, **Figure 3A**. The homogeneity across the hydrogel length after cure was studied *in vitro* in tubing and after dissection from the vein in the *ex vivo* studies. We also evaluated the potential impact of mixing with blood *in vitro* by pre-filling the tubing with porcine blood. Hydrogel swelling ratios were consistently higher in the *ex vivo* system or when combined with blood (**Figure 3B**). Differences in gel fractions between the *in vitro* gel cured in the empty tube and ex vivo studies were not significant. Swelling ratios were not significantly different from the proximal to the distal end indicating that the injectable gel was homogenous across its length.

A Impact of viscosity and cure rate on gel retention in cardiac veins

η = 41 mPa-s
t_{gel} = 5.8 min

η = 83 mPa-s
t_{gel} = 2.2 min

B Hydrogel homogeneity assessment

In vitro

Ex vivo

Figure 3. Hydrogel retention and homogeneity assessment in an *ex vivo* model. (A) Delivery of the hydrogel to main cardiac vein using an ex vivo porcine model. Poor gel retention was observed with lower viscosity precursor solution and slower cure rates. **(B)** Hydrogel segmental homogeneity assessment (mass swelling fraction, gel fraction) after cure in tubing, tubing filled with blood, and ex vivo porcine model (n = 3).

Flow simulations and rheological testing were used to further investigate the mechanistic role of these variables on hydrogel retention. When using the mixing head, there was minimal effect of precursor solution viscosity on the mixing of the two solutions upon delivery, **Figure 4**. Given that mixing of the two components is directly related to the rate of initiator decomposition and crosslinking, this flow simulation indicated that the mixing head negated possible effects on the cure rate. Rheology indicated that the complex viscosity of the solution after delivery was initially

dependent on the polymer concentration and then increased over time as radical crosslinking was initiated until the gel point. Increasing the initiator concentration increased the rate of crosslinking with a corollary effect on the rate of viscosity build up. We hypothesized that poor hydrogel retention was due to venous draining of the precursor solution through the tributaries. To achieve sufficient hydrogel retention to form a contiguous hydrogel electrode in the vein, the system should be designed with sufficient precursor solution viscosity and a rapid cure to slow the dispersal prior to gelation, **Figure 4**. Gelation kinetics that were too rapid were found to clog the dispensing catheter. The hydrogel precursor formulation that balanced these parameters was 20 w% 20kDa PEUDAm with 1 wt% NAGA in saline solution at 0.75 mM APS and 1.5 mM IG. This composition resulted in sufficient hydrogel retention and homogeneity in the *ex vivo* model to support its evaluation in subsequent *in vivo* studies.

Figure 4. Role of mixing and *in situ* cure on hydrogel retention. Fluid simulations show adequate mixing and negligible differences in the volume fraction of the precursor solutions at both high and low viscosities. Time sweep indicates that high-viscosity precursor solutions with fast cure (higher initiator concentration) display a rapid increase in complex viscosity that was correlated with increased gel retention.

Ionic hydrogels display stable conductivity higher than native myocardium

The ionic PEUDAm hydrogels rely on salts dissolved in the precursor solution to serve as charge carriers to generate a current when connected to a power source. As a basic proof-of-principle, **Figure 5** demonstrates that a current could be generated by connecting the ionic hydrogel to a

battery and a light-emitting diode. Initially phosphate buffered saline was used to provide the ionic species (**Figure S11**); however, it was found that the phosphate led to iron precipitation which negatively impacted the redox-mediated initiation mechanism. To address this limitation, saline was used as an alternative source of ions in the hydrogel precursor solutions, **Figure S12**. In order to activate the myocardium beyond a single point of contact *in vivo*, the conductivity of the hydrogel must be above the myocardial conductivity that ranges from 0.1 to 6.0 mS/cm.(*15-17*) Initial conductivity performed with a four-point probe indicated that the ionic hydrogel more than doubled these target values (12.8 ± 1.5 mS/cm, **Figure S11**).

Figure 5. Conductivity of ionic hydrogel. (A) Demonstration of hydrogel conductivity after application of direct current. (**B**) Bode plot from EIS of hydrogel after gelation and equilibration in DI water (water gel) and PBS (ionic gel). (**C**) Effect of subcutaneous implantation on conductivity as calculated from EIS data, area highlighted in red represent the ranges of myocardium conductivity. Individual data points, mean, and standard deviation are presented (n = 4). Significant differences between groups are marked with * ($p < 0.05$). (**D**) Current generated at different hydrogel lengths after 5V pulses. Pulses were 5 ms width at a frequency of 2 Hz (equivalent to 120 beat per minute). As hydrogel length increased, the increase in impedance caused a decrease in the current generated. (**E**) Hydrogel voltage response to stimuli remains stable after multiple pulses.

Electrochemical impedance spectroscopy was used to assess the ionic gel resistance and capacitance response to an applied voltage, **Figure 5B**. The slope on the Bode plots show the effects of ions transport and double-layer capacitance resulting in an increase in impedance at lower frequencies. Significantly, the ion concentration of the ionic gel resulted in reduced overall impedance of this gel relative to the water gel. The conductivity as measured by the impedance in the plateau regions was 18.3 ± 0.5 mS/cm for the ionic gel as compared to <0.1 mS/cm for the water gel control. Given that the conductivity was dependent on ion concentration in these gels, we hypothesized that there would be rapid ion diffusion in the hydrogels after implantation with corollary effects on gel conductivity. To determine the effect of ion flux on gel conductivity, both water and ionic gels were implanted subcutaneously into rats and the effect on conductivity measured after 7 days. Both gels reached a conductivity value of approximately 16 mS/cm (**Figure 5C**), which supports that the ion equilibration higher than myocardial conductivity. The swelling

ratio and modulus of the gels were not significantly different before and after implantation (**Figure S13**), demonstrating network properties did not change significantly during that period.

In the injectable hydrogel system, the pacing lead is connected to the proximal end of the hydrogel extending into the vein reaching the midmyocardium away from the power source. To assess how the current varies with hydrogel length, hydrogels were subjected to monophasic pulses of 5 milliseconds at a frequency of 2Hz, correlating to a pacing of 120 beats per minute (**Figure 5D**). A 5V power supply was used to function as a circuit that was comparable to standard pacemakers (*18*) (**Figure S14**). As expected, due to increased hydrogel impedance with length, the current decreased from 3.4 to 0.7 mA as the length increases from 1 to 5 cm (**Figure 5D**). These studies indicate that the hydrogel electrode was able to provide stable electrical stimuli over many cycles and across a substantial length of the cardiac vein (**Figure 5E**).

Successful deployment and safety of injectable hydrogel electrode in a porcine model

Although the *ex vivo* model indicated strong promise of the injectable hydrogel electrode, it does not fully replicate clinical deployment in a beating heart with vessels filled with blood. To this end, precursor solutions were injected into the anterior intraventricular vein using a porcine model and the *in vivo* gel formation and retention in the vein and tributaries characterized (**Figure 6A**). Explanted hydrogels displayed excellent segmental uniformity with minimal differences in equilibrium swelling ratio and gel fraction across the hydrogel length (proximal to distal). It should be noted that the calculated gel fraction does not consider any potential leaching from the gel between the time of injection until explant. The equilibrium swelling ratio provides a measure of the hydrogel network structure and indicated that the gel formation was uniform across the length of the vein (**Figure 6B**).

Once the hydrogel is injected into the vein and connected to a pacing device, it will remain in that location during the patient's lifetime. We conducted a preliminary safety assessment to confirm that the occlusion of the vein does not result in acute adverse effects on the animal and to evaluate the host response to the gel. In future clinical settings the hydrogel will be delivered via a catheter. In this initial study, however, an epicardial incision was made in order to access the vein and deliver the hydrogel. In this specific case, the hydrogel was injected into the middle cardiac vein and retained for 2 weeks. Tissue slices of location proximal to the injection site, middle of the vein, and distal to the injection site were analyzed to assess the cardiac injury and inflammatory responses after two weeks (**Figure 6C**). As expected in an open chest procedure with direction instrumentation, fibrosing epicarditis was observed across all locations, with moderate to mild perivascular and interstitial fibrosis closer to the incision/injection sites. Foreign body giant cells were also limited to these sites. Troponin levels were monitored on the day of injection and 2 weeks after injection. Troponin levels were high (3313 ng/L) directly after the procedure involving direct instrumentation but returned to baseline (34 ng/L) at the end of two weeks, indicating that the hydrogel did not induce ischemia or myocardial necrosis.

Figure 6. *In vivo* **assessment of injectable electrode in coronary vein of a porcine model**. **(A)** Postmortem confirmation of *in situ* cure of hydrogel in AIV (H1) and tributaries (H2). **(B)** Hydrogel homogeneity assessment of gel fraction and swelling ratio. Means and standard deviation are presented (n = 3). **(C)** Host response after 2 weeks implantation using slices of the location

proximal to the injection site, middle of the vein, and distal to the injection site (green arrow). The white arrow indicates the approximate location of the hydrogel in the vein. The response includes damage induced at the hydrogel injection site, with 1) moderate perivascular and interstitial fibrosis with focal replacement fibrosis (blue arrows), and 2) fibrosing epicarditis with foreign body giant cell reaction (red arrows). Mid branch of the hydrogel injection show 3) mild perivascular and interstitial fibrosis, 4) mild focal replacement fibrosis, and fibrosing epicarditis. Distal branch indicated 5) preserved myocardium with 6) only fibrosing epicarditis with focal extension into myocardium.

Hydrogel electrode mimics physiologic conduction by capturing midmyocardial tissue

A key aspect of the development of the injectable hydrogel electrode is that we should be able to pace the heart from this electrode on the cardiac veins. At a constant pulse width of 5 ms, unipolar pacing thresholds were measured to be 2.2 ± 1.2 mA for the metal electrode, 2.3 ± 0.6 mA, 1.5 ± 1.28 mA for an epicardial hydrogel point source (0.8 cm disk), 1.5 ± 1.3 mA for the epicardial hydrogel line source and 2.7 ± 1.8 mA with the hydrogel in the AIV. There was no significant difference in the capture thresholds of the different electrode configurations at 5 ms and 10 ms pulse widths. Pacing thresholds measured at other pulse widths are detailed in (**Figure S15**). Two-way ANOVA revealed significant differences only for capture thresholds with pulse widths below or equal to 1 ms. As seen in **Figure 7**, epicardial pacing using a metal electrode, hydrogel point, or hydrogel line, created an inverted QRS morphology indicative of a deviation from normal conduction. Remarkably, the QRS morphology by pacing through the hydrogel in the AIV was comparable to the QRS at sinus rhythm (**Figure 7B**). No inversion of the QRS morphology and the short isoelectric window before initial activation indicate the hydrogel electrode captures the

deep septal bundle branches and Purkinje fibers. Supplementary figures (**Figure S16 – S30**) show signs of mid septal capture as evidenced by negative deflection in the anterior leads similar to baseline ECG.

Figure 7. Schematic of myocardial activation with different stimulation sources. (A) Native baseline, epicardial metal electrode, and hydrogel in myocardium activation pathways. (B) Impacts on tracing are shown in the left. (B) Tracing of different leads of surface ECG showing electrical activation at baseline (sinus rhythm) and when paced (metal electrode and hydrogel in AIV). Morphologies of paced impulses in the epicardial and hydrogel electrodes were compared. Baseline activation during sinus rhythm starts from the sinoatrial node and travels through the atrioventricular node into the Purkinje fibers in the ventricles. Pacing the epicardium causes the QRS morphology to invert indicative of deviation from normal conduction. Point pacing with metal electrode creates a dog-bone shaped activation that radiates outward from the source of the

electrical stimulus. During pacing from the hydrogel in the AIV, clear capture is seen with a QRS morphology existing after every pacing spike. Peak of the baseline QRS morphology was matched with the peak of QRS morphology obtained from pacing with hydrogel in AIV. No observed inversion of the QRS morphology and a short isoelectric window before initial activation is suggestive of capture of the deep septal bundle branches and Purkinje fibers.

DISCUSSION

Re-entry is the leading cause of lethal arrhythmias (*1*). Its most common mechanism is delayed cardiac conduction in scarred tissue, predominantly observed after coronary artery occlusion during a myocardial infarction (*19*). The ability to directly pace and capture these regions would enable targeted correction of the local activation delay; however, development and testing of such an approach has been limited by the inability to directly access and pace these regions. Given that the venous structures drain the same region where the arterial occlusion caused ischemic damage, these vessels and their tributaries provide a direct approach to the scarred tissue. However, there are no pacing leads small enough to navigate these smaller tributaries and epicardial approaches have limited efficacy due to the greater mass and thickness of ventricular tissue. In this work, we developed an injectable hydrogel electrode that can fill coronary veins to convert them into flexible electrodes that directly pace the inaccessible mid-myocardium. Hydrogels are being explored as potential flexible electrodes due to their tunable properties and ease of incorporating conductive elements (*20*). An important criterion for translation was identifying a system that was biostable, biocompatible, and can be used with current endovascular approaches. We designed a new hydrogel chemistry that could meet the multifaceted needs of this application. The PEUDAm hydrogels displayed stiffness similar to the myocardium, necessary to avoid adverse tissue

responses(*21*), and increased biostability as compared to traditional PEG diacrylate. Uniaxial tensile testing indicated that the PEUDAm gels could withstand the strains imposed during a cardiac cycle; however, additional studies are needed to evaluate the long-term fatigue resistance. These hydrogel electrodes also needed adequate conductivity to ensure stimulation of the myocardium along the length of the hydrogel. Many of the conductive polymers are not suitable as an injectable electrode due to insolubility, toxicity, or processing limitations. Recent efforts to create injectable conductive hydrogels have utilized side-chain functionalization, grafting approaches, or dispersal strategies to increase solubility or maintain mild cure conditions; however, many of the resulting gels only attained conductivities at the lower end of reported myocardium values (*22*). Metallic nanoparticles are an alternative option but require high concentrations to achieve target conductivity with corollary effects on solution viscosity that can impede mixing and delivery (*23*). Our ionic hydrogel utilizes salts added to the precursor solutions to confer conductivity without detriment to other hydrogel properties, solution viscosity, or cure rate. The resulting ionic hydrogel electrode achieved conductivity more than double target myocardium values that was stable over numerous cycles and retained after implantation.

For this application, an injectable system was needed that could be delivered via catheter in a minimally invasive deployment strategy and cure *in situ* without external stimuli. Multiple strategies have been used to achieve *in situ* cure, ranging from photopolymerization, ionic interactions, and temperature or pH-induced transitions (*22, 24, 25*). Our two-component system relies on delivering hydrogel precursor solutions through a dual lumen catheter to provide direct access to the diseased regions with relative precision and ease. Mixing of the two solutions upon injection into the vein activates redox-initiated crosslinking of the gel for rapid *in situ* cure. In

contrast to other systems that rely on reversible bonds or physical crosslinking, our approach relies on inducing chemical crosslinking between the polymer chains via *in situ* radical polymerization to ensure long-term stability of the hydrogel. Ammonium persulfate (APS) and TEMED are one of the more common pairs used for in-situ polymerization of hydrogels (*26-31*), including cardiac applications (*26, 32*). Our lab has shown that iron gluconate (IG), a common supplement, can be used as an alternative reducing agent with a lower toxicity profile (*14*). This system is highly tunable and components can be adjusted to achieve a broad range of hydrogel properties and cure rates without changing the chemistry of the macromers. An *ex vivo* porcine model was used to identify the requisite properties for gel retention and homogeneity. A critical finding from these studies was that the initial precursor solution viscosity in combination with the hydrogel cure rate strongly influenced gel retention in the vein. A mixing head was necessary to ensure adequate mixing of the two components for hydrogel homogeneity but modeling of the mixing behavior displayed minimal effects of precursor solution viscosity. To apply these findings to endovascular delivery, we are using computationally-directed design of a mixing head at the distal end of the dual lumen catheter to ensure adequate mixing for uniform gel formation and facile release of the hydrogel at the distal end upon catheter retrieval.

Regarding clinical safety, there may be some concern regarding cardiac engorgement consequent to the occlusion of the epicardial vessels with the hydrogel electrode. However, extensive clinical experience strongly argues in favor of the safety of this practice—hundreds of thousands of occlusive leads are placed in these same vessels each time a CRT device is implanted—with no reports of complications. Longstanding literature also supports the safety of total occlusion of the coronary sinus (into which all venous tributaries drain) - a vessel much larger than the branch

vessels we will be targeting.(*33-35*) Pilot large animal data shows that occluding the epicardial venous tributaries was associated with minimal, if any, myocardial necrosis or inflammation. Although histopathologic studies demonstrated myocardial inflammation, it was confined to the regions directly instrumented in this open chest model. Clinically, such a traumatic delivery system is unlikely to be utilized. Our delivery system is currently being modified for transcatheter delivery using a dual lumen catheter. Gaining access to these venous branches using a closed-chest transcatheter approach is a standard skillset utilized by clinical electrophysiologists when placing a cardiac resynchronization device (a pacemaker with a lead in the ventricular branches of the epicardial venous system).

With this new injectable hydrogel electrode, we are able to explore the possibility of *prevention* of re-entry by directly capturing zones of delayed conduction, thus eliminating one of the necessary conditions for re-entry to arise. Unlike past efforts that used conductive hydrogels or other materials as scaffolds to regenerate myocardial tissue and restore conductivity,(*36-38*) these hydrogel electrodes will interface with current cardiac resynchronization devices to provide an immediate intervention to prevent VA. By injecting into the corresponding venous branches with the conductive hydrogels, electrophysiologists can target specific locations corresponding to each patient's scar location(s). The advantage of direct midmyocardial tissue capture is not limited to prevention of re-entry. Recent progress in atrial defibrillation technology has demonstrated that using multiple electrodes to deliver multi-stage electrotherapy can reduce the energy required defibrillation (*39*). Delivering multiple low-energy pulses over a larger area has shown benefits for terminating fibrillation with energies below pain thresholds (*40-42*). Moreno et al. performed an *ex vivo* study with epicardial line electrodes that generated a planar activation wavefront that

depolarized a large ventricular area at energy levels below the human pain threshold (*43*). These reports indicate that planar wavefront propagation from the hydrogel electrode as it spans the mid myocardium also provides the opportunity for painless ventricular defibrillation. Other multi-site pacing approaches to prevent and treat VA rely on epicardial placement of flexible electrodes. The lack of a minimally invasive deployment strategy of these hydrogels as flexible electrodes represents a large translational challenge for this technology. Furthermore, epicardial stimuli are less effective for defibrillation due to the greater mass and thickness of ventricular tissue. In contrast, this endovascular approach provides unprecedented simultaneous multi-site pacing in the mid-myocardium while also integrating with current clinical practices.

Cardiac pacing studies using these injectable hydrogel electrodes in a large animal model provided proof-of-concept for this new approach. Although we did not assess the effect of such pacing in diseased myocardium, our work demonstrates for the first time the ability to directly pace these erstwhile unapproachable regions of the heart. The evidence to support this is the nearly identical nature of surface electrocardiographic morphologies of the paced and native heartbeats. A similar effect normalization of QRS vectors has been observed in clinical reports of deep septal pacing (*44-46*). Furthermore, there is a marked latency between the pacing stimulus and initiation of the surface ECG. This is because the pacing output at least in part captures and conducts via the native conduction system before exiting into myocardial tissue. Thus, there is a delay between the pacing stimulus and onset of the ECG signal. These features were reproducibly observed and are strongly suggestive of direct mid-myocardial tissue capture, something that has been technologically vexing and, to the best of our knowledge, unreported. This first evidence that these electrodes can mimic physiologic conduction could alter the landscape of cardiac rhythm management.

Collectively, these injectable hydrogel electrodes show strong promise as a new cardiac resynchronization therapy to provide direct stimulation of wide areas of the heart.

MATERIALS AND METHODS

Materials. All materials were purchased from Sigma Aldrich unless otherwise indicated.

Synthesis of polyether urethane diacrylamide (PEUDAm). A new PEUDAm macromer was synthesized in a three-step process with a PEG-diamine intermediate. *Synthesis of PEG-CDI:* Carbodiimidizole (CDI, Alfa Aeser, 15 equiv.) was weighed in a dry, 500 mL round bottom flask and dissolved in anhydrous DCM (20 mL). PEG (Mn = 17600, 1 equiv.) was dissolved in anhydrous DCM (10 wt%) and placed under nitrogen atmosphere. The PEG solution was added dropwise to CDI over 2 hours at 30-35°C, then left to react at room temperature overnight, stirring vigorously. The reaction was quenched with the addition of deionized water and the solution was then washed with deionized water three times in a separatory flask, dried with sodium sulfate, and collected via vacuum filtration. Solvent was removed under vacuum with additional drying of the product at 65°C under vacuum for 1 hour to remove residual solvent and water prior to use. *Synthesis of PEG-Diamine:* Ethylenediamine (15 equiv.) was added to a dry 500 mL round bottom flask and a PEG-CDI solution in anhydrous DCM (12 wt%, 1 equiv.) was added dropwise at RT over 3 hours (1 drop per 1-3 seconds) under nitrogen. The reaction was allowed to proceed overnight under vigorous mixing. The solution was then washed with deionized water three times in a separatory flask, dried with sodium sulfate, and collected via vacuum filtration. The product was precipitated in 10x volume of ice-cold ether and collected via suction filtration. The product

was then dried under vacuum at 80°C for 3 hrs to remove residual solvent and water. *Synthesis of PEUDAm*: PEG-diamine (1 equiv.) was weighed in a dried 500 mL 3-neck round bottom flask and dissolved in anhydrous DCM at 15 wt% under nitrogen atmosphere. Triethylamine (2 equiv.) was added dropwise. Acryloyl chloride was diluted in DCM (~4 mL) and added dropwise (1 drop per 4-5 s). The reaction was allowed to proceed for 48 hours at room temperature, then quenched with potassium carbonate (2 M, 8 equiv.). The solution was washed twice with water, dried with sodium sulfate, and collected via suction filtration. The product was precipitated in 10x volume ice-cold diethyl ether and collected via suction filtration. The final product was dried at atmospheric pressure overnight, then briefly under vacuum.

Macromer characterization. Functionalization of PEG at each step was determined by analyzing proton nuclear magnetic resonance (H^1 NMR) spectra collected on a 300 MHz Bruker NMR. H^1 NMR (CDCl$_3$): *PEG-CDI*: δ = 8.12 (s, 2H, -N-C*H*-N-), 7.40 (s, 2H, -N-C*H*=CH), 7.02 (s, 2H, -CH=C*H*-N), 4.51 (m, 4H, -O-C*H*$_2$-CH$_2$-), 3.70 (m, 1720H, -O-C*H*$_2$-CH$_2$-). *PEG-EDA*: δ = 4.18 (t, 4H, -O-C*H*$_2$-CH$_2$-), 3.60 (m, 1720H, -O-C*H*$_2$-CH$_2$-), 3.19 (m, 4H, -CH$_2$-C*H*$_2$-NH), 2.78 (m, 4H, -NH$_2$-C*H*$_2$-CH$_2$-), 3.70 (m, 1720H, -O-C*H*$_2$-CH$_2$-). *PEUDAm*: δ = 6.94 (broad s, 2H, -C-N*H*-CH$_2$-), 6.28 (dd, 2H, *H*$_2$C=CH-C-), 6.17 (m, 2H, H$_2$C=C*H*-C-), 5.85 (broad s, 2H, -C-N*H*-CH$_2$-), 5.61 (dd, 2H, *H*$_2$C=CH-C-), 4.20 (m, 4H, -H$_2$C-C*H*$_2$-O-), 3.65 (m, 1720H, -O-C*H*$_2$-CH$_2$-), 3.34 (m, 4H, -CH$_2$-C*H*$_2$-NH-). Macromer molecular weight was determined by gel permeation chromatography (GPC) using an Agilent Technologies 1260 Infinity II GPC system. Macromer specimens were dissolved in 0.1 M lithium bromide in dimethylformamide at a concentration of 5 mg/mL at room temperature and syringe filtered. Injections (100 μL) were passed through 2 Phenogel columns with molecular weight ranges of 1–75 kDa at a flow rate of 1.0 mL/min and temperature of 25 °C.

The separated components were passed through an Agilent Multi Detector system (RI, RALS, LALS, UV) at 30 °C, and the molecular weight distribution was determined using the Agilent GPC/SEC software. Molecular weight was reported relative to polyethylene glycol standards. Viscosity was measured in triplicate using a Brookfield Viscometer equipped with a CP-40Z flat plate.

Synthesis of *N*-acryloyl glycinamide. *N*-acryloyl glycinamide was synthesized according to established protocols with minor variations (*10*). Glycinamide hydrochloride (Alfa Aeser, 6.3 g, 1 equiv.) was added to a 250 mL 3-neck RBF, dissolved in 18 mL diethyl ether and 40 mL of 2 M potassium carbonate (1.18 equiv.), and cooled over ice. Acryloyl chloride (0.98 equiv) was diluted in 24 mL cold diethyl ether and added dropwise to the reaction mixture over 30 minutes with vigorous stirring. The reaction was allowed to proceed for 4 hours at room temperature while protected from light. The solution was reduced to pH = 2 with 6 M HCl and then washed with diethyl ether (50 mL) three times. The aqueous solution was then placed under vacuum to remove residual ether, adjusted to neutral pH with 2 M NaOH, then frozen and lyophilized. Warm acetone (200 mL, 40°C) was added to the powdered product and allowed to stir for 8 min to extract the product. This process was repeated four times and approximately half of the solvent was removed via rotary evaporation. The dissolved product was crystallized on an ice salt bath. The crystals were collected and washed then dissolved in minimal warm methanol. Undissolved particles were removed, and the product was recrystallized on ice salt. The product was then dissolved in deionized water, frozen, and lyophilized before analysis with NMR. H^1 NMR (300 MHz, D_2O): δ = 6.31 (dd, 1H, -HC=CH_2), 6.23(dd, 1H, -HC=CH_2), 5.80 (dd, 1H, -HC=CH$_2$), 3.97 (s, 2H, HN-CH_2-CONH$_2$)

Hydrogel fabrication and gelation kinetics. Hydrogels were fabricated with the PEUDAm macromer (20 wt%), NAGA monomer (1 wt%) and a combination of redox initiators ammonium persulfate (APS, 0.75 mM) and iron gluconate (IG, 1.5 mM) in saline solution (1.35% NaCl). Hydrogel solutions were loaded into disposable double-barrel syringes with half the solution containing APS and the other half containing IG. The cure profiles of the injectable hydrogel was characterized by determining gelation onset and complete network formation using a Discovery Hybrid HR-2 rheometer. Hydrogel precursor solutions were prepared at an initiator concentration of 0.5, 0.75, or 1.0 mM with an APS:IG molar ratio of 1.0:2.0, loaded into double-barrel syringes, and injected through a mixing head onto a parallel-plate (20 mm) configuration heated to 37 °C. Storage and loss moduli were measured every 6 seconds with a 1.3 mm gap and 0.5% strain. To determine the effect of blood on cure rate, 50 μL of porcine blood was added at the center of the plate, and the hydrogel was injected directly into the blood. The onset of gelation was measured as the crossover point of loss and storage modulus. Complete network formation was measured as the third point after which there was a less than 1% change in storage modulus. Values were reported as the average of three specimens. To determine tensile properties, dog bone specimens (n = 4) with a gauge width of 5 mm width were cut from swollen hydrogel slabs (thickness ~1.5mm) using a modified 3D printed punch developed by Nelson *et al.* (*47*). Specimens were strained to failure using a dynamic mechanical analyzer (RSAIII, TA Instruments) at a rate of 0.1 mm/s. All mechanical tests were performed at room temperature.

Characterization of network formation. Macromer solutions were prepared as described above, injected through a mixing head into glass spacers (0.75mm), and allowed to cure for at least 30

minutes at 37°C prior to characterization. Network formation of the injectable hydrogel was determined by measuring gel fraction and swelling ratio. To measure the gel fraction, four specimens (10mm diameter, 0.75 mm height) were dried under vacuum for 24 hours and weighed. The salt content of the precursors was considered in the calculation of initial polymer mass (W_0). Specimens were swelled for 24 hours in 2 mL of DI water and weighed. The specimens were dried again for 24 hours to assess the remaining dry polymer mass (W_f). The ratio of the final to initial mass (W_f /W_0) was reported as the gel fraction. To measure equilibrium swelling ratio, four specimens (10mm diameter, 0.75 mm height) were swelled in DI water for 24 hours and weighed (W_s). The specimens were then dried under vacuum for 24 hours and weighed (W_D). The equilibrium mass swelling ratio was calculated as $Q=W_s/W_D$. To determine the homogeneity of gels after injection, hydrogel solutions were prepared and injected into a 3/16" ID PVC Tygon tubing (McMaster-Carr) using the double-barrel syringe with mixing head. The gels were allowed to cure for at least 10 minutes at 37°C. The gel was cut into five segments for characterization of gel fraction and equilibrium swelling as described above.

Accelerated degradation testing: To confirm the hydrolytic biostability of the new hydrogel composition, accelerated degradation testing of the PEUDAm hydrogels was performed in comparison to a polyethylene glycol diacrylate (PEGDA, 20 kDa, 20wt%) control. Hydrogel specimens (10 mm, n = 4) were placed in PBS (3 mL) in glass vials and incubated at elevated temperatures (70°C) or physiological temperature (37°C) with weekly solution changes. Increases in equilibrium swelling and reductions in compressive storage modulus were used as a measure of hydrolytic degradation. The hydrogels were weighed weekly for 8 weeks and the change in equilibrium swelling was calculated as: equilibrium swelling mass at each time point divided by

initial equilibrium swelling mass. For modulus measurements, specimens were tested monthly for 3 months using a dynamic mechanical analyzer (RSAIII, TA Instruments) equipped with a parallel-plate compression clamp. Testing was performed under unconstrained compression at room temperature. Compressive modulus was determined from a frequency sweep at 1 % strain. The compressive storage modulus was taken at 0.25 Hz in the linear region. Four specimens were tested per gel.

Cytocompatibility. An indirect assay was used to determine the cytocompatibility of hydrogel leachables immediately following cure. Human umbilical vein endothelial cells (HUVECs) were expanded in Endothelial Cell Growth Medium-2 (Lonza), harvested for use between passage P4-P6, and seeded at 30,000 cells per well in 24 well plates. Human induced pluripotent stem cells (hiPSCs, WTC 11), a gift from Dr. Bruce Conklin, were cultured via feeder-free conditions on vitronectin coated well plates in complete Essential 8 (E8) medium (ThermoFisher). Cardiomyocyte (CM) differentiation was conducted via a modified protocol based on the established Wnt GiWi method (*48, 49*). Briefly, cells were detached with Accutase ® and reseeded at 25,000 cells/cm^2 on Matrigel® (Corning) coated plates. Following incubation in E8 media supplemented with 10μM ROCK inhibitor for 24 hours, cells were expanded for additional 48 hours. CM differentiation was initiated using 12μM of CHIR99021 (LC Laboratories), in RPMI 1640 supplemented with B27™ without insulin (RB-, Life Technologies). After 24 hours incubation at 37°C, CHIR99021 was removed and replaced with RB- medium for 48 hours. Fresh RB- medium, supplemented with 5μM IWP-2 (Tocris), was then exchanged, and cells were incubated for 48 hours. Next, medium was exchanged for fresh RB- for 48 hours. Finally, medium was replaced with RPMI 1640 supplemented with B27™ and insulin (RB+, Life Technologies).

On Day 11, spontaneously beating CMs were reseeded onto Matrigel coated 24 well plates at 50,000 cells per well. After 24-hours, HUVECs and CMs were indirectly exposed to freshly-made hydrogels using a Transwell® insert. The hydrogels were prepared by adding all reagents into a 96 well plate (100μL hydrogels), mixing via pipette, and curing for 30 minutes at 37°C. Upon curing, the hydrogels were transferred to the Transwell® insert and the viability of the cells was determined after 24 hours of contact with leachables from the hydrogel. Cells were stained green-fluorescent calcein-AM and red-fluorescent ethidium homodimer-1 (Live/Dead Viability/Cytotoxicity Kit, Thermofisher). Imaging was conducted with a fluorescence microscope (Nikon Eclipse TE2000-S) and cells were manually counted (n = 3 images per well, 3 replicates).

***Ex-vivo* gel formation and retention.** Hydrogel retention and network formation were tested in an *ex vivo* pig heart model to identify requisite properties of the injectable hydrogel prior to *in vivo* studies. The effect of viscosity was investigated by varying the concentration of the PEUDAm macromer solution from 10 wt% (42 mPa-s) to 20 wt% (83 mPa-s). The effect of cure rate was investigated by varying the initiator concentration from 0.5:1.0 mmol APS:IG to 0.75:1.5 mmol APS:IG. Pig hearts were purchased from Animal Technologies. Increasing the APS concentration up to 1 mM resulted in premature gelation in the catheter and was not tested in the *ex vivo* model. Porcine hearts were submerged in a water bath at 37°C until thermal equilibrium was achieved. An incision was made to access the cardiac vein and porcine blood was used to fill the veins prior to injecting the hydrogel solution. Hydrogel precursor solutions were prepared as described above and loaded into the double-barrel syringe. A catheter was connected to the mixing head to deliver the hydrogel into the veins and the catheter and mixing head were purged with saline to avoid air bubbles. After injection, the hydrogel was allowed to cure for at least 30 minutes and then removed

using a scalpel to access the vein. The gel was divided into five segments for characterization of gel fraction and equilibrium swelling as described above. To investigate the key determinants of gel retention, the effect of viscosity on mixing was modeled and the increase in complex viscosity as a function of macromer viscosity and cure rate was measured. The complex viscosity was determined from the gelation kinetics experiments described above. The precursors were dispensed through a mixing head onto a parallel-plate (20 mm) configuration heated to 37 °C. Complex viscosity was measured every 6 seconds.

Flow simulations. Simulations were performed using SolidWorks Flow Simulation on a branched vessel constructed in SolidWorks 2021 (Dassault Systèmes, Vélizy-Villacoublay, France) with dimensions in the range of typical cardiac vessel sizes (main vessel ID=3mm). A standard spiral mixing head was modeled and inserted into the main branch of the vessel model. Ambient pressure conditions were present at each outlet, and a zero flow condition was applied to the inlet. The vessel model was filled with liquid matching blood viscosity to begin the simulation study. A transient analysis was performed in which the precursor solutions were injected through the mixing head at a rate of 3.5 mL/s for 0.5s, after which the study continued until t=15 seconds. The viscosity of the precursors was adjusted to assess the impact on mixing and initial flow into the vessels.

Electrical characterization. The resistance of the hydrogel specimens (n = 4) was determined using a four-point probe and electrochemical impedance spectroscopy. The gels were placed over a glass slide and the 4-point probe (Signatone 4Pro) was placed in contact with the center of the specimen. The Voltage-Current characteristics of the hydrogels were recorded by varying the

current from 0.001mA to 0.2mA. The resistance (R) of the sample was obtained from the slope of the V-I plot. For impedance analysis, the hydrogel constructs were placed between two electrodes with an alternate current sweep from 1 to 10^6 Hz. The resistance (R) of the sample was obtained when the impedance plateau in the bode plot. Conductivity was calculated as the inverse of the resistance multiplied by the gel thickness (L) and divided by the area (A). Additionally, the conductivity of ionic species present was determined using an Oakton CON 450 portable conductivity meter. Specimens were submerged in 5mL of DI water and allowed to reach equilibrium. The recorded conductivity value was compared to a standard curve to determine the hydrogel's relative ion concentration and ionic conductivity. Monophasic pulse signals (5 msec) were generated by a function generator (Rigol DG4000) at a frequency of 2 Hz. An oscilloscope (Tektronix MDO3000) was used to measure the voltage drop to a resistor connected in series with the hydrogel in order to calculate the current as a function of hydrogel length. The circuit was connected to a 5 volts power supply. The schematics of the circuit setup are shown in Fig S30.

In-vivo assessment of gel conductivity retention. All procedures were approved by the IACUC at the Texas Heart Institute. Hydrogels prepared in deionized water and or saline (1.35% NaCl) were implanted subcutaneously in a rat model to evaluate the effect of implantation on the ionic conductivity. Hydrogels were prepared and maintained under sterile conditions until implantation. All precursor and redox solutions were filtered using a 0.22-micron size filter. Specimens (5 mm by 10 mm) were cut and enclosed in sterile cylindrical stainless steel mesh cages (16 mm length, 8 mm diameter) prepared at the UT Austin laboratory as describe above. The wire cages and glass plates were cleaned with 1% alkaline detergent and sterilized with UV light. Gels were submerged into DI water or phosphate buffered saline (PBS 1X). Solution was changed three times and gels

were allowed to reach equilibrium swelling for at least 24 hours. Gels were kept submerged until implantation. Fischer rats were acquired and quarantined for 7 days prior to the procedure. The rats were anesthetized using 1.5% Isoflurane through a face mask that was secured to the operating table, hair was removed using clippers and depilatory cream. Incisions were made on the side of the rats, and two subcutaneous pockets were prepared above the fascia of the external oblique muscle to implant one cage with a hydrogel in DI water and another with a hydrogel in PBS on each rat. The incisions were then closed with 4-0 Vicryl sutures. After 7 days, rats were euthanized under deep anesthesia via intraperitoneal injection of Pentobarbital solution (Euthasol). The stainless steel cages were explanted, and hydrogel specimens were carefully removed from the cages. The ionic conductivity was assessed using a hand-held conductivity probe. The sol fraction and swelling ratio were measured as described above.

***In vivo* deployment and pacing in a porcine model.** Animal procedures were approved by the IACUC at the Texas Heart Institute. Yorkshire pigs (n=3) between 40kg and 50kgs were used for these studies. The animal was sedated and prepared for a non-survival study. The animal's vitals were constantly monitored using a 12-lead ECG for the entirety of the procedure. A median sternotomy was performed on the pig to expose the anterior surface of the heart. The anterior intraventricular vein (AIV) branch that would eventually be utilized for injection was identified prior to pacing, and four decapolar catheters were sutured around the target site in a square for sensing the local epicardial electrograms. The pacing target was selected close to the expected injection site on the AIV. Unipolar cathodal pacing using a bare metal electrode, hydrogel point, and hydrogel line were performed with the indifferent electrode placed on the animal's chest. Three different modes of pacing were attempted – point pacing, line pacing, and pacing with the

hydrogel in the AIV. Hydrogel precursors were prepared as described above and loaded into a double-barrel syringe with a mixing head. To prepare the hydrogel for point pacing, approximately 500μL of hydrogel was dispensed into a 2 mL tube containing an exposed stainless-steel wire such that that the gel cured (~ 8 mm diameter) around the end of the wire. To prepare the hydrogel line source, the hydrogel precursors were injected into a 14G angiocatheter and allowed to cure for 2 minutes to generate a cylindrical hydrogel line source for epicardial pacing (1.6 mm in diameter, 8 cm in length). The minimum current required to achieve reliable pacing capture (called capture threshold) at a constant pulse width (0.5, 1, 5, and 10ms) was recorded for each pacing mode, yielding different strength-duration curves. After completion of epicardial pacing, the hydrogel was injected into the AIV, using a 14G angiocatheter inserted directly into the AIV with a suture at the proximal end to minimize bleeding and backflow of the hydrogel precursor solutions upon injection. A double barrel syringe with a mixing head was loaded with 2 mL of hydrogel precursor solutions and attached to the angiocatheter via the luer lock system described above. After allowing the gel to cure for ~2 minutes, the catheter was and a pacing lead was attached to the hydrogel at the proximal end using an alligator clip. Capture threshold and strength-duration curves were again assessed with the intravascular hydrogel. At the end of the study, all animals were humanely euthanized per standard IACUC approved practice. After euthanasia, the AIV was carefully opened using a scalpel and the cured hydrogel was carefully excised from the AIV for inspection. The hydrogel was then sectioned and characterized for sol fraction and swelling ratio as described above.

Safety assessment in a porcine model. A Yorkshire pig was utilized for the chronic assessment of the safety of the injected hydrogel vein occlusion. The pig was sedated and prepared for a

chronic study, with vitals and ECG recorded and monitored throughout the study. A thoracotomy was performed to expose the left side of the heart. The largest vein in that window, the MCV, was identified by the surgeon and accessed using a 16G angiocatheter. The sterile hydrogel precursor solutions, prepared as described above, were injected via the double barrel syringe into the vein. The gel was allowed to cure for 5 minutes before catheter removal. A single prolene suture was placed at the injection site to minimize any potential bleeding and as a marker for histological analysis after 2 weeks. Echo cardiograms and blood enzymes were drawn at the beginning and end of the study to assess myocardial changes. The animal was allowed to recover and was observed for 2 weeks prior to euthanization. The animal was prepared for termination, and ECGs, echocardiogram, and enzyme blood draws were repeated before termination. After euthanized was complete, the heart was removed and placed in fixative for histological analysis. A small section of hydrogel was also removed for sol fraction and swelling ratio analysis prior to fixation. The heart was sectioned along the length of the hydrogel injection to include the proximal vein for site of injection, the mid-vein, and the distal vein, all approximately 2 cm apart. Paraffin-embedded sections were stained with either hematoxylin and eosin or Masson's Trichrome to assess cardiac injury and inflammatory responses of the tissue to the chronic hydrogel implant.

Statistical Analysis: Quantitative characterization of hydrogel properties is expressed as mean $\pm$ standard deviation. Statistical comparisons were made using the Student's t test for paired data and analysis of variance (ANOVA) for multiple comparisons with Tukey post hoc analysis for parametric data. Computations were performed using GraphPad Prism version 9.0.2 at the significance levels of $p < 0.05$.

Supplementary Materials

Fig. S1. Synthetic route for PEUDAm.

Fig. S2. H1 NMR Spectra of PEG-CDI.

Fig. S3. H1 NMR Spectra of PEG-EDA.

Fig. S4. H1 NMR Spectra of PEUDAm.

Fig. S5. GPC plot of PEUDAm.

Fig. S6. H1 NMR spectra of NAGA.

Fig. S7. PEUDAm stability under hydrolytic conditions. PEGDA 20kDa, 20% + 1% NAGA was used as a positive control to assess improved biostability of PEUDAm 20kDa, 20% + 1% NAGA gels. Hydrogel swelling of (A) PEGDA and (B) PEUDAm in PBS was recorded as a function of time at 37°C and under accelerated conditions (70°C). Means and standard deviations are presented (n = 4). In most cases, the standard deviation is smaller than the bullets' size and was not displayed by the software.

Fig. S8. PEUDAm stability under hydrolytic conditions. Storage modulus of hydrogels under 1% strain at 0.5 Hz (compression).

Fig. S9. Rheological data: Time sweep of PEUDAm hydrogels. Time sweep of PEUDAm 20kDa, 15% cured at a concentration at 0.5, 0.75, and 1.0mM APS. The concentration of IG was twice the APS concentration: 1.0, 1.5, and 2.0, respectively.

Fig. S10. Impact of macromer molecular weight and adding NAGA on full network formation. A) Impact on time to full network formation of PEUDAm 15%. B) Impact of adding NAGA into a PEUDAm 20k, 15% solution.

Fig. S11. Conductivity of hydrogel determined by 4-point probe and conductivity probe. A) Comparison of hydrogels conductivity using the four-point probe and the hand-held conductivity probe of hydrogels in equilibrium in deionized water (water gel) or PBS (ionic gel). B) Standard curve of conductivity vs. phosphate buffered saline solution concentration.

Fig. S 12. Hydrogel conductivity at different NaCl concentrations. Initial hydrogel conductivity at different NaCl concentrations and conductivity after reaching swelling equilibrium in PBS.

Fig. S13. Network properties prior and after implantation in rats.

Fig. S14. Circuit used to test the impact of hydrogel length on current. An oscilloscope measured the voltage across the sense resistor to determine the current.

Fig. S15. Capture thresholds for each acute porcine study (n=3).

Fig. S16. 12-lead ECG of the heart at sinus rhythm.

Fig. S17. 12-lead ECG of unipolar pacing with a metal electrode placed directly on the left ventricular myocardium. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Fig. S18. 12-lead ECG of unipolar pacing with a small blob of hydrogel electrode placed directly on the left ventricular myocardium. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Fig. S19. 12-lead ECG of unipolar pacing with a linear hydrogel electrode (~4 cm) placed directly on the left ventricular myocardium close to the interventricular septum. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Fig. S20. 12-lead ECG of unipolar pacing from hydrogel cured in situ in the AIV. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Fig. S21. Animal02: 12-lead ECG of the heart at sinus rhythm.

Fig. S22. Animal02: 12-lead ECG of unipolar pacing with a metal electrode placed directly on the left ventricular myocardium. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Fig. S23. Animal02: 12-lead ECG of unipolar pacing with a small blob of hydrogel electrode placed directly on the left ventricular myocardium. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Fig. S24. Animal02: 12-lead ECG of unipolar pacing with a linear hydrogel electrode (~4 cm) placed directly on the left ventricular myocardium close to the interventricular septum. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Fig. S25. Animal02: 12-lead ECG of unipolar pacing from hydrogel cured in situ in the AIV. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Fig. S26. Animal03: 12-lead ECG of the heart at sinus rhythm.

Fig. S 27. Animal03: 12-lead ECG of unipolar pacing with metal electrode placed directly on the left ventricular myocardium. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Fig. S28. Animal03: 12-lead ECG of unipolar pacing with a small blob of hydrogel electrode placed directly on the left ventricular myocardium. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Fig. S29. Animal03: 12-lead ECG of unipolar pacing with a linear hydrogel electrode (~4 cm) placed directly on the left ventricular myocardium close to the interventricular septum. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Fig. S30. Animal03: 12-lead ECG of unipolar pacing from hydrogel cured in situ in the AIV. Pacing artifacts can be seen as spikes in each of the channels. Pacing was performed at 120 bpm.

Table S1. Gel fraction effect of incorporating NAGA.

Table S2. Capture thresholds (in mA) for each material and acute animal tested (n=3). NC = no capture.

Movie S1. Mixing of precursors solutions using a dual lumen catheter for dispensing.

Movie S2. Mixing of precursors solutions when dispensed using a double-barrel syringe with a mixing head.

References and Notes

1.   R. M. John *et al.*, Ventricular arrhythmias and sudden cardiac death. *The Lancet* **380**, 1520-1529 (2012).

2.   R. Derksen *et al.*, Tissue discontinuities affect conduction velocity restitution: a mechanism by which structural barriers may promote wave break. *Circulation* **108**, 882-888 (2003).

3.   E. Patterson, B. J. Scherlag, E. J. Berbari, R. Lazzara, Slow conduction through an arc of block: A basis for arrhythmia formation postmyocardial infarction. *J Cardiovasc Electrophysiol* **28**, 1203-1212 (2017).

4.   Z. Qu, J. N. Weiss, Mechanisms of ventricular arrhythmias: from molecular fluctuations to electrical turbulence. *Annu Rev Physiol* **77**, 29-55 (2015).

5.   W. G. Stevenson, J. N. Weiss, I. Wiener, K. Nademanee, Slow conduction in the infarct scar: relevance to the occurrence, detection, and ablation of ventricular reentry circuits resulting from myocardial infarction. *Am Heart J* **117**, 452-467 (1989).

6.   H. J. Wellens, P. Brugada, J. Farre, Ventricular arrhythmias: mechanisms and actions of antiarrhythmic drugs. *Am Heart J* **107**, 1053-1057 (1984).

7.   E. S. Williams, M. N. Viswanathan, Current and emerging antiarrhythmic drug therapy for ventricular tachycardia. *Cardiol Ther* **2**, 27-46 (2013).

8.   R. Cardoso, F. R. Assis, A. D'Avila, Endo-epicardial vs endocardial-only catheter ablation of ventricular tachycardia: A meta-analysis. *J Cardiovasc Electrophysiol*, (2019).

9.   H. Versteeg, D. A. Theuns, R. A. Erdman, L. Jordaens, S. S. Pedersen, Posttraumatic stress in implantable cardioverter defibrillator patients: the role of pre-implantation distress and shocks. *International journal of cardiology* **146**, 438-439 (2011).

10.   J. Seuring, F. M. Bayer, K. Huber, S. Agarwal, Upper Critical Solution Temperature of Poly(N-acryloyl glycinamide) in Water: A Concealed Property. *Macromolecules* **45**, 374-384 (2012).

11.   S. Ramadan, N. Paul, H. E. Naguib, Standardized static and dynamic evaluation of myocardial tissue properties. *Biomedical Materials* **12**, 025013 (2017).

12.   M. S. Amzulescu *et al.*, Myocardial strain imaging: review of general principles, validation, and sources of discrepancies. *European Heart Journal - Cardiovascular Imaging* **20**, 605-619 (2019).

13.   X. Guo, Y. Liu, G. S. Kassab, Diameter-dependent axial prestretch of porcine coronary arteries and veins. *Journal of applied physiology* **112**, 982-989 (2012).

14.   M. Whitely *et al.*, Improved in situ seeding of 3D printed scaffolds using cell-releasing hydrogels. *Biomaterials* **185**, 194-204 (2018).

15.   M. Solazzo, F. J. O'Brien, V. Nicolosi, M. G. Monaghan, The rationale and emergence of electroconductive biomaterial scaffolds in cardiac tissue engineering. *APL bioengineering* **3**, 041501 (2019).

16.   D. M. Pedrotty *et al.*, Three-dimensional printed biopatches with conductive ink facilitate cardiac conduction when applied to disrupted myocardium. *Circulation: Arrhythmia and Electrophysiology* **12**, e006920 (2019).

17.   M. Potse, B. Dubé, A. Vinet, Cardiac anisotropy in boundary-element models for the electrocardiogram. *Medical & biological engineering & computing* **47**, 719-729 (2009).

18.   J. G. Webster, I. E. i. Medicine, B. Society, *Design of Cardiac Pacemakers*. (IEEE, 1995).

19.   M. J. Janse, A. L. Wit, Electrophysiological mechanisms of ventricular arrhythmias resulting from myocardial ischemia and infarction. *Physiological Reviews* **69**, 1049-1169 (1989).

20. G. Camci-Unal, N. Annabi, M. R. Dokmeci, R. Liao, A. Khademhosseini, Hydrogels for cardiac tissue engineering. *NPG Asia Materials* **6**, e99 (2014).

21. J. Prager *et al.*, Stiffness-matched biomaterial implants for cell delivery: clinical, intraoperative ultrasound elastography provides a 'target' stiffness for hydrogel synthesis in spinal cord injury. *Journal of Tissue Engineering* **11**, 204173142093480 (2020).

22. C. Yu, F. Yao, J. Li, Rational design of injectable conducting polymer-based hydrogels for tissue engineering. *Acta Biomaterialia*, (2021).

23. M. Shin, K. H. Song, J. C. Burrell, D. K. Cullen, J. A. Burdick, Injectable and Conductive Granular Hydrogels for 3D Printing and Electroactive Tissue Support. *Advanced Science* **6**, 1901229 (2019).

24. M. Zyl *et al.*, Injectable conductive hydrogel restores conduction through ablated myocardium. *Journal of Cardiovascular Electrophysiology* **31**, 3293-3301 (2020).

25. J. Huang, X. Jiang, Injectable and Degradable pH-Responsive Hydrogels via Spontaneous Amino–Yne Click Reaction. *ACS Applied Materials & Interfaces* **10**, 361-370 (2018).

26. J. L. Ifkovits *et al.*, Injectable hydrogel properties influence infarct expansion and extent of postinfarction left ventricular remodeling in an ovine model. *Proceedings of the National Academy of Sciences* **107**, 11507-11512 (2010).

27. J. Shi *et al.*, Cell-compatible hydrogels based on a multifunctional crosslinker with tunable stiffness for tissue engineering. *Journal of Materials Chemistry* **22**, 23952-23962 (2012).

28. N. J. Steinmetz, E. A. Aisenbrey, K. K. Westbrook, H. J. Qi, S. J. Bryant, Mechanical loading regulates human MSC differentiation in a multi-layer hydrogel for osteochondral tissue engineering. *Acta biomaterialia* **21**, 142-153 (2015).

29. J. S. Temenoff *et al.*, Thermally cross-linked oligo (poly (ethylene glycol) fumarate) hydrogels support osteogenic differentiation of encapsulated marrow stromal cells in vitro. *Biomacromolecules* **5**, 5-10 (2004).

30. T. S. Wilems *et al.*, Effects of free radical initiators on polyethylene glycol dimethacrylate hydrogel properties and biocompatibility. *Journal of Biomedical Materials Research Part A* **105**, 3059-3068 (2017).

31. S. A. Young *et al.*, Mechanically resilient injectable scaffolds for intramuscular stem cell delivery and cytokine release. *Biomaterials* **159**, 146-160 (2018).

32.   S. M. Dorsey *et al.*, MRI evaluation of injectable hyaluronic acid-based hydrogel therapy to limit ventricular remodeling after myocardial infarction. *Biomaterials* **69**, 65-75 (2015).

33.   A. Ido, N. Hasebe, H. Matsuhashi, K. Kikuchi, Coronary sinus occlusion enhances coronary collateral flow and reduces subendocardial ischemia. *American Journal of Physiology-Heart and Circulatory Physiology* **280**, H1361-H1367 (2001).

34.   B. Merkely *et al.*, Chronic implantation of intravascular cardioverter defibrillator in a canine model: device stability, vascular patency, and anchor histology. *Pacing and Clinical Electrophysiology* **36**, 1251-1258 (2013).

35.   J. J. Thornton, D. E. Gregg, Effect of chronic cardiac venous occlusion on coronary arterial and cardiac venous hemodynamics. *American Journal of Physiology-Legacy Content* **128**, 179-184 (1939).

36.   J. H. Traverse *et al.*, First-in-Man Study of a Cardiac Extracellular Matrix Hydrogel in Early and Late Myocardial Infarction Patients. *JACC: Basic to Translational Science* **4**, 659-669 (2019).

37.   S. Liang *et al.*, Paintable and rapidly bondable conductive hydrogels as therapeutic cardiac patches. *Advanced Materials* **30**, 1704235 (2018).

38.   A. Navaei *et al.*, Gold nanorod-incorporated gelatin-based conductive hydrogels for engineering cardiac tissue constructs. *Acta biomaterialia* **41**, 133-146 (2016).

39.   S. Buchan, R. Kar, M. John, A. Post, M. Razavi, Electrical Stimulation for Low-Energy Termination of Cardiac Arrhythmias: a Review. *Cardiovascular Drugs and Therapy*, (2021).

40.   L. Jin, J. Wang, B. Song, X. Wu, Z. Fang, in *2015 37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*. (2015), pp. 5688-5691.

41.   X. Zheng *et al.*, Reduction of Atrial Defibrillation Threshold With an Interatrial Septal Electrode. *Circulation* **102**, 2659-2664 (2000).

42.   X. Zheng, M. E. Benser, G. P. Walcott, R. E. Ideker, Right Atrial Septal Electrode for Reducing the Atrial Defibrillation Threshold. *Circulation* **104**, 1066-1070 (2001).

43.   A. Moreno *et al.*, Wide-area low-energy surface stimulation of large mammalian ventricular tissue. *Scientific reports* **9**, 1-11 (2019).

44.   S. K. Padala, K. A. Ellenbogen, Left bundle branch pacing is the best approach to physiological pacing. *Heart Rhythm O2* **1**, 59-67 (2020).

45.   L. I. B. Heckman *et al.*, Electrical characteristics of deep septal vs. left bundle branch (area) pacing. *European Heart Journal* **41**, (2020).

46.   A. Di Marco *et al.*, Deep septal pacing to upgrade patients with pacing-induced cardiomyopathy. *HeartRhythm Case Reports*.

47.   S. J. Nelson, J. J. Creechley, M. E. Wale, T. J. Lujan, Print-A-Punch: A 3D printed device to cut dumbbell-shaped specimens from soft tissue for tensile testing. *Journal of Biomechanics* **112**, 110011 (2020).

48.   N. Huebsch *et al.*, Automated Video-Based Analysis of Contractility and Calcium Flux in Human-Induced Pluripotent Stem Cell-Derived Cardiomyocytes Cultured over Different Spatial Scales. *Tissue Engineering Part C: Methods* **21**, 467-479 (2015).

49.   X. Lian *et al.*, Directed cardiomyocyte differentiation from human pluripotent stem cells by modulating Wnt/β-catenin signaling under fully defined conditions. *Nature Protocols* **8**, 162-175 (2013).

Acknowledgments:

The authors would like to thank the THI Animal Lab staff, particularly Dr. Abdelmotagaly Elgalad and Dr. Angel Moctezuma-Ramirez, as well as the THI Pathology Lab, particularly Dr. Pamela Potts and Dr. Ana Segura, for their contributions.

Funding:

- Ford Pre-Doctoral Fellowship, administered by the National Academy of Science, Engineering and Medicine (GJRR)
- Ford Dissertation Fellowship, administered by the National Academy of Science, Engineering and Medicine (GJRR)
- Office of Vice President for Research, The University of Texas at Austin, (ECH)
- The Roderick D. MacDonald Research Fund Award 19RDM004 (MR)

- The Sultan Qaboos Chair in Cardiology at the St. Luke's Foundation (MR)

Author contributions: Each author's contribution(s) to the paper should be listed [we encourage you to follow the CRediT model]. Each CRediT role should have its own line, and there should not be any punctuation in the initials.

- Conceptualization: ECH, MR

- Methodology: ECH, MR, GJRR, MJ, AP, SB, MW, MA

- Investigation: GJRR, MW, AP, MJ, ECH

- Visualization: GJRR, CW, MJ

- Supervision: ECH, MR

- Writing—original draft: GJRR, MW, AP, MJ, ECH

- Writing—review & editing: ECH, MR

Competing interests:

The authors have filed a provisional patent titled: "*Electrically conductive hydrogels usable as pacemaker lead extensions, apparatus for delivery of a hydrogel into the venous system, and methods of treating ventricular arrhythmia with electrically conductive hydrogels injected in the venous system*".

Data and materials availability:

All data are available in the main text or the supplementary materials. All the raw data will be uploaded into The University of Texas at Austin UTBox cloud storage. Access to the data could be provided upon request.