A comprehensive LFQ benchmark dataset on modern day acquisition strategies in proteomics.

Bart Van Puyvelde[1], Simon Daled[1], Sander Willems[2], Ralf Gabriels[3,4], Anne Gonzalez de Peredo[5], Karima Chaoui[5], Emmanuelle Mouton-Barbosa[5], David Bouyssié[5], Kurt Boonen[6,7], Christopher J. Hughes[8], Lee A. Gethings[8], Yasset Perez-Riverol[9], Nic Bloomfield[10], Stephen Tate[10], Odile Schiltz[5], Lennart Martens[3,4], Dieter Deforce[1], Maarten Dhaenens[1*]

1. ProGenTomics, Laboratory of Pharmaceutical Biotechnology, Ghent University, Ghent, Belgium
2. Department of Proteomics and Signal Transduction, Max Planck Institute of Biochemistry, Martinsried, Germany
3. VIB-UGent Center for Medical Biotechnology, VIB, 9000 Ghent, Belgium
4. Department of Biomolecular Medicine, Ghent University, 9000 Ghent, Belgium
5. Institut de Pharmacologie et de Biologie Structural (IPBS), Université de Toulouse, CNRS, UPS, Toulouse, France
6. VITO Health, Mol, Belgium
7. Centre for Proteomics, University of Antwerpen, Antwerp, Belgium
8. Waters Corporation, Wilmslow, UK
9. European Molecular Biology Laboratory, European Bioinformatics Institute, Wellcome Genome Campus, Hinxton, UK
10.Sciex, Concord, Ontario, Canada

*Corresponding author: Maarten Dhaenens (maarten.dhaenens@ugent.be)

In the last decade, a revolution in liquid chromatography-mass spectrometry (LC-MS) based proteomics was unfolded with the introduction of dozens of novel instruments that incorporate additional data dimensions through innovative acquisition methodologies, in turn inspiring specialized data analysis pipelines. Simultaneously, a growing number of proteomics datasets have been made publicly available through data repositories such as ProteomeXchange, Zenodo and Skyline Panorama. However, developing algorithms to mine this data and assessing the performance on different platforms is currently hampered by the lack of a single benchmark experimental design. Therefore, we acquired a hybrid proteome mixture on different instrument platforms and in all currently available families of data acquisition. Here, we present a comprehensive Data-Dependent and Data-Independent Acquisition (DDA/DIA) dataset acquired using several of the most commonly used current day instrumental platforms. The dataset consists of over 700 LC-MS runs, including adequate replicates allowing robust statistics and covering over nearly 10 different data formats, including scanning quadrupole and ion mobility enabled acquisitions. Datasets are available via ProteomeXchange (PXD028735).

Background & Summary

Hypothesis-driven biochemical assays have been the foundation of molecular biology for well over a century, with great success. However, the lack of a more holistic view on the biomolecular complexity requires trial-and-error experimentation. Therefore, the past few decades were characterized by a shift towards an experimental design wherein a broader biomolecular perspective of the system is first generated in order to contextualize the hypothesis and the targeted biochemical assays beforehand. These "omics" approaches were enabled by two technical revolutions, i.e. the sequencing of nucleotides and the accurate mass measurement of biomolecules by mass spectrometry (MS).

In its barest form, the output from an MS instrument is merely a list of m/z's with intensities measured at very precise moments in time. However, MS is quickly evolving towards capturing the full complexity of a biological sample. To this end, not only the accuracy of instruments has improved greatly, they now also incorporate analytical techniques that select or separate analytes based on other physico-chemical properties. In proteomics nowadays, a

mass spectrometer thus rarely only measures the *m/z* coordinate and intensity of (fragment)
ions. The ion coordinates are mostly supplemented with precursor *m/z*, retention time (t_R)
and/or ion mobility coordinates, depending on acquisition strategy. This creates a
multidimensional data matrix that captures the complexity of the sample to an unprecedented
depth [1].

The field of mass spectrometry has diversified greatly, driven by a fast sequence of
innovations from many vendors. Modern MS instruments allow the manipulation of ions in
countless of different ways, including different ionization methodologies, fragmentation
techniques, multipoles, time-of-flight tubes, ion mobility separation devices and trap designs,
including the now very dominant orbitrap. The way in which these different ion manipulations
are combined has ballooned the number of different acquisition strategies available to the
end user today but all these instrumental and strategic innovations are futile if no data
analysis pipeline is available to translate the data back into biology. For bottom-up
proteomics this implies reconstructing the peptide backbone sequences from their fragment
ions because the latter encompasses the specificity for identifying the hundreds of millions of
different protein sequences that make up the biotic world.

Conventionally, MS instruments have been operated using data dependent acquisition (DDA)
wherein the data from a precursor scan at low energy is used to pinpoint potentially
interesting analytes which are then sequentially selected for fragmentation at high energy.
These fragmentation spectra can then be identified by a plethora of different algorithms [2,3].
Data-independent acquisition (DIA) however, is the more intuitive way of analyzing a sample,
because it captures all (fragment) ions without any instrumental bias. However, interpreting
such complex data matrices has proven difficult and an additional separation dimension,
such as ion mobility, was added to increase the discriminating power [4–8]. Alternatively,
configuring a quadrupole to sequentially scan the entire mass range - while still operating
"data independent" - alleviates the complexity of the resulting fragmentation spectra even
more [9,10]. This has opened up the way for the many different spectrum-centric and peptide-
centric data analysis strategies available today [11–16]. The latest reduction in complexity or
"chimericy" of DIA spectra encompasses continuously scanning the quadrupole as is done
with SONAR [17] and Scanning SWATH [18] and combining quadrupole selection and ion
mobility separation, as is done with diaPASEF [19]. Unsurprisingly, machine learning is taking
center stage in mining the various resulting data architectures [20–27].

Here, we created a comprehensive dataset on a single benchmark experimental design
adapted from Navarro et al. [28]. It contains a ground truth that serves as a quality control for
bioinformatics algorithm development and evaluation. This sample was acquired with
adequate replicates on many of the current day instrumental platforms – partially in nano flow
LC and partially in capillary flow LC - by most of the available acquisition strategy families,
covering all commonly measured ion coordinates (**Figure 1**). Far from being complete, it still
is the most comprehensive repository of its kind for algorithmic development and validation,
both at the level of identification and quantification. Instead of being yet another way of
attaining the highest number of identified peptides, we hope it becomes a resource for
compatibility assessment and data analysis quality control. Above all, it is a snapshot of
current day completeness of our digital image of the protein world.

Figure 1. Schematic overview of the different acquisition strategies/instruments applied in the study. A comprehensive LC-MS/MS dataset was generated using samples composed of commercial Human, Yeast and E.coli full proteome digests. Two hybrid proteome samples A and B containing known quantities of Human, Yeast and E.coli tryptic peptides, as described by Navarro et al. were prepared in three consecutive times to include handling variability. Additionally, a QC sample was created by mixing one sixth of each of the six master batches (65% w/w Human, 22.5% w/w Yeast and 12.5% w/w E.coli). These commercial lysates were measured individually and as triple hybrid proteome mixtures each in triplicate using DDA and DIA acquisition methodologies available on six LC-MS/MS platforms, i.e. SCIEX TripleTOF5600 and TripleTOF 6600+, Thermo Orbitrap QE HF-X, Waters Synapt G2-Si and Synapt XS and Bruker timsTOF Pro. The complete dataset was made publicly available to the proteomics community through ProteomeXchange with dataset identifier: PXD028735. In addition, a system suitability workflow (AutoQC) was incorporated on each instrument using commercial E.coli lysate digest which were acquired at multiple timepoints throughout each sample batch. The

109 AutoQC data was automatically imported in Skyline and uploaded to the Panorama AutoQC server using AutoQC
110 loader, enabling system suitability assessment of each LC-MS/MS system used in the dataset.

Methods

Sample preparation

Mass spectrometry-compatible Human (P/N: V6951) and Yeast (P/N: V7461) protein digest
extracts were purchased from Promega (Madison, Wisconsin, United States). Lyophilised
MassPrep E.coli digest standard (P/N:186003196) was purchased from Waters Corporation
(Milford, Massachusetts, United States). The extracts were reduced with dithiothreitol (DTT),
alkylated with iodoacetamide (IAA) and digested with sequencing grade Trypsin(-Lys C) by
the respective manufacturers. The digested protein extracts were reconstituted in a mixture
of 0.1% Formic acid (FA) in water (Biosolve B.V, Valkenswaard, The Netherlands) and
spiked with iRT peptides (Biognosys, Schlieren, Switzerland) at a ratio of 1:20 v/v. Two
master samples A and B were created similar to Navarro et al., each in triplicate, as shown in
Figure 1. Sample A was prepared by mixing Human, Yeast and E.coli at 65%, 30% and 5%
weight for weight (w/w), respectively. Sample B was prepared by mixing Human, Yeast and
E.coli protein digests at 65%, 15%, 20% w/w, respectively. The resulting samples have
logarithmic fold changes (log2FCs) of 0,-1 and 2 for respectively Human, Yeast and E.coli.
One sixth of each of the triplicate master batches of A and B were mixed to create a QC
sample, containing 65% w/w Human, 22.5% w/w Yeast and 12.5% w/w E.coli.

LC-MS/MS

In this section, a detailed description of the different LC-MS/MS parameters is given for each
LC-MS/MS instrumental setup applied to generate this comprehensive dataset. All
instruments were operated according to the lab's best practice, i.e. not necessarily the best
attainable, but rather most realistic data quality. Sample load was chosen based on LC setup
(nano flow = 1 μg on column vs capillary flow = 5 μg on column) and instrument sensitivity.
Thus, differences in absolute number of identified peptides and proteins can be attributed to
sample load, LC flow rate, MS instrumentation, operator's choices and search algorithmic
compatibility; direct conclusions on MS instrument performance can therefore not be drawn
from this dataset.

As a rule of thumb, data-dependent acquisition (DDA) methods use high energy
fragmentation spectra (MS2) of narrow mass selections for identification and use the area
under the curve of the precursor (MS1) for quantification. Therefore, a cycle time needs to be
attained wherein enough datapoints across the precursor elution peak are sampled for
accurate quantification. In most data-independent acquisition (DIA) strategies, a broader
precursor selection window is used and both identification and quantification can be done at
the fragment level, taken that the cycle time for both MS1 and MS2 is adapted to the LC
gradient. Finally, Precursor Acquisition Independent From Ion Count (PAcIFIC) is a method
that is acquired solely to extend the size of the peptide library for detecting peptides in DIA
data and is therefore not strictly dependent on the cycle time. Of note, by scanning the
quadrupole instead of acquiring different mass windows separately, acquisition strategies like
SONAR and Scanning SWATH create an additional dimension in the data matrix, akin to
how ion mobility separation is perceived. Since these are similar to PAcIFIC acquisition, we
also acquired gas phase (GP) fractions in SONAR and Scanning SWATH for library building,
i.e. with no emphasis on cycle time.

1) SCIEX TripleTOF 5600 (Capilary flow LC)

A TripleTOF 5600 mass spectrometer (Sciex, Concord, Ontario, Canada) fitted with a
Duospray ion source operating in positive ion mode, was coupled to an Eksigent NanoLC
400 HPLC system (Eksigent, Dublin, CA). 5 µL of each sample was loaded at 5µL/min with
0.1% Trifluoroacetic acid (TFA) in water and trapped on a YMC TriArt C18 guard column (id
500µm, length 5mm, particle size 3 µm) for 5 minutes. Peptides were separated on a
microLC YMC TriArt C18 column (id 300 µm, length 15 cm, particle size 3 µm) maintained at
55°C at a flow rate of 5µL/min by means of trap-elute injection. Mobile phase A consisted of
UPLC-grade water with 0.1% (v/v) FA and 3% (v/v) DMSO, and mobile phase B consisted of
UPLC-grade ACN with 0.1% (v/v) FA. Peptide elution was performed at 5µL/min using the
following gradient: i) 2% to 30% mobile phase B in 120 min, ii) ramp to 90% mobile phase B
in 1 min. The washing step at 90% mobile phase B lasted 4 min and was followed by an
equilibration step at 2% mobile phase B (starting conditions) for 10 min. Ion source
parameters were set to 5.5 kV for the ion spray voltage, 30 psi for the curtain gas, 13 psi for
the nebulizer gas and 80°C as source temperature.

a. Data-Dependent Acquisition

For DDA (a cycle time of 3.5 s), MS1 spectra were collected between 399-1200 m/z for 500
ms. The 20 most intense precursors ions with charge states 2-5 that exceeded 250 counts
per second were selected for fragmentation, and the corresponding fragmentation MS2
spectra were collected between 50-2000 m/z for 151 ms. After the fragmentation event, the
precursor ions were dynamically excluded from reselection for 20 s.

b. PAcIFIC

For PAcIFIC (a cycle time of 4 s), the TripleTOF5600 was configured to acquire eight gas
phase fractionated acquisitions with isolation windows of 4 m/z using an overlapping window
pattern from narrow mass ranges, as described by Searle et al (i.e., 396.43 – 502.48; 496.48
– 602.52;596.52 – 702.57;696.57 – 802.61; 796.61 – 902.66; 896.6 – 1002.70; 996.70 –
1102.75; 1096.75 – 1202.80) [29]. See **Supplementary Table 2** for the actual windowing
scheme. MS2 spectra were collected in high-sensitivity mode from 360-1460 m/z, for 75 ms.
An MS1 survey scan was recorded per cycle from 360-1460 m/z for 50ms.

c. SWATH 64 variable windows

For SWATH (a cycle time of 3.4 s), a 64 variable window acquisition scheme as described by
Navarro et al. was used for all samples (**Supplementary Table 1**) [28]. Briefly, SWATH MS2
spectra were collected in high-sensitivity mode from 50-2000 m/z, for 50 ms. Before each
SWATH MS cycle an additional MS1 survey scan in high sensitivity mode from 400-1200 m/z
was recorded for 150 ms.

2) SCIEX TripleTOF 6600+ (Capilary flow LC)

A TripleTOF 6600+ mass spectrometer (Sciex, Concord, Ontario, Canada) fitted with an
Optiflow ion source operating in positive ion mode, was coupled to an Eksigent NanoLC 425
HPLC system (Eksigent, Dublin, CA). 5 µL of each sample was loaded at 5µL/min with 0.1%
FA in water by means of direct injection. Peptides were separated on a Phenomenex Luna
Omega Polar C18 column (150 x 0.3 mm, particle size 3 µm) at a column temperature of
30°C. Mobile phase A consisted of UPLC-grade water with 0.1% (v/v) FA, and mobile phase
B consisted of UPLC-grade ACN with 0.1% (v/v) FA. Peptide elution was performed at
5µL/min using the following gradient: i) 2% to 30% mobile phase B in 120 min, ii) ramp to
90% mobile phase B in 1 min. The washing step at 90% mobile phase B lasted 4 min and
was followed by an equilibration step at 2% mobile phase B (starting conditions) for 10 min.
Ion source parameters were set to 4.5 kV for the ion spray voltage, 25 psi for the curtain gas,

10 psi for nebulizer gas (ion source gas 1), 20 psi for heater gas (ion source gas 2) and
100°C as source temperature.

a. Data-Dependent Acquisition

203 For DDA acquisition (a cycle time of 3.3 s), MS1 spectra were collected between 400-1200
204 m/z for 250 ms. The 30 most intense precursor ions with charge states 2-4 that exceeded
205 300 counts per second were selected for fragmentation, and the corresponding
206 fragmentation MS2 spectra were collected between 100-1500 m/z for 100 ms. After the
207 fragmentation event, the precursor ions were dynamically excluded from reselection for 10 s.

b. SWATH 99 Variable windows

209 For SWATH (a cycle time of 4 s), a 99 variable window acquisition scheme was used (see
210 **Supplementary Table 3**) [30]. Briefly, SWATH MS2 spectra were collected in high sensitivity
211 mode from 100-1500 m/z, for 37.5 ms. Before each SWATH MS cycle an additional MS1
212 survey scan in high sensitivity mode was recorded for 250 ms.

c. Scanning SWATH GP 1Da

A Scanning SWATH beta version was installed on the Analyst TF control software in
collaboration with Sciex Research. Scanning SWATH Q1 calibration was confirmed by
directly infusing a tuning solution (ESI Positive Calibration Solution for the SCIEX X500B
System - P/N: 5049910) and by acquiring a pre-built calibration batch (SSCalibration.dab).
Afterwards, the calibration was verified by i) running a verification calibration batch and
inspect the data in PeakView. Calibration of the Q1 after initial calibration was controlled
automatically by internal recalibration of each data file at run time removing the need for
subsequent calibration events.

The gas-phase fractionation approach usually acquired in PAcIFIC, were also acquired by
Scanning SWATH because this uniquely allows to apply DIA annotation algorithms for library
building of subsequent full mass range DIA acquisition. Precursor isolation window was set
to 1 m/z and a mass range of 100 m/z was covered in 6 s (average accumulation time per
precursor: 59.57 ms). An MS1 scan was included and data was acquired in high resolution
mode. Raw data was converted to standard SCIEX data files with an effective precursor
isolation of 0.2 m/z bins and Q1 calibration was obtained by running rawSSProcessor.exe.

d. Scanning SWATH Fixed 5Da Window

230 The precursor mass range of 400-900 m/z was covered in 4 s with a 5 m/z isolation window
231 (average accumulation time per precursor: 37.5 ms). A TOF MS scan was included and data
232 was acquired in high sensitivity mode. Raw data was converted to standard SCIEX data files
233 with an effective precursor isolation of 1 m/z bins and Q1 calibration was obtained by running
234 the rawSSProcessor.exe.

3) Thermo Orbitrap QE HF-X (Nano flow LC)

A Thermo Orbitrap QE HF-X (Thermo Fisher Scientific, Waltham, Massachusetts, United
States) was coupled to an UltiMate 3000 LC-system (NCS-3500RS Nano/Cap System,
Thermo Fisher Scientific). Peptides were separated on an Acclaim PepMap C18 column (id
75 μm, length 50 cm, particle size 2 μm, Thermo Fisher Scientific ref 164942) at a flow rate
of 350 nL/min by means of trap-elute injection (Acclaim PepMap C18, id. 300 μm x 5mm)
after 3min desalting on a nano-trap cartridge (id.300 μm, length 5mm, Thermo Fisher
Scientific ref 160454).

Mobile phase A consisted of UPLC-grade water with 0.1% (v/v) FA, and mobile phase B
consisted of UPLC-grade ACN with 0.1% (v/v) FA. Peptide elution was performed at 350
nL/min using the following gradient: i) 2% to 30% mobile phase B in 120 min, ii) ramp to 90%
mobile phase B in 1 min. The washing step at 90% mobile phase B lasted 4 min and was
followed by an equilibration step at 2% mobile phase B (starting conditions) for 21 min.

a. Data-Dependent Acquisition

The data-dependent acquisition runs on the Q Exactive HF-X were acquired with MS survey
scans (350-1400 m/z) at a resolution of 60,000, and an AGC target of 3e6. The 12 most
intense precursor ions, were selected for fragmentation by high-energy collision–induced
dissociation, and the resulting fragments were analyzed at a resolution of 15,000 using an
AGC target of 1e5 and a maximum fill time of 22 ms. Dynamic exclusion was used within 30
s to prevent repetitive selection of the same peptide.

b. Narrow Window Gas-Phase fractionation (GP) DIA

Narrow-window GP-DIA data was acquired as described by Searle et al [29]. Briefly, 6 GP runs
(400-500, …, 900-1000 m/z) using staggered 4 m/z DIA spectra (4 m/z precursor isolation
windows at 30,000 resolution, AGC target 1e6, maximum inject time 60 ms, NCE 27, +3H
assumed charge state) were acquired using an overlapping window pattern, described by
Pino et al [31]. In each run, full MS scans matching each part of the fractionated mass range
(i.e., either 395-505, 495-605, 595-705, 695-805, 795-905, or 895-1005 m/z), acquired at a
resolution of 60,000 using an AGC target of 1e6 and a maximum inject time of 60ms, were
interspersed every 25 MS/MS spectra.

c. All Ion Fragmentation (AIF)

The AIF DIA data was acquired using a staggered pattern of 75x8 m/z isolation windows
over the mass range 400-1000 m/z as described by Pino et al (**Supplementary Table 4**) [31].
DIA MS/MS scans were acquired at 15,000 resolution, with an AGC target of 1e6, a
maximum inject time 20ms, and a NCE of 27. Full MS scans over the range 390-1010 m/z at
60,000 resolution, AGC target 1e6, maximum inject time 60 ms were interspersed every 75
MS/MS spectra.

272
4) Waters Synapt G2-Si (Capillary flow LC)

An M-class LC system (Waters Corporation, Milford, MA) was equipped with a 1.7 μm CSH
130 C18 300 μm x 100 mm column, operating at 5 μL/min with a column temperature of 55
°C. Mobile phase A was UPLC-grade water containing 0.1% (v/v) FA and 3% DMSO, mobile
phase B was ACN containing 0.1% (v/v) FA. Peptides were separated using a linear gradient
of 3–30% mobile phase B over 120 minutes. All experiments were conducted on a Synapt
G2-Si mass spectrometer (Waters Corporation, Wilmslow, UK). The ESI Low Flow probe
capillary voltage was 3 kV, sampling cone 60 V, source offset 60 V, source temperature 80
°C, desolvation temperature 350 °C, cone gas 80 L/hr, desolvation gas 350 L/hr, and
nebulizer pressure 2.5 bar. A lock mass reference signal of GluFibrinopeptide B (m/z
785.8426) was sampled every 30 s.

a. HD-DDA

Data was acquired according to Helm et al. with minor adaptations [7]. Briefly, in data-
dependent mode, the MS automatically switches between MS survey and MS/MS scans

based upon a set of switching criteria, including ion intensity and charge state. Full scan MS
and MS/MS spectra (m/z 50 - 5000) were acquired in sensitivity mode. MS survey spectra
were acquired using a fixed acquisition time of 250 ms and the ions present in each scan
were monitored for the following criteria: more than 3000 intensity/sec and only 2,3,4,5+
charge states. Once criteria were satisfied, the precursor ion isolation width of the
quadrupole was set to 1.0 Th around each precursor sequentially. Tandem mass spectra of
up to 12 precursors were generated in the trapping region of the ion mobility cell by using a
collisional energy ramp from 6/9 V (low mass 50 Da, start/end) to up to 147/183 V (high
mass 5000 Da, start/end), with actual values applied dependent upon the precursor m/z. The
MS2 scan time was set to 100 ms and the "TIC stop" parameter was set to 100,000
intensity/s allowing a maximum accumulation time of 300 ms (i.e. up to three tandem MS
spectra of the same precursor). IMS wave velocity was ramped from 2400 m/s to 450 m/s
(start to end) and the pusher/ion mobility synchronized for singly charged fragment ions in
MS/MS spectra, with up to 85% duty cycle efficiency.

b. UDMS[e]

Data was acquired according to Distler et al. with minor adaptations [32]. Briefly, Two data
functions were acquired over a mass range of m/z 50 to 2000 in alternating mode, differing
only in the collision energy applied to the gas cell. In low-energy MS1 mode, data was
collected at a constant gas cell collision energy of 4 eV. In elevated energy MS2 mode, the
gas cell collision energy was ramped from 10 to 60 eV according to a collision energy look up
table in function of drift time. The spectral acquisition time in each mode was 0.6 s with a
0.015 s interscan delay.

5) Waters Synapt XS (Capilary flow LC)

An M-class LC system (Waters Corporation, Milford, MA) equipped with a 1.7 µm CSH 130
C18 300 µm x 100 mm column, operating at 7 µL/min with a column temperature of 55 °C
was coupled to a Synapt XS quadrupole oa-ToF mass spectrometer (Waters Corporation,
Wilmslow, UK) operating at a mass resolution of 30000, FWHM. The ESI Low Flow probe
capillary voltage was 1.8 kV, sampling cone 30 V, source offset 4 V, source temperature 100
°C, desolvation temperature 300 °C, cone gas disabled, desolvation gas 600 L/hr, and
nebulizer pressure 3.5 bar was used. The time-of-flight (TOF) mass analyzer of the mass
spectrometer was externally calibrated with a NaCsI mixture from m/z 50 to 1990. A lock
mass reference signal of GluFibrinopeptide B (m/z 785.8426) was sampled every two
minutes. Mobile phase A was water containing 0.1% (v/v) FA, while mobile phase B was
ACN containing 0.1% (v/v) FA. The peptides were eluted and separated with a gradient of
5–40% mobile phase B over 120 minutes.

a. HD-DDA

In data-dependent mode, the MS instrument automatically switches between MS survey and
MS/MS scans based upon a set of switching criteria, including ion intensity and charge state.
Full scan MS and MS/MS spectra (m/z 50 - 5000) were acquired in resolution mode. MS
survey spectra were acquired using a fixed acquisition time of 200 ms and the ions present in
each scan were monitored for criteria intensity more than 5000 intensity/sec and 2,3,4+
charge states. Once criteria were satisfied, the precursor ion isolation width of the
quadrupole was set to 1.0 Th around each precursor sequentially. Tandem mass spectra of
up to 15 precursors were generated in the trapping region of the ion mobility cell by using a
collisional energy ramp from 6/9 V (low mass 50 Da, start/end) to up to 147/183 V (high
mass 5000 Da, start/end), with actual values applied dependent upon the precursor m/z. The

MS2 scan time was set to 70 ms and the "TIC stop" parameter was set to 100,000 intensity/s
allowing a maximum accumulation time of 100 ms (i.e. up to two tandem MS spectra of the
same precursor). IMS wave velocity was ramped from 2450 m/s to 550 m/s (start to end) and
the pusher/ion mobility synchronized for singly charged fragment ions in MS/MS spectra, with
up to 85% duty cycle efficiency.

b. SONAR GP

As for Scanning SWATH, the GP fractionation approach which is usually acquired in PAcIFIC
was also analysed by SONAR purely to extend the peptide library. Therefore, the mass scale
from m/z 400 to 1200 was divided into 100 Da sections, thus requiring 8 injections for each
sample. The quadrupole was continuously scanned from the start mass to end mass of each
section and a transmission window of 4 Da was used. In low-energy MS1 mode, data were
collected at constant gas cell collision energy of 6 eV. In elevated energy MS2 mode, the gas
cell collision energy was ramped with values calculated from the start and end mass of the
100 Da mass range being scanned by the quadrupole, and are shown in **Supplementary**
Table 5. The spectral acquisition time in each mode was 0.5 s with a 0.02 s interscan delay.

c. SONAR

The quadrupole was continuously scanned between m/z 400 to 900, with a quadrupole
transmission width of ~20 Da. Two data functions are acquired in an alternating mode,
differing only in the collision energy applied to the gas cell. In low-energy MS1 mode, data
were collected at constant gas cell collision energy of 6 eV. In elevated energy MS2 mode,
the gas cell collision energy was ramped from 16 to 36 eV (per unit charge). The spectral
acquisition time in each mode was 0.5 s with a 0.02 s interscan delay.

356
6) *Bruker TimsTOF Pro (Nano flow LC)*

An Acquity UPLC M-Class System (Waters Corporation) was fitted with a nanoEase[TM] M/Z
Symmetry C18 trap column (100Å, 5 μm, 180 μm x 20 mm) and a nanoEase[TM] M/Z HSS C18
T3 Column (100Å, 1.8 μm, 75 μm x 250 mm, both from Waters Corporation). The sample
was loaded onto the trap column in 2min at 5μl/min in 94% mobile phase A and 6% mobile
phase B. Mobile phase A is UPLC-grade water with 0.1% FA, while mobile phase B is 80%
ACN with 0.1% FA. The Acquity UPLC M-Class system was coupled online to a TimsTOF
Pro via a CaptiveSpray nano-electrospray ion source (Bruker Daltonics, Bremen, Germany),
with an ion transfer capillary temperature at 180°C. Liquid chromatography was performed at
40 °C and with a constant flow of 400 nL/min. Peptides were separated using a linear
gradient of 2–30% mobile phase B over 120 minutes. The TimsTOF Pro elution voltages
were calibrated linearly to obtain reduced ion mobility coefficients (1/K0) using three selected
ions of the Agilent ESI-L Tuning Mix (m/z, 1/K_0: 622.0289 Th, 0.9848 Vs cm^{-2} ; 922.0097 Th,
1.1895 Vs cm^{-2}; 1222.9906 Th, 1.3820 Vs cm^{-2}).

a. PASEF

Parallel Accumulation–Serial Fragmentation DDA (PASEF) was used to select precursor ions
for fragmentation with 1 TIMS-MS scan and 10 PASEF MS/MS scans, as described by Meier
et al. in 2018 [33]. The TIMS-MS survey scan was acquired between 0.70 - 1.45 V.s/cm^2 and
100 - 1700 *m/z* with a ramp time of 100 ms. The 10 PASEF scans contained on average 12
MS/MS scans per PASEF scan with a collision energy of 10 eV. Precursors with 1 – 5
charges were selected with the target value set to 20 000 a.u and intensity threshold to 2500
a.u. Precursors were dynamically excluded for 0.4 min.

b. diaPASEF

The diaPASEF method was implemented as described by Meier et al. in 2019 [19]. The DIA
parameters that define the windows can be found in **Supplementary Table 6**. The DIA range
was set to 400-1200 m/z with 16 diaPASEF scans of 25 m/z isolation windows, including an
overlap of 1 Da. Each diaPASEF scan consisted of two steps (measuring two 25 Da
intervals), with each step spanning an IMS range of 0.3 V.s/cm². The lower IMS value
increased linear from 0.6 to 0.834375 for the diaPASEF scans. The TIMS-MS scan was
identical to the PASEF method.

387
388
389
390
Data Records

Data record 1. The mass spectrometry DDA and DIA-MS proteomics data including
instrument raw files (.wiff, .raw, .d) have been deposited to the ProteomeXchange
Consortium via the PRIDE partner repository with the dataset identifier, PXD028735 [34,35]. For
every instrument, a separate Sample and Data Relationship File (SDRF) and an
Investigation Description File (IDF) have been uploaded to ProteomeXchange. Both the
SDRF and IDF file formats are relatively new in Proteomics and were developed in a
collaboration between EuBIC and the Proteomics Standards Initiative (PSI) [36,37]. These files
are used to annotate the sample metadata and link the metadata to the corresponding data
file(s) and thus will improve the reproducibility and reanalysis of this comprehensive
benchmark dataset.

Data record 2. The AutoQC data analysed in Skyline is available from Panorama Public with
the link https://panoramaweb.org/LFQBenchmark.url.

Technical Validation

We continuously performed system suitability procedures to monitor LC-MS/MS performance
in a longitudinal fashion. Therefore, we ran an AutoQC complex lysate, i.e. a commercial
E.coli protein digest extract (**Supplementary Table 7**), every 3 to 4 samples over all
acquired runs on all instruments. All the AutoQC samples were acquired in DDA on each LC-
MS/MS instrument, except for the Synapt XS, Orbitrap QE-HF and the timsTOF Pro, where
incidentally DDA and DIA acquisitions were alternated. The same mobile phase A and B
composition as for the benchmark samples was used as for the benchmark samples, but the
gradient applied was modified to reduce the time required to acquire the complete sample
batch: linear 3-40% B in 60 minutes, up to 85% B in 2 minutes, isocratic at 85% B for 7
minutes, down to 3% B in 1 minute and isocratic at 3% B for 10 minutes. Note that the
timsTOF Pro AutoQC samples were acquired using the 120min gradient similar to the actual
hybrid proteome samples.

System suitability assessment was performed by monitoring peptide-identification free
metrics (i.e. retention time, peak area, mass accuracy, etc.) extracted with the vendor neutral
Panorama AutoQC framework[38,39]. To isolate a set of peptides that can be used for this,
triplicate AutoQC samples acquired on each instrument were peak picked using MSConvert
(version 3.0.20070) and the corresponding .MGF files were searched against an E.coli
FASTA database using Mascot Daemon (v2.7). The searches were performed with following
settings: (i) 20 ppm peptide mass tolerance, (ii) 50 ppm fragment mass tolerance and (iii) two
allowed missed cleavages. The peptide and protein identification results were exported as
Mascot .DAT file and imported into Skyline Daily (version 21.1.1.160). The five highest
ranked proteins were retained in the target list and after importing one of the AutoQC .raw

files, we manually verified and removed each precursor with co-eluting peptides and low
MS1 signal intensity before a Skyline file was saved as template file. Finally, a configuration
file for each setup was created with the AutoQC Loader software (version 21.1.0.158) which
leads to the automatic import of every sample, with the pattern "AutoQC" in the .raw file or
folder structure, in the Skyline template .The data and skyline reports were published to the
PanoramaWeb folder "U of Ghent Pharma Biotech Lab - LFQBenchmark across Instrument
Platforms" containing six subfolders for each instrumental platform.

For each instrument, peak area, retention time and mass accuracy were manually checked
by plotting these metrics in Levey-Jennings plots. For almost every instrument a few outliers
were detected, as can be expected on a dataset of over 600 LCMS runs. Fortunately, most
of these can be explained by inspecting the raw data and by personal communication with
the technicians acquiring the respective datasets. **Figure 2** illustrates one such case. More
specifically, two AutoQC samples display a near-complete loss in peak area in the
TripleTOF6600+ DDA dataset. Indeed, these were caused by (a) a wrong vial in the sample
tray and (b) an empty vial. When these two samples are removed from the QC plot, a more
coherent perspective on the variation in standard deviation is seen in the Levey-Jennings
plot. Other instances that we have already found include (i) a significant shift in standard
deviation in peak area reported for both the Orbitrap, timsTOF Pro and Synapt XS dataset
because AutoQC samples were incidentally acquired in two different acquisition
methodologies, i.e. in DDA and DIA; (ii) For the timsTOF Pro, a drift in retention time was
seen, indicating LC related technical variation which could have been caused by e.g. too
short column equilibration times; (iii) In the Orbitrap QE HF-X AutoQC data one peptide
(EEAIIK) was undetectable in all the AutoQC samples acquired in AIF. Manual inspection of
the acquisition in Skyline (easily accessible through the Panorama QC pipeline) surfaced that
it fell out of the precursor m/z range (351.7053) that was acquired.

As expected in such a massive MS proteomics experiments, it seems that for every
instrument some outliers were recorded, most of which have explanations common to the
field. Above all, this demonstrates the necessity of a performant system suitability workflow
to increase the reproducibility and quality of LC-MS/MS proteomics datasets[40].

Figure 2. Levey-Jennings plot of the standard deviation in peak area for 50 selected precursors acquired in
DDA with the TripleTOF6600+. The upper chart shows two distinct outliers, acquired respectively on the 2[nd] and
12[th] of December (red boxes). Manual inspection of the data shows these were caused by **(a)** a wrong vial in the

462 sample tray and **(b)** an empty vial. When these two samples are excluded from the Levey-Jennings plot (lower
463 chart), a significant drop in standard deviation over the time period of data acquisition is seen.

Usage Notes

A comprehensive dataset such as the one presented here is inspired by the bioinformatics
need to cope with the recent expansion of novel acquisition strategies and data dimensions.
Apart from being a repository for validating both performance and compatibility of (new)
bioinformatic pipelines, it can serve as a reference for general proteomics courses (e.g.
Skyline tutorials, SWATH/DIA course) and be applied for training and validating
machine/deep learning algorithms. As such, it is intended to facilitate our understanding of
the impact of instrumentation on the perspective that is generated on protein biology and to a
larger extent to help unify the field of proteomics.

To demonstrate the applicability of this data repository, we assessed the impact of
instrumentation on the most conventional data format acquired by all instruments i.e. DDA.
Therefore, for every instrumental platform, triplicate Human, Yeast and E.coli DDA runs were
peak picked with MSConvert (version 3.0.21285) and exported as a Mascot .MGF file
(**Figure 3A**). MS Convert software was chosen as it is vendor-independent and contains
each vendor's implementation for peak picking, with the exception of UNIFI i.e. Waters.
Therefore, Progenesis QI for Proteomics (version 4.2.7207) was used for the Waters DDA
data. By doing so, the MS1 precursor space was aligned in the retention time dimension
before peak picking. Subsequently, all MS/MS spectra were exported as .MGF file for
peptide identification. A standard search with carbamidomethylation of cysteine as fixed
modification was performed using a database containing the Human, Yeast and E.coli
protein sequences (downloaded from Uniprot on 19/01/2021) and with following parameters:
i) mass error tolerances for the precursor ions and the fragment ions were set at 20 ppm and
50 ppm, respectively; ii) enzyme specificity was set to trypsin, allowing up to one missed
cleavage. Next, the results were exported as .DAT file and imported into Skyline to create a
non-redundant spectral library with BiblioSpec. Afterwards, the .BLIB file was converted to a
.dlib and .msp file format respectively using EncyclopeDIA [29]. The resulting .msp file was
converted using a Python conversion tool (speclib_to_mgf.py) built-in MS²PIP, to create a
peptide record file (.PEPREC) and .MGF file. Next the proportion (amount) of peptide
identifications overlapping between the different instruments was assessed using a custom
Python script (**Figure 3B**).

Since each instrument analysed the same commercial protein digests, a large overlap in
peptide identifications would be expected. However, while the timsTOF Pro, TripleTOF
6600+ and Orbitrap QE HF-X roughly identify a similar number of peptide sequences
(approximately 40,000 over nine LCMS runs), the overlap in identified sequences is in the
order of 50%, with that between the Orbitrap QE HF-X and the TripleTOF5600 and 6600+
being overall 10% higher than the overlap with the timsTOF Pro. While at the protein level a
lot more coherency is found, it is still striking how the Orbitrap QE and especially the
timsTOF Pro have significantly more proteins from a similar number of annotated peptides
(**Figure 3C**).

A benchmark like the one presented here however, allows to investigate potential assertions
attributed to different stages of the workflow.

Acquisition. Even with DDA, the acquisition parameters can be modified substantially. A
first, crucial difference is the used flow rate of the LC system: the TripleTOF 6600+ was
coupled to a 5µl/min flow, while the Orbitrap and timsTOF Pro used nanoflow (350 and 400
nL/min, respectively). While nano flow is more sensitive, it is also more fragile, which is
reflected in some of the quality control samples uploaded to Panorama QC. A second

important parameter is dynamic exclusion, i.e. how long a certain precursor is excluded from
fragmentation to avoid redundancy. As found in the methods section, this was set to 10s, 24s
and 30s for the TripleTOF 6600+, the timsTOF Pro and the Orbitrap QE HF-X, respectively.
This parameter usually is a function of LC gradient length and of consistent MSMS spectral
quality, which can be facilitated through e.g. automatic gain control (AUC) in orbitrap
instruments, but not in QTOF designs. As a result, the orbitrap sampled the peptides most
efficiently, with 20% of the peptides only sampled once in a run, i.e. three times over triplicate
runs of the same proteome (**Figure 3D**). A third aspect of acquisition is the addition of ion
mobility separation (IMS) capabilities in Waters and Bruker instruments. Importantly, they
both use IMS in a very different way, forcing a deeper understanding of instrumental
architecture and how ion mobility is applied by both vendors in DDA relative to the
quadrupole (Q) selection and the collision-induced dissociation (CID). Briefly, the order of ion
manipulation is Q-CID-IMS for the so-called High Definition DDA (HD-DDA) in the Waters
series and it is IMS-Q-CID in the Bruker instruments. Therefore, Waters separates the
fragment ions in IMS and leverages the efficient charge state separation to synchronize the
pusher of the TOF tube with singly charged fragment ions in order to only and nearly one
hundred percent efficiently sample the single charged fragments [7]. This can be expected to
impact the fragment intensities, especially when detector saturation in MSMS occurs. In the
timsTOF Pro on the other hand, the IMS is in fact resolving the precursor ion space before
fragmentation, leading to a different selection of the peptide precursor space compared to
other devices that do not use IMS (in this way).

Raw data. Next, the differences in the fragmentation process between the instruments (all of
which are beam-type CID) can be assessed by mutually mapping the MS2 intensities. **Figure
3E** shows their median Pearson correlation coefficients (PCC). By definition, this only plots
commonly found peptides. The largest differences in fragment intensities are found between
timsTOF Pro and Orbitrap QE HF-X, potentially underlying part of the differences seen in
peptide and protein identification. Importantly, the recent introduction of machine learning
algorithms has illustrated the potential of fragment intensity prediction for proteomics [24,25,41–
43]. Therefore, to attain an even deeper understanding of these fragmentation differences, we
calculated the PCC of both b- and y-ions compared to two prediction models (i.e. HCD and
TTOF 5600) from MS²PIP [44] (**Figure 3F**). This confirms that the orbitrap and triple TOF
designs have a very similar fragmentation pattern. Note that the normalized collision energy
was not taken into account as input feature [25,45,46]. Additionally, the Synapt and timsTOF
instruments are sampling the ion beam more efficiently, through ion mobility (Waters) or
trapping and pusher design (Bruker), potentially causing more frequent detector saturation in
the fragment space, leading to skewed intensity patterns compared to the predicted spectra.
Indeed, augmented gain controlled in orbitrap devices can efficiently avoid this.

Sample Complexity. We also acquired the proteomes as mixtures in known ratios, and
selected a triplicate series of Condition A and B for a subsequent analysis (**Figure 3G**). This
higher sample complexity is considerably more challenging for DDA acquisition, because the
instrument needs to do a faster sampling of the precursor ion space. To create a more
comprehensive picture of the resulting data, we turned to AlphaPept, an alternative
annotation algorithm that also quantifies the number of features in MS1, i.e. potential peptide
species in the sample [47]. Strikingly, while the TripleTOF 6600+ generated the most features
(330k compared to 260k and 220k for the Orbitrap QE HF-X and the timsTOF Pro), it
annotated considerably less peptides as well as proteins compared to the two other
instruments (**Figure 3G**). Still, the peptide/protein bias persists for the Orbitrap and timsTOF
Pro comparison. Importantly, AlphaPept support for Sciex was only preliminarily
implemented through mzML preprocessing, potentially underlying the lower annotation rate

and making it a perfect example of how the benchmark can support the development of
compatibility in the future.

Data analysis: Quantification. Next, we used the benchmark design to verify to what extent
the annotated peptides and proteins were found in their respective ratios (**Figure 1**). Apart
from being a benchmark target for quantification algorithms, this ground truth can also be
considered a very rough empirical estimation of the false discovery rate (FDR)[1]. At the
peptide level the spread of especially Human peptides into other ratios is increasingly visible
in the order TripleTOF 6600+ < Orbitrap QE HF < timsTOF Pro, which alludes to increasing
misannotation (**Figure 3G**). While ratio estimation is known to become compromised with
decreasing intensities of the MS1 features selected, the relative position of these peptides on
the y-axis argues against this intensity explanation. Unfortunately, the exported ion counts of
the timsTOF Pro differ at least 1-2 orders of magnitude compared to both the Orbitrap and
the TripleTOF 6600+ and more advanced (conversion) algorithms are needed to compare
these. However, at the protein level there was no overall skewing in the AlphaPept results
that would point towards a significant increase in FDR in any of the instruments.

In conclusion, considerable bioinformatics challenges remain to be tackled before the
intricacies of the data from contempory instrumentation and their acquisition parameters can
truly be defined.

Figure 3. Comparing the DDA data of six different instruments. Experimental design (A) Triplicate measurements of three individual proteomes. (B) The overlap in uniquely identified peptide sequences and (C) proteins between the six instruments. (D) Number of PSMs per peptide identification throughout nine DDA runs on three different proteomes for three instruments. (E) Pearson Correlation Coefficient (PCC) of the fragment intensities were calculated between the shared identified peptides from the DDA replicates between each instrument. The numbers in each box correspond to the median spectrum PCC between the instrument on the x-axis and the instrument on the y-axis. Dark blue color indicates a higher degree of overlap or higher median PCC. (F) Boxplots of the Pearson correlation coefficients (PCC) between the MS²PIP predicted (HCD and TTOF5600 model) and experimental fragment intensities across the six different LC-MS instruments. (G) The benchmark design of mixed proteomes for three instruments as annotated and quantified using AlphaPept. Here, triplicate runs of Condition A and Condition B were used , resulting in the six bars depicted in the middle, respectively representing the number of MS1 features, the number of identified peptides and the number of identified proteins for each instrument. The log-fold plots to the left depict the distribution of the peptide ratios in the x-axis as a function of their intensity in the y-axis; protein log fold changes are depicted to the right.

In conclusion, it is clear from this preliminary data usage case on the simple and most commonly applied DDA strategies that a lot of insights on data structure and bias were still hidden. Especially in light of the more recent DIA strategies, which were not analysed in this Scientific Data report, a lot of work needs to be done before the field truly understands how data acquisition and data analysis effectively perform. Importantly, each step in the data processing can greatly impact the final outcome and we especially anticipate a renewed interest in (multidimensional) peak picking algorithms as a currently underappreciated preprocessing step of the complex DIA data [1]. Additionally, recent annotation tools have doubled the number of peptides that can be extracted from the same DIA data, an asset not expected for novel DDA annotation algorithms [21,28]. We would therefore like to invite the developers of all current bioinformatics tools to make use of this comprehensive dataset to benchmark their performance for each instrument individually and adapt their algorithms to increase the performance on all. This dataset was compiled from different labs and therefore also captures differences in instrument usage peculiar to the field, yet it all can be assessed in the Panorama QC metrics that are also publically available. Especially a better understanding of the detectable and annotatable ion space will move the field forward and help researchers make informed decisions on the best acquisition strategies for their application or biological question under investigation.

Code Availability

MS²PIP is open source, licensed under the Apache-2.0 License, and hosted on https://github.com/compomics/ms2pip_c. The Jupyter notebooks used to generate **Figure 3 (B,C,E and F)** are available through Zenodo, under DOI: 10.5281/zenodo.5714380.

Acknowledgements

This research was funded by grants from the Research Foundation Flanders (FWO) awarded to BVP (grant number: 11B4518N), RG (1S50918N), and MD (12E9716N). Hans Vissers is acknowledged for his assistance with data conversion and formatting. This work was supported in part by the French Ministry of Research with the Investissement d'Avenir Infrastructures Nationales en Biologie et Santé program (ProFi, Proteomics French Infrastructure project; ANR-10-INBS-08).

Author contributions

BVP, SW, SD and MD conceived the study, BVP performed the TripleTOF 5600 and 6600+ data acquisition, SD performed the Synapt G2-Si data acquisition, AGP, DB, EM and KC performed the Orbitrap data acquisition, KB performed the timsTOF Pro data acquisition. CH and LG performed the Synapt XS data acquisition. BVP and SD prepared the samples and RG and SW wrote the scripts to generate Figure 3. YPR organised the ProteomeXchange

627 submission. NB and ST helped us to set up the acquisition methodologies on the TripleTOF
628 6600+ and provided valuable insights on the differences seen for each instrument. BVP and
629 MD wrote the draft manuscript with contributions from all authors. MD supervised and DD
630 and LM co-supervised the experiment.

Competing interests

632 Chris Hughes and Lee Gethings are employed by Waters Corporation. Nic Bloomfield and
633 Stephen Tate are employed by Sciex.

References

1. Willems, S. *et al.* Ion-networks: A sparse data format capturing full data integrity of
data independent acquisition mass spectrometry. *bioRxiv* (2019) doi:10.1101/726273.

2. Vaudel, M. *et al.* PeptideShaker enables reanalysis of MS-derived proteomics data
sets. *Nat Biotechnol* **33**, 22–24 (2015).

3. Verheggen, K., Martens, L., Berven, F. S., Barsnes, H. & Vaudel, M. Database Search
Engines: Paradigms, Challenges and Solutions. in *Advances in Experimental Medicine*
and Biology 147–156 (2016). doi:10.1007/978-3-319-41448-5_6.

4. Geromanos, S. J., Hughes, C., Ciavarini, S., Vissers, J. P. C. & Langridge, J. I. Using
ion purity scores for enhancing quantitative accuracy and precision in complex
proteomics samples. *Anal. Bioanal. Chem.* **404**, 1127–1139 (2012).

5. Richardson, K. *et al.* A probabilistic framework for peptide and protein quantification
from data-dependent and data-independent LC-MS proteomics experiments. *OMICS*
16, 468–482 (2012).

6. Li, G. Z. *et al.* Database searching and accounting of multiplexed precursor and
product ion spectra from the data independent analysis of simple and complex peptide
mixtures. *Proteomics* **9**, 1696–1719 (2009).

7. Helm, D. *et al.* Ion Mobility Tandem Mass Spectrometry Enhances Performance of
Bottom-up Proteomics. *Mol. Cell. Proteomics* **13**, 3709–3715 (2014).

8. Shliaha, P. V, Bond, N. J., Gatto, L. & Lilley, K. S. Effects of traveling wave ion mobility
separation on data independent acquisition in proteomics studies. *J Proteome Res* **12**,
2323–2339 (2013).

9. Gillet, L. C. *et al.* Targeted data extraction of the MS/MS spectra generated by data-
independent acquisition: a new concept for consistent and accurate proteome
analysis. *Mol Cell Proteomics* **11**, O111 016717 (2012).

10. Ludwig, C. *et al.* Data□independent acquisition□based SWATH□MS for quantitative
proteomics: a tutorial. *Mol. Syst. Biol.* (2018) doi:10.15252/msb.20178126.

11. Ting, Y. S. *et al.* Peptide-Centric Proteome Analysis: An Alternative Strategy for the
Analysis of Tandem Mass Spectrometry Data. *Mol. Cell. Proteomics* **14**, 2301–7
(2015).

12. Li, Y. *et al.* Group-DIA: analyzing multiple data-independent acquisition mass
spectrometry data files. *Nat. Methods* **12**, 1105–1106 (2015).

13. Kuharev, J., Navarro, P., Distler, U., Jahn, O. & Tenzer, S. In-depth evaluation of
software tools for data-independent acquisition based label-free quantification.
Proteomics **15**, 3140–3151 (2015).

14. Teleman, J. *et al.* DIANA-algorithmic improvements for analysis of data-independent

acquisition MS data. *Bioinformatics* (2015) doi:10.1093/bioinformatics/btu686.

15. Peckner, R. *et al.* Specter: linear deconvolution for targeted analysis of data-
independent acquisition mass spectrometry proteomics. *Nat. Methods* **15**, 371–378
(2018).

16. Wang, J. *et al.* MSPLIT-DIA: sensitive peptide identification for data-independent
acquisition. *Nat. Methods* **12**, 1106–1108 (2015).

17. Moseley, M. A. *et al.* Scanning Quadrupole Data-Independent Acquisition, Part A:
Qualitative and Quantitative Characterization. *J. Proteome Res.* **17**, 770–779 (2018).

18. Messner, C. B. *et al.* Ultra-fast proteomics with Scanning SWATH. *Nat. Biotechnol.* **39**,
846–854 (2021).

19. Meier, F. *et al.* diaPASEF: parallel accumulation–serial fragmentation combined with
data-independent acquisition. *Nat. Methods 2020 1712* **17**, 1229–1236 (2020).

20. Van Puyvelde, B. *et al.* Removing the Hidden Data Dependency of DIA with Predicted
Spectral Libraries. *Proteomics* **20**, e1900306 (2020).

21. Demichev, V., Messner, C. B., Vernardis, S. I., Lilley, K. S. & Ralser, M. DIA-NN:
neural networks and interference correction enable deep proteome coverage in high
throughput. *Nat. Methods 2019 171* **17**, 41–44 (2019).

22. Bouwmeester, R., Gabriels, R., Van Den Bossche, T., Martens, L. & Degroeve, S. The
Age of Data-Driven Proteomics: How Machine Learning Enables Novel Workflows.
Proteomics (2020) doi:10.1002/pmic.201900351.

23. C. Silva, A. S., Bouwmeester, R., Martens, L. & Degroeve, S. Accurate peptide
fragmentation predictions allow data driven approaches to replace and improve upon
proteomics search engine scoring functions. *Bioinformatics* **35**, 5243–5248 (2019).

24. Zhou, X. X. *et al.* PDeep: Predicting MS/MS Spectra of Peptides with Deep Learning.
Anal. Chem. (2017) doi:10.1021/acs.analchem.7b02566.

25. Gessulat, S. *et al.* Prosit: proteome-wide prediction of peptide tandem mass spectra
by deep learning. *Nat. Methods* (2019) doi:10.1038/s41592-019-0426-7.

26. Bouwmeester, R., Gabriels, R., Hulstaert, N., Martens, L. & Degroeve, S. DeepLC can
predict retention times for peptides that carry as-yet unseen modifications. *Nat.*
Methods 2021 1811 **18**, 1363–1369 (2021).

27. Mann, M., Kumar, C., Zeng, W.-F. & Strauss, M. T. Artificial intelligence for proteomics
and biomarker discovery. *Cell Syst.* **12**, 759–770 (2021).

28. Navarro, P. *et al.* A multicenter study benchmarks software tools for label-free
proteome quantification. *Nat. Biotechnol.* **34**, 1130–1136 (2016).

29. Searle, B. C. *et al.* Chromatogram libraries improve peptide detection and
quantification by data independent acquisition mass spectrometry. *Nat. Commun.* **9**,
5128 (2018).

30. Improved Data Quality Using Variable Q1 Window Widths in SWATH ® Acquisition
Data Independent Acquisition on TripleTOF ® and X-Series QTOF Systems. (2019).

31. Pino, L. K., Just, S. C., MacCoss, M. J. & Searle, B. C. Acquiring and Analyzing Data
Independent Acquisition Proteomics Experiments without Spectrum Libraries. *Mol.*
Cell. Proteomics **19**, 1088–1103 (2020).

32. Distler, U. *et al.* Drift time-specific collision energies enable deep-coverage data-

independent acquisition proteomics. *Nat. Methods* **11**, 167–170 (2014).

33.   F, M. *et al.* Online Parallel Accumulation-Serial Fragmentation (PASEF) with a Novel
Trapped Ion Mobility Mass Spectrometer. *Mol. Cell. Proteomics* **17**, 2534–2545
(2018).

34.   JA, V. *et al.* 2016 update of the PRIDE database and its related tools. *Nucleic Acids
Res.* **44**, D447–D456 (2016).

35.   Perez-Riverol, Y. *et al.* The PRIDE database and related tools and resources in 2019:
Improving support for quantification data. *Nucleic Acids Res.* (2019)
doi:10.1093/nar/gky1106.

36.   Dai, C. *et al.* A proteomics sample metadata representation for multiomics integration
and big data analysis. *Nat. Commun. 2021 121* **12**, 1–8 (2021).

37.   Bittremieux, W. *et al.* The European Bioinformatics Community for Mass Spectrometry
(EuBIC-MS): an open community for bioinformatics training and research. *Rapid
Commun. Mass Spectrom.* e9087 (2021) doi:10.1002/RCM.9087.

38.   Bereman, M. S. *et al.* An Automated Pipeline to Monitor System Performance in Liquid
Chromatography-Tandem Mass Spectrometry Proteomic Experiments. *J Proteome
Res* **15**, 4763–4769 (2016).

39.   Sharma, V. *et al.* Panorama: A Targeted Proteomics Knowledge Base. *J. Proteome
Res.* **13**, 4205–4210 (2014).

40.   Bereman, M. S. Tools for monitoring system suitability in LC MS/MS centric proteomic
experiments. *Proteomics* **15**, 891–902 (2015).

41.   Degroeve, S. & Martens, L. MS2PIP: a tool for MS/MS peak intensity prediction.
Bioinformatics **29**, 3199–3203 (2013).

42.   Tiwary, S. *et al.* High-quality MS/MS spectrum prediction for data-dependent and data-
independent acquisition data analysis. *Nat. Methods 2019 166* **16**, 519–525 (2019).

43.   Yang, Y. *et al.* In silico spectral libraries by deep learning facilitate data-independent
acquisition proteomics. *Nat. Commun. 2020 111* **11**, 1–11 (2020).

44.   Gabriels, R., Martens, L. & Degroeve, S. Updated MS^{2}PIP web server delivers fast
and accurate MS^2 peak intensity prediction for multiple fragmentation methods,
instruments and labeling techniques. *Nucleic Acids Res.* **47**, W295–W299 (2019).

45.   JK, D., AF, P. & JR, Y. Energy dependence of HCD on peptide fragmentation: stepped
collisional energy finds the sweet spot. *J. Am. Soc. Mass Spectrom.* **24**, 1690–1699
(2013).

46.   Tarn, C. & Zeng, W.-F. pDeep3: Toward More Accurate Spectrum Prediction with Fast
Few-Shot Learning. *Anal. Chem.* **93**, 5815–5822 (2021).

47.   Strauss, M. T. *et al.* AlphaPept, a modern and open framework for MS-based
proteomics. *bioRxiv* 2021.07.23.453379 (2021) doi:10.1101/2021.07.23.453379.

750