FPCountR: Absolute protein quantification using fluorescence measurements

Eszter Csibra and Guy-Bart Stan

Department of Bioengineering, Imperial College Centre for Synthetic Biology (IC-CSynB),

Imperial College London, London, UK, SW7 2AY

Correspondence:
Eszter Csibra (e.csibra@imperial.ac.uk) and Guy-Bart Stan (g.stan@imperial.ac.uk)

Keywords: absolute quantification, calibration, engineering biology, fluorescent proteins,
microplate reader, synthetic biology

13
Abbreviations: BCA, bicinchoninic acid; BSA, bovine serum albumin; EC, extinction
coefficient; EDTA, ethylenediaminetetraacetic acid; FP, fluorescent protein, GFP, green
fluorescent protein, MEFL, molecules of equivalent fluorescein, MEFP, molecules of
equivalent fluorescent protein, OD, optical density; PEMS, particles of equivalent
microspheres, SDS-PAGE, sodium dodecyl sulphate polyacrylamide gel electrophoresis;
SEVA, Standardised European Vector Architecture
20

Abstract

This paper presents a generalisable method for the calibration of fluorescence readings on microplate readers, in order to convert arbitrary fluorescence units into absolute units. FPCountR relies on the generation of bespoke fluorescent protein (FP) calibrants, assays to determine protein concentration and activity, and a corresponding analytical workflow. We systematically characterise the assay protocols for accuracy, sensitivity and simplicity, and describe a novel 'ECmax' assay that outperforms the others and even enables accurate calibration without requiring the purification of FPs. To obtain cellular protein concentrations, we consider methods for the conversion of optical density to either cell counts or alternatively to cell volumes, as well as examining how cells can interfere with protein counting via fluorescence quenching, which we quantify and correct for the first time. Calibration across different instruments, disparate filter sets and mismatched gains is demonstrated to yield equivalent results. It also reveals that mCherry absorption at 600nm does not confound cell density measurements unless expressed to over 100,000 proteins per cell. FPCountR is presented as pair of open access tools (protocol and R package) to enable the community to use this method, and ultimately to facilitate the quantitative characterisation of synthetic microbial circuits.

Introduction

There is a growing awareness that tackling the challenge of synthetic circuit design requires the synthesis of empirical characterisation data on genetic parts with mathematical modelling approaches for predicting and realising desired behaviours (1–3). However, there are numerous challenges in integrating experimental data with quantitative frameworks, as experimental data is typically acquired in relative or arbitrary units (specific to instruments and their respective settings), which cannot be converted into useful units and therefore limits our ability to make comparisons between experiments and models.

Fluorescent proteins (FPs) are our most versatile tools for the assessment of synthetic genetic element performance. Since their discovery, FPs have rightly been recognised as uniquely valuable reporter proteins for quantitative characterisation (4,5), since they do not require the addition of exogenous components to fluoresce. This makes their use easy and cost effective. Various laboratory instruments allow the characterisation of fluorescent systems in a wide range of dimensions and scales – through direct visualisation (using fluorescence microscopy), via single-cell fluorescence analysis (using flow cytometry), or via timecourse kinetic data acquisition (using microplate readers).

The 'protein quantification problem' of reporting GFP levels acquired by such instruments in 'relative fluorescent units' (RFU) has been widely recognised in synthetic biology (6). This recognition has led to the adoption of calibration standards such as fluorescein, a small molecular fluorophore with similar excitation and emission characteristics to GFP. Fluorescein can be used to calibrate a given instrument by converting the instrument's

arbitrary RFU output into units of 'molecules of equivalent fluorescein' (MEFL). This

technique has been demonstrated to enable the comparison of GFP expression data

gathered from different instruments as well as across laboratories (6,7).

While it is now approaching mainstream usage in synthetic biology, the conversion of green

fluorescence values into MEFL is arguably not the most important type of quantification

required for building synthetic circuits. Three aspects of the protein quantification problem

remain elusive. First, fluorescein is only a good calibrant for green fluorescent proteins,

leaving blue, yellow, orange and red FPs uncalibrated. Second, fluorescein can only provide

a conversion to units of fluorescein, whereas what is actually needed is a conversion to units

of protein. Currently, most experiments cannot even reveal the order of magnitude at which

FPs are being expressed (ie. 10 vs 100,000 molecules per cell) – in contrast to what is

possible with RNA sequencing (8). Third, while fluorescein can allow the comparison of GFP

levels between instruments and laboratories, it cannot address the comparison between

two different FPs in the same circuit. This is only possible if RFUs from both FPs can be

separately converted into molecular units of protein. Some have attempted to tackle this by

attempting to predict the relative brightness of FPs using theoretical values (9,10), but such

calibrations make a number of assumptions, for instance about translation rate equivalence

across constructs, that require validation before they can be adopted.

Fortunately, there is a reasonably simple solution. The ideal calibrant in molecular biology is

considered to be a purified sample of the molecule to be measured – in this case, the

fluorescent protein itself. While purified FPs are not commercially available, they can be

produced 'in-house', thereby providing that crucial direct link between relative fluorescence

units and molecules of protein. Indeed, FP calibration has been proposed in the past,

though its use has been limited to microscopy (11–13) and remains rare for microplate

assays (14,15). We suspect this is due to (a) an underappreciation that absolute

quantification is possible without 'omics', (b) a reticence to try unfamiliar biochemical

protocols that are not usually part of the synthetic biology or microbiology repertoire, and

(c) doubt that such protocols could be accurate or sensitive enough for general usage. These

motivated us to develop a general, yet simple-to-use calibration protocol of this nature.

In what follows, we outline our optimised calibration method and present it as a pair of

resources: a wet lab protocol, called FPCount (available on protocols.io) and an

accompanying analysis package, called FPCountR (available on GitHub). We present data

showing the development of this protocol, and systematically characterise the biochemical

and analytical requirements of an accurate and sensitive calibration. We also present a

'hack', which acts both to simplify the method to remove the requirement for protein

purification, and to make it more sensitive and robust. Using FPCountR, we show that

conversion to molecular units can be used to calibrate across different instruments,

disparate filter sets and mismatched gains to yield equivalent results. We demonstrate that

conversion to absolute units allows the user of our method to compare the protein

production efficiency of different fluorescent proteins expressed from an otherwise

identical vector in molecules per cell, or as a molar concentration. Finally, we demonstrate

that this method can be used to quantitatively evaluate the experimental protocols

themselves, such as the extent to which red fluorescent proteins confound optical density

readings in timecourse assays.

Results

Our aim for this work was to develop a generalisable method for FP calibration that could be used by any group wishing to calibrate fluorescence readings on microplate readers to molecular units. To do this, we defined a number of key aims for our proposed method. First, it should be accurate and sensitive, as we need the method to correctly estimate molecule numbers within cells, and as protein yields from small-scale purifications are typically modest. Second, the calibration protocol should be as simple as possible and adapted ideally such that each respective assay may be carried out in 96-microwell plate format using the same plate reader that is being calibrated. This way, multiple fluorescent proteins may be calibrated at once, and end users do not require any additional instrumentation. Third, the method should be suitable for the particular characteristics of fluorescent proteins. These proteins are smaller and structurally distinct from typical protein calibrants such as bovine serum albumin (BSA), and are known to present certain challenges for quantification due to light absorption by their chromophores. Thus, any assay developed for non-fluorescent proteins requires a separate validation on FPs to demonstrate that they are also adequate for this class of proteins. Finally, we wanted to enable the easy analysis of the data, by (i) enabling easy conversion of raw calibration data into a conversion factor that links the arbitrary fluorescence output of a protein with its quantity in molecular units, and (ii) allowing easy conversion of data from all future timecourse data from that instrument to produce outputs in (e.g. GFP) molecules, rather than relative (e.g. green) fluorescence units. An overview of the FPCount fluorescent protein calibration protocol is illustrated in Fig. 1.

Design of purification protocol for obtaining protein calibrants

In order to obtain our fluorescent protein calibrants, they must first be produced by overexpression, and purified. Protein purification methodologies can often be highly complex, requiring specialist expertise and instrumentation. To alleviate this, our protocol was explicitly designed to be as straightforward as possible, involving only the minimum number of required steps, and using commonly available reagents. It was also designed to be amenable to small-volume purifications to enable the calibration of multiple proteins in parallel. The protocol is summarised in Fig. 2A. A standardised FP expression vector was constructed from an arabinose-inducible His-tagged FP construct in a high-copy SEVA vector (Fig. 2B). Overnight expression was used to maximise protein production, and cells were lysed using sonication to avoid the requirement to add chemical components that may interfere with downstream processes, such as EDTA (with His-tag purification) or detergents (with protein quantification). Insoluble proteins were removed via centrifugation and proteins were purified using cobalt resin to minimise the co-isolation of impurities. Expression yields and solubility were checked by SDS-PAGE (Fig. 2C) and the quality of purified FP calibrants was verified by SDS-PAGE and fluorescence excitation and emission scanning (Fig. 2D-E, Supplementary Fig. 1). Purified calibrants were of consistently good purity and yield.

Conducting a plate reader calibration

The calibration of plate readers with fluorescein has traditionally been conducted using a dilution series of known concentrations of fluorescein, subjected to a fluorescence assay in the plate reader whose calibration is desired (measurement of relative fluorescence units, RFU). The results are used to relate fluorescein molecule number to RFU to obtain a

conversion factor, which can in turn be used to convert RFU readouts from experimental

data into MEFL units (6). For protein calibrants, one additional step is required: protein

concentration determination.

In our initial protocols, we opted for the bicinchoninic acid (BCA) assay due to its sensitivity,

ease of use and low protein-to-protein variability (16). In addition, microplate-optimised

reagents for a 'microBCA' were available from ThermoFisher, with excellent reported

sensitivities (to 2 ng/µl). Pilot tests showed an inhibitory effect of the Tris and NaCl in the

elution buffer (Supplementary Fig. 2A) suggesting that buffer exchange would be necessary

for high assay sensitivity. Normalised measurements using a BSA standard obtained with the

kit were fitted to a polynomial equation to obtain a standard curve (Supplementary Fig. 2D,

ii). This was then used to predict the concentrations of FPs prepared as serial dilutions by

first removing values under the reported threshold of sensitivity and fitting a linear model

through the rest of the values (Supplementary Fig. 2D, ii and iii). An extra step was added to

the recommended protocol (Supplementary Fig. 2A, ii and Supplementary Fig. 2C) to

account for baseline absorbance of red FPs in the A562 range. These calculations are

handled in the FPCountR package by the get_conc_bca() function. Using the resultant

predicted protein concentrations and fluorescence assay data on the same FP dilution

series, an adapted version of the generate_cfs() function from flopR (7) was used to

generate conversion factors (RFU/molecule) for mCherry in a Tecan Spark plate reader for

the red FP-typical filter set (ex 560/20, em 620/20; Supplementary Fig. 3).

Development of a second assay for FP concentration – the A280 assay

We sought to verify the accuracy of the BCA assay by re-quantifying our FPs with a second

method that is likely to give reliable concentration estimates. While a wide variety of

protein assays exist, the only widely-used 'absolute' assay that does not require a calibrant

is the A280 assay. As the name suggests, it quantifies protein concentration via light

absorbance at 280 nm, where three amino acid residues are known to absorb light in a way

that has been shown to be approximately additive (17). This means that a reasonable

prediction of light absorbance at 280 nm can be made for any pure protein of known

primary sequence by way of an extinction coefficient (EC; expected light absorption for a

given concentration of protein). As sample absorbance relates to molecular concentration

according to Beer's law, i.e., $A = EC * C * L$, (where A is the absorbance, EC is the

extinction coefficient ($M^{-1}cm^{-1}$), C is the concentration (M), and L is the path length (cm)),

the protein concentration may be calculated from absorbance using only the extinction

coefficient and the path length. The most common formats for A280 measurements are

laborious, single-throughput cuvette- or Nanodrop-type measurements, requiring the

adaptation of the standard A280 protocols for use in 96-well microplates. We have

summarised the requirements for such an adaptation in Supplementary Note 1 and

Supplementary Fig. 4-8. In brief, the best results for A280 assays were obtained by using UV-

clear plastic, removing additives, correcting for path length variation, and correcting for light

scatter. This required the collection of an absorbance spectrum from 200-1000 nm rather

than just one reading at 280 nm, and is processed by two consecutive FPCountR functions,

plot_absorbance_spectrum() and get_conc_a280() (Supplementary Fig. 6-7).

Systematic testing of the calibration protocol allows method validation

We sought to conduct a systematic assessment of the BCA and A280 methods by testing

three spectrally distinct FPs in two buffers, assessed with both assays in parallel (Fig. 3). The

chosen FPs (mTagBFP2, mGFPmut3 and mCherry) are widely used, monomeric, reasonably

fast-maturing and bright. All three are almost identical on the protein level to their FPbase

entries (Fig. 2B; Materials & Methods; Supplementary Data), with the exception that they all

have a His$_6$ tag N-terminal extension, and mGFPmut3 includes a well-defined monomeric

A206K mutation (18,19). The chosen buffers (T5N15 (5mM Tris-HCl pH 7.5, 15 mM NaCl)

and T5N15 with protease inhibitors) were both compatible with the microBCA assay

(Supplementary Fig. 2B), however, pilot studies suggested they might have different effects

on the A280 assay. Following purification, FPs were initially dialysed to remove additives,

then re-dialysed into the respective assay buffer (Fig. 3A). Each FP:buffer combination was

then serially diluted, and subjected to an absorbance scan (200-1000 nm measurement, for

the A280), a fluorescence assay (fluorescence measurement with appropriate filters for

each FP) and a microBCA assay (reagent addition, incubation and A562 measurement). The

results of this comparative test are shown in Fig. 3B-C, Supplementary Fig. 9 and

Supplementary Tables 1-2. Broadly, the results from each assay validate those of the other

assay: the measured concentration of each FP using the microBCA and A280 assays are

within 2-fold of each other for most samples (Fig. 3C) and apparent linear ranges reach 1

ng/μl for most dilution series (Fig. 3B). For comparison, the reported sensitivity on the

Nanodrop is 100 ng/μl (20). We observed some buffer sensitivity for both assays – the

microBCA produces more linear results in the buffer containing protease inhibitors, whereas

the A280 does better in the buffer without them. Buffer effects were also apparent for the

fluorescence assay (Supplementary Note 2, Supplementary Fig. 10). Conversion factors

obtained from different purification batches gave similar estimates where concentration

estimates were made using optimal assay:buffer pairings, varying between 1.02 and 1.63-

fold (Supplementary Fig. 9B).

FPbase-enabled ECmax method performs better than conventional assays

We trialled a third protein assay during this experiment, designated here as the 'ECmax'

method. The principal idea behind this assay is that the A280 extinction coefficient is not

the only known extinction coefficient for FPs. FPs also possess an extinction coefficient ('EC')

corresponding to their light absorption at their peak ('max') excitation wavelength. As the

'ECmax' of most FPs is available on FPbase, we can automate its retrieval using the FPbase

API (21). The analytical processing steps for the ECmax assay (in get_conc_ecmax()) are

similar to those of the A280 assay (Supplementary Fig. 8), and require no further readings.

As the maximal absorbance peaks for all FPs tested were higher than those at 280 nm

(Supplementary Fig. 1B, Supplementary Table 3), we anticipated that the ECmax assay

would be more sensitive. Further, as protease inhibitors absorb at wavelengths under 300

nm, we hypothesised this assay may be less buffer-sensitive. Compellingly, both appear to

be true: we consistently found that the ECmax assay produced larger linear ranges and

lower limits of detection than the other assays (approaching 0.1 ng/µl, 10-fold better than

the A280 assay and 1000-fold better than a Nanodrop), and that it produced almost

indistinguishable results whether or not the buffer contained protease inhibitors (Fig. 3B-C,

Supplementary Fig. 9, Supplementary Table 1). In addition, predictions from the ECmax

assay closely match those from the A280, typically predicting concentrations matching at

80-100 % those of the expected result (rather than 170-220% for microBCA; Fig. 3C,

Supplementary Table 3). For these reasons, we propose that the ECmax assay would be the

most robust assay to include in a simple calibration protocol and will proceed using

concentrations calculated from ECmax assays in what follows.

ECmax assay enables FP calibration without protein purification

Having established that the ECmax assay, a protein quantification assay that relies only on

the peak light absorbance of each FP, is highly accurate and sensitive for purified proteins

measured using trusted methods, we asked whether this method could enable us to drop

the purification step altogether. Dropping the purification step was not possible for the

other assays as they are designed to quantify total protein concentrations, but the ECmax

should in principle be specific for the considered FP and may therefore be used to quantify

FPs in crude lysates. To investigate this, we harvested and lysed cells expressing our three

FPs, separated the soluble fractions and concentrated them. Putting these through an

ECmax assay and fluorescence assay, we observed it was possible to quantify FPs in crude

lysates with high sensitivity (to 1 ng/µl), and to obtain almost identical conversion factor

values as from our purified FPs (Fig. 3E-F). Of two lysis methods tested – sonication and

chemical lysis – the former produced more accurate results. It can therefore be

demonstrated that by establishing the ECmax assay for FP quantification, it is possible to

remove the purification step altogether without compromising calibration accuracy.

Calibration allows comparison across instruments for FPs other than GFP

While fluorescein enables the comparison of experimental results across laboratories and

instruments by converting arbitrary units into 'molecules of equivalent fluorescein' (MEFL)

units, FP calibration in principle offers the same capability by converting arbitrary units into

units of 'molecules of equivalent FP' (MEFP), which is carried out using FPCountR's

process_plate() function (Fig. 4A-B). This allows us to quantify the number of FP molecules

in each well of our microplates. To calculate uncalibrated 'per cell' values, typical studies

will divide the RFU values by the optical density (OD600 or OD700) of the culture, which

quantifies cell density. The calibration of optical density to particle number can be achieved

through a similar calibration process using microspheres of similar size to *E. coli* (22,23).

Using both calibrations, it is possible to quantify molecule number per cell in 'molecules of

equivalent FP per particles of equivalent microspheres' (MEFP/PEMS, Fig. 4C), units which

should allow cross comparison between different instruments, gains and filter sets. To test

this, overnight cultures of *E. coli* containing mCherry expression vectors were split into

separate but identical microplates containing arabinose, and were grown in two plate

readers using a range of settings. The results show that normalised values of relative

fluorescence differ by ~1.5, ~3 and ~130-fold without calibration, whereas such values

become reliably comparable after calibration, even for experiments conducted using

instrument settings that produce values that cannot be legibly plotted on the same axis (Fig.

4D).

Calibration allows estimation of absolute cellular protein concentration

We next asked if calibration to units of MEFP/PEMS was a reasonable approximation for

molecule number per cell (Fig. 5A). We carried out microsphere calibration using 1 cm

cuvettes and a standard spectrophotometer and obtained conversion factors

(Supplementary Table 4) that fell within the range quantified by empirical OD600-specific

cell counts (24). In addition, other authors have confirmed that values of fluorescent protein

per cell using fluorescein and microsphere calibrants (MEFL/PEMS) are approximately equal

to those obtained using fluorescein-calibrated single cell data on a flow cytometer (7,23),

suggesting particle counts were likely to approximate actual cell numbers. However, we

observed a major caveat to the use of microspheres as calibrants, which is that their

absorbance profiles differ from that of cells. This can frustrate their ability to provide

accurate conversions between 'per cell' data calculated using OD700 versus OD600

measurements (Supplementary Fig. 11 and Supplementary Tables 4-5) for which we discuss

solutions in Supplementary Note 3.

The question of whether the measured fluorescence of FPs in cells is the equivalent of

measured fluorescence of the same number of FPs *in vitro* is less clear. Some authors have

found that cells attenuate (or 'quench') fluorescence (15,25), but the magnitude of the

effect has not been systematically investigated, particularly for modest cell concentrations

found in a typical *E. coli* growth assay. We quantified the quenching properties of *E. coli* cells

on our three FPs by mixing an increasing concentration of non-fluorescent cells with purified

FPs, and quantifying the difference in apparent fluorescence with added cells (Fig. 5B and

Supplementary Fig. 12). Our results suggest that this 'quenching' effect amounts to less than

20 % of the fluorescence signal for moderate cell densities (OD600/cm under 0.5), but

increases to ~30 % for the highest cell densities typically observed in microplate-scale

cultures (OD600/cm around 2). This information was used to add a correction step into the

process_plate() calculations so as to compensate for the expected percentage loss of

fluorescence with increasing cell density (Fig. 5B, right panel). The complete analytical

workflow from calibration to experimental data processing is illustrated in Supplementary

Fig. 13.

Using these amendments, it is possible to convert response curve assay data into molecules

per cell. Figure 5C shows one experiment using mCherry expression construct in two vectors

with different origins of replication. Using these vectors, we obtain figures for mCherry

abundance that vary between about 3,000 to 80,000 molecules per cell for p15A, and 1,000

to 600,000 for colE1. Protein abundance information, available from proteomics and

ribosome profiling studies, suggests that the typical *E. coli* protein is present in the order of

10^2-10^3 copies per cell, and the most abundant are present in the order of 10^5 copies per

cell or higher (26–28). Over-induction using the high-copy (colE1) vector therefore appears

to allow synthetic protein expression to reach the level of the most abundant proteins in the

cell. This is supported by the fact that these vectors reliably overexpress FPs to a level

observable by SDS-PAGE in unpurified lysates (Fig. 1, Supplementary Fig. 1). Modest

expression (10^2-10^3 per cell) can be achieved by combining low arabinose concentrations

with either vector. Leaky expression could be quantified in the colE1 vector as contributing

hundreds of mCherry molecules per cell even without induction, whereas in the p15A vector

it was merely ten copies per cell. In other words, the colE1 vector allows us to utilise the full

spectrum of protein abundances from modestly expressed enzymes such as the RecBCD

helicase (~100 copies per cell) through to the most abundant ribosomal proteins (~100,000

copies per cell, (28)). We could also use this to compare the number of molecules produced

from the same vectors but two different FPs. Interestingly, measurements from identical

SEVA vectors revealed that while the FP abundances were in the same order of magnitude,

mTagBFP2 accumulated to higher levels per cell than mCherry by 1.83-fold on average,

despite sharing the same promoter, 5' untranslated region, ribosome binding site and N-

terminal protein sequences (Fig. 5C-D).

In 2011, Volkmer and colleagues noted that while OD-specific *E. coli* cell counts varied with

growth rate, the OD-specific total cell volume was approximately 3.6 µl per OD600/cm,

regardless of strain or growth condition (24). Using OD as a measure of the cumulative

cellular volume in a culture could therefore be used to convert fluorescence and OD

measurements into concentrations in molar units, instead of 'per cell' values, and such

conversions may be more appropriate for comparing experimental results with quantitative

modelling of cellular reaction networks, since they are unaffected by growth rate

differences. Using this method, we found that FP abundances using the same vectors

populate a range of concentrations between 0.01-100 µM (Fig. 5D).

Absolute quantification reveals hidden properties of fluorescent proteins

Finally, we were interested in testing whether absolute quantification could illuminate a

well-known source of error in bacterial assays. The presence of red FPs has been suggested

to interfere with bacterial cell density estimations at 600 nm since red FPs typically absorb

well at this wavelength (29) and has led to the conclusion that circuits using red FPs must be

quantified at 700 nm, which is unaffected by their presence. However, the number of

molecules of red FPs that might be required for this effect to occur has never been

quantified.

Calibrated timecourse data of mCherry overexpression in *E. coli* (Fig. 6) was examined to

quantify these effects, with mGFPmut3 and mTagBFP2 used as negative controls. The ratio

between OD600 and OD700 measurements was used to identify errors caused by red FP

absorbance. Linear models fitted to the relationship between measured OD600 and OD700

values confirmed that this relationship was very similar for all uninduced cells (Fig. 6B;

OD600 = 1.30 * OD700 – 0.02), but mCherry induction resulted in a measurable deviation

(OD600 = 1.37 * OD700 – 0.03, Fig. 6B-C). Looking at the relationship between this shift and

cellular protein copy number (Fig. 6D), our results indicated that the OD600 error for

mCherry was only apparent when mCherry levels per cell were high (over 100,000 per cell),

and that the magnitude of this error was only about 5 %. In contrast, mGFPmut3 expression

had no effect on cell density estimation using OD600, as expected. Surprisingly, we

observed the opposite trend for mTagBFP2, in which OD600 measurement appeared to

underestimate cell density where mTagBFP2 was expressed at high levels per cell. The

reasons for this are currently unclear and beyond the scope of this paper, but our results

suggest that this could be an interesting avenue for future work. Generally, this experiment

confirmed that the presence of low to moderate FP levels per cell (under 100,000) do not

perturb cell density estimates, and errors are of a lower magnitude in all cases than those

from cellular fluorescence quenching.

Discussion

Our aim for this work was to develop a generalisable method that allows fluorescence

readings on microplate readers to be calibrated to molecular units of fluorescent protein.

The method ought to be (1) accurate and sensitive, (2) as simple as possible, (3) suitable for

any fluorescent protein, and (4) easily analysed. To develop the method, we adopted the

principles of redundant experimental design, including the validation of multiple assay

types, characterisation of the method's consistency, and the assessment of its generality for

three different FPs (30).

Our initial method using a simple purification protocol (Fig. 2) and a commercial protein

assay allowed us to develop an analysis pipeline to obtain conversion factors from purified

FP calibrants. To demonstrate the accuracy and validity of our calibrations, we verified that

the absorbance and fluorescence spectra of the calibrants matched their counterparts on

FPbase (Fig. 1, Supplementary Fig. 2; (21,31)), and validated our initial protein assay

measurements by cross comparison with two further methods. To do this, we adapted the

low-throughput A280 assay into an accurate, high-throughput assay format, and showed

that these were suitable for use with FPs even though some absorb in the near-UV range.

This type of assay for FP quantification has not, to our knowledge, been demonstrated in

the literature before, and we contend that it is likely to be of particular interest since it

requires no calibrant or commercial reagent, no expensive quartz-based consumables and

exhibits a sensitivity that exceeds that of commercial systems such as the Nanodrop.

We also discovered a methodological 'hack' to obtaining FP concentrations using the

extinction coefficients at their maximum excitation wavelength, the ECmax assay, which was

both the simplest and the most robust of all the assays tested. Specifically, the ECmax assay

was the least affected by buffer conditions, and had the largest linear range (to almost 0.1

ng/μl; Fig. 3). We note that the assay is limited by the fact that it requires the used FP to be

documented on FPbase, and assumes that the documented ECmax measured by other

laboratories is accurate. Promisingly, our results suggest good inter-lab agreement for these

measurements (compare A280 and ECmax estimates, Fig. 3C). In addition, an analysis of all

FPs on FPbase, comparing the proteins' extinction coefficients at 280 nm versus at their

maximum excitation wavelength, supports the idea that the ECmax is a more sensitive assay

for most FPs (Supplementary Fig. 14, Supplementary Table 3). While the EC(280) values are

highly uniform (median: 27,400 $M^{-1}cm^{-1}$), likely because most FPs are very similar in size, the

ECmax values are mostly considerably larger (median: 64,200 $M^{-1}cm^{-1}$). Impressively, as the

ECmax assay specifically quantifies FP concentration rather than total protein concentration,

we were also able to show that it is possible to do these calibrations in crude lysate without

compromising on accuracy, demonstrating that calibrants may be produced without affinity

purification (Fig. 3D-F). Overall, we expect calibrations carried out using the ECmax method,

using either purified calibrants or (sonicated) cell lysates, to be equally accurate. We provide

both rigorous and expedient protocols for prospective users on protocols.io.

Such calibrations can then be used, first, to enable the comparison of experimental results

from different plate readers or across different settings, in molecules of fluorescent protein

per particles of equivalent microspheres (MEFP/PEMS; Fig. 4B). This is akin to using

fluorescein but with a broader application range, since using bespoke calibrants for each FP

allows us to calibrate instruments for any FP regardless of its spectral characteristics. If used

merely as comparative units, the precision (repeatability) of each calibration is important,

but their accuracy (whether conversions predict molecule numbers as closely as possible) is

not. Second, they can be used to express protein abundance as 'molecules per cell' (Fig. 5C).

Accuracy here is an important consideration, and will not only depend on the accuracy of

the FP calibration (discussed above), but also on the microsphere calibration, and the

removal of any interactions between absorbance and fluorescence characteristics of cells

expressing FPs. As to microsphere calibration, cross-comparison with flow cytometry data

suggests cell count estimates from microsphere calibrations are reasonable (7,23), although

Beal and colleagues used 0.961 μm microspheres whose size is closer to *E. coli* than those

used in this study (0.890 μm; the larger type are now unavailable). Our protein abundance

estimates (10^2-10^5 proteins per cell; Fig. 5B) are also within reasonable bounds (26,28),

corroborating their use, and indicating that FP calibrations may enable protein abundance

comparisons between microplate assays and proteomics experiments.

The consideration of whether the presence of cells interferes with fluorescence

quantification (or vice versa) is multifaceted. It is well known that cells interfere with

fluorescence measurements through autofluorescence. Cellular autofluorescence is known

to largely impact GFP quantification accuracy (32), and is corrected for in FPCountR by

normalising to the background fluorescence of control cells at a similar OD (akin to (7,15)).

The 'quenching' of apparent FP fluorescence by the presence of cells is more rarely

considered (14,15). We have found that the effect size is comparable for different FPs

(Supplementary Fig. 12), unlike for autofluorescence, supporting previous observations by

others (25). A correction for this attenuation in FPCountR compensates fluorescence

according to the expected percentage quenched at the measured OD (Fig. 5B). Both of these

corrections are included in the process_plate() function.

Overall, it seems likely that the accuracy of calibrated molecule counts per cell may be

limited by the accuracy of cell count calibration, since it is known that OD-specific cell

counts change with growth rate, and that this is due to the positive relationship between

growth rate and cell size for E. coli (24). We therefore expect these calculations of protein

counts per cell will be more accurate where cell size and microsphere size are as similar as

possible, and inaccuracies will arise if the differences become significant. Thus, these values

are likely to be approximately accurate so long as the following assumptions are true: that E.

coli absorbance (scatter) is well represented by microsphere properties, and that it doesn't

change significantly over time or between samples. These will likely to be true for many *E.*

coli experiments since cell shrinkage has been shown to take place only after several hours

in stationary culture, and only large growth rate differences maintained for prolonged

periods were observed to make a significant difference to cell size in exponentially growing

cells (24), but may fail if circuits impose significant burden on host cell resources that

impacts growth rate or cell size (33–35). In contrast, using a conversion from OD to total cell

volume in a given culture well allows us to remove the requirement for counting cells (Fig.

5D) and allows the expression of protein concentrations as 0.01-100 µM. These figures

should be approximately accurate under the assumption that the OD-specific cell volume

doesn't vary significantly between samples or over time. This is strongly supported by the

results of Volkmer and colleagues (24) whose data suggests this variation is within 2-fold

across a wide range of growth conditions, but also by others who have shown that as cell

volumes increase, their OD-specific cell counts decrease by approximately the same

magnitude (33,36). Units of concentration may also be more meaningful for reaction

modelling since ultimately it is molecular concentration that is critical for binding and

kinetics (9,37,38).

We used absolute protein quantification to investigate the problem first described by Hecht

and colleagues (29) in which an association between mCherry overexpression and

deviations in cell density measurements were made. Repeating these assays showed a clear

effect, but this effect was of modest magnitude (<5%) and was only apparent at very high

mCherry levels per cell (over 10^5 molecules per cell; Fig. 6). This suggests that for most

circuits that use moderate expression levels to minimise cellular burden, OD600 values

would remain an accurate way to quantify cell density. We also unexpectedly observed the

opposite effect in mTagBFP2 expressions. As this analysis was technically a quantification of

whether OD600 and OD700 measurements deviate from each other in the presence of

different FPs, these results might not suggest that OD600-based cell density readings in the

presence of high mTagBFP2 are inaccurate. It might instead suggest that, conversely for

mTagBFP2, the OD700 readings are inaccurate. We do not currently have an explanation for

this finding. One possible contributing factor may be that while it is often described as one

of the best blue FPs, mTagBFP2 has a higher propensity to aggregate than mCherry and

most GFPs (39). Cell stress is known to induce *E. coli* to elongate (33,40), which may affect

its scattering properties. It is possible that very high FP levels may frequently cause a small

but significant error in cell density estimates due to combinations of effects from light

absorption and scatter that warrants more study in order to allow us to further improve

molecular quantifications of FPs under those conditions.

While flow cytometry and mass spectrometry allow us to probe single-cell measurements or

the protein complement of an entire cell, respectively, microplate readers remain an

important screening platform in the Design-Build-Test-Learn cycle due to ease of use, low

cost and high iterative capabilities. This necessitates the development of methods for

extracting informative numbers from such data. The ability to extract absolute protein

abundance information from assays of engineered microbial cells is imperative to enable

the characterisation, optimisation and tuning of genetic circuits in a rigorous and

quantitative manner, and will allow for a deeper insight into how protein abundance affects

genetic construct behaviour, cellular burden and growth rate. Importantly, our approach

enables accurate, non-destructive, and easy protein abundance comparisons, even between

samples that differ in growth rate or cell size.

Further, such absolute quantification need not be limited to fluorescent proteins. The last

few years have seen a fantastic expansion of fluorogenic molecules, tools that have allowed

the specific quantification of localised proteins (41–43), proteins in anaerobic environments

(44,45), and the fluorescent quantification of RNAs (38,46,47). Calibration of these

molecules would be more complex to achieve but no less valuable. Equally, calibration of

alternative instruments such as flow cytometers are also of interest, but would require a

very different approach, requiring calibrants to be attached to particles of cell-like sizes. We

hope that our demonstration of how to achieve absolute FP quantifications from microplate

reader assays using FPCountR can contribute to the effort to develop more quantitative

approaches for the analysis of circuit behaviour in synthetic biology and beyond.

Materials and Methods

Materials

Primers and gblocks were obtained from IDT, and *E. coli* strain DH5-alpha (Invitrogen,

18265-017) was used for molecular cloning. Chemicals and protein reagents were purchased

from Merck, Sigma, ThermoFisher Scientific, and Bio-Rad, molecular biology reagents from

NEB and Life Technologies and general laboratory reagents from Corning, Greiner Bio-One,

Starlab and VWR.

Fluorescent proteins

The mCherry (48) protein sequence was based on the FPbase entry for mcherry (FPbase ID:

ZERB6) with the following changes: M1(MVHHHHHHGSG). The mGFPmut3 (49) protein

sequence was based on the FPbase entry for gfpmut3 (FPbase ID: A20WC) with the

following changes: M1(MVHHHHHHGSG), A206K (a substitution to make the protein

monomeric; (18,19)). The mTagBFP2 (50) protein sequence was based on the FPbase entry

for mtagbfp2 (FPbase ID: ZO7NN) with the following changes: M1(MVHHHHHHGSG). Full

protein sequences are provided in Supplementary Table 6.

DNA assembly

Vectors for fluorescent protein purification and growth curve assays were constructed

according to standard protocols, via Golden Gate and Gibson assembly techniques using *E.*

coli strain DH5α (Invitrogen, 18265-017). Constructs were assembled into Standardised

European Vector Architecture (SEVA) backbones (51): pS381 was generated from pS181

with chloramphenicol substitution; pS361 was generated similarly from pS161. In an effort

to approximately equalise expression levels between different proteins, the 5' untranslated

region (including ribosome binding site) and 5' region of each construct was set to be

identical up to residue 11 (coding region begins: *DNA*:

ATGGTTCACCATCATCATCACcacGGTtcgggc, *protein*: MVHHHHHHGSG). The second residue

was set to valine to reduce the effects of N-end mediated degradation (52). The affinity tag

chosen for purification was the His$_6$ tag, which was followed by a short unstructured linker

(GSG). Full DNA sequences of vectors are provided in Supplementary Table 6.

FPCount (wet lab) protocol

The protocols are available on protocols.io. In total, we have described three FPCount

calibration protocols: (1) the full protocol using FP purification and three protein assays for

cross-validation, as conducted in Fig. 3 available at https://www.protocols.io/view/fpcount-

protocol-full-protocol-bztsp6ne (53), (2) a short protocol requiring purification but using the

ECmax assay only, available at https://www.protocols.io/view/fpcount-protocol-short-

protocol-bzt6p6re (54), and (3) the purification-free protocol that uses FPs in lysates for

calibration, available at https://www.protocols.io/view/fpcount-protocol-in-lysate-

purification-free-proto-bzudp6s6 (55). It is anticipated that users will have differing

requirements concerning protocol time and validation, so we have provided all three

protocols.

Protein expression and harvesting

Fluorescent proteins were produced using pS381 (SEVA) expression vectors in *E. coli*

BL21(DE3) strains. Glycerol stocks were inoculated into 50 ml Luria Broth (Miller)

supplemented with 50 µg/ml chloramphenicol and 0.02 % arabinose and were grown

overnight at 30 $^\circ$C at 250 rpm. Cells were harvested after ~16 h by transferring them to

prechilled containers on ice. All further steps were conducted on ice. OD600 readings were

taken, and 40 OD of cells was transferred to fresh falcon tubes, washed once in T50N300 (50

mM Tris-HCl pH 7.5, 300 mM NaCl) and resuspended in lysis buffer (T50N300, 1X protease

inhibitors (EDTA-free, Pierce A32955), filter sterilised, supplemented with lysozyme 100

µg/µl). Cells were separated into 20 OD (2 ml) fractions and sonicated (Qsonica Q125

sonicator, 50% amplitude, 10s on, 10s off, 2 min). Lysates were supplemented with 5 mM

$CaCl_2$, 50 mM $MgCl_2$ and treated with DNase I (50 U/ml, bovine pancreas, MP Biomedicals,

219006210) for 30 min at 4 $^\circ$C. Soluble fractions were isolated by isolation of the

supernatant after centrifugation (16,000 xg, 30 min, 4 $^\circ$C), and both fractions were checked

by SDS-PAGE followed by staining with Coomassie-based dye according to the

manufacturer's instructions (Instant Blue Protein Stain, Sigma ISB1L-1L).

Protein purification

Fluorescent proteins were purified in batch using His-tag affinity chromatography at room

temperature according to the resin manufacturer's instructions (Thermo Fisher). Lysates

were supplemented with 10 mM imidazole and 600 µl was applied to HisPur Cobalt resin

(300 µl, ThermoFisher) equilibrated in Binding Buffer (T50N300+pi, 10 mM imidazole),

mixed and incubated at room temperature for ~15 min before removal (1,000 xg, 1 min).

This was repeated four times, before the resin was washed with 10 column volumes of

Binding Buffer. Protein was eluted in Elution Buffer (T50N300+pi, 150 mM imidazole). All

protein fractions for calibration were stored protected from light at 4 °C.

Preparation for calibration assays

603 Elution fractions were combined and concentrated approximately 10-fold using Amicon

604 centrifugal filter columns (Merck, UFC5010), followed by buffer exchange (1000x) into

605 T5N15 (5mM Tris-HCl pH 7.5, 15 mM NaCl) or T5N15+pi (T5N15, 1x protease inhibitors, filter

606 sterilised).

Microplate reader assays. All assays were carried out using a Tecan Spark microplate reader

609 except the fluorescence spectra assays, which were carried out using a BMG Clariostar Plus

610 microplate reader.

Calibration assays. For each FP calibration, both concentration and fluorescence assays

were carried out on the same dilution series of protein. Concentrated, buffer exchanged FP

(100 µl) was diluted in 900 µl buffer, then diluted 1:2 into 500 µl buffer in 1.5 ml

eppendorfs. A total of 11 dilutions were prepared in this way, distributed into UV-

transparent microplates (Greiner, 655801) as duplicates (225 µl). Bovine serum albumin

(BSA) standards (from Micro BCA Protein Assay kit, ThermoFisher, 23235) were prepared in

parallel with the same buffer(s). This dilution set was then subjected to the protein

concentration and fluorescence assays.

Protein concentration assays: A280 assay and ECmax assay

Absorbance assays were carried out on 225 µl protein in UV-transparent plates, using the

Spark absorbance scan method (see Supplementary Methods).

Fluorescence assays

Following absorbance scans, 200 µl from each well of the original plate was transferred into

clear polystyrene plates (Corning, 3370). This plate was sealed (Eppendorf Masterclear real-

time PCR film adhesive, 30132947) and used to run the Spark fluorescence methods (see

Supplementary Methods) on all relevant instruments, channels and gains.

Protein concentration assays: microBCA assay

BCA assays were carried out using the Micro BCA Protein Assay kit (ThermoFisher, 23235)

according to the manufacturer's instructions (microplate protocol). Briefly, 150 µl of

working reagent was dispensed into a clean microplate, and 150 µl from each well of the

fluorescence assay plate was mixed into the reagent with a multichannel pipettor. Reactions

were covered with a plate seal (BreatheEasy sealing membrane, Sigma, Z380059), and

subjected to the Spark microBCA method (see Supplementary Methods).

Calibration of OD600 and OD700 values using microspheres

Calibration of optical density readings used to quantify cell number (OD600 and OD700) was carried out according to published protocols (7). The microspheres used were monodisperse silica microspheres (Cospheric, SiO2MS-2.0, 2.0g/cc, d50=0.890um, CV=3.2%, <1% Doubles).

Bacterial timecourse assays

DH10B *E. coli* transformants were grown overnight in M9 medium (M9 salts (1X, Sigma M6030), casamino acids (0.2%), fructose (0.8%), thiamine HCl (0.25 mg/ml), $MgSO_4$ (2 mM), $CaCl_2$ (0.1 mM)) supplemented with 50 μg/ml chloramphenicol, in a deep-well plate (30 °C, 700 rpm), and diluted the following morning into fresh M9 with antibiotic (deep-well plate, 30 °C, 700 rpm) to an OD600 (cm^{-1}) of 0.05. After 1 hour, cultures were transferred into clear 96-well microplates (Corning, 3370) with pre-loaded arabinose (5 μl). Plates were sealed (BreatheEasy sealing membrane, Sigma, Z380059) and grown in a Tecan Spark plate reader in kinetic mode (see Supplementary Methods, Spark growth curve method).

Analytical methods

All data was analysed using R (56). The FPCountR package that was developed for the FP calibrations is available on GitHub at https://github.com/ec363/fpcountr (57). For a summary of the functions, see the Fig. 1 and Supplementary Fig. 13.

Fluorescence scans

Excitation and emission spectra of the fluorescent proteins were conducted using a BMG Clariostar Plus microplate reader in sealed plates (Corning, 3370; Eppendorf Masterclear

real-time PCR film adhesive, 30132947) at wavelengths appropriate to each FP (see

Supplementary Methods).

Figures

Figures were created using RStudio and Biorender.com.

Data availability

Computer code produced in this study is available on the following repositories:

• R package for data analysis: GitHub (https://github.com/ec363/fpcountr)

This study includes no data deposited in external repositories.

Acknowledgements

GBS and EC acknowledge funding from the Royal Academy of Engineering (RAEng CiET

1819\5).

Author contributions

EC and GBS conceived the study. EC conducted the experiments, analysed the data and

wrote the R package. EC and GBS wrote the manuscript.

Conflict of Interest

The authors declare no conflict of interest.

References
1. Delvigne F, Zacchetti B, Fickers P, Fifani B, Roulling F, Lefebvre C, et al. Improving control
in microbial cell factories: from single cell to large-scale bioproduction. FEMS Microbiol

Lett [Internet]. 2018 Sep 24 [cited 2021 Nov 26]; Available from:
https://academic.oup.com/femsle/advance-
article/doi/10.1093/femsle/fny236/5106341

2. Beal J, Goñi-Moreno A, Myers C, Hecht A, Vicente M del C, Parco M, et al. The long
journey towards standards for engineering biosystems: Are the Molecular Biology and
the Biotech communities ready to standardise? EMBO Rep [Internet]. 2020 May 6
[cited 2021 Nov 26];21(5). Available from:
https://onlinelibrary.wiley.com/doi/10.15252/embr.202050521

3. Perrino G, Hadjimitsis A, Ledesma-Amaro R, Stan G-B. Control engineering and synthetic
biology: working in synergy for the analysis and control of microbial systems. Curr Opin
Microbiol [Internet]. 2021 Aug 1 [cited 2021 Jun 3];62:68–75. Available from:
https://www.sciencedirect.com/science/article/pii/S1369527421000655

4. Chalfie M, Tu Y, Euskirchen G, Ward WW, Prasher DC. Green fluorescent protein as a
marker for gene expression. Science [Internet]. 1994 Feb 11 [cited 2021 Jan
18];263(5148):802–5. Available from:
https://science.sciencemag.org/content/263/5148/802

5. Rodriguez EA, Campbell RE, Lin JY, Lin MZ, Miyawaki A, Palmer AE, et al. The Growing
and Glowing Toolbox of Fluorescent and Photoactive Proteins. Trends Biochem Sci.
2017 Feb;42(2):111–29.

6. Beal J, Haddock-Angelli T, Baldwin G, Gershater M, Dwijayanti A, Storch M, et al.
Quantification of bacterial fluorescence using independent calibrants. PLOS ONE
[Internet]. 2018 Jun 21 [cited 2020 Jun 29];13(6):e0199432. Available from:
https://journals.plos.org/plosone/article?id=10.1371/journal.pone.0199432

7. Fedorec AJH, Robinson CM, Wen KY, Barnes CP. FlopR: An Open Source Software
Package for Calibration and Normalization of Plate Reader and Flow Cytometry Data.
ACS Synth Biol [Internet]. 2020 Sep 18 [cited 2021 Nov 26];9(9):2258–66. Available
from: https://pubs.acs.org/doi/10.1021/acssynbio.0c00296

8. Gorochowski TE, Chelysheva I, Eriksen M, Nair P, Pedersen S, Ignatova Z. Absolute
quantification of translational regulation and burden using combined sequencing
approaches. Mol Syst Biol. 2019 03;15(5):e8719.

9. Boada Y, Vignoni A, Alarcon-Ruiz I, Andreu-Vilarroig C, Monfort-Llorens R, Requena A, et
al. Characterization of Gene Circuit Parts Based on Multiobjective Optimization by
Using Standard Calibrated Measurements. ChemBioChem [Internet]. 2019 Oct 15
[cited 2021 Nov 26];20(20):2653–65. Available from:
https://onlinelibrary.wiley.com/doi/10.1002/cbic.201900272

10. Vignoni A, Boada Y, Boada Acosta L, Andreu-Vilarroig C, Alarcón I, Requena A, et al.
Fluorescence calibration and color equivalence for quantitative synthetic biology. IFAC-
Pap [Internet]. 2019 [cited 2021 Nov 26];52(26):129–34. Available from:
https://linkinghub.elsevier.com/retrieve/pii/S2405896319321305

11. Taniguchi Y, Choi PJ, Li G-W, Chen H, Babu M, Hearn J, et al. Quantifying E. coli
Proteome and Transcriptome with Single-Molecule Sensitivity in Single Cells.
2010;329:8.

12. Finan K, Raulf A, Heilemann M. A Set of Homo-Oligomeric Standards Allows Accurate
Protein Counting. Angew Chem Int Ed [Internet]. 2015 Oct 5 [cited 2021 Nov
26];54(41):12049–52. Available from:
https://onlinelibrary.wiley.com/doi/10.1002/anie.201505664

13. Tie HC, Madugula V, Lu L. The development of a single molecule fluorescence standard
and its application in estimating the stoichiometry of the nuclear pore complex.
Biochem Biophys Res Commun [Internet]. 2016 Sep [cited 2021 Nov 26];478(4):1694–
9. Available from: https://linkinghub.elsevier.com/retrieve/pii/S0006291X16314486

14. Canton B, Labno A, Endy D. Refinement and standardization of synthetic biological parts
and devices. Nat Biotechnol [Internet]. 2008 Jul [cited 2022 Jan 3];26(7):787–93.
Available from: https://www.nature.com/articles/nbt1413

15. Hirst CD, Ainsworth C, Baldwin G, Kitney RI, Freemont PS. Protocol for the
Standardisation of Transcriptional Measurements. In: McGenity TJ, Timmis KN, Nogales
B, editors. Hydrocarbon and Lipid Microbiology Protocols [Internet]. Berlin, Heidelberg:
Springer Berlin Heidelberg; 2015 [cited 2021 Nov 26]. p. 9–26. (Springer Protocols
Handbooks). Available from: http://link.springer.com/10.1007/8623_2015_148

16. Noble JE, Knight AE, Reason AJ, Di Matola A, Bailey MJA. A Comparison of Protein
Quantitation Assays for Biopharmaceutical Applications. Mol Biotechnol [Internet].
2007 Sep 13 [cited 2021 Nov 26];37(2):99–111. Available from:
http://link.springer.com/10.1007/s12033-007-0038-9

17. Pace CN, Vajdos F, Fee L, Grimsley G, Gray T. How to measure and predict the molar
absorption coefficient of a protein. Protein Sci [Internet]. 1995 Nov [cited 2021 Nov
26];4(11):2411–23. Available from:
https://onlinelibrary.wiley.com/doi/10.1002/pro.5560041120

18. Zacharias DA, Violin JD, Newton AC, Tsien RY. Partitioning of Lipid-Modified Monomeric
GFPs into Membrane Microdomains of Live Cells. Science [Internet]. 2002 May 3 [cited
2021 Nov 26];296(5569):913–6. Available from:
https://www.science.org/doi/10.1126/science.1068539

19. Shaner NC, Steinbach PA, Tsien RY. A guide to choosing fluorescent proteins. Nat
Methods [Internet]. 2005 Dec [cited 2021 Nov 26];2(12):905–9. Available from:
https://www.nature.com/articles/nmeth819

20. Thermo Fisher Scientific. NanoDrop 1000 Spectrophotometer, V3.8 User's Manual
[Internet]. 2010 [cited 2021 Nov 26]. Available from:
https://tools.thermofisher.com/content/sfs/manuals/nd-1000-v3.8-users-manual-
8%205x11.pdf

21. Lambert TJ. FPbase: a community-editable fluorescent protein database. Nat Methods
765   [Internet]. 2019 Apr [cited 2021 Nov 26];16(4):277–8. Available from:
766   http://www.nature.com/articles/s41592-019-0352-8

767   22. Stevenson K, McVey AF, Clark IBN, Swain PS, Pilizota T. General calibration of microbial
768   growth in microplate readers. Sci Rep [Internet]. 2016 Dec [cited 2021 Nov
769   26];6(1):38828. Available from: http://www.nature.com/articles/srep38828

770   23. Beal J, Farny NG, Haddock-Angelli T, Selvarajah V, Baldwin GS, Buckley-Taylor R, et al.
771   Robust estimation of bacterial cell count from optical density. Commun Biol [Internet].
772   2020 Dec [cited 2021 Nov 26];3(1):512. Available from:
773   https://www.nature.com/articles/s42003-020-01127-5

774   24. Volkmer B, Heinemann M. Condition-Dependent Cell Volume and Concentration of
775   Escherichia coli to Facilitate Data Conversion for Systems Biology Modeling. Langowski
776   J, editor. PLoS ONE [Internet]. 2011 Jul 29 [cited 2021 Nov 26];6(7):e23126. Available
777   from: https://dx.plos.org/10.1371/journal.pone.0023126

778   25. Zhang C, Liu M-S, Han B, Xing X-H. Correcting for the inner filter effect in measurements
779   of fluorescent proteins in high-cell-density cultures. Anal Biochem. 2009 Jul
780   15;390(2):197–202.

781   26. Ishihama Y, Schmidt T, Rappsilber J, Mann M, Hartl FU, Kerner MJ, et al. Protein
782   abundance profiling of the Escherichia coli cytosol. BMC Genomics [Internet]. 2008
783   [cited 2021 Nov 26];9(1):102. Available from:
784   http://bmcgenomics.biomedcentral.com/articles/10.1186/1471-2164-9-102

785   27. Arike L, Valgepea K, Peil L, Nahku R, Adamberg K, Vilu R. Comparison and applications of
786   label-free absolute proteome quantification methods on Escherichia coli. J Proteomics
787   [Internet]. 2012 Sep [cited 2021 Nov 26];75(17):5437–48. Available from:
788   https://linkinghub.elsevier.com/retrieve/pii/S1874391912004861

789   28. Li G-W, Burkhardt D, Gross C, Weissman JS. Quantifying Absolute Protein Synthesis
790   Rates Reveals Principles Underlying Allocation of Cellular Resources. Cell [Internet].
791   2014 Apr [cited 2021 Nov 26];157(3):624–35. Available from:
792   https://linkinghub.elsevier.com/retrieve/pii/S0092867414002323

793   29. Hecht A, Endy D, Salit M, Munson MS. When Wavelengths Collide: Bias in Cell
794   Abundance Measurements Due to Expressed Fluorescent Proteins. ACS Synth Biol
795   [Internet]. 2016 Sep 16 [cited 2019 Dec 10];5(9):1024–7. Available from:
796   https://doi.org/10.1021/acssynbio.6b00072

797   30. Casadevall A, Fang FC. Rigorous Science: a How-To Guide. mBio [Internet]. 2016 Dec 30
798   [cited 2021 Nov 3];7(6). Available from:
799   https://journals.asm.org/doi/10.1128/mBio.01902-16

800   31. Raynal B, Lenormand P, Baron B, Hoos S, England P. Quality assessment and
801   optimization of purified protein samples: why and how? Microb Cell Factories

[Internet]. 2014 Dec [cited 2021 Nov 26];13(1):180. Available from:
http://microbialcellfactories.biomedcentral.com/articles/10.1186/s12934-014-0180-6

32. Lichten CA, White R, Clark IB, Swain PS. Unmixing of fluorescence spectra to resolve
quantitative time-series measurements of gene expression in plate readers. BMC
Biotechnol [Internet]. 2014 Dec [cited 2021 Nov 26];14(1):11. Available from:
https://bmcbiotechnol.biomedcentral.com/articles/10.1186/1472-6750-14-11

33. Basan M, Zhu M, Dai X, Warren M, Sévin D, Wang Y, et al. Inflating bacterial cells by
increased protein synthesis. Mol Syst Biol [Internet]. 2015 Oct [cited 2021 Nov
26];11(10):836. Available from:
https://onlinelibrary.wiley.com/doi/10.15252/msb.20156178

34. Ceroni F, Algar R, Stan G-B, Ellis T. Quantifying cellular capacity identifies gene
expression designs with reduced burden. Nat Methods [Internet]. 2015 May [cited
2019 Mar 14];12(5):415–8. Available from:
http://www.nature.com/articles/nmeth.3339

35. Boo A, Ellis T, Stan G-B. Host-aware synthetic biology. Curr Opin Syst Biol [Internet].
2019 Apr [cited 2020 Jun 26];14:66–72. Available from:
https://linkinghub.elsevier.com/retrieve/pii/S245231001930006X

36. Schaechter M, Maaløe O, Kjeldgaard NOY 1958. Dependency on Medium and
Temperature of Cell Size and Chemical Composition during Balanced Growth of
Salmonella typhimurium. Microbiology [Internet]. 1958 [cited 2021 Nov 26];19(3):592–
606. Available from:
https://www.microbiologyresearch.org/content/journal/micro/10.1099/00221287-19-
3-592

37. Bloom RJ, Winkler SM, Smolke CD. A quantitative framework for the forward design of
synthetic miRNA circuits. Nat Methods [Internet]. 2014 Nov [cited 2021 Nov
26];11(11):1147–53. Available from: http://www.nature.com/articles/nmeth.3100

38. Siegal-Gaskins D, Tuza ZA, Kim J, Noireaux V, Murray RM. Gene Circuit Performance
Characterization and Resource Usage in a Cell-Free "Breadboard". ACS Synth Biol
[Internet]. 2014 Jun 20 [cited 2021 Nov 26];3(6):416–25. Available from:
https://pubs.acs.org/doi/10.1021/sb400203p

39. Cranfill PJ, Sell BR, Baird MA, Allen JR, Lavagnino Z, de Gruiter HM, et al. Quantitative
assessment of fluorescent proteins. Nat Methods [Internet]. 2016 Jul [cited 2020 Jul
3];13(7):557–62. Available from: http://www.nature.com/articles/nmeth.3891

40. Justice SS, Hunstad DA, Cegelski L, Hultgren SJ. Morphological plasticity as a bacterial
survival strategy. Nat Rev Microbiol [Internet]. 2008 Feb [cited 2021 Nov 26];6(2):162–
8. Available from: http://www.nature.com/articles/nrmicro1820

41. Svendsen S, Zimprich C, McDougall MG, Klaubert DH, Los GV. Spatial separation and
bidirectional trafficking of proteins using a multi-functional reporter. BMC Cell Biol

[Internet]. 2008 Dec [cited 2021 Nov 26];9(1):17. Available from:
https://bmcmolcellbiol.biomedcentral.com/articles/10.1186/1471-2121-9-17

42. Li C, Mourton A, Plamont M-A, Rodrigues V, Aujard I, Volovitch M, et al. Fluorogenic
Probing of Membrane Protein Trafficking. Bioconjug Chem [Internet]. 2018 Jun 20
[cited 2019 Mar 20];29(6):1823–8. Available from:
http://pubs.acs.org/doi/10.1021/acs.bioconjchem.8b00180

43. Csibra E, Renders M, Pinheiro V. Bacterial cell display as a robust and versatile platform
for the engineering of low affinity ligands and enzymes. ChemBioChem [Internet]. 2020
May 15 [cited 2020 May 26];cbic.202000203. Available from:
https://onlinelibrary.wiley.com/doi/abs/10.1002/cbic.202000203

44. Streett HE, Kalis KM, Papoutsakis ET. A Strongly Fluorescing Anaerobic Reporter and
Protein-Tagging System for *Clostridium* Organisms Based on the Fluorescence-
Activating and Absorption-Shifting Tag Protein (FAST). Kivisaar M, editor. Appl Environ
Microbiol [Internet]. 2019 Jul 15 [cited 2021 Nov 26];85(14). Available from:
https://journals.asm.org/doi/10.1128/AEM.00622-19

45. Charubin K, Streett H, Papoutsakis ET. Development of Strong Anaerobic Fluorescent
Reporters for Clostridium acetobutylicum and Clostridium ljungdahlii Using HaloTag
and SNAP-tag Proteins. Kelly RM, editor. Appl Environ Microbiol [Internet]. 2020 Oct
[cited 2021 Nov 26];86(20). Available from:
https://journals.asm.org/doi/10.1128/AEM.01271-20

46. Pothoulakis G, Ceroni F, Reeve B, Ellis T. The Spinach RNA Aptamer as a Characterization
Tool for Synthetic Biology. ACS Synth Biol [Internet]. 2014 Mar 21 [cited 2021 Nov
26];3(3):182–7. Available from: https://pubs.acs.org/doi/10.1021/sb400089c

47. Yerramilli VS, Kim KH. Labeling RNAs in Live Cells Using Malachite Green Aptamer
Scaffolds as Fluorescent Probes. ACS Synth Biol [Internet]. 2018 Mar 16 [cited 2021
Nov 26];7(3):758–66. Available from:
https://pubs.acs.org/doi/10.1021/acssynbio.7b00237

48. Shaner NC, Campbell RE, Steinbach PA, Giepmans BNG, Palmer AE, Tsien RY. Improved
monomeric red, orange and yellow fluorescent proteins derived from Discosoma sp.
red fluorescent protein. Nat Biotechnol [Internet]. 2004 Dec [cited 2020 May
22];22(12):1567–72. Available from: https://www.nature.com/articles/nbt1037

49. Cormack BP, Valdivia RH, Falkow S. FACS-optimized mutants of the green fluorescent
protein (GFP). Gene [Internet]. 1996 Jan [cited 2019 Oct 23];173(1):33–8. Available
from: https://linkinghub.elsevier.com/retrieve/pii/0378111995006850

50. Subach OM, Cranfill PJ, Davidson MW, Verkhusha VV. An Enhanced Monomeric Blue
Fluorescent Protein with the High Chemical Stability of the Chromophore. PLOS ONE
[Internet]. 2011 Dec 8 [cited 2021 Jan 13];6(12):e28674. Available from:
https://journals.plos.org/plosone/article?id=10.1371/journal.pone.0028674

51. Silva-Rocha R, Martínez-García E, Calles B, Chavarría M, Arce-Rodríguez A, de las Heras
A, et al. The Standard European Vector Architecture (SEVA): a coherent platform for
the analysis and deployment of complex prokaryotic phenotypes. Nucleic Acids Res
[Internet]. 2013 Jan [cited 2021 Nov 26];41(Database issue):D666–75. Available from:
https://www.ncbi.nlm.nih.gov/pmc/articles/PMC3531073/

52. Mogk A, Schmidt R, Bukau B. The N-end rule pathway for regulated proteolysis:
prokaryotic and eukaryotic strategies. Trends Cell Biol [Internet]. 2007 Apr 1 [cited
2021 Nov 26];17(4):165–72. Available from: https://www.cell.com/trends/cell-
biology/abstract/S0962-8924(07)00028-1

53. Csibra E, Stan G-B. FPCount protocol - Full protocol. protocols.io [Internet]. 2021 [cited
2021 Dec 5]; Available from: https://www.protocols.io/view/fpcount-protocol-full-
protocol-bztsp6ne

54. Csibra E, Stan G-B. FPCount protocol - Short protocol. protocols.io [Internet]. 2021 [cited
2021 Dec 5]; Available from: https://www.protocols.io/view/fpcount-protocol-short-
protocol-bzt6p6re

55. Csibra E, Stan G-B. FPCount protocol - in-lysate (purification free) protocol. protocols.io
[Internet]. 2021 [cited 2021 Dec 5]; Available from:
https://www.protocols.io/view/fpcount-protocol-in-lysate-purification-free-proto-
bzudp6s6

56. R Core Team. R: The R Project for Statistical Computing [Internet]. 2021 [cited 2021 Dec
5]. Available from: https://www.r-project.org/

57. Csibra E. FPCountR: Fluorescent protein calibration for plate readers [Internet]. Zenodo;
2021 [cited 2021 Dec 5]. Available from: https://zenodo.org/record/5760028

Figure 1. Overview of fluorescent protein calibration workflow using FPCountR. A calibration workflow is described *(left),* followed by a demonstration of how this calibration can be used to convert experimental data from arbitrary fluorescence units per optical density into molecules per cell *(right)*. The calibration workflow consists of a wet lab protocol (*top, available on protocols.io*) and an analysis package (*bottom, available on GitHub*). In brief, the protocol describes how to prepare fluorescent protein calibrants by expression and purification, though the latter step is optional as lysates allow accurate calibration without the need for purification. The protocol also describes how to collect data for the calibration for both the protein assay to determine protein concentration, as well as the fluorescence assay to determine protein activity. The analytical workflow is provided as an open source R package, complete with functions that enable the extraction of protein concentrations from protein assay data, conversion factors (arbitrary fluorescence units per molecule) from a combination of protein and fluorescence assay data, and functions that allow users to convert experimental data into absolute units.

A

calibrants in lysate

Expression

high copy
SEVA vector

overnight
expressions

Lysis

cell lysis
by sonication

removal of
insoluble fraction

purified calibrants

Purification

His tag purifications
small-scale
using Cobalt resin

FP calibrants

B

ara

pS381

araC   Para   His6-FP

**SEVA vector
(pS381)**

name	FPbase slug	changes from FPbase
mTagBFP2	mtagbfp2	M1(MVHHHHHHGSG)
mGFPmut3	gfpmut3	M1(MVHHHHHHGSG), A206K
mCherry	mCherry	M1(MVHHHHHHGSG)

C

mCherry

M   S   I

kDa

M – marker
S – soluble
I – insoluble
F – flowthrough
E – elution

D

mCherry

S   F1   F5   E1   E2   E3

kDa

E

data
fpbase

mTagBFP2

mGFPm3

mCherry

wavelength (nm)

Figure 2. Preparation of fluorescent protein calibrants.
A. Protocol summary. The use of high-copy vectors and overnight expression was designed to maximise protein production, and the temperature was dropped to 30 ºC to minimise misfolding. Cells were lysed using sonication, to avoid the requirement to add chemical components that may interfere with downstream processes, such as EDTA (with His-tag purification) or detergents (with protein quantification), or with unknown components of 'black box' commercial lysis reagents. Insoluble proteins were removed via centrifugation and SDS-PAGE was used to confirm that the majority of the expressed FP was in the soluble fraction. Proteins were purified using His-tag affinity purification. His tags are popular affinity tags that are known for their reliability and are small in size, making them unlikely to compromise fusion protein function. Cobalt resin was used as it has higher specifity than nickel resins and is expected to co-isolate fewer impurities. The arrows on the left represent the steps required to prepare purified calibrants (grey) vs. calibrants in lysate (black). **B. Vector and FP design.** A standardised FP expression vector was constructed from an arabinose-inducible His-tagged FP construct in a high-copy SEVA vector. Three commonly-used FPs from across the spectral range were chosen for testing this protocol: mTagBFP2, mGFPmut3, and mCherry. A table of the three proteins illustrates any changes in protein sequence compared to their FPbase counterparts, showing identical matches with the exception of affinity tags and a monomerising mutation for GFPmut3. **C. Expression and solubility verification.** SDS-PAGE analysis of lysates after separation of the insoluble fraction was used to make sure that most of the fluorescent protein was soluble. The displayed SDS-PAGE is from an mCherry purification, showing the separation of the soluble (S) vs. insoluble (I) fraction, next to the protein marker (M) on a 12 % gel. **D. Purification verification.** SDS-PAGE analysis after purification was used to confirm the success of purifications. The displayed SDS-PAGE is from an mCherry purification, showing the separation of the soluble (S) fraction, next to two flowthrough (F) fractions from the binding steps showing efficient FP binding to the cobalt resin, and three elution (E) fractions. **E. Fluorescence spectra.** Fluorescence spectral scans were used to confirm that the purified FP behaves as expected. The figure shows obtained spectra (normalised such that the highest value = 1) fitted to a loess model with a 95% confidence interval (orange) overlaid with FPbase spectra (black) for each FP. Displayed spectra represent one sample measured in duplicate, that is representative of at least 2 independently purified batches of calibrant.

A **Comparison of calibration protocols**

B

C

D **Calibration in lysate**

E

F

Figure 3. Systematic characterisation of calibration protocols reveals novel assay that allows calibration in crude lysates.

A-C. Characterisation of calibration methods using purified proteins. A. Sample preparation and assay workflow. Purified FPs in elution buffer (T50N300 with protease inhibitors (pi) and imidazole) were dialysed (1) into T5N15 and then again (2) into T5N15 or T5N15 with pi. Serial dilutions of each of these FPs in both buffers were then prepared, and each dilution was subjected to three protein assays (microBCA, A280 and ECmax assay) as well as the fluorescence assay. **B. Effect of buffer and assay on measured protein concentration of each dilution.** An example of how the buffer and choice of assay effect the measured protein concentrations using a dilution series of mTagBFP2 protein (for full results see Supplementary Fig. 9). The serial dilution was a 2-fold dilution, where 11 dilutions were prepared and measured in duplicate. Points represent the mean of the duplicate values. All dilutions were measured with the A280 and ECmax assays but only the top 8 dilutions were tested with the microBCA assay. The values for the A280 and ECmax assays were normalised for scatter as indicated in Supplementary Table 2. Any missing data points had concentrations recorded as being below 0.01 ng/μl. **C. Effect of assay on calculated protein concentration compared to the A280.** The raw data of each serial dilution, as displayed in B, was fitted to a linear model and used to estimate the concentration of the first sample in the series (where the dilution factor = 1), according to methods for each assay as set out in Supplementary Fig. 2, 7 and 8. For a given FP batch (set), the microBCA and ECmax assay error was calculated by taking the fold difference in concentration predicted by the named assay, versus that predicted by the A280 assay using T5N15 buffer. Each point therefore represents one value for each FP batch. The full data for this figure can be found in Supplementary Table 1.

D-F. FP calibration in crude lysates. D. Sample preparation and assay workflow. Calibration in lysate requires fewer steps: affinity purification is omitted, buffer exchange is not required (optionally, lysates may be concentrated to increase signal), and dilution series are subject to only two assays requiring no commercial reagents or incubation steps. **E. Measured protein concentrations of FPs in lysate with ECmax assay.** Dilution series using lysates obtained by sonication (orange) or chemical lysis using the commercial B-PER reagent (Thermo Fisher, green), were measured with an absorbance scan and get_conc_ecmax(). Data plotted as in B. The top dilution of each series using chemical lysis was removed due to excessive sample scatter. **F. Difference between conversion factors using lysate vs purified protein.** Conversion factors were compared by calculating the fold difference (lysate/purified protein) at each gain. Plots display mean fold differences across the gains and error bars represent the standard deviations.

Figure 4. FP calibration allows comparison across gains and instruments for FPs other than GFP.
A. Calibration of fluorescence per well in units of MEFP. Timecourse experimental data of *E. coli* protein expression may be processed using the process_plate() function using conversion factors obtained from FP calibration. This allows the conversion of normalised data (in relative fluorescence units, rfu) to be converted into calibrated units, of 'molecules of equivalent fluorescent protein' or 'MEFP'. **B. Comparison of normalised vs calibrated data in MEFP.** Starter cultures of *E. coli* DH10B containing pS381_ara_mCherry were transferred into a 96-well plate. mCherry expression was uninduced (grey) or induced 0.1% (black) arabinose at 0 minutes. Absorbance at OD700 and fluorescence was monitored every ten minutes. Data was collected from three biological replicates, each of which is plotted. *Left panel*: normalised mCherry in units of RFU. *Right panel*: calibrated mCherry in units of MEFP. Inset plot shows the same data as the parent plot on a zoomed axis. **C. Calibration of fluorescence per cell in units of MEFP/PEMS.** By combining the FP conversion factors with conversion factors from a microsphere calibration the data can be further processed using the calc_fppercell() function into 'per cell' data with units of MEFP per 'particles of equivalent microspheres'' (PEMS). **D. Comparison of normalised vs calibrated data in MEFP/PEMS.** Expressions were carried out as in B. Normalised and calibrated values are shown when compared across different filter sets (*top,* notation: emission wavelength/bandwidth), instruments (*middle*) and gains (*bottom*). Data for the filter and gain comparisons were taken using the same instrument.

A

conversion factors from
FP calibration

conversion factors from
microsphere calibration

process_plate()

correction for
cell 'quenching'

molecules of (*in vivo*) FP
(molecules)

calc_fppercell()   calc_fpconc()

molecules of FP
per cell

concentration
of FP (M)

B

rfu(+cells)/rfu(−cells)

...ed cells (OD600, cm−1)

calibrant
○ mCherry
△ mGFPm3
✕ mTagBFP2

mCherry (mefp/pems)

time (min)

without_correction   ara (%)
with_correction   ○ 0.001
○ 0.1

C *OD-specific cell concentration*
(eg. from microsphere calibration)

pS361   pS381

mCherry/cell
(molecules per cell)

time (min)   time (min)

pS361   pS381

FP/cell
(molecules per cell)

ara (%)

D *OD-specific cell volume*
(eg. from Volkmer 2011 data)

pS361   pS381

mCherry conc (nM)

time (min)   time (min)

pS361   pS381

FP conc (nM)

ara (%)

ara (%)
● 0
● 3e−05
● 1e−04
● 3e−04
● 0.001
● 0.003
● 0.01
● 0.03
● 0.1

mCherry
mTagBFP2

Figure 5. Absolute quantification of *E. coli* timeseries data in molecules per cell.
A. Functions to convert experimental data to absolute units. Diagram of modifications to the process_plate() function to (i) incorporate a compensation step based on a quantitative understanding of the impact of cell density on apparent fluorescence (this allows the units to be recorded as molecules per cell), and (ii) to calculate molecular concentration of each FP instead, in molar units. **B. Quantification of the quenching effect on fluorescence on three FPs.** Purified FPs were mixed with non-fluorescent *E. coli* at a range of concentrations, and OD600 and fluorescence intensity were recorded. After normalising for cellular autofluorescence, the fold differences between relative fluorescence intensity (rfu) with (+) and without (-) added cells was quantified *(left)*. Data was collected in duplicates, with both points plotted. A model was fit through this data to enable prediction of expected fluorescence quenching for a given cell density on experimental data. An example of the effect of the correction (*right panel*). An mCherry expression vector induced with low (open circles) and high (closed circles) concentration arabinose is presented without (grey) and with (black) correction. Data was collected from three biological replicates and all points are plotted. **C-D. Absolute protein quantification in molecules per cell (C) and molar concentration (D).** mCherry expression (*top*) from medium (pS361, p15A) and high (pS381, colE1) copy vectors was induced at a range of arabinose concentrations and quantified in a timecourse assay in a calibrated plate reader. Data was processed as described in (A) and cell estimates based on microsphere calibrations were used to calculate per cell values (C), or OD-specific cell volume data from Volkmer et al., 2011 was used to calculate molar concentrations (D). Data was collected from three biological replicates, each of which is plotted. (*Bottom*) mCherry expression from the top panel is compared with mTagBFP2 expression from an identical assay, plotted against arabinose concentration at the 420 min post induction. Data was collected from three biological replicates. Displayed points show the mean, and error bars indicate standard deviations.

Figure 6. Evaluation of OD600:OD700 ratios in cell growth assays.
A. Cell count accuracy with FPs. In bacterial assays, cell counts are obtained from calibrated OD600 or OD700 measurements on the assumption that the only contributor to absorbance at 600 or 700nm are cells. Cellular OD600:OD700 ratios (*top plot*, data from Supplementary Fig. 11) approximate 1.3. Red FPs like mCherry absorb light at wavelengths used for cell density assessments (*bottom plot*, data from Supplementary Fig. 1B) and may lead to an error in cell count estimates. **B-D. Timecourse expression assays of three FPs** from the pS381 vector were monitored for OD600, OD700 and fluorescence intensity every 30 minutes, without (black) and with (red) arabinose. Data was collected from three biological replicates, and all points are plotted. This is a representative experiment of at least two independent experiments for each FP. Note OD600:OD700 ratios once again approximate 1.3. **B. Relationship between OD700 and OD600 values.** Linear models were fitted to data from three FPs, and shown above each plot. **C. OD600 error plotted against time.** OD600 error was obtained by dividing observed values by expected values. Expected values were calculated from the OD700 values and the measured OD600~OD700 relationship of the 0% arabinose sample (see B). **D. Effect of FP abundance on OD600 error.** The OD600 error was plotted against abundance of each FP in molecules/cell. Grey shading indicates the mean of the samples without arabinose ± 2*standard deviations.