Estimating Microbial Population Data from Optical Density

Authors: Portia Mira[1], Pamela Yeh[1,2], Barry G. Hall[3]

Institutional Affiliation:

[1]Department of Ecology and Evolutionary Biology, University of California, Los Angeles,

U.S.A.

[2]Santa Fe Institute, Santa Fe, New Mexico, U.S.A.

[3]Bellingham Research Institute, Portland, OR, USA

Abstract
The spectrophotometer has been used for decades to measure the density of bacterial populations using the turbidity expressed as optical density – OD. However, the OD alone is an unreliable metric and is only proportionately accurate to cell titers to about an OD of 0.6. The relationship between OD and cell titer depends on the configuration of the spectrophotometer, the length of the light path through the culture, the size of the bacterial cells, and the cell culture density. We demonstrate the importance of plate reader calibration to identify the exact relationship between OD and cells/ml. We use four bacterial genera and two sizes of micro-titer plates (96-well and 384-well) and show that the cell/ml per unit OD depends heavily on the bacterial cell size and plate size. We applied our calibration curve to real growth curve data and conclude the cells/ml – rather than OD – is a metric that can be used to directly compare results across experiments, labs, instruments, and species.

Introduction

The Beer-Lambert law (1) relates the molar concentration of a solute to absorbance of light according to the equation $C = \epsilon A$ where ϵ is the molar extinction coefficient and A is the absorbance. Epsilon (ϵ) is given at a specific wavelength and specific light path, usually a 1 cm light path. That relationship is what allows us to monitor enzyme reactions by absorbance, to measure protein concentrations by absorbance, and to do enzyme-linked immunoassays (ELISA).

The Beer-Lambert law, however, applies only to solutions in which molecules of solute are uniformly distributed throughout the solvent. It does not apply to suspensions of particulate matter such as microbial cells. Rather than absorbing light, particles scatter light, which is why we express turbidity as OD (optical density) instead of A (absorbance). The relationship between cells/ml and OD is a complex one and depends on several factors including length of light path, size of the particles (cells), and number of particles. There is no simple factor equivalent to ϵ that relate number of cells/ml to OD.

It is not a trivial matter to determine the number of cells/ml or the mass of cells in a culture. The classic way was to dry a culture and weigh the cells, a method that does not lend itself easy measurement of cell densities in small cultures, (to say nothing of the fact that while weighing the dehydrated cells, they absorb moisture from the air and the weight increases even as the balance is watched!). It can be important to determine cell densities easily and quickly, i.e., when monitoring growth in fermenters to determine when to harvest cells.

The convenience of measuring cell populations in microtiter plate readers led us to determine the relationship between OD and cells/ml for several microbial species and for plates of different sizes. Given that relationship OD can be used to calculate cell numbers just as A is used to calculate concentration of a solute.

Spectrophotometers have been used for over 6 decades as a means of measuring the population density of microbial cultures (2-4). Population density is estimated from the turbidity of the culture and is typically expressed as OD (optical density), typically at a wavelength of 600 nm. OD is the negative log of transmittance, which is the fraction of the light that is detected when it is passed through a cuvette that contains a sample of the culture. The Beer-Lambert law states that OD is proportional to the concentration of a solution (1). However, this law does not apply to suspensions of particles (or bacterial cultures) because instead of absorbing light, light is scattered off the axis of the detector (5, 6). As a result, the OD is proportional to the cell titer only up to a limited point, typically an OD of about 0.6. Above that range, some of the light that is scattered away from the detector by one cell is subsequently scattered back to the detector by another cell (7). As a result, the OD does not increase as fast as does the cell titer and therefore, one cannot rely on OD alone to accurately measure bacterial population densities.

To precisely estimate cell titers from observed OD measurements, it is necessary to calibrate the spectrophotometer. The relationship of OD to cell titer depends on four components: 1) the configuration of the spectrophotometer, 2) the length of the light path through the suspension, 3) the size of the cells, and 4) the cell culture density. Therefore, it is necessary to calibrate each spectrophotometer model separately for each microbial species that is to be studied.

Until about a decade ago ODs were determined by putting a sample of the culture into a cuvette of, typically, a 1 cm light path. Determining the growth rate required sampling from a culture at timed intervals and recording the OD at each time point. In practical terms it was difficult to follow more than about 20 cultures simultaneously. The advent of using a microtiter plate reader to monitor the growth of cultures in the wells of a microtiter plate permits high throughput measurements of microbial growth kinetics. However, the same considerations of calibration apply to microtiter plate readers as to spectrophotometers (7). Microtiter plates have various sizes (i.e., 96-wells, 384-wells) which means each well has different depths. Therefore, it is necessary to calibrate a plate reader separately for each size plate.

Here, we demonstrate the importance of calibrating a plate reader using a Biotek Epoch 2 plate reader, both 96-well and 384-well microtiter plates and four bacterial species that span a wide range of cell sizes. We then apply the calibration to a set of growth curves for *Escherichia coli* and show that using cells/ml yields the same growth rates as using OD.

Results

Calibration curves

Standing overnight cultures for each organism (*E. coli DH5α , S. epidermidis, B. megaterium and P. putida)* were concentrated to about 2.5×10^9 cells/ml in mineral salts (M9) buffer, 2x serially diluted. Each dilution, plus a buffer blank, was distributed to four wells (96 well plate) or 6 wells (384 well plate) and the ODs were measured. For each dilution, the mean OD was corrected by subtracting the mean OD of the blank (buffer) well, and corrected ODs were graphed vs the number of viable cells. Stevenson et al (7) suggested that a quadratic relationship exists between cell number and OD. However, to identify the best possible fit, we wanted to explore other relationships. We fit curves to the resulting points based on assumption of four relations- a linear relationship, a quadratic relationship, a cubic relationship, and a polynomial of degree 4 relationship. *E. coli* fits are shown as representative data (Figure 1) and the corresponding R^2 values, the correlation coefficients of the fits, for the other genera measured are also shown in Table 1.

The linear fit is clearly inappropriate, with $R^2 = 0.95$ for both 96 and 384 well plates. To choose among the other fits, we considered R^2 as a measure (Table 1). We found that the R^2 criterion for the polynomial of degree 4 fit is the best for *E. coli.*

We similarly calibrated the plate reader with *Staphylococcus epidermidis*, *Pseudomonas putida*, and *Bacillus megaterium*. In each case the polynomial of degree 4 was the best fit. The approximate cells/ml with the polynomial of degree 4 using the general equation: A OD^4+ B OD^3 + C OD^2 + D OD + E where A, B, C, D and E are the coefficients of the terms. Table 2 shows the polynomial degree 4 equations for each organism and plate size. We also considered another criterion for goodness of fit, Root Square Mean Error (RMSE) (data not shown). The smaller is RMSE the better is the fit. By RMSE criterion, polynomial degree 4 was also consistently the best fit.

The equations are different for each species, and within a species for 96 and 384 well plates (Table 2). This emphasizes the need to calibrate each species and plate size separately. We

provide these equations solely as examples, and we emphasize that they should not be used for instruments other than the Biotek Epoch 2.

The Cells/ml @ OD = 1 decreases as CV (cell volume) increases according to a quadratic function in which Cells/ml @ OD = 1 is 2.1 e8 x CV^2 - 5.9 e9 x CV +4.0 e10, with R^2 = 0.998 for 96-well plates and 1.2 e8 x CV^2 - 3.4 e9 x CV +2.3 e10, with R^2 = 0.999 for 384-well plates. This is consistent with Koch's 1961 and Stevenson et al's 2016 finding (2, 7).

<u>Application of calibration curve to real growth curve data</u>
The growth of two *E. coli* strains and one *S. epidermidis* strain at 37˚ in LBD medium was monitored. Population density was measured as OD and cells/ml based on a quadratic-fit calibration curve. Figure 2 shows a plot of one well for *S. epidermidis* strain and highlights that the curves based on OD and cells/ml are almost identical. For the *S. epidermidis* culture in Figure 2, the growth rate based on OD was µ=0.01459 ± 0.000412 min^{-1} based on 6 points from 140 through 240 minutes, with R = 0.9984. Based on cells/ml, the growth rate was similar, µ=0.01354 ± 0.000285 min^{-1} based on 6 points from 140 through 240 minutes, with R= 0.9983.

The program GrowthRates (8) version 5.1 (https://bellinghamresearch.com/) was used to estimate the growth rates in 12 wells for *E. coli*K12 strain DH5, the uropathogenic *E. coli* strain CFT073 (9), and *S. epidermidis* strain. ATCC 12228 (10). We found the growth rate estimates similar when comparing OD to cell/ml using the polynomial degree 4 fit. The growth rate estimated from cells/ml was significantly different from the growth rate based on OD for *E. coli* CFT073 and *S. epidermidis* (Table 3).

The growth rates estimated from OD and from cells/ml are not the same. Which estimates should we trust more? We trust the rates based on cells/ml because at higher ODs the OD readings fall off as the true population density (cells/ml) increases.

Discussion

Our work highlights the importance of calibrating a microplate reader. We use four different bacterial genera to explore the relationships between OD and cells/ml. First, we compared a quadratic, cubic and polynomial degree 4 fit to bacterial growth data and show that for all four genera, the best calibration fit is a polynomial of degree 4 (Table 1). To highlight the importance of calibrating the plate reader separately for 96-well and 384-well plates, we show the differences in the polynomial degree 4 equations. This difference likely arises from the different culture depths, hence different light path lengths, in 96- vs 384- well plates (Table 2). We also emphasize the importance of separate calibrations for each genus (Table 2). The calibration coefficients depend upon the cell volume, with the sum of those coefficient decreasing as a cubic function as the microbial cell volume increases. Good calibration and application of the calibration curve clearly depends upon consistent well volumes, not only within a single experiment, but between experiments.

Growth rates estimated from OD and cells/ml are not identical (Table 3), but we trust the rates estimated from cells/ml more than those estimated from OD.

Why is it worth the effort to calibrate a plate reader? First, because it allows us to express the maximum population density, i.e., the carrying capacity of the medium, in terms of cells/ml rather than OD. Consider the maximum OD for *E. coli* CFT073 and *S. epidermidis* in 96 well plates (Table 3). These values are very similar (0.622 and 0.610 respectively), but for *E. coli,* the OD of 0.622 represents only 8.2×10^8 cells/ml. On the other hand, the OD of 0.610 for *S. epidermidis* represents 1.25×10^{10} cells/ml. This is a fifteen-fold difference in the number of cells per milliliter in each overnight culture. This difference is important to consider when performing experiments that depend on the number of cellular divisions or cells present, such as cellular communication (11, 12) and antibiotic susceptibility (13-16) and biofilms (17).

Knowing the relationship between OD and cells/ml is not just valuable during growth rate determinations. For instance, when monitoring the growth yield in a fermenter it is very valuable to know the actual population density to decide when to harvest the cells. For S. epidermidis if the yield according to OD is OD = 2.5 that is five times the yield when OD = 0.5. The cells/ml at OD = 2.5 is 1.2×10^{12}, which is 125 times the yield when OD = 0.5 (9.5×10^9 cells/ml).

Probably the most important reason to calibrate plate readers is to use a consistent metric for expressing population densities. By expressing population densities in cells/ml, rather than OD, experiments can be directly compared from different instruments, different labs, and even different genera. Our work shows that using cells/ml as a metric permits reliable measurements of growth rates as does using OD (Table 3) because cell/ml allows consistency when expressing population densities. To measure bacterial growth rates more precisely, we encourage all to calibrate their instruments and to express their results in cells/ml.

Methods
Bacterial Strains
We used four bacterial strains of different genera: *Escherichia coli* K12 strain DH5 (F– φ80lacZΔ M15 Δ (*lacZYA-argF*) *U169 recA1 endA1 hsdR17* (rK– mK+) *phoA supE44 λ- thi–1 gyrA96 relA1*) from ThermoFisher, *Escherichia coli* strain CFT073 O6:K2:H1 (9), *Pseudomonas putida* strain ATCC 12633*, Staphylococcus epidermidis* strain ATC12228 (10) and *Bacillus megaterium* strain ATCC 14581.

Plate Reader Calibration
To identify the colony forming units per genus, we inoculated four standing overnight cultures for each genus. Cultures were inoculated at 37ºC in 10mL of LB (10 g tryptone, 5 g yeast extract, 10 g NaCl per liter) and placed in 15mL culture tubes with tightly sealed caps that allowed no aeration for 16-18 hours. Cultures of each genus were combined into a 50-mL conical tube and spun down at 4,000 rpm for 15 minutes at 4 ºC then resuspended in 4mL M9 buffer, this led to a 4X concentrated starting bacterial culture. Two sets of dilutions were made. The first, starting cultures were diluted at 10^6, 10^7, 10^8 for *E. coli,* and 10^4 and 10^5 for *S. epidermidis, P. putida,* and *B. megaterium* and plated on LB agar, inoculated at 37 ºC overnight and counted the following day. The second set of dilutions were 15 two-fold dilutions from the starting culture. Each tube was plated in 96-well plates (4 replicates 200 µl per well) and 384-well (6 replicates, 80µl per well) plus blank wells (only M9) for each plate size. The OD_{600} was measured every 5

minutes for 30 minutes using the Biotech Epoch 2 plate reader. Aggregated data was used in combination with the colony counts to obtain the individual calibration for each genus.

<u>Growth Rate Experiments</u>
Standing overnight cultures of *E. coli* and *S. epidermidis* were diluted (1:20) to obtain a starting OD of 0.02 – 0.03. Cultures were then plated across the row of a 96-well plate (12 replicate wells) and the OD_{600} was measured every 20 minutes for 22 hours. The growth rates were calculated from the OD measurements using the program GrowthRates (8) version 5.1 (https://bellinghamresearch.com/).

Acknowledgements:
We are grateful for funding from a KL2 Fellowship (PJY) through the NIH/National Center for Advancing Translational Science (NCATS) UCLA CTSI Grant Number UL1TR001881 as well as the Presidential Postdoctoral Fellowship to Dr. Portia Mira.

References

1. Swinehart DF. 1964. The beer-lambert law. Journal of chemical education 39.7:333.
2. Koch A. 1961. Some calculations on the turbidity of mitochondria and bacteria. Biochemica et Biophysica Acta 51:429-441.
3. Ur A, Brown DF. 1975. Impedance monitoring of bacterial activity. J Med Microbiol 8:19-28.
4. Koch AL. 1970. Turbidity measurements of bacterial cultures in some available commercial instruments. Anal Biochem 38:252-9.
5. Zaccanti G, Del Bianco S, Martelli F. 2003. Measurements of optical properties of high-density media. Appl Opt 42:4023-30.
6. Koch AL, Ehrenfeld E. 1968. The size and shape of bacteria by light scattering measurements. Biochim Biophys Acta 165:262-73.
7. Stevenson K, McVey AF, Clark IBN, Swain PS, Pilizota T. 2016. General calibration of microbial growth in microplate readers. Sci Rep 6:38828.
8. Hall BG, Acar H, Nandipati A, Barlow M. 2014. Growth rates made easy. Mol Biol Evol 31:232-8.
9. Welch RA, Burland V, Plunkett G, 3rd, Redford P, Roesch P, Rasko D, Buckles EL, Liou SR, Boutin A, Hackett J, Stroud D, Mayhew GF, Rose DJ, Zhou S, Schwartz DC, Perna NT, Mobley HL, Donnenberg MS, Blattner FR. 2002. Extensive mosaic structure revealed by the complete genome sequence of uropathogenic Escherichia coli. Proc Natl Acad Sci U S A 99:17020-4.
10. Zhang YQ, Ren SX, Li HL, Wang YX, Fu G, Yang J, Qin ZQ, Miao YG, Wang WY, Chen RS, Shen Y, Chen Z, Yuan ZH, Zhao GP, Qu D, Danchin A, Wen YM. 2003. Genome-based analysis of virulence genes in a non-biofilm-forming Staphylococcus epidermidis strain (ATCC 12228). Mol Microbiol 49:1577-93.
11. Bassler BL. 2002. Small talk. Cell-to-cell communication in bacteria. Cell 109:421-4.
12. Salmond GP, Bycroft BW, Stewart GS, Williams P. 1995. The bacterial 'enigma': cracking the code of cell-cell communication. Mol Microbiol 16:615-24.
13. Andrews JM. 2001. Determination of minimum inhibitory concentrations. J Antimicrob Chemother 48 Suppl 1:5-16.
14. Bollenbach T, Quan S, Chait R, Kishony R. 2009. Nonoptimal microbial response to antibiotics underlies suppressive drug interactions. Cell 139:707-18.
15. Mira P, Barlow M, Meza JC, Hall BG. 2017. Statistical Package for Growth Rates Made Easy. Mol Biol Evol 34:3303-3309.
16. Seigal A, Mira P, Sturmfels B, Barlow M. 2017. Does Antibiotic Resistance Evolve in Hospitals? Bull Math Biol 79:191-208.
17. Kirby AE, Garner K, Levin BR. 2012. The relative contributions of physical structure and cell density to the antibiotic susceptibility of bacteria in biofilms. Antimicrob Agents Chemother 56:2967-75.

Figures:

Figure 1: Relationship between cells/ml and OD with different fits. The solid lines and round points represent 96-well measurements, and the dashed lines and square points represent 384-well measurements for *E. coli.*

Figure 2. Growth curves of *S. epidermidis* in one well based on different measures of population density. The natural log of OD (circles) and cells per mL (squares) is plotted over time (minutes).

Tables:

Table 1: R^2 values for the different fits for each of the four bacterial genera measured.

Organism	96 well			384 well		
	Quadratic	**Cubic**	**Polynomial**	**Quadratic**	**Cubic**	**Polynomial**
S. epidermidis	0.9987	0.99995	1.0	0.99989	0.99995	0.99999
E. coli	0.99452	0.99961	1.0	0.99815	0.99995	1.0
P. putida	0.99824	0.99996	0.99996	0.99821	0.99987	0.99992
B. megaterium	0.99977	099999	1.0	0.99882	0.99936	0.99997

Organisms are listed in order of cell volume.

Table 2: Calibration equations and cell size

Species	Plate Size	Polynomial Degree 4 Equation, Cells/ml =	Cells/ml @ OD = 1	Cell Vol
S. epidermidis	96	$4.3e10\ OD^4 -3.8e10\ OD^3 +1.2e10\ OD^2 +1.7e10\ OD +1.7e8$	3.42×10^{10}	$1\ \mu^3$
	384	$-2.5e10\ OD^4 +4.2e10\ OD^3 -1.2e10\ OD^2 +1.5e10\ OD +9.4e7$	2.01×10^{10}	
E. coli	96	$1.6e9\ OD^4 -2.3e9\ OD^3 +1.3e9\ OD^2 +1.0e9\ OD +5.1e5$	1.6×10^9	$9.8\ \mu^3$
	384	$3.3e8\ OD^4 -2.1e8\ OD^3 +4.9e8\ OD^2 +8.3e8\ OD + 4.2e4$	1.44×10^9	
P. putida	96	$2.4e8\ OD^4 -2.7e8\ OD^3 +6.4e7\ OD^2 +4.7e8\ OD +1.4e5$	5.04×10^8	$12.3\ \mu^3$
	384	$3.7e8\ OD^4 -6.8e8\ OD^3 +3.8e8\ OD^2 +3.7e8\ OD +8.2e5$	4.41×10^8	
B. megaterium	96	$4.4e8\ OD^4 -4.9e8\ OD^3 +3.2e8\ OD^2 +5.9e8\ OD +4.3e6$	8.64×10^8	$17\ \mu^3$
	384	$-1.2e9\ OD^4 +3.6e9\ OD^3 -2.7e9\ OD^2 +1.2e9\ OD -5.3e6$	8.95×10^8	

Table 3: Comparison of growth rates based on OD and cells/ml.

	E. coli K12 DH5α			*E. coli* CFT073			*S. epidermidis* ATCC 12228		
	μ	R	max OD or cells/ml	μ	R	max OD or cells/ml	μ	R	max OD or cells/ml
OD	0.0212	0.9968	0.348	0.0235	0.9989	0.622	0.0155	0.9989	0.610
cells/ml	0.0207	0.9984	4.36×10^8	0.0253	0.9990	8.21×10^8	0.0139	0.9986	1.25×10^{10}
p-value	0.55			0.0024			9.19e-6		

Values are means of 12 replicates. In all cases the S.E. was < 0.05 of the mean.