Vendor-neutral sequences (VENUS) and fully transparent workflows

improve inter-vendor reproducibility of quantitative MRI

Agah Karakuzu[1,2*], Labonny Biswas[3], Julien Cohen-Adad[1,4,5], Nikola Stikov[1,2]

[1] NeuroPoly Lab, Institute of Biomedical Engineering, Polytechnique Montréal, Montréal, QC, Canada

[2] Montréal Heart Institute, Montréal, QC, Canada

[3] Sunnybrook Research Institute, University of Toronto, Toronto, ON, Canada

[4] Functional Neuroimaging Unit, CRIUGM, Université de Montréal, Montréal, QC, Canada

[5] Mila - Quebec AI Institute, Montreal, QC, Canada

* Corresponding Author: Agah Karakuzu

2500 Chem. De Poyltechnique

H3T 1J4, Montreal, Quebec

Canada

Tel: +1 (514) 340-4711

E-mail: agah.karakuzu@polymtl.ca

Abstract

Purpose: We developed a transparent end-to-end qMRI workflow that starts with a vendor-neutral acquisition and tested the hypothesis that vendor-neutral sequences (VENUS) decrease inter-vendor variability of T1, MTR and MTsat measurements.

Methods: We developed and deployed a vendor-neutral 3D spoiled gradient-echo (SPGR) sequence on three clinical scanners by two MRI vendors and acquired T1 maps on the NIST phantom, as well as T1, MTR and MTsat maps in three healthy participants. We performed hierarchical shift function analysis in vivo to characterize the differences between scanners when VENUS is used instead of commercial vendor implementations. Inter-vendor deviations were compared for statistical significance to test the hypothesis.

Results: In the NIST phantom, VENUS reduced inter-vendor differences from 8 - 19.4% to 0.2 - 5% with an overall accuracy improvement, reducing ground truth T1 deviation from 7 - 11% to 0.2 - 4%. In vivo we found that the variability between vendors is significantly reduced ($p = 0.015$) for all maps (T1, MTR and MTsat) using VENUS.

Conclusion: We conclude that vendor-neutral workflows are feasible and compatible with clinical MRI scanners. The significant reduction of inter-vendor variability using VENUS has important implications for qMRI research and for the reliability of multicenter clinical trials.

Introduction

As the invention of MRI approaches its 50th anniversary[1], the notion of image acquisition has almost become synonymous with data collection. A major driving force in the transformation of MR images from mere pictures into mineable data[2] is attributing physiologically relevant physical parameters to the voxels, namely quantitative MRI (qMRI). MRI is not a quantitative measurement device by design. Nonetheless, systematic manipulation of effective micrometer-level MRI parameters via specialized acquisition methods, followed by fitting the resulting data to a signal representation or a biophysical model[3] can yield parametric maps, turning scanners into quantitative diagnostic tools. Despite being as old as MRI itself, most of the qMRI methods have not succeeded to find their widespread use in the clinic, at least in part due to a major multicenter reproducibility challenge.

The introduction is organized around two problems hampering multicenter reproducibility of qMRI, which this study seeks to address:

1. Lack of transparency in vendor implementations and multicenter consistency of pulse sequences that are commonly used in qMRI
2. Technical roadblocks in the way of deploying a standardized pulse sequence along with a unified user interface to multiple imaging sites

T1 relaxometry is a clear example of how availability, transparency and multicenter consistency of pulse sequences influence multicenter reproducibility. Several methods such as inversion-recovery spin-echo (IR-SE) [4] variable flip angle (VFA-T1)[5] Look-Locker IR [6] and magnetization-prepared two rapid-echoes [7] have gained popularity in MRI research. Although measured T1 values can exhibit up to 30% inter-sequence variability in the same scan session for the same participant [8], a selected T1 relaxometry method is much more reliable within-site [9,10]. As for the multicenter stability, MP2RAGE appears to be a promising T1 mapping method at 7T with a single vendor considered across 8 imaging sites [11]. On the other hand, substantial multicenter variability is reported for another popular whole-brain imaging method VFA-T1, both in-vivo and in the system phantom [12,13]. Several factors contribute to the variability of the VFA-T1 measurement, including B1 field inhomogeneity [8], incomplete spoiling [14], sequence parameters and bore temperature [13], and uncontrolled magnetization transfer (MT) effects [15]. Because of all these diverse confounders of T1 stability, the healthy range of in-vivo T1 values at 3T remains elusive [16,17]. This constitutes a critical problem for the potential use of T1 relaxometry in clinics.

Considerable amount of research has focused on measurement bias stemming from the imperfections

related to image acquisition. However, reproducibility of the developed techniques is often hindered by

problem 1. For example, a simple yet powerful B1 correction framework for VFA-T1 has been

established [18], but B1 mapping sequences are typically not available in commercial systems or available

sequences are different across vendors. This not only imposes a practical challenge in evaluating the

reliability of VFA-T1 measurements across vendors [13,19] but the differences between vendor-native B1

mapping methods can aggravate the instability [20]. Another example is the spoiling gradient area and RF

spoiling phase increment in the commercial implementations of 3D spoiled gradient-echo (3D-SPGR)

sequences. Both parameters determine the accuracy of VFA-T1 mapping [14] However, vendors are known

to set different defaults for these parameters, rendering stock 3D-SPGR sequences of two major MRI

manufacturers unfit for this application [21]. Note that neither parameter can be controlled or displayed at

the scanner console. Similarly, fundamental properties of the excitation pulse (e.g., pulse shape, time-

bandwidth product, duration) are not disclosed and it is not known how these properties are adjusted

under different SAR requirements. To achieve a standardized 3D-SPGR acquisition for T1 mapping,

such parameter configurations should be disclosed, made accessible and equalized across scanners for

eliminating systematic biases. Recently, Gracien et al. showed a successful example of how this solution

can reduce systematic biases in relaxometry mapping between two different scanner models from the

same vendor [22].

Addressing inadequacies of model assumptions constitutes another solution toward improving the

reliability of qMRI methods [3]. For example, balancing the total amount of RF power deposited by each

run of a VFA acquisition using multiband non-selective pulses, Teixeira et al. enforced two-pool MT

systems to behave like a single pool MT system [15]. They showed that making single pool assumption of

relaxometry methods valid through controlled saturation of MT increases the measurement reliability.

Although this technique holds important implications for multicenter reproducibility of qMRI, deploying

it to multiple sites is not a straightforward process. Moreover, proprietary programming libraries of

different scanner manufacturers may not be flexible to allow identical implementations, exemplifying

the constraints imposed by *problem 2*. Another model-related improvement has been recently introduced

to replace the fixed empirical correction factor of MT saturation (MTsat) B1 correction [23] with a

correction factor map to reduce B1 dependency of MTsat maps [24]. The proposed methodology calls for

the details on saturation and excitation pulses (e.g., pulse shape, frequency offset, pulse duration, etc.)

as the correction framework is based on simulations. From the standpoint of *problem 1,* such information

is not easily accessible in the stock sequence, so the correction cannot be applied. From the perspective

of *problem 2,* deploying sequences to multiple centers with known saturation and excitation pulse

parameters may not be realistic within the confines of vendor development environments. Even though

both studies made their code publicly available to facilitate the reproducibility of their work [25], black-

box vendor strategies thwart these valuable efforts.

Fortunately, there are several open-source pulse sequence development platforms to contend with

problem 2 [26-32]. These platforms have in common the ability to interpret and translate the same sequence

logic for multiple vendors, considerably reducing multi-center development efforts and minimizing

implementation variability. Another advantage of these tools is to attract community-driven

development. For example, Pulseq has received considerable community attention to motivate the

development of sequences in Python[29], or even going beyond code to graphically assemble [33] Pulseq

descriptions using Pulseq-GPI[34]. Currently, Pulseq can be operated on two major clinical scanners

(Siemens and GE) and three pre-clinical scanner platforms. There is recent literature showing the

feasibility of Pulseq for performing multicenter qMRI studies. For example, a standardized chemical

exchange saturation (CEST) protocol has been developed and deployed on three Siemens scanners,

where two of the systems had different vendor software versions[35]. Results by Herz et al. showed

multicenter consistency for an advanced CEST method, which has been made publicly available for both

Python and Matlab users. Another recent Pulseq study performed inversion-recovery T1 mapping and

multi-echo spin-echo T2 mapping on two Siemens scanners at 1.5 and 3T using the system phantom[36].

In that study the reference T1 mapping method[37] accurately estimated T1 values within an 8% error

band, whereas the T2 accuracy was slightly reduced. Taken together, these studies reveal the vital role

of vendor-neutral pulse sequences in standardizing qMRI across centers. However, whether a vendor-

neutral approach can improve quantitative agreement between scanners from different vendors has

remained an open question.

The focus of earlier open-source pulse sequence platforms was providing a rapid and unified prototyping

framework for facilitating interoperability, so some of the most adjusted scan parameters (e.g., field of

view) had to be fixed once the sequence was downloaded to the scanner. More recent solutions such as

GammaStar[32] can remove this limitation by enabling user interaction trough the vendor's user interface

to modify fundamental protocol settings during the imaging session. Offering a more complete solution

to *problem 2* through on-the-fly sequence updates, GammaStar eases the collaborative sequence

development process by providing a web-based interface that can be accessed online, executed in a

Docker container or locally. Although such technical improvements reduced the entrance barrier for free
sequence development, exchange and standardization, validation for safety and multi-center use has
remained elusive. Recently, Tong et al. (2021) proposed a framework for testing, documenting and
sharing open-source pulse sequences[36], which adds an important missing piece to the community-driven
MRI development puzzle.

RTHawk[38] is another vendor-neutral solution, which is a proprietary platform for MRI software
development. As it is utilizing the same infrastructure as an FDA approved (510(k), No: K1833274)
cardiac imaging platform, it ensures operation within MRI hardware and safety limits. Unlike the above-
mentioned solutions, RTHawk provides a remote procedure call (RPC) server that replaces the vendor's
pulse sequence controller to orchestrate vendor-specific low-level hardware instructions. The RPC pulse
sequence server receives control commands and relevant sequence components (i.e., RF and gradient
waveforms, ADC and timing events designed in SpinBench, as shown in Fig. 1e) directly from a
standalone Ubuntu workstation connected to the scanner network (Fig. 1c). This gives the flexibility to
issue synchronous or asynchronous updates to a sequence in real-time, such as scaling/replacing
waveforms between TRs or changing the volume prescription. As the sequence control manager is
decoupled from the vendor's workstation, RTHawk makes it possible to develop a vendor-neutral
unified user interface (UUI) per application (Fig. 1d). In addition, collected raw data is streamed over to
the standalone Ubuntu workstation through a real-time transport protocol (RTP). The RTP data manager
enables adding or changing the metadata associated with each observation, which enables exporting raw
and reconstructed images in community data standards (Fig. 1f). Together, these components offer a
complete solution to *problem 2* by enabling the design of a unified user console that can automate qMRI
protocols (e.g., MTsat), dispensing with the manual effort of keeping prescan calibrations constant
across multiple runs and removing the need for vendor-specific data handling. Although the application
programming interface (API) to develop and the runtime to execute RTHawk applications are
proprietary components, the GUI-based sequence design and simulation engine (SpinBench) is freely
available for non-commercial use (Fig. 1e). In addition, researchers are allowed to share source-code
that uses RTHawk libraries for the acquisition control and image reconstruction[39]. As a result, these
sequences can be made open source.

Aside from vendor-neutral experiments, researchers looked at improving qMRI stability by customizing
vendor-native implementations and equalizing parameters to the utmost extent possible [22,40,41]. However,
downstream data harmonization methods were still needed to correct for certain inter-vendor

differences[41], or some of the bias could not be removed altogether, as not all qMRI-relevant settings can

be unified using a vendor-native approach[40]. This approach uses intricate vendor programming interfaces

and becomes less feasible in a multicenter setting. In addition, there is no guarantee of inter-vendor

generalization even for scanners with comparable hardware specs, as the selection of sequence design

elements is unique to each vendor. Even within the confines of a single vendor, recent studies have

demonstrated the promise of a vendor-neutral approach to qMRI in standardizing methods between

centers[35,36].

In this study, we test the hypothesis that our vendor-neutral sequence (VENUS) reduces inter-vendor

variability of T1, MTR and MTsat measurements. To test this hypothesis, we compared vendor-native

T1, MTR and MTsat maps[42] with those obtained by developing and using VENUS in three healthy

participants, across three different scanners models from two manufacturers at 3T.

Methods

Vendor-neutral pulse sequence development

We deployed vendor-neutral pulse sequences developed in RTHawk v3.0.0 (rc4-28-ge3540dda19)

(HeartVista Inc., CA, USA) on three 3T systems: i) GE Discovery 750 software version DV25

(R02_1549.b) (GE Healthcare, Milwaukee, MI, USA), ii) Siemens Prisma software version

VE11C (N4_LATEST_20160120) (Siemens Healthineers, Erlangen, Germany) and iii) Siemens

Skyra with the same software version as (ii). Throughout the rest of this article, these scanners will

be referred to as G1, S1 and S2, respectively (Fig. 1a). In all the experiments, RTHawk was

connected to the vendor's host workstation through the local area network using the SSH protocol

to drive the scanner hardware and to collect the raw data (Fig 1b).

General design considerations

All vendor-neutral protocols are based on a 3D SPGR pulse sequence [43], with the RF, gradient

waveforms, and the readout scheme developed from scratch as independent sequence blocks in

SpinBench v2.5.2, a graphical user interface (GUI) based pulse sequence design and Bloch

simulation software (Fig. 1e). Sequence stubs that correspond to a combination of these blocks

were ordered in a configuration file (application.apd) and linked to a scan control manager script

(control.js). Specifically, this script was responsible for shifting the timing blocks and scaling or

enabling/disabling the gradient waveforms in the stub to play the sequence following the

parameters requested by the user. To enable such interactions, a UUI (control.ui) was developed

for quantitative imaging, allowing the user to iterate through relevant acquisition parameters (e.g.,

FA, TR, and MT pulse for MTsat) from one simple panel that is vendor-neutral (Fig.1d). The

RTHawk real-time acquisition controller was configured to synchronously update these

parameters at the end of each individual acquisition in the quantitative protocol (e.g., a flip angle

image in the VFA-T1 mapping protocol). Identical scan geometry and pre-acquisition settings

were transferred between each individual acquisition. To avoid signal clipping, the highest SNR

acquisitions (e.g., T1w acquisition of the MTsat protocol) were defined as the first iteration. A

simple sum-of-squares multi-coil reconstruction (recon.js) for fully sampled k-space was

developed with a Fermi filter (transition width = 0.01, radius = 0.48, both expressed as a proportion

of the FOV) (Fig. 1f).

All the metadata annotations, accumulation logic of the collected data and naming of the exported

images were designed in full compliance with the community data standards: the Brain Imaging

Data Structure (BIDS) for the reconstructed images[44,45] and ISMRM-RD[46] for the k-space data

(Fig. 1a). Moreover, a custom panel in the UUI was provided for users to define freeform BIDS

naming entities such as participant ID, session, and acquisition labels (Fig. 1d). Other naming

entities relevant to the quantitative acquisitions (e.g., *flip* and *mt* for FA and MT pulse,

respectively) were automatically governed.

The vendor-neutral protocol

A slab-selective (thickness = 50 mm, gradient net area = 4.24 cyc/thickness) SINC excitation pulse

(time-bandwidth product (TΔf) = 8, duration = 1.03ms, Hanning windowed) was implemented

with a quadratic phase increment of 117° for RF spoiling. This was followed by a fully-sampled

3D cartesian readout with outside-in ordering (i.e., sampling the center last), with a padding factor

of 1.2. The default geometry properties were 256x256 acquisition matrix, 25.6 cm FOV and 20

partitions in the slab-selection direction, yielding 1x1x3 mm resolution. The readout gradient had

a rewinder lobe with 2 cyc/pixel net area and was followed by a spoiling gradient with an area of

40 mT·ms/m.

For the magnetization transfer (MT) saturation, a Fermi pulse (duration = 12ms, B1rms = 3.64 uT,

frequency offset = 1.2kHz, transition width = 0.35, max B1 = 5uT, pulse angle= 490°) was

designed as an optional block that can be enabled/disabled by the user. Finally, a loop command

was defined for the sequence to iterate through three sets of parameters, defined by the user in the
UUI for a complete MTsat protocol.

From this protocol we acquired three images: (i) PD-weighted SPGR with no MT, FA = 6° and
TR = 32ms (ii) MT-weighted SPGR with MT, FA = 6° and TR = 32ms (iii) T1-weighted SPGR
without MT, FA = 20° and TR = 18ms. From images (i) and (iii) we computed a T1 map, from
images (i) and (ii) we computed an MTR map, and from images (i), (ii) and (iii) we computed an
MTsat map.

Data acquisition

Experimental procedures agreed with the guidelines and regulations concerning human welfare
and experimentation set by Montreal Heart Institute Ethics Committee, Montreal, Canada and
Sunnybrook Health Center, Toronto, Canada. Three healthy male participants (age = 37 ∓ 6, height
= 178 ∓ 6 cm and body mass = 81 ∓ 4 kg) volunteered for multi-center data collection. Following
a detailed explanation of the purpose and the methodology of the experiment, the participants gave
their written informed consent.

The participants and the ISMRM-NIST system phantom (manufactured by HPD Inc., SN = 42)
were scanned on three imaging systems at two imaging sites. In S1 and S2, the phantom was
scanned using a 20-channel head coil due to space constraints, whereas a 32-channel coil was used
in G1. For in-vivo imaging, 32-channel head coils were used in G1 and S1, whereas a 64-channel
coil was used in S2. S1 was equipped with an XR-K2309_2250V_951A (Siemens Healthineers,
Erlangen,Germany) gradient system (80 mT/m maximum amplitude and 200 T/m/s slew rate per
axis, 48 cm maximum FOV), S2 with an XQ-K2309_2250V_793A (Siemens Healthineers,
Erlangen,Germany) gradient system (45 mT/m maximum amplitude and 200 T/m/s slew rate per
axis, 48cm maximum FOV) and G1 with a 8920-XGD (GE Healthcare, Milwaukee, USA) gradient
system (50 mT/m maximum amplitude and 200 T/m/s slew rate per axis, 48cm maximum FOV).
The nominal field strengths on G1 and S1-2 were 3T and 2.89T, respectively. Before the scan, the
system phantom was kept in the imaging site for at least a day, and in the scanner room for at least
3 hours. The measured bore temperature in G1, S1 and S2 was 20.1°C, 20.2°C and 20.8°C,
respectively.

The acquisition parameters were set according to a generic protocol established for MTsat imaging

of neural tissue[47]. Vendor-neutral acquisition parameters were identical on all systems. However,

it was not possible to equalize all the parameters between the vendor-native protocols. Comparison

of vendor-native and vendor-neutral protocols are presented in Table 1. To scan the system

phantom, prescan measurements were performed as described by Keenan et al. (2021) and the

vendor-neutral acquisitions were configured to start the acquisitions with these calibrations[13]. For

all acquisitions, the prescan settings of the initial T1w acquisition were used for the subsequent

PDw and MTw acquisitions on all vendor systems. Necessary care was taken not to perform

additional calibration measurements between acquisitions. For the VENUS acquisitions, B0

shimming gradients were set using a spiral multi-echo (four echoes) gradient-echo sequence by

fitting the field map obtained after water/fat separation to the first order in 3D. Gradient non-

linearity correction was performed as part of the on-site reconstruction pipeline (Fig. 1f) and the

relevant warping coefficients are reported in the image metadata for offline reconstruction. For the

systems S1-2 the identical protocol was used by exporting the vendor-native protocol files from

S2, whereas the protocols for G1 were set on-site.

Data processing

Vendor-native DICOM images were converted to BIDS v1.6.0, whereas vendor-neutral images

were already exported in the BIDS format. All the processing was performed using data-driven

and container mediated qMRFlow pipelines comprised of two docker images

(qmrlab/antsfsl:v2.5.0b and qmrlab/minimal:v2.5.0b) that are published by qMRLab using a

continuous deployment pipeline. Quantitative fitting was performed in qMRLab[48] v2.5.0b. Pre-

processing steps were performed using ANTs[49] for registration and FSL[50] for automatic gray-

matter (GM) and white-matter (WM) segmentation.

The system phantom T1 pipeline consisted of linearized T1 mapping in the VFA framework by

accounting for varying TRs[23]. Resultant phantom T1 maps were then masked using spherical ROIs

as described in[51]. The center of each sphere was manually prescribed per dataset (3D Slicer v4.11)

at the 10 highest T1 reference spheres of the system phantom (Fig. 4a). For the in-vivo data,

between-scan motion correction was performed by aligning PDw and MTw images onto the T1w,

filled by MTsat fitting (Fig. 3b). Brain region segmentations were performed on the T1w images

and ROI masking was performed to prepare data for statistical analyses (Fig. 3c).

Statistical analyses

All the descriptive statistics were reported by the processing pipeline in tabular format for phantom

and in-vivo maps (available at https://osf.io/5n3c). Vendor-neutral and vendor-native phantom

measurement performances were compared against the reference (Fig. 4b,c) and percent deviations

from the ground truth were reported (Fig. 4d).

Kernel density estimates of the T1, MTR and MTsat distributions in WM and GM were visualized

as ridgeline plots for one participant (Fig, 5d-i). Before the statistical comparisons in WM, these

distributions were filtered to a relevant range discerned from the ridgeline plots (Fig. 5d-i) to

remove spurious values captured by masking. The range was 0 to 3s for T1, 35 to 70% for MTR

and 1 to 8 for MTsat. Filtered distributions were then subjected to random sampling (the random

number generator was seeded for reproducibility) to obtain an equal number of WM voxels (N =

37,000), enabling a balanced comparison.

Percentile bootstrap based shift function analysis[52] was performed to compare dependent

measurements of T1, MTR and MTsat in WM (N=37,000) between different systems (G1 vs S1,

G1 vs S2 and S1 vs S2) for VENUS and for the vendor-native implementations. Deciles of the

distributions were computed using Harrell-Davis quantile estimator [53] and the decile differences

were calculated using 250 bootstrap samples. Note that unlike traditional significance tests that

operate on a single point estimate of the central tendency, shift function analysis can characterize

differences at any location of the distributions (Fig. 6a). Even for a single participant, this provides

a powerful graphical tool to assess how vendor-neutral and vendor-native measurements differ

between different systems throughout the relevant range of the distributions. To supplement the

shift function plots, we annotated the 5th decile (median) differences using respective percent

differences (Fig. 6b-d). The shift function was then extended to a hierarchical design to

characterize the overall difference between scanners for multiple participants, for VENUS and for

the vendor-native WM distributions of all studied metrics. A hierarchical shift function chart

shows individual between-scanner differences for each participant, along with averaged group

differences plotted against a categorical axis at each decile (Fig. 7a). Similarly, percent deviations

between scanners at the median deciles were annotated per subject and the average percent

deviations were reported (Fig. 7b-d). Please note that the hierarchical shift function by default

calculates high density intervals to test the significance of averaged decile differences against the

null value of zero using a t-test at trimmed means. Given the high number of samples and the fact

that the hypothesis aims at comparing between-scanner differences in a VENUS vs vendor-native

setting, these intervals are excluded from the visualization for simplicity. The reader is welcome

to reproduce these figures online, where necessary changes can be made to visualize the intervals

at https://github.com/qMRLab/VENUS.

Finally, quantitative measurement discrepancies of vendor-native and VENUS implementations

between different vendors (G1-S1 and G1-S2) were compared using Wilcoxon signed rank test.

The comparison was performed on the G1-S1 and G2-S2 percent absolute differences of T1, MTR

and MTsat in white-matter between vendor-native and vendor-neutral implementations. The level

of significance was set at 0.05.

Results

Figure 2 compares basic image quality metrics across phantom images. The contrast transitions across 10 spheres included in the analysis are discernible and qualitatively comparable in all T1w images (Fig. 2h-m). In addition, VENUS and vendor-native peak SNR values are on a par with each other, ranging from 25.4-27.8 dB (PDw) and 28.5-31.6 dB (T1w, Fig. 2a). According to the American College of Radiology (ACR) guidelines, the number of visible bright spots within a high-contrast resolution insert determines the resolution performance[54]. In the system phantom, 4x4 1 mm coarse resolution arrays are drilled into an 8 mm thick polyphenylene sulfide plate (plate 4), which falls into the FOV when the T1 array (plate 5) is centered in a 60mm slab[55]. Overall, the resolution markers are conspicuous in vendor-neutral images (Fig. 2b-d) with a slightly less horizontal resolution compared to the S1-2$_{\text{NATIVE}}$ (Fig. 2f,g). Although the number of spots is discernible (Fig. 2e), the insert pattern in the G1$_{\text{NATIVE}}$ image appears smoother.

Figure 4 shows the T1 mapping results from the NIST phantom. Shades of blue (light blue, cyan, teal) stand for the vendor-native T1 measurements, and shades of red (red, orange, yellow) stand for the VENUS T1 measurements. The ground truth NIST phantom values are marked with a cross.

The vendor-native implementations result in overestimation of the ground truth T1 values at different rates, particularly for the physiologically relevant T1 values in the interval from 0.7 to 1.9s (Fig. 4b). Ground truth T1 deviations (ΔT1) calculated by percent error indicate that G1$_{\text{NATIVE}}$ and S2$_{\text{NATIVE}}$ exhibit a persistent overestimation trend, whereas S1$_{\text{NATIVE}}$ shows better accuracy approaching that of VENUS T1 estimations in the relevant range (Fig. 4c). Within the same interval, the highest deviation is observed for G1$_{\text{NATIVE}}$, ranging from 9.7 to 30.4% for ΔT1. For R4-6, G1$_{\text{NATIVE}}$ and G1$_{\text{VENUS}}$ T1 measurements straddle the reference, where G1$_{\text{VENUS}}$ shows 5.1-13.8% underestimation and the G1$_{\text{NATIVE}}$ overestimation remains within the 3.4-10.5% interval (Fig. 4c). For lower T1 reference values (T1 < 170ms), all measurements indicate higher deviations, with S1-2$_{\text{NATIVE}}$ performing better than S1-2$_{\text{VENUS}}$. When the measured T1 values are averaged over S1-2 ($\bar{\text{S}}$), the differences between G1$_{\text{NATIVE}}$ and $\bar{\text{S}}_{\text{NATIVE}}$ are 8, 11, 12.5 and 19.4%, whereas the differences between G1$_{\text{VENUS}}$ and $\bar{\text{S}}_{\text{VENUS}}$ are 5, 2, 2 and 0.2% for R7-10, respectively (Fig. 4d). This reduction in between-vendor differences brought by VENUS is coupled with an improvement in accuracy. When averaged according to the implementation type, average VENUS

deviation ($\overline{\Delta T1_{VENUS}}$) falls within the 0.2 - 4% range and $\overline{\Delta T1_{NATIVE}}$ ranges from 7 to 11%. Even

though $G1_{NATIVE}$ has the dominant contribution to the higher $\overline{\Delta T1_{NATIVE}}$ values, Fig. 4d shows that

$\overline{S}_{VENUS}$ is closer to the reference than $\overline{S}_{NATIVE}$ for the majority of the R7-10 (ΔT1 of 7.6, 3.5, 5.4,

0.7% and 3.2, 0.9, 2, 1.3% for $\overline{S}_{NATIVE}$ and $\overline{S}_{VENUS}$, respectively). As a result, VENUS reduces

between-vendor differences with an overall accuracy improvement.

Figure 5 shows in vivo T1, MTR and MTsat maps from a single participant (P3). While most of

the improvements are evident from the maps (5a - 5c), the ridgeline plots (5d – 5i) make it easier

to appreciate the distribution range of values and VENUS vs vendor-native between-scanner

agreement in the GM and WM per metric. Conforming to the higher myelin content in WM, T1

values are lower (around 1.1 ∓ 0.2s, Fig. 5d), whereas MTR and MTsat values are higher (around

$50 \mp 8\%$ and 3.8 ∓ 0.9 a.u., Fig. 5e,f) in comparison to those in GM (1.9 ∓ 0.4s, $40 \mp 2\%$ and 1.8

∓ 0.5, for T1, MTR and MTsat, respectively, Fig. 5g-i). The general trend observed in the images

is captured by ridgeline plots, showing better agreement between VENUS distributions of G1, S1

and S2. This is further supported by between-scanner coefficient of variation (CoV) per metric

(Table 2), showing that VENUS reduces the CoV from 16.5, 10.1 and 12.5% to 6.1, 4.1 and 4.1%

for T1, MTR and MTsat, respectively. This indicates a sizable decrease in between-scanner

variability using VENUS compared with vendor-native measurements and the trend is consistent

across participants. However, CoV is a relative statistic of dispersion and does not characterize the

nature and absolute magnitude of changes between the measurements. Such characterization is

important to evaluate the between-scanner performance of a quantitative measure in distinguishing

biological variability. To enable such comparisons, Fig. 6b-d shows pairwise difference of each

metric between scanners using shift functions, which are constructed by absolute differences at

each decile of the compared distributions (Fig. 6a). VENUS and vendor-native between-scanner

differences for all metrics are annotated by percent differences at the median (fifth) decile for

simplicity.

Going from vendor-native (top rows, blue panels) to VENUS (bottom rows, red panels), Fig. 6b-

d indicate a decrease in T1, MTR and MTsat WM differences between scanners from different

vendors (G1 vs S1 and G1 vs S2) for P3, without exception and throughout the deciles. As for

within-vendor comparisons (S1vs S2), VENUS reduces difference scores for T1 and MTsat by 5.8

and 7.8% while increasing that for MTR by 5.3%. One can also appreciate the changes in shift

function shapes. For example, shift function for G1$_{NATIVE}$ vs S2$_{NATIVE}$ MTsat comparison in Fig.

6d shows a positive linear trend, indicating that WM voxels with higher MTsat values tend to show

a higher between-vendor difference. On the other hand, G1$_{VENUS}$ vs S2$_{VENUS}$ MTsat shift function

appears flatter, describing a more uniform (and reduced) bias throughout the WM distribution.

Figure 7 expands on Figure 6 for multiple participants by overlaying individual shift functions

(shades of pink) and illustrating the across-participants trend using group shift functions that are

red for VENUS and blue for vendor-native differences (Fig. 7a). A general observation is that

individual shift function shapes are consistent across participants, indicating that the nature of

between-scanner VENUS and vendor-native differences are not affected by anatomical

differences. However, the magnitude of the difference is participant-specific. For example, P3

shows the highest G1$_{NATIVE}$ vs S1$_{NATIVE}$ T1 difference of 31.2%, which is followed by 20.9 and

14.2% for P2 and P1, respectively (Fig. 7b). The same participant order is preserved for G1$_{VENUS}$

vs S1$_{VENUS}$; however, with consistently lower decile differences, which is what stands out in Fig.

7: participant-averaged between-vendor T1, MTR and MTsat WM differences are smaller for

VENUS than they are for vendor-native implementations, without exception and throughout the

deciles. Overall, VENUS G1 – S1 and G1 – S2 differences are on the order of 2.3 to 7.9%, which

starts from 13.8% and extends up to 25.6% for vendor-native implementations, averaged across

participants. The reduction in between-vendor differences achieved by VENUS is significant after

correction for multiple comparisons for all maps (p=0.015). Within-vendor difference trends

observed for P3 (Fig. 6b-d) hold true across participants, indicating a 3.7% increase in S1 – S2

MTR difference achieved by VENUS, coupled with a 3.2% and 2% decrease in S1 – S2 T1 and

MTsat differences, respectively (Fig. 7b-d).

Discussion

In this study, we developed and deployed a vendor-neutral qMRI protocol (VENUS) for T1, MTR

and MTsat mapping on three 3T commercial scanners by two vendors. Three healthy participants

were scanned to calculate all the metrics, as well as the ISMRM/NIST phantom for T1 mapping

and fundamental quality assessments. Our findings confirm the hypothesis that vendor-neutral

sequences decrease inter-vendor variability of T1, MTR and MTsat measurements. This key

improvement addresses *problem 1*, as stated in the Introduction, with transparent and publicly

available sequence descriptions. The developed sequence can be run on most GE and Siemens

scanners through a UUI that allows users to prescribe customized file naming entities for exporting

reconstructed images in the brain imaging data structure (BIDS)[44,45] and k-space data in the

ISMRM-RD format[46]. Conforming with community data standards, providing a user-friendly end-

to-end solution with a simplified inter-vendor deployment, this work offers a complete solution

for *problem 2* and shows a way forward for the standardization of qMRI.

Developing an end-to-end qMRI workflow

First, we created a vendor-native qMRI protocol that is unified across vendors to the greatest extent

possible, by keeping contrast, timing, and acquisition geometry identical (Table 1). However, other

vendor-native implementation details such as RF spoiling, MT and excitation pulse characteristics

were different, as it is commonly the case in multicenter studies[14,40]. Trying to address these issues

is difficult with a vendor-native sequence given that the implementations of commercial stock

sequences commonly used for qMRI are not open (*problem 1)*. One candidate solution for this

problem is modifying sequences on the vendor's proprietary development environment to equalize

implementations as much as possible, which has been shown to improve reproducibility to some

extent[22]. However, this requires familiarity with multiple sequence development environments and

still may fall short in unifying all the aspects of on-site acquisitions (e.g., different parallel imaging

implementations). Not only is this approach impractical for the developers, but it is also not a user-

friendly solution for clinical use. As we mention in the context of *problem 2,* reproducibility

solutions unifying inter-vendor implementations become more favorable if they are designed with

clinicians' needs in mind. To that end, we aimed at providing a unified and smooth user experience

by developing VENUS as an RTHawk application, which allows implementation details to be

shared publicly starting at the pulse sequence level.

Second, we built from scratch a vendor-neutral sequence that was developed and tested on a single

site and then ported to two more scanners from different vendors. In doing so, we adapted a system

that is primarily geared toward real-time imaging (RTHawk) to perform quantitative MRI

measurements. For example, absolute gradient limits have been allowed to achieve higher spoiling

gradient moments and string-valued customized metadata injection has been made possible to

comply with community data standards[44-46]. Nonetheless, there is still room for improvement to

make the most out of the hardware specifications in using RTHawk for qMRI. For example, the

gradient duty cycle specifications of all the scanners in the present work were suitable for playing

high spoiling gradients concurrently long enough to achieve spoiling levels suggested for an
accurate AFI B1 mapping[56]. However, it is not possible to implement gradient durations up to
80ms on RTHawk v3.0.0 (rc4-28-ge3540dda19) due to conservative hardware safety checks on
the minimum TR, as the acquisition controller assumes that the designed TRs will repeat for an
indefinite period. This limitation will be removed in a future release to regulate hardware and
patient safety limits by taking a finite sequence run into account.

Third, we created a fully transparent, container-mediated and data-driven workflow[57,58] that
automates the processing and reduces variability introduced by the operators. By design, the
workflow operates according to the BIDS qMRI standard[59] for picking up all the necessary data
and metadata, and generates outputs following a consistent derivative hierarchy. Moreover, the
raw data is exported in the ISMRM-RD format by our vendor-neutral sequence, allowing the use
of community developed reconstruction tools by simply adding another container at the beginning
of our modular workflow. We envision that using open-source reconstruction tools would be
highly favourable for vendor-neutral sequences employing under-sampled k-space with complex
trajectories to guarantee reproducibility[60-63]. In the present work, we analyzed the images
reconstructed using the RTHawk pipeline (https://github.com/qMRLab/VENUS.), given that we
only performed a simple multi-coil reconstruction.

Reducing inter-vendor variability

Stock sequences are optimized for reliable clinical imaging. These optimizations do not necessarily
serve for accuracy when the sequences are used for qMRI experiments. For example, phase
increment value of S1-2$_{NATIVE}$ sequences (Table 1) are hardcoded to maximize in-vivo signal
stability[64], not T1 accuracy in phantoms[65]. On the other hand, that of G1 has been shown to be
unsuitable for T1 mapping, exhibiting severe overestimations[21]. In this study, we set this value for
T1 accuracy[66] while unifying all other aspects of the vendor-neutral acquisition between scanners.
Results from phantom analysis clearly demonstrate that VENUS achieves higher accuracy and a
notable reduction in inter-vendor variability compared to its native counterparts (Fig. 4).

Our in vivo measurements are more difficult to evaluate because there is no ground truth
measurement. In the absence of an inversion recovery T1 map, we only looked at the agreement
between the three implementations and explored whether VENUS brought the T1 values closer
across vendors when compared to the vendor-native sequences. Visually (Fig. 5a), the reduction

in T1 variability can be appreciated for VENUS within the dynamic range of T1 adjusted for

WM/GM. As supported by the ridgeline plots (Fig. 5d,g), the G1$_{\text{NATIVE}}$ T1 distribution is globally

shifted towards higher values compared to S1-2$_{\text{NATIVE}}$, and their central tendency differs. As

observed in phantom, G1$_{\text{VENUS}}$ alleviates this discrepancy, shifting the T1 distribution closer to

those of S1-2$_{\text{VENUS}}$. Interestingly, WM T1 distributions show a more unimodal attribute on G1

than they do on S1-2 both for VENUS and vendor-native, with a more pronounced bimodal

appearance for S1-2$_{\text{VENUS}}$. A plausible explanation for that is vendor and implementation specific

differences resulting in discrepancies between B1+ field inhomogeneity. Nevertheless, the

VENUS shift functions for G1 vs S1 and G1 vs S2 comparisons are flatter than the vendor-native

shift functions (Fig. 6b), indicating that the inter-vendor WM T1 statistical distribution

characteristics are more similar using VENUS.

Table 2 indicates that reduction in inter-vendor variability is not limited to T1 but persists for all

the metrics across all participants. The inter-vendor variability in MTR and MTsat is relatively

easier to appreciate visually (Fig. 5b,c). The three MTR and MTsat maps from VENUS are in

better agreement, and this is most likely because our unified implementation compensated for MT

saturation pulse differences (Table 1).

S1-2 run the exact same vendor-native software versions, so we do not expect a big intra-vendor

difference for any of the parameters. However, T1 and MTsat agreements slightly improved using

VENUS (Fig. 7b,d). On the other hand, within-vendor agreement was better using native

sequences for MTR. Even in that case, transforming a vendor-native bias to a vendor-neutral one

is more advantageous to establish a standardized measurement. Knowing more about the vendor-

native implementations (e.g., why a particular MT pulse was selected as the default stock option

for that vendor) would help pinpoint the underlying causes of variability. This appears particularly

important for multiparametric mapping methods as the combined effect of different variability

sources from individual maps leads to a higher between-vendor disagreement in combined

quantitative maps (Fig. 7d).

As reducing variability matters, so does how we assess it. Shift functions[52] take the comparison

beyond differences in point estimates of centrality and relative spread (CoV) to a robust

characterization of differences on the absolute scale of the measurement. Figure 6 makes use of

80% of the data collapsed into a single relative statistic by Table 2 to explain how distributions

differ for P3. For example, Table 2 shows that VENUS reduces CoV from 12.1 to 4.1% for MTsat.
Figure 6d explains that majority of that reduction is achieved by decreasing absolute G1 vs S2
MTsat difference from 1.1 to 0.1 (a.u.), corresponding to a reduction from 25.7% to 3.2% inter-
vendor difference. In addition, Fig. 6d indicates that higher deciles benefit from the G1 vs S2
variability reduction more compared to the lower deciles, yielding a flatter shift function for
VENUS. This suggests that VENUS not only brings averaged MTsat values closer, but also
matches their distribution shape (Fig. 5f).

Implications of vendor-neutrality and the importance of transparency

The most important contribution of this article is the vendor-neutral solution it provides for multi-
center reliability of qMRI by significantly reducing inter-vendor variability. This has been a
plaguing issue for standardization of qMRI methods in validation[67,68], multi-center clinical trials[69],
establishing protocols[17], applied neuroimaging studies[70], determining the range of parameters in
disease typing [71,72] and health[16,40], between scanner upgrades[51] and even in phantom studies[12,13].
By reducing such variabilities, VENUS approach can bring qMRI closer to teasing out the true
biological variability in quantifying in-vivo tissue microstructure[73].

We recognize that part of the RTHawk workflow is proprietary. Nevertheless, we emphasize the
importance of the transparency to inter-vendor reproducibility at the level of sequence definitions.
RTHawk allows sharing open-source sequences (https://github.com/qMRLab/mt_sat). Indeed,
moving from open-source to open-execution[74] of pulse sequences is a long distance to cover, but
for that we also need open-hardware[75]. Neither RTHawk nor open-source solutions can access
under the hood of vendor-specific drivers to directly interface with the hardware. Although
RTHawk's pulse sequence and data management servers give more flexibility to the scanner
operation at multiple levels of the workflow (e.g., UUI, customized raw data stream, asynchronous
real-time updates to sequences, standalone workstation etc.), conversion of the open-source
sequence descriptions to vendor-specific hardware instructions is not transparent. We argue that
this is a reasonable trade-off to peel another layer from a vendor-specific ecosystem, as it does not
sacrifice the transparency of sources relevant to a pulse sequence description. The accuracy and
reliability of the parameter estimation methods depend on these descriptions; therefore, for qMRI
to work we need to be able to access, modify, and share the methods[57]. Fortunately, VENUS
approach to qMRI is not framework exclusive and satisfies this key requirement.

Namely, using community developed tools such as Pulseq, GammaStar, SequenceTree, ODIN or

TOPPE, interoperable qMRI applications can be developed. A critical step to achieve this is an

effective communication between method developers to foster compatibility between frameworks.

This is nicely exemplified by GammaStar and JEMRIS as both applications can export Pulseq

descriptions. Enabling a similar feature by developing a SpinBench plugin is among our future

goals. To facilitate discussions on this topic with vendor-neutral framework developers, we created

a forum page on the code repository of this article (https://github.com/qMRLab/VENUS).

Limitations and future directions

RF transmission systems were different between the all the scanners used for data collection. This

is indeed a likely cause of differences between the agreement and accuracy of T1 and MTsat maps.

Therefore, another obvious limitation of this study is the lack of B1+ mapping. Unfortunately, a

vendor-native B1+ mapping sequence was not available on G1, and it is also well-known that

discrepancies between vendor-native B1+ mapping contribute to between-scanner bias in T1

mapping[40]. Due to the RTHawk gradient limitations described above, developing an accurate AFI

sequence was not possible using the current version of RTHawk for a robust characterization of

the B1+ effect on T1 and MTsat (note that MTR does not require B1 correction). Therefore, further

investigation is needed to compare vendor-neutral B1+ maps across vendors for isolating the

specific contribution of transmit field inhomogeneity.

In addition to the non-uniformity of the B1+ field, another critical factor affecting the accuracy is

the calculation of a global RF scaling factor. Vendor-native systems set the transmit gain using

their own prescan routine, which may lead to a systematic bias in quantitative mapping. In this

work, we implemented prescan for G1 and S1-2 as described by[13] and configured RTHawk to use

the same calibration measurements. Nevertheless, it is possible to make this step vendor-neutral

as well. For future work, we plan to develop a double-angle VENUS prescan using the same

excitation pulses as the qMRI sequences that follow, to determine a global RF scaling factor.

Coupled with the use of anatomy-mimicking quantitative MRI phantoms[76], this would offer qMRI-

ready adaptive prescan routines and help investigate the effect of standardizing calibration

measurements on multicenter accuracy and agreement.

Finally, given that the data was acquired using full k-space sampling, we used vendor-native and

RTHawk resources to perform image reconstruction. The details of the RTHawk reconstruction

pipeline are available (Fig. 1f), which incorporates gradient non-linearity correction and exports

relevant warping coefficients. However, raw data from vendor-native acquisitions were not

available. Open-source reconstruction tools[60-62] are an important asset to investigate the potential

effect of reconstruction pipeline differences on image characteristics, such as observed in Fig. 2b-

g. Therefore, future work will enable raw data export from vendor-native systems and add a

containerized reconstruction node to the qMRFLow[58] pipeline for investigating potential sources

of reconstruction variability.

Conclusion

In this article we have demonstrated that vendor-neutral sequences (VENUS) and transparent

workflows improve reproducibility in quantitative MRI. Additionally, these workflows can be

deployed on an FDA-approved device, which demonstrates the potential for wide clinical

adoption. Quantitative MRI needs to bypass the vendor black boxes to make an impact in the clinic,

and this work shows the way forward.

Tables

Table 1 Comparison of acquisition parameters between vendor-native and vendor-neutral protocols. Parameters that are hardcoded on vendor-native systems are denoted by an asterisk (*).

	Common acquisition parameters			
FA (°) PDw/MTw/T1w	6/6/20			
MT PDw/MTw/T1w	off/on/off			
Voxel size (mm)	1x1x3			
TR (ms) PDw/MTw/T1w	32/32/18			
TE (ms)	4			
FOV (cm)	25.6			
Readout sampling rate (kSamp/s)	62.5			
MT frequency offset (Hz)*	1200			
Scanner ID and sequence type	G1$_{NATIVE}$	S1$_{NATIVE}$	S2$_{NATIVE}$	G1, S1-2$_{VENUS}$
Sequence name	3D SPGR	3D FLASH	3D FLASH	mt_sat (v1.1.0)
MT pulse shape*	Fermi	Gaussian	Gaussian	Fermi
MT pulse duration (ms)*	8	10	10	12
RF phase increment (°)*	115.4	50	50	117

572
573

575 Table 2 Coefficient of variation (%) of vendor-neutral (VENUS) and vendor-native quantitative measurements
576 between the scanners for each and across participants.

Participants	P1		P2		P3		Across	
Protocol	NATIVE	VENUS	NATIVE	VENUS	NATIVE	VENUS	NATIVE	VENUS
T1	9.8	1.3	11.7	4.0	16.5	6.1	11.36	4.3
MTR	8.5	3.4	8.5	3.4	10.1	4.1	7.9	3.2
MTsat	13.6	5.9	11.9	3.4	12.1	4.1	10.7	4.2

577
578

Acknowledgement

580 The authors would like to acknowledge Graham Wright, PhD for his help in organizing the
581 multicenter experiment, Paule Samson for her help in data collection, Juan Santos, PhD and
582 William R. Overall, PhD for their technical support in deploying RTHawk on multiple sites.

Data availability statement

All the vendor-neutral pulse sequences are publicly available as git submodules at
https://github.com/qmrlab/pulse_sequences and can be run on RTHawk systems v3.0.0 and later.
The RF and gradient waveforms (spv files) can be inspected and simulated using SpinBench
(https://www.heartvista.ai/spinbench). As per the general design principles of fully reproducible
qMRFlow pipelines, we adhered to a one-process one-container mapping for the processing of this
dataset. Docker images, BIDS and ISMRM-RD compliant dataset from the current study are freely
available at https://doi.org/10.17605/osf.io/5n3cu. Finally, the whole analysis and interactive
version of all the figures in this article will be available and executable online at
https://github.com/qmrlab/venus.

Figures

Figure 1 - Schematic illustration of the experimental design for multicenter data collection using vendor-native and vendor-neutral pulse sequences and pulse sequence development components: a) 3 MRI systems are located at 2 different sites and are labeled G1 (GE 750w), S1 (Siemens Prisma) and S2 (Siemens Skyra). Vendor "Native" systems export data in the DICOM format. The proposed vendor-agnostic "Neutral" system can export a complete set of reconstructed images in BIDS and the k-space data in ISMRM-RD format, synchronized across MRI systems. **b)** Connecting to the MRI system(s) over the local network, RTHawk (red workstation) can play open-source qMRI pulse sequences under version control (qMRPullseq). All the sequences are publicly available at https://github.com/qmrlab/pulse_sequences. **c)** Fully containerized qMRFlow data-driven pipelines can connect to the scanner data stream for post-processing on the RTHawk workstation (red workstation). The same pipelines can be reproduced on a local computer, supercomputing clusters or on the cloud. **d)** The acquisitions are controlled using a unified user interface (UUI), providing a consistent user experience across vendors. **e)** RF and gradient waveform stub blocks together with the readout logic is developed using SpinBench. **f)** RTHawk reconstruction pipeline nodes are illustrated for an 8-channel receiver, also indicating how raw and reconstructed data are exported and forwarded to the display tools for on-site visualization.

Figure 2 – Image quality assessment using the system phantom: a) Peak SNR values (PSNR) from T1w and PDw phantom images are displayed for vendor-neutral (red, orange, and yellow) and vendor-native (blue, cyan, and teal) G1, S1 and S2 scans, respectively. The same color coding is used in the following panels. **b-g)** Coronal PDw phantom images, with an inset zoom on two 4x4 grids with 1mm spacing. The brightness of the zoomed-in insets is increased by 30% for display purposes. **h-m)** Coronal T1w phantom images showing the center of the reference T1 arrays. The fine resolution (<0.6mm) inserts located at the center of the T1 array (rectangular area) are not relevant for the present resolution level. These inserts are colored following the same convention described in a) for convenience.

624
Figure 3 - Overview of the analysis workflow for phantom scans (a) and in vivo scans (b, c). File
collection (MTS) and output map names (T1map, MTsat, MTRmap) follow the BIDS standard
v1.6.0. **a)** Vendor-neutral and vendor-native ISMRM system phantom images were acquired at
two flip angles and two repetition times. The output data are then subjected to T1 fitting using
qMRLab (Docker container image: qmrlab/minimal:v2.5.0b). The resulting T1 maps are masked
using manually prescribed 10 spherical ROIs (reference T1 ranging from 0.9 to 1.9s). **b)** PDw and
MTw images are aligned to the T1w image to correct for between-scan motion. The aligned
dataset is then subjected to MTsat and MTR fitting in qMRLab to generate T1map, MTRmap and
MTsat. **c)** Brain extraction and tissue type segmentation is performed on the T1w images using
FSL. Following region masking and outlier removal for each map, vector outputs are saved for
statistical analysis and visualization in an online-executable Jupyter Notebook (R-Studio and
Python) environment. The tabular summary and the Nextflow pipeline execution report are
exported. The pipeline execution report is available at https://qmrlab.org/VENUS/qmrflow-exec-
report.html.

639
640

Figure 4 - Comparison of vendor-native and vendor-neutral T1 measurements in the studied range
of ISMRM-NIST system phantom reference values, from 0.09 to 1.99s **(a)**. T1 values from the
vendor-native acquisitions are represented by solid lines and square markers in cold colors, and
those from VENUS attain dashed lines and circle markers in hot colors. **b)** Vendor-native
measurements, especially $G1_{NATIVE}$ and $S2_{NATIVE}$, overestimate T1. $G1_{VENUS}$ and $S1-2_{VENUS}$ remain
closer to the reference. **c)** For VENUS, $\Delta T1$ remains low for R7 to R10, whereas deviations reach
up to 30.4% for vendor-native measurements. **d)** T1 values are averaged over S1-2 ($\overline{S}_{NATIVE\ and}$
$\overline{S}_{VENUS}$, green square and orange circle) and according to the acquisition type ($\overline{NATIVE}$ and
$\overline{VENUS}$, black square and black circle). Inter-vendor percent differences are annotated in blue
(native) and red (VENUS). Averaged percent measurement errors ($\overline{\Delta T1}$) are annotated on the plot
(black arrows).

Figure 5- Vendor-native and VENUS quantitative maps from one participant are shown in one axial slice (a-c). Distributions of quantified parameters in white matter (d-f) and gray matter (g-i) are shown using ridgeline plots of kernel density estimations. **a-c)** Inter-vendor images (G1 vs S1 and G1 vs S2) are appear more similar in VENUS (lower row) than in native (upper row). **d-f)** Distribution shapes and locations agree with visual inspection from (a), indicating closer agreement among VENUS sendistributions. **g-i)** Superior between-scanner agreement of VENUS persists in GM as well. Compared to WM, GM distributions are in the expected range (higher T1, lower MTR and MTsat values).

Figure 6 - Shift function analysis of T1, MTR and MTsat results from a single participant in white-matter (WM). a) Shift function analysis is a graphical tool for analyzing differences between two (dependent in this case) measurements at any location of the distributions. It shows 9 markers dividing the distribution into 10 equal chunks; hence the markers represent deciles. The shape of the curve (shift function) obtained by plotting decile differences against the first decile characterizes how distributions differ from each other. **b-d)** Here, shift function plots compare the agreement between different scanners for VENUS (bottom row) and vendor-native (top row) implementations in quantifying T1, MTR and MTsat. Across all the comparisons, the apparent trend is that the VENUS inter-vendor variability is lower than for the vendor-native implementations.

Figure 7 – Hierarchical shift function analysis of T1, MTR and MTsat results from three participants in the white-matter (WM). a) Hierarchical shift function repeats Figure 6 for all participants (shades of pink). Group deciles (red and blue markers for VENUS and vendor-native, respectively) show the average trend of inter-scanner differences across participants. **b-d)** G1-S1 and G1-S2 (inter-vendor) agree in VENUS better than they do in vendor-native for all quantitative maps of T1, MTR and MTsat.

References

1.   Lauterbur PC. Image formation by induced local interactions. Examples employing nuclear magnetic resonance. 1973. *Clin Orthop Relat Res*. Jul 1989;(244):3-6.

2.   Gillies RJ, Kinahan PE, Hricak H. Radiomics: Images Are More than Pictures, They Are Data. *Radiology*. Feb 2016;278(2):563-77. doi:10.1148/radiol.2015151169

3.   Novikov DS, Kiselev VG, Jespersen SN. On modeling. *Magn Reson Med*. Jun 2018;79(6):3172-3193. doi:10.1002/mrm.27101

4.   Hahn EL. An Accurate Nuclear Magnetic Resonance Method for Measuring Spin-Lattice Relaxation Times. http://dx.doi.org/10.1103/PhysRev.76.145. *Phys Rev*. 1949;76(1):145-146.

5.   Gupta RK. A new look at the method of variable nutation angle for the measurement of spin-lattice relaxation times using fourier transform NMR. *Journal of Magnetic Resonance (1969)*. 1977/01/01/ 1977;25(1):231-235. doi:https://doi.org/10.1016/0022-2364(77)90138-X

6.   D.C. Look, Locker DR. Time saving in measurement of NMR and EPR relaxation times. *Review of Scientific Instruments*. 1970;41(2):250-251. doi:http://link.aip.org/link/doi/10.1063/1.1684482

7.   Marques JP, Kober T, Krueger G, van der Zwaag W, Van de Moortele PF, Gruetter R. MP2RAGE, a self bias-field corrected sequence for improved segmentation and T1-mapping at high field. *Neuroimage*. Jan 15 2010;49(2):1271-81. doi:10.1016/j.neuroimage.2009.10.002

8.   Stikov N, Boudreau M, Levesque IR, Tardif CL, Barral JK, Pike GB. On the accuracy of T1 mapping: searching for common ground. *Magn Reson Med*. Feb 2015;73(2):514-22. doi:10.1002/mrm.25135

9.   Grafe D, Frahm J, Merkenschlager A, Voit D, Hirsch FW. Quantitative T1 mapping of the normal brain from early infancy to adulthood. *Pediatr Radiol*. Mar 2021;51(3):450-456. doi:10.1007/s00247-020-04842-7

10.   Okubo G, Okada T, Yamamoto A, et al. MP2RAGE for deep gray matter measurement of the brain: A comparative study with MPRAGE. *J Magn Reson Imaging*. Jan 2016;43(1):55-62. doi:10.1002/jmri.24960

11.   Voelker MN, Kraff O, Goerke S, et al. The traveling heads 2.0: Multicenter reproducibility of quantitative imaging methods at 7 Tesla. *Neuroimage*. May 15 2021;232:117910. doi:10.1016/j.neuroimage.2021.117910

12.   Bane O, Hectors SJ, Wagner M, et al. Accuracy, repeatability, and interplatform reproducibility of T 1 quantification methods used for DCE-MRI: Results from a multicenter phantom study. *Magnetic Resonance in Medicine*. 2018;79(5):2564-2575. doi:10.1002/mrm.26903

13.   Keenan KE, Gimbutas Z, Dienstfrey A, et al. Multi-site, multi-platform comparison of MRI T1 measurement using the system phantom. *PLOS ONE*. 2021;16(6):e0252966. doi:10.1371/journal.pone.0252966

14.   Yarnykh VL. Optimal radiofrequency and gradient spoiling for improved accuracy of T1 and B1 measurements using fast steady-state techniques. *Magnetic Resonance in Medicine*. 2010;63(6):1610-1626. doi:10.1002/mrm.22394

15.   A.G. Teixeira RP, Malik SJ, Hajnal JV. Fast quantitative MRI using controlled saturation
magnetization   transfer.   *Magnetic   Resonance   in   Medicine*.   2019;81(2):907-920.
doi:10.1002/mrm.27442

16.   Bojorquez JZ, Bricq S, Acquitter C, Brunotte F, Walker PM, Lalande A. What are normal
relaxation   times   of   tissues   at   3   T?   *Magn   Reson   Imaging*.   Jan   2017;35:69-80.
doi:10.1016/j.mri.2016.08.021

17.   Cohen-Adad J, Alonso-Ortiz E, Abramovic M, et al. Open-access quantitative MRI data of
the spinal cord and reproducibility across participants, sites and manufacturers. *Sci Data*. Aug 16
2021;8(1):219. doi:10.1038/s41597-021-00941-8

18.   Liberman G, Louzoun Y, Ben Bashat D. T(1) mapping using variable flip angle SPGR data
with flip angle correction. *J Magn Reson Imaging*. Jul 2014;40(1):171-80. doi:10.1002/jmri.24373

19.   Leutritz T, Seif M, Helms G, et al. Multiparameter mapping of relaxation (R1, R2*), proton
density and magnetization transfer saturation at 3 T: A multicenter dual-vendor reproducibility
and   repeatability   study.   *Hum   Brain   Mapp*.   Oct   15   2020;41(15):4232-4247.
doi:10.1002/hbm.25122

20.   Boudreau M, Stikov N, Pike GB. B1 -sensitivity analysis of quantitative magnetization
transfer imaging. *Magn Reson Med*. Jan 2018;79(1):276-285. doi:10.1002/mrm.26673

21.   Yarnykh VL. Effect of the phase increment on the accuracy of T1 measurements by the
variable flip angle method using a fast RF spoiled gradient echo sequence. 2007:

22.   Gracien R-M, Maiworm M, Brüche N, et al. How stable is quantitative MRI? – Assessment
of intra- and inter-scanner-model reproducibility using identical acquisition sequences and data
analysis programs. *NeuroImage*. 2020;207:116364. doi:10.1016/j.neuroimage.2019.116364

23.   Helms G, Dathe H, Dechent P. Quantitative FLASH MRI at 3T using a rational
approximation of the Ernst equation. Research Support, Non-U.S. Gov't. *Magn Reson Med*. Mar
2008;59(3):667-72. doi:10.1002/mrm.21542

24.   Rowley CD, Campbell JSW, Wu Z, et al. A model-based framework for correcting B 1 +
inhomogeneity effects in magnetization transfer saturation and inhomogeneous magnetization
transfer saturation maps. *Magn Reson Med*. Oct 2021;86(4):2192-2207. doi:10.1002/mrm.28831

25.   Stikov N, Trzasko JD, Bernstein MA. Reproducibility and the future of MRI research. *Magn
Reson Med*. Dec 2019;82(6):1981-1983. doi:10.1002/mrm.27939

26.   Jochimsen TH, Von Mengershausen M. ODIN—object-oriented development interface for
NMR. *Journal of Magnetic Resonance*. 2004;170(1):67-78.

27.   Stöcker T, Vahedipour K, Pflugfelder D, Shah NJ. High-performance computing MRI
simulations. *Magnetic resonance in medicine*. 2010;64(1):186-193.

28.   Magland JF, Li C, Langham MC, Wehrli FW. Pulse sequence programming in a dynamic
visual environment: SequenceTree. *Magnetic resonance in medicine*. 2016;75(1):257-265.

29.   Ravi KS, Geethanath S, Vaughan JT. PyPulseq: A python package for mri pulse sequence
design. *Journal of Open Source Software*. 2019;4(42):1725.

30.   Layton KJ, Kroboth S, Jia F, et al. Pulseq: a rapid and hardware-independent pulse
sequence prototyping framework. *Magnetic resonance in medicine*. 2017;77(4):1544-1552.

31.   Nielsen JF, Noll DC. TOPPE: A framework for rapid prototyping of MR pulse sequences.
Magnetic resonance in medicine. 2018;79(6):3128-3134.

32.   Cordes C, Konstandin S, Porter D, Günther M. Portable and platform-independent MR
pulse sequence programs. *Magnetic resonance in medicine*. 2020;83(4):1277-1290.

33.   Zwart NR, Pipe JG. Graphical programming interface: A development environment for
MRI methods. *Magnetic resonance in medicine*. 2015;74(5):1449-1460.
34.   Ravi KS, Potdar S, Poojar P, et al. Pulseq-Graphical Programming Interface: Open source
visual environment for prototyping pulse sequences and integrated magnetic resonance imaging
algorithm development. *Magnetic resonance imaging*. 2018;52:9-15.
35.   Herz K, Mueller S, Perlman O, et al. Pulseq-CEST: Towards multi-site multi-vendor
compatibility and reproducibility of CEST experiments using an open-source sequence standard.
Magnetic resonance in medicine. 2021;
36.   Tong G, Gaspar AS, Qian E, et al. A framework for validating open-source pulse sequences.
Magnetic resonance imaging. 2021;
37.   Barral JK, Gudmundson E, Stikov N, Etezadi-Amoli M, Stoica P, Nishimura DG. A robust
methodology for in vivo T1 mapping. Research Support, N.I.H., Extramural
Research   Support,   Non-U.S.   Gov't.   *Magn   Reson   Med*.   Oct   2010;64(4):1057-67.
doi:10.1002/mrm.22497
38.   Santos JM, Wright GA, Pauly JM. Flexible real-time magnetic resonance imaging
framework. IEEE; 2004:1048-1051.
39.   Overall WR, Pauly JM. An extensible, graphical environment for pulse sequence design
and simulation. 2007:
40.   Lee Y, Callaghan MF, Acosta-Cabronero J, Lutti A, Nagy Z. Establishing intra- and inter-
vendor reproducibility of T1 relaxation time measurements with 3T MRI. *Magn Reson Med*. Jan
2019;81(1):454-465. doi:10.1002/mrm.27421
41.   Leutritz T, Seif M, Helms G, et al. Multiparameter mapping of relaxation (R1, R2*), proton
density and magnetization transfer saturation at 3 T: A multicenter dual-vendor reproducibility
and repeatability study. *Human brain mapping*. 2020;41(15):4232-4247.
42.   Helms G, Dathe H, Kallenberg K, Dechent P. High-resolution maps of magnetization
transfer with inherent correction for RF inhomogeneity and T1 relaxation obtained from 3D
FLASH MRI. Evaluation Studies
Research   Support,   Non-U.S.   Gov't.   *Magn   Reson   Med*.   Dec   2008;60(6):1396-407.
doi:10.1002/mrm.21732
43.   Haase A, Frahm J, Matthaei D, Hanicke W, Merboldt KD. FLASH imaging. Rapid NMR
imaging using low flip-angle pulses. *Journal of Magnetic Resonance (1969)*. 1986;67(2):258-266.
44.   Gorgolewski KJ, Auer T, Calhoun VD, et al. The brain imaging data structure, a format for
organizing and describing outputs of neuroimaging experiments. *Scientific data*. 2016;3(1):1-9.
45.   Karakuzu A, Appelhoff S, Auer T, et al. qMRI-BIDS: an extension to the brain imaging data
structure   for   quantitative   magnetic   resonance   imaging   data.   *medRxiv*.
2021:2021.10.22.21265382. doi:10.1101/2021.10.22.21265382
46.   Inati SJ, Naegele JD, Zwart NR, et al. ISMRM Raw data format: A proposed standard for
MRI raw datasets. *Magnetic resonance in medicine*. 2017;77(1):411-421.
47.   Cohen-Adad J, Alonso-Ortiz E, Abramovic M, et al. Generic acquisition protocol for
quantitative MRI of the spinal cord. *Nature protocols*. 2021;16(10):4611-4632.
48.   Karakuzu A, Boudreau M, Duval T, et al. qMRLab: Quantitative MRI analysis, under one
umbrella. *Journal of Open Source Software*. 2020;5(53):2343.

49.   Avants BB, Tustison NJ, Song G, Cook PA, Klein A, Gee JC. A reproducible evaluation of
ANTs similarity metric performance in brain image registration. *Neuroimage*. 2011;54(3):2033-
2044.
50.   Jenkinson M, Beckmann CF, Behrens TEJ, Woolrich MW, Smith SM. Fsl. *Neuroimage*.
2012;62(2):782-790.
51.   Keenan KE, Gimbutas Z, Dienstfrey A, Stupic KF. Assessing effects of scanner upgrades for
clinical studies. *Journal of Magnetic Resonance Imaging*. 2019;50(6):1948-1954.
52.   Rousselet GA, Pernet CR, Wilcox RR. Beyond differences in means: robust graphical
methods to compare two groups in neuroscience. *European Journal of Neuroscience*.
2017;46(2):1738-1748.
53.   Harrell FE, Davis CE. A new distribution-free quantile estimator. *Biometrika*.
1982;69(3):635-640.
54.   American College of R. Phantom Test guidance for use of the large MRI phantom for the
ACR MRI accreditation program. 2019.
55.   Stupic KF, Ainslie M, Boss MA, et al. A standard system phantom for magnetic resonance
imaging. *Magnetic Resonance in Medicine*. 2021;86(3):1194-1211. doi:10.1002/mrm.28779
56.   Yarnykh VL. Actual flip-angle imaging in the pulsed steady state: A method for rapid three-
dimensional mapping of the transmitted radiofrequency field. *Magnetic Resonance in Medicine*.
2007;57(1):192-200. doi:10.1002/mrm.21120
57.   Karakuzu A, Boudreau M, Cohen-Adad J, Stikov N. Thinking outside the blackbox: A fully
transparent T1 mapping pipeline. ISMRM; 2020:
58.   Karakuzu A, Boudreau M, Duval T, et al. The qMRLab workflow: From acquisition to
publication. ISMRM; 2019:
59.   Karakuzu A, Hollander G, Appelhof S, et al. A joint-community effort to standardize
quantitative MRI data: Updates from the BIDS extension proposal. ISMRM; 2020:
60.   Hansen MS, Sørensen TS. Gadgetron: an open source framework for medical image
reconstruction. *Magnetic resonance in medicine*. 2013;69(6):1768-1776.
61.   Assländer J, Cloos MA, Knoll F, Sodickson DK, Hennig J, Lattanzi R. Low rank alternating
direction method of multipliers reconstruction for MR fingerprinting. *Magnetic resonance in
medicine*. 2018;79(1):83-96.
62.   Knopp T, Grosser M. MRIReco. jl: An MRI reconstruction framework written in Julia.
Magnetic Resonance in Medicine. 2021;
63.   Maier O, Baete SH, Fyrdahl A, et al. CG-SENSE revisited: Results from the first ISMRM
reproducibility challenge. *Magnetic resonance in medicine*. 2021;85(4):1821-1839.
64.   Preibisch C, Deichmann R. Influence of RF spoiling on the stability and accuracy of T1
mapping based on spoiled FLASH with varying flip angles. *Magnetic Resonance in Medicine: An
Official Journal of the International Society for Magnetic Resonance in Medicine*. 2009;61(1):125-
135.
65.   Heule R, Ganter C, Bieri O. Variable flip angle T1 mapping in the human brain with reduced
t2 sensitivity using fast radiofrequency-spoiled gradient echo imaging. *Magnetic Resonance in
Medicine*. 2016;75(4):1413-1422. doi:10.1002/mrm.25668
66.   Yarnykh VL. Optimal radiofrequency and gradient spoiling for improved accuracy of T1
and   B1   measurements   using   fast   steady-state   techniques.
http://dx.doi.org/10.1002/mrm.22394. *Magn Reson Med*. 2010;63(6):1610-1626.

67.   Mancini M, Karakuzu A, Cohen-Adad J, Cercignani M, Nichols TE, Stikov N. An interactive
meta-analysis of MRI biomarkers of myelin. *eLife*. 2020;9doi:10.7554/elife.61523
68.   Lazari A, Lipp I. Can MRI measure myelin? Systematic review, qualitative assessment, and
meta-analysis of studies validating microstructural imaging with myelin histology. *Neuroimage*.
2021:117744.
69.   Ashton E. Quantitative MR in multi-center clinical trials. *Journal of Magnetic Resonance*
Imaging: An Official Journal of the International Society for Magnetic Resonance in Medicine.
2010;31(2):279-288.
70.   Boshkovski T, Kocarev L, Cohen-Adad J, et al. The R1-weighted connectome:
complementing brain networks with a myelin-sensitive measure. *Network Neuroscience*.
2021;5(2):358-372.
71.   Young IR. Significant events in the development of MRI. *Journal of Magnetic Resonance*
Imaging: An Official Journal of the International Society for Magnetic Resonance in Medicine.
2004;20(2):183-186.
72.   Goebel J, Seifert I, Nensa F, et al. Can native T1 mapping differentiate between healthy
and diffuse diseased myocardium in clinical routine cardiac MR imaging? *PLoS One*.
2016;11(5):e0155591.
73.   Weiskopf N, Edwards LJ, Helms G, Mohammadi S, Kirilina E. Quantitative magnetic
resonance imaging of brain anatomy and in vivo histology. *Nature Reviews Physics*. 2021/08/01
2021;3(8):570-588. doi:10.1038/s42254-021-00326-1
74.   Salah K, Damiani E, Al-Fuqaha A, Martin T, Taha K, Khan MK. Open execution—The
blockchain model. *IEEE Blockchain Technical Briefs (December 2018)*. 2018;
75.   Winter L, Haopeng H, Barghoorn A, Hoffmann W, Hetzer S, Winkler S. Open source
imaging initiative.
76.   Gopalan K, Tamir JI, Arias AC, Lustig M. Quantitative anatomy mimicking slice phantoms.
Magnetic Resonance in Medicine. 2021;86(2):1159-1166.
880