Brain-wide continuous functional ultrasound imaging for real-time monitoring of hemodynamics during ischemic stroke

Clément Brunner[1,2,3,4], Nielsen Lagumersindez Denis[5,6,7], Karen Gertz[5,6,7], Micheline Grillet[1,2,3,4], Gabriel Montaldo[1,2,3,4], Matthias Endres[5,6,7] and Alan Urban[1,2,3,4,+]

[1] Neuro-Electronics Research Flanders, Leuven, Belgium.
[2] Vlaams Instituut voor Biotechnologie, Leuven, Belgium.
[3] Interuniversity microelectronics centre, Leuven, Belgium.
[4] Department of neurosciences, KU Leuven, Leuven, Belgium.
[5] Department of Neurology and Center for Stroke Research Berlin, Charité-Universitätsmedizin, Berlin, Germany.
[6] German Center for Neurodegenerative Diseases (DZNE), Berlin, Germany.
[7] German Centre for Cardiovascular Research (DZHK), Berlin, Germany.

+Corresponding author: Alan Urban, alan-urban@nerf.be

Abstract

Ischemic stroke occurs with no warning, and therefore, very little is known about hemodynamic perturbations in the brain immediately after stroke onset. Here, functional ultrasound imaging was used to monitor variations in relative cerebral blood volume (rCBV) compared to baseline. rCBV levels were analyzed brain-wide and continuously at high spatiotemporal resolution (100µm, 2Hz) until 70mins after stroke onset in rats. We compared two stroke models, with either a permanent occlusion of the middle cerebral artery (MCAo) or a tandem occlusion of both the common carotid and middle cerebral arteries (CCAo+MCAo). We observed a typical hemodynamic pattern, including a quick drop of the rCBV after MCAo, followed by spontaneous reperfusion of several brain regions located in the vicinity of the ischemic core. The severity and location of the ischemia were highly variable between animals. Still, both parameters were, on average, in good agreement with the final ischemic lesion volume measured 24hrs after stroke onset for the MCAo but not the CCAo+MCAo model. For the latter, we observed that the infarct was extended to regions that were initially not ischemic and located rostrally of the ischemic core. These regions strongly colocalize with the origin of transient hemodynamic events associated with spreading depolarizations.

Keywords
Ischemic stroke, functional ultrasound imaging, spreading depolarization, ischemic lesion prediction

Introduction

Acute ischemic stroke is most often caused by thrombotic or embolic occlusion of a cerebral artery. It is characterized by a sudden loss of blood circulation to an area of the brain, resulting in a corresponding loss of neurologic function. Ischemic stroke triggers a variable decrease of blood flow in the affected parenchyma that depends on the degree of collateral circulation. In brain regions with poor collaterals, lack of oxygen and glucose may disrupt neuronal activity and ultimately cause a drop in energy metabolism leading to tissue infarction[1,2]. Although both the ischemic core and penumbra are dysfunctional, the penumbra remains viable during a given time window upon restoring blood flow[3]. The penumbra will eventually grow into the ischemic core in the absence of reperfusion. The rate of infarct growth is highly variable between individuals and strongly depends on, e.g., the extensiveness of collateral circulation[4], the ischemic lesion location, and metabolic factors.

A large number of ischemic stroke models have been developed in rodents to recapitulate the pathophysiology in patients. A particular challenge is the heterogeneity of individual strokes in patients that make it difficult to be modeled in rodents[5]. A significant difference between humans and animal models is the presence of different penumbral and infarct flow thresholds, leading to varying rates and dynamics in infarct evolution[6,7]. The middle cerebral artery (MCA) is commonly affected in ischemic stroke, and hence, occluding this major vessel downstream of the internal carotid artery in rodents may replicate cerebral ischemia in stroke patients. The MCA can be accessed via an approach through either the internal or the external carotid artery and be occluded temporarily or permanently[8,9]. Infarcts induced by this approach often comprise both striatal and cortical damage. Alternatively, blocking the artery distal to the lenticulostriate branches (i.e., distal MCAo) with various strategies (vascular clip[10], suture[11], electrocoagulation[12], photothrombosis[13], endothelin-1[14], ferric-chloride[15], or magnetic nanoparticles[16]) results in a permanent occlusion with mainly cortical infarcts. The pros and cons of stroke models are extensively discussed in the literature by Macrae[17] and Flury et al.[18]. In short, most models of ischemic stroke involving only the occlusion of the MCA have a large variability[19,20]. Still, it has been shown that a tandem permanent occlusion of the distal MCA and ipsilateral common carotid artery (CCA) can alternatively be used as a surgically simple method for causing large neocortical infarcts with reproducible topography and volume in rat[21].

Hemodynamics and characteristics of lesion growth following immediately after vessel occlusion are essential issues and represent a new target area for promising therapeutic interventions. Spreading depolarization (SD) describes a slowly propagating wave of intense depolarization that travels across the cortical tissue and is typically coupled with local changes in cerebral blood flow (CBF). SD is associated with a functional hyperemic response when the neurovascular coupling is intact and hypoperfusion when the neurovascular coupling is compromised[22–24]. Moreover, SD may exacerbate ischemic injury[25–29]. The features of SDs induced by ischemia have been widely examined, but studies generally failed to identify their site of elicitation, where they are triggered. Therefore, developing new strategies for monitoring brain perfusion at a large scale and in-depth is crucial to better understand the mechanisms linking perfusion defects and tissue infarction; and improve preclinical drug development[30]. Precise tracking of the hemodynamic responses associated with SDs could help understand their impact on pathophysiology.

The current consensus relies on a multi-modal neuroimaging strategy to provide sufficient information to identify ischemic tissue, biomarkers of changes in tissue metabolism, potentially salvageable tissue, or monitor its reorganization[31,32]. Traditional imaging modalities such as magnetic resonance imaging (MRI), specifically perfusion- and diffusion-weighted imaging (PWI-DWI, pCASL, DSC)[33–35], positron emission tomography (PET)[32] and single-photon emission computed tomography (SPECT)[36] have limited spatial and temporal resolution to resolve vascular responses dynamically. Optical imaging techniques have a high resolution and are well suited to image changes in CBF and blood oxygenation. Laser Doppler flowmetry (LDF)[37,38] can measure blood flow at a low spatiotemporal resolution. Laser speckle contrast imaging (LSCI)[39] and optical intrinsic signal imaging (OIS)[40] are full-field imaging techniques. They can rapidly measure CBF and blood oxygenation, respectively, but only at the surface of the brain. Two-photon microscopy (2PM)[41] provides vascular morphology and flow information, but it requires exogenous contrast agents. Photoacoustic microscopy (PAM)[42–44] can measure these parameters but with a limited depth of field. Moreover, the latter two techniques are laser-scanning based, limiting their respective field-of-view and achievable frame-rate, making real-time observation of a larger area challenging.

As an alternative, functional ultrasound imaging used high-frame-rate plane-wave imaging to measure
cerebral blood volume (CBV) changes in response to neuronal activity[45–49]. Briefly, functional ultrasound
imaging is based on a power Doppler estimator, proportional to the energy of the signals coming from
moving scatterers[46]. This technology has a spatial resolution of ~100×300×100µm³, a temporal
resolution of ~0.1s, and a depth of field of ~1.5cm and is, therefore, ideal for preclinical stroke
research[10,45,50,51]. Moreover, it does not require contrast agents and can be adapted to freely moving
animals[52–54].

In this study, we used functional ultrasound imaging to monitor hemodynamic changes during stroke,
including perfusion and SDs at the brain-wide scale. We compared two rat stroke models, i.e., a model
consisting of permanent occlusion of the distal branch of the middle cerebral artery (MCAo) resulting in
small ischemic lesions and a model with a tandem occlusion of the common carotid and middle cerebral
arteries (CCAo+MCAo) resulting in larger ischemic lesions typically both comprising large parts of the
MCA territory. Hemodynamic changes before were monitored continuously until 70mins after stroke.
Data were then registered, segmented, and clustered[55,56] into 115 anatomical regions, including cortical,
thalamic, striatal, and hippocampal areas from a custom rat brain atlas based on Paxinos[57]. Comparing
the ischemic territory and the infarcted tissue at 24hrs revealed a mismatch in both its size and location
for both stroke models. Notably, we observed that brain regions located rostrally of the ischemic core
are not affected by the ischemia up to 70mins after stroke but became part of the ischemic lesion within
the 24hrs post-stroke. These regions are also from where SDs emerge quickly after the stroke onset.

Material and Methods
Animals
Experimental procedures were approved by the Committee on Animal Care of the Catholic University of
Leuven, in accordance with the national guidelines on the use of laboratory animals and the European
Union Directive for animal experiments (2010/63/EU). The manuscript was written according to the
ARRIVE Essential 10 checklist for reporting animal experiments[58]. Adult male Sprague-Dawley rats
(n=24; Janvier Labs, France) with an initial weight between 200-300g were housed in standard
ventilated cages and kept in a 12:12hrs reverse dark/light cycle environment at a temperature of 22°C
with *ad libitum* access to food and water. To avoid selection bias, animals were randomly distributed in
three experimental groups: (i) MCAo, in which the distal branch of the left MCA was permanently
occluded (n=10), (ii) CCAo+MCAo, in which the left CCA was permanently occluded shortly before the
MCAo (n=10), (iii) a sham group where four rats were imaged as in any other group, but the
microvascular clip was inserted next to the MCA without occluding it.

Cranial window for brain-wide imaging
A cranial window was performed in all rats under isoflurane anesthesia (Iso-Vet, Dechra, Belgium)
continuously delivered at 0.6l/min through a nose mask. A mixture of 5% isoflurane in compressed dry
air was used to induce anesthesia, subsequently reduced to 2.0-2.5% during surgery and to 1.5% for
imaging. Body temperature was monitored using a rectal probe and maintained at 36.5±0.5°C using a
digitally controlled heating blanket (CODA, Kent Scientific Corp., USA). Intraperitoneal injection of 5%
glucose solution was provided every 2hrs to prevent dehydration. As pre-operative analgesia, Xylocaine
(0.5%, AstraZeneca, England) was injected subcutaneously into the head skin. The scalp was shaved
and cleaned with iso-betadine before removing the entire dorsal skull. The cranial window extended
from bregma (ß) +4.0 to −7.0mm anterior to posterior, laterally 6.0mm to the right side, until the parietal-
squamosal suture on the left side to expose the distal branch of the MCA was performed as previously
described[10,48,50]. Sterile saline was regularly added during drilling sessions to avoid overheating the
tissue. The skull was carefully removed without damaging the dura. Finally, the brain was covered with
a low-melting 2% agarose (Sigma-Aldrich, USA) and ultrasound gel (Aquasonic Clear, Parker
Laboratories Inc, USA) to ensure a proper acoustic coupling with the ultrasound probe. At the end of the
imaging session, Metacam (0.2mg/kg, Boehringer Ingelheim, Canada) was injected subcutaneously for
postoperative analgesia, and rats were placed in a warm cage and monitored until waking up.

MCAo and CCAo+MCAo procedures
Rats from the CCAo+MCAo group underwent CCA dissection before cranial window surgery. Briefly,
the rat was placed in a supine position to shave and clean neck hairs. A vertical incision was made on
the neck, and the left CCA was carefully dissected from the surrounding tissue and exposed, avoiding
damaging the vagus nerve[21,59]. A 3-0 surgical silk thread (Ethicon, France) was placed around the CCA
and remained untightened until occlusion during the imaging session. After the imaging section, the
neck skin was sutured with a 2-0 surgical silk thread (Ethicon, France). Rats from MCAo and

CCAo+MCAo groups were subjected to a permanent MCA occlusion directly during the imaging session
using a microvascular clip (#18055-03, micro-serrefine, FST GmbH, Germany) that was manually
refined by grinding for optimal serration of the MCAo in adults rats.
158
Functional ultrasound imaging
The data acquisition was performed using a functional ultrasound imaging scanner equipped with
custom acquisition and processing software described by Brunner et al.[56]. In short, the scanner is
composed of a linear ultrasonic transducer (15MHz, 128 elements, Xtech15, Vermon, France)
connected to 128-channel emission-reception electronics (Vantage, Verasonics, USA) that are both
controlled by a high-performance computing workstation (fUSI-2, AUTC, Estonia). The transducer was
motorized (T-LSM200A, Zaber Technologies Inc., Canada) to allow anteroposterior scanning of the
brain. Imaging is performed on an anti-vibration table to minimize external sources of vibration. Each
coronal Doppler image is 12.8-mm width and 9-mm depth and comprises 300 compound images
acquired at 500Hz. Each compound image is computed by adding nine plane-wave (4.5kHz) angles
from -12° to 12° with a 3° step. The blood signal was extracted from 300 compound images using a
single value decomposition filter and removing the 30 first singular vectors[60]. The Doppler image is
computed as the mean intensity of the blood signal in these 300 frames that is an estimator of the
cerebral blood volume (CBV)[45,46]. This sequence enables a temporal resolution of 0.6s, an in-plane
resolution of 100×110µm, and an off-plane (thickness of the image) of 300µm[56].
174
2D scan of brain vasculature
Before occlusion, we performed a high-resolution 2D scan of the brain vasculature consisting of 89
coronal planes from bregma (ß) +4.0 to -7.0mm spaced by 125µm. This scan was used for data
registration. During the 90-min imaging session, the same brain area was scanned at lower resolution
with only 23 planes with a step of 500µm that took approximately 23s to complete. The number of
imaging planes was chosen to maximize the number of brain volumes scanned per min while preserving
a good resolution in the anteroposterior axis. For the MCAo group, a 20-min baseline was recorded
before occluding the MCA, whereas the CCAo+MCAo group has a 10-min baseline before CCAo,
followed by the MCAo 10mins later. Notably, the procedures were performed live with a real-time display
of the Doppler images, allowing for monitoring the brain perfusion and the direct confirmation of the CCA
and/or MCA occlusions (Movies 1 to 3). The imaging session was stopped after 90mins of recording.
186
Registration and segmentation
We developed a custom digital rat brain atlas for registration of the data based on the one stereotaxic
atlas of Paxinos[57]. In short, the spatial transformation matrix was computed on the high-resolution 2D
scan that was manually aligned on the atlas by affine transformation (i.e., translation, rotation, and
scaling). This matrix was then applied to the low-resolution 2D scan. The dataset was segmented into
115 anatomical regions/hemispheres that were subsampled from the 332 brain regions of the original
atlas (see Supplementary Table 1). The hemodynamic signals were averaged in each area. The
software for data registration and segmentation is open-access here[61]. The unrolled-cortex
representation corresponds to the maximum intensity projection of the signal located 250µm under the
cortical surface (see Supplementary Figure 1).
197
198 **Relative cerebral blood volume (rCBV)**
199 We used the relative cerebral blood volume (rCBV; expressed in % compared to baseline) to analyze
200 ischemia defined as the signal in each voxel compared to its average level during the baseline period
201 (Figure 1). After registration and segmentation, the rCBV signal was averaged in each individual region.
202
203 **Clustering of the rCBV loss**
204 Once registered and segmented, the brain regions were clustered based on the average rCBV loss after
205 the MCA occlusion and sorted from the most significant loss of rCBV signal to the smallest (from 1 to
206 115) as compared to baseline level. The list of the brain regions clustered can be found in
207 Supplementary Table 1.
208
Analysis of spreading depolarizations
The detection of spreading depolarizations (SDs) was performed based on the temporal analysis of the
rCBV signal. The rCBV signal was averaged in an area of interest (10×10 voxels) located in the left side
cortex, laterally to the ischemic territory. An SD is defined as a transient increase of rCBV signal above
50% compared to baseline. This procedure allowed to measure the occurrence of SDs over the
recording period and their propagation pattern in the field of view. The velocity of each SDs was

calculated as the time spent to propagate between two areas of interest located in the left side cortex,
along with the ischemic territory. The origin of each SDs was located by tracking back the trajectory to
the region of the first detection of the transient increase of rCBV. The occurrence, the propagation
pattern, and the velocity of SDs could be efficiently visualized using the unrolled-cortex projection as
presented in Movie 3.
220

Figure 1

Figure 1. Experimental workflow and timeline. From left to right. An 11×13-mm cranial window was performed to access the whole brain by functional ultrasound imaging before occluding the middle cerebral artery occlusion (MCAo) or MCAo combined with the common carotid artery (CCAo+MCAo). Image acquisition has been performed continuously for 70mins by scanning the brain repeatedly in the anteroposterior direction (23secs/scan) before, during, and after stroke onset. After the experiments, rats returned to their home cage and were euthanized 24hrs after occlusion to quantify the infarct size using cresyl violet staining. We developed a digital version of the rat Paxinos atlas[57] for the registration, segmentation, and temporal analysis of ischemia using a dedicated software solution[61].

Histopathology
Rats were killed 24hrs after the occlusion for histological analysis of the infarcted tissue. Rats received
a lethal injection of pentobarbital (100mg/kg i.p. Dolethal, Vetoquinol, France). Using a peristaltic pump,
they were transcardially perfused with phosphate-buffered saline followed by 4% paraformaldehyde
(Sigma-Aldrich, USA). Brains were collected and post-fixed overnight. 50-µm thick coronal brain
sections across the MCA territory were sliced on a vibratome (VT1000S, Leica Microsystems, Germany)
and analyzed using the cresyl violet (Electron Microscopy Sciences, USA) staining procedure. Slices
were mounted with DPX mounting medium (Sigma-Aldrich, USA) and scanned using a brightfield
microscope. Ischemic lesion areas were manually delineated and measured by two researchers
independently in double-blind conditions using the Fiji software[62]. Brain slices were then registered on
the rat atlas utilizing a set of landmarks, including the overall size of the slice, the cortical thickness, the
positions of ventricles, and the position/size of the hippocampus. The ischemic lesion was projected in
the unroll-cortex using the same approach as defined previously (see above and Supplementary Figure
1).
236

237   **Statistical analysis**
238   The dataset successfully passed Shapiro-Wilk and Kolmogorov-Smirnov tests (significance level
239   α=0.05) for normal distribution before being subjected to statistical analysis, see details in results and
240   figure captions. All data are shown as mean±standart deviation(sd). Statistical analysis were performed
241   using Prism9.3.1 (GraphPad Software).
242

Results
Animals
Three rats from the MCAo group were excluded from the analysis because we observed an extensive
postmortem hemorrhage extending in the cortex and subcortical regions. It was probably caused by a
displacement of the microvascular clip during the 24hrs recovery period. Therefore, the number of rats
in this group was reduced to n=7 for the analysis.
249

Brain-wide continuous monitoring of hemodynamics
The functional ultrasound imaging scanner allows repeated 2D scans in the anteroposterior direction.
For the first time, we demonstrate the applicability of this technology for large-scale and continuous
monitoring of hemodynamics in deep brain tissue before, during, and after stroke onset. After data
acquisition, every 23 cross-sections spaced of 500µm were registered and segmented on a custom-

developed digital rat atlas (see Materials and Methods) to provide a volumetric and dynamic view of the
changes in perfusion at high spatiotemporal resolution.
257
Typical plots of the temporal evolution of the hemodynamic signal are shown in Figure 2(a) for both the
MCAo (green plot) and CCAo+MCAo (orange plot) stroke models. In this example, the variations in the
cerebral perfusion were measured as rCBV compared to the baseline level in two regions of interest
located either in the ischemic core in the left hemisphere (S1BF) or in the opposite control hemisphere
(black ROI and plots). As shown, for the MCAo stroke model (Figure 2(a), green plot), the rCBV signal
dropped immediately by ~ -50% after MCAo and remained around this value during the entire duration
of the experiment. On the contrary, we observed a two-step and a more profound decrease of the rCBV
signal (Figure 2(a), orange line) in the CCAo+MCAo stroke model. First, we observed a transient drop
of the rCBV to approximately -60% of the baseline level caused by the CCA occlusion that quickly
recovered but not up to the initial values. Then, the MCAo triggers a massive reduction of ~ -95% of the
rCBV that remains at this level during the entire recording. Interestingly, in the CCAo+MCAo stroke
model, the contralesional hemisphere also shows a transient drop of ~ -15-20% of the rCBV signal after
CCAo that is not exacerbated after MCAo (Figure 2(a), black line). Note that the rCBV in the control ROI
located in the opposite hemisphere remains stable all along with the experiment in both stroke models
(Figure 2(a), dark line).
273
274 Figure 2(b) provides a detailed view of hemodynamic changes in 115 brain regions in the affected
275 hemisphere (See Supplementary Table 1) averaged for all animals in each group (MCAo group, n=7;
276 CCAo+MCAo group, n=10). Each region is sorted from the most significant loss of rCBV signal to the
277 smallest (from 1 to 115).
278
Three rCBV levels (-60, -40, and -20%) were defined to better compare the MCAo and CCAo+MCAo
stroke models. At the level above -60% rCBV, we noticed that more regions were affected in the
CCAo+MCAo than in the MCAo stroke model (~20 vs. ~10, respectively), confirming the cumulative
effect of MCAo+CCAo on the reduction of the perfusion. Note that all these regions were exclusively
located in the cortex. Between -40 and -60%, approximately the same number of regions (~30) were
affected in the two models, but we observed that those regions tend to be slightly reperfused naturally
over time (Figure 2(b)). Furthermore, we identified 50-60 regions located in the cortex with a drop of
rCBV signal below 20% of the baseline level in the MCAo group, whereas 80-90 regions were affected
for the same levels in the MCAo+CCAo group also including subcortical structures (Figure 2(b) and
Supplementary Table 1).
289
To evaluate the relevance of the rCBV signal to precisely assess the perfusion status and differentiate
each stroke model, we quantified the affected brain volume (in mm^3) for various levels of rCBV decrease
(Figure 2(c)). We observed that ischemic volume differs significantly between the MCAo and the
CCAo+MCAo groups for values ranging from -30 to -75% (Figure 2(c)), left panel; mean±sd, Unpaired
t-test, *p<0.05). The most significant difference between the two groups was observed for the value -
60% that was chosen for further analysis (Figure 2(c) to (e)).
296
The total volume of brain tissue showing an rCBV reduction of -60% was then analyzed during the entire
imaging session. We observed that the volume of ischemic tissue for the CCAo+MCAo group was
41±10mm^3 and did not show a significant change between the MCA occlusion and the end of the
imaging period (35±14mm^3; Figure 2(c), middle panel; mean±sd; p>0.44, GLM-based repeated
measures one-way ANOVA). On the contrary, the volume of tissue affected for the MCAo group tends
to shrink over time from 24±14mm^3 quickly after MCAo to 15±14mm^3 after 70mins afterward (Figure
2(c), middle panel; mean±sd; p<0.0001, GLM-based repeated measures one-way ANOVA).
Interestingly, the differences between the two groups for an rCBV level of -60% were already statistically
significantly different at 5mins after stroke onset. It became even more pronounced during the 70mins
imaging period (Figure 2(c), right panel; p=0.013; Unpaired t-test).
307
Figure 2(d) presents a typical case for each model in which we overlaid the ischemic territory on top of
each microvascular image collected during brain scans for each stroke model. We projected and
unrolled the cortical regions into a flat surface overlaid with the anatomical landmarks from the reference
rat brain atlas to appreciate the differences better. Such a view is advantageous for accurately
visualizing the regions with reduced rCBV, their spatial extent (Figure 2(e) and Supplementary Figure
1), and comparing animals and models.
314

Figure 2. Brain-wide continuous and real-time monitoring of hemodynamics during ischemic stroke. (a) A typical image of the brain microvasculature for one coronal cross-section from the 2D scan before the stroke. Only the left hemisphere is imaged entirely here because of the too-small size of the ultrasound transducer used for this study. All images were registered and segmented based on a digital version of the rat Paxinos atlas[57] (white outlines). Time series plots of the average signal in the primary somatosensory barrel-field cortex (S1BF; green and orange region) in the affected hemisphere and the control ROI located in the opposite hemisphere (black region) for each stroke model (MCAo in green; CCAo+MCAo in orange). (b) Global hemodynamic changes (rCBV) in 115 regions located in the ischemic hemisphere for both the MCAo (left and green panel) and CCAo+MCAo groups (right and orange panel). Regions were clustered according to the extent of rCBV signal loss (from more to less pronounced) (see Supplementary Table 1). (c) Left to right. Plot showing the total volume of brain region corresponding to a given rCBV decrease (%; mean±sd, multiple unpaired t-test, *p<0.05; left panel). Evolution of brain volume with 60% rCBV reduction during the 70-min post-stroke period (mean±sd; middle panel). Brain volume with a rCBV reduction of -60% at 5mins after stroke onset at a signal threshold of 60% (Unpaired t-test, *p=0.013; right panel) between MCAo (green, n=7) and CCAo+MCAo groups (orange, n=10). (d) Typical 2D brain scan of hemodynamics showing the loss in rCBV induced by either MCAo (left panel) or CCAo+MCAo (right panel). 22 cross-sections out of 23 are represented here. (e) Unrolled-cortex projection (see Supplementary Figure 1) showing the loss in rCBV induced by either MCAo (left panel) or CCAo+MCAo (right panel). D: dorsal, R: right, A: anterior, P: posterior, ß: Bregma reference point. Scale bars: 100µm.

315
316
317 **A large part of the infarct is located in regions not ischemic at 70mins after stroke onset.**
318 Cresyl violet staining of coronal brain slices was performed 24hrs after the experiments. A typical
319 example is presented in Figure 3(a). It shows that the infarcted area is localized only in the cortex in
320 both stroke models, but MCAo caused a much smaller and less extensive lesion (Figure 3(a), left panel;
321 green area) than CCAo+MCAo (Figure 3(a), right panel; orange area).
322
The statistical analysis of all rats confirmed that the infarct volume is significantly larger for the
CCAo+MCAo group (75±28mm³, n=10, mean±sd; Figure 3(b), in orange) when compared to the MCAo
group (24±10mm³, n=7, mean±sd; Figure 3(b), in green; ***p=0.0003, unpaired t-test). Note that these
results are in agreement with rCBV data, also showing a larger ischemic territory in the CCAo+MCAo
group. Taken together, this demonstrates that functional ultrasound images acquired at early time points
after stroke onset (from 5 to 70mins) have an excellent predictive value to predict brain infarction. To
confirm this hypothesis, we compared the results obtained by histopathology with those from functional
ultrasound imaging in each stroke model that were both presented using the same unrolled-cortex
projection as described previously (see Supplementary Figure 1). Figure 3(c) shows a typical case for
each group. We observed in the MCAo model (Figure 3(c), left panel) that the infarcted region (green
circle) only partially coincides with regions showing reduced rCBV (green halo). Interestingly, much of
the infarcted tissue is located rostrally to the region with low rCBV at 70mins, in an area that was not
initially ischemic. On the contrary, we observed a considerable overlap between the infarct region and

the initially ischemic region at 70mins in the CCAo+MCAo model (Figure 3(c), right panel) even though, here again, non-overlapping regions were observed in the anterior part of the brain. Nevertheless, when total infarct size was compared with the initial ischemic territory, it was preserved in the MCAo but not in the CCAo+MCAo model, in which the infarct volume was much smaller than the initial ischemic region.

Although we observed large variability in terms of the location and size of the infarct at 24hrs after stroke compared with the ischemic areas at 70mins after stroke onset for each animal (Figure 3), these results were confirmed at the group level, showing that the infarcted territory extends rostrally to the ischemic region by 1.53±1.68 and 1.37±0.66mm, respectively in MCAo and CCAo+MCAo groups (p=0.49, Unpaired t-test). This mismatch is consistent for the two groups (6/7 for MCAo and 7/7 for CCAo+MCAo rats). Additionally, we noticed that the MCAo group also shows, on average, a mismatch between the position of the infarct and the areas of low rCBV in the posterior part of the brain, where we previously observed good reperfusion during the 70-min period of the functional ultrasound imaging recording (Figure 2(b) and (c)). In both groups, these results indicate that the lack of blood supply may not be the only factor contributing to the infarction of the brain.

These findings prompted us to evaluate other important factors, such as SDs that have been shown to exacerbate focal ischemic injury by converting zones of the viable but non-functional ischemic penumbra to the core region beyond rescue[63].

Figure 3. Size and location of the region with a low rCBV signal at 70mins do partially overlap with the infarct at 24hrs. (a) Typical rat brain cross-section stained by cresyl violet to evaluate the infarct size at 24hrs after MCAo (left) or CCAo+MCAo (right). The infarcted territory is highlighted for each stroke model (green and orange, respectively). (b) Comparison of the infarct volume (mm^3) between the two models, showing the CCAo+MCAo (orange) display a statistically more extensive infarct than the MCAo group (green; Unpaired t-test, ***p=0.0003). (c) The infarct (colored halos) and the region with low rCBV levels (colored circles) are overlaid on the unrolled-cortex projection for one typical rat. (d) Same as in (c) but for all rats used in the study. D: dorsal, R: right, A: anterior, P: posterior, ß: Bregma reference point.

Spreading depolarizations start quickly after MCAo.
As shown in previous studies, functional ultrasound imaging is suitable for whole-brain tracking of transient hyperemic events associated with SDs[45,64]. The imaging protocol that we developed for this study allows for detection, monitoring, and quantification of SDs in real-time before, during, and after stroke onset. The spatiotemporal profile of SDs was extracted by averaging the rCBV signal in all voxels from the retrosplenial granular cortex (RSG) that is located in the vicinity of the region with the low level of rCBV (Figure 4(a), left panel). As already observed in previous studies, SDs were triggering transient hyperemia characterized by a rapid and massive increase of the rCBV signal (>+150% rCBV), followed by a sustained reduction of the baseline level (Figure 4(a), right panel, and Movies 2 and 3). We did not observe a significant difference in the number (4.3±2.3 and 4.7±2.9, mean±sd, p=0.7557, Unpaired t-test) and the frequency (1 SD each 17.5mins and 15.2mins, central frequency) of SDs between the MCAo and CCAo+MCAo groups (Figure 4(b)). A detailed analysis of one individual animal revealed that all SDs are generated within the same brain region (Movie 3). The sham controls had a maximum of 1 SD occurring in the first 5mins but did not show any infarcted tissue in the *postmortem* analysis (n=4, data not shown). It has been reported that the initial SD could be due to the mechanical pressure during the mechanical clipping to occlude the MCA as previously reported[65]. If we exclude this initial SD from

our analysis, we observed that the second SD appears rapidly within ~10mins for both models (MCAo:
6±11mins, CCAo+MCAo: 11±9mins). The group analysis of functional ultrasound imaging data for all
animals revealed that most SDs were originating from the primary somatosensory cortex (72.2% for
MCAo versus 83.3% for CCAo+MCAo) between S1ULp and S1FL (centroid coordinates: ß+1.5/2mm,
Lateral 4.5/5mm) in the close vicinity of the area with low rCBV signal (Figure 4(c)). SDs are propagating
at a constant velocity. We did not observe differences in the velocity of SDs between MCAo and
CCAo+MCAo groups (respectively 5.8±1.0 and 6.4±2.4mm/min, mean±sd; p=0.5683, Unpaired t-test;
Figure 4(d)), but the values are in agreement with others studies[66]. Figure 4(e) presents a typical color-
coded map overlaid on the unroll cortex projection showing the trajectory and progression of a standard
SD from the anterior to the posterior part of the brain by contouring the ischemic territory (Figure 4(e)
and Movie 3).
385

Figure 4. Real-time monitoring of spreading depolarizations using functional ultrasound imaging. (a) A coronal cross-section extracted from the 2D functional ultrasound imaging scan after stroke onset shows the substantial reduction of the rCBV signal in the cortex (red area). The time series plot of the average hemodynamic signal (rCBV) in the retrosplenial granular cortex (green and orange dotted line) from the ischemic hemisphere for each stroke model (MCAo in green; CCAo+MCAo in orange). (b) Monitoring of SDs. Each horizontal line represents one animal; each dot corresponds to the transient hemodynamic increase associated with an SD. The average SDs are not statistically different between the two groups (Unpaired t-test, ns=0.7557). (c) Location of the epicenter of SDs for the MCAo (green dots) and CCAo+MCAo (orange dots) models with respective centroids (mean±sd). (d) A plot of the distribution of the average velocity of SDs for MCAo (green) and CCAo+MCAo models (orange). Each point represents one rat. The average SD velocity is not statistically different between the two groups (Unpaired t-test, ns=0.5683). (e) Color-coded map showing the propagation of an SD in a concentric manner around the ischemic core from its epicenter located in the anterior part of the brain to the posterior part. D: dorsal, R: right, ß: Bregma reference point. Scale bar: 100µm.

Discussion

388 To recapitulate the diversity of stroke observed in patients, various experimental models and surgical
389 procedures are employed in rodents, including transient or permanent occlusion of MCA and/or CCA
390 with a filament, photo- or chemo-thrombosis[15–18]. However, each strategy is often associated with
391 complex and multifaceted stroke mechanisms, limiting their comparability[17,18].
392

393 In this work, we used two stroke models, in nearly identical conditions, using permanent occlusion (with
394 a vascular micro-clip) of the MCAo combined with or without CCAo (ligature) to reduce collateral
395 reperfusion, causing either small or large tissue infarction, respectively. This procedure was chosen to
396 allow a fair comparison of brain-wide hemodynamic monitoring with functional ultrasound imaging
397 between the two models, from early time point after stroke onset and during 70mins afterward.
398

Several imaging technologies such as magnetic resonance angiography (MRA), computed tomography
perfusion, contrast-enhanced CT angiography, and transcranial Doppler sonography (TCD) are used to
seek evidence of plausible stroke mechanisms and assess the efficiency of recanalization strategies[67].

However, hemodynamic patterns are highly variable since either improvements or deterioration related
to dynamic changes in brain perfusion may occur early during the clinical course of ischemic stroke.
These changes are often associated with spontaneous thrombolysis, re-occlusion, micro-embolism,
thrombus propagation, and collateralization[68]. CT perfusion can be performed within 30mins after
hospital admission, but most often information on brain hemodynamics is available only after several
hours to days after the stroke onset, which may be too late for significantly improving the diagnosis,
care, and prognosis of stroke patients[69]. Early, continuous and repeatable monitoring of cerebral
hemodynamics may offer new insights into acute ischemic stroke pathogenesis. In theory, precise
recording of brain perfusion may even guide therapeutic interventions to prevent neurological
deterioration or plan conservative treatment in response to particular hemodynamic patterns.

This study demonstrated that functional ultrasound imaging is a suitable technology for continuous
monitoring of brain-wide hemodynamics during stroke, allowing for real-time assessment of the ischemic
insult[10,50,51]. For the first time, we analyzed the changes in rCBV at high spatiotemporal resolution in
both MCAo and CCAo+MCAo stroke rat models, which are used mainly for preclinical stroke research[70].
Technically, we performed repeated 2D scans in the anteroposterior direction (23 positions, 11×12.8×9-
mm^3 total brain volume) to continuously monitor the rCBV signal from stroke onset to 70mins afterward.
Functional ultrasound imaging data were then registered and segmented across 115 regions based on
a digital version of the rat Paxinos atlas[57] using a custom software solution previously developed for
mice[55,56,71]. This analytics pipeline is open-access and available for research purposes[61]. It is a powerful
tool for efficiently comparing hemodynamic changes in the entire brain between animals of the same
group and between groups. Volumetric (Figure 2(d)) and cortical projection (Figure 2(e)) maps of the
ischemic territory for each rat were reconstructed from the stack of the 2D slices to analyze the variability
between animals and to compare it with histopathological data on the infarct (Figure 3(d)).

As expected, we observed that the rCBV loss and its spatial extend are more pronounced in the
CCAo+MCAo than in the MCAo model, probably due to a reduction of the collateral flow in MCA territory
after CCAo[72–74]. We defined four thresholds (above -60%, -40%, -20%, and below as compared to the
baseline) to compare the two models and showed that the identity and number of regions in each
category strongly differ between the two models. Moreover, we observed complex hemodynamic
patterns throughout the entire imaging period, including pre-occlusion, early before and after
CCAo+MCAo, to 70mins afterward. Thanks to the high spatiotemporal resolution of functional
ultrasound imaging and its ability to image hemodynamics in deep brain regions, we measured
decreasing rCBV levels in the ischemic territory from the cortical surface to deep subcortical regions
with typical rCBV values ranging from -100 to 60% in the cortex, ~ -40% in the hippocampus and ~ -
20% in the striatum (Figure 2(b)).

Even if we observed considerable variability in the size and location of the ischemic regions between
animals (Figure 3(d)), we demonstrated that the MCAo and CCAo+MCAo stroke models are statistically
significantly different when comparing the number of regions with an rCBV level below -60% (Figure
2(c) to (e)). For this threshold, we observed that all ischemic regions were located in the cortex (Figure
2(b)) and that the volume of tissue with an rCBV value below -60% progressively decreases from 24mm^3
5mins after stroke onset to 15mm^3, 70mins post-stroke in the MCAo model (pvalue=0.0117, Paired t-
test; Figure 2(c); middle panel). Such an effect was not observed for the CCAo+MCAo model
(pvalue=0.1022, Paired t-test; Figure 2(c), middle panel). When comparing only the regions with an
rCBV level above -60%, we showed a slow but continuous perfusion increase following the MCAo in
both stroke models (Figure 2(b)).

We quantified the infarct size and found it statistically significantly smaller for the MCAo than for the
CCAo+MCAo model (25 vs. 75mm^3) despite considerable inter animal variability in both models, as
previously reported for various preclinical stroke models[75] and patients[76]. Notably, we observed a
significant dispersion of infarct sizes in the CCAo+MCAo than in the MCAo model, possibly because of
spontaneous reperfusion that may occur outside the imaging period (Figure 2(b)), even though this
model has been shown to provide more reproducible stroke lesions[21,77,78]. Notably, the infarct volume
measured at 24hrs *postmortem* did match those of low rCBV measured *in vivo* using functional
ultrasound imaging for the MCAo but not the CCAo+MCAo model. For the latter, the infarct size was
approximately twice as large (80 vs. 40mm^3, respectively) as expected from the rCBV data. Hence, brain
regions with a significant reduction of rCBV as measured by functional ultrasound imaging did not fully
overlap with infarct size and location, especially for the CCAo+MCAo stroke model, in contrast to what
was previously observed in a thromboembolic mice model[51].

462
Nevertheless, we hypothesize that functional ultrasound imaging may predict more accurately the size
and location of the constituted infarct at 24hrs by merging information on the level of rCBV with analysis
of hemodynamic changes associated with SDs[45,64]. Indeed, we showed that the final brain infarct was
located more rostrally as compared to the initial ischemic territory measured 70mins after stroke in both
models (Figure 3(c) and (d)). Additionally, we demonstrate that the SDs propagate from one point of
origin in a rostrocaudal direction with a median velocity of ~6mm/min, which agrees with the literature[79–
83]. Interestingly, all SDs originated within brain regions that are part of the infarct but are not ischemic in
the first 70mins. It is important to note that SDs may occur up to 6hrs after stroke onset, which has been
shown to cause an expansion of core-infarcted tissue[66,84]. We did not observe differences in the number
of the SDs between the MCAo and CCAo+MCAo stroke during the 70mins imaging period confirming
previous results, which demonstrated that the number of SDs was not directly associated with the
expansion of the ischemic insult[81]. In short, our work shows the relevance of functional ultrasound
imaging to better understand early hemodynamic changes in preclinical stroke research. It enlightens
their high complexity, which may make stroke a heterogeneous disorder. Moreover, our results suggest
that the primary somatosensory cortex in rats is highly sensitive during a stroke and becomes a hot spot
for generating spreading depolarization, which may contribute to delayed infarction of this region.
479
We used two rat stroke models to evaluate the relevance of functional ultrasound imaging in two
ischemic conditions that may recapitulate the variability of infarct size and location observed in the clinic.
However, functional ultrasound imaging requires a craniotomy, especially for imaging deep brain
regions, since the skull is strongly attenuating high-frequency ultrasound signals[85]. Importantly, it has
been demonstrated that both neurological behavior and infarction size were significantly better in rats
treated very early by decompressive craniectomy (4hrs) after endovascular MCAo[86]. Therefore, it is
likely that our results may underestimate those in close skull since acute ischemia in the MCA territory
may lead to cerebral edema with raised intracranial pressure and reduce collateral blood flow, which is
associated with larger infarct size[87].
489
Two other limitations of our study relate to i) the lack of monitoring of physiological parameters (i.e.,
blood pressure, blood gas analysis,…) and ii) the anesthesia. It has been demonstrated that several
anesthetics have dose-specific effects on cerebral blood flow[88–91] but also affects neurovascular
coupling, autoregulation, ischemic depolarizations[91–94], excitotoxicity, inflammation, neural networks,
and numerous molecular pathways relevant for stroke outcome[91,95]. Additionally, we did not analyze the
behavioral outcome, which the STAIR committee recommends for improving the translation value of
preclinical stroke research[70]. These limitations could be further addressed using volumetric functional
ultrasound imaging (vfUSI) that has been recently developed and based on a 2D-array transducer to
acquire 3D images of brain activity in awake conditions[55].
499
500 The versatility, ease-of-use, contrast-free, low cost, portability, and recent success in the clinic[60,96,97]
501 makes functional ultrasound imaging a promising tool to monitor and map large-scale and in-depth
502 hemodynamics for a better diagnosis and prognosis of various brain vascular disorders.

Author contributions

	Brunner	Lagumersindez Denis	Gertz	Grillet	Montaldo	Endres	Urban
Concept	X	X	X			X	X
Methodology	X	X	X				X
Software					X		X
Imaging	X						
Histochemistry		X		X			
Data Analysis - Interpretation	X	X			X		X
Manuscript	X				X	X	X
Proofreading	X	X	X	X	X	X	X
Supervision			X		X	X	X
Funding						X	X

Funding
This work is supported by grants from the Fondation Leducq (15CVD02) and KU Leuven (C14/18/099-STYMULATE-STROKE). The functional ultrasound imaging platform is supported by grants from FWO (MEDI-RESCU2-AKUL/17/049, G091719N, and 1197818N), VIB Tech-Watch (fUSI-MICE), Neuro-Electronics Research Flanders TechDev fund (3D-fUSI project). The Endres lab received funding from DFG under Germany's Excellence Strategy – EXC-2049 – 390688087, BMBF, DZNE, DZHK, EU, Corona Foundation, and Fondation Leducq

Acknowledgements
The authors thank all the members of the Fondation Leducq network #15CVD02 for helpful comments and discussions. We thank all NERF animal caretakers, including I. Eyckmans, F. Ooms, and S. Luijten, for their help with the management of the animals.

Disclosure
A.U. is the founder and a shareholder of AUTC company commercializing functional ultrasound imaging solutions for preclinical and clinical research.

521 **Supplementary Materials**

Supplementary Figure 1

Supplementary Figure 1. Unrolled-cortex projection.

Supplementary Table 1. List of the 115 brain regions extracted from a digital version of the rat Paxinos atlas organized by main anatomical structures (left) and clustered based on the average rCBV loss after the MCAo (right).

Movie 1. Whole-brain scans before (left) and 70mins after CCA and MCA occlusions (right).

Movie 2. Movie of hemodynamic changes induced by concomitant CCA and MCA occlusions observed in a single coronal plane (ß+3mm) extracted from the whole-brain scan.

Movie 3. Movie of hemodynamic changes induced by concomitant CCA and MCA occlusions observed with the unrolled-cortex projection.

References

Astrup J, Siesjö BK, Symon L. Thresholds in cerebral ischemia - the ischemic penumbra. *Stroke* 1981; **12**: 723–725.

Maas MB, Lev MH, Ay H, Singhal AB, Greer DM, Smith WS *et al.* Collateral vessels on CT angiography predict outcome in acute ischemic stroke. *Stroke* 2009; **40**: 3001–3005.

Saver JL. Time is brain--quantified. *Stroke* 2006; **37**: 263–266.

Vagal A, Aviv R, Sucharew H, Reddy M, Hou Q, Michel P *et al.* Collateral clock is more important than time clock for tissue fate. *Stroke* 2018; **49**: 2102–2107.

Krafft PR, Bailey EL, Lekic T, Rolland WB, Altay O, Tang J *et al.* Etiology of stroke and choice of models. *Int J Stroke* 2012; **7**: 398–406.

Ginsberg MD. Adventures in the pathophysiology of brain ischemia: penumbra, gene expression, neuroprotection: the 2002 Thomas Willis Lecture. *Stroke* 2003; **34**: 214–223.

Heiss W-D, Kracht LW, Thiel A, Grond M, Pawlik G. Penumbral probability thresholds of cortical flumazenil binding and blood flow predicting tissue outcome in patients with cerebral ischemia. *Brain* 2001; **124**: 20–29.

Howells DW, Porritt MJ, Rewell SSJ, O'Collins V, Sena ES, van der Worp HB *et al.* Different strokes for different folks: the rich diversity of animal models of focal cerebral ischemia. *J Cereb Blood Flow Metab* 2010; **30**: 1412–1431.

Longa EZ, Weinstein PR, Carlson S, Cummins R. Reversible middle cerebral artery occlusion without craniectomy in rats. *Stroke* 1989; **20**: 84–91.

Brunner C, Korostelev M, Raja S, Montaldo G, Urban A, Baron J-C. Evidence from functional ultrasound imaging of enhanced contralesional microvascular response to somatosensory stimulation in acute middle cerebral artery occlusion/reperfusion in rats: A marker of ultra-early network reorganization? *J Cereb Blood Flow Metab* 2018; **38**: 1690–1700.

Lay CC, Jacobs N, Hancock AM, Zhou Y, Frostig RD. Early stimulation treatment provides complete sensory-induced protection from ischemic stroke under isoflurane anesthesia. *Eur J Neurosci* 2013; **38**: 2445–2452.

Imai H, Saito N, Macrae IM. Rat middle cerebral artery (MCA) occlusion models which involve a frontotemporal craniectomy. In: *Neuromethods*. Springer New York: New York, NY, 2016, pp 55–71.

Yao H, Sugimori H, Fukuda K, Takada J, Ooboshi H, Kitazono T *et al.* Photothrombotic Middle Cerebral Artery Occlusion and Reperfusion Laser System in Spontaneously Hypertensive Rats. *Stroke* 2003; **34**: 2716–2721.

Nikolova S, Moyanova S, Hughes S, Bellyou-Camilleri M, Lee T-Y, Bartha R. Endothelin-1 induced MCAO: dose dependency of cerebral blood flow. *J Neurosci Methods* 2009; **179**: 22–28.

Karatas H, Erdener SE, Gursoy-Ozdemir Y, Gurer G, Soylemezoglu F, Dunn AK *et al.* Thrombotic distal middle cerebral artery occlusion produced by topical FeCl(3) application: a novel model suitable for intravital microscopy and thrombolysis studies. *J Cereb Blood Flow Metab* 2011; **31**: 1452–1460.

Jia J-M, Chowdary PD, Gao X, Ci B, Li W, Mulgaonkar A *et al.* Control of cerebral ischemia with magnetic nanoparticles. *Nat Methods* 2017; **14**: 160–166.

MacRae I. Preclinical stroke research - Advantages and disadvantages of the most common rodent models of focal ischaemia. *Br J Pharmacol* 2011; **164**: 1062–1078.

Fluri F, Schuhmann MK, Kleinschnitz C. Animal models of ischemic stroke and their application in clinical research. *Drug Des Devel Ther* 2015; **9**: 3445–3454.

Ingberg E, Dock H, Theodorsson E, Theodorsson A, Ström JO. Method parameters' impact on mortality and variability in mouse stroke experiments: a meta-analysis. *Sci Rep* 2016; **6**: 21086.

Ström JO, Ingberg E, Theodorsson A, Theodorsson E. Method parameters' impact on mortality and variability in rat stroke experiments: a meta-analysis. *BMC Neurosci* 2013; **14**: 41.

Brint S, Jacewicz M, Kiessling M, Tanabe J, Pulsinelli W. Focal brain ischemia in the rat: methods for reproducible neocortical infarction using tandem occlusion of the distal middle cerebral and ipsilateral common carotid arteries. *J Cereb Blood Flow Metab* 1988; **8**: 474–485.

Hansen AJ, Quistorff B, Gjedde A. Relationship between local changes in cortical blood flow and extracellular K+ during spreading depression. *Acta Physiol Scand* 1980; **109**: 1–6.

Dreier JP, Körner K, Ebert N, Görner A, Rubin I, Back T *et al.* Nitric oxide scavenging by hemoglobin or nitric oxide synthase inhibition by N-nitro-L-arginine induces cortical spreading ischemia when K+ is increased in the subarachnoid space. *J Cereb Blood Flow Metab* 1998; **18**: 978–990.

Shin HK, Dunn AK, Jones PB, Boas DA, Moskowitz MA, Ayata C. Vasoconstrictive neurovascular coupling during focal ischemic depolarizations. *J Cereb Blood Flow Metab* 2006; **26**: 1018–1030.

25 Mies G, Iijima T, Hossmann KA. Correlation between peri-infarct DC shifts and ischaemic neuronal
damage in rat. *Neuroreport* 1993; **4**: 709–711.
26 Busch E, Gyngell ML, Eis M, Hoehn-Berlage M, Hossmann KA. Potassium-induced cortical
spreading depressions during focal cerebral ischemia in rats: contribution to lesion growth assessed
by diffusion-weighted NMR and biochemical imaging. *J Cereb Blood Flow Metab* 1996; **16**: 1090–
1099.
27 Back T, Ginsberg MD, Dietrich WD, Watson BD. Induction of spreading depression in the ischemic
hemisphere following experimental middle cerebral artery occlusion: effect on infarct morphology.
J Cereb Blood Flow Metab 1996; **16**: 202–213.
28 Dijkhuizen RM, Beekwilder JP, van der Worp HB, Berkelbach van der Sprenkel JW, Tulleken KAF,
Nicolay K. Correlation between tissue depolarizations and damage in focal ischemic rat
brain1Published on the World Wide Web on 12 July 1999.1. *Brain Res* 1999; **840**: 194–205.
29 von Bornstädt D, Houben T, Seidel JL, Zheng Y, Dilekoz E, Qin T *et al.* Supply-demand mismatch
transients in susceptible peri-infarct hot zones explain the origins of spreading injury
depolarizations. *Neuron* 2015; **85**: 1117–1131.
30 Herson PS, Traystman RJ. Animal models of stroke: translational potential at present and in 2050.
Future Neurol 2014; **9**: 541–551.
31 Donahue MJ, Achten E, Cogswell PM, De Leeuw F-E, Derdeyn CP, Dijkhuizen RM *et al.* Consensus
statement on current and emerging methods for the diagnosis and evaluation of cerebrovascular
disease. *J Cereb Blood Flow Metab* 2018; **38**: 1391–1417.
32 Baron J-C. Recent advances in mesoscopic-scale imaging in animal models of ischemic stroke.
Curr Opin Neurol 2016; **29**: 104–111.
33 Reid E, Graham D, Lopez-Gonzalez MR, Holmes WM, Macrae IM, McCabe C. Penumbra detection
using PWI/DWI mismatch MRI in a rat stroke model with and without comorbidity: comparison of
methods. *J Cereb Blood Flow Metab* 2012; **32**: 1765–1777.
34 Baskerville TA, McCabe C, Weir CJ, Macrae IM, Holmes WM. Noninvasive MRI measurement of
CBF: evaluating an arterial spin labelling sequence with 99mTc-HMPAO CBF autoradiography in a
rat stroke model. *J Cereb Blood Flow Metab* 2012; **32**: 973–977.
35 Chen F, Suzuki Y, Nagai N, Sun X, Coudyzer W, Yu J *et al.* Delayed perfusion phenomenon in a
rat stroke model at 1.5 T MR: an imaging sign parallel to spontaneous reperfusion and ischemic
penumbra? *Eur J Radiol* 2007; **61**: 70–78.
36 Szigeti K, Horváth I, Veres DS, Martinecz B, Lénárt N, Kovács N *et al.* A novel SPECT-based
approach reveals early mechanisms of central and peripheral inflammation after cerebral ischemia.
J Cereb Blood Flow Metab 2015; **35**: 1921–1929.
37 Morris GP, Wright AL, Tan RP, Gladbach A, Ittner LM, Vissel B. A comparative study of variables
influencing ischemic injury in the longa and koizumi methods of intraluminal filament middle cerebral
artery occlusion in mice. *PLoS One* 2016; **11**. doi:10.1371/journal.pone.0148503.
38 Beretta S, Versace A, Carone D, Riva M, Dell'Era V, Cuccione E *et al.* Cerebral collateral
therapeutics in acute ischemic stroke: A randomized preclinical trial of four modulation strategies.
J Cereb Blood Flow Metab 2017; **37**: 3344–3354.
39 Yuan L, Li Y, Li H, Lu H, Tong S. Intraoperative laser speckle contrast imaging improves the stability
of rodent middle cerebral artery occlusion model. *J Biomed Opt* 2015; **20**: 096012.
40 Grinvald A, Lieke E, Frostig RD, Gilbert CD, Wiesel TN. Functional architecture of cortex revealed
by optical imaging of intrinsic signals. *Nature* 1986; **324**: 361–364.
41 Shih AY, Driscoll JD, Drew PJ, Nishimura N, Schaffer CB, Kleinfeld D. Two-photon microscopy as
a tool to study blood flow and neurovascular coupling in the rodent brain. *J Cereb Blood Flow Metab*
2013; **33**: 319–319.
42 Razansky D, Klohs J, Ni R. Multi-scale optoacoustic molecular imaging of brain diseases. *Eur J*
Nucl Med Mol Imaging 2021. doi:10.1007/s00259-021-05207-4.
43 Gottschalk S, Degtyaruk O, Mc Larney B, Rebling J, Hutter MA, Deán-Ben XL *et al.* Rapid
volumetric optoacoustic imaging of neural dynamics across the mouse brain. *Nat Biomed Eng* 2019;
3: 392–401.
44 Lv J, Li S, Zhang J, Duan F, Wu Z, Chen R *et al.* In vivo photoacoustic imaging dynamically monitors
the structural and functional changes of ischemic stroke at a very early stage. *Theranostics* 2020;
10: 816–828.
45 Macé E, Montaldo G, Cohen I, Baulac M, Fink M, Tanter M. Functional ultrasound imaging of the
brain. *Nature methods* 2011; **8**: 662–664.
46 Macé E, Montaldo G, Osmanski B-F, Cohen I, Fink M, Tanter M. Functional ultrasound imaging of
the brain: theory and basic principles. *IEEE Trans Ultrason Ferroelectr Freq Control* 2013; **60**: 492–
506.

47 Montaldo G, Tanter M, Bercoff J, Benech N, Fink M. Coherent plane-wave compounding for very
high frame rate ultrasonography and transient elastography. *IEEE Trans Ultrason Ferroelectr Freq*
Control 2009; **56**: 489–506.
48 Urban A, Mace E, Brunner C, Heidmann M, Rossier J, Montaldo G. Chronic assessment of cerebral
hemodynamics during rat forepaw electrical stimulation using functional ultrasound imaging.
Neuroimage 2014; **101**: 138–149.
49 Urban A, Golgher L, Brunner C, Gdalyahu A, Har-Gil H, Kain D *et al.* Understanding the
neurovascular unit at multiple scales: Advantages and limitations of multi-photon and functional
ultrasound imaging. *Adv Drug Deliv Rev* 2017; **119**: 73–100.
50 Brunner C, Isabel C, Martin A, Dussaux C, Savoye A, Emmrich J *et al.* Mapping the dynamics of
brain perfusion using functional ultrasound in a rat model of transient middle cerebral artery
occlusion. *J Cereb Blood Flow Metab* 2017; **37**: 263–276.
51 Hingot V, Brodin C, Lebrun F, Heiles B, Chagnot A, Yetim M *et al.* Early Ultrafast Ultrasound
Imaging of Cerebral Perfusion correlates with Ischemic Stroke outcomes and responses to
treatment in Mice. *Theranostics* 2020; **10**: 7480–7491.
52 Sieu L-A, Bergel A, Tiran E, Deffieux T, Pernot M, Gennisson J-L *et al.* EEG and functional
ultrasound imaging in mobile rats. *Nat Methods* 2015; **12**: 831–834.
53 Tiran E, Ferrier J, Deffieux T, Gennisson J-L, Pezet S, Lenkei Z *et al.* Transcranial Functional
Ultrasound Imaging in Freely Moving Awake Mice and Anesthetized Young Rats without Contrast
Agent. *Ultrasound Med Biol* 2017; **43**: 1679–1689.
54 Urban A, Dussaux C, Martel G, Brunner C, Mace E, Montaldo G. Real-time imaging of brain activity
in freely moving rats using functional ultrasound. *Nature methods* 2015; **12**: 873–878.
55 Brunner C, Grillet M, Sans-Dublanc A, Farrow K, Lambert T, Macé E *et al.* A Platform for Brain-
wide Volumetric Functional Ultrasound Imaging and Analysis of Circuit Dynamics in Awake Mice.
Neuron 2020. doi:10.1016/j.neuron.2020.09.020.
56 Brunner C, Grillet M, Urban A, Roska B, Montaldo G, Macé E. Whole-brain functional ultrasound
imaging in awake head-fixed mice. *Nat Protoc* 2021. doi:10.1038/s41596-021-00548-8.
57 Paxinos G. *The rat brain in stereotaxic coordinates*. 6th ed. Academic Press, 2014.
58 Percie du Sert N, Hurst V, Ahluwalia A, Alam S, Avey MT, Baker M *et al.* The ARRIVE guidelines
2.0: Updated guidelines for reporting animal research. *PLoS Biol* 2020; **18**: e3000410.
59 Wayman C, Duricki DA, Roy LA, Haenzi B, Tsai S-Y, Kartje G *et al.* Performing permanent distal
middle cerebral with common carotid artery occlusion in aged rats to study cortical ischemia with
sustained disability. *J Vis Exp* 2016; : 53106.
60 Urban A, Brunner C, Dussaux C, Chassoux F, Devaux B, Montaldo G. Functional Ultrasound
Imaging of Cerebral Capillaries in Rodents and Humans. *Jacobs Journal of Molecular and*
Translational Medicine 2015; **1**: 007.
61 ClemBrunner, nerf-common. *nerf-common/whole-brain-fUS: whole-brain-fUS v1.0.* 2021
doi:10.5281/zenodo.4585348.
62 Schindelin J, Arganda-Carreras I, Frise E, Kaynig V, Longair M, Pietzsch T *et al.* Fiji: an open-
source platform for biological-image analysis. *Nat Methods* 2012; **9**: 676–682.
63 Menyhárt Á, Zölei-Szénási D, Puskás T, Makra P, Orsolya MT, Szepes BÉ *et al.* Spreading
depolarization remarkably exacerbates ischemia-induced tissue acidosis in the young and aged rat
brain. *Sci Rep* 2017; **7**. doi:10.1038/s41598-017-01284-4.
64 Rabut C, Correia M, Finel V, Pezet S, Pernot M, Deffieux T *et al.* 4D functional ultrasound imaging
of whole-brain activity in rodents. *Nature methods* 2019; **16**: 994–997.
65 Akerman S, Holland PR, Goadsby PJ. Mechanically-induced cortical spreading depression
associated regional cerebral blood flow changes are blocked by Na+ ion channel blockade. *Brain*
Res 2008; **1229**: 27–36.
66 Ayata C, Lauritzen M. Spreading depression, spreading depolarizations, and the cerebral
vasculature. *Physiol Rev* 2015; **95**: 953–993.
67 Zhang Z, Pu Y, Mi D, Liu L. Cerebral hemodynamic evaluation after cerebral recanalization therapy
for acute ischemic stroke. *Front Neurol* 2019; **10**: 719.
68 Akopov S, Whitman GT. Hemodynamic studies in early ischemic stroke. *Stroke* 2002; **33**: 1274–
1279.
69 Tsivgoulis G, Apostolidou N, Giannopoulos S, Sharma VK. Hemodynamic causes of deterioration
in acute ischemic stroke. *Perspect Med* 2012; **1**: 177–184.
70 Fisher M, Feuerstein G, Howells DW, Hurn PD, Kent TA, Savitz SI *et al.* Update of the stroke
therapy academic industry roundtable preclinical recommendations. *Stroke* 2009; **40**: 2244–2250.

71 Macé É, Montaldo G, Trenholm S, Cowan C, Brignall A, Urban A *et al.* Whole-Brain Functional
Ultrasound Imaging Reveals Brain Modules for Visuomotor Integration. *Neuron* 2018; **100**: 1241-
1251.e7.
72 Bederson JB, Pitts LH, Tsuji M, Nishimura MC, Davis RL, Bartkowski H. Rat middle cerebral artery
occlusion: evaluation of the model and development of a neurologic examination. *Stroke* 1986; **17**:
472–476.
73 Tanaka K, Ogawa N, Asanuma M, Kondo Y, Nomura M. Relationship between cholinergic
dysfunction and discrimination learning disabilities in Wistar rats following chronic cerebral
hypoperfusion. *Brain Res* 1996; **729**: 55–65.
74 Farkas E, Luiten PGM, Bari F. Permanent, bilateral common carotid artery occlusion in the rat: a
model for chronic cerebral hypoperfusion-related neurodegenerative diseases. *Brain Res Rev*
2007; **54**: 162–180.
75 Durand A, Chauveau F, Cho T-H, Bolbos R, Langlois J-B, Hermitte L *et al.* Spontaneous reperfusion
after in situ thromboembolic stroke in mice. *PLoS One* 2012; **7**: e50083.
76 Barber PA, Davis SM, Infeld B, Baird AE, Donnan GA, Jolley D *et al.* Spontaneous reperfusion after
ischemic stroke is associated with improved outcome. *Stroke* 1998; **29**: 2522–2528.
77 Chen ST, Hsu CY, Hogan EL, Maricq H, Balentine JD. A model of focal ischemic stroke in the rat:
reproducible extensive cortical infarction. *Stroke* 1986; **17**: 738–743.
78 Coert BA, Anderson RE, Meyer FB. Reproducibility of cerebral cortical infarction in the wistar rat
after middle cerebral artery occlusion. *J Stroke Cerebrovasc Dis* 1999; **8**: 380–387.
79 Nakamura H, Strong AJ, Dohmen C, Sakowitz OW, Vollmar S, Sué M *et al.* Spreading
depolarizations cycle around and enlarge focal ischaemic brain lesions. *Brain* 2010; **133**: 1994–
2006.
80 Takeda Y, Zhao L, Jacewicz M, Pulsinelli WA, Nowak TS Jr. Metabolic and perfusion responses to
recurrent peri-infarct depolarization during focal ischemia in the Spontaneously Hypertensive Rat:
dominant contribution of sporadic CBF decrements to infarct expansion. *J Cereb Blood Flow Metab*
2011; **31**: 1863–1873.
81 Bere Z, Obrenovitch TP, Kozák G, Bari F, Farkas E. Imaging reveals the focal area of spreading
depolarizations and a variety of hemodynamic responses in a rat microembolic stroke model. *J
Cereb Blood Flow Metab* 2014; **34**: 1695–1705.
82 Farkas E, Pratt R, Sengpiel F, Obrenovitch TP. Direct, live imaging of cortical spreading depression
and anoxic depolarisation using a fluorescent, voltage-sensitive dye. *J Cereb Blood Flow Metab*
2008; **28**: 251–262.
83 Bogdanov VB, Middleton NA, Theriot JJ, Parker PD, Abdullah OM, Ju YS *et al.* Susceptibility of
primary sensory cortex to spreading depolarizations. *J Neurosci* 2016; **36**: 4733–4743.
84 Risher WC, Ard D, Yuan J, Kirov SA. Recurrent spontaneous spreading depolarizations facilitate
acute dendritic injury in the ischemic penumbra. *J Neurosci* 2010; **30**: 9859–9868.
85 Tsai P-C, Gougheri HS, Kiani M. Skull impact on the ultrasound beam profile of transcranial focused
ultrasound stimulation. *Annu Int Conf IEEE Eng Med Biol Soc* 2019; **2019**: 5188–5191.
86 Doerfler A, Forsting M, Reith W, Christian Staff, Heiland S, Schäbitz W-R *et al.* Decompressive
craniectomy in a rat model of "malignant" cerebral hemispheric stroke: experimental support for an
aggressive therapeutic approach. *Neurosurg Focus* 1996; **1**: E1.
87 Beard DJ, Logan CL, McLeod DD, Hood RJ, Pepperall D, Murtha LA *et al.* Ischemic penumbra as
a trigger for intracranial pressure rise - A potential cause for collateral failure and infarct
progression? *J Cereb Blood Flow Metab* 2016; **36**: 917–927.
88 Newberg LA, Michenfelder JD, Chalon J. Cerebral protection by isoflurane during hypoxemia or
ischemia. *Surv Anesthesiol* 1984; **28**: 83.
89 Seto A, Taylor S, Trudeau D, Swan I, Leung J, Reeson P *et al.* Induction of ischemic stroke in
awake freely moving mice reveals that isoflurane anesthesia can mask the benefits of a
neuroprotection therapy. *Front Neuroenergetics* 2014; **6**. doi:10.3389/fnene.2014.00001.
90 Hillman TC, Matei N, Tang J, Zhang JH. Developing a standardized system of exposure and
intervention endpoints for isoflurane in preclinical stroke models. *Med Gas Res* 2019; **9**: 46–51.
91 Hoffmann U, Sheng H, Ayata C, Warner DS. Anesthesia in Experimental Stroke Research. *Transl
Stroke Res* 2016; **7**: 358–367.
92 Kudo C, Toyama M, Boku A, Hanamoto H, Morimoto Y, Sugimura M *et al.* Anesthetic effects on
susceptibility to cortical spreading depression. *Neuropharmacology* 2013; **67**: 32–36.
93 Kudo C, Nozari A, Moskowitz MA, Ayata C. The impact of anesthetics and hyperoxia on cortical
spreading depression. *Exp Neurol* 2008; **212**: 201–206.

94 Saito R, Graf R, Hübel K, Fujita T, Rosner G, Heiss WD. Reduction of infarct volume by halothane:
effect on cerebral blood flow or perifocal spreading depression-like depolarizations. *J Cereb Blood*
Flow Metab 1997; **17**: 857–864.
95 Masamoto K, Fukuda M, Vazquez A, Kim S-G. Dose-dependent effect of isoflurane on
neurovascular coupling in rat cerebral cortex. *Eur J Neurosci* 2009; **30**: 242–250.
96 Imbault M, Chauvet D, Gennisson J-L, Capelle L, Tanter M. Intraoperative Functional Ultrasound
Imaging of Human Brain Activity. *Sci Rep* 2017; **7**: 7304.
97 Soloukey S, Vincent AJPE, Satoer DD, Mastik F, Smits M, Dirven CMF *et al.* Functional Ultrasound
(fUS) During Awake Brain Surgery: The Clinical Potential of Intra-Operative Functional and
Vascular Brain Mapping. *Front Neurosci* 2020; **13**. doi:10.3389/fnins.2019.01384.