IDSL.UFA assigns high confidence molecular formula annotations for untargeted LC/HRMS datasets in metabolomics and exposomics

Sadjad Fakouri Baygi[1], Sanjay K Banerjee[2], Praloy Chakraborty[2], Yashwant Kumar[2], Dinesh Kumar Barupal[1]*

[1] Department of Environmental Medicine and Public Health, Icahn School of Medicine at Mount Sinai, New York, NY, 10029, USA

[2] Non-communicable Diseases Division, Translational Health Science and Technology Institute, Faridabad, Haryana, 121001, India

* Corresponding author: Address: CAM Building, 3rd floor, 17 E 102nd St, New York, NY 10029 Email: dinesh.kumar@mssm.edu, phone: +1-530-979-4354

Abstract

Untargeted LC/HRMS assays in metabolomics and exposomics aim to characterize the small molecule chemical space in a biospecimen. To gain maximum biological insights from these datasets, LC/HRMS peaks should be annotated with chemical and functional information including molecular formula, structure, chemical class and metabolic pathways. Among these, molecular formulas may be assigned to LC/HRMS peaks through matching theoretical and observed isotopic profiles (MS1) of the underlying ionized compound. For this, we have developed the Integrated Data Science Laboratory for Metabolomics and Exposomics – United Formula Annotation (IDSL.UFA) R package. In the untargeted metabolomics validation tests, IDSL.UFA assigned 54.31%-85.51% molecular formula for true positive annotations as the top hit, and 90.58%-100% within the top five hits. Molecular formula annotations were also supported by MS/MS data. We have implemented new strategies to 1) generate formula sources and their theoretical isotopic profiles 2) optimize the formula hits ranking for the individual and the aligned peak lists and 3) scale IDSL.UFA-based workflows for studies with larger sample sizes. Annotating the raw data for a publicly available pregnancy metabolome study using IDSL.UFA highlighted hundreds of new pregnancy related compounds, and also suggested presence of chlorinated perfluorotriether alcohols (Cl-PFTrEAs) in human specimens. IDSL.UFA is useful for human metabolomics and exposomics studies where we need to minimize the loss of biological insights in untargeted LC/HRMS datasets. The IDSL.UFA package is available in the R CRAN repository https://cran.r-project.org/package=IDSL.UFA. Detailed documentation and tutorials are also provided at www.ufa.idsl.me.

Introduction

Untargeted LC/HRMS analyses of human specimens enable studying the metabolome and exposome in an unbiased manner[1, 2].They have delivered many novel biomarkers and mechanisms for diseases and have improved our understanding of basic metabolic pathways[3-5]. These assays are unique in nature since they record all the mass to charge (m/z) ratio signals above the limit of detection of an instrument for ionized compounds in a sample[6]. This makes the collected data a rich source of information with great opportunities to generate novel hypotheses about metabolome and exposome. It is critical for the promises that untargeted assay offers, that the data are utilized in an inclusive way to not miss any discovery opportunities.

A key post-data acquisition step in the untargeted LC/HRMS assays is to annotate the detected peaks with a range of structural and functional information which can enable biological interpretations[3, 7, 8]. This information includes a chemical structure, molecular formula, chemical class and metabolic pathway[1, 9, 10]. These annotations may help in understanding the nature, origin and function of the chemical structure underlying a peak. Among these information, molecular formula can be assigned to a LC/HRMS peak using the observed and theoretical isotopic profiles for a chemical compound.[11] Isotopic profiles are distinguishable mass spectral signature that represent atomic masses and their natural abundances in the molecular formulas of a compound.[12] Despite the known limitations of high-resolution mass spectrometry instruments, observed experimental isotopic profiles for an ionized compound may sufficiently match the theoretical counterpart within instrument errors in many instances[11, 13], allowing to annotate LC/HRMS peaks with molecular formula[14]. Peak annotation by isotopic profile matching should be performed using efficient computational strategies to account for instrumental errors, multi-sample studies, biological plausibility and chemical diversity.[15]

There has been a great deal of efforts to develop computational tools for annotating peaks in a LC/HRMS dataset with MS1 only data. In a MS1 peak list, a series of m/z values representing different isotopes, ESI adducts, and in-source fragments can belong to one compound. Grouping these m/z values are normally performed by retention time and elution profile similarities within a single file, for example by *xcms-CAMERA*[16, 17], and peak intensity correlations across multiple samples such as MS-FLO[18] and CliqueMS[9] tools. Clustered isotopologues from these tools can be used by the Rdisop R package[19] to assign molecular formulas in a 'database independent' manner. But this approach may miss expected compounds for a sample due to MS instrument's sensitivity and specificity. MetDNA[7] can search for theoretical isotope profiles for a list of molecular formulas from a metabolic reaction network database in

the MS1 peak list. However, their 'database dependent' approach is prone to miss 1) exposure
compounds that are poorly represented in such biochemical databases and 2) compounds which may not
have any transformation products because of their bioaccumulative nature and 3) compounds that were
filtered out by the detection frequency and intensity thresholds while generating MS1 peak table for a
study. Moreover, MetDNA[7] and other tools including SIRIUS[20] and ZODIAC[21], NetID[22] are mainly
designed for assigning molecular formulas to peaks having MS/MS fragmentation data. Furthermore,
implementing these tools for larger studies where only MS1 data are available for every sample remains
to be challenging due to the ranking of formula hits on individual and aligned peak tables, scalable
computation and various sources for formulas which need to be covered for exposomics projects.

There is a need to develop new tools to compute and to compare theoretical and experimental isotopic
profiles for chemical lists from larger databases and chemical spaces for molecular formula annotations.
Here, we have developed a scalable, user-friendly, thoroughly tested R package, the IDSL.UFA to assign
molecular formulas with high confidence to peaks in untargeted LC/HRMS datasets from large-scale
studies. IDSL.UFA covers major possible situations in which a molecular formula can be assigned to
LC/HRMS peaks. We propose that processing LC/HRMS data with IDSL.UFA can find new opportunities
for hypotheses and biomarker discoveries for studying the role of metabolism and exposome in human
diseases.

Methods
Publicly available LC/HRMS test datasets: To test and develop the IDSL.UFA R package, we have
utilized the raw LC/HRMS data for human and mouse biospecimen studies (MTBLS1684[23],
MTBLS2542[24], ST001683[25], ST001430[26], ST001154[27], ST002044 and reference authentic standards
(MSV000088661) available from Metabolomics WorkBench (https://www.metabolomicsworkbench.org/),
MassIVE (https://www.massive.ucsd.edu), and MetaboLights (https://www.ebi.ac.uk/metabolights)
repositories. Data processing results that we have generated for these studies have been submitted to
the Zenodo.org repository and corresponding entry pages are provided in Table S.1. Sample preparation
and data collection procedures are available at entry pages for these studies in the repositories.

Data analysis setup: IDSL.UFA R package is available in the R-CRAN repository (https://cran.r-
project.org/package=IDSL.UFA). The package was installed using the '*install.packages("IDSL.UFA")*' R
command. The IDSL.MXP package (https://cran.r-project.org/package=IDSL.MXP) was used to read
mzML/mzXML/netCDF mass spectrometry data in the centroid mode. mzML files were generated from
the vendor specific data format using the ProteoWizard MSConvert utility[28] when needed. All data files
related to only one type of analysis such as "reverse phase - electrospray ionization negative mode" were
stored in a single file folder. Figure 1 (simplified) and Figure S.1 (detailed) show the workflow steps to
assign molecular formulas for a study. Data processing parameters for IDSL.UFA were provided in a
Microsoft excel file (https://zenodo.org/record/6466688) which was created for individual test studies. We
have provided the parameters files used in this manuscript in the Zenodo.org repository at
(https://zenodo.org/record/6466684). To run the IDSL.UFA workflow, only a single R command
'*UFA_workflow*(*spreadsheet* = *"address of the parameter xlsx file"*)' was needed. Tutorials to create the
parameter files for different scenarios are available at (https://ufa.idsl.me). For each individual peak list in
a study, a formula annotation list with rank, score and other peak properties was generated and exported
to a csv file. Likewise, for each peak in the aligned peak table, top 5-20 formulas with detection frequency
and median ranks across all samples were exported to a csv file for each test study.

Generating the isotopic profile database (IPDB): An IPDB is a digital collection of theoretical isotopic
profiles computed by the IDSL.UFA R package for a list of candidate molecular formulas. IDSL.UFA
queries and matches the experimental isotopic profile against this collection to annotate a LC/HRMS
peak. To compute the isotopic profile for a molecular formula, we have utilized the reference stable
isotope masses and abundances for elements in the periodic table from the PubChem database entries[29]
which have been sourced from International Union of Pure and Applied Chemistry (IUPAC)[30]. We have
also provided an online tool (https://.ipc.idsl.me) to compute an isotopic profile for a single molecular

formula. IDSL.UFA generates centroid isotopic profiles using a dynamic intensity threshold and a peak-
spacing criterion to merge adjacent isotopologues within a mass accuracy window. In this work, we have
covered two sources of molecular formulas.

Source A (databases): Chemical compound lists for four key databases in metabolomics and
exposomics including the blood exposome (chemicals expected in a mammalian blood specimen),
RefMet (measured and expected small molecules in biological organisms)[31], Lipid Maps (known lipid
molecules)[32] and the US-Food and Drug Administration substance registry[33] were obtained from their
online web addresses. These four databases were combined into a single compound list referenced as
IDSL.ExposomeDB in this manuscript and also provided at the Zenodo repository
(https://zenodo.org/record/5823455). Charged compounds, isotope-labeled compounds and multi-
components were excluded. Unique molecular formulas from this consolidated database were used for
computing IPDB. IPDBs for these four databases and the environmental protection agency (EPA)
CompTox Chemicals Dashboard[34] are available at the Zenodo repository
(https://zenodo.org/record/5823455).

Source B (enumerated chemical space with constraints): Molecular formulas were enumerated using
a set of combinatorial and filtering rules using C, H, As, B, Br, Cl, F, I, K, N, Na, O, P, S, Se, and Si
elements. These 16 elements were able to cover 93.76% of carbon-containing compounds (50 ≤ mass ≤
2000) in the IDSL.ExposomeDB combined with EPA chemistry Dashboard[34]. An enumerated chemical
space (ECS) can be represented using equation (1).

$C_cH_hAs_{as}B_bBr_{br}Cl_{cl}F_fI_iK_kN_nNa_{na}O_oP_pS_sSe_{se}Si_{si}$	(1)

where the subscripts of elements represent the number of atoms. A fully combinatorial chemical space
from above-mentioned 16 elements is impractical to be managed by current computational resources.
Therefore, we derived and coded in R a set of four rules which were inspired from the seven golden rules
approach[35] to constrain ECSs. These rules included **1) C/N chemical space rule** '((c/2-n-1) ≤
(h+cl+br+f+i) ≤ (2c+3n+6))' was used to set elemental boundaries for the organic compounds to ensure
entire moieties are bond to carbon and nitrogen atoms. **2) Extended SENIOR rule** was used to ensure
that the molecular formulas completely filled s- and p- valence electron shells.[35] **3) Maximum number of**
halogens thresholds was used to constrain halogenated compounds. For example, we have used the
maximum number of (br+cl) ≤ 8 and the maximum number of ((br+cl+f+i) ≤ 31) thresholds to cover
halogenated compounds in the blood exposome database. **4) Maximum number of elements rule** was
used to skip unrealistically complex molecular formulas generated through molecular formula
enumeration. For example, the maximum number of elements for glucose ($C_6H_{12}O_6$) is three (C, H, and
O). The ECS boundaries and rules for the MTBLS1684 study are provided in the Zenodo repository
(https://zenodo.org/record/5838603).

MS1 peak detection and alignment: IDSL.IPA[36] R package (https://cran.r-
project.org/package=IDSL.IPA) was used to generate individual peak lists for each sample and the
aligned peak table (m/z-RT pairs across all samples) for each study. Data processing parameter files and
IDSL.IPA results for each test study are provided in the Zenodo repository (see Table S.1). Details and a
tutorial for IDSL.IPA data processing can be found at (https://ipa.idsl.me) site.

Isotopic profile matching for individual sample: First, IDSL.UFA software accessed the peak
boundaries, ^{12}C m/z, ^{13}C m/z and ratio of cumulated intensity of ^{12}C to ^{13}C (R^{13}C) for each peak in an
IDSL.IPA generated peak list for a sample. Next, it finds all the theoretical isotopic profiles in an IPDB that
matches the ^{12}C and ^{13}C m/z for a peak. Then, for each matched theoretical profile, experimental profiles
are retrieved from raw data using a mass accuracy threshold within the peak boundaries for a peak. If a
compound formula has three isotopologues in the IPDB and only two were observed in the raw data, the
formula will not be annotated. IDSL.UFA requires that a minimum one MS1 scan across the peak should
have the full isotope profile for a formula in the IPDB.

For the experimental isotopic profiles, the IDSL.UFA software calculates cumulated intensities and
intensity-weighted average masses for each isotopologue using equations (2) and (3) across the
chromatographic peak to minimize the effect of fluctuations such as peak saturation.

$$\overline{Int} = \sum_{t=t_0}^{t=t_{end}} Int_t \tag{2}$$

$$\overline{m/z} = \frac{\sum_{t=t_0}^{t=t_{end}} m/z_t * Int_t}{\overline{Int}} \tag{3}$$

160   where m/z_t and Int_t represent mass and intensity of the matched isotopologue in individual scans across
161   the chromatographic peak from t_0 to t_{end}.

162   We have used the Profile cosine similarity ($\overline{PCS}$) to quantify profile similarity between experimental and
163   theoretical isotopic profiles using equation (4). To assess mass accuracy error for whole isotopic profile,
164   Normalized Euclidean mass error ($\overline{NEME}$) was calculated using the equation (5).[11]

$$\overline{PCS} = \sum_{i=1}^{S} \frac{I_i^{theor} I_i^{exptl}}{\sqrt{\sum_{i=1}^{S}(I_i^{theor})^2} \sqrt{\sum_{i=1}^{S}(I_i^{exptl})^2}} \tag{4}$$

$$\overline{NEME} = \sqrt{\frac{\sum_{i=1}^{S}(M_i^{theor} - M_i^{exptl})^2}{S}} \tag{5}$$

where I_i, M_i, and S represent the intensity of the isotopologue, mass of the isotopologues, and number of
isotopologues in the isotopic profile, respectively. Superscripts of *theor* and *exptl* also represent
theoretical and experimental isotopic profiles, respectively.

Candidate formulas were then filtered using thresholds for 1) $\overline{PCS}$ 2) $\overline{NEME}$ 3) the top 80% of number of
scans with the confirmed whole isotopic profile (NDCS) and 4) minimum percentage of NDCS within a
chromatography peak (RCS (%)). These linear cutoffs allow eliminating false positives; however, they can
reject true positive peaks with poor isotopic profiles.

Next, a matching score for each candidate filtered formula was computed using equation (6).

$$Score = \left(\frac{S^{coeff[1]} * \left(\frac{\overline{PCS}}{100}\right)^{coeff[2]} * \left(\frac{RCS}{100}\right)^{coeff[3]}}{\left(\frac{\overline{NEME}}{maxNEME}\right)^{coeff[4]} * \left(\exp\left(\left|\ln\left(\frac{\overline{R^{13}C_{PL}}}{R^{13}C_{IP}}\right)\right|\right)\right)^{coeff[5]}} \right) \tag{6}$$

where $\overline{R^{13}C_{PL}}$ and $R^{13}C_{IP}$ indicate experimental and theoretical $R^{13}C$ values, respectively. $R^{13}C$ values
represent the ratio of the general ^{13}C isotopologue [M+1] relative to ^{12}C isotopologue [M] on the most
abundant mass. *coeff[1-5]* are powers of the parameters to apply different magnitudes of each variable in
different studies. Using this score, a ranking for candidate formula was determined. By default, IDSL.UFA
utilized a value of 1 for *coeff[1-5]* to rank candidate molecular formulas in the equation (6). However, we
have provided a score coefficient optimization strategy in the section S.1 which can be helpful for
improving the ranking when larger size IPDB are utilized.

Summary statistics of molecular formulas annotation in the aligned peak table: It is quite common
to have more than 50 samples in metabolomics and exposomics projects, which can be leveraged to
compute a statistic for formula annotations across all the samples. For each peak (*m/z*-RT pair) in the

aligned peak table, corresponding molecular formula lists across all the samples were retrieved using the
peak indices provided by the IDSL.IPA data processing. We then aggregated these formula lists and
computed two properties 1) the detection frequency and 2) median rank for each formula assigned for a
peak across all the samples (individual peak list). Then we generated a new sort order for each molecular
formula at the aligned peak table level using the following formula: $\frac{\sqrt{frequency}}{median\ rank}$. For each peak in the
aligned peak table, top 5-20 formulas with detection frequency and median ranks across all samples were
exported to a csv file for each test study.

Molecular formula class detection: Many compounds belong to a chemical class with a distinct sub-
structure pattern such as polychlorinated biphenyl (PCBs), polybrominated diphenyl ethers (PBDEs),
polycyclic aromatic hydrocarbons (PAHs), perfluoroalkyl substances (PFAS), lipids and phthalates etc.
The formula annotations generated via the enumerated chemical space (ECS) approach were processed
to detect such classes within a list of formulas. The IDSL.UFA function '*detect_formula_sets*' was used to
detect 1) constant ΔH/ΔC ratios for polymeric (ΔH/ΔC = 2) and cyclic (ΔH/ΔC = 1/2) chain progressions
within polymeric and cyclic classes (Table S.2- S.4) and 2) a constant number of carbons and fixed
summation of hydrogens and halogens (Σ(H+Br+Cl+F+I)) representing classes similar to PCBs, PBDEs
(Table S.5).

Correlation analysis for gestational age: The ST001430 study[26] includes weekly blood samples of 30
pregnancies. The study has 781 total samples each processed in positive and negative modes to predict
gestational age. To reduce batch effects, the peak heights were adjusted by raw total ion chromatograms
(TICs) in each sample, and then the positive and negative aligned peak height tables were stacked to
generate a comprehensive list of peaks. We computed a Spearman correlation coefficient between
gestational age and peak height data for each pregnancy. A schematic of this workflow is presented in
Figure S.2.

Results and discussion
We have engineered a new software, IDSL.UFA, to annotate LC/HRMS peaks with molecular formulas for
an untargeted metabolomics or exposomics study. In this approach, IDSL.UFA computes theoretical
isotopic profiles for molecular formulas, matches theoretical isotopic profiles against experimental
LC/HRMS data in individual data file using a set of matching parameters and then summarizes the
formula annotations using detection frequency and median ranks in multiple samples (aligned annotated
peak table) in a study. The IDSL.UFA software has been implemented as an R package and made
publicly available via the R-CRAN repository and www.ufa.idsl.me site.

Section 1) Development and validation of IDSL.UFA results: To demonstrate the validity of our
approach to assign molecular formulas, we have utilized datasets with true positive annotations and show
their ranks in the IDSL.UFA result matrices.

Analysis of authentic reference standards: First, we evaluated performance of the IDSL.UFA software
to detect molecular formulas in LC/HRMS data for authentic reference standards. We found that the
average $\overline{NEME}$ (indicator of mass difference) was 0.70 mDa and $\overline{PCS}$ (indicator of isotope profile
similarity) were 99.968% between experimental and theoretical isotopic profiles for 367 authentic
standard compounds of common metabolites. This indicated that the observed isotopic profiles were very
similar to the theoretical counterparts for these reference standards and suggested that molecular
formulas can be reliably assigned to untargeted data generated by the commonly used LC/HRMS
instruments. The theoretical and experimental integrated isotopic profile spectra across chromatography
for these standards are provided at Zenodo repository accession (https://zenodo.org/record/5803968) and
an example compound (Kynurenine ion [$C_{10}H_{13}N_2O_3$]$^+$) is shown in Figure 2 and Figure S.3 ($\overline{NEME}$ ≤ 0.61
mDa and $\overline{PCS}$ = 100.000%).

Analysis of untargeted LC/HRMS data with structurally annotated peaks:

We selected four publicly available studies (ST001154[27], ST001683[25], MTBLS1684[23], and
MTBLS2542[24]). These studies have reported annotations with MSI 1-3 confidence levels
(https://zenodo.org/record/5838709) that were obtained using retention time, accurate mass and MS/MS
spectra matching. For these studies, the IDSL.UFA software assigned 61.85%, 54.31%, 70.58% and
85.51% molecular formula as the top hit, and 96.90%, 90.58%, 100% and 99.29% molecular formulas in
the top five hits in the aligned table. These results were generated using the IPDB of the
IDSL.ExposomeDB with 209,592 and 129,122 ion formulas in positive and negative modes from multiple
ionization pathways, respectively representing 83,951 unique intact molecular formulas
(http://zenodo.org/deposit/5838709).

For each selected study, an ECS IPDB was generated using the element boundaries that covered the
formula list of true positive annotations for the study. When IDSL.UFA software was used for each study
using those specific ECS IPDBs, the assignment rates were – 52.74%, 53.36%, 79.41% and 51.08%
molecular formula as the top hit, and 95.60%, 84.45%, 100% and 91.66% molecular formula in the top 5
hits in the aligned table (Figure S.4). Generally, the IDSL.UFA software annotated 924 (90.14%) and 877
(85.56%) molecular formulas across all four studies using IDSL.ExposomeDB and ECS IPDBs,
respectively. These findings demonstrate that IDSL.UFA is a sensitive approach to cover the majority of
formulas for chemicals detectable in a biospecimen.

There is a tradeoff of coverage and the confidence in annotation while choosing chemical space for
molecular formula annotation. We have noticed that the rank of true positive hits degrades when we have
used a larger chemical space (Figure S.5). However, when compounds that are known and expected to
be found in a blood specimen are used, we have observed that formulas for true positives are often
ranked top hits. Therefore, we recommend a chemical prioritization strategy by sample type and to first
match the compounds that are expected for that sample type and then expand the chemical space to
cover additional peaks.

Summary of the formula annotations in the aligned peak table: Our raw data processing generates
both a separate list of m/z-RT pairs for each sample (individual peak list) and a single combine list
(aligned-table) of m/z-RT pairs for all samples. IDSL.UFA annotates molecular formulas only to individual
peak lists, then, it computes the detection frequency and median rank for all formulas annotated for the
same peak across all samples using the aligned peak table (See methods). Our hypothesis is that the
most probable formula of the underlying ionized compound will have a higher detection frequency and
median rank across all the samples. For example, for the MTLS1684 study, 24/35 (69%) of the reported
annotations had a median rank of 1 and 8/35 (23%) had a median rank of 2 across all 499 samples
(https://zenodo.org/record/5838709). We propose that the summary of detection frequencies and ranks
across individual data files can be helpful in boosting the confidence for formula assignments in multi-
sample studies. It should be noted that IDSL.UFA does not group related peaks to flag them as potential
ESI adducts or in-source fragments. Such grouping of peaks can be achieved by existing solutions such
as MS-FLO[18] online tool or CliqueMS[9] R package.

Additional validation of molecular formula assignment by MS/MS: To further ensure that IDSL.UFA
can assign high confidence molecular formulas for untargeted LC/HRMS data, we utilized data from
ST002044 study which has high quality MS/MS data collected in the data dependent mode. A total 73 hits
were confirmed by matching their spectra to the NIST 2020 MS/MS library (https://chemdata.nist.gov) and
public mass spectral libraries (https://zenodo.org/record/6416108). IDSL.UFA assigned 78.75% of these
hits within a median rank of ≤ 2 in the aligned peak table generated using the IDSL.ExposomeDB formula
IPDB (Table S.6 and Figure S.6). These results provided additional supports to confidence in the
molecular formula assignment by the IDSL.UFA software using the IDSL.ExposomDB IPDB.

Rank score optimization: IDSL.UFA utilized a number of chromatographic-mass spectrometry
parameters to compute the rank of a molecular formula for a peak in the individual peak list. By default, a
score coefficient of 1 is used which works sufficiently in most situations. However, the rank can be further
improved by an optimization strategy that utilizes the true positive, curated and high-quality structure

annotations for each data file as input. This can be achieved by running a mixture of reference standards
using the same analytical method or by annotating peaks using MS/MS, RT and isotopic profile matching
using stringent criteria. For metabolite standards (MSV000088661) and blood specimens (ST002044)
studies, we have observed a significant improvement in the ranking of molecular formulas when
optimized score coefficients were utilized in the IDSL.UFA software (Table S.7).

Section 2) Application of IDSL.UFA for a pregnancy study

To demonstrate an application of IDSL.UFA software to characterize the metabolome and exposome for
blood specimens, we have re-processed a publicly available study ST001430[26] (n=781) which has weekly
blood samples analyzed for 30 pregnancies to accurately predict gestational age (GA in weeks). Raw
data were processed using the IDSL.IPA software to generate the individual peak lists and the aligned
peak table (https://zenodo.org/record/5804527). On average, (3,416 ESI$^-$ and 6,978 ESI$^+$) peaks were
detected across individual peak lists for this study and a total of (89,174 ESI$^-$ and 143,712 ESI$^+$) peaks
were reported in the aligned peak table. The IDSL.UFA software using the IDSL.ExposomeDB IPDB
annotated (80,957 ESI$^-$ and 124,647 ESI$^+$) peaks in the aligned peak table with at least one molecular
formula having a median rank of ≤ 5.

We identify the peaks that were associated with GA by computing a spearman correlation coefficient
between normalized peak-height for each peak and GA. On a spearman cutoff of (p-value ≤ 0.05, $|\rho|$ ≥
0.65, "two.sided" alternative), 274 peaks with a detection frequency of ≥ 5 within each subject were found
to be significantly associated with GA (only ≤ 36 weeks). We observed 242 (red) and 32 (blue) ascending
and descending correlations patterns with GA, which were consistent with the patterns reported in the
original paper[26] and corresponded to chemicals related to steroid hormone biosynthesis and long-chain
fatty acids. These results show the potential the IDSL.UFA approach to characterize the pregnancy
related metabolic changes (Figure 3).

To flag the potential peaks related to chemical exposures in the pregnancy study (ST001430), we first
assigned a molecular formula using an ECS that may cover diverse halogenated compounds that were
not found in the IDSL.ExposomDB formula list. IDSL.UFA resulted with 199,837 unique molecular
formulas on the aligned table (top rank ≤ 30 and number of hits ≤ 30) in the ST001430 study. Grouping
these formulas by a class detection approach (see method) highlighted that 7,615, 18,452, and 32,107
distinct formula classes. For instance, a class of heavily halogenated compounds, $C_nHClF_{2n}O_4$ (n=10-12),
known as chlorinated perfluorotriether alcohols (Cl-PFTrEAs) was detected for human specimens in this
study. Cl-PFTrEAs was previously only reported in air samples from eastern China[37] and may represent a
new ubiquitous global contaminant class. IDSL.UFA can only confirm isotopic profiles match (Figure S.6);
however, a confirmatory in-source fragment ([M-C_3F_6O]$^-$) was consistent with the published MS/MS
fragmentation (Figure S.8).[37] Authentic standards for Cl-PFTrEAs are not readily available; therefore a
confidence level 3b (isotopic profile match combined with fragmentation-based candidate) is suggested
for these annotations according to a recently proposed PFAS identification confidence level by
Charbonnet et al.[38] Levels of Cl-PFTrEAs were similar to the commonly known legacy halogenated
compounds[14] for human serum samples (Figure 4). These findings also show that IDSL.UFA software can
potentially detect chemicals of public health concerns in a human biospecimen and can be helpful in
expanding the existing database of exposome chemicals.[39]

Section 3) Performance benchmarking and comparison with existing tools

IDSL.UFA processed one file (D115_NEG.mzml from the ST2044 study) in ~10 minutes on a computer
with 6 cores, indicating the pipeline can be used in normally available computing resources.

To check how IDSL.UFA performed for low abundant signals, we utilized data from the MTBLS1040 study
which has a seven-point calibration curve for the analyzed compounds. For the hippuric acid standard in
the MTBLS1040 study, IDSL.UFA correctly assigned the molecular formula to the corresponding peak in

samples analyzed at up to 8 fmol concentration level (second-lowest point)
(https://zenodo.org/record/6466668).

IDSL.UFA software is designed to cover commonly used LC-HRMS instruments for human biospecimens
studies in the EBI MetaboLights and Metabolomics Workbench repositories. A mass resolution of 20,000
and mass accuracy of 5 ppm is often found for these instruments. We compared the results for publicly
available two raw data files for a BioRec human plasma sample analyzed for a lipidomic assay by
QToF(ST001843) and Orbitrap instruments(ST001264) using the same chromatography method in the
same lab. Our workflow generated 1752 peaks with 2855 formulas for the QToF data file and 1328 peaks
with 2209 formulas for the Orbitrap data file. A list of 151 true positive annotations from the ST110054[27]
study (MS/MS matches were inspected by an expert user from the same lab and chromatography
method) was utilized for these test data files (https://zenodo.org/record/6621138). For these true
positives, 35% were found to be top hits in the QToF data file and 53% in the Orbitrap data file. It seems
our approach works slightly better for Orbitrap data. However, an even higher resolution and better mass
accuracy can be helpful in removing several false positive annotations, and in improving the ranking of
the true positive annotations.

When we imported a MS1 only data file in the SIRIUS[20] tool, it did not process the file, which was
expected since SIRIUS only processes data files with MS/MS spectra. For a data file (D115_NEG.mzml
from the ST002044 study) with MS/MS spectra in the Data Dependent Acquisition (DDA) mode, SIRIUS
processed 885 MS/MS spectra and suggested formula annotations for 221 spectra, whereas IDSL.UFA
assigned molecular formula to 9303 peaks in this data file.

IDSL.UFA natively uses IUPAC isotope table data[29] to calculate theoretical isotopic profiles and
calculated almost identical isotopic profiles to that obtained from the *envi*Pat package[40] (Table S.8).
Negligible mass and profile similarity differences (NEME ≤ 0.69 mDa and PCS ≥ 99.999%) were
observed for formula $[C_8F_{17}O_3S]^-$ between IDSL.UFA and *envi*Pat[40].

Next, we compared the IDSL.UFA against Rdisop[19] R package to show the advantages of a database-
dependent approach (IDSL.UFA) over a database independent approach (Rdisop) for molecular formula
annotation. For kynurenine authentic standard (MSV000088661), both IDSL.UFA and Rdisop[19] ranked
the M+H adduct formula as the top hit(Section S.2 and Table S.9). But Rdisop's ranking for PFOS
isomers were >20 in the studies ST001430 and ST002044 (both human blood samples). Whereas
IDSL.UFA annotated both isomers of PFOS as top hit for these studies (Table S.10-11 and Figure S.9-
10). This suggests that Rdisop may miss important expected compounds when a complex chemical
space (CHBrClFNOPS) is targeted, but IDSL.UFA will be able to annotate them for human blood
specimens. Next, we extended the comparison to the lipidomics analysis with 151 true positive
annotations. Rdisop annotated 12%, whereas IDSL.UFA reported 53% of true annotations as top hits for
the Orbitrap data file (Figure S.11). These comparisons suggest that a database dependent approach for
formula annotation, such as IDSL.UFA should be used first to screen for expected compounds in HRMS
data before looking for unknown-unknowns. We also provide a comparison (Table S.12) between
IDSL.UFA and Rdisop[19] R packages, highlighting new features that IDSL.UFA is introducing into R
computing workflows for metabolomics and exposomics studies.

Our approach to obtain homologous series with polymeric chain increment from a list of input molecular
formulas is different from the prior approaches[41-43-40] in which molecular formulas are enumerated only for
a known series or chain increment rule. Therefore, our approach has the flexibility to discover new types
of homologous series among a collection of formulas.

Conclusion
IDSL.UFA enabled a comprehensive characterization of the chemical space that was detected by an
untargeted LC/HRMS assay to study the metabolome and exposome and its role in human health. The
unique feature of the IDSL.UFA software is to utilize the summary statistics for the rank and frequency of
detected molecular formulas in the aligned annotated molecular formula table. It can complement the

other peak annotation efforts that use mainly MS/MS data to annotate peaks, and thus lower the number of false negative reporting of peaks and minimize the under-utilization of the untargeted LC/HRMS datasets. We provided various scenarios to obtain molecular formulas from a known database and enumeration strategies to assign a formula to peaks in a LC/HRMS dataset. These new computational strategies for molecular formula assignment can greatly expand the quality of untargeted LC/HRMS data matrices and their analyses especially when MS/MS data are not available.

Funding: The research is in part supported by NIH grants U2CES026561, R01ES032831, U2CES026555 P30ES023515, U2CES030859 and UL1TR001433.
Author's contribution: SFB and DKB planned the study, prepared the results and drafted the manuscript. SFB and DKB coded the IDSL.UFA package. YK, SB and PC provided test LC/HRMS data for authentic standards of metabolites and human biospecimens. All authors have reviewed the manuscript content.

References

References
(1) Needham, B. D.; Adame, M. D.; Serena, G.; Rose, D. R.; Preston, G. M.; Conrad, M. C.; Campbell, A.
S.; Donabedian, D. H.; Fasano, A.; Ashwood, P.; et al. Plasma and Fecal Metabolite Profiles in Autism
Spectrum Disorder. *Biol Psychiatry* **2021**, *89* (5), 451-462. DOI: 10.1016/j.biopsych.2020.09.025
(2) Gonzalez-Dominguez, R.; Jauregui, O.; Queipo-Ortuno, M. I.; Andres-Lacueva, C. Characterization of
the Human Exposome by a Comprehensive and Quantitative Large-Scale Multianalyte Metabolomics
Platform. *Anal Chem* **2020**, *92* (20), 13767-13775. DOI: 10.1021/acs.analchem.0c02008
(3) Shen, B.; Yi, X.; Sun, Y.; Bi, X.; Du, J.; Zhang, C.; Quan, S.; Zhang, F.; Sun, R.; Qian, L.; et al.
Proteomic and Metabolomic Characterization of COVID-19 Patient Sera. *Cell* **2020**, *182* (1), 59-72 e15.
DOI: 10.1016/j.cell.2020.05.032
(4) Wozniak, J. M.; Mills, R. H.; Olson, J.; Caldera, J. R.; Sepich-Poore, G. D.; Carrillo-Terrazas, M.; Tsai,
C. M.; Vargas, F.; Knight, R.; Dorrestein, P. C.; et al. Mortality Risk Profiling of Staphylococcus aureus
Bacteremia by Multi-omic Serum Analysis Reveals Early Predictive and Pathogenic Signatures. *Cell*
2020, *182* (5), 1311-1327 e1314. DOI: 10.1016/j.cell.2020.07.040
(5) Wang, L. B.; Karpova, A.; Gritsenko, M. A.; Kyle, J. E.; Cao, S.; Li, Y.; Rykunov, D.; Colaprico, A.;
Rothstein, J. H.; Hong, R.; et al. Proteogenomic and metabolomic characterization of human
glioblastoma. *Cancer Cell* **2021**, *39* (4), 509-528 e520. DOI: 10.1016/j.ccell.2021.01.006
(6) Franzosa, E. A.; Sirota-Madi, A.; Avila-Pacheco, J.; Fornelos, N.; Haiser, H. J.; Reinker, S.; Vatanen,
T.; Hall, A. B.; Mallick, H.; McIver, L. J.; et al. Gut microbiome structure and metabolic activity in
inflammatory bowel disease. *Nat Microbiol* **2019**, *4* (2), 293-305. DOI: 10.1038/s41564-018-0306-4
(7) Shen, X.; Wang, R.; Xiong, X.; Yin, Y.; Cai, Y.; Ma, Z.; Liu, N.; Zhu, Z. J. Metabolic reaction network-
based recursive metabolite annotation for untargeted metabolomics. *Nat Commun* **2019**, *10* (1), 1516.
DOI: 10.1038/s41467-019-09550-x
(8) Wang, L.; Xing, X.; Chen, L.; Yang, L.; Su, X.; Rabitz, H.; Lu, W.; Rabinowitz, J. D. Peak Annotation
and Verification Engine for Untargeted LC-MS Metabolomics. *Anal Chem* **2019**, *91* (3), 1838-1846. DOI:
10.1021/acs.analchem.8b03132
(9) Senan, O.; Aguilar-Mogas, A.; Navarro, M.; Capellades, J.; Noon, L.; Burks, D.; Yanes, O.; Guimera,
R.; Sales-Pardo, M. CliqueMS: a computational tool for annotating in-source metabolite ions from LC-MS
untargeted metabolomics data based on a coelution similarity network. *Bioinformatics* **2019**, *35* (20),
4089-4097. DOI: 10.1093/bioinformatics/btz207
(10) Uppal, K.; Walker, D. I.; Jones, D. P. xMSannotator: An R Package for Network-Based Annotation of
High-Resolution Metabolomics Data. *Anal Chem* **2017**, *89* (2), 1063-1067. DOI:
10.1021/acs.analchem.6b01214
(11) Fakouri Baygi, S.; Fernando, S.; Hopke, P. K.; Holsen, T. M.; Crimmins, B. S. Automated Isotopic
Profile Deconvolution for High Resolution Mass Spectrometric Data (APGC-QToF) from Biological
Matrices. *Anal Chem* **2019**, *91* (24), 15509-15517. DOI: 10.1021/acs.analchem.9b03335
(12) Fakouri Baygi, S.; Crimmins, B. S.; Hopke, P. K.; Holsen, T. M. Comprehensive Emerging Chemical
Discovery: Novel Polyfluorinated Compounds in Lake Michigan Trout. *Environ Sci Technol* **2016**, *50* (17),
9460-9468. DOI: 10.1021/acs.est.6b01349
(13) Fakouri Baygi, S.; Fernando, S.; Hopke, P. K.; Holsen, T. M.; Crimmins, B. S. Decadal Differences in
Emerging Halogenated Contaminant Profiles in Great Lakes Top Predator Fish. *Environ Sci Technol*
2020, *54* (22), 14352-14360. DOI: 10.1021/acs.est.0c03825
(14) Fakouri Baygi, S.; Fernando, S.; Hopke, P. K.; Holsen, T. M.; Crimmins, B. S. Nontargeted Discovery
of Novel Contaminants in the Great Lakes Region: A Comparison of Fish Fillets and Fish Consumers.
Environ Sci Technol **2021**, *55* (6), 3765-3774. DOI: 10.1021/acs.est.0c08507
(15) Fakouri Baygi, S.; Hutinet, S.; Cariou, R.; Fernando, S.; Hopke, P. K.; Holsen, T. M.; Crimmins, B. S.
Comparison between Automated and User-Interactive Non-Targeted Screening Tools: Isotopic Profile
Deconvoluted Chromatogram (IPDC) Algorithm and HaloSeeker. *Int J Environ Sci Technol* **2022**, *27*, 1-
12.
(16) Li, Z.; Lu, Y.; Guo, Y.; Cao, H.; Wang, Q.; Shui, W. Comprehensive evaluation of untargeted
metabolomics data processing software in feature detection, quantification and discriminating marker
selection. *Anal Chim Acta* **2018**, *1029*, 50-57. DOI: 10.1016/j.aca.2018.05.001
(17) Tautenhahn, R.; Bottcher, C.; Neumann, S. Highly sensitive feature detection for high resolution
LC/MS. *BMC Bioinformatics* **2008**, *9*, 504. DOI: 10.1186/1471-2105-9-504
(18) DeFelice, B. C.; Mehta, S. S.; Samra, S.; Cajka, T.; Wancewicz, B.; Fahrmann, J. F.; Fiehn, O. Mass
Spectral Feature List Optimizer (MS-FLO): A Tool To Minimize False Positive Peak Reports in Untargeted

Liquid Chromatography-Mass Spectroscopy (LC-MS) Data Processing. *Anal Chem* **2017**, *89* (6), 3250-3255. DOI: 10.1021/acs.analchem.6b04372

(19) Neumann, S.; Pervukhin, A.; Böcker, S. Mass decomposition with the Rdisop package. **2022**, 1.

(20) Duhrkop, K.; Fleischauer, M.; Ludwig, M.; Aksenov, A. A.; Melnik, A. V.; Meusel, M.; Dorrestein, P. C.; Rousu, J.; Bocker, S. SIRIUS 4: a rapid tool for turning tandem mass spectra into metabolite structure information. *Nat Methods* **2019**, *16* (4), 299-302. DOI: 10.1038/s41592-019-0344-8

(21) Ludwig, M.; Nothias, L. F.; Duhrkop, K.; Koester, I.; Fleischauer, M.; Hoffmann, M. A.; Petras, D.; Vargas, F.; Morsy, M.; Aluwihare, L.; et al. Database-independent molecular formula annotation using Gibbs sampling through ZODIAC. *Nat Mach Intell* **2020**, *2* (10), 629-+. DOI: 10.1038/s42256-020-00234-6

(22) Chen, L.; Lu, W.; Wang, L.; Xing, X.; Chen, Z.; Teng, X.; Zeng, X.; Muscarella, A. D.; Shen, Y.; Cowan, A.; et al. Metabolite discovery through global annotation of untargeted metabolomics data. *Nat Methods* **2021**, *18* (11), 1377-1385. DOI: 10.1038/s41592-021-01303-3

(23) Alfano, R.; Chadeau-Hyam, M.; Ghantous, A.; Keski-Rahkonen, P.; Chatzi, L.; Perez, A. E.; Herceg, Z.; Kogevinas, M.; de Kok, T. M.; Nawrot, T. S.; et al. A multi-omic analysis of birthweight in newborn cord blood reveals new underlying mechanisms related to cholesterol metabolism. *Metabolism* **2020**, *110*, 154292. DOI: 10.1016/j.metabol.2020.154292

(24) Wu, P.; Chen, D.; Ding, W.; Wu, P.; Hou, H.; Bai, Y.; Zhou, Y.; Li, K.; Xiang, S.; Liu, P.; et al. The trans-omics landscape of COVID-19. *Nat Commun* **2021**, *12* (1), 4543. DOI: 10.1038/s41467-021-24482-1

(25) Han, S.; Van Treuren, W.; Fischer, C. R.; Merrill, B. D.; DeFelice, B. C.; Sanchez, J. M.; Higginbottom, S. K.; Guthrie, L.; Fall, L. A.; Dodd, D.; et al. A metabolomics pipeline for the mechanistic interrogation of the gut microbiome. *Nature* **2021**, *595* (7867), 415-420. DOI: 10.1038/s41586-021-03707-9

(26) Liang, L.; Rasmussen, M. H.; Piening, B.; Shen, X.; Chen, S.; Rost, H.; Snyder, J. K.; Tibshirani, R.; Skotte, L.; Lee, N. C.; et al. Metabolic Dynamics and Prediction of Gestational Age and Time to Delivery in Pregnant Women. *Cell* **2020**, *181* (7), 1680-1692 e1615. DOI: 10.1016/j.cell.2020.05.002

(27) Barupal, D. K.; Zhang, Y.; Shen, T.; Fan, S.; Roberts, B. S.; Fitzgerald, P.; Wancewicz, B.; Valdiviez, L.; Wohlgemuth, G.; Byram, G.; et al. A Comprehensive Plasma Metabolomics Dataset for a Cohort of Mouse Knockouts within the International Mouse Phenotyping Consortium. *Metabolites* **2019**, *9* (5). DOI: 10.3390/metabo9050101

(28) Chambers, M. C.; Maclean, B.; Burke, R.; Amodei, D.; Ruderman, D. L.; Neumann, S.; Gatto, L.; Fischer, B.; Pratt, B.; Egertson, J.; et al. A cross-platform toolkit for mass spectrometry and proteomics. *Nat Biotechnol* **2012**, *30* (10), 918-920. DOI: 10.1038/nbt.2377

(29) Kim, S.; Gindulyte, A.; Zhang, J.; Thiessen, P. A.; Bolton, E. E. PubChem Periodic Table and Element Pages: Improving Access to Information on Chemical Elements from Authoritative Sources. *Chem Teach Int* **2021**, *3* (1), 57-65. DOI: 10.1515/cti-2020-0006

(30) Currie, L. A. Nomenclature in Evaluation of Analytical Methods Including Detection and Quantification Capabilities (Iupac Recommendations 1995). *Pure Appl Chem* **1995**, *67* (10), 1699-1723. DOI: DOI 10.1351/pac199567101699

(31) Fahy, E.; Subramaniam, S. RefMet: a reference nomenclature for metabolomics. *Nat Methods* **2020**, *17* (12), 1173-1174. DOI: 10.1038/s41592-020-01009-y

(32) Fahy, E.; Sud, M.; Cotter, D.; Subramaniam, S. LIPID MAPS online tools for lipid research. *Nucleic Acids Res* **2007**, *35* (Web Server issue), W606-612. DOI: 10.1093/nar/gkm324

(33) *FDA structured product labelling*. https://www.fda.gov/industry/fda-resources-data-standards/structured-product-labeling-resources (accessed.

(34) Williams, A. J.; Grulke, C. M.; Edwards, J.; McEachran, A. D.; Mansouri, K.; Baker, N. C.; Patlewicz, G.; Shah, I.; Wambaugh, J. F.; Judson, R. S.; et al. The CompTox Chemistry Dashboard: a community data resource for environmental chemistry. *J Cheminform* **2017**, *9* (1), 61. DOI: 10.1186/s13321-017-0247-6

(35) Kind, T.; Fiehn, O. Seven Golden Rules for heuristic filtering of molecular formulas obtained by accurate mass spectrometry. *BMC Bioinformatics* **2007**, *8*, 105. DOI: 10.1186/1471-2105-8-105

(36) Fakouri Baygi, S.; Kumar, Y.; Barupal, D. K. IDSL.IPA Characterizes the Organic Chemical Space in Untargeted LC/HRMS Data Sets. *J Proteome Res* **2022**, *21* (6), 1485-1494. DOI: 10.1021/acs.jproteome.2c00120

(37) Yu, N.; Wen, H.; Wang, X.; Yamazaki, E.; Taniyasu, S.; Yamashita, N.; Yu, H.; Wei, S. Nontarget Discovery of Per- and Polyfluoroalkyl Substances in Atmospheric Particulate Matter and Gaseous Phase

Using Cryogenic Air Sampler. *Environ Sci Technol* **2020**, *54* (6), 3103-3113. DOI:
10.1021/acs.est.9b05457
(38) Charbonnet, J. A.; McDonough, C. A.; Xiao, F.; Schwichtenberg, T.; Cao, D.; Kaserzon, S.; Thomas,
K. V.; Dewapriya, P.; Place, B. J.; Schymanski, E. L.; et al. Communicating Confidence of Per- and
Polyfluoroalkyl Substance Identification via High-Resolution Mass Spectrometry. *Environmental Science*
& Technology Letters **2022**. DOI: 10.1021/acs.estlett.2c00206
(39) Barupal, D. K.; Fiehn, O. Generating the Blood Exposome Database Using a Comprehensive Text
Mining and Database Fusion Approach. *Environ Health Perspect* **2019**, *127* (9), 97008. DOI:
10.1289/EHP4713
(40) Loos, M.; Gerber, C.; Corona, F.; Hollender, J.; Singer, H. Accelerated isotope fine structure
calculation using pruned transition trees. *Anal Chem* **2015**, *87* (11), 5738-5744. DOI:
10.1021/acs.analchem.5b00941
(41) Schum, S. K.; Brown, L. E.; Mazzoleni, L. R. MFAssignR: Molecular formula assignment software for
ultrahigh resolution mass spectrometry analysis of environmental complex mixtures. *Environ Res* **2020**,
191, 110114. DOI: 10.1016/j.envres.2020.110114
(42) Hughey, C. A.; Hendrickson, C. L.; Rodgers, R. P.; Marshall, A. G.; Qian, K. Kendrick mass defect
spectrum: a compact visual analysis for ultrahigh-resolution broadband mass spectra. *Anal Chem* **2001**,
73 (19), 4676-4681. DOI: 10.1021/ac010560w
(43) Jacob, P.; Barzen-Hanson, K. A.; Helbling, D. E. Target and Nontarget Analysis of Per- and
Polyfluoralkyl Substances in Wastewater from Electronics Fabrication Facilities. *Environ Sci Technol*
2021, *55* (4), 2346-2356. DOI: 10.1021/acs.est.0c06690

Figure 1. A simplified flowchart of the IDSL.UFA software.

Figure 2. a) A chromatographic peak generated by the IDSL.IPA pipeline for Kynurenine ion ($[C_{10}H_{13}N_2O_3]^+$ = $[M+H]^+$) to detect peak boundaries. b) Comparison between the theoretical isotopic profile and integrated spectra across the chromatographic peak after molecular formula annotation using IDSL.UFA.

Figure 3. Trends of 274 peaks associated with pregnancy dynamics. Molecular formula annotated interactive plots are available at https://ufa.idsl.me/st001430 for an enumerated chemical space and IDSL.ExposomeDB IPDBs.

Figure 4. Peak area of halogenated contaminants in human blood ($[C_nF_{2n+1}O_3S]^-$ (n = 4, 6, 8), $[C_8F_{15}O_2]^-$, $[C_6Cl_5O]^-$, $[C_9H_{14}Cl_6O_4P]^-$, $[C_{18}H_{14}Cl_3O_8]^-$, $[C_{12}H_6Cl_3O_2]^-$) and Cl-PFTrEAs $[C_nClF_{2n}O_4]^-$ (n = 10-12) across 781 negative samples in the ST001430 study.