1 **Comparative assessment of quantification methods for tumor tissue phosphoproteomics**
2
3

Yang Zhang[1-3], Benjamin Dreyer[4], Natalia Govorukhina[1], Alexander M. Heberle[2,3], Saša Končarević[5],

Christoph Krisp[4], Christiane A. Opitz[6], Pauline Pfänder[6,7], Rainer Bischoff[1], Hartmut Schlüter[4], Marcel

Kwiatkowski[2,8,9,*], Kathrin Thedieck[2,3,10,*], Peter L. Horvatovich[1,*]

[1]Department of Analytical Biochemistry, Groningen Research Institute of Pharmacy, University of

Groningen, 9713 AV Groningen, The Netherlands

[2]Institute of Biochemistry and Center for Molecular Biosciences Innsbruck, University of Innsbruck, 6020

Innsbruck, Austria

[3]Laboratory of Pediatrics, Section Systems Medicine of Metabolism and Signaling, University of

Groningen, University Medical Center Groningen, 9713 AV, Groningen, The Netherlands

[4]Section/Core Facility Mass Spectrometry and Proteomics, Institute of Clinical Chemistry and Laboratory

Medicine, University Medical Center Hamburg-Eppendorf, Martinistraße 52, 20246 Hamburg, Germany

[5]Proteome Sciences R&D GmbH & Co. KG, Altenhöferallee 3, 60438 Frankfurt/Main, Germany

[6]Brain Cancer Metabolism Group, German Cancer Research Center (DKFZ), Heidelberg, Germany

[7]Faculty of Bioscience, Heidelberg University, Heidelberg, Germany

[8]Department of Molecular Pharmacology, Groningen Research Institute for Pharmacy,
University of Groningen, Groningen 9700 AD, The Netherlands

[9]Groningen Research Institute for Asthma and COPD, University Medical Center Groningen,

University of Groningen, Groningen 9700 AD, The Netherlands.

[10]Department of Neuroscience, School of Medicine and Health Sciences, Carl von Ossietzky University

Oldenburg, 26129 Oldenburg, Germany

*Corresponding authors

Abstract

With increasing sensitivity and accuracy in mass spectrometry, the tumor phosphoproteome is getting into reach. However, the selection of quantitation techniques best-suited to the biomedical question and diagnostic requirements remains a trial and error decision as no study has directly compared their performance for tumor tissue phosphoproteomics. We compared label-free quantification (LFQ), spike-in-SILAC (stable isotope labeling by amino acids in cell culture) and TMT isobaric tandem mass tags technology for quantitative phosphosite profiling in tumor tissue. TMT offered the lowest accuracy and the highest precision and robustness towards different phosphosite abundances and matrices. Spike-in-SILAC offered the best compromise between these features but suffered from a low phosphosite coverage. LFQ offered the lowest precision but the highest number of identifications. Both spike-in-SILAC and LFQ presented susceptibility to matrix effects. Match between run (MBR)-based analysis enhanced the phosphosite coverage across technical replicates in LFQ and spike-in-SILAC but further reduced the precision and robustness of quantification. The choice of quantitative methodology is critical for both study design such as sample size in sample groups and quantified phosphosites, and comparison of published cancer phosphoproteomes. Using ovarian cancer tissue as an example, our study builds a resource for the design and analysis of quantitative phosphoproteomic studies in cancer research and diagnostics.

Introduction

Phosphorylation events mediated by oncogenic kinases are widely recognized as major drivers of tumorigenesis. Despite more than 50 kinase inhibitors approved for cancer treatment by Food and Drug Administration (FDA) and European Medicines Evaluation Agency (EMEA)[1], immediate and acquired drug resistance remains a key medical challenge. Toward personalized treatments, genome and transcriptome analysis based detection of somatic mutations and mRNA changes has led to breakthroughs such as the identification of HER2-positivity that guides therapies with HER-targeting antibodies[2,3] or BCR-ABL fusion indicating a response to Imatinib and its analogues[4]. Although these therapies significantly improve the survival rates e.g. in breast cancer and leukemia, they suffer from significant initial and acquired resistance[5–7]. For most cancers and kinase-directed drug therapies reliable predictive markers of initial drug response and acquired resistance are largely missing. Pathway analysis in cancer cell lines highlights the potential to predict drug sensitivity[8], and calls for translation to patient tumor tissues. Measuring the phosphorylation events and networks directly targeted by kinase inhibitors in the patient's tumor may offer direct access to tumor- and patient-specific alterations governing drug resistance. Comprehensive quantitative coverage of the cancer-

and patient-specific tumor phosphoproteome may identify patient subgroups with common response or escape mechanisms and may open new avenues to precision oncology by revealing a specific kinase signature in individual patients. Antibody-based techniques such as immunohistochemistry (IHC) and reverse-phase protein arrays (RPPA) are used in diagnostics and research, but they are limited by the availability of reliable antibodies and the low quantitative accuracy and robustness of these assays[9,10]. Mass spectrometry (MS)-based phosphoproteomic methods that combine the latest generation of instruments[11–14] with advanced computational tools[15–18] allow us to detect and quantify thousands of phosphosites. Label-free quantification (LFQ), stable isotope labeling by amino acids in cell culture (SILAC)[19] and chemical labeling e.g. with isobaric tandem mass tags (TMT)[20] or isobaric tags for relative and absolute quantitation (iTRAQ)[21] are widely used for quantitative (phospho)proteomics and have been characterized in detail regarding their performance for studies in *in vitro* systems such as yeast or mammalian cell cultures[22–27]. But we do not yet know whether the results translate to clinical phosphoproteomics in tumors. Tumor tissue differs from cell cultures in many aspects that are relevant for sample preparation and MS analysis. For instance, tumor tissue is typically snap-frozen and must be powderized prior to lysis, whereas cultured cells can be directly taken up in lysis buffer. Tumor tissue also contains different cell types as well as extracellular matrix and is thus more heterogeneous than cultured cells[28]. In the last five years, a series of studies have used LFQ, spike-in-SILAC and TMT based quantification separately for tumor proteomics[29–34] or phosphoproteomics[35–37] highlighting the need for a comparative assessment of quantitative phosphoproteomic methodologies in tumor tissue. The classical SILAC approach requires full metabolic labeling of the entire proteome, and has been mainly limited to cell culture analysis with few exceptions[38]. Mann and colleagues introduced spike-in-SILAC[39] whereby a SILAC-labelled reference cell line is used to generate thousands of isotopically labeled peptides as internal standards for tissue proteome quantification. Ovarian cancer is well-accessible to clinical proteomics and presents a high unmet medical need due to limited treatment options and short survival[40]. Using ovarian cancer as an example, we compared the performance of LFQ-, spike-in-SILAC - and TMT-based phosphoproteomics regarding accuracy, precision and robustness towards variation of the matrix and the phosphosite abundance. We use the term robustness in this article as the relative independence of phosphosites quantification from the number of replicates, phosphorylation levels and matrix effects. The SKOV3 cell line, derived from ascites of a human ovarian adenocarcinoma[41] was used to compare our results from tumor tissue to a matched cell line.

Results

Sample preparation and MS measurements.

The snap-frozen ovarian tumor tissue was powderized and taken up in ice-cold lysis buffer. SKOV3 cells were cultured in full medium. After a wash step, ice-cold lysis buffer was added directly to the tissue culture plates, and the cells were scraped off and transferred to tubes. Samples were sonicated and centrifuged, and the protein concentrations were adjusted to 1 mg/mL. All steps were performed on ice. To obtain samples with known ratios of phosphopeptide quantities (**Figure S1A, panel A**), we dephosphorylated tissue and cell lysates containing 20 mg protein with alkaline phosphatase. The dephosphorylation efficiency was assessed by LC-MS/MS. After alkaline phosphatase treatment, the overall intensity of identified phosphopeptides was reduced to 1.6% for SKOV3 cells and 3% for ovarian tumor tissue (**Figure S2A**). We mixed the non-treated lysate with the dephosphorylated lysate at ratios of 1:5 (termed in the following as sample 1X), 1:1 (3X), and 1:0 (6X) (**Figure S1B**). Thus, the sample with the lowest phosphosite quantity was 1X. The phosphosite quantity in the 3X sample was thrice as high, and it was 6 times as high in the 6X sample. This resulted in a known fold change of 2 (FC2) for the 6X versus the 3X sample. The 1X versus 3X samples exhibited a known FC of 3 (FC3), and the 1X versus 6X samples had a known FC of 6 (FC6). The known fold changes served as the ground truth based on which we assessed in the following the performance of spike-in-SILAC, TMT, and LFQ based phosphosite quantification. In order to assess the technical variability, we performed the experiments in six technical replicates, which were prepared separately from the lysate samples (**Figure S1**). For this purpose, each sample was divided into six aliquots of 1 mg for LFQ, 500 µg for SILAC, and 300 µg for TMT.

To generate the spike-in-SILAC standard, the SKOV3 cells were cultured for 5 passages in medium containing heavy lysine and arginine[42]. For the cell lysate to be compared with tissue sample, SKOV3 cells were cultured in light SILAC media. The labeling efficiency was >98% and the arginine-to-proline conversion was 1% or below (**Figure S2B, C**). 500 µg of the heavy SILAC labelled cell lysate was mixed 1:1 with 500 µg of the 1X, 3X, or 6X tissue lysates or of the 1X, 3X, or 6X light SILAC-labelled cell lysates, respectively (**Figure S1 B, C**), resulting in a total amount of 1 mg for each sample. All samples were reduced, alkylated, digested with trypsin for 16 h, and desalted using reversed phase solid phase extraction. TMT 10-plex was used to label 6 technical replicates of the 6X, 3X, and 1X tissue and cell line samples, which we assigned to two batches (**Figure S1C**). Each batch contained in addition a reference sample that was composed of 100 µg of the 6X, the 3X, and the 1X sample. For TMT analysis, two TMT analysis batches were prepared, each consisting of 10 TMT-channels composed of 9 samples and one reference sample **Figure S1C (bottom)**. For each channel, 300 µg of TMT-

derivatized sample was used. This results in a total sample amount of 3 mg per TMT batch.
Thus, we stayed below the maximum binding capacity of the Fe-NTA (nitrilotriacetic acid)-
based IMAC columns (5 mg). Before IMAC enrichment, 50 µg of each sample were set apart
for total protein concentration analysis. The rest of the samples were subjected to
phosphopeptide enrichment by IMAC. All samples were dried and analyzed by LC-MS/MS
using an Orbitrap Fusion instrument. The gradient length was kept the same for all samples (2
h). The LFQ and SILAC samples were analyzed by data dependent acquisition (DDA)[43]
tandem MS (MS2). The TMT samples were analyzed by synchronous precursor selection MS3
(SPS-MS3)[44,45], to minimize reporter ion cross-contamination from peptides co-isolated for
fragmentation in the same isolation window. The data was analysed using the MaxQuant[15]
suite. Shotgun proteomics by DDA is a widely used technology[43]. Although DDA achieves fast
and accurate large-scale proteome analysis, missing values of 50% or higher are inherent
characteristics of the stochastic precursor selection[46,47]. Data pre-processing with match
between runs (MBR) is a common approach to reduce the number of missing values in DDA
and perform identification transfer by matching non-identified single-stage MS (MS1) features
to identified ones[27,48]. We compared data processing without and with the MBR[49,50] option for
LFQ and spike-in-SILAC, including the requantify (req) option that is typically applied in SILAC
experiments to extract the signal of non-detected peaks, which are complementary to
successfully identified light or heavy SILAC singleton peaks[15].
152

Phosphosite and phosphopeptide identifications.

We compared the total measurement time, the number of recorded MS2 spectra and MS3
spectra, the number of identified MS2 spectra and the number of identified phosphosites
between LFQ, SILAC and TMT (**Table 1**). In TMT, the MS3 spectra were used for quantification
and the MS2 spectra were used for identification. Therefore, we compared the number of
acquired and identified MS2 spectra to assess the identification performance. The threshold
for the phosphosite localization probability was set to 0.75[49].

For tumor tissue, SILAC provided the highest number of detected ($n_{spectra}$=613,676) and
identified (n_{id}=240,688) MS2 spectra, followed by LFQ ($n_{spectra}$=471,216, n_{id}=171,752) and TMT
($n_{spectra}$= 63,421 and n_{id}=8,122). However, the number of phosphopeptide identifications was
higher with LFQ (8,932) than with SILAC (6,923). The discrepancy between the number of
MS2 spectra and the number of identified phosphopeptides may be explained by the SILAC
pair-induced increased complexity at the MS1 level. The isolation and fragmentation of the
light and heavy phosphopeptide leads to a redundant identification, which reduces the number
of identified phosphopeptides compared to LFQ[51]. TMT resulted in the lowest number of

identified phosphopeptides (n= 4,281). Similar observations were made for cell lysate (**Table**
1).

In line with the number of identified phosphopeptides, LFQ provided the highest
number of identified phosphosites (n=5,812) followed by SILAC (n=4,828) and TMT (n=3,578)
for tumor tissue. Similar results were obtained for cell lysate (**Table 1**). 1,640 (20%) and 1,205
(18.6%) of the phosphosites were identified by all three methods in the tumor and cell line
samples, respectively (**Figure 1A and B**). SILAC and LFQ showed an overlap of ca. 40%
(40.2% for tumor tissue, 43.2% for cell line). Lower overlap was observed between TMT and
LFQ or SILAC for both tissue (TMT-LFQ: 25.9%, TMT-SILAC: 21%) and cell line (TMT-LFQ:
20.7%, TMT-SILAC: 21.7%). This may be explained by TMT-derivatization due to which the
resulting phosphopeptides differ in their chemical composition and physico-chemical
properties[20] from phosphopeptides analyzed by LFQ and SILAC that are chemically equivalent
and differ only in their isotopic composition. We compared the frequency of b- and y-fragment
ions of all identified phosphosites (**Figure 1C, D**). LFQ and SILAC exhibited similar fragment
ion frequencies in both tissue and cell line samples, with a higher frequency of y-ions (up to
75%) than b-ions (up to 25%), as expected for non-derivatized tryptic peptides[52]. For TMT-
derivatized phosphopeptides, the frequency of y- and b-fragment ions was equal. Similar
results were obtained for the non-phosphorylated peptides in the input sample (**Figure S3A,**
B) which is in agreement with earlier observations by Shen et al[53]. The higher frequency of b-
ions may be explained by the introduction of a tertiary amine at the N-terminus as a result of
TMT labeling. The tertiary amine of the TMT-tag has a higher gas phase basicity than the
primary amine of the N-terminus of a tryptic peptide. As a result, the proton at the piperidine
ring of the TMT-tag linked to the peptide N-terminus is less mobile, and b- and y-fragment ions
at are generated at similar frequencies[54]. The TMT-label impacts both the ionization and the
fragmentation of the derivatized phosphopeptide[55], which may explain the low overlap of
identifications in TMT with LFQ and spike-in-SILAC. Our results illustrate that a derivatization
strategy can shift the identifications to different subpopulations of phosphosites and peptides,
which needs to be considered when comparing datasets acquired without and with chemical
labelling.

We next compared the reproducibility of phosphosite identifications across the six
technical replicates in tumor tissue (**Figure 1E**). The LFQ and SILAC data sets were processed
with or without activation of the MBR feature in MaxQuant. The reproducibility of the
phosphosite identification is reflected by the distribution of the numbers of identified
phosphosites across the replicates. MBR can be considered as identification transfer step and
therefore influences the reproducibly of phosphosite identification. For TMT, MBR has recently
been implemented for MS2 methods[56]. To enhance the quantification accuracy, we opted for

SPS (synchronous precursor selection)–MS3 technology[44,45]. As expected, MBR increased the

number of reproducibly identified phosphosites for both LFQ and SILAC. The number of

phosphosites identified across six replicates decreased with decreasing phosphosite quantity

(i.e., 6X vs. 3X vs. 1X) for LFQ and spike-in-SILAC with and without MBR. This indicates that

with LFQ and spike-in-SILAC reproducible identification increases with phosphosite quantity

and the resulting signal intensity. This was in contrast to TMT for which the number of

phosphosites reproducibly identified across 6 replicates was constant over the entire range of

phosphorylation quantity (n=1593 in 6X, n=1591 in 3X and n=1582 in 1X). This likely comes

from the multiplexing of the 1X, 3X and 6X samples due to which low intensity signals in the

1X sample can be identified based on the higher average intensity from the pooled 1X, 3X and

6X samples[20]. The higher average phosphosite quantity and signal intensity in the pooled TMT

sample also explains the better reproducibility of identifications for samples with low

phosphosite quantity (i.e. 1X and 3X) as compared to LFQ and spike-in-SILAC. For the high

phosphosite quantity (6X) TMT yielded also more reproducibly identified phosphosites than

spike-in-SILAC, but the MBR option increased this number beyond that of TMT. LFQ without

and with MBR yielded higher numbers of reproducible identifications than TMT for high

phosphosite quantities (6X). The reproducible identification of phosphosites in six or three

replicates in TMT (with other replicate numbers largely missing) arises from the fact that the

TMT samples were measured in two batches with three replicates per batch (**Figure S1C**).

Therefore, a given phosphosite was identified in either of the batches (i.e. in 3 replicates), or

in both batches (i.e. in 6 replicates). Comparable results were obtained for the reproducibility

of the identification of phosphopeptides, with higher overall numbers as phosphosites can be

assigned to several peptides (**Figure S4A**).

The overall numbers of reproducibly identified phosphosites and phosphopeptides from

cell lysate were on average lower than from tissue lysate (**Figures 1F, S4B**). Also here, TMT

yielded a similar number of reproducible identifications across the different phosphorylation

quantities. TMT yielded a higher number of reproducible identifications than spike-in-SILAC

without and with MBR and LFQ across all phosphosite quantities. Only LFQ with MBR in a

sample with a high phosphosite quantity (6X) performed better than TMT in terms of

reproducibly identified phosphorylations. In both matrices, tumor tissue und cultured cells, TMT

exhibited better robustness than LFQ and spike-in-SILAC regarding reproducibility of

identifications across the whole range of phosphosite quantities. In tumor tissue with high

phosphorylation quantity, spike-in-SILAC and LFQ outperformed TMT regarding the number

of reproducible identifications, but this advantage was lost for lower phosphosite quantities.

Reproducibility, accuracy and precision of phosphosite quantification.

We assessed the reproducibility of phosphosite quantification by calculating the coefficient of variation (CV) distribution (**Figures 2A, B and S5A-D**). We observed no major difference in CV distribution in datasets analyzed with or without MBR both for LFQ (**Figure S5A, B**) or spike-in-SILAC (**Figure S5C, D**). TMT provided the lowest CVs, followed by spike-in-SILAC+MBRreq and LFQ+MBR (**Figure 2A, B)**. For TMT the CV distribution was similar across phosphosite quantities (1X, 3X, 6X). Thus, TMT exhibits the highest reproducibility for quantification of low-abundant phosphosites. TMT multiplexing reduces the technical variability as multiple samples are pooled and processed together during sample preparation (e.g. phosphopeptide enrichment and desalting by solid-phase extraction) and LC-MS/MS measurements. A reference channel in each TMT-batch further reduced the measurement variability between the two batches. In spike-in-SILAC, the spiked-in heavy-labeled phosphopeptides serve a similar purpose as the TMT reference channel, namely to reduce variability introduced during sample preparation and LC-MS/MS analysis. This likely explains the lower CV values with spike-in-SILAC as compared to LFQ, for which all experimental and measurement steps were conducted independently and no reference sample was spiked-in. TMT shows lower CVs compared to spike-in-SILAC, although mixing of the multiplexed samples in spike-in-SILAC occurs earlier than in TMT. The smaller CVs of TMT are most likely due to the peak quantification in smoothed MS3-orbitrap mass spectra, whereas spike-in-SILAC used MS1-based 2D peak quantification in an LC-MS map. In these, the chromatographic peaks have a higher noise level than the signals in smoothed MS3-orbitrap mass spectra[57]. This may enable a more precise quantification in TMT data as compared to spike-in-SILAC data.

The CV distribution reflects the variability based on the mean of six replicates for each phosphosite. To assess the agreement between datasets in a pairwise manner, we analyzed the linear correlation between two datasets by calculating the pairwise correlation of phosphosite quantities for phosphopeptides identified and quantified in both samples (**Figures 2C-H and S5E-H**). Similar to the CV plots, no major difference was observed in the correlation heatmaps with and without MBR/MBRreq for LFQ and spike-in-SILAC (**Figures 2C-F and S5E-H**). For tissue and cell lysates, both TMT (**Figure2G, H**) and spike-in-SILAC (**Figure 2E, F**) showed Pearson correlation coefficients above 0.9 within a sample group for all phosphosite quantities (1X, 3X, 6X). In TMT, the correlation was always above 0.9 and did not change between different phosphosite quantities, although a batch effect was visible that broadened the correlation coefficient distribution. For spike-in-SILAC, lower coefficients were observed for low phosphosite quantities (1X). This was even more pronounced for LFQ, resulting in a Pearson correlation coefficient below 0.9 for tumor lysate with low abundant phosphosites (1X,

Figure 2C). For all comparisons between sample groups (i.e., 6X vs. 3X, 3X vs. 1X, 6X vs. 1X), TMT yielded Pearson correlation coefficients above 0.9 (**Figure 2G, H**), reflecting the high reproducibility of TMT across different phosphosite quantities. For comparisons between sample groups, spike-in-SILAC and LFQ showed considerably lower Pearson correlation coefficients than TMT. The correlation coefficients were particularly low when samples with low phosphosite quantity (1X) were part of the comparison, reflecting the drop in reproducibility and higher error in the quantification of low abundant peaks for spike-in-SILAC and LFQ as the phosphosite quantity decreased. In summary, TMT exhibited the lowest variability and highest correlation for tissue and cell lysates, especially when the phosphosites were low-abundant.

Accuracy and precision of the quantification of phosphosite fold changes.

We evaluated the precision and accuracy of the determination of the fold-changes between samples with known relative phosphosite quantities (1X, 3X, 6X) (**Figure 3**). We assessed the distribution of detected fold changes relative to the theoretical fold changes (FC2, FC3 and FC6) by violin plots (**Figure 3A, B**) and by mean squared errors (MSE)[58] of the sum of positive deviation or variance (**Figure 3C, D**), indicative of the quantification error in accuracy and precision, respectively. LFQ and spike-in-SILAC determined all fold-changes with high accuracy and yielded more accurate quantifications than TMT in tumor and tissue lysates (**Figure 3A-D**). The use of MBR decreased the accuracy (**Figure 3C, D**). MBR also reduced the precision of all determined fold changes, and this effect was more pronounced in spike-in-SILAC than in LFQ. Matched features with low abundance or false feature matching may be reasons for lower precision and accuracy in MBR[57]. This shows that MBR trades off a smaller number of missing values for accuracy and precision. TMT showed the highest precision and the lowest accuracy, consistently underestimating the expected fold change (**Figure 3A-D**).

We also investigated the performance in detecting significant differences in the abundance of phosphosites between the different fold changes. Phosphosites had to be quantified in at least three out of six replicates. Statistical analysis was performed using a two-sided t-test and Benjamini-Hochberg procedure to correct for multiple testing. Across all fold-changes, TMT determined the highest number of phosphosites as being significantly different for both tissue (**Figure 4A**) and cell (**Figure 4B**) lysates. The next-best results were yielded by LFQ+MBR and spike-in-SILAC+MBR. LFQ and spike-in-SILAC performed worst. As multiplexing in TMT reduces the issue of missing values, almost all identified phosphosites (> 98% for all sample groups) could be used for differential analysis. As stated earlier, also quantification based on smoothed MS3-orbitrap mass spectra yields more precise quantification and lower CVs, likely enhancing the discrimination of small differences. Spike-in-SILAC and LFQ resulted in a considerable proportion of phosphosites identified in less than

three replicates, which prevented them from being used for differential analysis. The use of MBR in LFQ and spike-in-SILAC increased the number of significantly different phosphosites, in particular for the higher fold changes (FC3, FC6), which could be detected despite the lower precision (**Figure 3C, D**).

We calculated true-positive-rates (TPR) and false-positive-rates (FPR) of the differentially quantified phosphosites for each quantification method and visualized them in a receiver operating characteristic (ROC) curve[59] (**Figure 4C, D**). We also calculated the area under the ROC (AUROC), which provide classification performance of TPR and FPR for whole thresholds range (**Table 2**). At an FPR threshold of 0.05, the TPR was above 0.95 for all quantification approaches and the largest fold-change (FC6), indicating that all methods were able to correctly determine a fold-change of six (FC6) between the phosphorylation quantities (**Fig. 4C, Table 2**). These results were also reflected by the AUROC values showing value higher than 0.98 for all methods for both tissue and SKOV3 cell lines (**Table 2**). In general, TMT showed the highest TPRs and AUROCs for all fold-changes (**Table 2**). For the tissue lysate and the smallest fold-change (FC2), TMT (TPR: 0.99) and LFQ (TPR: 0.99) showed a TPR higher than 0.95. In contrast, TMT (TPR: 0.97) and spike-in-SILAC (TPR: 0.95) showed the highest TPRs for cell lysate, whereas LFQ yielded TPRs below 0.95 for FC2 (TPR: 0.59) and FC3 (TPR: 0.92). AUROC values showed similar trends then TPR at FPR level of 0.05 (**Table 2**). MBR or MBRreq generally resulted in lower TPRs and AUROCs, further highlighting that MBR/MBRreq traded the numbers of identified phosphosites (**Figure 1E, F**) and of significantly different phosphosites (**Figure 4A, B**), for accuracy, precision (**Figure 3C, D**), and robustness of the quantification (**Figure 4C, D**).

Taken together, TMT showed the lowest accuracy but the highest precision and the highest number of significantly differential phosphosites for all fold-changes in tumor and cell lysate. TMT also provided the highest TPR regardless of sample matrix. In contrast, the matrix affected the TPRs for LFQ and spike-in-SILAC, with spike-in-SILAC exhibiting a TPR below 0.95 for tissue and LFQ yielding a TPR below 0.95 for cell lysate.

Coverage of kinase targets by the quantified phosphosites.

As TMT differs from LFQ and spike-in-SILAC regarding the identified phosphosite profiles (**Figure 1A, B**), we asked whether different quantification methods also introduce a bias in the coverage of quantified kinase substrate sites. For this purpose, an enrichment analysis[60] for kinase substrates assigned to major oncogenic signaling pathways was conducted using the kinase-substrates database from PhosphoSitePlus[61] (**Figure 5A, B**). TMT showed a higher coverage of most pathways, as compared to LFQ and spike-in-SILAC. This was mitigated by MBR and MBRreq, likely due to the higher number of quantified phosphosites through identification transfer from samples with higher phosphorylation quantities (**Figure 4A, B**).

However, only 10% of the quantified phosphosites were annotated in the kinase substrate database, limiting the power of this analysis and raising the possibility that further biases may have been missed. We therefore analyzed the coverage of kinase target motifs among the significantly quantified phosphosite. Whereas LFQ and spike-in-SILAC without or with MBR/MBRreq covered a similar array of kinase target motifs, their coverage differed for TMT (**Figure S6A, B**). We conclude that the profiles of both identified and quantified phosphosites differ between LFQ, spike-in-SILAC and TMT.

Discussion

We report the first comprehensive comparative analysis of the quantitative performance of LFQ, spike-in-SILAC and TMT for the analysis of the tumor tissue phosphoproteome.

In summary, LFQ yielded the highest number of phosphosite identifications. MBR and MBRreq increased the number of identified phosphosites for LFQ and spike-in-SILAC (**Figure 1**). The lower number of phosphosites identified in TMT is likely explained by the SPS-MS3 approach where MS2 was used for phosphopeptide identification and phosphosite localization, while MS3 for quantification of the reporter ions requires a longer duty cycle and ion trap resonance CID. In ion-trap resonance CID used for TMT analysis, the depth of the potential prevents ion ejection and precursors can only be excited to a few electron volts, requiring long activation times to build up sufficiently high internal energy for fragmentation[62]. Gas-phase rearrangement reactions can occur prior to dissociation and have been reported for phosphate moieties of phosphoserine- and phosphothreonine-containing peptides[63]. Furthermore, only the precursor is excited in ion-resonance CID[64]. For phosphopeptides containing phosphoserines and phosphothreonines, this can result in abundant non-sequence informative fragment ions corresponding to the neutral loss of the phosphate moiety from the precursor[65,66]. In contrast, in beam-type HCD, which was used for fragmentation in LFQ and spike-in-SILAC, all ions are activated, and fragments including the phosphate neutral loss can undergo several consecutive fragmentation events[67]. Therefore, Beam-type HCD spectra contain more sequence informative fragment ions than resonance CID spectra, and the shorter activation times reduced potential gas-phase rearrangements. As beam-type HCD spectra are better suited for accurate phosphosite identification than resonance CID[68], TMT yields in total less identifications than LFQ or spike-in-SILAC.

Nevertheless, TMT provided the highest number of reproducible identifications across all fold-changes and phosphosite quantities, whereas for LFQ and spike-in-SILAC the number of reproducibly identified phosphosites decreased with decreasing phosphosite quantities, which is mainly related to DDA precursor selection stochasticity. This suggests that in these approaches, reproducible identification is highly dependent on phosphosite quantities. The

isobaric and multiplexing nature of TMT tags increases the signal intensity of phosphorylated
peptides at the MS1 level, reducing the number of missing values. Although the number of
reproducibly identified phosphosites could be increased with MBR and MBRreq for LFQ and
spike-in-SILAC, TMT outperformed the other approaches in this respect for the analysis of
both tissue and cell lysates.

Regarding phosphosite quantification, LFQ without MBR provided the highest accuracy
and the lowest precision, whereas TMT showed the lowest accuracy and the highest precision,
irrespective of the matrix (**Figure 3)**. MBR and MBRreq sacrifice precision and accuracy for a
higher number of quantified phosphosites, with spike-in-SILAC suffering more severely from
this issue than LFQ. In both tissue and cell lysate TMT exhibited the highest TPRs for the
differentially quantified phosphosites (**Figure 4, Table 2**). In contrast, we found matrix effects
for both LFQ and spike-in-SILAC. LFQ yielded high TPRs for the phosphoproteome in tissue
but not cell lysate, whereas spike-in-SILAC showed high TPRs in cell lysate but not for tumor
tissue. MBR decreased the TPRs and could not mitigate these matrix-effects. However, MBR
increased the number of significantly different quantified phosphosites in LFQ and spike-in-
SILAC, albeit not to the same amount as reached by TMT. This further highlights that MBR
trades the number of phosphosites identified and quantified for quantification accuracy,
precision, and robustness.

We analyzed the tumor phosphoproteome side-by-side with a matched cancer cell line
to ensure comparability of our data with previous studies on cell lines and assess the
robustness of the quantitative methods towards matrices as different as tumor and cell lysate.
Our analysis of the cell lysate reflects overall the findings of Hogrebe et al. [26], although there
are some differences in study design that need to be considered. Hogrebe et al. [26] used the
protease Lys-C whereas we used trypsin as the most widely used protease for bottom-up
proteomics[69,70]. TMT labeling after Titanium dioxide (TiO_2) enrichment, as done by Hogrebe et
al. [26], may reduce the advantage of TMT as multiplexing occurs late in the protocol and does
not compensate technical variability introduced in the enrichment step. Therefore, we
performed TMT labeling and multiplexing directly after trypsin digestion and used IMAC instead
of TiO_2 due to its higher selectivity, identification numbers and quantitative reproducibility[71].

The quantification method significantly determines the study result in terms of
identified, quantified and differentially assigned phosphosites, as well as quantification
accuracy and last but not least the tumor related kinase phosphosite target profiles covered.
For study design, the choice of the quantification method depends on the study aims. Although
TMT outperformed spike-in-SILAC regarding quantification precision and was least amenable
to matrix effects, it exhibited the lowest accuracy. If accuracy is a key requirement, TMT is

therefore not the method of choice. Spike-in-SILAC provided the best compromise between
accuracy, precision and robustness towards differences in phosphosite quantities. However,
spike-in-SILAC suffered from a low phosphosite coverage and did not reliably detect low fold
changes between phosphosite quantities in tissue. Its performance should therefore be
assessed for the matrix to be analyzed (i.e. the tumor tissue) in order to determine the
suitability of spike-in-SILAC regarding the TPR of low fold changes. LFQ was susceptible to
matrix effects but exhibited a higher TPR in tumor tissue than in cell lysate. However, LFQ
offered low precision, which was not counterbalanced by MBR. Rather, MBR degraded the
precision and robustness of quantification. Recently, Bekker-Jensen et al. showed that data-
independent acquisition (DIA) LFQ may overcome some of the limitations in DDA LFQ, which
we observed in our study[72]. Also, practical considerations such as instrument time may
influence the choice of method. TMT allows multiplexing which reduced the measuring time in
our study by a factor of nine as compared to the LFQ and spike-in-SILAC which represents an
inherent advantage of TMT regarding time and costs.

In conclusion, different quantification methods offer the highest accuracy, precision and
phosphosite coverage in tumor tissue proteomics and thus the choice of quantification method
is critical. The different behavior of TMT, LFQ and spike-in-SILAC as well as the influence of
MBR should be also taken into account when comparing published tumor tissue
phosphoproteomes. We advocate careful annotation of tissue phosphoproteomes to enable
meaningful comparison.

Figures and tables

Figure 1: Comparison of phosphosite identifications by LFQ, spike-in-SILAC and TMT in tumor tissue lysate or cell lysate samples. A, B: Venn diagram showing the number of identified phosphosites for the three different quantification methods (LFQ, spike-in-SILAC and TMT) in the tumor tissue lysates (A) and cell lysates (B). C, D: Density plots showing for each quantification method the distribution of b- and y-ions of the identified phosphopeptides in tumor tissue lysates (C) and cell lysates (D). E, F: Bar plots showing the number of phosphosites identified in each sample group (6X, 3X and 1X) for the different quantification methods used in tumor tissue lysates (E) and cell lysates (F). The color intensity indicates the number of replicates in which the phosphosites were identified. Only phosphosites with a localization probability of at least 0.75 were considered for the analysis.

Figure 2: Evaluation of the reproducibility of phosphosite quantification. A, B: Density plots showing the distribution of CV values for phosphosites quantified with LFQ+MBR, spike-in-SILAC+MBRreq and TMT in the tumor tissue lysates (A) and cell lysates (B). Different line types (solid, dotted, dashed) indicate different sample groups (6X, 3X and 1X). C-H: Boxplots

visualizing the pairwise correlation coefficients within the same sample group (i.e. 6X vs. 6X, 3X vs. 3X, 1X vs. 1X) and between the different sample groups (i.e., 6X vs. 3X, 3X vs. 1X, 6X vs. 1X) for each method in tumor tissue lysates (C: LFQ+MBR, E: spike-in-SILAC+MBRreq, G: TMT) and cell lysates (D: LFQ+MBR, F: spike-in-SILAC+MBRreq, H: TMT). The boxes show the first, second (median) and third quartile.

Figure 3: Evaluation of precision and accuracy of phosphosite quantification. (A, C) tumor tissue lysates. (B, D) cell lysates. A, B: Violin plots showing log$_2$-transformed fold changes to evaluate quantification precision and accuracy errors of the methods. The boxes show the first, second (median) and third quartile. Whiskers show the minimum/maximum value within the 1.5 interquartile range. Expected log$_2$-transformed fold changes were highlighted by colored lines. C, D: Bar plots showing mean squared errors of accuracy and precision. Mean squared errors were calculated as the sum of the square of positive deviation and variance for each method and all replicates.

Figure 4: Performance in detecting quantitative phosphosite differences. A, B: Barplots showing the number of identified and quantified phosphosites in tumor tissue lysates (A) and cell lysates (B). Identified only: phosphosites identified in less than three replicates. Quantified non-significant: phosphosites identified and quantified in at least three out of six replicates, with FDR > 0.05. Quantified significant: phosphosites identified and quantified in at least three out of six replicates, FDR ≤ 0.05. C, D: Receiver operating characteristic (ROC) curves for evaluating the ability to diagnose the different phosphorylation quantities in tumor tissue lysates (C) and cell lysates (D). Zoomed-in ROC curves are presented in the lower right corner of each graph.

Figure 5: Kinase substrate sites covered by the quantitative phosphoproteomes. Dot-plots visualizing the enrichment analysis of quantified phosphosites from tumor tissue lysates (A) and cell lysates (B). The kinase-substrate database from PhosphositesPlus was used as background dataset. The size of the dots represents the number of phosphosites quantified for the corresponding kinase. The color intensity corresponds to the adjusted p-value using Benjamini-Hochberg approach.

	method	measurement time (hour)	number of MS2	number of identified MS2	number of MS3	number of identified phosphopeptides	number of identified unique phosphopeptides (localization probability > 0.75)
tumor tisse lysate	LFQ	36	471,216	171,753	n.a	8,932	5,812
	siSILAC	36	613,576	240,688	n.a	6,923	4,828
	TMT	4	63,421	8,122	63,345	4,281	3,578
cell lysate	LFQ	36	498,505	206,819	n.a	7,439	4,163
	siSILAC	36	597,456	236,793	n.a	5,881	3,621
	TMT	4	56,149	7,059	56,062	3,683	3,023

Table 1: Key features of tumor tissue and cell phosphoproteomes quantified by LFQ, spike-in-SILAC (siSILAC) and TMT. Shown are the measurement time, the number of MS2 spectra, the number of identified MS2 spectra, the number of MS3 spectra, the number of identified phosphopeptides and the number of identified phosphosites (minimum localization probability of 0.75).

	tumor tissue lysate					
	true positive rate			AUROC		
fold change	FC2	FC3	FC6	FC2	FC3	FC6
LFQ+MBR	0.7500	0.9700	0.9900	0.9394	0.9852	0.9938
LFQ	0.9900	0.9940	0.9990	0.9953	0.9971	0.9990
siSILAC+MBRreq	0.7700	0.8800	0.9800	0.9442	0.9633	0.9899
siSILAC	0.9200	0.9700	0.9900	0.9702	0.9881	0.9960
TMT	0.9900	0.9970	0.9980	0.9942	0.9988	0.9990
	cell lysate					
	true positive rate			AUROC		
fold change	FC2	FC3	FC6	FC2	FC3	FC6
LFQ+MBR	0.5990	0.9400	0.9690	0.8923	0.9749	0.9916
LFQ	0.5910	0.9200	0.9700	0.8992	0.9813	0.9900
siSILAC+MBRreq	0.9300	0.9000	0.9950	0.9817	0.9673	0.9931
siSILAC	0.9500	0.9800	0.9970	0.9815	0.9925	0.9988
TMT	0.9680	0.9930	0.9990	0.9879	0.9933	0.9985

Table 2: True-positive-rates (TPRs) for the different quantification methods at a false-positive-rate (FPR) threshold of 0.05, and the area under the receiver operating characteristic (AUROC).

Methods

SKOV3 cell culture and SILAC labeling. The SKOV3 cells were purchased from ATCC (HTB-77, Homo sapiens ovary). Cell were cultivated in lysine- and arginine-free Dulbecco's modified Eagle's medium (DMEM) with 1 g/L glucose, (PAN Biotech, cat. no. P04-02506), supplemented with 10% (v/v) dialyzed fetal bovine serum (FBS) (Gibco, cat. no. 10270-106), D-(+)-Glucose solution (Sigma-Aldrich, cat. no. G8644-100 ML) to a final concentration of 4.5 g/L, 3 mM L-glutamine (Gibco, cat. no. 25030-024), 84 mg/L lysine and 146 mg/L arginine. For SILAC labeling "heavy" lysine ($^{13}C_6$, $^{15}N_2$, Cambridge Isotope Laboratories Inc., CNLM-291-H-PK) and arginine ($^{13}C_6$, $^{15}N_4$, Cambridge Isotope Laboratories Inc., CNLM-539-H-PK) (heavy labeled medium) or "light" lysine (Cambridge Isotope Laboratories Inc., ULM-8766-PK) and arginine (Cambridge Isotope Laboratories Inc., ULM-8347-PK)[19] (light labeled medium) was used. For TMT and LFQ analysis cells cultivated in light labeled medium were used. All cells were maintained at 37ºC and 7.5% CO_2 for at least five passages and regularly tested for mycoplasm. Arginine-to-proline conversion and SILAC labeling efficiency was tested by LC-MS/MS. The arginine-to-proline conversion was at or below 1% and the SILAC labeling efficiency was higher than 98%.

SKOV3 cell line and ovarian cancer tissue lysis. For serum and amino acid starvation, SKOV3 cells were washed with phosphate-buffered saline (PBS, PAN Biotech, cat. no. P0436500) and cultured for 16 h in amino acids free DMEM medium (PAN Biotech, cat. no. P0401507). For stimulation, the medium was exchanged to light labeled medium for 90 min. Cells were washed three times with ice-cold PBS and lysed in protein lysis buffer [8 M urea, 75 mM NaCl, 50 mM Tris, pH 8.2] supplemented with Complete Mini (Roche, cat. no. 04 693 124 001) and PhosSTOP (Roche, cat. no. 04 906 845 001).

The ovarian cancer tissue was provided by PROMETOV and analyzed under the ethics board approval S496/2014. For tissue lysis, frozen tumor samples were crushed on dry ice and subsequently pulverized at 35 Hz for 2 min in liquid nitrogen precooled teflon milling cups using a ball mill (Retsch Mixer Mill MM 400) and a metal ball. The well-mixed tissue powder was aliquoted on dry ice before taking up in protein lysis buffer.

Cell lysates and tissue lysates were sonicated at 20% amplitude with a 20 x 1 sec pulse on ice (1 sec sonication and 1 sec cooling down), before centrifugation at 12,500 × g, 4°C for 10 min. The supernatant was transferred to new tubes. The protein concentrations were measured using the BCA Protein Assay Kit (Thermo-Fisher Scientific, cat.no. 23235) according to the manufacturer's protocol and adjusted to 1 mg/mL.

Dephosphorylation and sample pooling. For dephosphorylation of the samples, it was required to exchange the protein lysis buffer with a phosphatase reaction buffer (10 mM Tris-

HCl, 5 mM MgCl$_2$, 100 mM KCl, 0.02% Triton X-100, pH 8.0), as alkaline phosphatases are

inactive in the urea-containing protein lysis buffer. For this purpose, lysates of light labeled

cells and tissues were loaded on a centrifugal filter unit (AMICON ULTRA-15 15ML - 10 KDa,

cat. no. UFC901024) and centrifuged at 3,000 g, RT for 1 h. Subsequently, the centrifugal filter

unit was washed five times with 5 ml phosphatase reaction buffer. To remove the phosphatase

groups of the proteins[73,74], alkaline phosphatase (protein: TSAP, 100:1, w/w, FastAP

Thermosensitive Alkaline Phosphatase, cat. no. EF0651) was added to the lysates. Samples

were incubated at 37°C for 1 h. After inactivation of the reaction at 74°C for 15 min, the buffer

was changed back to the protein lysis buffer, as described before using the centrifugal filter

unit. Subsequently, the protein concentrations were determined using the BCA Assay kit. The

intensities of the phosphopeptides before and after phosphatase treatment were measured

using LC-MS/MS. The dephosphorylation efficiency was calculated as the ratio of the total

intensity of the phosphopeptides before and after phosphatase treatment. To achieve three

sample groups with different phosphorylation levels, the samples with and without

phosphatase treatment were pooled according to the scheme in **Figure S1**. The original non-

treated lysates were mixed with the dephosphorylated samples lysate at ratios of 1:5 (1X), 1:1

(3X), and 1:0 (6X). Next, each sample was divided into six aliquots containing 1 mg for LFQ,

500 μg for spike-in-SILAC, and 300 μg for TMT, each. 500 μg of the heavy SILAC labelled cell

lysate was mixed 1:1 with 500 μg of the 1X, 3X, or 6X tissue samples or the 1X, 3X, or 6X of

the light SILAC-labelled cell line samples, respectively (Figure S1, panel 3).

Protein reduction, alkylation and digestion. To reduce disulfide bonds, lysates were

incubated with 5 mM DTT at 37°C under gentle shaking for 1 h. To alkylate cysteine residues,

iodoacetamide was added to a final concentration of 14 mM and incubated for 30 min at RT in

the dark. Subsequently, DTT was added to a final concentration of 5 mM and the samples

were incubated for an additional 15 min at RT in the dark. Next, the samples were diluted with

a ratio of 1:5 with 25 mM Tris-HCl, pH 8.2 to reduce the urea concentration to 1.6 M. For tryptic

digestion, the samples were incubated with trypsin (trypsin to substrate ratio of 1:50, Promega,

cat. no. V5111) and 1 mM CaCl$_2$ for 16 h at 37°C. The tryptic digestion was stopped by adding

TFA to a final concentration of 0.4% (v/v), respectively. The samples were desalted using

SepPak tC18 cartridges (Sep-Pak tC18 1 cc Vac Cartridge, 100 mg Sorbent per Cartridge, 37

- 55 μm, cat. no. WAT036820). The cartridges were conditioned using two times 1 mL condition

buffer (95% ACN, 5% H$_2$O, 0.1% TFA), before washing two times with 1 ml of wash buffer (5%

ACN,95% H$_2$O, 0.1% TFA). Samples were loaded to the cartridges followed by washing three

times with 1 mL wash buffer. The peptides were eluted using two times 1 mL elution buffer

(50% ACN, 50% H$_2$O). For the LFQ and spike-in-SILAC measurements, 50 μL of the eluent

was used for the LC-MS/MS analysis and 950 µL for the subsequent IMAC enrichment. All the

samples were dried using a SpeedVac.

TMT labeling. For TMT labeling the dried peptides were re-solubilized in 849 µL TEAB/ACN

buffer (80 % H_2O, 10% 1M TEAB, 10% ACN). The TMT (TMT, Thermo Scientific™,

TMT10plex™ Isobaric Label Reagent Set, cat. no. 90406) reagents were solubilized in 100%

ACN to a final concentration of 100 mM. 150 µL of the TMT stock reagents were added to the

peptides according to **Figure S1** and incubated for 1 h at RT. To prevent side reactions,

hydroxylamine (Thermo Scientific™, 50% Hydroxylamine for TMT experiments, cat. no.

90115) was added to a final concentration of 0.25% [w/v], followed by incubation for 15 min at

RT. Next, samples were pooled according to the scheme in **Figure S1** and incubated for an

additional 15 min. To reduce the concentration of ACN below 5% the samples were diluted 1:1

with 2% TFA, before dilution with H_2O. Samples were desalted using SepPak tC18 cartridges

(Sep-Pak tC18 1 cc Vac Cartridge, 200 mg Sorbent per Cartridge, 37 - 55 µm, cat. no.

WAT054925). The desalting steps were performed as described before increasing the used

buffer volumes to 2 mL. 50 µg of the eluent was used for the LC-MS/MS analysis and 2950 µg

for the following IMAC enrichment. All the samples were dried using a SpeedVac.

Phosphopeptides enrichment. The phosphopeptides were enriched using the Immobilized

metal affinity chromatography (IMAC) method (High-Select™ Fe-NTA Phosphopeptide

Enrichment Kit, cat. no. A32992) according to the manufacturer protocol. In brief, the

lyophilized peptide samples were suspended in 200 µL of IMAC Binding/Wash Buffer. After

removing the bottoms closure of the IMAC spin columns and unscrewing the screw caps, the

columns were centrifuged at 1000 × g for 30 sec to remove the storage buffer. For equilibration,

the columns were washed twice with IMAC Binding/Wash Buffer. 200 µL of the suspended

peptide samples were loaded on the equilibrated spin columns, and the columns were closed

with the screw caps. The resin was mixed with the sample gently until the resin was in

suspension. The suspension was incubated for 30 min, with gentle mixing every 10 min.

Columns were washed three times with Binding/Wash Buffer and one time with HPLC grade

H_2O. The phosphopeptides were eluted by adding two times 100 µL of Elution Buffer to the

column and centrifuged at 1000 × g for 30 sec. Samples were dried in a SpeedVac.

LC-MS/MS measurement. For LC-MS/MS analysis, samples were injected on an ultrahigh

performance nano liquid chromatography system (Dionex UltiMate 3000 RSLCnano, Thermo

Scientific, Bremen, Germany) coupled to an Orbitrap mass spectrometer (Fusion™, Thermo

Fisher Scientific) with a nano electrospray source.

The samples were loaded (3 µL/min) with the buffer A (0.1% formic acid (FA) in HPLC grade

H_2O) on a trapping column (Acclaim PepMap µ-precolumn, C18, 300 µm × 5 mm, 5 µm, 100

Å, Thermo Scientific, Bremen, Germany). After sample loading, the trapping column was

washed with 30 µL buffer A (3 µL/min) and the peptides were eluted (300 µL/min) onto

separation column (Acclaim PepMap 100, C18, 75 µm × 500 mm, 2 µm, 100 Å, Thermo

Scientific, Bremen, Germany). The column temperature was kept constant at 45°C and the

peptides were separated with a gradient from 5−25% buffer B in 90 min. The spray was

generated from a silica emitter with conductive coating (O.D. 360 µm, I.D. 20 µm, Tip I.D. 10

µm New Objective, Littleton, USA) at a capillary voltage of 1800 V. MS analyses were

performed in positive ion mode and data-dependent acquisition mode (DDA). LC-MS/MS

analysis was carried out in a cycle time of 3 s and the dynamic exclusion duration was set to

30 s. For the spike-in-SILAC and LFQ samples, an intensity threshold of $2 \cdot 10^4$, an isolation

width of 0.7 m/z and an HCD collision energy of 30 NCE was used for MS/MS experiments.

Precursor MS scans were performed over a m/z range from 380-1500, with a resolution of

120,000 FWHM at m/z 200 (RF Lens = 60%, maximum injection time= 50 ms, AGC target=

$2 \cdot 10^5$) and MS/MS spectra were recorded with a resolution of 7,500 FWHM at m/z 200

(maximum injection time= 22 ms, AGC target= $2 \cdot 10^5$).

For TMT, we used multi-notch synchronous precursor selection (SPS)-MS3 technology to co-

isolate and co-fragment multiple MS2 fragment ions ion the linear ion trap, followed by reporter

ions detection in the orbitrap (MS3). The SPS-MS3 approach reduces reporter ion variance

and enhance quantification accuracy[44,45]. An intensity threshold of $5 \cdot 10^3$ and an isolation width

of 0.7 m/z was used for MS/MS experiments. Precursor MS scans were performed over a m/z

range from 380-1500, with a resolution of 120,000 FWHM at m/z 200 (RF Lens = 60%,

maximum injection time= 50 ms, AGC target = $2 \cdot 10^5$) and the precursors were fragmented by

collision-induced dissociation (CID) (AGC $1 \cdot 10^5$, normalized collision energy (NCE) = 35, q-

value = 0.25, maximum injection time = 50 ms, isolation window = 0.7) and measured in the

ion trap (ion trap scan rate = turbo). For each MS2 spectrum, we collected an MS3 spectrum

in which multiple MS2 fragment ions are captured using isolation waveforms with multiple

frequency notches. MS3 precursors were fragmented by HCD and analyzed using the Orbitrap

(NCE = 65, AGC = $1 \cdot 10^5$, maximum injection time = 120 ms, resolution = 60,000, m/z range =

120-500).

Raw data processing. All LC-MS/MS data were processed with MaxQuant[15] version 1.6.5.

Peptides and proteins identification was performed using the Andromeda search engine[18] and

searched against the Swissprot database (Uniprot, downloaded 2019-09-20, 20430 entries)

and a contamination database (cRAP-database, http://www.thegpm.org/crap, 298 entries). For

analysis of LFQ measurements, data were searched with and without "match between runs"

(MBR). For analysis of the spike-in-SILAC measurements, the data were searched with and

without MBR and re-quantify (req) option. TMT correction factors were added in MaxQuant

using the values provided by the manufacturer. Carbamidomethylation of cysteines was specified as fixed modification and oxidation of methionine, and N-terminal protein acetylation and phosphorylation of serine, threonine and tyrosine residues were defined as variable modifications. Only phosphopeptides with a localization probability larger than 0.75 were considered for further analysis. Raw and MaxQuant processed data is available at ProteomeXchange under PXD030450 identifier. The identified phosphosite with quantitative values are available in Supporting information (Table S1).

Data pre-processing and statistics analysis. The pre-processed LC-MS/MS data were further processed and statistically analyzed using R 3.6.5[75]. The phosphosites intensity of LFQ and spike-in-SILAC was normalized to the total protein intensity[76]. The phosphosites intensity of TMT samples was normalized to the reference sample to remove the variance from batches[77], before further normalization on the total protein intensity. For the correlation matrix heatmap, the correlation matrix of normalized intensity between each sample were firstly calculated and then plotted by ggplot2[78]. Linear regression model were applied to classify the sample groups and the performance of regression models are visualized by the ROC curve in ggplot2[78]. To visualize kinase substrate sites that were enriched among the quantified phosphosites, enrichment analysis of predictive kinases was performed. The kinase-substrate relationship was extracted from the thePhosphoSitePlus[61] database. For the sequence motif analysis of the quantified phosphorylation sites, the ggseqlogo[79] package was applied to generate the sequence logos. The probability of the amino acid around the identified phosphosites is shown in the sequence motif plot, with annotating the residue physicochemical properties by color. The barplot, distribution plot and violin plots were plotted using the ggplot2[78] package. The VennDiagram[80] package was used to plot Venn diagrams. R script used for data analysis and visualization is available on GitHub at https://github.com/functional-proteo-metabolomics/quantitative_phosphoproteomics.

Supporting information

The Supplemental_figures.pdf contains supplementary figures and "Table S1_Quantitative_Phosphosites table.xlsx" contains the list of identified phosphosites with quantitative information (Table S1).

Acknowledgments

We thank Ulrike Rehbein and Alienke van Pijkeren for critical reading of the manuscript.

We acknowledge support from the PROMETOV project that is supported by the national
funding organizations and the EC under the framework of the ERA-NET TRANSCAN-2
initiative (to C.O., P.H., K.T); the MESI-STRAT project (grant agreement 754688 to K.T. and
C.O.) and PoLiMeR Innovative Training Network (Marie Skłodowska-Curie grant agreement
812616 to K.T.) which both received funding from the European Union Horizon 2020 Research
and Innovation Program; the German Tuberous Sclerosis Foundation (to K.T.); Stichting TSC
Fonds (to K.T.); the German Research Foundation (TH 1358/3- 1 to K.T.); a Rosalind Franklin
Fellowship of the University of Groningen (to K.T.); the Netherlands X-omics Initiative (NWO,
project 184.034.019 to P.H.), the European Respiratory Society (ERS, RESPIRE3, project
reference: R3201703-00121, to M.K.), the University of Innsbruck (project no: 316826, to M.K.)
the Tyrolian Research Fund (project no: 18903, to M.K.) and the Deutsche
Forschungsgemeinschaft (DFG) (INST 337/15-1, INST 337/16-1 & INST 152/837-1, to H.S.).
Apart from the first author and the corresponding authors, all other authors are listed in
alphabetical order.
687

Author contributions
Study design and concept was made by Y.Z., N.G., M.K., K.T., P.H., sample preparation was
performed by Y.Z. and LC-MS/MS analysis by Y.Z., B.D., C.K., the bioinformatics analysis was
performed by Y.Z., P.H., clinical tissue collection was performed by C.A.O., P.P., project
supervision and progress discussion were made by A.M.H., N.G., S.K., H.S., C.A.O., R.B.,
M.K., K.T., P.H., funding acquisition was made by S.K., H.S., C.A.O., K.T., P.H. The paper
was written and revised by all authors.
695

696 **Competing interests**
697 The authors have no competing interests.

References

1. Cohen, P., Cross, D. & Jänne, P. A. Kinase drug discovery 20 years after imatinib:
progress and future directions. *Nat. Rev. Drug Discov.* **20**, 551–569 (2021).

2. England, T. N. USE OF CHEMOTHERAPY PLUS A MONOCLONAL ANTIBODY
AGAINST HER2 FOR METASTATIC BREAST CANCER THAT OVEREXPRESSES
HER2. *Lancet Oncol.* **344**, 783–792 (2001).

3. Moja, L. *et al.* Trastuzumab containing regimens for early breast cancer. *Cochrane*
Database Syst. Rev. **2021**, (2012).

4. An, X. *et al.* BCR-ABL tyrosine kinase inhibitors in the treatment of Philadelphia
chromosome positive chronic myeloid leukemia: A review. *Leuk. Res.* **34**, 1255–1268
(2010).

5. Salkeni, M. A., Rizvi, W., Hein, K. & Higa, G. M. Neu Perspectives, Therapies, and
Challenges for Metastatic HER2-Positive Breast Cancer. *Breast Cancer Targets Ther.*
Volume 13, 539–557 (2021).

6. Kaehler, M. & Cascorbi, I. Pharmacogenomics of Impaired Tyrosine Kinase Inhibitor
Response: Lessons Learned From Chronic Myelogenous Leukemia. *Front.*
Pharmacol. **12**, 1–12 (2021).

7. Bauer, S., George, S., von Mehren, M. & Heinrich, M. C. Early and Next-Generation
KIT/PDGFRA Kinase Inhibitors and the Future of Treatment for Advanced
Gastrointestinal Stromal Tumor. *Front. Oncol.* **11**, 1–11 (2021).

8. Ben-Hamo, R. *et al.* Predicting and affecting response to cancer therapy based on
pathway-level biomarkers. *Nat. Commun.* **11**, 1–16 (2020).

9. Buckley, T. *et al.* Comparison of Novel Immunoassay With Liquid
Chromatography/Tandem Mass Spectrometry (LC-MS/MS) for Therapeutic Drug
Monitoring of Clozapine. *Ther. Drug Monit.* **42**, 771–777 (2020).

10. Sauer, F. J., Gerber, V., Frei, S., Bruckmaier, R. M. & Groessl, M. Salivary cortisol
measurement in horses: immunoassay or LC-MS/MS? *Domest. Anim. Endocrinol.* **72**,
106445 (2020).

11. Michalski, A. *et al.* Mass spectrometry-based proteomics using Q exactive, a high-
performance benchtop quadrupole orbitrap mass spectrometer. *Mol. Cell. Proteomics*
10, 1–12 (2011).

12. Senko, M. W. *et al.* Novel parallelized quadrupole/linear ion trap/orbitrap tribrid mass

spectrometer improving proteome coverage and peptide identification rates. *Anal.*
Chem. **85**, 11710–11714 (2013).

13. Kelstrup, C. D., Young, C., Lavallee, R., Nielsen, M. L. & Olsen, J. V. Optimized fast
and sensitive acquisition methods for shotgun proteomics on a quadrupole orbitrap
mass spectrometer. *J. Proteome Res.* **11**, 3487–3497 (2012).

14. Kelstrup, C. D. *et al.* Rapid and deep proteomes by faster sequencing on a benchtop
quadrupole ultra-high-field orbitrap mass spectrometer. *J. Proteome Res.* **13**, 6187–
6195 (2014).

15. Cox, J. & Mann, M. MaxQuant enables high peptide identification rates, individualized
p.p.b.-range mass accuracies and proteome-wide protein quantification. *Nat.*
Biotechnol. **26**, 1367–72 (2008).

16. Eng, J. K., McCormack, A. L. & Yates, J. R. An approach to correlate tandem mass
spectral data of peptides with amino acid sequences in a protein database. *J. Am.*
Soc. Mass Spectrom. **5**, 976–989 (1994).

17. Tyanova, S. *et al.* The Perseus computational platform for comprehensive analysis of
(prote)omics data. *Nat. Methods* **13**, 731–740 (2016).

18. Cox, J. *et al.* Andromeda: A peptide search engine integrated into the MaxQuant
environment. *J. Proteome Res.* **10**, 1794–1805 (2011).

19. Ong, S. E. *et al.* Stable isotope labeling by amino acids in cell culture, SILAC, as a
simple and accurate approach to expression proteomics. *Mol. Cell. Proteomics* **1**,
376–386 (2002).

20. Thompson, A. *et al.* Tandem mass tags: A novel quantification strategy for
comparative analysis of complex protein mixtures by MS/MS. *Anal. Chem.* **75**, 1895–
1904 (2003).

21. Ross, P. L. *et al.* Multiplexed protein quantitation in Saccharomyces cerevisiae using
amine-reactive isobaric tagging reagents. *Mol. Cell. Proteomics* **3**, 1154–1169 (2004).

22. Merl, J., Ueffing, M., Hauck, S. M. & von Toerne, C. Direct comparison of MS-based
label-free and SILAC quantitative proteome profiling strategies in primary retinal Müller
cells. *Proteomics* **12**, 1902–1911 (2012).

23. O'Connell, J. D., Paulo, J. A., O'Brien, J. J. & Gygi, S. P. Proteome-Wide Evaluation of
Two Common Protein Quantification Methods. *J. Proteome Res.* **17**, 1934–1942
(2018).

24. Megger, D. A. *et al.* Comparison of label-free and label-based strategies for proteome
analysis of hepatoma cell lines. *Biochim. Biophys. Acta - Proteins Proteomics* **1844**,
967–976 (2014).

25. Li, Z. *et al.* Systematic comparison of label-free, metabolic labeling, and isobaric
chemical labeling for quantitative proteomics on LTQ orbitrap velos. *J. Proteome Res.*
11, 1582–1590 (2012).

26. Hogrebe, A. *et al.* Benchmarking common quantification strategies for large-scale
phosphoproteomics. *Nat. Commun.* **9**, (2018).

27. Hogrebe, A. *et al.* Benchmarking common quantification strategies for large-scale
phosphoproteomics. *Nat. Commun.*)2018(43, العدد الحا.

28. Bodzon-Kulakowska, A. *et al.* Methods for samples preparation in proteomic research.
J. Chromatogr. B Anal. Technol. Biomed. Life Sci. **849**, 1–31 (2007).

29. Hsu, C. H. *et al.* Identification and characterization of potential biomarkers by
quantitative tissue proteomics of primary lung adenocarcinoma. *Mol. Cell. Proteomics*
15, 2396–2410 (2016).

30. Zhou, Q. *et al.* Quantitative proteomics identifies brain acid soluble protein 1 (BASP1)
as a prognostic biomarker candidate in pancreatic cancer tissue. *EBioMedicine* **43**,
282–294 (2019).

31. Johansson, H. J. *et al.* Breast cancer quantitative proteome and proteogenomic
landscape. *Nat. Commun.* **10**, 1–14 (2019).

32. Zhou, B. *et al.* Quantitative proteomic analysis of prostate tissue specimens identifies
deregulated protein complexes in primary prostate cancer. *Clin. Proteomics* **16**, 1–18
(2019).

33. Coscia, F. *et al.* Multi-level Proteomics Identifies CT45 as a Chemosensitivity Mediator
and Immunotherapy Target in Ovarian Cancer. *Cell* **175**, 159-170.e16 (2018).

34. Pozniak, Y. *et al.* System-wide Clinical Proteomics of Breast Cancer Reveals Global
Remodeling of Tissue Homeostasis. *Cell Syst.* **2**, 172–184 (2016).

35. Kohale, I. N. *et al.* Quantitative analysis of tyrosine phosphorylation from FFPE tissues
reveals patient-specific signaling networks. *Cancer Res.* **81**, 3930–3941 (2021).

36. Francavilla, C. *et al.* Phosphoproteomics of Primary Cells Reveals Druggable Kinase
Signatures in Ovarian Cancer. *Cell Rep.* **18**, 3242–3256 (2017).

37.   Dazert, E. *et al.* Quantitative proteomics and phosphoproteomics on serial tumor
794   biopsies from a sorafenib-treated HCC patient. *Proc. Natl. Acad. Sci. U. S. A.* **113**,
795   1381–1386 (2016).

796   38.   Krüger, M. *et al.* SILAC Mouse for Quantitative Proteomics Uncovers Kindlin-3 as an
797   Essential Factor for Red Blood Cell Function. *Cell* **134**, 353–364 (2008).

798   39.   Geiger, T. *et al.* Use of stable isotope labeling by amino acids in cell culture as a
799   spike-in standard in quantitative proteomics. *Nat. Protoc.* **6**, 147–157 (2011).

800   40.   Nero, C. *et al.* Patient-derived organoids and high grade serous ovarian cancer: from
801   disease modeling to personalized medicine. *J. Exp. Clin. Cancer Res.* **40**, 1–14
802   (2021).

803   41.   Tumor, H. New Human Tumor Cell Lines. 115–116 (1975).

804   42.   Geiger, T., Cox, J., Ostasiewicz, P., Wisniewski, J. R. & Mann, M. Super-SILAC mix
805   for quantitative proteomics of human tumor tissue. *Nat. Methods* **7**, 383–385 (2010).

806   43.   Gao, Y. & Yates, J. R. Protein analysis by shotgun proteomics. *Mass Spectrom.*
807   *Chem. Proteomics* 1–38 (2019). doi:10.1002/9781118970195.ch1

808   44.   McAlister, G. C. *et al.* MultiNotch MS3 enables accurate, sensitive, and multiplexed
809   detection of differential expression across cancer cell line proteomes. *Anal. Chem.* **86**,
810   7150–7158 (2014).

811   45.   Ting, L., Rad, R., Gygi, S. P. & Haas, W. MS3 eliminates ratio distortion in isobaric
812   multiplexed quantitative proteomics. *Nat. Methods* **8**, 937–940 (2011).

813   46.   Albrecht, D., Kniemeyer, O., Brakhage, A. A. & Guthke, R. Missing values in gel-based
814   proteomics. *Proteomics* **10**, 1202–1211 (2010).

815   47.   Karpievitch, Y. *et al.* A statistical framework for protein quantitation in bottom-up MS-
816   based proteomics. *Bioinformatics* **25**, 2028–2034 (2009).

817   48.   Lim, M. Y., Paulo, J. A. & Gygi, S. P. Evaluating False Transfer Rates from the Match-
818   between-Runs Algorithm with a Two-Proteome Model. *J. Proteome Res.* **18**, 4020–
819   4026 (2019).

820   49.   Monetti, M., Nagaraj, N., Sharma, K. & Mann, M. Large-scale phosphosite
821   quantification in tissues by a spike-in SILAC method. *Nat. Methods* **8**, 655–658 (2011).

822   50.   Cox, J. *et al.* Accurate proteome-wide label-free quantification by delayed
823   normalization and maximal peptide ratio extraction, termed MaxLFQ. *Mol. Cell.*

Proteomics **13**, 2513–2526 (2014).

51.   Pino, L. K., Baeza, J., Lauman, R., Schilling, B. & Garcia, B. A. Improved SILAC
quantification with data independent acquisition to investigate bortezomib-induced
protein degradation. *J Proteome Res.* **20**, 1918–1927 (2021).

52.   Kuchibhotla, B. *et al.* Combinatorial Labeling Method for Improving Peptide
Fragmentation in Mass Spectrometry. *J. Am. Soc. Mass Spectrom.* **28**, 1216–1226
(2017).

53.   Anderson, Deborah K., Liang, J. W. and C. L. Spectral Library Search Improves
Assignment of TMT Labeled MS/MS Spectra. *Physiol. Behav.* **176**, 139–148 (2017).

54.   Dongré, A. R., Jones, J. L., Somogyi, Á. & Wysocki, V. H. Influence of peptide
composition, gas-phase basicity, and chemical modification on fragmentation
efficiency: Evidence for the mobile proton model. *J. Am. Chem. Soc.* **118**, 8365–8374
(1996).

55.   Everley, R. A., Huttlin, E. L., Erickson, A. R., Beausoleil, S. A. & Gygi, S. P. Neutral
Loss Is a Very Common Occurrence in Phosphotyrosine-Containing Peptides Labeled
with Isobaric Tags. *J. Proteome Res.* **16**, 1069–1076 (2017).

56.   Cox, J., Yu, S. H. & Kyriakidou, P. Isobaric matching between runs and novel PSM-
level normalization in maxquant strongly improve reporter ion-based quantification. *J.*
Proteome Res. **19**, 3945–3954 (2020).

57.   Sánchez Brotons, A. *et al.* Pipelines and Systems for Threshold-Avoiding
Quantification of LC-MS/MS Data. *Anal. Chem.* **93**, 11215–11224 (2021).

58.   'Mean Squared Error (MSE)'. *www.probabilitycourse.com.*

59.   Fawcett, T. An introduction to ROC analysis. *Pattern Recognit. Lett.* **27**, 861–874
(2006).

60.   Yu, G., Wang, L. G., Han, Y. & He, Q. Y. ClusterProfiler: An R package for comparing
biological themes among gene clusters. *Omi. A J. Integr. Biol.* **16**, 284–287 (2012).

61.   Hornbeck, P. V *et al.* PhosphoSitePlus , 2014 : mutations , PTMs and recalibrations.
43, 512–520 (2015).

62.   Potel, C. M., Lemeer, S. & Heck, A. J. R. Phosphopeptide Fragmentation and Site
Localization by Mass Spectrometry: An Update. *Anal. Chem.* **91**, 126–141 (2019).

63.   Palumbo, A. M. & Reid, G. E. Evaluation of gas-phase rearrangement and competing

fragmentation reactions on protein phosphorylation site assignment using collision
induced dissociation-MS/MS and MS3. *Anal. Chem.* **80**, 9735–9747 (2008).

64.   Schultz, K. R. *et al.* Improved early event-free survival with imatinib in Philadelphia
chromosome - Positive acute lymphoblastic leukemia: A Children's Oncology Group
Study. *J. Clin. Oncol.* **27**, 5175–5181 (2009).

65.   Palumbo, A. M. *et al.* Tandem mass spectrometry strategies for phosphoproteome
analysis. *Mass Spectrom. Rev.* **30**, 600–625 (2011).

66.   Boersema, P. J., Mohammed, S. & Heck, A. J. R. Phosphopeptide fragmentation and
analysis by mass spectrometry. *J. Mass Spectrom.* **44**, 861–878 (2009).

67.   Nagaraj, N., D'Souza, R. C. J., Cox, J., Olsen, J. V. & Mann, M. Feasibility of large-
scale phosphoproteomics with higher energy collisional dissociation fragmentation. *J.*
Proteome Res. **9**, 6786–6794 (2010).

68.   Cui, L., Yapici, I., Borhan, B. & Reid, G. E. Quantification of competing H3PO4 Versus
HPO 3 + H2O neutral losses from regioselective 18O-labeled phosphopeptides. *J. Am.*
Soc. Mass Spectrom. **25**, 141–148 (2014).

69.   Burkhart, J. M., Schumbrutzki, C., Wortelkamp, S., Sickmann, A. & Zahedi, R. P.
Systematic and quantitative comparison of digest efficiency and specificity reveals the
impact of trypsin quality on MS-based proteomics. *J. Proteomics* **75**, 1454–1462
(2012).

70.   Olsen, J. V., Ong, S. E. & Mann, M. Trypsin cleaves exclusively C-terminal to arginine
and lysine residues. *Mol. Cell. Proteomics* **3**, 608–614 (2004).

71.   Ruprecht, B. *et al.* Comprehensive and reproducible phosphopeptide enrichment using
iron immobilized metal ion affinity chromatography (Fe-IMAC) columns. *Mol. Cell.*
Proteomics **14**, 205–215 (2015).

72.   Bekker-Jensen, D. B. *et al.* Rapid and site-specific deep phosphoproteome profiling by
data-independent acquisition without the need for spectral libraries. *Nat. Commun.* **11**,
1–12 (2020).

73.   Tsai, C. F. *et al.* Large-scale determination of absolute phosphorylation
stoichiometries in human cells by motif-targeting quantitative proteomics. *Nat.*
Commun. **6**, 1–8 (2015).

74.   Imamura, H., Sugiyama, N., Wakabayashi, M. & Ishihama, Y. Large-scale
identification of phosphorylation sites for profiling protein kinase selectivity. *J.*

Proteome Res. **13**, 3410–3419 (2014).

75.   Foundation, R. C. T. R: A language and environment for statistical computing. R for
889   Statistical Computing, Vienna, Austria. URL https://www.R-project.org/. (2021).

890   76.   Wu, R. *et al.* Correct interpretation of comprehensive phosphorylation dynamics
891   requires normalization by protein expression changes. *Mol. Cell. Proteomics* **10**, 1–12
892   (2011).

893   77.   Nusinow, D. & Gygi, S. A Guide to the Quantitative Proteomic Profiles of the Cancer
894   Cell Line Encyclopedia. (2020). doi:10.1101/2020.02.03.932384

895   78.   Wickham., H. ggplot2: Elegant Graphics for Data Analysis. *Springer-Verlag New York*
896   (2016).

897   79.   Wagih, O. ggseqlogo: A 'ggplot2' Extension for Drawing Publication-Ready Sequence
898   Logos. R package version 0.1. https://CRAN.R-project.org/package=ggseqlogo.
899   (2017).

900   80.   Chen, H. VennDiagram: Generate High-Resolution Venn and Euler Plots. R package
901   version 1.6.20. https://CRAN.R-project.org/package=VennDiagram. (2018).

A Tumor tissue lysate

B Cell lysate

C Tumor tissue lysate

D Cell lysate