Title: Aerosol Formation During Processing of Potentially Infectious Samples on

Roche Immunochemistry Analyzers (cobas e analyzers) and in an End-to-End

Laboratory Workflow to Model SARS-CoV-2 Infection Risk for Laboratory Operators

Running head (<50 characters inc. spaces): Aerosol formation in laboratory

sample processing

Authors: Géza V. Burghardt,[a] Markus Eckl,[b] Doris Huether,[b] Oliver H.D. Larbolette,[b]

Alessia Lo Faso,[b] Beatus R. Ofenloch-Haehnle,[b] Marlene A. Riesch,[b] and Rolf A.

Herb[b,*]

Affiliations:

[a]Roche Diagnostics International Ltd, Rotkreuz, Switzerland; [b]Roche Diagnostics

GmbH, Penzberg, Germany.

*Address correspondence to this author at: Roche Diagnostics GmbH, 2

Nonnenwald, 82377 Penzberg, Germany; fax +49885660794090; phone

+498856604090; email rolf.herb@roche.com

Keywords: laboratory management, laboratory methods and tools, infectious

disease

Abbreviations (in order of mention): COVID-19, coronavirus disease 2019; SARS-

CoV-2, severe acute respiratory syndrome coronavirus 2; HBsAg, hepatitis B surface

antigen; LoD, limit of detection; LoB, limit of blank; CCM, cobas connection modules;

MO1, manual output station 1; TT, TeraTerm; PPE, personal protective equipment

Summary (100/100 words)

This study investigated the formation of potentially infectious aerosols during

processing of infectious samples in a model using hepatitis B surface antigen

(HBsAg) as a marker for infectious virus particles. The risk to laboratory operators of

severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) infection was then

inferred. Air sampling was performed around cobas e analyzers and in an end-to-end

laboratory workflow, after which HBsAg was quantified and applied to SARS-CoV-2

using SARS-CoV-2 infection-specific parameters. The maximum aspiration rate of

<16 viral RNA copies/8-hour shift, when applied to a SARS-CoV-2 context, poses a

remote risk of SARS-CoV-2 infection to laboratory operators.

Abstract (247/250 words):

Background: This study assessed formation of potentially infectious aerosols during

processing of infectious samples in a real-world laboratory setting, which could then

be applied in the context of severe acute respiratory syndrome coronavirus 2 (SARS-

CoV-2).

Methods: This two-part study assessed aerosol formation when using cobas e

analyzers only and in an end-to-end laboratory workflow. To estimate aerosol

formation, recombinant hepatitis B surface antigen (HBsAg) was used as a surrogate

marker for infectious virus particles to evaluate the potential risk of SARS-CoV-2

infection to laboratory operators. Using the HBsAg model, air sampling was

performed at different positions around the cobas e analyzers and in four scenarios

reflecting critical handling and/or transport locations in an end-to-end laboratory

workflow. Aerosol formation of HBsAg was quantified using the Elecsys® HBsAg II

quant II assay. The model was then applied to a SARS-CoV-2 context using SARS-

CoV-2 infection-specific parameters to calculate viral RNA copies.

Results: Following application to SARS-CoV-2, the mean HBsAg uptake per hour

when recalculated into viral RNA copies was 1.9 viral RNA copies across the cobas

e analyzers and 0.87 viral RNA copies across all tested scenarios in an end-to-end

laboratory workflow. This corresponds to a maximum aspiration rate of <16 viral RNA

copies during an 8-hour shift when using cobas e analyzers and/or in an end-to-end

laboratory workflow.

Conclusions: The low production of marker-containing aerosol when using cobas e

analyzers and in an end-to-end laboratory workflow is consistent with a remote risk

of laboratory-acquired SARS-CoV-2 infection for laboratory operators.

Introduction

Coronavirus disease 2019 (COVID-19), which was declared a worldwide pandemic in March 2020 (1), is caused by severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) (2). Widespread testing for SARS-CoV-2 helps prevent the spread of disease by identifying people in the early stages of infection, facilitating disease surveillance and thus a targeted approach for dealing with any outbreaks, and ensures appropriate, timely treatment for patients if required (3, 4). This highlights the importance of laboratory operators carrying out tests for SARS-CoV-2 as key workers in the context of the current pandemic; however, operators could be at a heightened risk of SARS-CoV-2 infection due to exposure to potentially infectious specimens (5).

Aerosols and droplets are the main forms of transmission of SARS-CoV-2 (6); therefore, all laboratory practices should be performed in a way to minimize aerosol and droplet formation to reduce the risk of SARS-CoV-2 infection to laboratory operators (7-9). The processing of samples for SARS-CoV-2 testing by laboratory operators could result in the formation of aerosols, which, if the samples are positive, represents a potentially increased risk of SARS-CoV-2 infection to laboratory operators. There is no clear understanding of infection risk in the laboratory setting (10) and aerosol formation during processing of potentially infectious samples within a SARS-CoV-2 context has yet to be investigated.

The Elecsys® SARS-CoV-2 Antigen test (Roche Diagnostics International, Rotkreuz, Switzerland) is an electrochemiluminescence immunoassay used for in vitro qualitative detection of the SARS-CoV-2 nucleocapsid protein in nasopharyngeal and oropharyngeal swab samples. The procedure involves the processing of

potentially infectious SARS-CoV-2 samples on cobas e analyzers (Roche

Diagnostics International), often as part of an automated end-to-end laboratory

workflow. The end-to-end laboratory workflow involves key processes of pre-

analytical sample preparation, sample transport to analyzers and automated output

or archive solutions, where laboratory operators could be at risk of SARS-CoV-2

infection (10).

One challenge of investigating aerosol formation in a SARS-CoV-2 context is

identifying a surrogate marker for SARS-CoV-2-positive samples. Recombinant

hepatitis B surface antigen (HBsAg) was deemed a suitable surrogate for use in this

study as it self-assembles to form virus-like particles that are assumed to distribute

via aerosols, which allows comparison with typical airborne infectious agents (11, 12).

Recombinant HBsAg is also comparable, in order of magnitude, to the size of the

SARS-CoV-2 virus particle (11, 13). Furthermore, highly positive non-infectious

HBsAg samples are stable over time and HBsAg can be detected with very high

sensitivity over a wide dynamic range using the Elecsys HBsAg II quant II assay

(Roche Diagnostics International).

The aim of this study was to assess formation of potentially infectious aerosols when

processing infectious samples using an HBsAg model on a panel of cobas e

analyzers and in an automated end-to-end laboratory workflow, similar to the

conditions in which the Elecsys SARS-CoV-2 Antigen test would be run. The model

was then used to assess the risk to laboratory operators of SARS-CoV-2 infection

due to the formation of contaminated aerosols.

Materials and Methods

This study comprised two sub-studies, the first of which assessed aerosol formation

when handling potentially infectious samples using a panel of cobas e analyzers

(conducted at Roche Diagnostics, Penzberg, Germany). The second sub-study

assessed aerosol formation when handling potentially infectious samples using

cobas e analyzers in an end-to-end laboratory workflow (conducted at Roche

Diagnostics International, Rotkreuz, Switzerland).

THE HBsAg MODEL SYSTEM

Recombinant HBsAg (Roche Diagnostics International) was used as a marker for

aerosol formation. Two different batches of HBsAg were prepared for the cobas e

analyzer panel sub-study and for the end-to-end laboratory workflow sub-study. In an

area with a high prevalence of SARS-CoV-2 infection, it would be expected that 10%

of samples submitted to a clinical laboratory would be positive and 1% would be

highly positive. In the cobas e analyzer panel sub-study, 10% of samples contained

HBsAg marker, which simulated a worst-case scenario. In the end-to-end laboratory

workflow sub-study, only samples containing HBsAg were used. The highly positive

HBsAg samples had a concentration of 1.2–2.5 x10^6 IU/mL (determined by

measurements of dilution series across the panel of cobas e analyzers) in the cobas

e analyzer panel sub-study and a concentration of approximately 1x10^6 IU/mL

(determined by measurement of a dilution series on a cobas e 801 analytical unit) in

the end-to-end laboratory workflow sub-study. For extraction purposes, negative and

positive samples contained an equivalent volume of HBsAg-specific diluent

([HBSAGQ2 Dil HepB] from the Elecsys HBsAg II quant II immunoassay kit (Roche

Diagnostics International) and were used to exclude unknown contamination within

the experimental setup and laboratory environment. The negative sample in Elecsys

SARS-CoV-2 assay runs was viral transport medium (final concentration: 2% fetal

bovine serum, 100 µg/mL gentamycin and 0.5 µg/mL amphotericin B in sterile Hanks

balanced salt solution 1X with calcium and magnesium ions) produced at Roche

Diagnostics, Penzberg for internal use only. To establish a baseline for HBsAg

measurements, blank filters were eluted and then quantified with the Elecsys HBsAg

II quant II assay.

133 AEROSOL FORMATION DURING USE OF COBAS E ANALYZERS

Prior to the main experiments, two air sampling plausibility checks (details provided

in the Supplemental Methods; Supplemental Fig. 1) were carried out in an open,

aerosol-allowing environment. This was for the purposes of demonstrating that (i)

intense handling of HBsAg-positive samples can produce HBsAg aerosol particles

and (ii) that the air sampling technique used in all experiments was able to capture

HBsAg aerosol particles, including the elution of HBsAg with HBsAg-specific diluent

(HBSAGQ2 Dil HepB) from the filters and quantification with the Elecsys HBsAg II

quant II assay.

Aerosol formation was investigated on the cobas e 801 analytical unit (cobas$^®$ 8000

configuration)**,** cobas e 402 analytical unit, cobas e 601 module, cobas e 411

analyzer and the cobas e 801 in combination with cobas pro integrated solution

(quattro-BB-configuration, equipped with four cobas e 801 analytical units)

analyzers; these units were selected because they all include testing modules for

respiratory pathogens. Samples were processed using the pipetting sequences and

application of the Elecsys SARS-CoV-2 Antigen test. All cobas e analyzers were

operated under routine laboratory conditions.

Air sampling was conducted using Gilian GilAir Plus Air Sampling Pumps (Sensidyne,

St. Petersburg, FL) equipped with polycarbonate filters (0.8 µm pore size; SKC Inc.,

Omega Division, Eighty Four, PA). The pumps were operated at a sampling air

stream of typically 2000 mL/min according to manufacturer's instructions for 4.5–8.3

hours, in combination with the IOM sampler (SKC Inc.), and were calibrated at the

beginning of each day with the Mesalabs Defender 520-M (Mesa Laboratories

Lakewood, Lakewood, CO). The air sampling devices were placed around the fan

outlets of the cobas e analyzers at shoulder/head height of the laboratory operator

and at further distances from the cobas e analyzers through placement around the

different laboratory rooms and computer workstations. Further details of the

experimental conditions for air sampling around the cobas e analyzers are provided

in Supplemental Table 1.

After a fixed time of air sampling (Supplemental Table 1), the content of the filters

was eluted with 1 mL HBsAg dilution buffer (HBSAGQ2 Dil HepB). The HBsAg

concentration was then quantified using the Elecsys HBsAg II quant II assay on the

cobas e 411 analyzer and were then standardized for comparable performance. The

effective concentration of HBsAg was determined using the unprocessed signal

results from the cobas e 801 analytic unit in conjunction with the corresponding

calibration data. The limit of detection (LoD) and the limit of blank (LoB) for the

Elecsys HBsAg II quant II assay were 0.05 IU/mL and 0.005 IU/mL, respectively.

Under routine laboratory conditions, swab testing was performed for all cobas e

analyzers included in the panel, except the cobas e 801 analytical unit in

combination with cobas pro integrated solution, to assess surface contamination due

to aerosol formation (the swab testing locations are shown in Supplemental Fig. 2).

To replicate a worst-case scenario, swab testing assessed the accumulation of

HBsAg following 8-hours of full operation (the typical length of a shift for a laboratory

operator) without any cleaning. Filter elution and subsequent quantification of HBsAg

was performed as previously described.

AEROSOL FORMATION IN AN END-TO-END LABORATORY WORKFLOW

Aerosol formation in an end-to-end laboratory workflow was assessed across four

different test scenarios (Supplemental Table 2) that reflect critical processes and

locations within this workflow (Figs. 1 and 2). In brief, the instruments assessed

across the four test scenarios were the cobas connection modules (CCM, scenario

1; Roche Diagnostics International), manual output station 1 (MO1, scenario 2 and

4b; Hitachi, Tokyo, Japan), the cobas p 612 pre-analytical unit (scenario 3 and 4;

Roche Diagnostics International), and the cobas p 501 post-analytical unit (scenario

4a; Roche Diagnostics International).

The test scenarios were chosen to allow assessment of aerosol formation during

pre-analytical sample preparation, sample transport to the analyzers and automated

output or archive solutions. Within the test scenarios, different sample tubes

(Supplemental Table 2) were used to represent the different closure types and the

most common sample tube dimensions. In scenarios 1 and 2, samples tubes were

also selected for their relatively small filling volume, which allows a high filling level

with a limited amount of HBsAg solution. A TeraTerm (TT) (Hitachi, Tokyo, Japan)

was used in scenarios 1 and 2 to enable processing of 5-position racks directly on a

linear conveyor without the use of the cobas p 612 pre-analytical unit as an entry

point for the sample tube to the transport system. Similar to the cobas e analyzer

panel sub-study, samples were processed using the pipetting sequences and

application of the Elecsys SARS-CoV-2 Antigen test.

The air sampling pump and operating procedure previously described for the cobas

e analyzer panel sub-study was also applied to this sub-study. Air sampling devices

were positioned to measure aerosol formation at handling and transport locations

(Fig. 2). Following air sampling, the elution and quantification of HBsAg

concentration was performed in the same way as previously described for the cobas

e analyzer panel sub-study but was conducted on the cobas e 801 analytical unit.

APPLICATION TO A SARS-COV-2 CONTEXT

The quantified HBsAg values were applied to a SARS-CoV-2 context using the

following assumptions: the typical air intake of a human is 600 L/h; the positive

sample used in the experimental model represents a sample from a highly infectious

SARS-CoV-2 individual with a viral load of 5×10^8 viral RNA copies/mL (14); the

effective dose needed to cause infection by SARS-CoV-2 is between 100–1000 viral

RNA copies/mL (14).

For air sampling in both studies, the HBsAg uptake for an individual was calculated

by dividing a typical human air intake of 600 L/h by the volume of air drawn through

the air sampling filter and then multiplied by the HBsAg reduction factor (calculated

by dividing the concentration of aerosol HBsAg particles measured during air

sampling by the positive HBsAg sample concentration). The HBsAg uptake for an

individual, in terms of viral RNA copies, was then calculated by multiplying the

HBsAg uptake for an individual by the viral load (assumed to be 5×10^8 viral RNA

copies/mL). In the end-to-end laboratory workflow sub-study, as only samples

containing HBsAg were used, the HBsAg uptake by an individual, in terms of viral

RNA copies, was divided by 10 as it was assumed that 10% of the SARS-CoV-2

samples in a laboratory would be positive; therefore, mirroring the cobas e analyzer

panel sub-study and replicating a worst-case real-world laboratory setting. An

overview of the calculations used for air sampling in the study is provided in Table 1.

For swab testing, the estimated maximum viral RNA copies were calculated by

multiplying the HBsAg reduction factor (calculated in the same way as the air

sampling data), by 5×10^8 (representing a viral load of 5×10^8 viral RNA copies/mL;

Table 2).

DATA PROCESSING AND ANALYSIS

230   Internally validated software (OASEpro HetIA, Roche Diagnostics International) was

231   used to calculate the concentration (IU/mL) of HBsAg from the signal detected by the

232   cobas e analyzers and data were processed using Microsoft Excel (Microsoft,

233   Redmond, WA).

Results

COBAS E ANALYZER PANEL SUB-STUDY

Air sampling points were set up at various positions around the cobas e analyzers,

as shown in Fig. 3. The maximum concentration of aerosol HBsAg particles was 0.01

IU/mL (recorded from a filter positioned around a fan outlet of the cobas e 801

analytical unit; Table 1). To improve the robustness of the study, 0.01 IU/mL was

used as the maximum concentration of HBsAg particles for all cobas e analyzers

when calculating the HBsAg reduction factor (Table 1). The cobas e 801 analytical

unit in combination with cobas pro integrated solution recorded the highest HBsAg

uptake per hour and the highest HBsAg uptake per hour when recalculated into viral

RNA copies (6.4 ppb/h and 3.2 particles, respectively; Table 1). The cobas e 402

analytical unit recorded the lowest HBsAg uptake per hour and the lowest HBsAg

uptake per hour when recalculated into viral RNA copies (2.4 ppb/h and 1.2 particles,

respectively; Table 1). Across all cobas e analyzers, the mean HBsAg uptake per

hour when recalculated into viral RNA copies was 1.9 particles, which results in a

maximum aspiration rate of <16 viral RNA copies during a typical 8-hour shift for a

laboratory operator (Table 1).

Swab testing around the cobas e 801 analytical unit, after the HBsAg assay run,

recorded the highest mean estimated maximum viral RNA copies (25.5 particles;

Table 2). Swab testing around the cobas e 402 analytical unit, after the HBsAg assay

run, recorded the lowest mean estimated maximum viral RNA copies (0.32 particles;

Table 2). For the cobas e 402 analytical unit, cobas e 601 module and cobas e 411

analyzers, there was minimal difference in the mean estimated maximum viral RNA

copies before and after the HBsAg assay run (Table 2).

END-TO-END LABORATORY WORKFLOW SUB-STUDY

Air sampling points were set up at various positions across the four test scenarios,

as shown in Supplemental Fig. 3. In scenario 1 only, the eluate from the extracted

filters exhibited an effective concentration of HBsAg particles above the LoD (0.05

IU/mL) of the Elecsys HBsAg II quant II assay (Fig. 4). For scenario 1, HBsAg levels

above the LoD were also evident in the corresponding negative control and this

result was not observed in the subsequent repetitions of the experiment (Fig. 4).

Scenario 1 produced the highest mean concentration of aerosol HBsAg particles

(0.04975 IU/mL) across the tested scenarios (Table 1). Scenarios 4 and 2 produced

the highest and lowest HBsAg uptake per hour (34.57 ppb/h and 4.07 ppb/h,

respectively; Table 1). With the assumption that 10% of the SARS-CoV-2 samples in

a laboratory are positive, scenarios 4 and 2 also produced the highest and lowest

HBsAg uptake per hour when recalculated into viral RNA copies (1.73 particles and

0.20 particles, respectively; Table 1), while the corresponding mean across the test

scenarios was 0.87 particles (Table 1). As such, there is a mean maximum

aspiration rate of <16 viral RNA copies during an 8-hour shift.

Discussion

To our knowledge, this is the first study to assess formation of potentially infectious

aerosols when processing infectious samples using a panel of cobas e analyzers

and in an automated end-to-end laboratory workflow. While viral agents transmitted

through blood and bodily fluids are the most common source of infection for

personnel within a diagnostic laboratory (15), it is timely to assess aerosol infection

risk given the present COVID-19 pandemic. In this study, data on potentially

infectious aerosol formation were applied to a SARS-CoV-2 context and the low

number of viral RNA copies detected inferred a minimal risk of SARS-CoV-2

infection to laboratory operators when processing potentially infectious samples.

In the end-to-end laboratory workflow sub-study, in scenario 1 (the CCM/TT units),

elevated levels of HBsAg above the LoD were evident, elevated levels were also

evident in the corresponding negative control; however, such levels were not present

in the subsequent repetitions of the experiment. This can be considered a

contamination-driven effect and was not considered to originate from the samples

processed on the CCM/TT units.

Hepatitis B is spread through contact with infected bodily fluids; however, HBsAg can

self-assemble to virus-like particles that are assumed to distribute via aerosols;

therefore, the use of recombinant HBsAg in the model system allows the model to be

applied to all airborne pathogens and not just SARS-CoV-2. This addresses an

important unmet clinical need as aerosol formation has previously been undervalued

in terms of transmission of respiratory viral diseases due to a lack of understanding

of how infectious aerosols are produced and transported (16).

In this study, the mean maximum aspiration rate detected was <16 viral RNA copies

during an 8-hour shift when using cobas e analyzers and in an end-to-end laboratory

context. Furthermore, for swab testing, the greatest number of estimated maximum

viral RNA copies was limited to 25.5 particles over 8-hours of full operation and there

was only a minimal difference in mean estimated maximum viral RNA copies before

and after the HBsAg assay run. The infectious dose of a virus can vary greatly

among respiratory viruses (17). For SARS-CoV-2, the minimum dose of virus

particles necessary to cause infection remains an active research question; however,

it has been estimated to be between 100 and 1000 particles (18-21). Therefore, the

risk of SARS-CoV-2 infection for laboratory operators when using cobas e analyzers

and in an end-to-end laboratory workflow can be considered remote. Applying the

outcomes of studies of other viruses that have evaluated risk of infection to

laboratory operators may not be relevant in a SARS-CoV-2 context due to the highly

infectious nature of SARS-CoV-2; generally inhaled viruses require 1950–3000

particles to cause infection, compared with between 100 and 1000 particles for

SARS-CoV-2 (18-21).

When applying the findings of the study to a SARS-CoV-2 context, the assumption

regarding viral load did not consider the effect of variants of SARS-CoV-2. For

example, in the first known community transmission event of the delta variant in

mainland China, it was shown that the viral load associated with the delta variant

was approximately 1000 times higher than with the alpha or beta strain (22). In

addition, vaccination status has also been shown to affect viral load kinetics,

whereby vaccinated individuals have a faster mean rate of viral load decline relative

to unvaccinated individuals (23). Some European countries have already made

vaccination against COVID-19 mandatory for all healthcare workers (24, 25).

A strength of this study is that a worst-case scenario (10% of samples were highly

positive) was tested in several different analyzers using a clinically relevant end-to-

end laboratory setup; thus, replicating a real-world testing laboratory for SARS-CoV-

2. A further strength of this study is that swab testing of the instrument surfaces was

performed, which, in addition to aerosol formation, could represent a source of

infection for laboratory workers (26-29). In a similar study, Farnsworth et al. used a

fluorescent marker to assess instrument and specimen contamination and found that

while there is a low risk of instrument contamination, the handling of infectious

specimen containers can contaminate laboratory surfaces (30). Relative to a

fluorescent marker, the use of the HBsAg model in this study was more

representative of the mass: volume ratio seen in the aerosol distribution of virus

particles. Correct adherence in the use of personal protective equipment (PPE) can

be very effective in preventing infection through contact with contaminated surfaces

(30) and inhalation of droplets associated with SARS-CoV-2 infection (31, 32). Van

Doremalen et al. found that under experimental conditions, SARS-CoV-2 is viable in

aerosols for 3 hours (9); therefore, PPE should always be worn in a laboratory

setting. Healthcare workers caring for patients with COVID-19 report adherence

rates of 98.6% to PPE protocols (33); similar adherence rates are important to

reduce SARS-CoV-2 infection risk in a laboratory context. Future work should

assess molecular modules, to expand upon the findings for analyzers shown here.

In conclusion, the low production of potentially infectious aerosols when using cobas

e analyzers and in an end-to-end laboratory workflow (which includes processes of

pre-analytical sample preparation, sample transport to analyzers and automated

output or archive solutions) is consistent with a remote risk of SARS-CoV-2 infection

for laboratory operators. These results can be considered reassuring to laboratory

operators involved in the processing of potentially infectious samples.

Data Availability: The data supporting this manuscript can be requested in writing

from the corresponding author.

Ethical Guidelines: The artificial samples used for this study were produced using

purified, vendor purchased hepatitis B surface antigen (HBsAg) protein and dilution

media; therefore, no ethical approval was needed.

Author Contributions: All authors confirmed they have contributed to the

intellectual content of this paper and have met the following 4 requirements: (a)

significant contributions to the conception and design, acquisition of data, or analysis

and interpretation of data; (b) drafting or revising the article for intellectual content;

(c) final approval of the published article; and (d) agreement to be accountable for all

aspects of the article thus ensuring that questions related to the accuracy or integrity

of any part of the article are appropriately investigated and resolved.

Authors' Disclosures or Potential Conflicts of Interest:

Employment or Leadership: Markus Eckl, Doris Huether, Oliver Larbolette, Alessia

Lo Faso, Beatus Ofenloch-Haehnle, Marlene Riesch and Rolf Herb are employed by

Roche Diagnostics GmbH. Géza Burghardt is employed by Roche Diagnostics

International Ltd.

Consultant or Advisory Role: Markus Eckl participated in the Roche Diagnostics SARS-CoV-2 Biosafety, Antigen Testing & Infectivity Advisory Board Meeting (8[th] December 2020)

Stock Ownership: Géza Burghardt, Markus Eckl, Oliver Larbolette, Beatus Ofenloch-Haehnle and Rolf Herb hold non-voting equities in F. Hoffman-La Roche

Honoraria: N/A

Research Funding: Roche Diagnostics International Ltd funded the study and third-party medical writing assistance

Expert Testimony: N/A

Patents: N/A

Other Renumeration: N/A

Funding: The cobas e analyzer panel sub-study was funded by Roche Diagnostics GmbH (Penzberg, Germany) and the end-to-end laboratory workflow sub-study was funded by Roche Diagnostics International Ltd (Rotkreuz, Switzerland).

Acknowledgements: The authors wish to thank Marie Christine Weiss, PhD, for providing additional support in planning/reviewing experimental concepts, Thomas Grau, PhD, (as part of Portfolio Solution Integration at Roche Diagnostics International Ltd) for financing of the end-to-end laboratory workflow sub-study and Ursula Giesen, PhD, for a comprehensive review of both the work conducted in the cobas e analyzer panel sub-study and subsequent reports. Third-party medical writing support for the development of this manuscript, under the direction of the authors, was provided by Luke Edmonds, MSc, and Heather Small, PhD, of Ashfield MedComms (Macclesfield, UK), an Ashfield Health company, and was funded by

389   Roche Diagnostics International Ltd (Rotkreuz, Switzerland). COBAS, COBAS E,

390   COBAS P and ELECSYS are trademarks of Roche. All other product names and

391   trademarks are property of their respective owners.

References

1. WHO. WHO Director-General's opening remarks at the media briefing on COVID-19 - 11 March 2020. https://www.who.int/director-general/speeches/detail/who-director-general-s-opening-remarks-at-the-media-briefing-on-covid-19---11-march-2020 (Accessed October 26, 2021).

2. Zhou P, Yang XL, Wang XG, Hu B, Zhang L, Zhang W, et al. A pneumonia outbreak associated with a new coronavirus of probable bat origin. Nature 2020;579:270-3.

3. Salathé M, Althaus CL, Neher R, Stringhini S, Hodcroft E, Fellay J, et al. COVID-19 epidemic in Switzerland: on the importance of testing, contact tracing and isolation. Swiss Med Wkly 2020;150:w20225.

4. Lippi G, Plebani M. The critical role of laboratory medicine during coronavirus disease 2019 (COVID-19) and other viral outbreaks. Clin Chem Lab Med 2020;58:1063-9.

5. Pambuccian SE. The COVID-19 pandemic: implications for the cytology laboratory. J Am Soc Cytopathol 2020;9:202-11.

6. Wang J, Du G. COVID-19 may transmit through aerosol. Ir J Med Sci 2020;189:1143-4.

7. CDC. Interim laboratory biosafety guidelines for handling and processing specimens associated with coronavirus disease 2019 (COVID-19). https://www.cdc.gov/coronavirus/2019-nCoV/lab/lab-biosafety-guidelines.html (Accessed October 26, 2021).

8. WHO. Laboratory biosafety guidance related to coronavirus disease (COVID-19): Interim guidance, 28 January 2021.

https://www.who.int/publications/i/item/WHO-WPE-GIH-2021.1 (Accessed

October 26, 2021).

9.   van Doremalen N, Bushmaker T, Morris DH, Holbrook MG, Gamble A,

Williamson BN, et al. Aerosol and surface stability of SARS-CoV-2 as

compared with SARS-CoV-1. N Engl J Med 2020;382:1564-7.

10.   Karthik K, Aravindh Babu RP, Dhama K, Chitra MA, Kalaiselvi G, Alagesan

Senthilkumar TM, et al. Biosafety concerns during the collection,

transportation, and processing of COVID-19 samples for diagnosis. Arch Med

Res 2020;51:623-30.

11.   Yamaguchi M, Sugahara K, Shiosaki K, Mizokami H, Takeo K. Fine structure

of hepatitis B virus surface antigen produced by recombinant yeast:

comparison with HBsAg of human origin. FEMS Microbiol Lett 1998;165:363-

7.

12.   Ludwig C, Wagner R. Virus-like particles-universal molecular toolboxes. Curr

Opin Biotechnol 2007;18:537-45.

13.   Laue M, Kauter A, Hoffmann T, Möller L, Michel J, Nitsche A. Morphometry of

SARS-CoV and SARS-CoV-2 particles in ultrathin plastic sections of infected

Vero cell cultures. Sci Rep 2021;11:3515.

14.   Lelieveld J, Helleis F, Borrmann S, Cheng Y, Drewnick F, Haug G, et al.

Model calculations of aerosol transmission and infection risk of COVID-19 in

indoor environments. Int J Environ Res Public Health 2020;17.

15.   Sewell DL. Laboratory-associated infections and biosafety. Clin Microbiol Rev

1995;8:389-405.

16.   Wang CC, Prather KA, Sznitman J, Jimenez JL, Lakdawala SS, Tufekci Z, et

al. Airborne transmission of respiratory viruses. Science 2021;373:eabd9149.

17.   Di Girolamo P. Assessment of the potential role of atmospheric particulate

443   pollution and airborne transmission in intensifying the first wave pandemic

444   impact of SARS-CoV-2/COVID-19 in Northern Italy. Bull Atmos Sci Technol

445   2020;1:515-50.

446   18.   Evans MJ. Avoiding COVID-19: aerosol guidelines. Preprint at

447   https://www.medrxiv.org/content/medrxiv/early/2020/06/05/2020.05.21.20108

448   894.full.pdf (2020).

449   19.   Karimzadeh S, Bhopal R, Nguyen Tien H. Review of infective dose, routes of

450   transmission and outcome of COVID-19 caused by the SARS-COV-2:

451   comparison with other respiratory viruses. Epidemiol Infect 2021;149:e96.

452   20.   Barr GD. A model showing the relative risk of viral aerosol infection from

453   breathing and the benefit of wearing masks in different settings with

454   implications for Covid-19. Preprint at

455   https://www.medrxiv.org/content/medrxiv/early/2020/08/18/2020.04.28.20082

456   990.full.pdf (2020).

457   21.   Basu S. Close-range exposure to a COVID-19 carrier: transmission trends in

458   the respiratory tract and estimation of infectious dose. Preprint at

459   https://www.medrxiv.org/content/medrxiv/early/2020/07/29/2020.07.27.20162

460   362.full.pdf (2020).

461   22.   Li B, Deng A, Li K, Hu Y, Li Z, Xiong Q, et al. Viral infection and transmission

462   in a large, well-traced outbreak caused by the SARS-CoV-2 Delta variant.

463   Preprint at

464   https://www.medrxiv.org/content/medrxiv/early/2021/07/23/2021.07.07.21260

465   122.full.pdf (2021).

23. Singanayagam A, Hakki S, Dunning J, Madon KJ, Crone MA, Koycheva A, et

al. Community transmission and viral load kinetics of the SARS-CoV-2 delta

(B.1.617.2) variant in vaccinated and unvaccinated individuals in the UK: a

prospective, longitudinal, cohort study. [Epub ahead of print] Lancet Infect Dis

October 29, 2021 as doi:10.1016/S1473-3099(21)00648-4.

24. Paterlini M. Covid-19: Italy makes vaccination mandatory for healthcare

workers. BMJ 2021;373:n905.

25. Wise J. Covid-19: France and Greece make vaccination mandatory for

healthcare workers. BMJ 2021;374:n1797.

26. Boone SA, Gerba CP. Significance of fomites in the spread of respiratory and

enteric viral disease. Appl Environ Microbiol 2007;73:1687-96.

27. Lopman B, Gastañaduy P, Park GW, Hall AJ, Parashar UD, Vinjé J.

Environmental transmission of norovirus gastroenteritis. Curr Opin Virol

2012;2:96-102.

28. Repp KK, Keene WE. A point-source norovirus outbreak caused by exposure

to fomites. J Infect Dis 2012;205:1639-41.

29. Thornley CN, Emslie NA, Sprott TW, Greening GE, Rapana JP. Recurring

norovirus transmission on an airplane. Clin Infect Dis 2011;53:515-20.

30. Farnsworth CW, Wallace MA, Liu A, Gronowski AM, Burnham C-AD,

Yarbrough ML. Evaluation of the risk of laboratory microbial contamination

during routine testing in automated clinical chemistry and microbiology

laboratories. Clin Chem 2020;66:1190-9.

31. Dos Santos WM. Use of personal protective equipment reduces the risk of

contamination by highly infectious diseases such as COVID-19. Evid Based

Nurs 2021;24:41.

32.   Verbeek JH, Rajamaki B, Ijaz S, Sauni R, Toomey E, Blackwood B, et al.

492   Personal protective equipment for preventing highly infectious diseases due to

493   exposure to contaminated body fluids in healthcare staff. Emergencias

494   2021;33:59-61.

495   33.   Zhao Y, Liang W, Luo Y, Chen Y, Liang P, Zhong R, et al. Personal protective

496   equipment protecting healthcare workers in the Chinese epicentre of COVID-

497   19. Clin Microbiol Infect 2020;26:1716-8.

Tables

Table 1. Overview of the calculations and results for air sampling[a].

Overview of experimental results	cobas e analyzer panel sub-study					End-to-end laboratory workflow sub-study			
	cobas e 801 (cobas 8000 configuration)	cobas e 402	cobas e 601	cobas e 411	cobas e 801 with cobas pro integrated solution	Scenario 1	Scenario 2	Scenario 3	Scenario 4
HBsAg concentration (IU/mL)[b]	0.01	0.00167	0.003	0.004	0.004	0.04975	0.00244	0.00585	0.02074
Corrected maximum HBsAg concentration (IU/mL)	0.01	0.01	0.01	0.01	0.01	N/A	N/A	N/A	N/A
A) HBsAg concentration used in	0.01	0.01	0.01	0.01	0.01	0.04975	0.00244	0.00585	0.02074

calculation (IU/mL)									
B) Positive HBsAg sample concentration (IU/mL)	$\sim 2 \times 10^6$	$\sim 2.5 \times 10^6$	$\sim 1.2 \times 10^6$	$\sim 2 \times 10^6$	$\sim 1.5 \times 10^6$	1×10^6	1×10^6	1×10^6	1×10^6
C) HBsAg reduction factor (ppb) (C = A/B)	5	4	8.3	5	6.7	49.75	2.44	5.85	20.74
D) Volume of air through each filter (L)	900	998	1900	700	627	1800	360	240	360
E) typical human air intake	10 L/minute or 600 L/h								
F) HBsAg-uptake of a human per hour (ppb/h) (F = C x E/D)	3.3	2.4	2.6	4.3	6.4	16.58	4.07	14.62	34.57
Mean HBsAg-uptake of a human per hour	3.8					17.46			

(ppb/h)									
G) HBsAg-uptake of a human per hour recalculated into viral RNA copies (G = F x 5×10^8)	1.7	1.2	1.3	2.14	3.2	8.29	2.03	7.31	17.28
HBsAg-uptake of a human per hour recalculated into viral RNA copies in a real-world laboratory setting[c]	1.7	1.2	1.3	2.14	3.2	0.83	0.20	0.73	1.73
Mean	**1.9 viral RNA copies per hour** This corresponds to a maximum personal aspiration rate of <16 viral RNA copies in 8 hours					**0.87 viral RNA copies per hour** This corresponds to a maximum personal aspiration rate of <16 viral RNA copies in 8 hours			

[a]The differences in the decimal places shown in this table is due to the different readout capabilities of the analyzers. [b]In the cobas e analyzer panel sub-study, the maximum HBsAg concentration was recorded, whereas the mean HBsAg concentration was recorded in the

end-to-end laboratory workflow sub-study. [c]To apply the results to a worst-case real-world laboratory setting (10% of SARS-CoV-2 samples submitted to a laboratory are positive), 10% of samples contained HBsAg marker in the cobas e analyzer panel sub-study; whereas, in the end-to-end laboratory workflow sub-study, the assumption that 10% of the SARS-CoV-2 samples in a laboratory are positive was applied when calculating the uptake of viral RNA copies. Abbreviations: HBsAg, hepatitis B surface antigen; IU, international unit; N/A, not applicable; ppb, parts per billion; RNA, ribonucleic acid.

Table 2. Overview of the calculations and results for swab testing[a,b].

Overview of experimental results	cobas e analyzer panel sub-study							
	cobas e 801 (cobas 8000 configuration)[c]		cobas e 402		cobas e 601		cobas e 411	
	Before HBsAg assay run	After HBsAg assay run	Before HBsAg assay run	After HBsAg assay run	Before HBsAg assay run	After HBsAg assay run	Before HBsAg assay run	After HBsAg assay run
A) HBsAg concentration (IU/mL)[d]	-	0.102	0.0024	0.0016	0.0024	0.022	0.004	0.004
B) Positive HBsAg sample concentration (IU/mL)	~2 x 10^6	~2 x 10^6	~2.5 x 10^6	~2.5 x 10^6	~1.2 x 10^6	~1.2 x 10^6	~2 x 10^6	~2 x 10^6
C) HBsAg reduction factor (ppb) (C = A/B)[d]	-	51.0	0.96	0.64	2.0	18.3	2.0	2.0

| Estimated maximum viral RNA copies $(C \times 5 \times 10^8)^d$ | - | 25.5 | 0.48 | 0.32 | 1.0 | 9.2 | 1.0 | 1.0 |

[a]Swab testing assessed the accumulation of HBsAg over one day of full operation without any cleaning to replicate a worst-case scenario. [b]The differences in the decimal places shown in this table is due to the different readout capabilities of the analyzers. [c]Data were not collected before the HBsAg assay run for the cobas e 801 analytical unit (cobas 8000 configuration). [d]For each cobas e analyzer, the values presented are the mean results from all different swab testing locations. Abbreviations: HBsAg, hepatitis B surface antigen; IU, international unit; ppb, parts per billion; RNA, ribonucleic acid.

Figures

Fig. 1. Floor plan of the end-to-end laboratory workflow sub-study. Arrows indicate

the manual transport of sample tubes.

Abbreviations: AOB, add-on buffer unit; BLM, bulk loader module; cPRO, cobas pro;

c503, cobas 503 clinical chemistry analyzer; c701, cobas 701 clinical chemistry

analyzer; c702, cobas 702 clinical chemistry analyzer; c6800, cobas 6800 molecular

analyzer; c8000, cobas 8000 modular analyzer; e602, cobas 602 immune analyzer;

511 e801, cobas e 801 analytical unit; ISE, ion selective electrode module; MO1, manual

512 output station 1; PRIME, cobas PRIME, pre-analytical system; p471, cobas p 471

513 centrifuge unit; p501, cobas p 501 post-analytical system; p612, cobas p 612 pre-

514 analytical unit.

Fig. 2. Placement of air sampling units in the end-to-end laboratory workflow.

Abbreviations: AOB, add-on buffer unit; BLM, bulk loader module; cPRO, cobas pro;

c503, cobas 503 clinical chemistry analyzer; c701, cobas 701 clinical chemistry

analyzer; c702, cobas 702 clinical chemistry analyzer; c6800, cobas 6800 molecular

analyzer; c8000, cobas 8000 modular analyzer; e602, cobas 602 immune analyzer;

e801, cobas e 801 analytical unit; ISE, ion selective electrode module; MO1, manual

output station 1; PRIME, cobas PRIME, pre-analytical system; p471, cobas p 471

centrifuge unit; p501, cobas p 501 post-analytical system; p612, cobas p 612 pre-

analytical unit.

Fig. 3. Selection of air sampling positions included in the cobas e analyzer panel sub-study[a]. (A) cobas e 801 analytical unit (cobas 8000 configuration). (B) cobas e 402 analytical unit (C) cobas e 601 module. (D) cobas e 411 analyzer. (E) cobas e 801 analytical unit in combination with cobas pro integrated solution.

530   *a*Positions of air sampling around fan outlets are indicated by red arrows and the

531   sample pipette location is indicated by a blue arrow.

532 **Fig. 4.** The effective HBsAg analyte concentration from each of the testing scenarios

533 investigated in the end-to-end laboratory workflow sub-study.

534
535

536 Quantification of HBsAg was performed on the cobas e 801 analytical unit.

537 Abbreviations: CCM, cobas connection module; HBsAg, hepatitis B surface antigen;

538 IU, international unit; MO1, manual output station 1; TT, TeraTerm.