An Optrode Array for Spatiotemporally Precise Large-Scale Optogenetic Stimulation of Deep Cortical Layers in Non-human Primates

Andrew M. Clark[1], Alexander Ingold[1], Christopher F. Reiche[2], Donald Cundy III[1], Justin L. Balsor[1], Frederick Federer[1], Niall McAlinden[3], Yunzhou Cheng[3], John D. Rolston[4,√], Loren Rieth[5,6], Martin D. Dawson[3], Keith Mathieson[3], Steve Blair[2*], and Alessandra Angelucci[1*]

[1] Department of Ophthalmology and Visual Science, Moran Eye Institute, University of Utah, Salt Lake City, UT

[2] Department of Electrical and Computer Engineering, University of Utah, Salt Lake City, UT

[3] SUPA, Institute of Photonics, Department of Physics, University of Strathclyde, Glasgow, UK

[4] Departments of Neurosurgery and Biomedical Engineering, University of Utah, Salt Lake City, UT

[5] Mechanical and Aerospace Engineering, West Virginia University, Morgantown, WV

[6] Feinstein Institute for Medical Research, Manhasset, NY

[√] Present address: Department of Neurosurgery, Brigham & Women's Hospital, Harvard Medical School, Boston, MA

*Corresponding authors.

Corresponding authors' address:

A. Angelucci:

65 Mario Capecchi Drive
Salt Lake City, UT 84132, USA
Tel: (801) 5857489
Email: alessandra.angelucci@hsc.utah.edu

S. Blair:

50 S. Central Campus Dr., Rm. 2110
Salt Lake City, UT 84112, USA
Tel: (801) 5856157
Email: blair@ece.utah.edu

ABSTRACT

Optogenetics has transformed studies of neural circuit function, but remains challenging to apply in non-human primates (NHPs). A major challenge is delivering intense and spatially precise patterned photostimulation across large volumes in deep tissue. Here, we have developed and validated the Utah Optrode Array (UOA) to meet this critical need. The UOA is a 10×10 glass waveguide array bonded to an electrically-addressable μLED array. *In vivo* electrophysiology and immediate early gene (c-fos) immunohistochemistry demonstrated that the UOA allows for large-scale spatiotemporally precise neuromodulation of deep tissue in macaque primary visual cortex. Specifically, the UOA permits both focal (confined to single layers or columns), and large-scale (across multiple layers or columns) photostimulation of deep cortical layers, simply by varying the number of simultaneously activated μLEDs and/or the light irradiance. These results establish the UOA as a powerful tool for studying targeted neural populations within single or across multiple deep layers in complex NHP circuits.

Keywords

Optogenetics has transformed the study of neural circuit function by allowing for the

selective modulation of neural activity on a physiologically relevant timescale[1]. Progress in

applying optogenetics to non-genetically tractable models, such as the non-human primate

(NHP), has lagged behind that in the mouse[2]. Extending optogenetics to NHP studies is crucial,

as, due to their similarity to humans, NHPs represent critical models for understanding neural

circuit function and dysfunction[3-6], and provide an essential technology testbed towards the

potential application of optogenetics as therapeutic interventions in humans [7,8]. The continuing

refinement of viral methods for selectively delivering opsins to particular circuits [9,10] or cell

types[11-13], is opening up new opportunities to study neural circuits in NHPs[2,14]. Despite these

advances, a significant remaining obstacle is the lack of devices for reliably delivering light of

sufficient intensity to deep neural tissue across relatively large brain volumes with sufficient

spatial resolution to selectively modulate relevant circuit elements.

There are several features of cortical networks that provide both impetus and design

requirements for such a device.   For example, cortico-cortical feedback connections, which are

critical for the contextual modulation of sensory processing[9,15], as well as various cognitive

phenomena[16,17], and cortico-thalamic projections, arise from deep cortical layers[18,19]. Dissecting

these circuits requires selective perturbation of deep layer neurons with high spatiotemporal

precision. Moreover, the columnar architecture of the NHP cortex, which extends throughout the

cortical layers[20], requires optogenetic perturbations at the spatial scale of cortical columns

through the cortical depth. Methods for high-spatial resolution optogenetics recently developed

in smaller animals[21,22] only allow for stimulation of the superficial layers in the NHP.

Currently, NHP optogenetic experiments mainly follow two light delivery approaches:

through-surface illumination and penetrating probes. Surface photostimulation utilizes either a

laser- or LED-coupled optical fiber positioned above the cortex[9], or chronically-implantable

surface LED arrays[23]. These approaches enable photoactivation of a large area, but only to a

depth of < 1mm, due to light attenuation and scattering in tissue, as well as to unintended

superficial layer neuron activation and even heating damage at the higher intensities required to

reach deep layers[9,24]. In contrast, penetrating optical fibers, integrated with single[25,26] or

multiple[27] recording probes, allow photoactivation at depths >1mm, but only of a volume a few

hundred microns in diameter, and, due to their size and shape, can cause significant superficial

layer damage.

To overcome the above limitations, we developed the Utah Optrode Array (UOA), a 10x10 array of glass needle shanks tiling a 4x4 mm^2 area bonded to an electrically-addressable µLED array independently delivering light through each shank[28,29]. *In vivo* testing in macaque primary visual cortex (V1) demonstrated the UOA allows for spatio-temporally patterned photostimulation of deep cortical layers with sub-millimeter resolution (at the scale of single layers and columns) over a large volume. This selectivity can be scaled up to multiple layers and columns by varying the number of simultaneously activated µLEDs and/or the light irradiance. These results establish the UOA as a powerful tool for studying local and large-scale populations of deep layer neurons in NHP cortex.

                                                    RESULTS

98   **The UOA: Geometry and Optical Properties**

99   **Figure 1 about here**

The UOA is based on the geometry of the Utah Electrode Array (UEA)[30]. It is a 10x10 array of

penetrating glass optical light guides (needles), with customizable length (up to 2.5mm) and

shank width (80-120µm) on a 400µm pitch tiling 16mm². A custom µLED array fabricated on a

GaN on Sapphire wafer is directly integrated with the device, with each electrically addressable

80 x 80µm µLED delivering 450nm light through a single needle (**Fig. 1A-E**). A second 9x9

array of "interstitial" µLEDs is interleaved on the same device for independent surface

stimulation (as shown in **Fig. 1B**, but not used in this study). To limit the spatial spread of

coupled light, the first generation UOA used a metal pinhole array[28]. Bench testing

demonstrated the capacity of this device for delivering patterned light at irradiances in excess of

activation threshold across a range of commonly employed depolarizing[31] and hyperpolarizing[32]

opsins, with a 50% decrease in irradiance within tissue about 200µm from a needle tip[28]. These

initial results suggested that direct optogenetic activation through the UOA is on a spatial scale

commensurate with the functional architecture of primate cortex.

113   Here we have developed the second generation UOA, which incorporates an optically

opaque interposer layer with optical "vias" to eliminate unwanted surface illumination and inter-

needle crosstalk (**Fig. 1A,C**; see Online Methods for manufacturing details). This device (**Fig.**

1A-E) was first bench tested (**Fig. 1F**); *in vivo* optical performance was then estimated via ray

tracing (**Fig. 1G**). Maps of output power (mW) at each needle tip at different drive voltages are

shown in **Fig. 1F** (**Extended Data Fig. 1**, also shows the estimated output irradiances). At 3V,

output power and estimated irradiance levels are below the 1mW/mm² threshold for the

excitatory opsin *Channelrhodopsin-2* (ChR2) (**Extended Data Fig. 1 and Extended Data Table**

1). Note that defining the irradiance emitted from faceted optrode tips is challenging. For

simplicity, in **Extended Data Figure 1B**, we define the irradiance as the emitted optical power

divided by the area of the emission surface; however, optical modeling indicates that the

emission is non-uniform, with higher irradiance near the tip apex (**Fig. 1G**). There is also

variation in emission across the array, due primarily to variations in the resistance (and therefore

slope efficiency) of each µLED. At 3.5V, about 30% of the stimulation sites reach or exceed

ChR2   threshold   (mean   optical   power±SD   =   0.022±0.013mW;   mean   irradiance

=0.82±0.49mW/mm^2), while at 5V, more than 90% of the sites emit above threshold

(0.1±0.056mW; 3.79±2.08mW/mm^2). In principle, software modifications in the matrix driver

interface can be made to better equalize stimulation levels across the array.

Using optical ray tracing, we estimated the direct neural stimulation volume (based upon

the local irradiance in tissue) as a function of drive voltage and pattern of activated needles to

facilitate interpretation of the *in vivo* results (see Online Methods). The left column panels in

Figure 1G show the stimulation volume along the first UOA column as produced by the needle

(column 1, row 8) nearest one of the electrode penetrations (penetration 2 –P2) in the *in vivo*

experiments; the right column panels show the activation volume when all of column 1 is

activated. At low drive voltage (~3V, equal to 38% of the maximum input voltage used), highly

localized stimulation in tissue near the needle tips is produced (note also that the irradiance

across the tip surface is non-uniform – concentrated near the apex – explaining why above-

ChR2-threshold irradiance levels can be achieved at 3V). At higher voltages (≥ 5V/64% max

intensity), the stimulation volume overlaps that of adjacent needles, while also extending deeper

into tissue. When driving an entire column, at 3V, stimulation localized near each tip is mostly

retained, whereas a nearly continuous stimulation volume is obtained at 3.2V due to overlapping

fields. At 5V (64% of max intensity), the depth of this continuous volume increases, both above

and below the tips.

In Vivo **Testing: Electrophysiology**

148   **Figure 2 about here**

We used *in vivo* linear electrode array (LEA) recordings to assess the utility of UOAs for precise

modulation of activity in deep layer neurons expressing ChR2. ChR2 and tdTomato (tdT) were

expressed in macaque V1 via a mixture of Cre-expressing and Cre-dependent adeno-associated

viral vectors (AAV9)[9]. Following a survival period, we recorded multi-unit spiking activity

(MUA) using a 24-contact LEA inserted nearby an active UOA (i.e., fully integrated with a

µLED array, as described in **Fig. 1A-E**) implanted into a region of dense tdT expression (**Fig.**

2A-C; Extended Data Figs. 2-3A). We performed 3 LEA penetrations (P1-P3), but modulation

of neural activity via UOA photostimulation was only detected for P2 and P3 (likely because P1

was farthest from the region of tdT/ChR2 expression; see **Extended Data Fig. 3A**). Below we
report data from P2 and P3.

Comparison of Surface and UOA Photostimulation

Figure 2D shows neural responses recorded in P2 to simultaneous activation of μLEDs at all

UOA sites (whole array condition) at an irradiance level of 0.82 ± 0.49 (SD) mW/mm^2 induced by

an input intensity of 3.5V (see **Extended Data Table 1**) roughly equivalent to ChR2 activation

threshold[31]. To examine the spatiotemporal distribution of responses to UOA stimulation across

V1 layers, we first performed a current source density (CSD) analysis of the local field potential

(LFP) recorded across the LEA around the time of a UOA pulse (see Online Methods). The CSD

reveals the location of current sinks (negative voltage deflections reflecting neuronal

depolarization) and sources (positive voltage deflections reflecting return currents) throughout

the cortical depth. Current sinks and strong phasic MUA in response to UOA stimulation were

mostly localized to layer (L) 4C and the lower part of the deep layers, with L4C activation

leading that in deeper layers (**Fig. 2D**). This suggests that the UOA needle tips closest to P2

terminated in L4C, and that at these low photostimulation intensities light spread nearby the

UOA tips. At the highest intensity, instead, light spread farther into deeper layers (**Extended**

Data Fig. 4A-B). Importantly, this qualitatively distinct laminar pattern of neural activation

could not be explained by thermal artifacts (**Extended Data Figs. 5-6**). Additional analysis

demonstrating that response onset latency and onset reliability were lowest and highest,

respectively, for the P2 contacts located in L4C, together with postmortem histological

assessment, confirmed the UOA needle tips closest to P2 were located in L4C (**Extended Data**

Fig. 3A, B Right). Comparison of the above laminar patterns of response to UOA

photostimulation with that elicited by direct surface photostimulation in a different animal at a

slightly higher irradiance (2.2mW/mm^2) revealed a sharp dissociation. Specifically, surface

stimulation of ChR2 evoked responses starting in superficial layers and terminating in L4C **(Fig.**

2E).

UOA Stimulation Parameters Can Be Tuned to Achieve Laminar Specificity

To assess the impact of UOA stimulation on MUA we varied: (i) the spatial pattern of UOA stimulation, from single μLED sites, to entire columns, to the entire device, and (ii) stimulation intensity across different spatial patterns. In all conditions, we used phasic stimulation (5Hz, 100 msec pulses for 1 sec with 1.5-21 sec inter-trial intervals, with the longer intervals used at the higher stimulation intensities) with a slow on/off ramping to eliminate the potential of any electrical artifacts induced by capacitive coupling at the array/tissue interface[33]. As an example, **Figure 2F-I** shows responses from P2. As indicated by an analysis of firing rate increase across layers induced by activating a single μLED at different sites along column 1, the UOA needles closest to P2 were those in rows 8 and 9 (C1-R8, C1-R9), and their tips terminated into L4C (**Extended Data Fig. 3B Left**). The laminar distribution of MUA in P2 varied in amplitude across conditions, but was reliably confined to deeper layers. By varying the spatial pattern of stimulation and/or the stimulation intensity, MUA could be confined to single layers or spread across multiple layers. For example, activation of the whole UOA (**Fig. 2F**) at intensities >2.8V and up to 5V evoked a MUA peak within L4C (where the needle tips nearest to P2 terminated). This peak increased in magnitude with increasing stimulation intensity. Moreover, at and near the upper end of this intensity range (4-5V), a second, smaller, MUA peak was present in L6 (but not L5). In macaque V1, L4C projects to both L5 and L6[34], but its net effect is to suppress the former[35] and activate the latter[36], consistent with the interpretation that at the higher light intensities lack of L5 responses and increases in L6 responses may have resulted from synaptic spread from optogenetically-activated L4C neurons. Below, we provide evidence supporting this interpretation. At even higher intensities neural activity increased in L4C through L6 likely via direct activation of the deeper layers due to light scattering through a larger volume (**Extended Data Fig. 4C Left**). Although thermal artifacts could not explain the findings at the highest intensity tested with our stimulation parameters (**Extended Data Figs. 5B,6**), lower stimulation intensities should be used for neuroscience applications, particularly when the entire UOA is activated and shorter intertrial intervals are used. This is because heat-induced small increases in firing rates can occur at these higher intensities during the inter-trial period (**Extended Data Fig. 5A**) and potentially affect trial-specific responses at shorter inter-trial intervals than used in our study.

Activation at 5V evoked similar laminar patterns and magnitudes of MUA irrespective of whether a single μLED, an entire column nearest the LEA, or the whole UOA were illuminated

(**Fig. 2F,G,H**). However, at lower photostimulation intensities, firing rate increased with the

number of activated µLEDs (e.g., compare blue curves in **Fig. 2F,G,H**), and higher intensities

(>3.2V) were required to modulate neural activity via a single µLED (**Fig. 2H**). Moving the

µLED activation column a distance of 1.6mm on the UOA (from column 1 to 5) resulted in a 10-

fold reduction in MUA amplitude (**Fig. 2I**), and increases in firing rates in L4C were observed

only at the highest intensities used (7.8V; **Extended data Fig. 4C Right**). No increase in firing

rate could be evoked by activation of an entire column beyond this distance or of a single µLED

in column 1 beyond a similar distance on the UOA (row 4; corresponding to a distance from the

LEA of 2.6-2.7mm estimated on postmortem histology) even at the highest intensity used (7.8V,

Extended Data Table 1).

Tangential Extent of Responses Induced by Photostimulation Via the UOA

229   **Figure 3 about here**

An analysis similar to that performed for P2 allowed us to determine the location of P3 relative

to the UOA, and to establish that µLED C1-R7 was the closest to P3 and its tip terminated in the

superficial layers (**Extended Data Fig. 3C**).

We next asked whether the MUA across LEA contacts was tuned for the spatial site of

UOA stimulation. To estimate MUA selectivity for stimulation at UOA sites between columns 1-

5 and rows 3-10, we fit a multiple linear regression model to the MUA recorded at each LEA

contact, with row, column, and intensity (V) as independent variables (see Online Methods). We

included in this analysis only contacts on which there was a significant difference in firing rates

during the stimulation and control periods for at least one of the row or column conditions

(ANOVA, p<0.01). On average, including a quadratic term explained more of the variance in the

MUA response (mean R^2±SD: 0.58±0.14 vs. 0.31±0.11 for a linear model; Kolmogorov-

Smirnov, $p<10^{-7}$). **Figure 3A, E** shows plots of fitted MUA for 3.5V single-µLED

photostimulation for the contact in P2 and P3 that showed the greatest relative response

modulation. We normalized each contact's fitted responses to the peak, and averaged across

contacts to determine whether MUA preferred stimulation at different UOA sites on different

LEA penetrations (**Fig. 3B,F**). Consistent with our prior assessment (**Extended Data Fig. 3A-**

C), the peaks for P2 contacts tended to cluster mostly near C1-2/R8-9, while those for P3

contacts clustered mostly near C1-3/R4-7. The spatial pattern of peak activity across the LEA

suggested that, particularly for P3, the LEA was inserted at a slightly oblique angle. Peak

locations differed significantly across the two penetrations (ANOVA, p<0.01).

The data in **Figures 2G,I, 3A,B,E,F** indicated MUA amplitude decreased with increasing

distance between photostimulation and recording sites. To quantify this observation, and better

characterize the extent of photostimulation-evoked responses across the tangential domain of

cortex, we examined MUA amplitude as a function of distance on the UOA (in a straight line

extending along either the row or column axis) from the site that evoked the peak response (**Fig.**

3C-H). As is evident from the steeper decrease in responses along the column versus the row

axis, as well as the difference in relative response across stimulus intensities, there was a

significant main effect of UOA axis and input intensity on relative response (ANOVA, both

$p<10^{-21}$), as well as a significant difference across penetrations (ANOVA, $p<10^{-14}$). Finally, there

was a significant interaction between intensity and UOA axis as well as UOA axis and

penetration (ANOVA, both p<0.01). These results indicate that the response decrease from peak

is greater in the column versus the row direction, that intensity has a different effect on this drop-

off in the row versus column directions, and that this differed across penetrations. For example,

in the column direction, at 2.8V intensity MUA dropped to 16% of peak at a distance of 1.6 mm

from peak, but at ≥5V it dropped to 50% at the same distance (**Fig. 3C-G**). Instead, in the row

direction, at 2.8V MUA dropped to 80% of peak at a distance of 2.8mm, and to 90% at ≥5V

(**Fig. 3D,H**). The difference in response drop-off with distance in the column vs. row

directions is likely explained by the greater differences in irradiance, for a given input intensity,

along the column as compared to the row axis (see **Extended Data Fig. 1**).

In summary, the spatial spread of MUA along the tangential domain of cortex varied

according to UOA stimulation site and intensity. Importantly, the extent of this spread was more

limited at lower intensities, suggesting that increasing intensity increased the volume over which

cells were optogenetically activated, consistent with the model simulations in **Fig. 1G**.

UOA Activation Parameters Can Be Tuned to Activate Distinct Cortical Networks

Given the spatial separation between the LEA and the UOA (~1-1.1mm for P2 and 700-800µm

for P3, based on histology; **Extended Data Fig. 3A**), the reported sharp falloff in light intensity

over short distances in tissue[37,38], and our bench estimates of light spread from the UOA tips[28]

(see also **Fig. 1G**), we reasoned that the evoked MUA we recorded was likely relayed to the

recorded neurons indirectly, via activation of ChR2-expressing cells nearby UOA needle tips.

To examine this possibility, we measured the onset latency of evoked MUA across layers.

282   **Figure 4 about here**

Example latency data from P2 are shown in **Figure 4A**. Here, the UOA stimulus was a

single μLED (C1/R8/5V) nearest the recording location. The fastest evoked response occurred in

mid layers with an onset latency (see Online Methods) of about 15ms. Deep layer response onset

(mean±s.e.m: 30±7ms) lagged that in mid-layers, as would be expected if optogenetic activation

first propagated through L4C before being synaptically relayed to deeper layers, via L4C-to-L5/6

connections. Averaged PSTHs for the peri-pulse period on one example L4C and one L6 contact

are shown in **Figure 4B**. There was a significant pulse-by-pulse difference in onset latency

across contacts (ANOVA, $p<10^{-30}$, as well as a significant pairwise difference across these two

LEA recording sites (Tukey HSD test, $p<10^{-6}$; **Fig. 4B Right**).

Figure 4C shows average peri-pulse PSTHs across all LEA contacts as a function of

normalized cortical depth for exemplary whole array (top panels), single column (middle panels),

and single μLED (bottom panels) stimulation at different intensities or μLED-LEA distances.

Increasing total stimulus area at lower intensities (panels in the left column of **Fig. 4C**) increased

the number of responsive contacts and the amplitude of driven responses, and shortened onset

latencies. At higher intensities (5V, middle column), there was little change in these measures

across large differences in total stimulated area. Decreasing the stimulus intensity for a fixed area

(middle to left columns in **Fig. 4C**), or increasing the separation between the stimulated UOA

site/s and the LEA for a fixed stimulus intensity (middle to right panels in the center and bottom

rows of **Fig. 4C**) increased onset latencies across all contacts (mean latency±s.e.m at 5V and

32.V: 17±1.7ms and 25.4±2ms, respectively, whole array condition;   19.8±1.4ms and

37.5±1.9ms, C1 condition; 21.4±2.3ms and 74.1±1.6ms, C1-R8 condition; mean latency±s.e.m at

5V: 47.6±4.3ms and 59.4±4.1ms for C3 and C1-R6 conditions, respectively). Calculating onset

latency on a pulse-by-pulse basis and looking at the effects on latency of cortical depth,

stimulation pattern, and stimulation intensity, we observed significant main effects of pattern and

intensity, as well as significant two-way and three-way interactions between all three factors

(ANOVA, all $p<10^{-4}$). Limiting our analysis to each pattern, we observed a significant main

effect on latency of intensity and distance from the LEA for the single column conditions in **Fig.**

4C (ANOVA, all $p<10^{-4}$), and a significant main effect of distance for the single µLED

conditions (ANOVA, p=0.03). Furthermore, in many conditions, pairwise comparisons across

contacts revealed a significantly delayed response onset in deep layers relative to mid-layers for

most conditions in **Figure 4C** at 5V, and for some conditions at 3.2V (Tukey HSD, all p< 0.01;

Extended Data Fig. 7); this time-lag varied with intensity and separation between stimulation

and recording sites, increasing at lower intensities and greater distances. There was also a

significant difference in onset latency between mid- and superficial layers in some conditions

(C1 at 5V, whole array at 5V and 3.2V; Tukey HSD, all p<0.01; **Extended Data Fig. 7**).

Notably, however, when the whole µLED array was stimulated at the highest intensity (7.8V),

there was no significant difference in onset latencies between deep and middle layers, again

suggesting the former were directly activated by light spreading through deeper tissue (**Extended**

Data Figs. 4D and 7).

323　 **Figure 5 about here**

To quantify these effects across the population (n= 33 significantly responsive contacts,

across 2 LEA penetrations), we first calculated the distance between each LEA contact and the

contact with the shortest onset latency, and plotted this distance versus onset latency, separately

for each unique combination of UOA stimulation site(s) and intensity. Similar to the P2 data

shown in **Fig. 4C**, the population data showed 2 main effects. (1) Onset latency decreased

significantly across all contacts with increasing stimulation intensity (ANOVA, main effect of

intensity, all p<0.01; **Fig. 5A, 5B Left, 5C Left**) and proximity to the recording LEA site

(ANOVA, main effect of row or column on UOA, all $p<10^{-4}$; right panels in **Fig. 5B and 5C**).

(2) Onset latency increased significantly with contact distance on the LEA from the fastest

contact (**Fig. 5A-C**, main effect of distance on the LEA, ANOVA all p<0.01), suggesting that the

more distant contacts were activated indirectly via interlaminar networks. However, for

stimulation of the whole UOA at higher intensity (7.8V), evoked responses had similar onset

latencies across the LEA (thus, across V1 layers; **Extended Data Figs. 4E,7** top right).

Across the three categories of UOA stimulation (whole array, column, and single µLED),

only for the whole array and single µLED conditions did we observe a significant interaction

between the effects of distance along the LEA and UOA photostimulation intensity on onset

latency (**Fig. 5A, 5C Left**; both p<0.05, ANOVA). In these conditions, lowering

photostimulation intensity decreased the slope of the curves, indicating that the difference in

onset latency with distance on the LEA increased at lower intensity. Additionally, for the single

μLED condition, we also observed a significant decrease in the slope of the curves when

photostimulating at increasing UOA-LEA separation, but only when we moved the single μLED

stimulus to sites that were far enough from the LEA to necessitate stimulation at the very highest

powers used to elicit any response (dashed lines in **Fig. 5C Right**, μLED in rows 4-7; ANOVA,

LEA distance × UOA row × intensity interaction, $<10^{-3}$). For the single column condition, there

was no significant interaction between contact distance and either photostimulation intensity or

UOA-LEA separation (**Fig. 5B**; ANOVA, all p>0.09). Importantly, across all three

photostimulation patterns (whole array, single columns, and single μLEDs) there was remarkable

similarity in the timing of the fastest responses (**Fig. 5D**). Both increasing stimulus area and

stimulating at UOA sites closer to the recording locations reduced the light intensity necessary to

evoke responses at this latency, but did not result in shorter latencies. This is further evidence

that the evoked MUA nearby LEA contacts was relayed indirectly following optogenetic

activation at UOA tips, and that the timing of this activation depended upon both the location

and area of optogenetically-activated inputs.

In summary, by varying photostimulation intensity and/or number of stimulated sites, the

UOA allows activation of single or multiple layers, while by varying the spatial separation

between the site of UOA stimulation and that of the recording, the UOA allows investigations of

local vs long-range intra and interlaminar circuits.

In Vivo **Testing: c-Fos Expression**

365 　　　　　　　　　　　　　　　　　　　　　　　　　　　　　　　　**Figure 6 about here**

To validate the performance of the UOA for large-scale photostimulation, we measured changes

in c-fos expression, an immediate early gene whose expression rapidly increases when neurons

are stressed or activated[39,40]. C-fos protein expression can be used as an indirect measure of the

spatial pattern of neural activation. We analyzed patterns of c-fos expression using

immunohistochemistry (IHC) (see Online Methods) in two control and two experimental

hemispheres from 3 animals.

In one experimental case (MK414-RH), a "passive" UOA (lacking an integrated μLED array)

was implanted in a ChR2/tdT-expressing region of V1 (**Fig. 6A-B**). We photostimulated the deep

layers through a subset of needles, using a collimated, fiber-coupled, 473nm laser, while

shielding from light surrounding cortex and portions of the UOA (see Online Methods).

Histological analysis revealed that the UOA in this case was inserted at an angle (due to brain

curvature), its needle tips ending at the bottom of the superficial layers, anteriorly, and in

progressively deeper layers, posteriorly (most tips being in L4C, only the most ventral ones

reaching L6) (**Fig. 6A-B**). C-fos positive (c-fos+) cells were found throughout V1 (**Fig. 6A,C,D**),

as well as in V1 recipient extrastriate areas, including V2 (**Fig. 6A,C,D**), V3, and MT (not

shown)). This extensive pattern of elevated c-fos expression was likely induced by direct

optogenetic activation and indirectly via synaptic activation. To test this hypothesis, we repeated

the experiment in a different animal (MK422-RH) in which we greatly reduced glutamatergic

neurotransmission via application of the AMPA receptor antagonist NBQX to ChR2-expressing

cortex prior to passive-UOA insertion and photostimulation. Most of the UOA's needle tips, in

this case, only reached the bottom of the superficial layers (**Fig. 6E-F**). We also performed two

additional experiments, to control for the potential of elevated c-fos expression being induced by

either UOA insertion or stray photostimulation, respectively. In case MK414-LH, we inserted a

passive UOA in the supplementary motor area (SMA) not expressing ChR2, and euthanized the

animal 4 hours later without photostimulating. Histological analysis revealed that the UOA was

fully inserted in this case (tips reaching L5; **Fig. 6I**). In case MK421-RH, instead, we only

performed surface photostimulation of SMA cortex not expressing ChR2 and no UOA insertion

(**Fig. 6K**).

To quantify c-fos expression across our various manipulations, we counted c-fos+ cells in

3 regions of interest (ROIs) encompassing all cortical layers, one centered in the region of UOA

insertion and/or light stimulation, the other two located 4 and 8 mm, respectively, from the first

(*white boxes* numbered 1-3 in **Fig. 6A-L**; see Online Methods). **Figure 6M** plots the average

number of c-fos+ cells across samples, as a function of distance from the UOA insertion site,

while **Figure 6N** shows the laminar distributions of c-fos+ neurons at each distance. We found

significant local (involving all layers) and long-range c-fos expression only when

photostimulation of ChR2-expressing cortex was performed via the UOA (MK414-RH; **Fig.**

6C-D, M-N). Blocking glutamate neurotransmission prior to photostimulation prevented long-

range c-fos expression, and reduced its expression by 5 fold in the area of UOA stimulation,

where it was largely confined to the directly photostimulated layers (mostly superficial) near the

UOA tips (MK422-RH; **Fig. 6G-H,M-N**). UOA insertion-only led to as much local c-fos

expression as the glutamate block case, but to greater interlaminar (involving all layers), as well

as intra- and inter-areal long-range spread (MK414-LH; **Fig. 6J,M-N**), suggesting that neurons

activated by the insertion trauma also indirectly activated downstream networks. Finally, surface

photostimulation of cortex not-expressing ChR2, without UAO insertion, caused virtually no c-

fos expression, except for a few cells in L1 and upper L2 (MK421-RH; **Fig. 6L-N**). Statistical

analysis (one way ANOVA with Bonferroni corrected post hoc comparisons) revealed a

significant difference in the number of c-fos+ cells at each distance between the full

experimental case (MK414-RH) and all others ($p < 0.001$ at all distances for all pairwise

comparisons). There was no significant difference between the glutamate-block and UOA-

insertion-only cases at any distance ($p > 0.23$ at all distances), and both these cases differed

significantly from the light-only case at 0mm distance ($p < 0.05$ for all comparisons). Finally, the

number of c-fos+ cells decreased significantly with distance for cases MK414-RH ($p < 0.001$),

MK422-RH ($p = 0.001$), and MK414-LH ($p = 0.003$), but not for case MK421-RH ($p = 0.079$).

421   **DISCUSSION**

We have developed and validated a novel device, the UOA, which has the potential to

further optogenetic research in NHPs. Current optogenetic approaches in NHPs permit light

delivery either over a large superficial area[9,23], or to deeper tissue but over a small area[25-27,38].

Multi-site probes for larger volume stimulation have also been developed, and combined with

single[41] or multisite[42,43] electrical recordings, but these approaches are typically cumbersome to

assemble and don't easily scale to precisely target multiple small tissue volumes. The UOA

combines the advantages of all these approaches. It allows for both focal and larger-scale

neuronal activation of single or multiple deep layers simply by varying the number of

simultaneously activated µLEDs and/or the light irradiance. Moreover, although here we only

used the needle-aligned µLED array for deeper layer activation, the integrated interleaved

interstitial μLED array allows for selective photostimulation of superficial layers, either
independently or in conjunction with deep layers.

By design, the UOA is intended to achieve spatial resolution in cortical application in
NHPs, and eventually humans, and is, thus, ideal for addressing neuroscience questions that
require large-scale manipulations of deep and/or superficial cortical layers. Even though it
currently lacks recording capability, here we have demonstrated that the UOA, used as a
stimulation-only device in conjunction with LEA recordings, can be used to study inter-laminar
interactions. We were able to localize photostimulation to single or multiple cortical layers by
varying light intensity. Similarly, varying insertion depth (or shank length) offer the possibility
to select targeted layers. Relative differences in onset latency of evoked responses could be used
to distinguish distinct network activity patterns following different patterns of UOA stimulation.
For example, at low light irradiance, direct neuronal activation was initially localized to layers
nearest optrode tip termination before spreading trans-synaptically to other layers. Increasing
light irradiance reduced or eliminated these latency differences. Similarly, firing rates in L4C
increased less at higher versus lower intensities, suggesting response amplitude can be used to
identify local activation of higher threshold inhibitory networks.

We showed that by varying the distance between the stimulation site/s on the UOA and
the recording electrode, local versus long-distance intra-areal interactions can be studied.
Moreover, used in conjunction with c-fos IHC, we were able to identify multisynaptic
interactions within and beyond the photostimulated area. Photostimulation via the UOA
increased c-fos expression over distances much > 8mm (well beyond the stimulated cortical
area), but spiking activity could not be evoked beyond ~3 mm from the stimulated site,
indicating c-fos expression revealed subthreshold activity induced by network interactions. This
is consistent with previous demonstrations of c-fos expression several synapses away from an
electrically stimulated site. Thus the UOA in conjunction with c-fos IHC can be used for
functional mapping of neuronal circuits[39].

We also investigated whether our results could have been affected by local increase in
brain temperature caused by the μLEDs heating up when activated. This concern arises with
implantable devices[44] both in terms of temperature-induced tissue damage[45] and changes in
spiking activity[24,46]. It is generally assumed that tissue damage is negligible for temperature
increases < 1°C[29,47]. One difference of the UOA compared to other implantable μLED devices is

that the heat-generating μLEDs are mounted on the topside of the device and external to tissue,

compensating for the fact that the low optical coupling efficiency requires higher drive currents

than for optogenetic devices based upon embedded μLEDs on implantable shanks[44]. Detailed

thermal simulations showed that the intervening thermally-insulating layers of dura-gel and brain

tissue (combined thickness ~1.5 mm) caused a ~1 second delay in the temperature ramp at the

stimulation site in L4, so that the bulk of the temperature rise (and subsequent fall) occurred

during the inter-trial interval and not during the trial period. These simulations also showed that

peak temperature rise can be held below 1°C. Additional analysis of spiking rates during the

inter-trial interval showed some modulation from background activity, which could be

temperature mediated, but only when the whole array was activated at the highest intensity; and

even for this condition, spiking activity had returned to baseline by the end of the inter-trial

interval prior to the next trial. These results strongly suggests that our results were not affected

by thermal increases. However, additional *in vitro*, *in vivo* and *in silico* studies are planned to

assess, and minimize, temperature increases in tissue.

Future applications, beyond what shown here, could involve functional investigations of

inter-areal circuits, when UOA stimulation in one cortical area is coupled with recordings in a

different area. Importantly, despite its limited shank length (2.5 mm max), the UOA can also be

employed to study cortico-subcortical interactions, e.g., through modulation of axon terminals of

deep nuclei within cortex, and recordings of postsynaptic cortical neurons in the same cortical

area and/or layer.

In conclusion, the UOA as currently conceived, and especially following planned

addition of multi-color functionality and dual stimulation/recording capability, will enable

studies addressing fundamental questions in neuroscience, e.g., regarding the role of cortico-

cortical feedback and cortical layers in the model system closest to humans. As many human

neurological and psychiatric disorders have been linked to abnormalities in cortical circuits[4,5],

this technology can improve our understanding of the circuit-level basis of human brain

disorders, and will pave the way for a new generation of precise neurological and psychiatric

therapeutic interventions via cell type-specific optical neural control prosthetics.

494   **ONLINE METHODS**

Device Fabrication, Characterization, and Benchmarking

Fabrication and testing of the first generation UOA devices was previously reported[28,48]. The second-generation devices used in this study included an optical interposer layer that limits emission from the µLED array to the shank sites for illumination of deep cortical tissue.

Fabrication. A 2 mm-thick, 100mm diameter Schott Borofloat 33 glass wafer used to construct the optrode needles was anodically bonded to a freshly cleaned 0.1mm thick, 100 mm diameter intrinsic Si wafer serving as an optical interposer. The Si and Borofloat wafers were coarsely aligned, and bonding performed using an EVG 520 anodic bonder. The optical vias were patterned in the Si interposer by deep reactive ion etching (DRIE) using a Bosch process. A 10-µm-thick AZ9260 soft mask was photolithographically patterned to define the array of 80×80 µm^2 optical vias for shank and interstitial illumination for the DRIE process. The bonded wafer was then sub-diced into *modules* of 9 to 16 UOAs using a DISCO 3220 dicing saw.

UOA modules were mounted to a carrier wafer using WaferGrip™ (Dynatex International, Santa Rosa, CA). The glass shanks were cut with the DISCO 3220 using the previously reported process[28,48]. Briefly, beveled blades were first used to generate pyramidal tips on the surface, followed by standard profile blades to form the shanks. The shanks on a module were then etched to a nominal 110 µm thickness using a mixture of hydrofluoric (49%) and hydrochloric (37%) acid in a 9:1 ratio. The die was then demounted and cleaned, and the shanks were smoothened to decrease light scattering using a 725 °C heat treatment for 2 hours in a vacuum furnace. UOA modules were then singulated into individual 4×4 mm^2 UOAs using the DISCO 3220.

Arrays of µLEDs on thinned (150µm) sapphire substrates, from the Institute of Photonics at University of Strathclyde, were integrated with the UOA using closed-loop optical alignment to the optical vias on individual UOAs at Fraunhofer IZM (Berlin, Germany)[28], and bonded using index-matched epoxy. At the University of Utah, passive matrix µLED pads were wire bonded to an ICS-96 connector (Blackrock Microsystems, Salt Lake City, UT) using insulated gold alloy wire. The wire bundle and back-side of the UOA were then potted in NuSil MED-4211 silicone, respectively, followed by overcoating with a 6µm-layer of Parylene C.

Bench Testing. To characterize the electrical and optical performance of the finalized devices, the latter were attached to a custom switch board for matrix addressing the individual

optrode shanks. The switch board consisted of a matrix arrangement of parallel connected mechanical switches and electrical relays, 10 sets for the anodes and 10 sets for the cathodes. This enabled both manual and automated activation of individual optrode shanks or optrode patterns. For the automated activation and testing, the relays were connected to Arduino boards which received commands from the lab computer. To prevent voltage spikes originating from the switching of the channels from damaging the µLEDs, the anode paths also contained a small filter circuit consisting of capacitors and Zehner diodes (break-down voltage: 8.2V). For the automated testing, the UOAs were inserted into the opening of an integrating sphere that was, in turn, connected to a photodetector and power meter (Newport 2832-C Dual-Channel Power Meter). The calibration factor of the integrating sphere was determined using a fiber coupled LED prior to the experiment. Then the UOAs were connected to the switch board, and the latter was connected to a source measure unit (Keithley 236 Source Measure Unit) for the measurement. The automated characterization was conducted as follows: the switch board's Arduino boards received the command to switch to an individual optrode shank using the relays. Then the source measure unit applied a voltage pulse measurement pattern (pulse length 100ms, pause between pulses 1900ms to prevent heat buildup) sweeping the voltage from 0 to 7.2V (or until the compliance current of 100mA was reached) with each pulse increasing by 100mV. For each pulse, the resulting current and the output optical power were recorded; the optical power was then corrected using the integrating sphere calibration factor. This was repeated for each individual optrode shank of the device for a full characterization.

To ensure the stability of the device for an acute *in vivo* experiment, additional voltage transient measurements were made before and after a 48-hour soak test in phosphate-buffered saline (PBS) at 37 °C. Further, an electrode was immersed in solution to verify encapsulation integrity, as evidenced by lack of shorting to solution.

For the *in vivo* experiments, the switch board was upgraded two-fold: first, transistors were added to the cathode channels to allow for turning the device on and off based on an external TTL trigger. However, we found that turning on the optrodes using the trigger signal directly induced too strong a capacitively-coupled voltage signal in the recording. Therefore, as a second upgrade, an additional Arduino board with digital-analog-converter was added that received the external trigger and introduced rise and fall times to the square wave. This reduced the capacitively-coupled interference to a level below measurable when both the LEA and the

557 UOA were in close proximity in 1xPBS solution prior to the *in vivo* experiment. During the

558 experiment, the voltage for the UOA was supplied by a lab power supply via the switch board,

559 and the switches were operated manually to define the required patterns.

560    *Modeling.* To understand light spread in tissue, the optical output of the device was

modeled using ray-tracing software (Optics Studio 12, in non-sequential mode). This model has

been described previously[28]. Brain tissue was modeled using a Henyey-Greenstein scattering

model, with a scattering coefficient of 10 mm^{-1}, absorption coefficient of 0.07 mm^{-1}, and

anisotropy of 0.88[47]. Each needle was modelled individually using its measured optical output at

the given voltage level. To generate the cross-section images from a simultaneously illuminated

column (**Fig. 1G**), the light output from the 10 needles in that column were summed.

Animals

A total of 3 adult female Cynomolgus monkeys (*Macaca fascicularis*) were used in this study.

The left hemisphere of one animal (case MK421-LH) was used for the *in vivo*

electrophysiological testing of the active UOA (integrated with the µLED array). The right

hemisphere from the same animal (MK42-RH), and 3 hemispheres from 2 additional animals

(MK414RH and LH, and MK422-RH) were used for c-fos testing of the passive UOA (i.e.,

without an integrated µLED array). All procedures conformed to the National Institutes of

Health Guide for the Care and Use of Laboratory Animals and were approved by the University

of Utah Institutional Animal Care and Use Committee.

Survival Surgical Procedures and Viral Injections

Animals were pre-anesthetized with ketamine (10 mg/kg, i.m.), intubated, placed in a

stereotaxic apparatus, and artificially ventilated. Anesthesia was maintained with isoflurane (1–

2.5% in 100% oxygen). Heart rate, end tidal CO_2, oxygen saturation, electrocardiogram, and

body temperature were monitored continuously. I.V. fluids were delivered at a rate of 3-

5/cc/kg/hr. The scalp was incised and a craniotomy and durotomy were performed over area V1

(n=2 animals, MK421-LH and MK414-RH), or rostral to the precentral gyrus, roughly above the

supplementary motor area (SMA; n=1, MK422-RH). We injected a 1:1 viral mixture of

AAV9.CamKII.4.Cre.SV40 and AAV9.CAG.Flex.ChR2.tdTomato (Addgene Catalog #s:

105558, and 18917, respectively). We have previously found that this method nearly eliminates

retrograde expression of transgenes[9]. The viral mixture was slowly (~15nl/min) pressure-injected

(250-350nl repeated at 2 or 3 cortical depths between 0.5 and 1.5 mm from the cortical surface)

using a picospritzer (World Precision Instruments, FL, USA) and glass micropipettes (35-45µm

tip diameter). After each injection, the pipette was left in place for 5-10 min before retracting, to

avoid backflow of solution. A total of 5-6 such injections, each 500-750nl in total volume, and

spaced 1.5-2mm apart, were made in two animals (MK421-LH,MK414-RH) while the third

animal (MK422-RH) received 2 x 1,050nl injections. These injections resulted in a region of

high viral expression roughly 4-6 mm in diameter (as an example see **Extended Data Fig. 3A**

Right). Following viral injections, a sterile silicone artificial dura was placed on the cortex, the

native dura was sutured and glued onto the artificial dura, covered with Gelfoam to fill the

craniotomy, and the latter was sealed with sterile parafilm and dental acrylic. Anesthesia was

discontinued and the animal returned to its home cage. After a survival period of 5-10 weeks, to

allow for robust ChR2 expression, the animals were prepared for a terminal UOA

photostimulation procedure.

Terminal Surgical Procedures and UOA Insertion

Monkeys were pre-anesthetized and prepared for experiments as described above. Anesthesia

and paralysis were maintained by continuous infusion of sufentanil citrate (5–10□µg/kg/h) and

vecuronium bromide (0.3□mg/kg/h), respectively. Vital signs were continuously monitored for

the duration of the experiment, as described above. Following suture removal and scalp incision,

the craniotomy and durotomy were enlarged to allow space for device implantation, and ChR2

expression was verified *in* vivo using a custom fluorescent surgical microscope (Carl Zeiss,

GmbH; **Fig. 2B**). UOAs were positioned over cortical regions of high tdT/ChR2 expression (e.g.

Figs. 2B,6B,F), and then inserted using a high speed pneumatic hammer typically used for

insertion of Utah Electrode Arrays[34] (Blackrock MicroSystems, UT). Parameters used for

insertion were 20 psi for 30ms, using a 1 mm-long inserter, in order to achieve partial insertion

of the UOA, so as to minimize insertion trauma on the cortex. In two animals used for c-fos

experiments after partial insertion with the pneumatic inserter, the UOA was gently pushed down

to achieve deeper insertion.

Photostimulation

We implanted two types of UOA devices: (i) a 10x10 UOA with fully integrated µLED arrays (also referred to as "active" device; n=1 device in 1 animal, MK421-LH; see **Fig. 2A-C**), and (ii): 10x10 UOAs with an optical interposer integrated into the sapphire backplane, but with no µLED array for light delivery (referred to as "passive" devices; n=3 devices in 3 hemispheres from 2 animals, MK414-RH, MK414-LH, MK422-RH). The active device was used for electrophysiological testing experiments, while the passive devices were used for the c-fos experiments.

Active Device (Electrophysiology). Photostimulation with the active UOA occurred via the integrated µLED array. Photostimulation parameters were 5Hz, 100 msec-pulse duration for 1 sec, followed by 1.5-21sec inter-trial interval (longer intervals were used at the higher photostimulation intensities). We varied the spatial pattern (single µLED along column 1, whole single columns, and all µLEDs across the entire UOA) and intensity (from 2.8 to 7.8V input intensity) of photostimulation as described in the Results section.

Passive Devices (c-Fos). Selective photostimulation via passive devices was obtained by illuminating a subset of UOA needles with an appropriately positioned fiber-coupled 473nm laser (400 µm multimode optic fiber, ThorLabs Newton, NJ; laser: Laserwave, Beijing, China) held in place with a stereotaxic tower. We used a collimating lens (ThorLabs, Newton, NJ) to restrict spot size to ~1.5mm in diameter. To shield stray light, we covered any exposed tissue around the illuminated area, as well as the non-illuminated portions of the UOA, with an opaque (black) artificial dura. For each UOA we stimulated 2 or 3 separate sites. At each site we used phasic photostimulation (50Hz for 2.5 min, 2.5 min pause, and 20Hz for an additional 2.5 min; pulse duration was 10ms) at 3.8mW power output (corresponding to an estimated irradiance of 15-19mW/mm^2).

Electrophysiological Recordings

Extracellular recordings were made in V1 with 24-channel linear electrode arrays (LEAs; V-Probe, Plexon, Dallas, TX; 100µm contact spacing, 300µm from tip to first contact, 20µm contact diameter). The LEAs were inserted into the cortex next to the UOA to a depth of 2.4-2.6mm, slightly angled laterally (towards the UOA) and posteriorly. We made a total of 3 penetrations (P1-P3; **Extended Data Fig. 3A**), of which only P2 and P3 provided useful data.

After UOA and LEA were inserted into the cortex, we applied a layer of Dura-Gel

(CambridgeNeuroTech, Cambridge, UK) over the cortex and UOA, to prevent the cortex from

drying and stabilize the recordings. A 128-channel recording system (Cerebus, Blackrock

Microsystems, Salt Lake City, UT) was used for standard signal amplification and filtering.

Multi-unit spiking activity was defined as any signal deflection that exceeded a voltage threshold

(set at 4 x the SD of the signal on each channel). Threshold crossings were timestamped with

sub-millisecond accuracy. We did not record responses to visual stimuli but only to UOA

photostimulation performed as described above; thus, the monkey's eyes were closed during the

duration of the experiment.

Analysis of Electrophysiological Data

We analyzed MUA responses from a total of 45 contacts deemed to lie within the parafoveal

representation of V1 in two penetrations (out of 3 total, see above) for which neural activity was

modulated by photostimulation via the active UOA. For the results presented in **Figures 3-5**,

quantitative analysis was limited to contacts on which MUA was stimulus modulated (one-way

ANOVA comparing spike rates during full one-second photostimulation trials with spike rates

during control periods of equivalent duration, $p < 0.01$).

To quantify the change in MUA firing rates, relative to background, during

photostimulation we calculated firing rates for all pulse epochs within all trials and then

compared them to the average background rate. To estimate the preference at each recording site

for stimulation across the full range of tested UOA locations (**Fig. 3**), we regressed average

evoked-responses on UOA stimulation site and intensity. Preliminary analyses had revealed a

non-monotonic relationship between stimulation intensity and response on many contacts (cf.

Fig. 2F), thus we included a quadratic term in the regression model.

CSD analysis. For the CSD analysis shown in **Fig. 2D-E,** current source density (CSD)

was calculated from the band-pass filtered (1-100Hz) and pulse-aligned and averaged LFP, using

the kernel CSD toolbox (kCSD_Matlab)[49]. CSD was calculated as the second spatial derivative

of the LFP signal, reflecting the net local transmembrane currents generating the LFP. The depth

profile of the CSD was estimated by interpolating every 10μm. To facilitate comparisons across

conditions, CSDs from different conditions were normalized to the standard deviation (SD) of

the baseline (50ms prior to pulse onset) after subtraction of the baseline mean.

Onset Latency. To quantify the onset latency of MUA responses, we either: (i) calculated the average peri-stimulus time histogram (PSTH) from all pulse-aligned responses (e.g., **Fig. 4**) or (ii) estimated a PSTH separately for the response to each pulse (**e.g., Extended Data Fig. 7**). Peristimulus time histograms (PSTHs) were estimated via an adaptive algorithm in which the MUA raster was first convolved with a Gaussian kernel of fixed width (3ms bandwidth), kernel width was then adapted so that the number of spikes falling under the kernel was the same on average across the response (http://chronux.org[50]). We then subtracted the mean baseline response from the stimulus-evoked response. For each response measure, i.e., either the average or pulse-by-pulse PSTHs, we took the time at which the response reached 25% of the peak as the onset latency (results were qualitatively similar using 15% and 35% criteria; data not shown). We report the former measure as the mean onset latency in **Figures 4-5**. We used the latter measure to test for differences in onset latency across contacts within and across UOA stimulation parameters (**Figs. 4-5 and Extended Data Fig. 7**).

Statistical Analysis. Stimulus-evoked firing rates were calculated from pulse-aligned or trial-aligned responses and baseline corrected (mean baseline activity subtracted). We determined responsiveness to stimulation via a one-way ANOVA comparing firing rates during the full 1-second trial period with inter-leaved control periods of equivalent duration; MUA at an LEA recording site was deemed responsive if there was a significant difference between stimulation and control trials at the $p=0.01$ level. To estimate the selectivity of MUA for stimulation at different UOA sites we performed a multiple linear regression, with UOA column, row, and intensity as independent variables and pulse-aligned, baseline corrected, firing rates as the dependent measure. To test for differences in the goodness-of-fit of models with- and without a quadratic term, we used a two-sample Kolmogorov-Smirnov test. We assessed the effects of varying UOA stimulation site and intensity on response amplitude or onset latency using ANOVA models followed by the Tukey-Kramer test for post-hoc comparisons.

c-Fos Experiments

We used 4 hemispheres from 3 animals for these experiments (MK414-RH and LH, MK422-RH, and MK421-RH). Two of these animals (MK422 and MK414) were prepared for a terminal experiment (as described above) 5 or 10 weeks, respectively, after the viral injections, and a passive UOA was inserted in regions of high tdT/ChR2 expression in the injected hemisphere. In

one of these animals (MK422-RH), UOA insertion was preceded by glutamate block (see

below). After UOA insertion, photostimulation was performed via an optical fiber-coupled laser

through the UOA, as described above. Two additional hemispheres in 2 animals (MK414-LH

and MK421-RH) were used as controls. Specifically, case MK414-LH received insertion of a

passive UOA in non-opsin expressing SMA cortex, and was euthanized 4 hrs following UOA

insertion without receiving any photostimulation. As a separate control, in case MK421-RH we

performed surface photostimulation of SMA cortex not expressing opsins, using a fiber-coupled

laser and a collimating lens and the same photostimulation protocol described above for other c-

fos experiments; no UOA was inserted in this case. In all animals, UOA insertion and/or

photostimulation were performed after a 10-14-hour period of eye closure and at least 5 hours

after completion of surgery, and the animals were euthanized 75 minutes after completion of the

photostimulation protocol.

Pharmacological Blockade of Local Glutamate Signaling. To compare changes in c-fos

expression due to direct local optogenetic activation with indirect local and long-range changes

due to synaptic increases in excitatory glutamatergic neurotransmission downstream of the

directly-activated neurons, in one case (MK422-RH) we applied the selective glutamate AMPA

receptor antagonist 2,3-dihydroxy-6-nitro-7-sulfamoyl-benzoquinoxaline-2,3-dione (NBQX,

5mM) (Tocris BioSciences). NBQX was applied topically prior to UOA insertion, by soaking a

piece of Gelfoam placed over ChR2-expressing SMA cortex with 1ml of the drug solution. The

drug was allowed to passively diffuse through the cortical layers for 90 minutes, during which

100-200µl of the solution were applied every 15 minutes to ensure saturation of the Gelfoam,

after which the Gelfoam was removed and the passive UOA inserted over the region of

glutamate block. Photostimulation was performed as described above for the passive device.

Histology

On completion of the experiments, the animals were euthanized by an overdose of Beuthanasia

(0.22 ml/kg, i.v.) and perfused transcardially with saline for 2–3 min, followed by 4%

paraformaldehyde (PFA) in 0.1M phosphate buffer for 20 min to fix the brain. The brains were

post-fixed overnight in the same fixative, sunk in cryoprotectant 30% sucrose solution, and

sectioned at 40µm on a freezing microtome. The hemisphere used for electrophysiological

testing of the active UOA (MK421-LH) was sectioned tangentially. One in 3 sections were wet-

mounted and imaged for fluorescent tdT-label at 10x magnification. The same sections were then reacted for cytochrome oxidase (CO) to reveal cortical layers and the location of UOA and LEA insertions visible as discolorations in CO staining (**Extended Data Fig. 3A Left**).

All other hemispheres used for c-fos experiments were sectioned sagittally. One full series of sections (1:3) were immunoreacted for c-fos by overnight incubation in primary antibody (1:500 rabbit anti-c-fos, Ab 19089, Abcam, MA, USA) at room temperature, followed by 2 hrs incubation in near-infrared secondary antibody (1:200 donkey anti-rabbit IgG-AF647, Jackson ImmunoResearch, PA, USA) at room temperature. Sections were then wet-mounted, counterstained with blue fluorescent Nissl (1:100 N21479, Thermo Fisher Scientific, MA, USA), by dripping the solution onto the slide-mounted sections every 5 min for 20 min, rinsed, and coverslipped and sealed with CoverGrip™ Coverslip Sealant (Biodium, CA, USA).

Tissue Imaging

Imaging of tissue sections was performed on a Zeiss Axio Imager.Z2 fluorescent microscope (Zeiss, Germany) with a Zeiss X-cite 120 LED Boost light source, using a 10x objective and an Axiocam 506 mono camera (Zeiss, Germany). Image files were created and analyzed using Zen 2.6 Blue Software (Zeiss, Germany). The light intensity was set to 100%, and the exposure time for each channel was kept the same between images. The tangentially-sectioned hemisphere (MK421-LH) was imaged as described above. In all other cases, each sagittal section was imaged in 3 channels simultaneously, one channel for tdT/ChR2 (red- but note the color was artificially changed to green in **Fig. 6B,F**), one channel for Alexa-647-c-Fos (far-red), and the third channel for 435-455 Nissl (blue).

Analysis of c-Fos Expression

To quantify c-fos expression, c-fos+ cells were plotted and counted in sampled areas, using Neurolucida software 2006 (Microbrightfield Bioscience, VT, USA). For each case, we selected for counts 5 sections spaced 1 mm apart encompassing the area of UOA insertion and/or photostimulation (for the light-only case). In each section, we plotted and counted cells within three 200μm-wide windows spanning all cortical layers, one positioned at or near the center of the UOA insertion region (or of phtostimulation-only), and the other two located at distances of 4mm and 8mm, respectively, from the center of the UO insertion (**Fig. 6**). Thus, a total of 15

773   regions of interest (ROIs) were counted for each case. The laminar distribution of c-fos+ cells

774   was analyzed by tracing the layers on the Nissl stain and counting the number of c-fos+ cells

775   within each layer in Neurolucida. Statistical differences in c-fos+ cell counts among

776   experimental and control cases, and across distances were estimated using a one-way ANOVA

777   with Bonferroni corrected post hoc comparisons.

REFERENCES

1.   Deisseroth, K. Optogenetics: 10 years of microbial opsins in neuroscience. *Nat. Neurosci.*
18, 1213-1225 (2015).

2.   El-Shamayleh, Y. & Horwitz, G.D. Primate optogenetics: Progress and prognosis. *Proc.*
Natl. Acad. Sci. U.S.A. 116, 26195-26203 (2019).

3.   Goldberg, M.E. The neurology clinic needs monkey research. *Proc. Natl. Acad. Sci.*
U.S.A. 116, 26255-26258 (2019).

4.   Lynall, M.-E. et al. Functional Connectivity and Brain Networks in Schizophrenia. *J.*
Neurosci. 30, 9477-9487 (2010).

5.   Belmonte, M.K. et al. Autism and Abnormal Development of Brain Connectivity. *J.*
Neurosci. 24, 9228-9231 (2004).

6.   Vitek, J.L. & Johnson, L.A. Understanding Parkinson's disease and deep brain
stimulation: Role of monkey models. *Proc. Natl. Acad. Sci. U.S.A.* 116, 26259-26265
(2019).

7.   Picaud, S. et al. The primate model for understanding and restoring vision. *Proc. Natl.*
Acad. Sci. U.S.A. 116, 26280-26287 (2019).

8.   Jarvis, S. & Schultz, S.R. Prospects for Optogenetic Augmentation of Brain Function.
Front. Syst. Neurosci. 9, 157 (2015).

9.   Nurminen, L., Merlin, S., Bijanzadeh, M., Federer, F. & Angelucci, A. Top-down
feedback controls spatial summation and response amplitude in primate visual cortex.
Nature Commun. 9, 2281 (2018).

10.   Siu, C., Balsor, J., Merlin, S., Federer, F. & Angelucci, A. A direct interareal feedback-
to-feedforward circuit in primate visual cortex. *Nat Commun* 12, 4911 (2021).

11.   Mehta, P. et al. Functional Access to Neuron Subclasses in Rodent and Primate
Forebrain. *Cell Rep.* 26, 2818-2832 e2818 (2019).

12.   Vormstein-Schneider, D.C. et al. Viral manipulation of functionally distinct interneurons
in mice, non-human primates and humans. *Nat. Neurosci.* 23, 1629-1636 (2020).

13.   Dimidschstein, J. et al. A viral strategy for targeting and manipulating interneurons
across vertebrate species. *Nat. Neurosci.* 19, 1743-1749 (2016).

14.   Tremblay, S. et al. An Open Resource for Non-human Primate Optogenetics. *Neuron*
108, 1075-1090 e1076 (2020).

15.   Angelucci, A. et al. Circuits and mechanisms for surround modulation in visual cortex.
Ann. Rev. Neurosci. 40, 425-451 (2017).

16.   Moore, T. & Zirnsak, M. Neural Mechanisms of Selective Visual Attention. *Annu. Rev.*
Psychol. 68, 47-72 (2017).

17.   Barbas, H., J., W., Joyce, M.K.P. & Garcia-Cabezas, M.A. Pathway mechanism for
excitatory and inhibitory control in working memory. *J. Neurophysiol.* 120, 2659-2678
(2018).

18.   Rockland, K.S. & Pandya, D.N. Laminar origins and terminations of cortical connections
of the occipital lobe in the Rhesus monkey. *Brain Res.* 179, 3-20 (1979).

19.   Usrey, W.M. & Sherman, S.M. Corticofugal circuits: Communication lines from the
cortex to the rest of the brain. *J. Comp. Neurol.* 527, 640-650 (2019).

20.   Hubel, D.H. & Wiesel, T.N. Receptive fields and functional architecture of monkey
striate cortex. *J. Physiol. (Lond.)* 195, 215-243. (1968).

21.   Huang, X., Elyada, Y.M., Bosking, W.H., Walker, T. & Fitzpatrick, D. Optogenetic
assessment of horizontal interactions in primary visual cortex. *J. Neurosci.* 34, 4976-4990
(2014).

22.   Roy, A. et al. Optogenetic spatial and temporal control of cortical circuits on a columnar
scale. *J. Neurophysiol.* 115, 1043-1062 (2016).

23.   Rajalingham, R. et al. Chronically implantable LED arrays for behavioral optogenetics in
primates. *Nat. Methods.* 18, 1112-1116 (2021).

24.   Owen, S.F., Liu, M.H. & Kreitzer, A.C. Thermal constraints on in vivo optogenetic
manipulations. *Nat. Neurosci.* 22, 1061-1065 (2019).

25.   Wang, J. et al. Integrated device for combined optical neuromodulation and electrical
recording for chronic in vivo applications. *J. Neural Eng.* 9, 016001 (2012).

26.   Ozden, I. et al. A coaxial optrode as multifunction write-read probe for optogenetic
studies in non-human primates. *J. Neurosci. Methods* 219, 142-154 (2013).

27.   Anikeeva, P. et al. Optetrode: a multichannel readout for optogenetic control in freely
moving mice. *Nat. Neurosci.* 15, 163-170 (2011).

28.   McAlinden, N. et al. Multisite microLED optrode array for neural interfacing.
Neurophotonics 6, 035010 (2019).

29.   Scharf, R. et al. Depth-specific optogenetic control in vivo with a scalable, high-density
µLED neural probe. *Sci. Rep.* 6, 28381 (2016).

30.   Maynard, E.M., Nordhausen, C.T. & Normann, R.A. The Utah intracortical Electrode
Array: a recording structure for potential brain-computer interfaces. *Electroencephalogr.*
Clin. Neurophysiol. 102, 228-239 (1997).

31. Lin, J.Y. Optogenetic excitation of neurons with channelrhodopsins: light
instrumentation, expression systems, and channelrhodopsin variants. *Prog. Brain Res.*
196, 29-47 (2012).

32. Wiegert, J.S., Mahn, M., Prigge, M., Printz, Y. & Yizhar, O. Silencing Neurons: Tools,
Applications, and Experimental Constraints. *Neuron* 95, 504-529 (2017).

33. Kim, K. et al. Artifact-free and high-temporal-resolution in vivo opto-electrophysiology
with microLED optoelectrodes. *Nat. Commun.* 11, 2063 (2020).

34. Callaway, E.M. & Wiser, A.K. Contributions of individual layer 2-5 spiny neurons to
local circuits in macaque primary visual cortex. *Vis. Neurosci.* 13, 907-922 (1996).

35. Pluta, S. et al. A direct translaminar inhibitory circuit tunes cortical output. *Nat.*
Neurosci. 18, 1631-1640 (2015).

36. Qi, G. & Feldmeyer, D. Dendritic Target Region-Specific Formation of Synapses
Between Excitatory Layer 4 Neurons and Layer 6 Pyramidal Cells. *Cereb. Cortex* 26,
1569-1579 (2016).

37. Acker, L., Pino, E.N., Boyden, E.S. & Desimone, R. FEF inactivation with improved
optogenetic methods. *Proc. Natl. Acad. Sci. U.S.A.* 113, E7297-E7306 (2016).

38. Andrei, A.R., Pojoga, S., Janz, R. & Dragoi, V. Integration of cortical population signals
for visual perception. *Nat. Commun.* 10, 3832 (2019).

39. Sagar, S.M., Sharp, F.R. & Curran, T. Expression of c-fos protein in brain: metabolic
mapping at the cellular level. *Science* 240, 1328-1331 (1988).

40. Kovacs, K.J. c-Fos as a transcription factor: a stressful (re)view from a functional map.
Neurochem. Int. 33, 287-297 (1998).

41. Tamura, K. et al. A glass-coated tungsten microelectrode enclosing optical fibers for
optogenetic exploration in primate deep brain structures. *J. Neurosci. Methods* 211, 49-57
(2012).

42. Stark, E., Koos, T. & Buzsaki, G. Diode probes for spatiotemporal optical control of
multiple neurons in freely moving animals. *J. Neurophysiol.* 108, 349-363 (2012).

43. Royer, S. et al. Multi-array silicon probes with integrated optical fibers: light-assisted
perturbation and recording of local neural circuits in the behaving animal. *Eur. J.*
Neurosci. 31, 2279-2291 (2010).

44. Iseri, E. & Kuzum, D. Implantable optoelectronic probes for in vivo optogenetics. *J.*
Neural Eng. 14, 031001 (2017).

45. Galvan, A. et al. Nonhuman Primate Optogenetics: Recent Advances and Future
Directions. *J. Neurosci.* 37, 10894-10903 (2017).

46. Kim, T. et al. Thermal effects on neurons during stimulation of the brain. *J. Neural Eng.* 19 (2022).

47. Bernstein, J.G. et al. Prosthetic systems for therapeutic optical activation and silencing of genetically-targeted neurons. *Proc. SPIE Int. Soc. Opt. Eng.* 6854, 68540H (2008).

48. Scharf, R. et al. A compact integrated device for spatially-selective optogenetic neural stimulation based on the Utah Optrode Array. *Proc. SPIE* 10482, 104820M (2018).

49. Potworowski, J., Jakuczun, W., Leski, S. & Wojcik, D. Kernel current source density method. *Neural. Comput.* 24, 541-575 (2012).

50. Mitra, P. & Bokil, H. Observed Brain Dynamics. (Oxford University Press, New York; 2008).

ACKNOWLEDGMENTS

We thank Kesi Sainsbury for histological assistance, Seminare Ta'afua for help with experiments, and Julian Haberland and Christine Kallmayer at the Fraunhofer-Institut für Zuverlässigkeit und Mikrointegration (IZM) for µLED-to-interposer bonding . This work was supported primarily by a BRAIN grant (U01 NS099702) from the National Institute of Health (NIH) to S.B., A.A. L.R and S.M. Additional support was provided by grants from the NIH (R01 EY026812, R01 EY019743, R01 EY031959), the National Science Foundation (IOS 1755431), and the Mary Boesche endowed Professorship, to A.A, and an unrestricted grant from Research to Prevent Blindness, Inc. and a core grant from the NIH (EY014800) to the Department of Ophthalmology, University of Utah.

AUTHOR CONTRIBUTIONS

Conceptualization: A.C., A.I., C.F.R., N.M., L. R., K.M., S.B. A.A. Device Fabrication and Bench Testing: C.F.R., Y.C., N.M., L.R.,K.M., M.D.D., S.B. Modeling: N.M., K.M. *In Vivo* Electrophysiology Testing: A.C., D.C., C.F.R., F.F., A.A. *In Vivo* cFos Testing: A.I., J.B., F.F., J.D.R., A.A., A.C. Analysis of Electrophysiology Data: A.C. Analysis of cFos Data: A.I., J.B., F.F., A.A. Writing-Original Draft: A.C., L.R., S.B., A.A. Writing–Review/Editing: all authors. Supervision & Funding Acquisition: A.A., S.B., K.M., L.R.

COMPETING INTERESTS STATEMENT

The authors declare no competing interests.

 FIGURE LEGENDS

Figure 1. UOA Design and Optical Properties

(A) Schematics of UOA design superimposed to a Nissl-stained coronal section of macaque V1

showing cortical layers. The UOA consists of 3 main components: a µLED array (B), an optical

interposer (C) and a glass needle array (D). **(B)** Two interleaved µLED arrays on a sapphire

substrate are shown in this image; the first 10x10 array is needle-aligned for deep layer

stimulation, the second 9x9 interstitial array lies in-between the first for surface stimulation. The

interstitial array, although built into the UOA, was not used in this study. Scale bar: 1mm. **(C)** A

region of the silicon optical interposer corresponding to approximately the size of the *white box*

in (B); the optical "vias" are etched through the silicon and matched to the size of a µLED

(80x80µm^2). Scale bar: 200µm. **(D)** High magnification image of the glass needle shanks bonded

to the interposer. Scale bar: 200µm. **(E)** **Left**: The µLED on sapphire and needle array

components are integrated into the final device, wirebonded, and encapsulated. The image shown

is a representative device. The integrated UOA used in this study consisted of 10x10 glass needle

shanks, 1.6 mm long (to target deeper layers) and 100-110µm wide, with tip apex angles about

64º. An image of the actual device used in the *in vivo* testing studies, after completion of the

experiment and explantation is shown in **Extended Data Fig. 2**. Scale bar: 1mm. **Right**:

Example spatial patterns of device operation. **(F)** Average output optical power (in mW) across

each needle tip at different drive voltages (currents), when the entire UOA was turned on (*top*

left inset). *Blue and gray bars*: needle shanks with estimated tip irradiances above and below,

respectively, the 1mW/mm^2 threshold for ChR2 activation. **(G)** **Left**: Ray trace model of light

spread in cortical tissue when a single µLED (in column 1 and row 8, i.e., the closest to the linear

electrode array —LEA– in penetration 2 –P2– used for the electrophysiological testing

experiment, and indicated as a *black dot*) is activated at various input voltages (% of maximum

intensity used), with power output calibrated to the bench tests. **Right**: Model of light spread in

tissue when all of column 1 (the nearest to the LEA in P2 and P3) is activated at various input

voltages. *Green contour* encloses tissue volume within which the light irradiance is above

1mW/mm^2, the threshold for ChR2 activation. Scale bars: 400µm.

Figure 2. Laminar Distribution of Responses Induced by UOA Photostimulation.

(**A**) The UOA used in the *in vivo* experiments inserted in macaque V1. (**B**) Same field of view as in (A) shown under fluorescent illumination to reveal expression of the red fluorescent protein tdTomato (*arrow*). (**C**) Preparation for recording electrophysiological responses to photostimulation. A 24 channel linear electrode array (LEA) was inserted next to the UOA (guide tube protecting array marked "LEA") slightly angled laterally (towards the UOA) and posteriorly. Here the UOA is partially covered with a piece of Gelfoam. (**D**) Current Source Density analysis (CSD; **Left**) and MUA (**Right**) signals recorded through the depth of V1 in P2 in response to phasic UOA photostimulation (pulse parameters: 100ms pulse duration, 5Hz, 0.82mW/mm^2; pulse periods denoted as *blue bars* above MUA plot). Here, all 100 needle-aligned µLED sites ("whole µLED array" condition) were activated simultaneously. CSD responses to each 100ms pulse were zero-aligned, while MUA is shown for the full 5Hz pulse train. *The dashed lines in the CSD panel* demarcate the borders of layer 4C (L4C)*; the gray shaded region in the MUA panel* delimits the extent of L4C. (**E**) Same as in (D), but following surface photostimulation of V1 via a laser-coupled optical fiber with pulse parameters of 10ms, 5Hz, 2.2mW/mm^2. (**F-I**) <u>**Top**</u>: Relative cortical depth of each contact on P2 (*black dot in the insets*) is plotted versus the relative response (% firing rate increase over baseline) to UOA stimulation for different 450nm µLED illumination patterns (*top insets*). Different colored traces are data for different photostimulation intensities (expressed as voltage or percent of max intensity used). *Gray area:* extent of L4C; *dashed lines*: approximate location of the L4A/4B (upper) and L5/L6 (lower) borders. <u>**Bottom**</u>: PSTHs with and without µLED activation are shown for the same contact on the LEA in L4C (marked by the *black circle* in the graphs above*)* across conditions. *Dashed line in the PSTH*: pulse periods.

Figure 3. Tangential Extent of Responses Induced by UOA Photostimulation.

A) Examples of model fits to single µLED photostimulation for the contact from P2 showing the largest relative response increase across these stimulation conditions. This contact preferred stimulation in the proximal UOA columns 1-2, at sites closer to the top of the device (rows 9-7). The schematics on the left of the UOA and of the LEA-P2 indicates as *blue shading* the UOA sites represented in the heat map, and as a *red dot* the contact on the LEA whose response is mapped on the right. The *horizontal lines* and *gray shading* on the LEA schematics mark the pial and white matter, and L4C boundaries, respectively. Color scale applies to panels (A-B, E-F).

(B) Average normalized fitted responses across all responsive contacts in P2 (*red dots* in schematics of LEA to the left). **(C)** Change in response in the column direction for P2. Average relative response amplitude (% of peak model-fitted response) is plotted as a function of stimulation intensity and distance along a straight line extending from the preferred UOA site in the column direction, and sorted by input intensity. Data averaged across all responsive contacts. **(D)** Change in response in the row direction for P2. Average relative response amplitude (% of peak response) is plotted as a function of stimulation intensity and distance along a straight line extending from the preferred UOA site in the row direction. Data averaged across all contacts. **(E-H)** Same as in (A-D) but for P3.

Figure 4. Onset Latencies Reveal Local Networks Activated by Focal Optogenetic Modulation. (A) <u>Left</u>: Schematics of UOA stimulation through a single μLED site (C1-R8) and of LEA in P2. **<u>Right</u>:** Pulse-aligned raster plots for all 21 channels on the LEA through the depth of V1. *Black lines* separate data from different channels. *Gray shaded region:* channels in L4C. *Blue line above plot*: 100ms pulse period at the input voltage (irradiance) indicated. *Red and black arrows denote example contacts in L4C and 6, respectively*. A graded shift in MUA onset latency is apparent. **(B) <u>Left</u>:** Pulse-aligned PSTHs for the two channels indicated by arrows in the raster plot in (A). Responses are plotted as baseline-subtracted firing rate versus time. Response onset latency at the L6 contact (35ms) clearly lagged that on the L4C contact nearest the UOA needle tips (17ms). **<u>Right</u>:** Histograms of pulse-by-pulse onset latencies for the two example contacts. **(C)** Heatmaps of MUA (firing rate) through the depth of V1 during the peri-pulse period, for the UOA stimulation condition indicated by the insets at the top left of each plot. Stimulation intensity (average irradiance) is reported above each plot. The firing rate color scale applies to all panels. *White dots* mark the onset latency (estimated from the mean PSTH-see Online Methods) for each contact that was significantly responsive to UOA stimulation.

Figure 5. Population Onset Latencies as a Function of UOA Stimulation Intensity and Spatial Pattern.

(A) Distance on the LEA of each contact from the contact with the fastest onset latency is plotted against onset latency; lines are linear fits. Each line is from simultaneous stimulation throughout the whole μLED array at each indicated intensity. **(B) <u>Left</u>:** Effect of varying photostimulation

intensity for a fixed column (C1). **Right**: Effect of varying stimulated column (C1 to C4) for a fixed photostimulation intensity (5V). Either lowering intensity for a given column or increasing the distance between an activated column and the LEA had similar effects on the latency of network activation.   **(C)** As in (B), but for a single μLED stimulation condition. **Left**:, photostimulation intensity was varied for a fixed μLED (C1-R8); **Right**: the stimulated μLED was varied along column 1 (from row 3 to 9) at a fixed intensity (5V for μLEDS in rows 8-10, but 7.8V for those in rows 4-7, as lower intensities did not evoke a response from many of these latter μLEDs). **(D)** The shortest onset latency across all intensities (here expressed as percent of max- see legend in **Extended Data Fig. 4C** for corresponding input voltage) is plotted for the whole array condition (**Left**), and selected columns (**Middle**) or μLEDs (**Right**).

Figure 6. Local Optogenetic Activation Through the UOA Spreads Through Cortico-Cortical Networks.

(A-C) Case MK414-RH (UOA activation). The same sagittal section encompassing parts of V1 and V2 is shown under 3 different fluorescent illuminations, to reveal Nissl stain (A), tdT/ChR2 expression (B; the red tdT fluorescence was converted to green for purpose of illustration), and c-fos IHC (C). *White solid contour*: V1/V2 border; *dashed contours:* layer boundaries (layers are indicated); *white boxes:* ROIs (numbered 1-3 in panel C) where c-fos+ cells were counted. *White Arrows in (B)* point to the visible damage caused by each UOA needle, while the *gray arrow* points to the likely location of one of the UOA needles which did not cause visible damage in this section. *Asterisks in (B)* mark the core of the viral injections, and sites of highest tdT/ChR2 expression. *P:* posterior; *D:* dorsal. C-fos expression in this case is observed throughout all layers (local) and across cortical areas (long-range). Scale bar in (A): 1mm (valid for A-C). **(D)** Higher magnification of c-fos IHC in and around each ROI. Scale bar: 0.2mm. **(E-H)** Case MK422-RH (Glutamate block). Same as in (A-D) but for a different case in which an AMPA receptor antagonist was injected into the SMA prior to UOA insertion and photostimulation. The sagittal section is from the SMA. *D:* dorsal; *A:* anterior. Scale bars: 1mm (E, valid for E-G); 0.2 mmm (H). Blocking AMPA receptors demonstrates that initial optogenetic activation is limited to the stimulated layers in the region of UOA insertion. **(I-J)** Case MK414-LH (UOA insertion-only). C-fos IHC in a sagittal section of SMA cortex (I) and at higher magnification in and around each ROI used for cell counts (J), in a case which only received UOA insertion in cortex

not expressing ChR2, and no photostimulation. Scale bars: 1mm (I), 0.2mm (J). **(K-L)** Case

MK421-RH (Light-only). Same as in (I-J), but for a control case in which SMA cortex not

expressing ChR2 only received surface photostimulation via an optical fiber-coupled laser and

no UOA insertion. Here only one ROI is shown at higher magnification to reveal the few labeled

cells in L1. Scale bars: 0.5mm (K), 0.2mm (L). The increases in cFos expression seen after full

UOA activation of ChR2-expressing cortex cannot quantitively be explained by device insertion

or surface illumination. **(M)** Average number of c-fos+ cells across sections used for

quantification, as a function of distance from the center of UOA insertion for the 4 different

cases. Error bars: s.e.m. **(N)** Distribution of c-fos+ cells across layers at each distance.

Figure 1

Figure 2

Figure 3

Figure 4

Figure 5

UOA + ChR2 + light (MK414-RH)

UOA + ChR2 + Glu block + light (MK422-RH)

UOA insertion only (MK414-LH)

Light only (MK421-RH)

M

Number of c-fos+ cells

■ UOA activation – MK414RH
■ Glutamate block – MK422RH
■ UOA insertion only – MK414LH
■ Light only – MK421RH

Distance (mm)

N

% of c-fos+ cells

■ L1
■ SL
□ L4C
■ DL

□ UOA activation – MK414RH
□ Glutamate block – MK422RH
□ UOA insertion only – MK414LH
□ Light only – MK421RH

Distance (mm)

Figure 6