An Optrode Array for Spatiotemporally Precise Large-Scale Optogenetic Stimulation of Deep Cortical Layers in Non-human Primates

Andrew M. Clark[1], Alexander Ingold[1], Christopher F. Reiche[2], Donald Cundy III[1], Justin L. Balsor[1], Frederick Federer[1], Niall McAlinden[3], Yunzhou Cheng[3], John D. Rolston[4], Loren Rieth[5,6], Martin D. Dawson[3], Keith Mathieson[3], Steve Blair[2*], and Alessandra Angelucci[1*]

[1] Department of Ophthalmology and Visual Science, Moran Eye Institute, University of Utah, Salt Lake City, UT

[2] Department of Electrical and Computer Engineering, University of Utah, Salt Lake City, UT

[3] SUPA, Institute of Photonics, Department of Physics, University of Strathclyde, Glasgow, UK

[4] Department of Neurosurgery, University of Utah, Salt Lake City, UT

[5] Mechanical and Aerospace Engineering, West Virginia University, Morgantown, WV

[6] Feinstein Institute for Medical Research, Manhasset, NY

*Corresponding authors.

Corresponding authors' address:

A. Angelucci:

65 Mario Capecchi Drive
Salt Lake City, UT 84132, USA
Tel: (801) 5857489
Email: alessandra.angelucci@hsc.utah.edu

S. Blair:

50 S. Central Campus Dr., Rm. 2110
Salt Lake City, UT 84112, USA
Tel: (801) 5856157
Email: blair@ece.utah.edu

ABSTRACT

Optogenetics has transformed studies of neural circuit function, but remains challenging to apply in non-human primates (NHPs). A major challenge is delivering intense and spatially precise patterned photostimulation across large volumes in deep tissue. Here, we have developed and tested the Utah Optrode Array (UOA) to meet this critical need. The UOA is a 10×10 glass waveguide array bonded to an electrically-addressable μLED array. *In vivo* electrophysiology and immediate early gene (c-fos) immunohistochemistry demonstrate that the UOA allows for large-scale spatiotemporally precise neuromodulation of deep tissue in macaque primary visual cortex. Specifically, the UOA permits either focal (confined to single layers or columns), or large-scale (across multiple layers or columns) photostimulation of deep cortical layers, simply by varying the number of simultaneously activated μLEDs and/or the light irradiance. These results establish the UOA as a powerful tool for studying targeted neural populations within single or across multiple deep layers in complex NHP circuits.

Keywords

Optogenetics, macaque, neocortex, feedback, AAV, cFos, current source density, multi-channel electrophysiology

A central goal in neuroscience is to understand how neural circuits generate the computations underlying perception and behavior. Optogenetics has transformed the study of neural circuit function by allowing for the selective modulation of neural activity on a physiologically relevant timescale[1]. While very successful in mice, its application to non-genetically tractable model species, such as the non-human primate (NHP), has lagged behind[2]. Extending optogenetics to NHP studies is crucial, as, due to their similarity to humans, NHPs represent the most important model for understanding neural circuit function and dysfunction[3-6], and provide an essential technology testbed towards human use[7] and the potential application of optogenetics as therapeutic interventions in humans[8,9]. The continuing refinement of viral methods for selectively delivering opsins to particular neural circuit elements, either defined projections[10,11] or particular cell types[12-14], is rapidly opening up new opportunities to study neural circuits in NHPs[2,15]. Despite these advances, one of the most significant remaining obstacles in the application of optogenetics to NHPs is the lack of devices for reliably delivering light to deep neural tissue across relatively large brain volumes with both: (i) sufficient intensity to reach optogenetic activation threshold, and (ii) sufficient spatial resolution to selectively modulate relevant circuit elements. For example, it is presently challenging to optogenetically perturb the activity of neuronal populations in the deep layers of the cerebral cortex across large cortical volumes.

There are several features of cortical networks that both make development of such a device critical and provide design requirements for the device First, a subset of cortico-cortical feedback connections, which are critical for the contextual modulation of sensory processing[10,16], as well as cognitive phenomena such as selective attention[17] and working memory[18], arises from deep cortical layers[19]. Selective optogenetic modulation of these projections would greatly inform work examining the role of cortico-cortical feedback in neural computation. Second, cortico-thalamic projections, which are part of thalamo-cortico-thalamic loops (ubiquitous features of mammalian neural networks), arise exclusively from deep layers[20]. Dissecting these circuits will require selective perturbation of deep layer cortical neurons with high spatiotemporal precision. Finally, in NHP sensory cortex, specific stimulus features are mapped in an orderly, columnar, fashion. That is, nearby neurons prefer similar stimulus features, and this pattern of selectivity extends throughout the cortical depth[21]. While methods for selective illumination of the cortical surface at the spatial scale of these columns have recently been

developed in smaller animals[22,23], they unfortunately only allow for stimulation of the superficial

layers in the NHP.

State-of-the-art optogenetic experiments in NHPs have mainly followed two light

delivery approaches: through-surface illumination and penetrating probes. Surface

photostimulation utilizes either a laser- or LED-coupled optical fiber positioned above the

cortical surface[10], or chronically-implantable surface LED arrays[24]. These approaches enable

photoactivation of a large area, but only to a depth of <1mm, due to light attenuation and

scattering in tissue, as well as to unintended superficial layer neuron activation and even damage

caused by the heat generated by the light intensity required to reach deeper layers[10,25]. In

contrast, penetrating optical fibers, integrated with single[26,27] or multiple[28] recording probes,

allow photoactivation at depths >1mm, but only of a small tissue volume (a few hundred microns

in diameter), and due to their size and shape, can cause significant superficial layer damage.

To overcome the limitations of these light delivery methods, we have developed the Utah

Optrode Array (UOA), a 10x10 array of glass needle shanks tiling a 4x4 mm^2 area and bonded to

an electrically-addressable μLED array independently delivering light through each shank[29,30].

Device bench testing and *in vivo* testing in macaque primary visual cortex (V1) demonstrate that

the UOA allows for spatiotemporally patterned photostimulation of deep cortical layers with sub-

millimeter resolution (at the scale of single layers and columns) over a large expanse of cortical

tissue. This spatial selectivity can be scaled up to multiple layers and columns simply by varying

the number of simultaneously activated μLEDs and/or the light irradiance. These results

establish the UOA as a powerful tool for studying local and large-scale populations of deep layer

neurons in NHP cortex.

                                      RESULTS

The UOA: Geometry and Optical Properties

The spatiotemporally multiplexed UOA was developed based on the geometry of the widely utilized Utah Electrode Array (UEA)[31]. The UOA is a 10x10 array of penetrating optical light guides made in glass, with customizable length (up to 2.5mm) and shank width (80-120μm diameter) on a 400μm pitch tiling a 4x4mm area. For deep layer light delivery, a custom blue (~450nm) μLED array grown on sapphire is directly integrated with the device, with each electrically addressable 80μm μLED delivering light through a single needle shank (**Fig. 1A-E**). A second 9x9 array of "interstitial" μLEDs can be interleaved on the same device and matrix-addressed for independent surface stimulation (as shown in **Fig. 1B**, but the interstitial μLED array was not used in this study). To limit the spatial spread of light coupled into the optrode, the first generation UOA used a metal pinhole array[29]. Bench testing of this first generation device demonstrated the capacity for delivering irradiances in excess of activation threshold across a range of commonly employed depolarizing[32] and hyperpolarizing[33] opsins. Estimates from bench testing suggested a 50% decrease in irradiance within tissue within approximately 200μm of a needle tip and a beam cross section of approximately 150-200μm (full-width at half-maximum - FWHM)[29]. These initial results suggested that direct optogenetic activation through the UOA is on a spatial scale commensurate with the functional architecture of primate cortex.

Building upon these initial results, we have developed the second generation UOA reported here, which incorporates an optically opaque (silicon) interposer layer with optical "vias" on the array backplane to eliminate uncontrolled surface illumination and inter-needle crosstalk, and facilitate alignment during bonding to the μLED-array (**Fig. 1A, C**; see Online Methods for manufacturing details). This second-generation device (**Fig. 1A-E**) was bench tested (**Fig. 1F**) and its *in vivo* optical performance was estimated via ray tracing (**Fig. 1G**). From optical measurements, the output power (in mW) emitted through the tip of each needle in the 10x10 array at different drive voltages is shown in **Fig. 1F** and in **Extended Data Fig. 1**, where we also show the estimated output irradiance (mW/mm^2) for each needle tip (assuming a pyramidal tip and uniform light output). At 3V, estimated irradiance levels are below the 1 mW/mm^2 threshold for the excitatory opsin Channelrhodopsin-2 (ChR2); average output optical

power across all needles ± SD = 0.0057±0.004 mW, corresponding to an average estimated

irradiance of 0.21±0.14 mW/mm^2 (**Extended Data Table 1**). There is variation across the array,

due primarily to variations in the resistance (and therefore slope efficiency) of each μLED. At

3.5V, about 30% of the stimulation sites reach or exceed this threshold (average optical power ±

SD = 0.022 ± 0.013 mW; average irradiance: 0.82 ± 0.49 mW/mm^2), while at 5V, more than

90% of the sites emit above threshold (average optical power ± SD = 0.1 ± 0.056 mW; average

irradiance: 3.79 ± 2.08 mW/mm^2). In principle, software modifications in the matrix driver

interface could be made to better equalize stimulation levels across the array.

We used optical ray tracing to estimate the performance of the device *in vivo* based upon

calibration curves obtained from bench testing (see Online Methods for details). In particular, we

were interested in estimating the direct neural stimulation volume (based upon the local

irradiance in tissue) as a function of drive voltage and pattern of lit needles in order to interpret

the *in vivo* results. These results are shown in **Figure 1G**, where the left column panels show the

stimulation volume along the first UOA column as produced by the needle (column 1, row 8)

nearest one of the electrode penetrations (penetration 2, or P2) in the *in vivo* experiments, and the

right column panels show the activation volume when the entire UOA column 1 is lit. Each row

of figures is for a different drive voltage. At low drive voltage (~3V, equal to 38% of the

maximum input voltage used), highly localized stimulation in tissue near the needle tips is

produced (note also that the irradiance across the tip surface is non-uniform – concentrated near

the apex – explaining why above-threshold irradiance levels can be achieved at 3V as

demonstrated in the *in vivo* experiments). At higher voltages (~5V/64% max intensity and

above), the stimulation volume overlaps that of adjacent needles, while also extending deeper

into tissue. When driving an entire column, at 3V, stimulation localized near each tip is mostly

retained, whereas a nearly continuous stimulation volume is obtained at 3.2V due to overlapping

fields. At 5V (64% of max intensity) and 7.8V (100% max intensity), the depth of this

continuous volume increases, both above and below the tips.

In Vivo **Testing: Electrophysiology**

173   **Figure 2 about here**

We used *in vivo* linear electrode array (LEA) recordings to assess the utility of UOAs for precise

modulation of neural activity in deep cortical layers of monkey cortex expressing the excitatory

opsin *Channelrhodopsin-2* (ChR2). ChR2 was expressed in area V1 of a macaque monkey by

injecting a mixture of Cre-expressing and Cre-dependent adeno-associated virus (AAV9)

carrying the genes for ChR2 and the red reporter protein tdTomato (tdT). We have previously

shown that this viral mixture results in almost exclusive anterograde neuronal infection at the

injected site[10]. Following a survival period to allow for viral expression, we recorded neural

spiking activity using a 24-contact LEA inserted nearby a UOA (fully integrated with a μLED

array, as described above and in **Fig. 1A-E**) that was implanted into a region of dense ChR2

expression in V1 (**Fig. 2A-C; Extended Data Figs. 2-3A**). We performed 3 LEA penetrations

(P1-P3), but for only 2 of them (P2, P3) photostimulation via the UOA modulated neural

activity, likely because P1 was located farthest from the region of tdT/ChR2 expression, as

revealed by postmortem histology (**Extended Data Fig. 3A**). Below we report data from P2 and

P3.

Comparison of Surface and UOA Photostimulation

Figure 2D shows neural responses recorded in P2 to simultaneous stimulation of μLEDs at all

UOA sites at an irradiance level (average ± SD across the whole UOA = 7.4 ± 3.19 mW/mm^2,

induced by an input intensity of 7.8V; see **Extended Data Table 1**) almost an order of

magnitude greater than ChR2 activation threshold[32]. To examine the spatiotemporal distribution

of responses to UOA stimulation across V1 layers, we first performed a current source density

(CSD) analysis of the local field potential (LFP) recorded across the LEA around the time of a

UOA pulse. The CSD, computed as the second spatial derivative of the LFP, reveals the location

of current sinks (negative voltage deflections reflecting neuronal depolarization) and sources

(positive voltage deflections reflecting return currents) throughout the cortical depth. Current

sinks in response to UOA stimulation were confined to L4C and below (**Fig. 2D,** Left),

suggesting that direct optogenetic activation was confined to a region below the UOA needle

tips. Analysis of the multiunit (MU) spiking activity across contacts demonstrated strong phasic

responses to UOA stimulation that extended from L4C to the white matter boundary (**Fig. 2D**,

Right). **Figure 2E**, shows the CSD and MU signals under the same photostimulation and

recording conditions as in **Figure 2D**, but at lower light irradiance (0.82 ± 0.49 mW/mm^2 -

3.5V). At this lower intensity, CSD signals and MU activity were mostly localized to L4C and

the lower part of the deep layers, with L4C activation preceding in time that of deeper layers.

These laminar patterns of neural activity suggest that the UOA needle tips closest to P2

terminated in L4C and that at low photostimulation intensities light spread nearby the UOA tips.

Additional analysis demonstrating that response onset latency and light activation threshold were

lowest for the P2 contacts located in L4C, together with postmortem histological assessment,

further confirmed that the UOA needle tips closest to P2 were indeed located in L4C (**Extended**

Data Fig. 3A, B Right). Comparison of the above laminar patterns of response with that elicited

by direct surface stimulation in a different animal at a lower suprathreshold irradiance

(2.2mW/mm^2) revealed a sharp dissociation. Specifically, surface stimulation of ChR2 evoked

responses starting in superficial layers and terminating in L4C (**Fig. 2F**).

UOA Stimulation Parameters Can Be Tuned to Achieve Laminar Specificity

To systematically assess the impact of UOA stimulation on neural spiking responses we varied:

(i) the spatial pattern of UOA stimulation, from single μLED sites, to entire columns, to the

entire device, and (ii) stimulation intensity across different spatial patterns. In all conditions, we

used phasic stimulation (5Hz, 100 msec pulses for 1 sec with 1-10 sec inter-trial intervals, with

the longer intervals used at the higher stimulation intensities) with a slow on/off ramping to

eliminate the potential of any electrical artifacts induced by capacitive coupling at the

array/tissue interface[34]. As an example, **Figure 2G-J** shows responses from P2. An analysis of

firing rate increase across layers induced by systematically stimulating a single μLED along

column 1, indicated that the UOA needles in column 1 closest to P2 were those in rows 8 and 9

(C1-R8, C1-R9), and that the tips of these needles terminated into L4C (**Extended Data Fig. 3B**

Left). The laminar distributions of MU activity in P2 varied considerably in strength across

conditions, but were reliably confined to deeper layers. By varying the spatial pattern of

stimulation and/or the stimulation intensity, MU activation could be confined to single layers or

spread across multiple layers. For example, activation of the whole μLED array (**Fig. 2G**) at low

light intensity (>2.8 and up to 5V), evoked a peak of MU activation localized to L4C, the layer

of needle tip terminations nearest P2, and this peak increased in magnitude with increasing

stimulation intensity. At and near the peak of this intensity range, a second smaller peak of MU

activation was also present in L6 but not in L5. However, when the whole UOA was activated at

the highest intensity used (7.8V), the MU spiking peak in L4C decreased in magnitude, while
activation of L5 and L6 increased. In macaque V1, L4C projects to both L5 and L6[35], but its net
effect is to suppress the former[36] and activate the latter[37], consistent with the interpretation that at
lower light intensities lack of L5 responses and increases in L6 responses may have resulted
from synaptic spread from L4C neurons directly activated by UOA stimulation. In contrast, at
the higher photostimulation intensity (black curve in **Fig. 2G**) light scattering through a larger
volume may have directly contributed to firing rate increase in deeper layers 5 and 6, while the
reduced MU peak in L4C could have resulted from activation of higher threshold inhibitory
networks. In a later section of the Results, we provide evidence supporting this interpretation.
Intensities $\geq$ 5V evoked similar laminar patterns and magnitude of neural activity irrespective of
whether a single µLED or an entire column nearest the LEA or the whole UOA were illuminated
(**Fig. 2G, H, J**). However, at lower photostimulation intensities, for a given intensity, firing rate
increased with the number of µLEDs activated (e.g. compare blue curves in **Fig. 2G, H, J**), and
higher intensities (>3.2V) were required to modulate neural activity via a single active µLED
(**Fig. 2J**). Moving the µLED activation column from 1 to 5 resulted in a 10 times reduction in the
amplitude of MU activity (**Fig. 2I**). No increase in firing rate could be evoked by stimulation of
an entire column beyond column 5 or a single µLED in column 1 beyond row 4 (distances from
the LEA of about 2.6-2.7 mm – estimated on the postmortem histology) even at the highest drive
voltage used, i.e. 7.8V (corresponding to an average irradiance of 6-11 mW/mm^2, see **Extended**
Data Table 1).
256
257   *Tangential Extent of Responses Induced by Photostimulation Via the UOA*
258   **Figure 3 about here**
259
260   Having determined that C1-R8 and C1-R9 were the UOA needles in column 1 closest to P2, and
261   that these needles' tips terminated in L4C, we then performed a similar analysis to determine the
262   location of P3 relative to the UOA. The results of this analysis, reported in **Extended Data Fig.**
263   **3C**, suggested that C1-R7 was the UOA needle in column 1 closest to P3 and that this needle tip
264   terminated in the superficial layers.
265   We then asked whether MU responses across LEA contacts were tuned for the spatial site
of UOA stimulation, how such tuning varies with intensity, and whether UOA stimulation can be

tuned to selectively activate volumes on the scale of functional columns in NHP cortex.
Moreover, different spatial tuning across LEA contacts would provide us with information about
the angle of LEA insertion into cortex relative to the UOA. We performed this analysis
separately for each penetration.

To estimate the MU response to stimulation at UOA sites between columns 1-5 and rows
3-10, we fit a multiple linear regression model to the MU spiking recorded at each LEA contact,
with row, column, and intensity (volts) as independent variables (see Online Methods). We
included in this analysis only contacts on which there was a significant difference in MU firing
rates during the stimulation and control periods for at least one of the row or column conditions
(ANOVA, p < 0.01). On average, including a quadratic term explained more of the variance in
the MU response (mean R^2 ± SD: 0.58 ± 0.14 vs. 0.31 ± 0.11 for a linear model; Kolmogorov-
Smirnov, p < 10^{-7}). **Figure 3A, E** shows plots of MU responses, evoked by a mid-level
photostimulation intensity (3.5V), estimated from the fitted data for the contact in P2 and P3 that
showed the greatest relative response modulation. We normalized each contact's fitted responses
to the peak and averaged across contacts to get a sense of whether MU responses preferred
stimulation at different UOA sites on different LEA penetrations (**Fig. 3B**, **F**). Consistent with
postmortem histological assessment (**Extended Data Fig. 3A**) and the analysis in **Extended
Data Figure 3B-C**, the peaks for P2 contacts tended to cluster mostly near columns 1-2 and
rows 8-9, while those for P3 contacts clustered mostly near columns 1-3 and rows 4-7, and the
spatial pattern of peak activity suggested that for both penetrations, particularly P3, the LEA was
inserted at a slightly oblique angle. Statistical analysis revealed that peak locations differed
significantly across the two penetrations (ANOVA, p < 0.01).

The data in **Figures 2H-I, 3A, B, E, F** suggested that the magnitude of evoked MU
responses decreased with increasing distance between photostimulation and recording sites. To
quantify this observation and better characterize the extent of photostimulation-evoked responses
across the tangential domain of the cortical tissue, as well as how this extent is affected by
photostimulation intensity, we examined response amplitude as a function of distance on the
UOA (in a straight line extending along either the row or column axis) from the LED site that
evoked the peak response. Plots of relative response (percent of the peak fitted response) as a
function of distance in either the column or the row direction, sorted by input intensity, are
shown in **Figure 3C-H** separately for each penetration. As is evident from the steeper decrease

in responses along the column versus the row axis, as well as the difference in relative response

across stimulus intensities, there was a significant main effect of UOA axis and input intensity on

this relative response measure (ANOVA, both $p < 10^{-21}$) as well as a significant difference across

penetration (ANOVA, $p < 10^{-14}$). Finally, there was a significant interaction between intensity

and UOA axis as well as UOA axis and penetration (ANOVA, both $p < 0.01$). These results

indicate that the response decrease from peak is greater in the column versus the row direction,

that intensity has a different effect on this drop-off when looking at the row and column

directions, and that this differed across penetrations. In the column direction, at the lowest

stimulation intensity (2.8V), evoked responses dropped to 16% of the peak at a distance of about

1.6mm from the location evoking peak response, but at the higher intensities (5-7.8V) responses

only dropped to about 50% of the peak at the same distance (**Fig. 3C,G**). In contrast, in the row

direction, at the lowest stimulation intensity (2.8V), evoked responses only dropped to about

80% of the peak at a distance of about 2.8mm from the peak location, and to about 90% at ≥5V

intensity (5-7.8V) (**Fig. 3D,H**). The difference in response drop-off with distance in the column

vs. row directions is likely explained by the greater differences in irradiance, for a given input

intensity, along the column as compared to the row axis (see **Extended Data Fig. 1**).

In summary, the spatial spread of MU activation along the tangential domain of cortex

varied according to UOA stimulation site and intensity. Importantly, the extent of this spread was

more limited at lower intensities, suggesting that increasing intensity increased the volume over

which cells were optogenetically activated, consistent with the model simulations in **Fig. 1G**.

UOA Activation Parameters Can Be Tuned to Activate Distinct Cortical Networks

Given the spatial separation between the LEA and the UOA (estimated to be about 1-1.1mm for

P2 and 700-800μm for P3 on the histological sections; **Extended Data Fig. 3**), the reported

sharp falloff in light intensity over short distances in tissue[38,39], and our bench estimates of light

spread from the UOA tips (≤ 600μm laterally and radially from the needle tip at intensities ≤ 5V,

and 0.9-1.0mm at 7.8V intensity; **Fig. 1G** and[29]), we reasoned that the UOA-driven MU activity

we observed, at least at ≤ 5V photostimulation intensities across all contacts and at any intensity

in the deeper layer contacts, was not caused by direct activation of ChR2-expressing cells nearby

the LEA recording site, but rather by indirect activation of ChR2-expressing cells in the vicinity

of the UOA needle tips. Given the greater spread of light at the highest intensity used, direct

activation of neurons recorded at LEA contacts closest to the UOA tips was, however, possible.

Thus, to estimate the tangential spread of direct as well as indirect activation to the LEA

recording sites, we measured the onset latency of UOA-driven MU responses across layers.

Figure 4 about here

Example latency data from P2 are shown in **Figure 4A**. Here, the UOA stimulus was a

single μLED (C1-R8) nearest the recording location with an input voltage of 5V. Pulse-aligned

MU rasters on each contact are shown next to a schematic of the recording LEA. The fastest

response occurred in mid layers with an onset latency of about 15 msec, suggesting the recorded

neurons in these layers were activated indirectly via ChR2-expressing cells in mid-layers nearby

the C1-R8 needle tip. Moreover, deep layer response onset (mean ± s.e.m: 30 ± 7 msec) lagged

that in mid-layers, as would be expected if optogenetic activation first propagated through L4C

before being synaptically relayed to deeper layers, via L4C-to-L5/6 connections. Averaged

PSTHs for the peri-pulse period on one example L4C and one L6 contacts are shown in **Figure

4B**. There was a significant pulse-by-pulse difference in onset latency across contacts (ANOVA,

$p < 10^{-30)}$, as well as a significant pairwise difference across these two LEA recording sites

(Tukey HSD test, $p < 10^{-6}$; **Fig. 4B**, Right).

To better visualize onset latency of evoked responses throughout V1, **Figure 4C** shows

average peri-pulse PSTHs across all LEA contacts as a function of normalized cortical depth for

an example whole μLED array (top panels), single column (middle panels), and single μLED

(bottom panels) experiment at different photostimulation intensities and different spatial

separations between UOA stimulation and LEA sites. The primary effects of increasing total

stimulus area at the lower intensities were to increase the number of responsive contacts and the

amplitude of driven responses, as well as to shorten onset latencies (e.g. compare panels in the

left column of **Fig. 4C**). At higher intensities (e.g. 5V, middle column), there was little change in

these parameters across these large differences in total stimulated area. The main effect of

decreasing the stimulus intensity for a fixed area (**Fig. 4C** compare middle to left columns), or

increasing the separation between the stimulated UOA site/s and the LEA for a fixed stimulus

intensity (**Fig. 4C** compare middle to right panels in the center and bottom rows) was an

increased delay in onset latencies across all contacts. For example, mean onset latency at 5V and

3.2V across all contacts was 17 ± 1.7 msec and 25.4 ± 2 msec for the whole array condition, 19.8

± 1.4 msec and 37.5 ± 1.9 msec for the C1 condition, and 21.4 ± 2.3 msec and 74.1 ± 1.6 msec

for the single μLED (C1-R8) condition. Mean onset latency at 5V for the C3 and C1-R6

conditions was 47.6 ± 4.3 msec and 59.4 ± 4.1 msec, respectively. Calculating onset latency on a

pulse-by-pulse basis and looking at the effects on latency of cortical depth, stimulation pattern,

and stimulation intensity, we observed significant main effects of pattern and intensity,

significant two-way interactions between depth and pattern, depth and intensity, pattern and

intensity, and a significant three-way interaction between all three factors (ANOVA, all $p < 10^{-4}$

). Limiting our analysis to each pattern, we observed a significant main effect of intensity and

distance from the LEA on onset latency for the single column conditions in **Fig. 4C** (ANOVA,

all $p < 10^{-4}$), and a significant main effect of distance for the single μLED conditions (ANOVA,

$p = 0.03$). Furthermore, in many conditions, pairwise comparisons across contacts revealed a

delayed response onset in deep layers (and in superficial layers in some conditions) relative to

mid-layers; this time lag varied with intensity and distance of the stimulation site from the LEA,

increasing at lower intensities and greater distances. These differences in onset latency between

the fastest mid-layer response and those in L6 were statistically significant for most conditions

shown in **Fig. 4C** at 5V, and for some at 3.2V (Tukey Kramer, all $p < 0.01$; see **Extended Data**

Fig. 4 for specific comparisons). There was also a significant difference in onset latency between

mid- and superficial layers in some conditions, namely C1 at 5V and the whole array at 5V and

3.2V (Tukey Kramer, all $p < 0.01$; **Extended Data Fig. 4**). Notably, however, when the whole

μLED array was stimulated at the highest intensity used (7.8V), there was no difference in onset

latencies between the deep and middle layers, suggesting the former were directly activated by

light spreading through deeper tissue in this condition (**Fig. 4C** top right, and **Extended Data**

Fig. 4). The fastest onset latency in this condition was 11 msec, possibly a long enough delay to

suggest that light did not directly activate the L4C neurons recorded at the LEA site; however, a

shorter latency, indicative of direct light activation of L4C contacts in this condition, is also

possible given the slow on ramping of the photostimulating current used in our experiments, and

the possibility that a lower threshold criterion would yield shorter latencies.

388   **Figure 5 about here**

To quantify these effects across the population (n= 33 significantly responsive contacts,
across 2 LEA penetrations), we first calculated the distance between each LEA contact and the
contact with the shortest onset latency, and plotted this distance versus onset latency, separately
for each unique combination of UOA stimulation site(s) and intensity. Similar to the P2 data
shown in **Fig. 4C**, the population data showed 2 main effects. (1) Onset latency decreased
significantly across all contacts with increasing stimulation intensity (ANOVA, main effect of
intensity, all $p < 0.01$; **Fig. 5A, 5B Left, 5C Left**) and proximity to the recording LEA site
(ANOVA, main effect of row or column on UOA, all $p < 10^{-4}$; **Fig. 5B Right, 5C Right**). (2)
Onset latency increased significantly with contact distance on the LEA from the fastest contact
(**Fig. 5A-C**, main effect of distance on the LEA, ANOVA all $p < 0.01$), except for the whole
array condition for which, post-hoc comparisons revealed that at the highest intensity there was
no significant difference in latency across contacts (as also shown in **Extended Data Fig. 4**).
This suggests that at lower irradiance the increase in firing rate caused by direct light started near
the site of UOA tip termination (in L4C in P2, but in L3 in P3) and then spread to more distant
sites indirectly via interlaminar networks; in contrast, at higher irradiances, lack of difference in
onset latency across contacts suggests that light spreads through a larger tissue volume.

Across the three categories of UOA stimulation (whole array, column, and single μLED),
only for the whole array and single μLED conditions did we observe a significant interaction
between the effects of distance along the LEA and UOA photostimulation intensity on onset
latency (**Fig. 5A and C Left**; both $p < 0.05$, ANOVA). Specifically, in these conditions,
lowering photostimulation intensity decreased the slope of the curves, indicating that the
difference in onset latency with distance on the LEA increased at lower intensity. Additionally,
for the single μLED condition, we also observed a significant decrease in the slope of the curves
when photostimulating at increasing UOA-LEA separation, but only when we moved the single
μLED stimulus to sites that were far enough from the LEA to necessitate stimulation at the very
highest powers to elicit any response (dashed lines in **Fig. 5C Right**, μLED in rows 4-7;
ANOVA, LEA distance x UOA row x intensity interaction, $< 10^{-3}$). For the single column
condition, there was no significant interaction between contact distance and either
photostimulation intensity or UOA-LEA separation (**Fig. 5B**; ANOVA, all $p > 0.09$).

In summary, by varying photostimulation intensity and/or number of stimulated sites, the
UOA allows activation of single or multiple layers, while by varying the spatial separation

between the site of UOA stimulation and that of the recording, the UOA allows investigations of

local vs long-range intra and interlaminar circuits.

In Vivo **Testing: c-Fos Expression**

Figure 6 about here

As an additional approach to validate the performance of the UOA for large scale

photostimulation of deep cortical tissue, we measured changes in c-fos expression, an immediate

early gene whose expression rapidly increases when neurons are stressed or activated[40,41]. The

product of this gene, c-fos protein, can be identified using immunohistochemistry (IHC) and

used as an indirect measure of the spatial pattern of neural activation in post-mortem brains. We

analyzed patterns of c-fos expression using IHC (see Online Methods) in two control and two

experimental hemispheres from 3 animals. In one of the experimental cases (MK414-RH), a

"passive" UOA (a device similar to the one used for the electrophysiological experiments but

without an integrated μLED array) was implanted in area V1 of the right hemisphere in a region

of high ChR2 expression (**Fig. 6A-B**), 10 weeks after injecting a mixture of Cre-expressing and

Cre-dependent AAV9 vectors carrying the genes for ChR2 and tdT. Photostimulation of deep

cortical layers via the passive UOA was performed by illuminating a subset of UOA needles

using a fiber-coupled blue laser and a collimating lens, while shielding from light the non-

stimulated cortex and UOA needles (see Online Methods). About 75 minutes after completion of

the photostimulation protocol, the animal was euthanized, and the brain processed for histology

and c-fos IHC. Results from this case are shown in **Figure 6A-D**. Histological analysis revealed

that the UOA was inserted at an angle (due to the brain curvature), so that its needle tips ended at

the bottom of the superficial layers, dorsally, and in progressively deeper layers ventrally, with

most of the needle tips ending in L4C and only the most ventral ones reaching into L6 (**Fig. 6A-**

B). There was extensive c-fos expression encompassing all cortical layers, and extending far

beyond the site of UOA insertion and photostimulation. Specifically, c-fos positive (c-fos+) cells

were found throughout V1 (**Fig. 6A, C**), as well as in extrastriate cortical areas known to receive

inputs from V1, including V2 (**Fig. 6A, C**), the third-tier cortex and area MT (not shown).

Qualitatively, c-fos expression appeared densest at the site of UOA photostimulation in the

superficial and mid-layers, and decreased with distance from this site (**Fig. 6C, D**). This extensive pattern of c-fos+ cells suggested c-fos expression was induced by both direct neural activation by light as well as indirect activation via synaptic activity. To test this hypothesis, as well as to determine the extent of neural activation directly induced by UOA-photostimulation, we repeated the same experiment in a different animal (MK422-RH) in which the AMPA receptor antagonist NBQX was applied to ChR2-expressing cortex prior to passive-UOA insertion and photostimulation, in order to block most glutamatergic synaptic transmission. The UOA was only partially inserted in this case, so that most of its needle tips only reached the bottom of the superficial layers (**Fig. 6E-F**). As expected, C-fos expression in this case was much less extensive than in case MK414-RH, being largely confined to the superficial layers at the site of UOA photostimulation (**Fig. 6G-H**).

To control for potential c-fos activation induced by UOA insertion independent of photostimulation and/or ChR2 activation of neural tissue, in case MK414, we inserted a passive UOA in the supplementary motor area (SMA) of the hemisphere contralateral to the experimental one; the control hemisphere (LH) was not injected with virus, thus did not express opsins. The animal was euthanized 4 hrs following UOA insertion without receiving any photostimulation. Histological analysis revealed that the UOA was fully inserted in this case, its tips reaching into the upper part of the deep layers (L5). C-fos expression encompassed all layers, but was largely confined to the site of insertion (**Fig. 6I-J).** To control for potential c-fos activation induced by light, independent of ChR2-induced neural activation and/or UOA insertion, in a different animal (MK421-RH) we performed surface photostimulation of SMA cortex not expressing ChR2, using a fiber-coupled laser and a collimating lens; no UOA was inserted in this case. We found a few c-fos+ cells immediately underneath the illuminated area largely within L1 (**Fig. 6K-L**). To quantify these results, for each experimental and control case we counted c-fos+ cells in 3 regions of interest (ROIs) encompassing all cortical layers, one centered in the region of UOA insertion and/or light stimulation, the other two located 4 and 8 mm, respectively, from the first one (*white boxes* numbered 1-3 in **Fig. 6A-L**; see Online Methods for details). **Figure 6M** plots the average number of c-fos+ cells across samples, as a function of distance from the UOA insertion site, while **Figure 6N** shows the laminar distributions of c-fos+ neurons at each distance. Confirming our qualitative observations, we found significant local and long-range c-fos expression only in case MK414-RH, which received

photostimulation of ChR2-expressing cortex via the UOA. Application of the glutamate blocker

prior to photostimulation prevented long-range c-fos expression, and reduced its expression by 5

fold in the area of UOA stimulation, where it was largely confined to the directly

photostimulated layers near the UOA tips. UOA insertion-only led to as much local c-fos

expression as the glutamate block case, but to greater interlaminar, as well as intra- and inter-

areal long-range spread, suggesting that neurons activated by the insertion trauma, in turn

indirectly activated downstream networks. Finally, surface photostimulation of cortex not-

expressing ChR2, without UAO insertion, caused virtually no c-fos expression, except for a few

cells in L1 and upper L2. Statistical analysis (one way ANOVA with pairwise comparisons and

Bonferroni correction for multiple comparisons) revealed a significant difference in the number

of c-fos+ cells at each distance between the experimental case (MK414-RH) and the glutamate

block case (MK422-RH; $p<0.001$, at all distances) as well as between MK414-RH and each of

the control cases (MK414-LH: $p<0.001$, at all distances; MK 421-RH: $p<0.001$, at all distances).

There was no significant difference between the glutamate-block and UOA-insertion-only

(MK414-LH) cases at any distance ($p=1$ at 0mm, $p=0.23$ at 4 mm, and $p=0.44$ at 8mm distance),

but both cases differed significantly from the light-only case (MK421-RH) at 0mm distance

($p<0.05$ for all comparisons). Finally, the number of c-fos+ cells decreased significantly with

distance for cases MK414-RH ($p<0.001$), MK422-RH ($p=0.001$), and MK414-LH ($p=0.003$), but

not for case MK421-RH ($p=0.079$).

DISCUSSION

We have developed and tested *in vivo* in the NHP cortex a novel device, the UOA, a

10x10 array of penetrating light waveguides with integrated µLEDs, which has the potential to

open novel avenues for furthering optogenetic research in larger brain species, particularly

NHPs. Current optogenetic approaches in NHPs allow for light delivery either over a large area

but at limited depth[10,24], or to deeper tissue but over a small area[26-28,39]. Multi-site optical

stimulation probes for larger volume stimulation have also been developed and combined with

singe[42] or multisite[43,44] electrical recordings, but these multi-fiber-based approaches are typically

cumbersome to assemble and don't easily scale to precisely target multiple small tissue volumes.

The UOA combines the advantages of all these approaches. Namely, it allows for both focal and

larger-scale neuronal activation of single or multiple deep layers simply by varying the number
of simultaneously activated µLEDs and/or the light irradiance. Moreover, although here we only
used the needle-aligned µLED array for deeper layer activation, the integrated interleaved
interstitial µLED array allows for selective photostimulation of superficial layers either
independently or in conjunction with deep layers.

This novel device has the potential to significantly impact neuroscience allowing
dissection of neural circuit function in animal species capable of complex behavior. By design,
the UOA is intended to achieve spatial resolution in cortical application in NHPs, and eventually
humans, and is, thus, ideal for addressing neuroscience questions that require large-scale
manipulations of deep and/or superficial cortical layers. Even with its current limitation of
lacking recording capability, we have demonstrated that the UOA used as a stimulation-only
device in conjunction with LEA recordings can be used to study inter-laminar interactions. By
varying light intensity, the UOA allowed us to localize photostimulation to single or multiple
cortical layers, and by varying the depth of insertion (or the customizable shank length) it is
possible to target distinct layers. Differences in onset latency of light-evoked responses across
the cortical depth could be used to distinguish distinct network activity patterns following
different patterns of UOA stimulation. For example, at low light irradiance, direct neuronal
activation was initially localized to the layers nearest optrode tip termination before spreading
trans-synaptically to other layers. Increasing light irradiance directly contributed to firing rate
increases in deep layers below the UOA tips while firing rates in L4C increased less compared to
firing rates evoked at lower intensities (suggesting local activation of higher threshold inhibitory
networks).

We also demonstrated that by varying the distance between the stimulation site/s on the
UOA and the recording electrode, the UOA can be used to study local versus long-distance intra-
areal interactions, while when coupled to recordings in a different cortical area (using LEAs or
UEAs) the UOA allows investigations of the functions of inter-areal feedforward and feedback
circuits. Moreover, used in conjunction with c-fos IHC, we were able to identify multisynaptic
interactions within and beyond the photostimulated area. Photostimulation via the UOA
increased c-fos expression over distances much > 8mm (well beyond the stimulated cortical
area), but spiking activity could not be evoked beyond ~3 mm from the stimulated site,
indicating that c-fos expression can be used to reveal subthreshold activity induced by network

interactions. This is consistent with previous demonstrations of c-fos expression several synapses

away from the electrically stimulated site, indicating c-fos can be used for functional mapping of

neuronal circuits[40].

Importantly, despite its limited shank length (up to 2.5 mm), the UOA can still be

employed to study cortico-subcortical interactions, e.g. through modulation of axon terminals of

deep nuclei within cortex, and recordings of postsynaptic cortical neurons in the same cortical

area and/or layer.

Future developments of this device will involve: (i) addition of red µLEDs for dual color

optogenetic neuromodulation through each needle shank and interstitial site, (ii) addition of

electrical recording capabilities and (iii) optimization for chronic use in NHPs and humans.

In conclusion, the UOA will enable studies addressing long-standing fundamental

questions in neuroscience, e.g., regarding the role of cortico-cortical feedback and cortical layers

in the model system closest to humans. As many human neurological and psychiatric disorders

have been linked to abnormalities in cortical circuit[4,5], this technology can improve our

understanding of the circuit-level basis of human brain disorders, and will pave the way for a

new generation of precise neurological and psychiatric therapeutic interventions via cell type-

specific optical neural control prosthetics.

564   **ONLINE METHODS**

Device Fabrication, Characterization, and Benchmarking

Fabrication and testing of the first generation UOA devices was previously reported[29,45]. The

second-generation devices used in this study included an optical interposer layer that limits

emission from the µLED array to the shank sites for illumination of deep cortical tissue.

Fabrication. A 2 mm-thick, 100mm diameter Schott Borofloat 33 glass wafer used to

construct the optrode needles was anodically bonded to a freshly cleaned 0.1mm thick, 100 mm

diameter intrinsic Si wafer serving as an optical interposer. The Si and Borofloat wafers were

coarsely aligned and bonding performed using an EVG 520 anodic bonder. The optical vias were

patterned in the Si interposer by deep reactive ion etching (DRIE) using a Bosch process. A 10-

µm-thick AZ9260 soft mask was photolithographically patterned to define the array of 80×80

μm^2 optical vias for shank and interstitial illumination for the DRIE process. The bonded wafer

was then sub-diced into *modules* of 9 to 16 UOAs using a DISCO 3220 dicing saw.

UOA modules were mounted to a carrier wafer using WaferGrip™ (Dynatex

International, Santa Rosa, CA). The glass shanks were cut with the DISCO 3220 using the

previously reported process[29,45]. Briefly, beveled blades were first used to generate pyramidal

tips on the surface, followed by standard profile blades to form the shanks. The shanks on a

module were then etched to a nominal 110 µm thickness using a mixture of hydrofluoric (49%)

and hydrochloric (37%) acid in a 9:1 ratio. The die was then demounted and cleaned, and the

shanks were smoothened to decrease light scattering using a 725 °C heat treatment for 2 hours in

a vacuum furnace. UOA modules were then singulated into individual 4×4 mm^2 UOAs using the

DISCO 3220.

Arrays of µLEDs on thinned (150µm) sapphire substrates, from the Institute of Photonics

at University of Strathclyde, were integrated with the UOA using closed-loop optical alignment

to the optical vias on individual UOAs at Fraunhofer IZM (Berlin, Germany)[29], and bonded

using index-matched epoxy. At the University of Utah, passive matrix µLED pads were wire

bonded to an ICS-96 connector (Blackrock Microsystems, Salt Lake City, UT) using insulated

gold alloy wire. The wire bundle and back-side of the UOA were then potted in NuSil MED-

4211 silicone, respectively, followed by overcoating with a 6µm-layer of Parylene C.

Bench Testing. To characterize the electrical and optical performance of the finalized

devices, the latter were attached to a custom switch board for matrix addressing the individual

optrode shanks. The switch board consisted of a matrix arrangement of parallel connected

mechanical switches and electrical relays, 10 sets for the anodes and 10 sets for the cathodes.

This enabled both manual and automated activation of individual optrode shanks or optrode

patterns. For the automated activation and testing, the relays were connected to Arduino boards

which received commands from the lab computer. To prevent voltage spikes originating from the

switching of the channels from damaging the µLEDs, the anode paths also contained a small

filter circuit consisting of capacitors and Zehner diodes (break-down voltage: 8.2V). For the

automated testing, the UOAs were inserted into the opening of an integrating sphere that was, in

turn, connected to a photodetector and power meter (Newport 2832-C Dual-Channel Power

Meter). The calibration factor of the integrating sphere was determined using a fiber coupled

LED prior to the experiment. Then the UOAs were connected to the switch board, and the latter

was connected to a source measure unit (Keithley 236 Source Measure Unit) for the
measurement. The automated characterization was conducted as follows: the switch board's
Arduino boards received the command to switch to an individual optrode shank using the relays.
Then the source measure unit applied a voltage pulse measurement pattern (pulse length 100
msec, pause between pulses 1900 msec to prevent heat buildup) sweeping the voltage from 0 to
7.2V (or until the compliance current of 100mA was reached) with each pulse increasing by
100mV. For each pulse, the resulting current and the output optical power were recorded; the
optical power was then corrected using the integrating sphere calibration factor. This was
repeated for each individual optrode shank of the device for a full characterization.

To ensure the stability of the device for an acute *in vivo* experiment, additional voltage
transient measurements were made before and after a 48-hour soak test in phosphate-buffered
saline (PBS) at 37 °C. Further, an electrode was immersed in solution to verify encapsulation
integrity, as evidenced by lack of shorting to solution.

For the *in vivo* experiments, the switch board was upgraded two-fold: first, transistors
were added to the cathode channels to allow for turning the device on and off based on an
external TTL trigger. However, we found that turning on the optrodes using the trigger signal
directly induced too strong a capacitively-coupled voltage signal in the recording. Therefore, as a
second upgrade, an additional Arduino board with digital-analog-converter was added that
received the external trigger and introduced rise and fall times to the square wave. This reduced
the capacitively-coupled interference to a level below measurable when both the LEA and the
UOA were in close proximity in 1xPBS solution prior to the *in vivo* experiment. During the
experiment, the voltage for the UOA was supplied by a lab power supply via the switch board,
and the switches were operated manually to define the required patterns.

Modeling. To understand light spread in tissue, the optical output of the device was
modeled using ray-tracing software (Optics Studio 12, in non-sequential mode). This model has
been described previously[29]. Brain tissue was modeled using a Henyey-Greenstein scattering
model, with a scattering coefficient of 10 mm^{-1}, absorption coefficient of 0.07 mm^{-1}, and
anisotropy of 0.88[46]. Each needle was modelled individually using its measured optical output at
the given voltage level. To generate the cross-section images from a simultaneously illuminated
column (**Fig. 1G**), the light output from the 10 needles in that column were summed.

Animals

A total of 3 adult female Cynomolgus monkeys (*Macaca fascicularis*) were used in this study. The left hemisphere of one animal (case MK421-LH) was used for the *in vivo* electrophysiological testing of the active UOA (integrated with the μLED array). The right hemisphere from the same animal (MK42-RH), and 3 hemispheres from 2 additional animals (MK414RH and LH, and MK422-RH) were used for c-fos testing of the passive UOA (without an integrated μLED array). All procedures conformed to the National Institutes of Health Guide for the Care and Use of Laboratory Animals and were approved by the University of Utah Institutional Animal Care and Use Committee.

Survival Surgical Procedures and Viral Injections

Animals were pre-anesthetized with ketamine (10 mg/kg, i.m.), intubated, placed in a stereotaxic apparatus, and artificially ventilated. Anesthesia was maintained with isofluorane (1–2.5% in 100% oxygen). Heart rate, end tidal CO_2, oxygen saturation, electrocardiogram, and body temperature were monitored continuously. I.V. fluids were delivered at a rate of 5/cc/kg/hr. The scalp was incised and a craniotomy and durotomy were performed over area V1 (n=2 animals, MK421-LH and MK414-RH), or rostral to the precentral gyrus, roughly above the supplementary motor area (SMA; n=1, MK422-RH). We injected a 1:1 viral mixture of AAV9.CamKII.4.Cre.SV40 and AAV9.CAG.Flex.ChR2.tdTomato (Addgene Catalog #s: 105558, and 18917, respectively). We have previously found that this method nearly eliminates retrograde expression of transgenes[10]. The viral mixture was slowly (~15nl/min) pressure-injected (250-350nl repeated at 2 or 3 cortical depths between 0.5 and 1.5 mm from the cortical surface) using a picospritzer (World Precision Instruments, FL, USA) and glass micropipettes (35-45μm tip diameter). After each injection, the pipette was left in place for 5-10 min before retracting, to avoid backflow of solution. A total of 5-6 such injections, each 500-750nl in total volume, and spaced 1.5-2mm apart, were made in two animals (MK421-LH, MK414-RH) while the third animal (MK422-RH) received 2 x 1,050nl injections. These injections resulted in a region of high viral expression roughly 4-6 mm in diameter (as an example see **Extended Data Fig. 3A Right**). Following viral injections, a sterile silicone artificial dura was placed on the cortex, the native dura was sutured and glued onto the artificial dura, covered with Gelfoam to fill the craniotomy, and the latter was sealed with sterile parafilm and dental acrylic. Anesthesia

669 was discontinued and the animal returned to its home cage. After a survival period of 5-10

670 weeks, to allow for robust ChR2 expression, the animals were prepared for a terminal UOA

671 photostimulation procedure.

Terminal Surgical Procedures and UOA Insertion

Monkeys were pre-anesthetized and prepared for experiments as described above. Anesthesia

and paralysis were maintained by continuous infusion of sufentanil citrate (5–10□µg/kg/h) and

vecuronium bromide (0.3□mg/kg/h), respectively. Vital signs were continuously monitored for

the duration of the experiment, as described above. Following suture removal and scalp incision,

the craniotomy and durotomy were enlarged to allow space for device implantation, and ChR2

expression was verified *in* vivo using a custom fluorescent surgical microscope (Carl Zeiss,

GmbH; **Fig. 2B**). UOAs were positioned over cortical regions of high tdT/ChR2 expression (e.g.

Figs. 2B,6B,F), and then inserted using a high speed pneumatic hammer typically used for

insertion of Utah Electrode Arrays[34] (Blackrock MicroSystems, Salt Lake City, UT). Parameters

used for insertion were 20 psi for 30 msec, using a 1 mm-long inserter, in order to achieve partial

insertion of the UOA, so as to minimize insertion trauma on the cortex. In two animals used for

c-fos experiments after partial insertion with the pneumatic inserter, the UOA was gently pushed

down to achieve deeper insertion.

Photostimulation

We implanted two types of UOA devices: (i) a 10x10 UOA with fully integrated µLED arrays

(also referred to as "active" device; n=1 device in 1 animal, MK421-LH; see **Fig. 2A-C**), and

(ii): 10x10 UOAs with an optical interposer integrated into the sapphire backplane, but with no

µLED array for light delivery (referred to as "passive" devices; n=3 devices in 3 hemispheres

from 2 animals, MK414-RH, MK414-LH, MK422-RH). The active device was used for

electrophysiological testing experiments, while the passive devices were used for the c-fos

experiments.

Active Device (Electrophysiology). Photostimulation with the active UOA occurred via

the integrated µLED array. Photostimulation parameters were 5Hz, 100 msec-pulse duration for

1 sec, followed by 1-10sec inter-trial interval (longer intervals were used at the higher

photostimulation intensities). We varied the spatial pattern (single µLED along column 1, whole

single columns, and all µLEDs across the entire UOA) and intensity (from 2.8 to 7.8V input

intensity) of photostimulation as described in the Results section.

Passive Devices (c-Fos). Selective photostimulation via passive devices was obtained by

illuminating a subset of UOA needles with an appropriately positioned fiber-coupled 473nm

laser (400 µm multimode optic fiber, ThorLabs Newton, NJ; laser: Laserwave, Beijing, China)

held in place with a stereotaxic tower. We used a collimating lens (ThorLabs, Newton, NJ) to

restrict spot size to ~1.5mm in diameter. To shield stray light, we covered any exposed tissue

around the illuminated area, as well as the non-illuminated portions of the UOA, with an opaque

(black) artificial dura. For each UOA we stimulated 2 or 3 separate sites. At each site we used

phasic photostimulation (50Hz for 2.5 min, 2.5 min pause, and 20Hz for an additional 2.5 min;

pulse duration was 10 msec) at 3.8mW power output (corresponding to an estimated irradiance

of 15-19mW/mm^2).

713   Electrophysiological Recordings

Extracellular recordings were made in V1 with 24-channel linear electrode arrays (LEAs; V-

Probe, Plexon, Dallas, TX; 100µm contact spacing, 300µm from tip to first contact, 20µm

contact diameter). The LEAs were inserted into the cortex next to the UOA to a depth of 2.4-

2.6mm, slightly angled laterally (towards the UOA) and posteriorly. We made a total of 3

penetrations (P1-P3; **Extended Data Fig. 3A**), of which only P2 and P3 provided useful data.

After UOA and LEA were inserted into the cortex, we applied a layer of Dura-Gel

(CambridgeNeuroTech, Cambridge, UK) over the cortex and UOA, to prevent the cortex from

drying and stabilize the recordings. A 128-channel recording system (Cerebus, Blackrock

Microsystems, Salt Lake City, UT) was used for standard signal amplification and filtering.

Multi-unit spiking activity was defined as any signal deflection that exceeded a voltage threshold

(set at 4 x the SD of the signal on each channel). Threshold crossings were timestamped with

sub-millisecond accuracy. We did not record responses to visual stimuli but only to UOA

photostimulation performed as described above; thus, the monkey's eyes were closed during the

duration of the experiment.

729   Analysis of Electrophysiological Data

We analyzed MU spiking responses from a total of 45 contacts deemed to lie within the

parafoveal representation of V1 in two penetrations (out of 3 total, see above) for which neural

activity was modulated by photostimulation via the active UOA. For the results presented in

Figures 3-5, quantitative analysis was limited to contacts on which MU activity was stimulus

modulated (one-way ANOVA comparing spike rates during full one-second photostimulation

trials with spike rates during control periods of equivalent duration, $p < 0.01$).

To quantify the change in MU firing rates, relative to background, during

photostimulation we calculated firing rates for all pulse epochs within all trials and then

compared them to the average background rate. To estimate the preference at each recording site

for stimulation across the full range of tested UOA locations (**Fig. 3**), we regressed average

evoked-responses on UOA stimulation site and intensity. Preliminary analyses had revealed a

non-monotonic relationship between stimulation intensity and response on many contacts (cf.

Fig. 2G), thus we included a quadratic term in the regression model.

CSD analysis. For the CSD analysis shown in **Fig. 2D-F,** current source density (CSD)

was calculated from the band-pass filtered (1-100Hz) and pulse-aligned and averaged LFP, using

the kernel CSD toolbox (kCSD_Matlab)[47]. CSD was calculated as the second spatial derivative

of the LFP signal, reflecting the net local transmembrane currents generating the LFP. The depth

profile of the CSD was estimated by interpolating every 10μm. To facilitate comparisons across

conditions, CSDs from different conditions were normalized to the standard deviation (SD) of

the baseline (50 msec prior to pulse onset) after subtraction of the baseline mean.

Onset Latency. To quantify the onset latency of MU responses, we either: (i) calculated

the average peri-stimulus time histogram (PSTH) from all pulse-aligned responses (e.g. **Fig. 4**)

or (ii) estimated a PSTH separately for the response to each pulse (**e.g. Extended Data Fig. 4**).

Peristimulus time histograms (PSTHs) were estimated via an adaptive algorithm in which the

MU raster was first convolved with a Gaussian kernel of fixed width (3 msec bandwidth), kernel

width was then adapted so that the number of spikes falling under the kernel was the same on

average across the response (http://chronux.org[48]). We then subtracted the mean baseline

response from the stimulus-evoked response. For each response measure, i.e. either the average

or pulse-by-pulse PSTHs, we took the time at which the response reached 25% of the peak as the

onset latency (results were qualitatively similar using 15% and 35% criteria; data not shown).

We report the former measure as the mean onset latency in **Figures 4-5**. We used the latter

measure to test for differences in onset latency across contacts within and across UOA

stimulation parameters (**Figs. 4-5 and Extended Data Fig. 4**).

Statistical Analysis. Stimulus-evoked firing rates were calculated from pulse-aligned or

trial-aligned responses and baseline corrected (mean baseline activity subtracted). We

determined responsiveness to stimulation via a one-way ANOVA comparing firing rates during

the full 1-second trial period with inter-leaved control periods of equivalent duration; MU

activity at an LEA recording site was deemed responsive if there was a significant difference

between stimulation and control trials at the $p=0.01$ level. To estimate the selectivity of MU

activity for stimulation at different UOA sites we performed a multiple linear regression, with

UOA column, row, and intensity as independent variables and pulse-aligned, baseline corrected,

firing rates as the dependent measure. To test for differences in the goodness-of-fit of models

with- and without a quadratic term, we used a two-sample Kolmogorov-Smirnov test. We

assessed the effects of varying UOA stimulation site and intensity on response amplitude or

onset latency using ANOVA models followed by the Tukey-Kramer test for post-hoc

comparisons.

c-Fos Experiments

We used 4 hemispheres from 3 animals for these experiments (MK414-RH and LH, MK422-RH,

and MK421-RH). Two of these animals (MK422 and MK414) were prepared for a terminal

experiment (as described above) 5 or 10 weeks, respectively, after the viral injections, and a

passive UOA was inserted in regions of high tdT/ChR2 expression in the injected hemisphere. In

one of these animals (MK422-RH), UOA insertion was preceded by glutamate block (see

below). After UOA insertion, photostimulation was performed via an optical fiber-coupled laser

through the UOA, as described above. Two additional hemispheres in 2 animals (MK414-LH

and MK421-RH) were used as controls. Specifically, case MK414-LH received insertion of a

passive UOA in non-opsin expressing SMA cortex, and was euthanized 4 hrs following UOA

insertion without receiving any photostimulation. As a separate control, in case MK421-RH we

performed surface photostimulation of SMA cortex not expressing opsins, using a fiber-coupled

laser and a collimating lens and the same photostimulation protocol described above for other c-

fos experiments; no UOA was inserted in this case. In all animals, UOA insertion and/or

photostimulation were performed after a 10-14-hour period of eye closure and at least 5 hours

after completion of surgery, and the animals were euthanized 75 minutes after completion of the

photostimulation protocol.

794    ***Pharmacological Blockade of Local Glutamate Signaling.*** To compare changes in c-fos

expression due to direct local optogenetic activation with indirect local and long-range changes

due to synaptic increases in excitatory glutamatergic neurotransmission downstream of the

directly-activated neurons, in one case (MK422-RH) we applied the selective glutamate AMPA

receptor antagonist 2,3-dihydroxy-6-nitro-7-sulfamoyl-benzoquinoxaline-2,3-dione (NBQX,

5mM) (Tocris BioSciences, Minneapolis, MN). NBQX was applied topically prior to UOA

insertion, by soaking a piece of Gelfoam placed over ChR2-expressing SMA cortex with 1ml of

the drug solution. The drug was allowed to passively diffuse through the cortical layers for 90

minutes, during which 100-200µl of the solution were applied every 15 minutes to ensure

saturation of the Gelfoam, after which the Gelfoam was removed and the passive UOA inserted

over the region of glutamate block. Photostimulation was performed as described above for the

passive device.

807  **Histology**

On completion of the experiments, the animals were euthanized by an overdose of Beuthanasia

(0.22 ml/kg, i.v.) and perfused transcardially with saline for 2–3 min, followed by 4%

paraformaldehyde (PFA) in 0.1M phosphate buffer for 20 min to fix the brain. The brains were

post-fixed overnight in the same fixative, sunk in cryoprotectant 30% sucrose solution, and

sectioned at 40µm on a freezing microtome. The hemisphere used for electrophysiological

testing of the active UOA (MK421-LH) was sectioned tangentially. One in 3 sections were wet-

mounted and imaged for fluorescent tdT-label at 10x magnification. The same sections were then

reacted for cytochrome oxidase (CO) to reveal cortical layers and the location of UOA and LEA

insertions visible as discolorations in CO staining (**Extended Data Fig. 3A Left**).

817    All other hemispheres used for c-fos experiments were sectioned sagittally. One full

series of sections (1:3) were immunoreacted for c-fos by overnight incubation in primary

antibody (1:500 rabbit anti-c-fos, Ab 19089, Abcam, MA) at room temperature, followed by 2

hrs incubation in near-infrared secondary antibody (1:200 donkey anti-rabbit IgG-AF647,

Jackson ImmunoResearch, PA) at room temperature. Sections were then wet-mounted,

counterstained with blue fluorescent Nissl (1:100 N21479, Thermo Fisher Scientific, MA), by

dripping the solution onto the slide-mounted sections every 5 min for 20 min, rinsed, and

coverslipped and sealed with CoverGrip™ Coverslip Sealant (Biodium, CA).

Tissue Imaging

Imaging of tissue sections was performed on a Zeiss Axio Imager.Z2 fluorescent microscope

(Zeiss GmbH, Germany) with a Zeiss X-cite 120 LED Boost light source, using a 10x objective

and an Axiocam 506 mono camera (Zeiss GmbH, Germany). Image files were created and

analyzed using Zen 2.6 Blue Software (Zeiss GmbH, Germany). The light intensity was set to

100%, and the exposure time for each channel was kept the same between images. The

tangentially-sectioned hemisphere (MK421-LH) was imaged as described above. In all other

cases, each sagittal section was imaged in 3 channels simultaneously, one channel for tdT/ChR2

(red- but note the color was artificially changed to green in **Fig. 6B, F**), one channel for Alexa-

647-c-Fos (far-red), and the third channel for 435-455 Nissl (blue).

Analysis of c-Fos Expression

To quantify c-fos expression, c-fos+ cells were plotted and counted in sampled areas,

using Neurolucida software 2006 (Microbrightfield Bioscience, VT). For each case, we selected

for counts 5 sections spaced 1 mm apart encompassing the area of UOA insertion and/or

photostimulation (for the light-only case). In each section, we plotted and counted cells within

three 200μm-wide windows spanning all cortical layers, one positioned at or near the center of

the UOA insertion region (or of photostimulation-only), and the other two located at distances of

4mm and 8mm, respectively, from the center of the UO insertion (**Fig. 6**). Thus, a total of 15

regions of interest (ROIs) were counted for each case. The laminar distribution of c-fos+ cells

was analyzed by tracing the layers on the Nissl stain and counting the number of c-fos+ cells

within each layer in Neurolucida. Statistical differences in c-fos+ cell counts among

experimental and control cases, and across distances were estimated using a one-way ANOVA

with post-hoc comparisons and Bonferroni correction).

REFERENCES

1. Deisseroth, K. Optogenetics: 10 years of microbial opsins in neuroscience. *Nat. Neurosci.* 18, 1213-1225 (2015).

2. El-Shamayleh, Y. & Horwitz, G.D. Primate optogenetics: Progress and prognosis. *Proc. Natl. Acad. Sci. U.S.A.* 116, 26195-26203 (2019).

3. Goldberg, M.E. The neurology clinic needs monkey research. *Proc. Natl. Acad. Sci. U.S.A.* 116, 26255-26258 (2019).

4. Lynall, M.-E. et al. Functional Connectivity and Brain Networks in Schizophrenia. *J. Neurosci.* 30, 9477-9487 (2010).

5. Belmonte, M.K. et al. Autism and Abnormal Development of Brain Connectivity. *J. Neurosci.* 24, 9228-9231 (2004).

6. Vitek, J.L. & Johnson, L.A. Understanding Parkinson's disease and deep brain stimulation: Role of monkey models. *Proc. Natl. Acad. Sci. U.S.A.* 116, 26259-26265 (2019).

7. Andersen, R.A., Aflalo, T. & Kellis, S. From thought to action: The brain-machine interface in posterior parietal cortex. *Proc. Natl. Acad. Sci. U.S.A.* 116, 26274-26279 (2019).

8. Picaud, S. et al. The primate model for understanding and restoring vision. *Proc. Natl. Acad. Sci. U.S.A.* 116, 26280-26287 (2019).

9. Jarvis, S. & Schultz, S.R. Prospects for Optogenetic Augmentation of Brain Function. *Front. Syst. Neurosci.* 9, 157 (2015).

10. Nurminen, L., Merlin, S., Bijanzadeh, M., Federer, F. & Angelucci, A. Top-down feedback controls spatial summation and response amplitude in primate visual cortex. *Nature Commun.* 9, 2281 (2018).

11. Siu, C., Balsor, J., Merlin, S., Federer, F. & Angelucci, A. A direct interareal feedback-to-feedforward circuit in primate visual cortex. *Nat Commun* 12, 4911 (2021).

12. Mehta, P. et al. Functional Access to Neuron Subclasses in Rodent and Primate Forebrain. *Cell Rep.* 26, 2818-2832 e2818 (2019).

13. Vormstein-Schneider, D.C. et al. Viral manipulation of functionally distinct interneurons in mice, non-human primates and humans. *Nat. Neurosci.* 23, 1629-1636 (2020).

14. Dimidschstein, J. et al. A viral strategy for targeting and manipulating interneurons across vertebrate species. *Nat. Neurosci.* 19, 1743-1749 (2016).

15. Tremblay, S. et al. An Open Resource for Non-human Primate Optogenetics. *Neuron*

108, 1075-1090 e1076 (2020).

16.   Angelucci, A. et al. Circuits and mechanisms for surround modulation in visual cortex.
Ann. Rev. Neurosci. 40, 425-451 (2017).

17.   Moore, T. & Zirnsak, M. Neural Mechanisms of Selective Visual Attention. *Annu. Rev.*
Psychol. 68, 47-72 (2017).

18.   Barbas, H., J., W., Joyce, M.K.P. & Garcia-Cabezas, M.A. *J. Neurophysiol.* 120, 2659-
2678 (2018).

19.   Rockland, K.S. & Pandya, D.N. Laminar origins and terminations of cortical connections
of the occipital lobe in the Rhesus monkey. *Brain Res.* 179, 3-20 (1979).

20.   Usrey, W.M. & Sherman, S.M. Corticofugal circuits: Communication lines from the
cortex to the rest of the brain. *J. Comp. Neurol.* 527, 640-650 (2019).

21.   Hubel, D.H. & Wiesel, T.N. Receptive fields and functional architecture of monkey
striate cortex. *J. Physiol. (Lond.)* 195, 215-243. (1968).

22.   Huang, X., Elyada, Y.M., Bosking, W.H., Walker, T. & Fitzpatrick, D. Optogenetic
assessment of horizontal interactions in primary visual cortex. *J. Neurosci.* 34, 4976-4990
(2014).

23.   Roy, A. et al. Optogenetic spatial and temporal control of cortical circuits on a columnar
scale. *J. Neurophysiol.* 115, 1043-1062 (2016).

24.   Rajalingham, R. et al. Chronically implantable LED arrays for behavioral optogenetics in
primates. *Nat. Methods.* 18, 1112-1116 (2021).

25.   Owen, S.F., Liu, M.H. & Kreitzer, A.C. Thermal constraints on in vivo optogenetic
manipulations. *Nat. Neurosci.* 22, 1061-1065 (2019).

26.   Wang, J. et al. Integrated device for combined optical neuromodulation and electrical
recording for chronic in vivo applications. *J. Neural Eng.* 9, 016001 (2012).

27.   Ozden, I. et al. A coaxial optrode as multifunction write-read probe for optogenetic
studies in non-human primates. *J. Neurosci. Methods* 219, 142-154 (2013).

28.   Anikeeva, P. et al. Optetrode: a multichannel readout for optogenetic control in freely
moving mice. *Nat. Neurosci.* 15, 163-170 (2011).

29.   McAlinden, N. et al. Multisite microLED optrode array for neural interfacing.
Neurophotonics 6, 035010 (2019).

30.   Scharf, R. et al. Depth-specific optogenetic control in vivo with a scalable, high-density
muLED neural probe. *Sci. Rep.* 6, 28381 (2016).

31.   Maynard, E.M., Nordhausen, C.T. & Normann, R.A. The Utah intracortical Electrode

Array: a recording structure for potential brain-computer interfaces. *Electroencephalogr. Clin. Neurophysiol.* 102, 228-239 (1997).

32. Lin, J.Y. Optogenetic excitation of neurons with channelrhodopsins: light instrumentation, expression systems, and channelrhodopsin variants. *Prog. Brain Res.* 196, 29-47 (2012).

33. Wiegert, J.S., Mahn, M., Prigge, M., Printz, Y. & Yizhar, O. Silencing Neurons: Tools, Applications, and Experimental Constraints. *Neuron* 95, 504-529 (2017).

34. Kim, K. et al. Artifact-free and high-temporal-resolution in vivo opto-electrophysiology with microLED optoelectrodes. *Nat. Commun.* 11, 2063 (2020).

35. Callaway, E.M. & Wiser, A.K. Contributions of individual layer 2-5 spiny neurons to local circuits in macaque primary visual cortex. *Vis. Neurosci.* 13, 907-922 (1996).

36. Pluta, S. et al. A direct translaminar inhibitory circuit tunes cortical output. *Nat. Neurosci.* 18, 1631-1640 (2015).

37. Qi, G. & Feldmeyer, D. Dendritic Target Region-Specific Formation of Synapses Between Excitatory Layer 4 Neurons and Layer 6 Pyramidal Cells. *Cereb. Cortex* 26, 1569-1579 (2016).

38. Acker, L., Pino, E.N., Boyden, E.S. & Desimone, R. FEF inactivation with improved optogenetic methods. *Proc. Natl. Acad. Sci. U.S.A.* 113, E7297-E7306 (2016).

39. Andrei, A.R., Pojoga, S., Janz, R. & Dragoi, V. Integration of cortical population signals for visual perception. *Nat. Commun.* 10, 3832 (2019).

40. Sagar, S.M., Sharp, F.R. & Curran, T. Expression of c-fos protein in brain: metabolic mapping at the cellular level. *Science* 240, 1328-1331 (1988).

41. Kovacs, K.J. c-Fos as a transcription factor: a stressful (re)view from a functional map. *Neurochem. Int.* 33, 287-297 (1998).

42. Tamura, K. et al. A glass-coated tungsten microelectrode enclosing optical fibers for optogenetic exploration in primate deep brain structures. *Journal of neuroscience methods* 211, 49-57 (2012).

43. Stark, E., Koos, T. & Buzsaki, G. Diode probes for spatiotemporal optical control of multiple neurons in freely moving animals. *J. Neurophysiol.* 108, 349-363 (2012).

44. Royer, S. et al. Multi-array silicon probes with integrated optical fibers: light-assisted perturbation and recording of local neural circuits in the behaving animal. *Eur. J. Neurosci.* 31, 2279-2291 (2010).

45. Scharf, R. et al. A compact integrated device for spatially-selective optogenetic neural stimulation based on the Utah Optrode Array. *Proc. SPIE* 10482, 104820M (2018).

951   46.   Bernstein, J.G. et al. Prosthetic systems for therapeutic optical activation and silencing of
952   genetically-targeted neurons. *Proc. SPIE Int. Soc. Opt. Eng.* 6854, 68540H (2008).

953   47.   Potworowski, J., Jakuczun, W., Leski, S. & Wojcik, D. Kernel current source density
954   method. *Neural. Comput.* 24, 541-575 (2012).

955   48.   Mitra, P. & Bokil, H. Observed Brain Dynamics. (Oxford University Press, New York;
956   2008).
957

958   **ACKNOWLEDGMENTS**

We thank Kesi Sainsbury for histological assistance, Seminare Ta'afua for help with experiments, and
Julian Haberland and Christine Kallmayer at the Fraunhofer-Institut für Zuverlässigkeit und
Mikrointegration (IZM) for µLED-to-interposer bonding. This work was supported primarily by a
BRAIN grant (U01 NS099702) from the National Institute of Health (NIH) to S.B., A.A. L.R and S.M.
Additional support was provided by grant from the NIH (R01 EY026812, R01 EY019743, R01
EY031959) and the National Science Foundation (IOS 175543) to A.A, and an unrestricted grant from
Research to Prevent Blindness, Inc. and a core grant from the NIH (EY014800) to the Department of
Ophthalmology, University of Utah.

967
968   **AUTHOR CONTRIBUTIONS**

Conceptualization: A.C., A.I., C.F.R., N.M., L. R., K.M., S.B. A.A. Device Fabrication and
Bench Testing: C.F.R., N.M., L.R.,K.M., M.D.D., S.B. Modeling: N.M., K.M. *In Vivo*
Electrophysiology Testing: A.C., D.C., C.F.R., F.F., A.A. *In Vivo* cFos Testing: A.I., J.B., F.F.,
J.D.R., A.A., A.C. Analysis of Electrophysiology Data: A.C. Analysis of cFos Data: A.I., J.B.,
F.F., A.A. Writing-Original Draft: A.C., L.R., S.B., A.A. Writing–Review/Editing: all authors.
Supervision & Funding Acquisition: A.A., S.B., K.M., L.R.

975
976   **COMPETING INTERESTS STATEMENT**
977   The authors declare no competing interests.

 FIGURE LEGENDS

Figure 1. UOA Design and Optical Properties

(A) Schematics of UOA design superimposed to a Nissl-stained coronal section of macaque V1 showing the layers. The UOA consists of 3 main components: a µLED array (B), an optical interposer (C) and a glass needle array (D). **(B)** Two interleaved µLED arrays on a sapphire substrate are shown in this image; the first 10x10 array is needle-aligned for deep layer stimulation, the second 9x9 interstitial array lies in-between the first for surface stimulation. The interstitial array, although built into the UOA, was not used in this study. Scale bar: 1mm. **(C)** A region of the silicon optical interposer corresponding to approximately the size of the *white box in (B)*; the optical "vias" are etched through the silicon and matched to the size of a µLED (80x80µm^2). Scale bar: 200µm. **(D)** High magnification image of the glass needle shanks bonded to the interposer. Scale bar: 200µm. **(E)** <u>LEFT:</u> The µLED on sapphire and needle array components are integrated into the final device, wire-bonded, and encapsulated. The image shown is a representative device. The integrated UOA used in this study consisted of 10x10 glass needle shanks, 1.6 mm long (to target deeper layers) and 100-110µm wide, with tip apex angles about 64º. An image of the actual device used in the *in vivo* testing studies, after completion of the experiment and explantation is shown in **Extended Data Fig. 2**. Scale bar: 1mm. <u>RIGHT:</u> Example spatial patterns of device operation. **(F)** Average output optical power (in mW) across each needle tip at different drive voltages (currents), when the entire UOA was turned on (*top left inset*). *Blue and gray bars*: needle shanks with estimated tip irradiances above and below, respectively, the 1mW/mm^2 threshold for ChR2 activation. **(G)** <u>LEFT:</u> Ray trace model of light spread in cortical tissue when a single µLED (in column 1 and row 8, i.e. the closest to the linear electrode array —LEA– in penetration 2 –P2– used for the electrophysiological testing experiment, and indicated as a *black dot*) is activated at various input voltages (% of maximum intensity used), with power output calibrated to the bench tests. <u>RIGHT:</u> Model of light spread in tissue when column 1 (the nearest to the LEA in P2 and P3) is activated at various input voltages. *Green contour* encloses tissue volume within which the light irradiance is above 1mW/mm^2, the threshold for ChR2 activation. Scale bars: 400µm.

Figure 2. Laminar Distribution of Responses Induced by UOA Photostimulation.

(A) The UOA inserted in macaque V1. **(B)** Same field of view as in (A) shown under fluorescent illumination to reveal expression of the red fluorescent protein tdTomato (*arrow*). **(C)** Preparation for recording electrophysiological responses to photostimulation. A 24 channel linear electrode array (LEA) was inserted next to the UOA (guide tube protecting array marked "LEA") slightly angled laterally (towards the UOA) and posteriorly. Here the UOA is partially covered with a piece of Gelfoam. **(D)** Current Source Density analysis (CSD; Left) and multiunit (MU) spiking activity (Right) signals recorded through the depth of V1 in P2 in response to phasic UOA photostimulation (pulse parameters: 100 msec pulse duration, 5Hz, 7.4mW/mm^2; pulse periods denoted as *blue bars* above MU plot). Here, all 100 needle-aligned µLED sites ("whole µLED array" condition) were activated simultaneously. CSD responses to each 100 msec pulse were zero-aligned, while MU activity is shown for the full 5Hz pulse train. *The dashed lines in the CSD panel* demarcate the borders of layer 4C (L4C)*; the gray shaded region in the MU activity panel* delimits the extent of L4C. **(E)** Same as in (D), but for photostimulation irradiance of 0.82 mW/mm^2. **(F)** Same as in (D-E), but following surface photostimulation of V1 via a laser-coupled optical fiber with pulse parameters of 10 msec, 5Hz, 2.2mW/mm^2. **(G-J)** Left: Relative cortical depth of each contact on P2 (*black dot in the insets*) is plotted versus the relative response (% firing rate increase over baseline) to UOA stimulation for different 450nm µLED illumination patterns (*insets*). Different colored traces are data for different photostimulation intensities (expressed as voltage or percent of max intensity used). *Gray area: extent of L4C; dashed lines*: approximate location of the L4A/4B (upper) and L5/L6 (lower) borders. Right: PSTHs with and without µLED activation are shown for the same contact on the LEA in L4C (marked by the *black circle)* across conditions. *Dashed line in the PSTH*: pulse periods.

Figure 3. Tangential Extent of Responses Induced by UOA Photostimulation.

A) Examples of model fits to single µLED and single column photostimulation for an example contact from P2, the one that showed the largest relative response increase across these stimulation conditions. This contact preferred stimulation in the proximal UOA columns 1-2, at sites closer to the top of the device (rows 9-7). The schematics on the left of the UOA and of the LEA-P2 indicates as *blue shading* the UOA sites represented in the heat map, and as a *red dot* the contact on the LEA whose response is mapped on the right. The *horizontal lines* and *gray*

shading on the LEA schematics mark the pial and white matter, and L4C boundaries, respectively. Color scale applies to panels (A-B, E-F). **(B)** Average normalized fitted responses across all responsive contacts in P2 (*red dots* in schematics of LEA to the left). **(C)** Change in response in the column direction for P2. Average relative response amplitude (% of peak response) is plotted as a function of stimulation intensity and distance along a straight line extending from the preferred UOA site in the column direction. Data averaged across all contacts. **(D)** Change in response in the row direction for P2. Average relative response amplitude (% of peak response) is plotted as a function of stimulation intensity and distance along a straight line extending from the preferred UOA site in the row direction. Data averaged across all contacts. **(E-H)** Same as in (A-D) but for P3. (E-F) P2 preferred stimulation in the proximal UOA columns 1-3, at sites closer to the middle of the UOA (rows 4-7).

Figure 4. Onset Latencies Reveal Local Networks Activated by Focal Optogenetic Modulation. (A) <u>Left:</u> Schematics of UOA stimulation through a single μLED site (C1-R8) and of LEA in P2. <u>Right:</u> Pulse-aligned raster plots for all 21 channels on the LEA through the depth of V1. *Black lines* separate data from different channels. *Gray shaded region:* channels in L4C. *Blue line above plot*: 100ms pulse period at the input voltage (irradiance) indicated. *Red and black arrows denote example contacts in L4C and 6, respectively*. A graded shift in MU onset latency is apparent. **(B)** <u>Left:</u> Pulse-aligned PSTHs for the two channels indicated by arrows in the raster plot in (A). Responses are plotted as baseline-subtracted firing rate versus time. Response onset latency at the L6 contact (35 msec) clearly lagged that on the L4C contact nearest the UOA needle tips (17 msec). <u>Right:</u> Histograms of pulse-by-pulse onset latencies for the two example contacts. **(C)** Heatmaps of MU response (firing rate) through the depth of V1 during the peri-pulse period, for the UOA stimulation condition indicated by the insets at the top left of each plot. Stimulation intensity (average irradiance) is reported above each plot. The firing rate color scale applies to all panels. *White dots* mark the onset latency (estimated from the mean PSTH- see Online Methods) for each contact that was significantly responsive to UOA stimulation.

Figure 5. Population Onset Latencies as a Function of UOA Stimulation Intensity and Spatial Pattern.

(A) Distance on the LEA of each contact from the contact with the fastest onset latency is plotted

against onset latency; lines are linear fits. Each line is from simultaneous stimulation throughout

the whole µLED array at each indicated intensity. **(B)** <u>Left:</u> Effect of varying photostimulation

intensity for a fixed column (C1). <u>Right:</u> Effect of varying stimulated column (C1 to C4) for a

fixed photostimulation intensity (5V). Either lowering intensity for a given column or increasing

the distance between an activated column and the LEA had similar effects on the latency of

network activation. **(C)** As in (B), but for a single µLED stimulation condition. On the left

panel, photostimulation intensity was varied for a fixed µLED (C1-R8), while on the right panel,

the stimulated µLED was varied along column 1 (from row 3 to 9) at a fixed intensity (5V for

µLEDS in rows 8-10, but 7.8V for those in rows 4-7, as lower intensities did not evoke a

response from many of these latter µLEDs). **(D)** The shortest onset latency across all intensities

(here expressed as percent of max- see legend in **Fig. 2G** for corresponding input voltage) is

plotted for the whole array condition (Left), and selected columns (Middle) or µLEDs (Right).

1082

Figure 6. Local Optogenetic Activation Through the UOA Spreads Through Cortico-

Cortical Networks.

(A-C) Case MK414-RH. The same sagittal section encompassing parts of V1 and V2 is shown

under 3 different fluorescent illuminations, to reveal Nissl stain (A), tdT/Chr2 expression (B; the

red tdT fluorescence was converted to green for purpose of illustration), and c-fos IHC (C).

White solid contour: V1/V2 border; *dashed contours:* layer boundaries (layers are indicated);

white boxes: ROIs (numbered 1-3 in panel C) where c-fos+ cells were counted. *White Arrows in*

(B) point to the visible damage caused by each UOA needle, while the *gray arrow* points to the

likely location of one of the UOA needles which did not cause visible damage in this section.

Asterisks in (B) mark the core of the viral injections, and sites of highest tdT/ChR2 expression.

P: posterior; *V:* ventral. C-fos expression in this case is observed throughout all layers (local)

and across cortical areas (long-range). Scale bar in (A): 1mm (valid for A-C). **(D)** Higher

magnification of c-fos IHC in and around each ROI. Scale bar: 0.2mm. **(E-H)** Case MK422-RH.

Same as in (A-D) but for a different case in which an AMPA receptor antagonist was injected

into the SMA prior to UOA insertion and photostimulation. The sagittal section is from the

SMA. *D:* dorsal; *A:* anterior. Scale bars: 1mm (E and valid for E-G); 0.2 mmm (H). Blocking

AMPA receptors demonstrates that initial optogenetic activation is limited to the stimulated

layers in the region of UOA insertion. **(I-J)** Case MK414-LH. C-fos IHC in a sagittal section of

SMA cortex (I) and at higher magnification in and around each ROI used for cell counts (J), in a

case which only received UOA insertion. Scale bars: 1mm (I), 0.2mm (J). **(K-L)** Case MK421-

RH. Same as in (I-J), but for a control case in which SMA cortex only received surface

photostimulation via an optical fiber-coupled laser. Here only one ROI is shown at higher

magnification to reveal the few labeled cells in L1. Scale bars: 0.5mm (K), 0.2mm (L). Increases

in cFos expression cannot be explained by device insertion or surface illumination. **(M)** Average

number of c-fos+ cells across sections used for quantification, as a function of distance from the

center of UOA insertion for the 4 different cases. Error bars: s.e.m **(N)** Distribution of c-fos+

cells across layers at each distance.

Figure 1

Figure 2

Figure 3

Figure 4

Figure 5

UOA + ChR2 + light (MK414-RH)

UOA + ChR2 + Glu block + light (MK422-RH)

UOA insertion only (MK414-LH)

Light only (MK421-RH)

Figure 6