Multi-channel deep learning with intracranial neurostimulation can localize seizure onset zones in humans

*Graham W. Johnson,[1,7,8] Leon Y. Cai,[1,8] Derek J. Doss,[1,7,8] Jasmine W. Jiang,[7,8,9] Aarushi S. Negi,[9] Saramati Narasimhan,[1,7,8,9] Danika L. Paulo,[9] Hernán F. J. González,[1,7,8] Shawniqua Williams Roberson,[1,10] Sarah K. Bick,[9] Catie E. Chang,[2] Victoria L. Morgan,[1,7,8,9,10] Mark T. Wallace,[3,4,5,6] Dario J. Englot[1,2,7,8,9,10,11]

Departments of [1]Biomedical Engineering, [2]Electrical and Computer Engineering, [3]Hearing & Speech Sciences, [4]Psychology, [5]Psychiatry and Behavioral Sciences, and [6]Pharmacology at Vanderbilt University; [7]Vanderbilt University Institute of Imaging Science at Vanderbilt University Medical Center; [8]Vanderbilt Institute for Surgery and Engineering, Vanderbilt University; Departments of [9]Neurological Surgery, [10]Neurology, [11]Radiology and Radiological Sciences at Vanderbilt University Medical Center

Corresponding author: Graham Johnson

graham.w.johnson@vanderbilt.edu

Village at Vanderbilt
1500 21st Avenue South, Suite 4333
Nashville, Tennessee 37212

Keywords: temporal lobe epilepsy, seizure onset zone, single-pulse electrical stimulation, cortico-cortical evoked potentials, deep learning

Number of text pages:

Number of words: 1800

Number of figures: 1

Number of tables: 1

Abstract

In drug resistant temporal lobe epilepsy, automated tools for seizure onset zone (SOZ) localization using brief interictal recordings would supplement presurgical evaluations and improve care. Thus, we sought to localize SOZs by training a multi-channel convolutional neural network on stereo-EEG (SEEG) cortico-cortical evoked potentials. We performed single pulse electrical stimulation with 10 drug resistant temporal lobe epilepsy patients implanted with SEEG. Using the 500,000 unique post-stimulation SEEG epochs, we trained a multi-channel one-dimensional convolutional neural network to determine whether an SOZ was stimulated. SOZs were classified with a mean leave-one-patient-out testing sensitivity of 78.1% and specificity of 74.6%. To achieve maximum accuracy, the model requires a 0-350 ms post stimulation time period. Post-hoc analysis revealed that the model accurately classified unilateral vs bilateral mesial temporal lobe seizure onset, and neocortical SOZs. This is the first demonstration, to our knowledge, that a deep learning framework can be used to accurately classify SOZs using cortico-cortical evoked potentials. Our findings suggest accurate classification of SOZs relies on a complex temporal evolution of evoked potentials within 350 ms of stimulation. Validation in a larger dataset could provide a practical clinical tool for the presurgical evaluation of drug resistant epilepsy.

Introduction

Epilepsy affects over 50 million people worldwide, with temporal lobe epilepsy (TLE) being the most common focal epilepsy.[1] Approximately 30-40% of TLE patients continue to have debilitating seizures despite maximal therapy with anti-seizure medications.[2] Drug resistant patients may undergo presurgical evaluation ahead of resection, ablation, or neurostimulation therapies. A major goal of presurgical workup is to find the areas of the brain responsible for seizure generation, i.e. the seizure onset zones (SOZs). However, precise localization of SOZs can be challenging with non-invasive modalities such as scalp electroencephalography (EEG), MRI, and PET. Therefore, invasive intracranial monitoring with stereo-EEG (SEEG) is often pursued to provide direct electrographic recordings of seizures to localize SOZs. Monitoring after SEEG implantation often requires long hospital stays of days to weeks to record multiple ictal events.[3] Thus, it has been proposed that inter-ictal single-pulse electrical stimulation (SPES) of the SEEG contacts to elicit cortico-cortical evoked potentials (CCEP) can help localize SOZs more efficiently.[4-6]

A challenge of interpreting CCEPs using SEEG is that the foundational work in this field was done using subdural electrode grids that measured consistent electrographic phenomena after stimulation, e.g. N1 (10-50 ms) and N2 (50-300 ms) responses.[4] However, N1 and N2 wave polarity and morphology are defined based on the consistent electrode orientation of subdural electrode grids relative to the cortical surface, and thus orthogonal to cortical pyramidal neurons. In contrast, SEEG has less consistent orientation relative to cortical structures, and translation of N1 and N2 terminology for subcortical gray matter is even more challenging due to heterogenous cytoarchitecture.[7] Thus, it is difficult to predict the pattern of CCEP wave morphology for any given SEEG contact. Accordingly, most groups rely on coarse metrics for CCEPs in SEEG such as root-mean-squared power.[6,8] However, these metrics may miss important electrographic features that could help characterize the epileptogenic network. We propose that a multi-channel one-dimensional convolutional neural network (CNN) is well-suited for recognizing variable evoked wave morphology from multiple SEEG contacts simultaneously. This could be a useful tool to delineate whether a given set of CCEPs resulted from an SOZ or non-SOZ being stimulated. Further, by probing various time windows post-stimulation, we can systematically determine which post-stimulation time periods contain the most important classifying features.

Methods

Participants and single-pulse electrical stimulation

We collected over 500,000 post-stimulation 900 ms SEEG epochs from 10 patients with drug resistant TLE who underwent presurgical evaluation (**Table 1**). Clinical data were collected through chart review, and seizure outcomes were assigned using the Engel scale.[9] This study received Institutional Review Board approval and informed subject consents were obtained. We conducted single pulse electrical stimulation (SPES) with every adjacent bipolar pair of contacts in gray matter for each patient. We used a 10 second, 1 Hz, 300 microsecond, biphasic pulse at 3.0 milliamps with a recording sampling rate of 512 Hz.

Preprocessing and SOZ labeling

We filtered raw SEEG data using Matlab's filtfilt function (MathWorks inc., Natick, MA, USA) with Butterworth filter passbands of 1-59, 61-119 and 121-150 Hz. We then parsed the data into 900 ms epochs following each 1 Hz stimulation. This resulted in over 500,000 preprocessed epochs for training our model. SOZs were defined as regions containing any contacts involved in ictal onset of one or more seizures after epileptologist interpretation of all ictal data. Using custom SEEG planning software, CRAnial Vault Explorer (CRAVE), we automatically localized every contact for each patient and created a table of all inter-contact Euclidean distances.[10]

Deep learning

Using post-stimulation EEG epochs, we trained a one-dimensional multi-channel multi-scale CNN (**Fig. 1A**). To accomplish this, we modified the Multi-Scale-1D-ResNet developed by Fei Wang (https://github.com/geekfeiw/Multi-Scale-1D-ResNet) to input 40 SEEG channels simultaneously. To avoid stimulation artifact and implantation bias, the epochs were distance thresholded to exclude any SEEG channels within 20 mm of the stimulation pair.[11] For each training pass, we randomized the subset of 40 channels chosen from a patient's entire available channels. We utilized a weighted binary cross entropy loss function and stopped training after five model epochs. We implemented a leave-one-patient-out testing strategy across all patients. We first tested the ability of the trained model to classify SOZs using the entire 0-900 ms window. Next, we tested the model with only a non-overlapping 50 ms sliding window over the post-

stimulation period. We also trained the model on three separate randomized region labels to serve as a

control.

Post-hoc testing:

We conducted post-hoc testing to determine 1) which post-stimulation time period is best for SOZ

classification, 2) can the model classify unilateral vs. bilateral mesial temporal onset, and 3) can the model

accurately classify neocortical temporal SOZs? We accomplished (1) by nulling the data outside the desired

time window before training. For (2), we calculated the accuracy of left and right mesial temporal SOZ

classification for patients with bilateral mesial temporal seizures (n=4) vs. patients with a) unilateral mesial

temporal seizures on ictal SEEG, b) a bilateral SEEG implant, and c) seizure-free surgical outcomes (n=3). To

accomplish (3), we calculated the accuracy of neocortical temporal SOZ classification in all patients.

Statistical methods

We calculated the sensitivity and specificity for the leave-one-patient-out testing across all 10 patients for the

various time window analyses. We also report the Youden index (sensitivity + specificity – 100) to summarize

the usefulness of the model at a given time window; Youden index values above 50 are generally considered

to be a very useful model for classification, and values close to 0 are considered useless even if sensitivity or

specificity is individually high.[12] We compared Youden indexes with paired t-tests using Bonferroni-Holm

multiple comparison correction.

Data availability

Data and computer code are available upon reasonable request.

Results

CNN trained on long-range CCEPs accurately classifies SOZs

As outlined in **Fig. 1B,** the CNN trained on the entire 900 ms post-stimulation period correctly classified the stimulation pair as SOZ with a mean leave-one-patient-out testing sensitivity of 78.1% (95% confidence interval [CI] 67.8 to 88.4%) and specificity of 74.6% (95%CI 68.7 to 80.5%), resulting in an average Youden index of 52.7 (95%CI 43.7 to 61.8). In comparison, when the model was trained using regions randomly labeled as SOZ or non-SOZ, the average Youden index was significantly decreased to 16.5 (95%CI 9.62 to 23.4, t-test p=4.88e-6). Furthermore, the model achieved significantly improved Youden indexes when training on 50 ms sliding windows ranging from 0-350 ms compared to the same windows with random labels (**Fig. 1C**). Interestingly, the specificity and sensitivity of the model peaked during different periods in the initial 350 ms post-stimulation – the sensitivity peaked for the time window that spans 100 to 150 ms, while the specificity peaked for the 0-50 ms time window. This suggests that delayed responses are most *sensitive* for classifying SOZs, whereas early responses are most *specific* for classifying SOZs.

Important features are temporally distributed within the initial post-stimulation window

We performed post-hoc analyses to assess which early post-stimulation window was most effective at classifying if an SOZ was stimulated (**Fig. 1D**). Using a time window of 0-350 ms we observed a leave-one-patient-out average testing sensitivity of 74.0% (95%CI 63.3 to 84.7%) and specificity of 78.5% (95%CI 75.9 to 81.1%) with an average Youden index of 52.5 (95%CI 42.1 to 62.9) – very similar to when the model was trained on the entire 0-900 ms. When we divided the 0-350 ms period into 0-175 ms and 175-350 ms, the leave-one-patient-out testing Youden index dropped significantly, suggesting both early and late portions of this time window contribute to model performance.

The model can classify unilateral vs. bilateral onset mesial temporal lobe epilepsy, and can detect neocortical temporal SOZs

We observed that the bilateral onset patients had left mesial temporal structures correctly classified as SOZs for 68.9% (95% CI 58.7 to 79.1%) of the CCEP epochs, and right mesial temporal structure epochs classified as SOZs for 67.9% (95%CI 45.4 to 90.4%) (**Fig. 1E**). For unilateral patients, the model correctly classified

mesial temporal structures ipsilateral to the seizure onset hemisphere as SOZs for 91.5% (95%CI 89.7 to

93.3%) of the epochs with a low false positive rate of 35.1% (95%CI 16.7 to 53.5%) for non-SOZs on the

contralateral side. This sub analysis provides evidence that the model was not simply classifying all mesial

temporal structures as SOZs, but rather provides accurate classification for unilateral vs. bilateral mesial

temporal onset patients. Furthermore, the model correctly classified neocortical temporal SOZs at a rate of

64.4% (95%CI 44.3 to 84.5%), and misclassified neocortical temporal non-SOZs at only 26.0% (95%CI 19.7 to

32.3%) (**Fig. 1F**).

Discussion

We have demonstrated that a CNN trained entirely on SEEG-derived CCEPs farther than 20 mm from the site

of stimulation can classify an SOZ with high sensitivity and specificity in TLE. A strength of this approach is that

the model accurately classified SOZs despite the variable morphology of CCEPs during stimulation of SEEG

electrodes. Further, the most important post-stimulation features for classification are contained within 0-350

ms. This is not surprising considering that most previous findings using RMS have centered around N1 and N2

responses within 300 ms.[4,13,14] However, separating this window into smaller segments significantly reduces

model accuracy. This suggests that there is a complex pattern of CCEPs occurring at various periods post-

stimulation that must be considered in an ensemble to accurately classify the stimulation of ictogenic tissue –

this observation could be due to varied phenotypes of evoked responses.[15] Additionally, through our sub

analyses, we conclude that this model was not classifying all mesial temporal structures as SOZs and can

accurately distinguish unilateral vs bilateral mesial temporal onset. Finally, the model can also accurately

classify neocortical temporal SOZs.

Limitations and future work

Although 500,000 non-overlapping SEEG epochs were used to train the CNN, training and testing datasets

were divided at the patient level. Thus, our relatively small sample size of 10 patients limits our assessment of

generalizability and motivated our conservative strategy of leave-one-patient-out testing across the entire

cohort. Also, mean follow-up was 15.4 months, and future seizure recurrences may decrease the confidence in

clinical SOZ localization and change labels for the CNN. We also did not include any focal extratemporal-lobe

epilepsy patients and thus cannot comment on the extension of these techniques to that population. Our future

work is aimed at addressing these limitations by collaborating with other institutions that collect these rare

datasets. We also hope to test this model on patients with surgical outcomes of Engel 2-4. Perhaps, previously

unidentified SOZs, including bilateral seizure onset, could be elucidated in Engel 2-4 patients with a model

trained on Engel 1 patients.

Conclusions

This work serves as the first demonstration, to our knowledge, that a one-dimensional multi-channel multi-

scale CNN can learn highly non-linear features of SEEG-derived CCEPs occurring across multiple SEEG

channels simultaneously to classify when an SOZ is stimulated. Furthermore, we demonstrated the importance

of utilizing the entire 0-350 ms time window for classification. We hope that future work will consider using

deep learning as a tool to explore the complex CCEPs generated with SEEG.

Acknowledgements

215 We would like to thank the patients participating in our investigation.

Author contributions

G.W.J. contributed to project conception, data collection, data preprocessing, data analysis, and manuscript
preparation. L.Y.C. contributed to data analysis and manuscript preparation. D.J.D. contributed to data
analysis, and manuscript preparation. J.W.J. contributed to data collection, data preprocessing, data analysis
and manuscript preparation. A.S.N. contributed to data preprocessing, data analysis, and manuscript
preparation. S.N., D.P., H.F.J.G., S.W.R., S.K.B., C.E.C., V.L.M., M.T.W. contributed to data analysis and
manuscript preparation. D.J.E. contributed to project conception, data analysis, and manuscript preparation.

Competing interests

226 The authors disclose no conflicts of interest

Funding

229 This work was supported by the following funding sources: NINDS R01-NS112252-02, NINDS R01-NS075270,
230 NINDS R01-NS110130, NINDS R01-NS108445, NINDS F31-NS106735, NINDS F31-NS120401-01, and NIH
231 Training Grants: NIGMS T32-GM007347, NIBIB T32-EB021937, and NIBIB T32EB001628.

Ethics Statement

234 We confirm that we have read the journal's position on issues involved in ethical publication and affirm that this
235 report is consistent with those guidelines.

References

1. Behr, C., et al., *Epidemiology of epilepsy.* Revue Neurologique, 2016. **172**(1): p. 27-36.

2. Engel, J., Jr., *What can we do for people with drug-resistant epilepsy? The 2016 Wartenberg Lecture.* Neurology, 2016. **87**(23): p. 2483-2489.

3. Englot, D.J., *A modern epilepsy surgery treatment algorithm: Incorporating traditional and emerging technologies.* Epilepsy & Behavior, 2018. **80**: p. 68-74.

4. Matsumoto, R., T. Kunieda, and D. Nair, *Single pulse electrical stimulation to probe functional and pathological connectivity in epilepsy.* Seizure, 2017. **44**: p. 27-36.

5. Van 'T Klooster, M.A., et al., *Time–frequency analysis of single pulse electrical stimulation to assist delineation of epileptogenic cortex.* Brain, 2011. **134**(10): p. 2855-2866.

6. Guo, Z.-h., et al., *Epileptogenic network of focal epilepsies mapped with cortico-cortical evoked potentials.* Clinical Neurophysiology, 2020(131): p. 2657-2666.

7. Buzsáki, G., C.A. Anastassiou, and C. Koch, *The origin of extracellular fields and currents — EEG, ECoG, LFP and spikes.* Nature Reviews Neuroscience, 2012. **13**(6): p. 407-420.

8. Prime, D., et al., *Considerations in performing and analyzing the responses of cortico-cortical evoked potentials in stereo-EEG.* Epilepsia, 2018. **59**(1): p. 16-26.

9. Engel, J., *Update on surgical treatment of the epilepsies: Summary of The Second International Palm Desert Conference on the Surgical Treatment of the Epilepsies (1992).* 1993. **43**(8): p. 1612-1612.

10. D'Haese, P.-F., et al., *CranialVault and its CRAVE tools: A clinical computer assistance system for deep brain stimulation (DBS) therapy.* 2012. **16**(3): p. 744-753.

11. Prime, D., et al., *Quantifying volume conducted potential using stimulation artefact in cortico-cortical evoked potentials.* J Neurosci Methods, 2020. **337**: p. 108639.

12. Youden, W.J., *Index for rating diagnostic tests.* Cancer, 1950.

13. Mouthaan, B.E., et al., *Single Pulse Electrical Stimulation to identify epileptogenic cortex: Clinical information obtained from early evoked responses.* 2016. **127**(2): p. 1088-1098.

14. Zhang, N., et al., *The effectiveness of cortico-cortical evoked potential in detecting seizure onset zones.* Neurological research, 2018. **40**(6): p. 480-490.

15. Crocker, B., et al., *Local and distant responses to single pulse electrical stimulation reflect different forms of connectivity.* Neuroimage, 2021. **237**: p. 118094.

Figure Legends

Figure 1: **Deep learning on distant SEEG CCEPs can accurately classify SOZs. A)** We conducted single-pulse electrical stimulation on all gray matter bipolar pairs of contacts for 10 patients undergoing SEEG – this resulted in over 500,000 post-stimulation epochs from the recording channels. To avoid stimulation artifact and biases relating to contact implantation density due to clinical hypotheses, we excluded recordings from contacts that were within 20 mm of the stimulation site. We then trained a convolutional neural network (CNN) to classify if a clinically defined seizure onset zone (SOZ) or non-seizure onset zone (non-SOZ) was stimulated. **B)** We first trained the model using the entire 0-900 ms post-stimulation window. This resulted in a sensitivity of 78.1% and specificity of 74.6% with a significantly improved Youden index compared to training the model on random labels. **C)** For 50 ms sliding windows, the model performed better than random labels for the 0-350 ms post-stimulation 50 ms intervals. Paired t-tests with Bonferroni-Holm multiple comparison correction were conducted between the Youden index (blue), and the random-label Youden index (gray). Note: values on the x-axis represent <u>ending</u> time of the 50 ms window. **D)** Using only the 0-350 ms window resulted in a model accuracy comparable to using the 0-900 ms window. However, dividing this window into 0-175 ms or 175-350 ms resulted in a significant reduction in Youden indexes. **E)** The model was not simply classifying all mesial temporal structures as SOZs. For bilateral patients, the model classified left and right mesial temporal structures as SOZs with comparable confidence around 70%. For unilateral patients, the model correctly classified ipsilateral mesial temporal structures as SOZs at a rate of 91.5% and contralateral (i.e. non-SOZs) at a low rate of 35.1% - suggesting that the model can accurately classify unilateral vs. bilateral seizure onset. **F)** The model was also able to correctly classify neocortical temporal SOZs 64.4% of the time with a low false positive rate of 26.0%. White dot in violin plots is median, horizontal bar is mean.

A Data aquisition and neural network training

20 mm Exclusion

Red: Stimulation electrodes (excluded)
Orange: Near-field electrodes (excluded)
Blue: Distant electrodes (used in model)

Time post-stimulation, ms

Neural network

1D multi-channel multi-scale ResNet

40 EEG input channels

1D convolutional kernels of size 3, 5, and 7 samples

Windowed data →

Classify stimulation pair as:

SOZ or non-SOZ

B 0 - 900 ms window
Average leave-one-out testing

Spec Sens

p = 4.88e-6

Youden Youden random

Percent

C 50 ms sliding window
Average leave-one-out testing

Model useful

Specificity
Sensitivity
Youden
Youden: random labels

50 ms window end time (post-stim), ms

* p < 0.05
** p < 0.01

D Early time window sub analysis
Average leave-one-out testing

Specificity
Sensitivity
Youden

**
***
**

0 - 350 ms 0 - 175 ms 175 - 350 ms

Post-stimulation time window, ms

E Mesial temporal lateralization
Average leave-one-out testing 0-900 ms

Percent of epochs classified as SOZ

**

L mesial temporal (SOZ)
R mesial temporal (SOZ)
Ipsilateral mesial temporal (SOZ)
Contralateral mesial temporal (non-SOZ)

Bilateral patients Unilateral patients

F Neocortical temporal localization
Average leave-one-out testing 0-900 ms

Percent of epochs classified as SOZ

**

Neocortical temporal SOZ
Neocortical temporal non-SOZ

* p < 0.05
** p < 0.01
*** p < 0.001

Table 1: Patient demographics

Subject ID	Age	Sex	Racial identity	Epilepsy duration, years	Preoperative seizure frequency, per month	FBTC	MRI Findings	Number of contact pairs stimulated	Surgery	Surgical outcome	Postsurgical duration, months
Pat1	51	F	White	20	3.5	No	Normal	69	B/L RNS	>50 % reduction	25
Pat2	39	F	White	15	60.5	Yes	Normal	69	B/L RNS	<50 % reduction	21
Pat3	30	F	Two or more	14	9.5	Yes	Normal	42	R SAH	Engel 1d	21
Pat4	58	F	Black	8	1.0	Yes	Normal	45	R SAH	Engel 1a	13
Pat5	24	F	Asian	7	10.0	No	Normal	72	B/L RNS	<50 % reduction	12
Pat6	41	M	White	18	1.0	No	L MTS	71	L SAH	Engel 1a*	10
Pat7	28	M	Black	14	1.0	Yes	L temporal encephalocele	63	L ATL	Engel 1a*	10
Pat8	23	F	White	14	32.0	Yes	Normal	48	B/L RNS	>50 % reduction	13
Pat9	23	F	White	7	12.0	No	Normal	53	R ATL	Engel 1a	15
Pat10	23	M	White	18	1.0	Yes	Normal	65	R ATL	Engel 1a	14

R right; L left; B/L bilateral; FBTC focal to bilateral tonic-clonic seizures; RNS responsive neurostimulation; SAH selective amygdalohippocampectomy; ATL anterior temporal lobectomy; MTS mesial temporal sclerosis; <>50% reduction refers to reduction in monthly seizure frequency compared to preoperative baseline. *Outcome consistent with Engel 1a if persists >1yr.