1

MSclassifR: an R Package for Supervised Classification of Mass Spectra with Machine Learning Methods

Alexandre Godmer[1,2,*], Yahia Benzerara[1], Emmanuelle Varon[3], Nicolas Veziris[1,2], Karen Druart[4], Renaud Mozet[5], Mariette Matondo[4], Alexandra Aubry[2,6], Quentin Giai Gianetto[4,7,*]

[1] *AP-HP.Sorbonne université, Hôpital Saint-Antoine, Département de Bactériologie, Paris, France*

[2] *Sorbonne Université, INSERM, U1135, Centre d'Immunologie et des Maladies Infectieuses, Cimi-Paris, Paris, France*

[3]*Laboratory of Medical Biology and National Reference Centre for Pneumococci, Intercommunal Hospital of Créteil, 94000 Créteil, France*

[4] *Institut Pasteur, Université Paris Cité, Proteomics Platform, Mass Spectrometry for Biology Unit, UAR CNRS 2024*

[5] *UFR de Mathématiques et Informatique, Faculté des Sciences - Université Paris Cité*

[6] *AP-HP, Sorbonne Université, Hôpital Pitié-Salpêtrière, Centre National de Référence des Mycobactéries et de la Résistance des Mycobactéries aux Antituberculeux, Paris, France*

[7] *Institut Pasteur, Université Paris Cité, Bioinformatics and Biostatistics HUB*

[*] Corresponding author(s).

E-mail: alexandre.godmer@aphp.fr (Godmer A), quentin.giaigianetto@pasteur.fr (Giai Gianetto Q).

Abstract

MSclassifR is an R package that has been specifically designed to improve the
classification of mass spectra obtained from MALDI-TOF mass spectrometry. It offers
a comprehensive range of functions that are focused on processing mass spectra,
identifying discriminant m/z values, and making accurate predictions. The package
introduces innovative algorithms for selecting discriminating m/z values and making
predictions. To assess the effectiveness of these methods, extensive tests were
conducted using challenging real datasets, including bacterial subspecies of the
Mycobacterium abscessus complex, virulent and avirulent phenotypes of *Escherichia*
coli, different species of Streptococci and nasal swabs from individuals infected and
uninfected with SARS-CoV-2. Additionally, multiple datasets of varying sizes were
created from these real datasets to evaluate the robustness of the algorithms. The results
demonstrated that the Machine Learning-based pipelines in MSclassifR achieved high
levels of accuracy and Kappa values. On an in-house dataset, some pipelines even
achieved more than 95% mean accuracy, whereas commercial system only achieved
62% mean accuracy. Certain methods showed greater resilience to changes in dataset
sizes when constructing Machine Learning-based pipelines. These simulations also
helped determine the minimum sizes of training sets required to obtain reliable results.
The package is freely available online, and its open-source nature encourages
collaborative development, customization, and fosters innovation within the
community focused on improving diagnosis based on MALDI-TOF spectra.

Keywords: [MALDI-TOF MS]; [Machine Learning]; [R package]; [microbiology]

Key points:

- MSclassifR is a comprehensive R package enabling the construction of data analysis pipelines for the precise classification of mass spectra.

- Our R package contains an innovative method for variable selection from random forests, which delivered excellent results on real data.

- In-depth analysis of various machine learning-based pipelines using our package allowed us to make conclusions about the optimal m/z selection and prediction methods depending on the size of the training dataset.

- Using a publicly available dataset of mass spectra obtained from various MALDI-TOF instruments across different countries, MSclassifR is able to build robust pipelines capable of adapting to different instruments in an automatic way.

- When tested on an in-house dataset, MSclassifR pipelines consistently outperformed a commercial software in terms of prediction accuracy.

Introduction

In clinical microbiology, the accurate identification of pathogens is crucial for proper patient management. Matrix Assisted Laser Desorption Ionization - Time of Flight (MALDI-TOF) mass spectrometry has become widely adopted for pathogen identification in biological cultures due to its speed, affordability, and user-friendly nature [1]. This technique involves comparing the mass spectrum obtained from ionizing intact proteins of a viral or bacterial culture with reference mass spectrum profiles. Commercial software provided by spectrometer manufacturers uses algorithms to match the measured mass spectrum with the spectra of known microorganisms. While this technique is effective in distinguishing many species, it often falls short in differentiating spectrally close organisms, such as subspecies within the *Mycobacterium abscessus* complex [2] [3] or species from the *Enterobacter cloacae* complex [4] [5].

To overcome these limitations, several studies have explored advanced machine learning (ML) techniques to fully exploit the information contained in mass spectra and accurately classify them into distinct categories, such as detecting resistant or virulent microorganisms in closely related mass spectra [6] [7] [8]. However, many of these studies focus on a single ML method [9] [10] [11] [12] and others lack the provision of corresponding computer code, making reproducibility of results challenging [8].

While open-source R packages like MALDIquant and MALDIrppa already offer well-documented methods for pre-processing MALDI-TOF mass spectra [13] [14], additional steps are necessary to classify the pre-processed mass spectra, including the selection of discriminative peaks and an optimal mass spectra classification model. These steps require advanced knowledge and specialized R packages for variable selection and ML methods, such as the caret and mixOmics R packages, which further demand expertise in parameter optimization [15] [16]. On the other side, non-coders have access to all-in-one software solutions, but they are often expensive, do not

provide the computer code used, and provide limited user control over parameters (e.g.

[17] [18]).

In response to these challenges, we have developed MSclassifR, a comprehensive

package that enables the construction of data analysis pipelines for the precise

classification of mass spectra. Within this article, we introduce the functionalities

offered by MSclassifR and assess their effectiveness. We conducted tests using

challenging real datasets and created additional datasets with varying sizes from them

to evaluate the performance of data analysis pipelines built with MSclassifR.

Notably, our findings highlight the pivotal role of mass-over-charge (m/z) value

selection in achieving accurate classifications. We have devised an innovative statistical

method for variable selection, taking advantage of variable importances derived from

Random Forests (RF) and employing a mixture model that distinguishes between

discriminant and irrelevant features. This unique approach has led to superior

performance in terms of accuracies and Cohen's Kappa coefficient values compared to

alternative feature selection methods used before a prediction model.

Methods

112 The MSclassifR package offers R functions that concerns three main topics [*Figure*

113 *1A*]: (i) pre-processing of mass spectra, (ii) selection of discriminative m/z values, (iii)

114 estimation of a classification model to predict the category of a mass spectrum from

115 shortlisted m/z values.

Pre-processing of mass spectra methods

Many of the algorithms already available in the MALDIquant and MALDIrppa

packages are implemented in MSclassifR [13] [14]. For example, the smoothing step

can be performed by using the "Savitzky Golay" algorithm or the "Undecimated

Wavelet Transform" (UDWT) algorithm [19] [20] ; baseline subtraction can be

performed using the "statistics-sensitive non-linear iterative peak-clipping" (SNIP)

algorithm [21], the TopHat algorithm [22], or the Convexfull algorithm [23]; intensities

can be normalized using the "Total Ion Current" (TIC) algorithm or the "Probabilistic

Quotient Normalization" (PQN) algorithm [24]; mass spectra can be aligned using the

"Locally Weighted Scatterplot Smoothing" (LOWESS) method [25] or by employing

linear, quadratic, or cubic functions [26]. Finally, the peak selection step can be

repeated using the "Median Absolute Deviation" (MAD) method [27] or the "super

smoother" method [28]. However, we have defined a pipeline by default in MSclassifR

including several algorithms that are widely used for processing MALDI-TOF MS

spectra [29] [*Figure 1A*] . It consists of the following steps: (i) square root intensity

transformation, (ii) spectrum smoothing with UDWT algorithm, (iii) baseline

processing with SNIP algorithm, (iv) intensity calibration using TIC algorithm, (v)

spectrum alignment with the LOWESS method and selection of peaks with signal-to-

noise ratio (S/N).

Selecting discriminant *m/z* values

The selection of discriminant *m/z* values is analogous to the variable selection problem

in statistics. It is crucial to choose the right variables to avoid incorporating irrelevant

ones into prediction models, which can lead to inaccurate predictions. Additionally,

variable selection helps reduce the dataset size, which is advantageous from a

computational time perspective. MSclassifR provides several methods for variable

selection, which can be categorized into two groups [30]:

(a) the first aims to identify a minimum number of variables that yield the highest

possible performance criterion (e.g., Accuracy or Cohen's Kappa), often resulting in

the elimination of correlated variables;

(b) the second aims to identify all variables with potential predictive power (e.g. without

deletion of correlated variables).

On one hand, the first group is commonly employed in ML literature [30] where the

primary objective is to achieve the highest accuracy with the smallest set of variables.

In omics sciences, this approach is used to identify a small set of potential biomarkers

(e.g. genes, proteins, metabolites) associated with a specific biological condition. A

popular method of this type of approach is "Recursive Feature Elimination" (RFE)

which consists of recursively selecting smaller and smaller sets of variables. This type

of method typically involves setting a specific number (or range for that number) of

variables to select.

On the other hand, the second group enables the filtration of variables that lack

predictive power (e.g. background noise) and retains only those with predictive

capabilities. For example, this approach is used to extract the set of genes or proteins

that exhibit differential expression between different biological conditions. Thus, this

type of method adapts to the dataset to select all the predictive variables without

specifying the number of variables to select.

In addition to existing methods, we have implemented an original approach to identify

all variables with predictive power using RF. This method involves estimating the

distribution of variable importances for non-discriminant (useless) variables from

negative importances or simulated non-discriminant variables [*Figure 1B*].

Subsequently, a mixture model is assumed, where the cumulative distribution of

variable importances for all variables (F) is a combination of the cumulative

distributions for useless variables (F_u) and discriminant variables (F_d), with π

representing the proportion of useless variables in the dataset:

169
$$F(x) = \pi \times F_u(x) + (1 - \pi) \times F_d(x)$$

By estimating the distribution of useless variables ($F_u(x)$), we can determine its

quantile values (x_q) and compute $\epsilon_q = F(x_q)/q$ for a given quantile (q) of useless

variables [*Figure 1C*]. For high quantile values of q, it is expected that $F_u(x_q)$ tends

to 1. Thus, $\epsilon_q = \pi + (1 - \pi) \times F_d(x_q)$ for such quantiles when $F_d(x_q)$ is not null. It

is assumed that F_d becomes zero when the variable importance is small. If this

assumption also holds, ϵ_q tends to π. The minimum value of the ϵ_q corresponds to an

estimated proportion of useless variables in the dataset, i.e. $\tilde{\pi} = \min_{q} \epsilon_q$ [*Figure 1C*],

enabling us to estimate a number of discriminant variables (N_d) as floor ($N \times (1 - \tilde{\pi})$),

where N represents the total number of variables and "floor()" is the integer value

function. Finally, the N_d variables with the highest variable importances are selected

as the set of discriminant variables. Here, the π estimation is analogous to finding the

proportion of true null hypotheses in the context of multiple hypothesis testing with the

difference that we use the variable importances instead of p-values. A major difference

lies in the absence of a reliable theoretical framework for estimating the distribution of

variable importances for useless variables, whereas an assumption of independence

between the tests and therefore a uniform distribution of p-values under the null

hypothesis can be assumed in the multiple testing context [31]. It is therefore necessary

to estimate the distribution of useless variables either by generating such variables or

using the negative importance variables and assuming that $F_u(x)$ is a more or less

symmetrical distribution with respect to zero. Both approaches are possible through our

"SelectionVar()" function of our package ("*nbf*" argument). Because variable

importances can be estimated with different approaches using RF, two ways of

estimating them are implemented in "SelectionVar()": "mda" for the "mean decrease in

accuracy" variables importances, and "cvp" for "cross-validated permutation" variables

importances [32].

A

Preprocessing

Selecting discriminant mass-over-charge values between different categories of mass spectra

Estimating classification models

Predicting the category of a mass spectrum

B   **C**

Figure 1. MSclassifR workflow and plots presenting the method developed for thresholding the

variable importances of Random Forests

A. Overview of the proposed pipeline in MSclassifR, encompassing mass spectrum preprocessing with

various methods, selection of discriminative *m/z*, and the estimation of a classification model to predict

mass spectrum categories based on the shortlisted *m/z* values. **B.** Empirical cumulative distribution

functions of the variable importances (V.I.) for the useless *m/z* (red) and all the *m/z* (black) estimated by

the "SelectionVar()" function with the "mda" method from a real dataset [33] **C.** Estimation of $\tilde{\pi}$ using

the same V.I. as in **B.** and the quantities $\epsilon_q = F(x_q)/q$ where q are quantiles of useless variables.

206 **Predicting the category of a mass spectrum**

207 Once we selected *m/z* values, a model can be estimated to predict a category from these

208 shortlisted values. In supervised classification, a commonly used parametric model is

209 multinomial linear logistic regression. It consists in estimating the probability that a

210 mass spectrum belongs to a category from the following model:

211
$$P(MS \in j) = \frac{e^{X\beta_j}}{1 + \sum_{k=1}^{K-1} e^{X\beta_k}}$$

212 where $P(MS \in j)$ is the probability that a mass spectrum belongs to the j^{th} category

213 (among K categories), X are intensities of the mass spectrum measured at shortlisted

214 *m/z* values, and β_j are coefficients related to the j^{th} category. Once all the coefficients

215 are estimated, a mass spectrum is associated to a category by

216
$$MS \in j \text{ if } \underset{k \in \{1,...,K\}}{\text{argmax}} P(MS \in k) = j$$

In addition to multinomial linear logistic regression, popular ML algorithms are

available in MSclassifR like RF, neural networks, support vector machines with linear

kernels and extreme gradient boosting. Each of these algorithms can be used to estimate

a prediction model via the "LogReg()" function. Of note, the "LogReg()" function is

able to deal with imbalanced datasets displaying skewed class proportions. This is a

characteristic present in many real datasets when some categories are rarer than other.

Different approaches are implemented to resample the data before training the

algorithms via the "*Sampling*" argument: up-sampling, down-sampling and SMOTE

(Synthetic Minority Oversampling Technique). Next, the "PredictLogReg()" function

can be used to estimate the probabilities $P(MS \in j)$ from one or several prediction

models. In addition to provide the estimated $P(MS \in j)$ of each model in input,

"PredictLogReg()" includes two methods to aggregate the predictions of several

models: the max-voting method, and the Fisher's combined probability test to merge all

the estimated $P(MS \in j)$ of the models in input.

Another approach using parametric models involves conducting multiple linear

regressions, one for each category, and selecting the category associated with the

regression that has the lowest Akaike Information Criterion:

235
$$MS \in j \text{ if } \underset{k \in \{1,\dots,K\}}{\operatorname{argmin}} \text{AIC}_k(X = \alpha_k + \sum_{i=1}^{I} \beta_{ik} Z_{ik} + \varepsilon_k) = j$$

Here, X represents the intensities of the mass spectrum measured at m/z values, Z_{ik}

represents the intensities of the i^{th} mass spectrum of the k^{th} category used to estimate

the model at the same m/z values, ε_k denotes the Gaussian white noise corresponding

to the kth linear regression with parameters α_k and β_{ik}, and AIC_k represents the

Akaike Information Criterion of the kth linear regression. This approach has been

implemented in an original function called "PredictFastClass()" in MSclassifR. It is

important to note that in order to perform the regression, the number I of mass spectra

used within a category has to be lower than the number of m/z values. Additionally, for

each category, a Fisher test can be conducted to determine if the β_{ik} coefficients of the

linear model are zero, indicating significant correlation between the mass spectrum and

at least one of the mass spectra in a category. By performing this test for all categories

and selecting the minimum p-value, we can assess if the mass spectrum significantly

belongs to at least one of the input categories. If the hypothesis is rejected for all

performed regressions, the mass spectrum can be considered as not belonging to any of

the input categories.

Experimental datasets used to assess the performance of MSclassifR

We used MSclassifR on openly accessible datasets to assess pipelines for accurately

predicting the category of a given mass spectrum. These datasets are of variable sizes:

• Dataset 1: 362 nasal swab mass spectra obtained from 312 patients categorized

as either SARS-CoV-2 infected (211 mass spectra) or SARS-CoV-2 uninfected

(151 mass spectra) from the Nachtigall *et al*. study [33]. These data are available

at https://www.ebi.ac.uk/pride/archive/projects/PXD021388.

• Dataset 2: 882 mass spectra from 294 strains of *Escherichia coli*, classified into

two categories: Shiga-toxin producer (190 strains, 570 spectra) or non-producer

(104 strains, 312 spectra) from the of Christner *et al.* study [6]. These data are

available at https://zenodo.org/record/4996418.

- Dataset 3 (in-house): 1001 mass spectra from 41 clinical strains representing the three subspecies of the *Mycobacterium abscessus* complex: *M.* subsp. *abscessus* (15 strains, 633 spectra), *M.* subsp. *bolletii* (7 strains, 164 spectra), and *M.* subsp. *massiliense* (9 strains, 201 spectra). These data are available at https://zenodo.org/record/5793313.

- Dataset 4 (in-house): 1890 mass spectra from 79 clinical strains from *Streptococcus pneumoniae* (60 strains, 1434 spectra) and from the viridans streptococci group including *S. mitis* (11 strains, 264 spectra) and *S. pseudopneumoniae* (8 strains, 192 spectra). These data are available at https://zenodo.org/record/8419282.

General design of numerical experiments

In our experiments, we employed the previously discussed datasets and used the mass spectra from each dataset where we separated each into a training set of mass spectra, consisting of 70% of randomly chosen mass spectra, and we used the rest of the mass spectra (30%) to test the prediction models and evaluate the accuracy and the Kappa values. This was repeatedly performed several times to assess the variations of accuracy and Kappa values. For the training datasets, pre-processing of mass spectra was performed similarly on all datasets with the "SignalProcessing()" function of our R package by the following steps: (1) transforming intensities with sqrt method; (2) smoothing with UDWT algorithm; (3) remove baseline with SNIP method; (4) calibrate intensity with TIC method; (5) aligning spectra with LOWESS method and with a tolerance parameter of 5e-04 ppm. Next, intensity peaks of mass spectra were detected using the "PeakDectection()" function of our R package with default parameters (tolerance parameter of 2e-03 ppm, signal-to-noise ratio (SNR) of 3, and halfWindowSize parameter of 11). Mass spectra were normalized according to the maximum intensity value for each mass spectrum. For the testing datasets, the same pre-processing of mass spectra was used except that they were aligned on a reference mass spectrum created from the ones of the training dataset using a tolerance parameter

of 5e-04 ppm, a minFrequency parameter of 70% and the "strict" method for the

"*binPeaks_method*" argument of the "SignalProcessing()" function. Indeed, the

alignment to a reference mass spectrum is important to ensure that the mass spectra in

the testing dataset closely match those used during the training of the prediction model.

We conducted extensive comparisons by employing the previously discussed datasets

2 to 4 and generating subsets of randomly chosen mass spectra of various sizes among

them. Each subset has been created by respecting the original proportions of each

category in the datasets. To avoid overfitting, we ensured that when multiple replicates

of a particular sample were present, they were consistently grouped together within the

same subset. Here, our main focus revolved around the training dataset sizes and the

advantages of using contemporary ML methods (RF, neural networks, support vector

machines with linear kernels and extreme gradient boosting) compared to conventional

parametric approaches such as decision rules based on multinomial linear logistic

regression or multiple linear regressions, as presented earlier. Additionally, it is

important to evaluate the number of mass spectra needed to construct an accurate

classifier.

Results and discussion

Extensive comparison from randomly generated subsets of mass spectra

For the datasets (2 to 4) previously described, we randomly drew sets of mass spectra of different sizes. These generated datasets range from 50 mass spectra to 1890 mass spectra. For each of them, a training set consisting of 70% of the mass spectra from the generated dataset and a test set (composed of the remaining 30% of mass spectra) were constituted several times by random draws. For each generated dataset, we evaluated 6 different methods available in MSclassifR to select discriminative peaks, which we combined with 6 different methods to predict the different categories of mass spectra on the test sets. All the parameters used for the various functions are in table S4. The 36 pipelines were assessed on the test datasets, using both Accuracy and Cohen's Kappa criteria, leading to 8604 assessments (Table S1).

Using this methodology, we have made a number of observations on (i) the impact of the *m/z* selection method on the performances of classification models (ii) the impact of dataset size on the selection of *m/z* values and (iii) the global performances of the used pipelines (*m/z* selection + prediction algorithms).

The first observation (i) suggests the importance of performing a variable selection step before estimating a classification model. Regardless of the prediction algorithm used and the dataset size, the "cvp" and "mda" *m/z* selection methods significantly improve both Accuracy and Cohen's Kappa when compared to not selecting *m/z* (e.g. estimating the classification model on all the *m/z*) or using the "SelectionVarStat", "RFERF", or "sPLSDA" methods [*Figure 2A*]. This observation must nevertheless be put into perspective on data sets composed of many mass spectra (more than 1000). In such cases, "sPLSDA" and "RFERF" demonstrate similar performance levels to "cvp" and "mda", while the "SelectionVarStat" method's performance declines with increasing dataset size, leading to performances similar to not selecting *m/z* values [*Figure 2A*].

The second observation (ii) pertains to the number of selected *m/z* values. When there

are more mass spectra in a training dataset, the selection methods tend to select a greater

number of potentially discriminant *m/z* [*Figure 2B*]. In general, "SelectionVarStat"

tends to select the most *m/z* values among the different methods, followed by "mda"

and "cvp", while "sPLSDA" and "RFERF" tend to select very few *m/z* compared to the

other methods [*Figure 2B*]. Selecting more *m/z* can potentially lead to reduced

performance due to errors in mass spectra classification as previously observed with

"SelectionVarStat". That is why "sPLSDA" and "RFERF" are relatively robust to the

dataset size increasing [*Figure 2A*].

Figure 2. Performances of pipelines implemented in MSclassifR on 239 datasets of various sizes

A. Violin plots of Cohen's Kappa depending on the variable section method used. The dataset sizes are

in different colors. **B.** Violin plots of the number of selected *m/z* depending on the variable section method

used. The dataset sizes are in different colors. **C.** Violin plots of the Cohen's Kappa depending on the

pipeline (variable selection + prediction algorithm) used. Pipelines are ranked from the highest average

Cohen's Kappa (left) to the lowest (right) estimated on all the generated datasets. The dataset sizes are

in different colors. **Notes:** cvp: for cross-validated permutation variables importances; mda: mean

decrease in accuracy; multinom: logistic regression; nnet: single-hidden-layer neural network; No

selection: no *m/z* selection method; Predict_FastClass: multiple regression estimator; RFE-RF: Recursive

Feature Elimination with Random Forests; SelectionVarStat: ANOVA test; RF: random forests;

sPLSDA: Sparse Partial Least Squares Discriminant Analysis;; svmLinear2: linear Support Vector;

xgbTree: eXtreme Gradient Boosting. * *p-value* $< 5.10^{-2}$, ** *p-value* $< 10^{-2}$, *** *p-value* $< 10^{-3}$ (estimated

using univariate Wilcoxon rank sum test).

Concerning the global performance of the MSclassifR pipelines (observation (iii)), RF

globally outperform other prediction models [*Figure 2C*]. When assessing the

effectiveness of variable selection methods in combination with RF, "sPLSDA" and

"RFERF" exhibit significantly lower accuracy and Cohen's Kappa values compared to

alternative methods [*Figure 2C*]. Importantly, there is no significant difference in

performance when combining RF with the other methods "cvp," "mda",

"SelectionVarStat", and "No selection". These similar performances can be attributed

to the inherent functioning of the RF algorithm: it employs a random variable selection

process during the construction of each decision tree that can make the selection made

upstream obsolete. Interestingly, after performing *m/z* selection with "mda" or "cvp",

support vector machines with linear kernels "svmLinear2" perform similarly to RF and

could also be an interesting strategy to apply [*Figure 2C*]. Thus, using "mda" or "cvp"

m/z selection methods combined with either "rf" or "svmLinear2" often deliver some of

the best performances across different dataset sizes, with no significant differences

between them. We have also observed that certain strategies perform well on small size

datasets but exhibit poor performance on larger ones. Specifically, "Predict_FastClass"

associated with "mda", "cvp", "SelectionVarStat" or "No selection", as well as

"svmLinear2" combined with "SelectionVarStat" or "No selection", fall into this

category [*Figure 2C*]. While these strategies exhibit competitiveness when confronted

with constrained numbers of mass spectra in a training dataset, their performance

should be approached cautiously in broader contexts.

Application on another independent dataset

We aimed to determine if the promising pipelines, identified through extensive

comparisons, could perform well on a separate dataset (Dataset 1) with its own unique

challenges [33]. This particular dataset presents the added complexity of spectra

obtained from various MALDI-TOF instruments across different countries. It is worth

noting that the original authors applied a calibration function to mitigate inter-

laboratory differences in mass spectra, a step we did not replicate in our computational

experiments.

In our experiments, we applied the same preprocessing steps to the mass spectra as

before. We then rigorously evaluated the models by randomly dividing the spectra into

70% for training and 30% for testing. This process was repeated 15 times, resulting in

an average Accuracy of 0.875 and a maximum of 0.925 for the "mda rf" pipeline, which

had previously shown good performance (Table S2). Notably, these precision levels

closely match those reported in the article related to the dataset, even though we did not

perform any calibration. Thus, it shows the ability of MSclassifR to build robust

pipelines capable of adapting to different instruments in an automatic way.

Figure 3. Performances of pipelines implemented in MSclassifR on dataset 1 using 70% of the mass

spectra to train the pipelines and 30% to test (15 repeats for each).

A. Violin plots of the Accuracy depending on the pipelines used on Dataset 1. **B.** Violin plots of the

Accuracy depending on the resampling method used. **C.** Violin plots of the Accuracy depending on the

merging technique used. * *p-value* $< 5.10^{-2}$, ** *p-value* $< 10^{-2}$, *** *p-value* $< 10^{-3}$ (estimated using

univariate Wilcoxon rank sum test). **Notes:** comb_fisher: combine predictions with Fisher's method;

down: down resampling method; max_vote: combine predictions with hard voting system; mda: mean

decrease in accuracy; RF: random forests; up: up-resampling method; smote: SMOTE resampling

method.

To get the most out of MSclassifR, we assessed its capacity to combine prediction

probabilities through ensemble-learning methods like maximum voting and Fisher's

method. Simultaneously, we examined the potential enhancements that MSclassifR's

resampling techniques could bring to our results. For this purpose, we derived five

models from the "mda rf" pipeline and merged them after the application of diverse

dataset resampling approaches. Our findings reveal that employing up-sampling and

merging with either the Fisher's method or the maximum vote technique leads to a

significant improvement in overall performance on average (average Accuracy of 0.895

and 0.894 respectively, with maximums of 0.926 and 0.935) [*Figure 3A*]. In contrast,

the down-sampling method yielded inferior results when compared to the absence of

resampling [*Figure 3B*]. In addition, these ensemble-learning approaches make it

possible to reduce the variability of Accuracy, and therefore to obtain more robust

results [*Figure 3A-C*].

Comparison with a commercial MALDI-TOF MS system

Finally, we also conducted a comparison between all the pipelines within MSclassifR

using dataset 3 [34] and a commercial MALDI-TOF MS system (Microflex® LT

instrument (Bruker® Daltonics) IVD MALDI Biotyper®). This dataset comprises mass

spectra from three genetically narrowly related subspecies of the *Mycobacterium*

abscessus complex. The data set is particularly challenging because these subspecies

are spectrally very close and difficult to distinguish. The subspecies were characterized

using commercial molecular methods [34]. The commercial system provided

subspecies information for only 60% of the mass spectra, with a correct identification

rate of 62% (Table S3). In comparison, using only 70% of all the 1001 mass spectra to

train the pipelines and 30% to test them, this is randomly repeated 5 times, the pipelines

of MSclassifR provided a subspecies for all the mass spectra of the test set. The best

average accuracy was achieved by the "sPLSDA rf" pipeline, with an average accuracy

of 96.4% and a standard deviation of 0.029 from 157.4 selected m/z values in average

(Table S1). This highlights the ability of MSclassifR to achieve robust classification

accuracy, overcoming some limitations of commercial systems.

Conclusion

MSclassifR is a freely available R package, accessible via the CRAN repository network. It is designed for the automatic classification of mass spectra into predefined categories. In this paper, we have rigorously tested the capabilities of our package on datasets acquired from various sources and instruments, including both publicly available data and MALDI-TOF mass spectra obtained in-house. The results from our investigation reveal that MSclassifR can yield prediction models with excellent accuracy and Kappa values. Furthermore, when we compared MSclassifR predictions to those generated by commercial software on in-house dataset, we observed significant performance advantages for MSclassifR prediction models. Our article also delves into an extensive exploration of diverse pipelines for constructing prediction models with our package. From this exploration, we have drawn some important conclusions:

(a) A selection of discriminant *m/z* values prior to the estimation of the prediction model makes it possible to improve the accuracy of predictions.

(b) The original selection method that we developed and based on thresholding the variable importances of RF leads to better accuracy in predictions, whatever the size of the training dataset.

(c) Overall, the prediction method providing the best average accuracies on the used datasets is the RF algorithm.

(d) Some *m/z* selection methods perform poorly on small size datasets (less than 100 mass spectra) and provide good performances on large datasets (more than 1000 mass spectra).

(e) Some prediction methods provide good performances on small size datasets (less than 100 mass spectra) but perform badly on large datasets (more than 1000 mass spectra).

(f) Using ensemble learning and resampling techniques can provide better and more robust performances.

These outcomes are specific to the datasets and the function parameters we used, and their general applicability to any dataset might vary. Nonetheless, they provide valuable

insights into creating reliable prediction models from new coming mass spectra
datasets.

It should be emphasized that MSclassifR can be used in broader contexts than MALDI-
TOF mass spectrometry, for example when differentially analyzing quantified values
of analytes (e.g. genes, proteins, etc.) between biological conditions. Notably,
MSclassifR can be used for the analysis of liquid chromatography coupled to tandem
mass spectrometry (LC-MS/MS) data [35]. This technique can also help accurately
distinguish closely related microbial variants, which could promote a deeper
understanding of resistance mechanisms to antibiotics [36].

In conclusion, MSclassifR offers an attractive solution for biology laboratories seeking
to identify pathogens from mass spectra. Its open-source nature fosters collaboration,
customization, and innovation within the community dedicated to enhancing diagnostic
capabilities using mass spectra. We have provided online vignettes that demonstrate
how to use the package effectively. We are committed to regular maintenance and
continuous improvement of our package.
479

Data and availability

The MSclassifR R package is available online on the CRAN: https://cran.r-project.org/web/packages/MSclassifR/index.html

Links towards different vignettes illustrating how to use the functions of this package from real data sets are also available from this webpage.

All datasets used in this paper are available online on PRIDE and zenodo websites.

Authors' contributions

A.G. and Q.GG. conceived the project. A.G. and Q.GG. designed the algorithms and developed all the programs. M.M., A.A., N.V., E.V. and Y.B. have participated in design of the study and coordination and helped to draft the manuscript. A.G., Q.GG. and K.D performed computational work and analysed the data. All the authors reviewed the manuscript. All authors read and approved the final manuscript.

Competing interests

The authors have declared no competing interests.

References

1. Hou T, Chuan C, Teng S. Current status of MALDI-TOF mass spectrometry in clinical microbiology. JOURNAL OF FOOD AND DRUG ANALYSIS 2019; 27:404–414

2. Brown-Elliott BA, Fritsche TR, Olson BJ, et al. Comparison of Two Commercial Matrix-Assisted Laser Desorption/Ionization-Time of Flight Mass Spectrometry (MALDI-TOF MS) Systems for Identification of Nontuberculous Mycobacteria. Am J Clin Pathol 2019; 152:527–536

3. Suzuki H, Yoshida S, Yoshida A, et al. A novel cluster of Mycobacterium abscessus complex revealed by matrix-assisted laser desorption ionization-time-of-flight mass spectrometry (MALDI-TOF MS). Diagn Microbiol Infect Dis 2015; 83:365–370

4. Pavlovic M, Konrad R, Iwobi AN, et al. A dual approach employing MALDI-TOF MS and real-time PCR for fast species identification within the Enterobacter cloacae complex. FEMS Microbiology Letters 2012; 328:46–53

5. Godmer A, Benzerara Y, Normand AC, et al. Revisiting Species Identification within the Enterobacter cloacae Complex by Matrix-Assisted Laser Desorption Ionization-Time of Flight Mass Spectrometry. Microbiol Spectr 2021; 9:e0066121

6. Christner M, Trusch M, Rohde H, et al. Rapid MALDI-TOF Mass Spectrometry Strain Typing during a Large Outbreak of Shiga-Toxigenic Escherichia coli. PLOS ONE 2014; 9:e101924

7. Flores-Treviño S, Garza-González E, Mendoza-Olazarán S, et al. Screening of biomarkers of drug resistance or virulence in ESCAPE pathogens by MALDI-TOF mass spectrometry. Sci Rep 2019; 9:18945

8. Weis C, Cuénod A, Rieck B, et al. Direct antimicrobial resistance prediction from clinical MALDI-TOF mass spectra using machine learning. Nat Med 2022; 28:164–174

9. Bruyne KD, Slabbinck B, Waegeman W, et al. Bacterial species identification from MALDI-TOF mass spectra through data analysis and machine learning. Systematic and Applied Microbiology 2011; 34:20–29

10. Khot PD, Fisher MA. Novel approach for differentiating shigella species and escherichia coli by matrix-assisted laser desorption ionization-time of flight mass spectrometry. Journal of Clinical Microbiology 2013; 51:3711–3716

11. Desaire H, Hua D. Adaption of the Aristotle Classifier for Accurately Identifying Highly Similar Bacteria Analyzed by MALDI-TOF MS. Analytical Chemistry 2020; 92:1050–1057

12. Asakura K, Azechi T, Sasano H, et al. Rapid and easy detection of low-level resistance to vancomycin in methicillin-resistant Staphylococcus aureus by matrix-assisted laser desorption ionization time-of-flight mass spectrometry. PLOS ONE 2018; 13:e0194212

13. Gibb S, Strimmer K. Maldiquant: A versatile R package for the analysis of mass spectrometry data.
Bioinformatics 2012; 28:2270–2271

14. Palarea-Albaladejo J, Mclean K, Wright F, et al. MALDIrppa: quality control and robust analysis for
mass spectrometry data. Bioinformatics 2017; 34:522–523

15. Kuhn M. Building Predictive Models in R Using the caret Package. Journal of Statistical Software
2008; 28:1–26

16. Rohart F, Gautier B, Singh A, et al. mixOmics: an R package for 'omics feature selection and multiple
data integration. PLOS Computational Biology 2017; 13:1–19

17. . Clover Data Analysis. Clover MS Data Analysis

18. . Bruker Daltonics ClinProTools 2.0 Download (Free trial)...

19. Savitzky Abraham, Golay MJE. Smoothing and Differentiation of Data by Simplified Least Squares
Procedures. Anal. Chem. 1964; 36:1627–1639

20. Mallat S. A Wavelet Tour of Signal Processing Academic Press, New York. 1998;

21. Ryan CG, Clayton E, Griffin WL, et al. SNIP, a statistics-sensitive background treatment for the
quantitative analysis of PIXE spectra in geoscience applications. Nuclear Inst. and Methods in Physics
Research, B 1988; 34:396–402

22. Sauve AC, Speed TP. Normalization, baseline correction and alignment of high-throughput mass
spectrometry data. Proceedings Gensips 2004; 1–4

23. Andrew AM. Another efficient algorithm for convex hulls in two dimensions. Information Processing
Letters 1979; 9:216–219

24. Dieterle F, Ross A, Schlotterbeck G, et al. Probabilistic Quotient Normalization as Robust Method
to Account for Dilution of Complex Biological Mixtures. Application in 1H NMR Metabonomics. Anal.
Chem. 2006; 78:4281–4290

25. Cleveland WS. Robust locally weighted regression and smoothing scatterplots. Journal of the
American statistical association 1979; 74:829–836

26. Eilers PH, Marx BD. Flexible smoothing with B-splines and penalties. Statistical science 1996;
11:89–121

27. Leys C, Ley C, Klein O, et al. Detecting outliers: Do not use standard deviation around the mean, use
absolute deviation around the median. Journal of experimental social psychology 2013; 49:764–766

28. Friedman JH. A variable span smoother. 1984;

29. Stanford TE, Bagley CJ, Solomon PJ. Informed baseline subtraction of proteomic mass spectrometry
data aided by a novel sliding window algorithm. Proteome Science 2016; 14:19

30. Díaz-Uriarte R, Alvarez de Andrés S. Gene selection and classification of microarray data using
random forest. BMC Bioinformatics 2006; 7:3

31. Storey JD. A direct approach to false discovery rates. Journal of the Royal Statistical Society: Series
B (Statistical Methodology) 2002; 64:479–498

32. Janitza S, Celik E, Boulesteix A-L. A computationally fast variable importance test for random forests
for high-dimensional data. Advances in Data Analysis and Classification 2018; 12:885–915

33. Nachtigall FM, Pereira A, Trofymchuk OS, et al. Detection of SARS-CoV-2 in nasal swabs using
MALDI-MS. Nat Biotechnol 2020; 38:1168–1173

34. Godmer A, Aubry A, Gianetto QG, et al. Spectral database of the subspecies of the Mycobacterium
abscessus complex (MALDI-TOF Mass Spectrometry). 2021;

35. https://agodmer.github.io/MSclassifR_examples/Vignettes/Vignettemsclassifr_DAv3.html

36. Roux-Dalvai F, Gotti C, Leclercq M, et al. Fast and accurate bacterial species identification in urine
specimens using LC-MS/MS mass spectrometry and machine learning. Mol Cell Proteomics 2019;
18(12): 2492-2505

577