A Rapid, High Throughput, Viral Infectivity Assay using Automated Brightfield Microscopy with Machine Learning

Rupert Dodkins[1], John R. Delaney[1], Tess Overton[2], Frank Scholle[2], Alba Frias[3], Elisa Crisci[3], Nafisa Huq[4], Ingo Jordan[5], Jason T. Kimata[6], and Ilya G. Goldberg[1]

[1]ViQi Inc., Santa Barbara, CA, 93117, United States
[2]Department of Biological Sciences North Carolina State University Raleigh, NC 27695, USA
[3]College of Veterinary Medicine, Department of Population Health and Pathobiology, North Carolina State University, Raleigh, NC 27695, United States
[4]Melbec Microbiology Ltd, Rossendale, Lancashire, BB4 4QJ, United Kingdom
[5]ProBioGen AG, Goethestr. 54, 13086 Berlin, Germany
[6]Department of Molecular Virology and Microbiology, Baylor College of Medicine, Houston, TX, United States

Abstract

Infectivity assays are essential for the development of viral vaccines, antiviral therapies and the manufacture of biologicals. Traditionally, these assays take 2–7 days and require several manual processing steps after infection. We describe an automated assay (AVIA™), using machine learning (ML) and high-throughput brightfield microscopy on 96-well plates that can quantify infection phenotypes within hours, before they are manually visible, and without sample preparation. ML models were trained on HIV, influenza A virus, coronavirus 229E, vaccinia viruses, poliovirus, and adenoviruses, which together span the four major categories of virus (DNA, RNA, enveloped, and non-enveloped). A sigmoidal function, fit to virus dilution curves, yielded an R^2 higher than 0.98 and a linear dynamic range comparable to or better than conventional plaque or $TCID_{50}$ assays. Because this technology is based on sensitizing AIs to specific phenotypes of infection, it may have potential as a rapid, broad-spectrum tool for virus identification.

Vaccination is the most efficacious and economic intervention for prevention of infectious diseases, spillover events from animals to humans, and (in the case of veterinary vaccines) food security [1–3]. Importance of development of novel vaccines for pandemic preparedness, or adequate stockpiles of proven vaccines, has been drastically demonstrated in the coronavirus disease 2019 (COVID-19) outbreak that resulted in hundreds of millions of cases, and almost 6 million deaths worldwide to date[4].

Research on vaccines and antiviral therapies depend on infectivity assays for quantification of vector design decisions, purification processes and efficacy. However, infectivity assays are also central to the development and release of biologicals, medicines made from living cells taken from animals and thus at risk of contamination with adventitious agents [5,6]. As much as a third of all infectivity assays currently being done are for monitoring viral clearance in the manufacturing process for biological therapies [7,8]. For such assays a set of standard viral stocks is used to spike a manufacturing process and the infectious titer is measured before and after various process steps. The FDA outlines a reduction in infectivity of 10^{15} for all manufacturing steps

combined, for 11 different viruses, and these assays are typically done with conventional plaque assays[9].

All combined, any improvement in the pace of measuring viral infectivity, its cost, or the reliability of the result can thus have profound consequences for the development of conventional, lipid-encapsulated or viral-vectored vaccines, antiviral drugs and biocides, biological therapies that require clearance assays, as well as therapies that use viral vectors such as gene therapy, cancer immunotherapy, and oncolytic therapies.

Infectivity assays operate by observing and quantifying the effects of viral replication on a culture of susceptible or permissive cells. For example, in plaque assays, cells are infected on 6-well or 24-well plates and overlaid with an immobilizing medium that localizes viral spread to neighboring cells only [10,11]. This causes infected cells to form localized clusters of cytopathic effects (CPEs; dead or fused cells) that can be identified and counted to measure viral infectivity. The problem is that it takes 2-14 days to form plaques large enough to be identified by eye [12,13,14]. The long incubation time slows down research, the overlays interfere with automation, which limits throughput, and the large format of the assay contributes to costs, especially in high containment facilities[15].

The Tissue Culture Infectious Dose (TCID$_{50}$) assay is an endpoint dilution assay used to find a specific dilution of virus stock where 50% of the infected wells show signs of being infected after a set incubation period[16]. As with plaque assays, an extended incubation time is required for multiple rounds of infection to occur so that a well of cells initially infected with a single infectious viral particle will become visibly infected. Unlike plaque assays, the count of infection events is much lower, which contributes to reduced precision. The TCID$_{50}$ assay is typically done on 96-well or 384-well plates, but the number of samples per plate is limited by the requirement for multiple wells (usually 6-8 per replicate) for each dilution[17]. For obvious infection phenotypes like cell lysis, a well can be identified as infected vs. uninfected reliably without additional reagents, while more subtle phenotypes or more reliable automation may require more expensive reagents to detect infection reliably. This assay can be largely automated and it provides an estimate of the titer range but it is not as precise as a standard plaque assay[18]. The limited sensitivity to small differences in infectivity can cause promising lead compounds or vaccines to be missed. Typically TCID$_{50}$ assays determine titer to within a factor of 5 or 10, while plaque assays can be accurate to within a factor of 2 [19,20].

The fluorescent focus assay or focus forming assay (FFA) is a variation of the plaque assay, but instead of relying on multiple rounds of infection and cell lysis in order to detect plaque formation, FFA employs immunostaining with detection of particular viral antigen expressed in the infected cells and is developed either using fluorescently labeled antibodies or immunocytochemistry techniques[21]. Fluorescent protein genes can also be expressed by introducing them into a recombinant viral genome. FFA can be used to detect individual infected cells before a plaque is formed, allowing this technique to yield results in less time than plaque or TCID$_{50}$ assays. The labeled infected cells can also be counted by flow cytometry [22]. However, FFA is dependent on a consistent supply of antibody reagents or transgenic viral strains, and the experimental analysis requires multiple manipulations that can make the final quantification subjective. The requirement for reagents and processing steps necessary for antibody-based detection can be avoided by developing viral strains carrying genes for fluorescent proteins, luciferase or beta-galactosidase. Including marker genes in therapeutic viral vectors may be problematic, and not all viruses are available in recombinant formats[13]. Laser Force Cytology uses a flow cell and backscatter imaging to discern infected from uninfected cells [23,24]. This technique possesses many of the benefits of FFA, such as rapid turnaround time, without the

requirement of cell labeling. However, this method requires specialist hardware, has a low limit of detection, and has challenges in scalability across many samples due to its reliance on a flow cell. Using deep learning and convolutional neural networks has been previously used on brightfield images of infected cells to detect CPEs[27,28]. These studies focused on a single virus each rather than proposing a generalized infectivity platform, and unlike this study, the CPEs are manually visible at the timepoints that were used.

A key difference between AI-based image processing and conventional image analysis that is used in FFA processing or automated plaque counting, is how variability in the assay is accounted for. This includes variation in cell health and density, variation in virally-induced infection phenotypes, and variation from well to well, plate to plate, and day to day. Conventional image processing relies on choosing parameters that are sufficiently tolerant of the variability without adversely affecting the assay's sensitivity or precision. The parameter choice is typically in the hands of the experimentalist, and this can make the assay results dependent on the parameters chosen or the person choosing them. This can be even more subjective when the assay is directly read out by the experimentalist without software. When using AI-based techniques, the variability present in the assay is trained into the AI, since there are no image processing parameters for the experimentalist to choose. While different experimentalists may introduce bias by consistently growing slightly different looking cells and producing slightly different viral infectivity phenotypes, this difference can be quantified by having the two researchers titer the same viral sample, and importantly, the AI can be trained to ignore these variations by being trained with samples from multiple researchers.

We present a novel infectivity assay that solves the shortcomings of the traditional approaches listed above. The automated viral infectivity assay (AVIA) uses computer vision methods based on machine learning (ML) to analyze brightfield images collected from 96-well plates infected with limited dilutions of virus-containing samples. AIs can be trained to detect subtle cell morphological differences associated with virus infection prior to any manually visible cytopathic effects like swelling, fusion, or cell death [25,26,23]. This is made possible by these algorithms having access to the full 12-16 bits of dynamic range in modern sensors (as opposed to the approximately 6-7 bits of the human retina). These models detect trends across a vast range of images simultaneously and are specialized to ignore the large amounts of irrelevant image information in complex images such as brightfield. The use of brightfield microscopy obviates cell fixation and staining, which allows the technique to be more easily automated as well as making it faster, safer, more reproducible, and more cost effective. Due to the quantitative assay readout for each well, fewer dilutions are required per sample than $TCID_{50}$, freeing up valuable plate real estate, which increases sample throughput and further reduces cost per sample. Furthermore, the assay does not rely on plaque formation and is therefore not limited to plaque forming viruses (see Supplementary Fig. 2 with HIV for example). Lastly, the increased automation, reduction of sample preparation steps, and fully automated analysis ensures that results are consistent from day to day and from researcher to researcher. Table 1 summarizes the key metrics of the described infectivity assays, where the values for the traditional techniques are compiled from publications that made these comparisons [29,13,30,19,31, 32,33].

This paper presents the development and validation of the AVIA assay platform. In the Automated Viral Infectivity Assay section, we describe the assay and the metrics of success used in the investigations. The Viral versatility and Example inactivation assay sections showcase the training performance of AVIA on a selection of viruses investigated so far as well as an example inactivation assay. In Supplementary Materials, we explore optimization of training parameters, and generalizability of models across experimental conditions.

Automated viral infectivity assay

The challenge in training machine learning models on images with changes that cannot be observed by eye is setting up experimental conditions so that the images used for training have a well-defined ground truth. To accomplish this, we start with the assumption that a successful infectious viral particle will induce a detectable morphological change in the host cell. This way, a binary classifier model could be trained using cells infected with a high multiplicity of infection (MOI) to serve as the positive class, and mock-infected cells with no exposure to viral particles to serve as the negative class. The MOI of the positive class must be sufficiently high that all cells can be assumed to be infected synchronously as any uninfected cells will contaminate the positive class with false positives and place a limit on the accuracy of the trained model. For example, for an MOI of 3, $1-e^{-3}$ = 95% of cells will be infected according to Poisson statistics. Several identical wells of infected and uninfected cells are imaged at multiple locations to satisfy the large number of images needed to train an AI model based on deep-learning/convolutional neural networks[34].

Fig. 1 | Model workflow for AVIA in training (A) and conducting an assay (B) using a trained model. After the cells are incubated, the remainder of the process is entirely automated improving reliability, cost, and scaling potential. The machine learning workflow includes training an ensemble of machine learning models to improve accuracy and generalizability.

The workflow for training an AVIA model is as follows – a cell culture is first plated and infected, just as in a traditional assay. The cells are then left to incubate while a series of brightfield images are captured with a high-throughput microplate imager [35–38]. If the microplate imager includes an incubator (temperature control, CO_2), then this can happen automatically, otherwise the plate has to be manually transferred to the microplate reader at the selected timepoints. Then, the images and plate layout are transferred to ViQi Inc. servers, at which point the processing steps are complete for the user.

Currently, when first operating on a new virus, cell line, or imaging device, a new AI needs to be trained using the same cells, virus, plates and imaging instrument as will be used routinely for infectivity assays (Fig. 1A). Once a model is created, the user can conduct various infectivity assays such as inactivation assays, clearance assays or routine virus titers (Fig. 1B). For a

routine assay, a 96-well plate can contain 3-4 replicates and 3-4 virus dilutions, allowing 8-12 independent samples per 96-well plate.

Fig. 2 shows the input data for training and validating an AVIA model. A single 96-well plate is populated with a cell monolayer and infected at a range of MOIs according to the layout in Fig 2A, in which a quarter of the wells are infected at a high MOI and a quarter of the wells are uninfected (the training protocol), and half of the plate has a dilution series to validate the model performance on partially infected wells (validation protocol). To further increase the spatial resolution and maximize the number of training samples, the images are split up into square subregions, or 'tiles'. Fig. 2B and 2C show example images from the two training quadrants and a potential tile separation scheme – here the images would each yield 256 tiles of size 128x128px.

While training, the model is exposed to many examples of images from either class, along with their labels, and the achieved binary accuracy is monitored for train (seen, blue) and test (unseen, red) data (Fig. 2D and E). When the test curve falls below the training curve, it demonstrates that the model is overtraining to the training data (Fig. 2D). The deep neural networks used in AVIA contain millions of free parameters, and with a finite amount of training data, some amount of overfitting to the specific examples in training is inevitable. Occasionally test data shows better performance than training data (Fig. 2E), indicating that the dropout and regularization training parameters used to suppress overtraining may be too aggressive.

A successfully generalized model would be able to identify tiles as infected or uninfected in images at a range of MOIs between the two extremes used in training. A sigmoidal trend is expected with plateaus towards the two extreme MOIs used in training and a linear range along which an infectivity assay would be sensitive (Fig. 2F). To quantify this, infectivity is defined as the fraction of tiles with a raw AI score above the threshold of 0.5 in a given image. These image scores are grouped according to MOI and their means and standard error define the predicted infectivity for a given MOI. This curve is fit with the four parameter logistic model (4PL)

$$y_j = a + \frac{d-a}{1+\left(\frac{x_j}{c}\right)^d},$$

where y_j is the infectivity at MOI x_j, a is the infectivity of the upper asymptote, d is the infectivity of the lower asymptote, c is the MOI at the inflection point of the curve, and b is related to the slope at the inflection point[39]. The datapoints from the training protocol (e.g. MOI 0.0 and 2.0), that were reserved for testing the AI, are included in these fits. To avoid using multiple points at the high and low plateaus in fitting Eq. 4PL, a series of fits were performed by cumulatively masking points from the high and low MOIs until the minimum valid number of datapoints was reached. The fits are evaluated by their R^2 and the top 3 are reported. This curve also serves as the calibration for the assay, which allows converting the infectivity output of the AI (Y axis in Fig. 2F) to a multiplicity of infection (MOI) estimate. The quality of the fit indicates the quality of the predicted MOI when the AI's infectivity output falls within the linear range of the assay (green zone in Figure 2F).

Fig. 2 | Training and validation of AVIA ML models. A, the plate is partitioned into training MOIs (mock infected and saturated infection) and validation MOIs (serial dilutions). **B** and **C**, brightfield images are split into tiles, which are used as inputs to train, test and validate the model. The performance of the model on test data relative to train tiles determines the degree of overtraining (**D** and **E**). **F**, a successful model predicts infectivity with plateaus at low and high MOI and a linear range at intermediate MOIs on validation data. The error bars show the standard error of the images for a given MOI.

The kinetics of each virus and cell line combination vary. It is not clear *a priori* when the ideal time is to train and conduct an assay after infection. As time progresses, the signals from infection will generally get stronger, until they are so strong that they can be observed by eye. There is therefore an optimization of choosing a timepoint as early as possible to minimize the turnaround time of model training and assaying, but still retaining sufficient accuracy that the results are trustworthy and interpretable. Training the AI on phenotypes associated with productive infection rather than cell death or necrosis should also help with specificity. We therefore train models at several timepoints during the infection and compare the results. These experiments can consist of several binary models that then get compared or one multiclass model with infected and uninfected classes for the entire timecourse. This time analysis is typically performed before training and validating at a single timepoint.

Results

Viral versatility

With the model workflow optimized, and the degree of model generalizing power quantified, we are now in a position to demonstrate the application of AVIA on a set of viruses. For each virus, a separate model was trained using the MobileNet architecture, 128px tiles, 40% dropout, 1e-4.5 learning rate, but different selected assay timepoint and number of base CNNs that

compromise the model – single (S), 8 CNNs (E8), or 16 CNNs (E16). These performances could be improved by gathering more training examples, exploring more of the hyperparameter space, or incorporating more hyperparameter combinations into an ensemble model architecture if necessary. The assay timepoint was determined by selecting the minimum timepoint in the productive phase of the infection (i.e., not the first timepoint) with the highest test accuracy. This timepoint was used for training binary AIs for each virus (infected vs. uninfected), and validating their predictions using serial dilutions with the indicated MOIs determined by conventional means (plaque assay or TCID$_{50}$).

Fig. 3 shows the results of training and validation for each of the seven viruses. For each column, the top panel shows the development of the model performance on training and testing data as the training progresses through its cycles (epochs). The middle panel separates test performance into infected and uninfected classes to reveal any biases in the model towards false positives or false negatives. The bottom panel demonstrates model performance on MOIs from serial dilutions.

After training hundreds of models on different experimental setups and hyperparameter conditions a few metrics of success have been developed for model training and validation. These are:

1. A test accuracy plateau (or subsequent downtrend after peak accuracy) within the range of training epochs.
2. A peak test accuracy above 75%.
3. A peak test accuracy greater than 95% of the peak training accuracy.
4. A ratio of true positive rate to true negative rate (and vice versa) of less than 1.10.
5. A linear range of the assay, as determined by the sigmoid fit, within the bounds defined by the MOI dilutions in the calibration curve.
6. A linear range spanning more than a 10-fold range (comparable to that used in a traditional plaque assay with a countable range of 10-100 plaques per well) or equivalently containing ≥ 3 MOI dilutions.
7. A difference in the predicted MOI plateaus of ≥ 0.5.

Applying these seven metrics to the seven virus models shown, it is observed that all models satisfy criteria #1 and #2. These models were all trained with the EfficientNet-based CNN architecture which has smooth monotonic curves on AVIA data tested, as opposed to the VGG-based architecture that tended to find multiple local minima per 100 epochs for the selected hyperparameters. HIV would have failed criteria #3, but with the aid of the ensemble model method, all models pass. Three models narrowly fail criteria #4 achieving a ratio of 1.12 for HIV, MVA and adenovirus.

All but one model (vaccinia) satisfy criteria #5. This criteria is meant to ensure that the upper and lower bounds of the linear range are not beyond the MOIs in the calibration curve. Vaccinia appears to not have a well defined upper bound. This is either caused by an overestimate of the virus titer, or could potentially be addressed by improving the AI accuracy, thereby improving the false positive rate and shifting the calibration curve to the left. An improvement can be achieved using an ensemble AI or potentially larger tiles, which is currently under investigation.

Three models fail criteria #6, and have an assay range < 10-fold. In two cases (HIV and adenovirus), their linear ranges are still enough to cover three 2-fold dilutions for an 8-fold assay range. Vaccinia has a reduced range of 6.7. This can be addressed experimentally in the assays by using 2 or 3-fold dilutions rather than 5 or 10-fold ones. The AI can also be potentially improved as discussed above. Lastly, two models fail criteria #7. HIV has a linear prediction range of 0.5, passing narrowly, and the two MVA strains have a range of 0.4. Both MVA strains were already using a 16-model ensemble, so further improvements to AI accuracy could potentially come from larger tiles, or using both larger tiles and dilution data for training regressor AIs (see discussion). A reduced numerical range for valid predictions doesn't appear to adversely affect the other aspects of assay performance discussed above, however.

Fig. 3 | Model training and validation performance on six viruses. The top row shows the progress of model accuracy on train (seen by the AI, blue) and test (unseen by the AI, red) data, the middle row shows the test data accuracy for each of the two classes, the bottom row shows the model predictions on a range of intermediate MOIs from serial dilutions. The green box defines the assay's linear range from the best fit for a sigmoidal function as described in the text. The error bars show the standard error of the images for a given MOI.

Table 2 shows the versatility of AVIA. It demonstrates that a model could be trained for a virus/cell line combination within each major virus category of RNA, DNA, enveloped, and non-enveloped. Each model's performance on dilution validation data achieves a clear sigmoidal trend correlated with MOI with an $R^2 > 0.99$. These AI models were captured on four different microplate reader models, at five different laboratories, on seven different cell lines.

Example inactivation assay

The primary goal of AVIA is to create a viable method of infectivity quantification to increase the rate of vaccine development and other assays dependent on quantifying live virus titers. In this section, we demonstrate the applicability of AVIA in the context of an inactivation assay. Specifically, the model trained on Influenza A virus-infected MDCK cells at 16h after infection was used to quantify the effect of a biocide on infectivity of this virus.

The virus sample used in this experiment was treated with either media or the biocide product. The cells were then infected at a range of dilutions of virus, incubated, and imaged at 24h, 30h and 48h. If the biocide was successful, the dilution curve would be translated left towards lower MOIs.

Fig. 4 shows the AVIA predicted dilution curves for the treated and untreated populations at the three timepoints after infection. The same sigmoidal four-parameter model was used to fit treated and untreated dilutions at each timepoint. One of the four parameters, the inflection point, marks the location where the slope changes direction. The difference in the inflection point between the two curves quantifies the magnitude of the effect of the biocide. On a logarithmic scale, this parameter is called the log-difference or log-inactivation. This value is displayed in the legend for each timepoint. The log-reduction average and standard deviation are -1.48 and 0.03, respectively. These findings were consistent with the log-inactivation measured by conventional means[40] (personal communication).

Fig 4 Inactivation assay on influenza A virus-infected cells treated with media or biocide.
Infections were incubated for different times (24h, 30h, 48h; **A–C** respectively), imaged, and analyzed with a pre-trained AI (Influenza A at 16 hours post infection). X Axis: known virus dilutions; Y-axis: AI predictions of infectivity. Pink range represents the difference between IC50s for the fits to the media and product dilution series. The error bars show the standard error of the images for a given MOI.

It is noteworthy that the inactivation assay successfully quantified the log-inactivation at three incubation times well after the timepoint used to train the model. This demonstrates that the infectivity phenotypes present at 16 h were still being expressed in the cell culture at those later timepoints. The higher dilutions and longer timepoints allowed for multiple rounds of infection to be observed, expanding the range of log-inactivation observable in a single experiment.

Importantly, the log reduction measurement was stable across the three incubation times, with the standard deviation of 0.03 being approximately 2% of the mean. This provides further confidence for the robustness and precision of AVIA, and its applicability in real-world inactivation assays.

Discussion

Incubation time occupies the vast majority of the total turnaround time of traditional assays for determination of infectious units. The timecourse analysis with AVIA revealed that infection phenotypes could be detected within hours after infection. This represents a dramatic reduction in the required incubation time for ML-assisted assays over traditional methods, in addition to the improvements in reproducibility, cost per sample, and throughput.

In HIV/TZMbl and Influenza/MDCK virus-cell line combinations tested, the model was able to distinguish infected from uninfected cells at the first measurement with relatively good accuracy. It is estimated that the time between infection and imaging the first time point would be between 15 and 45 minutes. This is likely to be well below the amount of time required for the infection to take hold and for the cells to show any early signs of virus production. One potential mechanism for these early stage phenotypes is cell signaling as a result of viral infection [41,42]. We have not investigated the potential biological significance of these rapidly changing early phenotypes, but this could be of interest for further investigation. Artefactual shifts in phenotypes can be caused by CO_2 and temperature shifts resulting from imaging plates outside on an incubator, and uneven evaporation during incubation. Differential effects from virus-containing media can be caused by cytokines or other factors carried over with the media at high-titer infections. These effects will be addressed in future studies addressing the mechanism of these early phenotypes.

Splitting images into tiles allows for a larger batch of training examples to fit into memory during each training step. Larger batch sizes expose the machine learning model to a greater variety of training examples simultaneously and decrease the volatility in each step of the training progression due to a larger sample size. However, a smaller area per tile means each positive tile is less likely to contain legitimate signals of infection. A greater number of smaller tiles also splits signals at tile boundaries and reduces contextual information. It has been argued that tiling images only serves to hurt performance [43]. In this application, we believe there is more of a balance because a model trained on large tiles may perform better on the saturated infection in the training set, but we expect that the loss of ability to discriminate individual cells in large tiles and the reduced sampling will hurt performance in the sub-saturated infections used in real assays. The reduced number of training samples may also hurt the robustness of the trained models on unseen replicates. As expected, our preliminary observations are that larger tiles improve AI accuracy on the infected vs. uninfected cells (not shown). However this area deserves further investigation to determine the effect of reduced sampling, and of scoring large tiles that contain both infected and uninfected cells. It is also possible to train regressor models on intermediate dilutions of virus, so that AIs are exposed during training to the intermediate infection levels that they will be scoring when performing assays. Using large tiles with intermediate infections during training may overcome the effects of reduced sampling with larger tiles.

In most of the experiments described here, we could not detect differences between infected and uninfected cells by visual inspection, so the AIs appear to base their decisions on properties that a human observer cannot directly perceive. Some work has gone into trying to visualize

what trained AIs perceive [44,45], including generative models that produce images to illustrate the differences they detect [46,47]. We have not pursued trying to visualize or describe what our AIs use to differentiate infected from uninfected cells, primarily because our interest is currently limited to the utility of this tool in practical assays. We can however compare phenotypic similarities between cells infected with different viruses and time points (see Supplementary Materials sec. Timecourse Analysis for examples) and determine if these can be clustered by phenotype as perceived by an AI. We have used this technique previously to cluster phenotypes induced by RNA interference [48], age-related phenotypes in *C. elegans*[49] and visual trends in modern art[50].

As more viruses are characterized, we expect there to emerge clusters of similar phenotypes induced by viral infection in a certain host cell, which may correlate with infectivity pathways and antiviral defenses used by the different viruses. A map of infectious virus phenotypes may become useful when infectious viral load monitoring becomes more important in the context of control of seasonal and epidemic infectious events. This can be an important research tool for prediction of virulence, or potentially be used in a diagnostic setting such as environmental monitoring as an early indicator of the presence of infectious viruses.

The technique was validated on influenza A virus, HIV, MVA, poliovirus, vaccinia virus, adenovirus, and an adenovirus variant, spanning the four major types of viruses. On coronavirus 229E and MVA-infected cells, the trained models were robust across unseen biological replicates, without significant loss in accuracy. The consistency of this technique was demonstrated in an inactivation assay when the log-inactivation of an unknown biocide was measured at three different times after infection, achieving reproducibility within 2%.

We anticipate that this tool could be further developed into a broad and rapid diagnostic assay from infectivity phenotypes observed for a variety of viruses.

Technique	Reproducibility	Time	Labor	Cost	Automation/ Scalability
AVIA (ours)	excellent	hours	low	low	high
Plaque Assay	good	days	high	low	low
$TCID_{50}$	medium	days	high	medium - high	high
FFA: Antibodies (Ab) or GFP	medium	day	medium[GFP] - high[Ab]	medium[GFP] - high[Ab]	low[Ab] - medium[GFP]

Tab. 1 | A summary of three traditional techniques of conducting a viral infectivity assay as well as the proposed AVIA technique. By eliminating many of the human intensive steps of the traditional techniques, AVIA enables greater reproducibility, scalability, and lower costs.

Virus	Virus type	Assay	Max AI	Single model	Linear assay	Predicted	Sigmoid

		timepoint	accuracy	or Ensemble	range (MOI)	MOI range	R^2
Coronavirus 229E	RNA/enveloped	4h	80%	S	N/A	N/A	N/A
HIV	RNA/enveloped	24h	74%	E8	0.2 - 1.6	0.4 - 0.9	0.996
Influenza A virus	RNA/enveloped	10h	89%	S	0.03 - 0.5	0.2 - 0.9	0.973
MVA-CR	DNA/enveloped	4h	81%	E16	0.1 - 0.9	0.3 - 0.7	0.990
MVA-WT	DNA/enveloped	4h & 6h	73%	E16	0.2 - 1.6	0.4 - 0.8	1.000
Vaccinia	DNA/enveloped	24h	78%	S	0.3 - 2.0	0.3 - 0.9	0.995
Polio	RNA/ non-enveloped	21h	97%	S	0.02 - 0.2	0.2 - 1.0	1.000
Adenovirus	DNA/ non-enveloped	24h	87%	S	0.01 - 0.3	0.3 - 0.8	0.996

Tab. 2 | Model performances for viruses in AVIA assays. Max AI accuracy is the test accuracy at the peak of the training curve. The linear assay range (MOI) is the difference in predicted MOI between where the linear range meets the upper lower asymptotes. Corona 229E has no corresponding validation data.

Acknowledgements

This work was supported by NSF grant 2029707 to ViQi Inc. Support for JTK was provided by the Basic Science Core of the Texas Developmental Center for AIDS Research (AI161943). Support for TO, FS, AF, and EC was provided by the NCSU Comparative Medicine Institute, CMI, and the NCSU Center for Applied Virus Experimentation, CAVE. ProBioGen thanks Lina Hanisch and Nicole Seehase for expert technical assistance and Benjamin Werdelmann from Synentec for support in image acquisition with the Synentec Scientific device. ViQi and NCSU thank BioTek for their loan of a Cytation™ 1 to NCSU

Author contributions

I.G.G. designed the study. T.O., F.S., A.F., E.C., N.H., I.J., and J.T.K. acquired the data. R.D., J.R.D., and I.G.G. wrote the code. R.D. and J.R.D. performed the analysis. I.G. supervised the work. All authors read and corrected the final manuscript.

Competing interests

R.D., J.R.D., and I.G.G. are employees and stockholders of ViQi Inc, which is commercializing this assay. This work was supported by NSF SBIR Phase I grant 2029707 to ViQi Inc. There has been no other significant financial support for this work that could have influenced its outcome. The remaining authors declare no competing interests.

Methods

Viral cultures

Datasets were prepared for 7 different virus/cell lines pairs from 6 different laboratories according to the training protocol described in Fig. 2 (mock-infected, synchronous high-MOI

infection, and infections with 2-fold dilutions of virus). Except as noted in the text, infections and imaging were conducted on 96-well plates with flat plastic bottoms. Because the incubation time is low, the seeding density of the cells was relatively high (typically > 85% confluence), except as noted in the text. The dilution MOIs were predetermined in the different laboratories using $TCID_{50}$ or plaque assay and assumed to be accurate to within a factor of 2.

The cultures for Corona 229E-infected Huh7 cells and Influenza A-infected MDCK cells were prepared as described in et al.[51,] and Crisci et al.[52], respectively. The 229E images used in Supplementary Fig. 1 were captured at 20x and 40x using a manual microscope. Other 229E and influenza/MDCK images (see Supplementary Fig. 9) were captured on a Cytation 1[37] at 20x magnification. Modified vaccinia ankara (MVA) wildtype (WT) and CR-infected CR.pIX cells[53] were prepared as described in Jordan et al.[54], and the images were captured with a Synentec NyONE[36] at 20x. Human immunodeficiency virus (HIV)-infected TZMbl cells were prepared as described by Wang et al.[55] and the images captured with a Cytation 5[38] at 20x. Adenovirus-infected HeLa cells were prepared as described in Jogler et al.[56], vaccinia-infected BHK-21 cells were prepared as described in Earl et al.[57], polio-infected HeLa cells were prepared as described in et al.[58]. Influenza-infected MDCK cells (see Supplementary Fig. 9) were prepared as described by Kim et al.[59] and the images were captured with a Synentec CellaVista[35] at 20x.

Imaging
Each well on the 96-well plates was imaged 13-25 times using a grid of non-overlapping images. Images were collected every 2-4 hours before the onset of visible CPEs. Imaging was done with a high-throughput microplate imager or manual microscope (ZEISS Axio Imager/AxioCam HRc) as noted above. The sensor sizes were as noted in Table [2; virus table]. The focal plane of the images was set to the interior of the cells as opposed to the tops of the cells as is commonly done for brightfield to optimize cell segmentation tasks. The images were stored as 16-bit grayscale TIFF-files or 48-bit RGB (16-bit per channel), in which case they were converted to grayscale using the NTSC formula[60].

AI training
The classification model is a convolutional neural network (CNN), which uses a series of convolution kernels to extract features from input images at increasing levels of abstraction [61]. This model was implemented using Tensorflow[62]. The weights of these kernels are initialized to those pretrained on ImageNet[63]. Bootstrapping on the knowledge acquired from those models allowed the AVIA models to converge faster and at a higher accuracy [64]. After each cycle through the full training data (an epoch) the data is transformed with flips and rotations to increase the scope of the training examples.

Depending on the strength of the morphological signals from the infection, and the allowed time budget for training, models of varying complexity are trained. A single CNN based on EfficientNetB0[65] can be used as a rapid model with modest accuracy as a stand-alone model for training. On the other end of the scale, a total of sixteen CNNs can be trained with different base models and learning parameters (hyperparameters), and their predictions fed into an ensemble classifier such as a support vector machine (SVM) or random forest classifier [66,67]. In our experiments, three other CNN architectures were used including MobileNet[68],

InceptionResNetV2[69], and VGG[70]; and hyperparameters with two options for dropout[71]; and two options for learning rate. Future work will also include tiles with multiple sizes.

In initial experiments, cells were seeded at a lower confluence more typical of conventional infectivity assays that have long incubation times to allow for cell growth. Because of the short incubation time for AVIA, in subsequent experiments, cells were seeded at higher density or > 85% confluence to maximize the number of cells and thus the amount of information available to the AI when the cells are imaged. In the earlier, more sparsely seeded datasets, we trained a prefilter network to eliminate tiles that did not have sufficient cellular material. The image data for this network was the same as for the infectivity network, except tiles were manually labeled that had little to no cellular material so that they would be eliminated by the prefilter. This prefilter network was easily trained to approximately 92% accuracy on each occasion (not shown).

Ensemble AI
In ensemble models, numerical outputs of multiple CNN models (marginal probabilities of a given tile being infected) are fed as input features into an automated feature classifier trainer developed by ViQi. This trainer uses scikit-learn[72] to try several different feature normalization, scoring and selection techniques coupled to several different high-performing classifiers. The trainer optimizes appropriate parameters for the algorithms at each stage and automatically selects the best-performing model[73] and corresponding parameter set.

References

https://scikit-learn/stable/modules/generated/sklearn.pipeline.Pipeline.html

1. Moghadas, S. M. *et al.* The Impact of Vaccination on Coronavirus Disease 2019 (COVID-19) Outbreaks in the United States. *Clin. Infect. Dis.* **73**, 2257–2264 (2021).

2. Sangster, A. V. & Barratt, J. M. Towards Ending Immunization Inequity. *Vaccines* **9**, 1378 (2021).

3. Häsler, B. *et al.* The Economic Value of One Health in Relation to the Mitigation of Zoonotic Disease Risks. in *One Health: The Human-Animal-Environment Interfaces in Emerging Infectious Diseases: The Concept and Examples of a One Health Approach* (eds. Mackenzie, J. S., Jeggo, M., Daszak, P. & Richt, J. A.) 127–151 (Springer, 2013). doi:10.1007/82_2012_239.

4. COVID-19 Map - Johns Hopkins Coronavirus Resource Center. https://coronavirus.jhu.edu/map.html.

5. Dumont, J., Euwart, D., Mei, B., Estes, S. & Kshirsagar, R. Human cell lines for biopharmaceutical manufacturing: history, status, and future perspectives. *Crit. Rev.*

Biotechnol. **36**, 1110–1122 (2016).

6. Berting, A., Farcet, M. R. & Kreil, T. R. Virus susceptibility of Chinese hamster ovary (CHO) cells and detection of viral contaminations by adventitious agent testing. *Biotechnol. Bioeng.* **106**, 598–607 (2010).

7. Goussen, C. *et al.* Viral clearance capacity by continuous Protein A chromatography step using Sequential MultiColumn Chromatography. *J. Chromatogr. B* **1145**, 122056 (2020).

8. Viral Clearance Market by Method, Application & End User 2023 | MarketsandMarkets. https://www.marketsandmarkets.com/Market-Reports/viral-clearance-market-62681197.html.

9. Research, C. for D. E. and. Q5A Viral Safety Evaluation of Biotechnology Products Derived From Cell Lines of Human or Animal Origin. *U.S. Food and Drug Administration* https://www.fda.gov/regulatory-information/search-fda-guidance-documents/q5a-viral-safety-evaluation-biotechnology-products-derived-cell-lines-human-or-animal-origin (2020).

10. Dulbecco, R. & Vogt, M. Some problems of animal virology as studied by the plaque technique. *Cold Spring Harb. Symp. Quant. Biol.* **18**, 273–279 (1953).

11. Bracci, N., Pan, H.-C., Lehman, C., Kehn-Hall, K. & Lin, S.-C. Improved plaque assay for human coronaviruses 229E and OC43. *PeerJ* **8**, e10639 (2020).

12. Baer, A. & Kehn-Hall, K. Viral Concentration Determination Through Plaque Assays: Using Traditional and Novel Overlay Systems. *J. Vis. Exp. JoVE* 52065 (2014) doi:10.3791/52065.

13. Keiser, P. T., Anantpadma, M., Staples, H., Carrion, R. & Davey, R. A. Automation of Infectious Focus Assay for Determination of Filovirus Titers and Direct Comparison to Plaque and TCID50 Assays. *Microorganisms* **9**, 156 (2021).

14. Troisi, M. *et al.* Ocular Surface Infection by SARS-CoV-2 in COVID-19 Pneumonia Patients Admitted to Sub-Intensive Unit: Preliminary Results. *Microorganisms* **10**, 347 (2022).

15. Salje, H. *et al.* Variability in Dengue Titer Estimates from Plaque Reduction Neutralization Tests Poses a Challenge to Epidemiological Studies and Vaccine

Development. *PLoS Negl. Trop. Dis.* **8**, e2952 (2014).

16. Lei, C., Yang, J., Hu, J. & Sun, X. On the Calculation of TCID50 for Quantitation of Virus Infectivity. *Virol. Sin.* **36**, 141–144 (2020).

17. Darling, A. J., Boose, J. A. & Spaltro, J. Virus assay methods: accuracy and validation. *Biol. J. Int. Assoc. Biol. Stand.* **26**, 105–110 (1998).

18. Cresta, D. *et al.* Time to revisit the endpoint dilution assay and to replace the TCID50 as a measure of a virus sample's infection concentration. *PLOS Comput. Biol.* **17**, e1009480 (2021).

19. Smither, S. J. *et al.* Comparison of the plaque assay and 50% tissue culture infectious dose assay as methods for measuring filovirus infectivity. *J. Virol. Methods* **193**, 565–571 (2013).

20. Frias-De-Diego, A. & Crisci, E. Use of Crystal Violet to Improve Visual Cytopathic Effect-based Reading for Viral Titration using TCID50 Assays. *J. Vis. Exp. JoVE* (2022) doi:10.3791/63063.

21. Elahi, S. M. *et al.* A rapid Focus-Forming Assay for quantification of infectious adenoviral vectors. *J. Virol. Methods* **297**, 114267 (2021).

22. McKinnon, K. M. Flow Cytometry: An Overview. *Curr. Protoc. Immunol.* **120**, 5.1.1-5.1.11 (2018).

23. Hayes, P. R., Przybycien, T. M. & Schneider, J. W. Viral adventitious agent detection using laser force cytology: Intrinsic cell property changes with infection and comparison to in vitro testing. *Biotechnol. Bioeng.* **119**, 134–144 (2022).

24. Hebert, C. G., Rodrigues, K. L., DiNardo, N. & Hachmann, A.-B. Viral Infectivity Quantification and Neutralization Assays Using Laser Force Cytology. *Methods Mol. Biol. Clifton NJ* **2183**, 575–585 (2021).

25. Francica, J. R., Matukonis, M. K. & Bates, P. Requirements for cell rounding and surface protein down-regulation by Ebola virus glycoprotein. *Virology* **383**, 237–247 (2009).

26.   Li, G. *et al.* Characterization of cytopathic factors through genome-wide analysis of the Zika viral proteins in fission yeast. *Proc. Natl. Acad. Sci.* **114**, E376–E385 (2017).

27.   Wang, T.-E. *et al.* Differentiation of Cytopathic Effects (CPE) induced by influenza virus infection using deep Convolutional Neural Networks (CNN). *PLOS Comput. Biol.* **16**, e1007883 (2020).

28.   Werner, J. *et al.* Deep Transfer Learning Approach for Automatic Recognition of Drug Toxicity and Inhibition of SARS-CoV-2. *Viruses* **13**, 610 (2021).

29.   Hebert, C. G., DiNardo, N., Evans, Z. L., Hart, S. J. & Hachmann, A.-B. Rapid quantification of vesicular stomatitis virus in Vero cells using Laser Force Cytology. *Vaccine* **36**, 6061–6069 (2018).

30.   Joelsson, D. *et al.* Rapid automation of a cell-based assay using a modular approach: Case study of a flow-based Varicella Zoster Virus infectivity assay. *J. Virol. Methods* **166**, 1–11 (2010).

31.   Malenovska, H. Virus quantitation by transmission electron microscopy, TCID50, and the role of timing virus harvesting: A case study of three animal viruses. *J. Virol. Methods* **191**, 136–140 (2013).

32.   Payne, A. F., Binduga-Gajewska, I., Kauffman, E. B. & Kramer, L. D. Quantitation of flaviviruses by fluorescent focus assay. *J. Virol. Methods* **134**, 183–189 (2006).

33.   Masci, A. L. *et al.* Integration of Fluorescence Detection and Image-Based Automated Counting Increases Speed, Sensitivity, and Robustness of Plaque Assays. *Mol. Ther. Methods Clin. Dev.* **14**, 270–274 (2019).

34.   LeCun, Y., Bengio, Y. & Hinton, G. Deep learning. *Nature* **521**, 436–444 (2015).

35.   CELLAVISTA. *SYNENTEC GmbH* https://synentec.com/cellavista-4/.

36.   NYONE. *SYNENTEC GmbH* https://synentec.com/nyone/.

37.   Cytation 1 Cell Imaging Multi-Mode Reader - Technical Details. https://www.biotek.com/products/imaging-microscopy-cell-imaging-multi-mode-readers/cytati

on-1-cell-imaging-multi-mode-reader/technical-details/.

38.   Cytation 5 Cell Imaging Multi-Mode Reader - Technical Details.

https://www.biotek.com/products/imaging-microscopy-cell-imaging-multi-mode-readers/cytati

on-5-cell-imaging-multi-mode-reader/technical-details/.

39.   Azadeh, M. *et al.* Calibration Curves in Quantitative Ligand Binding Assays:

Recommendations and Best Practices for Preparation, Design, and Editing of Calibration

Curves. *AAPS J.* **20**, 22 (2017).

40.   Standards, E. BS EN 14476:2013+A2:2019 Chemical disinfectants and antiseptics.

Quantitative suspension test for the evaluation of virucidal activity in the medical area. Test

method and requirements (Phase 2/Step 1). *https://www.en-standard.eu*

https://www.en-standard.eu/bs-en-14476-2013-a2-2019-chemical-disinfectants-and-antiseptic

s-quantitative-suspension-test-for-the-evaluation-of-virucidal-activity-in-the-medical-area-test-

method-and-requirements-phase-2-step-1/.

41.   Greber, U. F. Signalling in viral entry. *Cell. Mol. Life Sci. CMLS* **59**, 608–626 (2002).

42.   Fitzgerald, K. A. & Kagan, J. C. Toll-like Receptors and the Control of Immunity. *Cell* **180**,

1044–1066 (2020).

43.   Reina, G. A., Panchumarthy, R., Thakur, S. P., Bastidas, A. & Bakas, S. Systematic

Evaluation of Image Tiling Adverse Effects on Deep Learning Semantic Segmentation. *Front.

Neurosci.* **14**, (2020).

44.   Kim, I., Rajaraman, S. & Antani, S. Visual Interpretation of Convolutional Neural Network

Predictions in Classifying Medical Image Modalities. *Diagnostics* **9**, 38 (2019).

45.   Fan, F., Xiong, J., Li, M. & Wang, G. On Interpretability of Artificial Neural Networks: A

Survey. *ArXiv200102522 Cs Stat* (2021).

46.   Zhao, T. & Murphy, R. F. Automated learning of generative models for subcellular

location: Building blocks for systems biology. *Cytometry A* **71A**, 978–990 (2007).

47.   Peng, T. & Murphy, R. F. Image-derived, three-dimensional generative models of cellular

organization. *Cytometry A* **79A**, 383–391 (2011).

48.   Shamir, L. *et al.* Wndchrm – an open source utility for biological image analysis. *Source Code Biol. Med.* **3**, 13 (2008).

49.   Johnston, J., Iser, W. B., Chow, D. K., Goldberg, I. G. & Wolkow, C. A. Quantitative Image Analysis Reveals Distinct Structural Transitions during Aging in Caenorhabditis elegans Tissues. *PLOS ONE* **3**, e2821 (2008).

50.   Shamir, L., Macura, T., Orlov, N., Eckley, D. M. & Goldberg, I. G. Impressionism, expressionism, surrealism: Automated recognition of painters and schools of art. *ACM Trans. Appl. Percept.* **7**, 8:1-8:17 (2010).

51.   Peddinti, B. S. T. *et al.* Rapid and Repetitive Inactivation of SARS-CoV-2 and Human Coronavirus on Self-Disinfecting Anionic Polymers. *Adv. Sci.* **8**, 2003503 (2021).

52.   Crisci, E., Fraile, L. & Montoya, M. Cellular Innate Immunity against PRRSV and Swine Influenza Viruses. *Vet. Sci.* **6**, E26 (2019).

53.   Jordan, I. *et al.* An avian cell line designed for production of highly attenuated viruses. *Vaccine* **27**, 748–756 (2009).

54.   Jordan, I. *et al.* A Deleted Deletion Site in a New Vector Strain and Exceptional Genomic Stability of Plaque-Purified Modified Vaccinia Ankara (MVA). *Virol. Sin.* **35**, 212–226 (2020).

55.   Wang, W. *et al.* CCR5 Gene Disruption via Lentiviral Vectors Expressing Cas9 and Single Guided RNA Renders Cells Resistant to HIV-1 Infection. *PLOS ONE* **9**, e115987 (2014).

56.   Jogler, C. *et al.* Replication Properties of Human Adenovirus In Vivo and in Cultures of Primary Cells from Different Animal Species. *J. Virol.* **80**, 3549–3558 (2006).

57.   Earl, P. L., Cooper, N., Wyatt, L. S., Moss, B. & Carroll, M. W. Preparation of cell cultures and vaccinia virus stocks. *Curr. Protoc. Mol. Biol.* **Chapter 16**, Unit16.16 (2001).

58.   Fiszman, M., Bucchini, D. & Girard, M. Purification of the Sabin Strain of Poliovirus Type I Through Treatment with Sarkozyl. *J. Virol.* **7**, 687–689 (1971).

59.   Kim, E.-H. *et al.* Generation of a High-Growth Influenza Vaccine Strain in MDCK Cells for

Vaccine Preparedness. *J. Microbiol. Biotechnol.* **28**, 997–1006 (2018).

60.   Studio encoding parameters of digital television for standard 4:3 and wide-screen 16:9

aspect ratios. 20.

61.   LeCun, Y. & Bengio, Y. Convolutional networks for images, speech, and time series. in

The handbook of brain theory and neural networks 255–258 (MIT Press, 1998).

62.   TensorFlow. *TensorFlow* https://www.tensorflow.org/.

63.   Deng, J. *et al.* ImageNet: A large-scale hierarchical image database. in *2009 IEEE

Conference on Computer Vision and Pattern Recognition* 248–255 (2009).

doi:10.1109/CVPR.2009.5206848.

64.   Weiss, K., Khoshgoftaar, T. M. & Wang, D. A survey of transfer learning. *J. Big Data* **3**, 9

(2016).

65.   Tan, M. & Le, Q. V. EfficientNet: Rethinking Model Scaling for Convolutional Neural

Networks. *ArXiv190511946 Cs Stat* (2020).

66.   Noble, W. S. What is a support vector machine? *Nat. Biotechnol.* **24**, 1565–1567 (2006).

67.   Full article: Random forest classifier for remote sensing classification.

https://www.tandfonline.com/doi/full/10.1080/01431160412331269698?casa_token=aH82lqF

RnqUAAAAA%3AXp5UBT_-XLLF0KdKfdHzEPVR1XIomA20PsORVIDguRj9mSil3YAx9RyRP

auj9uB2q6FG1IKKU0mmNQ.

68.   Howard, A. G. *et al.* MobileNets: Efficient Convolutional Neural Networks for Mobile

Vision Applications. *ArXiv170404861 Cs* (2017).

69.   Szegedy, C., Ioffe, S., Vanhoucke, V. & Alemi, A. Inception-v4, Inception-ResNet and the

Impact of Residual Connections on Learning. *ArXiv160207261 Cs* (2016).

70.   Simonyan, K. & Zisserman, A. Very Deep Convolutional Networks for Large-Scale Image

Recognition. *ArXiv14091556 Cs* (2015).

71.   Srivastava, N., Hinton, G., Krizhevsky, A., Sutskever, I. & Salakhutdinov, R. Dropout: A

Simple Way to Prevent Neural Networks from Overfitting. 30.

72.   Pedregosa, F. *et al.* Scikit-learn: Machine Learning in Python. *Mach. Learn. PYTHON* 6.

73.   sklearn.pipeline.Pipeline. *scikit-learn*

    https://scikit-learn/stable/modules/generated/sklearn.pipeline.Pipeline.html.