Research article

# Effect of biomaterial stiffness on cardiac mechanics in a biventricular infarcted rat heart model with microstructural representation of *in situ* intramyocardial injectate

YD Motchon[1], KL Sack[1,2], MS Sirry[3], M Kruger[4], E Pauwels[5,6], D Van Loo[5,7], A De Muynck[5], L Van Hoorebeke[5], NH Davies[4], T Franz[1,8,]*

[1]Division of Biomedical Engineering, Department of Human Biology, University of Cape Town, Observatory, South Africa

[2]Department of Surgery, University of California at San Francisco, San Francisco, California, USA

[3]Department of Biomedical Engineering, School of Engineering and Computing, American International University, Al Jahra, Kuwait

[4]Cardiovascular Research Unit, MRC IUCHRU, University of Cape Town, Observatory, South Africa

[5]Centre for X-ray Tomography, Department of Physics and Astronomy, Ghent University, Ghent, Belgium

[6]Current affiliation: Nuclear Medicine, University Hospitals Leuven, Leuven, Belgium

[7]Current affiliation: XRE nv, Bollebergen 2B box 1, 9052 Ghent, Belgium

[8]Bioengineering Science Research Group, Faculty of Engineering and Physical Sciences, University of Southampton, Southampton, United Kingdom

*Corresponding author: thomas.franz@uct.ac.za

## Abstract

Intramyocardial delivery of biomaterials is a promising concept to treat myocardial infarction. The delivered biomaterial provides mechanical support and attenuates wall thinning and elevated wall stress in the infarct region. This study aimed at developing a biventricular finite element model of an infarcted rat heart with a microstructural representation of an *in situ* biomaterial injectate, and a parametric investigation of the effect of the injectate stiffness on the cardiac mechanics.

A three-dimensional subject-specific biventricular finite element model of a rat heart with left ventricular infarct and microstructurally dispersed biomaterial delivered one week after infarct induction was developed from *ex vivo* microcomputed tomography data. The volumetric mesh density varied between 303 mm$^{-3}$ in the myocardium and 3,852 mm$^{-3}$ in the injectate region due to the microstructural intramyocardial dispersion. Parametric simulations were conducted with the injectate's elastic modulus varying from 4.1 to 405,900 kPa, and myocardial and injectate strains were recorded.

With increasing injectate stiffness, the end-diastolic median myocardial fibre and cross-fibre strain decreased in magnitude from 3.6% to 1.1% and from -6.0% to -2.9%, respectively. At end-systole, the myocardial fibre and cross-fibre strain decreased in magnitude from -20.4% to -11.8% and from 6.5% to 4.6%, respectively. In the injectate, the maximum and minimum principal strains decreased in magnitude from 5.4% to 0.001% and from -5.4% to -0.001%, respectively, at end-diastole and from 38.5% to 0.06% and from -39.0% to -0.06%, respectively, at end-systole.

With the microstructural injectate geometry, the developed subject-specific cardiac finite element model offers potential for extension to cellular injectates and *in silico* studies of mechanotransduction and therapeutic signalling in the infarcted heart with an infarct animal model extensively used in preclinical research.

**Keywords**: myocardial infarction; biomaterial injection therapy; finite element method; cardiac mechanics

# Nomenclature

| Symbol | Description |
| --- | --- |
| a | Material parameter, dimension of stress |
| $a_{fs}$ | Material parameter defining coupling from in the fibre and sheet directions, with a dimension of stress |
| $a_i$ | Material parameter, defined for i = f and s in the fibre and sheet directions, respectively, with stress dimension |
| B | Governs the shape of peak isometric tension-sarcomere length relation |
| b | Dimensionless material parameter in Holzapfel model |
| $b_{fs}$ | Material parameter defining coupling from in the fibre and sheet directions, dimensionless |
| $b_i$ | Material parameter, defined for i = f and s in the fibre and sheet directions, respectively, dimensionless |
| $C_{10}$ | Coefficient used in Abaqus to describe the material stiffness in a Neo-Hookean strain energy density function |
| $Ca_0$ | Peak intracellular calcium concentration |
| $D_1$ | Parameter for elastic materials defining the compressibility of the material |
| E | Elastic modulus |
| $ECa_{50}$ | Length-dependent calcium sensitivity |
| h | Parameter to define the pathological degree of the tissue |
| $I_{4f}$ | Transversely isotropic invariant in the fibre direction |
| $I_{4s}$ | Transversely isotropic invariant in the sheet direction |
| $I_{8fs}$ | Orthotropic invariant from coupling in fibre and sheet direction |
| $I_i$ | Isotropic invariants in principal directions |
| J | Measures the volume change of compressible materials |
| p | Parameter scaling the isotropic response of the diseased tissue |
| T | Stress tensor |
| $T^{(a)}$ | Active stress tensor |
| $T^{(p)}$ | Passive stress tensor |

| | |
|---|---|
| $u_c$ | Unit vector in the circumferential direction |
| $u_l$ | Unit vector in the longitudinal direction |
| $W$ | Strain energy density function |
| $\alpha_h$ | Helix angle |
| $\alpha_t$ | Transverse angle |
| $\varepsilon_{EL}$ | Element strain; mean value of strains at integration points in an element |
| $\varepsilon_{IP,i}$ | Strain at integration point i in an element with i = 1 to 4 |
| $\kappa$ | Bulk modulus |
| $\nu$ | Poisson's ratio |
| $\sigma$ | Stress |

# 1. **Introduction**

Cardiovascular disease (CVD) is the leading cause of death worldwide. About 17.3 million individuals are concerned per year globally (Mendis *et al.* 2011). The current trends predict an increase in CVD related deaths from 17.5 million (i.e. 31% of global deaths) in 2012 to 22.2 million by 2030, representing an increase of 26.9%. Alarmingly, in the working-age population of low- and middle-income countries, including South Africa, the rate of people affected is becoming considerably high (Finegold *et al.* 2013, Opie and Mayosi 2005). Unfortunately, access to healthcare facilities in most sub-Saharan countries is not affordable for most populations. This situation affects the global economy and therefore influences the social cohesion in the communities.

Myocardial infarction (MI) originates from coronary occlusion, causing a lack of oxygenated blood supply to a specific myocardial region. In the long term, the cardiomyocyte death is followed by scar formation, causing excessive load to the heart, dysfunction and heart failure. Current treatments for MI include reperfusion (drug administration or surgery), mechanical device assistance (e.g. left ventricle assist device) and heart transplantation. The main objective of these treatments is to help arteries for dilatation, assist the left ventricle in pumping blood, or replace the damaged heart. New therapy approaches based on biomaterial injections have emerged. Clinical studies showed promising results addressing adverse ventricular remodelling following post-MI inflammatory response (Christman *et al.* 2004, Fan *et al.* 2019, Johnson and Christman 2013, Silveira-Filho *et al.* 2021).

Limitations of experimental studies are the cost of resources and the invasiveness of *in vivo* protocols. Computational modelling appeared as an alternative to investigate the mechanical mechanisms and parameters involved in therapeutic biomaterial injections in infarcted hearts, such as biomaterial stiffness, delivery location, injection volume, and injection pattern (Cai *et al.* 2017, Kortsmit *et al.* 2013a, Kortsmit *et al.* 2013b, Wall *et al.* 2006, Wang *et al.* 2017, Wenk *et al.* 2011, Wenk *et al.* 2009, Wise *et al.* 2016).

Wall *et al.* (2006) reported a reduction in the wall stress, affecting the ventricular function, using an ovine heart finite element (FE) model with a non-contractile biomaterial injectate in the left ventricular (LV) wall implemented by modification of the FE mesh. The results depended on the volume, the location, and the material stiffness of the injected biomaterial.

The morphology and dispersion of the intramyocardial injectates are challenging to control. Wang *et al.* (2017) studied the impact of injectate volume and stiffness on LV myofiber stress and wall thickness. They showed that a larger volume and stiffer material contributes to myofibre stress reduction and wall thickness increase in the LV, which is essential to alter ventricular remodelling.

Computational models have been used to optimize the pattern (Sepantafar *et al.* 2016, Wenk *et al.* 2009) and the volume of biomaterial injectates (Wise *et al.* 2016).

The current study aimed to develop a biventricular finite element model of a rat heart with an antero-apical infarct and a biomaterial injectate in the infarct region. Particular emphasis was placed on the realistic microstructural geometrical representation of the *in situ* intramyocardial dispersion of the biomaterial delivered one week after infarct induction. A parametric study on the effect of the injectate stiffness on the cardiac mechanics was undertaken to demonstrate the feasibility of the developed model.

# 2. Materials and methods

In the current study, the mechanical response of cardiac tissue and intramyocardially injected biomaterial were investigated in the context of a treatment for myocardial infarction. The therapy consists of intramyocardially delivering a biomaterial into the impaired heart, especially around the infarct region. Several studies have reported biomaterial injection to alter the remodelling following the infarction. The developed FE model was designed to predict the mechanical responses of the myocardium and the injected biomaterial.

## 2.1 Volumetric image data of infarcted rat heart with intramyocardial injectate

*Ex vivo* microcomputed tomography (μCT) image data of an infarcted rat heart with polymeric intramyocardial injectate from an unrelated study (unpublished data) were used for geometric reconstruction. In brief, male Wistar rats (body mass: 180-220 g) were anaesthetized, and the heart was exposed via left thoracotomy along the 4th intercostal space. Myocardial infarction was induced by permanent ligation of the left anterior descending coronary artery 3 mm distal to the auricular appendix. The discolouration of the anterior ventricular wall and reduced contractility were hallmarks of a successful occlusion of the artery. The chest was stepwise closed, and buprenorphine was administered for pain management. Seven days later, the heart was accessed via the 4th intercostal space and 100 μl of radiopaque silicone rubber containing lead chromate (Microfil® MV-120 Flow-Tech, Carver, MA, USA) diluted 1.5:1 with MV-diluent was injected into the infarct area. Dispersion and *in situ* polymerization of the Microfil® material were allowed for 30 min after the injection. The animals were then humanely killed, and the hearts carefully harvested, thoroughly rinsed with saline, fixed in a 4% paraformaldehyde solution and transferred to saline for microcomputed tomography (μCT) scanning. All animal experiments were authorized by the Institutional Review Board of the

University of Cape Town and performed according to the National Institutes of Health (NIH, Bethesda, MD, USA) guidelines.

The μCT scans were performed with a custom-made scanner with a Feinfocus X-ray tube and a Varian 2520V Paxscan a-Si flat panel detector (CsI screen, 1920 x 1536, 127 μm pixel size) at the Centre for X-ray Tomography of Ghent University (UGCT) (Masschaele *et al.* 2007). For each scan, 1,801 projections were captured with an exposure time of 0.8 s. The resulting scan images had a voxel pitch of 10 μm. Reconstruction was performed using the UGCT software package Octopus (Vlassenbroeck *et al.* 2007).

## 2.2 Three-dimensional reconstruction and meshing of a biventricular cardiac geometry

Only the two ventricles were considered for the FE model, resulting in a biventricular (BV) geometry truncated at the base. Before segmentation, the orientation of the image stack was aligned with the longitudinal cardiac axis. The image segmentation involved region-growing, level-set thresholding and manual segmentation (Simpleware ScanIP, Synopsys). Two masks were created, distinguishing the cardiac tissue and the injectate.

The resulting geometry captured the essential morphology of the left (LV) and right ventricle (RV) and the dispersed intramyocardial injectate in the LV free wall (Figure 1 a and Table 1). The geometry was meshed with 206,142 quadratic tetrahedral elements (injectate: 58,902 elements, myocardium: 147,240 elements). The mesh density varied between 302.8 $mm^{-3}$ in the myocardium and 3,852.3 $mm^{-3}$ in the injectate region (Figure 1 c).

The meshed geometry was imported in Abaqus 6.14-3 CAE (Dassault Systèmes, Providence, RI, USA). After that, the infarct was approximated in the antero-apical region of the LV and defined as a node-set.

Figure 1. a) Biventricular cardiac geometry of the rat heart developed from µCT image data with LV, RV, and dispersed intramyocardial injectate in the infarcted region of the LV. b) Infarct region shown with nodes set, the infarct was defined around the injectate. c) Increase in mesh density from the myocardium to the injectate to accommodate the microstructural dispersion of the injected material. d) Nodes on base with boundary condition of zero displacement to prevent rigid body motion (red points).

Table 1. Morphometric data of the biventricular geometry of the rat heart.

| Parameter | Value |
|---|---|
| Myocardial volume (mm$^3$) | 1,451 |
| LV wall thickness (mm) | 4.69 - 4.93 |
| Septal wall thickness (mm) | 2.53 - 4.54 |
| RV free wall thickness (mm) | 1.74 - 3.24 |
| Apex-base length (mm) | 13.94 |

## 2.3 Finite element model development

The meshed BV geometry was imported in Abaqus CAE (Abaqus 6.14-3, Dassault Systèmes, Providence, RI, USA) to implement the different mechanical properties and investigate cardiac tissue mechanics. A 10-node tetrahedral element type (C3D10M) was used to obtain a discretized FE problem.

### 2.3.1 Myofibre structure

The rule-based approach was used to describe the myofibre angle distribution in the myocardium. A Matlab code developed by Sack *et al.* (2018) was used to generate fibre orientation values through the myocardium. The primary function of the algorithm was to find the different components of the fibre vector within each element in the meshed region of the myocardium. The projected values of the fibre vector were obtained with two principal angles, including (i) the helix angle $\alpha_h$, formed by the projection of the fibre on the plane created by the circumferential-longitudinal unit vector ($u_c$, $u_l$) and the circumferential unit vector $u_c$ and (ii) the transverse angle $\alpha_t$ between the projection of the fibre on the plane formed by the radial-circumferential unit vector ($f_{p1}$) and the circumferential unit vector. In the current study, the fibre angle was -50° to 80° from the epicardial to the endocardial surfaces (Chen *et al.* 2003, Sirry 2015). The obtained fibre orientation data were incorporated in the FE model in Abaqus.

### 2.3.2 Constitutive laws

**Healthy and infarcted myocardium**

The healthy myocardium during the passive filling was treated as a hyperelastic anisotropic material using a strain energy density function adopted from Holzapfel and Ogden (2009):

$$W = \frac{a}{2b}e^{b(I_1-3)} + \sum_{i=f,s}\frac{a_i}{2b_i}\left\{e^{b_i(I_{4i}-1)^2} - 1\right\} + \frac{a_{fs}}{2b_{fs}}\left\{e^{b_{fs}(I_{8fs})^2} - 1\right\} + \frac{1}{D}\left(\frac{J^2-1}{2} - \ln(J)\right) \quad (1)$$

The passive mechanical properties of the infarcted myocardium depend on the stage of the infarct (Holmes *et al.* 2005, Holmes *et al.* 1997, Tyberg *et al.* 1974, Villarreal *et al.* 1991). A one-week infarct stage was considered in the current study. Hence, increased stiffness (Hood *et al.* 1970) in fibre, circumferential and longitudinal direction was implemented by a parameter adjustment as shown in Eqn **Error! Reference source not found.**). The passive mechanical behaviour of infarcted myocardium was described with a similar formulation based on a validated constitutive model developed by Sack *et al.* (2018):

$$W = \frac{\bar{a}}{2b}e^{b(I_1-3)} + \sum_{i=f,s}\frac{\bar{a}_i}{2b_i}\left\{e^{b_i(I_{4i}-1)^2} - 1\right\} + \frac{\bar{a}_{fs}}{2b_{fs}}\left\{e^{b_{fs}(I_{8fs})^2} - 1\right\} \tag{2}$$

with

$$
\begin{aligned}
\bar{a} &= a[h + (1-h)p], \\
\bar{a}_i &= a_i[h + (1-h)p], \text{ and} \\
\bar{a}_{fs} &= a_{fs}[h + (1-h)p]
\end{aligned}
\tag{3}
$$

The active contraction of the myocardium was implemented with a time-varying elastance approach with an additional term to consider the passive response during the contraction (Guccione and Mcculloch 1993, Guccione *et al.* 1993, Walker *et al.* 2005). The merit of this model resides in the fact that it predicts the active tension developed by the myocardium by including the deactivation of calcium channels, the time of onset of contraction and the sarcomere length history (Guccione *et al.* 1993).

The mathematical formulation of the active tension model is (Guccione *et al.* 1993, Sack *et al.* 2018, Walker *et al.* 2005):

$$T_a(t, E_{ff}) = \frac{T_{max}}{2}\frac{Ca_0^2}{Ca_0^2 + EC_{a_{50}}^2(E_{ff})}\left(1 - \cos\left(\omega(t, E_{ff})\right)\right) \tag{4}$$

with

$$ECa_{50}^2(E_{ff}) = \frac{Ca_{0\,max}}{\sqrt{e^{B(l(E_{ff})-l_0)} - 1}}, \tag{5}$$

and

$$
\begin{aligned}
\omega(t, E_{ff}) &= \pi\frac{t}{t_0}, && \text{for } 0 \le t \le t_0 \\
\omega(t, E_{ff}) &= \pi\frac{t - t_0 + t_r(l)}{t_r(l)}, && \text{for } t_0 \le t \le t_r \\
\omega(t, E_{ff}) &= 0, && \text{for } t \ge t_0 + t_0
\end{aligned}
\tag{6}
$$

The description and values of the passive constitutive parameters are provided in Supplemental Tables 1 and 2.

**Injectable biomaterial**

The injectable biomaterial, e.g. polyethylene glycol (PEG) hydrogel, was described as hyperelastic isotropic incompressible material with a Neo-Hookean material model:

$$W = C_{10}(I_1 - 3) \tag{7}$$

where $I_1$ is the first deviatoric strain invariant, and $C_{10}$ characterizes the material stiffness obtained from the elastic modulus $E_{inj}$ and the Poisson's ratio $v_{inj}$:

$$C_{10} = \frac{E_{inj}}{4(1 + v_{inj})} \tag{8}$$

with $v_{inj} = 0.5$ to represent incompressibility and $E_{inj}$ between 4.1 and 405,900 kPa (see section 2.4 Finite element simulations for further details).

### 2.3.3 Boundary conditions

A zero displacement was applied to the nodes at the base in the longitudinal direction to prevent the rigid body motion of the geometry, see Figure 1(d). The passive filing was implemented by a linearly increasing pressure load on both LV and RV cavity surfaces. Several studies reported a higher pressure in the LV cavity (LV cavity pressure is four times the RV cavity pressure). The measured systolic pressure is 30-40 mmHg and 100-140 mmHg in normal human hearts, respectively, in the RV and LV cavity (Mininni *et al.* 1996, Reichek and Devereux 1982). Pacher *et al.* (2004) measured left ventricle end-diastolic and left ventricular end-systolic pressure in rats and found $3.8 \pm 0.9$ mmHg and $133.8 \pm 8.1$ mmHg, respectively. In the current study, the cavity pressure was taken from 0 to 3.0 mmHg and 0 to 0.75 mmHg, respectively, for LV and RV cavities. This choice agrees with the range of the experimental end-diastolic pressure findings.

### 2.3.4 Determination of computation time for end-systole and end-diastole

The end-diastolic time point was determined by applying a linearly increasing pressure on the LV and RV endocardial surfaces until the LV cavity volume matched a target experimental ED volume by Pacher *et al.* (2004). This step was performed on the BV model with injectate elastic modulus $E_{inj} = 73.8$ kPa.

A second simulation was performed to determine the end-systolic time point. The obtained time corresponded to the contraction and was determined from the ED time point defined previously until the active tension declined. The LV volume was calculated at this time point and compared with end-systolic volume values in the literature (Pacher *et al.* 2004).

## 2.4 Finite element simulations and data acquisition

A parametric study to determine the impact of the injectate stiffness on the deformation of the myocardium and the biomaterial injectate was conducted with a range of values for the elastic modulus of injectate, i.e. $E_{inj}$ = 4.1, 7.4, 40.6, 73.8, 405.9, 738, 4,059, 7,380, 40,590, 73,800 and 405,900 kPa.

The strain in the myocardium and the biomaterial injectate was recorded at the end-diastolic and end-systolic time point of the cardiac cycle for each $E_{inj}$ value. The myofibre and cross-fibre strain were used for the healthy and infarcted myocardium, whereas the maximum principal and minimum principal strain were used for the injectate. For each finite element in the mesh, the element strain, $\varepsilon_{EL}$, was determined as the arithmetic mean of the strain values at the four integration points, $\varepsilon_{IP,i}$, of the 10-node tetrahedral elements.

## 2.5 Statistical analysis

Descriptive statistical analysis was performed on the strain data to determine the normality (Shapiro-Wilk normality test) and variability (SciPy, https://scipy.org/ and NumPy, https://numpy.org/). Data were presented using box and whisker plots (MatPlotLib, Python, https://matplotlib.org/) indicating median and interquartile ranges.

## 3. Results

Representative end-diastolic and end-systolic strain distributions are shown for the myofibre and cross-fibre strain in the BV wall in Figure 2 and the maximum and minimum principal strain in the injectate in Figure 3, for an elastic modulus of the injectate of $E_{inj}$ = 73.8 kPa.

Figure 2. Short-axis and longitudinal contour plots showing myofibre strain (a, b) and cross-fibre strain (c, d) in the BV model at end-diastolic (ED, left column) and end-systolic time point (ES, right column) for an injectate elastic modulus $E_{inj}$ = 73.8 kPa. The strain distribution in the injectate (black areas) is shown in Figure 3.

Figure 3. Short-axis and longitudinal contour plots of maximum (a, b) and minimum principal strain (c, d) in the injectate at end-diastolic (ED, left column) and end-systolic time point (ES, right column) for an injectate elastic modulus $E_{inj}$ = 73.8 kPa.

## 3.1 Effect of injectate stiffness on myocardial deformation

The median end-diastolic myofibre and cross-fibre strain decreased from 3.5% to 1.0% and -5.9% to -2.7% (decrease in magnitude) for the increase in the injectate elastic modulus (Figure 4 a and b). These changes in strain appear to be more pronounced in the low modulus region $E_{inj}$ = 4.1 to 738 kPa and only marginal for $E_{inj}$ > 738 kPa.

At end-systole, the median myofibre strain decreased in magnitude from -20.4% to -11.8%, with increasing elastic modulus. The median cross-fibre strain decreased from 6.5% to 4.6% with an increasing elastic modulus with an intermittent marginal increase for $E_{inj}$ = 405.9 kPa and 738.0 kPa (Figure 4 c and d).

Figure 4. Myocardial deformation. End-diastolic (ED) myofibre (a) and cross-fibre (b) strain and end-systolic (ES) myofibre (c) and cross-fibre (d) strain versus injectate elastic modulus. The box and whiskers indicate the median (red line in box), interquartile range (IQR) from first and third quartile (lower and upper bound of the box), 1.5x IQR (lower and upper whisker), and data outside the 1.5x IQR ranges (open circles).

## 3.2 Effect of injectate stiffness on injectate deformation

As a general observation, the magnitude (i.e. median), range (difference between the highest and lowest element strain) and interquartile range of maximum and minimum principal strain decreased considerably with increasing injectate stiffness for lower $E_{inj}$ = 4.1 - 738 kPa. However, it only

decreased marginally for the higher $E_{inj}$ = 4,059 to 405,900 kPa, both at end-diastolic and end-systolic time points.

At end-diastole, the median maximum principal strain decreased from 5.4% to 0.001%, and the median minimum principal strain decreased in magnitude from -5.4% to -0.001% for increasing injectate stiffness (Figure 5 a and b).

At end-systole, the median maximum principal strain decreased from 38.5% to 0.06%, and the median, minimum principal strain decreased in magnitude from -39.0% to -0.06% with increasing injectate stiffness (Figure 5 c and d).

Figure 5. Injectate deformation. End-diastolic (ED) maximum (a) and minimum (b) principal strain and end-systolic (ES) maximum (c) and minimum (d) principal strain. The box and whiskers indicate the median (red line in box), interquartile range (IQR) from first and third quartile (lower and upper bound of the box), 1.5x IQR (lower and upper whisker), and data outside the 1.5x IQR ranges (open circles).

## 4. Discussion

The present computational study involved the development of computational models for cardiac mechanics. The effect of the therapeutic biomaterial injectate stiffness on the cardiac mechanics during a cardiac cycle was shown using the developed models. At a given injectate elastic modulus, the end-diastolic myofibre strain decreased from the endocardial to the epicardial surface, as also

reported in other studies (Costa *et al.* 1999, Takayama *et al.* 2002, Waldman *et al.* 1988). The biomaterial delivery 7 days after infarct induction was chosen since significant improvements in cardiac function and scar tissue mechanics were reported *in vivo* for the injection of PEG hydrogel one week following the infarction compared to biomaterial delivery immediately after the infarction (Kadner *et al.* 2012).

Many computational studies involving animal species used simplified cardiac geometries (Dieudonné 1969, Wall *et al.* 2006, Wenk *et al.* 2009). The current study was the first to develop a subject-specific biventricular geometry of a rat heart with *in situ* microstructural details of an intramyocardial injectate in the LV free wall. The morphological details at the micrometre length scale resulted in a considerably higher and heterogeneous mesh density. The reconstruction of the biventricular FE model used a validated method developed by Sack *et al.* (2018); hence, a mesh sensitivity study was not necessary.

Different strain energy density functions were used for the discrete representation of myocardial and injectate regions, contrary to studies that used the homogenization approach that combines the myocardium and biomaterial injectate (Wall *et al.* 2006, Wenk *et al.* 2011). This approach allowed capturing the distinctly different mechanical properties of the myocardium and injected biomaterial.

The current study adopted passive myocardial material parameters from Dương *et al.* (2018), who developed and validated the constitutive model for infarcted rat myocardium with experimental data from biaxial tension and uniaxial compression (Sirry *et al.* 2016). The active contraction in the healthy myocardial regions was modelled with a validated time-varying elastance model (Guccione *et al.* 1993). The parameters estimated by Guccione *et al.* (1993) have been used in several computational studies involving rat (Wise *et al.* 2016) and porcine models (Sack *et al.* 2018).

Therapeutic intramyocardial biomaterial injectates with different mechanical properties have been reported to induce different mechanical and functional responses in the infarcted heart (Wall *et al.* 2006).

The current study showed that an increasing elastic modulus (i.e. stiffness) of the biomaterial injectate reduces the ES and ED myocardial strains in the myofibre and cross-fibre directions in the healthy and infarcted regions of the BV geometry.

In several studies that investigated the impact of injectates on the infarcted heart, a reduction in the wall stress from non-treated to treated infarct was observed and identified as a critical factor for ventricular biomechanical and function improvement (Kichula *et al.* 2014, Kortsmit *et al.* 2013a,

Miller *et al.* 2013, Sirry 2015). Additionally, increasing the injectate stiffness further reduced ventricular wall stress (Kichula *et al.* 2014, Sirry 2015).

In the current study investigating the strain in the entire BV geometry, the increase of the injectate elastic modulus led to reduced myofibre and cross-fibre strain in the healthy region of the BV geometry. This finding suggests that an increasing elastic modulus of the injectate for MI treatment may result in a beneficial effect of reducing the myocardial strains in the remote healthy regions of the BV geometry.

The biomaterial injectate is often used as a scaffold for therapeutically delivered cells for MI treatment. Thus, investigating the mechanical response of the biomaterial injectate delivered to the infarct is crucial for embedded cells' mechanics and associated signalling.

The current study showed that the mechanical response of the injectate depends on its stiffness. The maximum and minimum principal strains decreased for an increasing injectate elastic modulus. The upper range of the injectate's elastic modulus (i.e. $E_{inj} > 738$ kPa) used in the study is unrealistic for PEG hydrogels and other injectable biomaterials (Rizzi *et al.* 2006, Singelyn *et al.* 2012). However, the current investigations aimed to explore the injectate response for an extensive range of injectate stiffness, and it was shown that the injectate deformations were negligible for these unrealistically high values of the injectate elastic modulus.

Limitations of the current study included a lack of advanced interaction definitions at the interfaces between the myocardium and injectate. The continuous mesh used for the myocardium and injectate resulted in fixed contact properties at the domain interfaces. More realistic contact interactions will improve the simulation of the interface mechanics but require experimental data that currently do not exist.

## 5. Conclusions

This is the first computational study that generated and used a high-resolution microstructurally geometry of an *in situ* biomaterial injectate delivered one week after infarct induction in a biventricular cardiac finite element model of a rat heart with antero-apical infarct in the left ventricular free wall. The model's feasibility was demonstrated by assessing the effect of injectate stiffness on myocardial and injectate mechanics. With the microstructural injectate geometry, the computational model offers vast potential for extension to cellular injectates and *in silico* studies of mechanotransduction and therapeutic signalling in the infarcted heart with a myocardial infarction animal model extensively used in preclinical research.

# Funding

This work was supported by financially supported by the National Research Foundation of South Africa (IFR14011761118 to TF), the South African Medical Research Council (SIR328148 to TF), and the CSIR Centre for High Performance Computing (CHPC Flagship Project Grant IRMA9543 to TF), and the Dr. Leopold und Carmen Ellinger Stiftung (UCT Three-Way PhD Global Partnership Programme Grant DAD937134 to TF). The funders had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript. Any opinion, findings, conclusions, and recommendations expressed in this publication are those of the authors, and therefore, the funders do not accept any liability.

# Conflict of Interests

Conflicts of interest do not exist.

# Data availability

Data supporting the results presented in this article are available on the University of Cape Town's institutional data repository (ZivaHub) under http://doi.org/10.25375/uct.19630203 as YD Motchon, KL Sack, MS Sirry, E Pauwels, D Van Loo, A De Muynck, L Van Hoorebeke, NH Davies, T Franz. Effect of biomaterial stiffness on cardiac mechanics in a biventricular infarcted rat heart model with microstructural representation of in situ intramyocardial injectate. Cape Town, ZivaHub, 2022, DOI: 10.25375/uct.19630203.

# References

Cai L, Wang Y, Gao H, Li Y, Luo X. *A mathematical model for active contraction in healthy and failing myocytes and left ventricles*. PLoS One 2017, **12**(4): e0174834.

Chen J, Song S-K, Liu W, Mclean M, Allen JS, Tan J, Wickline SA, Yu X. *Remodeling of cardiac fiber structure after infarction in rats quantified with diffusion tensor mri*. American Journal of Physiology-Heart and Circulatory Physiology 2003, **285**(3): H946-H54.

Christman KL, Fok HH, Sievers RE, Fang Q, Lee RJ. *Fibrin glue alone and skeletal myoblasts in a fibrin scaffold preserve cardiac function after myocardial infarction*. Tissue Eng 2004, **10**(3-4): 403-9.

Costa KD, Takayama Y, Mcculloch AD, Covell JW. *Laminar fiber architecture and three-dimensional systolic mechanics in canine ventricular myocardium*. Am J Physiol 1999, **276**(2): H595-607.

Dieudonné J-M. *The left ventricle as confocal prolate spheroids*. bulletin of mathematical biophysics 1969, **31**(3): 433-9.

Dương MTN, Ach T, Alkassar M, Dittrich S, Leyendecker S. *Numerical simulation of cardiac muscles in a rat biventricular model*. 6th European Conference on Computational Mechanics (ECCM 6), 7th European Conference on Computational Fluid Dynamics (ECFD 7). UK, 2018.

Fan C, Shi J, Zhuang Y, Zhang L, Huang L, Yang W, Chen B, Chen Y, Xiao Z, Shen H, Zhao Y, Dai J. *Myocardial-infarction-responsive smart hydrogels targeting matrix metalloproteinase for on-demand growth factor delivery*. Adv Mater 2019, **31**(40): e1902900.

Finegold JA, Asaria P, Francis DP. *Mortality from ischaemic heart disease by country, region, and age: Statistics from World Health Organisation and United Nations*. Int J Cardiol 2013, **168**(2): 934-45.

Guccione J, Mcculloch A. *Mechanics of active contraction in cardiac muscle: Part i—constitutive relations for fiber stress that describe deactivation*. Journal of Biomechanical Engineering 1993, **115**(1): 72-81.

Guccione J, Waldman L, Mcculloch A. *Mechanics of active contraction in cardiac muscle: Part ii—cylindrical models of the systolic left ventricle*. Journal of Biomechanical Engineering 1993, **115**(1): 82-90.

Holmes JW, Borg TK, Covell JW. *Structure and mechanics of healing myocardial infarcts*. Annu Rev Biomed Eng 2005, **7**: 223-53.

Holmes JW, Nunez JA, Covell JW. *Functional implications of myocardial scar structure*. Am J Physiol 1997, **272**(5 Pt 2): H2123-30.

Holzapfel GA, Ogden RW. *Constitutive modelling of passive myocardium: A structurally based framework for material characterization*. Philos Trans A Math Phys Eng Sci 2009, **367**(1902): 3445-75.

Hood WB, Bianco JA, Kumar R, Whiting RB. *Experimental myocardial infarction: Iv. Reduction of left ventricular compliance in the healing phase*. The Journal of Clinical Investigation 1970, **49**(7): 1316-23 % @ 0021-9738.

Johnson TD, Christman KL. *Injectable hydrogel therapies and their delivery strategies for treating myocardial infarction*. Expert opinion on drug delivery 2013, **10**(1): 59-72.

Kadner K, Dobner S, Franz T, Bezuidenhout D, Sirry MS, Zilla P, Davies NH. *The beneficial effects of deferred delivery on the efficiency of hydrogel therapy post myocardial infarction*. Biomaterials 2012, **33**(7): 2060-6.

Kichula ET, Wang H, Dorsey SM, Szczesny SE, Elliott DM, Burdick JA, Wenk JF. *Experimental and computational investigation of altered mechanical properties in myocardium after hydrogel injection*. Ann Biomed Eng 2014, **42**(7): 1546-56.

Kortsmit J, Davies NH, Miller R, Macadangdang JR, Zilla P, Franz T. *The effect of hydrogel injection on cardiac function and myocardial mechanics in a computational post-infarction model*. Comput Methods Biomech Biomed Engin 2013a, **16**(11): 1185-95.

Kortsmit J, Davies NH, Miller R, Zilla P, Franz T. *Computational predictions of improved of wall mechanics and function of the infarcted left ventricle at early and late remodelling stages: Comparison of layered and bulk hydrogel injectates*. Advances in Biomechanics and Applications 2013b, **1**(1): 41-55.

Masschaele BC, Cnudde V, Dierick M, Jacobs P, Van Hoorebeke L, Vlassenbroeck J. *Ugct: New x-ray radiography and tomography facility*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment 2007, **580**(1): 266-9.

Mendis S, Puska P, Norrving B. Global atlas on cardiovascular disease prevention and control, World Health Organization, 2011.

Miller R, Davies NH, Kortsmit J, Zilla P, Franz T. *Outcomes of myocardial infarction hydrogel injection therapy in the human left ventricle dependent on injectate distribution*. Int J Numer Method Biomed Eng 2013, **29**(8): 870-84.

Mininni S, Diricatti G, Vono MC, Giglioli C, Margheri M, Olivo G, Gensini G, Galanti G. *Noninvasive evaluation of right ventricle systolic pressure during dynamic exercise by saline-enhanced doppler echocardiography in progressive systemic sclerosis*. Angiology 1996, **47**(5): 467-74.

Opie LH, Mayosi BM. *Cardiovascular disease in sub-saharan africa*. Circulation 2005, **112**(23): 3536-40.

Pacher P, Mabley JG, Liaudet L, Evgenov OV, Marton A, Hasko G, Kollai M, Szabo C. *Left ventricular pressure-volume relationship in a rat model of advanced aging-associated heart failure*. Am J Physiol Heart Circ Physiol 2004, **287**(5): H2132-7.

Reichek N, Devereux RB. *Reliable estimation of peak left ventricular systolic pressure by m-mode echographic-determined end-diastolic relative wall thickness: Identification of severe valvular aortic stenosis in adult patients*. Am Heart J 1982, **103**(2): 202-3.

Rizzi SC, Ehrbar M, Halstenberg S, Raeber GP, Schmoekel HG, Hagenmüller H, Müller R, Weber FE, Hubbell JA. *Recombinant protein-co-peg networks as cell-adhesive and proteolytically degradable hydrogel matrixes. Part ii: Biofunctional characteristics*. Biomacromolecules 2006, **7**(11): 3019-29.

Sack KL, Aliotta E, Ennis DB, Choy JS, Kassab GS, Guccione JM, Franz T. *Construction and validation of subject-specific biventricular finite-element models of healthy and failing swine hearts from high-resolution dt-mri*. Front Physiol 2018, **9**(539): 539.

Sepantafar M, Maheronnaghsh R, Mohammadi H, Rajabi-Zeleti S, Annabi N, Aghdami N, Baharvand H. *Stem cells and injectable hydrogels: Synergistic therapeutics in myocardial repair*. Biotechnol Adv 2016, **34**(4): 362-79.

Silveira-Filho LM, Coyan GN, Adamo A, Luketich SK, Menallo G, D'amore A, Wagner WR. *Can a biohybrid patch salvage ventricular function at a late time point in the post-infarction remodeling process?* Basic to Translational Science 2021, **6**(5): 447-63.

Singelyn JM, Sundaramurthy P, Johnson TD, Schup-Magoffin PJ, Hu DP, Faulk DM, Wang J, Mayle KM, Bartels K, Salvatore M, Kinsey AM, Demaria AN, Dib N, Christman KL. *Catheter-deliverable hydrogel derived from decellularized ventricular extracellular matrix increases endogenous cardiomyocytes and preserves cardiac function post-myocardial infarction*. J Am Coll Cardiol 2012, **59**(8): 751-63.

Sirry MS. *Computational biomechanics of acute myocardial infarction and its treatments*. Human Biology, Univerity of Cape Town, 2015, **PhD**: 145.

Sirry MS, Butler JR, Patnaik SS, Brazile B, Bertucci R, Claude A, Mclaughlin R, Davies NH, Liao J, Franz T. *Characterisation of the mechanical properties of infarcted myocardium in the rat under biaxial tension and uniaxial compression*. J Mech Behav Biomed Mater 2016, **63**: 252-64.

Takayama Y, Costa KD, Covell JW. *Contribution of laminar myofiber architecture to load-dependent changes in mechanics of lv myocardium*. Am J Physiol Heart Circ Physiol 2002, **282**(4): H1510-20.

Tyberg JV, Forrester JS, Wyatt HL, Goldner SJ, Parmley WW, Swan HJ. *An analysis of segmental ischemic dysfunction utilizing the pressure-length loop*. Circulation 1974, **49**(4): 748-54.

Villarreal FJ, Lew WY, Waldman LK, Covell JW. *Transmural myocardial deformation in the ischemic canine left ventricle*. Circ Res 1991, **68**(2): 368-81.

Vlassenbroeck J, Dierick M, Masschaele B, Cnudde V, Van Hoorebeke L, Jacobs P. *Software tools for quantification of x-ray microtomography at the ugct*. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment 2007, **580**(1): 442-5.

Waldman LK, Nosan D, Villarreal F, Covell JW. *Relation between transmural deformation and local myofiber direction in canine left ventricle*. Circulation research 1988, **63**(3): 550-62.

Walker JC, Ratcliffe MB, Zhang P, Wallace AW, Fata B, Hsu EW, Saloner D, Guccione JM. *Mri-based finite-element analysis of left ventricular aneurysm*. Am J Physiol Heart Circ Physiol 2005, **289**(2): H692-700.

Wall ST, Walker JC, Healy KE, Ratcliffe MB, Guccione JM. *Theoretical impact of the injection of material into the myocardium: A finite element model simulation*. Circulation 2006, **114**(24): 2627-35.

Wang H, Rodell CB, Lee ME, Dusaj NN, Gorman JH, 3rd, Burdick JA, Gorman RC, Wenk JF. *Computational sensitivity investigation of hydrogel injection characteristics for myocardial support*. J Biomech 2017, **64**: 231-5.

Wenk JF, Eslami P, Zhang Z, Xu C, Kuhl E, Gorman Iii JH, Robb JD, Ratcliffe MB, Gorman RC, Guccione JM. *A novel method for quantifying the in-vivo mechanical effect of material injected into a myocardial infarction*. The Annals of Thoracic Surgery 2011, **92**(3): 935-41.

Wenk JF, Wall ST, Peterson RC, Helgerson SL, Sabbah HN, Burger M, Stander N, Ratcliffe MB, Guccione JM. *A method for automatically optimizing medical devices for treating heart failure: Designing polymeric injection patterns*. J Biomech Eng 2009, **131**(12): 121011.

Wise P, Davies NH, Sirry MS, Kortsmit J, Dubuis L, Chai CK, Baaijens FP, Franz T. *Excessive volume of hydrogel injectates may compromise the efficacy for the treatment of acute myocardial infarction*. International journal for numerical methods in biomedical engineering 2016, **32**(12): e02772.