Clinical evaluation of the Multimapping technique for simultaneous myocardial T_1 and T_2 mapping

Charlotta Jarkman[1], Carl-Johan Carlhäll[1,2,3] †, Markus Henningsson[1,2]†*

1. Department of Clinical Physiology in Linköping, Department of Health, Medicine and Caring Sciences, Linköping University, Linköping, Sweden
2. Division of Diagnostics and Specialist Medicine, Department of Health, Medicine and Caring Sciences (HMV), Linköping University, Linköping, Sweden
3. Center for Medical Image Science and Visualization (CMIV), Linköping University, Linköping, Sweden

* Correspondence:
Markus Henningsson
markus.henningsson@liu.se

† Equal contributors

Keywords: T_1 mapping, T_2 mapping, ECV, quantitative CMR, simultaneous multiparametric CMR

Abstract

The Multimapping technique was recently proposed for simultaneous myocardial T_1 and T_2 mapping. In this study we evaluate its correlation with clinical reference mapping techniques in patients with a range of cardiovascular diseases (CVD), compare image quality and inter- and intra-observer repeatability. Multimapping consists of a ECG-triggered, 2D single-shot bSSFP readout with inversion recovery and T_2 preparation modules, acquired across 10 cardiac cycles. The sequence was implemented at 1.5T and compared to clinical reference mapping techniques, Modified Look-Locker inversion recovery (MOLLI) and T_2 prepared bSSFP with four echo times (T_2bSSFP), and compared in 47 patients with CVD (of which 44 were analyzed). In diseased myocardial segments (defined as presence of late gadolinium enhancement) there was a high correlation between Multimapping and MOLLI for native myocardium T_1 (r^2=0.73), ECV (r^2=0.91) and blood T_1 (r^2=0.88), and Multimapping and T_2bSSFP for native myocardial T_2 (r^2=0.80). In healthy myocardial segments a bias for native T_1 (Multimapping=1116±21 ms, MOLLI=1002±21, P<0.001), post-contrast T_1 (Multimapping=479±31 ms, MOLLI=426±27 ms, 0.001), ECV (Multimapping=21.5±1.9%, MOLLI=23.7±2.3%, P=0.001) and native T_2 (Multimapping=48.0±3.0 ms, T_2bSSFP=53.9±3.5 ms, P<0.001) was observed. The image quality for Multimapping was scored as higher for all mapping techniques (native T_1, post-contrast T_1, ECV and T_2bSSFP) compared to the clinical reference techniques. The inter- and intra-observer agreement was excellent (intraclass correlation coefficient, ICC>0.9) for all measurements, except for inter-observer repeatability of Multimapping native T_1 (ICC=0.87), post-contrast T_1 (ICC=0.73) and T_2bSSFP native T_2 (ICC=0.88). Multimapping show high correlations with clinical reference mapping techniques for T_1, T_2 and ECV in a diverse cohort of patients with different cardiovascular diseases. Multimapping enables simultaneous T_1 and T_2 mapping and can be performed in a short breath-hold, with image quality superior to that of the clinical reference techniques.

1   Introduction

Myocardial T_1 and/or T_2 values are altered in many cardiovascular diseases (1). T_1 and T_2
quantification, along with disease-specific patterns of regional and global distribution, can be
captured with myocardial mapping techniques (2). In the last 15-20 years a number of T_1 and T_2
mapping techniques have been published, with different strengths and weaknesses in terms of
quantification accuracy, precision, scan time, spatial resolution and coverage (3). Despite being one
of the first T_1 mapping techniques, the modified Look-Locker inversion recovery (MOLLI) remains
the most clinically used method due to its high precision and availability on all major scanner
platforms (4,5). However, MOLLI T_1 accuracy is relatively low and the quantification is susceptible
to confounding effects from heart-rate or T_2-dependent variability, magnetization transfer effects,
motion artifacts and system imperfections (5,6). Different T_1 mapping methods have been proposed
to address these shortcomings, yet have failed to make significant inroads in the market share of
clinical use (7–11). Myocardial T_2 mapping can be performed with multi-echo spin-echo or T_2-
prepared balanced steady-state free precession (T_2bSSFP) techniques (12–14). The latter approach is
likely the most widely used clinically due to its relative robustness to physiological motion.

In recent years, there has been a growing interest in techniques to simultaneously map T_1 and T_2 in a
single scan (15–21). Advantages of this approach compared to conventional mapping, which is
performed separately for T_1 and T_2, is that the images are intrinsically spatially aligned, scan time is
typically shorter and the confounding effects of T_1 or T_2 on the quantification of the opposite
parameter is minimized. Despite the many theoretical advantages of simultaneous T_1 and T_2
mapping, there is a paucity of translational studies using these techniques in patients with
cardiovascular disease (22,23). This may be due to the more sophisticated acquisition, reconstruction
and mapping strategies necessary for such techniques, which pose challenges for clinical translation.
Recently, a new technique for simultaneous T_1 and T_2 mapping, termed Multimapping, was proposed
using a standard Cartesian trajectory and evaluated (primarily) in healthy subjects (24). Due to its
simplicity Multimapping may be readily applied in a clinical setting to enable evaluation of
simultaneous T_1 and T_2 mapping in patients with cardiovascular disease.

The primary aim of this study is to validate Multimapping T_1 and T_2 values against clinical reference
techniques in patients with different cardiovascular diseases in terms of parameter quantification and
image quality. A secondary study aim is to evaluate Multimapping intra- and inter-observer
variability.

2   Materials and Methods

2.1   Study population

All patients provided written informed consent prior to participation and the study was approved by
the local ethics committee (Linköping Regional Ethics Committee, 2015/396–31) and conducted
according to the Declaration of Helsinki. Patients referred for CMR at Linköping University Hospital
between June and November 2021 were considered for inclusion in this study. In total, 47 patients
were recruited. Datasets from three patients were excluded, two because no late gadolinium
enhancement (LGE) images were acquired and one due to excessive fold-over artifacts. Clinical
characteristics of the remaining patients can be seen in Table 1. Of the included patients, normal
cardiac MRI scan was found in 15 (34.1%) patients, myocarditis in 11 (25%) patients, dilated
cardiomyopathy (DCM) in 6 (13.6%) patients, ventricular hypertrophy (hypertrophic
cardiomyopathy or hypertrophy of unknown origin) in 5 (11.4%) patients, ischemic myocardial
injury (acute/recent or old) in 3 (6.8%) patients, arrhythmogenic right ventricular cardiomyopathy in
2 (4.5%) patients, pericarditis in 1 (2.3%) patient and congenital heart disease in 1 (2.3%) patient.

2.2   Data acquisition and reconstruction

All scans were performed on a 1.5T Philips clinical CMR scanner (Philips Healthcare, Best, The
Netherlands) using a 28-channel torso coil. The Multimapping pulse sequence and post-processing
steps are illustrated in Figure 1. Ten single-shot images are acquired across consecutive cardiac
cycles using balanced-steady state free precession (bSSFP) readouts, triggered to the mid-diastolic
rest period. Adiabatic inversion radiofrequency (RF) pulses with an assumed inversion efficiency of
0.94 and delay times of 300 ms are performed in the 1^{st} and 5^{th} cardiac cycle to improve T_1
sensitization. T_2 preparation modules with hard 90º RF pulses and four adiabatic refocusing RF
pulses are performed in the 8^{th}, 9^{th}, and 10^{th} cardiac cycles to improve T_2 sensitization using echo
times of 30, 50 and 70 ms, respectively. The Multimapping imaging parameters for all experiments
are: field-of-view = 320×320 mm, spatial resolution = 2×2 mm, slice thickness = 10 mm, nominal
flip angle = 50°, bandwidth = 1076 Hz/pixel, TR = 2.3 ms, TE = 1.2 ms, SENSE factor = 2, linear
profile order. Ten startup RF pulses are used with linearly increasing flip angles. The Multimapping
scan was acquired in a mid-ventricular short axis slice (except in one patient which was mistakenly
acquired in a four-chamber view) during a breath-hold. Native Multimapping was acquired in all 47
patients, and post-contrast Multimapping was performed in 31 patients approximately 15 to 20
minutes after contrast agent administration (0.2 mmol/kg gadobutrol). Due to clinical prioritization,
the post-contrast Multimapping was performed after the acquisition of post-contrast MOLLI and
LGE.

All Multimapping source images were reconstructed on the scanner and transferred to an offline
workstation (Intel Core i7-8565U 1.80 GHz processor with 16Gb RAM) to generate T_1 and T_2 maps
using MATLAB R2021b (The MathWorks, Natick, MA). The MATLAB code used to generate the
maps, including example Multimapping source images from one subject, is available at
https://github.com/Multimapping/Matlab_files. Since blood samples were not available for all
patients, Multimapping synthetic ECV maps were generated using synthetic hematocrit values, based
on the native MOLLI left ventricular blood pool measurements, as previously outlined (25). Image
registration using a rigid body transformation was applied to spatially align the native and post-
contrast T_1 maps prior to ECV calculation.

MOLLI was acquired in all 47 patients and T_2bSSFP was acquired in 45 patients, in the same slice as
Multimapping and used as clinical reference techniques for T_1 and T_2 mapping, respectively. All
imaging parameters for the reference techniques (field-of-view, spatial resolution etc.) were the same
as for Multimapping, except for the flip angle which was 35º. MOLLI was acquired with the 5(3s)3
scheme and used the same adiabatic inversion pulse as Multimapping. T_2bSSFP was acquired with
four images at different T_2 preparation echo times (0, 23, 46 and 70 ms) and used 3 pause cardiac
cycles between each image. Furthermore, T_2bSSFP used the same RF pulse types for the T_2
preparation module as Multimapping. The reference maps were reconstructed on the scanner using
vendor-provided inline mapping algorithms, except for the ECV maps which were generated offline
using MATLAB. Similar to Multimapping, synthetic ECV maps were generated using a synthetic
hematocrit value derived from the left ventricular blood pool T_1 measured in the native MOLLI
image. As for Multimapping, image registration was applied to native and post-contrast MOLLI T_1
maps before ECV was calculated. LGE imaging parameters were TR/TE = 5.6/2.0 ms, flip angle =
25º, FOV = 350×350 mm^2, spatial resolution = 1.8×1.8 mm^2.

2.3 Image Analysis

T_1 (native and post-contrast), T_2 (native) and ECV measurements were made by drawing manual
regions-of-interest (ROI) in all datasets. To compare Multimapping to the clinical reference
techniques, two sets of myocardial measurements were performed, one targeting any diseased
myocardium and one targeting healthy myocardium. For the measurements of diseased myocardium
in all maps, ROIs were drawn in the area corresponding with the most prominent positive LGE
findings of each patient. Since only a subset of patients had positive LGE findings in the imaged
slice, this resulted in 21 measurements for native T_1 and T_2, and 12 measurements for post-contrast
T_1 and ECV. For the measurements of healthy myocardium in all maps, ROIs were drawn in the area
remote of any LGE abnormality and preferentially in the interventricular septum if it was free of
abnormal LGE. Patients were excluded from this analysis if there were indications suggestive of
global or diffuse myocardial disease. The measurements in the healthy myocardium were performed
in a total of 19 patients for native T_1, 12 patients for T_1 post-contrast and ECV, and 18 patients for
T_2.

Measurements were performed in all patients by one observer (CJ, 1 year of CMR experience). To
allow intra-observer variability analysis, measurements were repeated in 23 patients by the same
observer two weeks later. For inter-observer variability analysis, the same 23 patients were measured
by two additional observers (MH and CJC with 14 and 21 years of CMR experience, respectively).
Furthermore, to compare blood T_1 (native and post-contrast), ROIs were drawn in the left ventricular
blood pool (avoiding any papillary muscles) in the Multimapping T_1 and MOLLI images.

The image quality of the acquired maps were qualitatively compared using a Likert scale as devised
by Jaubert et al. (22) with the following categories: 1 = uninterpretable, 2 = poor definition of edges,
significant noise and/or residual artifacts, 3 = mildly blurred edges, mild noise and/or residual
artifacts, 4 = slightly blurred edges, minor residual artifacts, 5 = negligible blurring or residual
artifacts. Visual scoring was performed for T_1 (native and post-contrast), T_2 (native) and ECV
separately using the different mapping techniques and this analysis was performed in 20 patients. The
visual scoring was performed by consensus of two blinded observers (CJ and CJC).

2.4   Statistical Analysis

Continuous variables are expressed as mean ± SD. Categorical variables are expressed as counts and
percentages. A P-value of < 0.05 was considered statistically significant. For the remote
measurements, two-tailed Student paired t-tests were performed to compare Multimapping to MOLLI
for native T_1, post-contrast T_1 and ECV, and Multimapping and T_2bSSFP for native T_2. For the
remote measurements, all parameters tested positive for normality using a Shapiro-Wilk test. Bland-
Altman and correlation plots were used to evaluate the agreement and correlation, respectively,
between Multimapping and the reference techniques of the measurements in diseased myocardium
for native T_1, post-contrast T_1, ECV and native T_2. To investigate any heart-rate dependency for the
mapping techniques, the measurements of the remote myocardium was correlated with the heart-rate
at the time of the scan. Similarly, dependency on T_2 for T_1 (and vice-versa) was evaluated by
correlating remote T_1 with T_2 for both Multimapping and the reference techniques, and testing for
statistical significance ($P < 0.05$). Intra-observer repeatability and inter-observer repeatability was
assessed with intraclass correlation coefficient (ICC) analysis. ICC was calculated using absolute
agreement two-way mixed model. Statistical analysis was performed using IBM SPSS Statistics,
version 27.0.

3   Results

Clinical evaluation of simultaneous myocardial T_1 and T_2 mapping

Representative parameter maps acquired with Multimapping, reference techniques and LGE in a
patient with no cardiovascular disease findings are shown in Figure 2. Parameter maps from a patient
with myocarditis are shown in Figure 3, with prominently altered quantitative values seen in both
Multimapping and reference techniques. The final example, in Figure 4, shows parameter maps from
a patient with myocardial infarction with a clearly delineated area of infarction in the Multimaps,
correlating with LGE. Multimaps for all patients can be downloaded from
https://github.com/Multimapping/Patient_study/raw/main/MapReconstructions.pdf.

3.1   Comparison of Multimapping and reference techniques in remote myocardium

The Multimapping native T_1 was 1116±21 ms and for MOLLI 1002±21 ms, resulting in a statistically
significant bias of 114 ms ($P < 0.001$). Multimapping post-contrast T1 was 479±31 ms and for
MOLLI 426±27 ms, yielding a bias of 53 ms which was statistically significant ($P < 0.001$).
Multimapping ECV was 21.5±1.9%, MOLLI ECV was 23.7±2.3%, resulting in a bias of -2.2%
which was statistically significant ($P = 0.001$). Multimapping native T_2 was 48.0±3.0 ms while
T_2bSSFP was 53.9±3.5 ms, a statistically significant bias of -5.9 ms ($P < 0.001$).

There was no correlation between native T_1 and T_2 for neither Multimapping nor MOLLI and
T_2bSSFP. Multimapping T_1 (native and post-contrast), T_2 or ECV and MOLLI T_1 (native and post-
contrast) did not correlate with heart-rate either. However, MOLLI ECV showed a correlation with
heart rate ($P = 0.03$), as did T_2bSSFP ($P < 0.001$). Plots for the correlation analysis between T_1 and
T_2 for Multimapping and reference techniques, and heart-rate and all techniques are shown in Figure
5.

3.2   Comparison of Multimapping and reference techniques for diseased myocardium

Correlation and Bland-Altman plots for Multimapping and reference techniques for native and post-
contrast T_1 (both myocardium and blood), ECV and native T_2 are shown in Figure 6. In general, the
correlation between Multimapping and the clinical reference techniques was very strong ($r^2 > 0.7$) for
most variables. A strong correlation coefficient ($r^2 > 0.5$) was found between Multimapping and
MOLLI for myocardial T_1 post-contrast ($r^2 = 0.66$) and blood T_1 post-contrast ($r^2 = 0.53$).

3.3   Inter- and intra-observer variability

The myocardial measurements and measurements of the left ventricular blood pool for intra-
repeatability assessment showed excellent repeatability (myocardial ICC > 0.97, LV blood pool ICC
= 1.00). The myocardial measurements for inter-repeatability showed moderate to excellent
repeatability (ICC > 0.73) for all mapping techniques. The native and post-contrast T_1 measurements
of the blood pool for inter-repeatability showed good to excellent repeatability (ICC > 0.92). ICC
values for each method can be seen in Table 2.

3.4   Image quality assessment

The image quality was scored significantly higher for Multimapping compared to T_2bSSFP ($P < $
0.001), MOLLI native T_1 ($P = 0.007$), MOLLI post-contrast T_1 ($P < 0.001$), and MOLLI ECV ($P < $
0.001). The distributions for the visual scores for all mapping techniques are shown in Figure 7.

4   Discussion

In this study, a new method for simultaneous T_1 and T_2 mapping was compared to clinical reference
mapping technique in a cohort of patients with cardiovascular disease. We found a strong to very
strong correlation between the methods for all measured parameters (native T_1, post-contrast T_1, ECV
and native T_2), while the image quality was considered better using the proposed Multimapping
technique compared to the reference methods. Furthermore, intra- and inter-observer variability of
Multimapping parameter measurements were in general low and similar to those obtained with the
clinical reference techniques.

In segments of healthy/remote myocardium we measured a mean native T_1 of 1116 ms using
Multimapping, more than 100 ms higher than for MOLLI. However, MOLLI is known to
significantly underestimate T_1 when compared to more accurate methods such as SASHA (5), which
typically yields native T_1 of around 1200 ms at 1.5T (7,26). The native T_1 Multimapping values are
also in line with the previous study using this technique in healthy volunteers which measured 1114
ms (24). For post-contrast T_1 mapping there was also a significantly longer T_1 using Multimapping
(479 ms) compared to MOLLI (427 ms). Although post-contrast T_1 values are more difficult to
compare between studies due to differences in contrast agents and the timing of acquisition after
injection, previous studies have shown underestimation of post-contrast T_1 for MOLLI compared to
more accurate techniques such as SASHA (27,28). The study by Nordlund et al. (28) also
demonstrated that MOLLI overestimates ECV by approximately 4% in healthy volunteers compared
to SASHA, the latter technique correlating more closely with radioisotopes in pigs. This suggests that
the significantly lower ECV measured in this study with Multimapping (22%) may be more accurate
compared to MOLLI (24%). However, further studies comparing Multimapping to more accurate
techniques such as SASHA are required to evaluate these claims. Correlation of T_1 and T_2 values
with potential confounding variables such as heart-rate or the opposite (T_2 or T_1) parameter did not
show any particular dependency for Multimapping in this regard. However, both T_2bSSFP and
MOLLI ECV appeared to be inversely correlated with heart-rate. This suggest additional delay
cardiac cycles may be required to yield less biased T_2 and MOLLI ECV values for high heart-rates.
Conversely, Multimapping may be a more robust approach for T_1 and T_2 mapping at higher heart
rates as no additional modification of the pulse sequence is required.

In the measurements of myocardial segments with disease, we found a high correlation between
Multimapping and the clinical reference techniques for native T_1 (blood and myocardium), T_2, and
ECV. While correlations for post-contrast T_1 (blood and myocardium) was more moderate, this may
be explained by the confounding factor of time after injection, which affect the T_1 measurements.
Furthermore, measured post-contrast T_1 in this study had a narrower range for both Multimapping
and MOLLI compared to native T_1 which can contribute to a weaker correlation between techniques.
Nevertheless, a very strong correlation between Multimapping and reference techniques for native
T_1, T_2 and ECV indicates that Multimapping is a useful technique that can be used to detect disease.

The image quality was superior using Multimapping compared to all clinical reference techniques.
This could be due to the higher flip angle of 50º using Multimapping, compared to MOLLI and
T_2bSSFP which both use a flip angle 35º, with otherwise identical imaging parameters to
Multimapping. A higher flip angle for bSSFP-based mapping techniques leads to a higher signal-to-
noise ratio which typically contributes to improved image quality. The shorter duration of
Multimapping (10 beats) compared to both MOLLI (11 beats) and T_2bSSFP (16 beats) means that
Multimapping is less prone to respiratory motion-induced misalignment, which may also contribute
to a better image quality. While the Multimapping and MOLLI pulse sequences are very similar
(both inversion recovery with bSSFP readouts), Multimapping benefits from phase sensitive
inversion recovery post-processing step which has been shown to improve T_1 map image quality

compared to fitting with magnitude images (29), used in the vendor provided fitting algorithm for
MOLLI. Compared to Multimapping, T$_2$bSSFP uses significantly fewer source images for T$_2$
mapping, and while only three T$_2$ preparation modules are included in the Multimapping pulse
sequence, the bSSFP readout is intrinsically T$_2$/T$_1$ weighted which contributes to the T$_2$ sensitivity
and may explain the improved image quality.

The intra- and inter-observer repeatability analysis showed an excellent repeatability for most
measurements using both Multimapping and reference techniques. While Multimapping post-contrast
myocardial T$_1$ inter-observer ICC of 0.73 was relatively low compared to that of MOLLI (ICC =
0.95), post-contrast T$_1$ mapping is primarily used to generate ECV maps, and here Multimapping and
MOLLI yielded near identical ICC values of 0.94 and 0.93, respectively.

4.1   Comparison with other simultaneous T$_1$ and T$_2$ mapping techniques

Several simultaneous T$_1$ and T$_2$ mapping techniques have been proposed over the last years,
comparable to Multimapping. Published studies using similar methods in healthy volunteers are
summarized in Table 3, including Multimapping (24). Multimapping has a shorter scan duration than
nearly all other simultaneous T$_1$ and T$_2$ mapping techniques, requiring 10 beats, which is also shorter
than both MOLLI and T$_2$bSSFP. As many patients with cardiovascular diseases struggle to hold their
breath for an extended period, reducing the scan time of mapping techniques is important and has
been the focus of several studies (11,30). This is also in line with the endeavour of utilizing less time-
consuming CMR protocols in order to improve adoption of CMR in routine cardiovascular practice.
Inversion recovery magnetization preparation pulses are often used for myocardial T$_1$ mapping as
they increase quantification precision compared to saturation recovery (5), using the full dynamic
range of the longitudinal magnetization, at the expense of accuracy as inversion pulses are more
sensitive to confounding elements such as magnetization transfer and transverse relaxation during the
pulses, which reduce their efficiency (6,31,32). Therefore, saturation recovery technique
measurements are generally considered to be closer to the "true" in-vivo T$_1$ times, typically several
100 ms higher than MOLLI on either 1.5T and 3T scanners. In this regard, Multimapping which uses
inversion recovery, generates T$_1$ values in healthy/remote myocardium of 1116 ms, which is closer to
the saturation recovery based techniques (of approximately 1200 ms) than MOLLI (approximately
1000 ms) or the most comparable simultaneous T$_1$ and T$_2$ mapping studies, Blume et al. (15) and
Jaubert et al. (33), which report a myocardial T$_1$ of 1017 ms and 1045 ms, respectively. This may be
due to the assumed lower inversion efficiency of 0.94 for the inversion pulses, incorporated into the
Multimapping signal model, which is likely closer to the true inversion efficiency than assuming
perfect efficiency. However, the inversion efficiency potentially varies between field strengths and
vendors, or even spatially across an image due to B$_0$ and B$_1$ inhomogeneity. Furthermore, the current
Multimapping technique does not consider magnetization transfer. To yield more accurate T$_1$ values,
reproducible across scanner platforms, these confounding effects should be included in the
Multimapping signal model, preferably on a pixel-wise basis, although this will likely negatively
impact the precision.

It can be difficult to precisely pinpoint the sources of differences in T$_1$ and T$_2$ values for the different
techniques outlined in Table 3, particularly as many techniques rely on unconventional acquisition,
reconstruction and mapping strategies. These include, for example, non-Cartesian (radial or spiral)
trajectories with iterative reconstruction algorithm, coupled with sophisticated and advanced mapping
techniques which may be difficult to reproduce. In contrast, the Multimapping pulse sequence
consists of a MOLLI-like acquisition scheme (inversion recovery with Cartesian single-shot 2D
bSSFP readout) which are available on all major vendor platforms, with the addition of T$_2$prep

modules which have also been implemented on all vendor platforms. For transparency, the
Multimapping parameter mapping method using dictionary matching has been provided open source
to enable reproduction of this technique by other investigators which may also enable direct
comparison of Multimapping with other simultaneous T_1 and T_2 mapping techniques.

4.2   Limitations

This study has several limitations: no respiratory motion correction was performed. Correcting for
respiratory-induced image misalignment is important even for breath-held scans and can be achieved
using image registration (34). Although image registration could be readily applied to Multimapping
source image to this end, this was not performed in order to have a fair comparison with MOLLI and
T_2bSSFP maps which were generated using inline vendor algorithm without motion correction. A
second technical limitation of Multimapping is that manual interaction is required to define a ROI in
the myocardial septum for the B_1 estimation. However, this is a relatively simple step, comparable to
the input required to define ROIs for ECV maps. Further work is required to automatize this step or
to incorporate B_1 in the high resolution T_1 and T_2 dictionary matching, which would obviate the need
for any manual interaction but with a potential penalty to the precision. A study limitation is that the
patient cohort consisted of a small, heterogeneous population of patients with various cardiovascular
diseases, and performed on a single 1.5T Philips scanner. Further studies are required to evaluate the
performance of Multimapping at 3T and using other vendors.

5   Conclusions

319 Multimapping T_1 and T_2 values show high correlations with clinical reference mapping techniques in
320 a diverse cohort of patients with different cardiovascular diseases. Multimapping enables
321 simultaneous T_1 and T_2 mapping and can be performed in a short (10 cardiac beats) breath-hold, with
322 image quality superior to that of the clinical reference techniques.

6   Conflict of Interest

324 The authors declare that the research was conducted in the absence of any commercial or financial
325 relationships that could be construed as a potential conflict of interest.

7   Author Contributions

327 MH and CJC conceived of the study. MH developed the methods, acquired the data, performed
328 image reconstruction and processing. MH, CJ, and CJC analyzed the data. All authors participated in
329 revising the manuscript, read and approved the final manuscript.

8   Funding

331 The research leading to these results has received funding from the Swedish Medical Research
332 Council [2018-02779], the Swedish Heart and Lung Foundation [20170440], ALF Grants Region
333 Östergötland [LIO-797721] and the Swedish Research Council [2018-04164].

9   Acknowledgments

335 Not applicable.

10   Data Availability Statement

The Multimapping post-processing pipeline (Matlab code) can be downloaded from:
https://github.com/Multimapping/Matlab_files. Images for all patients included in this study can be
downloaded from:
https://github.com/Multimapping/Patient_study/raw/main/MapReconstructions.pdf.

11   References

1.   Messroghli DR, Moon JC, Ferreira VM, Grosse-Wortmann L, He T, Kellman P, et al. Clinical
recommendations for cardiovascular magnetic resonance mapping of T1, T2, T2 and
extracellular volume: A consensus statement by the Society for Cardiovascular Magnetic
Resonance (SCMR) endorsed by the European Association for Cardiovascular Imaging
(EACVI). Vol. 19, Journal of Cardiovascular Magnetic Resonance. 2017.

2.   Kim PK, Hong YJ, Im DJ, Suh YJ, Park CH, Kim JY, et al. Myocardial T1 and T2 mapping:
Techniques and clinical applications. Vol. 18, Korean Journal of Radiology. 2017. p. 113–31.

3.   Ogier AC, Bustin A, Cochet H, Schwitter J, van Heeswijk RB. The Road Toward
Reproducibility of Parametric Mapping of the Heart: A Technical Review. Front Cardiovasc
Med. 2022;9:876475.

4.   Messroghli DR, Radjenovic A, Kozerke S, Higgins DM, Sivananthan MU, Ridgway JP.
Modified look-locker inversion recovery (MOLLI) for high-resolution T 1 mapping of the
heart. Magn Reson Med. 2004;52(1):141–6.

5.   Kellman P, Hansen MS. T1-mapping in the heart: Accuracy and precision. J Cardiovasc Magn
Reson. 2014;16(1).

6.   Robson MD, Piechnik SK, Tunnicliffe EM, Neubauer S. T1 measurements in the human
myocardium: The effects of magnetization transfer on the SASHA and MOLLI sequences.
Magn Reson Med. 2013;70(3):664–70.

7.   Chow K, Flewitt JA, Green JD, Pagano JJ, Friedrich MG, Thompson RB. Saturation recovery
single-shot acquisition (SASHA) for myocardial T 1 mapping. Magn Reson Med.
2014;71(6):2082–95.

8.   Weingärtner S, Akçakaya M, Basha T, Kissinger K V., Goddu B, Berg S, et al. Combined
saturation/inversion recovery sequences for improved evaluation of scar and diffuse fibrosis in
patients with arrhythmia or heart rate variability. Magn Reson Med. 2014;71(3):1024–34.

9.   Weingärtner S, Roujol S, Akçakaya M, Basha TA, Nezafat R. Free-breathing multislice native
myocardial T1 mapping using the slice-interleaved T1 (STONE) sequence. Magn Reson Med.
2015;74(1):115–24.

10.   Marty B, Coppa B, Carlier PG. Fast, precise, and accurate myocardial T1 mapping using a
radial MOLLI sequence with FLASH readout. Magn Reson Med. 2018;79(3):1387–98.

11.   Piechnik SK, Ferreira VM, Dall'Armellina E, Cochlin LE, Greiser A, Neubauer S, et al.
Shortened Modified Look-Locker Inversion recovery (ShMOLLI) for clinical myocardial T1-
mapping at 1.5 and 3 T within a 9 heartbeat breathhold. J Cardiovasc Magn Reson.
2010;12(1).

12.   Baeßler B, Schaarschmidt F, Stehning C, Schnackenburg B, Maintz D, Bunck AC. Cardiac
376   T2-mapping using a fast gradient echo spin echo sequence - First in vitro and in vivo
377   experience. J Cardiovasc Magn Reson. 2015;17(1).

378   13.   Ding H, Fernandez-De-Manuel L, Schär M, Schuleri KH, Halperin H, He L, et al. Three-
379   dimensional whole-heart T2 mapping at 3T. Magn Reson Med. 2015;74(3):803–16.

380   14.   Sprinkart AM, Luetkens JA, Träber F, Doerner J, Gieseke J, Schnackenburg B, et al. Gradient
381   Spin Echo (GraSE) imaging for fast myocardial T2 mapping. J Cardiovasc Magn Reson.
382   2015;17(1).

383   15.   Blume U, Lockie T, Stehning C, Sinclair S, Uribe S, Razavi R, et al. Interleaved T1 and T2
384   relaxation time mapping for cardiac applications. J Magn Reson Imaging. 2009;29(2):480–7.

385   16.   Shao J, Zhou Z, Nguyen KL, Finn JP, Hu P. Accurate, precise, simultaneous myocardial T1
386   and T2 mapping using a radial sequence with inversion recovery and T2 preparation. NMR
387   Biomed. 2019;32(11).

388   17.   Kvernby S, Warntjes MJ a. B, Haraldsson H, Carlhäll CJ, Engvall J, Ebbers T. Simultaneous
389   three-dimensional myocardial T1 and T2 mapping in one breath hold with 3D-QALAS. J
390   Cardiovasc Magn Reson. 2014;16:102.

391   18.   Akçakaya M, Weingärtner S, Basha TA, Roujol S, Bellm S, Nezafat R. Joint myocardial T1
392   and T2 mapping using a combination of saturation recovery and T2-preparation. Magn Reson
393   Med. 2016;76(3):888–96.

394   19.   Christodoulou AG, Shaw JL, Nguyen C, Yang Q, Xie Y, Wang N, et al. Magnetic resonance
395   multitasking for motion-resolved quantitative cardiovascular imaging. Nat Biomed Eng.
396   2018;2(4):215–26.

397   20.   Milotta G, Bustin A, Jaubert O, Neji R, Prieto C, Botnar RM. 3D whole-heart isotropic-
398   resolution motion-compensated joint T1/T2 mapping and water/fat imaging. Magn Reson
399   Med. 2020;84(6):3009–26.

400   21.   Mao X, Lee H-L, Hu Z, Cao T, Han F, Ma S, et al. Simultaneous Multi-Slice Cardiac MR
401   Multitasking for Motion-Resolved, Non-ECG, Free-Breathing T1–T2 Mapping. Front
402   Cardiovasc Med. 2022;9.

403   22.   Jaubert O, Cruz G, Bustin A, Hajhosseiny R, Nazir S, Schneider T, et al. T1, T2, and Fat
404   Fraction Cardiac MR Fingerprinting: Preliminary Clinical Evaluation. J Magn Reson Imaging.
405   2021;53(4):1253–65.

406   23.   Cavallo AU, Liu Y, Patterson A, Al-Kindi S, Hamilton J, Gilkeson R, et al. CMR
407   Fingerprinting for Myocardial T1, T2, and ECV Quantification in Patients With Nonischemic
408   Cardiomyopathy. Vol. 12, JACC: Cardiovascular Imaging. 2019. p. 1584–5.

409   24.   Henningsson M. Cartesian dictionary-based native T1 and T2 mapping of the myocardium.
410   Magn Reson Med. 2022 Jan 5;87(5):2347–62.

25.   Fent GJ, Garg P, Foley JRJ, Swoboda PP, Dobson LE, Erhayiem B, et al. Synthetic

Myocardial Extracellular Volume Fraction. Vol. 10, JACC: Cardiovascular Imaging. 2017. p.
1402–4.

26. Roujol S, Weingärtner S, Foppa M, Chow K, Kawaji K, Ngo LH, et al. Accuracy, precision,
and reproducibility of four T1 mapping sequences: A headto-head comparison of MOLLI,
ShMOLLI, SASHA, and SAPPHIRE. Radiology. 2014;272(3):683–9.

27. Weingärtner S, Meßner NM, Budjan J, Loßnitzer D, Mattler U, Papavassiliu T, et al.
Myocardial T1-mapping at 3T using saturation-recovery: reference values, precision and
comparison with MOLLI. J Cardiovasc Magn Reson. 2016;18(1):84.

28. Nordlund D, Xanthis C, Bidhult S, Jablonowski R, Kanski M, Kopic S, et al. Measuring
extracellular volume fraction by MRI: First verification of values given by clinical sequences.
Magn Reson Med. 2020;83(2):662–72.

29. Xue H, Greiser A, Zuehlsdorff S, Jolly MP, Guehring J, Arai AE, et al. Phase-sensitive
inversion recovery for myocardial T1 mapping with motion correction and parametric fitting.
Magn Reson Med. 2013;69(5):1408–20.

30. Guo R, El-Rewaidy H, Assana S, Cai X, Amyar A, Chow K, et al. Accelerated cardiac T1
mapping in four heartbeats with inline MyoMapNet: a deep learning-based T1 estimation
approach. J Cardiovasc Magn Reson. 2022;24(1).

31. Kellman P, Herzka DA, Hansen MS. Adiabatic inversion pulses for myocardial T1 mapping.
Magn Reson Med. 2014;71(4):1428–34.

32. Chow K, Kellman P, Spottiswoode BS, Nielles-Vallespin S, Arai AE, Salerno M, et al.
Saturation pulse design for quantitative myocardial T1 mapping. J Cardiovasc Magn Reson.
2015;17(1).

33. Jaubert O, Cruz G, Bustin A, Schneider T, Lavin B, Koken P, et al. Water–fat Dixon cardiac
magnetic resonance fingerprinting. Magn Reson Med. 2020;83(6):2107–23.

34. Roujol S, Foppa M, Weingärtner S, Manning WJ, Nezafat R. Adaptive registration of varying
contrast-weighted images for improved tissue characterization (ARCTIC): Application to T1
mapping. Magn Reson Med. 2015;73(4):1469–82.

35. Santini F, Kawel-Boehm N, Greiser A, Bremerich J, Bieri O. Simultaneous T1 and T2
quantification of the myocardium using cardiac balanced-SSFP inversion recovery with
interleaved sampling acquisition (CABIRIA). Magn Reson Med. 2015;74(2):365–71.

36. Hamilton JI, Jiang Y, Chen Y, Ma D, Lo WC, Griswold M, et al. MR fingerprinting for rapid
quantification of myocardial T1, T2, and proton spin density. Magn Reson Med.
2017;77(4):1446–58.

37. Shao J, Ghodrati V, Nguyen KL, Hu P. Fast and accurate calculation of myocardial T1 and T2
values using deep learning Bloch equation simulations (DeepBLESS). Magn Reson Med.
2020;84(5):2831–45.

38. Guo R, Cai X, Kucukseymen S, Rodriguez J, Paskavitz A, Pierce P, et al. Free-breathing

simultaneous myocardial T1 and T2 mapping with whole left ventricle coverage. Magn Reson
Med. 2021;85(3):1308–21.

39.   Hermann I, Kellman P, Demirel OB, Akçakaya M, Schad LR, Weingärtner S. Free-breathing
452   simultaneous T1, T2, and T2☐ quantification in the myocardium. Magn Reson Med.
453   2021;86(3):1226–40.

454   40.   Chow K, Hayes G, Flewitt JA, Feuchter P, Lydell C, Howarth A, et al. Improved accuracy and
455   precision with three-parameter simultaneous myocardial T1 and T2 mapping using
456   multiparametric SASHA. Magn Reson Med. 2022;87(6):2775–91.

12   Figures

Figure 1. Multimapping pulse sequence (A) and post-processing (B). The pulse sequence consists of
ECG-triggered 2D single-shot acquisitions in 10 consecutive cardiac cycles, preceded by inversion
pulses in cycles 1 and 5, and T$_2$prep with different echo times (TE) in cycles 8, 9 and 10. The post-
processing involves first performing a phase sensitive inversion recovery (PSIR) operation to restore
the polarity of the signal, followed by an initial dictionary matching (Dict match 1) to estimate a
global B$_1$ correction factor based on a ROI in the myocardial septum. The dictionary matching finds
the pixel-wise closest match between acquired data and the simulated acquisition (with scan-specific
delays) for different T$_1$, T$_2$ and RF scaling factor (B$_1$) combinations. A second dictionary matching
step (Dict match 2) is then performed for only T$_1$ and T$_2$ with high temporal granularity (1 ms
resolution) to generate the final T$_1$ and T$_2$ maps.

Figure 2. LGE and parameter maps (native T_2, native T_1, post-contrast T_1 and ECV) are shown for a patient with no cardiovascular disease findings. The parameter maps were acquired using either Multimapping (top row) or reference techniques (bottom row).

Figure 3. LGE and parameter maps in a patient with myocarditis, as indicated by the increased signal in the lateral wall in the LGE and also apparent as altered values in the parameter maps (Multimapping top row, reference techniques bottom row).

481   Figure 4. LGE and parameter maps in a patient with myocardial infarction, clearly visualized using
482   all techniques, both native and post-contrast.

484 Figure 5. The two top rows show the correlation analysis of heart-rate dependency to T_1 (native and
485 post-contrast), ECV and T_2. The bottom row shows the correlation analysis between native T_1 and T_2.

Correlation Plots

Bland Altman Plots

■ Native   ■ Post Contrast

Figure 6. Correlation and Bland-Altman plots comparing Multimapping (MM) to MOLLI and
T_2bSSFP. For the correlation plots, black and grey lines indicate line of best fit, and the dottet lines
show the identity lines. The black lines indicate bias in the Bland-Altman plots. Correlation
coefficient (r^2) is reported in the correlation plots and mean difference, lower and upper limits of
agreement (1.96 SD) in the Bland-Altman plots.

Figure 7. Distribution of image quality scores for Multimapping (MM) and in-vivo reference
techniques.

13   Tables

Table 1. Clinical Characteristics

Patients (n)	44
Age (years)	49±20
Male sex, n (%)	28 (64)
BMI (kg/m2)	25±4
Height (cm)	176±10

Weight (kg)	78±12
Heart rate (bpm)	67±14
LVEF (%)	53±11
LVSV (ml)	94±18
LVEDV (ml)	186±54

499   BMI, body mass index; LVEF, left ventricular ejection fraction; LVSV, left ventricular stroke
500   volume; LVEF, left ventricular end-diastolic volume.

502   Table 2. Inter and intra-observer ICC (95% confidence interval).

	Myocardium		Blood pool	
	Intra-repeatability	**Inter-repeatability**	**Intra-repeatability**	**Inter-repeatability**
MM T$_1$ Native	0.99 (0.97-1.00)	0.87 (0.76-0.94)	1.00 (1.00-1.00)	0.92 (0.85-0.96)
MOLLI T$_1$ Native	0.99 (0.97-0.99)	0.93 (0.88-0.97)	1.00 (1.00-1.00)	0.97 (1.00-1.00)
MM T$_1$ PC	0.99 (0.98-1.00)	0.73 (0.54-0.86)	1.00 (1.00-1.00)	1.00 (1.00-1.00)
MOLLI T$_1$ PC	0.97 (0.93-0.99)	0.95 (0.90-0.98)	1.00 (1.00-1.00)	1.00 (1.00-1.00)
MM ECV	0.99 (0.98-1.00)	0.94 (0.89-0.97)		
MOLLI ECV	0.99 (0.97-0.99)	0.93 (0.87-0.97)		
MM T$_2$	0.99 (0.96-0.99)	0.91 (0.82-0.95)		
T$_2$bSSFP	0.98 (0.94-0.99)	0.88 (0.78-0.95)		

503 MM, Multimapping; ECV, extracellular volume fraction; MOLLI, modified Look-Locker inversion
504 recovery; T$_2$bSSFP, T$_2$ prepared balanced steady-state free precession.

506   Table 3. Comparable published simultaneous T$_1$ and T$_2$ mapping techniques.

Study	Scan time	FB/ BH	IR/SR	Readout	Subjects	Additional mapping	Field strength	Native T1 (ms)	Native T2 (ms)
Blume et al. (15)	Around 3 minutes	FB	IR	2D Cartesian bSSFP	19 HV	-	1.5T	1017±91	50±4
Kvernby et al. (17)	15 beats	BH	IR	3D Cartesian GRE	10 HV	-	3T	1083±43	50.4±3.6
Akçakaya et al. (18)	13 beats	BH	SR	2D Cartesian bSSFP	10 HV	-	1.5T	1210±24	48.2±2.8
Santini et al. (35)	8 beats	BH	IR	2D Cartesian bSSFP	5 HV	-	3T	1227±68	37.9±2.4
Hamilton et al. (36)	15 beats	BH	IR	2D spiral GRE	11 HV	-	3T	1235, range 1199 - 1316	38, range 32 – 43
Jaubert et al. (33)	15 beats	BH	IR	2D radial ME-GRE	10 HV	PDFF	1.5T	1045±32	42.8±2.6
Christodoulou et al. (19)	88 seconds	FB	IR	2D radial GRE	10 HV	Cardiac motion	3T	1216±67	47.8±4.9
Shao et al. (37)	11 beats	BH	IR	2D radial GRE	10 HV	-	3T	1366±31	37.4±0.9
Guo et al. (38)	1.4±0.3 minutes (WH)	FB	SR	M2D Cartesian bSSFP	13 HV	-	3T	1373±50	48.7±2.5
Hermann et al. (39)	18.5 seconds (+resp gating)	FB	SR	2D Cartesian ME-GRE	10 HV	T2*	3T	1573±86	33.2±3.6
Chow et al. (40)	11 beats	BH	SR	2D Cartesian bSSFP	10 HV	-	3T	1523±18	36.7±1.1
Jarkman et al.	10 beats	BH	IR	2D Cartesian bSSFP	Remote myocardium 19 patients	-	1.5T	1116±21	48.0±3.0

507 WH, whole-heart; FB, free-breathing; BH, breath-hold; IR, inversion recovery; SR, saturation
508 recovery; bSSFP, balanced steady-state free precession; GRE, spoiled gradient echo; ME-GRE,
509 multi-echo spoiled gradient echo; M2D, multi-slice 2D; HV, healthy volunteer; PDFF, proton density
510 fat fraction.