Title: All-optical closed-loop voltage clamp for precise control of muscles and neurons in live animals

Authors: Amelie C.F. Bergs[1,2], Jana F. Liewald[1,2], Silvia Rodriguez-Rozada[3], Qiang Liu[4, 5], Christin Wirt[1,2], Artur Bessel[6], Nadja Zeitzschel[1,2], Hilal Durmaz[1,2], Adrianna Nozownik[3], Johannes Vierock[7], Cornelia I. Bargmann[4,8], Peter Hegemann[7], J. Simon Wiegert[3], Alexander Gottschalk[1,2,*]

Affiliations:

[1] Buchmann Institute for Molecular Life Sciences, Goethe University; Max-von-Laue-Strasse 15, D-60438 Frankfurt, Germany.

[2] Institute of Biophysical Chemistry, Goethe University; Max-von-Laue-Strasse 9, D-60438 Frankfurt, Germany.

[3] Research Group Synaptic Wiring and Information Processing, Center for Molecular Neurobiology Hamburg, University Medical Center Hamburg-Eppendorf, D-20251 Hamburg, Germany.

[4] Lulu and Anthony Wang Laboratory of Neural Circuits and Behavior, The Rockefeller University, New York, NY 10065, USA.

[5] Department of Neuroscience, City University of Hong Kong, Tat Chee Avenue, Kowloon Tong, Hong Kong, China.

[6] independent researcher: Melatener Strasse 93, D-52074 Aachen, Germany.

[7] Institute for Biology, Experimental Biophysics, Humboldt University, D-10115 Berlin, Germany.

[8] Chan Zuckerberg Initiative, Palo Alto, USA.

* Corresponding author. Email: a.gottschalk@em.uni-frankfurt.de

Abstract: Optogenetics are used to stimulate or inhibit neurons. Because optogenetic stimulation is typically static, neurons and circuits can quickly adapt, allowing perturbation, but not true control. To overcome this, we established an optogenetic voltage-clamp (OVC). The genetically encoded voltage-indicator QuasAr2 provides information for fast, closed-loop optical feedback to the bidirectional optogenetic actuator BiPOLES. Voltage-dependent fluorescence is held within tight margins, thus clamping the cell at a distinct potential. We established the OVC in muscles and neurons of *Caenorhabditis elegans*, and transferred it to hippocampal neurons in rat brain slices. We calibrated fluorescence signals to electrically measured membrane potentials, showed that the OVC reports on homeostatically altered cellular physiology in mutants affecting neurotransmission, and that it can dynamically clamp spiking. The OVC combines non-invasive imaging with the control capabilities of electrophysiology. Its applicability to individual cells facilitates high-throughput contact-less electrophysiology and paves the way for true optogenetic control in behaving animals.

One-Sentence Summary: Next generation optogenetic control of neuron function *via* optical feedback for non-invasive substitution of electrophysiology.

Introduction

Identifying connections between distinct neurons, and their contribution to driving behavior, is a central issue in neuroscience [1, 2]. To explore such relations, methods to record and concurrently regulate neural activity are needed [3]. Also, high-throughput screening approaches in excitable cell physiology require the use of such methods. To date, diverse approaches to control or observe excitable cell function are used. Patch-clamp electrophysiology provides superior temporal accuracy and sensitivity, yet is limited by its invasiveness [4-6]. Ca^{2+} imaging, applicable in intact living organisms, enables integrating cell physiology and behavioral output [7-9]. However, due to sensor Ca^{2+} buffering and the weak, non-linear correlation of cytosolic Ca^{2+} concentration and membrane voltage, this technique suffers from comparably low temporal resolution, and fails to resolve subthreshold voltage transients or high-frequency action potentials (APs). Further, as it does not directly measure membrane voltage and since Ca^{2+} concentration does not fall below basal cytosolic levels, Ca^{2+} imaging is unsuited to reveal synaptic inhibition. This is offered by genetically encoded (fluorescent) voltage indicators (GEVIs), e.g. rhodopsin-based GEVIs, among others [10-17]. The fluorescence of retinal, embedded in rhodopsins, reliably monitors voltage dynamics at millisecond time scales. Since rhodopsin-based GEVIs emit near-infrared light, they can be multiplexed with optogenetic actuators of membrane currents, to selectively photostimulate or -inhibit activity of excitable cells with high spatiotemporal precision [18-20]. Single optogenetic actuators were employed together with imaging tools, which allows to unidirectionally perturb and observe, but not to fully control activity within a neural circuit [11, 12, 21]. This approach was implemented as "(i)-Optopatch", comprising the voltage indicator QuasAr2 and the blue-light activated Channelrhodopsin (ChR) CheRiff [11, 12]. Bidirectional optical modulation and readout of voltage was implemented as light-induced electrophysiology (LiEp) for drug screening, however, this approach lacked a feedback loop [22]. A bidirectional approach, using ChR2 and halorhodopsin (NpHR) as actuation pair with feedback control – "optoclamp" – employed extracellular microelectrode arrays instead of GEVIs [23]. This approach observes (and clamps) neuronal ensembles and their average firing rates, but not single neurons, by indirect voltage readout.

True all-optical control over excitable cell activity with closed-loop feedback, as in voltage-clamp electrophysiology, should combine two opposing optogenetic actuators for de- and hyperpolarization, as well as a GEVI. Such a system could respond to intrinsic changes in membrane potential. It would also prevent inconsistent activity levels arising from cell-to-cell variation in optogenetic tool expression levels, not taken into account when static stimulation

patterns are used. The system should synergize the non-invasive character of imaging methods with the control capabilities of electrophysiology.

Here, we established an <u>o</u>ptogenetic <u>v</u>oltage-<u>c</u>lamp (OVC) in *Caenorhabditis elegans* and rat hippocampal organotypic slices. The OVC uses QuasAr2 for voltage read-out [11, 20], and BiPOLES, a tandem protein comprising the depolarizer Chrimson and the hyperpolarizer *Gt*ACR2, stimulated by orange and blue light, respectively, for actuation [19, 24-27]. Spectral separation and balanced 1:1 expression of BiPOLES enable gradual transitions from depolarized to hyperpolarized states, and *vice versa*. QuasAr2 fluorescence is used to compute a feedback of color-adapted light signals transmitted to BiPOLES, in closed-loop at 90 Hz, thus keeping the voltage-dependent fluorescence at a desired level. We characterized the system in body-wall muscle cells (BWMs), as well as in cholinergic and GABAergic motor neurons. Simultaneous measurements allowed calibrating fluorescent signals to actual membrane voltages. In *unc-13* mutants, the OVC readily detected altered excitability of muscle, as a response to the reduced presynaptic input. In rodent neurons, the OVC also modulated fluorescence and voltage, yet with a smaller range and acting more as a conductance-clamp, due to the resting potential being close to Cl⁻ reversal potential. Last, in spontaneously active tissues, i.e., pharyngeal muscle and the motor neuron DVB [28-31], the OVC could dynamically follow and counteract native APs.

Results

Reading GEVI fluorescence to steer optogenetic actuators with light feedback in closed-loop

An OVC should measure voltage-dependent fluorescence of an excitable cell, e.g. *via* a rhodopsin-GEVI [10, 11], and provide adjusted light-feedback to optogenetic actuators of membrane voltage, e.g. cation- and anion-selective rhodopsin channels [19, 27, 32, 33] (**Fig. 1A**). The OVC needs to work in closed-loop, to quickly counteract intrinsic activity. We first implemented suitable hard- and software (**Fig. 1B**): To excite fluorescence of rhodopsin GEVIs (e.g. QuasAr2, emitting in the far-red) [10, 11], we expanded a 637 nm laser to cover a larger field of view. GEVI fluorescence of a region of interest (ROI) was monitored by a sensitive camera, and compared to a target value (**Fig. 1C, Extended data Fig. 1A-C**). Light feedback is sent to the sample from a monochromator, whose wavelength limits (400-600 nm, near where the chosen hyperpolarizing and depolarizing optogenetic tools are maximally activated) could be accessed at 100 µs and 0.1 nm temporal and spectral resolution, respectively.

Communication between camera and monochromator is provided by a custom-written script in Beanshell (part of µManager interface [34]), processing incoming gray values into relative changes

of fluorescence ($\Delta F/F_0$) (**Fig. 1C, Extended data Fig. 1A-C**), and operating at ca. 90 fps speed (10 ms exposure). Since GEVIs exhibit photobleaching, $\Delta F/F_0$ values would gradually deviate from actual voltage levels. Thus, an initial calibration phase is used to calculate correction parameters (**Fig. Extended data Fig. 1D, E**). Once the system can access bleach-corrected $\Delta F/F_0$ values, it feeds them into a decision tree algorithm (**Fig. Extended data Fig. 1A**), where they are compared to a desired holding $\Delta F/F_0$ value. Deviation between target and actual $\Delta F/F_0$ determines the wavelength change of the monochromator (I(ntegral)-controller; **Extended data Fig. 1A**). An alternative algorithm, using a PID controller [35] and Kalman filter [36] for sensor smoothing, reduced the control deviation, but did not increase overall system performance (see methods; **Extended data Fig. 2**).

Fig. 1. Components, setup and functionality of the OVC. (A) Molecular OVC components: Two counteracting optogenetic actuators for de- and hyperpolarization, and a voltage indicator. **(B)** Hardware setup and communication: Membrane voltage is monitored via a fast and sensitive (sCMOS or EMCCD) camera; gray values are processed by live control software. Based on the difference to the set holding value, a feedback of light from a monochromator is sent to the optogenetic actuators. **(C)** OVC feedback mechanism: After selection of clamping parameters and calibration for bleaching correction, $\Delta F/F_0$ is calculated and compared with the set value, to adapt the wavelength accordingly, in closed-loop.

Combining single actuators with QuasAr2 for unidirectional steering of membrane voltage

First, we expressed QuasAr2 in *C. elegans* body wall muscles (BWMs). 20 s calibration under 637 nm laser light sufficed to estimate photobleaching parameters. 300 µW/mm² blue light or presenting the full spectrum (400 to 600 nm) caused no additional bleaching and did not affect

QuasAr2 fluorescence (**Extended data Fig. 1D-F**). Thus, monochromator light does not affect bleaching-corrected fluorescence and estimated membrane voltages. To test the feedback loop, we expressed either *Chlamydomonas reinhardtii* ChR2(H134R), or *Guillardia theta* anion channelrhodopsin *Gt*ACR2, in cholinergic motor neurons (**Extended data Fig. 3A**), to manipulate muscle voltage indirectly via light-induced (de-)activation of motor neurons (**Extended data Fig. 3B, C**). ChR2 or *Gt*ACR2 activation allowed adjusting QuasAr2 fluorescence to values between +20 and -15 % $\Delta F/F_0$, respectively (**Extended data Fig. 3D**). However, fluorescence (i.e., voltage) returned to baseline only by intrinsic membrane potential relaxation (**Extended data Fig. 3D-G**), thus limiting temporal resolution: For ChR2, fluorescence reached the target range after 150 ± 12.7 ms and relaxed within 678 ± 148.9 ms (**Extended data Fig. 3G**; 254 ± 20.4 ms and 239.5 ± 29.9 ms, respectively, for *Gt*ACR2).

BiPOLES enables bidirectional voltage-clamping in *C. elegans* muscle

Optogenetic de- or hyperpolarization of cholinergic motor neurons affects muscle activation and body contraction or elongation [32, 37]. We used this to assess combinations of spectrally distinct, opposing actuator pairs, like ChR2 (470 nm) and *Gt*ACR1 (515 nm), or *Natronomonas pharaonis* halorhodopsin NpHR (590 nm) (**Extended data Fig. 4A-D**). Yet, as separate expression precluded control over relative expression levels, one actuator was typically outperformed by the other, impeding precise control of membrane potential. To overcome this problem, we used the tandem protein BiPOLES, in which the de- and hyperpolarizers Chrimson and GtACR2 (590 nm and 470 nm, respectively) are linked as one sequence [26] (**Fig. 2A**).

BiPOLES activation with a 400-600 nm ramp evoked robust bidirectional effects on body length (**Extended data Fig. 4E**). We thus co-expressed BiPOLES and QuasAr2 in BWMs (**Fig. 2A, B**). BiPOLES expressed at comparably low levels, at the plasma membrane (PM) and in few intracellular aggregates, while QuasAr2 localized mainly to the PM. QuasAr2 excitation (637 nm) causes some activation of Chrimson (their spectra overlap [25, 38]; **Extended data Fig. 5A**). As BiPOLES enables bidirectional voltage steering, we counteracted Chrimson currents by setting the monochromator to a compensatory wavelength. This was adjusted until -5 and 5 % $\Delta F/F_0$ could each be maintained ≈5 s. In animals expressing the OVC in BWMs and/or cholinergic neurons, 637 nm laser light caused reduced motor behavior, as this is sensitive to perturbation of muscle and/or motor neuron function; however, adding compensation light (521 ± 1.68 nm) restored body length, as well as swimming behavior, (**Extended data Fig. 5C-F, Extended data video 1**). BWMs expressing the OVC showed voltage fluctuations comparable to animals expressing QuasAr2 only (**Extended data Fig. 4F**). Also, despite 637 nm laser light, Chrimson

could be further activated by 590 nm (300 µW/mm²) light, causing a ca. 2.3 % ΔF/F$_0$ increase of

QuasAr2 fluorescence (**Extended data Fig. 5B**).

Fig. 2. Bi-directional clamping of membrane voltage in BWMs. (A) OVC (QuasAr2 and BiPOLES; inset, ★: mCerulean) in BWMs. **(B)** Confocal fluorescence z-projection. BiPOLES (mCerulean) and QuasAr2 in BWM membranes, scheme (dashed lines represent muscle plasma membranes). Scale bar, 50 µm. **(C)** OVC three-step protocol (0, -5 and 5 % ΔF/F$_0$) in BWMs, insets for close-up. Wavelength shown in the respective color, holding values and tolerance range (grey boxes) are indicated for each step. Orange shade in upper panel: transition period to reach tolerance range. **(D)** Upper panel: Overlay of mean (± S.E.M.) wavelength and (lower panel) fluorescence traces (n=22; holding values: 0, -5, 5 % ΔF/F$_0$). **(E)** Times required for the indicated 5 and 10 % ΔF/F$_0$ transitions. **(F)** Membrane resistance of BWM cells, before, and during 637 nm laser illumination (orange), and during laser + calibration light; n=6. **(G)** APs in simultaneous patch-clamp and fluorescence recordings during OVC calibration phase (637 nm laser + calibration wavelength). Inset for close-up. **(H)** Example trace of simultaneous patch-clamp and fluorescence recording of an OVC three-step protocol (-3 to +3 % ΔF/F$_0$). Upper panel: Fluorescence recording with indicated wavelength adaptation and tolerance ranges. Lower panel: Corresponding simultaneous voltage recording. Membrane potential was recorded in whole-cell patch-clamp mode. **(I)** Overlay of mean (± S.E.M.) wavelength, fluorescence (upper panel) and voltage traces (lower panel) of N=4, n=8 measurements as in (H). **(J)** Statistical analysis of data in (I); voltage modulation between transition events during simultaneous OVC measurements (left panel: absolute voltages; right panel: voltage modulation for the respective steps of the OVC protocol). One-way ANOVA with Bonferroni correction. ***P ≤ 0.001, **P ≤ 0.01. In (E, F and J), box plots (median, 25-75th quartiles); mean: open dot; whiskers: 1.5x IQR.

After bleaching correction (R^2 of the exponential fit was always higher than 0.8, in most cases even >0.95; **Extended data Fig. 6B**), the OVC generated incremental changes in wavelength that closely followed the fluctuating fluorescence signals (i.e., membrane voltage; **Fig. 2C**), as soon as the tolerance range for holding $\Delta F/F_0$ was exceeded. Cells were reliably (**Extended data Fig. 6C**), and quickly (**Extended data Fig. 6A**), clamped to holding values between -5 and 5 % $\Delta F/F_0$ (**Fig. 2C, D**), sometimes even -8 to 8 % $\Delta F/F_0$ could be reached (**Extended data Fig. 6F**). This range is smaller than for single-tool experiments (**Extended data Fig. 3D**), since Chrimson and GtACR2 action spectra partially overlap, and the channels have different conductance [19, 25, 27]. However, due to the bidirectionality and live feedback, the BiPOLES-OVC acted significantly faster, particularly for the return towards resting potential: Transition times were only 147.7 ± 25.5 ms (5→0 %) and 88.0 ± 5.7 ms (-5→0 %) for BiPOLES, compared to 678.6 ± 148.9 ms for ChR2 and 239.5 ± 29.9 ms for GtACR2 (**Fig. 2E, Extended data Fig. 3G, S6A**). The OVC allowed continuous bidirectional clamping (up to minutes; **Extended data Fig. 5G**), something not achievable with single tools. Once the closed-loop control was interrupted, membrane voltage returned to baseline and higher fluorescence fluctuations were observed (1.87 ± 0.1 vs. 0.78 ± 0.05 % $\Delta F/F_0$ during -5 % clamping; 1.89 ± 0.14 % $\Delta F/F_0$ for animals expressing only QuasAr; **Extended data Fig. 4F**). To extend the applicability of the OVC, we wrote software allowing to select holding $\Delta F/F_0$ values during a running acquisition (**Extended data Fig. 6D-F**), i.e. to respond to observations and to dynamically adapt clamping parameters. The software provides a live status update, whether the OVC system is on hold, adapting or if it has reached its limits (**Extended data Fig. 6D, Extended data video 2**).

Calibrating QuasAr2 fluorescence and membrane potential in BWMs

Next, we determined the actual voltage range that can be adjusted by the OVC. First, we measured membrane voltages and QuasAr2 fluorescence levels in response to wavelength ramps within the margins of the OVC (400-600 nm and *vice versa*). The range of relative fluorescence changes ($\Delta\Delta F/F_0$ at 400-410 and 590-600 nm) was around 16 % (**Extended data Fig. 7A-D**), in line with the range of holding values stably achieved by the OVC (± 8 % $\Delta F/F_0$, **Extended data Fig. 6F**). The corresponding voltage range was around 18 mV (**Extended data Fig. 7D**). The QuasAr2 fluorescence-voltage relation is linear [11], thus we estimated that during -5 to +5 % $\Delta F/F_0$ OVC operation, about 11 mV membrane potential change is achieved. However, these measurements were without OVC feedback. Thus, to directly calibrate the OVC operation range, we concomitantly measured voltage and fluorescence (**Extended data Fig. 7E**). Membrane resistance was not significantly altered by 637 nm laser or compensation light (**Fig. 2F**), and corresponding APs could be observed by fluorescence and patch-clamp before the OVC

clamp phase started (**Fig. 2G**). Concomitant laser (637 nm) and compensation illumination (520 nm) did not alter APs (**Extended data Fig. 7F, G**), and it did not alter membrane potential; however, adding 470 nm light could hyperpolarize the cell by about 16 mV (**Extended data Fig. 7H**). Also, small voltage fluctuations were similar in dark and light conditions by amplitude and frequency (**Extended data Fig. 7I-K**), arguing that membrane resistance of the cells was not altered by the compensation light (**Fig. 2F**). Thus, BiPOLES activation does not lead to general shunting of membrane potential due to open channels. Finally, OVC-induced fluorescence changes (-3 to +3 % $\Delta F/F_0$) modulated voltage by ca. -6 to +7 mV (**Fig. 2H-J**), i.e. ca. 13 mV range per 6 % $\Delta F/F_0$ (22 mV per 10 % $\Delta F/F_0$). This was larger than what we observed by indirect comparison (**Extended data Fig. 7D**), probably due to the live-feedback of the OVC.

Demonstrating homeostatic changes in muscle excitability using the OVC

May the OVC be used to assess divergent cell physiology? *unc-13(n2813)* mutants lack an essential synaptic vesicle priming factor [39] and thus exhibit largely reduced postsynaptic currents upon ChR2-stimulation of motor neurons [40]. Yet, *unc-13* and other neurotransmission mutants showed enhanced muscle contraction when BWMs were directly photostimulated. We hypothesized that muscles homeostatically change their excitability. To explore if the OVC can detect this, we induced a +5 % $\Delta F/F_0$ depolarization step: Indeed, in *unc-13(n2813)* mutants the OVC required significantly blue shifted light (i.e., less Chrimson activation) to induce the same level of depolarization as in wild type (534.2 ± 2.8 vs. 550.4 ± 3.1 nm; **Fig. 3A, B**). Muscles in *unc-13* animals may exhibit higher excitability to balance the lower excitatory input they receive. Though membrane resistance was unaltered (**Fig. 3C**), the amplitude of induced voltage increases was higher in *unc-13* mutants when we injected current ramps (**Fig. 3D-F**). Similarly, APs measured by voltage imaging were significantly increased by amplitude and duration (**Fig. 3G-J**). This indicates that in response to a lack of presynaptic input, ion channels shaping muscle voltage / APs, i.e. voltage-gated Ca^{2+} (EGL-19) and K^+ channels (SLO-2, SHK-1) [41, 42], may undergo homeostatic changes to enable the observed higher muscle excitability.

Fig. 3. Assessing altered cell physiology in mutants, using the OVC. (A) *unc-13(n2813)* mutant and wild type

animals were subjected to OVC depolarization, to +5 % ΔF/F₀. Mean wavelength traces (± S.E.M.), n = 9-10. **(B)** Mean

wavelength required during the depolarization step, required to hold +5 % ΔF/F₀. **(C)** Membrane resistance was

analyzed in wild type and *unc-13* mutant muscles. n = 9-10. **(D, E)** Current ramps (0-20 pA over 100 ms; red) were

injected into wild type and *unc-13* mutant muscles and the induced voltage increases were aligned to the peak. n = 5-

11. **(F)** Group data of induced voltage increases in (D and E). **(G, H)** Optical recordings of spontaneous APs in *unc-13*

mutants and wild type animals. Overlay of single (each two are highlighted in cyan) and mean (± S.E.M.) traces. n =

26-27. **(I, J)** Statistical analysis of peak AP amplitude and AP duration (at FWHM) of data in (G and H). Two-sided t-

test with Bonferroni correction in (B, C, F, I, J). ***P ≤ 0.001, **P ≤ 0.01, *P ≤ 0.05 and box plots (median, 25-75th

quartiles); mean: open dot; whiskers: 1.5x IQR.

OVC in cholinergic and GABAergic motor neurons

We tested if the OVC works also in *C. elegans* neurons. Cholinergic and GABAergic motor neurons are small (ca. 2-3 µm cell body, BWMs ca. 50 µm) and exhibit low absolute fluorescence. Both mCerulean (BiPOLES) and QuasAr2 fluorescence were observed in respective ganglia, including the anterior nerve ring and ventral nerve cord (**Fig. 4A, B**). Calibration parameters could be adopted from muscle experiments (**Fig. 4C**). The OVC could readily clamp neuronal voltage-dependent fluorescence between -5 and 5 % $\Delta F/F_0$ (**Fig. 4C-G**), and also adaptive experiments were possible (**Extended data Fig. 6G, H**).

Fig. 4. Bi-directional clamping of voltage-dependent fluorescence in *C. elegans* neurons. (A) OVC in cholinergic or GABAergic neurons. **(B)** Confocal fluorescence z-projections of BiPOLES (mCerulean) and QuasAr2 expression. Arrowheads: Neuronal cell bodies. Scale bars: 50 µm. **(C)** OVC protocol (holding values 0, -5, 5 % $\Delta F/F_0$) in cholinergic neurons, single recording. **(D, E)** Mean (± S.E.M.) traces of the OVC in cholinergic (D; n = 14), or GABAergic (E, n = 7) neurons, respectively, holding values: 0, -5, 5 % $\Delta F/F_0$. **(F, G)** Transition times for 5 and 10 % $\Delta F/F_0$ steps (F: cholinergic, G: GABAergic neurons); box plots (median, 25-75th quartiles), mean: open dot, whiskers: 1.5x IQR.

In mammalian neurons, the OVC acts as an optogenetic conductance-clamp

Rodent neurons are larger, more hyperpolarized cells and display faster AP kinetics than *C. elegans* muscles and neurons. Since such cells are relevant to studies of human-like neurophysiology, we explored the utility of the OVC in rat hippocampal pyramidal neurons (organotypic slice culture; **Fig. 5A**), expressing QuasAr2 and soma-targeted (som)BiPOLES.

Fig. 5. Establishing the OVC in rat hippocampal pyramidal neurons. (A) Fluorescence micrographs of somBiPOLES (CFP co-expression marker) and QuasAr2 expression in a neuron (asterisk). Scale bar: 10 µm. **(B)** OVC three-step protocol (0, -3 and 3 % ΔF/F$_0$). Right: Wavelength shown in the respective color (upper), holding values and tolerance range (grey boxes, lower panel) are indicated for each step. Left: Orange shade marks transition/calibration period to reach tolerance range. **(C)** Overlay of mean (± S.E.M.) wavelength (upper panel) and fluorescence traces (lower panel; n=12; holding values: 0, -3, 3 % ΔF/F$_0$). **(D)** Average fluorescence data (lower panel) for cells expressing QuasAr2 only, while the OVC attempts to run a 0, -3, 3 % ΔF/F$_0$ protocol (n = 5; upper panel: Monochromator wavelength). **(E)** QuasAr2 fluorescence during electrically evoked 100 mV depolarization step, from -74.5 mV to +25.5 mV holding voltage. **(F)** Simultaneous patch-clamp (voltage, lower panel) and fluorescence recording (upper panel) with indicated wavelength adaptation and tolerance ranges. **(G)** Statistical analysis of voltage modulation between transition events during simultaneous patch clamp / OVC measurements (n = 3-7). **(H)** Modulation of membrane voltage in cells expressing somBiPOLES only, in response to different light application, as indicated. One-way ANOVA with Bonferroni correction. ***P ≤ 0.001, **P ≤ 0.01, *P ≤ 0.05; n = 3-7. In (G, H), box plots (median, 25-75th quartiles); mean: open dot; whiskers: 1.5x IQR.

After bleaching calibration (**Fig. 5B**), optical clamping could be achieved between ± 3 % $\Delta F/F_0$ (**Fig. 5B, C**) with the OVC protocol as used in *C. elegans*. Importantly, in the absence of somBiPOLES, monochromator light did not lead to modulation of QuasAr2 fluorescence (**Fig. 5D**), while electrically evoked potential shifts (100 mV depolarization step, from -74.5 mV to +25.5 mV holding voltage) caused clear increases of QuasAr2 fluorescence (ca. 21 % $\Delta F/F_0$; **Fig. 5E**). In cells expressing QuasAr2 and somBiPOLES, the ± 3 % $\Delta F/F_0$ OVC protocol caused hyperpolarizing and depolarizing potential jumps (ca. - 4 mV and + 3 mV, respectively; **Fig. 5F, G**). These were small, compared to earlier experiments using somBiPOLES, where 595 nm light application caused depolarization of up to 30 mV, while 400 nm light clamped cells to the Cl⁻ reversal potential [26]. Possibly, due to the more negative resting potential in mammalian neurons compared to *C. elegans* cells, 637 nm activation of Chrimson causes stronger effects, and thus the OVC triggers more compensatory GtACR2 currents. When we expressed somBiPOLES only, 637 nm laser light depolarized hippocampal neurons (via Chrimson) by ca. 21.5 mV, which could be partially counteracted by GtACR2 activation using 530 nm compensation light (**Fig. 5H**). However, due to the high conductance of GtACR2, shunting effects likely prevented further optical hyper- or depolarization. Thus, in mammalian neurons, the OVC works with a limited range, likely due to different voltage and ion conditions in the resting state, and acts more as a 'conductance-clamp'.

Dynamic suppression of muscle APs

Thus far, we imposed fixed voltage steps to cells. To see if the OVC can dynamically counteract spontaneous activity, as opposed to shunting hyperpolarization, we turned to a periodically active muscular pump, the *C. elegans* pharynx (**Fig. 6A**). This feeding organ exhibits periodic, ca. 4 Hz APs [20], triggered by serotonin. Closing of the grinder, a structure used to crush bacteria, was triggered by activation of *Gt*ACR2 (400 nm), and could be observed by QuasAr2 fluorescence (**Extended data Fig. 8A**), while illumination with only the 637 nm laser caused grinder opening. The OVC could clamp the pharynx between -5 and 5 % $\Delta F/F_0$ (**Fig. 6B, C, Extended data Fig. 8B-D**). Optically observed APs show a rise time constant of around 15 ms and duration of ca. 150 ms (at full-width-half-maximum - FWHM) [20]. To clamp these APs dynamically, i.e., to counteract de- and repolarization phases quickly enough, we altered the OVC software to respond with more pronounced wavelength changes to a given fluorescence change (increased integral gain; methods, **equation 6**), and reduced the tolerance range to ± 0.005 %. During calibration, APs showed 30.1 ± 1.1 % $\Delta F/F_0$ mean amplitude and 120 ± 4 ms duration at FWHM (**Fig. 6B, C, E, F**). Upon clamping, all subsequent attempts to spike were dynamically counteracted and almost completely suppressed (**Fig. 6B-F**). AP amplitude was significantly reduced to ca. 2 % (**Fig. 6E,**

F), as was AP duration (to 82.6 ± 3.3 ms at FWHM). Clamping was rapid: APs occurring at clamping onset were shortened by 65 %, while the OVC traversed the full wavelength spectrum within 20 ms (**Fig. 6D**).

Fig. 6. All-optical clamping of APs in pharyngeal muscle and the enteric motor neuron DVB. (A) The *C. elegans* pharynx, a muscular pump used for feeding, expressing the OVC (QuasAr2 fluorescence). **(B-D)** OVC experiment in pharyngeal muscle, to hold fluorescence at 0% $\Delta F/F_0$ and suppress APs. (B) Original trace, calibration, and clamping phase. Inset shows close-up of boxed region, transition from calibration to clamp phase. (C) Overlay of wavelength and $\Delta F/F_0$ traces during clamp phase. (D) Close-up of region boxed in (C): OVC counteracting first AP during clamp phase. **(E)** Aligned traces of spontaneous (light grey, n=20) and clamped (blue, n=20) pharyngeal fluorescence signals. Mean wavelength chosen by the system is shown in the respective color. **(F)** Statistical analysis of data in (E), fluorescence voltage signal amplitude and duration at FWHM. **(G)** The GABAergic motor neuron DVB (cyan) innervates enteric

muscles (grey, upper panel). Below: Confocal z-projection of BiPOLES and QuasAr2 fluorescence in DVB, and overlay;
scale bar: 50 μm. **(H)** Voltage fluorescence signals of spontaneous APs in DVB (blue trace). Expulsion muscle
contraction, deduced from DVB cell body movement, shown as gray trace. Right panel: Close-up of boxed region. **(I)**
Overlay of spontaneous (light gray traces, n=8; mean: black) and clamped (blue traces, n=5; mean: orange) DVB
voltage signals. Mean wavelength set by the OVC shown in the respective color. **(J)** Statistical analysis of data in (I),
mean voltage fluorescence signal amplitude and total duration. **(K)** Current clamp recordings (n=5, lower panel) of DVB
in dissected animals, following indicated (upper panel) hyperpolarizing (grey) and depolarizing (cyan) injection of
current in 1 pA steps. **(L)** Group data of (K): resting potential, threshold, peak amplitude, duration at FWHM, and
afterhyperpolarization (AHP). In (F, J and L): box plots (median, 25-75th quartiles), mean: open dot, whiskers: 1.5x
IQR. In (F and L), two-sided t-test with Bonferroni correction. ***P ≤ 0.001, **P ≤ 0.01.

Clamping APs in the GABAergic motor neuron DVB

C. elegans exhibits a rhythmic motor program to move and expel gut contents [28-31]. Expulsion

muscle contraction is regulated by an intestinal pacemaker and by GABAergic motor neurons

DVB and AVL (**Fig. 6G**). Ca^{2+} imaging of DVB [43] revealed activity reminiscent of APs, thus far

observed in only few *C. elegans* neurons [44]. DVB voltage imaging showed APs (7.7 ± 1 % $\Delta F/F_0$,

500 ± 60 ms at FWHM) at regular time intervals of 40-50 s, that were followed by posterior body

contraction (**Fig. 6H**). Using parameters as for the pharynx, DVB APs were clamped to

significantly reduced amplitude and duration (**Fig. 6I, J**; see **Extended data Fig. 8E** for single

trace). DVB allowed calibrating the neuronal OVC by patch-clamp electrophysiology [45]: Resting

potential was -49.0 ± 8.9 mV (**Fig. 6K, L**), and increasing current steps (1 pA) evoked APs (-23.4

mV threshold, depolarization to 26.3 mV, 359 ms duration (FWHM), -38.5 mV

afterhyperpolarization; n=5 cells; **Fig. 6L**). DVB AP amplitude thus is ca. 50 mV. Although

fluorescence baseline may be altered by the 637 nm laser and Chrimson activation, this did not

evoke APs. Thus, the QuasAr2 signal in DVB, i.e., 7.7 % $\Delta F/F_0$ for an AP, corresponds to ca. 50

mV (calculating from threshold, or up to 75 mV, from resting potential).

Discussion

Here, we established the OVC, the first all-optical voltage-clamp approach to date. We

demonstrated its performance in various excitable cell types in intact animals (*C. elegans*) and

tested it in mammalian hippocampal neurons. In *C. elegans*, the OVC allowed reliable clamping

of voltage in two muscular organs and three neuron types *via* QuasAr2 fluorescence read-out and

bidirectional optogenetic actuation *via* BiPOLES. The OVC could further detect the altered

postsynaptic excitability of *unc-13* mutants, a homeostatic response to reduced presynaptic input.

Using the OVC, we did not only control resting membrane potential, but could also dynamically

clamp spontaneous rhythmic activity and APs in pharyngeal muscle and in the motor neuron DVB.

The OVC operates at rates up to 90 Hz on a typical PC with moderate computing power, which

orchestrates the communication between camera and monochromator, while in parallel running image acquisition and providing live computation of bleaching-corrected $\Delta F/F_0$ values, fed into a decision tree and Integral-control algorithm. An alternative algorithm with PID controller and Kalman filter provided more accurate clamping, though it did not increase system speed. While the OVC at present is slower and less sensitive than standard patch-clamp electrophysiology, it outperforms it in terms of non-invasiveness, throughput and ease of application. In mammalian neurons, likely due to different physiology, the OVC in its current composition of light-gated ion channels and QuasAr2 worked as a conductance clamp. However, in *C. elegans*, though activation of Chrimson by the QuasAr2 excitation light affects baseline physiology, this effect could be counterbalanced by utilizing the bidirectional control properties of BiPOLES, using a compensation wavelength. No adverse effects on membrane resistance were observed (**Fig. 2F**), it did not affect membrane potential, intrinsic neuronal activity and cellular excitability, as judged by voltage fluctuations in muscle (**Extended data Fig. 7H-K**), muscles fired normal APs (**Extended data Figs. 4F, 7F, G**), locomotion behavior was unaltered (**Extended data Fig. 5E, F**), and also spontaneous activity in pharyngeal muscle or the DVB neuron was unaffected.

The voltage range covered by the OVC differed between muscles and neurons: 10 % $\Delta F/F_0$ change corresponded to about 22 mV in BWMs, and to about 65 mV in DVB. Differences in membrane potential and AP amplitude depend on the different ion channels and gradients present in the two cell types. The different range of $\Delta F/F_0$ fluorescence of QuasAr2 in muscles and neurons may correspond to a different surface to volume ratio of the two cell types, and possibly, different levels of QuasAr2 protein in plasma membrane vs. intracellular membranes. Fluorescence from the latter would not contribute to voltage-dependent $\Delta F/F_0$, while it would increase overall fluorescence at rest.

In mammalian neurons, optical voltage control was currently limited to a few mV. This may in part be due to the high expression levels of BiPOLES, and in part due to the low electrochemical gradient for Cl⁻, as the resting potential is close to the Nernst potential for Cl⁻. Thus, in comparison to *C. elegans*, more activation of GtACR2 is required to counteract Chrimson-based depolarization, causing membrane leak. Minimizing the optical crosstalk of voltage sensor and BiPOLES is especially important in mammalian neurons. Further red-shifted voltage sensors and blue-shifted BiPOLES variants are thus required for future OVC experiments. The membrane leak in mammalian neurons makes the OVC a 'conductance-clamp'. Light-driven ion pumps, provided they become as powerful as the channels used here, could be a future alternative.

The speed of the OVC is currently limited by the sensitivity, fluorescence yield, dynamic range of the GEVI (requiring 10 ms exposure times) and its time constant, the camera frame rate, and the computer hardware. With higher fluorescence change in response to membrane potential and

a larger absolute GEVI fluorescence (i.e., improved signal to noise ratio), higher frame rates could
potentially be achieved. The software component currently runs within µManager on a PC.
Running the software on an integrated circuit, like a field programmable gate array (FPGA), and
using a faster camera may enable faster operation speeds up to 1 kHz. GEVI kinetics is not likely
to become a limiting factor, as QuasAr2 has been shown to faithfully report on single APs in
mammalian neurons (2 ms), clearly showing AP rise and repolarization phases [11]. OFF kinetics
of the actuators may pose a limit to the maximal response times of the OVC: These are ca. 21-
45 ms for Chrimson, for which, however, a 10 times faster variant has been isolated [24, 25], and ca.
40 ms for *Gt*ACR2; also here, a 20 times faster ZipACR was described [46]. However, due to
BiPOLES, fast OFF kinetics are less important, as it enables active counteraction. Since channel
ON kinetics are generally faster, they are also not likely to limit the maximal speed achievable by
the OVC.

The OVC detected altered muscular excitability in *C. elegans unc-13* mutants, which is likely
caused by changes in ion channel physiology. This shows that the OVC represents an approach
to enable all-optical, contact-less high-throughput screening applications, e.g. of compound
libraries targeted at ion channels. Testing its use in mammalian cell types other than neurons,
where such ion channels can be expressed individually, and where the resting membrane
potential is less depending on the Cl⁻ gradient, will facilitate this approach.

The establishment of the OVC paves the way for all-optical control of individual neurons in
freely behaving transparent animals like *C. elegans*. Accuracy of fluorescence quantification (and
thus OVC feedback) requires the cell to remain in focus, which can be achieved with the OVC,
while it is impossible to keep a patch-clamp electrode physically attached to such small animala.
Using the OVC on neurons controlling behavior will allow fine-tuning of, and enable understanding
how activity of the individual cell regulates, behavioral aspects. Online behavioral analysis (e.g.
extent of body bending, locomotion velocity), may be used to feedback to the neuron such that
behavior can be dynamically controlled.

Transferring the OVC to other model organisms may require modifications depending on the
respective cellular properties. Extending electrophysiology applications, the OVC will enable
efficient space clamp, as well as dynamic local clamping in neuronal processes. The OVC
software is universally applicable, as it can be easily adapted to other GEVI-optogenetic actuator
combinations. The OVC broadens applicability of optogenetics as it allows modulation in closed-
loop, as opposed to static optical perturbation that cannot adapt to the variable activity patterns
found in living organisms. Dynamic responsiveness is also advantageous with regard to future
therapeutic applications, for example in acute control of seizures [47], or in adaptive deep brain
stimulation, as it would allow adjusting the therapy to the need of the patient [48].

References

1. Bargmann, C.I. & Marder, E. From the connectome to brain function. *Nat Methods* **10**, 483-490 (2013).
2. Anderson, D.J. Circuit modules linking internal states and social behaviour in flies and mice. *Nat Rev Neurosci* **17**, 692-704 (2016).
3. Rajasethupathy, P., Ferenczi, E. & Deisseroth, K. Targeting Neural Circuits. *Cell* **165**, 524-534 (2016).
4. Francis, M.M., Mellem, J.E. & Maricq, A.V. Bridging the gap between genes and behavior: recent advances in the electrophysiological analysis of neural function in Caenorhabditis elegans. *Trends Neurosci* **26**, 90-99 (2003).
5. Liu, P., Chen, B. & Wang, Z.-W. GABAergic motor neurons bias locomotor decision-making in C. elegans. *Nature communications* **11**, 5076 (2020).
6. Noguchi, A., Ikegaya, Y. & Matsumoto, N. In Vivo Whole-Cell Patch-Clamp Methods: Recent Technical Progress and Future Perspectives. *Sensors (Basel)* **21** (2021).
7. Akerboom, J. et al. Genetically encoded calcium indicators for multi-color neural activity imaging and combination with optogenetics. *Frontiers in molecular neuroscience* **6**, 2 (2013).
8. Chen, T.W. et al. Ultrasensitive fluorescent proteins for imaging neuronal activity. *Nature* **499**, 295-300 (2013).
9. Lin, M.Z. & Schnitzer, M.J. Genetically encoded indicators of neuronal activity. *Nat Neurosci* **19**, 1142-1153 (2016).
10. Kralj, J.M., Douglass, A.D., Hochbaum, D.R., Maclaurin, D. & Cohen, A.E. Optical recording of action potentials in mammalian neurons using a microbial rhodopsin. *Nat Methods* **9**, 90-95 (2012).
11. Hochbaum, D.R. et al. All-optical electrophysiology in mammalian neurons using engineered microbial rhodopsins. *Nat Methods* **11**, 825-833 (2014).
12. Fan, L.Z. et al. All-Optical Electrophysiology Reveals the Role of Lateral Inhibition in Sensory Processing in Cortical Layer 1. *Cell* **180**, 521-535 e518 (2020).
13. Piatkevich, K.D. et al. A robotic multidimensional directed evolution approach applied to fluorescent voltage reporters. *Nat Chem Biol* **14**, 352-360 (2018).
14. Knopfel, T. & Song, C. Optical voltage imaging in neurons: moving from technology development to practical tool. *Nat Rev Neurosci* **20**, 719-727 (2019).
15. Gong, Y., Wagner, M.J., Zhong Li, J. & Schnitzer, M.J. Imaging neural spiking in brain tissue using FRET-opsin protein voltage sensors. *Nature communications* **5**, 3674 (2014).
16. Gong, Y., Li, J.Z. & Schnitzer, M.J. Enhanced Archaerhodopsin Fluorescent Protein Voltage Indicators. *PLoS One* **8**, e66959 (2013).
17. Maclaurin, D., Venkatachalam, V., Lee, H. & Cohen, A.E. Mechanism of voltage-sensitive fluorescence in a microbial rhodopsin. *Proc Natl Acad Sci U S A* **110**, 5939-5944 (2013).
18. Flytzanis, N.C. et al. Archaerhodopsin variants with enhanced voltage-sensitive fluorescence in mammalian and Caenorhabditis elegans neurons. *Nature communications* **5**, 4894 (2014).
19. Bergs, A. et al. Rhodopsin optogenetic toolbox v2.0 for light-sensitive excitation and inhibition in Caenorhabditis elegans. *PLoS One* **13**, e0191802 (2018).
20. Azimi Hashemi, N. et al. Rhodopsin-based voltage imaging tools for use in muscles and neurons of Caenorhabditis elegans. *Proc Natl Acad Sci U S A* **116**, 17051-17060 (2019).
21. Kannan, M. et al. Fast, in vivo voltage imaging using a red fluorescent indicator. *Nat Methods* **15**, 1108-1116 (2018).
22. Streit, J. & Kleinlogel, S. Dynamic all-optical drug screening on cardiac voltage-gated ion channels. *Scientific reports* **8**, 1153 (2018).
23. Newman, J.P. et al. Optogenetic feedback control of neural activity. *eLife* **4**, e07192 (2015).
24. Klapoetke, N.C. et al. Independent optical excitation of distinct neural populations. *Nat Methods* **11**, 338-346 (2014).
25. Oda, K. et al. Crystal structure of the red light-activated channelrhodopsin Chrimson. *Nature communications* **9**, 3949 (2018).
26. Vierock, J. et al. BiPOLES is an optogenetic tool developed for bidirectional dual-color control of neurons. *Nature communications* **12**, 4527 (2021).
27. Govorunova, E.G., Sineshchekov, O.A., Janz, R., Liu, X. & Spudich, J.L. NEUROSCIENCE. Natural light-gated anion channels: A family of microbial rhodopsins for advanced optogenetics. *Science* **349**, 647-650 (2015).
28. Shtonda, B. & Avery, L. CCA-1, EGL-19 and EXP-2 currents shape action potentials in the Caenorhabditis elegans pharynx. *J Exp Biol* **208**, 2177-2190 (2005).
29. Beg, A.A. & Jorgensen, E.M. EXP-1 is an excitatory GABA-gated cation channel. *Nat Neurosci* **6**, 1145-1152 (2003).
30. McIntire, S.L., Jorgensen, E., Kaplan, J. & Horvitz, H.R. The GABAergic nervous system of Caenorhabditis elegans. *Nature* **364**, 337-341 (1993).
31. Avery, L. & Thomas, J.H. in C. elegans II. (eds. nd, D.L. Riddle, T. Blumenthal, B.J. Meyer & J.R. Priess) (Cold Spring Harbor (NY); 1997).
32. Zhang, F. et al. Multimodal fast optical interrogation of neural circuitry. *Nature* **446**, 633-639 (2007).
33. Nagel, G. et al. Light activation of channelrhodopsin-2 in excitable cells of Caenorhabditis elegans triggers rapid behavioral responses. *Curr Biol* **15**, 2279-2284 (2005).

34. Edelstein, A., Amodaj, N., Hoover, K., Vale, R. & Stuurman, N. Computer Control of Microscopes Using µManager.
(John Wiley & Sons, Inc., 2010).
35. Foellinger, O., Doerrscheidt, F. & Klittich, M. Regelungstechnik, Edn. 12th. (Vde Verlag, 2016).
36. Kalman, R.E. A New Approach to Linear Filtering and Prediction Problems. *J. Basic Eng.* **82**, 35-45 (1960).
37. Nagel, G. et al. Channelrhodopsins: directly light-gated cation channels. *Biochem Soc Trans* **33**, 863-866 (2005).
38. Zou, P. et al. Bright and fast multicoloured voltage reporters via electrochromic FRET. *Nature communications* **5**,
4625 (2014).
39. Richmond, J.E., Davis, W.S. & Jorgensen, E.M. UNC-13 is required for synaptic vesicle fusion in C. elegans. *Nat*
Neurosci **2**, 959-964 (1999).
40. Liewald, J.F. et al. Optogenetic analysis of synaptic function. *Nat Methods* **5**, 895-902 (2008).
41. Liu, P. et al. Genetic dissection of ion currents underlying all-or-none action potentials in C. elegans body-wall
muscle cells. *J Physiol* **589**, 101-117 (2011).
42. Gao, S. & Zhen, M. Action potentials drive body wall muscle contractions in Caenorhabditis elegans. *Proc Natl*
Acad Sci U S A **108**, 2557-2562 (2011).
43. Wang, H. & Sieburth, D. PKA controls calcium influx into motor neurons during a rhythmic behavior. *PLoS Genet*
9, e1003831 (2013).
44. Liu, Q., Kidd, P.B., Dobosiewicz, M. & Bargmann, C.I. C. elegans AWA Olfactory Neurons Fire Calcium-Mediated
All-or-None Action Potentials. *Cell* **175**, 57-70 e17 (2018).
45. Jiang, J. et al. C. elegans enteric motor neurons fire synchronized action potentials underlying the
defecation motor program. *bioRxiv*, 2021.2008.2025.457687 (2021).
46. Govorunova, E.G. et al. The Expanding Family of Natural Anion Channelrhodopsins Reveals Large Variations in
Kinetics, Conductance, and Spectral Sensitivity. *Scientific reports* **7**, 43358 (2017).
47. Paz, J.T. et al. Closed-loop optogenetic control of thalamus as a tool for interrupting seizures after cortical injury.
Nat Neurosci **16**, 64-70 (2013).
48. Widge, A.S., Malone, D.A., Jr. & Dougherty, D.D. Closing the Loop on Deep Brain Stimulation for Treatment-
Resistant Depression. *Frontiers in neuroscience* **12**, 175 (2018).
49. Adam, Y. et al. Voltage imaging and optogenetics reveal behaviour-dependent changes in hippocampal dynamics.
Nature **569**, 413-417 (2019).
50. Gradinaru, V. et al. Molecular and cellular approaches for diversifying and extending optogenetics. *Cell* **141**, 154-
165 (2010).
51. Mahn, M. et al. High-efficiency optogenetic silencing with soma-targeted anion-conducting channelrhodopsins.
Nature communications **9**, 4125 (2018).
52. Schindelin, J., Rueden, C., Miura, K., Hiner, M. & Tinevez, J.-Y. CorrectBleach: upgrade with Exponential fitting
method. (2016).
53. Schindelin, J. et al. Fiji: an open-source platform for biological-image analysis. *Nature Methods* **9**, 676-682 (2012).
54. Kim, P. Kalman Filter for Beginners: with MATLAB Examples, Edn. 1st. (CreateSpace Independent Publishing
Platform, 2011).
55. Gee, C.E., Ohmert, I., Wiegert, J.S. & Oertner, T.G. Preparation of Slice Cultures from Rodent Hippocampus.
Cold Spring Harbor protocols **2017** (2017).
56. Wiegert, J.S., Gee, C.E. & Oertner, T.G. Viral Vector-Based Transduction of Slice Cultures. *Cold Spring Harbor*
protocols **2017** (2017).
57. Wiegert, J.S., Gee, C.E. & Oertner, T.G. Single-Cell Electroporation of Neurons. *Cold Spring Harbor protocols*
2017 (2017).

Acknowledgements

548 We thank Jiajie Shao for support in confocal microscopy, Adam Cohen for reagents, and Sven

549 Plath, Regina Wagner, Hans-Werner Müller, Franziska Baumbach, Nico Sturman, Adrian

550 Breicher, Timotheus Kozlowski, and Kathrin Sauter for expert technical assistance.

Funding

552 Goethe University (AG); Deutsche Forschungsgemeinschaft, SFB807/B02 (AG), SPP1926-XIb /

553 WI 4485/3-2 (JSW), and SFB1315/C01 (JV, PH); Max Planck Society, PhD fellowship (ACFB);

554 Chan Zuckerberg Initiative, support (QL, CIB)

Author contributions

Conceptualization: ACFB, AG, JSW

Methodology: ACFB, AB, JV, PH, JSW

Investigation: ACFB, JFL, SRR, QL, AN, CW, NZ, HD

Visualization: ACFB, AG

Funding Acquisition: AG, ACFB, JSW, PH, CIB

Project administration: AG

Supervision: AG, JSW, CIB

Writing – original draft: ACFB, AG

Writing – review & editing: ACFB, AG, SRR, JSW, JV, CIB, PH

Competing Interests

A patent application has been filed: EP 21162331.9 (ACFB, AG).

Data and materials availability

The datasets generated during and/or analyzed during the current study are available from the corresponding author on reasonable request. This also applies to materials described in the study.

Software availability

The software / code used for operating the OVC is written in beanshell, the scripting language used by μManager freeware. Two scripts ("OVC_main_script", "OVC_on_the_run_mode") are available as a supplementary files. They will be made available upon publication under the license XYZ.

Extended data

Online Methods

Extended data Figures 1 to 8

Extended data videos 1 and 2

Online Methods

Transgenic *C. elegans* strains

The following strains were used or generated, and are available upon request: **ZX2476**: *zxEx1139[pmyo-3::QuasAr2; pmyo-2::CFP]*, **ZX2482**: *zxEx1145[pmyo-3::QuasAr2; pmyo-2::CFP]; zxIs5[punc-17::ChR2(H134R)::yfp;lin-15+]*, **ZX2483**: *zxEx1146[punc-17::ACR2::eYFP; pmyo-3::QuasAr2; pelt-2::GFP]*, **ZX2586**: *zxEx1228[punc-17::GtACR2::mCerulean::βHK::Chrimson; pelt-2::GFP]*, **ZX2714**: *zxEx1250[punc-17::GtACR2::mCerulean::βHK::Chrimson; pmyo-3::QuasAr2; pelt-2::GFP]*, **ZX2753**: *zxEx1266[pmyo-3::GtACR2::mCerulean::βHK::Chrimson; pmyo-3::QuasAr2; pmyo-2::CFP]*, **ZX2755**: *zxEx1268[punc-47::QuasAr2::GFP; pmyo-2::CFP]*, **ZX2826**: *zxEx1282[pmyo-2::QuasAr2; pmyo-2::GtACR2::mCerulean::βHK::Chrimson; pmyo-3::CFP]*, **ZX2827**: *zxEx1283[punc-17::GtACR2::mCerulean::βHK::Chrimson; punc-17::QuasAr2; pelt-2::GFP]*, **ZX2828**: *zxEx1284[punc-47::QuasAr2::GFP; punc-47::GtACR2::mCerulean::βHK:: Chrimson; pmyo-2::CFP]*, **ZX2876**: *zxIs139[pmyo-3::GtACR2::mCerulean::βHK::Chrimson; pmyo-3::QuasAr2; pmyo-2::CFP]*, **ZX2935**: *unc-13(n2813); zxIs139[pmyo-3::GtACR2::mCerulean::βHK::Chrimson; pmyo-3::QuasAr2; pmyo-2::CFP]*.

Molecular biology

Plasmids pAB4 (p*unc-17*::ACR2::eYFP), pAB16 (p*myo-3*::QuasAr; Addgene plasmid #130272), pAB17 (p*unc-17*::QuasAr), pAB23 (p*tdc-1s*::QuasAr::GFP) and pNH12 (p*myo-2*::MacQ::mCitrine) were described earlier [19, 20]. **pAB26 (p*unc-17*::GtACR2::mCerulean::βHK::Chrimson)** was generated by Gibson Assembly based on RM#348p (p*unc-17*; a gift from J. Rand) and pAAV-hSyn-BiPOLES-mCerulean (Addgene plasmid #154944), using *Nhe*I and primers 5'-attttcaggaggacccttggATGGCATCACAGGTCGTC-3' and 5'-ataccatggtaccgtcgacgTCACACTGTGTCCTCGTC-3'. **pAB27 (p*myo-3*::GtACR2::mCerulean::βHK::Chrimson)** was generated *via* Gibson Assembly based on pDD96.52 (p*myo-3*, Addgene plasmid #1608) and pAAV-hSyn-BiPOLES-mCerulean, using *Bam*HI and primers 5'-actagatccatctagagATGGCATCACAGGTCGTC-3' and 5'-ttggccaatcccgggCACTGTGTCCTCGTCCTC-3'. **pAB28 (p*unc-47*::QuasAr::GFP)** was generated by Gibson Assembly based on pMSM08 (p*unc-47*::eGFP::MmBoNTB) and pAB23 (p*tdc-1s*::QuasAr::GFP), using *Xma*I, *Msc*I and primers 5'-ttacagcaccggtggattggATGGTAAGTATCGCTCTG-3' and 5'-ttctacgaatgctcctaggcCTATTTGTATAGTTCATCCATGC-3'. **pAB29 (p*myo-2*::QuasAr)** was generated by Gibson Assembly based on pNH12 (p*myo-2*::MacQ::mCitrine) and pAB16 (p*myo-*

3::QuasAr), using primers 5'-caccgagtgaGAAGAGCAGGATCACCAG-3', 5'-tgcagagcgatacttaccatCCCCGAGGGTTAAAATGAAAAG-3', 5'-ATGGTAAGTATCGCTCTGCAG-3' and 5'-cctgctcttctcaCTCGGTGTCGCCCAGAATAG-3'. **pAB30 (p*myo-2*::GtACR2::mCerulean::bHK::Chrimson)** was generated by Gibson Assembly based on pNH12 (p*myo-2*::MacQ::mCitrine) and pAAV-hSyn-BiPOLES-mCerulean, using *Bam*HI, *Hind*III and primers 5'-ggacgaggacacagtgtgaaAAGAGCAGGATCACCAGC-3' and 5'-agacgacctgtgatgccatgCCCCGAGGGTTAAAATGAAAAG-3'. **pAB31 (p*unc-47*::GtACR2::mCerulean::bHK::Chrimson)** was made by Gibson Assembly based on pAB28 (p*unc-47*::QuasAr::GFP) and pAAV-hSyn-BiPOLES-mCerulean, using *Age*I, *Eco*RI and primers 5'-acatttatttcattacagcaATGGCATCACAGGTCGTC-3' and 5'-agcgaccggcgctcagttggTCACACTGTGTCCTCGTC-3'.

For neuronal expression, the coding sequence of QuasAr2 [49] (Addgene #107705) was cloned together with a trafficking signal (ts: KSRITSEGEYIPLDQIDINV) and an ER-export signal (ER: FCYENEV) from the Kir 2.1 channel [50, 51] into an AAV2-backbone behind a human synapsin (hSyn) promoter resulting in **pAAV-QuasAr2-ts-ER.**

OVC voltage imaging experiments

For voltage imaging experiments, animals were supplemented with ATR (Sigma-Aldrich, USA): One day prior to experiments, transgenic L4 stage animals were transferred to NGM plates, seeded with OP50 bacterial suspension mixed with ATR (stock in ethanol). To avoid interfering fluorescence of unbound ATR, its concentration was adjusted for each tissue. Final ATR concentrations (mM): BWMs (0.01), pharynx (0.03), cholinergic neurons (0.1), GABAergic neurons (0.005). Animals were immobilized with polystyrene beads (0.1 µm diameter, at 2.5 % w/v, Sigma-Aldrich) on 10 % agarose pads (in M9 buffer). Voltage-dependent fluorescence of QuasAr2 was excited with a 637 nm red laser (OBIS FP 637LX, Coherent) at 1.8 W/mm^2 and imaged at 700 nm (700/75 ET Bandpass filter, integrated in Cy5 filter cube, AHF Analysentechnik), while optogenetic actuators (BiPOLES, *Gt*ACR2 or ChR2(H134R)) were activated using a monochromator (Polychrome V, Till Photonics / Thermo Scientific), set to emit light from 400 to 600 nm at 300 µW/mm^2. Imaging was performed on an inverted microscope (Zeiss Axio Observer Z1), equipped with a 40x oil immersion objective (Zeiss EC Plan-NEOFLUAR 40x/ N.A. 1.3, Oil DIC ∞ / 0.17), a laser beam splitter (HC BS R594 lambda/2 PV flat, AHF Analysentechnik), a galilean beam expander (BE02-05-A, Thorlabs) and an EMCCD or an sCMOS camera (Evolve 512 Delta, Photometrics, or Kinetix 22, Teledyne). All OVC experiments were performed at ≈ 90 fps with 10 ms exposure and a binning of 4x4 (computer specifications: 24 GB RAM, AMD FX-8150 Octa-core processor (3.6 GHz), NVIDIA GeForce GT 520). To induce

pharyngeal pumping, animals were incubated in 3 µl serotonin (20 mM, in M9) for 3 min prior to

experiments.

OVC software

The OVC control software was written in Beanshell, the scripting language used by the open

source microscopy software µManager v. 1.4.22 [34]. Scripts are provided in supplementary

information. Experiments were initiated via the µManager script panel. For support of Polychrome

V related commands, a copy of the Java archive file *TILLPolychrome.jar* must be placed into the

MMPlugins folder and the dynamic link library file *TILLPolychromeJ.dll* into the Sys32 folder (both

provided by Till Photonics). The OVC software is compatible with all cameras supported by

µManager. Before running the software, simple rectangular ROIs (to save computation capacity,

Extended data Fig. 1B) must be selected in the live image mode by using the ROI button in the

µManager main window. Once a ROI is selected, the script can be executed via the script panel

GUI. An input tab prompts the user to select the OVC- and acquisition parameters. The software

allows to define a holding $\Delta F/F_0$ value (in %) and the number of frames for each of the steps

(three-step-protocol). One can further select the number of frames for the calibration period, the

tolerance range (in %), the algebraic sign of the increment factor that decides whether to increase

or decrease the wavelength with respect to the current $\Delta F/F_0$, the starting wavelength and the

wavelength limits (in nm). The free choice of these parameters ensures that the program can also

be used for other and/or future combinations of optogenetic tools with different spectral properties.

For acquisition, light intensity of the Polychrome V (in %), exposure time of the camera (in ms)

and binning can be selected.

During calibration, acquired grey values are stored in an array and used to evaluate the

parameters for the exponential decay function to correct for bleaching of the voltage sensor, live

during the clamping phase (**Equation 1-3**; based on the ImageJ Plugin "Correctbleach" [52, 53]). At

the last time point of the calibration phase, the first 50 bleaching corrected gray values of the

recording are used to calculate F_0. Subsequently, bleaching corrected $\Delta F/F_0$ values are calculated

at each time point of the clamping phase (**Equation 4**). Once the system has access to the

bleaching-corrected $\Delta F/F_0$ values, it feeds them into a decision tree algorithm, where they are

compared to a desired holding $\Delta F/F_0$ value. Depending on the sign of the difference, it decides,

whether to shift the wavelength blue or red (mirrored by the sign of the increment i[k]). Wavelength

adaptation occurs *via* an implemented I-controller (integral gain K_I was empirically chosen to fit

the desired system-behavior) that maps the "I" dynamics by buffering the manipulated variable

(wavelength λ) to compensate for permanent control deviation, which would otherwise occur by

using a sole P-controller (**Extended data Fig. 1A**). That means that the difference (error) between

command and current $\Delta F/F_0$ determines the wavelength change of the monochromator – hence, the current wavelength results from the previous plus the respective wavelength change at each time point (**Equation 5 and 6**). The software comes with a three-step protocol, where among others, the desired holding values (within a tolerance range) and number of frames for each step can be selected by the user (**Extended data Fig. 1C**). The system changes the wavelength only if the tolerance range is exceeded.

$$exp.offset = a * e^{-b*frame} + c \qquad \text{(Eq. 1)}$$

$$ratio = exp.offset_0 \div exp.offset_{frame} \qquad \text{(Eq. 2)}$$

$$F_{correct} = F_{frame} * ratio \qquad \text{(Eq. 3)}$$

$$\Delta F/F_{0\,correct} = ((F_{correct} - F_0) / F_0) * 100 \qquad \text{(Eq. 4)}$$

$$error = \left| \Delta F/F_{0\,holding} - \Delta F/F_0 \right| \qquad \text{(Eq. 5)}$$

$$\lambda = \lambda_{frame-1} + K_I * error * increment \qquad \text{(Eq. 6)}$$

Exp. offset: exponential offset. a, b, c: exponential fitting parameters. Ratio: ratio between first and current exponential offset. $F_{correct}$: Bleaching-corrected gray values. F_{frame}: Gray value of current frame. $\Delta F/F_{0\,correct}$: Bleaching-corrected relative change in fluorescence. F_0: Resting fluorescence value. λ: wavelength. K_I: integral gain. Equations 1-3 based on the ImageJ Plugin "Correctbleach" [52].

When the OVC acquisition is complete, a results text file is given out, and the respective image stack is saved automatically as TIFF file. The results file comprises the selected parameters, the bleach correction function and its quality, the framerate, as well as time, (bleaching corrected-) raw grey- and $\Delta F/F_0$ values, and wavelength traces. In addition, status information (system on hold, adapting, or wavelength limits reached) is recorded for each time point of the clamp phase. For the "on-the-run" mode, an additional control window opens as soon as the measurement starts. Here, the holding $\Delta F/F_0$ values can be selected live, either via a range slider or with help of the keyboard arrow keys, while a live status update is displayed.

A further software version involves a PID-controller to adjust the wavelength of the OVC. At each time point, the new wavelength is calculated by the addition of the output (u) of the PID equation to the calibration wavelength (**Equations 7-9**).

$$error = \Delta F/F_{0\,holding} - \Delta F/F_0 \qquad \text{(Eq. 7)}$$

$$u = K_P * e + K_I * T_a * e_{sum} + K_D * \frac{e - e_{t-1}}{T_a} \qquad \text{(Eq. 8)}$$

$$\lambda = \lambda_0 + u \qquad \text{(Eq. 9)}$$

u: output. K_P: proportional gain. K_I: integral gain. K_D: derivative gain. e: error. T_a: sampling rate. e_{sum}: summed up error. λ_0: calibration wavelength.

719   Parameter tuning was performed by applying the Ziegler-Nichols method to empirically achieve

720   system stability (**Extended data Fig. 2C**). Therefore, K_P was adapted until the system started to

721   oscillate at the critical proportional gain ($K_{P\ crit}$). The oscillation period was noted as T_{crit}. PID

722   coefficients were calculated according to **Equations 10 and 11**:

723   *Ziegler-Nichols parameter tuning.*

K_P	T_n	T_v
0.6 $K_{P\ crit}$	0.5 T_{crit}	0.12 T_{crit}

K_P: proportional gain. $K_{P\ crit}$: critical proportional gain. T_{crit}: period of oscillation. T_n: integration-

coefficient. T_v: derivation-coefficient [35].

$K_D = K_P * T_v$ (Eq. 10)

$K_I = K_P * \frac{1}{T_n}$ (Eq. 11)

To also achieve a better controller performance, a Kalman filter for sensor smoothing was added

(**Extended data Fig. 2A**). Here, the error covariances of the measurement- and system noise of

the Kalman filter were empirically adjusted so that the OVC controller dynamics achieved a

desired performance. After initializing the system state with an a priori estimate (variables $\hat{x}_0$:

estimated fluorescence at the beginning of the measurement and P_0: estimated error covariance

of measured fluorescence vs. estimated fluorescence at the beginning of the measurement),

prediction and correction are performed at each time step, alternately propagating the system

state in time, and correcting it with new observations (**Equations 12-17**) [54]:

1. Prediction of the fluorescence value and the error covariance:

$x^*(k + 1|k) = \mathbf{A}\hat{x}(k)$ (Eq. 12)

$\mathbf{P}^*(k + 1|k) = \mathbf{A}\mathbf{P}(k)\mathbf{A}^\top + \mathbf{Q}$ (Eq. 13)

k: discrete time-Step. A: system matrix, was here set to scalar value one (which is sufficient for

error smoothing). x^*: predicted fluorescence. P^*: predicted error covariance of the estimate. Q:

Error covariance of the system noise. $\mathrm{P}(k)$: estimated error covariance.

2. Calculation of the Kalman gain:

$\mathbf{S}(k + 1) = \mathbf{H}\mathbf{P}^*(k + 1|k)\mathbf{H}^\top + \mathbf{R}$ (Eq. 14)

$\mathbf{K}(k + 1) = \mathbf{P}^*(k + 1|k)\mathbf{H}^\top\mathbf{S}^{-1}(k + 1)$ (Eq. 15)

S: auxiliary quantity for determining the Kalman gain. H: system output matrix. R: error covariance

matrix of the measurement noise. K: Kalman gain.

3. Correction of the state estimate:

$\hat{x}(k + 1) = x^*(k + 1|k) + \mathbf{K}(k + 1)[z(k + 1) - \mathbf{H}x^*(k + 1|k)]$ (Eq. 16)

$z(k + 1)$: current measurement. $x^*(k + 1|k)$: predicted measurement. $\hat{x}(k + 1)$: current estimate.

4. Correction of the covariance estimation

$$\mathbf{P}(k+1) = [\mathbb{1} - \mathbf{K}(k+1)\mathbf{H}]\mathbf{P}^*(k+1|k) \qquad \text{(Eq. 17)}$$

The system matrices A and H were set as constant scalars (A = 1, H = 1; constant system dynamics) since this assumption is perfectly sufficient for signal smoothing for the PID controller. Here, the sampling rate is fast enough that the Kalman filter can respond to membrane potential changes occurring in this time period with reasonable estimates.

Quality assessment of the OVC

The predictive accuracy for bleaching correction had R^2 values above 0.95 in most cases, and at least above 0.8 (**Extended data Fig. 6B**). Experiments with lower R^2 values were attributable to APs occurring during the calibration phase (e.g. in pharyngeal muscle). As expected, the fraction of system saturation decreased once optogenetic tools for de- and hyperpolarization were simultaneously used, which stresses the importance of a second, opposing actuator (**Extended data Fig. 6C**). While for the single-tool approaches, the percentage of system saturation was relatively high with 16.3 ± 2.8 %, this relation was significantly reduced for all bidirectional BiPOLES combinations. This was also reflected in the speed of the system, where compared to single-tool combinations, BiPOLES was significantly faster, particularly for driving excited or hyperpolarized states back towards resting potential (**Extended data Fig. 6A**).

Electrophysiology in *C. elegans*

Electrophysiological recordings from body wall muscle cells were conducted in immobilized and dissected adult worms as described previously [40]. Animals were immobilized with Histoacryl L glue (B. Braun Surgical, Spain) and a lateral incision was made to access neuromuscular junctions (NMJs) along the anterior ventral nerve cord. The basement membrane overlying body wall muscles was enzymatically removed by 0.5 mg/ml collagenase for 10 s (C5138, Sigma-Aldrich, Germany). Integrity of BWMs and nerve cord was visually examined *via* DIC microscopy. Recordings from BWMs were acquired in whole-cell patch-clamp mode at 20-22°C using an EPC-10 amplifier equipped with Patchmaster software (HEKA, Germany).

The head stage was connected to a standard HEKA pipette holder for fire-polished borosilicate pipettes (1B100F-4, Worcester Polytechnic Institute, Worcester, MA, USA) of 4-10 MΩ resistance. The extracellular bath solution consisted of 150 mM NaCl, 5 mM KCl, 5 mM $CaCl_2$, 1 mM $MgCl_2$, 10 mM glucose, 5 mM sucrose, and 15 mM HEPES (pH 7.3 with NaOH, ≈330 mOsm). The internal/patch pipette solution consisted of K-gluconate 115 mM, KCl 25 mM, $CaCl_2$ 0.1 mM, $MgCl_2$ 5 mM, BAPTA 1 mM, HEPES 10 mM, Na_2ATP 5 mM, Na_2GTP 0.5 mM, cAMP 0.5 mM, and

cGMP 0.5 mM (pH 7.2 with KOH, ≈320 mOsm). For some experiments, BAPTA was replaced by EGTA (5 mM).

For light activation a monochromator (Polychrome V, Thermo Scientific) was used, ranging from 400 to 600 nm at 300 µW/mm^2. To create conditions as similar as possible to those for the OVC experiments, additional excitation was performed with a laser at 637 nm. Subsequent analysis and graphing were performed using Patchmaster, and Origin (Originlabs).

Electrophysiological recordings from the DVB neuron were conducted in immobilized and dissected adult worms as described previously [44] with minor modifications. Both dissection and recording were performed at room temperature. Briefly, an animal was immobilized with cyanoacrylate adhesive (Vetbond tissue adhesive; 3M) along the ventral side of the posterior body on a Sylgard 184-coated (Dow Corning) glass coverslip. A small longitudinal incision was made using a diamond dissecting blade (Type M-DL 72029 L; EMS) in the tail region along the glue line. The cuticle flap was folded back and glued to the coverslip with GLUture Topical Adhesive (Abbott Laboratories), exposing DVB. The coverslip with the dissected preparation was then placed into a custom-made open recording chamber (~1.5 ml volume) and treated with 1 mg/ml collagenase (type IV; Sigma) for ~10 s by hand pipetting. The recording chamber was subsequently perfused with the standard extracellular solution using a custom-made gravity-feed perfusion system for ~10 ml. The standard pipette solution was (all concentrations in mM): K-gluconate 115; KCl 15; KOH 10; MgCl$_2$ 5; CaCl$_2$ 0.1; Na$_2$ATP 5; NaGTP 0.5; Na-cGMP 0.5; cAMP 0.5; BAPTA 1; Hepes 10; Sucrose 50, with pH adjusted to 7.2, using KOH, osmolarity 320–330 mOsm. The standard extracellular solution was: NaCl 140; NaOH 5; KCl 5; CaCl$_2$ 2; MgCl$_2$ 5; Sucrose 15; Hepes 15; Dextrose 25, with pH adjusted to 7.3, using NaOH, osmolarity 330–340 mOsm. Liquid junction potentials were calculated and corrected before recording. Electrodes with resistance of 20–30 MΩ were made from borosilicate glass (BF100-58-10; Sutter Instruments) using a laser pipette puller (P-2000; Sutter Instruments) and fire-polished with a microforge (MF-830; Narishige). Recordings were performed using single-electrode whole-cell current-clamp (Heka, EPC-10 USB) with two-stage capacitive compensation optimized at rest, and series resistance compensated to 50 %.

Preparation of organotypic hippocampal slice cultures and transgene delivery

All procedures were in agreement with the German national animal care guidelines and approved by the independent Hamburg state authority for animal welfare (Behörde für Justiz und Verbraucherschutz). They were performed in accordance with the guidelines of the German Animal Protection Law and the animal welfare officer of the University Medical Center Hamburg-Eppendorf.

Organotypic hippocampal slices were prepared from Wistar rats of both sexes (Jackson-No. 031628) at post-natal day 5-7 as previously described [55]. In brief, dissected hippocampi were cut into 350 μm slices with a tissue chopper and placed on a porous membrane (Millicell CM, Millipore). Cultures were maintained at 37°C, 5% CO_2 in a medium containing 80% MEM (Sigma M7278), 20% heat-inactivated horse serum (Sigma H1138) supplemented with 1 mM L-glutamine, 0.00125% ascorbic acid, 0.01 mg ml^{-1} insulin, 1.44 mM $CaCl_2$, 2 mM $MgSO_4$ and 13 mM D-glucose. No antibiotics were added to the culture medium. Pre-warmed medium was replaced twice per week.

For transgene delivery, organotypic slice cultures were transduced at DIV 3-5 with a recombinant adeno-associated viral vector (rAAV) encoding a soma-targeted version of BiPOLES [26] (AAV9-CaMKII-somBiPOLES-mCerulean, Addgene #154948). The rAAV was locally injected into the CA1 region using a Picospritzer (Parker, Hannafin) by a pressurized air pulse (2 bar, 100 ms) expelling the viral suspension into the slice [56]. During virus transduction, membranes carrying the slices were kept on pre-warmed HEPES-buffered solution. At DIV 14-16, individual CA1 pyramidal cells in the previously somBiPOLES-transduced slices were transfected by single-cell electroporation [57] with the plasmid pAAV-QuasAr2-ts-ER at a final concentration of 10 ng/μl in K-gluconate-based solution consisting of (in mM): 135 K-gluconate, 10 HEPES, 4 Na_2-ATP, 0.4 Na-GTP, 4 $MgCl_2$, 3 ascorbate, 10 Na_2- phosphocreatine (pH 7.2). An Axoporator 800A (Molecular Devices) was used to deliver 50 hyperpolarizing pulses (-12 V, 0.5 ms) at 50 Hz. During electroporation slices were maintained in pre-warmed (37°C) HEPES-buffered solution (in mM): 145 NaCl, 10 HEPES, 25 D-glucose, 2.5 KCl, 1 $MgCl_2$ and 2 $CaCl_2$ (pH 7.4, sterile filtered). A plasmid encoding hSyn-mCerulean (at 50 ng/μl) was co-electroporated and served as a marker to rapidly identify cells putatively co-expressing QuasAr2 and somBiPOLES under epifluorescence excitation.

OVC voltage imaging and electrophysiology experiments in rat organotypic slices

At DIV 19-21, voltage imaging and/or whole-cell patch-clamp recordings of transfected CA1 pyramidal neurons were performed. Experiments were done at room temperature (21-23°C) using a BX51WI upright microscope (Olympus) equipped with an EMCCD camera (Evolve 512 Delta, Photometrics), dodt-gradient contrast and a Double IPA integrated patch amplifier controlled with SutterPatch software (Sutter Instrument). QuasAr2 was excited with a 637 nm red laser (OBIS FP 637LX, Coherent) at 0.52 W/mm^2 via a dichroic mirror (660nm, Edmund Optics) and voltage-dependent fluorescence was detected through the objective LUMPLFLN 60XW (Olympus) and through an emission filter (732/68 nm Bandpass filter, Edmund Optics). somBiPOLES was activated using a monochromator (Polychrome V, Till Photonics / Thermo Scientific), set to emit

light from 400 to 600 nm at 1.7 mW/mm^2. Laser and monochromator light were combined using

a 605nm dichroic mirror (Edmund Optics). Irradiance was measured in the object plane with a

PM400 power meter equipped with a S130C photodiode sensor (Thorlabs) and divided by the

illuminated field of the Olympus LUMPLFLN 60XW objective (0.134 mm^2). As for *C.elegans*, OVC

experiments were performed at ≈ 90 fps with 10 ms exposure and a binning of 4x4.

For electrophysiology, patch pipettes with a tip resistance of 3-4 MOhm were filled with

intracellular solution consisting of (in mM): 135 K-gluconate, 4 MgCl$_2$, 4 Na$_2$-ATP, 0.4 Na-GTP,

10 Na$_2$-phosphocreatine, 3 ascorbate, 0.2 EGTA, and 10 HEPES (pH 7.2). Artificial cerebrospinal

fluid (ACSF) consisted of (in mM): 135 NaCl, 2.5 KCl, 2 CaCl$_2$, 1 MgCl$_2$, 10 Na-HEPES, 12.5 D-

glucose, 1.25 NaH$_2$PO$_4$ (pH 7.4). Synaptic transmission was blocked by adding 10 µM CPPene,

10 µM NBQX, and 100 µM picrotoxin (Tocris) to the extracellular solution. Measurements were

corrected for a liquid junction potential of -14,5 mV. Access resistance of the recorded neurons

was continuously monitored and recordings above 30 MΩ were discarded.

Statistical methods

Statistical methods used are described in the figure legends. No sample was measured multiple

times, and no samples were excluded from analysis. Statistical analysis was performed in Prism

(Version 8.0.2, GraphPad Software, Inc.), OriginPro 2021 (OriginLab Corporation, Northampton,

USA), Microsoft Excel 2016, 2019, SutterPatch V2 (Sutter Instrument) MATLAB 2016b, 2019b

(Mathworks), or ImageJ v1.53c. No statistical methods were applied to predetermine sample size.

However, sample sizes reported here are consistent to data presented in previous publications,

and which typically allow distinction of significant differences in the types of experiments used.

Data was tested for normality prior to statistical inference. Data is given as means ± SEM, if not

otherwise stated. Significance between data sets after paired or two-sided t-test or ANOVA, with

Bonferroni correction for multiple comparisons, is given as p-value (* $p \leq 0.05$; ** $p \leq 0.01$; *** $p \leq$

0.001), when not otherwise stated. Group data are presented as box plots with median (line), 25-

75th quartiles (box), mean (open dot) and whiskers, representing 1.5x inter quartile range (IQR).

Single data points are shown for n≤10. The R^2 coefficient of determination was determined as: R^2

= 1 - SSE/SSD (SSE = sum of the squared errors; SSD = sum of the squared deviations about

the mean).

Extended data

Extended data Figure 1. Setup of the OVC software and control parameters. (A) Wiring diagram of I-controller: I-controller generates a digital control value u[k] based on error e[k] for the monochromator, which in turn generates analog light output u[t]. BiPOLES modulates membrane voltage y(t) upon analog light activation that in turn defines QuasAr2's fluorescence z(t) (transduced analog actual value). Fluorescence is monitored via camera and $\Delta F/F_0$ is calculated (digital actual value). w[k]: Set point. **(B)** Rectangular ROI selection in µManager "live mode" prior to OVC measurement. **(C)** Input tab of the standard OVC software, implemented in µManager. **(D)** Bleaching behavior and exponential fits of the fluorescence of QuasAr2 in BWMs, in control animals expressing no actuators. Animals were illuminated with the 637 nm laser for the entire duration, and in addition, with 470 nm (300 µW/mm²) for the last 20 s (n = 8). Fits were performed for the first (red) or second 20 sec period (blue), or for the entire duration (grey). Lower panel: Statistical analysis of the deduced fit parameters (n = 9). **(E)** Exponential fits of QuasAr2 fluorescence as in (D) but illuminated with either 637 nm laser or 637 nm laser and 470 (300 µW/mm²) for the entire duration. Middle and right

panels: Statistical analysis of the fit parameters of data in left panel (n = 11-12). **(F)** Average data for animals expressing
only QuasAr2 in BWMs, while the OVC attempts to run a 0, -5, 5 % $\Delta F/F_0$ protocol (n = 8). Upper panel: Monochromator
wavelength; lower panel: mean fluorescence traces. Statistically significant differences analyzed by One-way ANOVA
or two-sided t-test with Bonferroni correction (in D, E). Box plots (median, 25-75th quartiles); mean: open dot; whiskers:
1.5x inter-quartile range (IQR).

Extended data Figure 2. Implementation of a PID controller with Kalman filter. (A) Wiring diagram of PID-controller
with Kalman filter. Signal flow as in Fig. S1A with the addition of a Kalman filter. **(B)** Comparison of single OVC
experiments with PID control but without Kalman filter. **(C)** Ziegler-Nichols parameter tuning. **(D)** OVC three-step
protocol (0, -5 and 5 % $\Delta F/F_0$) in BWMs, insets for close-up. Wavelength shown in the respective color, holding values
are indicated for each step, yellow trace represents data processed with Kalman filter for sensor smoothing. Orange
shade in left panel: transition period to reach tolerance range. **(E)** Upper panel: Overlay of mean (± S.E.M.) wavelength
and (lower panel) fluorescence traces (n=8; holding values: 0, -5, 5 % $\Delta F/F_0$). **(F)** Times required for the indicated 5
and 10 % $\Delta F/F_0$ transitions for PID- (plus Kalman filter) and I-controller (n=8-22). Statistically significant differences
analyzed by Two-sided t-test with Bonferroni correction. Box plots (median, 25-75th quartiles); mean: open dot;
whiskers: 1.5x inter-quartile range (IQR).

Extended data Figure 3. Unidirectional steering of membrane voltage. (A) Cholinergic neurons express either ChR2 or *Gt*ACR2, while QuasAr2 is expressed in BWMs. **(B)** Three-step protocol in ChR2 animals. Left: Raw and bleaching corrected data of calibration (red: exponential fit) and clamping phase. Right panels: Corresponding wavelength and $\Delta F/F_0$ traces. Holding values 15, 20 and 25 % $\Delta F/F_0$; grey shade: tolerance range for each step. **(C)** Three-step (-5, -10 and -15 % $\Delta F/F_0$) protocol in *Gt*ACR2 animals, as in (B). **(D)** Single traces for both strains, 0 %, then -15 to 20 % $\Delta F/F_0$ (in 5% increments), and return to baseline. **(E)** Mean (± S.E.M.) traces (n = 17); holding values: 0, 5, 0 % $\Delta F/F_0$, in ChR2 animals. **(F)** As in E, for *Gt*ACR2 animals (n = 14; 0, -5, 0 % $\Delta F/F_0$). **(G)** Transition time, respective 5 % $\Delta F/F_0$ steps; box plots (median, 25-75th quartiles); mean: open dot; whiskers: 1.5x inter-quartile range (IQR).

Extended data Figure 4. Testing different combinations of depolarizing and hyperpolarizing actuators. (A-E)
Body length measurements to test functionality of optogenetic actuator combinations, expressed in cholinergic neurons.
(A) NpHR and ChR2 (n = 8, monochromator wavelength ramp from 400 to 600 nm, 300 µW/mm²). (B) *Gt*ACR1 and
ChR2 (n = 10). (C) ChR2-only (n = 8). (D) *Gt*ACR1 and ChR2. Upper panel: 5 s light pulses, wavelength as indicated,
300 µW/mm² (n = 5-7). Lower panel: Mean (± S.E.M.) body length changes. Statistically significant differences were
analyzed by one-way ANOVA with Bonferroni correction (***P ≤ 0.001, **P ≤ 0.01, *P ≤ 0.05). (E) Upper panel: Body
length measurements of animals expressing BiPOLES in cholinergic neurons (mean ± S.E.M., wavelength ramp 400 -

600 nm, n = 8). Below: Representative still images of animals for each phase of the experiment: (a) before light, (b) *Gt*ACR2 effect / muscle relaxation, (c) transition from hyper- to depolarization, (d) Chrimson effect / muscle contraction. **(F)** Left panel: OVC experiment in BWMs (BiPOLES and QuasAr2) with clamp interruption (protocol: 1) holding value - 5 %, wavelength determined by OVC, 2) interruption at calibration wavelength, 3) holding value - 5 % $\Delta F/F_0$). Middle panel: Close-up, with comparison to voltage fluorescence activity of unstimulated animals expressing only QuasAr2. Right panel: Group data of $\Delta F/F_0$ fluorescence amplitudes of typical, action potential-based fluctuations during -5 % clamping and no OVC action, as well as in animals expressing only QuasAr (n=30). Box plots (median, 25-75th quartiles); mean: open dot; whiskers: 1.5x inter-quartile range (IQR).

Extended data Figure 5. BiPOLES activation by 637 nm laser light and calibration wavelength have no adverse effects on muscle function and locomotion. (A) Normalized absorption spectra of Chrimson and QuasAr2 [25, 38], laser wavelength used for QuasAr2 imaging is indicated by a red line. **(B)** Mean $\Delta F/F_0$ QuasAr2 signal (n=6) in BWMs (co-expressed with BiPOLES) in the presence of laser light (637 nm) and additional signal in response to Chrimson excitation light (590 nm). **(C)** Body length measurements of animals expressing BiPOLES and Quasar in BWMs (mean ± S.E.M., light pulses: 637 nm laser and 637 nm laser plus calibration wavelength (ca. 521 nm), n = 14). **(D)** Statistical analysis of data in (C). **(E, F)** Analysis of swimming activity of animals expressing BiPOLES and Quasar in BWMs (E, n = 9-20) or cholinergic neurons (F, n = 10-18). Illumination parameters as in (D), or in the absence of all-*trans* retinal (ATR). In (D, E, and F): Statistically significant differences were analyzed by two-sided t-test or One-way ANOVA with Bonferroni correction (***P ≤ 0.001, **P ≤ 0.01, *P ≤ 0.05). In (B, D, E, and F): Box plots (median, 25-75th quartiles); mean: open dot; whiskers: 1.5x inter-quartile range (IQR). **(G)** Two examples of long-term OVC experiments (BWMs: BiPOLES and QuasAr2, holding values 5 and -5 % $\Delta F/F_0$).

Extended data Figure 6. Performance assessment of the OVC and software version for live voltage adjustment.
(A) Summary of OVC speed, reflected by the time needed for a 5 % $\Delta F/F_0$ step for different configurations (single tools, same-cell approach; n=10-22). **(B)** Quality of the exponential fit for bleach correction (coefficient of determination, R^2), for the different QuasAr2 / optogenetic actuator combinations, as indicated (n=7-22). **(C)** Percentages of the clamping status (hold, adapting, system saturation), for the different QuasAr2 / optogenetic actuator combinations, as indicated (n=7-22). **(D-H)** Time-varying OVC live control. (D) User interface for software version allowing live control of membrane voltage fluorescence. (E) Scheme: Holding values can be selected using arrow keys. Live status (system on hold, adapting or exceeding limits) is shown, enabling adjustment. (F-H) Example traces of wavelength (upper panels) and holding values as % $\Delta F/F_0$ (lower panels) in BWMs (F) cholinergic (G) and GABAergic neurons (H). Statistically significant differences were analyzed by one-way ANOVA with Bonferroni correction (***P ≤ 0.001, **P ≤ 0.01, *P ≤ 0.05). Box plots (median, 25-75th quartiles); mean: open dot; whiskers: 1.5x inter-quartile range (IQR).

Extended data Figure 7. Calibrating voltage and fluorescence, and normal membrane voltage behavior in BiPOLES-activated BWMs. (A-C) Calibration of fluorescence (upper panels, n=9) and voltage in BWMs (lower panels, n = 8), corresponding to wavelength ramps from 600-400 nm (B) or 600-400 nm (C), with simultaneous 637 nm laser excitation. Changes in QuasAr2 fluorescence ($\Delta F/F_0$) shown as mean (± S.E.M.), and membrane potential was recorded in whole-cell patch-clamp mode. Single voltage traces are presented in (A). **(D)** Mean relative fluorescence change, or membrane potential (between 401-410 and 590-600 nm), upper and lower panels, respectively. **(E)** Original record of simultaneous voltage and fluorescence measurement during calibration and clamp phase (3 step OVC protocol, -3, +3 % $\Delta F/F_0$). Note the fluorescence trace was subsequently bleaching-corrected for the calibration phase. **(F)** APs in simultaneous patch-clamp and fluorescence recordings during OVC calibration phase (637 nm laser and 521 nm calibration wavelength). **(G)** Statistical analysis of AP amplitude. N = 4 recordings, n = 24-27 APs. **(H)** Membrane potential in muscle was measured by patch-clamp under the indicated light conditions. **(I, J)** Analysis of small, subthreshold voltage fluctuations observed during patch-clamp, without or with 637 nm laser and calibration wavelength. Differences (arrows) of actual peaks to shifted moving average (blue), as a proxy for base line (I), were statistically analyzed in (J). n=14 recordings, with ca. 55.000 single data points per measurement. **(K)** Frequency distribution of distinct voltage signals, incl. APs, observed during patch-clamp recordings. Mean number of events

(+S.E.M.) per second is shown as a function of peak amplitudes (in mV), for the different illumination conditions, as
indicated below. N = 14, n = 3118 peaks. None of the respective amplitude distributions showed significant differences
to any other bins. Statistically significant differences were analyzed by one-way ANOVA with Bonferroni correction (in
H, J) or paired, two-sided t-test (in I). Box plots (median, 25-75th quartiles); mean: open dot; whiskers: 1.5x inter-quartile
range (IQR).

Extended data Figure 8. Voltage imaging and OVC measurements in pharyngeal muscle and the motor neuron
DVB. (A) Voltage imaging and analysis of grinder opening in animals expressing QuasAr2 and BiPOLES in pharyngeal
muscle, stimulated with consecutive 400 nm light pulses (300 µW/mm^2). Corresponding open or closed state of the
grinder as indicated. **(B)** Transition time required by the OVC in the pharynx to execute 5 and 10 % ΔF/F$_0$ steps. **(C)**
Mean (± S.E.M.) traces (n = 16) of OVC protocol in pharyngeal muscle (0, -5, 5 % ΔF/F$_0$). **(D)** "On-the-run"- experiment
of OVC in pharyngeal muscle. **(E)** Example DVB voltage fluorescence (lower panel) and wavelength (upper panel)
traces upon suppression of an AP by the OVC. Inset: Close-up and overlay of spontaneous (light gray trace) and
clamped (black trace) DVB voltage signal. Monochromator wavelength is presented in the respective color. Voltage
signal and movement artefact highlighted in blue and grey, respectively.

Extended data video 1. *C. elegans* **expressing BiPOLES in cholinergic motor neurons, during a wavelength ramp from 400-600 nm.** Animal crawling on solid substrate, with inactive (relaxed) muscle, transition to normal movement, followed by activated (contracted) muscle and respective paralysis, as indicated.

Extended data video 2: 'On the run' mode of the OVC, enabling adjusting membrane voltage during a running acquisition. Left panels, upper: Monochromator wavelength, lower: voltage fluorescence trace. Arrow keys indicated the live settings chosen by the experimenter.