Microfluidics-free single-cell genomics with templated emulsification

Iain C. Clark[1], Kristina M. Fontanez[2], Robert H. Meltzer[2], Yi Xue[2], Corey Hayford[2], Aaron

May-Zhang[2], Chris D'Amato[2], Ahmad Osman[2], Jesse Q. Zhang[2], Pabodha Hettige[2], Jacob S.A.

Ishibashi[2], Cyrille L. Delley[3], Daniel W. Weisgerber[3], Joseph M. Replogle[4], Marco Jost[4,5], Kiet

T. Phong[6], Vanessa E. Kennedy[7], Cheryl A. C. Peretz[8], Esther A. Kim[9], Siyou Song[9], William

Karlon[10], Jonathan S. Weissman[4,11], Catherine C. Smith[7], Zev J. Gartner[6,12], Adam R. Abate[3,*]

[1]Department of Bioengineering, University of California, Berkeley, California Institute for

Quantitative Biosciences, Berkeley, California, USA.

[2]Fluent Biosciences, Watertown, MA, USA.

[3]Department of Bioengineering and Therapeutic Sciences, University of California, San

Francisco, San Francisco, California, USA.

[4]Whitehead Institute for Biomedical Research, Massachusetts Institute of Technology;

Cambridge, MA, USA.

[5]Present address: Department of Microbiology, Harvard Medical School, Boston, Massachusetts,

USA.

[6]Department of Pharmaceutical Chemistry, University of California San Francisco, San

Francisco, CA, USA.

[7]Department of Medicine, University of California San Francisco, San Francisco, CA, USA.

[8]Department of Pediatrics, University of California San Francisco, San Francisco, CA, USA.

[9]Division of Plastic and Reconstructive Surgery, University of California San Francisco, San

Francisco, CA, USA.

[10]Departments of Pathology and Laboratory Medicine, University of California San Francisco,

San Francisco, CA, USA.

[11]Howard Hughes Medical Institute, Massachusetts Institute of Technology, Cambridge, MA

02142, USA.

[12]Chan Zuckerberg Biohub, San Francisco, CA, USA.

[*]Corresponding author: Adam R. Abate, adam@abatelab.org

Abstract

Single-cell RNA sequencing is now a standard method used to reveal the molecular details of

cellular heterogeneity, but current approaches have limitations on speed, scale, and ease of use

that stem from the complex microfluidic devices or fluid handling steps required for sample

processing. We, therefore, developed a method that does not require specialized microfluidic

devices, expertise, or hardware. Our approach is based on particle-templated emulsification,

which allows single-cell encapsulation and barcoding of cDNA in uniform droplet emulsions

with only a vortexer. PIP-seq accommodates a wide range of emulsification formats, including

microwell plates and large-volume conical tubes, enabling thousands of samples or millions of

cells to be processed in minutes. We demonstrate that PIP-seq produces high-purity

transcriptomes in mouse-human mixing studies, is compatible with multi-omics measurements,

and can accurately characterize cell types in human breast tissue when compared to a

commercial microfluidic platform. Single-cell transcriptional profiling of mixed phenotype acute

leukemia using PIP-seq revealed the emergence of heterogeneity within chemotherapy-resistant

cell subsets that were hidden by standard immunophenotyping. PIP-seq is a simple, flexible, and

scalable next-generation workflow that extends single-cell sequencing to new applications,

including screening, diagnostics, and disease monitoring.

Introduction

Single-cell RNA sequencing (scRNA-seq) is an essential technology in the biological sciences because it reveals how the properties of tissues arise from the transcriptional states of numerous interacting cells. Defining the gene expression signatures of individual cells allows cell type classification, the discovery of unique cell states during development and disease, and the prediction of regulatory mechanisms that control these states. As a result, bulk sequencing is being rapidly replaced by single-cell methods. The first single-cell approaches isolated cells and prepared them individually for sequencing[1-4]. While improvements in molecular biology increased data quality[5,6], the requisite isolation and processing of separate cells ultimately limited throughput. Implementation of valve-based microfluidics reduced hands-on time[7], but failed to significantly increase cell number, and thus could not capture the heterogeneity intrinsic to most tissues. Advances in high throughput droplet microfluidic barcoding have expanded single-cell sequencing to tens of thousands of cells[8,9] and fueled biological discovery, but require expensive instruments located in core facilities and therefore remain inaccessible to many labs. Methods for direct combinatorial indexing of cells[10,11], the use of nanowell arrays [12], or sample multiplexing[13,14] have overcome some limitations of microfluidics, but no current method simultaneously accommodates both low (10) and high ($>10^6$) cell numbers, can be applied to hundreds of independent samples and can be rapidly implemented without custom equipment.

The scalability of single-cell methods is important for many applications, including tissue atlas projects[15-18], million cell perturbation experiments[19], drug development pipelines[20], and developmental studies[21]. Droplet microfluidics has an intrinsic disadvantage at high cell numbers due to the upper limit on drop generation speed. At high fluid velocities, droplet generation becomes uncontrolled, resulting in polydispersed emulsions and poor bead loading that reduces

single-cell data quality[22,23]. Therefore, to sequence millions of cells requires long run times,

parallel droplet generators with complex designs that are prone to clogging, or implementation of

additional barcoding steps before encapsulation[24]. More generally, droplet microfluidics relies on

an expensive instrument usually located in a core facility, which necessitates sample transport or

fixation that can alter RNA profiles. Centralized processing also reduces access to many labs and

does not fit experiments that need rapid or point-of-collection sample handling, like remote

fieldwork or studies using infectious samples requiring biosafety precautions[12,25].

Much effort has thus gone into developing microfluidic-free single-cell methods. Split-

pool ligation[10,11] and tagmentation[26,27] perform direct combinatorial barcoding of bulk

suspensions, and significantly increase cell number; however, these laborious workflows require

enormous numbers of pipetting operations and are poorly suited for low cell inputs. Moreover,

while scalable, these methods require substantial expertise[28], and broad adoption of split-pool

barcoding will likely require robotic automation in a centralized facility. Alternatively, methods

based on nanowells prioritize simplicity and cost-effectiveness[12,25]. No microfluidics are required

and wells are loaded by sedimentation, providing an instrument-free and point-of-use solution.

However, nanowell array chips do not efficiently scale in cell or sample number; the planar

arrays capture cells on a 2D surface and, thus, cannot compete with emulsions or combinatorial

indexing utilizing a 3D volume that easily scales to millions of cells. Moreover, unless combined

with multiplexing[13,14], nanowell chips are poorly suited for processing many separate samples

because they require one array per sample, and thus hundreds of arrays for hundreds of samples.

To advance the field of single-cell genomics, next-generation technologies must simultaneously

innovate on speed, scale, and ease of use. An ideal system would be compatible with the

barcoding of separate samples in well plates, accommodate orders-of-magnitude differences in

cell number, be completed in minutes, and be easy to run at the bench or in the field without

specialized instrumentation. An approach that achieved these innovations simultaneously would

significantly enhance the usefulness of single-cell genomics, impact basic and clinical research,

and facilitate new diagnostics.

Here we describe a flexible, scalable, and instrument-free scRNA-seq method based on

rapid templated emulsification of cells and barcoded hydrogel templates without microfluidics[29].

In contrast to microfluidic emulsification, in which droplets are created sequentially and thus

their number scales with instrument run time, templated emulsification generates monodispersed

droplets in parallel by bulk self-assembly and thus the number of droplets (and cells that can be

barcoded) scales only with container volume. The result is an extremely scalable, user-friendly

scRNA-seq method that we call PIP-seq (**P**article-templated **I**nstant **P**artition-seq). Templated

emulsification produces drops that are equivalent to those generated with microfluidics and

compatible with the latest innovations in multi-omic measurements. Here, we show that PIP-seq

generates accurate single-cell gene expression profiles from human tissues and is compatible

with multi-modal measurements of RNA and sgRNA (CROP-seq), or RNA and protein (CITE-

seq). Finally, we demonstrate the use of PIP-seq to monitor the response of patients with mixed

phenotype acute leukemia (MPAL) to chemotherapy, revealing heterogeneity within cells with

similar immunophenotypes. In summary, PIP-seq fills an unmet technical need by improving the

speed, scalability, and ease of use of single-cell sequencing.

Results

Overview of the technology

PIP-seq uses particle templating to compartmentalize cells, barcoded hydrogel templates, and

lysis reagents in monodispersed water-in-oil droplets (**Fig. 1a**). Rapid emulsification with a

standard vortexer allows cells to be encapsulated at the bench or point of collection in minutes.

The cells are lysed by increasing the temperature to 65°C, which activates Proteinase K,

releasing cellular mRNA that is captured on polyacrylamide beads decorated with barcoded

polyT sequences (**Fig. 1b**). PIP-seq emulsions can be stored for days at 0°C without change in

data quality **(Extended data Fig. 1)**, allowing samples to be banked for future processing. Upon

resuming, oil is removed, beads are transferred into a reverse transcription buffer, and full-length

cDNA is synthesized, amplified, and prepared for sequencing (**Fig. 1c-d**).

A unique and valuable feature of PIP-seq is that cell encapsulation in droplets is

performed in parallel using bead size to control droplet volume. In contrast to microfluidics, the

number of droplets scales with total container volume, not emulsification time. For example, at a

6% collision rate that includes cell doublets and barcode reuse, we estimate that 3.5k cells can be

barcoded with 35 µL of barcoded hydrogel templates in a 500 µL tube, 225k cells can be

barcoded with 2 mL of barcoded hydrogel templates in a 15mL conical tube, and 1M cells can be

barcoded with 10mL of barcoded hydrogel templates in a 50mL conical tube (**Fig. 1e**).

Regardless the tube size, only two minutes of vortexing is required for cell capture. PIP-seq is

equally scalable to large sample numbers. Encapsulation can be performed directly in 96, 384, or

1536 well plates (**Fig. 1f, Extended data Fig. 2**), greatly simplifying experiments testing

hundreds of different conditions and streamlining integration with robotic handling systems.

Thus, compared to current single-cell RNA-sequencing technologies, PIP-seq has the greatest

flexibility to cover combinations of cell and sample numbers (**Fig. 1g**).

Single-cell RNA-sequencing with particle-templated emulsification

High-throughput single-cell sequencing requires efficient cell lysis and reverse transcription of

mRNA using barcoded primers. In the absence of microfluidics, barcoded hydrogel templates,

cells, and lysis reagents must be combined before emulsification. To prevent cell lysis before

compartmentalization, we use Proteinase K (PK), a protease that has minimal activity at 4°C but

can be activated at higher temperatures. After emulsification, the sample is heated to efficiently

lyse cells. To illustrate this process, we stained cells with calcein, performed templated

emulsification at 4°C with PK, and imaged the droplets before and after thermal activation. Intact

cells appeared as compact puncta before lysis, but rapidly released calcein into the bulk of the

droplets after the temperature was increased (**Fig. 2a, Extended data Fig. 2a,b**). Thus, cells can

be mixed with PK in bulk before emulsification, and thermal activation triggers the release of

mRNA for barcoding after emulsification.

To ensure that temperature-activated lysis and bulk agitation do not pre-lyse cells and

result in mRNA cross-contamination, we performed mouse-human cell line mixing studies. We

synthesized barcoded polyacrylamide beads with polyT sequences, using split-pool ligation of

four 6 bp randomers[30]. Beads contained ~10^8 (96^4) unique barcodes, providing ample sequence

space to label a million cells. PIP-seq barcode rank plots for mixed mouse-human cell

suspensions allowed cell identification by UMI abundance (**Fig. 2b**). The fraction of mouse

reads in human transcriptomes was below 3%, and transcriptomes containing multiple cells were

rare and consistent with Poisson encapsulation of two cells **(Fig. 2c,d)**. These results illustrate

that PIP-seq yields high-purity single-cell RNA-seq data with minimal transcriptome mixing and

low doublet formation.

Accurate and scalable reconstruction of single-cell phenotypes in complex tissue

An important application of single-cell sequencing is atlasing cell types in heterogeneous tissue.

To investigate the feasibility of atlasing studies, we applied PIP-seq to samples derived from

healthy breast tissue. In tandem, we performed scRNA-seq on tissues from the same patients

using a commercially available scRNA-seq technology (10x Genomics, Chromium v3). We

integrated PIP-seq data across participants and recovered expected cell types by dimensionality

reduction, including the two lineages of luminal epithelial cells (LEP1 and LEP2), myoepithelial

cells, fibroblasts, vascular cells, and immune cells **(Fig. 3a, Extended data Fig. 3a,b)**[31]. To

compare transcriptome capture between platforms, we downsampled the 10x Chromium and

PIP-seq datasets to an equivalent number of cells and reads (2400 cells and 36,500 reads per

cell). Chromium detected more unique genes (2298 *vs*. 1757, median) and transcripts (7491 *vs*.

3394) per cell, with similar percentages of reads assigned to mitochondrial transcripts (2.34% *vs*.

1.32%) **(Extended data Fig. 3c)**. To compare the transcriptome accuracy of PIP-seq, we

downsampled each dataset to an equivalent number of unique molecular identifiers (UMIs) per

cell (2400 cells, 1500 UMIs), integrated the data, performed dimensionality reduction, and

identified clusters **(Fig. 3b,c)**. We compared marker genes and the correlation between gene

expression profiles by cluster. Predicted marker genes were concordant between methods **(Fig.**

3d), gene expression was highly correlated **(Fig. 3e, Extended data Fig. 4a)**, and breast tissue

markers from previous reports were segregated identically within integrated clusters **(Extended**

data Fig. 4b). Comparison of PIP-seq to publicly available data from 10x (v3, v2) and

previously published scRNA-seq workflows demonstrated that PIP-seq produced high-quality

transcriptomes across a range of sequencing depths **(Extended data Fig. 5)**. Next, we validated

the scalability of PIP-seq, capturing and scRNA sequencing 138,146 breast tissue cells in a

single tube reaction, as well as 65,000 peripheral blood mononuclear cells (PBMC) **(Extended**

data Fig. 6a-c). At high cell numbers, we identified a population of CD34 hematopoietic

stem/progenitor cells in the PBMC sample, highlighting the importance of scalability in

detecting rare cell types **(Extended data Fig 6b,c).** Last, we validated that PIP-seq is compatible

with antibody-based cell hashing **(Extended data Fig 6d,e)**. Hashing can be used to further

increase the number of cells and conditions processed**.** Thus, PIP-seq is an easy-to-use, accurate

and scalable method to profile complex tissues.

PIP-seq for single-cell pooled CRISPR screens

CRISPR perturbations combined with single-cell sequencing allow unbiased discovery of

genotype-phenotype relationships[32-34]. Expanding this approach to genome-wide sgRNA

libraries can elucidate gene function on an unprecedented scale. However, such studies require

sequencing millions of cells to characterize all perturbations in libraries with tens or hundreds of

thousands of individual sgRNA[19]. To demonstrate how the throughput of PIP-seq enables

perturbation studies at scale, we profiled the transcriptional changes associated with a CRISPR

interference allelic series CROP-seq library[35]. This library expressed sgRNA and a

polyadenylated copy of the guide sequence from separate promoters. Guide RNAs were captured

and barcoded with the cell's polyadenylated mRNA, making this approach immediately

compatible with PIP-seq. The library is designed to quantitatively titrate gene expression using

sgRNAs with target site mismatches[35], allowing us to compare measured gene expression to

expected knockdown efficiency across each gene's allelic series **(Fig. 4a).** We transduced K562

cells containing a stable dCas9-KRAB with the CRISPRi lentiviral library and performed PIP-

seq to capture the transcriptional profiles and sgRNA identity of individual cells **(Fig. 4b-c)**. For

cells with single gRNA assignments, previously reported knockdown efficiencies[35] correlated

with the normalized counts of targeted genes **(Fig. 4d)** and were most significant for highly

expressed genes **(Extended data Fig. 7a,b).** In addition, the knockdown of genes produced

known transcriptional changes. For example, gRNA targeting HSPA5 resulted in endoplasmic

reticulum stress and increased unfolded protein response **(Fig. 4e).** These results validate the use

of PIP-seq for CROP-seq experiments, paving the way for routine million-cell experiments that

map genotype-phenotype relationships at genome-scale.

PIP-seq identifies unique single-cell transcriptomic signatures correlated with the relapse

of mixed phenotype acute leukemia (MPAL)

Monitoring of cancer in response to therapy is an emerging application of single-cell sequencing

that benefits from rapid sample processing at the point of collection, and the ability to delay

cDNA synthesis and library preparation until multiple samples have been collected. We

investigated the utility of PIP-seq for understanding cancer dynamics by first validating the

single-cell transcriptional responses of two cancer cell lines (H1975, PC9) to Gefitinib, an

epidermal growth factor receptor (EGFR) tyrosine kinase inhibitor. We treated H1975 and PC9

with DMSO (vehicle control) or 1 μM Gefitinib overnight and performed PIP-seq **(Fig. 5a)**. A

transcriptional response in H1975, which is resistant to Gefitinib due to EGFR mutations L858R

and T790M, was not observed, while Gefitinib sensitive PC9 cells showed a substantial shift in

gene expression **(Fig. 5b).** Differential gene expression analysis revealed increased levels of

TACSTD2 (*TROP2*) in PC9 cells, consistent with its known modulation during lung

adenocarcinoma tumor growth[36] **(Fig 5c)**, and decreased expression of *CDK4*, which is known to

enhance sensitivity to EGFR inhibitors[37] (**Extended data Fig. 8**). In addition, drug-resistant

H1975 cells spiked into a background of sensitive cells (1:9 H1975:PC9) could be detected

solely by their single-cell phenotypes, and at roughly the expected frequency (4.7%) **(Fig. 5d)**.

Thus, PIP-seq recovered genes with reported roles in lung-cancer drug resistance and could

identify resistance phenotypes within a background of drug-sensitive cells.

Next, we applied PIP-seq to study mixed phenotype acute leukemia (MPAL), a high-risk

disease characterized by multiple hematopoietic lineages[38,39]. Recurrence and changes in

immunophenotype with chemotherapy are typically monitored using flow cytometry of surface

markers during diagnosis, treatment, and relapse, but this provides limited insight into the drivers

of relapse after drug treatment. Like other single-cell RNA-seq methods, PIP-seq can be

multiplexed to simultaneously characterize single-cell gene expression and surface

immunophenotype[40]. Using PIP-seq, we performed antibody-derived tag (ADT) sequencing

(CITE-seq) on longitudinal samples collected from MPAL patients treated with chemotherapy.

PIP-seq confirmed the diagnosis of these samples as B/myeloid MPAL and identified aberrant

expression of immune and stem cell markers that matched with clinical immunophenotypes

determined by flow cytometry **(Supplemental Table 2, Extended data Fig. 9)**. However, PIP-

seq revealed an additional layer of complexity undetectable by traditional immunophenotyping.

Dimensionality reduction identified new cell clusters that emerged after drug treatment **(Fig.**

5e,f, Extended data Figs. 10, 11). These clusters had similar immunophenotypes **(Fig. 5g,h)**,

but contained significant transcriptional heterogeneity **(Fig. 5i,k, Supplemental Tables 4,5)**.

Cell populations up-regulating genes and pathways (oxidative phosphorylation, G2M checkpoint

modulation, ribosome biogenesis) implicated in a variety of cancers including acute

lymphoblastic leukemia[41-50], but not previously linked to MPAL, were observed **(Fig. 5j,l)**.

Taken together, our results highlight the value of single-cell methodologies for studying the

heterogeneous response of cancer subpopulations to chemotherapy and the potential for the

integration of simple and reliable scRNA-seq workflows into clinical research.

Discussion

Genomics has progressed rapidly to high throughput, multi-modal, single-cell analysis[40,51-54].

Further improvements in data quality, the ability to measure additional cellular properties, and

new computational approaches for understanding and integrating single-cell information[55-57] will

continue to refine our understanding of cell states. At the same time, there remains an unmet

need for simplified workflows that scale in cell number and sample size, and that allow for

breaks in processing after initial sample collection. PIP-seq is a microfluidics-free, single-cell

RNA-sequencing method that produces high-quality data using a drastically simplified

emulsification technique. Like other high-throughput single-cell approaches, PIP-seq is

fundamentally a strategy to barcode mRNA from cells so that material can be pooled and

sequenced. The core advantage of PIP-seq is the speed and simplicity of sample processing.

Particle-templated emulsification forms monodispersed bead-containing emulsions in minutes

with a standard laboratory vortexer, removing the need for instrumentation located in core

facilities or hours of multichannel pipetting to perform split-pool indexing in plates. This

expands access to single-cell technologies in several important ways. It reduces the need for

sample transport, enabling immediate processing by technicians without prior training, and

collection and banking of samples from remote locations, including field sites. It allows

infectious samples that require special precautions to be processed at the point of collection or in

the biosafety facilities where they are stored. More generally, rapid sample processing eliminates

the need for fixatives and minimizes transcriptional perturbations and batch artifacts associated

with processing many samples in series.

In addition to workflow simplicity, PIP-seq is intrinsically scalable, handling cell inputs

over five orders of magnitude ($10-10^6$), making it well suited for screening, genome-wide

Perturb-seq experiments, and large cell-atlas studies. While methods based on combinatorial

indexing scale efficiently to large cell numbers, PIP-seq has a substantially simpler workflow

and is also compatible with high throughput processing of samples in plates, allowing many

conditions and replicates to be run simultaneously. This has important implications for data

quality and biological discovery in single-cell experiments since the detection of true positives

and reduction in false positives in differential expression analysis is substantially improved by

incorporating replicates and statistical methods that account for biological variability[58].

Increased flexibility in the number of samples that can be processed also enables previously

difficult experimental designs, like dose-response curves, time-course studies, combinatorial

perturbations, single-cell sequencing of organoids, and large drug screens. In addition, because

PIP-seq can directly emulsify in plates, it integrates with robotic fluid handling and therefore

comprises a drop-in solution for single-cell readouts in high-throughput experiments in academia

or industry.

We confirmed the accuracy of PIP-seq as a single-cell genomics tool by profiling

heterogeneous tissue and directly comparing our results to a commercial scRNA-seq platform

(10x Genomics). PIP-seq cell type classification, marker identification, and gene expression

levels were tightly matched with 10x data but detected fewer genes per cell. We attribute these

differences to the extensive optimization that the commercial platform has undergone and

suspect that, like other single-cell techniques[3-6,12,25,59], further improvements to PIP-seq

molecular biology will increase sensitivity. In addition, because PIP-seq emulsions are

functionally equivalent to those made with microfluidics, our approach is immediately

compatible with emerging advances, including improvements to the molecular biology of myriad

multi-omic profiling methods developed for other droplet microfluidic barcoding systems[40,57,60].

Finally, we demonstrated the utility of PIP-seq in processing clinical samples. In

combination with barcoded antibodies, we profiled the relapse of mixed phenotype acute

leukemia (MPAL) after chemotherapy. MPAL is a subtype of leukemia characterized by poor

prognosis[61], lineage ambiguity, lack of consensus regarding therapy, and significant intratumoral

genetic and immunophenotypic heterogeneity[62,63]. The molecular mechanisms underlying

treatment resistance in this complex disease remain undefined. Changes in gene expression have

been linked to prognosis and treatment resistance in multiple cancers. However, tumor

heterogeneity makes it unlikely that bulk sequencing methods would identify strong gene

signatures associated with resistance in clinical samples. Using PIP-seq of longitudinal samples

from two individuals with MPAL with disease progression after initial therapy, we identified

transcriptional heterogeneity beyond that observed by immunophenotype and speculate that this

heterogeneity may play a role in MPAL treatment resistance. We observed up-regulation of

genes and pathways previously associated with acute lymphoblastic leukemia in several cell

subsets that emerged after chemotherapy, as well modulation of ribosomal genes in both patients.

Control of translation has been previously implicated in many cancers[41-46,64], including leukemia,

but has not yet been linked to MPAL progression and drug resistance, suggesting that new

therapeutics targeting ribosomal biogenesis and/or protein translation may also have therapeutic

potential in MPAL[65]. Our results motivate the use of single-cell technologies for understanding

MPAL tumor heterogeneity and response to chemotherapy and suggest that the broad adoption

of such technologies for monitoring cancer progression (and tailoring treatment) is within reach.

In summary, single-cell RNA-sequencing provides unparalleled insight into cell heterogeneity

but remains underutilized in many settings. PIP-seq addresses this with a simple, rapid, and

scalable workflow that can be used by any lab containing standard molecular biology equipment.

Methods

Proteinase K triggered cellular lysis and mRNA capture. Mammalian cells were stained with

Calcein AM (ThermoFisher #C3099) in 1 mL of PBS with 0.04% BSA according to

manufacturer's instructions. After 30 min of incubation at room temperature on a rotisserie

incubator (Isotemp, Fisher Scientific), cell suspensions were quantified with a Luna-FL

automated cell counter and diluted in 1X PBS with 0.04% BSA. 1500 calcein-stained cells in 5

µL of were added to 35 µL of barcoded hydrogel templates with 29 U/mL Proteinase K (NEB

#P8107S) and 70 mM DTT (Sigma #D9779) and mixed for 10 pipet strokes. Care was taken to

avoid generating bubbles when mixing cells with 280 µL of 0.5% ionic Krytox in HFE 7500 oil[66]

was added to the cell-bead mixture and vortexed at 3000 RPM for 15 seconds horizontally and

then 2 minutes vertically with a custom vortexer (Fluent BioSciences, #FB0002776). Oil was

removed from below the emulsion such that less than 100 µL remained. The PIP emulsion was

subsampled on a C-Chip disposable hemacytometer (Fisher Scientific # DHCN015) before lysis

with each subsample consisting of 3.5 µL of PIP emulsion per field of view. The C-chip was

imaged in bright-field at 2X magnification. The remaining PIP emulsion was subjected to

enzymatic lysis at 65°C for 35 min on a PCR thermocycler (Eppendorf Mastercycler Pro) with

lid temperature set to 105°C. After lysis was complete, fluorescence images were captured using

a Nikon 2000 microscope with 470 nm excitation (Thorlab M470L5).

Synthesis of barcoded bead templates. Prototype barcode bead fabrication proceeded

according to previous reports[30]. Briefly, a simple co-flow microfluidic device was used to

combine acrylamide premix (6% w/v Acrylamide, 0.1% Bis-Acrylimide, 0.3% w/v ammonium

persulfate, 0.1□×□Tris-buffered saline–EDTA (TBSET: 10 mM Tris–HCL pH 8.0, 137 mM

NaCl, 20 mM EDTA, 1.4 mM KCl, 0.1% v/v Triton-X100), 50 uM acrydited primer

(/5Acryd/TTTTTTTAAGCAGTGGTATCAACGCAGAGTACGACTCCTCTTTCCCTACACG

ACGCTCTTCC) with oil (HFE-7500, 3 M Novec) containing 2% (w/v) surfactant (008-Fluoro-

surfactant, Ran Technologies) and 0.4% v/v Tetramethylethylenediamine (TEMED). The

emulsion was solidified at room temperature for 12 hours and beads were removed using

1H,1H,2H,2H-Perfluoro-1-octanol (Sigma Aldrich), washed thrice with Tris–EDTA-Tween

buffer (TET: 10 mM Tris–HCl pH 8.0, 10 mM EDTA, 0.1% v/v Tween-20) followed by two

washes with 30 mM NaCl, 10 mM Tris–HCl pH 8.0, 1 mM MgCl2, 0.1% Tween-20. The final

bead size was 80 microns. Split-pool barcode assembly utilized the ligation assembly approach

as described previously[30]. Beads were resuspended in T4 ligation buffer (NEB #B0202S), heated

with a complementary oligonucleotide to 75°C for 2 min, and cooled to room temperature to

anneal. 100 microliters of beads were distributed into each well of a 96-well plate containing a

unique barcode with 1x T4 ligation buffer and 1.9 U/ul T4 DNA ligase (NEB #M0202M).

Ligations were incubated at 25C for 1 hour and heat inactivated at 65C for 10 minutes. Well

contents were combined and washed 5 times in 15mL TET. The process was repeated to add 4

barcodes and a UMI with polyT (NNNNNNNNNNNNTTTTTTTTTTTTTTTTTTTV). Quality

control steps were identical to previous reports[30]. Bead manufacturing methods were transferred

to Fluent BioSciences for scaled production, validation, and distribution. Commercially

produced beads were used for several experiments, as noted.

Varied format emulsification. PIP emulsification in varied formats was performed in 0.5 mL

microcentrifuge tubes, 15 mL conical tubes and 50 mL conical tubes. Briefly, PIP particles were

suspended in buffer with 29 U/mL Proteinase K Proteinase K (NEB #P8107S) and 70 mM DTT

(Sigma #D9779) and pelleted through centrifugation. Barcoded hydrogel templates were then

distributed at 35 µL, 0.5 mL and 8 mL volumes in 0.5 mL, 15 mL and 50 mL tubes, respectively.

Fluorinated oil with surfactant (Fluent Biosciences # FB0001804) was added to each tube at 200

µL, 8 mL and 32 mL volumes, respectively. Emulsification was conducted on a Vortex Genie 2

with a custom adapter (Fluent #FBS-SCR-8VX) at maximum rpm for 1 min. After

emulsification, the samples were allowed to settle for 30 seconds, and excess oil was removed

via syringes using G22 blunt needles. The emulsion was subsampled, loaded on a C-Chip

disposable hemacytometer (Fisher Scientific #DHCN015), and imaged under brightfield

microscopy (DIAPHOT300, Nikon) at 2x and 4x magnification.

Emulsification in well plates was tested using two bead buffer conditions. First, to test

emulsification in 96, 384, and 1536 well plates, PIP particles were suspended in 2% (v/v) Triton

X-100 (Sigma, X100-5ML) in 10 mM TrisHCl (Teknova, T1075), then centrifuged at 6000 rcf,

and the supernatant removed **(Fig. 1, Extended data Fig. 2c)**. Depending on the well plate

working volume, 38 µL, 8 µL, or 3 µL of the centrifuged barcoded hydrogel templates were

added to 96, 384, or 1532 well plates respectively. For 96 and 384 well plates 2 µL of sample

was added to each well, for 1532 well plates, 1 µL was added to each well. PIP and sample

volumes totaled 25% of the volume of each well. The plate was then sealed (Applied

Biosystems, 4306311) and shaken for 5 min (IKA, 253614 and 3426400) to ensure complete

mixing. Well plates were centrifuged at 200 rcf for 1 min before removing the seal. Then 80 µL,

20 µL, or 8 µL 2% (w/w) fluorosurfactant (Ran BioTechnologies, 008 Fluorosurfacant) in HFE

oil (3M, Novec 7500) was added to each well in 96 (Applied Biosystems, N8010560), 384

(Applied Biosystems, A36931), or 1532 (Nunc, 253614) well plates respectively. The addition of

oil represented 50% of the volume of each well for a total volume of 75% consisting of PIP,

sample, and oil. After resealing, PIP emulsification was performed by vortexing for 30 seconds

at 3200 rpm (Benchmark Scientific, BV1003). The emulsified plate was centrifuged at 200 rcf

for 1 min before removing the seal and imaging droplets from individual wells on a fluorescence

microscope (EVOS FL Auto).

Second, to test well plate emulsification with cells in 96 and 384 well plates, PIP particles

were suspended in buffer with 29 U/mL Proteinase K (NEB #P8107S) and 70 mM DTT (Sigma

#D9779) and pelleted through centrifugation. For 96-well plates (Eppendorf #0030129300), 25

µL of barcoded hydrogel templates were then distributed into each well with 4000 cells per well

(2000 cells/µL x 2 µL). 150 µL of fluorinated oil with surfactant (Fluent Biosciences #

FB0001804) was added to each well. Emulsification was conducted on a Vortex Genie 2 with

flat-head adapter at 3000 rpm for 2 minutes. For 384-well plates (Corning #3347), 15 µL of

barcoded hydrogel templates were then distributed into each well with 3000 cells per well (2000

cells/µL x 1.5 µL). 105 µL of fluorinated oil with surfactant (Fluent Biosciences # FB0001804)

was added to each well. Emulsification was conducted on a Vortex Genie 2 with a flat-head

adapter at 3000 rpm for 2 minutes **(Fig. 1, Extended data Fig. 2a,b)**.

PIP-seq protocol. Unless otherwise noted, cells were centrifuged 300 x g for 5 min, washed

twice in 1X PBS without calcium or magnesium (ThermoFisher #70011044) with 0.04% BSA,

filtered with a 70 µm cell strainer, and resuspended in 1X PBS with 1% Pluronic F127 (Sigma

#P2443). Pre-aliquoted barcoded hydrogel templates were thawed on ice. Volumes of barcoded

hydrogel templates, cells, and oil varied based on the number of cells as noted in each

experimental subsection below. A standard small-format run was as follows: 5 µL of 500

cells/µL was added to 35 µL of barcoded hydrogel templates with 29 U/mL Proteinase K and 70

mM DTT (Fluent BioSciences # FB0001876) and mixed for 10 strokes. Care was taken to avoid

generating bubbles when mixing cells with barcoded hydrogel templates. 280 µL of oil (Fluent

Biosciences # FB0001804) was added to the cell-bead mixture and vortexed (Vortex Genie 2,

Scientific Industries) using a custom adapter (Fluent BioSciences, #FB0002100) at the maximum

rpm for 15 seconds horizontally and 2 minutes vertically. 230 µL of excess oil was removed and

the emulsion and enzymatic lysis was completed at 65°C for 35 minutes with a 4°C hold on a

PCR thermocycler with lid temperature set to 105°C. The remaining oil was removed. The

emulsion was broken as follows. Using a multichannel pipet, 180 µL of room temperature high

salt buffer (250 mM Tris-HCl pH 8, 375 mM KCl, 15 mM MgCl$_2$, 50 mM DTT) was added to

the top of the emulsion followed by 40 µL of 100% 1H,1H,2H,2H-perfluoro-1-octanol (Sigma

Aldrich, 370533). The samples were vortexed for 3 seconds, briefly centrifuged, and the bottom

oil phase was removed. Barcoded hydrogel templates were transferred into a 1.5 ml Eppendorf

tube and washed 3 times with 2x RT buffer (100 mM Tris-HCL pH 8.3, 150 mM KCl, 6 mM

MgCl2, 20 mM DTT) with 1% Pluronic F68 (Gibco #24040032). After washing, the beads were

pelleted, aqueous removed, and the remaining bead and buffer volume was 25 ul. To this mix, 25

µL of reverse transcription (RT) master mix. The master mix consisted of 4.8% PEG8000, 4%

PM400, 2.5 µM template switch oligo (PIPS_TSO), 1 mM dNTPs (NEB), 1 U/µL RNase

Inhibitor (NxGen, Lucigen), 1 U/µL reverse transcriptase (ThermoFisher, Maxima H-minus

EP0751). RT master mix was added, mixed, and cDNA synthesis was completed for 30 minutes

at 25°C, 90 minutes at 42°C, followed by 10 minutes at 85°C and a 4°C hold. Whole

transcriptome amplification (WTA) was performed directly on RT product without purification

by adding 50 µL of 1x KAPA Hifi master mix, 0.25 µM primer (PIPS_WTA_primer), and

thermocycling 95°C for 3 min, then 16 cycles of (98°C 15 s, 67°C 20s, 68°C 4 min) followed by

72°C 5 min and 4°C hold. After WTA, barcoded hydrogel templates were removed using

Corning Spin-X filter columns (1 min at 13,000 xg), and amplified cDNA was purified using

0.6x Ampure XP. Libraries were generated from WTA amplified material using the Nextera XT

DNA Library Preparation Kit with a custom primer (PIPS_P5library) and standard Nextera P7

indexing primers (N70x). Libraries were pooled and sequenced using an Illumina NextSeq 2000

instrument with 15% PhiX. Oligonucleotides used in this study are found in **Supplemental**

Table 1.

Human-mouse mixing studies. Human HEK 293T cells (ATCC # CRL-3216) were grown in

DMEM (ThermoFisher #11995073) supplemented with 10% FBS (ThermoFisher #A3840001)

and 1% Penicillin-Streptomycin-Glutamine (ThermoFisher #10378016). Murine NIH/3T3 cells

(ATCC # CRL-1658) were grown in DMEM (ThermoFisher #11995073) supplemented with

10% bovine calf serum (ATCC # 30-2030) and 1% Penicillin-Streptomycin-Glutamine. Cells

were grown to a confluence of ~70% and treated with TrypLE Express with Phenol red

(ThermoFisher #12605010) for 3 minutes, quenched with an equal volume of growth medium,

and centrifuged at 5 min at 200 x g. The supernatant was removed, and the cells were

resuspended in 1X DPBS without calcium or magnesium. Cells were diluted to their final

concentration in 1X DPBS with 0.04% BSA and mixed evenly to create a 50:50 Human:Mouse

mixture. Cell viability was evaluated using acridine orange/propidium iodide stain (Logos Bio #

F23001) and quantified with a Luna-FL automated cell counter. Cells were processed using the

PIP-seq protocol as described above.

72-hour hold experiments. 5 µL of a 50:50 mixture of human HEK 293T cells and murine

NIH/3T3 cells (800 cells/µL) was added to 35 µL of barcoded hydrogel templates Fluent

BioSciences Part # FB0003067) with 29 U/mL Proteinase K and 70 mM DTT and mixed for 10

strokes. 280 µL of oil (Fluent Biosciences # FB0001804) was added to the cell-bead mixture,

which was vortexed on a digital vortexer using a custom adapter (Fluent BioSciences

#FB0002084) at 3000 rpm for 15 seconds horizontally and 2 minutes vertically. 230 µL of

excess oil was removed and the emulsion was placed in a pre-heated digital dry bath at 66°C for

38 min and 4°C for 11 minutes. Control samples proceeded to emulsion breaking while 0°C hold

samples were placed in an ice bucket in the 4°C refrigerator for 72 hours prior to breaking

emulsions. Breaking, mRNA extraction, reverse transcription, WTA, and cDNA isolation,

adapter ligation-based library preparation and Illumina sequencing were performed as previously

described.

Healthy breast tissue comparison to 10X. Fresh reduction mammoplasty tissue was processed

as previously described[31,67]. Bulk mammary tissues were mechanically processed into a slurry

and digested overnight with collagenase type 3 (200U/mL, Worthington Biochem CLS-3) and

hyaluronidase (100U/mL, Sigma-Aldrich H3506) in medium containing charcoal:dextran

stripped FBS (GeminiBio 100-119). The digested fragments were size filtered into a sub-40

micron fraction and an above 100-micron fraction and cryopreserved. For PIP-seq, cells were

thawed and then resuspended in PBS + 0.04% BSA and passed through a 70-micron FlowMi cell

strainer (Sigma # BAH136800070). For 10x Genomics data, the 100-micron fraction were

thawed, then further digested with trypsin, followed by dispase (Stemcell Technologies #07913)

and DNAseI (Stemcell Technologies #07469) digestion to achieve single-cell suspension. For

PIP-seq, 20 µL of cells (1500 cells/µL in PBS + 0.04% BSA) was added to 200 µL of barcoded

hydrogel templates (Fluent BioSciences, #FB0002617) and mixed for 10 strokes. 1000 µL of oil

(Fluent Biosciences # FB0001804) was added to the cell-bead mixture and it was vortexed on a

digital vortexer using a custom adapter (Fluent BioSciences, #FB0002100) at 3000 rpm for 15

seconds horizontally and 2 minutes vertically. 800 µL of excess oil was removed and the

emulsion was placed on a pre-heated digital dry bath at 66°C for 38 min and 4°C for 11 minutes.

Breaking, mRNA extraction, reverse transcription, WTA, and cDNA isolation was performed

under standard conditions. Adapter ligation-based library preparation was performed according

to manufacturer's instructions (Watchmaker Genomics, #7K0019-024). Samples were sequenced

on the Illumina NextSeq 2000, with four patient samples pooled per P3 cartridge and sequenced

at a read depth of approximately 36,500 reads/cell. For 10X Genomics, cells from each patient

were labeled with MULTIseq barcodes[13], then pooled and stained with DAPI to be sorted for

DAPI-live cells. Single-cell libraries were prepared according to the 10X Genomics Single Cell

V3 protocol (v3.1 Rev D) with the standard MULTIseq sample multiplexing protocol. The

libraries were sequenced on a NovaSeq S4 lane at a read depth of about 70,000 reads/cell. To

compare platforms, we downsampled PIP-seq and 10x data, which had different numbers of cells

and sequencing depth per cell. The PIP-seq data had 54,825 cells, sequenced at approximately

36,500 reads per cell, while the 10x data had 2420 cells sequenced at approximately 70k reads

per cell. Data were downsampled to 2400 cells and 36,500 reads in R (downsampleReads,

DropletUtils). For correlation and marker gene comparisons, data were downsampled to 2400

cells and 1500 UMIs in R (SampleUMI, Seurat v4.1.0). Markers used for breast tissue cluster

cell type calling are found in **Supplemental Table 2**.

Single-tube large format breast tissue study. PIP-seq was performed as previously described,

except that cells were counted and diluted with PBS + 0.04% BSA to a concentration of 10,000

cells/µL. 40 µL of cell suspension was input into 800 µL of barcoded hydrogel templates (Fluent

BioSciences Part # FB0003067). 4000 µL of oil (Fluent Biosciences # FB0001804) was added to

the cell-bead mixture and vortexed on a digital vortexer using a custom adapter (Fluent

BioSciences # FB0002659) at 3000 rpm for 15 seconds horizontally and 2 minutes vertically.

Excess oil was removed using a 3 mL syringe with a G22 blunt bottom syringe needle. Lysis

proceeded using 3300 µL of a lysis emulsion (Fluent BioSciences #FB0003039) added to the

cell-bead emulsion. The mixture was placed in a pre-heated digital dry bath at 37°C for 45 min

and 4°C for 10 minutes. Breaking, mRNA extraction, reverse transcription, WTA, and cDNA

isolation were performed under the same conditions as described previously. Adapter ligation-

based library preparation was performed according to manufacturer's instructions (Watchmaker

Genomics, #7K0019-024). 80 ng of cDNA was used to prepare four replicate library

preparations which were pooled and sequenced on two Illumina NextSeq 2000 P3 cartridges at a

read depth of 13,025 reads/cell, after concatenation.

CROP-seq. K562 CRISPRi cells were cultured in RPMI-1640 (Gibco #11875093) with 10%

FBS (Thermo Fisher Scientific, #10438026) and 1% penicillin/streptomycin (Thermo Fisher

Scientific, #15140148) in an incubator at 37°C with 5% CO2. K562 CRISPRi cells were

transduced with a lentivirus library containing 138 sgRNA[35] at a multiplicity of infection of 0.1.

Lentivirus-infected cells (BFP+) were sorted to high purity using a BD FACS Aria III (100 µM

nozzle) and processed according to the PIP-seq scRNA-seq workflow. 3 µL of cells (333

cells/µL) was added to 28 µL of barcoded hydrogel templates with 29 U/mL Proteinase K and 70

mM DTT and mixed for 10 strokes. 150 µL of 0.5% ionic Krytox in HFE 7500 oil was added to

the cell-bead mixture and it was vortexed at 3000 RPM for 1 minute on a Vortex Genie 2 with a

custom tube adapter. cDNA was processed according to the standard PIP-seq protocol to obtain

sequence ready libraries containing transcriptome information. To recover sgRNA sequences, we

implemented an additional amplification step. We amplified 1 ng cDNA in a 50 µL reaction

using of primers P5-PE1 (0.5uM) and Weissman_U6 (0.25uM) **(Supplementary Table 1)** with

1x Kappa HIFI. Reactions were thermocycled at 95°C for 3 min followed by 10 cycles of [95°C

for 20s, 70°C for 30s [-0.2°C per cycle], 72°C for 20s], followed by 8 cycles of [95°C for 20s,

68°C for 30s, 72°C for 20s], followed by 72°C for 4 minutes, and 4°C forever. Library PCR

product enriched in sgRNA sequences was purified with a double sided 0.5x/0.8x Ampure XP

bead cleanup and the size was determined (Agilent Tapestation).

Transcriptome and sgRNA libraries were pooled at 20:1 before sequencing. Reads were

first processed to extract sgRNA sequences. The bioinformatics pipeline was run using a custom

index built from the full human transcriptome (GENCODE v32) and guide RNA sequences

(salmon v1.2.0.). This approach led to the recovery of >14,000 unique gRNA counts across all

cell-associated barcodes. Cells were assigned to gRNA groups using a previously reported

approach[32]. Briefly, cells were classified as uniquely expressing a single gRNA species if the

guide's expression was at least 10-fold higher than the sum of all other gRNAs. Similarly, cells

were classified as containing multiple gRNAs in cases where the difference was smaller than

one. For the 581 single cells sequenced, 2 did not have any gRNA, 441 contained a single

gRNA, and 138 contained multiple gRNA. Cell barcodes were processed using Seurat v4.1.0. All

gRNAs in the list of features were excluded from the identification of variable transcripts

(feature selection) and in subsequent stages of dimensionality reduction and clustering. To

understand the relationship between gRNAs and mRNA expression, gRNAs were ranked

according to their expected level of knockdown, reported previously[35], and a generalized additive

model was used to assess groupwise trends for each set of gRNAs.

Lung adenocarcinoma cell line experiments. PC9 was obtained from the RIKEN Bio Resource

Center (#RCB4455). H1975 was obtained from the American Type Culture Collection (#CRL-

5908). Cells were cultured in RPMI-1640 (Gibco #11875093) with 10% fetal bovine serum

(FBS), penicillin, and streptomycin in an incubator at 37°C with 5% CO2. 1 µM Gefitinib

(Frontier Scientific, 501411677) or DMSO was added to culture flasks 24 hours before cells

were harvested for processing. PC9 and H1975 were both treated with Gefitinib and DMSO. To

perform the cell mixing study, Gefitinib treated H1975 and Gefitinib treated PC9 were mixed at

a ratio of 1:9 H1975:PC9. 5 µL of cells (400 cells/µL) were added to 28 µL of barcoded

hydrogel templates with 22.8 U/mL Proteinase K and 28 mM DTT and mixed for 10 pipet

strokes. 150 µL of 0.5% ionic Krytox in HFE 7500 oil[66] was added to the cell-bead mixture and

it was vortexed at 3000 RPM for 1 minute on a Vortex Genie 2 with a custom tube adapter.

Triplicate tubes of 400 cells were processed per treatment condition. Data was analyzed using

Seurat v4.1.0.

Healthy PBMCs. Cryopreserved PBMCs were obtained from a commercial provider (AllCells,

Lot #3052467). Cells were thawed and prepared for PIP-seq as previously described in the

MPAL study, except that the final cell dilution was made in 1X PBS + 0.04% BSA. *For the high*

cell count PBMC study. PIP-seq was performed as previously described in the high cell number

breast tissue study except that cells were counted and diluted with PBS + 0.04% BSA to a

concentration of 4300 cells/µL and 44 µL of cell suspension was input into 800 µL of barcoded

hydrogel templates (Fluent BioSciences Part # FB0003067). Cryopreserved PBMCs used for cell

hashing were obtained from a commercial provider (AllCells, and Lot #3082436) and prepared

for PIP-seq as described previously. *For the cell hashing study.* Cell staining and PIP-seq were

performed according to PIPseq Single Cell Epitope Sequencing User Guide

(FB0002079). Briefly, one million PBMCs were resuspended in 47.5 µL of Cell Staining Buffer

(BioLegend #420201) and 2.5 µL TruStain FcX block (BioLegend #422301) was added prior to

mixing and incubation for 10 min on ice. Next, 1 µg of TotalSeqA antibody was diluted in Cell

Staining Buffer and 50 µL of this antibody dilution was added to the blocked cells prior to

incubation on ice for 30 min. Stained cells were washed in Cell Staining Buffer three times and

resuspended in 1X PBS+0.04% BSA at 2000 cells/µL. For the PIP-seq, 20 µL of this cell

resuspension was added to 200 µL of barcoded hydrogel templates (Fluent BioSciences

#FB0002617) and processed through PIP-seq.

Mixed phenotype acute leukemia (MPAL). Patients whose samples were used in this study

were treated at the University of California San Francisco. Samples were collected in accordance

with the Declaration of Helsinki under institutional review board-approved tissue banking

protocols, and written informed consent was obtained from all patients. Sample clinical

characteristics are found in **Supplemental Table 3**. Cryopreserved peripheral bone marrow

mononuclear cell samples (PBMCs) were thawed by hand until approximately 85% of ice

remained. Using a 5 mL serological pipette, 1 mL of 4°C defrosting media (DMEM with 20%

FBS and 2 mM EDTA) was added dropwise to each sample and then, without disturbing the

remaining ice pellet, the sample was carefully transferred dropwise to a pre-prepared 40 mL

aliquot of 4°C defrosting media. This was repeated until the contents of the entire cryovial were

transferred into the 50 mL conical of defrosting media. The sample was inverted 4-5 times and

centrifuged at 114 x g for 15 min at 4°C with no brake. The supernatant was aspirated and 10 mL

of room temperature RPMI 1640 with 1% Penicillin-Streptomycin-Glutamine was used to gently

resuspend the cells. Cell clumps were manually removed, and if necessary cells were filtered

through a 70 micron cell strainer into a fresh 50 mL conical. The sample was inverted 2-3 times

and centrifuged at 114 x g for 10 minutes with low brake at room temperature. The supernatant

was aspirated and cells were resuspended in an appropriate volume of 1X PBS + 5% FBS. Cells

were quantified with AO/PI and viability evaluated on the Luna-FL. 1-2 million cells were

aliquoted into a new 15mL conical and then centrifuged at 350 x g for 4 min at 4°C and then

the supernatant was aspirated and the tube was placed on ice. 45 µL of cold Cell Staining

Buffer (BioLegend, 420201) was added per million cells and resuspended gently. 5 µL Trustain

FcX block (BioLegend, 422301) was added per million cells and gently mixed 10 times with a

wide-bore pipette tip. Cells were blocked on ice for 15 minutes. A custom pool of 19 TotalSeq

A antibodies was obtained from BioLegend. Immediately prior to use, antibodies were mixed

and centrifuged at 10,000 x g for 4 min at 4°C. 4.6 µL of 0.5 µg/µL antibody pool was added

per million of the blocked cells and gently mixed 10 times with a wide-bore pipette tip. The

samples were incubated on ice for 60 minutes. Next, 3.5 mL of cold Cell Staining Buffer was

added, gently mixed with a wide-bore pipette tip, and slowly inverted twice to mix. Cells were

centrifuged at 350 x g for 4 min at 4°C and then the supernatant was removed. The addition of

cold Cell Staining buffer was repeated twice for a total of 3 washes. After the final supernatant

aspiration, stained cells were resuspended in 1X PBS with 0.04% BSA and mixed 5-10 times

until cells were completely suspended without visible clumps. Cell concentration was determined

with AO/PI and viability was evaluated on the Luna-FL. Final dilutions were made in 1X PBS

with 0.04% BSA. 20 µL of cells was added to 200 µL of barcoded hydrogel templates (1000

cells/µL) and processed according to PIPseq Single Cell Epitope Sequencing User Guide

(FB0002079). Marker genes identified for patients 65 and 873 are found in **Supplemental**

Tables 4 and 5, respectively.

PIP-seq bioinformatic analysis. Analysis of sequencing data was performed using custom

scripts to generate gene expression matrices starting from processed FASTQ sequences. The

pipeline comprises 4 basic steps: barcode identification and error correction, mapping to

reference sequences, cell calling, and gene expression matrix generation. Briefly, after

demultiplexing the sequencing data each read in the FASTQ is matched against a "whitelist" of

known barcodes. Reads were matched with a hamming distance tolerance of 1, meaning that the

barcode portion of a read can differ from a whitelist entry by one base and still be matched to

that barcode. Reads that did not match any barcode in the whitelist were discarded from further

analysis. Matching reads were output to a new, intermediate FASTQ file that was then used for

mapping against an appropriate transcriptome reference. Reference transcriptomes matching the

species of each sample were prepared using the Salmon *index* function with the default k-mer

size of 31[68]. GENCODE references were used to build the transcriptome indexes including

GRCh38.p13 for human, GRCm38.p6 for mouse, and the combination thereof for HEK/3T3 cell

mixture studies. Following barcoding, Salmon *alevin* v1.2.0[69] was used to map reads to the full

transcriptome. The intermediate FASTQ generated during barcoding were provided as input into

alevin along with a list of all whitelisted barcodes contained in raw reads. After mapping, data

were output as UMI counts matrices (sparse matrix, gene list, barcode list) with dimensions of

all barcodes x all genes in index. An in-house python implementation of emptyDrops[70], a

standard scRNA-seq method to separate putative cells from background, was then applied. A

custom threshold for each experiment was set, beneath which no true cell barcodes were

expected to fall. As with emptyDrops, an estimated ambient profile across all barcodes beneath

that threshold was created. A p-value was computed by comparing the gene expression profile

for each barcode above the threshold against the ambient profile. Barcodes with a statistically

significant difference (Benjamini-Hochberg adjusted-p<0.001) from the ambient background

profile were categorized as cell-containing barcodes. The *alevin* output matrices were then subset

to only include called cell barcodes. Gene expression matrices were normalized prior to

performing unsupervised clustering and UMAP dimensionality reduction. First, gene expression

counts for each cell were divided by the total counts for that cell and multiplied by a scaling

factor of 10,000. Finally, the data were transformed to natural-log scale using log1p(). The

Seurat package (v4.1.0) was used to perform downstream clustering, marker gene determination,

and visualization in R. Seurat's FindClusters() and RunUMAP() commands were used with

default settings.

For saturation curve comparisons, PIPseq and 10x samples were downsampled to

matching depths of 5K-80K reads per called cell. Downsampling was performed using seqtk for

PIPseq samples and using DropletUtils read10xMolInfo() function with a molecule_info.h5 file

directly downloaded from the 10x website. Inflection point-based cell calling was used to

standardize cell calls across platforms. Median transcripts/cell and genes/cell were calculated

from the cell fraction of the resulting count matrices. For violin plot comparisons, samples were

prepared to match the same processing configuration used by Ding et al (2020)[28]. First, samples

were downsampled to 53K reads per called cell and trimmed to 50bp for read 2 prior to

processing, sampling in the same manner described above. Each violin plot represents the cell

fraction from a single replicate of an HEK/3T3 cell mixture, with human and mouse split out into

separate plots.

Analysis of PBMC data for the high cell count study was performed using custom scripts

as described above, until the completion of mapping. Cell calling, clustering, and differential

expression were performed using PIPseeker v1.0.0 (Fluent Biosciences) in *reanalyze* mode using

–force-cells 65000. Top differentially expressed genes from the PIPseeker graph-based

clustering result were used to determine cell types by comparing to a reference gene list

(Supplementary Table 7). Log-normalized expression for key genes (e.g. CD34) were overlaid

on the UMAP projection to highlight markers associated with specific cell types (color bar in

log10 scale). Analysis of PBMC data for the cell hashing study was performed using PIPseeker

v1.0.0 in *count* mode using STAR (v2.7.10a) and the PIPseeker human reference

(https://www.fluentbio.com/products/pipseeker-for-data-analysis/). ADT analysis was performed

by performing barcode error correction with PIPseeker v1.0.0 (*count* mode) and custom scripts

to trim R2 to the first 16bp. Error-corrected and trimmed FASTQs were input to CITE-seq Count

(v1.4.3) using the following settings: -t (hashtag whitelist) -cbf 1 -cbl 16 -umif 17 -umil 28 --

cells (# called cells from RNA cell calling). The hashtag whitelist contained two TotalSeq™-A

anti-human   Antibody   hashes   (A0253   -   TTCCGCCTCTCTTTG, A0255   -

AAGTATCGTTTCGCA). The filtered matrix output by PIPseeker for the RNA data was

merged with the umi-count matrix from CITE-seq Count on cell barcode to create a merged

matrix.   The   hashing   data   was   demultiplexed   in   Seurat   using   HTODemux

(positive.quantile=0.99). Downstream was performed in Seurat using SCTransform() along with

RunPCA(), FindNeighbors(dims=1:15) and RunUMAP(dims=1:15). Cell type annotation was

performed with singleR (v1.4.1) and used an annotated 10x Genomics v1 chemistry dataset as a

reference. Cells were classified by their max hash identity and projected in the RNA-based

UMAP space. The HTO data was subjected to clustering in Seurat using the HTOHeatmap()

function to visualize singlets, doublets, and unclassified cells.

For 72-hour hold experiments, analysis was performed using custom scripts as previously

described. Samples were normalized to the same depth (45,000 reads/cell). Cell types were then

annotated as human (HEK293T) or mouse (NIH-3T3) using a purity threshold of >85% single-

species content per barcode. Barcodes from each species were subset and transcript counts were

summed for each gene to generate two pseudobulk counts tables per sample. Samples were

aggregated separately for each species and then were analyzed with DESeq2. A contrast of 0 vs.

72 hours was performed for each species, while controlling for batch effects associated with

different users. For the correlation analysis, pseudobulk counts derived above were normalized to

transcripts per million and transformed to log1p scale. Pearson correlations (R) and slopes (m)

were calculated by fitting a linear model to the data. Data were then plotted in R with ggplot2

v3.3.5 and were aggregated into a grid using GGally v2.1.2. Additionally, the distribution of

cells in UMAP space at 0 and 72 hours post-lysis was examined. After processing data in Seurat

as described, harmony batch-correction was used to integrate datasets.

Figures

Figure 1. Rapid and scalable templated emulsification for single-cell genomics. (a-d) PIP-

seq enables the encapsulation, lysis, and barcoding of single cells. **(a)** Schematic of the

emulsification process. Barcoded particle templates, cells, and lysis reagents are combined with

oil and vortexed to generate monodispersed droplets. **(b)** Heat activation of Proteinase K results

in lysis and release of mRNA that is captured on bead-bound barcoded polyT oligonucleotides.

(c) Oil removal is followed by bulk reverse transcription of mRNA into cDNA. **(d)** Barcoded

whole transcriptome amplified cDNA is prepared for Illumina sequencing. **(e-g)** Efficient single

bead, single drop encapsulation at scale. **(e)** Particle templated emulsification in different sized

tubes (1.5 ml, 15 ml, 50 ml) produces monodispersed emulsions capable of barcoding orders of

magnitude different cell numbers. **(f)** PIP-seq is compatible with plate-based emulsification,

including 96-, 384- and 1536-well plate formats. **(g)** The estimated ability of different

technologies to easily scale with respect to cell and sample number.

Figure 2. Heat-activated enzymatic lysis yields high purity single-cell transcriptomes. **(a)**

Fluorescence microscopy (Brightfield, GFP) of calcein-stained cells emulsified with barcoded

bead templates before and after heat-activated lysis. Inset: images show cell puncta (left) and

release of calcein (right) after lysis. **(b-d)** Cell purity assessed with mouse-human mixing

studies. **(b)** The distribution of total UMIs as a function of cell barcode rank. **(c-d)** Purity

analysis of cell transcriptomes assessed using barnyard plots. Cells are colored by cell type: red

points are mouse reads, blue points are human reads, and green points are mixed reads.

Figure 3. Accurate single-cell transcriptional profiling of healthy breast tissue using PIP-

seq. (a) Clustering and identification of cell types from PIP-seq data (54,825 cells from 2

patients). **(b-e)** Comparison of PIP-seq to 10X Genomics data collected from the same tissue. **(b)**

Integration of PIP-seq and 10x data. **(c-d)** Cell clustering and comparison of marker genes

between platforms. **(d)** Heatmaps of marker gene expression show similar patterns in PIP-seq

and 10X data. **(e)** Correlations in normalized gene expression, by cluster, between platforms (see

also Extended data Fig. 4a).

Figure 4. Transcriptome and guide RNA sequencing using PIP-seq. (a) Schematic of the

CROP-seq sgRNA library designed with target mismatches to modulate the activity of essential

genes. **(b)** Lentiviral transduction of the CRISPRi library in K562 cells. **(c)** Schematic of the

capture and barcoding of polyadenylated mRNA and sgRNA using PIP-seq. RNA and sgRNA

libraries are prepared separately and pooled for sequencing. **(d)** Quantification of gene

expression of sgRNA within an allelic series. sgRNA is ordered from high to low predicted

knockdown efficiency (Jost et. al, 2020). Non-targeting sgRNA are denoted as "Null." **(e)** Pre-

ranked gene set enrichment analysis (GSEA) of scRNA-seq data comparing sgHSPA5

transduced cells to non-sgHSPA5 transduced cells shows enrichment in genes related to

endoplasmic reticulum stress and unfolded protein response.

Figure 5. Molecular signatures of drug-resistant cancer phenotypes in cell lines and patient

samples. **(a)** A two-by-two experimental study design using lung adenocarcinoma cell lines

(H1975, PC9) treated with Gefitinib or DMSO. **(b)** Clustering of single-cell RNA-seq after drug

treatment shows transcriptional perturbations in Gefitinib-sensitive PC9, but not Gefitinib-

resistant H1975 cells. **(c)** Increased expression of *TACSTD2* in PC9 challenged with Gefitinib.

(d) Identification of drug-resistant H1975 cells spiked into drug-sensitive PC9 cells based on

Gefitinib-induced transcriptional perturbation. **(e-l)** PIP-seq RNA and barcoded antibody (CITE-

seq) analysis of mixed phenotype acute leukemia (MPAL). **(e)** Clustering of single cells for

patient 65 before (left panel) and after (right panel) chemotherapy. **(f)** Clustering of single cells

for patient 873 before (left panel) and after (right panel) chemotherapy. **(g-h)** Antibody derived

tag (ADT) abundance, by cluster, before (t1) and after (t2) chemotherapy. ADTs change as a

function of chemotherapy but are consistent among clusters for both patients, with the exception

of T cell subsets. **(i-l)** Analysis of transcriptional heterogeneity in MPAL samples **(i)** Heatmap of

top differentially expressed marker genes by cluster after relapse in patient 63. **(j)** GSEA pre-

ranked analysis comparing transcriptomic differences between clusters 1 and 7 in patient 65

using gene sets: HP Acute Leukemia (M35856), Hallmark G2M Checkpoint (M5901), Hallmark

Oxidative Phosphorylation (M5936), and GOCC Ribosome (M17089) **(k)** Heatmap of top

differentially expressed marker genes by cluster after relapse in patient 873. **(l)** GSEA pre-ranked

analysis comparing transcriptomic differences between clusters 3 and 5 in patient 873 using gene

sets HP Acute Myeloid Leukemia (M36586)**,** GOCC Ribosome (M17089)**,** GOBP Oxidative

Phosphorylation (M17089), and Abnormal Myeloid Leukocyte Morphology (M37711).

Extended data Figure 1. Flexible sample processing with PIP-seq. (a) Storage of droplets

after emulsification for 72 hours at 0°C did not change quality metrics. **(b)** Data integration

between time points. **(c)** Correlations in normalized gene expression, by sample, between time

points for mouse and human cells.

Extended data Figure 2. (a-c) Microscope images of droplets and cells in plate emulsification

experiments. **(a,b)** Barcode bead templates, stained cells (puncta), and lysis reagents are

combined with oil and vortexed in **(a)** 96-well and **(b)** 384-well plates to generate monodispersed

droplets. Heat activation of Proteinase K results in lysis and release of calcein dye, and full-drop

fluorescence. **(c)** Microscope images of droplets from random wells of a 384-well plate

emulsification experiment.

Extended data Figure 3. Quality control analysis of PIP-seq using healthy breast tissue (a)

Integration and clustering of 54,825 cells from 2 patients with 2 replicates per patient. **(b)**

Coloring of UMAP by the number genes (nFeature RNA) for each cell. **(c)** The number of

unique genes (nFeature RNA), transcripts (nCount RNA), percent mitochondrial reads, and

percent ribosomal reads as a function of cluster. **(d)** Comparison between 10X Genomics' and

PIP-seq data after downsampling 2400 cells to 36,500 reads per cell.

Extended data Figure 4. Comparison, after down-sampling (2400 cells 1500 UMIs) of the

larger breast tissue PIP-seq dataset to 10x Genomics data generated from identical samples. **(a)**

Correlations in normalized gene expression, by cluster, between platforms. **(b)** Expression of

marker genes overlayed on clusters is consistent between platforms.

Extended data Figure 5. Data quality assessment of PIP-seq. **(a)** Representative distribution of

reads per cell. **(b)** Correlation between reads and genes per cell. **(c,d)** Comparison of (c)

UMIs/cell and (d) genes/cell in current single-cell methods. **(e,f)** Comparison of PIP-seq to 10X

Genomics across a range of sequencing depths (0-80,000 reads/cell) (e) UMIs/cell and (f)

genes/cell.

Extended data Figure 6. High cell number PIP-seq. **(a)** scRNA-seq of 138,146 single cells from

breast tissue using a single-tube emulsification in 2-minutes. **(b)** scRNA-seq of 65k peripheral

blood mononuclear cells (PBMCs) recovers a small population of CD34 cells. **(c)** Coloring of

UMAP by the normalized expression of *CD34* for each cell. **(d)** Hashing of PBMCs

demonstrates compatibility of PIP-seq with barcoded antibodies.

Extended data Figure 7. (a) Gene expression for each sgRNA within an allelic series for all

genes in the CRISPRi library. Each sgRNA is ordered from predicted high to low knockdown

efficiency. Non-targeting sgRNA are denoted as "Null." **(b)** The relationship between gene

expression and predicted knockdown of each gene. Expected changes in transcription across the

allelic series were prominent in highly expressed genes. p-value represents the significance of

the generalized additive model relating gRNA identity to knockdown efficiency for each gene.

Extended data Figure 8. Identification of Gefitinib-specific transcriptional responses in cancer

cell lines. **(a)** Violin plots of median expression values for selected differentially expressed

genes. **(b)** The expression of selected differentially expressed genes superimposed on H1975 and

PC9 cell clusters.

Extended data Figure 9. Clinical flow cytometry and corresponding antibody derived tag

(ADT) data for patients 65 and 873 with mixed phenotypical acute leukemia (MPAL).

Extended data Figure 10. Analysis of PIP-seq data from MPAL Patient #65. **(a)** Integration of

replicates and time points. **(b)** Correlation between the number of cells in each cluster before and

after relapse identifies expansion of clusters 1,4, and7. **(c)** Volcano plots showing differential

gene expression, by cluster, between t1 (before treatment) and t2 (after relapse).

Extended data Figure 11. Analysis of PIP-seq data from MPAL Patient #873. **(a)** Integration of

replicates and time points. **(b)** Correlation between the number of cells in each cluster before and

after treatment identifies the expansion of clusters 3 and 5. **(c)** Volcano plots showing differential

gene expression, by cluster, between t1 (before treatment) and t2 (after relapse).

Tables

Supplementary Table 1. Oligonucleotides used in this study.

Supplementary Table 2. Breast tissue marker genes **(a)** used for cluster cell type identification

and **(b)** identified using differential expression between clusters (FindAllMarkers Seurat v4.1.0,

only.pos=FALSE, min.pct = 0.25, logfc.threshold = 0.5, test.use='bimod').

Supplementary Table 3. Clinical characteristics of study participants.

Supplementary Table 4. Marker genes for patient 65 clusters identified using combined t1 and

t2 timepoints (FindAllMarkers Seurat v4.1.0, only.pos=TRUE, min.pct = 0.25, logfc.threshold =

0, test.use='bimod').

Supplementary Table 5. Marker genes for patient 873 clusters identified using combined t1 and

t2 timepoints (FindAllMarkers Seurat v4.1.0, only.pos=TRUE, min.pct = 0.25, logfc.threshold =

0, test.use='bimod').

Supplementary Table 6. Sequencing metrics.

Supplementary Table 7. PBMC marker genes (a) used for cluster-based cell type identification

and (b) identified using differential expression between clusters.

Acknowledgments: We thank members of the Abate laboratory for helpful advice and

discussions. **Funding:** ICC was supported by K22AI152644 and DP2AI154435 from the NIH.

This work was supported by grants 1R44GM145185, 1R43CA239978-01 and 1R43GM137648-

01A1 to Fluent BioSciences. MJ was supported by K99GM130964 from the NIH. JMR was

supported by F31NS115380 from the NIH. JSW was supported by NIH 1RM1 HG009490-01

and Howard Hughes Medical Institute. CCS was supported by the Damon Runyon Cancer

Research Foundation (CI-99-18), the American Cancer Society (132032-RSG-18-063-01-TBG),

and the Leukemia & Lymphoma Society.

Author contributions: Conceptualization, ICC, ARA; Methodology, ICC, KMF, RHM, CD,

ARA; Investigation, ICC, KMF, YX, DW, KTP, CDA, AO, JQZ, PH, JSAI, VK, CP; Formal

Analysis, ICC, CH, AM-Z; Resources KMF, RHM, JMR, JSW, CCS, ZJG, ARA; Manuscript

preparation, ICC, CCS, CACP, ARA; Supervision, ICC, KMF, RHM, CCS, ZJG, ARA.

Competing interests: ARA filed a patent related to templated emulsification and is a founder of

Fluent Biosciences. ICC consults for Fluent Biosciences and is on its Scientific Advisory Board.

KMF, RM, YX, CH, AO, PH, JSAI, JQZ, AM-Z, CDA, AO, JQZ, PH, and JI are employees at

Fluent Biosciences and are working to commercialize the PIP-seq technology. MJ consults for

Maze Therapeutics and Gate Biosciences. JMR consults for Maze Therapeutics and Waypoint

Bio. JSW declares outside interest in 5 AM Venture, Amgen, Chroma Medicine, DEM

Biosciences, KSQ Therapeutics, Maze Therapeutics, Tenaya Therapeutics, Tessera Therapeutics

and Velia Therapeutics.

Data and materials availability: Sequencing data were deposited into GEO SuperSeries

accession number GSE202919. The Reviewer access token is: ezipsugojfuvvql. All code will be

made available from the corresponding author upon reasonable request.

References

1   Eberwine, J. *et al.* Analysis of gene expression in single live neurons. *Proc Natl Acad Sci*
U S A **89**, 3010-3014 (1992). https://doi.org:10.1073/pnas.89.7.3010

2   Tang, F. *et al.* mRNA-Seq whole-transcriptome analysis of a single cell. *Nature Methods*
6, 377-382 (2009). https://doi.org:10.1038/nmeth.1315

3   Hagemann-Jensen, M. *et al.* Single-cell RNA counting at allele and isoform resolution
using Smart-seq3. *Nat Biotechnol* **38**, 708-714 (2020). https://doi.org:10.1038/s41587-
020-0497-0

4   Hahaut, V., Pavlinic, D., Cowan, C. & Picelli, S. Lightning Fast and Highly Sensitive
Full-Length Single-cell sequencing using FLASH-Seq. *bioRxiv*, 2021.2007.2014.452217
(2021). https://doi.org:10.1101/2021.07.14.452217

5   Picelli, S. *et al.* Full-length RNA-seq from single cells using Smart-seq2. *Nature*
Protocols **9**, 171-181 (2014). https://doi.org:10.1038/nprot.2014.006

6   Picelli, S. *et al.* Smart-seq2 for sensitive full-length transcriptome profiling in single
cells. *Nature Methods* **10**, 1096-1098 (2013). https://doi.org:10.1038/nmeth.2639

7   Ziegenhain, C. *et al.* Comparative Analysis of Single-Cell RNA Sequencing Methods.
Mol Cell **65**, 631-643 e634 (2017). https://doi.org:10.1016/j.molcel.2017.01.023

8   Macosko, E. Z. *et al.* Highly Parallel Genome-wide Expression Profiling of Individual
Cells   Using   Nanoliter   Droplets.   *Cell*   **161**,   1202-1214   (2015).
https://doi.org:10.1016/j.cell.2015.05.002

9   Zilionis, R. *et al.* Single-cell barcoding and sequencing using droplet microfluidics.
Nature Methods **12**, 44-73 (2016). https://doi.org:10.1038/nprot.2016.154

10   Rosenberg, A. B. *et al.* Single-cell profiling of the developing mouse brain and spinal
cord   with   split-pool   barcoding.   *Science*   **360**,   176-182   (2018).
https://doi.org:10.1126/science.aam8999

11   Cao, J. *et al.* Comprehensive single-cell transcriptional profiling of a multicellular
organism. *Science* **357**, 661-667 (2017). https://doi.org:10.1126/science.aam8940

12   Gierahn, T. M. *et al.* Seq-Well: portable, low-cost RNA sequencing of single cells at high
throughput. *Nat Methods* **14**, 395-398 (2017). https://doi.org:10.1038/nmeth.4179

13   McGinnis, C. S. *et al.* MULTI-seq: sample multiplexing for single-cell RNA sequencing
using lipid-tagged indices. *Nat Methods* **16**, 619-626 (2019).
https://doi.org:10.1038/s41592-019-0433-8

14   Stoeckius, M. *et al.* Cell Hashing with barcoded antibodies enables multiplexing and
doublet detection for single cell genomics. *Genome Biol* **19**, 224 (2018).
https://doi.org:10.1186/s13059-018-1603-1

15   Consortium, T. T. S. & Quake, S. R. The Tabula Sapiens: a multiple organ single cell
transcriptomic atlas of humans. *bioRxiv*, 2021.2007.2019.452956 (2022).
https://doi.org:10.1101/2021.07.19.452956

16   Haniffa, M. *et al.* A roadmap for the Human Developmental Cell Atlas. *Nature* **597**, 196-
205 (2021). https://doi.org:10.1038/s41586-021-03620-1

17   Melms, J. C. *et al.* A molecular single-cell lung atlas of lethal COVID-19. *Nature* **595**,
114-119 (2021). https://doi.org:10.1038/s41586-021-03569-1

18   Han, X. *et al.* Construction of a human cell landscape at single-cell level. *Nature* **581**,
303-309 (2020). https://doi.org:10.1038/s41586-020-2157-4

19   Replogle, J. M. *et al.* Mapping information-rich genotype-phenotype landscapes with
genome-scale Perturb-seq. *bioRxiv*, 2021.2012.2016.473013 (2021).
https://doi.org:10.1101/2021.12.16.473013

20   Heath, J. R., Ribas, A. & Mischel, P. S. Single-cell analysis tools for drug discovery and
development. *Nat Rev Drug Discov* **15**, 204-216 (2016).
https://doi.org:10.1038/nrd.2015.16

21   Cao, J. *et al.* The single-cell transcriptional landscape of mammalian organogenesis.
Nature **566**, 496-502 (2019). https://doi.org:10.1038/s41586-019-0969-x

22   Utada, A. S., Fernandez-Nieves, A., Stone, H. A. & Weitz, D. A. Dripping to jetting
transitions in coflowing liquid streams. *Physical review letters* **99**, 094502 (2007).
https://doi.org:10.1103/PhysRevLett.99.094502

23   Clark, I. C. & Abate, A. R. Microfluidic bead encapsulation above 20 kHz with triggered
drop formation. *Lab Chip* **18**, 3598-3605 (2018). https://doi.org:10.1039/c8lc00514a

24   Datlinger, P. *et al.* Ultra-high-throughput single-cell RNA sequencing and perturbation
screening with combinatorial fluidic indexing. *Nat Methods* **18**, 635-642 (2021).
https://doi.org:10.1038/s41592-021-01153-z

25 Aicher, T. P. *et al.* Seq-Well: A Sample-Efficient, Portable Picowell Platform for
Massively Parallel Single-Cell RNA Sequencing. *Methods Mol Biol* **1979**, 111-132
(2019). https://doi.org:10.1007/978-1-4939-9240-9_8

26 Amini, S. *et al.* Haplotype-resolved whole-genome sequencing by contiguity-preserving
transposition and combinatorial indexing. *Nature Genetics* **46**, 1343-1349 (2014).
https://doi.org:10.1038/ng.3119

27 Cusanovich, D. A. *et al.* Multiplex single cell profiling of chromatin accessibility by
combinatorial cellular indexing. *Science* **348**, 910-914 (2015).
https://doi.org:10.1126/science.aab1601

28 Ding, J. *et al.* Systematic comparison of single-cell and single-nucleus RNA-sequencing
methods. *Nat Biotechnol* **38**, 737-746 (2020). https://doi.org:10.1038/s41587-020-0465-8

29 Hatori, M. N., Kim, S. C. & Abate, A. R. Particle-Templated Emulsification for
Microfluidics-Free Digital Biology. *Anal Chem* **90**, 9813-9820 (2018).
https://doi.org:10.1021/acs.analchem.8b01759

30 Delley, C. L. & Abate, A. R. Modular barcode beads for microfluidic single cell
genomics. *Sci Rep* **11**, 10857 (2021). https://doi.org:10.1038/s41598-021-90255-x

31 Murrow, L. M. *et al.* Changes in epithelial proportions and transcriptional state underlie
major premenopausal breast cancer risks. *bioRxiv*, 430611 (2020).
https://doi.org:10.1101/430611

32 Datlinger, P. *et al.* Pooled CRISPR screening with single-cell transcriptome readout. *Nat*
Methods **14**, 297-301 (2017). https://doi.org:10.1038/nmeth.4177

33 Dixit, A. *et al.* Perturb-Seq: Dissecting Molecular Circuits with Scalable Single-Cell
RNA Profiling of Pooled Genetic Screens. *Cell* **167**, 1853-1866 e1817 (2016).
https://doi.org:10.1016/j.cell.2016.11.038

34 Replogle, J. M. *et al.* Combinatorial single-cell CRISPR screens by direct guide RNA
capture and targeted sequencing. *Nat Biotechnol* **38**, 954-961 (2020).
https://doi.org:10.1038/s41587-020-0470-y

35 Jost, M. *et al.* Titrating gene expression using libraries of systematically attenuated
CRISPR guide RNAs. *Nat Biotechnol* **38**, 355-364 (2020).
https://doi.org:10.1038/s41587-019-0387-5

36   Lin, J. C. *et al.* TROP2 is epigenetically inactivated and modulates IGF-1R signalling in
lung   adenocarcinoma.   *EMBO   Mol   Med*   **4**,   472-485   (2012).
https://doi.org:10.1002/emmm.201200222

37   Malumbres, M. CDK4/6 Inhibitors resTORe Therapeutic Sensitivity in HER(2)(+) Breast
Cancer. *Cancer Cell* **29**, 243-244 (2016). https://doi.org:10.1016/j.ccell.2016.02.016

38   Alexander, T. B. *et al.* The genetic basis and cell of origin of mixed phenotype acute
leukaemia. *Nature* **562**, 373-379 (2018). https://doi.org:10.1038/s41586-018-0436-0

39   Kotrova, M. *et al.* Distinct bilineal leukemia immunophenotypes are not genetically
determined. *Blood* **128**, 2263-2266 (2016). https://doi.org:10.1182/blood-2016-07-
725861

40   Stoeckius, M. *et al.* Simultaneous epitope and transcriptome measurement in single cells.
Nat Methods **14**, 865-868 (2017). https://doi.org:10.1038/nmeth.4380

41   Barna, M. *et al.* Suppression of Myc oncogenic activity by ribosomal protein
haploinsufficiency. *Nature* **456**, 971-975 (2008). https://doi.org:10.1038/nature07449

42   Kang, J. *et al.* Ribosomal proteins and human diseases: molecular mechanisms and
targeted   therapy.   *Signal   Transduct   Target   Ther*   **6**,   323   (2021).
https://doi.org:10.1038/s41392-021-00728-8

43   Kampen, K. R., Sulima, S. O., Vereecke, S. & De Keersmaecker, K. Hallmarks of
ribosomopathies.   *Nucleic   Acids   Res*   **48**,   1013-1028   (2020).
https://doi.org:10.1093/nar/gkz637

44   Fancello, L., Kampen, K. R., Hofman, I. J., Verbeeck, J. & De Keersmaecker, K. The
ribosomal protein gene RPL5 is a haploinsufficient tumor suppressor in multiple cancer
types. *Oncotarget* **8**, 14462-14478 (2017). https://doi.org:10.18632/oncotarget.14895

45   Rao, S. *et al.* Inactivation of ribosomal protein L22 promotes transformation by induction
of   the   stemness   factor,   Lin28B.   *Blood*   **120**,   3764-3773   (2012).
https://doi.org:10.1182/blood-2012-03-415349

46   De Keersmaecker, K. *et al.* Exome sequencing identifies mutation in CNOT3 and
ribosomal genes RPL5 and RPL10 in T-cell acute lymphoblastic leukemia. *Nat Genet* **45**,
186-190 (2013). https://doi.org:10.1038/ng.2508

47   Chen, C. *et al.* Oxidative phosphorylation enhances the leukemogenic capacity and
resistance to chemotherapy of B cell acute lymphoblastic leukemia. *Sci Adv* **7** (2021).
https://doi.org:10.1126/sciadv.abd6280

48   Nelson, M. A. *et al.* Intrinsic OXPHOS limitations underlie cellular bioenergetics in
leukemia. *Elife* **10** (2021). https://doi.org:10.7554/eLife.63104

49   Lobrich, M. & Jeggo, P. A. The impact of a negligent G2/M checkpoint on genomic
instability and cancer induction. *Nat Rev Cancer* **7**, 861-869 (2007).
https://doi.org:10.1038/nrc2248

50   Didier, C. *et al.* G2/M checkpoint stringency is a key parameter in the sensitivity of AML
cells to genotoxic stress. *Oncogene* **27**, 3811-3820 (2008).
https://doi.org:10.1038/sj.onc.1211041

51   Demaree, B. *et al.* Joint profiling of DNA and proteins in single cells to dissect genotype-
phenotype associations in leukemia. *Nat Commun* **12**, 1583 (2021).
https://doi.org:10.1038/s41467-021-21810-3

52   Shahi, P., Kim, S. C., Haliburton, J. R., Gartner, Z. J. & Abate, A. R. Abseq: Ultrahigh-
throughput single cell protein profiling with droplet microfluidic barcoding. *Sci Rep* **7**,
44447 (2017). https://doi.org:10.1038/srep44447

53   Gaiti, F. *et al.* Epigenetic evolution and lineage histories of chronic lymphocytic
leukaemia. *Nature* **569**, 576-580 (2019). https://doi.org:10.1038/s41586-019-1198-z

54   Ma, S. *et al.* Chromatin Potential Identified by Shared Single-Cell Profiling of RNA and
Chromatin. *Cell* **183**, 1103-1116 e1120 (2020). https://doi.org:10.1016/j.cell.2020.09.056

55   Hao, Y. *et al.* Dictionary learning for integrative, multimodal, and scalable single-cell
analysis. *bioRxiv*, 2022.2002.2024.481684 (2022).
https://doi.org:10.1101/2022.02.24.481684

56   Ghazanfar, S., Guibentif, C. & Marioni, J. C. StabMap: Mosaic single cell data
integration using non-overlapping features. *bioRxiv*, 2022.2002.2024.481823 (2022).
https://doi.org:10.1101/2022.02.24.481823

57   Hao, Y. *et al.* Integrated analysis of multimodal single-cell data. *Cell* **184**, 3573-3587
e3529 (2021). https://doi.org:10.1016/j.cell.2021.04.048

58   Squair, J. W. *et al.* Confronting false discoveries in single-cell differential expression.
Nat Commun **12**, 5692 (2021). https://doi.org:10.1038/s41467-021-25960-2

59   Hughes, T. K. *et al.* Second-Strand Synthesis-Based Massively Parallel scRNA-Seq
Reveals Cellular States and Molecular Features of Human Inflammatory Skin
Pathologies. *Immunity* **53**, 878-894 e877 (2020).
https://doi.org:10.1016/j.immuni.2020.09.015

1024 60 De Rop, F. V. *et al.* Hydrop enables droplet-based single-cell ATAC-seq and single-cell
1025   RNA-seq using dissolvable hydrogel beads. *Elife* **11** (2022).
1026   https://doi.org:10.7554/eLife.73971

1027 61 Mejstrikova, E. *et al.* Prognosis of children with mixed phenotype acute leukemia treated
1028   on the basis of consistent immunophenotypic criteria. *Haematologica* **95**, 928-935
1029   (2010). https://doi.org:10.3324/haematol.2009.014506

1030 62 Takahashi, K. *et al.* Integrative genomic analysis of adult mixed phenotype acute
1031   leukemia delineates lineage associated molecular subtypes. *Nat Commun* **9**, 2670 (2018).
1032   https://doi.org:10.1038/s41467-018-04924-z

1033 63 Granja, J. M. *et al.* Single-cell multiomic analysis identifies regulatory programs in
1034   mixed-phenotype acute leukemia. *Nat Biotechnol* **37**, 1458-1465 (2019).
1035   https://doi.org:10.1038/s41587-019-0332-7

1036 64 Ebright, R. Y. *et al.* Deregulation of ribosomal protein expression and translation
1037   promotes breast cancer metastasis. *Science* **367**, 1468-1473 (2020).
1038   https://doi.org:10.1126/science.aay0939

1039 65 Khot, A. *et al.* First-in-Human RNA Polymerase I Transcription Inhibitor CX-5461 in
1040   Patients with Advanced Hematologic Cancers: Results of a Phase I Dose-Escalation
1041   Study. *Cancer Discov* **9**, 1036-1049 (2019). https://doi.org:10.1158/2159-8290.CD-18-
1042   1455

1043 66 DeJournette, C. J. *et al.* Creating biocompatible oil-water interfaces without synthesis:
1044   direct interactions between primary amines and carboxylated perfluorocarbon surfactants.
1045   *Analytical Chemistry* **85**, 10556-10564 (2013). https://doi.org:10.1021/ac4026048

1046 67 Labarge, M. A., Garbe, J. C. & Stampfer, M. R. Processing of human reduction
1047   mammoplasty and mastectomy tissues for cell culture. *J Vis Exp* (2013).
1048   https://doi.org:10.3791/50011

1049 68 Patro, R., Duggal, G., Love, M. I., Irizarry, R. A. & Kingsford, C. Salmon provides fast
1050   and bias-aware quantification of transcript expression. *Nat Methods* **14**, 417-419 (2017).
1051   https://doi.org:10.1038/nmeth.4197

1052 69 Srivastava, A., Malik, L., Smith, T., Sudbery, I. & Patro, R. Alevin efficiently estimates
1053   accurate gene abundances from dscRNA-seq data. *Genome Biol* **20**, 65 (2019).
1054   https://doi.org:10.1186/s13059-019-1670-y

1055   70   Lun, A. T. L. *et al.* EmptyDrops: distinguishing cells from empty droplets in droplet-
1056   based   single-cell   RNA   sequencing   data.   *Genome   Biol*   **20**,   63   (2019).
1057   https://doi.org:10.1186/s13059-019-1662-y
1058

a Particle templated emulsification

Barcoded templates
Heat-activated lysis reagents
Single cells

Add oil   Vortex (~ 1 min)

b Triggered lysis, mRNA capture

Oil
barcoded templates
lysed cell
Lysis reagents

c Reverse transcription

polyA - mRNA
barcode UMI polyT   cDNA   cTSO

d Library preparation

Whole Transcriptome Amplification

TSO

Tagmentation

Library PCR

P5 Nextera Barcode UMI   cDNA   Nextera P7

e Cell Number Scalability

500ul tube (35ul PIPs ~ 4k cell input)   15 mL Falcon (2 mL PIPs ~ 250k cell input)   50 mL Falcon (10mL PIPs ~1M cells cell input)

f Sample Number Scalability

96 well   384 well   1536 well

g scRNA-seq Scalability

Number of samples processed

1k
100
10
1

PIPseq

Combinatorial Indexing

Microfluidics

Plates

Microwells

10 100 1k 10k 100k
Number of cells processed per sample

Figure 1

Figure 2

Figure 3

Figure 4

Figure 5

Extended data Figure 1

Extended data Figure 2

Extended data Figure 3

Extended data Figure 4

Extended data Figure 5

Extended data Figure 6

Extended data Figure 7

Extended data Figure 8

Extended data Figure 9

Patient #65

Extended data Figure 10

Patient #873

log2FoldChange (t2/t1)

up in t1 ⟵⟶ up in t2

Extended data Figure 11