Patterning High to Low Heterogenous Fluid Shear Stress Landscapes with Multiphoton Inner Laser Lithography (MILL) for Live Cell Adhesion and Translocation

Yean J. Lim[a,b], Junxiang Zhang[a,b], Hanqi Lin[a,b], Tienan Xu[a,b], Yongxiao Li[a,b], Zhiduo Zhang[a,b], Sarah M. Hicks[a], Ivan Chudinov[c,d], Dmitry Nechipurenko[c], Elizabeth E. Gardiner[a], Woei M. Lee[a,b]

[a] Division of Genomic Sciences and Cancer, The John Curtin School of Medical Research, The Australian National University, Canberra, ACT 2601, Australia

[b] ACRF INCITe Centre - ANU Node, The John Curtin School of Medical Research, The Australian National University, Canberra, ACT 2601, Australia

[c] Faculty of Physics, Lomonosov Moscow State University, Moscow, Russia

[d] Moscow Institute of Physics and Technology, Moscow, Russia

Abstract: Heterogenous fluid shear stress is known to provide mechanical cues for cell adhesion and translocation. To assemble 3D microstructures using current fabrication methods in a single channel and recapitulate in vivo heterogenous fluid flow would require hours of fabrication and specialized equipment. Inspired by the traditional art form of inside painting, we developed a technique for 3D fabrication of micro-patterned flow channels and mixed in vivo fluid flow in a matter of minutes. We termed this technique Multiphoton Inner Laser Lithography (MILL). We further showed that when combined with adaptive optics, MILL is compatible with both flat and curved channel shapes. MILL recapitulated in vivo tissue topology and 3D fluid flow within tissue stroma (low fluid shear, $0 - 3.5$ dynes/cm^2) and blood vessel (high fluid shear, $0 - 80$ dynes/cm^2). We demonstrate fibroblast cell and platelets adhere and translocate differently between laminar flow patterns that are homogenous versus heterogeneous in real time. Parallel strips of MILL channels were assembled for simultaneous platelet function test to quantify the efficacy of an antithrombotic GPVI Fab (~2000 microthrombi per test). The MILL technique can be readily reproduced *in vivo* fluid flow in minutes and benefit preclinical screening of drug pharmacokinetics.

Significant points

- *MILL channels* are made from commercially available off the shelf components using a standard multiphoton imaging system
- Varying degrees of *in vivo* heterogenous laminar flow is shown to directly influence cells translocating on thinly coated collagen surfaces
- Parallel strips of *MILL* channels were assembled for simultaneous platelet function tests (~2000 microthrombi per test).

Keywords: microfluidics, lithography, fluid shear, fibroblast, thrombus

Introduction

Extracellular fluid flowing through porous tissue matrix and blood vessels generates an uneven distribution of laminar flow rates and imposes heterogenous fluid shear stress on surrounding adherent cell population. Laminar flows are often assigned a single shear rate [1] because it exhibits a parabolic profile across a uniform smooth channel. *In vivo* flow through living tissue encounters multiple barriers (e.g. surface protrusions, cells and extracellular matrix) that alter the parabolic flow, creating different degrees of laminar flow in the interstitium and blood vessels. Hence, 'laminar flow' does not fully encompass *in vivo* flow complexity [2]. Fluid flowing from the heart to interstitial tissue generates a range of laminar fluid shear landscapes ranging from $0 - 80$ dynes/cm^2. We propose that laminar flow landscapes possess mixed laminar shear which can be referred to as 'heterogeneous flow'. Simple flow landscapes with fixed laminar shear are therefore referred as 'homogeneous flow'. While it has been shown that *in vivo* heterogenous fluid flow varies from the low fluid shear regimes ($0 - 3.5$ dynes/cm^2) in interstitial tissue matrix to high fluid shear regimes ($0 - 80$ dynes/cm^2) in vasculature [3], there has been limited studies into *in vivo* fluid motion that regulate cell adhesion and migration [2,4,5]. We argue that this is also because of lack of direct 3D flow channel fabrication methods to adequately reflect complex heterogeneous *in vivo* flow environment.

All cells have evolved mechanisms to sense and navigate through heterogenous fluid shear landscapes using mechanosensitive signaling receptors such as Piezo and transient receptor potential (TRP) family proteins [3,6]. More prominently, endothelial cells [3], fibroblast [6], platelets [7] and immune cells [8] residing in these landscapes require Piezo or TRP to migrate and activate. Heterogeneous fluid shear landscapes are responsible for triggering intracellular signaling, transduction[9] and cell migration[10] and are important in maintaining the homeostasis of tissue niches [11], initiating tissue repair [12], bone development [13], cancer metastasis [14,15], angiogenesis [4], regulating cardiac fibrosis and impacts on thrombosis [16]. Despite this, the majority of tissue-on-chip [17] or organ-on-chip [18] systems are generally designed with homogeneous fluid shear stress (FSS) in soft lithographic hollow channels lined by a single layer of cultured cells or tissue. Complex 3D microstructures fabricated in a single microfluidic channel are necessary to reproduce a mixture of fluid shear conditions present in *in vivo* microenvironments. Common fabrication methods to create microstructure patterning in a single flow channel either uses soft lithography (Figure 1 A) i)) or fused deposition (Figure 1 A) ii)). Both of the techniques require multiple stepwise deposition of curable resin or polydimethylsiloxane (PDMS) respectively, onto open two-dimensional surfaces [19] that often require hours of preparation and specialized instruments.

Aside from flexibility in 2D micropatterning, high optical transparency (>90%) and surface

functionalization are essential in molecular and cellular assays [20] for genomics, proteomics and

phenotypic quantification [21]. There are significant fabrication hurdles to develop 2D soft lithography

techniques [22] for fully enclosed 3D surfaces that has high optical transparency. (1) Soft lithography

processes are optimized for open and exposed flat surfaces that require passivating steps [23], (2) soft

lithography requires specialized equipment (UV lamp, thermal heating oven, desiccator, plasma

cleaner) [24], (3) PDMS variably absorbs different hydrophobic molecules and so confound studies

involving the delivery of small molecules into cells [25]. Taken together, standard soft lithography

involves hours of labor-intensive, repeated set of steps (Figure 1A) [26] to achieve 3D microstructures.

To advance cell biology questions incorporating complex heterogeneous *in vivo* flow environment, we

propose to develop a full 3D lithography method [27,28] that is directly compatible with enclosed 3D flow

channels [29]. The goal of direct 3D modification in an enclosed flow channel is to mimic *in vivo* flow

conditions of different microenvironments. In the tissue microenvironment, the 3D surface roughness

influences the trajectory of fluid flow [2,5] that is important in platelet production [1] , transmural nutrient

delivery [30] atherosclerosis [5] and tumor metastasis [14].

In this study, we describe a direct 3D fabrication method, Multiphoton Inner Laser Lithography

(*MILL*), that modifies glass capillaries [31,32] for use in a microfluidic flow assay within 1 hour (Figure

1A, iii) with micrometer resolution. In doing so, we, for the first time, gain access to *in vivo*, 3D

heterogenous fluid flow landscapes present in vasculature and stroma. *MILL* is inspired by the art form

of inside painting [33], that uses multiphoton absorption to sculpt the inner surfaces of transparent glass

capillaries using doped photocurable ultraviolet resins [34]. *MILL* is developed with a standard

multiphoton imaging system without the need for specialized 3D laser photolithography platforms [35].

We further showed that adaptive optics *MILL* removes spatial optical aberrations in curved thick

cylindrical glass capillary tubes. For *MILL* capillary assays to match with traditional microfluidics, we

also describe a rapid hermetical sealing protocol using standard UV adhesive. The UV assisted bonding

protocol allows standard silicone fluidic tubes to connect to *MILL* glass capillary tubes and establish a

robust fluidic flow assay. A soft silicone capillary gripper is developed to hold multiple channels for

high throughput measurements.

With *MILL* capillaries, we first recapitulate the dynamics of cell-tissue interactions induced by

homogeneous and heterogeneous fluid shear under low FSS ($0 - 3.5$ dynes/cm^2) in interstitial tissue and

high FSS ($0 - 81$ dynes/cm^2) in vasculature. We evaluate these fluid landscapes using cells resident to

these tissue (*i.e.* fibroblasts in interstitial tissue and platelets in blood vessels) to map the direction of

cell translocating after exposure to heterogenous flow. Fibroblasts adhered by integrin binding and

clustering on three-dimensional extracellular matrix (ECM) and initiated cytoskeletal reorganization to alter their physical shape [36] and migrate along mechanical and chemical cues. Heterogenous FSS serves as a distinctive mechanical cue (separated from material stiffness), which has previously been observed to stimulate migratory protomyofibroblast phenotype after integrin β1 activation on collagen fibrils [37-39]. *In vivo* results have shown that fibroblast cells are capable of sensing homogenous shear gradients [6]. However, little is known about how fibroblast cells proliferate and differentiate into protomyofibroblasts in irregular interstitial flow conditions. We first established a long-term (12 hrs) live cell imaging protocol using regular capillary tubes that pinpointed the emergence of migratory phenotype resembling protomyofibroblasts [40,41] under interstitial fluid flow (< 1 dyn/cm^2) first observed by Ng *et al* [39]. Due to lack of molecular markers in protomyofibroblasts (low α-smooth muscle actin (SMA)) [40], we focused on migratory responses of fibroblast and alignment of actin cytoskeleton stress fibers [42]. Using the same interstitial flow rates, we explored heterogeneous fluid flow conditions that mimic tissue microarchitecture or niche environments [11,43] and the relationship with fibroblast activation. Using *MILL*, we further explored how adherent fibroblast cells sense and respond to shear gradients [44].

Existing fluid shear assays are often conducted on rectangular flow channels that do not faithfully replicate the circular geometry in blood vessels [45]. We applied *MILL* to incorporate structures into cylindrical capillary tubes to mimic the appropriate heterogeneous fluid dynamics found in the vasculature [46]. Using adaptive optics, we showed that a *MILL* cylindrical tube can take on complex geometries [47]. To simulate the heterogeneous FSS in cylindrical flow channels, we modelled and designed asymmetrical microstructures [48] that mimic small and large stenoses. Since *MILL* capillary tubes are thin, optically transparent, and smooth, they are highly suited to perform a range of cellular imaging (holographic, confocal, multiphoton and total internal reflectance microscopy) to visualize and quantify several multiple biological markers to assess underlying biological functions.

Results

Fibroblast translocation requires integrin engagement and is directly modulated by homogenous FSS

While fibroblasts can proliferate through direct sensing of varying matrix stiffness (1 to 5 kPa [49]) and ligand concentrations, their ability to respond to FSS within interstitium has so far only been inferred [37,39,41,42]. To determine the extent to what extent FSS would influence cell translocation, we subjected a murine fibroblast L929 cell line to different levels of laminar fluid shear from 0 to 3.5 dynes/cm^2 and that mimics fluid shear stress found in interstitial spaces [39]. We then investigated the influence of ligand binding on cell shape and migratory behavior of fibroblast [40-42], where a receptor-ligand adhesion (e.g. integrin-collagen binding) supports increased cell resistance to shear forces over

non-specific (e.g. surface charge) interactions. In our experiments, the substrate coating thickness (Poly-

L-Lysine (PLL) and collagen fibrils) is a third of the cell thickness, which confers limited force for

mechanotransduction. Cells adhering onto PLL or collagen fibrils were exposed to 12 hours of either

continuous FSS exposure in transparent capillary tubes or no shear on a coated glass bottom culture

dish (see Methods). A portion of cell adhesion could also be accounted by the surface adsorption of

fibronectin derived from cell culture serum in both conditions. We quantified migratory patterns and

cell shape (determined by cell ellipticity) over 12 hours at 10 second intervals using label-free

quantitative phase microscopy (QPM) [50] (see Methods), which provides high resolution volumetric

morphological profiling. Cell ellipticity (Figure 2 B) i)) quantifies the transition of a cell state from

resting to adhesive [51]. Higher ellipticity values indicate proadhesive cells that are more engaged with

ligand-coated surfaces. Under all conditions of shear stress, immobilized collagen appeared to elicit

stronger adhesiveness than PLL. The results also suggest that adhesion dynamics can be directly

modulated by homogenous FSS alone and is independent of the type of ligand coating. At a shear stress

of 1.7 and 2.6 dynes/cm^2, cell adhesiveness was observed to be low with mean ellipticity of 0.3 and 0.5

for PLL and collagen coatings, respectively (Figure 2 B) ii)). On PLL, cell adhesiveness is highest at

0.9 dynes/cm^2 with peak mean ellipticity of 0.65. Conversely, cell adhesion on collagen remained high

with a peak mean ellipticity of 0.8 at 0.9 dynes/cm^2. We observed that mean ellipticity reached 0.75 at

3.5 dynes/cm^2 on collagen. Single cell tracking showed that 3.5 dynes/cm^2 FSS led to a rapid (<20 sec)

expansion of cell perimeter followed by a loss of motion, which we ascribed to cell rupture. Importantly,

cell rupture was observed when collagen was the ligand but not PLL (Supplementary Figure S1). In

fact, cells adherent on PLL did not rupture over the full 12 hrs. Based on the results in Figure 2 B) ii),

integrin adhesion to collagen under FSS [39] resulted in a 2- to 3-fold higher surface adhesion when

compared with PLL. After cells have adhered to the ligand surface, we tracked the mobility of cells.

Under FSS, we repeatedly observed formation of cell aggregates with either ligand coatings. Hence,

automated tracking of cell clusters was performed using Trackmate (ImageJ). The results in Figure 2B)

iii) shows that cell mobility reached a maximum speed of 9 µm/hr on collagen coating under shear stress

of 0.9 dynes/cm^2 that is around twice the speed when compared with PLL coating. This increase was

abrogated at higher FSS (1.7 dynes/cm^2).

Figure 2C) i) to iv) shows QPM images from a time lapse recording taken at 2-hour intervals.

On PLL coating without FSS, most cells (> 80%, n = 37) remained stationary, with a small population

that spread and migrated (as shown Figure 2 C) i)), often remaining separated. However, cells exposed

to 0.9 dynes/cm^2 shear stress would spontaneously cluster into multiple discrete aggregates, indicating

that intercellular adhesion increased under homogenous FSS. Within each cell aggregate, as shown in

Figure 2C) ii), we observed that individual cells would 'roll' and organize in the direction of flow.
Without flow, cells adhering on collagen-coated capillaries extends membrane protrusions (filopodial)
matching half its own cell diameter over several hours (Figure 2 C) iii)). In contrast, as shown in Figure
2 C) iv) and Supplementary Video M1, the first hour of shear exposure (0.9 dynes/cm^2) to collagen-
adherent cells stimulated an increase in motility of up to 9 µm/hr. Together the data indicates shear
stress of 0.9 dynes/cm^2 stimulates maximum adhesion with the highest rate of mobility, a biophysical
marker of migratory fibroblast types [12]. 0.9 dynes/cm^2 are fluid shear reported in interstitial fluid [39] and
lymph nodes [52], where fibroblasts reside. The presence of collagen and shear are both required to induce
a phenotype consistent to proto-myofibroblast formation.

As shown in Figure 2 A), changes to cell shape require actin polymerization to generate tense
actin fibers. We quantified actin fiber alignment using F-actin labeling (phalloidin) after exposing cells
to fluid shear stress for 12 hrs. We also compared cells that were not exposed to any flow gradient to
quantify the influence of fluid shear. On PLL coating, over 80% of the cells (n = 6) retained a spherical
shape and actin fibers were randomly arranged (*i.e.,* disordered), shown in Figure 2 D), indicating
minimal cytoskeletal rearrangement. Cells adhering to PLL coating that were exposed to homogenous
FSS of 0.9 dynes/cm^2 would generate "fin-like" lamellipodium protrusions whilst remaining weakly
aligned, as shown in Figure 2 D) ii). This is consistent with time lapse imaging results from cell shape
and motility shown in Figure 2 B) ii) and iii). PLL, in general, shows lower amount of cell adhesiveness
and cytoskeleton rearrangements.

On the other hand, cells adhering onto collagen coated surfaces displayed extended protrusions
driven by actin polymerization as shown in Figure 2 D) iii) through integrin engagement [12]. However,
under shear stress, membrane protrusions can span over 20 – 30 µm long, as shown Figure 2 D) iv).
This result indicates that even at low FSS, actin polymerization can be modulated significantly. Next,
we quantified the degree of actin alignment along a chosen axis using morphometric analysis program
developed by Lickert *et al* [51]. In our case, we measure the actin alignment along the flow axis (parallel)
as shown in Figure 2 D) v) and vi). A rainbow color code is assigned to the alignment direction. As
shown in Figure 2 D) v) and vi), the cells on both PLL and collagen coating were polarized to form
actin fibers that align along the flow axis. Actin fibers aligned in parallel to the flow axis will be assigned
value close to unity. By examining single cells (n=4) that are not clustered, we observed a two-fold
higher alignment of actin to fluid flow on collagen fibers (Figure 2 D) vii), P < 0.001), but no significant
increase in in alignment to PLL. Under both absence or presence of fluid shear, we observed ~2-fold
increased alignment in cells on collagen compared to PLL coating (P < 0.01). From these results – cell
shape, motility, and actin alignment along flow, fibroblast differentiation and translocation are both

regulated by integrin engagement and FSS. Additional data on cell viability and time lapse results
(Figure 2C) are available in Supplementary Figure S1 and Supplementary Video M1, respectively.

216
217 *Homogeneous and heterogeneous fluid shear modulates cell motility and actin bundling differently*
218
In living organs, fluid flow within tissue matrix and microvasculature are mostly heterogenous.
Cell-cell interactions are therefore subject to heterogenous shear stresses that can alter signal
transductions by a variety of receptor-ligand systems. In contrast to the homogenous FSS, here we asked
how heterogenous shear stress modulates fibroblast activation and motility. Based on our findings in
Figure 2, we expected heterogeneous shear to form groups of fibroblasts with different levels of
adhesion. To investigate this, we need to design and pattern three dimensional physical obstructions
within microfluidic channels (Figure 1 B) and 2). To test the throughput of *MILL*, we generated a 30
µm-thick herringbone structure used to generate turbulent flow for mixing solutions [53]. The structure
consists of 9 herringbones that span the entire width and 1 mm along the length of a square glass
capillary (lumen width and height: 200 µm) (Figure 1 B)), where one herringbone took ~9 minutes to
complete, demonstrating that *MILL* achieves millimeter scale structures within 1.5 hours.

Having proven the throughput of *MILL*, we designed and constructed rectangular obstructions
on opposite sides of a capillary (Supplementary Figure S2). The obstructions are three-dimensional
microscopic rectangular blocks (W×L: 60 µm × 40 µm) that are spaced at 16 or 24 µm apart with
different heights ranging from 6 to 18 µm as shown in Figure 3 A) i). To visualize the lithography
structures, we dope the NOA81 optical adhesive with the fluorescent dye rhodamine B. From
optimizing dye concentration, we chose a higher dosage of 20 mg/mL, which improved flatness of a
lithography layer (Supplementary Figure S3). The flow profile of a standard capillary tube is first
calibrated with 1 µm fluorescent microspheres suspended 1× phosphate buffered saline (PBS) and
drawn into the *MILL* capillary tube with an automated syringe pump (Harvard Apparatus). Figure 3 A)
ii) shows a cross section of the flow profile indicating a parabolic velocity distribution ranging from 0.2
mm/s to 0.8 mm/s that is typical of laminar flow in a capillary tube. Conversely, in the *MILL* capillary
tubes, we showed heterogenous flow velocity that ranges from a 0.5 to 1.5 mm/s with shear stress ~1.2
dynes/cm^2 (Figure 3 A) iii)). Fluid flow within the cluster of microscopic obstructions was minimal
(<0.5 mm/s) and fastest between obstructions and capillary walls (1.5 mm/s). The same obstructions
were also affected by a shear gradient around 100 µm upstream and downstream. The micro-obstruction
created a highly heterogenous fluid shear that range from 0.2 to 1.5 dynes/cm^2, resembling *in vivo* tissue
niches within interstitial tissue [54].

Next, we determined if the localized regions of shear gradient modulate shape, motility, and
actin cytoskeleton of adherent cells. For this, fibroblast cells (L929) were first seeded and left to adhere
onto thinly-coated collagen channels for the first hour as shown in Figure 3 B) i) and ii) (collagen
coating is shown in Supplementary Figure S2). After which, cells were exposed continuously for up to
12 hours of continuous homogenous and heterogenous fluid shear conditions as shown Figure 3 B) iii)
and iv), respectively. Representative images of live cell QPM images shown in Figure 3 B) iii) and iv)
identified changes in cell shape and more importantly, mobility. Regions of heterogenous fluid shear
showed large changes in cell shape that appear to be sensitive to shear gradient as illustrated in Figure
3 B) v). In Supplementary Video M2, there is a distinct difference in morphology between cells residing
within low shear region ($\sim\Delta0$ dynes/cm^2) and on margins of high shear stresses at $\Delta0.9$ dynes/cm^2. We
then quantified cell motility under fluid shear change ($\Delta0$ and $\Delta0.9$ dynes/cm^2) in Figure 3 B) iii) and
iv), which showed a significant, 2-fold increase in cell motility in heterogenous fluid shear (Figure 3 B)
vi) $P < 0.05$)). Here, we measured motility in stationary cells that actively formed protrusions (as in B)
iv)) and cells that were actively migrating. Next, we used multiphoton imaging to capture the *MILL*
microstructures (Rhodamine B), actin cytoskeleton arrangement (F-actin Phalloidin-Green) as well as
collagen fibrils (white) along homogenous and heterogenous shear gradients. We observed that cells
organize into distinct clusters around collagen fibers (Figure 3 B) vii-ix)). Importantly, extended actin
cytoskeleton appears to coincide with changes in shear stress as shown in Figure 3 B) vii) and viii) and
minimal actin reorganization of the cells that resides within rectangular obstruction Figure 3 B) ix)
(Supplementary Video M2).

We also correlate the direction in which bundles of collagen fibrils align before and after 12
hours FSS exposure, shown in Figure 3 C) i) and ii). Our results indicate both collagen matrix and
adherent fibroblast cells align to fluid shear gradients. Homogeneous FSS modulated fibroblasts appear
to exert more force on the collagen matrix than those exposed to heterogeneous FSS Figure 3 C) i) and
ii)). This behavior is consistent with the proto-myofibroblast phenotype that has been known to exert
forces onto collagen matrix. Because of the high density of clustered cells, it was not possible to quantify
single actin fibers as shown in Figure 2 D). Instead, we measured the distribution of fluorescence
intensity at different compartments of the cell body (Figure 3 C) iii-vi)). The magnitude of fluorescence
intensity indicates bundling of actin fibers in each cell [55]. Without any FSS stimulus, the actin
fluorescence signals appear to be evenly distributed in adherent cells with minimal actin bundling.
Adherent cells exposed to either homogenous and heterogenous FSS possessed a significant ($P < 0.05$)
increase in actin bundles along the cell periphery and elongated cell shapes, where homogeneous shear
induced 58% higher peripheral actin compared to no shear ($P < 0.001$). On the other hand, homogeneous

FSS (shear gradient of Δ0 dynes/cm^2, Figure 3 C) vii), stimulated slightly higher increase (~13 %) of

actin bundles than heterogenous FSS (shear gradient of Δ0.2 and Δ1.2dynes/cm^2, Figure 3 C) vii)).

These findings indicate that lower heterogenous FSS can lead to an increase in actin bundling and cell

adhesion.

284
285 *Aberration correction needed for MILL cylindrical capillary tubes*
286
The effects of FSS are most pronounced on cells marginating along the cylindrical walls of

blood vessels under flow regimes of high FSS (> 10 dynes/cm^2). We next determined if *MILL* could

mimic the influence of homogenous and heterogeneous FSS on a cylindrical wall, where an irregular

microstructure creates a non-symmetrical shear disruption resembling a vascular stenosis [48]. We first

perform computational fluid dynamic (CFD) modelling to calculate the expected flow rates in a

cylindrical tube with inner diameter of 200 µm as shown in Figure 4 A) i) smooth wall and ii) 20%

narrowing (stenosis). Along the walls of the 20% narrowing, the fluid flow velocity increased from 0

to 0.3 mm/s within 5 µm from the surface, compared to 24 µm on a smooth wall. Alteration of shear

forces along an injured wall not only regulates movement of cells and platelets, it also influences the

coagulation pathways involved in thrombus formation [56]. Cylindrical glass tubes are known to degrade

optical performance [57]. To counter this, we used our Raster Adaptive Optics (RAO) method [58] to achieve

diffraction limited *MILL* performance which we termed as Adaptive Optics (AO) *MILL*. Figure 4 B) i)

shows the anticipated written structure of a single pillar without (orange) and with AO (red). We

identified degraded optical performance shown in Figure 4 B) ii) and the amount of optical aberrations

using RAO as shown in Figure 3 B) iii). Figure 4 B) iv) shows that aberration-corrected laser writing

that display sharp edges, measured with QPM. AO *MILL* enables robust photolithography with a ~15%

error in the structure's dimensions as shown Figure 4 C) v) and vi) in width and thickness. We show

that the structure with AO *MILL* width and thickness was not significantly different from writing on a

flat surface with minimal aberration (n=5), demonstrating that AO *MILL* achieves the writing resolution

of the system. We conducted particle tracking velocimetry (PTV), as shown Figure 4 D) at a flow rate

of 1 µl/min to experimentally show that the flow velocities match the expected CFD simulation (Figure

4 A)). We also measured the influence of *MILL* microstructures on the organization of collagen matrix

as shown in Figure 4 E) i) and ii). While collagen fibrils adhered to the stenosis structure (red), the

overall collagen alignment and distribution across the channel without and with stenosis do not differ

significantly (Figure 4 D) iii)).

312
313
314

315   *Platelet aggregate and move under heterogeneous FSS*
316
The relationship between stenosis and FSS is crucial to our understanding of thrombus
formation and blood clotting. Platelets in the vasculature must resist the high FSS force (>10 dynes/cm^2)
by adhering and aggregating to form a thrombus. While platelets are known to be uniquely sensitive to
homogenous FSS, the rolling and mobility of adherent platelets around a stenosis is not well studied.
Platelet adhesion and shape change remain important biophysical markers for thrombus stability and
bleeding disorders [56]. Thus, we next show that AO-*MILL* can reveal adhesion and motility of individual
platelets as they aggregate onto an adhesive ligand (e.g. collagen). We first coated a thin layer collagen
on smooth and AO-*MILL* cylindrical capillary tubes as shown in Figure 5 A) i) and iv), respectively.
Flow velocity of both smooth and AO-*MILL* tubes are shown in Figure 5 A) ii) and v), where AO-*MILL*
generates heterogeneous shear around the stenosis. Each capillary tube inlet and outlet are connected to
a reservoir containing citrated whole blood and a syringe pump, respectively. To form thrombi, the
pump draws whole blood across the collagen-coated channel at arterial shear stress of 81 dynes/cm^2
(shear rate: 1800 s^{-1}). To visualize single platelets during thrombi formation, we incubated whole blood
with anti-CD42a antibody (ThermoFisher Scientific) conjugated to AlexaFluor 594 to label platelet
membrane. To observe thrombus formation in real time and in 3D, we used a video rate multiphoton
microscope system [59] to record a single volume every second and quantify platelets aggregating. Figure
5 A) iii) and vi) shows a maximum projection of a representative image of several island of thrombus
formed after 10 minutes of exposure to fluid shear in two different cylindrical tubes, smooth and stenotic
capillary tubes respectively. Results in Figure 5 A) vi) indicates that heterogenous FSS affect the spatial
organization of the aggregating platelets. Platelet aggregates appear to organize along the shear
direction and within the shear gradient adjacent to the stenosis. The spatial and temporal resolution (1
μm and 20 milliseconds, respectively) [58] in our system permits tracking of individual platelets adhering
on the coated capillary walls (Figure 5 B)). Tracking results show platelets aggregating ~10 μm away
from the stenosis possess an average motility of less than 0.1 µm/s. However, at regions within ~10
µm and downstream to the stenosis, platelets aggregating along the shear stress gradient (55-60
dynes/cm^2) moved along the flow direction at velocities of up to 2 µm/s (Figure 5 B)). We also captured
the 3D trajectory of rolling platelet aggregates as shown in Figure 5 C) i) and ii) and Supplementary
Movie M3.
345
Degree of laminar flow triggers stenosis triggers platelet adhesion and aggregates
Real time evaluation of platelet adhesion and aggregating profile under different vascular FSS
conditions are important functional parameters to determine clotting speed, and platelet dysfunction

under a given fluid shear stress. Next, we extended *MILL* capillary tubes to determine the effects of a

larger stenosis with greater heterogeneity in fluid shear stress. We anticipated the FSS gradients from

the stenosis and formed platelet aggregate collectively amplify the clotting process to create a large

thrombus at the center of the stenosis. To test this, we fabricated a concave structure that mimics an

arteriosclerotic vessel [60] (Figure 6 A) i)) that disrupts flow velocity by an order of magnitude. Using

CFD, we first simulate an input flow of 1 µl/min (1.5 mm/s) that results in a steep gradient of flow

velocity of approximately ~ 0.9 mm/s (Figure 6 A) ii)). In contrast with the small 20% stenosis in the

previous section, the wall flow velocity with concave structure generates a 3-fold higher acceleration.

Using multiphoton microscopy, we measured thrombus formation shown in Figure 6 A) iii) that

displays a single large platelet aggregate formed between the concave structure. Using the image, we

conducted a second CFD model to estimate the velocimetry along with the thrombus as shown in Figure

6 iv) and a follow up PTV flow measurement in Figure 6) v). The formation of a single large thrombus

on an existing stenosis exacerbates the changes in fluid shear to 144 dynes/cm^2 that would

consequentially increase the size of the thrombus and drastically the overall flow gradient. Both

simulation and experimental flow velocities confirm a steep acceleration of flow velocity to 1.8 mm/s

at the thrombus. For the sake of completeness, we further compared the volume of thrombi formed

between homogenous and heterogenous FSS. We observed that homogenous FSS creates a landscape

of evenly distributed thrombus across three fields of view spanning across 1 mm in length (Figure 6 B)

i)). However, in the case of heterogenous FSS, a single platelet aggregate can reach up to 25 µm thick

as shown in the center image of Figure 6 B) ii). The measured distribution of thrombus height also

doubled (4 µm) before the stenosis rather than after the stenosis. The overall volume of thrombi in

heterogenous flow is around 4-fold higher than homogenous flow as shown in Figure 6 B) iii).

Increasing throughput of MILL capillaries for thrombosis screening

Microfluidic devices have both throughput and technical advantages over traditional parallel-

plate chambers and cone-and-plate viscometer [61], such as the precise control of blood flow and the

ability to perform multiple experiments with small blood sample volumes. In particular, microfluidics

have demonstrated a major role for heterogenous FSS in the initiation and proliferation of platelet

aggregation as well as affect antiplatelet therapy on platelet aggregation [61]. To increase the throughput

of these *MILL* capillaries, we developed a simple PDMS clamp (Figure 7 A)) to mount multiple *MILL*

capillaries and perform homogenous and heterogenous FSS assays in parallel. The role of the PDMS

clamp is to secure the capillaries onto a glass slide to prevent sample drifting and rotation, while

providing an imaging window and reservoir for immersion objectives. Capillary inlet and outlet tubings

are connected and sealed with UV glue. The inner fluid shear profile can be modified with structures

using computer aided design and *MILL*. Following the steps in the previous sections, incorporating

multiple capillaries into a single chip allows coating individually or in parallel. Here we focus on

collagen coating as shown in Figure 7 B) that is used to evaluate the variation in coating uniformity.

Figure 7 C) demonstrates thrombus formed within capillaries generating homogenous and heterogenous

FSS using anticoagulated blood treated with an antibody against CD42a, a platelet-specific membrane

protein. Using QPM and automated stage scanning, we can rapidly identify and verify the total thrombus

volume:area ratio, which is an indication of thrombus height and spread. Figure 7 D) shows the

distribution of thrombi formed under shear stress of 81 dynes/cm^2 for 10 mins in four different capillary

tubes, two which have a pair of *MILL* structures described in Figure 5. These structures were spaced

apart by 100 µm, where homogeneous flow is recovered (Figure 4 A) i)).

We subjected preformed thrombi in two tubes to flow of phosphate-buffered saline alone or

containing a monoclonal Fab fragment (10 µg/ml, clone 12A5) to human glycoprotein VI (GPVI) at 81

dynes/cm^2. GPVI is a platelet-specific membrane protein that binds to collagen and fibrin to mediate

platelet adhesion and aggregation[62]. Hence, anti-GPVI Fab treatment will interfere with platelet GPVI-

collagen adhesion and remove newly bound platelets at the periphery of the thrombus [63]. Hence without

GPVI Fab treatment, we expect more platelet aggregation and thrombi with larger volume and area than

treated thrombi, where platelet aggregates are looser. Our results show that homogeneous flow formed

thrombi with area spanning from 10 µm^2 to 1000 µm^2 and volumes from 1 µm^3 to 500 µm^3. GPVI Fab

treatment to preformed thrombi resulted in the majority (76%) of thrombi with area and volume below

10 µm^2 and 100 µm^3, respectively. The results also indicated that GPVI-treated thrombi exhibited a

higher volume:area ratio, which we predict is due to lower adhesion forces altering thrombus

contraction. Under heterogeneous shear, GPVI Fab treatment did not alter thrombi aggregation

compared to homogeneous shear, where thrombi volume:area ratio and spreading were similar, with

87% of thrombi below 10 µm^2. Anti-GPVI Fabs are currently being evaluated as antithrombotic

therapeutics as these reagents can interfere with aggregation[64] or disaggregate formed thrombi [63,65] but

with minimal bleeding risk [66,67]. These differences between homogeneous and heterogeneous shear

indicate shear disruption could modulate platelet adhesion and hence, alter the efficacy of treatments.

411   Discussion

In this study, we demonstrated that *MILL* outpaces (< 1 hour, Figure 1A) and overcomes the

oversimplified laminar flow, 2D geometries of current fabrication techniques. *MILL* capillaries are

accessible *in vitro* systems for studying live cell responses under a uniform (homogeneous) or mixed

(heterogeneous) FSS landscape and are tailored for flat or curved flow chambers. These FSS landscapes

represent diverse *in vivo* physiological environments from low interstitial FSS in tissue interstitial niche

to high fluid shear in vasculature. Importantly, our long-term live cell imaging results showed that

heterogeneous FSS landscape produced fibroblasts with increased motility (Figure 3 B) vi)), but

reduced cell adhesion (lower peripheral actin) when compared to homogeneous FSS (Figure 3 C)).

Heterogenous FSS modulates spatial temporal patterning of platelet aggregates and thrombus formation

(Figure 5 and 6). This result reveals fibroblast sensitivity to gradients in FSS. Overall, *MILL* capillaries

can be easily designed to suit conventional CFD modeling and enable quantitative measurement of

heterogeneous fluid shear stress on adherent cells. The demonstration of parallel multi-*MILL* in Chip

device also shows that it is possible to tailor different fluid shear stress in a high throughput fashion that

could benefit screening of drug efficacy (Figure 7).

Implication to fibrosis and thrombosis

Our results demonstrate that the FSS landscape regulates cell adhesion in fibrosis and

thrombosis. Homogeneous shear increased fibroblast surface adhesion (Figure 2 B) ii)) and actin fiber

alignment to the shear direction (Figure 2 D)) by 2-fold in fibroblasts—a response similar to endothelial

cells exposed to 15 dynes/cm^2 of FSS [68]. However, fibroblasts under heterogeneous FSS resembling

interstitial flow ($\Delta 0.2 - 1.2$ dynes/cm^2) were ~ 3 times more motile (Figure 3 B) vi)) and exhibited 50%

increase in peripheral actin (Figure 3 C)) compared to homogeneous shear. Likewise, heterogeneous

FSS increased platelet translocation (Figure 5 B)) and thrombus contraction by at least 2-folds (Figure

6 B) iii)) compared to homogeneous flow. Therefore, exposing cells to homogeneous or heterogeneous

FSS can lead to different conclusions on how fibroblasts and platelets mobilize during fibrosis and

thrombosis *in vitro* and *in vivo*. As fluid shear in the interstitium and vasculature are more than an order

of magnitude difference, we expect large differences in how cells in both environments sense fluid shear.

To gain better insight into how heterogeneous shear controls these cell decisions, future efforts will

identify the membrane receptors (e.g. transient receptor potential (TRP) and Piezo protein families) and

signaling pathways responsible for sensing FSS in fibroblasts and platelets.

Implication for organ on chip culture: substrate mechanics and nutrient exchange in glass versus
PDMS

An extension of *MILL* is to generate customized niches for organ on a chip approaches using

UV-curable hydrogels [69,70] in glass capillaries. *In vivo*, organs consist of stromal tissue with complex a

micrometer-scale matrix that produces a heterogeneous FSS landscape. Maintaining the fabrication

resolution is important for reproducing this matrix. Hence, we further improve on MILL with adaptive

optics to correct for sample aberrations during fabrication. The 2-fold improvement in precision that

AO MILL (Figure 4 C)) provides will be important for irregular niches such as bone sinusoids [11] and

mechanically flexible scaffolds for cardiac cells contraction [71]. In this study we created a niche

mimicking interstitial tissue with heterogeneous FSS using *MILL* structures spaced apart by 16 or 24

μm (Figure 3 A)). By opting for glass capillaries instead of PDMS, we avoid the non-specific

sequestration of hydrophobic molecules in PDMS, which constitute up to 60% of small molecule drugs

[18], but at the expense of gas permeability [72]. To support long-term cell culture growth in this niche, we

used a HEPES-based buffer that does not require CO_2 and non-sterile conditions for 12-hour flow

experiments (Figure 2 and 3). Restricted gas exchange within the glass capillary tube means fibroblast

differentiation and organ development studies lasting >12 hours will require dedicated fluidic setups to

maintain nutrient supply, sterility, pH, CO_2 and O_2 levels in the culture medium. Existing commercial

fluidic pump systems circumvent these demands by integrating a closed fluid circuit within CO_2

incubators.

Implication for cell adhesion in metastatic diseases.

When fibroblasts were cultured in heterogeneous FSS (*i.e.,* $\Delta 0.2$-1.2 dynes/cm^2, Figure 3 C)),

we observed cell translocation against the flow direction (see Supplementary Video M2). This response

to flow (rheotaxis) is critical for leukocyte rolling [73] in vasculature and metastatic cancer cells [74], which

can migrate toward a blood vessel and against a flow gradient. Epithelial tumors (e.g. breast cancer)

undergo an epithelial to mesenchymal transition (EMT) characterized by gene expression and cell

morphology resembling mesenchymal cells [74]. Cancer-associated fibroblasts (CAF) similarly facilitate

cancer migration by remodeling the surrounding ECM [75]. Elucidating the direct effect of FSS and the

contribution of secondary cell mediators to cancer metastasis will be important questions that *MILL*

capillaries can address. A key question is whether cancer cells sense interstitial, heterogeneous FSS to

trigger EMT and heterogeneity is regulated by other cells including myofibroblasts [12] and macrophages

[76], which can produce, remodel and pull the ECM.

Identifying shear conditions that commit cells to divergent cell fates

Adherent and differentiated cells exist in a niche that supports their survival and proliferation,

but how does the niche mechanically select for the differentiated cell? *MILL* capillaries recreate an

environment resembling interstitial flow with 'niches' of accelerating and decelerating FSS gradients

up to $\Delta 1.2$ dynes/cm^2 (Figure 3 A)). *In vivo* shear conditions feature a mix of diverging and converging

fluid paths, that is insufficiently described as laminar. Hence, we describe these fluid shear landscapes

in *MILL* capillaries as homogeneous or heterogeneous (Figure 8). We show that these niches regulate

fibroblast transition to a proto-myofibroblast state that were up to 3 times more motile (Figure 3 B) vi))

but exhibited 20% less actin bundling (Figure 3 C)) than under homogeneous FSS. Hence, these flow

niches act as a mechanical stimulus and generate a shear 'map' that regulates fibroblast differentiation.

ECM stiffness in cell differentiation is well established [12], but evidence shows that FSS is also essential

for the differentiation of embryonic stem cells [13], endothelial cells [4] and fibroblasts [39]. Future efforts

will assess how shear and chemical (e.g. TGF-β) affects the rate of differentiation using molecular

markers (e.g. α-smooth muscle actin and cadherin-11) and ECM production (e.g. collagen) as markers

of differentiation.

Scaling for high throughput screening of pathophysiological responses

Multiwell plates remain the workhorse platform for *in vitro* high throughput screening (HTS)

assays of cell behavior. However, current fabrication techniques require hours to replicate the in vivo

FSS landscape of stroma and vasculature, that *MILL* achieves in less than an hour. The fabrication speed

sets *MILL* as a versatile tool to prototype complex *in vivo* FSS landscapes within a single capillary tube,

that could not be easily achieved using soft lithography or fused deposition techniques (Figure 2 A).

Using *MILL* we address these restrictions using a standard multiphoton imaging system (Figure 2) to

precisely (~15% error) modify FSS within commercial glass capillary tubes (Figure 4 B) v) and vi)).

We employed 4 capillary tubes in a single chip assay (Figure 7), which can be scaled for even greater

throughput. Incorporating *in vivo* flow parameters to assays will benefit preclinical screening of drug

pharmacokinetic and pharmacodynamic properties [20]. However, as with current HTS methods, machine

automation to control flow and reagent input will be required to achieve larger biological scales.

By assembling unstructured and structured capillaries into a fluidic chip (Figure 7), we

demonstrate high throughput platelet function testing under laminar (*i.e.,* homogeneous, 81 dynes/cm^2)

and *in vivo* (*i.e.,* heterogeneous, $> \Delta 10$ dynes/cm^2) FSS. Considering heterogeneous FSS abolishes the

effect of an antithrombotic drug treatment (Figure 7 D)), coagulation under heterogeneous FSS could

be important to infer a patient's platelet activation status and response to vascular damage or pathology.

Current coagulation assays do not account for heterogeneous shear although these represent *in vivo*

vasculature (e.g. atherosclerosis [60]) that are clinically relevant for patients undergoing cardiac surgery.

Hence, clotting assays that incorporate heterogeneous flow and shear stress could be used as a screening

test for preoperative care and antithrombotic drug selection.

Conclusion

515   *MILL* was shown to recreate both low and high heterogeneous laminar flow rates that are

present in porous tissue matrix and blood vessels. We have shown that *MILL* enables us to study cell

adhesion and spreading that is specific to low FSS ($0 - 3.5$ dynes/cm^2) landscapes in tissue stroma and

high FSS ($0 - 80$ dynes/cm^2) landscapes in blood vessels. Furthermore, we demonstrated that *MILL*

fabrication exceeding current 3D fabrication techniques in terms of speed and cost because it can be

performed within glass capillaries using routine consumables. This makes *MILL* well-suited for high

throughput cell assays to organ-in-a-chip assays. In our study, we have identified differences of cell

translocating under homogeneous and heterogeneous FSS regimes. We first showed that fibroblasts,

which reside in porous tissue matrix, respond to low heterogeneous FSS ($\Delta 0.2$ dynes/cm^2), after

adhering to collagen fibers, by increasing translocation speed by 3-fold over speed under homogeneous

FSS. We also showed that platelets, which initiate repair vessel injury, respond to high heterogeneous

FSS ($\Delta 10$ dynes/cm^2) with increased aggregation by 10-fold. We built upon this knowledge of

heterogeneous FSS and created multi-*MILL* in Chip for drug screen that identified a possible effect of

heterogeneous shear on antithrombotic drug activity. Taken together, the use of *MILL* would expand

the exploration of heterogeneous physiological shear environments for other biological cells type i.e.

macrophages [8], cancer cells [6,77]. *MILL* assays can therefore capture *in vivo* cell responses under both

physiological and pathological fluid shear landscapes and so can be used as pharmacological assays

without resorting to complex microfluidic systems [78].

Methods

<u>Preparation of Capillary Chips</u>

537   For all experiments, borosilicate capillaries were cut to a length of 35 mm (rectangular tubes)

or 85 mm (circular tubes) and mounted onto 75 mm × 26 mm microscopy glass slides. Rectangular

capillaries (W × H; 0.3 mm × 0.1 mm, VitroCom, Mountain Lakes, NJ, USA) were secured onto the

glass slide using NOA81 UV adhesive (Norland Products, Cranbury, NJ, USA) with a UV curing

system (365 nm, Thorlabs, Newton, NJ, USA) at 20% intensity for 5 seconds. To secure circular tubes

(ID: 0.2 mm, VitroCom) onto glass slides, a mold formed with circular capillaries was used to create a

PDMS clamp as shown in Figure 6 A). Capillary ends were inserted into Tygon inlet and outlet tubing

(ID: 0.51 mm, Saint-Gobain, Courbevoie, France), which was sealed with NOA63 UV adhesive

(Norland Products) and UV-cured at 20% intensity for 30 seconds.

 MILL and RAO *MILL*

UV adhesive for *MILL* was prepared by dissolving Rhodamine B (20 mg/mL, Townson &
Mercer, Australia) in NOA81 UV adhesive (Norland Products) by manual mixing and through a
suspension mixer for 2 hours and then injected into borosilicate capillaries with a 1 mL syringe. *MILL*
was performed using a custom-built polygon scanning microscope [59] with a Ti-Sapphire pulse laser
(Spectra Physics, MKS Instruments, Inc., Andover, MA, USA) tuned to 810 nm with a pulse width of
100 fs and repetition rate of 82 MHz. We used an average laser power of 22 mW and 1.2 mW after the
20× water-immersion objective lens (W Plan Apochromatic, 1.00 NA, Zeiss, Germany) for *MILL* and
imaging, respectively. Patterning of microstructures using 2PP was achieved by restricting the scanning
range to 6 µm × 4 µm at the image plane and translating the scanned region laterally with galvo mirrors
to form the structure. Depth of the *MILL* structures was controlled with the sample stage (3DMS, Sutter
Instrument, Novato, CA, USA) and tested for lithography precision (see Figure 3 B) v-vi)). *MILL* was
also replicated on an Olympus FVMPE-RS multiphoton microscope controlled by the Fluoview
software (data not shown). Unpolymerized adhesive was washed out with acetone and ultrapure water
(Milli-Q, Merck, Rahway, NJ, USA).

RAO was performed as previously described [58]. Briefly, the sample aberrations were
determined at a depth matching the height of the structure to be written. The deformable mirror is
stepped through the first 11 Zernike modes (excluding tilt, piston and defocus) and amplitudes using
the fluorescence signal as feedback. The identified Zernike mask was then applied to the entire MILL
structure.

<u>Expression and Purification of Anti-GPVI Fab</u>

Mouse anti-human GPVI monoclonal antibodies were generated (The WEHI Antibody Facility,
Melbourne, Australia) and isolated from hybridoma supernatant as previously described [64]. Antibody
in the hybridoma supernatant was purified by passing through a column of DEAE Affigel blue (Bio-
Rad, Sydney, NSW, Australia), dialysed into a Tris-saline buffer (Tris 10 mM, NaCl 150 mM, pH 7.4)
then affinity purified on a protein A sepharose column (GE Healthcare, Chicago, IL, USA). Bound
antibody was eluted with 0.1 M Glycine pH 2.4 and neutralised with 1 M Tris pH 8.5. Purified antibody
was then dialysed into phosphate buffered saline (NaCl 137 mM, KCl 2.7 mM, KH2PO4 1.8 mM,
Na2HPO4 10 mM, pH 7.4). Fab fragments were generated from these antibodies using a Fab
preparation kit (Pierce Biotechnology, Rockford, IL, USA) according to manufacturer's instructions.

 Cell Culture and Shear Assays

All cell culture reagents were purchased from Thermo Fisher Scientific (Waltham, MA, USA).

Murine L929 fibroblasts (ATCC, Manassas, VA, USA) were maintained in low glucose (1 g/L) DMEM

supplemented with 10% fetal bovine serum, L-glutamine (4 mM) and pyruvate (1 mM) at 37°C and 5%

CO_2. Cells were split 1:6 at 80% confluence.

For shear assays, fibroblasts in growth phase were detached with trypsin-EDTA (0.05%) and

centrifuged at $300 \times g$ for 5 minutes. Supernatant was removed and cells were resuspended at 2.5×10^5

cells/mL in 1 mL of HEPES-Krebs buffer (120 mM NaCl, 22 mM HEPES, 4.6 mM KCl, 1 mM $MgSO_4$,

155 µM Na_2HPO_4, 412 µM KH_2PO_4, 5 mM $NaHCO_3$, 1 g/L glucose, 1.5 mM $CaCl_2$, pH 7.4)

supplemented with 10% (w/v) fetal bovine serum. Cells were seeded on glass-bottom dishes (no FSS

condition) or flowed into the capillary at 2.5 mL/min (with FSS) on a stage heated to 37°C and allowed

to settle for 1 hour. Krebs-HEPES buffer was then flowed through at different velocities equating to

shear stress values of $(0 - 3.5 \text{ dynes/cm}^2)$ for up to 12 hours and numbers and behavior of adherent cells

were monitored with label-free quantitative phase imaging.

Capillary Thrombus Assay

Circular capillaries (ID: 0.2 mm, 65 µm wall thickness) were precoated with Type-I Collagen

(HORM, Takeda Austria GmbH, Linz, Austria) for 1 hour and washed with 1× PBS at a flow rate of 80

µl/min for 5 minutes. Whole blood was collected in citrated (3.2% w/v, Sigma-Aldrich, St. Louis, MI,

USA) saline and incubated for 30 min with anti-CD42a antibody (1/100 dilution, clone FMC-25,

Thermo Fisher Scientific) conjugated to AlexaFluor 594. The blood was flowed using a syringe pump

(PhD Ultra, Harvard Apparatus, Holliston, MA, USA) on withdraw mode and at a shear stress of 1800

s^{-1} for 10 min, followed by PBS or anti-GPVI Fab (10 µg/ml in PBS, clone 12A5) for another 10 min.

Thrombi were fixed with 4% paraformaldehyde (Sigma-Aldrich) for 10 minutes and then washed with

PBS for 10 minutes under 80 µl/min flow. All blood samples were taken with consent from each donor

and approved by the ANU research ethics office (2022/372).

Multiphoton Imaging and Deconvolution

Rapid live imaging of thrombi formation was performed using a custom-built polygon scanning

microscope [59] at 20 frames per second and rapid volumetric scanning (21 Z-slices per stack) was

achieved by tuning the defocus Zernike amplitude on a deformable mirror conjugated to the back focal

plane of the objective. Fluorescence emission was measured through a detection unit consisting of 2

dichroic mirrors (FF-562-Di02, Semrock, IDEX Health & Science, Rochester, NY, USA), emission

filters (FF01-514/44 and FF01-624/40, Semrock) and three GaAsP photomultiplier tubes (H7422-40,

Hamamatsu, Hamamatsu City, Japan).

SHG and actin imaging were performed with an Olympus FVMPE-RS multiphoton microscope

and a 25× water-immersion objective lens (XL PLAN W MP, 1.05 NA, Olympus, Tokyo, Japan). The

excitation laser was tuned to 900 nm at a laser power setting of 15% and emission signal detected

through an FV30-FVG filter set (dichroic: SDM475, emission filters: BA410-455 and BA495-540) and

2 GaAsP photomultiplier detectors. Volumetric images were obtained with a resonant mirror scanning

at 15 frames per second with 8-line averaging.

A customized ImageJ macro based on the CLIJ2 plugin [79] was written for deconvolving

volumetric images obtained from the polygon 2P microscope. Images were deconvolved using the

Lucy-Richardson algorithm with 20 iterations and an experimental PSF obtained from imaging of 1 μm

yellow-green, fluorescent beads (Polysciences, Warrington, PA, USA).

<u>Confocal Imaging</u>

High resolution confocal imaging of actin was performing with a Leica SP5 microscope with

488 nm and 561 nm excitation lasers and 63× oil immersion objective (HCX PL APO, 1.4 NA, Leica

Microsystems, Wetzlar, Germany). Fluorescence emission was measured using hybrid GaAsP detectors

and emission cutoffs of 502-545 nm for Actin Green and 584-660 nm for rhodamine B. Scanning was

performed with a 4-line averaging and pinhole set to 1 Airy Unit.

<u>Label-free Quantitative Phase Microscopy</u>

Quantitative phase microscopy was performed using a custom-built imaging system with a 514

nm laser (OBIS 514nm LS 20mW Laser, Coherent Inc., Santa Clara, CA, USA). The laser coupled to

an optical fiber, which was then split into an object and reference path. Light from the object path is

collimated and passed through the sample and imaged through a 20× microscope objective (UCPlanFL

N, 0.4 NA, Olympus). The transmitted light is interfered with the reference light and was captured by

a CMOS camera (BFS-U3-32S4, Teledyne FLIR LLC, Wilsonville, OR, USA). The phase information

was reconstructed by an open-source MATLAB program (DHM_MATLAB_ANUAOLAB, V4.0.

Source code available at https://github.com/PurelyWhite/DHM_MATLAB_ANUAOLAB) and

visualized in Fiji (ImageJ).

Computational fluid dynamics simulations

Flow simulation was performed using COMSOL Multiphysics software (Burlington, MA,

USA). Briefly, blood or plasma were represented as a Newtonian fluid with the viscosity of water (1

mP s). The mass inflow of 1 mg/min (corresponding to 1 µl/min) in a cylindrical vessel of 200 µm

diameter and 800 µm length was considered as boundary condition for the inlet and zero pressure for

the outlet. No-slip boundary conditions were chosen for the rest of the system boundaries and laminar

flow module with incompressible fluid approximation was used for numerical analysis of a stationary

Navier-Stokes equations solution. Magnitudes of flow velocities were extracted in a regular grid with

1 µm spatial resolution. 3D reconstruction of the stenosis geometry was performed with Fiji (ImageJ)

capability to render a wavefront object from a volumetric scan of a MILL structure. The triangulation

parameters were kept default (threshold 50, resampling factor 2). The Wavefront objects were imported

into the Autodesk Fusion 360 (Autodesk, San Rafael, CA, USA) as the mesh object. Each mesh object

was cleaned from noise and smoothed. An array of surfaces was created to obtain a set of projection

sketches as an intersection of a mesh body with a surface. The capillary object was created as a

transitional shape between projection border profiles and saved in the appropriate format for the

COMSOL Multiphysics import.

Particle Tracking Velocimetry

1 µm yellow-green beads (Polysciences) were flowed through a capillary tube by pulling at 250

nL/min (rectangular capillary) or 1 µl/min (circular capillary) and imaging was performed by resonant

scanning in the Olympus multiphoton (scanning at 22 Hz) or polygon scanning microscope (25 Hz)

with AO, respectively. Beads were tracked using TrackMate [80] to obtain the velocity and trajectory

using the Kalman filter for homogeneous flow or simple LAP tracker for heterogeneous flow, and a

maximum spot displacement threshold of 20 µm.

Measurement of Collagen and Actin Morphometric Phenotyping

Collagen fiber alignment was measured using the MATLAB code CT-FIRE developed by Liu

and colleagues [81]. For this, a volumetric SHG scan was analyzed and collagen angle at each Z-slice

quantitated.

Actin alignment of high-resolution confocal images was measured using a MATLAB

(MathWorks, Natick, MA, USA) script written by Lickert *et al.* for cell segmentation and actin fiber

measurements [51]. Quantitation of peripheral actin density was performed using the Measure Object

Intensity Distribution module in Cell Profiler [82].

Cell and Platelet Tracking

Time lapse videos of L929 cells and platelets were tracked using TrackMate. Individual

platelets were identified using the Laplacian or Gaussian detector. Due to variations in the cell

morphology, cell segmentation was performed using the Thresholding detector or CLIJ2 Voronoi Otsu

Labeling, dependent on which distinguished individual or clustered cells more accurately. To remove

misidentified tracks, each result was optimized for maximum track length, and mean directional change.

The selection of filters on spots and tracks with the relevant parameters setting were tuned for each

dataset to achieve the best tracking performance. Analysis of cell motility and morphology was done

based on the tracking data provided by TrackMate.

Measurement of Cell Ellipticity

Reconstructed quantitative phase images with height information of the cells were plotted with

the Fiji function '3D Surface Plot' with a custom-developed user interface for parameter control

(viewing angle in XY plane, viewing angle in XZ plane, perspective, smoothing and global height

modification). The self-developed user interface was used to provide the side projections (XZ

projection) of the cells and to generate a mask of the cell surface. The surface was fitted to an ellipse,

where the height and length of the cells were used to calculate cell ellipticity according to the definition

of first flattening ($f = 1- b/a$), where 'b' is the longer axis and 'a' is the shorter axis.

Statistical Analysis

Data were analyzed using Prism (version 9.3.1, Graphpad Software, San Diego, CA, USA). Ordinary

one-way and two-way ANOVAs were performed with Tukey's test. Unpaired parametric t-tests were

performed with P-values calculated using two-tailed analysis.

References

703 1 Ito, Y. *et al.* Turbulence Activates Platelet Biogenesis to Enable Clinical Scale Ex Vivo
704   Production. *Cell* **174**, 636-648 e618, doi:10.1016/j.cell.2018.06.011 (2018).

705 2 King, R. B., Raymond, G. M. & Bassingthwaighte, J. B. Modeling blood flow
706   heterogeneity. *Annals of Biomedical Engineering* **24**, 352-372,
707   doi:10.1007/BF02660885 (1996).

708 3 Roux, E., Bougaran, P., Dufourcq, P. & Couffinhal, T. Fluid Shear Stress Sensing by
709   the Endothelial Layer. *Front Physiol* **11**, 861, doi:10.3389/fphys.2020.00861 (2020).

710 4 Haase, K. *et al.* Physiologic flow-conditioning limits vascular dysfunction in
711   engineered human capillaries. *Biomaterials* **280**, 121248,
712   doi:10.1016/j.biomaterials.2021.121248 (2022).

713 5 Saqr, K. M. *et al.* Physiologic blood flow is turbulent. *Sci Rep* **10**, 15492,
714   doi:10.1038/s41598-020-72309-8 (2020).

715 6 Yankaskas, C. L. *et al.* The fluid shear stress sensor TRPM7 regulates tumor cell
716   intravasation. *Science Advances* **7**, eabh3457, doi:doi:10.1126/sciadv.abh3457 (2021).

717 7 Agbani, E. O. *et al.* Aquaporin-1 regulates platelet procoagulant membrane dynamics
718   and in vivo thrombosis. *JCI Insight* **3**, doi:10.1172/jci.insight.99062 (2018).

719 8 Atcha, H. *et al.* Mechanically activated ion channel Piezo1 modulates macrophage
720   polarization and stiffness sensing. *Nat Commun* **12**, 3256, doi:10.1038/s41467-021-
721   23482-5 (2021).

722 9 Al-Tamimi, M. *et al.* Pathologic shear triggers shedding of vascular receptors: a novel
723   mechanism for down-regulation of platelet glycoprotein VI in stenosed coronary
724   vessels. *Blood* **119**, 4311-4320, doi:10.1182/blood-2011-10-386607 (2012).

725 10 Polacheck, W. J., Charest, J. L. & Kamm, R. D. Interstitial flow influences direction of
726   tumor cell migration through competing mechanisms. *Proceedings of the National
727   Academy of Sciences* **108**, 11115-11120, doi:doi:10.1073/pnas.1103581108 (2011).

728 11 Marturano-Kruik, A. *et al.* Human bone perivascular niche-on-a-chip for studying
729   metastatic colonization. *Proceedings of the National Academy of Sciences* **115**, 1256-
730   1261, doi:doi:10.1073/pnas.1714282115 (2018).

12 Pakshir, P. *et al.* The myofibroblast at a glance. *J Cell Sci* **133**, doi:10.1242/jcs.227900
(2020).

13 Yourek, G., McCormick, S. M., Mao, J. J. & Reilly, G. C. Shear stress induces
osteogenic differentiation of human mesenchymal stem cells. *Regen Med* **5**, 713-724,
doi:10.2217/rme.10.60 (2010).

14 Follain, G. *et al.* Fluids and their mechanics in tumour transit: shaping metastasis.
Nature Reviews Cancer **20**, 107-124, doi:10.1038/s41568-019-0221-x (2020).

15 Wirtz, D., Konstantopoulos, K. & Searson, P. C. The physics of cancer: the role of
physical interactions and mechanical forces in metastasis. *Nature Reviews Cancer* **11**,
512-522, doi:10.1038/nrc3080 (2011).

16 Nesbitt, W. S. *et al.* A shear gradient-dependent platelet aggregation mechanism drives
thrombus formation. *Nat Med* **15**, 665-673, doi:10.1038/nm.1955 (2009).

17 Yeung, C. K. *et al.* Tissue Chips in Space—Challenges and Opportunities. *Clinical and*
Translational Science **13**, 8-10, doi:https://doi.org/10.1111/cts.12689 (2020).

18 Ingber, D. E. Human organs-on-chips for disease modelling, drug development and
personalized medicine. *Nature Reviews Genetics*, doi:10.1038/s41576-022-00466-9
(2022).

19 Qin, D., Xia, Y. & Whitesides, G. M. Soft lithography for micro- and nanoscale
patterning. *Nature Protocols* **5**, 491-502, doi:10.1038/nprot.2009.234 (2010).

20 Hou, X. *et al.* Interplay between materials and microfluidics. *Nature Reviews Materials*
2, 17016, doi:10.1038/natrevmats.2017.16 (2017).

21 David J. Beebe, Glennys A. Mensing, a. & Walker, G. M. Physics and Applications of
Microfluidics in Biology. *Annual Review of Biomedical Engineering* **4**, 261-286,
doi:10.1146/annurev.bioeng.4.112601.125916 (2002).

22 Kajtez, J. *et al.* 3D-Printed Soft Lithography for Complex Compartmentalized
Microfluidic Neural Devices. *Advanced Science* **7**, 2001150,
doi:https://doi.org/10.1002/advs.202001150 (2020).

23 Beh, C. W., Zhou, W. & Wang, T.-H. PDMS–glass bonding using grafted polymeric
adhesive – alternative process flow for compatibility with patterned biological
molecules. *Lab on a Chip* **12**, 4120-4127, doi:10.1039/C2LC40315C (2012).

24   Fiorini, G. S. & Chiu, D. T. Disposable microfluidic devices: fabrication, function, and
application. *BioTechniques* **38**, 429-446, doi:10.2144/05383rv02 (2005).

25   Toepke, M. W. & Beebe, D. J. PDMS absorption of small molecules and consequences
in microfluidic applications. *Lab on a Chip* **6**, 1484-1486, doi:10.1039/B612140C
(2006).

26   Liao, C., Anderson, W., Antaw, F. & Trau, M. Two-Photon Nanolithography of
Tailored Hollow three-dimensional Microdevices for Biosystems. *ACS Omega* **4**, 1401-
1409, doi:10.1021/acsomega.8b03164 (2019).

27   Maruo, S. & Fourkas, J. T. Recent progress in multiphoton microfabrication. *Laser &*
Photonics Reviews **2**, 100-111, doi:https://doi.org/10.1002/lpor.200710039 (2008).

28   Qin, D. *et al.* Structural colour enhanced microfluidics. *Nat Commun* **13**, 2281,
doi:10.1038/s41467-022-29956-4 (2022).

29   S P Sutera, a. & Skalak, R. The History of Poiseuille's Law. *Annual Review of Fluid*
Mechanics **25**, 1-20, doi:10.1146/annurev.fl.25.010193.000245 (1993).

30   Miteva, D. O. *et al.* Transmural flow modulates cell and fluid transport functions of
lymphatic   endothelium.   *Circ   Res   **106**,   920-931,
doi:10.1161/CIRCRESAHA.109.207274 (2010).

31   Bretag, A. H. The glass micropipette electrode: A history of its inventors and users to
1950. *Journal of General Physiology* **149**, 417-430, doi:10.1085/jgp.201611634 (2017).

32   Utada, A. S. *et al.* Monodisperse Double Emulsions Generated from a Microcapillary
Device. *Science* **308**, 537-541, doi:doi:10.1126/science.1109164 (2005).

33   Guo, J. A new model of painting glass. *Glass Technology: European Journal of Glass*
Science and Technology Part A **59**, 54-57, doi:10.13036/17533546.59.2.022 (2018).

34   Eschenbaum, C. *et al.* Hybrid lithography: Combining UV-exposure and two photon
direct laser writing. *Opt. Express* **21**, 29921-29926, doi:10.1364/OE.21.029921 (2013).

35   Bernardeschi, I., Ilyas, M. & Beccai, L. A Review on Active 3D Microstructures via
Direct   Laser   Lithography.   *Advanced   Intelligent   Systems*   **3**,   2100051,
doi:https://doi.org/10.1002/aisy.202100051 (2021).

36   Katz, B.-Z. *et al.* Physical State of the Extracellular Matrix Regulates the Structure and
Molecular Composition of Cell-Matrix Adhesions. *Molecular Biology of the Cell* **11**,
1047-1060, doi:10.1091/mbc.11.3.1047 (2000).

37   Garanich, J. S., Mathura, R. A., Shi, Z.-D. & Tarbell, J. M. Effects of fluid shear stress
on adventitial fibroblast migration: implications for flow-mediated mechanisms of
arterialization and intimal hyperplasia. *American Journal of Physiology-Heart and*
Circulatory Physiology **292**, H3128-H3135, doi:10.1152/ajpheart.00578.2006 (2007).

38   Steward, R. L., Cheng, C.-M., Ye, J. D., Bellin, R. M. & LeDuc, P. R. Mechanical
stretch and shear flow induced reorganization and recruitment of fibronectin in
fibroblasts. *Scientific Reports* **1**, 147, doi:10.1038/srep00147 (2011).

39   Ng, C. P., Hinz, B. & Swartz, M. A. Interstitial fluid flow induces myofibroblast
differentiation and collagen alignment in vitro. *Journal of Cell Science* **118**, 4731-4739,
doi:10.1242/jcs.02605 (2005).

40   Pakshir, P. *et al.* The myofibroblast at a glance. *Journal of Cell Science* **133**,
doi:10.1242/jcs.227900 (2020).

41   Hinz, B. Formation and Function of the Myofibroblast during Tissue Repair. *Journal*
of Investigative Dermatology **127**, 526-537, doi:https://doi.org/10.1038/sj.jid.5700613
(2007).

42   Tomasek, J. J., Gabbiani, G., Hinz, B., Chaponnier, C. & Brown, R. A. Myofibroblasts
and mechano-regulation of connective tissue remodelling. *Nature Reviews Molecular*
Cell Biology **3**, 349-363, doi:10.1038/nrm809 (2002).

43   Vaghi, C., Fanciullino, R., Benzekry, S. & Poignard, C. Macro-scale models for fluid
flow in tumour tissues: impact of microstructure properties. *Journal of Mathematical*
Biology **84**, 27, doi:10.1007/s00285-022-01719-1 (2022).

44   Tschumperlin, D. J. Fibroblasts and the Ground They Walk On. *Physiology* **28**, 380-
390, doi:10.1152/physiol.00024.2013 (2013).

45   Zhou, J. & Papautsky, I. Viscoelastic microfluidics: progress and challenges.
Microsystems & Nanoengineering **6**, 113, doi:10.1038/s41378-020-00218-x (2020).

46   Popel, A. S. & Johnson, P. C. MICROCIRCULATION AND HEMORHEOLOGY.
Annual Review of Fluid Mechanics **37**, 43-69,
doi:10.1146/annurev.fluid.37.042604.133933 (2005).

47   Salter, P. S. & Booth, M. J. Adaptive optics in laser processing. *Light: Science &*
Applications **8**, 110, doi:10.1038/s41377-019-0215-1 (2019).

48   Tang, D., Yang, C., Kobayashi, S., Zheng, J. & Vito, R. P. Effect of Stenosis
Asymmetry on Blood Flow and Artery Compression: A Three-Dimensional Fluid-
Structure Interaction Model. *Annals of Biomedical Engineering* **31**, 1182-1193,
doi:10.1114/1.1615577 (2003).

49   Solon, J., Levental, I., Sengupta, K., Georges, P. C. & Janmey, P. A. Fibroblast
Adaptation and Stiffness Matching to Soft Elastic Substrates. *Biophysical Journal* **93**,
4453-4461, doi:10.1529/biophysj.106.101386 (2007).

50   He, X. *et al.* Automated Fourier space region-recognition filtering for off-axis digital
holographic   microscopy.   *Biomed.   Opt.   Express*   **7**,   3111-3123,
doi:10.1364/BOE.7.003111 (2016).

51   Lickert, S. *et al.* Morphometric analysis of spread platelets identifies integrin
alphaIIbbeta3-specific contractile phenotype. *Sci Rep* **8**, 5428, doi:10.1038/s41598-
018-23684-w (2018).

52   Birmingham, K. G. *et al.* Lymph Node Subcapsular Sinus Microenvironment-On-A-
Chip Modeling Shear Flow Relevant to Lymphatic Metastasis and Immune Cell
Homing. *iScience* **23**, 101751, doi:10.1016/j.isci.2020.101751 (2020).

53   Williams, M. S., Longmuir, K. J. & Yager, P. A practical guide to the staggered
herringbone mixer. *Lab Chip* **8**, 1121-1129, doi:10.1039/b802562b (2008).

54   Chary, S. R. & Jain, R. K. Direct measurement of interstitial convection and diffusion
of albumin in normal and neoplastic tissues by fluorescence photobleaching.
Proceedings   of   the   National   Academy   of   Sciences **86**,   5385-5389,
doi:doi:10.1073/pnas.86.14.5385 (1989).

55   Elineni, K. K. & Gallant, N. D. Regulation of cell adhesion strength by peripheral focal
adhesion distribution. *Biophys J* **101**, 2903-2911, doi:10.1016/j.bpj.2011.11.013 (2011).

56   Lin, J., Sorrells, M. G., Lam, W. A. & Neeves, K. B. Physical forces regulating
hemostasis and thrombosis: Vessels, cells, and molecules in illustrated review. *Res*
Pract Thromb Haemost **5**, e12548, doi:10.1002/rth2.12548 (2021).

57   Kostencka, J., Kozacki, T., Kus, A. & Kujawinska, M. Accurate approach to capillary-
supported   optical   diffraction   tomography.   *Opt   Express* **23**,   7908-7923,
doi:10.1364/OE.23.007908 (2015).

58   Li, Y. *et al.* Raster adaptive optics for video rate aberration correction and large FOV
multiphoton imaging. *Biomed Opt Express* **11**, 1032-1042, doi:10.1364/BOE.377044
(2020).

59   Li, Y. X. *et al.* Flexible polygon-mirror based laser scanning microscope platform for
multiphoton in-vivo imaging. *J Biophotonics* **10**, 1526-1537,
doi:10.1002/jbio.201600289 (2017).

60   Thondapu, V. *et al.* High spatial endothelial shear stress gradient independently predicts
site of acute coronary plaque rupture and erosion. *Cardiovasc Res* **117**, 1974-1985,
doi:10.1093/cvr/cvaa251 (2021).

61   Montague, S. J., Lim, Y. J., Lee, W. M. & Gardiner, E. E. Imaging Platelet Processes
and Function-Current and Emerging Approaches for Imaging in vitro and in vivo. *Front
Immunol* **11**, 78, doi:10.3389/fimmu.2020.00078 (2020).

62   Perrella, G., Nagy, M., Watson, S. P. & Heemskerk, J. W. M. Platelet GPVI
(Glycoprotein VI) and Thrombotic Complications in the Venous System. *Arterioscler
Thromb Vasc Biol* **41**, 2681-2692, doi:10.1161/ATVBAHA.121.316108 (2021).

63   Ahmed, M. U. *et al.* Pharmacological Blockade of Glycoprotein VI Promotes
Thrombus Disaggregation in the Absence of Thrombin. *Arterioscler Thromb Vasc Biol*
40, 2127-2142, doi:10.1161/ATVBAHA.120.314301 (2020).

64   Al-Tamimi, M. *et al.* Anti-glycoprotein VI monoclonal antibodies directly aggregate
platelets independently of FcgammaRIIa and induce GPVI ectodomain shedding.
Platelets **20**, 75-82, doi:10.1080/09537100802645029 (2009).

65   Gupta, S. & Brass, L. F. Glycoprotein VI Blockade: Not Just Targeting Collagen
Anymore? *Arterioscler Thromb Vasc Biol* **40**, 1964-1966,
doi:10.1161/ATVBAHA.120.315059 (2020).

66   Harbi, M. H., Smith, C. W., Nicolson, P. L. R., Watson, S. P. & Thomas, M. R. Novel
antiplatelet strategies targeting GPVI, CLEC-2 and tyrosine kinases. *Platelets* **32**, 29-
41, doi:10.1080/09537104.2020.1849600 (2021).

67   Borst, O. & Gawaz, M. Glycoprotein VI - novel target in antiplatelet medication.
Pharmacol Ther **217**, 107630, doi:10.1016/j.pharmthera.2020.107630 (2021).

68   van der Meer, A. D., Poot, A. A., Feijen, J. & Vermes, I. Analyzing shear stress-induced
alignment of actin filaments in endothelial cells with a microfluidic assay.
Biomicrofluidics **4**, 11103, doi:10.1063/1.3366720 (2010).

69 Batchelor, R. *et al.* Two in One: Light as a Tool for 3D Printing and Erasing at the
Microscale. *Adv Mater* **31**, e1904085, doi:10.1002/adma.201904085 (2019).

70 Schuster, B. *et al.* Automated microfluidic platform for dynamic and combinatorial
drug screening of tumor organoids. *Nat Commun* **11**, 5271, doi:10.1038/s41467-020-
19058-4 (2020).

71 Michas, C. *et al.* Engineering a living cardiac pump on a chip using high-precision
fabrication. *Sci Adv* **8**, eabm3791, doi:10.1126/sciadv.abm3791 (2022).

72 Leung, C. M. *et al.* A guide to the organ-on-a-chip. *Nature Reviews Methods Primers*
2, 33, doi:10.1038/s43586-022-00118-6 (2022).

73 Dominguez, G. A., Anderson, N. R. & Hammer, D. A. The direction of migration of T-
lymphocytes under flow depends upon which adhesion receptors are engaged. *Integr*
Biol (Camb) **7**, 345-355, doi:10.1039/c4ib00201f (2015).

74 Fuh, K. F., Shepherd, R. D., Withell, J. S., Kooistra, B. K. & Rinker, K. D. Fluid flow
exposure promotes epithelial-to-mesenchymal transition and adhesion of breast cancer
cells to endothelial cells. *Breast Cancer Res* **23**, 97, doi:10.1186/s13058-021-01473-0
(2021).

75 Sahai, E. *et al.* A framework for advancing our understanding of cancer-associated
fibroblasts. *Nat Rev Cancer* **20**, 174-186, doi:10.1038/s41568-019-0238-1 (2020).

76 Pakshir, P. *et al.* Dynamic fibroblast contractions attract remote macrophages in
fibrillar collagen matrix. *Nat Commun* **10**, 1850, doi:10.1038/s41467-019-09709-6
(2019).

77 Yeoman, B. *et al.* Adhesion strength and contractility enable metastatic cells to become
adurotactic. *Cell Rep* **34**, 108816, doi:10.1016/j.celrep.2021.108816 (2021).

78 Fuhrmann, A. & Engler, A. J. Acute shear stress direction dictates adherent cell
remodeling and verifies shear profile of spinning disk assays. *Phys Biol* **12**, 016011,
doi:10.1088/1478-3975/12/1/016011 (2015).

79 Haase, R. *et al.* CLIJ: GPU-accelerated image processing for everyone. *Nat Methods*
17, 5-6, doi:10.1038/s41592-019-0650-1 (2020).

80 Tinevez, J. Y. *et al.* TrackMate: An open and extensible platform for single-particle
tracking. *Methods* **115**, 80-90, doi:10.1016/j.ymeth.2016.09.016 (2017).

81   Bredfeldt, J. S. *et al.* Computational segmentation of collagen fibers from second-
harmonic generation images of breast cancer. *J Biomed Opt* **19**, 16007,
doi:10.1117/1.JBO.19.1.016007 (2014).

82   Stirling, D. R. *et al.* CellProfiler 4: improvements in speed, utility and usability. *BMC*
Bioinformatics **22**, 433, doi:10.1186/s12859-021-04344-9 (2021).

Author Contributions
W.M.L. initiated, developed and supervised the project. Y.J.L. led the project, performed experimental

work and carried out the analysis of the results. J.Z. prepared and performed cell experimental work.

H.L. and T.X. assisted in lithography. H.L. assisted in image analysis and deconvolution. Y. L. built

the AO system. Z. Z. advised on software development of phase imaging. S.M.H. and E.E.G. advised

on platelet experiments and S.M.H. prepared GPVI Fabs. I.C. and D.N. carried out flow simulation.

W.M.L. and Y.J.L. wrote the manuscript with input from all authors.

Disclosures

The adaptive optics technique used in this paper has been submitted for a provisional patent application,

Application No. 2019904929.

Funding

Australian Research Council (DE160100843, DP190100039, DP200100364)

Figure 1. Current approaches for fabricating microfluidic channels with 3D patterning. A) Diagram of steps and time requirement to fabricate a circular microchannel with an irregular-shaped structure using (i) soft lithography, (ii) fused deposition modeling or (iii) *MILL*. The lithography resolutions and time incurred are indicated below each process. **B)** Brightfield image and 2-photon fluorescence image (inset) of herringbone structure in a square glass capillary tube.

Figure 2. Fibroblast adhesion and proliferation under interstitial, homogenous fluid shear stress:
shape change, mobility, and actin reorganization. A, Experiment steps to measure cell response
(shape change and mobility) along different shear stress and adhesion ligand coatings. Cells (L929,
fibroblasts) are first seeded on glass surfaces (dish or capillary tube) that are precoated with collagen or
poly-L-Lysine (PLL). A range of homogeneous fluid shear stresses (0 to 3.5 dynes/cm^2) were imposed
over 12 hours period and cell motility monitored by microscopy. **B,** i) Cell adhesion and proliferation
are each measured based on cell ellipticity and motility, respectively, using automated cell tracking. ii)
Results of cell shape change (ellipticity) subjected to different shear stress (0-3.5 dynes/cm^2) on
collagen or PLL-coated capillaries were tracked over the first 3 hours (n=3 cells). iii) Cell cluster
motility at shear stress 00.1, 0.9 and 1.7 dynes/cm^2. **C)** i) - vi) Time lapse imaging of L929 cells with
quantitative phase microscopy over the full 12 hours for selected field of view. Cell cluster motility in
B) iii) was measured after cell segmentation (yellow outline). **D)** i) - vi) Confocal maximum intensity
projections of actin fiber (F-actin) in L929 cells after 12 hours at shear stress 0 and 0.9 dynes/cm^2 under
PLL and collagen coated glass surface. v) and vi) Segmentation and quantification of actin fiber
orientation, alignment along flow direction. vii) alignment of actin parallel to flow (n=4 cells). 0 means
completely orthogonal and unity means parallel to flow axis. Scale bars- C) 40 μm and 15 μm, D) 20
μm and 5 μm. Data are the means and SD with D) vi) 2-way ANOVA analysis (**P < 0.01; ****P <
0.001; ns, not significant)

Figure 3. Heterogenous fluid shear stimulates heterogenous cell adhesion, mobility and actin
cytoskeleton. A, i) orthogonal plots of structures formed by MILL as scanned by confocal microscopy,
ii) homogenous and iii) heterogenous flow profile determined by PTV. (**B,** i) and ii)) QPM images of
L929 fibroblasts after 1 hour of cell seeding, iii) and iv) after 12 hours of flow to v) determine how
heterogeneous shear stress controls cell motility. vi) Quantification of cell motility from time lapse
phase imaging (n=3). vii), viii), ix) 2-photon SHG and fluorescence imaging of fixed cells after 12 hours
of flow under laminar, homogenous flow rate. **C)** i) and ii) Shows regions with laminar and different
shear stress. iii-vi) confocal imaging of actin in L929 cells and segmentation of the cell to vii) identify
the proportion of actin at the cell periphery (n=6). Data are the means and SD with B) vi) unpaired t test
and C) vii) ordinary one-way ANOVA analysis (*P<0.05; ****P<0.001; ns, not significant).

Figure 4. Adaptive Optics Multiphoton Inner Laser Lithography for cylindrical tube. **A,** Non-
symmetrical disruption of flow from lithography of asymmetrical structure within a circular capillary
with flow velocity determined by CFD simulation. **B**, i) Aberration imposed by cylindrical tube. ii)
Lithography structures subjected to aberrations. iii) Identifying aberration mask and iv) lithography
structure formed after AO correction and (v) width and (vi) thickness of structures formed with or
without AO MILL measured with QPM. Results are the mean and SD of n=5 structures. **C,** i) and ii)
Flow velocities across an empty and fabricated AO MILL structure determined experimentally with
PTV. **D**, Second harmonic imaging of collagen distribution across a tube under (i) laminar or (ii)
heterogeneous shear and (iii) fiber alignment measured after thresholding and segmentation. Data are
the means and SD with B) v) and vi) ordinary one-way ANOVA analysis (*P < 0.05; **P < 0.01; ns,
not significance).

982
983

Figure 5. Asymmetrical AO-*MILL* structure generated heterogenous FSS that increases platelet translocation. A, (i) SHG imaging of collagen distribution (white) (ii) flow velocity determined by PTV. (iii) XY slice of platelets adhering on the stenosis, iv) SHG imaging of collagen distribution (white) surrounding an AO-*MILL* structure (red). (v) flow velocity around AO-*MILL* structure determined by PTV. **B** tracked to measure platelet motility across the capillary. Arrows indicate direction of platelet translocation. Platelet motility quantified across the capillary in color code and fluid shear stress boundaries are demarcated. **C**, i) Volumetric render of platelet velocities across the stenotic capillary and ii) cross sectional zoomed in image adjacent to the AO MILL structure showing platelet movement (white arrow).

Figure 6. Effect of vascular fluid shear stress increases thrombus formation. A, (i) Shear disruption from an irregular AO MILL structure that generates (ii) a heterogeneous flow profile simulated by CFD and promotes (iii) thrombus formation. iv) simulation and (v) PTV after thrombus is formed. **B** QPM imaging of thrombi formed in a capillary **(i)** without and **(ii)** with the AO-MILL structure of formed thrombi upstream and downstream to the stenosis. **iii),** QOM images quantified across three region.

Figure 7. Increasing FSS assay throughput with multiple MILL capillary chip. A, Design of a reusable PDMS-based clamp consisting 4 capillary slots and an imaging window, which is adhered onto a glass slide. Inlets and outlets are connected to tubing using an optical UV glue for rapid sealing and structures formed in selected capillaries using MILL. **B**, The collagen coverage is characterized by SHG imaging prior to **C.** thrombus formation under flow in each capillary tube. **D,** Scatter plot of thrombi volume over area as measured by QPM with homogenous and heterogenous flow and with and without GPIV Fab treatment.

1015 **Figure 8. Degrees of laminar flow: homogeneous or heterogeneous flow profiles.** Laminar
1016 flow is represented in different shear regimes that are classified as homogeneous vs
1017 heterogeneous shear distributions, which cells sense and respond to differently.

1018

1019