Complete sequence verification of plasmid DNA using the Oxford Nanopore Technologies' MinION device

Scott D. Brown[1], Lisa Dreolini[1], Jessica F. Wilson[1], Miruna Balasundaram[1], *Robert A. Holt[1,2,3]

[1] Canada's Michael Smith Genome Sciences Centre; BC Cancer Research Institute; 675 W 10th Ave, Vancouver, BC, V5Z 1L3; Canada
[2] Department of Molecular Biology and Biochemistry; Simon Fraser University; SSB8166 – 8888 University Drive, Burnaby, BC, V5A 1S6; Canada
[3] Department of Medical Genetics; University of British Columbia; C201 – 4500 Oak Street, Vancouver, BC, V6H 3N1; Canada
[*] Corresponding Author - Robert A. Holt
675 W 10th Ave
Vancouver, BC, Canada V5Z 1L3
Correspondence: rholt@bcgsc.ca

Abstract

Sequence verification is essential for plasmids used as critical reagents or therapeutic products. Here, we describe a cost-effective and accurate plasmid sequencing and consensus generation procedure using the Oxford Nanopore Technologies' MinION device as an alternative to commercial plasmid sequencing options. This procedure can verify the identity of a pure population of plasmid, either confirming it matches the known and expected sequence, or identifying mutations present in the plasmid if any exist. We use a full MinION flow cell per plasmid, maximizing available data and allowing for stringent quality filters. Pseudopairing reads for consensus basecalling reduces error rates to 0.53 %, and our pileup consensus approach provides per-base counts and confidence scores, allowing for interpretation of the resulting consensus sequences. For pure plasmid samples, we demonstrate 100 % accuracy in the resulting consensus sequence, and the sensitivity to detect small mutations such as insertions, deletions, and single nucleotide variants. In cases where the sequenced pool of plasmids contains subclonal templates, detection sensitivity is similar to that of traditional capillary sequencing. Our pipeline can provide significant cost savings compared to outsourcing clinical-grade sequencing of plasmids, making generation of high-quality plasmid sequence for clinical sequence verification more accessible.

Keywords: MinION, cell therapy, plasmid sequencing, long-read sequencing

Introduction

Sequence verification is essential for plasmids destined for downstream clinical applications. Plasmids used for gene delivery (i.e. for generating genetically modified Immune Effector Cells), mRNA synthesis, or those used directly as plasmid DNA vaccines, must have their sequence identity confirmed prior to use as critical reagents or therapeutic products (1, 2). Commercial services for clinical grade sequencing are available, but may offer longer than desired turnaround times and higher cost than some budgets may allow. A simple, fast, and inexpensive in-house option for plasmid sequence verification would be of potential utility. Here, we evaluate the Oxford Nanopore Technologies' (ONT) USB-powered MinION sequencing device for this purpose.

Conventionally, capillary-based Sanger sequencing approaches have been used for plasmid sequence verification (3). Long read sequencing is an attractive alternative as the sequence of the entire plasmid can be contained in a single read. Unlike plasmid sequence verification using short-reads (4, 5), long-read data is amenable to verify repetitive regions, check for inversions, and identify other large-scale structural errors. However, long-read data generally suffers from higher error rates than short-read data, making the detection of small errors, such as SNVs, more challenging (6–8). Despite these challenges, approaches have been developed to use ONT long read data to successfully sequence plasmids in a number of research applications (9–13), however, the existing tools and workflows are not appropriate for clinical applications as they do not provide information on the confidence of each basecall, generally do not provide quality measures for the resulting consensus sequence, and can introduce biases due to assumptions made when resolving ambiguous basecalls that are the result of systematic errors (14). Existing long read assemblers output a consensus sequence without any per-base quality metrics, and have difficulty processing the large numbers of reads required to gain confidence in the resulting sequence (15–17).

Here, we describe a cost-effective and accurate plasmid library preparation procedure and analysis pipeline that generates a high-quality consensus sequence of the plasmid from an ONT MinION sequencing run, allowing for complete sequence verification of a plasmid prior to downstream applications. Utilizing the full capacity of the MinION flowcell, we obtain millions of raw reads for each plasmid which are then quality filtered to the subset of most accurate reads. We prioritize interpretability of the resulting sequence, providing confidence measures for each base of the consensus. We demonstrate that the resulting quality-filtered and processed consensus sequence achieves 100 % accuracy for three exemplar plasmids, and that mutated and contaminating templates can be detected at a level comparable to that of Sanger sequencing. Importantly, our entire process can be established to run within an individual lab, giving the operator complete control over all steps of the sequencing process.

Results

Oxford Nanopore Technologies MinION library construction and sequencing of linearized plasmid.

Our novel approach was developed and validated using three unique plasmids previously generated in-house: BCRxV.TF.1, BCRxV.GagPolRev.1 and BCRxV.VSVG.1 (18) (**Supplementary File 1**). The sequence identity of each plasmid had been previously confirmed by Sanger sequencing following the NIH Human Genome finishing standard (> Phred30, double read coverage) (19). Prior to creation of the MinION sequencing libraries, we ran a restriction enzyme fingerprinting assay to verify the purity and identity of the plasmids (Materials and Methods).

Sequencing libraries were constructed using the SQK-LSK110 Ligation Sequencing kit following the Genomic DNA by Ligation protocol provided by ONT. Plasmids were linearized prior to library construction with restriction enzymes that generate 5' overhangs. This is essential to maintain full plasmid sequence during the library construction end-repair step. We loaded each plasmid library onto a separate R10.3 MinION flow cell, and ran each flow cell for the full 72 hour run time. Each run generated an average of 3.86 million reads.

Raw sequence basecalling and read filtering.

We performed basecalling of the raw .fast5 data produced by the MinION using the GPU version of Guppy v5.0.11 on a GPU cluster (Materials and Methods). We used the '--min_qscore' argument to filter reads by their lowest observed base quality score. We set this value to 12, only retaining basecalled reads where all bases had a quality score of 12 or higher. For our three plasmids, an average of 40.8 % of reads passed this quality filter (range 27.4 – 50.3 %). While the majority of reads were the expected sequence length for each plasmid (**Supplementary Figure 1**), we noted that a small subset of reads were dramatically shorter or longer than expected. These may have arisen from fragmented template, incomplete transit of template through the pore, or concatenation of template molecules during library construction. We surmised that these reads of abnormal length would not benefit the downstream analysis, and only used reads with length within 250 bases of the expected template length. We observed between 70 – 87 % of quality-filtered reads meeting the length filtering criteria for use in subsequent analysis.

Improving basecalling accuracy by pseudopairing reads.

Upon aligning these reads to the known reference sequence, we identified some cases where an incorrect base was consistently called predominantly in the forward or predominantly in the reverse read sets (**Figure 1**). This strand-specific bias may be the result of the specific sequence context within the pore at that position of the template, and differences in the ease that the basecaller has in decoding the forward vs. reverse sequence motif (14, 20, 21). Assessing the forward and reverse read sets independently, there was no way to determine which of the read sets contained the correct base call. When combining the forward and reverse read sets, there was a roughly equal number of reads supporting each base, resulting in the appearance of a heterozygous base call at that position. Therefore, another approach was needed to resolve this strand-specific bias.

Paired basecalling of matched forward and reverse strands of the same template molecule has shown great improvements in basecalling accuracy (22). This approach aligns the forward and reverse raw signals, and uses a consensus decoding algorithm to generate the basecalls. This requires knowledge of which reads originated from the same template molecule. Generating such data requires special sequencing chemistry, such as ONT's 2D or $1D^2$ kits, which were unavailable to us. Since we are sequencing a single plasmid, we know all reads should contain the same sequence. We hypothesized that we would be able to pseudopair forward and reverse reads to allow paired basecalling of the raw .fast5 data, generating a single consensus duplex basecall for every pseudopair of reads. We predicted that the consensus of the raw signal in these problematic strand-specific regions of the template would yield the correct base call. To generate these pseudopairs, we first aligned the Guppy basecalled reads to the expected reference as a way of annotating each read as coming from the sense or antisense strand of the template. Then, we paired sense and antisense reads, giving preference to pairs of reads with similar read lengths. These read pair lists were then used in conjunction with the raw .fast5 data as input into the Bonito basecaller (https://github.com/nanoporetech/bonito; v0.4.0) to perform duplex basecalling.

Overall, the Bonito duplex basecalled reads alleviated the strand-specific biases observed in the Guppy basecalled data. We noted, however, rare cases where miscalling was still occurring. Assuming that these miscalls were the result of modified bases, we used Bonito's model finetuning feature to fine-tune the R10.3 basecalling model using a random sample of 100 reads from one of our plasmids. This fine-tuning resulted in the elimination of systematic errors in our sequence reads for all three plasmids. Overall, our duplex basecalled reads had an order of magnitude reduction in error rate compared to the non-paired data (0.53 % compared to 5.3 % from the Guppy basecalled data, as reported by Qualimap).

Generating a high quality plasmid consensus sequence.

To obtain a consensus sequence for the plasmid from the high accuracy duplex basecalled reads, we implemented a "pileup" approach, taking the most-evidenced base at each position. To facilitate this, we generated a draft consensus by performing *de novo* assembly using a subset of the basecalled reads. We then aligned all reads to this assembly, producing an alignment matrix of all reads positioned in the same reference frame. This two-step procedure was more tractable, computationally, than performing a multiple sequence alignment of hundreds of thousands of reads. We then stepped through each position of the alignment matrix and enumerated the number of reads supporting each base. We handled insertions by tracking bases that aligned between two adjacent bases of the assembly. We visualized this data to get an overview of the strength of the signal and noise in our data (**Figure 2**).

Overall, we saw that the median signal:noise ratio across our three plasmids ranged from 3250 – 3770, while there were rare individual cases of it dropping to as low as 5.7. Generally, signal:noise ratios decreased at the extreme 5' and 3' ends of the template. Additionally, we observed drops in signal across homopolymer stretches (**Figure 3**). This is consistent with known difficulties in nanopore sequencing of homopolymer stretches (7), with reads underreporting the number of nucleotides in the homopolymer stretch. In one particularly long homopolymer from BCRxV.VSVG, only 17.3 % of the reads supported the full 17 bases.

To generate a high quality consensus sequence from this data, we specified criteria that each base, and the overall consensus, must pass. We took the top-evidenced base at each position as the

consensus base. The number of reads supporting this base must be at least 5× as high as the next most observed base at that position, and must be at least 10 % of the total number of aligned reads. Finally, the per-base confidence was calculated as the percent of reads at each position that support the consensus base, and the median per-base confidence across the entire plasmid must be greater than 99.9 %.

We observed rare cases where alignment artefacts near error-prone homopolymer regions resulted in two different consensus sequences depending on whether the *de novo* assembly reflected the sense or antisense strand (**Supplementary Figure 2**). In order to be confident that our resulting consensus sequence was accurate, we performed two alignment and consensus generation passes – once as described with alignment to the assembly, and once with alignment to the reverse compliment of the assembly (Materials and Methods). This yielded two consensus sequences, and if these sequences differ, this suggests that a variant template may be present (see next section, ***Detecting variant plasmids***).

Detecting variant plasmids.

Once we have obtained the high-quality consensus sequences (seeded from the sense and antisense versions of the assembly), we can determine if these sequences match the expected plasmid sequence. For our main goal of sequence verification, we simply compared the forward and reverse strand consensus sequences with the expected sequence (the "known" sequence of the plasmid based on cloning or synthesis designs) by sequence alignment and checked for any discrepancies. If any discrepancies were found in either consensus, we reported that the sequenced plasmid contains that variant. In the case where any position of the consensus did not meet our criteria, an "N" base is added to the consensus. The base signal strengths and confidence scores for these positions should be manually inspected, as they may be the result of a decreased signal to noise ratio in a difficult to sequence area, or could be the result of a variant in the plasmid.

As a secondary goal, we can also detect the presence of subclonal templates in the population of plasmid molecules used in library construction. It is rare that such a case would arise, but could exist if a mutation occurred early within the bacterial culture's growth phase, or the bacterial culture was seeded from two bacteria in close proximity. Based on the thresholds we have set for consensus sequence generation, we expected a theoretical lower limit of detection for subclonal variant templates to be 20 % for single nucleotide variants (SNVs) and insertions (a variant template present at a frequency of less than 20 % would be indistinguishable from noise). For deletions, since we allowed coverage of a single base to drop to as low as 10 % of all aligned reads (to detect all bases within homopolymers), we expected that the sensitivity to detect subclonal deletions would be low, requiring greater than 90 % of the templates to contain the deletion. We tested these expectations, experimentally, by designing a mutant version of BCRxV.VSVG containing three SNVs, three insertions, and three deletions (**Supplementary File 1**). This variant plasmid was sequenced as described for the other plasmids, and the raw .fast5 reads were mixed with raw wildtype BCRxV.VSVG reads *in silico* at varying frequencies, allowing random pseudopairing of wildtype and mutant reads to occur as they would have if the templates were mixed, physically, prior to sequencing. The analysis was run for each dataset, which demonstrated that the sensitivity for detecting subclonal variant templates was consistent with our predictions. Two of the three insertions were detected when their

template frequency was approximately 20 %, with the other (which resulted in a lengthening of the 17mer homopolymer, and so would be predicted to be harder to detect) was detected at 40 % frequency (**Figure 4**). The three SNVs were detected with template frequencies ranging from 25 – 30%. The deletions were not detectable until the variant template was at a frequency of 95 %. At lower variant template frequencies, deletions were indistinguishable from the drop in signal observed within homopolymer regions, and are not able to be reliably detected.

Contaminating, unrelated plasmids are another type of variant template that we may wish to detect in our sample. These are plasmids whose presence should be detected by fingerprinting assays performed prior to sequencing; however it is possible, though exceedingly unlikely, that an unrelated plasmid would have the same fingerprint and length as the expected plasmid. As reads from such a plasmid would not be expected to align well to the expected reference, we can count the fraction of reads which are successfully pseudopaired, since this step only uses reads which have a single continuous alignment to the expected reference that is within 250 bp of the reference length. In our three sequencing runs, 92.3 – 97.2 % of quality- and length-filtered reads were successfully pseudopaired. In a simulation where we ran our pipeline using a subset of 10,000 quality- and length-filtered reads, but used a length-matched incorrect reference, 0 % of the reads were successfully pseudopaired. Hence, this metric can be checked to ensure there is no significant contamination of an unrelated plasmid, with values less than 90 % being evidence of plasmid contamination.

Discussion

We have described a pipeline that generates a high-quality consensus sequence of linearized plasmid using ONT MinION sequencing, leveraging substantial sequencing depth and stringent quality filters to overcome the relatively high error rates associated with nanopore sequencing. The purpose of this method is to verify the identity of a pure sample of plasmid with known sequence. For pure samples of linearized plasmid, our pipeline produces the true sequence with a median per-base confidence of over 99.9 %.

Our pipeline improves typical read error rates by an order of magnitude by performing paired basecalling on pseudopaired reads, leveraging the fact that we are sequencing a single plasmid. True duplexing would only improve this result, whether from a technological advance to generate paired data natively at the time of sequencing, or barcoding of every template molecule with a unique molecular identifier (UMI) prior to sequencing. This would ensure that paired reads came from the same template molecule, and would increase the sensitivity for detection of subclonal variant templates.

This pipeline is first and foremost a sequence verification method. In this embodiment, the user must know the expected sequence of the plasmid. Due to this, the approach is not entirely unbiased. Additionally, we report a single consensus sequence of all plasmid templates present in the sample rather than separate consensus sequences for each plasmid, if there should be more than one. Future developments could extend this framework to be run completely independently of the expected reference sequence, allowing for multiple consensus sequences to be output if multiple templates are detected.

With our current pipeline, subclonal insertion and SNV variants are detectable when they are at roughly 25 % frequency. Due to the necessarily low coverage threshold needed for

homopolyer stretches (at least 10 %), our sensitivity to detect subclonal deletions is low. It is conceivable that one could set a more stringent coverage threshold (80 %) globally while having a relaxed threshold (10 %) specifically within homopolymer regions. This would increase the sensitivity of non-homopolymer deletions to be comparable with insertions and SNVs, however, deletions within homopolymers would still require greater than 90 % frequency to be detected. For the current implementation, we acknowledge this limitation in detecting deletions.

Homopolymers themselves present challenges for accurate basecalling. Initially, we tested R9.4.1 flowcells, but found the performance within homopolymer regions to be insufficient. R10 generation flowcells are known to improve accuracy within homopolymers, increasing the number of nucleotides that are read within the pore at once (23). There is certainly an upper limit to the length of a hompolymer that can be accurately sequenced using our approach, though we have not yet identified it. Within our test plasmids, we have demonstrated successful sequence determination for homopolymers up to and including an 18mer.

While we have benchmarked this procedure on R10.3 flowcells, we fully expect that future developments to this platform (R10.4 and beyond) will only improve the accuracy. Before running this analysis on R10.4+ data, it is likely that the basecalling model fine-tuning would need to be repeated for the appropriate flowcell-matched basecalling model.

A major benefit of our pipeline over assembly and polishing approaches is that the output of our pipeline includes, in addition to a consensus sequence, the underlying per-base data for each position of the consensus. This is somewhat analogous to traditional Sanger sequencing approaches used for clinical sequence verification, and allows for manual inspection and interpretation of the quality of the resulting consensus. This allows the confidence, per base, to be assessed, strengthening the ability to use this pipeline as a replacement for Sanger sequencing of plasmids for clinical use.

Our pipeline can provide significant cost savings compared to outsourcing clinical-grade Sanger sequencing of plasmids, with estimated costs an order of magnitude lower for a single plasmid. For labs with access to a MinION sequencer and compute infrastructure, this approach could make generation of high-quality plasmid sequence for clinical sequence verification more accessible.

Materials and Methods

Restriction enzyme fingerprinting assay. The purity of each plasmid sample to be sequenced was confirmed by restriction enzyme fingerprinting prior to library construction. A pair of restriction enzymes was chosen to generate a unique banding pattern for each construct. Reactions were set up with 2.5 µg plasmid, 180 total units of restriction enzyme (New England Biolabs), plus 1 unit of Topoisomerase I (New England Biolabs, M0301S) to promote complete digestion. Digestions were run according to the manufacturer's instructions. The resulting banding patterns were visualized on 0.8% E-gels (Thermofisher, G501808). A representative gel is shown in **Supplementary Figure 3**.

Sequence library construction. Six-hundred fmol of each plasmid was linearized with a restriction enzyme (New England Biolabs) chosen to generate a single cut site with 5' overhangs. Each reaction was supplemented with 1 unit of Topoisomerase I (New England Biolabs, M0301S). Complete digestion was confirmed by running 200 ng of undigested plasmid alongside 200 ng of digested sample on 1.2% E-gel (Thermofisher, G21801). Samples were purified using Qiaquick PCR

Cleanup kit (Qiagen, 19086) following the manufacturer's instructions, and eluting with 50 µL pre-warmed (50°C) EB (Qiagen, 19086).

Libraries were prepared using the ONT SQK-LSK110 kit following the Genomic DNA by Ligation protocol (GDE_9108_v110_revC_10Nov2020). Samples were quantified with Qubit dsDNA BR kit (Q32850). 300 fmol of linearized plasmid was repaired using NEBNext FFPE DNA Repair Mix (NEB, M6630S) and NEBNext Ultra II End Repair/dA Tailing Module (NEB, E7546S). Samples were purified with a 1× volume of Ampure XP beads (Beckman, A63880), and eluted with 61 µL pre-warmed (50°C) Qiagen EB. Adapter ligation was performed with NEB Quick Ligase (E6056S), ONT Ligation Buffer (LNB) and ONT adapter mix (AMX-F). Samples were purified with 0.5× Ampure XP beads, and washed with ONT Long Fragment Buffer, and eluted with 15 µL ONT Elution Buffer (included in SQK-LSK110 kit).

Sequencing using ONT MinION. MinION R10.3 (FLO-MIN111) flowcells were prepared and run according to ONT instructions: 50 fmol of prepared library was loaded per flowcell, and the run time was set to 72 hours. The MinION USB device was connected to a computer running Windows 10, with an Intel Core i7-9700 8 core CPU, 32GB RAM, and a 1TB SSD data drive. On-device basecalling was disabled, instead performing basecalling in a later step on our higher-powered computational resources.

Basecalling raw sequence data with Guppy. We assembled all steps of our pipeline into a Snakemake v5.7.4 workflow (**Supplementary Figure 4**). We performed initial basecalling of the raw .fast5 data using the GPU version of Guppy v5.0.11. We ran basecalling on a GPU cluster comprised of eight NVIDIA GeForce RTX 3090 GPUs, 256GB RAM, and two Intel Xeon Silver 4212 12 core CPUs. Each .fast5 file generated by MinKNOW was submitted to the cluster in its own job, requesting 12GB RAM and 1 GPU. We used the "dna_r10.3_450bps_hac.cfg" config file and "template_r10.3_450bps_hac.jsn" model file packaged with Guppy, and set "--min-qscore" to 12. Basecalling took roughly 1 minute per .fast5 file. We used a custom Python script to filter the resulting basecalled reads to those that had length within 250bp of the expected sequence length.

Pseudopairing reads. In order to pseudopair forward and reverse reads, we performed an initial alignment of the basecalled reads to the expected reference sequence. We used minimap2 v2.17, with the "-x map-ont" and "--cs" options, generating a .paf file. We then use a custom Python script to parse this .paf file, generating a list of forward and reverse read names which contain only a single alignment to the reference, and for which the number of aligned bases is no shorter than the reference length minus 500 bases. We sort these two lists by read length, and step through each simultaneously, outputting pseudopaired forward and reverse read names to a file.

Fine-tuning the Bonito basecalling model. We performed Bonito basecalling model fine-tuning as described in the Bonito manual. We fine-tuned the "dna_r10.3" model included with Bonito v0.4.0 using 100 reads of our BCRxV.VSVG plasmid, requesting 64GB of RAM and using a batch size of 32, 1 epoch, and a learning rate of 5e-4. We used this fine-tuned model for all Bonito duplex runs.

Paired basecalling with Bonito. We used a modified version of Bonito v0.4.0 to perform the duplex basecalling of the read pairs, making some minor adjustments to the Bonito code to prevent the process from using too much memory on our system (https://github.com/scottdbrown/bonito/releases/tag/v0.4.0a). We submitted the Bonito duplex basecalling jobs to our GPU cluster, requesting 40GB of RAM and 6 CPUs per job. We set "--max-cpus" to 4, and used our own fine-tuned basecalling model. On average, Bonito duplex basecalling jobs of around 750 read pairs took 20 minutes to complete.

De novo assembly to generate a reference. We combined all duplex basecalled reads into a single .fasta file. We subsampled 500 reads from this file to use for *de novo* assembly using Canu v2.0 (16). We set the "-corrected" and "-nanopore" options, set the "genomeSize" to the length of our expected reference, and used 16 threads. We extracted the first sequence in the resulting "unitigs.fasta" file to use as the reference sequence for subsequent read alignment and consensus sequence determination.

Consensus sequence generation. To generate a consensus sequence for our sequence data, we align the duplex basecalled data to the *de novo* assembled reference sequence. Because of the logic of our consensus sequence building, and due to the standard practice of left-aligning gaps in indels, we perform two separate consensus sequence determination steps using both the *de novo* assembly as provided, as well as the reverse compliment. This is to ensure that any variant near a homopolymer stretch does not have the variant support diluted due to alignment artefacts (**Supplementary Figure 2**).

We align the duplex basecalled reads to the *de novo* assemblies, separately, using minimap2 v2.17, with the "-x map-ont" and "--cs" options, generating a .paf file. We then use a custom Python script to parse the alignments. We build an array with positions for 5' of the reference sequence, every base of the reference sequence, every position in between bases of the reference sequence, and 3' of the reference sequence. We step through every read alignment and update a dictionary at each position recording the number of reads supporting each observed base at each position (with multiple bases represented as an array of dictionaries). We then build the consensus sequence by stepping through every position of the array and adding the most evidenced base if it meets our criteria: (1) the top base being at least 5× more abundant than the second top base, and (2) the coverage of the top base being at least 10 % of the max coverage. If only the first criterion is not met, an N base is added to the consensus, signifying insufficient read support to accurately call a base at this position. If the second criterion is not met, no base is added to the consensus.

Variant plasmid design. To test the detection of subclonal variant templates, we designed a 1040 bp variant insert for our BCRxV.VSVG plasmid with the following nine mutations: 1886delT, T2010A, 2205insA, 2379delT, 2448insA, G2517C, 2685insA, 2770delG, and C2840T (positions given relative to resulting sequence after linearization with BamHI). We had this fragment synthesized by Azenta Life Sciences, and cloned it into our wildtype BCRxV.VSVG plasmid using SfuI and BalI restriction sites.

Subclonal variant simulation. We sequenced our variant BCRxV.VSVG plasmid as described for the other plasmids. To simulate sequence datasets with BCRxV.VSVG variant plasmids at varying levels of abundance, we first performed Guppy basecalling on all wildtype and variant data to obtain the read names that passed quality filtering. We then created raw .fast5 read sets of 10,000 total reads *in silico*, ranging from 0 % to 100 % variant reads in 5 % increments. We then processed these read sets as described, allowing random pseudopairing of wildtype and variant reads to occur, and interrogated the resulting consensus sequences to determine at what variant read frequency each mutation was detected.

Quality reports. To obtain quality metrics of our duplex basecalled reads, we used Qualimap v2.2.2 to provide a summary report on these reads aligned to the expected reference sequence. We used minimap2 v2.17 with the "-ax map-ont --cs" options to generate a .sam file, used samtools v1.7 to convert the .sam file to a .bam file and to index this .bam file, then ran Qualimap bamqc with 32 threads and 16GB of memory. From this report, we obtain the number of aligned reads, average coverage, mean error rates of the reads, and view the plots to ensure there are no issues with the data.

End Matter

Data and Code Availability

Custom code is available at the following GitHub repository: https://github.com/scottdbrown/minion-plasmid-consensus. Raw .fast5 sequence data for our four sequencing runs is available upon request.

Author Contributions

R.A.H and S.D.B. designed research, L.D. performed plasmid preparation, sequence library construction, and plasmid sequencing, S.D.B. analyzed data and developed the pipeline, J.F.W. and M.B. oversaw quality control and data interpretation, and S.D.B., L.D., and R.A.H. wrote the paper.

Acknowledgments

This research was supported by the British Columbia Cancer Foundation, the Leon Judah Blackmore Foundation, and the BioCanRx network. We would like to thank Luka Culibrk for helpful review of the code.

References

1. FDA (2007) Guidance for industry: Considerations for plasmid DNA vaccines for infectious disease indications.

2. Health Canada Therapeutic Products Programme (2001) Guideline for Industry Quality of Biotechnological Products☐: Analysis of the Expression Construct in Cells Used for Production of r-DNA Derived Protein Products, ICH Topic Q5B.

3. Sanger,F., Nicklen,S. and Coulson,A.R. (1977) DNA sequencing with chain-terminating inhibitors. *Proc. Natl. Acad. Sci. U. S. A.*, **74**, 5463–5467.

4. Shapland,E.B., Holmes,V., Reeves,C.D., Sorokin,E., Durot,M., Platt,D., Allen,C., Dean,J., Serber,Z., Newman,J., *et al.* (2015) Low-Cost, High-Throughput Sequencing of DNA Assemblies Using a Highly Multiplexed Nextera Process. *ACS Synth. Biol.*, **4**, 860–866.

5. Gallegos,J.E., Rogers,M.F., Cialek,C.A. and Peccoud,J. (2020) Rapid, robust plasmid verification by de novo assembly of short sequencing reads. *Nucleic Acids Res.*, **48**, e106–e106.

6. Laver,T., Harrison,J., O'Neill,P.A., Moore,K., Farbos,A., Paszkiewicz,K. and Studholme,D.J. (2015) Assessing the performance of the Oxford Nanopore Technologies MinION. *Biomol. Detect. Quantif.*, **3**, 1–8.

7. Rang,F.J., Kloosterman,W.P. and de Ridder,J. (2018) From squiggle to basepair: Computational approaches for improving nanopore sequencing read accuracy. *Genome Biol.*, **19**, 1–11.

8. Tabata,Y., Matsuo,Y., Fujii,Y., Ohta,A. and Hirota,K. (2022) Rapid detection of single nucleotide polymorphisms using the MinION nanopore sequencer: a feasibility study for perioperative precision medicine. *JA Clin. Reports*, **8**, 1–8.

9. Wick,R.R., Judd,L.M., Wyres,K.L. and Holt,K.E. (2021) Recovery of small plasmid sequences via Oxford Nanopore sequencing. *Microb. genomics*, **7**.

10. González-Escalona,N., Allard,M.A., Brown,E.W., Sharma,S. and Hoffmann,M. (2019) Nanopore sequencing for fast determination of plasmids, phages, virulence markers, and antimicrobial resistance genes in Shiga toxin-producing Escherichia coli. *PLoS One*, **14**, e0220494.

11. Taylor,T.L., Volkening,J.D., DeJesus,E., Simmons,M., Dimitrov,K.M., Tillman,G.E., Suarez,D.L. and Afonso,C.L. (2019) Rapid, multiplexed, whole genome and plasmid sequencing of foodborne pathogens using long-read nanopore technology. *Sci. Reports 2019 91*, **9**, 1–11.

12. Prussing,C., Snavely,E.A., Singh,N., Lapierre,P., Lasek-Nesselquist,E., Mitchell,K., Haas,W., Owsiak,R., Nazarian,E. and Musser,K.A. (2020) Nanopore MinION Sequencing Reveals Possible Transfer of bla KPC-2 Plasmid Across Bacterial Species in Two Healthcare Facilities. *Front. Microbiol.*, **11**.

13. Peter,S., Bosio,M., Gross,C., Bezdan,D., Gutierrez,J., Oberhettinger,P., Liese,J., Vogel,W., Dörfel,D., Berger,L., *et al.* (2020) Tracking of Antibiotic Resistance Transfer and Rapid Plasmid Evolution in a Hospital Setting by Nanopore Sequencing. *mSphere*, **5**.

14. Currin,A., Swainston,N., Dunstan,M.S., Jervis,A.J., Mulherin,P., Robinson,C.J., Taylor,S., Carbonell,P., Hollywood,K.A., Yan,C., *et al.* (2019) Highly multiplexed, fast and accurate nanopore sequencing for verification of synthetic DNA constructs and sequence libraries. *Synth. Biol.*, **4**, 25.

15. Wick,R.R., Judd,L.M., Cerdeira,L.T., Hawkey,J., Méric,G., Vezina,B., Wyres,K.L. and Holt,K.E. (2021) Trycycler: consensus long-read assemblies for bacterial genomes. *Genome Biol.*, **22**, 1–17.

16. Koren,S., Walenz,B.P., Berlin,K., Miller,J.R., Bergman,N.H. and Phillippy,A.M. (2017) Canu: scalable and accurate long-read assembly via adaptive k-mer weighting and repeat separation. *Genome Res.*, **27**, 722–736.

17. Kolmogorov,M., Yuan,J., Lin,Y. and Pevzner,P.A. (2019) Assembly of long, error-prone reads using repeat graphs. *Nat. Biotechnol. 2019 375*, **37**, 540–546.

18. Rive,C.M., Yung,E., Dreolini,L., Brown,S.D., May,C.G., Woodsworth,D.J. and Holt,R.A. (2022) Selective B cell depletion upon intravenous infusion of replication-incompetent anti-CD19 CAR lentivirus. *Mol. Ther. - Methods Clin. Dev.*, **26**, 4–14.

19. Standard Finishing Practices and Annotation of Problem Regions for the Human Genome Project (2001).

20. Krishnakumar,R., Sinha,A., Bird,S.W., Jayamohan,H., Edwards,H.S., Schoeniger,J.S., Patel,K.D., Branda,S.S. and Bartsch,M.S. (2018) Systematic and stochastic influences on the performance of the MinION nanopore sequencer across a range of nucleotide bias. *Sci. Reports 2018 81*, **8**, 1–13.

21. Branton,D., Deamer,D.W., Marziali,A., Bayley,H., Benner,S.A., Butler,T., Di Ventra,M., Garaj,S., Hibbs,A., Huang,X., *et al.* (2008) The potential and challenges of nanopore sequencing. *Nat. Biotechnol. 2008 2610*, **26**, 1146–1153.

22. Silvestre-Ryan,J. and Holmes,I. (2021) Pair consensus decoding improves accuracy of neural network basecallers for nanopore sequencing. *Genome Biol.*, **22**, 1–6.

23. Oxford Nanopore Technologies (2020) R10.3: the newest nanopore for high accuracy nanopore sequencing – now available in store. *News Artic. Mon 13th January 2020.*

Figure 1. Strand-specific errors in basecalling. Sequence logos are shown for one stretch of sequence from forward (top) and reverse (bottom) reads. The height of each letter gives the fraction of reads at that position that supported that base. Position 6 (and position 3, to a lesser extent) shows a strand-specific error – the correct base is C, but in the reverse reads a T is the most-evidenced base. When forward and reverse reads are pooled, there appears to be two bases at this position.

Figure 2. Overview of consensus read support across BCRxV.GagPolRev. Filled points denote the most-evidenced base at each position (x-axis) and hollow points denote the second-most-evidenced base. The y-axis shows the number of reads supporting each base. In general, nearly all reads support the top base, and there is good separation between signal (filled points) and noise (hollow points). For BCRxV.GagPolRev, the lowest signal to noise ratio is between the C and T at position 7554. We see a decrease in coverage at the extreme 5' and 3' ends of the plasmid.

Figure 3. Decreased read support across homopolymer stretches. Points denote the most-evidenced base at each position across this subset of BCRxV.VSVG with a 17mer homopolymer. Due to difficulty basecalling the exact number of bases in homopolymer regions, many reads report fewer than 17 bases. This results in an apparent coverage drop off across the homopolymer. The horizontal dashed grey line at 10 % of the max coverage represents the minimum base coverage required to add base to the consensus. Note that for this visualization, indels were right-aligned to show coverage dropping from left to right.

Figure 4. Subclonal variant template detection sensitivity. For three different mutation classes (SNV, insertion, deletion), the minimum percent abundance of variant reads (x-axis) that allowed detection of the mutation in the *in silico* generated mixed datasets is plotted. SNVs and insertions are detectable when around 30 % of reads are from the variant template, while deletions were not detectable until 95 % of the reads were from the variant template.

Supplementary Figure 1. Raw read length distribution for BCRxV.VSVG. A density plot showing the distribution of quality-filtered read lengths obtained from BCRxV.VSVG. The region between the dashed vertical lines show the read lengths that pass length-filtering (are within 250 bp of the expected reference length). The x-axis has been truncated at 15,000 bp, removing 0.4 % of reads with length longer than 15,000 bp. The longest observed read had a length of 356,764 bp.

Supplementary Figure 2. An illustrative example of inhibited variant detection near a homopolymer due to an artifact of left-aligning indels. The reference sequence that reads are being aligned to is shown in blue. All reads (brown) are missing the "G" base at position 5 of the reference (supporting a deletion), and some reads are also missing "A" bases of the homopolymer (a common error in Nanopore reads). When aligning these reads to the reference sequence as provided (sense strand), due to left-aligning indels, the read support for the "G" base at position 3 gets spread across positions 3 and 5, resulting in the deletion not being reported by the pipeline (bottom left). If the same reads are aligned to the reverse compliment of the reference (antisense strand), left-aligning the indels allows the gaps to fall at the position of the deleted base, resulting in the deletion being detected and reported (bottom right).

Supplementary Figure 3. Representative plasmid fingerprinting gel. BCRxV.GagPolRev.1 was digested with PvuII-HF and SalI-HF to generate an expected pattern of 3.744, 3.047, 2.154, 1.044 and 0.804 kb bands.

Supplemental Figure 4. Flow diagram of Snakemake analysis pipeline. An overview of the Snakemake pipeline, showing how the inputs and outputs of different rules are connected. The top half of the plot (above "merge_fasta") is showing the flow for three raw .fast5 files.

Supplementary File 1. Reference sequences for the four plasmids sequenced in .fasta format.