Backscattering Mueller Matrix polarimetry on whole brain specimens shows promise for minimally invasive mapping of microstructural orientation features

Justina Bonaventura [1], **Kellys Morara** [2], **Rhea Carlson** [2], **Courtney Comrie** [2], **Noelle Daigle** [1], **Elizabeth Hutchinson** [2], and **Travis W. Sawyer** [1,2,3,*]

[1] *Wyant College of Optical Sciences, University of Arizona, Tucson, AZ, USA*
[2] *Department of Biomedical Engineering, University of Arizona, Tucson, AZ, USA*
[3] *College of Medicine, University of Arizona, Tucson, AZ, USA*

Correspondence*:
Travis W. Sawyer, University of Arizona, 1230 N Cherry Ave, Tucson AZ 85719
tsawyer9226@arizona.edu

2 ABSTRACT

Understanding microscale physiology and microstructural cellular features of the brain is key
to understanding the mechanisms of neurodegenerative diseases and injury, as well as the
prominent changes that neurons and glia in the brain undergo in development and aging which
reflect functional state. Non-invasive imaging modalities sensitive to the microscale - especially
diffusion magnetic resonance imaging (dMRI) - are extremely promising for three-dimensional
mapping of cellular microstructure of brain tissues and brain connectivity via tractography;
however, there is a need for robust validation techniques to verify and improve the biological
accuracy of fiber orientation information derived from these techniques. Recent advances in
dMRI acquisition and modeling have moved toward probing of the more complex grey matter
architecture, challenging current validation techniques, which are largely based on *ex vivo* staining
and microscopy and focused on white matter. Polarized light imaging (PLI) has been shown to be
a successful technique for high resolution, direct, microstructural imaging and has been applied
to dMRI validation with clear advantages over conventional staining and microscopy techniques.
Conventionally, PLI is applied to thin, sectioned samples in transmission mode, but unlike
histologic staining, PLI can be extended to operate with high sensitivity in reflectance mode and
even extended to 3D imaging to bridge the gap toward *in vivo* validation of dMRI measurements
of orientation features in both gray and white matter of the brain. In this report we investigate the
use of backscattering Mueller Matrix polarimetry to characterize the microstructural content of
intact Ferret brain specimens. The experimental results show that backscattering polarimetry can
probe white matter fiber coherence and fiber orientation in whole brain specimens, and show
promise for probing grey matter microstructure. Ultimately, these preliminary results motivate
further study to fully understand how backscattering polarimetry can best be used for validation
of *in vivo* microstructural imaging of the brain.

Keywords: backscattering polarimetry, Mueller matrix, brain tissue, white matter, microstructure

1 INTRODUCTION

Developing methods to probe microscale physiology and microstructural cellular features within the brain
is a rich topic of research both in the context of clinical care and neuroscientific studies. Understanding

structure-function relationships in the human brain is critical both to understanding normal brain devel-
opment and function, and also to elucidate the mechanisms of disorders and diseases that alter brain
physiology, for example neurodegenerative diseases such as Alzheimer's and Parkinson's disease (Lamptey
et al. (2022)).

Advances in neuroimaging technologies have enabled a wide range of research studies to explore
changes in the brain structure and function during healthy aging and under pathological conditions. In
particular, the last two decades have seen an explosion of studies using Magnetic Resonance Imaging
(MRI), which enables *in vivo*, depth-resolved, and minimally invasive 3D visualization of structural and
functional properties of the brain (Wang et al. (2021); Forouzannezhad et al. (2019)). Diffusion MRI
(dMRI) techniques, such as diffusion tensor imaging, are a subset of MRI that generate contrast using
the diffusion of water molecules (Novikov et al. (2019)). dMRI can be used to provide exquisite maps
of spatial organization and connectivity of white matter, as the cell structure of axons limits molecular
movement perpendicular to the axonal fibers, giving rise to anisotropic diffusion (Le Bihan (2014)). As
such, dMRI has become the modality of choice for investigating the neural architecture of the brain.

dMRI has been used extensively to study white matter changes during brain development and aging
(Rathi et al. (2014); Gunning-Dixon et al. (2009); Madden et al. (2009)), in the context of pathological
disorders and diseases (Nguyen et al. (2022); Zhao et al. (2022); Rosenkranz et al. (2022)), as well as from
traumatic brain injury (Hutchinson et al. (2018); Minchew et al. (2022); Guerrero-Gonzalez et al. (2022);
Liu et al. (2022)). However, the application of dMRI to cortical gray matter analysis has only recently
begun to be explored, with studies reporting age related changes using diffusion imaging based measures
(Bhagat and Beaulieu (2004); Pfefferbaum et al. (2010)). While dMRI is a powerful technique, one
technological limitation is a lack of sufficient specificity to directly measure microstructural and molecular
quantities, instead relying heavily on biophysical modeling or indirect inference; as a result, validation of
the "biological accuracy" of dMRI estimates is essential for advancing the technology. Validation of dMRI
metrics most commonly utilize histologic staining which is limited by stain specificity, tissue processing
and sectioning artifacts, and a lack of quantitative methods. While more sophisticated methods including
structural tensor analysis and electron microscopy have improved these efforts, these all remain limited to
analysis of fixed, sectioned, *ex vivo* tissues with mm-scale field of view.

Polarized light imaging (PLI) is a popular technique used for probing physiological structures in the brain
with high spatial resolution, which has shown great promise for dMRI validation (Mehta et al. (2013)). In
particular, the myelin sheath surrounding neurons is strongly birefringent (Morgan et al. (2021)), which
will cause a phase delay in polarized light along the axis as the neuronal fibers. Traditionally, PLI is
conducted by transmitting light of a specific polarization through a thin (on the order of 100 microns) slice
of fixed tissue. The polarized light will interact with the cellular and sub-cellular structures, altering the
polarization state, which is then measured by analyzing the transmitted light with a polarizer and camera
detector. Rotating the analyzing polarizer in discrete increments and measuring the resulting transmitted
power enables the estimation of the magnitude birefringence (also called "retardance") and the orientation
of the underlying structures (the "retardance angle") (Larsen et al. (2007)). While valuable for mapping out
fiber orientation, this approach is limited to only measuring specimens that are thin, fixed, and sectioned.
Moreover, retardance and retardance angle are only two of many different polarization properties that could
encode valuable information about the underlying tissue microstructure (Ghosh et al. (2008, 2009)).

Backscattering PLI, in particular those based on Mueller Matrix or Stokes polarimetry is emerging
as a valuable technique for comprehensive polarization characterization of tissue specimens. Multiple
approaches exist for Mueller Matrix polarimetry but in general, an optical system is constructed to cycle

through different input polarization states and sample different output polarization states (Qi et al. (2018)).
The collected data can then be used to reconstruct the Mueller Matrix, which provides a complete picture of
the polarization properties of the material. The Mueller Matrix can be further decomposed into individual
elements such as the linear retardance, diattenutation and circular polarization (Ghosh and Vitkin (2011)).
Some recent works have shown that multi-wavelength Stokes polarimetry has great promise for brain
microstructural mapping (Menzel et al. (2015); Borovkova et al. (2022)), but current studies remain limited
to sectioned specimens. There remains a need to advance this technology toward *in vivo* acquisition, the
first step of which is to demonstrate the utility of the technique on unsectioned whole brain specimens.

In this brief research report, we present a preliminary study to test the feasibility of multi-wavelength
backscattering Mueller Matrix polarization microscopy for microstructural characterization of whole brain
specimens in reflectance mode. For three specimens of ferret brain, we measure three different regions
of the brain with known microstructural geometry: two composed of coherent white matter, and one
composed of grey matter. We demonstrate that the technology is sensitive to orientation and fiber coherence
in white matter, and shows promise for measuring grey matter microstructure. These results motivate
further study of backscattering Mueller Matrix polarimetry in the context of *in vivo* characterization of
brain microstructure, particularly for grey matter regions.

2   MATERIALS AND METHODS

2.1   Specimen Preparation

90 Perfusion-fixed ferret brain specimens (n=3) were generated in a previous unrelated study and provided
91 for whole brain imaging in this study (Hutchinson et al. (2017)). Briefly, adult male ferrets underwent
92 perfusion fixation with 4% paraformaldehyde (PFA). The brains were removed and post-fixed in 4% PFA
93 for 10 days and then transferred for re-hydration and long-term storage under refrigeration in phosphate
94 buffered saline with 0.01% sodium azide.

2.2   Polarization Imaging

For fixed brain specimens (Figure 1A), images were collected for three regions with known physiological
and microstructural characteristics using a benchtop multispectral polarization microscope. Images were
acquired of the corticospinal tract (CST), the optic chiasm (OC) and the cerebellum. These regions were
selected since each has well known microstructural orientation. The corticospinal tract is a one-directional
white matter fiber bundle along the direction of the spinal tract; the optic chiasm is composed of two,
crossing white matter fiber bundles; and the cerebellum is a region of grey matter, with parallel fibers
located beneath the surface layer. Thus, these three regions provide increasing complexity to examine the
capabilities of the polarization imaging technology.

The polarization microscope is a custom instrument developed by Nikon (Oka and Saito (2006); Saito
et al. (2013)), which consists of an episcopic illumination system capable of dark-field illumination in five
wavelengths and a polarimetric imaging system for measuring the state of polarization of scattered light
from the tissue sample positioned at the microscope's specimen stage (Figure 1B,C). A Nikon bright-field
objective lens with minimal polarization distortion (Nikon XFI TU Plan Fluor EPI) was used to operate
within the epi-illumination and dark-field imaging constraints of this polarimeter. 5X magnification was
used to capture all images due to its larger field of view and depth of field. The instrument has been
previously characterized, validated, and used for human tissue imaging, and produces spatial pixel-wise

mappings of the sixteen Mueller Matrix components (Fujii et al. (2019); Saito et al. (2019)). Exposure
times were optimized using the instruments "auto-adjust" setting.

2.3 Image Processing and Analysis

Wavelengths where the Mueller Matrix was unsuccessfully reconstructed were excluded from analysis.
Incomplete reconstruction can be caused by tissue-dependent saturation, numerical instability, or other
sources of noise. Ultimately, the 405 nm and 442 nm were analyzed for the CST, all wavelengths except 442
nm were analyzed for the OC and all five wavelengths were analyzed for the cerebellum. The raw Mueller
Matrix maps were decomposed using the Lu-Chipman decomposition to extract polarization parameters
including retardance, and retardance angle (Figure 1D) (Lu and Chipman (1996)).

For each region that was imaged, a sample-specific region of interest (ROI) was selected using ImageJ
using the depolarization image as a reference (Rasband (2012)). Each region of interest was saved as a
separate binary mask. This process was done in order to quantify the specific tissues of interest. For the
CST, an ROI was selected for the spinal tract and the surrounding background region (known as the olivary
complex, which lacks significant white matter directionality); for the OC, an ROI was selected for each of
the two, crossing, coherent white matter bundles (referred to as left and right lobes), and for the cerebellum,
a single ROI was selected for a region of the surface with minimal curvature.

Pixel-wise distributions of retardance and retardance angle were generated to qualitatively assess the
polarizing characteristics of each ROI. For the CST, the total retardance was compared between the CST
and background by taking the mean value of the distribution for each sample and averaging across the
collected samples. We characterize the distributions of the retardance and retardance angle by measuring
the mean, standard deviation, and kurtosis and taking the ratio of the CST to the background. For the OC,
we compare the mean retardance of the distributions, as well as the difference in mean retardance angle
between the different samples. For the cerebellum, we quantify the mean, standard deviation, and kurtosis
of the retardance and retardance angle distributions as a function of wavelength.

2.4 Statistical Analysis

Statistical analysis of the quantitative comparisons for retardance and retardance angle for CST and OC
was done using a paired t-test, as each sample had two comparable regions of interest. The cerebellum only
had one region of interest and therefore no statistical comparisons were conducted. For the OC, we also
extracted a radial slice and examined the retardance angle along this slice. This was done to examine the
response of the imaging technique in a region of crossing fibers. To determine if the response is produced
as an averaging of the microstructure or responding to the dominant orientation, we fit the resulting data
along the slice to a linear curve. If backscattering PLI is sensitive to the averaging of the microstructure,
the retardance angle would manifest as a linear gradient across the crossing fiber region; alternatively,
sensitivity to the dominant microstructure would manifest as a step-function between the two crossing fiber
directions. To test the linearity of this relationship, we conduct a hypothesis test with a null hypothesis that
the slope is zero by using a Wald Test with t-distribution of the test statistic.

3 RESULTS

3.1 Corticospinal Tract

Figure 2 shows the results for measurements of the CST for whole brain specimens. Of the five wave-
lengths used for imqwaging, only 405 nm and 445 nm were unsaturated with the optimized exposure times.

Figure 2A is a photograph of the specimen, while Figure 2B and 2C show the retardance and retardance
angle maps for the region of interest showing the CST against the background olivary region (BKG).
Quantification of these maps (Figure 2D,E) shows that there is a significant increase in retardance for the
CST compared to the background (p<0.05) for both 405 nm and 442 nm. The retardance distributions for
both tissue types resemble a normal distribution with similar spreads, but the CST has a significantly higher
mean retardance. Distributions of retardance angle (Figure 2E) indicate that the CST has a significantly
more coherent distribution of angles, which can be shown quantitatively by nearly a four-fold decrease of
standard deviation (p<0.05 at 405 nm and 442 nm), without any significant change in mean or kurtosis.

3.2 Optic Chiasm

Figure 3 shows the results for measurements of the OC for whole brain specimens. Only the 442 nm
wavelength was excluded due to saturation and unsuccessful reconstruction of the Mueller Matrix. Figure
3A shows a photograph with the ROI outlines, while 3B and 3C show the retardance and retardance angle
maps for the region of interest showing the left and right lobes. The retardance angle along the radial slice
shown as a black curve in Figure 3C was extracted and the retardance angle was plotted against the radial
angle (Figure 3D), which was found to have a linear slope (p<0.001), suggesting that the retardance angle
is dictated by an average of the underlying microstructure, and not dictated by the dominant microstructural
influence.

Distributions of the retardance and retardance angle for the ROIs (Figure 3E, 3F) shows that the retardance
of both lobes is approximately equal, which is expected given that both lobes are composed of white matter.
However, the retardance angle distributions indicate that the difference in retardance angle becomes more
prominent at higher wavelengths. This may be due to the higher penetration depth of longer wavelengths,
enabling more sensitive measurements. Figures 3G and 3H show the quantification of these distributions,
which support the qualitative observations; where there is no significant different in retardance across all
wavelengths, but there is a statistically significant difference in the angle of the two lobes at a wavelength
of 632 nm (p<0.05). Some distributions such as the Right Lobe in Figure 3E,F seem to exhibit bimoodal
behavior, which may be caused by the influence of the curved geometry of the sample resulting in variations
in fiber orientation.

3.3 Cerebellum

Figure 4 shows the results for the measurements of the cerebellum. Figure 4A is a photograph of the
specimen, with Figure 4B demonstrating the expected underlying microstructural content of parallel fiber,
which runs along the length of the ridges. Figures 4C and 4D show maps of the retardance and retardance
angle, respectively. Distributions of the retardance and retardance angle for the ROI (Figure 4E, 4F) seems
to suggest that the retardance is relatively constant across the tissue, but decreasing with higher wavelength.
Conversely, the retardance angle seems to increase but become less tightly distributed. Quantification of the
characteristics of these distributions through mean, standard deviation and kurtosis (Figure 4G, 4H) shows
consistent observations - the mean retardance decreases with increasing wavelength and mean retardance
angle increases with increasing wavelength. Given that the ridges are not oriented horizontally, we would
expect a nonzero retardance angle if the measurements are sensitive to the underlying parallel fibers. These
results seem to suggest that higher wavelengths may be more sensitive to the underlying microstructure,
which could be due to the higher penetration depth. A further consideration regarding penetration depth is
that an additional orientation feature may arise due to the climbing fibers that reach beneath the molecular
layer of the cerebellum (Figure 4B).

4   DISCUSSION

Our preliminary results suggest that backscattering Mueller Matrix polarimetry could be a promising technique for noninvasive mapping of microstructural orientation features in the brain. We show that we are able to obtain measurements of fiber coherence and density (retardance), as well as fiber orientation (retardance angle) that are consistent with known microstructural characteristics. One advantage to the Mueller Matrix approach is that the measurement also provides information about other polarization parameters including depolarization, diattenuation, and circular polarization interaction. These properties may also contain valuable information for probing tissue microstructure, which could complement the known markers of retardance and retardance angle.

Primarily, other works investigating Mueller Matrix polarimetry in the context of brain microstructural imaging are limited to sectioned, fixed specimens of brain tissues (Borovkova et al. (2022); Schucht et al. (2020)). Our demonstration of this technology is valuable in that it evaluates the potential for *in vivo* use through the application to intact whole brain specimens in reflectance mode. However, other steps are necessary to conduct *in vivo* studies, for example optimizing acquisition times to mitigate motion artifacts. In addition, our study shows preliminary promise for using PLI to measure to grey matter microstructure, for which developing validation techniques is becoming increasing important as dMRI methods continue to advance. Further study is needed to validate the capabilities of this technology for grey matter microstructural assessment and likely hardware and image analysis algorithm optimizations are required to maximize the performance of the technology.

While promising, our results indicate that the selection of wavelength is critical. Wavelength-resolved light-tissue interaction is predominantly dictated by absorption, scattering, and penetration depth (Tuchin (2003)). Primary tissue absorbers are melanin, lipids and water, which give rise to an optical window in the visible spectrum (Smith et al. (2009)). Scattering will be influenced by wavelength and particle size but generally is higher for lower wavelengths, leading to decreased penetration depth. Given the heterogeneity of brain tissues, it may be necessary to optimize acquisition wavelength for specific anatomical structures and regions; we see that for the white matter regions such as the corticospinal tract, there is severe oversaturation at higher wavelength compared to the grey matter regions in the cerebellum. Our measurements may suggest that there is some sensitivity of polarized light to grey matter microstructure, as we observed an increased signature of retardance angle from the cerebellum, which is consistent with the angle of parallel fibers. This becomes more pronounced at higher wavelengths, which may be due to the requisite deeper penetration needed for the light to interact with the parallel fiber structures. In addition, the wavelength of light may influence the size of the tissue structures it primarily interacts with, providing more motivation to further understand how wavelength selection can be used as a tool to optimize the acquisition of microstructural information.

As anticipated, measuring non-flat geometries, such as with whole brain specimens, is particularly challenging and may introduce artifacts into the results. To account for this effect, further studies utilizing tissue phantoms will be undertaken to characterize how curved surfaces may influence the acquisition and reconstruction of Mueller Matrix data. The influence of topology may be correctable through calibration and data processing techniques; alternatively, hardware optimization could also be undertaken to increase the depth of field to enable in-focus imaging of samples with significant topological features. With a two-dimensional acquisition, the technology is only sensitive to in-plane orientation; further innovations could be made with the integration of three-dimensional acquisition through sample or hardware tilting, among other approaches Menzel et al. (2015); Borovkova et al. (2022).

5   CONCLUSION

Evaluation of microscale physiology and microstructural cellular features of the brain is becoming increasingly important to understand the mechanisms of neurodegenerative diseases, development and aging. While imaging modalities such as diffusion magnetic resonance imaging (dMRI) are extremely powerful for noninvasive, three-dimensional, microstructural characterization of tissues, there is a need for robust validation techniques to verify the biological accuracy of dMRI imaging data, particularly as these techniques are extended to probe grey matter. Polarization imaging is a promising technique for high resolution direct microstructural imaging, but application is typically limited to thin sectioned samples in transmission mode. We have shown for the first time that backscattering Mueller Matrix polarimetry in reflectance mode can be used to map microstructural features of intact brain tissue specimens with high biological accuracy. The experimental results show that backscattering polarimetry can probe white matter fiber coherence and fiber orientation in whole brain specimens, and show promise for probing grey matter microstructure. Ultimately, these preliminary results motivate further study to fully understand how backscattering polarimetry can best be used for validation of *in vivo* microstructural imaging of the brain.

CONFLICT OF INTEREST STATEMENT

The authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

AUTHOR CONTRIBUTIONS

KM and RC contributed to data collection. JB assisted with data collection and performed image analysis and processing. CC, RC, and EH contributed to specimen preparation. TS and EH contributed to the study design. TS wrote the first draft of the manuscript. ND contributed to data collection and manuscript preparation. All authors contributed to manuscript revision, read and approved the submitted version.

FUNDING

Funding for this work was provided by Wyant College of Optical Sciences Friends of Tucson Optics (FoTO) Endowed Scholarships.

ACKNOWLEDGMENTS

We would like to thank the Nikon Corporation and Nikon Research Corporation of America for the use of the polarimeter system, as well as Saito Naooki and Heather Durko for technical guidance. We also would like to thank Faith Rice and Dr. Jennifer Barton for assistance with operating the polarimeter.

REFERENCES

Bhagat, Y. A. and Beaulieu, C. (2004). Diffusion anisotropy in subcortical white matter and cortical gray matter: changes with aging and the role of CSF-suppression. *Journal of magnetic resonance imaging : JMRI* 20, 216–227. doi:10.1002/jmri.20102

Borovkova, M., Sieryi, O., Lopushenko, I., Kartashkina, N., Pahnke, J., Bykov, A., et al. (2022). Screening of Alzheimer's Disease With Multiwavelength Stokes Polarimetry in a Mouse Model. *IEEE transactions on medical imaging* 41, 977–982. doi:10.1109/TMI.2021.3129700

Forouzannezhad, P., Abbaspour, A., Fang, C., Cabrerizo, M., Loewenstein, D., Duara, R., et al. (2019). A
survey on applications and analysis methods of functional magnetic resonance imaging for Alzheimer's
disease. *Journal of neuroscience methods* 317, 121–140. doi:10.1016/j.jneumeth.2018.12.012

Fujii, T., Yamasaki, Y., Saito, N., Sawada, M., Narita, R., Saito, T., et al. (2019). Polarization characteristics
of dark-field microscopic polarimetric images of human colon tissue. In *Proc.SPIE*. vol. 10890.
doi:10.1117/12.2509000

Ghosh, N. and Vitkin, I. A. (2011). Tissue polarimetry: concepts, challenges, applications, and outlook.
Journal of Biomedical Optics 16, 110801. doi:10.1117/1.3652896

Ghosh, N., Wood, M. F. G., Li, S.-h., Weisel, R. D., Wilson, B. C., Li, R.-K., et al. (2009). Mueller matrix
decomposition for polarized light assessment of biological tissues. *Journal of Biophotonics* 2, 145–56.
doi:10.1002/jbio.200810040

Ghosh, N., Wood, M. F. G., and Vitkin, I. A. (2008). Mueller matrix decomposition for extraction of
individual polarization parameters from complex turbid media exhibiting multiple scattering, optical
activity, and linear birefringence. *Journal of Biomedical Optics* 13, 044036. doi:10.1117/1.2960934

Guerrero-Gonzalez, J. M., Yeske, B., Kirk, G. R., Bell, M. J., Ferrazzano, P. A., and Alexander, A. L.
(2022). Mahalanobis distance tractometry (MaD-Tract) - a framework for personalized white matter
anomaly detection applied to TBI. *NeuroImage* 260, 119475. doi:10.1016/j.neuroimage.2022.119475

Gunning-Dixon, F. M., Brickman, A. M., Cheng, J. C., and Alexopoulos, G. S. (2009). Aging of cerebral
white matter: a review of MRI findings. *International journal of geriatric psychiatry* 24, 109–117.
doi:10.1002/gps.2087

Hutchinson, E. B., Schwerin, S. C., Avram, A. V., Juliano, S. L., and Pierpaoli, C. (2018). Diffusion MRI
and the detection of alterations following traumatic brain injury. *Journal of neuroscience research* 96,
612–625. doi:10.1002/jnr.24065

Hutchinson, E. B., Schwerin, S. C., Radomski, K. L., Sadeghi, N., Jenkins, J., Komlosh, M. E., et al.
(2017). Population based MRI and DTI templates of the adult ferret brain and tools for voxelwise
analysis. *NeuroImage* 152, 575–589. doi:10.1016/j.neuroimage.2017.03.009

Lamptey, R. N. L., Chaulagain, B., Trivedi, R., Gothwal, A., Layek, B., and Singh, J. (2022). A Review of
the Common Neurodegenerative Disorders: Current Therapeutic Approaches and the Potential Role of
Nanotherapeutics. *International journal of molecular sciences* 23. doi:10.3390/ijms23031851

Larsen, L., Griffin, L. D., Grässel, D., Witte, O. W., and Axer, H. (2007). Polarized light imaging of white
matter architecture. *Microscopy research and technique* 70, 851–863. doi:10.1002/jemt.20488

Le Bihan, D. (2014). Diffusion MRI: what water tells us about the brain. *EMBO molecular medicine* 6,
569–573. doi:10.1002/emmm.201404055

Liu, Y., Lu, L., Li, F., and Chen, Y.-C. (2022). Neuropathological Mechanisms of Mild Traumatic Brain
Injury: A Perspective From Multimodal Magnetic Resonance Imaging. *Frontiers in neuroscience* 16,
923662. doi:10.3389/fnins.2022.923662

Lu, S.-Y. and Chipman, R. A. (1996). Interpretation of Mueller matrices based on polar decomposition.
Journal of the Optical Society of America A 13, 1106. doi:10.1364/JOSAA.13.001106

Madden, D. J., Bennett, I. J., and Song, A. W. (2009). Cerebral white matter integrity and cognitive
aging: contributions from diffusion tensor imaging. *Neuropsychology review* 19, 415–435. doi:10.1007/
s11065-009-9113-2

Mehta, S. B., Shribak, M., and Oldenbourg, R. (2013). Polarized light imaging of birefringence and
diattenuation at high resolution and high sensitivity. *Journal of optics (2010)* 15. doi:10.1088/2040-8978/
15/9/094007

Menzel, M., Michielsen, K., De Raedt, H., Reckfort, J., Amunts, K., and Axer, M. (2015). A Jones matrix
formalism for simulating three-dimensional polarized light imaging of brain tissue. *Journal of the Royal*
Society, Interface 12, 20150734. doi:10.1098/rsif.2015.0734

Minchew, H. M., Ferren, S. L., Christian, S. K., Hu, J., Keselman, P., Brooks, W. M., et al. (2022).
Comparing imaging biomarkers of cerebral edema after TBI in young adult male and female rats. *Brain*
research 1789, 147945. doi:10.1016/j.brainres.2022.147945

Morgan, M. L., Brideau, C., Teo, W., Caprariello, A. V., and Stys, P. K. (2021). Label-free assessment
of myelin status using birefringence microscopy. *Journal of neuroscience methods* 360, 109226.
doi:10.1016/j.jneumeth.2021.109226

Nguyen, A. T., Kouri, N., Labuzan, S. A., Przybelski, S. A., Lesnick, T. G., Raghavan, S., et al. (2022).
Neuropathologic scales of cerebrovascular disease associated with diffusion changes on MRI. *Acta*
neuropathologica doi:10.1007/s00401-022-02465-w

Novikov, D. S., Fieremans, E., Jespersen, S. N., and Kiselev, V. G. (2019). Quantifying brain microstructure
with diffusion MRI: Theory and parameter estimation. *NMR in biomedicine* 32, e3998. doi:10.1002/
nbm.3998

Oka, K. and Saito, N. (2006). Snapshot complete imaging polarimeter using Savart plates. In *Proc.SPIE*.
vol. 6295. doi:10.1117/12.683284

Pfefferbaum, A., Adalsteinsson, E., Rohlfing, T., and Sullivan, E. V. (2010). Diffusion tensor imaging
of deep gray matter brain structures: effects of age and iron concentration. *Neurobiology of aging* 31,
482–493. doi:10.1016/j.neurobiolaging.2008.04.013

Qi, J., He, H., Lin, J., Dong, Y., Chen, D., Ma, H., et al. (2018). Assessment of tissue polarimetric properties
using Stokes polarimetric imaging with circularly polarized illumination. *Journal of biophotonics* 11,
e201700139. doi:10.1002/jbio.201700139

Rasband, W. S. (2012). ImageJ. *U. S. National Institutes of Health, Bethesda, Maryland, USA* ,
//imagej.nih.gov/ij/

Rathi, Y., Pasternak, O., Savadjiev, P., Michailovich, O., Bouix, S., Kubicki, M., et al. (2014). Gray matter
alterations in early aging: a diffusion magnetic resonance imaging study. *Human brain mapping* 35,
3841–3856. doi:10.1002/hbm.22441

Rosenkranz, M. A., Dean, D. C. r., Bendlin, B. B., Jarjour, N. N., Esnault, S., Zetterberg, H., et al. (2022).
Neuroimaging and biomarker evidence of neurodegeneration in asthma. *The Journal of allergy and*
clinical immunology 149, 589–598. doi:10.1016/j.jaci.2021.09.010

Saito, N., Odate, S., Otaki, K., Kubota, M., Kitahara, R., and Oka, K. (2013). Wide field snapshot imaging
polarimeter using modified Savart plates. In *Proc.SPIE*. vol. 8873. doi:10.1117/12.2022829

Saito, N., Sato, K., Fujii, T., Durko, H. L., Goldstein, G. L., Phillips, A. H., et al. (2019). Multispectral
Mueller matrix imaging dark-field microscope for biological sample observation. In *Proc.SPIE*. vol.
10890. doi:10.1117/12.2508109

Schucht, P., Lee, H. R., Mezouar, H. M., Hewer, E., Raabe, A., Murek, M., et al. (2020). Visualization
of White Matter Fiber Tracts of Brain Tissue Sections With Wide-Field Imaging Mueller Polarimetry.
IEEE transactions on medical imaging 39, 4376–4382. doi:10.1109/TMI.2020.3018439

Smith, A. M., Mancini, M. C., and Nie, S. (2009). Bioimaging: second window for in vivo imaging. *Nature*
nanotechnology 4, 710–711. doi:10.1038/nnano.2009.326

Tuchin, V. V. (2003). Light-tissue interactions. In *Biomedical Photonics Handbook* (Boca Raton, Florida:
CRC Press). 1–27

Wang, Z., Xin, J., Wang, Z., Yao, Y., Zhao, Y., and Qian, W. (2021). Brain functional network modeling
and analysis based on fMRI: a systematic review. *Cognitive neurodynamics* 15, 389–403. doi:10.1007/
s11571-020-09630-5
Zhao, H., Tsai, C.-C., Zhou, M., Liu, Y., Chen, Y.-L., Huang, F., et al. (2022). Deep learning based
diagnosis of Parkinson's Disease using diffusion magnetic resonance imaging. *Brain imaging and
behavior* 16, 1749–1760. doi:10.1007/s11682-022-00631-y

FIGURE CAPTIONS

Figure 1. (A) Photograph of Ferret brain specimen along with regions of interest including corticospinal tract (CST); optic chiasm (OC) and cerebellum (CB). (B) Schematic diagram of the PLI system where light is emitted from a multi-wavelength light source (LS) which passes through a lens (L) and polarization state generator (PSG) before being directed to the sample through a beamsplitter (BS) and imaging lens. Backscattered light is collected and directed through a polarization state analyzer (PSA) before being focused onto a CCD camera. (C) Photograph of the polarimetry system. (D) Flow chart of data analysis including Mueller Matrix decomposition and calculation of distribution statistics.

Figure 2. (A) Photograph of corticospinal tract (CST), maps of (B) retardance and (C) retardance angle of black circular region of interest shown in (A). Wavelength-resolved (D) retardance and (E) retardance angle distributions for CST compared to background (BKG). (F) Quantification of distribution statistics for retardance shows significance increase for CST compared to background, and (G) relative standard deviation and kurtosis of angle distributions shows significant decrease for CST, showing sensitivity to coherent fibers over unorganized tissue.

Figure 3. Photograph of optic chiasm (OC), maps of (B) retardance and (C) retardance angle of black circular region of interest shown in (A). (D) Circular slice of retardance angle along black line shown in (C), which shows a statistically significant linear relationship with radial angle. Wavelength-resolved (E) retardance and (F) retardance angle distributions for CST compared to background (BKG). (G) Quantification of mean retardance shows no significance difference for both lobes, and (H) the mean retardance angle for both lobes increases with wavelength.

Figure 4. (A) Photograph of cerebellum, along with (B) diagram showing expected physiological structure of the grey matter. Maps of (C) retardance and (D) retardance angle of black circular region of interest shown in (A). Wavelength-resolved (E) retardance and (F) retardance angle distributions for ROI with minimal curvature. (G) Quantification of distributions statistics for retardance shows a decrease of the mean with increasing wavelength, and (H) retardance angle shows an increase of the mean with increasing wavelength, suggesting increased depth penetration of longer wavelengths may be more sensitive to parallel fiber structures.