Article title: Safety margin evaluations for transtympanic laser stimulation of the cochlea

from the outer ear in Mongolian gerbils (*Meriones unguiculatus*)

List of author names: Aya Okamoto[1], Miku Uenaka[1], Yuki Ito[1], Tomohiro Miyasaka[2],

Koji Toda[3], Shizuko Hiryu[1], Kohta I. Kobayasi[1*], Yuta Tamai[1,*]

Author affiliations:

[1]Neuroethology and Bioengineering, Graduate School of Life and Medical Sciences,

Doshisha University, Kyotanabe, Kyoto, Japan

[2]Neuropathology, Graduate School of Life and Medical Sciences, Doshisha University,

Kyotanabe, Kyoto, Japan

[3]Department of Psychology, Keio University, Tokyo, Japan

Corresponding author: Kohta I. Kobayasi and Yuta Tamai

(E-mail: kkobayas@mail.doshisha.ac.jp; ytamai@mail.doshisha.ac.jp)

Current address (Yuta Tamai): Neurobiology of Social Communication, Department of

Otolaryngology-Head and Neck Surgery, Hearing Research Centre, University of

Tübingen, Medical Center, Elfriede-Aulhorn-Strasse 5, 72076 Tübingen, Germany

Keywords

Auditory prosthesis, Infrared laser stimulation, Auditory perception, Head-fixed classical

conditioning

Abstract

Infrared laser stimulation of the cochlea has been proposed as a possible alternative

to conventional auditory prostheses. The safety margin of laser stimulation must be

determined for the practical application of the laser to a prosthesis. This study assessed

the effect of laser-induced damage to the cochlea on auditory perception. To examine the

effects of laser-induced injury of the auditory system on the perception of auditory stimuli,

we trained Mongolian gerbils (*Meriones unguiculatus*) using a head-fixed classical

conditioning task. White noise of 80 dB sound pressure level (SPL) was presented as a

conditioned stimulus for a reward (a drop of water), and licking behavior was recorded

as a conditioned response. After training, an optic fiber was inserted into the ear canal.

Each subject was exposed to continuous pulsed laser exposure of 1.6, 3.3, 6.6, 26.4, 52.8,

or 105.6 W/cm^2 for 15 h. White noise of various intensities was presented without pairing

it with water before and after laser exposure, to assess the decrease in auditory perception

due to laser-induced injury. The licking rate did not change after laser exposure of 6.6

W/cm^2 or weaker but drastically decreased after 26.4 W/cm^2 or higher. These results

suggest that the injury threshold in Mongolian gerbils for laser stimulation is between 6.6

and 26.4 W/cm^2; laser exposure over 6.6 W/cm^2 could be out of the safety margin. These

findings will contribute to the appropriate delimitation of safe and effective laser

stimulation parameters in future research.

Introduction

Hearing impairment is a prevalent sensory disorder worldwide. Approximately 466 million people, 5% of the world's population, suffer from hearing deficits, resulting in severe communication difficulties and social isolation (World Health Organization, 2018, 2019, 2021). Hearing aids are among the most widely used therapeutic options to compensate for these problems. These devices improve speech comprehension and social interaction in hard-of-hearing individuals (Arlinger, 2003; Mulrow et al., 1990; Mulrow, Tuley, & Aguilar, 1992). However, these patients require an intact middle ear to ensure sufficient and proper hearing. Hard-of-hearing individuals with middle ear dysfunction must use an implantable auditory prosthesis (e.g., bone-anchored hearing devices, middle ear implants, or cochlear implants) that bypasses the function of the middle ear to stimulate the peripheral auditory system. One of the most significant drawbacks preventing the widespread use of these devices is surgical intervention, which increases the risk of complications post-surgery (Ahmed, Saqulain, Khan, & Kausar, 2021; Håkansson, Eeg-Olofsson, Reinfeldt, Stenfelt, & Granström, 2008; Yeung, Griffin, Newton, Kenna, & Licameli, 2018).

Infrared laser stimulation has been discussed as a possible alternative to conventional auditory devices (Littlefield & Richter, 2021). The advantage of laser

stimulation is that it can activate spatially focused cell populations in a non-contact

manner without the need to deliver any exogenous agent into tissues, such as a viral vector

(Fekete, Horváth, & Zátonyi, 2020). Early research by Izzo *et al.* (2006) suggested that

pulsed laser stimulation of the auditory nerves induced a cochlear response. Several later

studies that focused on the detailed relationship between the irradiation site of the cochlea

and laser-evoked neural response indicated that laser stimulation could elicit spatially

specific neural activities (Izzo, Suh, et al., 2007; Matic, Walsh, & Richter, 2011; Moreno

et al., 2011; Richter et al., 2011). That is, precisely manipulating the laser stimulation

system can improve the spectral resolution of cochlear implants. Our previous studies

leveraged the benefit of the contactless feature of laser stimulation to reduce the

invasiveness of surgical implantation of the auditory prosthesis. They revealed that

transtympanic laser stimulation from the outer ear could evoke a cochlear response,

bypassing the middle ear function (Tamai, Horinouchi, Hiryu, & Kobayasi, 2019; Uenaka,

Nagamura, Hiryu, Kobayasi, & Tamai, 2022). The auditory cortical activity elicited by

transtympanic laser stimulation depends on stimulus intensity (Tamai et al., 2020). These

results imply that a hearing prosthesis with infrared laser stimulation could be a

noninvasive alternative to the conventional implantable auditory prosthesis.

For the clinical application of transtympanic laser stimulation to auditory

prostheses, the safety margin of laser irradiation should be assessed. Because the laser-

evoked neural response is probably related to the thermal elevation of the tissue (Albert

et al., 2012; Shapiro, Homma, Villarreal, Richter, & Bezanilla, 2012; Wells et al., 2007),

thermal build-up might be the main reason for laser-induced damage (Brown et al., 2020;

Chernov, Chen, & Roe, 2014; Thompson, Wade, Pawsey, & Stoddart, 2013). Therefore,

thermal damage to the peripheral auditory system can occur when a laser with high radiant

energy is continuously presented. Some studies have investigated laser irradiation trauma

in the cochlea (Izzo et al., 2006; Matic et al., 2013; Rajguru et al., 2010; Tamai et al.,

2019). Goyal *et al.* (2012) compared the relationship between radiant energy in

continuous laser exposure and the reduction of the laser-evoked electrophysiological

response in the cochlea, reporting that the injury threshold for laser stimulation at 250 Hz

was between 25 and 30 µJ/pulse. However, no study has described the influence of laser-

induced thermal elevation on cochlear tissues or the effect of thermal physiological

damage on auditory perception.

A comprehensive evaluation of laser irradiation trauma elucidates the effects of

laser-induced damage on auditory perception. It provides a precise picture of the effect of

laser stimulation on the auditory peripheral system. It assessed laser-induced damage

characteristics of the peripheral auditory system by comparing electrophysiological,

histological, and behavioral alterations following continuous laser exposure. Here, we

performed (1) histological examination of spiral ganglions with hematoxylin and eosin

(HE) staining, (2) electrophysiological recording of the laser-evoked cochlear response

during laser irradiation, and (3) head-fixed classical conditioning with auditory stimulus

(see Figure 1 for details).

Results

Before the histological examination, the surface temperature of the cochlea was

recorded to investigate the thermal elevation following continuous laser irradiation

(Figure 2). The surface temperatures did not increase from 24 °C (i.e., room temperature)

using the 6.6 W/cm^2 laser irradiation or less; the laser 26.4 W/cm^2 or more increased the

surface temperatures immediately within 1 min into laser irradiation. The temperature

rise reached a plateau approximately 5 min after the laser onset (Figure 2A). Average

surface temperature at 10 min after the laser irradiation onset was described in Figure 2B

and showed that the increasing temperature induced by 26.4, 52.8, and 105.6 W/cm^2

lasers were 31.6, 36.6, and 44.7 °C, respectively. The ANOVA revealed that the effect of

radiant energy was statistically significant ($F_{(5,17)}=21.26$, $P<0.001$, one-way ANOVA).

Figure 2C shows the lateral and medial surface temperatures induced by laser irradiation

at 105.6 W/cm^2 on the lateral side of the cochlea. The average temperatures of the lateral

and medial sides resulted in approximately 41.2 and 32.5 ℃, respectively, and the

temperature difference between the left and the right sides was statistically significant

(P<0.01, paired t-test). The medial side temperature was approximately in line with that

of the lateral side irradiated by the laser at 26.4–52.8 W/cm^2.

Horizontal sections of the cochlea were subjected to HE staining to evaluate the

effect of laser exposure on the spiral ganglion neuron density (Figure 3A). Lower spiral

ganglion neuron density was found in the cochlea continuously irradiated by the laser of

105.6 W/cm^2 than in less than 6.6 W/cm^2 (Figure 3B). The middle turn of the cochlea

irradiated by the laser at 105.6 W/cm^2 is depicted in Figure 3C. A comparison of the

lateral and medial sides showed a remarkable decrease in the cell density on the lateral

side, mainly irradiated by the transtympanic laser, compared to the medial side. On the

lateral side of the cochlea, the spiral ganglion density difference between the left and right

ears significantly decreased (r=-0.87, P<0.05, Pearson's correlation analysis) as radiant

energy intensified, whereas on the medial side, it did not (r=-0.45, P=0.23, Pearson's

correlation analysis) (Figure 3D). The cell density decrement on the medial side after

laser exposure of 105.6 W/cm^2 approximately coincides with that on the lateral side

irradiated by the laser with 26.4–52.8 W/cm^2. Figure 3E shows the spiral ganglion neuron

density difference on the lateral side of the apical-middle and basal turn of the cochlea.

The apical-middle turn of the cell density in the left ear was significantly lower than that

in the right ear (P<0.05, one-sample t-test). In contrast, the basal turn of the cell density

was less affected than the apical-middle side (P=0.19, one-sample t-test) (Figure 3E).

Figure 4 shows the cochlear response after continuous laser exposure for 15 h. The

manifest damage in cochlear response was not observed by laser exposure up to 6.6

W/cm^2, while the cochlear response elicited by laser exposure of 105.6 W/cm^2 decreased

(Figure 4A). The time course of the relative response amplitude showed that continuous

laser exposure at 105.6 W/cm^2 caused a drastic decrease in the laser-evoked response

amplitude within ca. 65 min (Figure 4B). In comparison, laser irradiation of 6.6 W/cm^2

or less did not. Figure 4C shows the relationship between the laser irradiation time and

percentage of response amplitude. Laser irradiation over 26.4 W/cm^2 drastically increased

the average amplitude decrement, while laser stimulation with up to 26.4 W/cm^2 did not

(Figure 4D).

Figure 5A shows the licking behavior evoked by white noise (80 dB SPL) before

and after 1.6, 6.6, and 105.6 W/cm^2 laser exposure. The laser exposure at 1.6 and 6.6

W/cm^2 did not cause an observable behavioral response change, while an apparent

decrement of the licking rate was observed after 105.6 W/cm^2 laser exposure. The

behavioral changes after laser exposure are depicted in Figure 5B. Laser irradiation of the

cochlea over 6.6 W/cm^2 drastically changed the behavioral level (Figure 5B). It caused

the delta behavioral level (DBL) to reach out of the safety margin, which was calculated

using three standard deviations of DBL in the control condition (Figure 5C).

Discussion

This study demonstrated that the laser stimulation with a higher average power

density triggered the thermal elevation of the cochlear surface (Figure 2). Because the

conventional mechanism of the laser-evoked cochlear response is the optothermal effect,

which causes an excessive temperature increase in the cells (Albert et al., 2012; Shapiro

et al., 2012; Wells et al., 2007), several studies indicated that laser-induced thermal

elevation and possible thermal-damage on the cochlea might constitute a significant

limitation in applying the laser to the auditory prosthesis (Goyal et al., 2012; Izzo, Walsh

Jr., et al., 2007; Matic et al., 2013; Rajguru et al., 2010; Thompson et al., 2013). Our study

demonstrated that cochlear surface temperature significantly increased as radiant energy

increased (Figure 2AB). The surface temperature continued to rise for approximately 5

min after laser onset, and then the elevated temperature reached a plateau (Figure 2A).

These results suggest that continuous laser irradiation (minute order) elevates cochlear

temperature compared with short-term irradiation. The continuous laser irradiation at

105.6 W/cm^2 caused an average surface temperature up to 44.7 °C, potentially causing

thermal tissue damage (Dewey, 1994). These results demonstrate that continuous laser

stimulation bears the risk of thermal damage to the cochlear tissues when a laser with

high radiant energy is used.

A comparison of the thermal elevation of the cochlear surface with histological

damage of spiral ganglion neurons revealed that the thermal distribution of the cochlear

surface affected the laser irradiation trauma of the spiral ganglion neurons. Our previous

study showed that transtympanic lasers mainly irradiated the middle turn of the lateral

side of the cochlea (Uenaka et al., 2022). This study implies that the highest peak of the

thermal distribution of the cochlear surface might be observed in the primary irradiated

area, and the temperature could decrease from the lateral to the medial side. It also showed

that the thermal elevation of the lateral side caused by laser irradiation of 105.6 W/cm^2

was significantly higher than that on the medial side. The 105.6 W/cm^2 laser-induced

temperature rise of the medial side reached 32.5 °C, which agreed with the lateral side

temperature caused by the laser at 26.4–52.8 W/cm^2 (Figure 2C). These results suggest

that approximately half of the radiant energy of laser irradiation reached the medial side

of the cochlea and agreed with the laser irradiation site in our previous study (Uenaka et

al., 2022). In addition, a more significant decrease in cell density was observed on the

lateral side of the apical-middle turn than on the basal turn (Figure 3E). These results

coincide with the radiant energy distribution (i.e., thermal distribution) of the cochlear

surface and suggest that the spatial injury profile of ganglion neurons depends on the

thermal distribution of the cochlear surface.

The electrophysiological cochlear response was recorded to evaluate the acute

damage threshold of continuous laser stimulation. This response is widely used as a

sensitive probe for physiological cochlear function (Prado-Guitierrez, Fewster, Heasman,

McKay, & Shepherd, 2006; Ramekers, Versnel, Strahl, Klis, & Grolman, 2015; van Eijl,

Buitenhuis, Stegeman, Klis, & Grolman, 2017). Previous studies, including ours (Izzo et

al., 2006; Rajguru et al., 2010; Tamai et al., 2019), have assessed laser-irradiation trauma

of the cochlea by recording the cochlear response induced by continuous laser stimulation.

These studies suggest that the laser safely evokes cochlear responses without causing

severe irradiation damage. However, Goyal *et al*. (2012) reported that continuous laser

irradiation with higher radiant energy induced thermal damage to the cochlear function.

This indicates that the injury threshold of laser exposure is between 25 and 30 µJ/pulse at

250 pulses/s (i.e., approximately 47.1 and 56.5 W/cm^2). Our results showed that the

cochlea kept responding to laser stimulation of 26.4 W/cm^2 or less, while the laser at 52.8

W/cm^2 or above remarkably decreased the cochlear response (Figure 4). This result

suggests that the electrophysiological damage threshold of the laser stimulation is

between 26.4 and 52.8 W/cm^2, and the injury threshold is in line with the previous study

by Goyal *et al*. (2012).

Measuring the impact of thermal-induced physiological deterioration in the cochlea

on auditory perception is crucial. Behavioral assessment of hearing loss generally

provides compelling evidence of auditory perceptual consequences following

compromised auditory processing. Therefore, evaluating auditory perception following

thermal injury in the cochlea is an essential step toward the clinical application of laser

stimulation to auditory prostheses. This study showed that the auditory-evoked licking

behavior did not change after laser exposure of 6.6 W/cm^2 or weaker, but drastically

decreased after 26.4 W/cm^2 or above. These results suggest that the thermal injury

threshold for transtympanic laser stimulation is between 6.6 and 26.4 W/cm^2. Therefore,

transtympanic laser stimulation over 6.6 W/cm^2 could be out of the safety margin.

A comparison of the behavioral and electrophysiological results revealed that the

thermal damage threshold for the auditory-evoked behavior alteration was slightly lower

than that for the laser-evoked cochlear response. Here, the laser irradiates the tympanic

membrane, auditory ossicles, and cochlear tissues from the outer ear to induce the

cochlear response (Uenaka et al., 2022); laser stimulation bears the risk of lesioning more

peripheral auditory function from the origin of the laser-evoked response. Our previous

study revealed that laser irradiation (λ=1871 nm) from the outer ear bypasses the

tympanic membrane and auditory ossicles because these tissues do not have higher

absorbance to the infrared laser (Uenaka et al., 2022). However, some studies investigated

the effect of the laser irradiation of the tympanic membrane and auditory ossicles in *vivo*

(Sorg et al., 2019) and in *vitro* (Pillong et al., 2020). This indicates that the laser irradiation

of 27.9 W/cm^2 or above at 532 nm could cause thermal damage. Although these results

are not directly comparable to our data because of the difference in wavelength-dependent

absorbance in the tympanic membrane and auditory ossicles, as some studies described

(Carr, Valdez, Bruns, & Bawendi, 2016; Goblet, Matin, Lenarz, & Paasche, 2021;

Heimann et al., 2021), thermal damage to the tympanic membrane and auditory ossicles

might be one reason why continuous laser exposure affects the auditory-evoked response

more than the laser-evoked response.

In addition to the tympanic membrane and auditory ossicles, hair cells in the

cochlea can be damaged by laser irradiation. The physiological mechanism of the laser-

evoked cochlear response is still debated whether the response is hair-cell-mediated

(Fekete et al., 2020). Hence, whether the laser-evoked response here characterized the

state of cochlear function is unclear, including that of hair cells. To reveal the

physiological mechanism of laser stimulation in the cochlea, some studies (Baumhoff,

Kallweit, & Kral, 2019; Schultz et al., 2012; Thompson et al., 2015) compared laser-

evoked auditory responses in normal hearing and chemically compromised cochlea and

reported that laser-evoked auditory responses were only observed in the normal hearing

cochlea. These studies suggest that the cochlear response elicited by laser stimulation is

mediated by hair cells.

In contrast, Richter *et al*. (2008) revealed that 30–40 dB chemical deafening of the

auditory threshold did not affect the threshold of the laser-evoked cochlear response. Tan

et al. (2018) demonstrated that laser stimulation could elicit auditory neural activity in

congenitally deaf mice in which the cochlea could not transmit synapses between the

inner hair cells and spiral ganglion neurons.

In addition, when animals have an impaired hearing threshold, auditory masking

does not affect laser-induced responses (Young, Tan, Xia, & Richter, 2015). These results

suggest that laser irradiation of the cochlea directly activates the auditory nerves,

bypassing hair cells. If the laser activates the cochlear nerves without needing hair cell

depolarization, the laser-evoked response does not characterize the state of the hair cells.

Therefore, our results show that the auditory-evoked response might be more vulnerable

to thermal injury than the laser-evoked response (Figures 4 and 5). The effect of thermal

damage to hair cells on laser-evoked cochlear responses must be considered in the future.

In summary, we assessed the laser-induced damage characteristics of the peripheral

auditory system through histological, electrophysiological, and behavioral tests.

Histological evaluation of the cochlea after continuous laser exposure revealed that the

laser-induced thermal distribution of the cochlear surface coincided with the spatial injury

profile of spiral ganglion neurons. The measurement of auditory-evoked behavioral and

laser-evoked electrophysiological responses indicate that cochlear response during the

laser exposure of 26.4 W/cm^2 or weaker did not decrease. In contrast, auditory-induced

behavioral response drastically decreased after 26.4 W/cm^2 or higher. This discrepancy

in the injury threshold suggests that laser stimulation bears the risk of lesioning more

peripheral auditory functions from the origin of the laser-evoked response. Evaluating

perceptual consequences indicated that the thermal injury threshold for transtympanic

laser stimulation in Mongolian gerbils is between 6.6 and 26.4 W/cm^2. Therefore,

transtympanic laser under 6.6 W/cm^2 could be available without concern for acute injury.

A comprehensive safety margin assessment is an essential step toward the practical

application of transtympanic laser stimulation in auditory prostheses.

Acknowledgement

The authors would like to thank Hiroshi Riquimaroux for his support during the

early stages of this work. We also thank Kazuki Tanaka, Mizuki Katayama, and Keito

Hosokawa for technical support and Editage (www.editage.com) for English language

editing. This research was financially supported by the Sasakawa Scientific Research

Grant No. 2022–6027 (A.O.), the Japan Society for the Promotion of Science (JSPS)

KAKENHI Grants Nos. 22H02946 (T.M.), 21H03469 (K.I.K.), 21K21322 (Y.T.), the

JSPS Overseas Research Fellowship (Y.T.), and Keio Academic Development Fund

(K.T.).

Conflict of interests

The authors have no conflicts of interest to declare.

Materials and methods

Animals

All experimental procedures were performed according to the guidelines

established by the Ethics Review Committee of Doshisha University. The experimental

protocols were approved by the Animal Experimental Committee of Doshisha University.

A total of 33 experimentally naive Mongolian gerbils (*Meriones unguiculatus*) aged 3–

18 months were used here. All gerbils were bred and reared in a laboratory. Each animal

was housed with 2–5 gerbils in a 20 (W) × 40 (L) × 17 cm (H) cage, with free access to

food and water. The animal room had a 12-h light-dark schedule, a temperature of 22–

23 °C, and relative humidity of approximately 50%.

Animal surgery

Mongolian gerbils were deeply anesthetized with ketamine (47 mg/kg i.m.) and

xylazine (9.3 mg/kg i.m.). Maintenance doses of ketamine (17.5 mg/kg) and xylazine (7.0

mg/kg) were injected every 30–50 min or if the animals showed signs of increasing

arousal (e.g., voluntary whisker movement). The body temperature of the gerbils was

maintained using heating pads placed beneath the animals. The left side of the bulla was

exposed by incision of the muscle and skin from the shoulder to the jaw, and a hole

(diameter: 1 mm) was made on the bulla. A silver electrode (Nilaco, Tokyo, Japan;

diameter: 0.13 mm; impedance < 20 kΩ) was inserted into the hole and hooked onto the

bony rim of the round window to record cochlear responses. The electrode was stabilized

with a bulla using acrylic glue and dental cement (Provinice; Shofu, Kyoto, Japan). The

left pinna was removed to provide a clear view of the tympanic membrane. A liquid of

acrylic glue and dental cement (Provinice; Shofu, Kyoto, Japan) was mounted into the

right ear for approximately 50 dB SPL attenuation of the sound pressure reaching the right

ear.

The parietal and temporal sides of the fur and the skin of the subject's head were

shaved. The temporal muscle was removed after applying a local anesthetic (xylocaine

gel; Aspen Japan, Tokyo, Japan). The skull was carefully cleaned with 0.3% sodium

hypochlorite solution and saline. A small hole was made in the dorsal skull 2 mm anterior

to the bregma and 1 mm lateral to the midline, and a reference electrode was inserted into

the hole. A V-shaped metal plate (F-911; Hilogik, Osaka, Japan) was attached to the skull

using acrylic dental resin (Super bond C&B; Sun Medical, Shiga, Japan) and dental

cement (Provinice; Shofu, Kyoto, Japan), allowing the subjects to be head-fixed during

the experiment. All subjects were singly housed during recovery for at least two days

before training began.

Auditory stimuli

White noise (700–44,800 Hz) was used as an auditory stimulus, with a duration of

500 ms. The auditory stimulus was presented via a doom tweeter (FT28D; Fostex, Tokyo,

Japan) from 15 cm above the subject's head. The stimulus was generated using a digital-

to-analog converter (Octa-Capture; Roland, Shizuoka, Japan) with a sampling rate of 192

kHz. The signal was intensified using an amplifier (A-10; Pioneer, Tokyo, Japan). The

sound pressure level was calibrated to 20, 35, 50, 65, 80, and 95 dB SPL using a

microphone (Type 7016; Aco, Tokyo, Japan).

Apparatus and procedure in the behavioral experiment

Behavioral experiments were conducted in a soundproof box 60 (W) × 60 (L) × 60

cm (H). Licking recordings in head-fixed animals were performed as previously described

(Guo et al., 2014; Kaneko et al., 2022; Toda et al., 2017; Yamamoto, Yamada, Yatagai,

Ujihara, & Toda, 2022). Each subject was positioned on a covered elevated platform

(custom-designed and 3D printed). A drinking spout was placed 1 mm in front of the

mouth of the animal, and the clamp and drinking spout positions were adjusted to detect

licking behavior easily. A touch sensor (DCTS-10; Sankei Kizai, Tokyo, Japan),

connected between the drinking spout and the metal sheet of the stage, recorded the timing

of licks at 2000 samples/s. The recorded licking timing was stored on a computer via

Arduino Uno (Arduino, Ivrea, Italy) controlled by Python 3.7 on Spyder 501. All

behaviors were recorded using a video camera (HD-5000; Microsoft, Washington, USA).

After at least 2 days of recovery from surgery, the gerbils were water-restricted

in their home cage. The training session typically comprised 400 trials daily. Head-fixed

participants were allowed to voluntarily lick the spout at all times. A 3.2 µl drop of water

(unconditioned stimulus: US) was delivered through the tube every 7–13 s after

presenting an 85 dB SPL of white noise (conditioned stimulus: CS) controlled by

MATLAB (MathWorks, Massachusetts, USA) and a microcontroller (Arduino Uno;

Arduino, Ivrea, Italy) with a custom-made relay circuit.

After animals showed high licking rates following the auditory cue, catch trials

were conducted to evaluate the subject's perception. The test session comprised rewarded

and non-rewarded trials (i.e., catch trials). Catch trials were included to assess the subjects’

perceptions without reward feedback. White noise of 20, 35, 50, 65, 80, or 95 dB SPL

was presented in the catch trials. During a test session, 280 rewarded trials and 120 catch

trials were conducted in a pseudo-random order. In the catch trials, each sound intensity

was tested 20 times. The test session was performed pre- and post-laser exposure.

 Recording and analysis of the behavioral data

363   An averaged histogram of licking behavior was used to evaluate the time course of

364 behavioral responses. The histograms were calculated in a 100 ms bin size and

365 smoothened using a 300-ms moving average window. Peak licking rates, durations, peak

366 periods, and latencies were extracted from the time course of licking behavior to assess

367 the behavioral response. Peak licking rates and periods were defined as the licking rate

368 and time when the licking rate reached the maximum value within 4 s after stimulus onset,

369 respectively. The baseline licking rate was defined as the pre-stimulus licking rate

370 obtained for 1 s, and duration was quantified as the temporal length of licking behavior

371 >50% of the licking rate from the mean baseline licking rate to the peak licking rate.

372 Latency was determined by measuring the time between stimulus onset and the behavioral

373 response reaching the mean plus three standard deviations of the baseline licking rate.

374 The delta licking rate was defined as the difference between the mean baseline licking

375 rate and the post-stimulus licking rate to evaluate the response amplitude. The analysis

376 range of post-stimulus licking behavior was 0.2–1.2 s because the behavioral response

377 was mostly observed 0.2 s after stimulus onset (see Supplemental Figure 1 for details).

378 The behavioral level was calculated by standardizing the highest delta licking behavior

379 in the pre-exposure test. The delta behavioral level (DBL) between the pre- and post-

exposure test was defined as the summation of the difference in the behavioral level for

all sound pressure levels to evaluate the effect of continuous laser exposure on auditory

perception.

Laser exposure

A repetitive pulsed infrared laser was used for laser exposure. The individual pulse

of the laser had a 100 μs rectangular shape in intensity. It was generated using a diode

laser stimulation system (BWF-OEM-1850; B&W TEK, Delaware, USA) with a

wavelength (λ=1871) comparable to those previously reported (Baumhoff et al., 2019;

Izzo et al., 2006; Izzo, Walsh Jr., et al., 2007). The voltage commands for the laser stimuli

were generated using a function generator (WF1974; NF, Kanagawa, Japan). The

repetition frequency of the pulsed laser was changed to 125, 250, 500, 2000, 4000, and

8000 Hz to adjust the amount of radiant energy reaching the cochlea to 1.6, 3.3, 6,6, 26.4,

52.8, and 105.6 W/cm^2, respectively. The radiant energy and beam width at the cochlea

(i.e., 2.2 mm from the fiber tip) were measured using a digital power meter (PM100D;

Thorlabs, Tokyo, Japan) with a thermal power sensor (S302C; Thorlabs, Tokyo, Japan)

and an InGaAs biased detector (DET10D/M; Thorlabs, Tokyo, Japan), respectively. Our

previous study measured the laser-beam profile (Uenaka et al., 2022).

An optic fiber (diameter: 100 μm; NA, 0.22) was inserted mediolaterally into the

left outer ear canal with the tip angled rostrally by 10° and dorsally by 5° using a

micromanipulator (MM-3; Narishige, Tokyo, Japan) in awake head-fixated subjects. The

optic fiber tip was set at c.a. 0.7 mm in front of the tympanic membrane to ensure that the

cochlea was irradiated transtympanically from the outer ear as described in our previous

reports (Tamai et al., 2019; Tamai et al., 2020; Uenaka et al., 2022).

Each subject's cochlea was continuously stimulated by laser irradiation from the

outer ear for 15 h. Continuous laser exposure was performed on the same covered elevated

platform in the soundproof box as in the behavioral experiment. The temperature in the

soundproof box was adjusted to approximately 26 °C, and a small pee pad was placed

beneath the subject to prevent the subject's temperature from decreasing. After laser

exposure and behavioral experiments, the subjects were overdosed with an intraperitoneal

injection of pentobarbital (200 mg/kg). Laser-induced cochlear response was

continuously recorded throughout the study period. The cochleae were histologically

investigated to clarify the thermal damage caused by laser irradiation.

Recording and analysis of the cochlear response during laser exposure

After the pre-exposure test, cochlear response was recorded in awake subjects

during continuous laser exposure. The cochlear response was amplified 1000 times using

a bio-amplifier (MEG-1200; Nihon Kohden, Tokyo, Japan) and stored on a computer via

an analog-to-digital converter (Octa-capture; Roland, Shizuoka, Japan) at a sampling rate

of 96,000 kHz.

The recorded signal was processed using a band-pass filter (100–10,000 Hz) and

separated at 10 ms intervals. The cochlear signal was averaged 1000 times every 10 s for

15 h, and the response amplitude of the averaged cochlear signal was evaluated using the

amplitude spectrum distribution. To obtain high-frequency resolution, 32,768 points of

zero paddings were performed on the head and end of the averaged signal. The amplitude

spectrum was extracted using fast Fourier transforms (FFT, 16,384 points with a Hanning

window); therefore, the frequency resolution was 5.9 Hz. The peak of the amplitude

spectrum at the stimulus repetition frequency was extracted as the response amplitude.

The mean response amplitude was calculated at 5 min intervals standardized by the first

5 min mean amplitude.

Preparation of histological evaluation

Following the post-exposure test session, the laser-irradiated subjects were

euthanized by an overdose of pentobarbital (200 mg/kg i.p.). After postmortem perfusion

with phosphate buffered saline (PBS) and 4% paraformaldehyde in PBS, the animals were

decapitated and fixed in the same fixative for two days in a refrigerator at 4 °C.

Subsequently, the tissues were stored in cooled PBS containing 0.02% sodium azide.

The paraformaldehyde-fixed tissues were decalcified by immersing in a mixture

solution of formic acid (0.28 g/ml), citric acid (0.1 g/ml), and sodium hydroxide (0.02

g/ml) for 7–10 days. The decalcifying solution was exchanged every 2–3 days. After the

decalcification, the samples were preserved in PBS with 0.02% sodium azide.

Decalcified head tissues were dehydrated using a gradually ascending

concentration of ethanol (70–100%) for 2 days followed by immersion in 100% xylene

and embedding into a paraffin block. Horizontally, paraffin-embedded sections with the

thickness of 6 μm was processed using a microtome (HM 430; Thermo, Tokyo, Japan).

Cochlear sections between 7 and 8 mm ventral to the sagittal midline from each

subject were examined. Sections were deparaffinized by immersion in 100% xylene.

After the deparaffinization, sections were hydrated with 100% ethanol and distilled water.

Sections were stained with Mayer's hematoxylin solution and rinsed with distilled water.

After differentiation with 70% ethanol containing 0.1% HCL, the sections were stained

with Putt's eosin solution. HE-stained sections were re-immersed in 100% ethanol and

cleared in xylene. All stained sections were mounted in mounting medium (HSR solution;

Sysmex, Hyogo, Japan). Specimens were photographed using a digital camera (BX53;

Olympus, Tokyo, Japan) under a light microscope (EX-F1; Casio, Tokyo, Japan).

Histological analysis was performed using ImageJ software (Abràmoff, Magalhães, &

Ram, 2004; Schneider, Rasband, & Eliceiri, 2012).

Behavioral data analysis for a safety margin evaluation

Each subject was placed on a covered elevated platform for 15 h in a temperature-

controlled soundproof box, similar to the laser exposure experiment without presenting

laser irradiation. We compared the performance of the classical conditioning task before

and after placement on the covered elevated platform. The mean and standard deviation

of the control condition were measured to determine the safety margin of continuous laser

irradiation. The safety margin was defined as the mean $\pm$ three standard deviations of

the DBL in the control condition (n=9).

Correlation between physiological damage in the cochlea and behavioral alternation

To determine whether these histological and functional deficits in the cochlea were

associated with behavioral alterations, we correlated the change in spiral ganglion neuron

density and amplitude of the cochlear response with the behavioral response. Further

details are provided in Supplemental Figure 2.

Recording thermal elevation of the cochlear surface

Eleven gerbils were euthanized (pentobarbital; 200 mg/kg) and decapitated for

cochlear surface thermal measurements. Both sides of the cochlea were extracted from

the skulls under a stereomicroscope. The cochlea was placed in a Petri dish and stabilized

with dental wax. The laser fiber tip was directed vertically to the surface of the cochlea

and was set to 2.2 mm from the cochlear surface. Continuous laser irradiation with 1.6,

3.3, 6.6, 26.4, 52.8, and 105.6 W/cm^2 was performed for 1 h, and the surface temperature

elevations were measured using a thermal camera (FLIR C3-X; Teledyne FLIR LLC,

Oregon, USA), which was positioned at 21 cm from the cochlear surface with vertically

45° angle.

To acquire the thermal elevation on the lateral and medial sides of the cochlear

surface, a thermocouple (MSM-3905; Biomedica, Osaka, Japan) was used. The tip of the

thermocouple sensor was in contact with the second turn of the cochlear surface on one

side for 1 h. After recording the temperature on one side, the specimen was left for at least

15 min to cool down to room temperature (~24 °C), and then the temperature on another

side was measured.

Flow of the behavioral, electrophysiological, and histological procedure

Surgery, behavior training, pre-exposure test, laser exposure for 15 h, post-exposure

test, and histological evaluation were performed in this order. The recovery period was

performed after surgery before and after laser exposure. Further details are provided in

Figure 1.

References

References

Abràmoff, M. D., Magalhães, P. J., & Ram, S. J. (2004). Image processing with ImageJ.
Biophotonics int., 11(7), 36-42.

Ahmed, J., Saqulain, G., Khan, M. I. J., & Kausar, M. (2021). Complications of cochlear
implant surgery: A public implant centre experience. *Pak. J. Med. Sci., 37*(5),
1519-1523. doi:10.12669/pjms.37.5.3960

Albert, E. S., Bec, J. M., Desmadryl, G., Chekroud, K., Travo, C., Gaboyard, S., . . .
Chabbert, C. (2012). TRPV4 channels mediate the infrared laser-evoked response
in sensory neurons. *J. Neurophysiol., 107*(12), 3227-3234.
doi:10.1152/jn.00424.2011

Arlinger, S. (2003). Negative consequences of uncorrected hearing loss-a review. *Int. J.
Audiol., 42 Suppl 2*, 2s17-20.

Baumhoff, P., Kallweit, N., & Kral, A. (2019). Intracochlear near infrared stimulation:
Feasibility of optoacoustic stimulation in vivo. *Hear. Res., 371*, 40-52.
doi:10.1016/j.heares.2018.11.003

Brown, W. G. A., Needham, K., Begeng, J. M., Thompson, A. C., Nayagam, B. A.,
Kameneva, T., & Stoddart, P. R. (2020). Thermal damage threshold of neurons
during infrared stimulation. *Biomed. Opt. Express, 11*(4), 2224-2234.
doi:10.1364/boe.383165

Carr, J. A., Valdez, T. A., Bruns, O. T., & Bawendi, M. G. (2016). Using the shortwave
infrared to image middle ear pathologies. *Proc. Natl. Acad. Sci. U S A, 113*(36),
9989-9994. doi:10.1073/pnas.1610529113

Chernov, M. M., Chen, G., & Roe, A. W. (2014). Histological assessment of thermal
damage in the brain following infrared neural stimulation. *Brain. Stimul., 7*(3),
476-482. doi:10.1016/j.brs.2014.01.006

Dewey, W. C. (1994). Arrhenius relationships from the molecule and cell to the clinic. *Int.
J. Hyperthermia, 10*(4), 457-483. doi:10.3109/02656739409009351

Fekete, Z., Horváth, Á. C., & Zátonyi, A. (2020). Infrared neuromodulation: A
neuroengineering perspective. *J. Neural. Eng., 17*(5), 051003. doi:10.1088/1741-
2552/abb3b2

Goblet, M., Matin, F., Lenarz, T., & Paasche, G. (2021). Optical absorbance of the
tympanic membrane in rat and human samples. *PLoS One, 16*(7), e0254902.
doi:10.1371/journal.pone.0254902

Goyal, V., Rajguru, S., Matic, A. I., Stock, S. R., & Richter, C. P. (2012). Acute damage
threshold for infrared neural stimulation of the cochlea: functional and

histological evaluation. *Anat. Rec. (Hoboken), 295*(11), 1987-1999.
doi:10.1002/ar.22583

Guo, Z. V., Hires, S. A., Li, N., O'Connor, D. H., Komiyama, T., Ophir, E., . . . Svoboda,
K. (2014). Procedures for behavioral experiments in head-fixed mice. *PLoS One,*
9(2), e88678. doi:10.1371/journal.pone.0088678

Håkansson, B., Eeg-Olofsson, M., Reinfeldt, S., Stenfelt, S., & Granström, G. (2008).
Percutaneous versus transcutaneous bone conduction implant system: A
feasibility study on a cadaver head. *Otol. Neurotol., 29*(8), 1132-1139.
doi:10.1097/MAO.0b013e31816fdc90

Heimann, L., Carlein, C., Sorg, K., Diller, R., Langenbucher, A., Schick, B., & Wenzel,
G. I. (2021). Wavelength-specific optoacoustic-induced vibrations of the guinea
pig tympanic membrane. *J. Biomed. Opt., 26*(3), 038001.
doi:10.1117/1.Jbo.26.3.038001

Izzo, A. D., Richter, C. P., Jansen, E. D., & Walsh Jr., J. T. (2006). Laser stimulation of
the auditory nerve. *Lasers Surg. Med., 38*(8), 745-753. doi:10.1002/lsm.20358

Izzo, A. D., Suh, E., Pathria, J., Walsh, J. T., Whitlon, D. S., & Richter, C. P. (2007).
Selectivity of neural stimulation in the auditory system: A comparison of optic
and electric stimuli. *J. Biomed. Opt., 12*(2), 021008. doi:10.1117/1.2714296

Izzo, A. D., Walsh Jr., J. T., Jansen, E. D., Bendett, M., Webb, J., Ralph, H., & Richter, C.
P. (2007). Optical parameter variability in laser nerve stimulation: A study of pulse
duration, repetition rate, and wavelength. *IEEE Trans. Biomed. Eng., 54*(6), 1108-
1114. doi:10.1109/TBME.2007.892925

Kaneko, S., Niki, Y., Yamada, K., Nasukawa, D., Ujihara, Y., & Toda, K. (2022). Systemic
injection of nicotinic acetylcholine receptor antagonist mecamylamine affects
licking, eyelid size, and locomotor and autonomic activities but not temporal
prediction in male mice. *Mol. Brain., 15*(1), 77. doi:10.1186/s13041-022-00959-
y

Littlefield, P. D., & Richter, C. P. (2021). Near-infrared stimulation of the auditory nerve:
A decade of progress toward an optical cochlear implant. *Laryngoscope Investig.*
Otolaryngol., 6(2), 310-319. doi:10.1002/lio2.541

Matic, A. I., Robinson, A. M., Young, H. K., Badofsky, B., Rajguru, S. M., Stock, S., &
Richter, C. P. (2013). Behavioral and electrophysiological responses evoked by
chronic infrared neural stimulation of the cochlea. *PLoS One, 8*(3), e58189.
doi:10.1371/journal.pone.0058189

Matic, A. I., Walsh, J. T., Jr., & Richter, C. P. (2011). Spatial extent of cochlear infrared

neural stimulation determined by tone-on-light masking. *J. Biomed. Opt., 16*(11),

118002. doi:10.1117/1.3655590

Moreno, L. E., Rajguru, S. M., Matic, A. I., Yerram, N., Robinson, A. M., Hwang, M., . . .

Richter, C. P. (2011). Infrared neural stimulation: Beam path in the guinea pig

cochlea. *Hear. Res., 282*(1-2), 289-302. doi:10.1016/j.heares.2011.06.006

Mulrow, C. D., Aguilar, C., Endicott, J. E., Tuley, M. R., Velez, R., Charlip, W. S., . . .

DeNino, L. A. (1990). Quality-of-life changes and hearing impairment. A

randomized trial. *Ann. Intern. Med., 113*(3), 188-194. doi:10.7326/0003-4819-

113-3-188

Mulrow, C. D., Tuley, M. R., & Aguilar, C. (1992). Sustained benefits of hearing aids. *J.*

Speech Hear. Res., 35(6), 1402-1405. doi:10.1044/jshr.3506.1402

Pillong, L., Stahn, P., Hinsberger, M., Sorg, K., Schick, B., & Wenzel, G. I. (2020).

Cytotoxicity studies of an optoacoustic stimulation strategy for the development

of laser-based hearing aids. *J. Biomed. Opt., 25*(6), 068002.

doi:10.1117/1.Jbo.25.6.068002

Prado-Guitierrez, P., Fewster, L. M., Heasman, J. M., McKay, C. M., & Shepherd, R. K.

(2006). Effect of interphase gap and pulse duration on electrically evoked

potentials is correlated with auditory nerve survival. *Hear. Res., 215*(1-2), 47-55.

doi:10.1016/j.heares.2006.03.006

Rajguru, S. M., Matic, A. I., Robinson, A. M., Fishman, A. J., Moreno, L. E., Bradley,

A., . . . Richter, C. P. (2010). Optical cochlear implants: evaluation of surgical

approach and laser parameters in cats. *Hear. Res., 269*(1-2), 102-111.

doi:10.1016/j.heares.2010.06.021

Ramekers, D., Versnel, H., Strahl, S. B., Klis, S. F., & Grolman, W. (2015). Recovery

characteristics of the electrically stimulated auditory nerve in deafened guinea

pigs: Relation to neuronal status. *Hear. Res., 321*, 12-24.

doi:10.1016/j.heares.2015.01.001

Richter, C. P., Bayon, R., Izzo, A. D., Otting, M., Suh, E., Goyal, S., . . . Walsh Jr., J. T.

(2008). Optical stimulation of auditory neurons: Effects of acute and chronic

deafening. *Hear. Res., 242*(1-2), 42-51. doi:10.1016/j.heares.2008.01.011

Richter, C. P., Rajguru, S. M., Matic, A. I., Moreno, E. L., Fishman, A. J., Robinson, A.

M., . . . Walsh, J. T. (2011). Spread of cochlear excitation during stimulation with

pulsed infrared radiation: Inferior colliculus measurements. *J. Neural. Eng., 8*(5),

056006. doi:10.1088/1741-2560/8/5/056006

Schneider, C. A., Rasband, W. S., & Eliceiri, K. W. (2012). NIH Image to ImageJ: 25

years of image analysis. *Nat. Methods., 9*(7), 671-675. doi:10.1038/nmeth.2089

Schultz, M., Baumhoff, P., Maier, H., Teudt, I. U., Kruger, A., Lenarz, T., & Kral, A.
(2012). Nanosecond laser pulse stimulation of the inner ear-a wavelength study.
Biomed. Opt. Express, 3(12), 3332-3345. doi:10.1364/boe.3.003332

Shapiro, M. G., Homma, K., Villarreal, S., Richter, C. P., & Bezanilla, F. (2012). Infrared
light excites cells by changing their electrical capacitance. *Nat. Commun., 3*, 736.
doi:10.1038/ncomms1742

Sorg, K., Stahn, P., Pillong, L., Hinsberger, M. P., Heimann, L., Foth, H. J., . . . Wenzel,
G. I. (2019). First biocompatibility margins for optical stimulation at the eardrum
via 532-nm laser pulses in a mouse model. *J. Biomed. Opt., 24*(8), 085003.
doi:10.1117/1.Jbo.24.8.085003

Tamai, Y., Horinouchi, K., Hiryu, S., & Kobayasi, K. I. (2019). Infrared laser stimulation
of cochlear nerve through a tympanic membrane. *Conf. Proc. IEEE Life Sci. Tech.*
2019, 1(1), 54-57. doi:10.1109/LifeTech.2019.8884023

Tamai, Y., Ito, Y., Furuyama, T., Horinouchi, K., Murashima, N., Michimoto, I., . . .
Kobayasi, K. I. (2020). Auditory cortical activity elicited by infrared laser
irradiation from the outer ear in Mongolian gerbils. *PLoS One, 15*(10), e0240227.
doi:10.1371/journal.pone.0240227

Tan, X., Jahan, I., Xu, Y., Stock, S., Kwan, C. C., Soriano, C., . . . Richter, C. P. (2018).
Auditory neural activity in congenitally deaf mice induced by infrared neural
stimulation. *Sci. Rep., 8*(1), 388. doi:10.1038/s41598-017-18814-9

Thompson, A. C., Fallon, J. B., Wise, A. K., Wade, S. A., Shepherd, R. K., & Stoddart, P.
R. (2015). Infrared neural stimulation fails to evoke neural activity in the deaf
guinea pig cochlea. *Hear. Res., 324*, 46-53. doi:10.1016/j.heares.2015.03.005

Thompson, A. C., Wade, S. A., Pawsey, N. C., & Stoddart, P. R. (2013). Infrared neural
stimulation: Influence of stimulation site spacing and repetition rates on heating.
IEEE Trans. Biomed. Eng., 60(12), 3534-3541.
doi:10.1109/TBME.2013.2272796

Toda, K., Lusk, N. A., Watson, G. D. R., Kim, N., Lu, D., Li, H. E., . . . Yin, H. H. (2017).
Nigrotectal stimulation stops interval timing in mice. *Curr. Biol., 27*(24), 3763-
3770.e3763. doi:10.1016/j.cub.2017.11.003

Uenaka, M., Nagamura, H., Hiryu, S., Kobayasi, K. I., & Tamai, Y. (2022). Feasibility
evaluation of transtympanic laser stimulation of the cochlea from the outer ear. *J.*
Acoust. Soc. Am., 152(3), 1850-1855. doi:10.1121/10.0014241

van Eijl, R. H., Buitenhuis, P. J., Stegeman, I., Klis, S. F., & Grolman, W. (2017).

Systematic review of compound action potentials as predictors for cochlear
implant performance. *Laryngoscope, 127*(2), 476-487. doi:10.1002/lary.26154

Wells, J., Kao, C., Konrad, P., Milner, T., Kim, J., Mahadevan-Jansen, A., & Jansen, E. D.
(2007). Biophysical mechanisms of transient optical stimulation of peripheral
nerve. *Biophys. J., 93*(7), 2567-2580. doi:10.1529/biophysj.107.104786

World Health Organization. (2018). Addressing the rising prevalence of hearing loss.
Available at: https://apps.who.int/iris/handle/10665/260336 [Accessed 7
September 2022].

World Health Organization. (2019). New WHO-ITU standard aims to prevent hearing
loss among 1.1 billion young people. *Available at:*
https://www.who.int/news/item/12-02-2019-new-who-itu-standard-aims-to-
prevent-hearing-loss-among-1.1-billion-young-people [Accessed 7 September
2022].

World Health Organization. (2021). Deafness and hearing loss. *Available at:*
https://www.who.int/news-room/fact-sheets/detail/deafness-and-hearing-loss
[Accessed 7 September 2022].

Yamamoto, K., Yamada, K., Yatagai, S., Ujihara, Y., & Toda, K. (2022). Spatiotemporal
Pavlovian head-fixed reversal learning task for mice. *Mol. Brain., 15*(1), 78.
doi:10.1186/s13041-022-00952-5

Yeung, J., Griffin, A., Newton, S., Kenna, M., & Licameli, G. R. (2018). Revision
cochlear implant surgery in children: Surgical and audiological outcomes.
Laryngoscope, 128(11), 2619-2624. doi:10.1002/lary.27198

Young, H. K., Tan, X., Xia, N., & Richter, C. P. (2015). Target structures for cochlear
infrared neural stimulation. *Neurophotonics, 2*(2), 025002.
doi:10.1117/1.NPh.2.2.025002

Figure 1. Time course of the experiment. Before head-fixed classical conditioning,

surgery was performed to set the head plate and recording electrode on the subjects and

to impair hearing sensitivity in the right ear. At least two days after the surgery, the

training session was conducted to associate the relationship between the auditory cue and

the liquid reward for 11.8 ± 3.9 days (Mean $\pm$ SEM, n=18). The pre-exposure test was

conducted after the training was completed. Subsequently, continuous laser exposure was

presented for 15 h after recovery (0.3 ± 0.2 days). The cochlear response was recorded

during the laser exposure. The post-exposure test was carried out 2.8 ± 0.9 (Mean$\pm$SEM)

days after the laser exposure. Finally, the subjects were euthanized on the same day as the

post-exposure test, and histological inspection was performed.

669
670

Figure 2. Thermal elevation of the cochlear surface caused by laser irradiation. (A)

Time course of the average temperature rise on the irradiation site of the cochlear surface.

(B) Surface temperature at 10 min after the laser onset in (A). Error bars indicate the

standard deviation (1.6 W/cm^2: n = 3, others: n = 4). (C) Temperature difference between

the lateral and medial sides of the cochlear surface following the laser irradiation of 105.6

W/cm^2 (n=5). Glay open circles show the temperature in each cochlea. In (A) and (B), the

gray dotted line shows the room temperature (24 °C). Error bars describe the standard

error of the mean. **: $P < 0.01$.

Figure 3. Laser-induced lesion in spiral ganglion neurons. (A) Horizontal section of

laser irradiated cochlea showing all turns (radiant energy: 6.6 W/cm^2). (B) Density of

spiral ganglion neurons in the lateral side of the basal turn after laser exposure of 1.6

W/cm^2 (left), 6.6 W/cm^2 (middle), and 105.6 W/cm^2 (right). (C) Lateral (left) and medial

(right) sides of the spiral ganglion neuron density in middle turn. The dotted line separates

the lateral and medial sides of the spiral ganglion area in the middle turn. (D) Change in

the cell density after laser exposure on lateral (left) and medial (right) sides of the cochlea.

The cell density difference was defined as the mean spiral ganglion density difference

between the laser irradiated (left ear) and the opposite (right ear) cochlea. Gray open

circles indicate individual data. Correlation coefficients were calculated using Pearson's

correlation analysis. (E) Mean cell density difference between the laser irradiated and

opposite cochlea in apical-middle (left) and basal (right) turn on the lateral side. Data

from all radiant energy conditions were pooled in the analysis. *: $P < 0.05$.

Figure 4. Cochlear function assessment during continuous laser exposure. (A)

Cochlear response during continuous pulsed laser irradiation of 1.6 (left), 6.6 (middle),

and 105.6 (right) W/cm^2 (i.e., 125, 500, and 8000 pulses/s). (B) Time course of the

cochlear response amplitude elicited by the 1.6 (left), 6.6 (middle), and 105.6 (right)

W/cm^2 laser stimulation. (C) Relationship between laser irradiation time and percent of

the response amplitude. (D) Average amplitude decrease caused by continuous laser

irradiation depending on the radiant energy. Gray open circles indicate individual data.

Gray lines in (B) and (C) show individual data.

Figure 5. Influence of continuous laser irradiation on the auditory-evoked

behavioral response. (A) Auditory-evoked licking behavior pre- and post-laser exposure

of 1.6 (left), 6.6 (middle), and 105.6 (right) W/cm². Individual trials and mean peri-

stimulus time histograms (PSTH) of the licking behavior elicited by auditory stimuli (80

dB SPL) are shown in top and bottom figures, respectively. The blue zones indicate

auditory stimulation periods. The light blue dotted lines show the reward timing in the

training trials. Black plots in the top figure show licking timing. (B) Behavioral level

alternation after continuous laser irradiation. Solid and dotted lines show pre- and post-

behavioral levels, respectively. The behavioral level was determined as the ratio of delta

licking rate over the maximum delta licking rate of the pre-laser exposure (1.6 W/cm^2:

n=1, 3.3 W/cm^2: n=2, 6.6 W/cm^2: n=2, 26.4 W/cm^2: n=2, 52.8 W/cm^2: n=2, 105.6 W/cm^2:

n=3). (C) Change in the delta behavioral level between pre- and post-laser exposure. In

the control condition, delta behavioral level was measured by comparing the pre- and

post-placing the subjects on the experimental setup for 15 h (n=9). Gray dotted line shows

the mean of the delta behavioral level in the control condition, and colored areas indicate

two (red) and three (orange) standard deviations. The safety margin was determined as

the mean ± three standard deviations (i.e., orange area). Gray open cycles represent delta

behavioral response in each subject (n = 18).

728
729
730
731

 Supplemental Figures

 Supplemental Figure 1. Auditory-induced licking behavior in head-fixed Mongolian

 gerbils and behavioral profile depending on the sound pressure level. (A) Schematic

 of the head-fixed setup. We investigated the perception of auditory stimuli by using head-

 fixed classical conditioning. A drop of water was delivered through the tube after

 presenting white noise, and the timing of licks was recorded using a touch sensor, as in

 previous studies. (B) Individual trial (top) and mean peri-stimulus time histogram (PSTH)

 (bottom) of the licking behavior elicited by auditory stimuli with various sound pressure

 levels. The blue zones indicate auditory stimulation periods. The light blue dotted lines

 show the reward timing in the training trials. Black plots in the top figure show licking

timing. Increasing the intensity of white noise elicited a systematic change in behavioral

responses. (C) Intensity dependence of the peak licking rate, duration, peak time period,

and latency. As the sound pressure level increased from 20 to 95 dB SPL, peak licking

rate was increased from 1.52 to 7.28 licks/s, duration was increased from 0.38 to 0.83 s,

peak period was decreased from 1.66 to 0.63 s, latency was decreased from 1.06 to 0.11

s. (D) Mean auditory-evoked delta licking rate. The delta licking rate was defined as the

response amplitude to quantify the response amplitude of licking behavior. The delta

licking rate significantly increased from 0.04 to 4.78 licks/s depending on the intensity of

the white noise. In (C) and (D), the open grey cycles show the individual data (n=12).

Supplemental Figure 2. Change in the delta behavioral level depending on the

decrease in the cell density difference (A) and neural response amplitude (B).

Histological and behavioral data (n=9) or electrophysiological and behavioral data (n=8)

were obtained from the subjects. The correlation coefficients were calculated using

Pearson's correlation analysis. As the spiral ganglion neuron density decreased, the delta

behavioral level significantly decreased (r=-0.80, P<0.05, Pearson's correlation analysis).

A similar tendency was observed in the relationship between the decrement of the

cochlear response and behavioral response; the decreasing amplitude of the cochlear

response significantly reduced the behavioral response (r=-0.83, P<0.05, Pearson's

correlation analysis). These results suggest that auditory perception is affected by the

physiological deterioration of the cochlea. The open circles indicate individual data. *: *P*

< 0.05.