OpenFIBSEM: an application programming interface for easy FIB/SEM automation.

Patrick Cleeve[1], David Dierickx[1], Genevieve Buckley[2], Sergey Gorelick[3], Lucile Naegele[1], Lachlan Burne[1], James C Whisstock[1,4,*], Alex de Marco[1,*]

1. Department of Biochemistry and Molecular Biology, Biomedicine Discovery Institute, Monash University, 3800 Clayton Victoria, Australia
2. Ramaciotti Centre for cryo-Electron Microscopy, Monash University, 3800 Clayton Victoria, Australia
3. Monash Centre for Electron Microscopy, Monash University, 3800 Clayton Victoria, Australia
4. EMBL Australia, Monash University, 3800 Clayton Victoria, Australia

* correspondence to alex.demarco@monash.edu; james.whisstock@monash.edu

Abstract

Automation in microscopy is the key to success in long and complex experiments. Most microscopy manufacturers provide Application Programming Interfaces (API) to enable communication between a user-defined program and the hardware. Although APIs effectively allow the development of complex routines involving hardware control, the developers need to build the applications from basic commands. Here we present a Software Development Kit (SDK) for easy control of Focussed Ion Beam Scanning Electron Microscopes (FIB/SEM) microscopes. The SDK, which we named OpenFIBSEM consists of a suite of building blocks for easy control that simplify the development of complex automated workflows.

Keywords

Focused Ion Beam microscopy, automation, Python, API, Microscopy, controller

Introduction

Focused Ion Beam Scanning Electron Microscopes (FIB/SEM) have been widely used across multiple fields: from semiconductors to materials and life sciences(Lepinay and Lorut, 2013; Phaneuf, 1999; Schaffer and Wagner, 2008; Villinger et al., 2012). The most common uses for FIB/SEMs are sample preparation for transmission electron microscopy (TEM), commonly known as lamella preparation (Rigort et al., 2012; Schaffer et al., 2012); FIB lithography (binary, grayscale, or additive) (Garg et al., 2019; Gorelick et al., 2019; Gorelick and De Marco, 2018), and three-dimensional imaging (Slice&View or metrology) (Kizilyaprak et al., 2014; Mammadi et al., 2020; Villinger et al., 2012). Regardless of the application, automation, customisation of the workflow, and real-time data analyses are critical. The effects of automation regarding throughput and success rate have been amply demonstrated on multiple occasions in microscopy. For example, in TEM imaging, the introduction of SerialEM, Leginon, and other open-access TEM control software (Carragher et al., 2000; de la Cruz et al., 2019; Potter et al., 1999; Suloway et al., 2009; Zheng et al., 2007) allowed high-throughput single particle and tomographic acquisition, as well as the introduction of complex and non-conventional acquisition methods such as the dose-symmetric tilt-series (Hagen et al., 2017). In light microscopy, Micro-manager (Edelstein et al., 2010; Edelstein et al., 2014) allowed easy control of commercial and custom-built setups and the development of automated procedures for complex data acquisition (Pitrone et al., 2013). Both the abovementioned cases effectively represent an easy interpreter of the microscope and detector manufacturers with both a graphical and a scriptable user interface.

Microscope manufacturers provide Application Programming Interfaces (API) to allow communication with the instrumentation. The classes in the APIs comprise basic commands and calls to low-level actions such as stage movements (absolute or relative) or basic image acquisition. In this work, we focus on the ThermoFisher API Autoscript. The library grants access to all the core functionalities of the microscopes, but it does not provide significantly advanced functionality. For example, it is easy to define image acquisition using Autoscript, but it is not possible to customise the autofocus routine with arbitrary parameters.

Currently, there are multiple commercial products available to run automated
procedures. For example, AutoSlice&View allows the guided setup of FIB/SEM
tomography data acquisition. The program has an easy-to-follow graphical user
interface and allows for a certain degree of flexibility in the setup. Still, due to how
data is stored in the memory, it shows significant lag on long acquisitions, and it
lacks the possibility to automatically re-align the Field of View (FOV) based on the
image content. The main consequence of this limitation is that upon acquisition of
large volumes, the FOV of interest moves out of the imaged region once a significant
depth has been milled. The abovementioned example highlights how easy it is for a
research-oriented commercial product not to anticipate a use case.

In most cases, the manufacturers will address these limitations, but between the
user feedback and the implementation of the desired feature, up to two years can
easily pass. Further, the commission may never occur for some features deemed
minor or niche. As such, due to the inability to alter the program flow or customise
how lower-level operations are performed, researchers are often unable to pursue a
research direction due to the cost and complexity of redeveloping solutions.

There are at least two examples in TEM imaging where open-source software was
developed and successfully pushed the development of custom complex workflows:
SerialEM (Mastronarde, 2005) and Leginon (Carragher et al., 2000). Here, both
packages are built on the API provided by the microscope manufacturers but provide
significantly advanced functionality and, most importantly, the possibility to easily
script and create novel pipelines. For example, the recent development of py-EM led
to the possibility controlling TEM microscopes using Python (Schorb et al., 2019).
This high level of flexibility associated with the increased accessibility given by the
library led to the development of other packages such as PACEtomo (Eisenstein F et
al., 2022). Finally, both, SerialEM and Leginon extend their support beyond a single
manufacturer, making it easier for a novice user to learn the control of multiple
instruments sharing a common interface.

Here we introduce OpenFIBSEM, a python-based Software Development Kit (SDK)
that builds on the manufacturer's API, providing a simplified way of interacting with
the microscope and designing custom workflows. The package comprises a
collection of classes that aid imaging, stage movement, manipulator control, system

calibration and easy integration with a graphical user interface. Currently,
OpenFIBSEM only supports ThermoFisher instruments, but we plan to extend this to
other manufacturers in the future to enable users to develop automated routines on
multiple instruments with a stable syntax; therefore, simplifying the adoption of
complex workflows across facilities and instruments.

Results and Discussion

OpenFIBSEM aims to enable fast workflow development while allowing the low-level
customisation required by researchers (Figure 1). The library is designed to provide
high-level abstractions of common operations, such as image acquisition and
movement, and reusable components, such as automated calibration procedures. In
addition, the library includes several reusable user interface components (PyQt /
Napari) for common user operations (e.g. movement). Ideally, the library should
abstract microscope operation from the underlying hardware, allowing the user to
focus on what they want to do rather than how to do it.

The library can be downloaded from www.github.com/demarcolab/fibsem and is
organised according to the following structure:

Core

• **Acquire**: A collection of helper functions for the simple definition of imaging
settings enabling image acquisition and post-processing.
• **Milling**: A collection of helper functions to set up and run focused ion beam
milling routines. This collection includes some patterning examples.
• **Movement**: Functionality for the stage and needle movement, corrected for both
the view perspective and stage orientation.
• **Structures**: A collection of classes defining the data structures for a correct
interpretation of the settings and types.
• **Utils**: A collection of general microscope utilities, including checks for network
connection and filesystem access.

Automation

• **Alignment**: Provides the tools for automated alignments of the sample position
to correct for drift or to correct the eucentric position.

• **Calibration**: Includes multiple automated calibration routines and helps manage
the microscope state.
• **Validation**: A collection of automated validation functions to check and set the
desired microscope settings. The user can configure all the settings.

Image Analysis

• **Imaging**: A collection of helper functions for simple image analysis and
manipulation, including filtering and masking.
• **Detection**: Comprises functions for the automated detection of common features
present in conventional workflows, such as the manipulator (NeedleTip class) or
the post of a half-moon grid (LandingPost class).
• **Segmentation**: A library providing tools for implementing a deep learning-based
segmentation workflow, including labelling, training and inference.

Utilities

• **Conversions**: Provides standard conversions between coordinate systems used
in the microscope and allows converting distances in images between pixels and
meters.
• **UI**: A collection of basic user interface windows in PyQt designed for basic
interaction with OpenFIBSEM. Two examples are the movement window and the
detection window.

Getting started
The main goal of the package is to simplify the performance of general tasks within
the microscope capabilities to enable rapid development. The library requires a
general familiarity with Python programming; however, the setup process is concise.
The example below shows how to connect to the microscope, define the imaging
parameters and acquire an image using the electron beam with the automated
contrast selection.

```
1 from fibsem import utils, acquire
2 from fibsem.structures import BeamType, ImageSettings

# connect to microscope
```

```
3 microscope = utils.connect_to_microscope(ip_address="10.0.0.1")

# set imaging settings
5 image_settings = ImageSettings(
        hfw=150e-6,
        resolution="1536x1024",
        dwell_time=1e-6,
        autocontrast=True,
        beam_type=BeamType.ELECTRON)

# acquire electron image
6 eb_image = acquire.new_image(microscope, image_settings)
```

Although acquiring an image is a simple case, it is the most commonly performed
microscope operation. By analysing the difference between the code required using
OpenFIBSEM and the one directly implemented from AutoScript, it is immediately
evident that the complexity has decreased and it is now possible to obtain the same
result without the need of traversing into three separate modules (imaging, beams,
and auto_functions).

```
1 from autoscript_sdb_microscope_client import SdbMicroscopeClient
2 from autoscript_sdb_microscope_client.structures import GrabFrameSettings,
      RunAutoCbSettings

# connect to microscope
3 microscope = SdbMicroscopeClient()
4 microscope.connect("10.0.0.1")

# set the active view
5 microscope.imaging.set_active_view(1)

# set frame settings
6 frame_settings = GrabFrameSettings(
    resolution="1536x1024",
    dwell_time=1e-6)

# set hfw
7 microscope.beams.electron_beam.horizontal_field_width.value = 150e-6
```

```
# autocontrast
8 cb_settings = RunAutoCbSettings(
     method="MaxContrast",
     resolution="768x512",
     number_of_frames=5 )
9 microscope.auto_functions.run_auto_cb(cb_settings)

# grab frame
10 eb_image = microscope.imaging.grab_frame(frame_settings)
```

Automating routines

One of the most common actions performed when using a FIB/SEM is focusing; this
is crucial not only to obtain optimal images but also to provide a measurement of the
effective position of the sample along the electron axis. Accordingly, in automated
procedures, a robust autofocus routine is critical.

The autofocus routine is an excellent example of an automated procedure that is
available but inefficient. Autoscript provides a standard autofocus routine that uses
the full image and focusses with a standard interval (+/-10μm). In many cases, this
routine is not flexible enough; for example, if the sample is highly dose-sensitive,
using a sub-region is critical to avoid damage because of repeated exposure.
Further, re-focusing is likely required if the stage movements are significant and the
sample is irregular. In this case, the standard procedure often fails because the
interval for the autofocus may not be large enough. Below is an example of a simple
routine developed with OpenFIBSEM enabling custom autofocus.

```
def auto_focus(microscope: SdbMicroscopeClient, settings:   ImageSettings,
     mode: str = "default",
     wd_delta: float = 0.05e-3, steps: int  = 5,
     reduced_area: Rectangle = Rectangle(0.3, 0.3, 0.4, 0.4) ) -> None:

     if mode == "default":
         microscope.imaging.set_active_device(BeamType.ELECTRON.value)
```

```
        microscope.imaging.set_active_view(BeamType.ELECTRON.value)
        focus_settings = RunAutoFocusSettings()
        microscope.auto_functions.run_auto_focus(focus_settings)

    if mode == "sharpness":
        prev_image_settings = image_settings
        settings.resolution = "768x512"  # "1536x1024"
        settings.dwell_time = 0.2e-6
        settings.hfw = 50e-6
        settings.autocontrast = True
        settings.beam_type = BeamType.ELECTRON
        settings.save = False

 # loop through working distances and calculate the sharpness (acutance)

        current_wd = microscope.beams.electron_beam.working_distance.value
        min_wd = current_wd - (steps*wd_delta/2)
        max_wd = current_wd + (steps*wd_delta/2)
        working_distances = np.linspace(min_wd, max_wd, steps+1)

 # loop through working distances and calculate the sharpness (acutance)
 # highest acutance is best focus
        sharpeness_metric = []
        for i, wd in enumerate(working_distances):
            microscope.beams.electron_beam.working_distance.value = wd
            img        =        acquire.new_image(microscope,        settings,
reduced_area=reduced_area)

# calculate sharpness (Acutance)
            out = gradient(skimage.filters.median(np.copy(img.data)), disk(5))
            sharpness = np.mean(out)
            sharpeness_metric.append(sharpness)
# select working distance with max acutance
        idx = np.argmax(sharpeness_metric)
# set new working distance
        microscope.beams.electron_beam.working_distance.value            =
working_distances[idx]
        settings = prev_image_settings
```

Enhanced Functionality

The AutoScript API provides functionality to move the stage in two coordinate

systems (relative and raw). The raw coordinate system is based on the physical

cartesian coordinates, while the relative coordinate system is based on the relative

distance from the beam (defined by the focus and link routine). When moving using

AutoScript or the xT-UI, the stage will move correctly and maintain eucentricity

between both the Electron and Ion beams. However, if the stage shuttle has a pre-
tilted angle (such as for lamella preparation to provide additional access for the Ion
beam), neither the API nor the UI will move correctly and degrade eucentricity over
time as there is no way to tell the microscope that the stage is pre-tilted.

OpenFIBSEM provides corrected movement functions for both the stage and
manipulator. The user can provide the pre-tilt of the shuttle in the settings, and the
library will calculate the corrected movements to maintain eucentricity. Moreover, the
user doesn't need to spend time understanding the implementation details regarding
different coordinate systems, stage movements, and hardware and, therefore, can
focus on research. Figure 2 shows the difference between baseline and corrected
movements for a stage with a 35-degree pre-tilt angle. The user clicks the same
position in xT-UI, and openFIBSEM and the actual positions after movement for each
beam are shown.

Common Workflow Tools
OpenFIBSEM provides a set of common tools enabling multiple workflows utilised in
imaging. For example, deep learning has been successfully implemented for
denoising and segmentation (a couple of examples repositories can be found here:
https://github.com/HenriquesLab/ZeroCostDL4Mic,
https://github.com/juglab/DenoiSeg). However, specific implementations often don't
provide the tools or flexibility for researchers to utilise models or only cover a specific
aspect of deep learning (usually training the model). This excludes the most time-
consuming and technically difficult aspects of using a deep learning model: labelling
and deployment. To overcome this problem, here we include a workflow for
collecting microscopy images, labelling them, training a segmentation model, and
deploying it in a workflow.

We have created a labelling user interface using napari (Sofroniew et al., 2022) and
Dask (Rocklin, 2015), allowing a user to label hundreds of images efficiently. The
user can then train and test their choice of prebuilt models from segmentation-
models-Pytorch (Paszke et al., 2019) by editing a simple configuration file (YAML).
We also provide a simple wrapper class for deploying the model into an
openFIBSEM workflow. An example of the proposed workflow can be found in figure
3, with example outputs of each stage of the pipeline.

User Interface

To enable non-expert users to utilise FIBSEM machines to pursue research, workflows often require image visualisation and user interaction. To this end, we developed a library of common user interface components and tools based on PyQt and Napari that can be used as standalone or as building blocks to be integrated into larger applications.

An example of a user interface provided with OpenFIBSEM, that can be used standalone or integrated into larger workflows and application, is the movement window. This application provides a user interface for users to double-click to move (mirroring the microscope user interface interaction). This applet allow navigation in two different movement modes, stable and eucentric. Stable movements (discussed above) consist of standard stage moves that maintain eucentricity between beams, while eucentric mode can correct the eucentricity between beams. The movements in both cases account for the sample pre-tilt and assume a flat sample, therefore in the case of a sample displaying large variations in height across the surface corrections will be required more often when moving. Over time, we aim to add additional components and improve the functionality and modularity of these interfaces.

Examples of workflows developed with OpenFIBSEM

To highlight the possibilities offered by the OpenFIBSEM package, we developed a few examples of automated routines covering multiple use cases (Figure 1). All the exemplary code can be found in the "Example" directory of the repository.

The most straightforward case use consists of performing grey-scale lithography using a pattern imported from images. This is common practice in microfabrication, and the example we present consists of fabricating the profile of a lens. The applet includes the interactive movement along the sample using the eucentric corrected movement and ensures that the image file format is compatible with the pattern definition. Patterns are created as 8-bit grayscale images and then converted to a 24-bit bitmap as required by the system.

Another case is the serial fabrication of lamellae   on grids. This use-case is extremely popular in the life sciences, and the automation of this procedure has been presented in a couple of articles (Buckley et al., 2020; Klumpe et al., 2021).

The first example of this procedure was published by our group and was a complex
package; with OpenFIBSEM, the core of the program can be scripted extremely
quickly. A working example of this procedure can be found in the Example directory
named: "autolamella.py". Although this is only a simple example, it highlights the
potential for new users and groups to develop their own automation routines.

The last example is the performance of FIB/SEM tomography. This is common
practice across multiple disciplines, from the material to the life sciences. The
workflow consists of imaging a surface with the SEM, polishing a defined amount of
material from the surface using the FIB and imaging the newly exposed surface. This
is then iterated until the desired volume has been acquired. The simplest setup does
not require any stage movement, and the SEM imaging is performed at an angle
using the dynamic focusing option to compensate for the focal gradient. In these
conditions, the code is reduced to a loop where the sample surface is first imaged,
and then the surface is polished with the FIB. A re-alignment step can be added in
the loop to compensate for any drift. Among the examples available in the repository,
we show a short version of the FIB/SEM tomography workflow developed in less
than 40 lines of code using OpenFIBSEM.

Although these examples were intended as simple demonstrations or re-
implementations of existing workflows to provide users with a set of templates to
start using OpenFIBSEM, we immediately found use cases and users for these
programs. Accordingly, these developed into standalone projects.

Vulcan
Vulcan is a project for simplifying and enhancing grayscale lithography
manufacturing using OpenFIBSEM. It contains tools for pattern conversion, milling
calibration, and workflow tools. It can be used, for example, in conjunction with the
optical simulation package Juno (https:github.com/DeMarcoLab/juno) to design,
develop, manufacture and validate micro-optical elements.

https://github.com/DeMarcoLab/vulcan

Salami
Salami is a custom slice and view project to enable custom data collection
strategies. The project is currently being used to screen and optimise data collection
strategies for serial FIB/SEM tomography of large samples. The flexibility of the

package allows for testing The iterative development for the project is shown in
Figure 4, and the repository can be accessed from
https://github.com/DeMarcoLab/salami.

Conclusions

In this article we introduced a high-level SDK enabling the quick and robust
development of automated pipelines for FIB/SEM data acquisition and
microfabrication. The increased level of abstraction provided by this package
compared to the microscope API does not only make the development of new
routines easier, but it is also integrated with numerous accessory functions that can
be used to extend the capability of the microscope by integrating complex image
processing pipelines.

The current version of OpenFIBSEM has been developed around the ThermoFisher
Autoscript API. We inted to maintain this package, while extending the support to
other manufacturers, with the goal of providing a common high-level programming
language across multiple instruments. The code is freely available under the MIT
licence agreement at www.github.com/demarcolab/fibsem. To request new features
or report bugs users are asked to submit the requests through GitHub.

Figures

Figure 1: The architecture concept. The diagram shows the lowest level and
therefore furthest away from the user is the hardware. The communication API
provides access and control of the hardware, the Backend SDK provides a user-
friendly control therefore enabling an easy development of user-specific applications.
The outermost ring provides a list of examples of the user-specific applications that
can easily be developed when a high-level API is present.

Figure 2: OpenFIBSEM Corrected Movement. Given the orientation of the beams relative to the sample the XY coordinate axes of the stage correspond to those of the Electron Beam view. Accordingly, as visible from the left column it is shown how the selection of a destination drives the correct movement of the stage to the target location. From the Ion Beam view a direct movement based on the distance between the starting point and the desired location provides an incorrect movement. This mismatch between the target and the actual reached position depends on the fact

that there is a difference between the coordinate system of the stage and the beam.
Further, if the sample is mounted with a tilt relative to the XY plane of the stage this
mismatch becomes particularly evident from the Ion Beam view and must be
accounted for. The last panel shows the effect that tilt correction has when
calculating the correct shift.

Figure 3: The U-Net assisted segmentation workflow integrated in OpenFIBSEM.
The toolkit includes a user interface based on napari to import and label (segment)
the images to train the model, a training setup and the code to easily deploy the

segmentation model to identify correctly features of interest. The example in the
figure shows the labelling and the recognition of the needle micromanipulator and a
lamella.

Figure 4: Diagram of an example of possible evolution of a protocol. Here we show 5
iterations during the development of a FIB/SEM tomography workflow. The colours
are representative of the components that have been added at each iteration. The
example shows how the initial iterative sequence of milling a surface, imaging, and
repositioning of the milling pattern can be improved by adding a step to neutralise
the charge left by the ion beam, then the imaging parameters can be changed, or
screened. Finally, the autofocus routine can be improved by using a custom
algorithm and finally a drift correction can be applied.

Acknowledgements

331 PC was supported by the Australian Research Data Commons project XN006. DD,
332 GB, JCW were supported by the Australian Research Council Laureate Fellowship
333 FL180100019. SG and AdM were supported by the Human Frontiers Science
334 Program grant RGP0017/2020.

Authors contributions

PC conceived and developed the library architecture and wrote the article, DD, GB,

SG LB, LN developed the library. AdM and JCW managed the project and wrote the

article.

References

Buckley, G., Gervinskas, G., Taveneau, C., Venugopal, H., Whisstock, J.C., and de Marco, A. (2020). Automated cryo-lamella preparation for high-throughput in-situ structural biology. J Struct Biol *210*, 107488.

Carragher, B., Kisseberth, N., Kriegman, D., Milligan, R.A., Potter, C.S., Pulokas, J., and Reilein, A. (2000). Leginon: an automated system for acquisition of images from vitreous ice specimens. J Struct Biol *132*, 33-45.

de la Cruz, M.J., Martynowycz, M.W., Hattne, J., and Gonen, T. (2019). MicroED data collection with SerialEM. Ultramicroscopy *201*, 77-80.

Edelstein, A., Amodaj, N., Hoover, K., Vale, R., and Stuurman, N. (2010). Computer control of microscopes using microManager. Curr Protoc Mol Biol *Chapter 14*, Unit14 20.

Edelstein, A.D., Tsuchida, M.A., Amodaj, N., Pinkard, H., Vale, R.D., and Stuurman, N. (2014). Advanced methods of microscope control using muManager software. J Biol Methods *1*.

Eisenstein F, Yanagisawa H, Kashihara H, Kikkawa M, Tsukita S, and Danev, R. (2022). Parallel cryo electron tomography on in situ lamellae. Biorxiv, 487557.

Garg, V., Chou, T., Liu, A., De Marco, A., Kamaliya, B., Qiu, S., Mote, R.G., and Fu, J. (2019). Weaving nanostructures with site-specific ion induced bidirectional bending. Nanoscale Adv *1*, 3067-3077.

Gorelick, S., De Jonge, M.D., Kewish, C.M., and De Marco, A. (2019). Ultimate limitations in the performance of kinoform lenses for hard x-ray focusing. Optica *6*, 790-793.

Gorelick, S., and De Marco, A. (2018). Fabrication of glass microlenses using focused Xe beam. Opt Express *26*, 13647-13655.

Hagen, W.J.H., Wan, W., and Briggs, J.A.G. (2017). Implementation of a cryo-electron tomography tilt-scheme optimized for high resolution subtomogram averaging. J Struct Biol *197*, 191-198.

Kizilyaprak, C., Bittermann, A.G., Daraspe, J., and Humbel, B.M. (2014). FIB-SEM tomography in biology. Methods Mol Biol *1117*, 541-558.

Klumpe, S., Fung, H.K., Goetz, S.K., Zagoriy, I., Hampoelz, B., Zhang, X., Erdmann, P.S., Baumbach, J., Muller, C.W., Beck, M.*, et al.* (2021). A modular platform for automated cryo-FIB workflows. Elife *10*.

Lepinay, K., and Lorut, F. (2013). Three-dimensional semiconductor device investigation using focused ion beam and scanning electron microscopy imaging (FIB/SEM tomography). Microsc Microanal *19*, 85-92.

Mammadi, Y., Joseph, A., Joulain, A., Bonneville, J., Tromas, C., Hedan, S., and Valle, V. (2020). Nanometric metrology by FIB-SEM-DIC measurements of strain field and fracture separation on composite metallic material. Mater Design *192*.

Mastronarde, D.N. (2005). Automated electron microscope tomography using robust prediction of specimen movements. Journal of Structural Biology *152*, 36-51.

Paszke, A., Gross, S., Massa, F., Lerer, A., Bradbury, J., Chanan, G., Killeen, T.,
Lin, Z.M., Gimelshein, N., Antiga, L*., et al.* (2019). PyTorch: An Imperative Style,
High-Performance Deep Learning Library. Adv Neur In *32*.
Phaneuf, M.W. (1999). Applications of focused ion beam microscopy to materials
science specimens. Micron *30*, 277-288.
Pitrone, P.G., Schindelin, J., Stuyvenberg, L., Preibisch, S., Weber, M., Eliceiri,
K.W., Huisken, J., and Tomancak, P. (2013). OpenSPIM: an open-access light-sheet
microscopy platform. Nat Methods *10*, 598-599.
Potter, C.S., Chu, H., Frey, B., Green, C., Kisseberth, N., Madden, T.J., Miller, K.L.,
Nahrstedt, K., Pulokas, J., Reilein, A*., et al.* (1999). Leginon: a system for fully
automated acquisition of 1000 electron micrographs a day. Ultramicroscopy *77*, 153-
161.
Rigort, A., Bauerlein, F.J., Villa, E., Eibauer, M., Laugks, T., Baumeister, W., and
Plitzko, J.M. (2012). Focused ion beam micromachining of eukaryotic cells for
cryoelectron tomography. Proc Natl Acad Sci U S A *109*, 4449-4454.
Rocklin, M. (2015). Dask: Parallel computation with blocked algorithms and task
scheduling. Paper presented at: Proceedings of the 14th python in science
conference.
Schaffer, M., Schaffer, B., and Ramasse, Q. (2012). Sample preparation for atomic-
resolution STEM at low voltages by FIB. Ultramicroscopy *114*, 62-71.
Schaffer, M., and Wagner, J. (2008). Block lift-out sample preparation for 3D
experiments in a dual beam focused ion beam microscope. Microchim Acta *161*,
421-425.
Schorb, M., Haberbosch, I., Hagen, W.J.H., Schwab, Y., and Mastronarde, D.N.
(2019). Software tools for automated transmission electron microscopy. Nat Methods
16, 471-477.
Sofroniew, N., Lambert, T., Evans, K., Nunez-Iglesias, J., Bokota, G., Winston, P.,
Peña-Castellanos, G., Yamauchi, K., Bussonnier, M., Doncila Pop, D*., et al.* (2022).
napari: a multi-dimensional image viewer for Python (Zenodo).
Suloway, C., Shi, J., Cheng, A., Pulokas, J., Carragher, B., Potter, C.S., Zheng,
S.Q., Agard, D.A., and Jensen, G.J. (2009). Fully automated, sequential tilt-series
acquisition with Leginon. J Struct Biol *167*, 11-18.
Villinger, C., Gregorius, H., Kranz, C., Hohn, K., Munzberg, C., von Wichert, G.,
Mizaikoff, B., Wanner, G., and Walther, P. (2012). FIB/SEM tomography with TEM-
like resolution for 3D imaging of high-pressure frozen cells. Histochem Cell Biol *138*,
549-556.
Zheng, S.Q., Keszthelyi, B., Branlund, E., Lyle, J.M., Braunfeld, M.B., Sedat, J.W.,
and Agard, D.A. (2007). UCSF tomography: an integrated software suite for real-time
electron microscopic tomographic data collection, alignment, and reconstruction. J
Struct Biol *157*, 138-147.

Autolamella v2
automated
cryo-lamella prep

PIE-Scope
Correlative Light and
Electron Microscopy
control

OpenFIBSEM (SDK)

Communication API (AutoScript)

Hardware
(FIB/SEM
Microscope)

Salami
Automated
FIB/SEM tomography

AutoLiftout
automated
cryo-liftout

**Automated
microscope
calibrations**

Vulcan
Grayscale
FIB lithography

Label

Train

Validation loss

Train loss

Epoch n

Epoch n

Deploy

Image

Segmentation mask

Iteration 1	Mill pattern (polish face)	EB, default autofocus		EB Imaging: single image		IB Move pattern
Iteration 2	Mill pattern (polish face)	EB, default autofocus	Charge neutralization	EB Imaging: single image		IB Move pattern
Iteration 3	Mill pattern (polish face)	EB, default autofocus	Charge neutralization	EB Imaging param sweep: energy, dwell t, current		IB Move pattern
Iteration 4	Mill pattern (polish face)	EB, custom autofocus	Charge neutralization	EB Imaging para sweep: energy, dwell t, current		IB Move pattern
Iteration 5	Mill pattern (polish face)	EB, custom autofocus	Charge neutralization	EB Imaging param sweep: energy, dwell t, current	Drift correction	IB Move pattern