MorphoAST – A Machine Learning Workflow for Sub-Doubling Time Single Cell Antimicrobial Susceptibility Testing

Kristel C. Tjandra[1]†, Nikhil Ram-Mohan[1]†, Manuel Roshardt[2]†, Zhaonan Qu[3], Elizabeth Zudock[1], Kathleen E. Mach[4], Okyaz Eminaga[4], Joseph C. Liao[4,5], Samuel Yang[1]*, and Pak Kin Wong[2,6]*

[1] Department of Emergency Medicine, Stanford University School of Medicine, Palo Alto, CA, USA

[2] Department of Biomedical Engineering, The Pennsylvania State University, University Park, PA, USA

[3] Department of Economics, Stanford University, Palo Alto, CA, USA

[4] Department of Urology, Stanford University School of Medicine, Palo Alto, CA, USA

[5] Veterans Affairs Palo Alto Health Care System, Palo Alto, CA, USA

[6] Department of Mechanical Engineering and Department of Surgery, The Pennsylvania State University, University Park, PA, USA

†These authors contributed equally to this work.

* Corresponding authors. Email: syang5@stanford.edu; pak@engr.psu.edu

Abstract

Background —

Multidrug-resistant bacteria are among the most urgent global public health threats. Rapid determination of antimicrobial resistance in a single bacterium is a major clinical unmet need in the diagnosis of bacterial infections.

Methods —

By capturing dynamic single-cell morphological features of over twenty-eight thousand cells, we evaluated strategies based on time and concentration differentials for classifying the susceptibility of *Klebsiella pneumoniae* to meropenem and predicting their minimum inhibitory concentrations (MIC).

Findings —

The classifiers achieved as high as 97% accuracy in 20 minutes (two-fifths of the doubling time) and reached over 99% accuracy within 50 minutes (one doubling time) in predicting the antimicrobial response. A regression model based on the concentration differential of individual cells predicted the MIC with >97% categorical agreement and 100% essential agreement. When tested against cells from an unseen strain, the regressor achieved a categorical agreement of 91.9% with a very major error of 0.1%.

Interpretation —

We report morphometric antimicrobial susceptibility testing (MorphoAST), an image-based machine learning workflow, for rapid determination of antimicrobial susceptibility by single-cell morphological analysis in a sub-doubling time. Our approach has the ability to predict bacterial antimicrobial responsiveness in a fraction of the organisms' doubling time. This innovation will have significant implications for the future management of bacterial infections.

46   Funding —

47   This work was supported in part by NIH NIAID (R01AI153133).

49   **Keywords**

50   multidrug-resistant bacteria, artificial intelligence, antibiotic susceptibility test, minimum

51   inhibitory concentration, microscopy

Research in Context

Evidence before this study —

Classic antimicrobial susceptibility methods rely heavily on bacterial replication, requiring at least 24-hour incubation, with minimum inhibitory concentration (MIC) reported as a quantitative measure of antimicrobial susceptibility. Recent insights into bacterial viability shed light on how bacteria undergo a wide variety of morphological changes when responding to various environments. This phenotypic adjustment led us to imaging-based techniques that could rapidly capture minute bacterial morphological changes at the earliest onset of antibiotic treatment.

Added value of this study —

We developed a machine learning workflow based on morphometric analysis of antibiotic-treated bacterial cells over time. By evaluating both time and concentration differential strategies, we were able to find an efficient workflow that combines single-cell imaging, computer vision feature extraction, and supervised learning models for predicting bacterial responses to antimicrobials. To our best knowledge, we are the first to show that such an approach can be implemented not only to classify but to quantify susceptibility using breakpoint MIC.

Implications of all the available evidence —

Our study demonstrates that single-cell morphological analysis is a promising avenue for accelerating AST and MIC quantification with the potential of achieving assay readouts in a fraction of the bacterial doubling time. Our approach is an advancement towards the development of innovative tools for rapidly determining phenotypic antimicrobial

susceptibility in bacteria. The clinical utility of this method will be further tested in prospective

studies.

Introduction

The emergence of multidrug-resistant pathogens is a worldwide calamity. Multidrug-

resistant bacteria, such as the carbapenemase-producing *Enterobacteriaceae*, pose an

increasing threat to public health due to their high mortality rate and rapid acquisition of

resistance to available antimicrobials.(1, 2) *Enterobacteriaceae* are common causes of

nosocomial infection and could lead to various life-threatening infections, including severe

lung, urinary tract, and bloodstream infections. Broad-spectrum carbapenem like meropenem,

an intravenous beta-lactam antibiotic, is often adopted for *Enterobacteriaceae* nosocomial

infections until the antimicrobial susceptibility test (AST) results are available to guide more

targeted therapy. However, in the absence of a rapid method to determine antimicrobial

susceptibility, empirical use of antibiotics is warranted. Inappropriate antibiotic treatment

results in almost twice higher mortality rate in infected patients and accelerates the emergence

and spread of superbugs.(3) Therefore, rapid determination of antimicrobial susceptibility to

guide treatment is crucial to save lives and curb the widespread of multidrug-resistant

pathogens.

Phenotypic AST, such as broth microdilution and Kirby-Bauer disk diffusion test,

evaluates the ability of an antimicrobial to inhibit bacteria growth and is the gold standard for

determining antimicrobial susceptibility. The turbidity of liquid media or the formation of

bacterial colonies provides a measure of bacteria growth with the presence of an antimicrobial

and generates quantitative minimum inhibitory concentration (MIC).(4) Nevertheless,

phenotypic AST, which relies on bacterial replication for 18 hours or more, is unable to

accommodate a rapid turnaround. This limitation of current AST approaches propelled the

development of novel strategies to improve both speed and accuracy.(5–9) Single-cell imaging

analysis is an emerging strategy that offers the possibility of reducing the turnaround time and

improving the diagnostic resolution. By visualising the replication of individual cells, the

response of bacteria to antimicrobials can be in principle reduced to one or a few doubling

times of the bacteria.(8–12) However, rapid AST techniques that rely on the area occupied by

the cells as a measure of the growth could mistake a transient increase in cell sizes due to

antimicrobial tolerance as growth.(9) Phenotypic variants, drug accumulation, and growth

phase can also introduce uncertainties in rapid AST assays.(13, 14) Because of these reasons,

existing imaging-based approaches often require at least 2 hours, if not more, to deliver reliable

results, especially for slow-growing and fastidious bacteria.(9, 15–18) While there are ample

examples of rapid AST, the search for an assay that can deliver AST results (1) rapidly in a

point-of-care timeframe, (2) quantitatively with MIC determination, (3) efficiently with a small

inoculum size, and (4) suitable for direct use with primary clinical samples is still ongoing.

A promising strategy for rapidly determining bacteria response to antimicrobials is to

monitor their morphological changes.(19) Bacteria undergo a wide variety of morphological

changes in response to the environment.(20) These changes, such as filamentation, bulging,

and lysis, indicate stress in bacteria and have been applied for investigating the mechanisms of

action of antimicrobials.(7, 21–23) For instance, distinct morphological transformations could

be induced in *Escherichia coli* depending on the type of beta-lactam antimicrobials.(24)

Pseudomonas aeruginosa, which are known to be highly tolerant against beta-lactams, undergo

a transition from rod-shaped to viable spherical cells when treated with meropenem.(25) The

minute change in the bacterial morphology could be indicative of the bacterial response to

antimicrobials before cell replication. While promising examples of bacterial analysis at the

single-cell level have emerged,(26–28) the potential of single-cell morphological analysis has

not been realised for rapid AST and MIC determination due to a lack of effective strategy for

analysing the dynamic, multiparametric morphological features of bacteria.

Herein, we demonstrate a machine learning workflow, termed morphometric

antimicrobial susceptibility test (MorphoAST), for single-cell AST and MIC quantification in

a fraction of the bacterial doubling time. The workflow combines single-cell imaging,

computer vision feature extraction, and supervised learning models for predicting the response

of bacteria to antimicrobials. We measured dynamic morphological features of individual

Klebsiella pneumoniae in the presence of meropenem. We processed the data using time and

concentration differential strategies and trained machine learning models to predict the

antimicrobial response and MIC. The models were validated by cross-validation and left-out

data to access their predictive power. A test dataset with cells from an unseen strain was also

tested to provide an unbiased evaluation of the trained model. The results were reported

according to the CLSI performance standards for antimicrobial susceptibility testing (M100)

guideline.

Methods

Bacteria culture

Klebsiella pneumoniae isolates (listed in table S1) from the CDC & FDA Antimicrobial

Resistance (AR) bank and American Type Culture Collective (ATCC) were grown in Mueller

Hinton Broth (MHB) at 37°C overnight. The next day, cells were sub-cultured in fresh MHB

media for three hours (growth phase) to a density equivalent to $OD_{600} = 0.5$. Bacteria cells (10

µL) were treated with 10 µL meropenem at varying final concentrations of 0 (control), 0.05,

0.5, 5, and 50 µg/mL to ascertain the physiological responses of cells. A subset of these isolates

was treated with finer range of antimicrobial concentrations (0.02, 0.04, 0.016, 0.032, 0.064,

0.125, 0.5, 1, 2 µg/mL) to train the algorithm for quantitative prediction of MIC. The MICs of

the isolates were obtained from the CDC and independently confirmed using the broth

microdilution method according to the CDC guideline.

Bacteria imaging

Bacteria cells and meropenem in liquid media are briefly vortexed before mounting on a 1%

(v/v) UltraPure agarose pad (10 x 20 mm, Invitrogen US) in M9 minimal media (Gibco, US)

on a micro slide covered with a glass coverslip (#1.5, 22 x 30 mm). Bacteria imaging was

performed using a Nikon Ti2-E inverted microscope equipped with a DS-Qi2 CMOS camera

and an Okolab stage-top temperature control chamber (37°C, 5% CO_2). Images were acquired

using a Nikon CFI Plan Apochromat l DM 100X oil objective lens and an external Phase

Contrast (Ph3) module. Each isolate was observed over a period of 75 minutes (15-90 minutes

after mixing with the antimicrobial).

Feature extraction

Time-series images were stacked and corrected for shifts in the time series using the Template

Matching and Slice Alignment ImageJ plugin.[37] Each stack was then analysed using the

MicrobeJ plugin for ImageJ.[29] The plugin, designed for the detection and analysis of

bacterial cells, uses computer vision algorithms to automatically identify a cell and determine

a suite of morphological features, such as its area, length, circularity, and perimeter (see

Supplementary Table 2). The tracking data were summarised as a table and exported as a .csv

file for further analysis.

Data analysis and machine learning

The morphological data were analysed to predict the susceptibility and MIC of the bacterial strain before the average doubling time (~50 minutes). Two differential strategies based on the time-dependent or antibiotic concentration-dependent changes of the bacteria were applied. The processed data were then applied for susceptibility classification and MIC prediction. A schematic of the workflow is shown in Figure 1.

In the time differential (or dynamic) approach, processed image stacks for *Klebsiella pneumoniae* isolates (KP016, KP0140, KP1705, KP0153, and KP0142) at various antimicrobial concentrations were evaluated. The cells were labeled as division (resistant) or no division (susceptible). Data analyses were performed using Python with data analysis and machine learning libraries.[38,39] To explore AST in a sub-doubling time, time-lapse data from the first 35 minutes (corresponding to 15 to 50 minutes after antimicrobial exposure) at a 5-minute interval were analysed. Erroneous data (e.g., missing bacteria) were detected and removed. Feature-changing rates of individual cells were determined by fitting the time-lapse data with the exponential function. The area, aspect ratio, circularity, length, and perimeter were the most relevant dynamic features. The changing rates of these features for each cell, along with the meropenem concentration, time of tracking, the bacterial strain, and the label for growing/not growing, were used to train the K-Nearest Neighbor and artificial neural network (MPLClassifier) models from the scikit-learn machine learning library.[40] For the K-Nearest Neighbor model, k was optimal at 20. The optimal structure of the artificial neural network was found at three hidden layers with five neurons each, even though a simple grid (two hidden layers with two neurons each) was sufficient for well-separated data (e.g., 50-minute data). The accuracy of each model was obtained by averaging the values of 10 runs.

In the concentration differential approach, the effect of each concentration of meropenem on individual cells was determined as a deviation from the population mean of untreated cells at the same time point for each of the 19 features, concentrations, and strains.

Only cells detected within the first three time points and with no missing data were included in

the analysis. If a cell died within the observation timeframe, the subsequent time points were

marked with zero values for the features. The best modeling strategy for the classification of

the normalised data was chosen among Random Forest, Naive Bayes, K-Nearest Neighbor,

and Support Vector Machine based on the accuracy of models trained at the first time point

(Supplementary Table 3). Random Forest classification models were trained on the normalised

dataset for each time point with 5-fold cross-validation and 10 repetitions using the R package

caret.[41] The models with the highest accuracy were retained.

210   In the regression model for MIC prediction, processed image stacks for all 12 *K.*

pneumoniae strains that spanned 4 resistant, 5 susceptible, and 2 intermediate against

meropenem were used to generate a regressor that can predict the MIC. The data were

normalised as described above in the random forest classifier and split into a training-to-

validation ratio of 80:20. The best modeling strategy for the prediction of MIC was chosen

among Linear Regression, Neural Net, Random Forest, K-Nearest Neighbor, Support Vector

Machine, and Gaussian Process Regression based on the lowest root mean square error

(RMSE) and highest coefficient of determination (R^2) for models trained at the first time point

(Supplementary Table 4). The training data were weighted to account for any class-level

imbalance. Random Forest regression models were trained for each time point with 5-fold

cross-validation and 5 repetitions using R packages ranger[42] and caret. [41] Models were

trained against the MIC reported by the CDC for these strains, and the model with the lowest

RMSE and highest R^2 was selected. The final models were tested against the left-out data.

Categorical agreement (CA) and essential agreement (EA) for the cross-validated test as well

as the left-out test were determined. The major error (ME) and very major error (VME) rates

were calculated for the model. The generalisability of the model was further tested against an

unseen strain (KP0139).

Results

The MorphoAST workflow for rapid AST

We developed a machine learning workflow for rapid AST in a sub-doubling time of the bacteria. The MorphoAST workflow started with imaging of individual bacteria under various antimicrobial concentrations (Figure 1A). Cells were grown to a log phase and treated with varying concentrations of meropenem. Cells without any antimicrobials were imaged in the same manner as controls for the experiment. Live bacteria were mounted on an agarose pad to minimise cell movement due to cell motility and Brownian motion. Time-lapse images were taken every 5 minutes. Images were aligned to match the location of individual cells at each frame and then subject to MicrobeJ analysis for extracting morphological features.[29] A total of 21 parameters (Supplementary Table 2) describing the cellular dimension and orientation were generated. The data have been analysed either fitting with an exponential function to extract the feature changing rates (Figure 1B) or normalised against the untreated control to extract the feature differentials (Figure 1C). The data were applied to train and validate machine learning classifiers to predict the antimicrobial response (Figure 1D) and regressors to predict the MIC of the bacterial strain against the antimicrobial (Figure 1E).

Single-cell imaging and feature extraction

A total of 12 *K. pneumoniae* strains bearing various carbapenem resistance genes and sensitivity towards meropenem were monitored over time (Supplemtnary Table 1). Changes in bacterial morphology were observed over 90 minutes post-antimicrobial incubation. Only the first 50 minutes, which is approximately the doubling time of *K. pneumoniae*, was analysed.

Each strain showed a unique morphological response to the varying antimicrobial

concentration based on their MIC. Figure 2A and B show *K. pneumoniae* (KP0142) treated

with and without meropenem. Bulging of the bacteria was only observed in the meropenem

case. Supplementary Figure 1 compares three strains (KP0016 – Susceptible, KP0142 –

Intermediate, and KP0143 – Resistant) at the early time points (see also movies S1-3).

Similarly, at 5 μg/mL meropenem, which is higher than the breakpoint MIC for meropenem,

susceptible and intermediate strains (KP0016 and KP0142) showed noticeable 'bulging' or

protrusion around the center of the cell that was not observed in the resistant strain (KP0143).

To automate the analysis and avoid subjectivity, the MicrobeJ plugin was applied for

extracting morphological features (Figure 2C). The cell behaviours were summarised by

extracting the changing rates of each feature from the data (Figure 2D and Supplementary

Figure 2). Figure 2E shows an example of the distributions of length and area changing rates

of a single strain (KP0142) under various meropenem concentrations. The centroids of the

changing rates varied with the antimicrobial concentration. However, there were large

variations among individual bacteria at the same concentrations and substantial overlaps

between different concentrations. It is challenging to accurately predict the bacteria response

based on one or two features. A statistical approach, specifically machine learning algorithms,

is required to analyse the multiparametric data for improving classification accuracy.

Classification of bacterial response to antimicrobial with dynamic features

We first evaluated the dynamic (or time differential) approach for predicting the bacterial

response to meropenem. The dynamic features of a total of 1338 bacterial cells across various

antimicrobial concentrations were measured. The cells were labeled as 1 (resistant or division)

or 0 (susceptible or no division) at the strain-concentration combination based on the CDC-

reported meropenem MIC for *K. pneumoniae*. Since the behaviors of individual bacteria were

highly diverse, one to seven bacteria under the same strain-concentration combinations were

randomly grouped. The average feature changing rates were calculated for each group. As

shown in the principal component analysis (PCA) plots (Figure 3A), grouping bacteria

substantially reduced the variation within a group and increased the separation between the

'division' and 'no division' groups. The separation between the groups increased with the

number of bacteria in the group. The grouped data were trained and validated using the training

and validation datasets with the K-Nearest Neighbors and artificial neural network classifiers

(Figure 3B). For the 35-min data (i.e., 15-50 minutes of exposure), groups of one bacterium

resulted in an accuracy of 89% and 90% with the K-Nearest Neighbors and artificial neural

network classifiers, respectively. The prediction accuracy was generally improved by

increasing the number of bacteria. With groups of seven bacteria, both classification algorithms

reached over 99% accuracy. The improvement can be understood by a reduction of the

statistical variation of individual cells by averaging data from multiple bacteria.

We then evaluated the accuracy of the classification model with sub-doubling times.

The feature change rates were extracted using different durations of the data (5 to 35 minutes),

corresponding to 15-50 minutes of antimicrobial incubation. Similarly, the data are

summarised and visualised by the PCA plots (Figure 4A). The data with a 5-minute duration

(i.e., between 15 minutes and 20 minutes after antimicrobial exposure) displayed a

considerable variance and overlapped substantially. Nevertheless, the centroids of division and

no division groups were distinct, and the separation improved with data of a longer duration.

Clear separations between the groups could be observed in data with 25-min or 35-min

durations. Again, we trained and validated the data using machine learning classifiers,

including the K-Nearest Neighbors and artificial neural network algorithms (Figure 4B-C).

These algorithms exhibited similar performances and had an accuracy of ~80% with the 5-

minute data. The results reached around 95% with the 25-minute data and over 99% with the 35-minute data. The confusion matrices indicated a false positive rate close to zero and a false negative rate of ~1.1% (Figure 4D). These results support the use of dynamic morphological features, i.e., the time differential approach, for predicting the antimicrobial response in a sub-doubling time.

Susceptibility classification with concentration differential features

We next evaluated the concentration differential approach for predicting antimicrobial susceptibility. A total of 5050 bacterial cells with various antimicrobial concentrations were included in the prediction of survivability and susceptibility to antimicrobial exposure. The deviations of morphological features from the population means of untreated cells were calculated at each time point. Since the bacterial features against various antimicrobial concentrations were measured, the differential data were applied to classify the interpretative categories (i.e., Susceptible 'S', Intermediate 'I', and Resistant 'R') according to the CLSI guideline. Figure 5A shows the PCA plots of the data. The centroids of the S, I, and R groups are separated after 20 minutes of antimicrobial exposure, and there was a considerable amount of overlap. The separation widened when the antimicrobial exposure time increased. We trained and cross-validated the data with various classification algorithms, including Random Forest, Naive Bayes, K-Nearest Neighbor, and Support Vector Machine. The results suggest the Random Forest algorithm outperformed other models (Supplementary Table 3). The categorical agreement (CA), which is the percentage of interpretive agreement between the predicted and true labels, was determined (Figure 5B). The Random Forest model achieved a 97% accuracy in as early as 20 minutes (two-fifths of the doubling time) and reached over 99% accuracy in 40 minutes (four-fifths of the doubling time). Grouping multiple bacteria, however, did not improve the accuracy (Supplementary Figure 3).

Next, the major error (ME: susceptible classified as resistant), very major error (VME:

resistant classified as susceptible), and the minor errors (mE: susceptible or resistant classified

as intermediate or intermediate classified as susceptible or resistant) were estimated by

examining the confusion matrices (Figure 5C). With 20 minutes of meropenem exposure, the

Random Forest model had a ME of 3.8%, VME of 0.7%, and a cumulative mE of 3.2%. The

performance of the Random Forest model further improved with an increase in the

antimicrobial exposure time. At 40 minutes of exposure, this model had a ME of 1.4%, VME

of 0.2%, and a cumulative mE of 1%.   These accuracy and error rates are within the FDA

acceptable limits (CA>90%, VME <1.5%, and ME<3%), supporting the use of morphological

features for rapid single-cell AST.

Minimum inhibitory concentration prediction in sub-doubling times

We further evaluated the use of the concentration differential data for MIC prediction. A

Random Forest regressor was trained from a total of 28,312 cells across the concentrations.

The model, including historic data (i.e., earlier time points), had a cumulative improvement in

performance measured by the RMSE and R^2 values with increased exposure time (Figure 6A).

The Random Forest regressor predicted the MIC in the 5-fold cross-validated training dataset

with an RMSE of 0.7 and an R^2 of 0.96. The performance of the model was further assessed by

the 20% validation dataset, which comprised 7,082 cells. The CDC reported MIC and predicted

MIC based on the Random Forest regressor are compared in Figure 6B (see also Supplementary

Figure 4). The data showed a strong correlation, and the regressor collectively predicted the

MIC within plus or minus, one two-fold dilution for all strains, resulting in a 100% essential

agreement (EA) with 30-50 minutes of antimicrobial exposure. The predicted and reported

MIC labels were also compared (Figure 6C). The model correctly predicted 99% of all the

resistant bacteria in as few as 20 minutes (Supplementary Figure 4). The performance of the

model increased with the antimicrobial exposure time and archived >97% CA with 0.6% ME

and 0% VME in 50 minutes.

In addition to the validation data, the model was tested against a test dataset with unseen

cells that were not in the training and validation datasets (Figure 6D-E). The CDC reported

MIC for the unseen strain (KP0139) is >8 µg/mL, i.e., resistant. Within 20 minutes of

meropenem treatment, the model correctly classifies 82.1% of cells as resistant with a VME

(i.e., predicted to be susceptible) of 0.5%. When the incubation time was increased to 50

minutes, the categorical agreement increases to 91.9% with a VME of only 0.1%.

Role of funders

The funders of this study were not involved in the study design, data collection, analysis,

interpretation and writing of the report.

Discussion

This study demonstrated a rapid workflow for determining the antimicrobial susceptibility of bacteria using time-lapse single-cell imaging, computer vision, and machine learning models. By tracking the antimicrobial-induced morphological changes of individual cells, the MorphoAST workflow predicted the susceptibility category and MIC with high accuracy in a fraction of the doubling time of the bacteria. We evaluated two data processing strategies based on the time and concentration differentials. Both strategies successfully predicted the susceptibility in as few as 20 minutes (two-fifths of the doubling time) of antimicrobial exposure with high accuracy. The concentration differential approach with the random forest classifier achieved an overall better accuracy and resolution for predicting the susceptibility category and MIC (CA >97% in as few as 20 minutes). Therefore, the concentration differential approach should be applied whenever possible. The time differential approach, however, could be useful when only a small number of bacteria is available (e.g., direct detection of bacteria from clinical samples) and measuring multiple antimicrobial concentrations is challenging (e.g., a point-of-care device that detects only a small number of conditions). Our data also suggested that the prediction accuracy was generally improved with the antimicrobial exposure time, the number of bacteria being analysed, and the number of testing conditions. These results underscore important considerations and tradeoffs in the design of the single-cell AST workflow and provide examples of future experimental designs.

Our results showed that morphological features are useful information for rapid AST. In particular, the formation of bulges among *K. pneumoniae* with varying meropenem MIC are distinct and allowed us to classify susceptibility in a sub-doubling time. Bulge formation has been associated with the disruption of peptidoglycan and cell wall degradation before cell lysis.[24] Prior studies have also reported the formation of cell-wall deficient spheroplasts in carbapenem-tolerant *K. pneumoniae* strains, after exposure to meropenem.[30,31] While

antimicrobial-induced shape changes in bacterial cells are not fully understood in general,

recent quantitative modeling reveals potential advantages of this physiological adaptation,

which include decreasing antimicrobial influx and diluting intracellular antimicrobials, leading

to a higher tolerance.[32] The size and shape regulation differs from one organism to another,

but importantly, provides a general reflection of their response in a short time frame.[33] As

demonstrated in this study, MorphoAST provides a workflow for utilising the potential of

morphological features for rapid AST. The workflow should be compatible with various

morphological features, such as filamentation, bulging, and lysis, induced by specific

antimicrobial-bacteria combinations. The methodical procedure identifies and distinguishes

key morphological features that can be captured before and after the treatment (time

differential) or with various antimicrobial dilutions (concentration differential).

A major advantage of the MorphoAST workflow is the short turnaround time. As the

morphological features can be captured in a fraction of the doubling time, it dramatically

reduces the assay time compared to the standard phenotypic AST (e.g., broth microdilution),

which typically requires one or more days. The approach bypasses the requirement of cell

replication in other single-cell AST techniques.[34] This characteristic will be particularly

useful for diagnosing slow-growing and difficult-to-culture bacteria in normal laboratory

conditions. Compared to other single-cell analysis techniques that measure the nanoscale

motion of bacteria and metabolic activities,[26–28] the MorphoAST workflow requires only a

small inoculum size and a relatively simple setup consisting mainly of agar pads and a

microscope. As imaging is common in AST[35,36] and low-cost microscopes are readily

available, the MorphoAST workflow can be integrated with existing systems and implemented

in a variety of settings. These advantages and characteristics will potentially increase the utility

of MorphoAST for direct sample AST testing, especially in diseases like sepsis where the

bacteria load in the blood is typically very low. The small inoculum will also considerably

reduce the time to AST results at the point of care.

MorphoAST will potentially accelerate microbiological analysis for combating

multidrug-resistant bacteria, such as carbapenemase-producing *Enterobacteriaceae*. In both

left-out data and unseen cells, our results demonstrate outstanding categorical agreement (97%)

and essential agreement (100%), which are above the FDA acceptable level ($\geq$90%) for new

AST devices. The machine learning models consistently obtained a low VME (0-0.1%), which

is substantially better than the acceptable VME rate ($\leq$1.5%). A low VME rate is particularly

important as missing a resistant strain could result in the prescription of an ineffective

antimicrobial for the patient. Further investigation will be required to elucidate the mechanisms

of action of meropenem and its effects on other bacteria. Toward the clinical translation of

MorphoAST, additional bug-drug combinations should be tested to evaluate the accuracy of

the approach in other settings. Automation of the drug mixing and bacteria trapping steps will

shorten the initial preparation time and capture changes in bacterial morphologies at earlier

time points. Implementing the workflow in an integrated, low-cost imaging system, instead of

a microscope, will also be important for disseminating the workflow for managing a wide

spectrum of infectious diseases.

Contributors

Conceptualisation: S.Y., P.K.W., K.C.T., N.R.M., M.R. which was refined by input from J.C.L and K.E.M. Methodology: S.Y., P.K.W., K.C.T., N.R.M., M.R. Investigation: K.C.T., N.R.M., M.R., E.Z. Imaging: K.C.T. Data analysis: K.C.T., N.R.M., M.R., Z.Q., O.E. Funding acquisition: S.Y., P.K.W., J.C.L. Project administration: S.Y., P.K.W., J.C.L. Supervision: S.Y., P.K.W., J.C.L. Writing – original draft: P.K.W., K.C.T., N.R.M., M.R. All authors discussed the results and contributed to reviewing and editing the manuscript

Declaration of interests

Authors declare that they have no competing interests.

Acknowledgment

We acknowledge technical support from Youngbin Lim from the Stanford Cell Sciences Imaging Facility. We are grateful for the insightful discussions and computational support from Dean Deng from Stanford University. This work was performed, in part, at the Stanford Cell Sciences Imaging Facility of Stanford University.

Data sharing and materials availability

The data, code, and materials that support the findings of this study will be openly available in a public repository and from the corresponding authors PW and SY on request.

References

456   [1]   Nordmann P, Naas T, Poirel L. Global spread of carbapenemase producing

457   Enterobacteriaceae. Emerg Infect Dis 2011;17:1791–8.

458   https://doi.org/10.3201/eid1710.110655.

459   [2]   Durante-Mangoni E, Andini R, Zampino R. Management of carbapenem-resistant

460   Enterobacteriaceae infections. Clin Microbiol Infect 2019;25:943–50.

461   https://doi.org/10.1016/j.cmi.2019.04.013.

462   [3]   Jernigan JA, Hatfield KM, Wolford H, Nelson RE, Olubajo B, Reddy SC, et al.

463   Multidrug-Resistant Bacterial Infections in U.S. Hospitalized Patients, 2012–2017. N

464   Engl J Med 2020;382:1309–19. https://doi.org/10.1056/nejmoa1914433.

465   [4]   Jorgensen JH, Ferraro MJ. Antimicrobial susceptibility testing: A review of general

466   principles and contemporary practices. Clin Infect Dis 2009;49:1749–55.

467   https://doi.org/10.1086/647952.

468   [5]   Vasala A, Hytönen VP, Laitinen OH. Modern tools for rapid diagnostics of

469   antimicrobial resistance. Front Cell Infect Microbiol 2020;10:308.

470   https://doi.org/10.3389/fcimb.2020.00308.

471   [6]   Tjandra KC, Ram-Mohan N, Abe R, Hashemi MM, Lee JH, Chin SM, et al. Diagnosis

472   of Bloodstream Infections: An Evolution of Technologies towards Accurate and Rapid

473   Identification and Antibiotic Susceptibility Testing. Antibiotics 2022;11.

474   https://doi.org/10.3390/antibiotics11040511.

475   [7]   Baltekin Ö, Boucharin A, Tano E, Andersson DI, Elf J. Antibiotic susceptibility testing

476   in less than 30 min using direct single-cell imaging. Proc Natl Acad Sci U S A

477   2017;114:9170–5. https://doi.org/10.1073/pnas.1708558114.

[8]   Li H, Torab P, Mach KE, Surrette C, England MR, Craft DW, et al. Adaptable

microfluidic system for single-cell pathogen classification and antimicrobial

susceptibility testing. Proc Natl Acad Sci 2019:201819569.

https://doi.org/10.1073/pnas.1819569116.

[9] Choi J, Yoo J, Lee M, Kim EG, Lee JS, Lee S, et al. A rapid antimicrobial

susceptibility test based on single-cell morphological analysis. Sci Transl Med 2014;6.

https://doi.org/10.1126/scitranslmed.3009650.

[10] Lu Y, Gao J, Zhang DD, Gau V, Liao JC, Wong PK. Single cell antimicrobial

susceptibility testing by confined microchannels and electrokinetic loading. Anal

Chem 2013;85:3971–6. https://doi.org/10.1021/ac4004248.

[11] Zhang F, Jiang J, McBride M, Yang Y, Mo M, Iriya R, et al. Direct Antimicrobial

Susceptibility Testing on Clinical Urine Samples by Optical Tracking of Single Cell

Division Events. Small 2020;16:1–10. https://doi.org/10.1002/smll.202004148.

[12] Choi J, Jung YG, Kim J, Kim S, Jung Y, Na H, et al. Rapid antibiotic susceptibility

testing by tracking single cell growth in a microfluidic agarose channel system. Lab

Chip 2013;13:280–7. https://doi.org/10.1039/c2lc41055a.

[13] Łapińska U, Voliotis M, Lee KK, Campey A, Stone MRL, Tuck B, et al. Fast bacterial

growth reduces antibiotic accumulation and efficacy. Elife 2022;11:1–33.

https://doi.org/10.7554/elife.74062.

[14] Brauner A, Fridman O, Gefen O, Balaban NQ. Distinguishing between resistance,

tolerance and persistence to antibiotic treatment. Nat Rev Microbiol 2016;14:320–30.

https://doi.org/10.1038/nrmicro.2016.34.

[15] Veses-Garcia M, Antypas H, Löffler S, Brauner A, Andersson-Svahn H, Richter-

Dahlfors A. Rapid phenotypic antibiotic susceptibility testing of uropathogens using

optical signal analysis on the nanowell slide. Front Microbiol 2018;9:1–10.

https://doi.org/10.3389/fmicb.2018.01530.

[16] Matsumoto Y, Sakakihara S, Grushnikov A, Kikuchi K, Noji H, Yamaguchi A, et al. A

microfluidic channel method for rapid drug-susceptibility testing of Pseudomonas

aeruginosa. PLoS One 2016;11:1–17. https://doi.org/10.1371/journal.pone.0148797.

[17] Avesar J, Rosenfeld D, Truman-Rosentsvit M, Ben-Arye T, Geffen Y, Bercovici M, et

al. Rapid phenotypic antimicrobial susceptibility testing using nanoliter arrays. Proc

Natl Acad Sci U S A 2017;114:E5787–95. https://doi.org/10.1073/pnas.1703736114.

[18] Kalashnikov M, Mueller M, McBeth C, Lee JC, Campbell J, Sharon A, et al. Rapid

phenotypic stress-based microfluidic antibiotic susceptibility testing of Gram-negative

clinical isolates. Sci Rep 2017;7:1–10. https://doi.org/10.1038/s41598-017-07584-z.

[19] Bourne CR. Bacterial Growth Mindset: Structural Plasticity in Defense Systems.

Structure 2021;29:97–8. https://doi.org/10.1016/j.str.2021.01.007.

[20] Yang DC, Blair KM, Salama NR. Staying in Shape: the Impact of Cell Shape on

Bacterial Survival in Diverse Environments. Microbiol Mol Biol Rev 2016;80:187–

203. https://doi.org/10.1128/mmbr.00031-15.

[21] Bruus H. Acoustofluidics 10: Scaling laws in acoustophoresis. Lab Chip

2012;12:1578. https://doi.org/10.1039/c2lc21261g.

[22] Zahir T, Camacho R, Vitale R, Ruckebusch C, Hofkens J, Fauvart M, et al. High-

throughput time-resolved morphology screening in bacteria reveals phenotypic

responses to antibiotics. Commun Biol 2019;2:1–13. https://doi.org/10.1038/s42003-

019-0480-9.

[23] Nonejuie P, Burkart M, Pogliano K, Pogliano J. Bacterial cytological profiling rapidly

identifies the cellular pathways targeted by antibacterial molecules. Proc Natl Acad Sci

U S A 2013;110:16169–74. https://doi.org/10.1073/pnas.1311066110.

[24] Yao Z, Kahne D, Kishony R. Distinct Single-Cell Morphological Dynamics under

Beta-Lactam Antibiotics. Mol Cell 2012;48:705–12.

https://doi.org/10.1016/j.molcel.2012.09.016.

[25] Monahan LG, Turnbull L, Osvath SR, Birch D, Charles IG, Whitchurch CB. Rapid

conversion of Pseudomonas aeruginosa to a spherical cell morphotype facilitates

tolerance to carbapenems and penicillins but increases susceptibility to antimicrobial

peptides. Antimicrob Agents Chemother 2014;58:1956–62.

https://doi.org/10.1128/AAC.01901-13.

[26] Rosłoń IE, Japaridze A, Steeneken PG, Dekker C, Alijani F. Probing nanomotion of

single bacteria with graphene drums. Nat Nanotechnol 2022;17:637–42.

https://doi.org/10.1038/s41565-022-01111-6.

[27] Scherer B, Surrette C, Li H, Torab P, Kvam E, Galligan C, et al. Digital electrical

impedance analysis for single bacterium sensing and antimicrobial susceptibility

testing. Lab Chip 2021;21:1073–83. https://doi.org/10.1039/d0lc00937g.

[28] Kaushik AM, Hsieh K, Mach KE, Lewis S, Puleo CM, Carroll KC, et al. Droplet-

Based Single-Cell Measurements of 16S rRNA Enable Integrated Bacteria

Identification and Pheno-Molecular Antimicrobial Susceptibility Testing from Clinical

Samples in 30 min. Adv Sci 2021;8:1–14. https://doi.org/10.1002/advs.202003419.

[29] Ducret A, Quardokus EM, Brun Y V. MicrobeJ, a tool for high throughput bacterial

cell detection and quantitative analysis. Nat Microbiol 2016;1:1–7.

https://doi.org/10.1038/nmicrobiol.2016.77.

[30] Cross T, Ransegnola B, Shin J-H, Weaver A, Fauntleroy K, VanNieuwenhze MS, et

al. Spheroplast-Mediated Carbapenem Tolerance in Gram-Negative Pathogens.

Antimicrob Agents Chemother 2019;63. https://doi.org/10.1128/AAC.00756-19.

[31] Murtha AN, Kazi MI, Schargel RD, Cross T, Fihn C, Cattoir V, et al. High-level

carbapenem tolerance requires antibiotic-induced outer membrane modifications.

PLoS Pathog 2022;18:1–19. https://doi.org/10.1371/journal.ppat.1010307.

[32] Ojkic N, Serbanescu D, Banerjee S. Antibiotic Resistance via Bacterial Cell Shape-

Shifting. MBio 2022. https://doi.org/10.1128/mbio.00659-22.

[33] Ojkic N, Banerjee S. Bacterial cell shape control by nutrient-dependent synthesis of

cell division inhibitors. Biophys J 2021;120:2079–84.

https://doi.org/10.1016/j.bpj.2021.04.001.

[34] Pancholi P, Carroll KC, Buchan BW, Chan RC, Dhiman N, Ford B, et al. Multicenter

evaluation of the accelerate PhenoTest BC kit for rapid identification and phenotypic

antimicrobial susceptibility testing using morphokinetic cellular analysis. J Clin

Microbiol 2018;56. https://doi.org/10.1128/JCM.01329-17.

[35] Song D, Lei Y. Mini-review: Recent advances in imaging-based rapid antibiotic

susceptibility testing. Sensors and Actuators Reports 2021;3.

https://doi.org/10.1016/j.snr.2021.100053.

[36] Salido J, Bueno G, Ruiz-santaquiteria J, Cristobal G. A review on low-cost

microscopes for Open Science 2022:1–14. https://doi.org/10.1002/jemt.24200.

[37] Qingzong T. Template Matching and Slice Alignment n.d.

https://sites.google.com/site/qingzongtseng/template-matching-ij-plugin#credit

(accessed June 28, 2022).

[38] Wes McKinney. Python for Data Analysis. vol. 71. 2017.

https://doi.org/10.1097/00007890-200105270-00005.

[39] Harris CR, Millman KJ, van der Walt SJ, Gommers R, Virtanen P, Cournapeau D, et

al. Array programming with NumPy. Nature 2020;585:357–62.

https://doi.org/10.1038/s41586-020-2649-2.

[40]   Pedregosa F, Michel V, Grisel O, Blondel M, Prettenhofer P, Weiss R, et al. Scikit-

learn: Machine Learning in Python. J Mach Learn Res 2011;12:2825–30.

https://doi.org/10.5555/1953048.2078195.

[41]   Kuhn M. Building Predictive Models in R Using the caret Package. J Stat Softw

2008;28:1–26. https://doi.org/10.18637/jss.v028.i05.

[42]   Wright MN, Ziegler A. ranger : A Fast Implementation of Random Forests for High

Dimensional Data in C++ and R. J Stat Softw 2017;77.

https://doi.org/10.18637/jss.v077.i01.

Figures

Figure 1. A schematic flowchart of the single-cell MorphoAST workflow for susceptibility classification and minimum inhibitory concentration (MIC) prediction. (A) The workflow starts with time-lapse imaging of individual bacteria under various antimicrobial concentrations. The morphological features of the bacteria are extracted automatically using MicrobeJ, an ImageJ plugin. (B) In the time differential approach, the feature changing rates are extracted by exponential curve fitting from the time-lapse data. (C) In the concentration differential approach, all features extracted from MicrobeJ are normalised against the feature mean of the control population. All datasets are visualised and separated into training and validation sets. (D-E) The feature-changing rate datasets are applied to train an artificial neural network to create a classification model. In parallel, the normalised dataset is trained using a random forest classifier with cross-validation. The trained models are then validated with the validation dataset for analysing the prediction accuracy for susceptibility classification. (F) A Random Forest regression model is generated by the normalised data. The regressor model is validated through cross-validation and tested against left-out data and unseen cells for evaluating categorical agreement and essential agreement of the assay.

Figure 2. Representative results of single-cell MorphoAST of *K. pneumoniae* treated with meropenem, an intravenous β-lactam antimicrobial. (A) Time-lapse images of *K. pneumoniae* (KP0142) treated with (top) 0 and (bottom) 50 μg/ml meropenem. Time-lapse imaging was performed 15 minutes after mixing with meropenem at a 5-minute interval. The MIC of the strain was 2 μg/ml, and cell division was completely inhibited at 50 μg/ml. Bulging the cells was observed only in the meropenem case. (B) Zoom-in views of KP0142 treated with 50 μg/ml meropenem. (C) Morphological analysis with MicrobeJ for extracting cell features, such as area, length, width, circularity, curvature, sinuosity, angularity, and solidity. (D) Examples of two bacterial growth curves for extracting the area changing rate. (E) Distributions of area and length changing rates of individual *K. pneumoniae* exposed to various meropenem concentrations.

Figure 3. Classification accuracy of bacteria groups using the time differential approach. (A) Principal component analysis plots of dynamic features of 1338 bacteria from various bacteria strains and antimicrobial concentrations. The data are labeled as 1 (division or resistant) or 0 (no division or susceptible). Bacteria are randomly selected from the same strain-concentration combination to form a group, and the dynamic features are averaged for groups of 1, 3, 5, and 7 bacteria. (B-C) Prediction accuracy was obtained by (B) K-Nearest Neighbors and (C) artificial neural network algorithms for different group sizes. For single bacteria, the accuracy was slightly higher for the artificial neural network (~90%). With groups of seven bacteria, both models obtained over 99% accuracy. The accuracy values are an average of 10 repetitions. (D) Evolution of confusion matrices of neural network classifiers with groups of 1 to 7 bacteria. Both the major error (true susceptible and predicted resistant) and the very major error (true resistant and predicted susceptible) were reduced with an increasing number of bacteria.

Figure 4. Classification accuracy of sub-doubling time susceptibility prediction using the time differential approach. (A) Principal component analysis plots of feature changing rates with 20-50 minutes of antimicrobial exposure, which correspond to 5 to 35 minutes duration of dynamic data. The data are labeled as 1 (division or resistant) or 0 (no division or susceptible) at the strain-concentration combination with groups of seven bacteria. (B-C) Prediction accuracy was obtained by (B) K-Nearest Neighbors and (C) artificial neural network algorithms for different durations of data. The accuracy was approximately 80% for a 5-min duration, and the value improved to over 99% accuracy with a 35-minute duration of data. The accuracy values are an average of 10 repetitions. (D) Evolution of confusion matrices of neural network classifiers with groups of 20-50 minutes of antimicrobial exposure. Both the major error (true susceptible and predicted resistant) and the very major error (true resistant and predicted susceptible) were reduced with the duration of the data.

Figure 5. Classification accuracy of sub-doubling time susceptibility prediction using the concentration differential approach. (A) Principal component analysis plots of the normalised data with 20-50 minutes of antimicrobial exposure. The data are labeled based on the susceptibility to meropenem - red for resistant, black for intermediate, and green for susceptible. (B) Prediction accuracy of random forest classifiers with different durations of antimicrobial exposure. Median accuracy was determined using 5-fold cross-validation and 10 repetitions. Multiclass prediction accuracy was >96% as early as 15 minutes of antimicrobial exposure with increasing accuracy over time. (C) Evolution of confusion matrices of Random Forest classifiers with groups of 20-50 minutes of antimicrobial exposure. Resistant bacteria were classified with 98.8% accuracy in 20 minutes. The accuracy increased to 99.7% in 50 minutes with a very major error of 0.2% and a major error of 1.1%.

648
649

Figure 6. Minimum inhibitory concentration prediction with a random forest regressor using the concentration differential approach. (A) Evaluation of the regression model with root-mean-square error (RMSE) and R-squared values. The regression model trained using 80% of the data was assessed for the lowest RMSE and highest R-squared values. With historical data and accumulated time, the model performs better as seen in both metrics. (B) Mode of predicted MIC for each strain with the 20% left-out data in comparison to the CDC reported MIC for each strain. The regressor achieved 100% essential agreement in the predicted mode MIC within 50 minutes. (C) Categorical classification of the cells from the 20% leave out based on the predicted MIC 20, 30, 40, and 50 minutes. After 50 minutes of exposure to meropenem, the model classifies susceptibility with a ME of 0.6%, VME of 0%, and mE of 2.2%. (D) MIC prediction and susceptibility classification of an unseen strain KP0139 with 1,317 cells imaged across multiple concentrations of meropenem. The model accurately classifieds ~1,080 (82.1%) cells KP0139 as resistant to meropenem within 20 minutes of exposure with a VME of 0.5%. This improves to 91.9% of cells being accurately classified as resistant within 50 minutes with a VME of 0.1%. (E) Histogram of the log2 dilution of the ratio between reported MIC (MIC_{rep}) and CDC reported MIC (MIC_{pred}). Within 50 minutes, 96% of the cells were predicted to have a MIC within plus or minus, one two-fold dilution, earning a 96% essential agreement for a completely unseen strain.