SLIPMAT: a pipeline for extracting tissue-specific spectral profiles from ^{1}H MR spectroscopic imaging data

Olivia Vella[a], Andrew P. Bagshaw[a] and Martin Wilson[a,*]

[a] Centre for Human Brain Health and School of Psychology, University of Birmingham, Birmingham, UK.

[*] Corresponding author. E-mail: wilsonmp@bham.ac.uk

1 Abstract

^{1}H Magnetic Resonance Spectroscopy (MRS) is an important non-invasive tool for

measuring brain metabolism, with numerous applications in the neuroscientific and clinical

domains. In this work we present a new analysis pipeline (SLIPMAT), designed to extract

high-quality, tissue-specific, spectral profiles from MR spectroscopic imaging data (MRSI).

Spectral decomposition is combined with spatially dependant frequency and phase

correction to yield high SNR white and grey matter spectra without partial-volume

contamination. A subsequent series of spectral processing steps are applied to reduce

unwanted spectral variation, such as baseline correction and linewidth matching, before

direct spectral analysis with machine learning and traditional statistical methods. The method

is validated using a 2D semi-LASER MRSI sequence, with a 5-minute duration, from data

acquired in triplicate across 8 healthy participants. Reliable spectral profiles are confirmed

with principal component analysis, revealing the importance of total-choline and scyllo-

inositol levels in distinguishing between individuals – in good agreement with our previous

work. Furthermore, since the method allows the simultaneous measurement of metabolites

in grey and white matter, we show the strong discriminative value of these metabolites in

both tissue types for the first time. In conclusion, we present a novel and time efficient MRSI

acquisition and processing pipeline, capable of detecting reliable neuro-metabolic

differences between healthy individuals, and suitable for the sensitive neurometabolic

profiling of in-vivo brain tissue.

Keywords: neurochemical; metabolism; MR spectroscopy; MRS; spant; machine learning.

1 Introduction

Proton (^{1}H) Magnetic Resonance Spectroscopy (MRS) is an established technique for the non-invasive measurement of metabolism, offering unique insight into brain function and health. While widely used clinically for characterisation of healthy and pathological brain tissue (Oz et al., 2014), MRS can also be used to investigate functional changes in brain metabolites. Using MRS, the levels of primary excitatory and inhibitory neurotransmitters: glutamate and GABA, have been shown to modulate in response to motor learning tasks (Kolasinski et al., 2019), visual stimulation (Bednařík et al., 2015) and other experimental paradigms (Mullins, 2018) – demonstrating direct links between neurotransmitter levels; brain activation and plasticity. Glutamate responses have also been recently shown to distinguish between perceived and unperceived visual stimuli in primary visual cortex (DiNuzzo et al., 2022), crucially in the absence of neurovascular changes measured with fMRI. The specific role and action of other MRS-detectable molecules are less well understood but equally implicated in cognitive processes. For example, N-Acetylaspartic acid (NAA), one of the most abundant metabolites in the human brain, has been shown to relate to fluid intelligence and creativity (Jung et al., 2009; Nikolaidis et al., 2017).

Whilst short term alterations in metabolites such as glutamate, GABA and lactate are reasonably well characterised in response to basic visual stimuli and motor tasks over periods of minutes, longer term changes are only beginning to be established. Recent work suggests that glutamate in the lateral prefrontal cortex accumulates over period of hours in response to hard cognitive work (Wiehler et al., 2022), highlighting an important relationship between cognitive fatigue; sleep and neuro-metabolism. Alternatively, we may query which aspects of neuro-metabolism are reliably different between individuals over a longer time-period of months – with recent work demonstrating the discriminative value of choline-containing metabolites and scyllo-inositol (Wu et al., 2022).

Despite the historical and recent scientific insights derived from MRS measures of neuro-metabolism, the technique remains relatively underused in the field of cognitive neuroscience – particularly when compared to other MR-based methods such as functional-MRI and diffusion-MRI. Difficulties in obtaining high data quality; accurate signal localisation and robust analysis methods likely explain this discrepancy. However the wider availability of modern methods established by the MRS research community is expected to address these technical challenges (Maudsley et al., 2021; Near et al., 2021; Öz et al., 2021; Wilson et al., 2019).

In this work, we present a novel analysis pipeline designed to ameliorate several technical challenges associated with obtaining high quality neuro-metabolic spectral profiles from

1. MRS. The most common MRS acquisition technique (single voxel MRS) is known to suffer
2. from partial volume effects, making it difficult to avoid spectral contamination from unwanted
3. tissue types. For instance, a 2 cm sided voxel placed in cortical grey matter will always
4. contain a degree of white matter tissue, and variations in cortical thickness will introduce
5. unwanted variance in the spectral data due to the metabolic differences between the two
6. tissue types. To address this limitation, we employ a 2D semi-LASER magnetic resonance
7. spectroscopic imaging (MRSI) sequence to acquire spectral data from an axial slice just
8. above the corpus collosum. These spectra, containing a mixture of white and grey matter
9. contributions, are first aligned in terms of phase and frequency (Wilson, 2019), before
10. applying spectral decomposition (Goryawala et al., 2018) to obtain "pure" white and grey
11. matter spectra with high SNR and spectral resolution. A series of spectral processing steps
12. are then applied to minimise the expected spectral variations related to experimental, rather
13. than metabolic, variability. Finally, the processed white and grey matter spectra may be
14. investigated separately, or in combination, to form a neuro-metabolic spectral profile –
15. amenable for further analysis with machine-learning or traditional statistical approaches. We
16. refer to this new analysis pipeline as SLIPMAT: SpectraL Image Processing for Metabolic
17. Analysis of Tissue-types, since metabolically distinct tissue types are extracted from a single
18. MRSI dataset. The method is applied to MRSI data and shown to extract high quality neuro-
19. metabolic spectral profiles, capable of distinguishing between individuals when combined
20. with machine learning – in good agreement with our previous work (Wu et al., 2022).

Methods

SLIPMAT algorithm

23. The objective of the SLIPMAT method is to extract high quality white and grey matter
24. spectra for analysis with machine-learning or conventional techniques. In practice, this
25. requires MRSI and volumetric MRI datasets to be acquired from the same participant,
26. ensuring the two datasets are spatially consistent. The simplest approach, as used here, is
27. to ensure both scans are acquired in the same session and the participant's head remains
28. static throughout the session.

29. **Step 1: intra-scan spectral correction and tissue decomposition**

30. The first analysis step involves extracting the useful region of interest (ROI) in the MRSI
31. dataset, for instance excluding voxels outside the pre-localised volume of interest (VOI)
32. region or brain. A reference spectrum is then defined for frequency and phase alignment.
33. Here, we used a single spectrum central to the acquisition region to minimise the potential
34. for scalp lipid contamination. However, in more challenging brain regions, an "ideal"

reference spectra could be simulated to match the acquisition protocol to ensure perfect

baseline, phasing, lineshape and SNR. The frequency and phase of all other spectra within

the ROI are adjusted to match the reference using the RATS method (Wilson, 2019).

Following alignment, spectra are intensity normalised based on the sum of the metabolite-

rich spectral data points between 4 and 0.2 ppm to reduce the influence of B1

inhomogeneity. Since broad signal components, such as residual water tails, can heavily

influence normalisation, the Asymmetric Least Squares (ALS) baseline correction method

(Eilers and Boelens, 2005) is applied prior to summation. Note, the baseline corrected

spectra are only used to derive the normalisation scaling values, which are subsequently

applied to the uncorrected spectra.

In parallel, the volumetric MRI is segmented into tissue types using standard tools (Zhang et

al., 2001), before being spatially co-registered to the MRSI data – assuming no movement

between the MRI and MRSI scans. Knowledge of the proportion of white and grey matter

contributing to each voxel allows the spectra of the two tissue types to be determined with

simple linear algebra (Goryawala et al., 2018) – eliminating partial volume effects and

significantly boosting the spectral SNR relative to individual MRSI voxels. These processing

steps are summarised graphically in Fig. 1.

Fig. 1. The initial phase of the SLIPMAT method, resulting in a grey and white matter spectral pair for further processing.

At this stage of the algorithm, two high quality tissue-specific spectra have been derived from the MRSI data and could be analysed with conventional fitting approaches (Near et al., 2021).

Step 2: inter-scan consistency correction and scaling

Once the tissue-specific spectra have been determined, a further stage of frequency and phase correction is applied (Wilson, 2019) to ensure consistency between all MRS scans involved in subsequent analysis. In our data the mean spectrum was calculated across all decomposed white and grey matter spectra and used as a reference. However, a good quality spectrum taken from a single scan, or a simulated spectrum, could also be used. A simple zero-order baseline correction step is applied, subtracting any intensity offset from zero as measured from a flat spectral region (0 to -1 ppm). A variable degree of gaussian line broadening is also applied to each spectrum to reduce the influence of inconsistent shimming between spectra. For our data, the largest FWHM of the tNAA resonance was 0.06004 ppm, therefore all spectra were broadened to 0.061 ppm.

Zero-filling to twice the original length is performed to improve spectral SNR (Bartholdi and Ernst, 1973), before discarding the imaginary component and removing spectral points outside the metabolite-rich spectral region (4 to 0.2 ppm). ALS baseline correction is applied to eliminate signals from residual water or broad out-of-volume scalp lipids. Finally, spectra are scaled by the summation of spectral data points and white / grey matter spectra from the same scan and may be used individually or concatenated to form a single feature vector for further analysis (Fig. 2).

2 *Fig. 2. The final phase of the SLIPMAT method, designed to produce high quality spectra*

3 *with consistent linewidth, baseline, phase and frequency offsets between scans.*

5 All spectral processing and co-registration steps are implemented in the open-source

6 spectral analysis package spant (Wilson, 2021a) and image segmentation is performed with

7 FSL (Woolrich et al., 2009).

MR acquisition

A validation MR dataset was acquired at the Centre for Human Brain Health from 8 healthy

adults (2 female, mean age 21 years) with a 3 T Siemens Magnetom Prisma (Siemens

Healthcare, Erlangen, Germany) system using a 32-channel receiver head coil-array. A T1-

weighted MRI scan was acquired with a 3D-MPRAGE sequence: FOV = 208 × 256 × 256

mm, resolution = 1 × 1 × 1 mm, TE / TR = 2 ms / 2000 ms, inversion time = 880 ms, flip

angle = 8 degrees, GRAPPA acceleration factor = 2, 4 minute 54 second scan duration.

Water supressed MRSI data were acquired with 2D phase-encoding: FOV = 160 × 160 × 15

mm, nominal voxel resolution 10 × 10 × 15 mm, TE / TR = 40 ms / 2000 ms, complex data

points = 1024, sampling frequency = 2000 Hz. The MRSI slice was aligned axially in the

subcallosal plane with an approximately 1 mm gap from the upper surface of the corpus

callosum. The semi-LASER method (Scheenen et al., 2008) was used localize a 100 × 100

× 15 mm VOI, central to the FOV, four saturation regions were placed around the VOI

prescribing a 100 × 100 mm interior, and an additional four saturation regions were

positioned to intersect the four corners of the semi-LASER VOI to provide additional scalp

lipid suppression. The total acquisition time for a single MRSI scan was 5 minutes and 6

seconds, and three MRSI datasets were acquired sequentially during the same session to

assess technical repeatability.

Algorithm validation

The SLIPMAT method was applied to 60 voxels taken from the central 8 × 8 grid of each

MRSI dataset, with the 4 corner voxels being excluded due to their proximity to the diagonal

saturation regions. MRSI scans, acquired three times each in 8 participants, resulted in 24

pairs of processed white and grey matter spectra for further analysis. The study was

conducted according to the principles expressed in the Declaration of Helsinki and

participants gave written informed consent before data collection. All T1 scans were defaced

(fsl_deface tool) as a first step for data sharing purposes.

To validate the approach the dataset was interrogated in three ways. Firstly, principal

component analysis (PCA) was applied to the processed 48 white and grey matter spectra to

explore the greatest sources of spectral variance. Secondly, the white and grey matter

spectral pairs were concatenated and PCA was applied to establish the reliability and

uniqueness of an individual's neuro-metabolic profile. Finally, one-way ANOVA was applied

to each spectral data point, with the participant ID used as a factor variable, to further

examine which spectral regions were most unique to an individual. Results were compared

with known spectral differences between white and grey matter, and a previous study

exploring individual differences in neurometabolism to assess consistency.

All MRS analysis, statistics and machine-learning was performed with the R statistical

computing platform (R Core Team, 2021) and analysis scripts used to generate the figures

and tables in this paper are available from https://github.com/martin3141/slipmat_paper

upon publication. MRS and MRI data are available in NIfTI format (Clarke et al., 2022) from

Zenodo in BIDS format, doi: 10.5281/zenodo.7189140.

Results

Compared to other MR techniques, MRS is particularly sensitive to inhomogeneities in the

static field across the VOI. Spatially dependant changes in the frequency and phase of

metabolite resonances results in destructive interference when combing spectra, either

through simple spatial averaging or spectral decomposition. To mitigate these effects, the

SLIPMAT method corrects the spatially dependant frequency and phase inconsistencies

using a method designed to be robust to confounding spectral features, such as baseline

instability (Wilson, 2019). Fig. 3 shows how the SNR and spectral linewidth of decomposed

white and grey matter spectra are significantly improved through the correction of spatial

phase and frequency inconsistencies, as performed by the first phase of the pipeline (Fig 1.).

Average improvements in SNR across all MRSI scans were 50% (174 to 261, paired t-test

p=2.2e-16) accompanied by a 17% reduction (0.058 to 0.048 ppm, paired t-test p=1.6e-4) in

spectral linewidth, measured from the tNAA resonance at 2.01 ppm. These improvements

are visually confirmed by comparing the grey matter lactate doublet at 1.3 ppm – which is

clearly distinct from the noise in the corrected data (Fig 3C) compared to the uncorrected

(Fig 3A).

Fig. 3. Representative improvements in spectral quality following spatial frequency and

phase correction for decomposed grey (A, C) and white (B, D) matter spectra. Corrected

spectra (C, D) show improved SNR and spectral resolution compared to uncorrected (A, B).

Abbreviations: LW - linewidth, SNR - signal to noise ratio, Ins - myo-inositol, tCho – total-

choline, Glu – glutamate, NAAG - N-Acetylaspartylglutamic acid, Lac – lactate.

The metabolite profiles of white and grey matter are known to be different (Pouwels and

Frahm, 1998) and our results support these observations. For example, glutamate and myo-

inositol are stronger in parietal grey matter, whereas NAAG and total-choline are stronger in

parietal white matter (Fig. C vs D).

PCA was applied to the full set of white and grey matter spectra across 8 individuals to

confirm the primary sources of spectral variance arise from genuine neuro-metabolic

differences, rather than spectral artifacts such as residual water signals and scalp lipid

interference. Fig. 4 shows 80% of the variance (PC1) can be explained by differences

between the two tissue types, with the loadings plot (part B) highlighting the expected

spectral regions associated with total-choline (3.2 ppm), NAAG (2.0 ppm) and glutamate (2.4

ppm).

Fig. 4. PCA scores (A) and loadings (B) of all white and grey matter spectra following

processing with SLIPMAT. Clear separation is seen between the two tissue types in

agreement with known metabolic characteristics.

In addition to tissue-types, Fig. 4 also demonstrates consistent spectral differences between

subjects, with some clustering of repeat measures observed in PC2. To explore these

differences further, the grey and white matter spectra for each MRSI dataset were combined

into a single feature vector – allowing both tissue types to contribute to an individual's neuro-

metabolic spectral profile. PCA scores of these combined profiles (Fig. 5 A) shows a clear

separation between subjects – with the variance between repeats (technical variance) being

smaller than the variance between individuals (biological variance). Inspection of the PCA

loadings (Fig. 5 B-E) highlights the importance of total-choline (3.21 ppm) and scyllo-inositol

(3.34 ppm) in discriminating between individuals, in strong agreement with our previous work

(Wu et al., 2022).

Fig. 5. PCA scores (A) and loadings (B-E) of concatenated white and grey matter spectral pairs following processing with SLIPMAT. Loadings for the PCs have been split into white and grey matter and displayed separately to aid spectral interpretation. Grey matter spectral scores are shown in parts B) and D) corresponding to PC 1 and PC 2 respectively. White matter spectral scores are shown in parts C) and E) corresponding to PC 1 and PC 2 respectively.

Unsupervised methods such as PCA are informative for validation purposes, but in practice the neuro-metabolic influence of a known intervention or disease state is more commonly sought. One simple approach is to apply a conventional statistical technique, such as a t-test in the case of two groups, to each spectral data-point to identify key frequencies and metabolites related to the known grouping. Here we use a one-way analysis of variance (ANOVA) to assess which spectral features are the most discriminatory between individuals for both white and grey matter.

Fig. 6. Spectral ANOVA results performed on each frequency domain point independently

and grouped over participants for decomposed (A) white and (B) grey matter spectra. Large

F-statistic values correspond to a strong difference in the spectral intensity between

participants. Dashed horizonal lines represent the Bonferroni corrected significance

threshold (p < 1e-4). Predominant spectral differences between participants are clear in both

tissue types at 3.35 and 3.2 ppm assigned as sIns and tCho respectively.

Strong individual differences in total-choline and scyllo-inositol are apparent from the high

ANOVA F-statistics in those spectral regions (Fig. 6), with larger differences observed in

white matter compared to grey matter. We note the strong similarities between Fig. 6 A) in

this report and Fig. 3 from (Wu et al., 2022), with both plots highlighting similar spectral

regions, despite differences in: scanner hardware and location; participant cohort; acquisition

protocol (single-voxel PRESS vs semi-LASER MRSI) and analysis pipeline (Wilson, 2021b).

We also note, for the first time, that total-choline and scyllo-inositol are strong discriminatory

features in white matter, whereas this was not possible to determine from our previous work,

where a single-voxel was placed in the ACC region containing an average tissue proportion

of 75% grey matter and 25% white matter.

1 Discussion

The primary goal of MRS analysis is to extract biologically meaningful measures, whilst

reducing the influence of variability from technical / experimental factors. In this work, we

demonstrate how the sequential application of discrete spectral processing steps can

achieve this aim. Partial volume effects; frequency and phase inconsistencies; baseline; and

linewidth variability are reduced using a variety of techniques (Eilers and Boelens, 2005;

Goryawala et al., 2018; Wilson, 2019), and combined into a novel and fully automated

analysis pipeline: SLIPMAT. The method is validated using MRSI data from 8 healthy

participants, highlighting subtle spectral features associated with individual differences in

neuro-metabolism in strong agreement with previous work (Wu et al., 2022).

Conventional MRS analysis relies heavily on parametric least-squares fitting of a known

basis-set of metabolite, macromolecule and lipid signals (Near et al., 2021). However, in this

work, we take the approach of extracting spectral "fingerprints" for further analysis with

multivariate and machine learning techniques, a strategy more commonly associated with

analytical chemistry – known as chemometrics (Wehrens, 2020). One significant advantage

of this approach includes the potential to discover spectral features that are not typically

incorporated into a fitting basis-set, e.g., novel metabolites. In addition, this methodology is

relatively simple in design compared to modern MRS fitting algorithms, where differing

algorithmic approaches result in weak agreement between software packages (Craven et al.,

2022). A modular, pipeline-based, design simplifies the evaluation of each individual

processing step, allowing improvements to be made in a more systematic and transparent

fashion.

To the best of the author's knowledge, this study is the first to investigate the efficacy of the

spectral decomposition method for conventional 2D semi-LASER MRSI, demonstrating

excellent spectral resolution and SNR (Fig. 3) when combined with spatial frequency and

phase correction (Wilson, 2019). Whilst 2D semi-LASER is limited to a single slice, the

acquisition protocol is time efficient, taking 5 minutes, compared to a typical whole-brain

MRSI acquisition duration of 26 minutes.

The importance of correcting spatially dependant frequency shifts in MRSI data processing

have been shown previously, with alignment methods including cross-correlation (Ebel and

Maudsley, 2003), water reference data (Maudsley et al., 2006) and magnitude-based

spectral correction (Le Fur and Cozzone, 2014). Here we use the RATS method (Wilson,

2019) – originally developed for correcting frequency and phase inconsistencies for single-

voxel MRS. The method is shown to perform well for MRSI data (Fig. 3) and may be

especially suited to this application due to its robustness to baseline instability – a common

artefact for MRSI datasets. More generally, this work supports the use of correcting spatially

dependant frequency and phase differences in MRSI to improve data quality, in agreement

with recent recommendations (Maudsley et al., 2021).

Single-voxel MRS remains more popular than MRSI due to reduced technical demands

(Wilson et al., 2019), however we envisage our proposed method may be an attractive

alternative for certain applications. In clinical MRS, reduced scan time is particularly crucial,

and the opportunity to acquire high quality white and grey matter spectra, free from partial

volume effects, within the same 5-minute scan represents a compelling advance. Potential

clinical applications include biomarker discovery for early-stage neurodegeneration or mental

illness – where changes in conventional MRI are either absent or occur too late in the

disease process for effective treatment. In cognitive neuroscience, where relevant changes

in neurometabolism are likely to be small, partial volume effects introduce unwanted

variability and have the potential to dominate over genuine metabolic modulations. We

anticipate the proposed method will be a more sensitive tool for detecting small changes in

metabolite levels due to the elimination of these partial volume effects with spectral

decomposition.

Whilst we chose a chemometric approach for spectral processing, we note that conventional

fitting methods could also be applied directly to the decomposed white and grey-matter

spectra (Fig. 1). This has been demonstrated previously with a high-resolution whole-brain

MRSI protocol (Goryawala et al., 2018). A dedicated comparison between chemometric and

conventional fitting approaches is warranted for larger MRS datasets, and we anticipate this

will be aided by recent effects to lower barriers associated with MRS data sharing (Clarke et

al., 2022; Soher et al., 2022).

Potential disadvantages of the proposed method include difficulties in combining datasets

acquired under different experimental conditions. For example, a dataset acquired at 1.5 and

3 T magnetic field strengths would have differing spectral features for any metabolites

containing J-coupled resonances, and these features may dominate over biological

variability. Fitting based methods are expected to be less susceptible to these differences,

since the basis set may be adapted to match the magnetic field strength and specific

acquisition protocol (Near et al., 2013). Future work could include the evaluation of more

modern machine learning approaches, such as deep learning, at reducing the impact of

combining spectral profiles acquired with differing acquisition protocols.

In conclusion, we have developed a novel and time efficient MRSI acquisition and

processing pipeline, capable of detecting reliable neuro-metabolic differences between

healthy individuals. Future work will investigate how rapid tissue-specific neuro-metabolic

profiling can be applied to neuroscientific investigation and clinical studies of the brain.

CRediT authorship contribution statement

Olivia Vella: Validation, Formal analysis, Investigation, Visualisation. Andrew P. Bagshaw:

Conceptualisation, Writing – Review & Editing. Martin Wilson: Conceptualisation,

Methodology, Software, Validation, Formal analysis, Investigation, Data curation, Writing –

Original Draft, Visualisation, Supervision.

Data and code availability

All MR data is available from Zenodo in BIDS format, doi: 10.5281/zenodo.7189140. Code

used to generate the results is available from GitHub,

https://github.com/martin3141/slipmat_paper.

References

Bartholdi, E., Ernst, R.R., 1973. Fourier spectroscopy and the causality principle. Journal of
        Magnetic Resonance (1969) 11, 9–19. https://doi.org/10.1016/0022-2364(73)90076-0
Bednařík, P., Tkáč, I., Giove, F., DiNuzzo, M., Deelchand, D.K., Emir, U.E., Eberly, L.E.,
        Mangia, S., 2015. Neurochemical and BOLD responses during neuronal activation
        measured in the human visual cortex at 7 Tesla. J Cereb Blood Flow Metab 35, 601–
        610. https://doi.org/10.1038/jcbfm.2014.233
Clarke, W.T., Bell, T.K., Emir, U.E., Mikkelsen, M., Oeltzschner, G., Shamaei, A., Soher,
        B.J., Wilson, M., 2022. NIfTI-MRS: A standard data format for magnetic resonance
        spectroscopy. Magn Reson Med. https://doi.org/10.1002/mrm.29418
Craven, A.R., Bhattacharyya, P.K., Clarke, W.T., Dydak, U., Edden, R.A.E., Ersland, L.,
        Mandal, P.K., Mikkelsen, M., Murdoch, J.B., Near, J., Rideaux, R., Shukla, D., Wang,
        M., Wilson, M., Zöllner, H.J., Hugdahl, K., Oeltzschner, G., 2022. Comparison of
        seven modelling algorithms for γ-aminobutyric acid-edited proton magnetic
        resonance spectroscopy. NMR Biomed 35, e4702. https://doi.org/10.1002/nbm.4702
DiNuzzo, M., Mangia, S., Moraschi, M., Mascali, D., Hagberg, G.E., Giove, F., 2022.
        Perception is associated with the brain's metabolic response to sensory stimulation.
        eLife 11, e71016. https://doi.org/10.7554/eLife.71016
Ebel, A., Maudsley, A.A., 2003. Improved spectral quality for 3D MR spectroscopic imaging
        using a high spatial resolution acquisition strategy. Magn Reson Imaging 21, 113–
        120. https://doi.org/10.1016/s0730-725x(02)00645-8
Eilers, P.H., Boelens, H.F., 2005. Baseline correction with asymmetric least squares
        smoothing. Leiden University Medical Centre Report 1, 5.
Goryawala, M.Z., Sheriff, S., Stoyanova, R., Maudsley, A.A., 2018. Spectral decomposition
        for resolving partial volume effects in MRSI. Magnetic Resonance in Medicine 79,
        2886–2895. https://doi.org/10.1002/mrm.26991
Jung, R.E., Gasparovic, C., Chavez, R.S., Flores, R.A., Smith, S.M., Caprihan, A., Yeo,
        R.A., 2009. Biochemical Support for the "Threshold" Theory of Creativity: A Magnetic
        Resonance Spectroscopy Study. J Neurosci 29, 5319–5325.
        https://doi.org/10.1523/JNEUROSCI.0588-09.2009

Kolasinski, J., Hinson, E.L., Divanbeighi Zand, A.P., Rizov, A., Emir, U.E., Stagg, C.J., 2019. The dynamics of cortical GABA in human motor learning. The Journal of Physiology 597, 271–282. https://doi.org/10.1113/JP276626

Le Fur, Y., Cozzone, P.J., 2014. FID modulus: a simple and efficient technique to phase and align MR spectra. MAGMA 27, 131–148. https://doi.org/10.1007/s10334-013-0381-8

Maudsley, A.A., Andronesi, O.C., Barker, P.B., Bizzi, A., Bogner, W., Henning, A., Nelson, S.J., Posse, S., Shungu, D.C., Soher, B.J., 2021. Advanced magnetic resonance spectroscopic neuroimaging: Experts' consensus recommendations. NMR in Biomedicine 34, e4309. https://doi.org/10.1002/nbm.4309

Maudsley, A.A., Darkazanli, A., Alger, J.R., Hall, L.O., Schuff, N., Studholme, C., Yu, Y., Ebel, A., Frew, A., Goldgof, D., Gu, Y., Pagare, R., Rousseau, F., Sivasankaran, K., Soher, B.J., Weber, P., Young, K., Zhu, X., 2006. Comprehensive processing, display and analysis for in vivo MR spectroscopic imaging. NMR Biomed 19, 492–503. https://doi.org/10.1002/nbm.1025

Mullins, P.G., 2018. Towards a theory of functional magnetic resonance spectroscopy (fMRS): A meta-analysis and discussion of using MRS to measure changes in neurotransmitters in real time. Scandinavian Journal of Psychology 59, 91–103. https://doi.org/10.1111/sjop.12411

Near, J., Evans, C.J., Puts, N.A.J., Barker, P.B., Edden, R.A.E., 2013. J-difference editing of gamma-aminobutyric acid (GABA): Simulated and experimental multiplet patterns. Magnetic Resonance in Medicine 70, 1183–1191. https://doi.org/10.1002/mrm.24572

Near, J., Harris, A.D., Juchem, C., Kreis, R., Marjańska, M., Öz, G., Slotboom, J., Wilson, M., Gasparovic, C., 2021. Preprocessing, analysis and quantification in single-voxel magnetic resonance spectroscopy: experts' consensus recommendations. NMR Biomed 34, e4257. https://doi.org/10.1002/nbm.4257

Nikolaidis, A., Baniqued, P.L., Kranz, M.B., Scavuzzo, C.J., Barbey, A.K., Kramer, A.F., Larsen, R.J., 2017. Multivariate Associations of Fluid Intelligence and NAA. Cereb Cortex 27, 2607–2616. https://doi.org/10.1093/cercor/bhw070

Oz, G., Alger, J.R., Barker, P.B., Bartha, R., Bizzi, A., Boesch, C., Bolan, P.J., Brindle, K.M., Cudalbu, C., Dinçer, A., Dydak, U., Emir, U.E., Frahm, J., González, R.G., Gruber, S., Gruetter, R., Gupta, R.K., Heerschap, A., Henning, A., Hetherington, H.P., Howe, F.A., Hüppi, P.S., Hurd, R.E., Kantarci, K., Klomp, D.W.J., Kreis, R., Kruiskamp, M.J., Leach, M.O., Lin, A.P., Luijten, P.R., Marjańska, M., Maudsley, A.A., Meyerhoff, D.J., Mountford, C.E., Nelson, S.J., Pamir, M.N., Pan, J.W., Peet, A.C., Poptani, H., Posse, S., Pouwels, P.J.W., Ratai, E.-M., Ross, B.D., Scheenen, T.W., Schuster, C., Smith, I.C.P., Soher, B.J., Tkáč, I., Vigneron, D.B., Kauppinen, R.A., MRS Consensus Group, 2014. Clinical proton MR spectroscopy in central nervous system disorders. Radiology 270, 658–679. https://doi.org/10.1148/radiol.13130531

Öz, G., Deelchand, D.K., Wijnen, J.P., Mlynárik, V., Xin, L., Mekle, R., Noeske, R., Scheenen, T.W.J., Tkáč, I., 2021. Advanced single voxel 1H magnetic resonance spectroscopy techniques in humans: Experts' consensus recommendations. NMR in Biomedicine 34, e4236. https://doi.org/10.1002/nbm.4236

Pouwels, P.J., Frahm, J., 1998. Regional metabolite concentrations in human brain as determined by quantitative localized proton MRS. Magn Reson Med 39, 53–60. https://doi.org/10.1002/mrm.1910390110

R Core Team, 2021. R: A Language and Environment for Statistical Computing.

Scheenen, T.W.J., Klomp, D.W.J., Wijnen, J.P., Heerschap, A., 2008. Short echo time 1H-MRSI of the human brain at 3T with minimal chemical shift displacement errors using adiabatic refocusing pulses. Magnetic Resonance in Medicine 59, 1–6. https://doi.org/10.1002/mrm.21302

Soher, B.J., Clarke, W.T., Wilson, M., Near, J., Oeltzschner, G., 2022. Community-Organized Resources for Reproducible MRS Data Analysis. Magnetic Resonance in Medicine 88, 1959–1961. https://doi.org/10.1002/mrm.29387

Wehrens, R., 2020. Chemometrics with R: Multivariate Data Analysis in the Natural and Life Sciences, Use R! Springer, Berlin, Heidelberg. https://doi.org/10.1007/978-3-662-62027-4

Wiehler, A., Branzoli, F., Adanyeguh, I., Mochel, F., Pessiglione, M., 2022. A neuro-metabolic account of why daylong cognitive work alters the control of economic decisions. Current Biology 32, 3564-3575.e5. https://doi.org/10.1016/j.cub.2022.07.010

Wilson, M., 2021a. spant: An R package for magnetic resonance spectroscopy analysis. JOSS 6, 3646. https://doi.org/10.21105/joss.03646

Wilson, M., 2021b. Adaptive baseline fitting for 1H MR spectroscopy analysis. Magn Reson Med 85, 13–29. https://doi.org/10.1002/mrm.28385

Wilson, M., 2019. Robust retrospective frequency and phase correction for single-voxel MR spectroscopy. Magn Reson Med 81, 2878–2886. https://doi.org/10.1002/mrm.27605

Wilson, M., Andronesi, O., Barker, P.B., Bartha, R., Bizzi, A., Bolan, P.J., Brindle, K.M., Choi, I.-Y., Cudalbu, C., Dydak, U., Emir, U.E., Gonzalez, R.G., Gruber, S., Gruetter, R., Gupta, R.K., Heerschap, A., Henning, A., Hetherington, H.P., Huppi, P.S., Hurd, R.E., Kantarci, K., Kauppinen, R.A., Klomp, D.W.J., Kreis, R., Kruiskamp, M.J., Leach, M.O., Lin, A.P., Luijten, P.R., Marjańska, M., Maudsley, A.A., Meyerhoff, D.J., Mountford, C.E., Mullins, P.G., Murdoch, J.B., Nelson, S.J., Noeske, R., Öz, G., Pan, J.W., Peet, A.C., Poptani, H., Posse, S., Ratai, E.-M., Salibi, N., Scheenen, T.W.J., Smith, I.C.P., Soher, B.J., Tkáč, I., Vigneron, D.B., Howe, F.A., 2019. Methodological consensus on clinical proton MRS of the brain: Review and recommendations. Magn Reson Med 82, 527–550. https://doi.org/10.1002/mrm.27742

Woolrich, M.W., Jbabdi, S., Patenaude, B., Chappell, M., Makni, S., Behrens, T., Beckmann, C., Jenkinson, M., Smith, S.M., 2009. Bayesian analysis of neuroimaging data in FSL. Neuroimage 45, S173-186. https://doi.org/10.1016/j.neuroimage.2008.10.055

Wu, B., Bagshaw, A.P., Hickey, C., Kühn, S., Wilson, M., 2022. Evidence for distinct neuro-metabolic phenotypes in humans. NeuroImage 249, 118902. https://doi.org/10.1016/j.neuroimage.2022.118902

Zhang, Y., Brady, M., Smith, S., 2001. Segmentation of brain MR images through a hidden Markov random field model and the expectation-maximization algorithm. IEEE Trans Med Imaging 20, 45–57. https://doi.org/10.1109/42.906424