Fast imaging of millimeter-scale areas with beam deflection transmission electron microscopy

Zhihao Zheng[1], Christopher S. Own[2], Adrian A. Wanner[1,3], Randal A. Koene[2], Eric W. Hammerschmith[1], William M. Silversmith[1], Nico Kemnitz[1], David W. Tank[1], H. Sebastian Seung[1,†]

[1]Princeton Neuroscience Institute, Princeton University, Princeton, NJ, USA
[2]Voxa, Seattle, WA, USA
[3]Present address: Paul Scherrer Institute, Villigen, Switzerland
[†]Correspondence to sseung@princeton.edu

Abstract

We have achieved a three fold increase in the speed of transmission electron microscopy by using a beam deflecting mechanism to enable highly efficient acquisition of multiple image tiles for each motion of the mechanical stage. For millimeter-scale areas, the duty cycle of imaging doubles and exceeds 30%, yielding a net average imaging rate of 0.3 gigapixels per second.

Main text

Volume electron microscopy (EM) is currently the only approach that has been used to reconstruct a connectome – a complete map of neural connectivity at synaptic resolution. Serial section transmission electron microscopy (ssTEM) is a classic volume EM technique used for 3D reconstruction of neural tissues at small scale[1–3]. In the 1980s, ssTEM was used for mapping the first whole-brain connectome of *C. elegans*[4]. Later on, ssTEM was used to acquire 3D EM datasets of multiple *C. elegans* brains[5,6], an entire fly larval brain[7], normal and pathological retinas[8,9], and mouse cortex[10,11]. Recently, automation and parallelization of TEM[10,12,13] have allowed imaging of ever larger volumes, such as a complete adult flybrain[14] and a cubic millimeter volume of mammalian cortex[15].

The previous state of the art in high throughput TEM was a system at the Allen Institute for Brain Science[12]. This achieves high imaging speed through several innovations. A reel-to-reel tape translation system allows automated delivery of sections based on GridTape technology[13]. A 50 MPixel CMOS camera with a low distortion lens achieves a burst imaging rate of 0.5 gigapixels per sec (GPix/s, imaging only). A scalable software infrastructure allows closed-loop workflow management based on real-time image processing[16]. However, imaging time for a cubic millimeter volume is still 6 months with multiple TEMs[12] and there is a growing need to image much larger datasets[17].

Imaging a square millimeter area requires thousands of x-y stage translations to acquire image tiles that are later stitched together to form a multi-tile image. Because the time to image a tile has become so fast the stage translations require more time than the image capture itself[12]. A massive number of positional translations – typically ~10,000 per mm^2 – also rapidly

44   consumes piezo stage life of ~10 billion cycles of the crystal per year when operated 24/7.
45   Therefore, to advance the state of the art, it is essential to reduce the overhead due to stage
46   translations.
47
48   Here we report the development of a beam deflection mechanism (Cricket[TM]) to
49   significantly increase TEM imaging throughput (bdTEM, beam deflection). In addition, the
50   microscopes are integrated with an automated reel-to-reel tape translation system[12,13] and a 50
51   MPixel CMOS camera[12]. The bdTEM achieves the fastest TEM imaging rate to date, a three
52   fold increase over the previous TEM with the same camera[12].
53
Beam deflection (Cricket). We repurposed electromagnetic lens deflectors in the TEMs,
above the object plane and below (Fig. 1a). The deflectors above the object plane shift the
electron beam over the specimen in a matrix of 3 x 3 tiles and simultaneously the deflectors
below the object plane precisely de-scans the image back onto the TEM camera. The beam
deflection allows acquisition of 9 image tiles without moving the stage, eliminating 8 out of 9
stage movements (Fig. 1b-c, Extended Data Fig. 1b). Overhead for stage motion is therefore
significantly reduced, accounting for only 7% of montage imaging time with an additional 9% of
Cricket settling time (Fig. 1d). As a result, the imaging duty cycle increases from a previous
peak performance of 15%[12] to over 30% of total time per section in our system (Methods).
63
Optics. The detector optics of the TEM is upgraded with a compact light-optical lens
assembly and a high speed CMOS camera that yields a frame size of 6,000 × 6,000 pixels (or
36 MPix). The sensor sits near the film plane of a conventional TEM (Extended Data Fig. 1a),
eliminating the impractical need of an extended vacuum column required by the past
generations of large pixel-count detectors[10,12]. Images acquired at an exposure time of 40 ms
have sufficient resolution for synapse identification and reconstruction of neuronal morphology
(Fig. 1e-i, Supplementary video 3). As a result, the burst imaging rate (0.9 GPix/s, imaging only)
of bdTEMs is 1.8 times that of the previous fastest imaging rate[12] (0.5 GPix/s).
72
Section transition and stage. The TEMs are integrated with a previously reported reel-to-
reel tape translation system and a two-axis piezo-driven stage[12,13]. This reel-to-reel tape
translation system allows automated section exchange and barcode-based slot identification
using GridTape technology[12,13]. In previous volume TEM techniques, stage step-and-settle time
ranges from under 50 ms to 100 ms for one axis[12,14]. The stage in our TEMs achieves a step-
and-settle time of ~57 ms on average on both axes (Extended Data Fig. 1c-d), increasing
efficiency on the remaining stage motions.
80
These modifications dramatically increase imaging throughput. Acquisition of a 1 mm^2
section takes 8.6 minutes including imaging, stage motion, and section transition (net average
imaging rate of 0.3 GPix/s at 3 nm/pix, Fig. 1d, Table 1, and Supplementary Video 1). The net
imaging rate is 3 times the peak imaging rate of the previous fastest TEM at 0.1 GPix/s with the
same 50 MPix camera[12]. The previous imaging rate[12] includes overhead for image quality
control and post-processing. These functionalities have not yet been integrated into our

acquisition software but in principle could be performed during the overhead for acquisition
software and section transition.
89
We set out to test whether the TEMs can be used to acquire a 3D EM volume. The reel-
to-reel sample delivery system requires serial ultrathin sections to be cut and collected on
GridTape[12,13]. Based on previous designs[12,13], we modified an ATUMtome (RMC/Boeckeler) to
be compatible with an ultramicrotome (Leica UC7) and GridTape, and used the automated
sectioning system to collect serial ultrathin sections of mouse hippocampus tissue. Imaging a
series of sections involves cycling through the following steps: (1) translate GridTape to the
targeted slot guided by index barcodes; (2) acquire an overview image of the section at low
magnification (Fig. 1c); (3) extract ROI from the overview image and set imaging parameters
(e.g. illumination correction and autofocus); (4) montage imaging of the ROI at high resolution.
Basic software has been implemented to run the workflow (Supplementary video 1). With 10%
overlap between supertiles and 15% overlap between subtiles within a supertile, all tiles of a
common section of 1 mm^2 in size can be stitched together (Fig. 1e-i). Consecutive sections are
aligned to assemble a volume (Supplementary Video 2-3). Ultrastructural features such as
postsynaptic densities and vesicles are clearly defined (Fig. 1i). Small-diameter axons and
dendrites can be followed across sections. These data demonstrate that sequential sections
imaged with the bdTEMs can be assembled into a 3D volume for connectomic application.
106
An array of four microscopes have been fully installed in our facility (Extended Data Fig.
2), some of which have been reliably acquiring data for many months with reasonable
microscope downtime[12]. By estimate, if the four bdTEMs are running 24/7, imaging a cubic
millimeter will take about a month (Table 1). However, non-stop long-term imaging requires
further development of automation software that can handle image quality control, workflow
management, image and section databases, and stitching and alignment. A scalable
architecture with these functionalities has been developed[16] but needs to be adapted to
integrate with bdTEMs. On raw imaging speed, breakdown of various overhead sources (Fig.
1d) points to several avenues of future improvements. Upgrade of the optics with a 100 MPix
camera will in principle double the burst imaging rate. Additional increase of imaging speed can
be gained by refactoring frequently used tasks in the acquisition software to low-level programs,
reducing tile overlap, enlarging pixel size, and increasing efficiency on section transition.
119
Volume EM is dichotomized into SEM-based and TEM-based methods, and both have
been shown to be capable of acquiring large-scale datasets (MultiSEM[18] and GridTape-based
TEM[15]). Here we have further improved the TEM imaging speed to 0.3 GPix/s. Similar to the
previous TEMs[12,13], our modifications are based on a retrofitted JEOL 1200EX-II microscope
and cost for the entire system (< $500,000) is significantly lower compared to equivalent
imaging systems, permitting acquisition of multiple TEMs for an imaging facility (six at the Allen
Institute for Brain Science[12] and four at Princeton University). The current system, however,
requires specialized films on GridTape, which scales linearly in cost with the number of sections
(USD $4 - $8 per slot); but improving film-making methods and economies of scale may
decrease the cost in the future. Crucially, both imaging methods currently depend on cutting and
collecting a large number of ultrathin sections. Sectioning-related artifacts (folds, cracks, knife

marks, etc.) are very common, though automated reconstruction methods that are robust to
some defects have been developed[19,20]. Wide-area milling of imaged tissue was recently
developed to circumvent ultrathin sectioning[21] and, if integrated with a MultiSEM, can potentially
scale up connectomic volumes beyond petascale. In addition, the reliability of serial sectioning
can be improved by magnetic section collection[22]. For TEM, one proposal to improve reliability
of sectioning is to increase the thickness of serial sections, which are then imaged with
tomography to recover a finer axial resolution[23]. Whether SEM or TEM-based techniques is the
method of choice for exascale datasets remains to be determined.
139
Methods
Sample preparation and sectioning. All procedures were carried out in accordance with the
Institutional Animal Care and Use Committee at Princeton University. Mice (C57BL/6J-Tg(Thy1-
GCaMP6f) GP5.3Dkim/J, Jackson Laboratories, 028280) aged 6 months were transcardially
perfused with a fixative mixture of 2.5% paraformaldehyde and 1.3% glutaraldehyde in 0.1M
Cacodylate with 2mM CaCl2 pH 7.4. The brain was extracted and post-fixed for 36 hrs at 4°C in
the same fixative solution. The perfused brain was subsequently rinsed in 0.1M Cacodylate with
2mM CaCl2 for 1 hr (3 x 20 mins) and 300 μm coronal sections were cut on a Leica Vibratome.
Sample blocks were cut out and stained based on a modified reduced osmium treatment
(rOTO) protocol with the addition of formamide[15,24]. The tissue blocks were first *en bloc* stained
with 8% formamide, 2% osmium tetroxide, 1.5% potassium ferrocyanide for 3 hours.
Subsequently, the samples were immersed in 1% TCH (94 mM) 50 °C for 50 mins, followed by
a second step of 2% osmium staining for 3 hours. The sample was placed in 1% uranyl acetate
overnight at 4°C, followed by lead aspartate (Walton's, 20 mM lead nitrate in 30 mM aspartate
buffer, pH 5.5) at 50 °C for 2 hours. After washed with water (3 x 10 mins), samples proceeded
through a graded acetonitrile dehydration series (50%, 75%, 90% w/v in acetonitrile, 10 minutes
each at 4 °C, then 4 x 10 minutes of 100% acetonitrile at room temperature). After a progressive
resin infiltration series (33% resin:acetonitrile, 50% resin:acetonitrile, 66% resin:acetonitrile, 8
hours each), the sample was incubated in fresh 100% resin overnight and the resin was cured
in the oven at 60 °C for at least 48 hrs.
GridTape (Luxel Corporation) contains regularly spaced apertures with plastic film
substrates for serial sections[12,13]. The films are transparent to electrons, and therefore are
compatible with TEM imaging. We combined an automated tape collecting system (ATUMtome,
RMC/Boeckeler) and an ultramicrotome (UC7, Leica) to create an automated tape-based
sectioning and collection system (Automated Tape Collecting Ultramicrotome, ATUM) for
GridTape, based on previous designs[12,13]. Our custom ultramicrotome setup also includes a
computer-controlled, high-precision motorized stage, monitoring of temperature and humidity,
and three cameras to monitor the collection process. Additionally, we have built closed-loop
control software to phase-lock aperture movement with cutting, in order to center sections
collected in the apertures on GridTape. After resin embedding, the ultrathin sections were then
cut at a nominal thickness of 40 nm and automatically collected onto a GridTape.
171
Reel-to-reel translation system and GridStage. The reel-to-reel translation system and
GridStage (Voxa) have been described previously[12]. Briefly, Gridstage Reel comprises three
major components: a stage cartridge, a reel storage and delivery system, and an airlock

assembly enabling cartridge loading. A 3D rendering of the GridStage Reel system is shown in
Extended Data Fig. 1, attached to a JEOL 1200EX-II microscope chassis. GridStage's two-axis
stage cartridge is set to a fixed height to enable the imaging environment to stay the same from
section to section as they are delivered into the imaging area by conveyor. Each axis is driven
by a precision closed-loop piezo motor. The precision is ~50 nm, and absolute positional
accuracy on the sample is typically between 100 - 200 nm. This mechanism positions the
sample accurately and reliably within the requirements of montage image overlap
reproducibility, even given electron-optical environment variability (e.g. charging and sample
morphology changes) of different samples in the volume. The cartridge incorporates an
accurate barcode reading and clamping mechanism to precisely position and identify slots on
GridTape, with read accuracy greater than 99.9% for fresh GridTape. This enables precise
tracking of the current tissue section under first-time montage imaging or during re-imaging
operations. The precision of the GridStage sample ID subsystem supports either sequential or
fully random-access sample delivery modes.
The stage has fast responsiveness, with roundtrip communication to the drive system on
the order of a few milliseconds, and typical step-and-settle times (>95% of moves) occurring in
under 60 ms (Extended Data Fig. 1c-d). The average step-and-settle time is similar to and
slightly improved over a previously reported fast Piezo-driven stage[14]. Voxa reels are housed in
robotic delivery and take-up conveyor systems connected to the microscope's vacuum, and can
accept a GridTape reel with up to 5,500 sections spaced 6 mm apart. A continuously-monitored
belt tension meter ensures the GridTape translates within specified operating parameters, and
supports safe transit of the samples into and out of the imaging area. The GridStage reel
sample delivery system can deliver sequential sections as fast as once every two seconds, for
quick survey modes.
199
Lens assembly and camera. The electron optics of each of our TEM systems consists of a
custom phosphor scintillator of 75 mm in diameter, a lens assembly (AMT, NanoSprint50M-AV),
and a high speed CMOS camera (XIMEA, CB500MG-CM). The pickup area of 47 x 47 mm on
the phosphor screen is near the film plane of the JEOL 1200EX-II TEM and therefore minimizes
image distortion at the bottom of the column and matches the nominal magnification setting of
the microscope platform. Conveniently, this setup eliminates the need for a lengthened column
that previously requires building scaffolds and is impractical for many facilities[10,12,14]. In addition,
compared to the previous design of an extended column, the smaller scintillator and compact
lens in our TEMs allows good signal-to-noise with lower dose or with shorter exposure times
leading to more efficient use of available beam electrons. The CMOS camera has a PCIe
interface that supports fast data transfer rates. Overall, the lens and camera supports nominal
image acquisition with a frame size of 6,000 x 6,000 pixels at a shortest exposure time of 40 ms.
In the future, an upgrade of a 100+ Mpix CMOS camera with backside illumination could
potentially lead to another couple-fold increase of imaging rate[12].
214
TEM imaging. Some components of the TEM systems have been briefly presented before[25].
The Beam deflection mechanism (Cricket) for TEM has been prototyped previously[12] but not
demonstrated for large-scale imaging. For imaging, a reel containing samples on a GridTape
are first loaded into the reel housing of the reel-to-reel system (Extended Data Fig. 1), which is

connected with the TEM column vacuum. After loading, the microscope is pumped down to
reach the vacuum level of ~1E-7 Torr. Lanthanum hexaboride crystal (LaB$_6$) filaments were
used due to their high electron flux per unit area and a longer lifetime (1,000 hrs or more as
opposed to 200 hrs for a tungsten source) reducing downtime needed for filament change.
Crucially LaB$_6$ filaments depend more sensitively on a good vacuum pressure for good stability
and lifetime (low 10^{-7} Torr). It typically takes about half a day when loading a new sample reel to
pump down the column to the base operating pressure for LaB6. We then increase the high-
tension voltage to 120 KV, turn on filament current and then perform alignments on various
components of the TEM, following routine technical instructions for TEM operations. These
procedures take half a day and only need to be performed after installation of a new filament or
loading of a new tape. Calibration of reel-to-reel systems involves tape and tension calibration
for reliable translation and barcode reading, which typically takes ~5 minutes and is only done
when first loading a reel. Cricket alignment is performed to ensure each sub-tile has sufficient
overlap (~15%) with neighboring sub-tiles and that there is minimal distortion across the field
between images from different sub-tiles to facilitate efficient stitching and reconstruction. Cricket
alignment is usually stable over several months.
The TEMs are compatible with GridTape[13], which has regularly spaced 2 x 1.5 mm^2
apertures in aluminum-coated polyimide tape. Each aperture can be identified by a unique
barcode milled on the tape. The tape is coated with a 50 nm-thick film (Luxel Corporation) that
spans the apertures and serves as support for sections. The steps to perform before imaging
every section include: deliver a new aperture into the field of view using the reel-to-reel system,
locate the aperture center using stage algorithmically, find the illumination center of the electron
beam (which is often off-center due to magnetic hysteresis from magnification change or
change in charge equilibrium state of the new aperture), adjust camera gain control, extract
ROI, and autofocus at high magnification. These are done for every aperture with a tissue
section, and typically take 2 - 3 minutes per aperture (Table 1). Software (Blade, Voxa Inc.) has
been developed that can control the Reel-to-reel system, GridStage, and TEM automatically.
The software has lower-level functions that execute each step in the workflow (e.g. travel to a
specific aperture in the GridTape reel, move the stage, adjust focus of the TEM). The low-level
software supports end-user scripting and control of application-specific microscope workflows
via the Blade API supplied to the user. Additionally, the Blade software allows initialization of
imaging remote control over IP of all operations and multiple user-defined imaging modes - e.g.
continuous batch imaging, re-imaging, and selective or random sampling.
The acquisition computer for each microscope has an intermediate storage of 16 - 32 TB
SSDs via a PCIe card with 4 USB3.0 ports, for ~24 hours of imaged data. Each USB port can
support up to 10 Gb/s of transfer speed. After caching on the temporary storage on SSD, the
data are then transferred via a 10 Gb/s network connection to an on-premise petascale cloud
storage in a separate building. Data transfer is done using CloudFiles (https://github.com/seung-
lab/cloud-files).
In general, imaging rates (Fig. 1d, Table 1) are computed as the number of pixels that
are acquired in a given amount of time to image a section. Burst imaging rates are calculated
from the number of pixels in a frame over camera exposure time per frame (6,000 x 6,000 pixels
in 40 ms). Montage imaging rates are calculated from the total number of pixels of all images
(i.e. number of pixels in a tile multiplied by number of tiles) over time, from the first image to the

last image of a montage acquisition. Sources of overhead for montage imaging rates include imaging, stage motion, beam deflection, computational overhead of the acquisition software, and computing basic image statistics (Fig. 1d). Net imaging rates are computed from the same number of pixels as in montage imaging rate, including all overhead from montage imaging, but additionally includes overhead for section transition (e.g. low-magnification overview of sections on GridTape, ROI extraction, autofocus, automatic beam centering). Effective imaging rates are calculated from the total number of pixels of a montage acquisition with overlap pixels only counted once, over the same overhead as in the net imaging rate. Time for a montage acquisition varies from section to section by ~30 MPix/s (s.d.) depending on variable computational overhead in the acquisition software. Peak imaging rate is taken from an acquisition with the shortest imaging time. Average imaging rates are taken from a typical production acquisition (an average or 10 or more acquisitions). Average transition time is the average taken from 10 consecutive automatic acquisitions.

Imaging duty cycle is computed as the fraction of time for image acquisition (number of tiles x exposure time per tile) out of total time per section. For example, at 3 nm/Pixel the image acquisition time is 2.9 minutes (4,356 tiles at 40 ms per tile), which is 33.7% out of total time per section (8.6 minutes). Similarly, the imaging duty cycle at 4 nm/Pixel is 31%. In comparison, the peak performance of the previous state of the art system[12] has an imaging duty cycle of 15% (2,600 tiles at 50 ms per tile out of 14 minutes of total time).

Stitching and alignment. Image stitching was performed with the AlignTK software[10] (https://mmbios.pitt.edu/aligntk-home). Image alignment was performed with a custom software pipeline[20].

Table 1. Performance metric for bdTEMs.

Supplementary video 1. Screen recording of an acquisition at a peak imaging rate.
Supplementary video 2. Aligned images across 55 sections acquired at 120 ms exposure time.
Supplementary video 3. Aligned images across 20 sections acquired at 40 ms exposure time.
https://doi.org/10.6084/m9.figshare.21614058

Acknowledgments
The Cricket system was initially prototyped and tested by Voxa with the Allen Institute for Brain Science[12]. We thank R. Clay Reid, Daniel J. Bumbarger, Wenjing Yin, Derrick Brittain, Marc Takeno, Nuno Macarico da Costa, David Hildebrand for advice on microscopy and serial sectioning setup; Stephan Y. Thiberge for help in the planning and construction of the TEM facility; Lawrence Own and Teddy DeRego for their development and support in GridStage and Cricket hardware and software; Wei-Chung Allen Lee and Jasper Phelps for advice on image stitching; John Price for modification of ATUMtome. This work was supported by the NIH grants 1S10OD023602-01A1 and 1U19NS104648 and the Simons Collaboration on the Global Brain. The authors acknowledge the use of Princeton's Imaging and Analysis Center, which is partially supported through the Princeton Center for Complex Materials (PCCM), a National Science Foundation (NSF)-MRSEC program (DMR-2011750).

307
Author contributions
C.S.O. and A.A.W. conceived the study. Z.Z. developed the conceptual framework for quantifying the performance. D.W.T. and H.S.S. acquired funding and supervised. A.A.W. planned the construction of the facility and managed the installation and testing of the first instrument. A.A.W. registered and stitched the initial datasets. Z.Z., C.S.O., and R.A.K. developed software for TEM imaging. Z.Z., A.A.W., and E.W.H. developed software and hardware for serial sectioning. W.M.S. developed software for data transfer. Z.Z. performed image stitching. Z.Z. and N.K. performed image alignment. Z.Z. acquired and analyzed data. Z.Z. and C.S.O. carried out validation and evaluation. Z.Z. and H.S.S. drafted the manuscript with input from all authors.

Competing interest
N.K. and H.S.S. disclose financial interests in Zetta AI LLC. C.S.O. and R.A.K. disclose financial interests in Voxa. A.A.W. is a founder and owner of ariadne.ai ag (Switzerland).

References

1. Harris, K. M. *et al.* Uniform serial sectioning for transmission electron microscopy. *J. Neurosci.* **26**, 12101–12103 (2006).

2. Stevens, J. K., Davis, T. L., Friedman, N. & Sterling, P. A systematic approach to reconstructing microcircuitry by electron microscopy of serial sections. *Brain Res.* **2**, 265–293 (1980).

3. Harris, K. M. & Weinberg, R. J. Ultrastructure of synapses in the mammalian brain. *Cold Spring Harb. Perspect. Biol.* **4**, (2012).

4. White, J. G., Southgate, E., Thomson, J. N. & Brenner, S. The structure of the nervous system of the nematode Caenorhabditis elegans. *Philos. Trans. R. Soc. Lond. B Biol. Sci.* **314**, 1–340 (1986).

5. Cook, S. J. *et al.* Whole-animal connectomes of both Caenorhabditis elegans sexes. *Nature* **571**, 63–71 (2019).

6. Witvliet, D. *et al.* Connectomes across development reveal principles of brain maturation. *Nature* **596**, 257–261 (2021).

7.   Eichler, K. *et al.* The complete connectome of a learning and memory centre in an insect

brain. *Nature* **548**, 175–182 (2017).

8.   Sigulinsky, C. L. *et al.* Network Architecture of Gap Junctional Coupling among Parallel

Processing Channels in the Mammalian Retina. *J. Neurosci.* **40**, 4483–4511 (2020).

9.   Pfeiffer, R. L. *et al.* A pathoconnectome of early neurodegeneration: Network changes in

retinal degeneration. *Exp. Eye Res.* **199**, 108196 (2020).

10. Bock, D. D. *et al.* Network anatomy and in vivo physiology of visual cortical neurons. *Nature*

471, 177–182 (2011).

11. Lee, W.-C. A. *et al.* Anatomy and function of an excitatory network in the visual cortex.

Nature **532**, 370–374 (2016).

12. Yin, W. *et al.* A petascale automated imaging pipeline for mapping neuronal circuits with

high-throughput transmission electron microscopy. *Nat. Commun.* **11**, 4949 (2020).

13. Phelps, J. S. *et al.* Reconstruction of motor control circuits in adult Drosophila using

automated transmission electron microscopy. *Cell* **184**, 759–774.e18 (2021).

14. Zheng, Z. *et al.* A Complete Electron Microscopy Volume of the Brain of Adult Drosophila

melanogaster. *Cell* **174**, 730–743.e22 (2018).

15. MICrONS Consortium *et al.* Functional connectomics spanning multiple areas of mouse

visual cortex. *bioRxiv* 2021.07.28.454025 (2021) doi:10.1101/2021.07.28.454025.

16. Mahalingam, G. *et al.* A scalable and modular automated pipeline for stitching of large

electron microscopy datasets. *Elife* **11**, (2022).

17. Abbott, L. F. *et al.* The mind of a mouse. *Cell* **182**, 1372–1376 (2020).

18. Shapson-Coe, A. *et al.* A connectomic study of a petascale fragment of human cerebral

cortex. Preprint at https://doi.org/10.1101/2021.05.29.446289 (2021).

19. Macrina, T. *et al.* Petascale neural circuit reconstruction: automated methods. *bioRxiv*

2021.08.04.455162 (2021) doi:10.1101/2021.08.04.455162.

20. Popovych, S. *et al.* Petascale pipeline for precise alignment of images from serial section

electron microscopy. *bioRxiv* 2022.03.25.485816 (2022) doi:10.1101/2022.03.25.485816.

21. Hayworth, K. J. *et al.* Gas cluster ion beam SEM for imaging of large tissue samples with

10 nm isotropic resolution. *Nat. Methods* **17**, 68–71 (2020).

22. Templier, T. MagC, magnetic collection of ultrathin sections for volumetric correlative light

and electron microscopy. *Elife* **8**, (2019).

23. Soto, G. E. *et al.* Serial section electron tomography: a method for three-dimensional

reconstruction of large structures. *Neuroimage* **1**, 230–243 (1994).

24. Mikula, S. & Denk, W. High-resolution whole-brain staining for electron microscopic circuit

reconstruction. *Nat. Methods* **12**, 541–546 (2015).

25. Own, C. S. *et al.* Multi-order Scaling of High-throughput Transmission Electron Microscopy.

Microsc. Microanal. **25**, 1040–1041 (2019).

Fig. 1

Fig. 1. Beam deflection for TEM imaging.
a-b, Cricket Schematic. At a stage position, Cricket sequentially scans each tile of a 3 x 3 matrix pattern on the sample. The tiles can be stitched together in real-time, producing a "supertile". Each square represents a single tile and 9 tiles compose a supertile with overlap between tiles. Neighboring supertiles (different colors) can be stitched together to form a montage image. **c,** Overview of a section on a 2 mm x 1.5 mm slot imaged using Cricket at low magnification. **d**, Imaging speed based on acquisitions at 3 nm/pix (*Methods*). The bar lengths are proportional to the contribution of each overhead to imaging speed. Computational overhead of acquisition ("acq software") includes communications between different components, software time jitters, computing image statistics, on-the-fly stitching of subtiles to supertiles, and image storage. **e,** A fully stitched montage of a hippocampus section (1 mm^2) imaged with a Cricket-equipped TEM. The montage includes 4320 tiles (or 480 supertiles). **f,** A montage of 6 stitched supertiles (from red outline in **e**). **g,** A Cricket supertile consisting of 9 tiles (from orange outline in **f**). Each tile has 6,000 x 6,000 pixels and the overlap between neighboring subtiles of a supertile is 15% (~900 pixels). **h,** A single image tile (from green outline in **g**). **i**, High-magnification EM images with 120 ms (upper, from blue outline in **h**) and 40 ms exposure time (lower).

Table 1

Performance metric for TEM with Cricket (1 mm² section of 45 nm thick)					
imaging parameters	avg. 3 nm/pix (120 ms exp.)	avg. 3 nm/pix (40 ms exp.)	peak 3 nm/pix (40 ms exp.)	avg. 3.6 nm/pix (40 ms exp.)	unit
resolution	3	3	3	3.6	nm/pixel
tile size	6000	6000	6000	6000	pix (aspect ratio of 1:1)
FOV size	18	18	18	21.6	µm
pixels per tile	36	36	36	36	Mpixel
supertile size (per side, 15% overlap b/w tiles)	16200	16200	16200	16200	pix (aspect ratio of 1:1)
overlap between supertiles	600	600	600	600	pix (i.e. 10% of tile size)
exposure time	120	40	40	40	ms
tiles per section (w/ overlap)	4356	4356	4140	2916	tile
supertiles per section (w/ overlap)	484	484	460	324	supertile
imgaging time per section	11.7	6.4	4.7	4	minute
transition overhead (tape translation, ROI definition, autofocus, etc.)	2.2	2.2	1.8	2.2	minute
total time per section	13.9	8.6	6.5	6.2	minute
burst imaging rate (imaging only)	900	900	900	900	Mpix/s
montage imaging rate (w/ stage, Cricket)	223	408	529	437	Mpix/s
net imaging rate (w/ stage, Cricket, transition)	188	304	382	282	Mpix/s
effective imaging rate (w/ stage, Cricket, transition, overlap)	142	229	288	213	Mpix/s
sections per day (1 scope at 24 hr)	103	167	221	232	section
time to image 1 mm³ (4 scopes at 24/7)	54	34	26	24	days
time to image 1 mm³ (4 scopes at 65% uptime #)	83	52	39	37	days
# 65% uptime accounts for microscope maintenance, imaging pause (Yin et al. 2020)					

Extended Data Fig. 1

a

Cricket: beam scanner for TEM

GridStage: automated reel-to-reel system

Cartridge: dual-axis piezo-driven fast stage

GridTape and housing

50 megapixel camera & lens system

b   supertile

one tile (6k x 6k pix)

beam deflection sequence

c

d

Extended Data Fig. 1. A bdTEM for high-throughput imaging.
a, The TEM is a refurbished JEOL 1200EX-II with 120 KV accelerating voltage. The custom modifications include a beam deflection mechanism (Cricket, Voxa), an advanced reel-to-reel tape translation system with a dual-axis piezo-driven fast stage (GridStage, Voxa), and a high speed CMOS camera (CB500MG-CM, XIMEA) with improved lens design (NanoSprint50M-AV, AMT). **b,** Scanning sequence of tiles in a supertile. **c-d,** Distributions of stage step-and-settle time in x (c) and y (d) axes.

Extended Data Fig. 2

a

b

Extended Data Fig. 2. A bdTEM facility.
a, Dimensions of a microscope room that houses an array of 4 bdTEMs with an inside room for sectioning.
b, A photograph of 4 bdTEMs that are installed in the same microscope room (a).