An automated respiratory data pipeline for waveform characteristic analysis

Savannah Lusk, Christopher S. Ward, Andersen Chang, Avery Twitchell-Heyne, Shaun Fattig,

Genevera Allen, Joanna Jankowsky, Russell Ray

Abstract

Comprehensive and accurate analysis of respiratory and metabolic data is crucial to modelling

congenital, pathogenic, and degenerative diseases converging on autonomic control failure. A

lack of tools for high-throughput analysis of respiratory datasets remains a major challenge. We

present Breathe Easy, a novel open-source pipeline for processing raw recordings and

associated metadata into operative outcomes, publication-worthy graphs, and robust statistical

analyses including QQ and residual plots for assumption queries and data transformations. This

pipeline uses a facile graphical user interface for uploading data files, setting waveform feature

thresholds, and defining experimental variables. Breathe Easy was validated against manual

selection by experts, which represents the current standard in the field. We demonstrate

Breathe Easy's utility by examining a 2-year longitudinal study of an Alzheimer's Disease mouse

model to assess contributions of forebrain pathology in disordered breathing. Whole body

plethysmography has become an important experimental outcome measure for a variety of

diseases with primary and secondary respiratory indications. Respiratory dysfunction, while not

an initial symptom in many of these disorders, often drives disability or death in patient

outcomes. Breathe Easy provides an open-source respiratory analysis tool for all respiratory

datasets and represents a necessary improvement upon current analytical methods in the field.

Introduction

Whole-body plethysmography (WBP) in rodent models is an invaluable tool for investigating the etiology and progression of many diseases including congenital respiratory disorders such as Sudden Infant Death Syndrome, contagious respiratory viruses such as SARS-CoV-2, respiratory dysfunction derived from spinal cord and brain injuries, and loss of airway protection and respiratory failure in neurodegenerative disorders such as Alzheimer's Disease. Elaborated WBP systems can measure the volume and composition of air that is inhaled and exhaled per breath offering novel and critical insights into disease etiology and progression. This technique has become integral in the analysis of respiratory outcomes leading to an average of over 600 articles every year published using this method for the last 20 years. Despite the importance of this technique, its application remains limited as few, if any, comprehensive, open-source tools exist for respiratory waveform analysis, leaving manual data annotation as the dominant approach. To aid in the consistent analysis of respiratory waveform and metabolic data, and to make possible high-throughput and immensely sized experimental designs, we have developed the Breathe Easy software package.

The analysis of plethysmography data consists of two components: 1) the conversion of collected waveforms into physiological measures, e.g., respiratory frequency and tidal volume, and 2) the statistical assessment and graphical display of those measures and their associative relationships with independent variables. Labor-intensive, time-intensive, and expertise-dependent manual selection of data remains the most common approach[1,2], which results in a significant bottleneck. Current annotation methods limit reproducibility within and between labs (Supplemental Table 1); is subject to observer bias (i.e., where humans tend to select for particularly slow, even breaths), especially when identifying and excluding movement artifacts; and typically includes only a tiny sample of all available data (usually only examining 30 seconds – 2 minutes of breathing per experimental condition[1–3]). Analysis of the selected data can be similarly encumbered by a heavy reliance on more manual approaches that involve copying and pasting manually-selected data into inherited spreadsheets with default formulae[2,3]. Each component of the analysis is manually driven, which makes analysis of data from even the simplest of experiments exceedingly time consuming. The few available commercial solutions that enable partially automated analysis are 1) expensive; 2) closed source; 3) remain limited in their overall capabilities; and 4) often require purchase of a commercial plethysmograph system for full analytical options, which precludes analysis of data collected on other systems and the use of custom WBP systems (Supplemental Table 2). These limitations further constrain the

complexity and magnitude of respiratory studies and dissuade multivariate experimental designs

examining interactions across multiple variables such as genotype, drug, pathogen, injury,

physiological challenges (e.g., hypercapnic, hypoxic, or room air breathing), sex, and age (in the

case of longitudinal studies). Thus, approaches are lacking that enable the customized

evaluation of large or complex (multivariate) datasets for cardiorespiratory and metabolic

physiology studies.

Here we present Breathe Easy, a novel, modular pipeline that can be run unsupervised in a

continuous fashion over large datasets or more flexibly for highly customized analyses, thereby

overcoming the above limitations. Breathe Easy consists of three modules: 1) a graphical user

interface for rapid and facile loading of data and customization of parameters surrounding

waveform feature extraction, desired statistical analysis, and graphical displays; 2) a breath

extraction module to segment and quantify pertinent respiratory variables using pressure,

temperature, and gas concentration waveforms filtered by exclusion criteria; and 3) an analysis

and graphing module that generates publication-worthy graphs and statistics accompanied by

QQ and residual plots to gauge adherence to assumptions. We first validated Breathe Easy to

show that it outperforms commonly used manual annotation techniques. We then used Breathe

Easy to analyze a large, longitudinal respiratory dataset collected over two years (more than

600 WBP sessions) using a genetic model of Alzheimer's Disease (AD) and demonstrated that

our software robustly handles a multitude of waveform anomalies and can readily manage large

terabyte-scale respiratory data sets.

Results

Overview of software

Figure 1 provides an overview of our new modular, comprehensive, and automated respiratory

analysis software pipeline, Breathe Easy, which is flexibly designed for use both by novices and

experts. More detailed descriptions of each component of Breathe Easy can be found in the

Methods section. To use Breathe Easy, we strongly recommend new users consult our User

Manual website, which can be accessed through the GitHub repository linked to this manuscript.

Briefly, the data and user-defined inputs for the pipeline are managed through the main

Graphical User Interface (GUI), increasing accessibility to novices both in respiratory waveform

analysis and coding (Fig. 1A). The overarching workflow outlined in Figure 1B for a typical

multivariate respiratory experiment is managed through a series of subGUIs (Fig. 1C) that allow

the user to define inclusion and exclusion criteria for waveform features, parameters for the

experimental design, and options for statistical analysis and graphical output. Screenshots of

each subGUI are shown in Supplemental Figure 1.

The raw plethysmography signal files, experimental metadata files, and user-defined filtering

parameters (i.e., configuration files generated by the GUI based on user selections) are passed

to the Breathing Analysis Selection and Segmentation for Plethysmography and Respiratory

Observations (BASSPRO) module. The configuration files inform BASSPRO how the user

defines a breath and exclusion criteria for artifacts like animal movement or flow obstructions.

The application of these criteria is demonstrated in an example trace in Figure 2A. BASSPRO

records instantaneous parameters for each identified breath that is selected for inclusion and

compiles the output in a single .json file for each recording.

Following quality control, waveform filtering, and feature detection by BASSPRO, the data is

sent to the STAtistics and Graph Generator (STAGG) module where statistics and graphs are

generated based on user-defined settings. The user selects independent, dependent, and

covariate variables for statistics; assignments for graphing respiratory variables; instructions for

graphs other than respiratory variables selected by the user (e.g., body weight, body

temperature, behavioral experiment outcomes, etc.); or different graphing settings for

respiratory variables.

A successful run of Breathe Easy generates publication-ready figures with statistical

significance marks added and tables of quantitative data representing the graphical outputs and

showing statistical testing results. The output of this pipeline and required efforts to run the

program represent a significantly improved and streamlined pipeline for plethysmography data
analysis.

Validation

To validate the automated selection of breaths by BASSPRO, we compared the breaths
selected by BASSPRO to those selected manually by experts in 16 representative traces from
the 14 months timepoint in our Alzheimer's dataset. Three experts (as defined by a record of
peer-reviewed publication in respiratory waveform analysis[4–6]) were engaged to select periods
of calm, quiet breathing from regions of interest (i.e., stable gas concentrations for room air and
challenge gas exposure) for analysis and comparison. The same files were then run through
Breathe Easy using default parameters. Out of 391,424 total breaths identified in the 16 traces,
312,515 breaths were from desired sections of the recording (e.g., animal habituated and room
air or fully equilibrated 5% CO_2). Each expert selected, on average, $18,854.3 \pm 6,678.7$ breaths
whereas BASSPRO selected 149,395 breaths from those same 16 files (Fig. 2B). $88.1\% \pm 3.8\%$
of breaths selected by each of the experts were also selected by BASSPRO, showing high
overlap between the two selection methods. In contrast, $11.9\% \pm 3.8\%$ of breaths selected by
experts were rejected by BASSPRO, where ~3% of this difference is due to experts selecting
outside of the desired periods of interest, e.g., 6 minutes before the end of the condition instead
of only selecting from the last 5 minutes. Thus, of breaths selected by experts as calm, quiet
breathing from the correct periods of interest, less than 9% of breaths selected by experts were
rejected by BASSPRO for not meeting quality standards.

There is concordance between BASSPRO and the experts and high discordance between
excluded breaths and those included by BASSPRO and experts when considering the average
of basic breathing variables calculated from breath selections as shown in Supplemental Table
3 including breath cycle duration (A), ventilatory frequency (B), and uncorrected tidal volume (C,
Fig. 2C). There is also a statistically significant difference in the irregularity scores for basic
variables when comparing breaths selected by BASSPRO to those excluded by BASSPRO as
shown for uncorrected tidal volume (Fig. 2D). We also see concordance between BASSPRO
and the experts with the averages of refined respiratory variables as shown in Supplemental
Table 3 including peak inspiratory flow (D), peak expiratory flow (E), ventilatory equivalents of
oxygen (F), corrected tidal volume (G), weight normalized ventilation (H), and oxygen
consumption (I).

Biological Application

AD is a chronic, neurodegenerative disease with only palliative treatments currently available.

As neurodegeneration worsens, key bodily functions including respiration begin to fail, which

directly or indirectly leads to death. Sleep-disordered breathing (SDB) and aspiration pneumonia

show strong associations with AD progression. There is a significantly increased prevalence of

SDB in patients with dementia compared to the general population (71% vs. 55.7%)[7–9], and the

E4 allele of apolipoprotein E, which is the strongest genetic risk factor for AD, is more highly

expressed in patients with sleep apnea[10]. Continuous positive airway pressure for treatment of

SDB in AD patients may improve symptoms[11]. Additionally, a leading cause of morbidity in AD

and other neurodegenerative pathologies is aspiration pneumonia, which is hypothesized to

result from disruption and degeneration of protective brainstem respiratory and swallowing

networks[12–19]. The summation of the evidence suggests that poor breathing function may play a

role in AD onset, progression, and severity, and that AD progression may create a mutually

destructive feedforward loop that disrupts respiratory networks.

To gain insight into respiratory function during AD progression, we used WBP on an inducible

amyloid precursor protein (APP) forebrain mouse model[20] to characterize metabolic and

respiratory function over nearly two years. Along with room air breathing (21% O_2/79% N_2),

hypercapnic (5% CO_2/21% O_2/74% N_2) and hypoxic (10% O_2/90% N_2) respiratory homeostatic

chemosensory functions were assayed. We focused on chemosensory respiratory reflexes as

these neurological reflexes are often perturbed in SDB and other respiratory pathophysiologies.

A total of 633 recordings were completed in the 2-year longitudinal Alzheimer's study following

55 mice across experimental and control genotypes. Beginning approximately 5 months after

cessation of Doxycycline treatment, immediately following the first appearance of amyloid

plaques, mice were recorded about every 2 months for 19 months. Timepoints reported in

figures and text refer to total time off Doxycycline treatment. Not all mice met inclusion criteria

for all timepoints; some died of age-related complications or recordings were not of sufficient

quality. Because our analyses are completed on a breath-by-breath level rather than average

values for each animal, we were able to include data from at least one period of the recording

for 44 animals, resulting in 302 recordings for hypercapnic ventilatory response and 301

recordings for hypoxic ventilatory response, which are included in the data shown in Figures 3-

6.

No statistically significant trend in respiratory variables was observed in the Alzheimer's group

compared to age-matched controls over time. While some timepoints showed significant

differences between experimental and control animals or within the same group over time, these

differences presented as random with no obvious biological cause. A number of the significant

differences were due to significant variability in one or more of the control groups: ventilatory

frequency (V_F) at 10 months in room air, and ventilation ($\dot{V}_E$) at 5 months in room air, 10 months

in room air, 12 months in room air, 19 months in room air, 7 months in 5% CO_2, and 10 months

in 5% CO_2 (Fig. 3A-B, Supplemental Table 4). There are also cases where just one or two of the

control groups differ from the experimental group: V_F at 5 months in 5% CO_2 and 12 months in

5% CO_2 (Fig. 3A, Supplemental Table 4).

Breathe Easy is designed so the user can easily assess the effects of grouped or individual

levels within independent variables across multiple variables in an experiment to uncover key

factors involved in differential outcomes. For example, in the case of V_F at 5 months in 5% CO_2,

we can see that this difference is driven completely by the males, whereas the difference at 12

months in 5% CO_2 no longer holds up when the control groups are combined, and sex is added

as a factor (Fig. 3C). Another example of deconvoluting outcomes can be found by examining

individual levels for sex and genotype variables in V_F and $\dot{V}_E$ at 7 months in 5% CO_2 where it

becomes clear that the difference is driven by a change in V_F, but not tidal volume, in the

females leading to a significant difference in $\dot{V}_E$ (Fig. 3C-E). However, this difference in $\dot{V}_E$

seems chiefly caused by the CaMKII-tTA control group when the control groups are split, and

sex is removed as a factor. This is in line with previous findings indicating phenotypes with the

CaMKII-tTA driver to be commonplace[21].

Upon further investigation into metabolism, we see similar trends for seemingly random

significant differences across a number of comparisons that do not have an obvious biological

cause; these significant results are also not sustained when independent variables are grouped

differently (Fig. 4A,C). However, we do note significant differences at 7 months as we saw with

$\dot{V}_E$ and V_F. In 5% CO_2 there is a significant difference between experimental and control animals

in ventilatory equivalents of oxygen ($\dot{V}_E/\dot{V}_{O_2}$), but not oxygen consumption ($\dot{V}_{O_2}$) (Fig. 4A-B). This

difference persists in both males and females for $\dot{V}_E/\dot{V}_{O_2}$ whereas no difference is noted in $\dot{V}_{O_2}$

(Fig. 4C-D). This suggests, combined with the data shown in Fig. 3B,E, that this difference is

driven by a change in $\dot{V}_E$ and not $\dot{V}_{O_2}$. Further, although the difference in $\dot{V}_E$ was not significant in

males (Fig. 3E), there is significance in males for $\dot{V}_E/\dot{V}_{O_2}$. This suggests that a trend in $\dot{V}_E$ for

males may be present, but not significant, and becomes clear when normalized to metabolic

demand.

On the other hand, in 10% O_2, we see no differences in $\dot{V}_E$ at 7 months (Fig. 3B,E), but we do

see significant differences in $\dot{V}_{O_2}$ and $\dot{V}_E/\dot{V}_{O_2}$ even when the data is split by sex (Fig. 4), which

suggests this difference is driven by a change in $\dot{V}_{O_2}$. Interestingly, we see a bimodal distribution

of the $\dot{V}_{O_2}$ response where males show a decrease and females show an increase. Despite this

differential response, the change in $\dot{V}_E/\dot{V}_{O_2}$ between control and experimental animals still holds

when sex is combined (Fig. 4B).

We also report the featured breathing phenotypes for apneas and sighs in this study where we

do not observe any meaningful changes in sigh rate that would be indicative of a phenotype

driven by this Alzheimer's model (Fig. 5B,D). For apnea rate, there are a number of significant

differences between controls and experimental animals in females under 5% CO_2 when split by

sex (Fig. 5A), but none of those changes remain significant when sex is removed (Fig. 5C).

There is also a difference in experimental animals between 5 and 7 months, and 5 and 16

months in 5% CO_2 (Fig. 5A), however this difference is also gone when sex is removed and

instead we see a difference in control animals between 7 and 12 months in 5% CO_2 (Fig. 5C).

Additionally, we can easily test for contributions of known related variables like weight and age.

No significant difference between experimental and control animals was seen in the number of

animals included or weight at the time of their recordings (Fig. 6A, Supplemental Table 5) aside

from the expected significant weight gain with age. There was a significant difference in age

between experimental and control animals at 12 months (E=experimental average±standard

error, C=control average±standard error, p-value) (E= 400.1±14.0, C= 393.6±12.4, p=0.0073),

14 months (E= 459.3±10.8, C= 452.8±12.3, p=0.021), 16 months (E= 518.1±8.67, C=

510.6±9.52, p=0.022), and 19 months (E= 601.2±6.46, C= 593.3±9.4, p=0.022) (Fig. 6B).

Although the ages are significantly different, we expect that this significance is derived from the

order by which the animals were run through our experimental paradigm. Unfortunately, it was

not considered that the genotypes and groups should be randomly recorded at each of the

timepoints, so the animals were run by group in the same order for each timepoint. This led to a

significant, but not biologically relevant difference in age at the final timepoints. This hypothesis

is supported by the consistent decrease in variation in age for both groups as age increases.

Discussion

Here we present a novel software platform that offers a significant improvement upon current

analysis methods for WBP data. While commercial options for partially automated analyses are

available, they are closed source and offer little flexibility for breath identification and filtering,

and subsequent analysis (Supplemental Table 2). While several approaches to analyzing

differing aspects of rodent respiratory data have been published in one form or another[22–26], no

available software suite, to our knowledge, has the capability of filtering for the typical standard

of calm, quiet breathing in rodent plethysmography experiments, i.e., automatically filters

movement and other waveform artifacts. In addition to programmatic benefits outlined in

Supplemental Table 2, Breathe Easy is open source, free to academic researchers, and offers a

novel, raw data-to-figure pipeline.

Conventional approaches to identification and selection of breaths for analysis fall into two

extremes: manual selections or inclusion of all breaths. The BASSPRO module of Breathe Easy

provides its users with the ability to identify all breaths within a recording and provide a filtered

selection of breaths that generally agree with expert selections of "calm, quiet" breathing. The

automated selections provided by BASSPRO have the advantage of being impartial and

tractably defined by user-adjustable settings that can easily be shared across multiple users,

eliminating the problem of inter-rater reliability when analyzing large datasets. Additionally,

BASSPRO avoids using the breathing rate of individual breaths as a selection filter. Instead, it

uses a sliding window that incorporates neighboring breaths to exclude breaths proximal to very

high frequency breathing, such as bouts of sniffing, while permitting inclusion of a larger range

of likely "calm, quiet" breaths throughout the full duration of the recording. This range is lost with

manual selection as experimenters are prone to selecting the slowest and steadiest breathing

intervals. BASSPRO also integrates information across other data channels to aid in defining

the filtering criteria for automated selections. BASSPRO additionally has the capability of

providing indirect calorimetry respiratory outcomes, which are essential as a means of

controlling for the effects of metabolic demand on breathing outcomes. Limitations of the current

implementation relate to the initial candidate breath identification settings as the same settings

are used for candidate breath detection for the full recording and for all recordings analyzed as

a batch. This may require a user to refine and repeat their analysis in cases where marked

differences in respiratory flow are present across animals, or across portions of the recording.

Breathe Easy's STAGG module is also unique in the application of appropriate statistical testing

for respiratory datasets. While many articles in the field apply ANOVA or t-tests, these tests are

mostly inappropriate for the structure of most data collected in plethysmography experiments,
which often involve repeated measurements on the same subjects, particularly for
chemosensory assays, and naturally involve multiple variables that influence breathing but are
not always considered, e.g., age, sex, and experimenter. Our application of linear mixed effects
models (LMEM) with Tukey post hoc analysis represents a needed standardization for analysis
in the field and, arguably, a significant improvement in statistical analysis, as the model
accounts for repeated measures and accommodates more complex experimental designs.
Several assumptions are made in the application of mixed effects models that should be
addressed by the researcher before interpreting results. We use a standard LMEM model for all
dependent variables, in which it is assumed that the random effect coefficients and the overall
model residuals follow normal distributions. Cases where the residuals are not independent and
identically distributed indicate potential invalidity of the model. While strict adherence to
assumptions with an LMEM is not always required[27], we do provide output to address these
assumptions. QQ and residual plots are generated for each dependent variable. Based on
these, the user can choose to include models with transformed dependent variables to address
adherence to these assumptions. Transformations that are currently available include $\log_{10}$,
natural log, and square root.

While the LMEM is widely used to analyze data in biological contexts, the standard model we
use in STAGG for all designated dependent variables has some potential limitations. One issue
is that the standard LMEM is only applicable for continuous dependent variables. If a user
wanted to analyze categorical or discrete dependent variables, this would require the usage of a
generalized linear mixed effects model, which is currently not implemented in STAGG. In cases
of optional graphs where this would be appropriate, an ANOVA is used instead of the primary
LMEM model. Additionally, the standard LMEM model does not account for any possible
dependencies between subjects or any correlations between observations outside of what is
accounted for in the random effects. For example, there could be autocorrelation effects from a
single session or possible additional repeated measures effects from the experimental design,
which the current implementation does not handle. However, it is possible for the user to run
STAGG to test these dependencies by selecting rig, experimental date, time of day, etc. from
the imported metadata as dependent variables to test for significant contributions.

Breathe Easy was validated against the typical standard of manual selection of recordings (Fig.
2). The concordance between BASSPRO selected breaths and all three experts demonstrate
the ability of the software to select breathing that an expert would deem calm, quiet breathing.

Furthermore, as sample size increases, we know the accuracy of estimated mean becomes

closer to the true mean. We see a significant difference between breaths selected by BASSPRO

and breaths excluded by BASSPRO; therefore, the accuracy of BASSPRO is not just due to an

increase in sample size and is instead due to the quality of selected breaths. We also see

significant variation between experts, which is not uncommon in respiratory experiments where

reproducibility both in values and units reported is questionable (Supplemental Table 1). This

problem can easily be solved by adoption of Breathe Easy across labs performing WBP

experiments to ensure consistent data analysis.

The high level of variance in agreement between experts, and between experts and BASSPRO-

selected breaths is encouraging as it supports and provides evidence for the argument of

increased consistency when using Breathe Easy compared to manual selection. It is not

uncommon in the field of respiratory studies for experiments to have unreliable reproducibility.

These data show that even with a blinded group of experts there is variation in the manual

selections that lead to significantly different results. Furthermore, while BASSPRO-selected

breaths are significantly different in many cases from expert-selected breaths, over 88% of the

breaths selected by experts were also selected by BASSPRO. This suggests that because

expert manual selection results in such a low sampling frequency of the data (~10% of the data

selected by BASSPRO), the variation between experts, and variation between experts and

BASSPRO is actually a result of the method of manual selection missing data leading to

undersampling. On the whole experts agree with BASSPRO as to what is quality breathing, but

not all of the quality breaths are part of the expert selections on the individual level.

Together this supports the hypothesis that Breathe Easy demonstrates exquisite accuracy not

solely because of the large number of breaths analyzed, but rather that BASSPRO's filtering

criteria extracts high quality breaths that includes the variety of calm, quiet breathing. Such

representative variation present during calm, quiet breathing is exceedingly difficult to capture

with manual selection.

Overall, the biological application of this software to our longitudinal Alzheimer's study failed to

reveal a consistent change over time, which suggests that forebrain APP expression alone is

not a likely significant driver of respiratory dysfunction. This finding is somewhat surprising as

forebrain streptozotocin AD models can drive changes in breathing rate[2,3]. Further, it is well

appreciated that forebrain injury from trauma or stroke can have lasting effects on breathing[28,29]

and that forebrain seizures are capable of perturbing breathing[30].

Of the results that were significant, we suspect that contributions of phenotypes previously

reported in the control mice for this animal model to play a significant role[21,31]. Given the

variability of the control groups and the lack of any biologically significant development at 7

months off doxycycline, the differences reported at 7 months are most likely outliers. On the

whole, there does not seem to be consistent evidence to support a direct link between the

forebrain pathology of this model and pathological respiratory phenotypes.

Many additional features and outputs are being developed for future versions of Breathe Easy.

The version presented here is a complete pipeline with calculation and analysis for the most

frequently reported basic and refined respiratory variables. Future versions will aim to provide

additional flexibility with user-defined breathing to include coughing and vocalization. We have

also derived iterations of Breathe Easy for exploratory and data mining functions to investigate

the dynamics of the waveform itself for predictive and diagnostic features. A component of these

dynamics will include analysis of first and second derivatives of each component of the

waveform for analysis of active vs. passive breathing and early/ late inspiration and expiration.

Using principal component analysis (PCA) of individual breath signals, it will be possible to

interrogate traces for unknown features and phenomena that inform respiratory outcomes. Many

groups are currently investigating the application of this analysis to raw waveforms, but the input

to their analyses is limited to all breaths and noise from a recording or hand-picked sections of

data[25,32]. Additionally, several exciting avenues of breathing research are opening up that would

benefit from modules to extract and merge other physiological signals such as ECG, EEG,

EMG, and video for automated generation of ethograms describing an animal's state as Breathe

Easy does not currently segregate breaths by state.

Methods

Animal ethics statement

All animal studies were conducted at the AALAC accredited animal facilities at Baylor College of

Medicine and all procedures were reviewed and approved by Baylor College of Medicine

Institutional Animal Care and Use Committee under protocol AN-6171.

Animal protocols

Breeding, genetic background, and maintenance of mice

Mice were maintained by mating a double Tg APP/tTA male on C57BL/6J background to a WT

FVB/NJ female resulting in offspring that were FVBB6 F1. All four groups of mice used in this

study were derived from this crossing: 1) wildtype (WT) FVBB6 (control), 2) Tg102_tetO-

APPswe/ind (tetO) (control, MMRRC #34845[20]), 3) TgCaMKIIα-tTA_LineB (CaMK) (control,

Jackson Laboratories #3010[33]), and 4) Tg102_tetO-APPswe/ind; TgCaMKIIα-tTA_LineB (Exp)

(experimental). WT control mice are included as background genetic controls. Groups 2 and 3

represent single transgenic controls wherein transgenic insertions included in the experimental

animals are present, but because only one of the transgenic alleles is included the resultant

APP phenotype is not present. Group 4 is the experimental group where both transgenic alleles

required for APP expression are present. Briefly, in the absence of doxycycline, tetracycline

transactivator expressed under the CaMKIIα promoter allows for the expression of the

tetracycline-responsive chimeric mouse APP. This version of APP contains a humanized Aβ

domain encoding the Swedish and Indiana mutations. Animals were housed on a 12h/12h light

cycle with food and water ad libitum. Offspring were started on a Dox diet during the first

postnatal week until 6 weeks of age when they were placed on regular chow. Animals were

measured at multiple timepoints from ~5 months after Dox withdrawal to ~19 months after Dox

withdrawal. There was no biologically significant difference in the weight or age of the mice in all

groups at each recorded timepoint (Supplemental Tables 5C-D).

Experimental design

A total of 4 genetic groups were used in this study as outlined above. See Supplemental Tables

5A-B for the number of mice and Supplemental Table 5D for the average age of mice included

for analysis for each timepoint, air exposure type, genotype, and sex. Briefly, we started with 11

WT controls, 12 tetO controls, 12 CaMK controls, and 11 Exp mice with sex evenly distributed

for each group. The number of animals recorded does not always reflect the number of animals

included in analysis as certain portions of the recording or entire recordings may not have been

of sufficient quality. These cases lead to the mismatching in number of mice at those timepoints

and conditions. Although halves of recordings may be included in analysis, our pipeline

evaluates the results based on all breaths rather than averages, so all accepted breaths for

each genotype at each time point for a given exposure are combined together to increase the

power and accuracy of our analyses.

Each timepoint consisted of recording the mice for baseline room air breathing followed by

either hypercapnic or hypoxic exposure. After a few days of recovery and reset, the same mice

were recorded a second time for baseline room air breathing followed by the opposite challenge

exposure, hypoxia or hypercapnia (Supplemental Fig. 2).

<u>Whole body plethysmography</u>

Whole body plethysmography on conscious, unrestrained mice was performed on ~7 month to

~21 month old adult mice[34]. Briefly, mice were habituated for 5 days and tested within 1 week of

the last day of habituation. Habituation consisted of ~5 minutes of handling, rectal temperature

and body weight measurement, and 30 minutes in a plethysmography chamber[35]. On the day of

testing, body weight and rectal temperature were recorded prior to the WBP experiment.

Animals were then placed in a water-jacketed, temperature-controlled (30°C), flow-through

plethysmography chamber and allowed to acclimate for 30 minutes in hydrated (via an inline

bubble chamber) room air gas (21% O_2/ 79% N_2). Following acclimation, the animal is recorded

in room air for up to 1 hour for baseline measurements. After room air, the gas is switched to

hydrated hypercapnic (5% CO_2/ 21% O_2/ 74% N_2) or hydrated hypoxic (10% O_2/ 90% N_2) gas

for a 20-30-minute exposure. Gas is then switched back to room air for up to one hour. Lengths

of recordings are adjusted as needed based on animal activity to ensure sufficient collection of

calm quiet breathing. The animal is removed from the chamber at the end of the last room air

period and rectal temperature is recorded upon removal, and 30 minutes after removal. Body

temperature from the beginning and end of the recordings are used to linearize body

temperature over the duration of the recording in BASSPRO for use in Drorbaugh and Fenn-

derived Bartlett-Tenney calculations[36,37]. Exiting gases are dried via a nafion tube embedded in

silicone desiccant prior to entry in gas concentration sensors.

<u>Data collection</u>

A Validyne DP45 differential pressure transducer was used to measure pressure changes

between the animal chamber and a reference chamber. This signal was converted by a CD15

carrier demodulator. Oxygen and carbon dioxide concentrations were measured using AEI

oxygen and carbon dioxide sensors and analyzers. Chamber temperature was measured using

a ThermoWorks MicroThermo 2 thermocouple. All data was recorded via an ADI PowerLab

analogue to digital recorder using LabChartPro in real time.

The raw signals produced during whole body plethysmography experiments are a time series of

voltages corresponding to the sensor readings over the course of the recording session. In the

experimental paradigm that we use, these sensors correspond to 1) the differential pressure

produced within the recording chamber due to inhalation and exhalation via the countercurrent

heating and cooling of respired air with a small contribution from compression and rarefaction

outside and within the animal that is equilibrated as a function of airway resistance, 2) the

temperature of the recording chamber, 3) the oxygen concentration of the gas leaving the

chamber, and 4) the carbon dioxide concentration of the gas leaving the chamber. While the

sampling frequency may vary, 1,000 Hz was used in the current data and has shown to be high

enough frequency to permit high resolution and low risk of aliasing. The breathing waveform can

be used to derive temporal and volumetric parameters related to breathing such as breath cycle

duration, breathing rate, and an uncorrected tidal volume. The use of a temperature-clamped

chamber to compare to body temperature provides a more accurate estimate of true tidal

volume[36]. The remaining channels of data corresponding to oxygen and carbon dioxide

concentrations can be used to calculate metabolic parameters such as oxygen consumption

($\dot{V}_{O_2}$), which can be combined with breathing parameters like ventilation ($\dot{V}_E$) to ascertain if

changes in breathing are consistent with metabolic demand.

<u>Consideration of data</u>

The whole-body chamber, where the animal is housed during plethysmography experiments, is

set up for near-balanced gas flow-through. During the experiment, the pressure fluctuations,

largely due to heating and cooling of inspired and expired air respectively with a small

contribution from compression and rarefaction of space inside and outside of the body in this

chamber are compared to that of a reference chamber. The reference chamber is a duplicate of

the containment device and does not house an animal during experimentation. Thus, any

pressure differentials recorded between the experimental chamber and the reference chamber

are from two sources. First, there may be slow changes in relative pressures due to inevitable

imbalances with inflow and outflow. To guard against chamber over pressurization, as we run at

a slightly higher inflow pressure, we connect the two chambers with a small gauge cannula to

act as a low frequency band pass filter that prevents rapid breathing and movement pressure

changes to transit the cannula. We also engineer a small leak to the atmosphere in the

reference chamber to maintain the overall system at room atmospheric pressure. The final

sources of pressure changes are respiration, respiration related behaviors (sniffing), and animal

movement. This simple design allows for precise measurement of animal respiratory patterns.

Optimization of flow characteristics ensures a stable baseline pressure waveform and minimizes

noise in the data to facilitate software analysis. When this system is combined with oxygen and

carbon dioxide sensors in line with the experimental chamber outflow, data can then be used to

determine even more key variables including, but not limited to, oxygen consumption and

carbon dioxide production, which enables the normalization of respiratory output to one or more

metabolic parameters. Data resulting from an experiment using this or similarly designed

custom systems are recorded using multi-channel voltage analog to digital recording devices

such as PowerLab or LabJack (open source). We record at 1,000 Hz for highest resolution and

minimal risk of aliasing, for each measured variable. Our software uses these multichannel

voltage outputs as the starting point of processing. A robust and well-thought-out respiratory

and metabolic system ensures high quality data with minimal artifacts for straight forward

analysis of outcomes.

Optimal plethysmography guidelines

While BASSPRO-STAGG can handle many common errors and mishaps that can occur during

a plethysmography experiment to minimize data loss, there are general guidelines that, if

followed, provide optimal recordings for analysis in BASSPRO-STAGG. Users need not re-

perform experiments to abide by these guidelines. However, if protocols are updated to adapt to

these considerations, then outcomes will most likely be far more accurate and reliable, and

fewer experiments resulting in subpar recordings.

<u>Adjustment of exposure protocol as needed</u>

It is common practice for labs to develop a standard operating procedure for gas exposures in

WBP. For example, 30 minutes of habituation, up to one hour of room air, 30 minutes of 5%

CO_2, and up to one hour of room air for challenge recovery, particularly for hypoxic challenges

and the examination of the post-hypoxic response. While this standardization provides an ideal

platform for ensuring consistency between recordings, unfortunately mice are not consistent. It

is imperative that the operator performing experiments peruse the recorded sections of each

condition of the experiment for quality breathing before proceeding to the next condition or

section of the experiment. Some animals may take longer to reach a calm state, evident when

movement artifacts are absent.

Habituation

While acclimation periods are most often included in a plethysmography run, habituation is less

frequently included in experimental design. Rodents do not fully habituate to new environments

within 30 minutes, particularly ones with significant differences from their home cages, e.g.,

size, foreign objects, and pressure. This lack of habituation results in increased noise as the

animal explores its new housing as well as aberrant recording values related to the animal

chewing on sensors and plugging gas intakes. To mitigate any contributions to noise derived

from lack of habituation, we recommend habituating rodents once a day for at least 5 days

where the last day of habituation is within a week of the plethysmography recordings[35].

Habituation should include anything performed during the actual experiment, e.g., temperature

recording, injections, handling, and housing inside a plethysmograph chamber.

Gas concentration calibration curve

To calculate gas concentrations accurately, the voltages recorded during the calibration phases

of the experiment are used to assign the voltage to a known concentration for each recording

session. This allows downstream calculations to be more accurate within the recording and

between recordings. Often, however, only room air and the challenge gas for a given

experiment are recorded during these calibration sessions. This creates a unique situation

where a completed calibration curve with two values is only possible for one type of gas output.

For example, if the user runs room air and 5% CO_2 balance room air calibration sessions for a

hypercapnic exposure, then only the CO_2 calibration curve will have two points, one at 0%

(room air) and one at 5% (5% CO_2 balance room air). Furthermore, the accuracy of the gases

used in these challenges is not 100%. Two different gas mixtures with an expected oxygen

concentration of 21% may have slightly different readings. To mitigate these issues, we

recommend that users 1) record calibration sessions for two types of challenge gases in

addition to room air and 2) do not change the "zero" set point following the first gas. This allows

the software to gather two calibration curve points for both types of gas that are analyzed, O_2

and CO_2. It also prevents user-derived voltage changes that deviate from the expected curves

and would influence concentration calculations later in the recording.

Breathing Analysis Selection and Segmentation for Plethysmography and Respiratory

Observations – BASSPRO

The overall purpose of the waveform analysis program, called BASSPRO, is to import signal

data and identify breaths within raw plethysmography data; calculate relevant respiratory

parameters from those breaths; filter and group breaths into selections; combine metadata with

the calculated respiratory parameters; and pass those data on to the statistics and graphing

component of the program. This is accomplished using a variety of packages and series of

operations that utilize the Python 3 programming language (packages summarized in

Supplemental Table 6A).

<u>Data import</u>

Metadata are provided through an externally generated .csv file. Metadata represent three main

categories: 1) periodic sampling during recording, 2) persistent animal metadata, and 3)

initializing experimental data (Supplemental Figure 3). Periodic sampling during the recording

are values obtained by the researcher during the recording that are specific to a given animal.

These data are typically intrasession drug administration or body temperature where continuous

body temperature is not always gathered in plethysmography experiments. Persistent animal

metadata are all the details about the animal being recorded that do not change (e.g., the

genotype, age, date of birth, sex, group (control or experimental), etc.). Initializing experimental

data refer to information about the apparatus and conditions on the day of the experiment that

may play a role in the outcomes measured. Examples of initializing data include the

identification number for the plethysmograph (if more than one is available in the lab), the gas

tank (if tanks are changed out while completing all recordings for a given experiment), body

weight, room temperature, and room barometric pressure. If an animal is run more than one

time, for example in a longitudinal study, then there will be initializing data for each run, but the

persistent animal metadata will remain the same for every run of that animal. An example .csv

with all required fields for BASSPRO can be accessed in our User Manual website hosted on

our project GitHub page. The inclusion of as many factors as possible in the metadata, such as

experimenter, specific gas tanks, and rigs, allows for the software-based evaluation of potential

recording artifacts that may impact outcomes, e.g., a particular tank or WBP rig having

significant error in the recorded data.

<u>Breath selection and filtering</u>

BASSPRO works by identifying every breath in a .txt file export of a WBP experiment followed

by filtering of those breaths for desired qualities. The inputs to BASSPRO are all .txt files for an

experiment and two or three configuration files, the number depends on the type of run the user

chooses.

BASSPRO uses the two or three configuration files generated within the GUI to select and filter

breaths including: 1) basic breathing parameters, 2) manual selections of breaths and 3)

automated selection criteria. Basic breathing parameters are requirements that are used by

BASSPRO to extract all breaths from the recording. The basic breathing parameters are
settings that denote whether a peak found within the breathing waveform is a breath based on
its characteristics, such as slope and duration, as well as using surrounding breaths to assess
the presence of artifacts. A series of quality control steps are also implemented for the non-
breathing signals such as chamber temperature and gas concentrations to identify data
collection artifacts. If artifacts are present, then automated corrections may be applied and
indicated in the log created by the run.

There are three ways BASSPRO-STAGG can be utilized to filter the identified breaths; manual
selection, automated selection and filtering, and manual selection and automated filtering. The
manual selection and automated selection criteria files inform BASSPRO on the three ways the
user can run our program. A manual selection file includes breaths that are selected by the user
from manual interrogation of the recorded plethysmography experiments. In this case, the user
will perform most of the work of BASSPRO themselves and the Breathe Easy run will provide
graphs and statistics. For automated selection, an automated selection criteria (autocriteria) file
will be provided where the user determines what qualifications they deem appropriate for a
breath to be considered calm, quiet breathing and which experimental conditions they would like
to include in analysis. For example, the user could want only breaths if they are within a
continuous group of at least 100 breaths that meet standards for calm, quiet breathing (based
on the parameters of current and neighboring breaths) during the experimental conditions of
room air and 5% CO_2. To accurately calculate tidal volumes, the criteria for selecting the
calibrated volume injections from room air and 5% CO_2 may also be denoted in the autocriteria
file. Default settings for a number of experimental paradigms come loaded as defaults in the
automated settings subGUI for the user to select from. For manual and automated runs, the
user provides their selections, and breathing based on both manual and automated filtering will
be identified and included in the output. Calibration information from either manual or automated
settings can be used, but if settings are provided in both configuration files, only the manual
settings will be applied.

For any type of run, calibration periods must be identified for any sections where respiratory
variables derived from Drorbaugh and Fenn, and Bartlett and Tenney formulae[36,37], which we
call refined variables, are desired as an outcome. If calibration injections are not provided, then
only basic breathing variables (breath cycle duration, uncorrected tidal volume, ventilatory
frequency) can be calculated. An example of both a manual selection file and an autocriteria file
are provided in the User Manual website found in the GitHub repository for this project. If

variables from the experimental or animal information metadata are not input by the user, then
most have default alternatives that are implemented so meaningful data can still be produced.
Generation of these settings files and choices that can be made are more explicitly outlined in
the User Manual website.

Export to STAGG

Multiple BASSPRO outputs are available including the main .json output that is utilized by the
STAGG portion of this software suite, a .csv output enumerating all detected breaths that can be
used to audit breath identification performance, and a .csv output providing aggregate summary
values. The .json files are tables where each row is one breath that includes basic breathing
variables, calculated respiratory variables, persistent animal metadata, initializing experimental
data, and periodically sampled data during the recording. Each column is a value from one of
those categories of data. One .json file is generated per animal per recording. A log file is also
produced that can be used to verify settings used during a run, observe if irregularities were
present in the signal files, or determine if corrections were applied during the processing of the
files. Once all files have been created, they are stored in the output folder specified in the GUI in
automatically generated folders that include the date and time of the run. Only the .json files are
passed along to STAGG.

Subprocesses of BASSPRO

BASSPRO performs all the filtering and selection of breaths as outlined above. However, there
are a number of subprocesses that BASSPRO also performs during a typical run. Some of
these processes are automatically performed and others are optional. It will be important for a
user to familiarize themselves with the implications of each of these subprocesses as most are
implemented to navigate a variety of common anomalies in plethysmography recordings. Also
outlined is a more detailed explanation of how quality breathing is extracted from the raw
waveforms.

Body temperature linearization

BASSPRO features a body temperature interpolation capability where body temperature
measurements taken at the beginning, middle, and end of a session can be used to interpolate
temperature values throughout the duration of the session for later use in calculating refined
respiratory variables derived from Drorbaugh and Fenn, and Bartlett and Tenney formulae[36,37].
Essentially, this function takes as many points as are available for temperature and creates a
linear interpolator that connects those measurements throughout the recording. This artificially
provides body temperature measurements for the entire recording rather than two to three

distinct temperatures copied throughout the recording as is currently applied[6,34]. Currently, this

interpolation cannot be turned off and is applied in all cases where temperature is not

continuously recorded.

<u>Signal spiking adjustments for temperature, flow, and gas concentration channels</u>

It is not uncommon for abrupt, high magnitude spikes or short square wave pulses to

sporadically contaminate the voltage signals on some of the recorded channels. BASSPRO can

apply a correction to the signal when this error is detected. For chamber temperature, when a

value outside of the range of 18-35°C is detected for an otherwise good breath, then the

chamber temperature is replaced either 1) with an interpolated value from surrounding

temperature values that fall within the correct range, or 2) with a default value, which is set to

30°C if no temperatures from surrounding breaths fall within the desired range. Both the range

of acceptable temperatures and the default value used for replacement can be changed in the

basic settings subGUI in the *Rig Specifications* tab.

It is also possible that the chamber temperature can exceed the recorded body temperature of

the animal, which will lead to errors in values calculated using the equations from Drorbaugh

and Fenn, and Bartlett and Tenney[36,37]. This particular error can be a result of drifting chamber

temperature voltages, or aberrant body temperature measurements. In either case, the user has

the option to exclude any breaths where the chamber temperature for that breath exceeds the

body temperature for that breath. These breaths are removed by default but can be changed by

editing the 'High chamber temperatures inclusion' variable in the *Additional Settings* tab of the

automated settings subGUI.

For aberrant spiking in flow and gas concentration channels, the user can apply a smoothing

filter, similar to flow smoothing filters commonly applied to WBP recordings. To apply the

smoothing filter to flow and gas concentration channels, simply change the value for 'Smoothing

Filter', which by default is 0=no smoothing on flow and 1=smoothing on flow, to "fg" in the basic

settings subGUI in the *Breath Inclusion* tab. This setting by default only applies to the flow

channel.

<u>Automatic identification, selection of breaths, and basic respiratory variables</u>

Breaths are identified through transitions between inspiratory and expiratory flows that are

above their respective and tunable thresholds. Flow signal used for this identification is

calculated as a simple first derivative of the original respiratory waveform as collected by a

differential pressure transducer used for whole-body plethysmography recordings. When a flow

is identified as crossing the threshold such that breathing has transitioned between inspiration
and expiration (or vice versa), then the point at which the waveform crosses zero immediately
preceding the threshold crossing is used as the time when the transition occurred. This permits
identification of timepoints defining inhalation and exhalation for each breath cycle.

An optional filter can be applied (or implemented in later steps for automated selection) that
utilizes the breathing frequency of neighboring breaths (Equation 1) and disagreement in
amplitude of inhaled and exhaled volumes (Equation 2) to select breaths.

Equation 1. Percentage determination of neighboring breaths above a set breathing
frequency where n is the instantaneous breath, w is ½ of the window surrounding the
breath, F is the instantaneous breathing frequency, and X is the threshold set by the
user (recommended as 500 for mouse breathing).

$$\sum_{n_{-w}-n_{-1},n_{+1}-n_w}^{K} K_n \begin{cases} 1 \; if \; F_n > X; \\ 0 \; otherwise \end{cases}$$

656   **Equation 2.** Calculation of inhalation/ exhalation volume disagreement.

$$\frac{|V_i - V_e|}{V_i/2 + V_e/2}$$

During the initial identification of breaths, several parameters are calculated for each breath,
which are available as part of the output. Some have been identified as useful for automated
selection criteria as well. The candidate breaths that are initially identified are then available for
manual or automated selection. Manual selection will place all breaths contained within
manually-defined intervals within the output. Automated selection relies on a series of filters to
refine the list of candidate breaths into the set of breaths contained within the largest continuous
interval for each experimental condition as well as the subset of those breaths that meet criteria
for "calm, quiet breathing".

Parameters that are defined for each breath include the duration of inhalation and exhalation,
the total breath cycle duration, the instantaneous breathing rate, uncorrected inhaled and
exhaled tidal volumes, peak inspiratory and expiratory flows, and scoring for featured breathing
such as being an apnea or sigh (1=True, 0=False). Apneas can be defined by duration relative
to neighboring breaths (at least twice as long recommended for mice) and minimum length of
apnea (0.5 seconds recommended for mice), and sighs are defined as breaths with increased

tidal volume relative to neighboring breaths (at least twice the amplitude is recommended for

mice). These definitions can be changed based on user requirements, but these defaults

represent standards of the field.

Irregularity scores are also calculated for these basic respiratory parameters relative to the

preceding breath. The calculation for irregularity scores is shown in Equation 3 where X is the

value for which an irregularity score is being calculated and N is the number of the current

breath for which the irregularity score is being calculated.

Equation 3. Irregularity score calculation.

680
$$\frac{|X_N - X_{N-1}|}{X_{N-1}}$$

To facilitate identification of calm quiet breathing, additional parameters are also calculated as

Boolean True/ False values based on the underlying parameters value relative to the thresholds

set in the autocriteria file.

<u>Automated sectioning by comments</u>

The user has the option of using comments inserted during the plethysmography recording as

experimental landmarks that BASSPRO can use to divide the recording. The initial release of

BASSPRO is designed to work with signal files exported from LabChart and can parse

timestamps placed using the LabChart software. Any recording that is exported with comments

in the style of the LabChart text file should also work given that comments in the autocriteria file

match those found in the recording. To use comments that are already contained in the signal

files, the user will include those as text strings in the autocriteria file, which is only required if the

user desires automated selection of breaths from their signal files. Each comment within the

autocriteria file is used to define the beginning of an experimental condition whereas the ending

of an experimental condition is defined by the next comment in the LabChart file that is also

listed in the autocriteria file, or by the end of the LabChart file if no suitable comment is found.

The first division of experimental conditions is into "verified conditions" that meet the basic

inclusion criteria based on their timestamps and gas concentrations. The user may specify

offsets from beginning or ending of an experimental condition as well as minimum and

maximum O_2 and CO_2 concentrations that can be used to refine the appropriate window for

selection of breaths (i.e., to account for and exclude ramping gas concentrations during

transition periods). These parameters are intended to be set by the user to define objective

criteria for inclusion. Verified experimental conditions that pass the inclusion criteria for the

condition are passed on to the next set of filters, verified breathing.

<u>Refined variable calculation</u>

Several respiratory parameters are automatically extracted from each identified breath whether

the breath is assigned a corresponding calibration period or not. These include, but are not

limited to, breath height, breath width, breath frequency, breath cycle duration, and uncorrected

tidal volume. Drorbaugh and Fenn-derived Bartlett-Tenney corrections are used to calculate

more refined estimates for respiratory volumes and metabolism[36,37]. To make those calculations,

calibration injections must be included so accurate volumes of gases inhaled and exhaled, and

baseline concentration voltages for oxygen and carbon dioxide concentrations can be

calculated. Variables that are calculated in this nature include corrected tidal volume (V_T),

minute ventilation ($\dot{V}_E$), oxygen consumption ($\dot{V}_{O_2}$), and ventilatory equivalents for oxygen

($\dot{V}_E/\dot{V}_{O_2}$).

It should be noted that the nature of calculating these values introduces a division transform

problem. This problem impacts the aggregate value calculated for an animal for a gas type or

condition. For example, you can either calculate these variables using the average of recorded

values for the animal during the gas exposure, or you can calculate these variables for each

breath and average at the end. These methods will yield similar, yet slightly different results with

their divergence typically increased as breathwise variation in breathing variable values

increases. With STAGG, our statistical model is built on the analysis of individual breaths, with

the ability to apply rank preserving transformations to the data to address non-normality in the

statistical model. We utilize the individual observations for the statistical analysis performed in

STAGG in order to account for the variability in individual measurements present in the

observed data. Using the individual breaths for analysis also increases the statistical power of

hypothesis tests relative to using aggregated mouse values. However, some users may prefer

the traditional approach of derivation of refined variables from pre-aggregated components, and

BASSPRO can provide an optional output file that makes these available.

Equations 4-9 elaborate the calculations made for each breath during a standard BASSPRO

run.

Equation 4. Vapor pressure calculation where T is the temperature (Kelvin) of the air

where vapor pressure is desired.

$$Vapor\ pressure = 1.142 + (0.8017 \times T) - (0.012 \times T^2) + (0.0006468 \times T^3)$$

Equation 5. Corrected tidal volume (V_T) calculation with Drorbaugh and Fenn-derived
Bartlett-Tenney correction where A is voltage from the breath, B is average voltage of
calibration injections, C is nominal volume of calibration injections, D is body
temperature (C), E is room barometric pressure (mmHg), F is vapor pressure of the
chamber, G is chamber temperature (C), and H is vapor pressure inside the mouse.

$$V_T = \left(\frac{A}{B}\right) \times C \times \frac{[(D + 273.15) \times (E - F)]}{[(D + 273.15) \times (E - F) - (G + 273.15) \times (E - H)]}$$

740   **Equation 6.** Minute ventilation calculation using corrected tidal volume (Equation 2).

$$\dot{V}_E = V_T \times V_F$$

742   **Equation 7.** Oxygen consumption ($\dot{V}_{O_2}$) calculation where oxygen in is taken from
743   voltage during calibration, and flowrate is recorded as initializing experimental data in
744   liters per minute (LPM) and converted to milliliters per minute (mLPM) in the equation. A
745   is flow rate (LPM), B is oxygen concentration in (%), and C is oxygen concentration out
746   (%).

$$\dot{V}_{O_2} = A \times 1000 \times \left(\frac{B}{100} - \frac{C}{100}\right)$$

748   **Equation 8.** Carbon dioxide production ($\dot{V}_{CO_2}$) calculation where carbon dioxide in is
749   taken from voltage during calibration, and flowrate is recorded as initializing
750   experimental data in LPM and converted to mLPM in the equation. A is flow rate (LPM),
751   B is carbon dioxide concentration out (%), and C is carbon dioxide concentration in (%).

$$\dot{V}_{CO_2} = A \times 1000 \times \left(\frac{B}{100} - \frac{C}{100}\right)$$

753   **Equation 9.** Ventilatory equivalents for oxygen ($\dot{V}_E/\dot{V}_{O_2}$) calculation using minute
754   ventilation (Equation 3) and oxygen consumption (Equation 4).

$$\frac{\dot{V}_E}{\dot{V}_{O_2}} = \frac{V_T \times V_F}{\dot{V}_{O_2}}$$

STAtistics and Graph Generator – STAGG

The overall purpose of the STAGG platform of Breathe Easy is to statistically analyze and graph
all of the desired variables for a given experiment. The flexibility coded into this component
allow the user to analyze almost any desired respiratory variable along with any metadata
external to the WBP recordings, like weight, age, or sex. Additional quantified data from paired

behavioral assessments or even immunohistological cell counts can also be included. STAGG

is comprised of a series of R scripts that process the data from BASSPRO. All packages can be

found in Supplemental Table 6B.

<u>Data import and statistical analysis</u>

First, the .json output of BASSPRO as well as the STAGG configuration files are loaded. Using

the rjson package the data import script reads in each .json file one at a time and prints to the

user the filename currently being imported as a progress bar. Once a .json is imported, it is

added to the end of the previously imported .json to create a large tibble containing all data from

every file. STAGG requires 3 configuration files: 1) variable configuration, 2) graph

configuration, and 3) other configuration. The variable configuration file outlines all the statistical

settings for a given experiment where independent, dependent, and covariate variables are

assigned. This file will also house permutations or settings specific to the variables that are set

by the user in the STAGG settings subGUI, which includes things like transformations,

minimums and maximums of the y-axis, and the generation of Poincaré plots. This file will also

include 'Alias' values that may be assigned by the user. For example, instead of VO2pg

appearing on a graph, the user may instead prefer Oxygen Consumption. In this case, the user

can reassign the name for VO2pg by replacing VO2pg in the 'Alias' column with Oxygen

Consumption. The graph configuration file contains instructions on how to plot the dependent

variables selected on the graph. For example, if you want to see your dependent variables

graphed by air condition and sex, then you would select air condition as xvar and sex as

pointdodge. Finally, the other configuration file outlines optional graphs that are selected by the

user. In optional graphs you can graph variables that may not be respiratory variables, but they

are metadata that you want to plot for the experiment. You can also plot your respiratory

variables with different settings than those in the primary model, which are selected in the graph

settings. Settings for optional graphs follow the same guidelines as those for variable and graph

configuration. Explicit instructions and more details for assigning values and making selections

for these files are included in our User Manual website hosted on our GitHub page for this

project.

Assignments for independent, dependent, and covariate variables are employed to fit separate

linear mixed effects models from the data for each of the designated dependent variables using

the *lmer* function from the *lme4* package in R as well as any user-selected transformations for

these variables. The independent variables are included in the model as fixed effects, along with

an interaction variable which involves all designated independent variables. On the other hand,

only the main effects for the covariates are included in the model as fixed effects. The unique

mouse identification number (MUID) is included in the model as a random effect. Once these

are set, a loop processes each of the dependent variables one at a time with the assigned

independent and covariate factors. Results of the LMEM then undergo post hoc analysis with a

Tukey's honestly significant difference (HSD) test using the *glht* function from the *multcomp*

package, in which pairwise differences between different interaction groups are tested for

statistical significance. The outputs of this final test are compiled into a single table. This is then

filtered to keep only biologically relevant significant results, i.e., only results which are both

statistically significant and for which there is only a single difference amongst the combined

group levels of the interaction variables. These results are used during the graph generation

process to place asterisks that signify significant differences where appropriate. Full tabular

results from the mixed effects model and the Tukey post-hoc tests are saved in an Excel

spreadsheet where each dependent variable is given a tab in the sheet that holds results for

each comparison made, and QQ and residual diagnostic plots for each individual mixed effects

model are saved in the same StatResults folder.

For optional graphs, the primary LMEM model are used for statistics unless the response

variable chosen either 1) is not included in that model or 2) is not a continuous variable. If 1,

then a new LMEM is run with the selected graphing options used in that model as independent

variables. If 2, then an ANOVA is used to generate statistics for that graph. Results of these

statistical tests are saved in a separate statistics folder from the primary LMEM.

<u>Graph generation</u>

All dependent variables assigned by the user are graphed with statistical analysis results added

to each. Settings from the GUI that are stored in the graph configuration file designate how the

variables are plotted, i.e., which variables the dependent variables will be graphed against.

Once these are assigned, scatterplots are generated via the *ggplot* function which show each

dependent variable plotted against the selected independent variables or covariates at the

individual mouse level. If there are significant results in the Tukey output, then asterisks are

added to the appropriate group based on the location identified for each comparison and dotted

lines are added to denote the groups that are statistically significantly different from one

another. After each component has been added to the graph, the graphs are saved as either

.svg or .jpeg files in the designated output folder.

 Optional graphs

The final graph making script allows the user to choose any combination of independent and/ or

dependent variables for plotting. This can include anything from body weight and body

temperature, to plotting the dependent variables by different independent or dependent

variables. This flexibility also opens the possibility for graphing results from behavioral studies

and other tests performed outside of the plethysmography recording.

R Markdown and ReadMe

STAGG is completed by the run of an R markdown script that grabs all .svg or .jpeg files that

are present in the selected output folder and compiles them into an .html page where the user

can view all graphs generated in one place. This feature is ideal for the user because it allows

perusal of data without needing to open multiple individual files.

In addition to an .html file with all figures produced, a ReadMe word document is produced that

houses all of the settings the user selected for that particular run of STAGG. It eliminates all

variables that were not assigned a role in the variable configuration file to make perusal easier

for the user. This ReadMe file is ideal for keeping track of different runs of the software and

taking notes on adjustments or changes that should be made from that particular run.

Graphical user interface – GUI

The GUI is a PyQt5-built program that integrates BASSPRO and STAGG into a portable

desktop application. The GUI's design follows the standard workflow of plethysmography

analyses. Lateral panels display the paths of selected input for both modules, providing

immediate feedback to the user on their selections. At the bottom of the GUI is the hangar

display, where more explicit feedback as well as terminal output is posted for user review. On

either side of the hangar display, the user can launch either module of Breathe Easy, or launch

them sequentially. The GUI addresses 3 main issues with running BASSPRO-STAGG: 1)

portability, 2) processing power/ speed, and 3) settings and formatting input.

Various subGUIs are used to provide specific input required by BASSPRO or STAGG. They are

designed with the user in mind in that they prompt the user to input settings that are coded

within BASSPRO-STAGG without needing to open, understand, or edit the code or

configuration files directly. These subGUIs are easy to navigate and walk the user through all

definitions of parameters required as well as recommended settings based on our experience.

There are five total subGUIs: 1) basic BASSPRO settings, 2) manual BASSPRO settings, 3)

automated BASSPRO settings, 4) metadata settings, and 5) STAGG settings. Explicit

instructions on using the GUI and subGUIs and expectations for their inputs can be found on the

User Manual website hosted in our GitHub repository.

<u>Portability</u>

A critical feature of the BASSPRO-STAGG application is its portability. The program employs a

variety of languages, libraries, and packages from the Python and R universes. To

accommodate novice users that are not skilled in setting up Python and R environments, a

distributable pre-compiled package was created permitting installation as easy as downloading

an archive file, unzipping the contents, and running the executable file contained within. Nuitka

and R-Portable were utilized to create packaging that allows the user to run the application

without needing to download or run Python or R. To facilitate this portability, the GUI creates

default directories and identifies relative paths that allow the program to run independent of local

environments. Currently, our GUI only works with PC machines. However, the code can be run

manually on a Mac, and we are in the process of creating a Mac executable to open and run the

GUI.

<u>Parallel processing</u>

Another critical global feature of the GUI is its implementation of parallel processing. The GUI

uses the *concurrent.futures* module to spawn separate processes for every signal file, allowing it

to implement asynchronous execution for signal file interrogation. The user can choose the

number of CPU cores they wish to devote to BASSPRO, which allows them to continue

experiencing adequate performance while running other programs on the same computer. This

parallel processing greatly reduces runtime, depending on configuration.

If parallel processing is not chosen by the user, then BASSPRO defaults to threading the signal

files sequentially through BASSPRO. On average, the interrogation of a single signal file

representing a 1.5 to 2 hour-long plethysmography run takes two to three minutes on

contemporary computer hardware. Running 16 signal files through BASSPRO one at a time

would take 32-48 minutes. Running those same 16 signal files through BASSPRO using parallel

processing, with at least eight processing cores, can finish the task within six minutes. On

average, we see an 81.8% reduction in run time with parallel processing compared to serial

processing. Note that the majority of consumer computers are limited to eight processing cores.

This is a dramatic contrast to the 50 minutes-3.15 hours spent by experts to manually select

breaths from the same 16 files. Either option, parallel or sequential processing, is an

improvement on the time and labor cost. It should be noted that although all cores can be sent

instructions during the parallel processing procedure, with contemporary machines, there should

be enough processing power to continue operating other applications while BASSPRO-STAGG
is running. This said, depending on configuration, running other applications during BASSPRO-
STAGG can potentially slow the processing time as the full processing power of each core
would not be fully dedicated to the pipeline.

Currently, BASSPRO is the only module of the pipeline that utilizes this feature. Utilizing parallel
processing for STAGG in its current form would not work or provide any benefit as there are
many steps requiring sequential reading and writing throughout the STAGG module that make
parallel processing impossible. However, once a dataset has been run through STAGG, one of
the outputs is an R environment created just after all of the input .json files are read into R. If an
R environment is selected as the start point for STAGG rather than all of the .json files, then
there is an 82.2% reduction in run time. So, there is still an opportunity to increase the speed at
which STAGG runs without parallel processing.

Sub-GUIs

Each subGUI discussed below provides specific input required by BASSPRO or STAGG. They
are designed with the user in mind in that they prompt the user in layman terms to input settings
that are coded within BASSPRO-STAGG in a less intuitive fashion. These subGUIs are easy to
navigate and walk the user through all definitions of parameters required as well as
recommended settings based on our experience. There are 5 total subGUIs: 1) basic
BASSPRO settings, 2) manual BASSPRO settings, 3) automated BASSPRO settings, 4)
metadata settings, and 5) STAGG settings. Not all the subGUIs are required for each type of
run (i.e., manual vs automated). Each of the subGUIs produces a uniquely formatted .csv file
that is sent to the appropriate program to provide the required instructions. Please see our User
Manual website for more information and screenshots.

<u>Basic BASSPRO settings subGUI</u>

There are four groups of data and user definitions input in the basic parameters subGUI. These
groups are separated into four tabs at the top of the subGUI with a fifth tab that summarizes all
of the selections made in the other four tabs. Please see the User Manual website for more
information on choosing settings in this subGUI.

Breath calling settings describe minimum quantitative guidelines for a peak in a recording to be
considered a breath. This does not define the breath as calm, quiet breathing, rather just
identifies that peak as breathing. Breath calling also defines featured breathing that may not
pass quality breathing filters set later, like apneas and sighs. Defining these features allow the

user to quantify these kinds of breaths separately from calm, quiet breathing. Rig specifications

are where settings specific to your plethysmograph will need to be detailed for use in respiratory

calculations from the raw waveform. Breath inclusion filters are very basic filters that can be set

for an initial pass of exclusion of breaths. Generally, the defaults set here are recommended for

a first pass through data. Output settings allows the user to indicate in which formats, other than

the .json files, they would like the output of BASSPRO saved. Options here can be helpful for

perusing the raw data in Excel.

<u>Manual BASSPRO settings subGUI</u>

The manual settings subGUI streamlines the creation of the manual selection file needed for

BASSPRO to analyze signal files with breaths that have been manually selected. The subGUI

requires the user to select a folder containing the .txt files produced by exporting LabChart

DataPad views. If a user is making selections with a software other than LabChart, then those

selections need only be formatted to match the style and headings of DataPad selections found

in LabChart for this subGUI to be used. Alternatively, they can directly generate a manual

selections file using our example file provided within the documentation for the User Manual.

These exported files are collated into a table presented in the subGUI window, allowing the user

to review their selection of signal files. The user specifies the type of experimental condition

applied during the plethysmography runs represented by the chosen signal files. Based on the

user's selection of experimental condition, the subGUI provides default settings for BASSPRO

to use during its analyses. These settings are presented as a table in the subGUI that the user

can review and edit, as needed. Upon merging the collated DataPad views with the settings

chosen, the subGUI creates the manual selection file by assigning the appropriate settings to

each row of the collated DataPad views. This table is then exported to BASSPRO for analysis.

There are a number of assumptions made based on the file names for the recording files and

exported DataPads. Please see the User Manual website for more information on how to export,

guidelines for naming signal and settings files, and what options are available for manual

selection.

<u>Automated BASSPRO settings subGUI</u>

The automated settings subGUI facilitates the creation and modification of the automated

criteria (autocriteria) file needed for BASSPRO to analyze signal files by identifying breaths and

features, filtering those breaths, and calculating output measures. The subGUI only asks the

user to specify the experimental paradigm relevant to the signal files they wish to analyze. Their

selection will populate the subGUI with values based on the default settings for that

experimental condition. The user can also choose to load a previously made autocriteria file that
will populate the subGUI with the settings defined therein. Alternatively, the user can derive all
new settings based on comments pulled directly from the signal files. This option requires more
work and should be done with caution and help from our User Manual. Automated selection
settings vary depending on the section of the plethysmography run, so each setting will have
several values, one for each of the sections. The user can review the values of all sections for
each setting and edit those values as needed. These modified settings can then be saved as a
new autocriteria file that will dictate how BASSPRO detects and filters the breaths in the signal
files. Once generated, these settings are collated into a .csv file that is exported to BASSPRO.
Suggested default settings can be found in a table hosted in our User Manual website.

<u>Metadata settings subGUI</u>

The metadata settings subGUI was designed to streamline the codification and reorganization
of variables found in the metadata. This subGUI allows the user to rename variables, convert
continuous variables into categorical variables, and create novel groups of variables. It modifies
the metadata intended as BASSPRO input without altering the original format, instead creating
a new file that can be used for subsequent BASSPRO analyses. Continuous variables can be
difficult to visualize in a graph, so creating bins for ranges of values can allow the user to
convert continuous variables into categorical variables for an easier to interpret graphical output,
e.g., one month and two months instead of P30, P32, P59, and P64. Novel groups of variables
allow the user to group animals together into meaningful biological groups for easier comparison
even if the values for their grouping are different, e.g., Control and Experimental instead of Tg,
WT, Het. Each of these permutations of the data are added to the existing data, which allows
the user to re-load the same data files and perform different runs with the same dataset based
on the exact question they would like to answer.

<u>STAGG settings subGUI</u>

The STAGG settings subGUI was designed to facilitate the configuration of statistical and
graphical parameters used in STAGG. It presents the user with variables sourced from the
metadata and analysis parameters provided as BASSPRO input as well as variables derived
from BASSPRO output. This collation allows the user to assign statistical and graphical roles to
variables of interest by designating them as independent, dependent, or covariate for statistical
analyses and choosing how they are presented in the figures produced by STAGG. This
interface also permits renaming of variables, allowing the user to create new naming
conventions for a variable that they would like to be displayed on a graph, e.g., Ventilatory

Frequency instead of VF. Please see our online User Manual for more information and help
understanding some of the statistics employed in our software.

Machine requirements

General machine requirements to run BASSPRO-STAGG vary based on the size of the user's
dataset. In general, we recommend twice as much RAM as the size of the full dataset. The user
should have a minimum of 5GB disk space to download and install and a minimum of 4 GB
DDR3 RAM, which should be sufficient for a small dataset. We recommend 50 GB disk space to
download, install, and store datasets with 16 GB DDR3 RAM. If datasets are highly complex,
i.e., more than 20 combinations of independent factors and more than 50 recordings, then 32
GB+ DDR3 RAM are ideal. Outside of this recommendation, minimum and recommended
operating systems and system architectures are described in Supplemental Table 7.

Later STAGG development was largely affected by the demands of working with very large
datasets. The size of the data table during importation, the size of the vectors produced during
modeling, and the number of pairwise comparisons each represented bottlenecks in
performance. STAGG loads the BASSPRO output into a single table that is actively maintained
in RAM while the program is building it during importation. A crude indication of whether or not
the computer running STAGG has sufficient RAM to handle the dataset is to compare that RAM
to the collective size of the input files. A computer with 16 GB of installed RAM cannot create a
table collating data from input files that collectively take up more than 16 GB of memory. There
are two workarounds: 1) use a computer with more RAM, or 2) reduce the size of the input files
by removing data that will not be used in the analysis. The removal of data could be removing a
continuously recorded channel or metadata that is not used.

Similarly, sufficient RAM is needed to maintain the vector(s) created when STAGG models the
data. This issue is again related to the size of the dataset with which STAGG is working.
However, an additional factor at play is the number of comparisons made in the model. You
may have insufficient RAM to maintain a few large vectors or many smaller vectors. Additionally,
one of the libraries used to build the model caps the number of pairwise comparisons at 1,000.
These related, but separate, issues were addressed by only modeling comparisons of interest,
which are defined as differing at only one level of the independent variables being compared.
This restriction has proven to be an acceptable and encompassing definition for biologically
relevant comparisons and, therefore, an acceptable filter.

Data Availability Statement

1019

All of our source code as well as a downloadable launcher for the GUI is available in our public

GitHub repository (https://github.com/MolecularNeurobiology/Breathe_Easy). Within the

ReadMe are multiple links to our online documentation (Sphinx and GitHub Pages). We highly

suggest first time users to install and launch the GUI, then walk through the vignette provided

that includes an example dataset and settings files.

Raw data for our complete Alzheimer's dataset is too large to store on Zenodo, which is our

default data hosting site. However, we will share the data via cloud-based services upon

request. Please send correspondence to Russell Ray at Russell.ray@bcm.edu for requesting

access.

References

1. Spinieli, R. L., Ben Musa, R., Kielhofner, J., Cornelius-Green, J. & Cummings, K. J. Orexin contributes to eupnea within a critical period of postnatal development. *Am. J. Physiol. Regul. Integr. Comp. Physiol.* **321,** R558–R571 (2021).

2. Brown, A. G., Thapa, M., Hooker, J. W. & Ostrowski, T. D. Impaired chemoreflex correlates with decreased c-Fos in respiratory brainstem centers of the streptozotocin-induced Alzheimer's disease rat model. *Exp. Neurol.* **311,** 285–292 (2019).

3. Ebel, D. L., Torkilsen, C. G. & Ostrowski, T. D. Blunted Respiratory Responses in the Streptozotocin-Induced Alzheimer's Disease Rat Model. *J. Alzheimers Dis.* **56,** 1197–1211 (2017).

4. Sun, J. J., Huang, T.-W., Neul, J. L. & Ray, R. S. Embryonic hindbrain patterning genes delineate distinct cardio-respiratory and metabolic homeostatic populations in the adult. *Sci. Rep.* **7,** 9117 (2017).

5. Ward, C. S. *et al.* Loss of mecp2 function across several neuronal populations impairs breathing response to acute hypoxia. *Front. Neurol.* **11,** 593554 (2020).

6. Lusk, S. *et al.* A CRISPR toolbox for generating intersectional genetic mouse models for functional, molecular, and anatomical circuit mapping. *BMC Biol.* (2022).

7. Liguori, C. *et al.* Sleep-disordered breathing and the risk of Alzheimer's disease. *Sleep Med. Rev.* **55,** 101375 (2021).

8. Liguori, C. *et al.* Orexinergic system dysregulation, sleep impairment, and cognitive decline in Alzheimer disease. *JAMA Neurol.* **71,** 1498–1505 (2014).

9. Macheda, T. *et al.* Chronic Intermittent Hypoxia Induces Robust Astrogliosis in an Alzheimer's Disease-Relevant Mouse Model. *Neuroscience* **398,** 55–63 (2019).

10. Osorio, R. S. *et al.* Interaction between sleep-disordered breathing and apolipoprotein E genotype on cerebrospinal fluid biomarkers for Alzheimer's disease in cognitively normal elderly individuals. *Neurobiol. Aging* **35,** 1318–1324 (2014).

11. Ancoli-Israel, S. *et al.* Cognitive effects of treating obstructive sleep apnea in Alzheimer's disease: a randomized controlled study. *J. Am. Geriatr. Soc.* **56,** 2076–2081 (2008).

12. Troche, M. S., Huebner, I., Rosenbek, J. C., Okun, M. S. & Sapienza, C. M. Respiratory-swallowing coordination and swallowing safety in patients with Parkinson's disease.

1059  *Dysphagia* **26,** 218–224 (2011).

13. Yagi, N. *et al.* Inappropriate Timing of Swallow in the Respiratory Cycle Causes Breathing-Swallowing Discoordination. *Front. Physiol.* **8,** 676 (2017).

14. Pitts, T. Airway protective mechanisms. *Lung* **192,** 27–31 (2014).

15. Fuzik, J. *et al.* Integration of electrophysiological recordings with single-cell RNA-seq data identifies neuronal subtypes. *Nat. Biotechnol.* **34,** 175–183 (2016).

16. Kalia, M. Dysphagia and aspiration pneumonia in patients with Alzheimer's disease. *Metab. Clin. Exp.* **52,** 36–38 (2003).

17. Won, J. H., Byun, S. J., Oh, B.-M., Park, S. J. & Seo, H. G. Risk and mortality of aspiration pneumonia in Parkinson's disease: a nationwide database study. *Sci. Rep.* **11,** 6597 (2021).

18. Wada, H. *et al.* Risk factors of aspiration pneumonia in Alzheimer's disease patients. *Gerontology* **47,** 271–276 (2001).

19. Suttrup, I. & Warnecke, T. Dysphagia in parkinson's disease. *Dysphagia* **31,** 24–32 (2016).

20. Jankowsky, J. L. *et al.* Persistent amyloidosis following suppression of Abeta production in a transgenic model of Alzheimer disease. *PLoS Med.* **2,** e355 (2005).

21. Jankowsky, J. L. & Zheng, H. Practical considerations for choosing a mouse model of Alzheimer's disease. *Mol. Neurodegener.* **12,** 89 (2017).

22. Wilson, C. *SWAN: Single Wave Analysis Notebook*. (GitHub, 2016).

23. Sun, Y. *et al.* INSMA: An integrated system for multimodal data acquisition and analysis in the intensive care unit. *J. Biomed. Inform.* **106,** 103434 (2020).

24. Hsieh, Y.-H. *et al.* Brainstem inflammation modulates the ventilatory pattern and its variability after acute lung injury in rodents. *J. Physiol. (Lond.)* **598,** 2791–2811 (2020).

25. Sunshine, M. D. & Fuller, D. D. Automated classification of whole body plethysmography waveforms to quantify breathing patterns. *Front. Physiol.* **12,** 690265 (2021).

26. Khurram, O. U., Gransee, H. M., Sieck, G. C. & Mantilla, C. B. Automated evaluation of respiratory signals to provide insight into respiratory drive. *Respir. Physiol. Neurobiol.* **300,** 103872 (2022).

27. Schielzeth, H. *et al.* Robustness of linear mixed-effects models to violations of distributional

assumptions. *Methods Ecol. Evol.* (2020). doi:10.1111/2041-210X.13434

28. Seiler, A. *et al.* Prevalence of sleep-disordered breathing after stroke and TIA: A meta-
analysis. *Neurology* **92,** e648–e654 (2019).

29. Koutsoukou, A. *et al.* Respiratory mechanics in brain injury: A review. *World J. Crit. Care*
Med. **5,** 65–73 (2016).

30. Dlouhy, B. J. *et al.* Breathing Inhibited When Seizures Spread to the Amygdala and upon
Amygdala Stimulation. *J. Neurosci.* **35,** 10281–10289 (2015).

31. Han, H. J. *et al.* Strain background influences neurotoxicity and behavioral abnormalities in
mice expressing the tetracycline transactivator. *J. Neurosci.* **32,** 10574–10586 (2012).

32. Janke, E. *et al.* Machine Learning to Classify the Emotional States of Mice from
Respiration. *SSRN Journal* (2022). doi:10.2139/ssrn.4106834

33. Mayford, M. *et al.* Control of memory formation through regulated expression of a CaMKII
transgene. *Science* **274,** 1678–1683 (1996).

34. Ray, R. S. *et al.* Impaired respiratory and body temperature control upon acute serotonergic
neuron inhibition. *Science* **333,** 637–642 (2011).

35. Martinez, V. K. *et al.* Off-Target Effects of Clozapine-N-Oxide on the Chemosensory Reflex
Are Masked by High Stress Levels. *Front. Physiol.* **10,** 521 (2019).

36. Bartlett, D. & Tenney, S. M. Control of breathing in experimental anemia. *Respir Physiol* **10,**
384–395 (1970).

37. Drorbaugh, J. E. & Fenn, W. O. A barometric method for measuring ventilation in newborn
infants. *Pediatrics* **16,** 81–87 (1955).

1109

Acknowledgements

We thank J. Sun, V. Martinez, and F. Saldana for technical support. Funding from NIH and NSF. The funders had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript.

Funding

S. Lusk: NIH R01HL130249, 44617-S4 (Diversity Supplement) and NIH 1F32HL160073-01A1

G. Allen: NSF NeuroNex-1707400

J. Jankowsky: NIH RF1 AG054160

R. Ray: NIH R01 HL130249, NIH R01 HL161142, McNair Foundation

This project was supported by the Mouse Metabolism and Phenotyping Core at Baylor College of Medicine which is subsidized through Baylor College of Medicine's Advanced Technology Cores and through funding from NIH (UM1HG006348, R01DK114356)

Author Information

Authors and Affiliations

Department of Neuroscience, Baylor College of Medicine, Houston TX

Savannah Lusk, Andersen Chang, Avery Twitchell-Heyne, Shaun Fattig, Joanna Jankowsky & Russell Ray

Department of Integrative Physiology, Baylor College of Medicine, Houston TX

Christopher Ward

Departments of Electrical and Computer Engineering, Statistics, and Computer Science, Rice University, Houston TX

Genevera Allen

Departments of Molecular and Cellular Biology, Neurology, and Neurosurgery, Baylor College of Medicine, Houston TX

Joanna Jankowsky

1140

1141 *Contributions*

1142 Experimental design (S.L., C.S.W., J.J., R.R.); software development (S.L., C.S.W., A.C.,

1143 A.T.H., S.F.); animal preparation (J.J.); data collection (S.L., R.R.); analysis design and

1144 implementation (S.L., C.S.W., A.C., A.T.H., S.F., G.A., J.J., R.R.); manuscript preparation (S.L.,

1145 C.S.W., R.R.).

1146

1147 *Corresponding author*

1148 Correspondence to Russell S. Ray.

1149

1150 **Ethics Declaration**

1151 The authors declare no competing interests.

1152 **Figures**

Figure 1. Overview of Breathe Easy components and workflow. A) Screenshot of the main GUI interface where the red numbering corresponds to displays shown in B-C. Text shown in white boxes are not what the user will see in the GUI, rather an example of what will populate those areas when using the GUI. B) A walk through of the full workflow of Breathe Easy with reference to each component in the GUI. C) Workflows and outputs of all of our subGUI interfaces.

Figure 2. Demonstration of breath selection validation where A) shows selection process behind the scenes of BASSPRO where gray overhead dots show all breaths identified and the red and green dots below the trace show the decision to include (green) or exclude (red) each breath and B) shows a comparison of breaths selected by BASSPRO to those manually selected by 3 experts from the same 16 files. C-D are uncorrected tidal volume and irregularity score of uncorrected tidal volume for breaths selected by BASSPRO vs manual selections by 3 experts. Error bars = standard error mean, * = $p < 0.05$.

Figure 3. Respiratory outcomes from our 2-year longitudinal Alzheimer's study. A) Ventilatory frequency and B) Ventilation plotted by timepoint, genotype, and condition with sex considered in the statistical model as a covariate. C) Ventilatory frequency, D) Tidal volume, and E) Ventilation plotted by timepoint, group, condition, and sex. Error bars = standard error mean, * = p<0.05.

Figure 4. Respiratory and metabolic outcomes from our 2-year longitudinal Alzheimer's study. A) Oxygen consumption and B) Ventilatory equivalents of oxygen plotted by timepoint, genotype, and condition with sex considered in the statistical model as a covariate. C) Oxygen consumption and D) Ventilatory equivalents of oxygen plotted by timepoint, group, condition, and sex. Error bars = standard error mean, * = $p<0.05$.

Figure 5. Apnea and sigh outcomes from our 2-year longitudinal Alzheimer's study. A) Apnea and B) Sigh plotted by timepoint, group, condition, and sex. C) Apnea and D) Sigh plotted by timepoint, group, and condition with sex considered in the model as a covariate. Error bars = standard error mean, * = p<0.05.

Figure 6. A) Body weight and B) Age of animals used in our longitudinal Alzheimer's study plotted by timepoint and group. Error bars = standard error mean, * = p<0.05.

Supplemental Information

Supplemental Figure 1. Screenshots of each of our subGUIs that can be launched from our main GUI: A) automated settings, B) manual settings, C) basic settings, D) metadata configuration, and E) STAGG settings.

4 Genotypes * 2 Sexes * 8 Timepoints * 3 Conditions = 192 levels

Supplemental Figure 2. Schematic of our longitudinal Alzheimer's study where average age at each timepoint across the length of the experiment is shown followed by a detailed protocol followed with each animal at each timepoint. Also provided is an explanation of variables included in analyses for this study.

Data collection

Continuous sampling from recordings

Respiratory trace (differential pressure)
Oxygen level (percent)
Carbon dioxide level (percent)
Chamber temperature (Celsius)

Raw recordings

Periodic sampling during recordings

Animal body temperature (Celsius)
*May be included in continuous sampling depending on set-up

Persistent animal metadata

Genotype
Group (control vs. experimental)
DOB
Age at time of experiment
Sex
etc.

Initializing experimental data

Gas tank
Rig number
Animal body weight (g)
Room temperature (Celsius)
Room barometric pressure (mmHg)
Researcher performing experiment
etc.

Metadata

Supplemental Figure 3. Schematic of variables collected for a plethysmography experiment that can/ must be included in Breathe Easy for analysis.

Supplemental Table 1. Non-exhaustive literature review of reported respiratory variables from 1998-2022 from a variety of laboratories across the world. Values were taken from graphs as most publications did not report raw values. Only values for control, adult mice were used to populate this table.

Ref	Year	Original units reported			Values reported after conversion		
		Respiratory rate/ Ventilatory frequency	Tidal volume	Minute volume/ Ventilation	Respiratory rate/ Ventilatory frequency (breaths min.$^{-1}$)	Tidal volume (mL breath^{-1} g^{-1})	Minute volume/ Ventilation (mL min.$^{-1}$ g^{-1})
1	2022	breaths min.$^{-1}$	mL breath^{-1} g^{-1}	mL min.$^{-1}$ g^{-1}	55-275	0.002-0.014	0.01-3
2	1998	Not explicitly reported.	μL g^{-1}	μL sec.$^{-1}$ g^{-1}	73-162	0.0065-0.0085	0.9-1.02
3	2019	breaths min.$^{-1}$	mL breath^{-1} g^{-1}	mL min.$^{-1}$ g^{-1}	125-260	0.004-0.013	0.01-3
4	2002	breaths sec.$^{-1}$	μL g^{-1}	mL min.$^{-1}$ g^{-1}	150-180	0.0057-0.0061	0.75-1.1
5	2021	breaths min.$^{-1}$	mL kg^{-1}	mL min.$^{-1}$ kg^{-1}	150-600	0.008-0.013	1.75-5
6	2012	breaths min.$^{-1}$	mL	mL min.$^{-1}$	170	0.011*	2.18*
7	2017	breaths min.$^{-1}$	mL	mL min.$^{-1}$	175-190	0.0036-0.0054	0.54-0.71
8	2010	cycles min.$^{-1}$	μL g^{-1}	mL min.$^{-1}$ g^{-1}	180	0.0064-0.007	1.3
9	2016	breaths min.$^{-1}$	*NR*	*NR*	185-210	*NR*	*NR*
10	2015	breaths min.$^{-1}$	*PO*	*NR*	187-211.9	*PO*	*NR*
11	2013	breaths min.$^{-1}$	μL g^{-1}	mL g^{-1}	231-253	0.0049-0.0061	1.2-1.38
12	2020	breaths min.$^{-1}$	*NR*	*NR*	250	*NR*	*NR*
13	2017	breaths min.$^{-1}$	mL breath^{-1} g^{-1}	mL min.$^{-1}$ g^{-1}	275-300	0.12-0.14	3.5-4.5
14	2018	breaths min.$^{-1}$	mL breath^{-1}	mL min.$^{-1}$	400	0.012-0.015	5
15	2018	breaths sec.$^{-1}$	*NR*	*NR*	420	*NR*	*NR*
16	2011	*NR*	*NR*	mL min.$^{-1}$ g^{-1}	*NR*	*NR*	3
17	2013	*NR*	*NR*	μL	*NR*	*NR*	3.11
18	2006	*PO*	*PO*	*PO*	*PO*	*PO*	*PO*
19	2010	*PO*	*PO*	mL min.$^{-1}$	*PO*	*PO*	1-1.6

NR = not reported and PO = percentage only. PO also includes percent change, arbitrary units, and delta values.

*Where units reported do not match the standard of Breathe Easy, conversions were made using reported body weight values or standard conversions to match the units to those of Breathe Easy to allow for easier across-publication comparisons. If body weight was not reported, an * marks those values and average body weight was assumed to be 27.5 g.

Supplemental Table 2. Feature comparison of commercially available products to Breathe Easy. This table was generated using publicly available information from each of the companies' websites.

Manufacturer: *System/Software:*	**Breathe Easy**	DSI *Finepointe*	DSI *Ponemah*	ADI *LabChart*	EMKA Scireq *iox2*
Features					
Free/ Open source?	Y	N	N	N	N
Data collection	Designed around ADI Powerlab data collection but can use outputs from other systems through conversion to .txt file input standard	Y (limited to Finepointe)	Y (limited to Ponemah)	Y (also has signal import if in compatible format)	Y (must purchase conversion software to use with recordings performed on other systems)
Breath detection	Y	Y	Y	Y	Y
Calculation of basic breath parameters (volume, frequency, flow, etc.)	Y	Y	-	Y (frequency and volume)	Y
Detection of breathing features (apnea, sigh, cough)	Y (apnea, sigh)	Y	-	-	-
Calculation and normalization with metabolic parameters (VO2, VCO2)	Y	Y	-	Y	-
Manual sub-selection of data for analysis	Y	Y	-	Y	Y
Automated selection of data for analysis (basic per breath)	Y	Y	-	-	Y
Automated selection of data for analysis (filtered quality breathing from desired conditions only)	Y	-	-	-	-
Summary figures	Y	Y	-	-	Y (requires separate software purchase)
Summary statistics	Y	Y	-	-	Y (requires separate software purchase)
Statistical comparisons between groups	Linear Mixed Effects Model with Tukey Post-hoc; ANOVA for some metadata comparisons when appropriate	ANOVA	-	-	-

Supplemental Table 3. Average values for basic respiratory variables in validation study comparing BASSPRO automated breath selections to manual selections from 3 experts. Values present as mean ± standard error mean. Symbols following averages indicate significant differences between indicated groups. Reference group for identified significance are outlined below each table.

A Breath Cycle Duration (log10)

Condition	Sex	Group	BASSPRO	Excluded	Expert 1	Expert 2	Expert 3
Room Air	Female	Control	-0.592 +/- 0.0136 &^#@	-0.61 +/- 0.0135 *^#@	-0.539 +/- 0.0142 *&@	-0.53 +/- 0.0141 *&	-0.518 +/- 0.014 *&^
Room Air	Female	Experimental	-0.631 +/- 0.0234 &^#@	-0.663 +/- 0.0235 *^#@	-0.586 +/- 0.0244 *&	-0.591 +/- 0.0239 *&	-0.593 +/- 0.024 *&
Room Air	Male	Control	-0.663 +/- 0.0135 &^#@	-0.714 +/- 0.0135 *^#@	-0.627 +/- 0.014 *&#	-0.591 +/- 0.0138 *&^@	-0.617 +/- 0.0139 *&#
Room Air	Male	Experimental	-0.676 +/- 0.0234 &^#	-0.705 +/- 0.0235 *^#@	-0.645 +/- 0.024 *&	-0.641 +/- 0.0239 *&	-0.664 +/- 0.024 &
5% CO2	Female	Control	-0.724 +/- 0.0136 &^#@	-0.748 +/- 0.0136 *^#@	-0.681 +/- 0.0139 *&	-0.68 +/- 0.0137 *&	-0.69 +/- 0.0139 *&
5% CO2	Female	Experimental	-0.707 +/- 0.0235 &#@	-0.755 +/- 0.024 *^#@	-0.702 +/- 0.0239 &	-0.683 +/- 0.0236 *&	-0.677 +/- 0.0241 *&
5% CO2	Male	Control	-0.699 +/- 0.0136 &^#@	-0.798 +/- 0.0136 *^#@	-0.685 +/- 0.0139 *&	-0.679 +/- 0.0137 *&	-0.683 +/- 0.0139 *&
5% CO2	Male	Experimental	-0.717 +/- 0.0235 &	-0.786 +/- 0.0239 *^#@	-0.726 +/- 0.0239 &	-0.711 +/- 0.0235 &	-0.707 +/- 0.0239 &

* p < 0.05 vs. BASSPRO

& p < 0.05 vs. Excluded

^ p < 0.05 vs. Expert 1

# p < 0.05 vs. Expert 2

@ p < 0.05 vs. Expert 3

B Ventilatory Frequency (sqrt)

Condition	Sex	Group	BASSPRO	Excluded	Expert 1	Expert 2	Expert 3
Room Air	Female	Control	15.6 +/- 0.235 &^#@	16.3 +/- 0.235 *^#@	14.6 +/- 0.247 *&@	14.4 +/- 0.246 *&	14.2 +/- 0.243 *&^
Room Air	Female	Experimental	16.1 +/- 0.407 &^#@	17.1 +/- 0.408 *^#@	15.2 +/- 0.425 *&	15.3 +/- 0.416 *&	15.4 +/- 0.418 *&
Room Air	Male	Control	16.8 +/- 0.235 &^#@	18.1 +/- 0.235 *^#@	16.1 +/- 0.243 *&#	15.4 +/- 0.24 *&^@	15.9 +/- 0.242 *&#
Room Air	Male	Experimental	17 +/- 0.407 &^#	17.9 +/- 0.407 *^#@	16.3 +/- 0.417 *&	16.2 +/- 0.415 *&@	16.7 +/- 0.418 &#
5% CO2	Female	Control	17.9 +/- 0.236 &^#@	18.7 +/- 0.237 *^#@	17 +/- 0.241 *&	17 +/- 0.238 *&	17.2 +/- 0.242 *&
5% CO2	Female	Experimental	17.5 +/- 0.408 &#@	18.7 +/- 0.418 *^#@	17.4 +/- 0.416 &#@	17 +/- 0.41 *&^	16.9 +/- 0.419 *&^
5% CO2	Male	Control	17.4 +/- 0.236 &^#@	19.6 +/- 0.237 *^#@	17.1 +/- 0.242 *&	17 +/- 0.237 *&	17 +/- 0.242 *&
5% CO2	Male	Experimental	17.7 +/- 0.408 &	19.3 +/- 0.416 *^#@	17.9 +/- 0.416 &	17.6 +/- 0.409 &	17.5 +/- 0.416 &

* $p < 0.05$ vs. BASSPRO

& $p < 0.05$ vs. Excluded

^ $p < 0.05$ vs. Expert 1

# $p < 0.05$ vs. Expert 2

@ $p < 0.05$ vs. Expert 3

C Uncorrected Tidal Volume (log10)

Condition	Sex	Group	BASSPRO	Excluded	Expert 1	Expert 2	Expert 3
Room Air	Female	Control	-3.01 +/- 0.0201 &^@	-3.06 +/- 0.0201 *#	-3.05 +/- 0.0205 *#	-3.02 +/- 0.0205 &^@	-3.05 +/- 0.0204 *#
Room Air	Female	Experimental	-2.97 +/- 0.0347 ^#@	-2.98 +/- 0.0348 ^#@	-3.01 +/- 0.0355 *&	-3.02 +/- 0.0351 *&	-3.02 +/- 0.0352 *&
Room Air	Male	Control	-2.96 +/- 0.0201 &^#@	-3 +/- 0.02 *^#@	-3.08 +/- 0.0204 *&#@	-3.05 +/- 0.0203 *&^	-3.04 +/- 0.0203 *&^
Room Air	Male	Experimental	-2.93 +/- 0.0347 &^#@	-2.94 +/- 0.0347 *^#@	-3 +/- 0.0351 *&@	-3 +/- 0.0351 *&@	-2.97 +/- 0.0352 *&^#
5% CO2	Female	Control	-2.89 +/- 0.0201 &	-2.91 +/- 0.0201 *#	-2.9 +/- 0.0203	-2.89 +/- 0.0202 &	-2.9 +/- 0.0203
5% CO2	Female	Experimental	-2.88 +/- 0.0348	-2.87 +/- 0.0352	-2.87 +/- 0.0351	-2.88 +/- 0.0349	-2.87 +/- 0.0352
5% CO2	Male	Control	-2.83 +/- 0.0201 &	-2.81 +/- 0.0201 *^#	-2.83 +/- 0.0203 &	-2.83 +/- 0.0202 &@	-2.81 +/- 0.0203 #
5% CO2	Male	Experimental	-2.82 +/- 0.0348 &	-2.79 +/- 0.0351 *^#	-2.83 +/- 0.0351 &	-2.82 +/- 0.0348 &	-2.8 +/- 0.0351

* $p < 0.05$ vs. BASSPRO

& $p < 0.05$ vs. Excluded

^ $p < 0.05$ vs. Expert 1

# $p < 0.05$ vs. Expert 2

@ $p < 0.05$ vs. Expert 3

D Peak Inspiratory Flow (log10)

Condition	Sex	Group	BASSPRO	Expert 1	Expert 2	Expert 3
Room Air	Female	Control	0.594 +/- 0.0267 ^#@	0.476 +/- 0.027 *@	0.478 +/- 0.027 *@	0.456 +/- 0.0269 *^#
Room Air	Female	Experimental	0.582 +/- 0.0463 ^#@	0.45 +/- 0.0467 *	0.453 +/- 0.0465 *	0.473 +/- 0.0465 *
Room Air	Male	Control	0.664 +/- 0.0267 ^#@	0.48 +/- 0.0269 *@	0.474 +/- 0.0269 *@	0.497 +/- 0.0269 *^#
Room Air	Male	Experimental	0.805 +/- 0.0463 ^#@	0.633 +/- 0.0465 *@	0.628 +/- 0.0465 *@	0.678 +/- 0.0466 *^#
5% CO2	Female	Control	0.696 +/- 0.0267 ^#@	0.628 +/- 0.0269 *@	0.632 +/- 0.0268 *@	0.659 +/- 0.0269 *^#
5% CO2	Female	Experimental	0.671 +/- 0.0463 #@	0.666 +/- 0.0465 #@	0.622 +/- 0.0463 *^	0.627 +/- 0.0466 *^
5% CO2	Male	Control	0.771 +/- 0.0267 ^#@	0.752 +/- 0.0269 *#	0.734 +/- 0.0268 *^@	0.75 +/- 0.0269 *#
5% CO2	Male	Experimental	0.844 +/- 0.0463	0.841 +/- 0.0465	0.843 +/- 0.0463	0.85 +/- 0.0465

* p < 0.05 vs. BASSPRO

^ p < 0.05 vs. Expert 1

# p < 0.05 vs. Expert 2

@ p < 0.05 vs. Expert 3

E Peak Expiratory Flow (sqrt)

Condition	Sex	Group	BASSPRO	Expert 1	Expert 2	Expert 3
Room Air	Female	Control	2.23 +/- 0.0952 ^#@	2.04 +/- 0.0958 *	2.06 +/- 0.0957 *	2.03 +/- 0.0956 *
Room Air	Female	Experimental	2.45 +/- 0.165 ^#@	2.25 +/- 0.166 *	2.26 +/- 0.165 *	2.25 +/- 0.165 *
Room Air	Male	Control	2.77 +/- 0.0952 ^#@	2.34 +/- 0.0956 *@	2.35 +/- 0.0954 *@	2.42 +/- 0.0955 *^#
Room Air	Male	Experimental	3.14 +/- 0.165 ^#@	2.65 +/- 0.165 *	2.63 +/- 0.165 *	2.69 +/- 0.165 *
5% CO2	Female	Control	2.71 +/- 0.0952 ^#@	2.64 +/- 0.0955 *	2.63 +/- 0.0953 *	2.62 +/- 0.0955 *
5% CO2	Female	Experimental	2.8 +/- 0.165 #@	2.78 +/- 0.165 #	2.68 +/- 0.165 *^	2.71 +/- 0.165 *
5% CO2	Male	Control	3.35 +/- 0.0952 #	3.32 +/- 0.0955 #	3.27 +/- 0.0953 *^@	3.32 +/- 0.0955 #
5% CO2	Male	Experimental	3.53 +/- 0.165	3.56 +/- 0.165 @	3.53 +/- 0.165	3.49 +/- 0.165 ^

* $p < 0.05$ vs. BASSPRO
^ $p < 0.05$ vs. Expert 1
# $p < 0.05$ vs. Expert 2
@ $p < 0.05$ vs. Expert 3

F Ventilatory Equivalents of Oxygen (log10)

Condition	Sex	Group	BASSPRO	Expert 1	Expert 2	Expert 3
Room Air	Female	Control	1.4 +/- 0.0227 ^#@	1.34 +/- 0.023 *#	1.36 +/- 0.0229 *^	1.35 +/- 0.0229 *
Room Air	Female	Experimental	1.38 +/- 0.0393 ^#@	1.32 +/- 0.0397 *	1.33 +/- 0.0395 *	1.33 +/- 0.0395 *
Room Air	Male	Control	1.49 +/- 0.0227 ^#@	1.38 +/- 0.0229 *	1.39 +/- 0.0228 *	1.38 +/- 0.0229 *
Room Air	Male	Experimental	1.58 +/- 0.0393 ^#@	1.49 +/- 0.0395 *	1.49 +/- 0.0395 *	1.5 +/- 0.0395 *
5% CO2	Female	Control	1.69 +/- 0.0227 ^#@	1.68 +/- 0.0228 *	1.67 +/- 0.0228 *	1.67 +/- 0.0228 *
5% CO2	Female	Experimental	1.62 +/- 0.0393 #@	1.62 +/- 0.0395 #@	1.59 +/- 0.0394 *^	1.59 +/- 0.0396 *^
5% CO2	Male	Control	1.7 +/- 0.0227 ^#@	1.69 +/- 0.0228 *	1.68 +/- 0.0227 *	1.68 +/- 0.0228 *
5% CO2	Male	Experimental	1.71 +/- 0.0393 @	1.69 +/- 0.0395 #@	1.71 +/- 0.0393 ^	1.73 +/- 0.0395 *^

* p < 0.05 vs. BASSPRO
^ p < 0.05 vs. Expert 1
# p < 0.05 vs. Expert 2
@ p < 0.05 vs. Expert 3

G Corrected Tidal Volume (sqrt)

Condition	Sex	Group	BASSPRO	Expert 1	Expert 2	Expert 3
Room Air	Female	Control	0.0741 +/- 0.00243 ^#@	0.07 +/- 0.00245 *#	0.0725 +/- 0.00245 *^@	0.0701 +/- 0.00244 *#
Room Air	Female	Experimental	0.0677 +/- 0.00422 ^#@	0.0638 +/- 0.00424 *	0.0637 +/- 0.00423 *	0.0633 +/- 0.00423 *
Room Air	Male	Control	0.0764 +/- 0.00243 ^#@	0.0656 +/- 0.00244 *#@	0.0676 +/- 0.00244 *^	0.0678 +/- 0.00244 *^
Room Air	Male	Experimental	0.0865 +/- 0.00422 ^#@	0.076 +/- 0.00423 *@	0.0754 +/- 0.00423 *@	0.0789 +/- 0.00423 *^#
5% CO2	Female	Control	0.0826 +/- 0.00244 ^#	0.0818 +/- 0.00244 *	0.082 +/- 0.00244 *	0.0823 +/- 0.00244
5% CO2	Female	Experimental	0.0758 +/- 0.00422 #	0.0756 +/- 0.00423 #	0.0743 +/- 0.00422 *^	0.0753 +/- 0.00423
5% CO2	Male	Control	0.0862 +/- 0.00244 #@	0.0862 +/- 0.00244 #@	0.0853 +/- 0.00244 *^@	0.0874 +/- 0.00244 *^#
5% CO2	Male	Experimental	0.0896 +/- 0.00422 #@	0.0896 +/- 0.00423 @	0.0903 +/- 0.00422 *@	0.0922 +/- 0.00423 *^#

* $p < 0.05$ vs. BASSPRO
^ $p < 0.05$ vs. Expert 1
# $p < 0.05$ vs. Expert 2
@ $p < 0.05$ vs. Expert 3

H Ventilation per gram Body Weight (log10)

Condition	Sex	Group	BASSPRO	Expert 1	Expert 2	Expert 3
Room Air	Female	Control	0.101 +/- 0.0301 ^#@	0.00141 +/- 0.0304 *@	0.0214 +/- 0.0304 *@	-0.0206 +/- 0.0303 *^#
Room Air	Female	Experimental	0.0667 +/- 0.0522 ^#@	-0.0273 +/- 0.0526 *	-0.0231 +/- 0.0524 *	-0.0296 +/- 0.0524 *
Room Air	Male	Control	0.193 +/- 0.0301 ^#@	0.0272 +/- 0.0303 *@	0.0229 +/- 0.0303 *@	0.045 +/- 0.0303 *^#
Room Air	Male	Experimental	0.314 +/- 0.0522 ^#@	0.183 +/- 0.0524 *@	0.175 +/- 0.0524 *@	0.224 +/- 0.0525 *^#
5% CO2	Female	Control	0.329 +/- 0.0302 ^#@	0.282 +/- 0.0303 *	0.281 +/- 0.0302 *	0.293 +/- 0.0303 *
5% CO2	Female	Experimental	0.242 +/- 0.0522 #@	0.235 +/- 0.0524 #@	0.2 +/- 0.0523 *^	0.204 +/- 0.0525 *^
5% CO2	Male	Control	0.349 +/- 0.0302 ^#	0.329 +/- 0.0303 *	0.318 +/- 0.0302 *@	0.339 +/- 0.0303 #
5% CO2	Male	Experimental	0.397 +/- 0.0522	0.404 +/- 0.0524	0.395 +/- 0.0522	0.407 +/- 0.0524

* $p < 0.05$ vs. BASSPRO

^ $p < 0.05$ vs. Expert 1

# $p < 0.05$ vs. Expert 2

@ $p < 0.05$ vs. Expert 3

I Oxygen Consumption (log10)

Condition	Sex	Group	BASSPRO	Expert 1	Expert 2	Expert 3
Room Air	Female	Control	-1.3 +/- 0.0277 ^#@	-1.34 +/- 0.0278 *@	-1.34 +/- 0.0278 *@	-1.37 +/- 0.0278 *^#
Room Air	Female	Experimental	-1.32 +/- 0.048 ^#@	-1.35 +/- 0.0481 *	-1.35 +/- 0.0481 *	-1.36 +/- 0.0481 *
Room Air	Male	Control	-1.3 +/- 0.0277 ^#@	-1.35 +/- 0.0278 *#@	-1.37 +/- 0.0277 *^@	-1.34 +/- 0.0278 *^#
Room Air	Male	Experimental	-1.26 +/- 0.048 ^#	-1.31 +/- 0.0481 *@	-1.31 +/- 0.0481 *@	-1.27 +/- 0.0481 ^#
5% CO2	Female	Control	-1.37 +/- 0.0277 ^#@	-1.4 +/- 0.0278 *#@	-1.39 +/- 0.0277 *^@	-1.37 +/- 0.0278 *^#
5% CO2	Female	Experimental	-1.38 +/- 0.048 #	-1.39 +/- 0.0481	-1.39 +/- 0.048 *	-1.39 +/- 0.0481
5% CO2	Male	Control	-1.35 +/- 0.0277 #	-1.36 +/- 0.0278 @	-1.36 +/- 0.0277 *@	-1.35 +/- 0.0278 ^#
5% CO2	Male	Experimental	-1.31 +/- 0.048 ^#@	-1.29 +/- 0.0481 *#@	-1.32 +/- 0.048 *^	-1.32 +/- 0.0481 *^

* $p < 0.05$ vs. BASSPRO
^ $p < 0.05$ vs. Expert 1
# $p < 0.05$ vs. Expert 2
@ $p < 0.05$ vs. Expert 3

Supplemental Table 4. Average values and p-values for select ventilatory frequency and ventilation comparisons showing control variation in the Alzheimer's dataset. *Presented as average ± SEM (p=p-value* Comparison Group).

Respiratory Outcome (transform)	Timepoint	Condition	Experimental*	CaMK*	tetO*	WT*
Ventilatory Frequency (log10)	10 months	Room Air	2.43±0.13 (0.00022*WT)	2.43±0.14 (0.00074*WT)	2.40±0.14	2.36±0.17
Ventilation (log10)	5 months	Room Air	0.174±0.25	0.258±0.27 (p=0.00024*tetO)	0.129±0.29	0.198±0.27
Ventilation (log10)	10 months	Room Air	0.260±0.26 (p=0.0010*tetO, p=0*WT)	0.289±0.27 (p=9.9e-5*tetO, p=0*WT)	0.140±0.26 (p=1.9e-11*WT)	-0.113±0.42
Ventilation (log10)	12 months	Room Air	0.291±0.26	0.202±0.27 (p=0.026*tetO)	0.319±0.21	0.235±0.22
Ventilation (log10)	19 months	Room Air	0.120±0.26 (p=0.00035*caMK)	-0.0190±0.28	0.0673±0.25	0.0283±0.27
Ventilation (log10)	7 months	5% CO_2	0.185±0.18 (p=3.6e-10*CaMK)	0.398±0.16	0.284±0.16 (p=0.010*CaMK)	0.275±0.14
Ventilation (log10)	10 months	5% CO_2	0.333±0.14 (p=2.0e-15*WT)	0.351±0.15 (p=0*WT)	0.299±0.13 (p=1.8e-9*WT)	0.0850±0.34
Ventilatory Frequency (log10)	5 months	5% CO_2	2.44±0.10	2.46±0.11	2.50±0.098 (p=0.036*Exp)	2.45±0.11
Ventilatory Frequency (log10)	12 months	5% CO_2	2.51±0.069	2.46±0.10 (p=0.010*Exp)	2.50±0.076	2.46±0.093 (p=0.0019*Exp)

*Where Experimental = Tg102_tetO-APPswe/ind; TgCaMKIIα-tTA_LineB, CaMK = TgCaMKIIα-tTA_LineB, tetO = Tg102_tetO-APPswe/ind, and WT = no genetic manipulation.

Supplemental Table 5. Number of mice included in analysis of hypercapnic (A) and hypoxic (B) ventilatory responses by age, gas condition, genotype, and sex. Average age (C) and weight (D) of all mice recorded by genotype, age group, hypercapnia or hypoxia exposure, and sex.

A Number of mice included for analysis in hypercapnia experiments by age, gas condition, genotype, and sex.

Sex	Genotype	Condition	5 mo.	7 mo.	10 mo.	12 mo.	14 mo.	16 mo.	19 mo.
						Age			
Female	CaM	5% CO_2	6	6	6	6	6	5	6
		Room Air	6	6	6	6	6	6	6
	Exp	5% CO_2	6	6	6	5	6	5	6
		Room Air	6	6	6	5	6	6	6
	tetO	5% CO_2	4	4	4	2	4	4	3
		Room Air	4	4	4	4	4	4	4
	WT	5% CO_2	6	6	6	5	6	6	5
		Room Air	6	6	6	5	6	6	6
Male	CaM	5% CO_2	6	6	6	5	6	5	5
		Room Air	6	6	6	5	6	5	5
	Exp	5% CO_2	5	5	5	3	5	5	5
		Room Air	5	5	5	4	5	5	5
	tetO	5% CO_2	6	6	6	6	6	5	5
		Room Air	6	6	6	6	6	5	5
	WT	5% CO_2	5	5	5	5	4	5	5
		Room Air	5	5	5	5	4	5	5

B Number of mice included for analysis in hypoxia experiments by age group, gas condition, genotype, and sex.

Sex	Genotype	Condition	5 mo.	7 mo.	10 mo.	12 mo.	14 mo.	16 mo.	19 mo.
						Age			
Female	CaM	10% O_2	4	3	4	4	4	4	3
		Room Air	4	3	4	4	4	4	3
	Exp	10% O_2	6	6	6	6	6	6	6
		Room Air	6	6	6	6	6	6	6
	tetO	10% O_2	6	6	6	6	6	6	6
		Room Air	6	6	6	6	6	6	6
	WT	10% O_2	6	6	6	6	3	6	5
		Room Air	6	6	6	6	3	6	5
Male	CaM	10% O_2	5	6	6	6	6	5	6
		Room Air	6	6	6	6	6	5	6
	Exp	10% O_2	5	5	5	5	5	4	4
		Room Air	5	5	5	5	5	5	5
	tetO	10% O_2	5	6	6	6	6	5	5
		Room Air	6	6	6	6	6	5	5
	WT	10% O_2	5	5	5	5	5	5	5
		Room Air	5	5	5	5	5	5	5

C Average age, in postnatal days, of mice by genotype, age group, hypercapnia or hypoxia exposure, and sex.

Sex	Genotype	Condition	5 mo.	7 mo.	10 mo.	12 mo.	14 mo.	16 mo.	19 mo.
						Age			
Female	CaM	Hypercapnia	172.2	235.0	320.8	403.3	463.7	521.3	604.5
		Hypoxia	176.2	241.0	322.3	412.0	468.0	525.0	607.0
	Exp	Hypercapnia	171.8	235.7	320.7	406.0	464.0	520.7	603.0
		Hypoxia	175.5	240.5	322.3	410.7	467.7	524.4	605.3
	tetO	Hypercapnia	175.3	228.0	307.8	379.8	440.8	501.5	585.0
		Hypoxia	180.3	230.5	312.0	388.5	444.5	504.5	588.0
	WT	Hypercapnia	173.2	229.0	308.8	380.2	440.8	502.3	585.7
		Hypoxia	178.2	231.7	313.2	387.7	446.7	505.3	588.6
Male	CaM	Hypercapnia	173.0	230.2	317.0	393.2	452.7	513.6	599.2
		Hypoxia	175.8	236.0	319.7	398.7	462.7	517.4	602.2
	Exp	Hypercapnia	169.6	233.0	317.0	390.3	450.4	512.2	597.8
		Hypoxia	173.8	237.4	320.0	396.8	457.8	515.7	600.7
	tetO	Hypercapnia	177.3	228.0	306.7	388.2	448.2	501.6	586.2
		Hypoxia	182.4	231.2	308.3	396.5	455.8	506.4	587.8
	WT	Hypercapnia	172.4	229.8	314.0	387.8	444.3	508.4	592.4
		Hypoxia	176.4	234.0	316.2	394.2	457.2	511.6	594.6

D Average weight, in postnatal days, of mice by genotype, age group, hypercapnia or hypoxia exposure, and sex.

Sex	Genotype	Condition	5 mo.	7 mo.	10 mo.	12 mo.	14 mo.	16 mo.	19 mo.
						Age			
Female	CaM	Hypercapnia	31.10	36.00	39.05	40.32	43.43	43.88	45.81
		Hypoxia	31.56	36.28	39.46	41.92	43.67	43.74	45.54
	Exp	Hypercapnia	33.35	37.68	41.76	41.98	46.12	45.40	47.97
		Hypoxia	32.99	37.80	42.00	42.34	44.55	45.85	47.91
	tetO	Hypercapnia	36.38	35.82	38.50	41.08	42.11	45.03	46.71
		Hypoxia	35.84	36.79	39.47	41.73	42.55	44.98	45.88
	WT	Hypercapnia	30.42	31.11	33.68	33.59	38.18	40.39	39.92
		Hypoxia	29.79	31.92	34.74	35.53	36.59	39.47	39.58
Male	CaM	Hypercapnia	42.91	44.40	47.46	47.86	47.87	49.98	51.05
		Hypoxia	41.34	43.00	46.74	48.15	48.63	49.52	50.25
	Exp	Hypercapnia	40.30	44.24	44.16	47.07	47.70	48.76	48.43
		Hypoxia	40.18	43.39	44.36	46.80	48.53	47.56	46.13
	tetO	Hypercapnia	44.37	45.59	49.66	49.37	52.73	52.21	54.69
		Hypoxia	43.15	45.10	48.94	50.02	52.67	51.91	54.07
	WT	Hypercapnia	43.48	46.33	48.33	48.41	51.30	49.30	52.42
		Hypoxia	41.94	45.10	48.01	48.99	50.11	49.71	52.57

Supplemental Table 6. All libraries and packages used by A) BASSPRO, B) STAGG, and C) the GUI and subGUIs. Citations are included for each library/ package.

A All packages used by BASSPRO.

Library	Application	Citation
argparse	This library collects arguments that are provided in the command line when executing the script so that they can be easily accessed as variables	[20]
csv	This library assists with reading in data from .csv files such as the metadata and auto_criteria settings files	[20]
datetime	This library provides a data structure for handling dates and time and applying mathematic operations to them – this is used as part of the logging system for tracking events through the execution of the script	[20]
logging	This library aids in generating contextualized logging feedback to the console as well as a log file.	[20]
math	This library provides additional functions for numerical operations and is currently used to aid in verifying if N/A values are present.	[20]
numpy	This library includes several classes and functions that are used including arrays for storing values as well as mathematic and filter generating functions	[21]
os	This library provides access to functions for file i/o and filepath creation and parsing	[20]
pandas	This library is used for its DataFrame class that is used to store several variables including the input signals and the detected breaths and their derived parameters. Several methods that act upon the DataFrame are utilized for filtering and applying calculations to the data.	[22]
re	This library is used for regular expression parsing of text strings and is used for parsing MUID and PlyUID information	[20]
scipy	This library provides functions used for signal processing including high and low pass filters that are used to reduce noise in the derived flow signal. Additionally, this library's *linregress* function is used for the creation of calibration curves (O_2 and CO_2 concentrations vs. voltages).	[23]
statistics	This library is used to provide access to statistical functions such as median	[20]
sys	This library is used to connect the systems stdout to logging functions, used for displaying logging messages in the console	[20]
tkinter	This library is used to provide a minimal GUI when running the BASSPRO script separately from the GUI	[20]

B All packages used by STAGG.

Package	Application	Citation
argparser	Allows for operation of the R module, STAGG, from the command line terminal	[24]
data.table	Reformatting of data tables in preparation for manipulations	[25]
ggpubr	Graphing and data visualization	[26]
ggthemes	Additional settings and configurations for graphing	[27]
kableExtra	Builds common complex tables and manipulates table styles	[28]
lme4	Fits linear and generalized linear mixed-effects models	[29]
lmerTest	Provides p-values for linear mixed-effects models	[30]
magrittr	Improves readability and intuitiveness of code	[31]
multcomp	Fits linear mixed-effects models and multiple comparisons	[32]
openxlsx	Reads, writes, and edits xlsx files	[33]
pandoc	Used for R markdown rendering	[34]
RColorBrewer	Generates colors for graphing	[35]
reshape2	Used to transform data table formats for some optional graphs	[36]
rjson	Allows import of .json formatted files	[37]
rmarkdown	Produces html summary file with all .svg output files in one page	[38–40]
svglite	Produces user-friendly editable .svg image files.	[41]
tidyselect	Allows for verbiage consistency between packages	[42]
tidyverse	Data management and formatting; include dplyr and ggplot2 packages	[43]
xtable	Generates tables from functional outputs	[44]

C All packages used by the GUI.

Library	Application	Citation
numpy	This library includes several classes and functions that are used including arrays for storing values as well as mathematic and filter generating functions	[21]
pandas	This library is used for its DataFrame class that is used to store several variables including the input signals and the detected breaths and their derived parameters. Several methods that act upon the DataFrame are utilized for filtering and applying calculations to the data.	[22]
pyodbc	This library provides methods and utilities for accessing ODBC databases in Python.	[45]
PyQt5	This library is used for rendering graphics and GUI components like buttons and input boxes. It also manages processes that run parallel to the main GUI thread.	[46]
PyQt5-Qt5	This library is a subset of a Qt installation that is required by PyQt5.	[46]
PyQt5-sip	This library provides Python bindings for C/C++ in support of PyQt5.	[47]
python-dateutil	This library extends the Python native package *datetime* with extra utilities for working with dates and times	[48]
pytz	This library provides Olson timezone calculations.	[49]
six	This library provides smooth compatibility between libraries written for both Python 2 and Python 3.	[50]

Supplemental Table 7. Minimum and recommended machine requirements for running Breathe Easy.

	Minimum Hardware	Recommended Hardware
Operating System	Windows 8	Windows 10+
System Architecture	64-bit x86	64-bit x86
Diskspace	Minimum 5 GB	Minimum 50 GB
RAM	Minimum 4 GB	Minimum 16 GB

References

1. Lusk, S. *et al.* A CRISPR toolbox for generating intersectional genetic mouse models for functional, molecular, and anatomical circuit mapping. *BMC Biol.* (2022).

2. Dauger, S., Nsegbe, E., Vardon, G., Gaultier, C. & Gallego, J. The effects of restraint on ventilatory responses to hypercapnia and hypoxia in adult mice. *Respir Physiol* **112,** 215–225 (1998).

3. Martinez, V. K. *et al.* Off-Target Effects of Clozapine-N-Oxide on the Chemosensory Reflex Are Masked by High Stress Levels. *Front. Physiol.* **10,** 521 (2019).

4. Tankersley, C. G., Haxhiu, M. A. & Gauda, E. B. Differential CO(2)-induced c-fos gene expression in the nucleus tractus solitarii of inbred mouse strains. *J. Appl. Physiol.* **92,** 1277–1284 (2002).

5. Silva, T. de M. E. *et al.* Machine learning approaches reveal subtle differences in breathing and sleep fragmentation in Phox2b-derived astrocytes ablated mice. *J. Neurophysiol.* (2021). doi:10.1152/jn.00155.2020

6. Bassi, M. *et al.* Central leptin replacement enhances chemorespiratory responses in leptin-deficient mice independent of changes in body weight. *Pflugers Arch.* **464,** 145–153 (2012).

7. Romer, S. H. *et al.* Accessory respiratory muscles enhance ventilation in ALS model mice and are activated by excitatory V2a neurons. *Exp. Neurol.* **287,** 192–204 (2017).

8. Gestreau, C. *et al.* Task2 potassium channels set central respiratory CO2 and O2 sensitivity. *Proc. Natl. Acad. Sci. USA* **107,** 2325–2330 (2010).

9. Huang, T.-W. *et al.* Progressive changes in a distributed neural circuit underlie breathing abnormalities in mice lacking mecp2. *J. Neurosci.* **36,** 5572–5586 (2016).

10. Fechtner, L., El Ali, M., Sattar, A., Moore, M. & Strohl, K. P. Fentanyl effects on breath generation in C57BL/6J and A/J mouse strains. *Respir. Physiol. Neurobiol.* **215,** 20–29 (2015).

11. Toward, M. A., Abdala, A. P., Knopp, S. J., Paton, J. F. R. & Bissonnette, J. M. Increasing brain serotonin corrects CO2 chemosensitivity in methyl-CpG-binding protein 2 (Mecp2)-deficient mice. *Exp. Physiol.* **98,** 842–849 (2013).

12. Ward, C. S. *et al.* Loss of mecp2 function across several neuronal populations impairs breathing response to acute hypoxia. *Front. Neurol.* **11,** 593554 (2020).

13. Sun, J. J., Huang, T.-W., Neul, J. L. & Ray, R. S. Embryonic hindbrain patterning genes delineate distinct cardio-respiratory and metabolic homeostatic populations in the adult. *Sci. Rep.* **7,** 9117 (2017).

14. Hill, R., Dewey, W. L., Kelly, E. & Henderson, G. Oxycodone-induced tolerance to respiratory depression: reversal by ethanol, pregabalin and protein kinase C inhibition. *Br. J. Pharmacol.* **175,** 2492–2503 (2018).

15. Vogelgesang, S., Niebert, M., Bischoff, A. M., Hülsmann, S. & Manzke, T. Persistent expression of serotonin receptor 5b alters breathing behavior in male mecp2 knockout mice. *Front. Mol. Neurosci.* **11,** 28 (2018).

16. Ray, R. S. *et al.* Impaired respiratory and body temperature control upon acute serotonergic neuron inhibition. *Science* **333,** 637–642 (2011).

17. Alhaddad, H. *et al.* Gender and strain contributions to the variability of buprenorphine-related respiratory toxicity in mice. *Toxicology* **305,** 99–108 (2013).

18. Lewanowitsch, T., Miller, J. H. & Irvine, R. J. Reversal of morphine, methadone and heroin induced effects in mice by naloxone methiodide. *Life Sci.* **78,** 682–688 (2006).

19. Goineau, S., Rompion, S., Guillaume, P. & Picard, S. Ventilatory function assessment in safety pharmacology: optimization of rodent studies using normocapnic or hypercapnic conditions. *Toxicol. Appl. Pharmacol.* **247,** 191–197 (2010).

20. Van Rossum, G. & Drake, F. L. *Python 3 Reference Manual.* (CreateSpace, 2009).

21. Harris, C. R. *et al.* Array programming with NumPy. *Nature* **585,** 357–362 (2020).

22. McKinney, W. Data Structures for Statistical Computing in Python. (2010).

23. Virtanen, P. *et al.* SciPy 1.0: fundamental algorithms for scientific computing in Python. *Nat. Methods* **17,** 261–272 (2020).

24. Shih, D. J. H. *argparser: Comman-Line Argument Parser*. (R, 2021).

25. Dowle, M. & Srinivasan, A. *data.table: Extension of "data.frame."* (R, 2021).

26. Kassambara, A. *ggpubr: "ggplot2" Based Publication Ready Plots*. (R, 2020).

27. Arnold, J. B. *ggthemes: Extra Themes, Scales, and Geoms for "ggplot2."* (R, 2021).

28. Zhu, H. *kableExtra: Construct Complex Table with "kable" and Pipe Syntax*. (R, 2021).

29. Bates, D., Maechler, M., Bolker, B. & Walker, S. Fitting Linear Mixed-Effects Models Using lme4. *J Stat Softw* **67,** 1–48 (2015).

30. Kuznetsova, A., Brockhoff, P. B. & Christensen, R. H. B. lmerTest Package: Tests in Linear Mixed Effects Models. *Journal of Statistical Sofware* **82,** 1–26 (2017).

31. Bache, S. M. & Wickham, H. *magrittr: A Forward-Pipe Operator for R*. (R, 2020).

32. Hothorn, T., Bretz, F. & Westfall, P. Simultaneous Inference in General Parametric Models. *Biom J* **50,** 346–363 (2008).

33. Schauberger, P. & Walker, A. *openxlsx: Read, Write and Edit xlsx Files*. (R, 2020).

34. Pandoc - About pandoc. at <https://pandoc.org/>

35. Neuwirth, E. *RColorBrewer: ColorBrewer Palettes*. (R, 2014).

36. Wickham, H. Reshaping Data with the reshape Package. *J. Stat. Softw.* **21,** (2007).

37. Couture-Beil, A. *rjson: JSON for R*. (R, 2018).

38. Allaire, J. J. *et al. rmarkdown: Dynamic Documents for R*. (R, 2021).

39. Xie, Y., Dervieux, C. & Riederer, E. *R Markdown Cookbook*. (Chapman and Hall/ CRC, 2020).

40. Xie, Y., Allaire, J. J. & Grolemund, G. *R Markdown: The Definitive Guide*. (Chapman and Hall/ CRC, 2018).

41. Wickham, H. *et al. svglite: An "SVG" Graphics Device*. (r-lib.org, 2022).

42. Henry, L. & Wickham, H. *tidyselect: Select from a Set of Strings*. (R, 2020).

43. Wickham, H. *et al.* Welcome to the tidyverse. *Journal of Open Source Software* **4,** 1686 (2019).

44. Dahl, D. B., Scott, D., Roosen, C., Magnusson, A. & Swinton, J. *xtable: Export Tables to LaTeX or HTML*. (R, 2019).

45. Kleehammer, M. *pyodbc*. (GitHub, Inc., 2020).

46. Riverbank Computing. *PyQt*. (Riverbank Computing, 2021).

47. Riverbank Computing. *PyQt-sip*. (Riverbank Computing, 1998). at <https://www.riverbankcomputing.com/software/sip>

48. Niemeyer, G., Pieviläinen, T., de Leeuw, Y. & Ganssle, P. *python-dateutil*. (GitHub, 2003). at <https://github.com/dateutil/dateutil>

49. Bishop, S. *pytz*. (PyPI, 2008).

50. Peterson, B. *six*. (PyPI, 2021). at <https://pypi.org/project/six/>