Differentiation of leukocytes in bronchoalveolar lavage fluid samples using higher harmonic generation microscopy and deep learning

L.M.G. van Huizen[1]*[¶], M. Blokker[1¶], Y. Rip[1], M. Veta[2], K.A. Kalverda[3], P.I. Bonta[3], J.W. Duitman[3,4,5], & M.L. Groot[1]

[1] LaserLab Amsterdam, Department of Physics, Faculty of Science, Vrije Universiteit Amsterdam, Amsterdam, The Netherlands

[2] Medical Image Analysis Group (IMAG/e), Department of Biomedical Engineering, University of Technology, Eindhoven, The Netherlands

[3] Department of Pulmonary Medicine, Amsterdam UMC Location University of Amsterdam, Amsterdam, The Netherlands

[4] Department of Experimental immunology, Amsterdam UMC Location University of Amsterdam, Amsterdam, The Netherlands

[5] Amsterdam Infection & Immunity, Inflammatory Diseases, Amsterdam, The Netherlands

* Corresponding author

E-mail: l.m.g.van.huizen@vu.nl (LMG)

[¶] These authors contributed equally to this work.

Abstract

In many diseases such as interstitial lung diseases (ILDs), patient diagnosis relies on diagnostic analysis of bronchoalveolar lavage fluid (BALF) and biopsies. In BALF the differentiation of neutrophils, eosinophils, lymphocytes, and macrophages can contribute to diagnose the underlying ILD entity. To analyze the BALF standard cytological techniques are labor-intensive and time-consuming. Studies have shown promising cell identification performance on blood fractions analyzed by third harmonic generation (THG) and multiphoton excited autofluorescence (MPEF) microscopy. Here, we extend this to BALF samples, and we trained a deep learning algorithm for automated analysis on the image level against reference cytology. We imaged blood fractions from three healthy individuals and one asthma patients, and six BALFs from ILD patients. We determined the leukocyte characteristics in terms of cellular and nuclear morphology, and THG and MPEF signal intensity. A deep learning model was trained on both blood fractions and BALF 2D images was used to estimate the leukocyte ratios by using only the standard cytology differential cell ratios at the image-level as reference. The deep learning network has learned to identify individual cells and was able to provide a reasonable estimate of the leukocyte percentage, coming within a 2 to 10% margin in BALF samples in the hold-out testing set. We suggest that the performance of the combined label free imaging and AI analysis can be improved further by collecting 3D data and data of additional fluid samples of various ILD diseases and healthy samples, THG/MPEF microscopy in combination with deep learning is a promising technique for instant differentiation and quantification of leukocytes. Immediate feedback on leukocyte ratios will not only speed-up the diagnostic process but can also reduce costs, the workload and reduce inter-observer variations.

Keywords

Immune cells, blood, bronchoalveolar lavage, third harmonic generation, multiphoton excited autofluorescence, deep learning regression, grad-cam

Introduction

Leukocytes are part of the immune system and are key players in the defense against infectious agents and injuries, for example in case of inflammation and tissue repair [1]. Most abundant leukocytes in the human body are neutrophils, eosinophils, lymphocytes, and macrophages, but also other subtypes such as plasma cells and basophils are recognized. In many diseases the leukocyte composition is altered, e.g. during infection, auto-immune diseases and other inflammation driven diseases. Information about the cell composition can contribute to the guidance of physicians in the diagnosis and treatment of a patient. Interstitial lung diseases (ILDs) encompass a broad variety of diseases, such as idiopathic pulmonary fibrosis (IPF), chronic hypersensitivity pneumonitis (cHP), non-specific interstitial pneumonia (NSIP), and connective tissue related ILD (CTD-ILD). Although rarely diagnostic, during endoscopic procedures bronchoalveolar lavage fluid (BALF) can be of added value to differentiate subsets of ILDs, specifically, in differentiating IPF from more inflammation driven ILDs such as cHP and NSIP [2,3]. Nowadays, to examine the ratio of immune cells in BALF the immune cells are collected and analyzed by cytological techniques and flow cytometry. This involves manual counting relying on standard cytological and microscopy techniques. Although these techniques have improved over the past decades, cell differentiation is still a labor-intensive and time-consuming process. Instant feedback on BALF immune cell content could be of specific interest as the decision to pursue for more invasive diagnostic techniques, such as transbronchial cryobiopsy, may depend on the BALF results.

A promising imaging technique for label-free rapid on-site feedback is third harmonic generation (THG) and multiphoton excited autofluorescence (MPEF) microscopy, in short label-free microscopy. This technique provides images of fresh specimens revealing histology information without processing the tissue [4–8]. Previous studies using blood samples have shown that THG signals reveal discriminative information for the different leukocytes, providing cellular and nuclear morphological information [9–11]. The THG signal

intensity has been shown to be higher for granulocytes (neutrophils and eosinophils) than

agranulocytes (lymphocytes and monocytes) because granulocytes contain THG generating

granules in their cytoplasm [9,10]. Cell cytoplasms generate MPEF signals which differ

between the different leukocyte types in both fluorescence pattern and intensity [12,13]. For

example, the granules in the cytoplasm of eosinophils generate strong autofluorescent

signals. Previous studies have focused on cell differentiation using label-free microscopy in

blood samples, while to the best of our knowledge, THG and MPEF microscopy has not

been applied to BALF.

Automatic identification and classification of the different leukocyte types in the label-free

images could potentially increase the added value in providing rapid feedback on BALF.

Manual counting of leukocytes in these images is time consuming, and prone to

interobserver variations. When only image-level annotations are available, the task of

leukocyte counting is a perfect candidate for deep learning to undertake. Deep learning

allows development of cell counters based on the image data. Studies have shown that the

label-free images are suitable for automatic image analysis using deep learning networks

[14–16]. These studies have used deep learning networks to classify malignant tissue from

healthy tissue and to classify activated lymphocytes, but not for cell counting. Automated cell

counting based on U-net segmentation was demonstrated on label-free stimulated Raman

scattering images of human brain tumors [17]. Regression based counting generates cell

counts based directly on the entire input image, therefore avoiding the need to annotate cells

beforehand. Deep learning regression was performed on microscopy images of human

osteosarcoma and human leukemia cell lines to count cells [18], where they used absolute

cell counts for only one cell type to train the model. In our study, the task is expanded to four

types of leukocytes, with solely fractional sample counts available as labels. The model

therefore must learn to differentiate and regress these four leukocyte classes in the label-

free images, based on cell density, relying on weak labels, in both blood fraction and BALF

samples.

We present a proof-of-principle study with the goal to characterize and quantify the different

immune cells (neutrophiles, eosinophiles, lymphocytes and macrophages) in label-free

images of BALF samples, using blood fraction samples as additional data and standard

cytological cell differentiation as the reference. We examined whether deep learning

networks can provide accurate immune cell counts of label-free BALF images based on the

current gold standard for cell differentiation.

Materials and methods

I.   Sample acquisition, preparation, and processing

Two types of samples were used to address the aim of this study, i.e. blood samples and

bronchoalveolar lavage fluid (BALF) samples (Fig 1A). Samples were collected at the

Amsterdam UMC (location AMC) and imaged with the label-free microscope directly on-site

or in case the mobile label-free microscope was at Amsterdam UMC location VUmc,

samples were transported on ice to the Imaging Centre at Amsterdam UMC, location VUmc

and imaged within three hours. All patients provided informed consent as part of studies

approved by a Medical Research Board under protocol number NTR NL7634 or FP02C-18-

001. The research was conducted in accordance with the Netherlands Code of Conduct for

Research Integrity and the Declaration of Helsinki.

Fig 1: Schematic representation of sample acquisition (A), label-free imaging using a

THG/MPEF microscope (B) and standard cytology processing (C). Images partly created with

BioRender.com.

Six BALF samples were obtained from ILD patients who underwent bronchoalveolar lavage

during an endoscopic procedure. Samples were collected by installing 8 fractions of 20 mL

saline solution, of which of the last fraction the last few milliliters were collected separately

for label-free imaging. Samples were immediately cooled on ice after the bronchoscopy

procedure, transported to the lab and processed according to the standard clinical

procedures. Venous blood (three healthy volunteers and one asthma patient) was collected

in heparin tubes and peripheral blood mononuclear cells (PBMCs) and granulocytes were

isolated using standard density gradient techniques. From healthy volunteers both PBMC

and granulocyte fractions were included, and from the asthma patient only the granulocyte

fraction. See Table 1 for an overview of included samples. Subsequently, samples (BALF

and blood cell fractions) were processed according to standard cytology protocols of the

Long Immunology (LONI) lab of Amsterdam UMC location AMC, resulting in Diff-Quick (DQ,

RAL Diagnostics, ref 720555-000) cytospin slides (Fig 1C). Differential cell counts

(macrophages/monocytes, lymphocytes, neutrophils, basophils, and eosinophils) were

analyzed by counting at least 500 cells using a standard light microscope at 40x

magnification, by two independent experienced observers. Results were calculated as

percentages of the total cell population (Table 2) and considered as the gold standard

reference values of these samples.

Table 1: Overview of the imaged samples, the acquired number of mosaics and total imaged

areas. Samples include blood fractions (granulocytes and PBMC) of 4 volunteers and BALF samples

of six ILD patients. Of the person with asthma only the granulocyte fraction was included. The

average number of cells per mm² was for granulocyte fractions 4808 cells, for PBMC fractions 6464

cells, and for BALF samples 549 cells.

Sample type	ID	Patient condition	Number of samples	Total mosaics	Total imaged area
Granulocytes	GR 1-3	Healthy	3	4	15.63 mm²
	GR 4	Asthma	1	1	6.25 mm²
PBMC	PBMC 1-3	Healthy	3	3	18.75 mm²
Bronchoalveolar lavage	BAL 1-6	ILD	6	11	68.75 mm²

160
161

162 **Table 2: Reference cytology differential cell percentages and standard deviation per sample.**

163 Highlighted are the samples used as validation set and testing set for deep learning cell

164 differentiation.

Sample type		ID	Neutrophils (%)	Eosinophils (%)	Lymphocytes (%)	Monocytes / macrophages (%)
Blood fraction	Granulocytes	GR1	94.5 (± 0.3)	5.0 (± 0.5)	0.5 (± 0.3)	0.0 (± 0.0)
		GR2	94.0 (± 1.3)	5.0 (± 0.5)	1.0 (± 0.8)	0.0 (± 0.0)
		GR3	88.3 (± 1.9)	10.6 (± 1.1)	1.1 (± 0.8)	0.0 (± 0.0)
		GR4	69.2 (± 0.6)	14.1 (± 0.4)	14.4 (± 0.6)	2.3 (± 0.3)
	PBMC	PBMC1*	0.4	0.0	98.1	1.5
		PBMC2	1.9 (± 0.8)	0.0 (± 0.0)	94.6 (± 0.6)	3.5 (± 1.3)
		PBMC3*	2.9	0.6	95.3	1.2
Bronchoalveolar lavage		BAL1	5.1 (± 0.6)	0.3 (± 0.1)	35.3 (± 2.0)	59.3 (± 1.5)
		BAL2	17.5 (± 2.3)	4.2 (± 0.1)	4.7 (± 0.8)	73.5 (± 2.9)
		BAL3	92.7 (± 2.8)	1.1 (± 0.2)	1.8 (± 0.5)	4.4 (± 2.3)
		BAL4	32.0 (± 0.9)	4.2 (± 0.5)	6.4 (± 0.6)	57.4 (± 1.0)
		BAL5	0.3 (± 0.1)	0.1 (± 0.1)	4.4 (± 0.8)	95.2 (± 0.8)
		BAL6	3.0 (± 0.1)	10.7 (± 0.9)	13.8 (± 2.0)	72.5 (± 1.0)

Training set   Validation set   Testing set

* For this fraction no standard deviation is provided because the fraction was counted by one

observer.

A few milliliters of each unprocessed sample/fraction was transferred to a sample holder dish

(µ-Dish 35 mm, high, Ibidi GmbH). After a waiting time of at least 15 minutes to let the cells

rest and sink to the bottom of the dish, the cells were imaged with the label-free microscope

(Fig 1B).

II. Label-free microscopy

To image the samples a portable higher harmonic generation (HHG) microscope was used (Flash Pathology B.V.). Details of the microscope have been described in our previous study [7]. In short, this microscope focuses a femtosecond-pulsed laser beam with a center wavelength of 1070 nm into the sample, which generates nonlinear multiphoton signals, relevant in this study: third harmonic generation (THG), two-photon excited autofluorescence (2PEF) and three-photon excited autofluorescence (3PEF). These three signals are detected in epi-direction and separated by their wavelength using appropriate dichroic mirrors and filters at 349-361 nm (THG), 550-665 nm (2PEF), and 380-420 nm (3PEF). Because of practical reasons not all signals could be detected at the same time, and the measurement was performed twice subsequently, once with detecting THG and 2PEF signals simultaneously and once with THG and 3PEF signals. The lateral resolution of the microscope is 0.4 µm and axial resolution 2.5 µm.

By scanning the laser beam with galvo mirrors over the sample an image was acquired with a field-of-view of 250 µm, and 2 or 4 pixels per micrometer, depending on the scan program. Subsequent images were stitched into a mosaic image (2D) by moving a translational stage in x- and y-direction, and into a stacked depth scan (3D) by moving the stage in z-direction.

III. Data acquisition and image processing

For each granulocyte fraction (n = 4), PBMC fraction (n = 3) and BALF sample (n = 6) one or two stitched mosaic scans and various stacked depth scans were acquired using the label-free microscope (Table 1). Each image was gamma corrected (factor 0.7) and saved to a 24-bit RGB BMP file, using built-in software (Flash Pathology B.V.).

IV.   Leukocyte characterization

The label-free images of the blood fractions were analyzed qualitatively to determine the

leukocyte characteristics, except for macrophages which are absent in blood. Both the cell

morphology and THG and MPEF (2PEF and 3PEF) signal intensities were compared,

including the largest cellular diameter (measured 100 cells per immune cell type using the

software ImageJ (Fiji version 1.53). The macrophage characteristics were evaluated using

the label-free images of BALF samples. Together with what is known from literature, a

complete set of leukocyte characteristics were formed to identify the different leukocytes in

the label-free images of the BALF samples.

V.   Deep learning cell counting

For deep learning cell counting, the available blood fraction and BALF cases were separated

on patient level into a training, validation, and testing set (see Table 2). Each split contained

at least one granulocyte, one PBMC and one BALF case. From each case, only 2D mosaics

that consist of THG and 2PEF images were included. Given that each mosaic covered a

large area of the respective sample, therefore including a large number of immune cells,

each mosaic was treated as an individual case. Accordingly, every mosaic was labelled with

the reference cytology percentages from their respective sample.

Each mosaic was processed using background subtraction and contrast limited adaptive

histogram equalization (CLAHE). Background subtraction allowed for removal of glass

interface interference in the THG channel, due to THG being an interface-sensitive

technique and cells localizing at the bottom of the imaging dish. CLAHE enhanced the

contrast of the large area images, evenly revealing all the immune cell-information present in

the image data. Pixel size was fixed across all images to 0.5 µm by down-scaling if

necessary. Due to the small sample size, larger mosaics (> 2500 x 2500 pixels) were divided

into four mosaics and treated as individual cases. This was possible thanks to the large

numbers of cells captured with our microscope, since we scanned regions up to multiple

millimeters squared depending on the field-of-view of the scan program. This also alleviated

the GPU memory constraints. In total, this resulted in 40 samples in the training set, 10

samples in the validation set, and 16 samples in the testing set.

The deep learning regression network was set up as a fully convolutional network (FCN) to

allow for dynamic image input size while keeping the pixel size fixed. The underlying network

architecture consisted of ResNet50 [19] pre-trained on ImageNet [20]. In addition,

EfficientNetV2B0 [21] and MobileNetV2 [22] were investigated. The last fully-connected layer

was replaced with global average pooling to obtain a 2D vector, a 4-unit fully-connected

layer and subsequent softmax activation. The network was optimized by minimization of the

mean absolute error, using Adam [23] and one cycle learning rate method [24] for a duration

of 200 epochs with an initial learning rate of 1e-05. The baseline ResNet50 model weights

were frozen for the first epoch and unfrozen for the remaining 199 epochs. Early stopping

was applied after 15 epochs without decrease of the validation loss.

Due to over- and underrepresentation of certain immune cells across the data, the neural

network might be inclined to favor one class of immune cells over the other during

optimization. Since the leukocyte counts are fractions which always sum to 1, the setup was

equivalent to multi-class classification with class imbalance. The imbalance was combated

by application of class weights in the calculation of the loss function during training. In

addition, potential overfitting due to a small training dataset was countered by the application

of data augmentation during training. Upon each epoch, each input image was randomly

rotated, flipped, and brightness, contrast, saturation, and hue adapted.

The top-performing model, based on the performance on the validation set, was evaluated

on the hold-out testing set. Each case in the testing set consisted of one mosaic; each

mosaic was fed into the model and testing regression percentages were saved and

compared to reference cytology leukocyte ratios.

Implementation of the deep learning cell counting was done in TensorFlow 2.10 [25], and

training took place on a GPU workstation (Lambda Quad, Lambda Labs, Inc.).

VI.   Deep learning class activation maps

The regression outputs of the ResNet50 model do not allow for further interpretation of the

model performance. Localizing model outputs to objects (i.e. immune cells) is a major benefit

for model validation. In order to validate deep learning regression outputs, gradient-weighted

class activation mapping (Grad-CAM) was implemented [26] using the tf-explain plugin [27].

Grad-CAM uses gradient flow at the last convolutional layer of the model, extracting the

gradients that contribute to the prediction of the class of interest, overlaying this as a heat-

map on the original input image. In practice, Grad-CAM gives a rough estimate of the

regions-of-interest in the input image that contribute the most to the class prediction. For

each testing set mosaic, a Grad-CAM heatmap was generated for each immune cell class,

resulting in four heatmaps per test mosaic. The deep learning cell counting pipeline is

visualized in Fig 2.

Fig 2: Deep learning cell counting pipeline. Pre-processed THG/MPEF mosaics are forward fed

into the pre-trained ResNet50 model, together with reference cytology leukocyte ratios as labels.

Class weights keep the model from overfitting on the ratio distribution. Mean absolute error is

calculated between the fractions input and softmax output. After optimization, Grad-CAM activation

maps can be extracted from the last convolutional layer for model validation. Scale bar 150 µm.

Results

I. Label-free microscopy images of leukocytes

Four granulocyte fractions, three PBMC fractions and six BALF samples were included and imaged using the label-free microscope. Each mosaic was generated within, on average, three minutes. Despite the difficulty of imaging thin transparent samples in epi-direction, the individual cells were clearly visible in the THG/MPEF images (see Fig 3 for an example of an imaged BALF sample). The THG signal (depicted in green) showed the morphological information with subcellular details and the 2PEF and 3PEF signals (depicted in blue and purple, respectively) revealed the autofluorescence of cell cytoplasms, generated by the compounds flavin adenine dinucleotide (FAD) and nicotinamide adenine dinucleotide (NAD(P)H), respectively.

Fig 3: Label-free imaging of a BALF sample showing cellular structures. The BALF images contain cells of different sizes and shapes, and varying THG and 2PEF signal intensity. Magnified images show cell nuclei (CN), nuclear lobes (NL), and cell cytoplasm with increased 2PEF signal (F), THG signal (T) or both (FT). Non-cellular structures were also present, indicated as dirt (D). Acquisition time of this quarter of the acquired 2D mosaic was 0.6 seconds.

II. Leukocyte characterization

To discriminate between the different leukocyte types, the label-free images of the granulocyte fractions and PBMC fractions were evaluated qualitatively and compared with literature findings. Fig 4A-C shows various examples of the different immune cells visualized by the different imaging modalities (THG, 2PEF and 3PEF).

Fig 4: Label-free images of different leukocytes showing the discriminating characteristics in
cellular and nuclear morphology and signal intensity. A-C: THG, 2PEF and 3PEF images of
neutrophils, eosinophils, and lymphocytes from various blood fractions (A) and one bronchoalveolar
lavage sample (B) show similar cell morphology as in standard cytology Diff-Quick (DQ) stained
images. Neutrophils and eosinophils were imaged from the same granulocyte fraction and the images
were saved with the same contrast settings. Two examples of macrophages are presented from two
different BALF samples (C). D: 8 µm depth scans (with steps of 1 µm) of two neutrophils show that
the visible number of nuclear lobes depends on the imaging plane. A-D: For all images the same
scale was used, for direct comparison of cell sizes.

The clearest difference between the granulocyte and PBMC fraction was the higher THG
signal intensity of the granulocyte fractions in comparison to the PBMC fractions. This is due
to strong THG signals generated by densely packed lipid granules inside the cytoplasm of
granulocyte cells [9], while PBMC cells have no cytoplasmic granules. THG signals also
revealed morphological differences between the different cell types, especially the cell size
and the cell nuclear morphology are important discriminating factors. Lymphocytes were
smaller in cell diameter, on average 7.1 µm (STD = 0.8 µm), compared to neutrophils and
eosinophils with an average cell diameter of 8.9 µm (STD = 0.7 µm) and 8.7 µm (STD = 0.6
µm), respectively. Additionally, while the cell nuclei of lymphocytes and monocytes are
mononuclear, the cell nuclei of neutrophils and eosinophils are multi-lobular, which is clearly
visible in the THG images (Fig 4A). According to literature, which is inconclusive about the
number of lobes, the neutrophils have 2 or 3 to 5 nuclear lobes [1,28], while eosinophils
usually have 2 or sometimes 3 nuclear lobes [1,29]. However, in the 2D label-free images
often only one or two lobes were visible, because of the position of the imaging plane, while
in the 3D imaging stacks all lobes could be revealed (Fig 4D).

To discriminate within the granulocyte fraction between neutrophils and eosinophils, the
MPEF images could be used, because the cell cytoplasm of eosinophils generated stronger

2PEF and 3PEF signals than the cell cytoplasm of neutrophils (Fig 4A). The 3PEF signals

were stronger than the 2PEF signal, suggesting that NADH was mostly involved in

generating the autofluorescence signals.

To determine the discriminative features of macrophages, and because macrophages are

absent in blood, the BALF samples were evaluated. Fig 4C shows two examples. Similar to

monocytes (the precursor cells of macrophages), macrophages are normally mononuclear

and do not have granules in their cytoplasm. However, macrophages engulf and digest

pathogens, and therefore their cytoplasm can contain inhomogeneities, such as pigments,

that generate THG and MPEF signals. In contrast to neutrophils, eosinophils and

lymphocytes that are round cells, macrophages were present in varying shapes. Also, the

average diameter of macrophages of 19.6 µm (STD = 3.0 µm) was larger and varied more

than of the other leukocytes.

Together, the cell size and shape, the nuclear shape, the presence of cytoplasmic granules,

and the THG and MPEF signal intensities, form a complete set of characteristics to

discriminate the different leukocytes (Table 3). Using these characteristics, all four cell types

were identified in the BALF samples (Fig 4B and C).

Table 3: Leukocyte characteristics to distinguish the different cell types, based on THG and MPEF (2PEF and 3PEF) images. Discriminative features include average cellular size (measured of 100 cells per leukocyte type) and shape, nuclear morphology, the presence of granules and the THG and MPEF signal intensity.

Leukocyte type	Cell diameter*	Cell shape	Cell nucleus	Cell cytoplasm	THG signal	MPEF signal
Neutrophils	8.9 µm (± 0.7)	Round	2 – 5 lobes (not round)	Granules present	High	Low
Eosinophils	8.7 µm (± 0.6)	Round	2 – 3 lobes (not round)	Granules present	High	High
Lymphocytes	7.1 µm (± 0.8)**	Round	Round/kidney-shaped mononuclear	No granules, little cytoplasm	Low	Low
Macrophages	19.6 µm (± 3.0)	Varying	Round/kidney-shaped mononuclear**	No granules, can have pigment/vesicles	Varying	Varying

* The average cell diameter of neutrophils, eosinophils and lymphocytes are measured in blood fractions, while of macrophages are measured in BALF samples

** Activated cells in BALF samples from ILD patients can differ in morphology, such as increased cell diameter up to 18 µm for lymphocytes, and multinuclear macrophages.

III.   Deep learning cell counting

To estimate immune cell percentages automatically, a pre-trained ResNet50 was used to learn a regression task. Instead of absolute counts of the four different types of immune cells, the deep learning model had to learn class density dependencies in order to produce immune cell percentages. Several regression models were evaluated as part of the hyperparameter optimization process: ResNet50, EfficientNetV2B0, and MobileNetV2. Each network differs in architecture and in size of weights: 3.5 million weights (MobileNetV2), 7.2 million weights (EfficientNetV2B0) and 25.6 million weights (ResNet50), making it interesting to investigate regression performance between these models based on the validation set. In practice, the models performed similarly on the blood fraction samples in the validation set, while ResNet50 outperformed the other models on the BALF samples, which are of main interest. This might be because of the difficulty of the task at hand, requiring a lot of weights

to properly model the underlying correlations between image pixels and cytospin fractions. In

addition to model architecture, the architecture weight initialization method, and learning rate

were optimized. This demonstrated that pre-training on ImageNet and a learning rate of 1e-

05 respectively resulted in the most optimal validation results.

In Fig 5 the top-performing deep learning model performance is compared to the reference

cytology cell percentages for both the validation and testing set. Despite the model's

difficulties to estimate leukocytes in blood fractions, it comes within a 2 to 10% margin in

BALF samples in both the validation and hold-out testing set. In general, the ResNet50

model tends to underestimate the neutrophil fraction, unless a very small number of

neutrophils is present in the sample. The model makes up the underestimation by

overestimating lymphocytes and monocytes/macrophages. For the BALF cases specifically,

the model underestimates eosinophils.

Fig 5: Deep learning regression outputs on the validation and testing set compared to

reference cytology counts. See Table 2 for included cases in validation and test sets. Note that

there is no standard deviation bar for the reference cytology of the PBMC in the testing set, because

this sample was counted once.

Since a blood fraction/BALF case is made up of multiple mosaics carrying the same

leukocyte fractions, a standard deviation of the model's predictions can be obtained for each

case. These indicate that in the majority of the cases, the model performs regression with

high certainty up to a standard deviation of 3%.

S1 Table lists the detailed fraction output of the top-performing ResNet50 model and

reference cytology scores, for both the validation and testing sets.

IV.   Deep learning activation maps

Using Grad-CAM, a class activation map was generated for each immune cell type for test case BALF 4. Fig 6 displays zoomed-in activation maps for each immune cell class for this BALF case. An activation map indicates the contribution of each mosaic pixel to the class prediction. Between immune cell classes, the difference in pixel activation can be compared.

Fig 6: Grad-CAM activation results for test case BALF 4. See S1 Table for the regression performance of the ResNet50 model in this case. The activation maps are focused on the yellow square inlet from Fig 3 in the original mosaic image to illustrate individual cell activation. The maps are displayed using a ramping lookup table, which expresses the pixel value in four colors ranging from minimum to maximum activation: blue, green, yellow and red. Noticeable is the low activation of background pixels (green) and medium-to-high activation of cellular objects (yellow to red). Between class activation maps, the same part of a cell might be less or more activated. Scale bar 50 μm.

As shown in Fig 6, all cellular structures in the BALF mosaic are taken into account during class regression as evident from the yellow to red activations. For different classes, different maximum activations occur, but they seem to be mostly focused on single cells or clumps of cells. No direct correlation can be drawn between the number of low/high activated cells per class map and the leukocyte fractions. Similarly, a single cell might be activated in each of the four class activation maps, although different parts of the cell might be less or more activated depending on the class.

Another observation from Fig 6 is the clear distinction between background pixels and foreground pixels. However, the background pixels are still of interest to the model, given that they are colored green instead of blue. Cells with low signal, consisting of a cloud of pixels instead of solid objects, are still recognized and deemed of (high) importance to various class activations. Interestingly, high activations of cells are accompanied with a halo

418   of low activations surrounding the cell, indicating a decision gradient with a clear distinction

419   between background and foreground pixels.

420   Discussion

This is the first study to image BALF samples using THG and MPEF (2PEF and 3PEF)

microscopy. Even though the samples were transparent, and the generated signals were

detected only in epi-direction by the microscope, the detected signal intensity was high

enough to reveal the individual immune cells. In accordance with previous studies

[9,10,12,13] the different leukocytes could be easily identified in the blood fractions by

cellular and nuclear morphology and distinctive THG and MPEF signal intensities. However,

identification of the cells in BALF samples by human observers was more challenging. We

trained deep learning algorithms on image-level annotations only, based on reference

cytology. The top-performing deep learning network trained was able to provide a leukocyte

percentage estimation in substantial agreement with corresponding reference cytology

percentages. In addition, class activation maps revealed that the model has learned to

detect leukocytes without pixel-level annotations, relying only on leukocyte ratios.

In this proof-of-principle study we encountered some challenges. First, the 2D mosaics were

acquired at a depth a few microns above the glass interface: close enough to the glass

surface to visualize the cells resting on the glass, including both large macrophages and

small lymphocytes, but distant enough to avoid interference with the strong THG generated

signals from the glass interface itself. As the intrinsic depth sectioning of the THG and MPEF

was 2.5 micrometer, the visible cell size and cell nuclear appearance depends on the

chosen imaging depth (as shown in Fig 4D). Depth scans, to reveal the full 3D information,

could improve discrimination of the different leukocytes. Second, the reference cytology

percentages were counted not from the exact same lung fluid as imaged with the

microscope, but were from the same patients and lung area, which could affect the leukocyte

ratios. Instead of training on leukocyte ratios, an improvement would be to have a one-to-

one comparison between cytology-stained slides and label-free images, with ideally

annotated cells, which was unfortunately not feasible in this study. Third, it was more

challenging to identify the leukocyte in BALF samples compared to blood fractions, because

of the presence of more different cell types, and the overlapping characteristics of

macrophages and the other leukocytes, especially in the 2D plane. Additionally, BALF

samples were obtained from patients suspected of ILD, while blood fractions were mainly

from healthy volunteers. Therefore, it is likely that activated immune cells were present in

BALF samples that differ in appearance from healthy cells, both in cell morphology as in

generated THG and MPEF signal intensities. For example, activated lymphocytes have

increased cell sizes up to 18 µm [1] and generate stronger THG signals [11]. In addition,

multinuclear macrophages can be present, called giant cells when the cell contains more

than four nuclei. An atlas on how the different leukocytes can appear on the THG and MPEF

images, from both inactive and active cells, from healthy and diseased patients, could help

to discriminate the different leukocytes in BALF samples. Finally, because of practical

reasons, only the THG and 2PEF images were included to train and test the deep learning

network, and not the 3PEF images. The 3PEF images showed higher signal intensities from

the cellular cytoplasm compared to the 2PEF images, and therefore it could be beneficial to

include the 3PEF images in future studies.

The ultimate goal was to perform leukocyte differentiation on the BALF samples, utilizing

blood fractions as supporting material. Looking at the obtained results in S1 Table and Fig 5,

regression by deep learning based on the label-free images is feasible. With only 40

samples, annotated on the case-level by reference cytology fractions only, the algorithm

learns to differentiate four different immune cell types and accurately detect cell density at

least on the same order of magnitude as the current cytology workflow. While human

observers are limited to training by a fixed set of cell morphology characteristics, the model

has the freedom to extract any features that are necessary to achieve the same output

fractions as the cytospin golden standard. The superior regression performance on BALF

samples, compared to blood fractions, might be explained by the twice as many numbers of

BALF samples available in our dataset. Another confounding factor could be the cell density

difference between blood fraction and BALF, with the latter containing 10 times less cells for

the same area (as shown in Table 1). Blood fraction data merely served as a support to

increase the dataset, while the main focus is on BALF case differentiation and counting. The

development and optimization of the deep learning model results in a fast analysis, faster

than any counting process by a human observer. Training the ResNet50 until convergence

took less than 10 minutes, leaving plenty of room for hyperparameter optimization. One

reason for why training of the deep learning network took a short amount of time was the

size of the training dataset which contained only 40 samples (12 blood fractions and 28

BALF samples) of 2500 x 2500 pixels sized mosaics.

When training on the raw, unprocessed (except for pixel size normalization) mosaics, the

regression performance was similar to our results shown in this manuscript. However, upon

inspection using Grad-CAM, most of the activations occurred on background pixels, glass-

interface noise, and tile-by-tile stitching artefacts. Apparently, by focusing on background

space, the model was able to correlate background pixels to cell densities. To mitigate,

background subtraction was implemented using a rolling ball. Upon re-training and class

activation visualization, the model was now focusing on cellular objects instead of

background pixels. Still, the model is sensitive to intensity drops in mosaic tile corners, which

was more enhanced in the validation blood fractions compared to the test samples. This

might explain the regression performance drop in Fig 5.

The Grad-CAM atlases allow for a unique view into the model's decision process. They allow

for thorough inspection of the gradients in the last convolutional layer of the ResNet50 model

upon processing of a BALF test case mosaic. The downside of this approach is that it is

unclear how much each pixel attributes to the actual model fraction output. Upon inspection

of the activation maps, each class heatmap indicated that the model focused on cellular

objects for the prediction of each class fraction. This underlines that each cell is of

importance for the actual regression outcome of each class. This makes sense since for

each leukocyte class, the first step is actual recognition of each single cell in the mosaic

before moving on to counting only the cells of interest. In addition, cells can be activated for

multiple leukocyte classes. This is an inherent flaw in the regression task design where we

are unable to limit pixel attribution to a single class. Implementing a semantic segmentation

strategy would solve this problem but would require pixel-level annotations. This is a general

limitation in our setup where we do not have the manpower or training abilities to acquire

annotations, which made the regression strategy a favorable one to start with.

The current experimental setup serves as a proof of principle study where we show the

capabilities of deep learning to perform regression on a limited and weakly-labelled dataset.

In a next iteration, the performance could benefit from more BALF samples, and a transition

to 3D by acquisition of multi-layered mosaics across a depth of 10 to 20 µm to capture the

full morphology of each cell.

Immediate feedback on immune cell content has potential for clinical guidance. Although

BALF rarely results in a specific diagnosis, evident lymphocytosis is inconsistent with IPF.

Thereby in selected cases, when BALF reveals lymphocytosis, tissue acquisition may be

omitted. Alternatively, when lymphocytosis is absent, tissue acquisition might be the next

diagnostic step, which could be performed with immediate feedback in the same endoscopic

procedure whereas currently BALF results are awaited before a patient is scheduled for a

biopsy procedure in a separate session. Additionally, identifying and discriminating

leukocytes in the label-free images can aid identifying and discriminating of immune cells in

label-free imaging of tissues, such as cryobiopsies obtained for ILD diagnosis. We are

currently working on a study employing label-free imaging on fresh lung biopsies in ILD.

Identifying the different leukocytes in the histology samples is important for ILD diagnosis.

Moreover, considering the evolvement of THG/MPEF microscopy as an endo-microscope to

evaluate tissue *in vivo*, identification of different cell types is crucial.

Considering the surmountable challenges discussed, we believe that THG/MPEF

microscopy in combination with deep learning is a promising technique to differentiate

leukocytes in fresh, label-free imaged BALF samples, with the potential to speed-up the

diagnostic process and reduce costs and workload across the board.

535   **Acknowledgments**

536   We like to thank Tamara Dekker and Barbara Dierdorp for manually counting the leukocytes

537   in the DQ cytospin slides.

References

1. Mescher A. Junqueira's Basic Histology: Text and Atlas, Fifteenth Edition. 15th ed. McGraw-Hill Education; 2018.

2. Adderley N, Humphreys CJ, Barnes H, Ley B, Premji ZA, Johannson KA. Bronchoalveolar lavage fluid lymphocytosis in chronic hypersensitivity pneumonitis: a systematic review and meta-analysis. European Respiratory Journal. 2020;56: 2000206. doi:10.1183/13993003.00206-2020

3. Davidson KR, Ha DM, Schwarz MI, Chan ED. Bronchoalveolar lavage as a diagnostic procedure: a review of known cellular and molecular findings in various lung diseases. J Thorac Dis. 2020;12: 4991–5019. doi:10.21037/jtd-20-651

4. Jain M, Narula N, Aggarwal A, Stiles B, Shevchuk MM, Sterling J, et al. Multiphoton Microscopy: A Potential "Optical Biopsy" Tool for Real-Time Evaluation of Lung Tumors Without the Need for Exogenous Contrast Agents. Arch Pathol Lab Med. 2014;138: 1037–1047. doi:10.5858/arpa.2013-0122-OA

5. Zhang Z, de Munck JC, Verburg N, Rozemuller AJ, Vreuls W, Cakmak P, et al. Quantitative Third Harmonic Generation Microscopy for Assessment of Glioma in Human Brain Tissue. Advanced Science. 2019; 1900163. doi:10.1002/advs.201900163

6. van Huizen LMG, Kuzmin N v., Barbé E, van der Velde S, te Velde EA, Groot ML. Second and third harmonic generation microscopy visualizes key structural components in fresh unprocessed healthy human breast tissue. J Biophotonics. 2019;12. doi:10.1002/jbio.201800297

7. van Huizen LMG, Radonic T, van Mourik F, Seinstra D, Dickhoff C, Daniels JMA, et al. Compact portable multiphoton microscopy reveals histopathological hallmarks of unprocessed lung tumor tissue in real time. Transl Biophotonics. 2020;2. doi:10.1002/tbio.202000009

8.   Yang L, Park J, Marjanovic M, Chaney EJ, Spillman DR, Phillips H, et al.

Intraoperative Label-Free Multimodal Nonlinear Optical Imaging for Point-of-

Procedure Cancer Diagnostics. IEEE Journal of Selected Topics in Quantum

Electronics. 2021;27: 1–12. doi:10.1109/JSTQE.2021.3054578

9.   Tsai C-K, Chen Y-S, Wu P-C, Hsieh T-Y, Liu H-W, Yeh C-Y, et al. Imaging granularity

of leukocytes with third harmonic generation microscopy. Biomed Opt Express.

2012;3: 2234–2243. doi:10.1364/BOE.3.002234

10. Wu C-H, Wang T-D, Hsieh C-H, Huang S-H, Lin J-W, Hsu S-C, et al. Imaging

Cytometry of Human Leukocytes with Third Harmonic Generation Microscopy. Sci

Rep. 2016;6: 37210. doi:10.1038/srep37210

11. Gavgiotaki E, Filippidis G, Zerva I, Kenanakis G, Archontakis E, Agelaki S, et al.

Detection of the T cell activation state using nonlinear optical microscopy. J

Biophotonics. 2019;12. doi:10.1002/jbio.201800277

12. Monici M. Cell and tissue autofluorescence research and diagnostic applications.

Biotechnol Annu Rev. 2005;11: 227–256. doi:10.1016/S1387-2656(05)11007-2

13. Teh SK, Zheng W, Li S, Li D, Zeng Y, Yang Y, et al. Multimodal nonlinear optical

microscopy improves the accuracy of early diagnosis of squamous intraepithelial

neoplasia. J Biomed Opt. 2013;18: 036001. doi:10.1117/1.JBO.18.3.036001

14. Wang ZJ, Walsh AJ, Skala MC, Gitter A. Classifying T cell activity in autofluorescence

intensity images with convolutional neural networks. J Biophotonics. 2020;13:

e201960050. doi:10.1002/jbio.201960050

15. Shi J, Tu H, Park J, Marjanovic M, Higham AM, Luckey NN, et al. Weakly supervised

identification of microscopic human breast cancer-related optical signatures from

normal-appearing breast tissue. bioRxiv. 2022. doi:10.1101/2022.05.24.493356

16. Blokker M, Hamer PC de W, Wesseling P, Groot ML, Veta M. Fast intraoperative

histology-based diagnosis of gliomas with third harmonic generation microscopy and

deep learning. Sci Rep. 2022;12: 11334. doi:10.1038/s41598-022-15423-z

17.   Zhang Q, Yun KK, Wang H, Yoon SW, Lu F, Won D. Automatic cell counting from

stimulated Raman imaging using deep learning. PLoS One. 2021;16: e0254586.

doi:10.1371/journal.pone.0254586

18.   Lavitt F, Rijlaarsdam DJ, van der Linden D, Weglarz-Tomczak E, Tomczak JM. Deep

Learning and Transfer Learning for Automatic Cell Counting in Microscope Images of

Human Cancer Cell Lines. Applied Sciences. 2021;11: 4912.

doi:10.3390/app11114912

19.   He K, Zhang X, Ren S, Sun J. Deep Residual Learning for Image Recognition. 2015.

20.   Deng J, Dong W, Socher R, Li L-J, Kai Li, Li Fei-Fei. ImageNet: A large-scale

hierarchical image database. 2009 IEEE Conference on Computer Vision and Pattern

Recognition. IEEE; 2009. pp. 248–255. doi:10.1109/CVPR.2009.5206848

21.   Tan M, Le Q v. EfficientNetV2: Smaller Models and Faster Training. 2021.

22.   Sandler M, Howard A, Zhu M, Zhmoginov A, Chen L-C. MobileNetV2: Inverted

Residuals and Linear Bottlenecks. 2018.

23.   Kingma DP, Ba J. Adam: A Method for Stochastic Optimization. 2014.

24.   Smith LN. A disciplined approach to neural network hyper-parameters: Part 1 --

learning rate, batch size, momentum, and weight decay. 2018.

25.   Abadi M, Barham P, Chen J, Chen Z, Davis A, Dean J, et al. TensorFlow: A System

for Large-Scale Machine Learning. Proceedings of the 12th USENIX Conference on

Operating Systems Design and Implementation. USA: USENIX Association; 2016. pp.

265–283.

26.   Selvaraju RR, Cogswell M, Das A, Vedantam R, Parikh D, Batra D. Grad-CAM: Visual

Explanations from Deep Networks via Gradient-based Localization. 2016.

doi:10.1007/s11263-019-01228-7

27.   Meudec R. tf-explain. Zenodo; 2021. doi:10.5281/zenodo.5711704

28.   Tigner A, Ibrahim SA, Murray I. Histology, White Blood Cell. Treasure Island (FL):

StatPearls Publishing; 2019.

29.   Luca DC. Eosinophils. 1 Aug 2012 [cited 9 Nov 2022]. Available:

https://www.pathologyoutlines.com/topic/bonemarroweosinophils.html

Supporting information

S1 Table: Evaluation of top-performing ResNet50 model on the validation set (top rows, light

grey) and testing set (bottom rows, dark grey). The average mean absolute error on the validation

set was 0.136. The average mean absolute error on the testing set was 0.087. For each case in the

two sets, the standard cytology percentages with standard deviation (except for the PBMC in the

testing set, because this sample was counted once) are given and compared to the model's output.

Given that a mosaic may be divided into multiple, single cases, an average regression output and

standard deviation can be derived. For the blood fraction data, the macrophages class is replaced by

monocytes, since monocytes are the precursor cells of macrophages present in blood.

A

Blood fraction

Centrifuge

PBMC   Granulo-
        cytes

Bronchoalveolar lavage

B Label-free imaging

2.5 mm

C Cytology procedure

Cytospin   DQ staining

Input mosaic & fractions

Neutrophils: 0.2%
Eosinophils: 0.2%
Lymphocytes: 5.2%
Macrophages: 94.4%

Activation map

MAX

MIN

Output fractions
% Neutrophils
% Eosinophils
% Lymphocytes
% Macrophages

Deep learning model (ResNet50)

2500

3 2500
Input

112 112 112

64 64
Conv1

× 3

64 64 112 256
112

128128 512 56

× 4

256 256 28 1024 28

× 6

512 512 14 2048

× 3

softmax

1 2048
avgpool

1 4

A Blood fractions

	THG	2PEF	3PEF	DQ
Neutrophils				
Eosinophils				
Lymphocytes				

B BALF

	THG	2PEF	3PEF	DQ
Neutrophils				
Eosinophils				
Lymphocytes				

C THG 2PEF 3PEF DQ

Macrophages

D

Neutrophil

z = 0 µm ← → z = 8 µm

Neutrophil

5 µm