Artificial intelligence-driven morphology-based enrichment of malignant cells

from body fluid

Anastasia Mavropoulos[1], Chassidy Johnson[1], Vivian Lu[1], Jordan Nieto[1], Emilie Schneider[1],

Kiran Saini[1], Michael L. Phelan[1], Linda Hsie[1], Maggie Wang[1], Janifer Cruz[1], Jeanette Mei[1], Julie

Kim[1], Zhouyang Lian[1], Nianzhen Li[1], Stephane C. Boutet[1], Amy Wong-Thai[1], Weibo Yu[2], Qing-

Yi Lu[2], Teresa Kim[2], Yipeng Geng[2], Maddison (Mahdokht) Masaeli[1], Thomas D. Lee[2*], Jianyu

Rao[2,3*]

[1]Deepcell, Inc., 4025 Bohannon Dr., Menlo Park, CA, 94025

[2]Pathology and Laboratory Medicine, University of California Los Angeles (UCLA), Los Angeles,

CA, 90095

[3]To whom correspondence may be addressed. Email: JRao@mednet.ucla.edu

* Co-corresponding authors

Classification(s): Biological Sciences, Physical Sciences, Biophysics and Computational Biology

Keywords: morphology, cytology, effusion, carcinoma, high-throughput, machine learning, deep

learning, image analysis, enrichment, computer vision

ABSTRACT

Cell morphology is a fundamental feature used to evaluate patient specimens in pathological analysis. However, traditional cytopathology analysis of patient effusion samples is limited by low tumor cell abundance coupled with high background of non-malignant cells, restricting the ability for downstream molecular and functional analyses to identify actionable therapeutic targets. We applied the Deepcell platform that combines microfluidic sorting, brightfield imaging, and real-time deep learning interpretations based on multi-dimensional morphology to enrich carcinoma cells from malignant effusions without cell staining or labels. Carcinoma cell enrichment was validated with whole genome sequencing and targeted mutation analysis, which showed higher sensitivity for detection of tumor fractions and critical somatic variant mutations that were initially at low-levels or undetectable in pre-sort patient samples. Combined, our study demonstrates the feasibility and added value of supplementing traditional morphology-based cytology with deep learning, multi-dimensional morphology analysis, and microfluidic sorting.

INTRODUCTION

Cell morphology is a fundamental cell feature associated with cell identity, state, function, and disease. In recent reports, cell morphology has been used as key measurements in diverse applications including small molecule screening in cancer and infectious diseases, evaluating drug resistance, cell profiling to infer mechanism of action, and assessing metastatic capacity [1-7]. With the advent of quantitative single cell imaging technology, morphology profiling has emerged as a fundamental readout to define the functional states, such as metastatic potential, of individual cells [7]. While strides have been made in diverse applications to utilize quantitative profiling to systematically define morphological features, pathology workflows still largely rely on manual morphological observations. Trained pathologists analyze malignant cells in patient samples based on cell morphology, with the aid of dyes/stains, but these assessments may vary depending on individual pathologists and yield limited quantitative data, thereby limiting standardized information derived from tumor specimens.

Importantly, these traditional processes are unable to detect specific molecular targets that are crucial for precision care. With the availability of targeted therapies for certain genetic alterations, molecular testing has now become standard of care in cancer treatment [8,9]. For instance, tyrosine kinase inhibitors are currently used to treat lung cancer in the presence of certain mutations [10], while tumors with high mutational burden (TMB) are responsive to immune checkpoint inhibitors [11]. Thus, in the era of precision medicine, detecting specific actionable molecular targets is more important than merely confirming the presence or absence of malignancy in patient samples.

Malignant effusions are exudates containing fluid and cells in the pleural, peritoneal, and pericardial body spaces, and constitute an unequivocal sign of widespread cancer. These effusions are most commonly found in the pleural cavity with lung and breast cancers [12] and

peritoneal cavity with ovarian cancer [13]. Effusions offer a non-invasive sampling source that can

be serially collected to capture disease evolution as it represents molecular alterations present

in real-time and tissue samples may not always be available for variety of reasons. However, a

significant limitation in the molecular analysis of effusions is the presence of benign background

cells resulting in low tumor fractions which have been reported down to <0.1% of the total cell

population [14]. Consequently, molecular analysis such as next generation sequencing (NGS)

assays are not performed in these specimens as the limit of detection is approximately 10%

tumor cell fractions. Moreover, the sensitivity and specificity of cytological examination to

identify malignant cells in effusions is often limited by the morphologic similarity between

reactive mesothelial cells and tumor cells. Additionally, samples with low cancer cell fraction

often result in low sensitivity (~60%), false negative, or inconclusive diagnoses (atypical or

equivocal diagnosis) and unsuitable for downstream molecular and biomarker characterizations

[15-18].

We applied the Deepcell platform, which uses image-based deep learning and multi-

dimensional morphology analysis, to a specific proof-of-concept use case in the field of

cytology, an approach we termed "SMART" (Single-cell, Multiplex, AI-based, Real-Time)

cytology. We set out to increase sensitivity of carcinoma cell detection in cytology laboratory

workflows by training a supervised AI model to identify and isolate carcinoma cells in malignant

effusion samples. Cells classified as carcinoma were sorted by microfluidic flow and enriched

cells were analyzed by common molecular characterization techniques to verify AI-enabled

enrichment. Whole genome sequencing and targeted mutation sequencing of pre- and post-

sorted patient samples demonstrated carcinoma cell enrichment resulted in higher sensitivity,

enabling more specific clinical insight and therapeutic options.

In this study, we utilize deep learning morphological interpretations to identify and sort

carcinoma cells from lung, breast, gastro-intestinal (GI) tract, and gynecological cancers to

uncover actionable molecular insights in an unlabeled and unperturbed manner. Integrating

clinical expertise and deep neural network models based on quantitative imaging has the

potential to ultimately save lives by detecting carcinoma cells cytologists miss, isolating low-

abundance cells for molecular workup, and enabling routine monitoring of minimally invasive

liquid biopsies [19-21]. While the goal of our AI methodology is not diagnosis, our study represents

a proof-of-concept for integrating machine learning and quantitative morphological profiling to

facilitate molecular cytology.

RESULTS

AI Model Training and Validation

To begin, we used a supervised deep learning approach and trained a model, termed the

'malignant effusion AI model', to identify carcinoma cells in a background of cells typically found

in effusions such as leukocytes, histiocytes, mesothelial cells, and other immune cells. For both

model training and validation, pure cell populations from benign and malignant body fluids

originating from the pleural and peritoneal cavities were immunostained with a panel of

antibodies to identify and distinguish cell types present in the fluids. Fluorescence-activated cell

sorting (FACS) was used to isolate breast carcinoma (EpCAM+/Claudin-4+/CD45-; herein

Claudin4), leukocyte (CD45+), mesothelial (CD45-/EpCAM-/Claudin4-/Calretinin+), histiocyte

(CD45+/CD68+), and a collection of biomarker-negative cell types "other" (CD45-/EpCAM-

/Claudin4-) to obtain pure cell populations (**Fig. 1A**). The model training set included FACS-

sorted cell populations from breast cancer samples and cell lines (SK-BR-3) imaged on the

Deepcell platform, yielding 2,597,910 high-resolution brightfield images of cells (**Fig. 1B)**. The

model training set of images were sent to the deep convolutional neural network for multi-

dimensional (>1,000 dimensions) quantitative morphological image analysis and model training,

as described previously[22]. A model validation set of images, separate from images used for

model training set, was created using 2,031,024 images of both cell lines (MeT-5A) and FACS

purified patient-derived cell types, as described above and in Methods (**Fig. 1C**).

Next, we evaluated the malignant effusion AI model performance in identifying and

distinguishing breast carcinoma cells from non-carcinoma background cells with *in silico* (i.e.,

through computational simulation) model validation. The receiver operating characteristic curve

(ROC) shows high probability (AUC=0.96) of the model identifying breast carcinoma cells from

non-carcinoma cells (**Fig. 1D**). To generate a stringent ROC curve, model validation was done

using all images from the validation set and each cell class was weighed such that respective

contributions to the area under the curve (AUC) is the same.

Quantitative morphological features extracted from cell images are summarized as

"embeddings", which are low dimensional representations of images in vector space [23]. These

embedding vectors are not generally interpretable in terms of conventional morphology metrics

but can be used to perform cluster analysis to group morphologically similar cells and visualized

with Uniform Manifold Approximation and Projection (UMAP)[24]. To further validate the AI model,

we generated a UMAP plot using an *in silico* mixture of known cell types including cell lines and

isolated carcinoma, leukocyte, mesothelial, and histocyte cells from the model validation set

(**Fig. 1C**). UMAP coloring is based on known cell identities ("ground truth"), while clustering

patterns are based on AI embeddings. Breast carcinoma cells are a morphologically distinct

class, which is demonstrated by separation from other cell types and clustering of carcinoma

cells together (**Supplementary Fig. 1A**). Mesothelial and carcinoma cells cluster closely

together with some overlap to each other, suggesting some shared morphological features

between these two cell types. Histiocytes, leukocytes, and biomarker-negative "other" cells

show broader distribution and higher morphological variation to carcinoma cells than

mesothelial cells, as demonstrated by clear separation of these cell classes from carcinoma

cells in embedding space. In addition, carcinoma cells are morphologically distinct from

leukocytes which are at opposite regions of the UMAP. These common and distinct visual

features detected by the AI model are consistent with morphological observations seen in

effusion cell types by cytologists [25]. Together, this UMAP of isolated known cell types comprised

of 15,000 images (randomly samples from >2 million validation cell images) verify morphological

features detected by the deep learning model are consistent with ground truth.

To further evaluate model performance against ground truth, we compared AI-predicted and

biomarker-based (flow cytometry) carcinoma cell percentages on a cohort of pleural and ascitic

effusion samples from patients diagnosed with different cancer indications. Cytologist-confirmed

malignant effusion samples from patients diagnosed with breast (n=3), GI tract (n=5),

gynecological (n=10), and lung (n=3) cancers were divided and processed in parallel for (1)

imaging on the Deepcell platform and (2) EpCAM+/Claudin4+/CD45- biomarker immunolabeling

by flow cytometry that served as ground truth. The AI model prediction for carcinoma cell

percentage was highly concordant (R^2=0.949) to flow cytometry quantification

(EpCAM+/Claudin4+/CD45-), with correlation remaining linear over a wide dynamic range

(0.0%-91.2%) (**Fig. 1E-F**). This indicates that although label-free, the malignant effusion AI

model can detect carcinoma cells and with similar accuracy as protein expression-based

biomarkers. These results further suggest that although the model was trained on breast

carcinoma cells, the model learned distinct morphological features linked to carcinoma cells

such that carcinoma cells from other origins (e.g., GI tract, lung, and gynecological) are

distinguished from benign cells. It has been reported that the frequency of metastasis in effusion

samples from GI tract (~4-6%) and lung (~6%) cancers are lower than in breast (~40%) and

ovarian (~22%) cancers [26]. Our results are relatively consistent with these clinical findings, with

predicted carcinoma cells frequencies in GI tract and lung cancers lower (mean, median, range)

compared to breast and gynecological cancers. Importantly, the AI model was able to correctly

predict the proportion of tumor cells at low (~0%) and high (~80%) ends of tumor frequency

(**Fig. 1E, Supplementary Fig. 1B-C**). Therefore, the lower carcinoma percentage reported in

these samples reflect true biology for these cancer types. Differences in carcinoma frequency

quantification between biomarker-based flow cytometry and AI model observed were not

statistically significant (**Supplementary Fig. 1B**). Together, these data suggest the AI model

can detect carcinoma cells from diverse cancer types and independently of its tissue of origin

over a wide dynamic range.

Morphologically Similar Cell Types Cluster Together in AI Embedding Space

To assess the morphology of carcinoma cells between patients, we applied the trained and

validated malignant effusion AI model to generate cell class predictions in specimens from two

breast cancer patients. Patient A (25.2% AI carcinoma prediction) was diagnosed with Her2-

metastatic breast cancer and Patient B (42% AI carcinoma prediction) was diagnosed with

ER+/PR+/Her2 indeterminate invasive lobular breast cancer. AI embedding UMAP analysis

shows carcinoma cells from Patient A and B cluster separately from mesothelial, leukocytes,

and histiocytes, which cluster together within each respective cell class and suggest stability of

embeddings across the two patients for non-carcinoma cells (**Fig. 2A**). Interestingly, carcinoma

cells from these two cases cluster separately from each other, indicating morphologic

dissimilarity of carcinoma cells between different patients (**Fig. 2A**). Indeed, representative

brightfield images of carcinoma cells confirm differences detected by the AI model, with Patient

A cells exhibiting more granularity and textured appearance compared to Patient B cells (**Fig.**

2A), reinforcing that the cells are morphologically distinct.

To gain more insights into cell features contributing to AI cell predictions, we analyzed the

embedding vectors and the respective contribution for each cell class (**Fig. 2B**). Each cell class

exhibits unique profiles, with heatmap color representing embedding value. Several AI feature

vectors (DCX19, DCX56, DCX45, DCX44, DCX21) differentially contributing to carcinoma

predictions between Patient A and B (**Fig. 2B**). Since carcinoma cells are generally larger than

background cells present in body fluids, we next asked whether cell size was the distinguishing

feature used in AI cell class predictions. Computer vision measurements of minimum radius

indicate overlap in size between all cell classes, with carcinoma cells having highest overlap

with mesothelial cells followed by histiocytes, biomarker-marker negative "other" cells, and

leukocytes (**Fig. 2C**). This indicates that the model learned morphology features other than cell

size to classify carcinoma from non-carcinoma cells.

To further compare the morphology of carcinoma cells from breast cancer patients, we added a

third breast cancer patient to the embeddings analysis (**Supplementary Fig. 2A**). Patient A and

C carcinoma cells plot as distinct clusters but close together, suggesting high relative

morphological similarity compared to Patient B carcinoma cells. Representative carcinoma cell

images from Patient A show cells exhibit more granularity and texture, while cell images from

Patient C show cells are relatively larger with medium granularity and texture (**Supplementary**

Fig. 2A). We applied computer vision morphometric measurements to gain further insight into

morphological profiles, namely spot measurements and texture characterization. The spot

measurement isolates large dark sports in a specific size range using the Laplacian of Gaussian

filter [27] and the area opening [28]. For texture characterization, we used the uniform local binary

pattern (LBP) variant [29,30]. Computer vision morphometrics confirmed Patient A cells had more

large dark spots compared to Patients B and C (**Supplementary Fig. 2B**). Additionally,

morphometrics verified larger cells in Patient C and a different texture in the cell's periphery in

Patients A and C compared to Patient B (**Supplementary Fig. 2B**). These results are consistent

with visual examination from representative images, and morphometric measurements hint at

morphological differences underlying interpatient heterogeneity (**Supplementary Fig. 2A-B**).

Processed Cells are Viable and Compatible with Cytology Workflows

One of the major challenges of using effusion samples as liquid biopsy material is retrieving

viable and intact cells for molecular and functional analyses. Many enrichment methods rely on

fixation, epitope tags, or other techniques that may alter and destroy cells during the process.

We previously reported cell lines and dissociated tumor tissue run through the Deepcell platform

retain viability and gene expression profiles[22], but have not tested clinical effusion samples. We

evaluated cell viability of breast and ovarian cancer pleural fluid samples by trypan blue

exclusion with and without instrument processing. Paired t-testing demonstrated no significant

differences in viability after instrument processing, which is consistent with our previous results[22]

(**Fig. 3A**).

To test whether sorted cells are intact and retain morphological architecture that cytologists use

to access cell identity, ascitic fluids were collected from two different cases of high-grade serous

ovarian cancer for Papanicolaou (PAP) staining after sorting with 0.85 AI confidence threshold.

Samples were processed with cytospin, mounted, and stained on slides (**Fig. 3B,**

Supplementary Fig. 3) then examined under a microscope by a board-certified pathologist.

Results demonstrated that cells retain morphological architecture and are compatible with

cytological staining, indicating feasibility for applying this framework in clinical workflows.

To assess whether cells loaded and flown-through the instruments are not adversely affected

and maintain gene expression profiles, we performed bulk RNA-Seq on non-cancerous disease

(benign), breast, and ovarian cancer pleural effusions (**Fig. 3C-D**). For each cancer type, gene

expression analysis shows minimal differences between the initial samples and after instrument

processing, suggesting cells run through the instrument are not affected or altered (**Fig. 3C**).

The major variability observed in gene expression is driven by sample-specific signatures rather

than instrument processing and sample handling (**Fig. 3D**). Gene expression correlation

coefficient (R^2) between gene expression profiles pre- and post- flow-through was 0.983,

signifying that cells processed through the Deepcell instrument for downstream analysis are

directly comparable with unprocessed cells (**Fig. 3E**).

Validation of AI-mediated Identification of Carcinoma Cells Using Molecular Analysis

To confirm this framework is feasible for cytology workflows, pancreatic adenocarcinoma, high-

grade serous endometrial carcinoma, and high-grade ovarian/fallopian tube carcinoma ascitic

fluids were collected for PAP staining before and after cell sorting (**Fig. 4A, Supplementary**

Fig. A-B). Pre- and post-sort samples were sent to a board-certified pathologist for analysis.

Results confirmed a higher purity of cells of interest and atypical cells, with depleted

inflammatory cells eliminating visual noise on the slides. This confirms AI carcinoma predictions

and platform sorting enables higher sensitivity for cytopathology analysis.

Thus far, we have shown (1) AI carcinoma cell predictions for all cancer types are highly

concordant to protein expression-based identification of carcinoma cells by flow cytometry, (2)

AI image analysis shows that carcinoma cells are morphologically distinct cells, and (3)

carcinoma cells cluster separately from non-carcinoma cells in embedding space. To more

definitively verify that sorted AI predicted carcinoma cells are malignant, we supplemented our

analysis with molecular methods to assess and characterize cells. We performed whole genome

sequencing (WGS) followed by copy number variation (CNV) analysis on patient samples with

low (0%) to mid (25-42%) carcinoma frequencies to cover a range of initial pre-sort effusion

compositions (**Fig. 4B**). Results show high concordance between AI carcinoma predictions and

%EpCAM+Claudin4+CD45-, reaffirming consistency of the AI model with protein expression

(ground truth). Further, pre-sort samples exhibit high concordance with genomic assessment of

tumor fraction based on CNV analysis (**Fig. 4B**). After sorting, all samples demonstrated an

increase in tumor fraction after carcinoma cell enrichment, as analyzed by ichorCNA

(**Supplementary Fig. 4C**)[31]. One exception is Sample #4, which had a primary cytology

diagnosis of <1% malignant and 0% by flow cytometry, suggesting the initial processed sample

may not have contained malignant cells. Of note are Samples #8 and #11, which were

diagnosed with invasive lobular breast carcinoma and high-grade serous ovarian carcinoma,

respectively. Sample #13 carcinoma frequency was estimated at 42% by AI and 6% by flow

cytometry (EpCAM+Claudin4+CD45-), but the estimated tumor fraction at 40% was more

concordant with AI. Sample #16 was estimated at 17% by AI, 5% by flow cytometry, and 18%

by tumor fraction. These data suggest deep morphological interpretation may overcome label

and epitope-based constraints such as biomarker specificity and provide a more comprehensive

screening for malignancy. Remarkably, all cases demonstrated enriched CNV profiles and at

least 2X increase in tumor fraction with undetectable[31] (<3%) tumor fractions and post-sort

increases to detectable levels (**Fig. 4B-C**). Together, this indicates AI carcinoma predictions are

concordant with protein and genomic expression and accurate in carcinoma cell classification

and sorting, as validated by WGS. Consequently, this augmented workflow enabled detection of

actionable molecular information that was not detectable in the original patient sample.

Targeted Mutation Analysis Reveals Increase Sensitivity for VAF Detection

Lastly, targeted mutation sequencing analysis was performed on select patient samples to verify

enriched populations yielded higher sensitivity for mutational detection. Sorted carcinoma cell

samples showed striking improvement in somatic variant allele frequency (VAF) in six of the

seven patient cases (**Fig. 5**). The seven patient cases were selected based on the availability of

primary tumor mutation profiles of donor samples processed on the Deepcell platform. The

primary mutation profile panel was obtained independently using a 50-gene cancer hotspot

panel. Using a different 56 oncogene panel, we established the somatic VAF before and after

carcinoma cell sorting of the effusion samples on the Deepcell platform. Results showed

mutations detected in post-sort samples match mutations detected in the primary tumor tissue

and/or liquid biopsy specimen, affirming that the platform accurately identifies and enriches for

carcinoma cells with clinically relevant somatic mutations from diverse cancer types.

Of note, mutations that were undetectable (<1%) in three pre-sort specimens were identified at

VAFs ranging from 15% to 47% after sorting. Two pre-sort specimens with mutations at VAFs

lower than the typical limit of detection for conventional next generation sequencing

demonstrated post-sort VAFs of 15% and 34%. For example, EGFR c.2573T>G (p.L858R), a

known pathogenic somatic mutation in lung adenocarcinoma, was detected at 21% VAF in the

post-sort sample but not in the pre-sort patient sample (**Fig. 5**). These cases with pathogenic

VAF detection in post-sort but not pre-sort samples correlating with primary tumor tissue further

supports integration of AI enriched effusion samples with cell cytology workflows to facilitate

more sensitive and efficient molecular analysis. Thus, the application of deep learning to

morphological characterization of malignant effusions has the potential to enable insight into

actionable molecular lesions and inform treatment options.

DISCUSSION

Cytology is a well-established first line of diagnostic and screening method for human diseases, especially cancer. The field underwent several generations of development, from traditional smear-based cytology to liquid-based (monolayer preparation) cytology to currently digital and computational cytology. However, even with advances of digital and molecular technologies, molecular-driving cytological analysis is still limited due to the lack of robust, standardized techniques for preparing cytology samples for molecular analysis. Additionally, current morphology-based cytological analysis is qualitative, subjective, and restricted by features the human eye can define, with a mean sensitivity of 60% depending on the primary tumor, sample preparation, and individual cytologist [32]. Furthermore, molecular analysis of carcinoma cells is needed to identify actionable targets [33-36]. Morphology-based cytology analysis alone, even with the addition of ancillary techniques such as immunohistochemistry or cytogenetic analysis [37-39], is insufficient for achieving this goal given the increasing number of targets in the context of tumor heterogeneity [40-43].

Methodologies currently exist to increase tumor cellularity before molecular testing is performed. With liquid-based cytology where effusions are directly transferred to glass slides or used to create cells blocks from which microscopic sections are subsequently cut and transferred to glass slides, select areas with focally increased tumor content can be manually scrapped off and submitted for downstream processing [44]. Laser capture microdissection is a more sophisticated method involving the use of a laser to dissect out small regions or even single cells more precisely. As these methods depend on morphologic assessment by a human operator, similar issues exist as with cytological evaluation, and they are inefficient when tumor cells are present in low quantities or difficult to distinguish from benign background cells. For instance, it is historically difficult to distinguish mesothelial cells from carcinoma cells in effusion samples based on morphology, immunohistochemistry (IHC), and/or cytological stains. Though

there may be some subtle features (e.g., mesothelial cells often show "windows" and "hugging"

between cells, etc.), an individual mesothelial cell, especially when it is reactive (e.g., stimulated

by inflammation) can be difficult to distinguish from a metastatic carcinoma cell, which may

result in an equivocal or atypical diagnosis [45]. As carcinoma cells may be limited in number,

even using a battery of IHC markers may not provide a definitive answer, posing a diagnostic

challenge. Accordingly, applying machine intelligence as a facilitating tool can aid in detecting

visually problematic cells such as mesothelial versus carcinoma cells in a robust manner, as

demonstrated in this study. Of note, we were able to enrich for unlabeled carcinoma cells to the

extent that molecular changes below the standard limit of detection of current next generation

sequencing (approximately 5%) were detected at high levels ranging from 14% to 47% allele

frequencies. These results highlight the potential of the methodology to isolate rare single

cancer cells from a background of benign cells with applications in minimal residual disease

detection [46,47].

One limitation of this study is variable levels of enrichment seen in several cases. We posit this

issue can be attributed to two factors: (1) physical hardware, such as the performance and

speed of pneumatic sorting valves and (2) model performance, such as whether enough

heterogeneity of cell types and effusion specimens were used to train the model. In future

iterations, mechanical improvements and adjustments in microfluidic flow rates can be made to

increase purity levels. Furthermore, additional biological diversity of patient specimens from

different cancer types and varying carcinoma frequencies may enhance model performance and

enrichment levels. For instance, in this study we observed partial overlapping of carcinoma and

mesothelial cells as observed in the AI embeddings UMAP clustering, which may result in

decreased depletion of mesothelial cells compared to other non-carcinoma cells. To reduce the

degree of overlapping, we expect that the model can be trained further with additional annotated

cell images of cell line and patient-derived mesothelial cells to discern carcinoma and

mesothelial morphological differences. We interpret this limitation as related to the same

cytopathologist challenge in distinguishing morphologically similar cells such as carcinoma cells

and mesothelial cell compared to smaller lymphocytes, even with aid of immunohistochemical

stains.

Applying advancements in artificial intelligence (AI), specifically deep learning, to tissue-based

images has become an intense area of research in digital pathology with applications including

cancer detection [48,49] and staging based on the tumor architecture [50,51], among others. Further

support for this paradigm shift in precision medicine was demonstrated in a recent study

showing the combination of AI and clinician detects more instances of breast cancer than either

alone [19]. However, AI has not been extrapolated to single-cell cytology analysis such as label-

free tumor cell enrichment. Here we present an AI-driven technology platform to classify and

isolate carcinoma cells from breast, lung, GI tract, and gynecological cancer effusions using

learned morphological representations. Interestingly, although the AI model was trained on

breast carcinoma cells (cell line and isolated patient cells), it performed remarkably well on other

cancer types assessed in this study (lung, GI tract, gynecological). This indicates carcinoma

cells residing in effusions share similar morphological properties, which may potentially be

associated with metastatic fitness and capacityl[6,7].

Importantly, as proof-of-concept for precision medicine applications, while critical molecular

events such as clinically relevant actionable mutational targets were undetectable without

processing, such mutations were detected after processing. Microfluidic sorting, although not

currently highly efficient, allows effusions to retain viability and compatibility with genomic

analysis such as RNA-seq and WGS, making sorted cells suitable for a comprehensive

assessment of patient's mutation profile. We additionally validated sorted cells are compatible

with downstream single-cell RNA-Seq and live cell culture for drug testing and other

experimental conditions (data not shown). Beyond the physical enrichment of target cells, high-

content images are captured and stored, allowing for continuous measurement and re-analysis

of captured images to detect additional phenotypes resulting from differing cell states/types,

drug treatments, or molecular aberrations at single-cell resolution. This technology provides the

foundation for the future practice of SMART cytology.

MATERIALS AND METHODS

Specimen Collection

Ascites and pleural effusion specimens were collected from patients admitted to UCLA Medical

Center (Los Angeles, CA) and reviewed by trained pathologists or cytologists. As byproducts of

diagnostic procedures, samples were de-identified and the study was exempted from the UCLA

Institutional Review Board. Specimens with >100mL leftover fluid following cytological

diagnostic analysis were shipped to Deepcell (Menlo Park, CA). Samples were stored at 4°C

within 24 hours of collection, shipped on ice, and processed within <48 hours from time of

collection. Specimen information including time of collection, treatment history, date of

treatment, cytopathology and molecular diagnostic results, histological diagnosis, and

percentage of malignant cells determined by cytology was collected and shared. Additional

patient specimens were commercially obtained from Bio-Options (Brea, CA, USA) and

Discovery Life Sciences (DLS; Los Osos, CA, USA).

Sample Preparation

Upon receipt, effusion samples were centrifuged to pellet the cells. The cell pellet was

resuspended in Ham's F-12K (Kaighn's) Medium (Gibco) + 10% Fetal Bovine Serum (Peak

Serum) and strained using a 40µm cell strainer (Corning), then inspected for cell clumps. If

clumps were present in the strainer, 1mg/mL collagenase was used for dissociation (Sigma-

Aldrich). Combined cells and dissociated clumps were then fixed with 4% paraformaldehyde

(EMS) for 20 minutes at room temperature with gentle agitation and resuspended with calcium-

and magnesium-free PBS (Gibco). If RBC were present, the cell suspension was stored

overnight at 4°C to allow spontaneous hemolysis then washed with PBS. Fixed samples were

stored at 4°C until use. Single cell suspensions and removal of red blood cells (RBC) was

performed with RBC lysis buffer diluted in water according to manufacturer's recommendations

(BioLegend). For cryopreservation, when applicable, all steps above were performed, cells were

then washed and resuspended in CryoStor® CS10 (Stemcell Technologies). For cell viability

assessment, pre-sort or post-sort cells were stained with trypan blue (Gibco). Cells were then

counted under a microscope (Leica).

Cytology

We used conventional cytology methods to perform cytomorphological analysis of body fluids

samples. Briefly, an aliquot (50 mL) of effusion fluid sample was transferred to a 50 mL conical

tube and centrifuged. The supernatant was decanted by manual pipetting, so as to retain 3-4

mL of pellet mixing well with supernatant remaining in the tube to make cytospin slides and cell

block. For samples containing red blood cells indicated by pink to red colored specimens, the

cell pellet was overlaid with 10 mL of CytoRich Red (BD), resuspended, and centrifuged

again. After assembling Shandon cytospin slide into the cytospin processing bin, one to three

drops of cell pellet suspension were added into each cytospin chamber. After securely fasten

the lid onto the processing bin ("Seated head roter"), the cystospin bin was placed in the

Cytospin processor (Richard-Allan Scientific) and centrifuged at 191 g for 5 minutes. Slides

were immediately fixed in 95% ethanol for 20 min. The staining process was carried out using a

Tissue-Tek Prisma® Plus Automated Slide Stainer (Prisma-P-AD, Sakura Finetek USA). For

cell block preparation, cells were fixed by adding 1 mL of zinc formalin to the remaining cell

pellet, mixed by pipetting, and centrifuged. After stood for 1 hour, the material was transferred

into a biopsy bag, and submitted to the histological laboratory for paraffin embedding, sectioning

(4 to 6 micron), and routine Hematoxylin and Eosin (H&E) staining. All prepared slides were

reviewed and examined by clinical technologists and board-certified pathologists under the light

microscope (BX53, Olympus).

Cell Viability

For cell viability assessment, pre-run (T = 0 hr), post-run (flown-through + T = 3 hr), and "un-run"

(room temperature for T = 3 hr) cells were stained with trypan blue (Gibco). Cells were then

counted under a microscope (Leica). Relative cell viability was calculated between post-run/pre-

run cells and un-run/pre-run cell viability.

Sample Preparation for Model Training and Validation

For AI model training and validation, pure populations of known cells were be obtained by FACS

isolation of CD45+ (leukocyte, Biolegend Cat#304006), CD45+/CD68+ (histiocytes, BioLegend,

Cat#333827), Calretinin (mesothelial, Santa Cruz Biotechnology Cat#sc-365956), and

EpCAM+/Claudin4+ (carcinoma cells, R&D Systems Cat#AAGV0320031). Fixed cells isolated

from patient samples were stained with respective primary antibodies for 30 minutes at 25°C

and washed twice with BD Stain buffer (Becton-Dickinson, Cat#554656). Cells were FACS

sorted on a BD Melody (Becton-Dickinson) using the following gating strategy: First, debris is

eliminated by FSC-A vs. SSC-A, then cell clumps and doublets are removed by SSC-H vs.

SSC-W and FSC-H vs. FSC-W. Leukocytes are identified by CD45 expression and gated further

by CD45+ vs. CD68 with histiocytes identified by CD68 expression. The CD45- population is

investigated by Claudin4 vs. EpCAM for cancer which is determined by both Claudin4 and

EpCAM expression and the population with no expression from Claudin4 and EpCAM being

defined as CD45- other population which includes other non-hematopoietic and epithelial cells.

The CD45- other population is then gated by CD45- vs. Calretinin with mesothelial cells

identified by Calretinin expression. In addition to patient samples, commercially available cell

lines SK-BR-3 (ATCC, Catt# HTB-30™) and MeT-5A (ATCC, Cat# CRL-9444™) maintained per

manufacturer's recommendation were used for model training and validation.

Model Training and Validation

Instrument

The Deepcell instrument consists of hardware (microfluidic chip, high-resolution imager, real-

time signal processing), deep learning convolutional neural network model[52], and proprietary

software. Captured images are cropped around the cell and fed into a convolutional neural

network for generation of multi-dimensional morphological descriptors and classification in real-

time. Based on cell classification, pneumatic sorting valves are activated for either target cell

collection or waste. Target cells are classified using a confidence threshold level over 0.85

(based on the softmax activation function[53], which transforms the raw neural network outputs

into probability vectors) and collected. Cells below the set confidence threshold and other cell

classes go to waste. Sorted cells are then retrieved for downstream analysis.

Dataset

To develop the 'malignant effusion AI model', pure cell populations (cell lines, FACS-sorted

patient samples) were divided into multiple aliquots and imaged on more than one instrument to

mitigate inter-instrument variations on the model training. Pure cell populations were used to

curate training and validation sets composed of the following cell classes: breast carcinoma,

CD45+, CD45- and histiocytes. Distinct samples and cell lines were used for the training and

validation sets. For each cell class, we manually annotated "out of focus" images (cell that is

blurred and intracellular features are not sharp or clearly defined) and debris/non-cellular

images (bubbles, blank frames, unhealthy cells, cell clumps) to exclude images that may cause

misclassification.

Classification model

All cell images were resized to 299x299x3 pixels to make them compatible with the InceptionV3

architecture[52]. The training progressed 30k steps with a batch size of 1024, and a 0.16 learning

rate on an Adam optimizer. The training algorithm was modified to consider the skewed

distribution of cell classes by providing class weights which are inversely proportional to the

class frequencies in the training data. The classifier thus learns equally from all classes.

Embedding model

We implemented a variant of the InceptionV3 model to construct the embedding model. From

one of the top layers proximal to the softmax output layer, we took one of the final dense layers

and input it into a size 64 embedding layer. This embedding InceptionV3 model was trained with

the same hyperparameters as in the supervised classification model training.

Bulk RNA-Seq

Total RNA was extracted from cells using the RNeasy Mini Kit (Qiagen). cDNA synthesis,

amplification, and library preparation were performed with the Quantseq 3'm RNA-seq Library

Prep Kit (Lexogen) according to the manufacturer's protocol. The final libraries were sequenced

on the Illumina Miniseq system using 150-cycle MiniSeq High Output Reagent Kit (FC-420-

1002). The demultiplexed reads in FASTQ format were trimmed for adapter sequence using the

Atropos[54] software with Illumina universal primer read sequence and the trim parameters, where

the maximum trimmed length is set at 50bp and quality score cutoff is set at 20, with the option

of discarding reads containing more than 5 N bases and removing flanking N bases from each

read [54]. The trimmed reads are then aligned to the human genome, GRCh38.p13 using STAR

aligner [55]. The raw count for each gene is generated, where the Ensembl (https://ensembl.org;

European Bioinformatics Institute) gene ID is used as an identifier. For downstream gene

expression analysis, the raw count data from each individual sample is aggregated into a single

raw count table. Differential expression analysis was done using DESeq2 [56].

WGA and CNV Preparation

Whole genome amplification (WGA) was performed using ResolveDNA™ WGA (BioSkryb),

according to manufacturer's protocol, on 500-1000 both presorted and Deepcell-sorted cells.

NGS libraries were constructed using WM kit (Watchmaker Genomics), 500 ng of WGA product

were used to prepare a library for sequencing with Watchmaker DNA Library Prep Kit with

Fragmentation. The library was sequenced for copy number variation (CNV) analysis on

Illumina NextSeq 1000 (Illumina) with NextSeq1000/2000 P2 Reagents (300 Cycles) v3 reagent

(Illumina).

CNV Analysis Method

The raw FASTQ files are aligned to the human reference genome, GRCh38.p13, using BWA-

MEM algorithm by bwa software, version 0.7.17 [57]. The aligned reads are checked for

duplication, and those reads with the alignment quality score 20 and above are selected for the

CNV analysis. The read count coverage files for each sample are generated in WIG format

using HMMCopy's readCounter command with 1Mb bin size option. The panels of normals are

prepared from the benign cell lines, GM12878, GM24143, GM24149, GM24385, and HFL1, to

be used as a reference during the CNV analysis, following the recommended processing steps

provided by ichorCNA manual. These files are used as an input to ichorCNA software

(bioconda r-ichorcna version 0.3.2), for CNV analysis and tumor fraction prediction [31], using the

ranges of ploidies of 1.75, 2, and 2.25 and initial normal contamination percentage ranging from

1% to 99%. The solutions ranked by the descending log-likelihood values by the ichorCNA are

manually reviewed to eliminate the suboptimal solutions and to identify the solution that best

explains the data, as recommended by ichorCNA.

Spike-in Data CNV Analysis

To evaluate the tumor fraction prediction performance from ichorCNA on the whole genome

sequencing data generated and processed at Deepcell, a benign cell line, GM12878 and lung

adenocarcinoma cell line, NCI-H23, were mixed at various ratio with expected tumor fractions

ranging at 5, 10, 20, 50% using 25, 50, and 100 cells. Whole genome amplification was

performed on the prepared mixture samples with BioSkryb ResolveDNA™ Whole Genome

Amplification Kit (BioSkryb Genomics) and Takara PicoPlex WGA v3 kit (Takara) following the

manufacture's protocols. The libraries were prepared using Roche KAPA HyperPlus Kit (Roche)

with 500 ng WGA product input and sequenced with illumina Nextseq 1000 (Illumina). The

observed tumor fraction and expected tumor fraction were compared and fitted using the linear

regression model.

Mutation Analysis with Ion AmpliSeq Cancer Hotspot Panel v2

Targeted amplification of multiple coding regions of 50 genes recurrently mutated in solid

tumors (*ABL1, AKT1, ALK, APC, ATM, BRAF, CDH1, CDKN2A, CSF1R, CTNNB1, EGFR,*

ERBB2, ERBB4, EZH2, FBXW7, FGFR1, FGFR2, FGFR3, FLT3, GNA11, GNAQ, GNAS,

HNF1A, HRAS, IDH1, IDH2, JAK2, JAK3, KDR, KIT, KRAS, MET, MLH1, MPL, NOTCH1,

NPM1, NRAS, PDGFRA, PIK3CA, PTEN, PTPN11, RB1, RET, SMAD4, SMARCB1, SMO,

SRC, STK11, TP53, and *VHL*) was performed using the Ion AmpliSeq Cancer Hotspot Panel v2

(ThermoFisher Scientific, Waltham, MA). Following next-generation sequencing on the Ion

Torrent PGM system (ThermoFisher), the raw sequencing data was processed using the

Torrent Suite Software and Ion Reporter Software with resultant bases aligned to targeted

regions (hg19). Variant call format (VCF) files were generated with the variantCaller plugin.

Detected alterations were filtered and reviewed manually to determine pathogenicity and

reportability. Depending upon the total amount of DNA present, the assay can detect ~5% of an

alternate allele in a background of reference alleles. This method is limited to the detection of

certain categories of mutations, including single nucleotide variants and small

insertions/deletions (≤18 nucleotides), within the regions of the genome targeted by this assay.

Mutation Analysis with xGen™ 56G Onco Amplicon Panel v2

Target amplification for mutation detection was performed using 20 ng of WGA product with

Integrated DNA Technologies xGen™ Amplicon Core kit and xGen™ 56G Onco Amplicon

Panel v2 (Integrated DNA Technologies), which offers comprehensive and hotspot coverage of

56 clinically-relevant oncology-related genes (ABL1, AKT1, ALK, APC, ATM, BRAF, CDH1,

CDKN2A, CSF1R, CTNNB1, DDR2, DNMT3A, EGFR, ERBB2, ERBB4, EZH2, FBXW7,

FGFR1, FGFR2, FGFR3, FLT3, FOXL2, GNA11, GNAQ, GNAS, HNF1A, HRAS, IDH1, IDH2,

JAK2, JAK3, KDR, KIT, KRAS, MAP2K1, MET, MLH1, MPL, MSH6, NOTCH1, NPM1, NRAS,

PDGFRA, PIK3CA, PTEN, PTPN11, RB1, RET, SMAD4, SMARCB1, SMO, SRC, STK11,

TP53, TSC1, VHL). Target amplicons were sequenced using Illumina Miniseq with Miniseq Mid

Output Kit (300-cycles) reagent (Illumina). Depending on DNA input and sequencing depth, the

assay can detect ~1% of an alternate allele in a background of reference alleles.

Mutation Analysis Method

The Illumina adapter-trimmed FASTQ files are aligned to the human reference genome,

GRCh38.p13, using BWA-MEM algorithm, followed by the clipping of amplicon primer-binding

regions by ampliconclip, template length correction by fixmates and *MD* and *NM* tag fix by

calmd, all available from samtools software, version 1.14 [58,59]. The germline calls are made

using the pileup data, using the aligned reads that overlap with the target regions of the given

assay. The initial calls are made against the reference genome sequences, followed by the

manual review using the IGV viewer visualization tool to eliminate the false positive calls. The

somatic calls are made using Mutect2 tools from GATK4 version 4.2.2.0 [60-62] on amplicon target

sites, using the matched normal sample. When a matched normal sample is unavailable,

Mutect2 is used with the panels of normals, generated from 1000 Genomes Project and

gnomAD population allele frequency VCF file. All somatic calls are further filtered using GATK4

FilterMutectCalls, with minimum unique reads supporting the alternate allele set to 2, minimum

alt reads required on both forward and reverse strands set at 2, and minimum allele fraction

required set at 0.01, followed by the manual review using IGV visualization tool.

Statistical Methods and Hypothesis Testing

When applicable, statistically significant differences between experimental conditions were

tested with GraphPad Prism Software (Version 9.4.0).

REFERENCES

Tanaka, M. *et al.* An unbiased cell morphology-based screen for new, biologically active small molecules. *PLoS Biol* **3**, e128 (2005). https://doi.org:10.1371/journal.pbio.0030128

Puspita, N. A. & Bedford, A. Morphological Changes of Cisplatin-resistant Human Breast Cancer MCF-7 Cell Line. *International Journal of Integrated Health Sciences* **5**, 8-14 (2017). https://doi.org:10.15850/ijihs.v5n1.960

Regev, A. *et al.* The Human Cell Atlas. *Elife* **6** (2017). https://doi.org:10.7554/eLife.27041

Nassiri, I. & McCall, M. N. Systematic exploration of cell morphological phenotypes associated with a transcriptomic query. *Nucleic Acids Res* **46**, e116 (2018). https://doi.org:10.1093/nar/gky626

Mirabelli, C. *et al.* Morphological cell profiling of SARS-CoV-2 infection identifies drug repurposing candidates for COVID-19. *Proc Natl Acad Sci U S A* **118** (2021). https://doi.org:10.1073/pnas.2105815118

Nguyen, A., Yoshida, M., Goodarzi, H. & Tavazoie, S. F. Highly variable cancer subpopulations that exhibit enhanced transcriptome variability and metastatic fitness. *Nature Communications* **7**, 11246 (2016). https://doi.org:10.1038/ncomms11246

Wu, P. H. *et al.* Single-cell morphology encodes metastatic potential. *Sci Adv* **6**, eaaw6938 (2020). https://doi.org:10.1126/sciadv.aaw6938

Cobain, E. F. *et al.* Assessment of Clinical Benefit of Integrative Genomic Profiling in Advanced Solid Tumors. *JAMA Oncology* **7**, 525-533 (2021). https://doi.org:10.1001/jamaoncol.2020.7987

El-Deiry, W. S. *et al.* The current state of molecular testing in the treatment of patients with solid tumors, 2019. *CA Cancer J Clin* **69**, 305-343 (2019). https://doi.org:10.3322/caac.21560

Lindeman, N. I. *et al.* Updated Molecular Testing Guideline for the Selection of Lung Cancer Patients for Treatment With Targeted Tyrosine Kinase Inhibitors: Guideline From the College of American Pathologists, the International Association for the Study of Lung Cancer, and the Association for Molecular Pathology. *Journal of Thoracic Oncology* **13**, 323-358 (2018). https://doi.org:https://doi.org/10.1016/j.jtho.2017.12.001

Klempner, S. J. *et al.* Tumor Mutational Burden as a Predictive Biomarker for Response to Immune Checkpoint Inhibitors: A Review of Current Evidence. *Oncologist* **25**, e147-e159 (2020). https://doi.org:10.1634/theoncologist.2019-0244

Light, R. W. Clinical practice. Pleural effusion. *N Engl J Med* **346**, 1971-1977 (2002). https://doi.org:10.1056/NEJMcp010731

Ferlay, J. *et al.* Cancer incidence and mortality worldwide: Sources, methods and major patterns in GLOBOCAN 2012. *International Journal of Cancer* **136**, E359-E386 (2015). https://doi.org:https://doi.org/10.1002/ijc.29210

Kipps, E., Tan, D. S. & Kaye, S. B. Meeting the challenge of ascites in ovarian cancer: new avenues for therapy and research. *Nat Rev Cancer* **13**, 273-282 (2013). https://doi.org:10.1038/nrc3432

15 Jerz, J. L. *et al.* Detection of malignancy in body fluids: a comparison of the hematology and
cytology laboratories. *Arch Pathol Lab Med* **138**, 651-657 (2014).
https://doi.org:10.5858/arpa.2013-0295-OA

16 Maskell, N. A. & Butland, R. J. BTS guidelines for the investigation of a unilateral pleural effusion
in adults. *Thorax* **58 Suppl 2**, ii8-17 (2003). https://doi.org:10.1136/thorax.58.suppl_2.ii8

17 Moriarty, A. T. *et al.* A liquid concept--do classic preparations of body cavity fluid perform
differently than ThinPrep cases? Observations from the College of American Pathologists
Interlaboratory Comparison Program in Nongynecologic Cytology. *Arch Pathol Lab Med* **132**,
1716-1718 (2008). https://doi.org:10.5858/132.11.1716

18 Pinto, D., Chandra, A., Crothers, B. A., Kurtycz, D. F. I. & Schmitt, F. The international system for
reporting serous fluid cytopathology-diagnostic categories and clinical management. *J Am Soc*
Cytopathol **9**, 469-477 (2020). https://doi.org:10.1016/j.jasc.2020.05.015

19 Leibig, C. *et al.* Combining the strengths of radiologists and AI for breast cancer screening: a
retrospective analysis. *The Lancet Digital Health* **4**, e507-e519 (2022).
https://doi.org:10.1016/s2589-7500(22)00070-x

20 Schaffter, T. *et al.* Evaluation of Combined Artificial Intelligence and Radiologist Assessment to
Interpret Screening Mammograms. *JAMA Netw Open* **3**, e200265 (2020).
https://doi.org:10.1001/jamanetworkopen.2020.0265

21 McKinney, S. M. *et al.* International evaluation of an AI system for breast cancer screening.
Nature **577**, 89-94 (2020). https://doi.org:10.1038/s41586-019-1799-6

22 Salek, M. *et al.* Realtime morphological characterization and sorting of unlabeled viable cells
using deep learning. *bioRxiv*, 2022.2002.2028.482368 (2022).
https://doi.org:10.1101/2022.02.28.482368

23 Neu, D. A., Lahann, J. & Fettke, P. A systematic literature review on state-of-the-art deep learning
methods for process prediction. *Artificial Intelligence Review* **55**, 801-827 (2021).
https://doi.org:10.1007/s10462-021-09960-8

24 McInnes, L., Healy, J. & Melville, J. UMAP: Uniform Manifold Approximation and Projection for
Dimension Reduction. *arXiv* (2020). https://doi.org:10.48550/arXiv.1802.03426

25 Cakir, E., Demirag, F., Aydin, M. & Unsal, E. Cytopathologic differential diagnosis of malignant
mesothelioma, adenocarcinoma and reactive mesothelial cells: A logistic regression analysis.
Diagn Cytopathol **37**, 4-10 (2009). https://doi.org:10.1002/dc.20938

26 Bedrossian, C. W. M. Malignant Effusions: A Multimodal Approach to Cytologic Diagnosis.
(1994).

27 Chen, J. S., Huertas, A. & Medioni, G. Fast Convolution with Laplacian-of-Gaussian Masks. *IEEE*
Transactions on Pattern Analysis and Machine Intelligence **PAMI-9**, 584-590 (1987).
https://doi.org:10.1109/TPAMI.1987.4767946

28 Vincent, L. M. Grayscale area openings and closings, their efficient implementation and
applications. *Proc. EURASIP Workshop on Mathematical Morphology and its Applications to*
Signal Processing, 22-27 (1993).

29   Ojala, T., Pietikäinen, M. & Harwood, D. A comparative study of texture measures with
classification based on featured distributions. *Pattern Recognition* **29**, 51-59 (1996).
https://doi.org:https://doi.org/10.1016/0031-3203(95)00067-4

30   Ojala, T., Pietikainen, M. & Maenpaa, T. Multiresolution gray-scale and rotation invariant texture
classification with local binary patterns. *IEEE Transactions on Pattern Analysis and Machine*
Intelligence **24**, 971-987 (2002). https://doi.org:10.1109/TPAMI.2002.1017623

31   Adalsteinsson, V. A. *et al.* Scalable whole-exome sequencing of cell-free DNA reveals high
concordance with metastatic tumors. *Nature Communications* **8**, 1324 (2017).
https://doi.org/10.1038/s41467-017-00965-y

32   Psallidas, I., Kalomenidis, I., Porcel, J. M., Robinson, B. W. & Stathopoulos, G. T. Malignant
pleural effusion: from bench to bedside. *Eur Respir Rev* **25**, 189-198 (2016).
https://doi.org/10.1183/16000617.0019-2016

33   Garraway, L. A. Genomics-Driven Oncology: Framework for an Emerging Paradigm. *Journal of*
Clinical Oncology **31**, 1806-1814 (2013). https://doi.org:10.1200/jco.2012.46.8934

34   Gokozan, H. N. *et al.* Combining molecular testing and the Bethesda category III:VI ratio for
thyroid fine-needle aspirates: A quality-assurance metric for evaluating diagnostic performance in
a cytopathology laboratory. *Cancer Cytopathology* **130**, 259-274 (2022).
https://doi.org:https://doi.org/10.1002/cncy.22542

35   Lindeman, N. I. *et al.* Updated Molecular Testing Guideline for the Selection of Lung Cancer
Patients for Treatment With Targeted Tyrosine Kinase Inhibitors: Guideline From the College of
American Pathologists, the International Association for the Study of Lung Cancer, and the
Association for Molecular Pathology. *Archives of Pathology & Laboratory Medicine* **142**, 321-346
(2018). https://doi.org:10.5858/arpa.2017-0388-CP

36   Salto-Tellez, M. Diagnostic Molecular Cytopathology - a further decade of progress.
Cytopathology **26**, 269-270 (2015). https://doi.org:https://doi.org/10.1111/cyt.12276

37   Capper, D. *et al.* Assessment of BRAF V600E mutation status by immunohistochemistry with a
mutation-specific monoclonal antibody. *Acta Neuropathol* **122**, 11-19 (2011).
https://doi.org/10.1007/s00401-011-0841-z

38   Mino-Kenudson, M. *et al.* A Novel, Highly Sensitive Antibody Allows for the Routine Detection of
ALK-Rearranged Lung Adenocarcinomas by Standard Immunohistochemistry. *Clinical Cancer*
Research **16**, 1561-1571 (2010). https://doi.org:10.1158/1078-0432.Ccr-09-2845

39   Yemelyanova, A. *et al.* Immunohistochemical staining patterns of p53 can serve as a surrogate
marker for TP53 mutations in ovarian carcinoma: an immunohistochemical and nucleotide
sequencing analysis. *Mod Pathol* **24**, 1248-1253 (2011).
https://doi.org/10.1038/modpathol.2011.85

40   Baloch, Z. W. & Gupta, P. K. Cytopathology Comes of Age. *Acta Cytologica* **64**, 5-6 (2020).
https://doi.org/10.1159/000502373

41   Fisher, R., Pusztai, L. & Swanton, C. Cancer heterogeneity: implications for targeted
therapeutics. *Br J Cancer* **108**, 479-485 (2013). https://doi.org/10.1038/bjc.2012.581

42   Salto-Tellez, M. More Than a Decade of Molecular Diagnostic Cytopathology Leading Diagnostic
and Therapeutic Decision-Making. *Arch Pathol Lab Med* **142**, 443-445 (2018).
https://doi.org/10.5858/apra.2017-0258-ED

43   Sottoriva, A. *et al.* Intratumor heterogeneity in human glioblastoma reflects cancer evolutionary
dynamics. *Proceedings of the National Academy of Sciences* **110**, 4009-4014 (2013).
https://doi.org:doi:10.1073/pnas.1219747110

44   Roy-Chowdhuri, S. *et al.* Factors affecting the success of next-generation sequencing in cytology
specimens. *Cancer Cytopathol* **123**, 659-668 (2015). https://doi.org:10.1002/cncy.21597

45   Shidham, V. B. The panorama of different faces of mesothelial cells. *Cytojournal* **18**, 31 (2021).
https://doi.org:10.25259/cmas_02_02_2021

46   Jongen-Lavrencic, M. *et al.* Molecular Minimal Residual Disease in Acute Myeloid Leukemia. *N
Engl J Med* **378**, 1189-1199 (2018). https://doi.org:10.1056/NEJMoa1716863

47   Pantel, K. & Alix-Panabières, C. Liquid biopsy and minimal residual disease - latest advances and
implications for cure. *Nat Rev Clin Oncol* **16**, 409-424 (2019). https://doi.org:10.1038/s41571-019-
0187-3

48   Ehteshami Bejnordi, B. *et al.* Diagnostic Assessment of Deep Learning Algorithms for Detection
of Lymph Node Metastases in Women With Breast Cancer. *Jama* **318**, 2199-2210 (2017).
https://doi.org:10.1001/jama.2017.14585

49   Saito, A. *et al.* Prediction of early recurrence of hepatocellular carcinoma after resection using
digital pathology images assessed by machine learning. *Mod Pathol* **34**, 417-425 (2021).
https://doi.org:10.1038/s41379-020-00671-z

50   Bulten, W. *et al.* Automated deep-learning system for Gleason grading of prostate cancer using
biopsies: a diagnostic study. *Lancet Oncol* **21**, 233-241 (2020). https://doi.org:10.1016/s1470-
2045(19)30739-9

51   Ström, P. *et al.* Artificial intelligence for diagnosis and grading of prostate cancer in biopsies: a
population-based, diagnostic study. *Lancet Oncol* **21**, 222-232 (2020).
https://doi.org:10.1016/s1470-2045(19)30738-7

52   Szegedy, C., Vanhoucke, V., Ioffe, S., Shlens, J. & Wojna, Z. in *2016 IEEE Conference on
Computer Vision and Pattern Recognition (CVPR).* 2818-2826.

53   Goodfellow, I., Yoshua, B. & Courville, A. 6.2.2.3 Softmax Units for Multinoulli Output
Distributions. *Deep Learning, MIT Press*, 180-184 (2016).

54   Didion, J. P., Martin, M. & Collins, F. S. Atropos: specific, sensitive, and speedy trimming of
sequencing reads. *PeerJ* **5**, e3720 (2017). https://doi.org:10.7717/peerj.3720

55   Dobin, A. *et al.* STAR: ultrafast universal RNA-seq aligner. *Bioinformatics* **29**, 15-21 (2013).
https://doi.org:10.1093/bioinformatics/bts635

56   Love, M. I., Huber, W. & Anders, S. Moderated estimation of fold change and dispersion for RNA-
seq data with DESeq2. *Genome Biology* **15**, 550 (2014). https://doi.org:10.1186/s13059-014-
0550-8

57   Li, H. & Durbin, R. Fast and accurate short read alignment with Burrows–Wheeler transform.
Bioinformatics **25**, 1754-1760 (2009). https://doi.org:10.1093/bioinformatics/btp324

58   Li, H. *et al.* The Sequence Alignment/Map format and SAMtools. *Bioinformatics* **25**, 2078-2079
(2009). https://doi.org:10.1093/bioinformatics/btp352

59   Danecek, P. *et al.* Twelve years of SAMtools and BCFtools. *Gigascience* **10** (2021).
https://doi.org:10.1093/gigascience/giab008

60   Benjamin, D. *et al.* Calling Somatic SNVs and Indels with Mutect2. *bioRxiv*, 861054 (2019).
https://doi.org:10.1101/861054

61   Cibulskis, K. *et al.* Sensitive detection of somatic point mutations in impure and heterogeneous
cancer samples. *Nature Biotechnology* **31**, 213-219 (2013). https://doi.org:10.1038/nbt.2514

62   Poplin, R. *et al.* Scaling accurate genetic variant discovery to tens of thousands of samples.
bioRxiv, 201178 (2018). https://doi.org:10.1101/201178

Figure 1

Figure 1: Development and Verification of AI Model for Carcinoma Cells in Effusion Samples

(A) Graphical schematic of experimental workflow. Top: Patient samples are labeled with biomarkers to

identify and isolate breast carcinoma (EpCAM+/Claudin4+/CD45-), leukocyte (CD45+), mesothelial

(CD45-/EpCAM-/Claudin4-/Calretinin+), histiocyte (CD45+/CD68+), and a collection of biomarker-

negative cell types "other" (CD45-/EpCAM-/Claudin4-) to obtain corresponding pure populations using

FACS. Resulting cell populations of known identity are imaged with the Deepcell instrument, and

brightfield images designated as "ground truth" are input into the deep learning neural network to train

the "malignant effusion AI model". Bottom: After model training and validation, patient samples

(without labels or stains) are imaged and sorted in real-time using the Deepcell instrument based on

model classification. AI model predictions of carcinoma cells are then confirmed with cytology and

molecular analysis.

(B) Summary table of training sets, including clinical specimens, cells lines, and some technical replicates

and number of images used for model training.

(C) Summary table of validation sets, including clinical specimens, cells lines, and some technical

replicates and number of images used for used for *in silico* model validation.

(D) Receiver operating characteristic (ROC) curve plotting true positive rate (TPR) against false positive

rate (FPR), illustrating performance of the classification model. The probability of the classification

model to detect breast carcinoma cells from non-carcinoma cells is represented by the area under the

curve (AUC) of 0.96.

(E) Proportion in percent of EpCAM+Claudin4+CD45- carcinoma cells quantified by flow cytometry (gray)

compared to the proportion in percent of AI model prediction of carcinoma cell (orange) in patient-

derived breast, GI tract, gynecological, and lung cancer effusion samples.

(F) Correlation plot showing proportion in percent of EpCAM+Claudin4+CD45- carcinoma cells

(quantified by flow cytometry) and the proportion in percent of AI model prediction of carcinoma

(quantified by AI) in effusion samples, related to **Fig. 1E**. Correlation coefficient of determination (R-

squared) is 0.949.

Figure 2

821	**Figure 2: Morphologically Similar Cell Types Custer Together in AI Embedding Space**
822	**(A)** Embeddings UMAP plot of AI prediction of cell classes. Embeddings are extracted from images of
823	unlabeled breast cancer effusion samples from two different patients. Classifications of cell types are
824	based on the malignant effusion AI model. Distinct cell classes are shown, including breast carcinoma
825	(red/orange color family), CD45- (mesothelial + others) (gold color family), CD45+ (leukocytes +
826	others) (blue color family), and histiocytes (pink color family). Representative brightfield images from
827	each cluster are shown; for display only, these images have been normalized for illumination
828	intensity, and the pixels outside the cells have been blanked out.
829	**(B)** Dendrogram heatmap of embedding vectors with highest contribution to AI prediction of cell classes.
830	Red box highlights vectors (DCX19, DCX56, DCX45, DCX44 and DCX21) contributing to distinct
831	clustering of carcinoma cells from Patient A and B.
832	**(C)** Histogram plot by isolated known cell types, as quantified by computer vision morphometric
833	measures of cell size (radius). Isolated known cell types ("ground truth") are as follows: breast
834	carcinoma (EpCAM+/Claudin4+/CD45-), leukocyte (CD45+), mesothelial (CD45-/EpCAM-/Claudin4-
835	/Calretinin+), histiocyte (CD45+/CD68+), and biomarker-negative cell "other" cells (CD45-/EpCAM-
836	/Claudin4-).
837	
838	
839	
840	
841	
842	
843	
844	
845	
846	
847	
848	

Figure 3

Figure 3: Processed Cells are Viable and Compatible with Cytology Workflows

(A) Relative cell viability of samples before (pre) and after (post) instrument flow-through.

(B) PAP stain performed and assessed by a board-certified pathologist on two high-grade serous ovarian

cancer ascitic effusion samples show sorted cells remain intact and amenable to cytology staining.

Scale bar as indicated.

(C) Bulk RNA-Seq heatmap showing normalized gene expression counts of pre (green) and post (blue)

instrument flow-through effusion samples from patients diagnosed with non-cancerous disease

(benign; dark blue), breast (purple), and ovarian (grey) cancers.

(D) Principal component analysis (PCA) plot of pre (circle) and post (triangle) instrument flow-through

non-cancerous disease (dark blue), breast (purple), and ovarian (grey) cancer samples. Principal

component 1 (PC1) represents 67% contribution of variation, and principal component 2 (PC2)

represents 25% contribution of variation. Related to **Fig. 3C**.

(E) Correlation plot showing pre versus post instrument flow-through sample gene expression based on

bulk RNA-Seq analysis. Each dot represents the transcript count from pleural fluid derived from

patient diagnosed with ovarian cancer fluid (X- and Y-axis is transformed (regularized logarithm; rlog)

count from pre and post flow-through samples, respectively). R square value is indicated in the graph.

Related to **Fig. 3C**.

Figure 4

Sample #	Malignancy	AI carcinoma prediction	EpCAM+Claudin4+CD45-	Tumor Fraction (CNV)	
		Pre-sort			**Post-sort**
01	Fallopian Tube Carcinoma Stage IIIC	7%	2%	<3%	9%
02	Mullerian Adenocarcinoma	4%	1%	<3%	19%
03	Pancreatic Adenocarcinoma	4%	1%	<3%	11%
04	Lung Adenocarcinoma	7%	0%	8%	9%
05	Breast Carcinoma	25%	26%	26%	40%
06	Lung Adenocarcinoma (Suspicious)	4%	1%	<3%	14%
07	Ovarian Carcinoma	2%	1%	<3%	30%
08	Breast Carcinoma (Invasive lobular)	42%	6%	40%	82%
09	Ovarian/Fallopian Carcinoma (High-Grade Serous)	11%	16%	25%	42%
10	Fallopian Carcinoma	18%	17%	20%	45%
11	Ovarian Carcinoma (High-Grade Serous)	17%	5%	18%	31%
12	Breast Carcinoma	3%	0%	<3%	26%

Figure 4: Validation of AI-Driven Identification of Carcinoma Cells

(A) PAP stain performed and assessed by a board-certified pathologist on pre- and post-sort pancreatic

adenocarcinoma ascitic fluid at different magnifications. Red arrows indicate cells of interest/atypical

cells, yellow arrow indicate inflammatory cells. Scale bar as indicated.

(B) Table of sample cases indicating pre-sort proportion of carcinoma cell in percent predicted by the AI

model (AI carcinoma prediction), proportion of carcinoma cell (EpCAM+/Claudin4+/CD45-) measured

by flow cytometry, and tumor fractions calculated from CNV analysis. Corresponding post-sort tumor

fractions are shown for samples after AI carcinoma cell enrichment.

(C) Representative CNV plots showing amplifications/deletions in pre- and post-sort patient samples.

909 **Figure 5**

Sample #	Malignancy	Allele Freq. %VAF		Primary Mutation Profile
		Pre-sort	Post-sort	
01	Fallopian Tube Carcinoma Stage IIIC	TP53 c.560G>T		MS-stable; TMB: 3Muts/Mb; TP53 c.560 G>T
		0%	15%	
02	Mullerian Adenocarcinoma	TP53 c.578A>G		IHC: PAX8, pan-keratin, p53, overexpressed ER. PR/LoH 8.8%, MS stable, 3muts/Mb, TP53 c.578A>G, copy number amp: CCND1, C11orf30, EPHB4, FGF19, FGF3, FGF4
		2%	34%	
03	Pancreatic Adenocarcinoma	KRAS c.35C>T		IHC: CK7, IMP3. MSS/MSS, 4 muts/Mb, AKT2 amp, CDKN2A loss, DNMT3A Q656*, KRAS amp, KRAS c.35C>T (p.G12D), MTAP loss
		0%	47%	
04	Lung Adenocarcinoma	N/A		IHC: TTF1, Napsin A, p40, PDL1 (5%). CDC6-RET, MS-stable. Cytology confirmed <1%
		no mutation found	no mutation found	
06	Lung Adenocarcinoma (Suspicious)	EGFR c.2573T>G		IHC: Claudin4+,TTF1 scattered+, ALK-, TP53+, PDL1, TPS 70%. EGFR c.2573T>G (p.L858R), BRCA2 p.I2296fs, TP53 p.E287* stop gain LOF, APC p.E918* stop gain, copy number amp: CCND1, FGF3, FGF4, KRAS
		0%	21%	
09	Ovarian/Fallopian Carcinoma (High-Grade Serous)	TP53 p.L201fs		IHC: MSS, PDL1 CPS0, TP53 c.578A>G (p.H193R, missense LOF; liquid biopsy specimen), TP53 p.L201fs (frameshift LOF; peritoneal fluid)
		28%	76%	
10	Fallopian Carcinoma	TP53 c.653T>A		IHC: PDL1 (~50%), Claudin4+, PAX8 +. Foundation Medicine: TPS >50%, TP53 c.653T>A (p.V218E), LOH>16%, MUTYH splice site mutation, MSS, TMB low
		34%	87%	

911 **Figure 5: Targeted Mutation Analysis Reveals Increased Sensitivity for VAF Detection**

912 Table of sample cases showing pre- and post-sort somatic variant allele frequency (VAF), as detected by

913 targeted mutation sequencing (56G oncogene panel). Primary mutation profiles from tissue and/or IHC for

914 corresponding sample cases are shown, with orange highlighted text indicating matching detected

915 mutation in post-sort specimens determined by Ampliseq Cancer Hotspot panel V2.

ACKNOWLEDGEMENTS

AUTHOR CONTRIBUTIONS

Research design and study concept: CJ, AM, JR

Experiments and data analysis: AM, JN, ES, KS, MLP, LH, MW, JC, JM, JK, ZL, AW-T

Data science and machine learning: KS, JK, ZL, AW-T

Figures and data visualization: VL, KS, JK

Manuscript drafting: VL, CJ

Manuscript review and editing: all authors

COMPETING INTEREST

AM, CJ, VL, JN, ES, KS, MLP, LH, MW, JC, JM, JK, ZL, NL, SCB, AW-T, MM are current or

former employees at Deepcell, Inc.

DATA AVAILABILITY

All processed molecular data (bulk RNA-Seq, whole genome sequencing, targeted mutation

data) and corresponding scripts to generate figures are available on an S3 bucket

(s3://deepcell_public/ mavropoulos_2022) and https://github.com/deepcell/mavropoulos_2022,

respectively. Raw data can be provided upon reasonable request.

FUNDING

This work was supported by UCLA Jonsson Comprehensive Cancer Center Impact Grant (WY,

QYL, TK, YG, TDL, JR) and Deepcell, Inc., Menlo Park, CA, USA (AM, CJ, VL, JN, ES, KS,

MLP, LH, MW, JC, JM, JK, ZL, NL, SCB, AW-T, MM).