Ultrasensitive detection of circulating LINE-1 ORF1p as a specific multi-cancer biomarker

Martin S. Taylor*,[1,‡], Connie Wu*,[2,3,a ‡], Peter C. Fridy[4], Stephanie J. Zhang[2,3], Yasmeen Senussi[2,3], Justina C. Wolters[5], Wen-Chih Cheng[6], John Heaps[6], Bryant D. Miller[6], Kei Mori[2,7], Limor Cohen[2,3,8], Hua Jiang[4], Kelly R. Molloy[9], Brian T. Chait[9], Michael Goggins[10], Irun Bhan[11], Joseph W. Franses[11], Xiaoyu Yang[12], Mary-Ellen Taplin[12], Xinan Wang[13], David C. Christiani[11,13], Bruce E. Johnson[12], Matthew Meyerson[12], Ravindra Uppaluri[14], Ann Marie Egloff[14], Elyssa N. Denault[11], Laura M. Spring[11], Tian-Li Wang[10], Ie-Ming Shih[10], Euihye Jung[15], Kshitij S. Arora[1], Lawrence R. Zukerberg[2], Osman H. Yilmaz[16], Gary Chi[11], Bryanna L. Norden[11], Yuhui Song[11], Linda Nieman[11], Aparna R. Parikh[11,] Matthew Strickland[11], Ryan B. Corcoran[11], Tomas Mustelin[17], George Eng[1,18], Ömer H. Yilmaz[1,18], Ursula A. Matulonis[12], Steven J. Skates[19], Bo R. Rueda[20], Ronny Drapkin[15], Samuel J. Klempner[11], Vikram Deshpande[1], David T. Ting[11], Michael P. Rout[4], John LaCava[4,21], David R. Walt[2,3,‡], and Kathleen H. Burns[2,6,‡]

1. Department of Pathology, Massachusetts General Hospital and Harvard Medical School, Boston, MA, USA.
2. Department of Pathology, Brigham and Women's Hospital and Harvard Medical School, Boston, MA, USA.
3. Wyss Institute for Biologically Inspired Engineering at Harvard University, Boston, MA, USA.
4. Laboratory of Cellular and Structural Biology, The Rockefeller University, New York, NY, USA.
5. Department of Pediatrics, University of Groningen, University Medical Center Groningen, Groningen, The Netherlands.
6. Department of Pathology, Dana Farber Cancer Institute and Harvard Medical School, Boston, MA, USA.
7. Healthcare Optics Research Laboratory, Canon U.S.A., Inc., Cambridge, MA, USA
8. Department of Chemistry and Chemical Biology, Harvard University, Cambridge, MA, USA
9. Laboratory of Mass Spectrometry and Gaseous Ion Chemistry, The Rockefeller University, New York, New York, USA
10. Johns Hopkins University School of Medicine, Baltimore, MD, USA.
11. Mass General Cancer Center and Department of Medicine, Massachusetts General Hospital and Harvard Medical School, Boston, MA, USA.
12. Department of Medical Oncology, Dana Farber Cancer Institute and Harvard Medical School, Boston, MA, USA.
13. Department of Environmental Health, Harvard T.H. Chan School of Public Health, Harvard University, Boston, MA, USA
14. Department of Surgery, Brigham and Women's Hospital and Harvard Medical School, Boston, MA, USA.
15. University of Pennsylvania Perelman School of Medicine, Philadelphia, PA, USA.
16. Department of Pathology, Beth Israel Deaconess Medical Center and Harvard Medical School, Boston, MA, USA.
17. Division of Rheumatology, Department of Medicine, University of Washington, Seattle, WA, USA.
18. The David H. Koch Institute for Integrative Cancer Research at MIT, Department of Biology, Massachusetts Institute of Technology, Cambridge, Massachusetts, USA
19. MGH Biostatistics, Massachusetts General Hospital and Harvard Medical School, Boston, MA, USA.

20. Department of Obstetrics and Gynecology, Massachusetts General Hospital, and Harvard Medical School, Boston, MA, USA.
21. European Research Institute for the Biology of Ageing, University Medical Center Groningen, The Netherlands.

* Equal Contribution
‡ Correspondence : mstaylor@mgh.harvard.edu, conniewu@umich.edu, dwalt@bwh.harvard.edu, kathleenh_burns@dfci.harvard.edu
Present addresses: [a] University of Michigan Life Sciences Institute, Department of Biomedical Engineering, Ann Arbor, MI, USA.

Abstract

Improved biomarkers are needed for early cancer detection, risk stratification, treatment selection,

and monitoring treatment response. While proteins can be useful blood-based biomarkers, many

have limited sensitivity or specificity for these applications. Long INterspersed Element-1 (LINE-

1, L1) open reading frame 1 protein (ORF1p) is a transposable element protein overexpressed in

carcinomas and high-risk precursors during carcinogenesis with negligible detectable expression

in corresponding normal tissues, suggesting ORF1p could be a highly specific cancer biomarker.

To explore the potential of ORF1p as a blood-based biomarker, we engineered ultrasensitive

digital immunoassays that detect mid-attomolar (10^{-17} M) ORF1p concentrations in patient plasma

samples across multiple cancers with high specificity. Plasma ORF1p shows promise for early

detection of ovarian cancer, improves diagnostic performance in a multi-analyte panel, and

provides early therapeutic response monitoring in gastric and esophageal cancers. Together,

these observations nominate ORF1p as a multi-cancer biomarker with potential utility for disease

detection and monitoring.

Statement of Significance (50 word)

LINE-1 ORF1p transposon protein is pervasively expressed in many cancers and a highly specific

biomarker of multiple common, lethal carcinomas and their high-risk precursors in tissue and

blood. Ultrasensitive ORF1p assays from as little as 25 µL plasma are novel, rapid, cost-effective

tools in cancer detection and monitoring.

Introduction

There is significant clinical need for non-invasive methods to detect, risk stratify, and monitor

cancers over time. Many malignancies are diagnosed at late stages when disease is widespread,

contributing significantly to cancer morbidity and mortality[1]. In contrast, there is a likely window

in early-stage disease when patients are typically asymptomatic, in which treatments can be much

more effective. Biomarkers are also needed to assess likelihood of progression in patients with

precursor lesions, to provide prognostic information, and to predict and monitor responses or

resistance to treatment[2]. Considerable advances have been made towards detecting circulating

tumor DNA, circulating tumor cells, microRNAs, and extracellular vesicles as non-invasive cancer

biomarkers[3]. However, achieving high sensitivities and specificities, particularly in affordable,

scalable, clinical grade screening assays for early cancer detection, remains a major challenge.

The plasma proteome provides a rich reservoir of potential biomarkers[4], which may be used

individually or in combination for Multi-Cancer Early Detection (MCED) assays[5]. However, most

readily detectable proteins, including CA125 and HE4[6], FDA-cleared markers for the differential

diagnosis of pelvic masses, are not sufficiently sensitive at the required high specificity[7] for

cancer screening and/or are expressed in normal tissues and therefore lack the requisite

specificity.

We have previously shown that expression of long interspersed element-1 (L1, LINE-1)-encoded

open reading frame 1 protein (ORF1p) is a hallmark of many cancers[8], particularly p53-deficient

epithelial cancers. These encompass many of the most commonly occurring and lethal human

cancers, including esophageal, colorectal, lung, breast, prostate, ovarian, uterine, pancreatic, and

head and neck cancers. L1 is the only active protein-coding transposon in humans. We each

inherit, dispersed throughout our genomes, a complement of active L1 loci encoding two proteins:

ORF1p, the highly expressed RNA binding protein(8), and ORF2p, an endonuclease and reverse

transcriptase with limited expression(9) that generates L1 insertions in cancer genomes(10-13).

L1 expression is repressed in normal somatic tissues, resulting in either very low or undetectable

levels of L1 RNA and protein that appear to originate from epithelium(9,14). Epigenetic

dysregulation of L1 and L1 ORF1p overexpression begin early in carcinogenesis, and histologic

precursors of ovarian, esophageal, colorectal, and pancreatic cancers studied all express ORF1p

at varying levels(8,15). ORF1p is thus a promising highly specific cancer biomarker.

Although elevated expression of ORF1p is readily detected by immunostaining in tumor tissue,

ORF1p is found in plasma at low concentrations, well below detection limits of conventional

clinical laboratory methods. We therefore applied the much more sensitive Single Molecule Arrays

(Simoa), a digital bead-based ELISA technology, and in preliminary studies detected ORF1p in

plasma at femtomolar levels in subsets of patients with advanced breast (33%, n=6)(16) and

colorectal (90%, n=32)(17) cancers, respectively. Here, we assess the landscape of ORF1p

plasma levels across multiple cancers, iteratively develop highly sensitive assays for potential

applications in early or minimal residual disease detection, and provide evidence that plasma

ORF1p may be an early indicator of therapeutic response.

Results

Because our preliminary survey of plasma ORF1p levels by Simoa in patients with advanced

stage colorectal cancer (CRC) indicated detectable ORF1p levels in 90% of cases(18), higher

than the proportion of CRCs we previously reported to express ORF1p by immunohistochemistry

(50%, n=18)(8), we first sought to benchmark ORF1p in tissues. Using a re-optimized protocol,

we stained 211 CRCs [178 sequential cases included on a tissue microarray (TMA) as well as an

additional 33 with matched plasma] and found 91% of CRC cases were immunoreactive for

Figure 1. ORF1p expression is early and pervasive in carcinomas. **a**, ORF1p immunostaining in a cohort of 211 colorectal cancers. **b**, Representative BE case: lesional cells overexpress p53, the L1 RNA, and ORF1p. **c**, L1 RNA and ORF1p overexpression across a cohort of 72 consensus BE cases and 51 carcinomas. **d**, Schematic of single-molecule protein detection by Simoa; a second generation assay is shown. Antibody/nanobody-coated magnetic beads, present in excess relative to target, capture single target ORF1p molecules. Enzyme-labeled detection reagent (here, a homodimeric nanobody) is added, forming an "immunosandwich", beads are loaded into microwells that each can hold at most one bead, and ORF1p molecules are then digitally detected using a fluorogenic substrate by counting "on" wells. First generation Simoa instead uses Nb5-coated beads and Ab6 detector. **e**, First-generation ORF1p Simoa detects plasma ORF1p with high specificity across major carcinomas. Pie charts indicate percentage of samples with detectable levels; dashed red line, LOD. ******, this control patient was thought to be 'healthy' at the time blood was donated to the biobank but was later found to have prostate cancer and lymphoma.

ORF1p (**Fig. 1a**). This result is consistent with genetic studies demonstrating somatic L1

retrotransposition in most CRCs(19), including activity in precancerous lesions antedating *APC*

tumor suppressor loss(20-22). Similarly, genetic evidence shows esophageal adenocarcinoma

(EAC) has high L1 activity(12), and L1 insertions occur in the highly prevalent Barrett's esophagus

(BE) precursor early in carcinogenesis(23,24). We therefore assembled a cross-sectional cohort

of 72 BE cases with consensus diagnosis reached by three expert gastrointestinal pathologists.

L1 RNA and ORF1p expression were pervasive in dysplastic BE and present in 100% of 51

esophageal carcinomas (**Fig. 1b,c**); all five BE cases indefinite for dysplasia and positive for

ORF1p and/or L1 RNA developed high grade dysplasia on subsequent biopsies (not shown).

Overall, this picture is similar to high grade serous ovarian cancers (HGSOC), where ORF1p is

expressed in 90% of cases and 90% of fallopian tube precursor lesions (serous tubal

intraepithelial carcinomas, STICs)(8,15,25). Taken together, ORF1p tissue expression is highly

prevalent in gastrointestinal and gynecologic carcinomas and high-risk precursor lesions.

We next sought to extend our tissue findings and explore plasma ORF1p. We optimized our

previously reported ORF1p Simoa assay and assessed the landscape of ORF1p levels in

pretreatment plasma from patients with advanced cancers. This "first-generation" assay uses a

recombinant, single-domain camelid nanobody (Nb5) as the capture reagent and a monoclonal

antibody (Ab6) as the detector reagent and has a limit of detection of 0.056 pg/mL (~470 aM

trimeric ORF1p), corresponding to 1.9 fM in plasma after correcting for sample dilution (**Fig. 1d,**

Table S1). With this assay, we surveyed multiple cancer types and >400 'healthy' control

individuals, who were without known cancer at the time blood was donated to the biobank. Plasma

ORF1p appears to be a highly specific cancer biomarker, with undetectable levels in ~99% of

controls (ages 20-90, **Fig. 1e, S1**). Of the five control patients with detectable ORF1p, the one

with the highest ORF1p was later found to have advanced prostate cancer and a cutaneous T

cell lymphoma; limited clinical information is available for the other four positive 'healthy'

individuals. With a cutoff set at 98% specificity in healthy controls, the highest proportions of

ORF1p(+) cases were observed in colorectal (58%, n=101) and ovarian cancers (71%, n=145).

While most of these patients had advanced-stage disease, plasma ORF1p remained detectable

in several early-stage patients in the cohort, including in those with ovarian and lung cancers and

in 5/18 with intraductal papillary mucinous neoplasms in the pancreas (IPMN, **Fig. S2-S4**).

Notably, four of eight stage I ovarian cancers in the cohort were positive (**Fig. S2**), suggesting

that plasma ORF1p may be an indicator of early-stage disease. As L1 expression is also

dysregulated in autoimmune disease and autoantibodies against ORF1p are prevalent in patients

with systemic lupus erythematosus (SLE), we measured plasma ORF1p in 30 SLE patients and

observed no detectable levels (**Fig. S5**)(26). Detectable ORF1p was seen in 1 of 30 patients with

chronic liver disease; the one positive patient was subsequently diagnosed with hepatocellular

carcinoma (**Fig. S5**). Size exclusion chromatography analysis of patient plasma further showed

that the majority of ORF1p resides outside extracellular vesicles (**Fig. S6**). Together, these

findings support the hypothesis that tumor-derived ORF1p can be found in the peripheral blood

of cancer patients and may act as a cancer-specific biomarker.

Given the gap between proportions of ORF1p(+) cancers by tumor immunohistochemistry (~90%

for CRC and HGSOC) versus by blood testing (~60-70%), we evaluated the possibility of

increasing plasma assay sensitivity by decreasing the assay's lower limit of detection. To this end,

we developed a panel of ORF1p affinity reagents, including new recombinant rabbit monoclonal

antibodies (RabMAbs) and engineered camelid nanobodies raised against recombinant human

ORF1p. Because ORF1p is homotrimeric, we engineered multimeric nanobody reagents with the

goal of enhancing binding affinity via increased avidity. These parallel development efforts

ultimately yielded both improved nanobody and rabbit monoclonal antibody reagents with at least

low-picomolar equilibrium dissociation constants (K_D) (**Fig. S7-S12, Table S2-S4**). Iterative

screening of these reagents with Simoa using recombinant antigen and select patient plasma

Figure 2. Improved detection of ORF1p with second- and third-generation assays. **a**, 34H7::Nb5-5LL second-generation assay measurements across a multi-cancer cohort. **b**, Ovarian cancer patients with age- and gender-matched controls in first- and second-generation assays; patients are a subset of those in 2a; red dots: stage I disease, orange dots: stage II disease. **c**, Schematic of affinity reagents used. 34H7 and 62H2 are custom mAbs; Nb5-5LL and Nb5-9 are an engineered homodimeric and heterodimeric nanobodies, respectively. **d**, ROC curves with single marker ORF1p across all healthy and ovarian cancer patients (top, n=128-132 cancer, 447-455 healthy), and multivariate models for ovarian (bottom, n=51-53 cancer, 50 healthy). **e**, Targeted proteomics measurements of plasma ORF1p from a gastric cancer patient using two quantotypic peptides (LSFISEGEIK and NLEECIR) with internal standards. **f**, correlation between measured ORF1p by Simoa and targeted proteomics assays; r=0.97 (Simoa vs LSFISEGEII) and r=0.99 (Simoa vs NLEECIR, t test), p<0.0001 for both. **g**, Comparison of 2nd and 3rd generation Simoa assays (25 μL) in 25 mostly undetectable gastroesophageal (GE) cancer and healthy control patients. **h**, Schematic of MOSAIC assays. Captured single molecule "immunosandwiches" are formed analogously to Simoa assays. DNA-conjugated streptavidin enables rolling circle amplification to be carried out, generating a strong local fluorescent signal on the bead surface, and then "on" and "off" beads are quantified by flow cytometry. **i**, 37H7::Nb5-5LL MOSAIC and Simoa assays in 10 previously-undetectable GE cancer and healthy control patients.

samples yielded three best-performing capture::detection pairs, termed "second-generation,"

which use rabbit monoclonal antibodies 34H7 and 62H12 as capture reagents and either Ab6 or

homodimeric form of Nb5 (Nb5-5LL) as detector (**Fig. 2a-c, S13-S16**). Adding detergent further

improved performance by limiting bead aggregation and improving bead loading into microwells.

These second-generation assays comprised capture::detection pairs of 34H7::Nb5-5LL,

62H12::Nb5-5LL, and 62H12::Ab6, achieving detection limits of 0.016-0.029 pg/mL (130-240 aM

trimeric ORF1p), and the four different reagents have predominantly non-overlapping epitopes in

binning experiments (34H7 and 62H12 partially overlap, **Fig. 2a-c, Table S1, S5-S6**). Somewhat

unexpectedly, analytical sensitivity did not perfectly correspond to clinical sensitivity. While the

second-generation assays demonstrated less than an order-of-magnitude improvement in

analytical sensitivity over the first-generation assay, they showed considerable improvement in

circulating ORF1p detectability over background in buffer in re-measured samples across a large

cohort of healthy and cancer patients (**Fig 2a, S17**). This difference may be due to differing

accessibilities of circulating ORF1p epitopes or to different nonspecific binding patterns in plasma.

Undetectable or extremely low ORF1p levels in healthy individuals could readily be discriminated

from measured ORF1p levels in ovarian cancer patients, resulting in a strong discriminatory ability

with single-marker models (area under the receiver operating characteristic curve, AUCs of 0.93

to 0.948, sensitivity of 41% to 81% at 98% specificity, **Fig. 2d top panel, Table S7**). This large

cohort included pre-treatment plasma samples from ovarian cancer patients (mostly high-grade

serous ovarian carcinoma) with age-matched controls (n=51-53 women, **Fig 2b**); again, second-

generation assays showed higher sensitivities while maintaining high specificities, notably

achieving detection of five out of six Stage I/II patients at >98% specificity. Furthermore,

multivariate models combining ORF1p (34H7::Nb5-5LL assay) with ovarian cancer biomarkers

CA125 and HE4 yielded improved diagnostic performance over these existing markers (CA125

and HE4 alone, AUC = 0.94, 59% sensitivity at 98% specificity; ORF1p, CA125, and HE4, AUC

= 0.98, 91% sensitivity at 98% specificity; **Fig 2d bottom panel, S18; Table S8**). While it is not

clear whether the low ORF1p levels detected in several healthy individuals is due to nonspecific

binding, true background levels of ORF1p, or an unappreciated pre-malignant state, several

positive healthy controls were positive by only one of the three second-generation assays (n=4

positive by only 62H12::Nb5-5LL and n=75 positive by only 62H12:Ab6), suggesting nonspecific

binding in at least some of these cases and the potential to improve specificity by combining data

from multiple assays. Our results indicate that by developing improved affinity reagents, we

achieved improved clinical sensitivity in detecting circulating ORF1p in cancer patients, with 83%

sensitivity at >98% specificity towards early detection of ovarian cancer.

To further validate our results, we developed a targeted proteomics approach to measure ORF1p

following affinity capture, with two distinct peptides measured vs. internal isotopically labeled

control peptides (**Fig. 2e**). With this assay, we applied much larger volumes of plasma (3-6 ml,

120-240 fold more than the 25 µL used in Simoa assays) from a cohort of 10 patients, including

2 gastroesophageal (GE) cancer patients and one healthy control with very high ORF1p (230-

1230 pg/ml), two healthy controls with high ORF1p, (3-5 pg/ml), and 5 healthy controls with low

ORF1p (undetectable – 0.2 pg/ml). The results (**Fig. 2f, S19**) show strong correlation with Simoa,

providing further confidence in our results (r=0.97-0.99, p<0.0001).

Building on the improvements made through nanobody engineering in our second-generation

assays, we developed an expanded set of homodimeric, heterodimeric, and heterotrimeric anti-

ORF1p nanobodies and screened them in combination with 34H7 and 62H12 capture antibodies,

resulting in "third-generation" assays (**Figs. S9, S12, S20-21**). We noticed that reagents

containing Nb2 performed very well in SPR but poorly in Simoa detection, and we hypothesized

this was because Nb2 contains a lysine in the CDR, which would be biotinylated in the procedure,

reducing affinity. We therefore engineered the new reagents to be C-terminally biotinylated on

cysteine residues and varied linker sequence. Five of these assays, which utilize Nb2- and Nb9-

containing constructs, outperform our second-generation assays in a cohort of 25 GE cancer

patients with ORF1p measurements that were mostly undetectable previously, while maintaining

high specificity versus healthy individuals (**Fig. 2g, S21**).

To leverage more sensitive assays for ORF1p detection, we next tested ORF1p affinity reagents

from one of the second-generation Simoa assays on our recently developed Molecular On-bead

Signal Amplification for Individual Counting platform (MOSAIC, **Fig. 2h**). MOSAIC develops

localized on-bead signal from single captured molecules, in contrast to the microwell array format

in Simoa, and improves analytical sensitivity by an order of magnitude over Simoa via increasing

the number of beads counted(27). Furthermore, as the developed Simoa assays used only 25 µL

plasma, we hypothesized that using larger plasma volumes would enhance ORF1p detectability

by increasing the number of analyte molecules present. By using a 20-fold higher sample volume

(500 µL plasma) and the MOSAIC platform, we achieved ten-fold higher analytical sensitivity, with

a limit of detection of 0.002 pg/ml ORF1p (17 aM trimer, **Fig. S22**). Indeed, in a pilot cohort of

gastroesophageal cancer and healthy patients, ORF1p levels in nine of ten previously

undetectable cancer patients were readily discriminated from healthy individuals (**Fig. 2i**). Thus,

in addition to improved affinity reagents, using larger sample volumes and more analytically

sensitive technologies can further enhance both sensitivity and discrimination of circulating

ORF1p levels between healthy controls and patients with cancer.

To test whether ORF1p might be useful for monitoring therapeutic response, 19 patients with

gastroesophageal cancer were identified who had both detectable plasma ORF1p at diagnosis

as well as subsequent samples available collected during or after treatment. Primary tumors were

all adenocarcinoma and located in the esophagus (n=7), gastroesophageal junction (n=7) and

stomach (n=5). All patients received systemic therapy. A smaller fraction of patients also received

Figure 3. ORF1p is an early predictor of response in 19 gastroesophageal patients undergoing chemo/chemoradiotherapy. **a**, Plasma ORF1p as measured by all three second-generation Simoa assays before and during/post treatment; Responders and Non-Responders were characterized by post-therapy, pre-surgery imaging; p<0.0001, Fisher's exact test. Non-Responders also have higher pre-treatment ORF1p than Responders (p=0.02, t-test). **b**, Representative CT and PET-CT from patients in the cohort.

radiation and/or surgery (Supplement, **Table S9)**. Clinical response ('Responders' and 'Non-

Responders') was determined by review of re-staging CT and PET-CT imaging. Over an average

of 465 days (range 98-1098), 12 patients died, six were alive at last follow-up (all 'Responders'),

and one was lost to follow-up. All 6 patients with detectable ORF1p at follow-up sampling, as

defined by positivity over background in two of three assays, were also Non-Responders by

imaging (**Fig. 3a**, p<0.0001, Fisher's Exact test) and had reduced survival (p = 0.001 log-rank test

for overall survival). In contrast, in all 13 Responders, circulating ORF1p dropped to undetectable

levels post-treatment. Representative PET and PET-CT images are shown (**Fig. 3b**). Thus,

reduction in circulating ORF1p paralleled treatment response and survival, while persistent

circulating ORF1p corresponded to patients with refractory disease, indicating the predictive

potential of this marker.

Discussion

Taken together, our data reveal for the first time that circulating ORF1p is a multi-cancer protein

biomarker with potential utility across clinical paradigms, including early detection, risk

stratification, and treatment response. These assays are enabled by ultrasensitive single-

molecule detection technologies and high-quality affinity reagents, which are both required due

to the attomolar-to-femtomolar circulating levels of ORF1p in cancer patients. Iterative

improvements including optimized affinity reagents, buffer, and assay design yield highly sensitive

and specific assays. A 20-fold volume scale-up to 500 µL appears promising for improving

sensitivity without obviously compromising specificity, and this volume remains much smaller than

a typical 5-10 mL blood draw and could be scaled further without limiting clinical applicability. The

data strongly suggest that these assays are measuring *bona fide* tumor-derived circulating ORF1p

for the following reasons: (1) four developed assays with predominantly non-overlapping high

affinity reagents all measure similar levels across hundreds of samples; (2) levels appear specific

to cancer patients, whose tumors overexpress ORF1p; (3) they correlate strongly with

measurements made by targeted proteomics, and (4), plasma levels pre- and on/post treatment

correlated with therapeutic response. Nonetheless, the low levels of circulating ORF1p makes

orthogonal confirmation in larger cohorts by any other method challenging, as even the most

sensitive mass spectrometry assays have limits of detection orders of magnitude higher.

The results expand our understanding that L1 expression is early and pervasive across

carcinomas from multiple organs and high-risk precursor lesions, including dysplastic Barrett's

esophagus, which is challenging to diagnose and manage. Circulating ORF1p shows promise in

early detection applications such as in ovarian cancer and may be more useful as part of a multi-

analyte detection test combined with, for example, cfDNA methylation, longitudinal CA125 in

ovarian cancer, or CEA in colorectal cancer(3,5,28). We demonstrate that ORF1p is an early

indicator of chemotherapeutic response in gastric and esophageal cancers at timepoints where

other parameters are often ambiguous, opening possibilities for monitoring minimal residual

disease or relapse. Importantly, ORF1p appears to provide a level of specificity for cancers not

achieved by other protein biomarkers, likely due to the unique biology of the retrotransposon, with

repression of L1 in normal somatic tissue(9,13,14). ORF1p is therefore attractive as a putative

"binary" cancer biomarker, in which a positive signal is highly specific for disease, with diagnostic

utility both in tissue and plasma.

The assays are cost-effective (<$3 in consumables), rapid (<two hours), simple to perform,

scalable, and have clinical-grade coefficients of variation (<15%). Flow cytometers for MOSAIC

are common in clinical reference laboratories, and the assay could be modified for DNA-based

readout by qPCR or sequencing. Limitations of the current work include the relatively small

numbers of early-stage samples and a small and heterogeneous gastroesophageal therapeutic

cohort. Larger cohorts will be needed for further validation. Further optimizations to both assay

design and reagents will likely be possible, and larger cohorts are needed to further validate and

develop third generation Simoa assays and MOSAIC assays. Finally, it is unclear how ORF1p,

which is normally cytosolic, enters the blood and what clinicopathologic factors might affect these

levels. Future work will also be needed to understand whether there is a normal baseline level of

circulating ORF1p, as implied by the trace amounts seen when ORF1p was measured from much

larger volumes of plasma using targeted mass spectrometry, and what factors affect this level.

Methods
259 **Methods**
260 Provided in detail in *Supplementary Information.*
261

Acknowledgements
We thank Jeni Fairman for illustrations and Bert Vogelstein for plasma samples from colorectal
cancer patients. We are grateful to Phil Cole for resources for protein expression and purification
and helpful discussions and to Andrew Kruse and Edward Harvey for helpful discussions
regarding nanobodies. We thank Zuzana Tothova for helpful discussion and review of the
manuscript. This work was supported by the National Institutes of Health grants R01GM130680
(KHB), K08DK129824 (MST), F32EB029777 (CW), R01CA240924 (DTT), U01CA228963 (DTT),
P41 GM109824 (MPR, BTC), T32CA009216 (MST, GE), U01CA233364, U2CCA271871,
U01CA152990 (SJS), R01GM126170 (JL), P50CA228991 Ovarian SPORE (EJ, T-LW, I-MS,
RD); Break *Through* Cancer (KHB); Earlier.Org (KHB and DRW); Minnesota Ovarian Cancer
Alliance (KHB); DOD W81XWH-22-1-0852 (EJ, RD); Canary Foundation (RD); Gray Foundation
(EJ, RD); The Concord (MA) Detect Ovarian Cancer Early Fund (SJS), Good Ventures (Open
Philanthropy Project); Friends of Dana-Farber Cancer Institute; Dana-Farber Cancer Institute;
and the Dana-Farber/Harvard Cancer Center (DF/HCC); ACD-Biotechne (DTT, VD); Robert L.
Fine Cancer Research Foundation (DTT); Worldwide Cancer Research grant 19-0223
(JL); Robertson Therapeutic Development Fund (JL); Nile Albright Research Foundation
(BRR); Vincent Memorial Research Foundation (BRR); SU2C Gastric Cancer Interception
Research Team Grant (SU2C-AACR-DT-30-20, SJK, DTT, administered by the American
Association for Cancer Research, the Scientific Partner of SU2C).

Author Contributions
MST, CW, ÖHY, SJK, VD, DTT, JL, DRW, and KHB formulated the research plan and interpreted
experimental results with assistance from SJZ, LC, YS, JCW, WCC, JH, BDM, and HJ. CW, SZJ,
LC, and YS performed Simoa and MOSAIC experiments. WCC, JH, HJ, and BDM performed
biochemical experiments. GE performed mouse experiments and interpreted results. PCF, MST,
CW, HJ, KRM, BTC, MPR, and JL developed and engineered nanobody constructs. PCF
performed SPR affinity measurements. MST, BDM, JCW, and JL designed and performed mass
spectrometry experiments. MG, IB, JWF, XY, MET, XW, DC, BEJ, MM, RU, AME, END, LMS,
TLW, IMS, EJ, BV, GC, BLN, ARP, MS, UAM, BRR, RD, SJK, and DTT provided patient samples
and data and interpreted clinical results. SJS and KM carried out bioinformatic analysis. MST,
LRZ, OHY, and VD diagnosed biopsies, scored cases, and interpreted results. MST, CW, DRW,
and KHB wrote the manuscript. All authors edited and approved the manuscript.

Competing Interests
295 **Competing Interests**

MST has received consulting fees from ROME Therapeutics and Tessera Therapeutics that are not related to this work. MST and JL have equity in ROME therapeutics. DTT has received consulting fees from ROME Therapeutics, Tekla Capital, Ikena Oncology, Foundation Medicine, Inc., NanoString Technologies, and Pfizer that are not related to this work. DTT is a founder and has equity in ROME Therapeutics, PanTher Therapeutics and TellBio, Inc., which is not related to this work. DTT receives research support from ACD-Biotechne, PureTech Health LLC, Ribon Therapeutics, and Incyte, which was not used in this work. LMS declares the following relationships: Consultant/advisory board: Novartis, Puma, G1 therapeutics, Daiichi Pharma, Astra Zeneca; Institutional research support: Phillips, Merck, Genentech, Gilead, Eli Lilly. SJK declares Consulting/advisory: Eli Lilly, Merck, BMS, Novartis, Astellas, AstraZeneca, Daiichi-Sankyo, Novartis, Sanofi-Aventis, Natera, Exact Sciences, Mersana. Stock/Equity: Turning Point Therapeutics, Nuvalent. BRR serves on SAB for VincenTech and receives research support from Novartis Institutes for Biomedical Research that are not related to this work. DRW has a financial interest in Quanterix Corporation, a company that develops an ultra-sensitive digital immunoassay platform. He is an inventor of the Simoa technology, a founder of the company and also serves on its Board of Directors. KHB declares relationships with Alamar Biosciences, Genscript, Oncolinea/PrimeFour Therapeutics, ROME Therapeutics, Scaffold Therapeutics, Tessera Therapeutics, and Transposon Therapeutics. MST and KHB receive royalties from sales of ORF1p antibodies and MST, CW, PCF, KRM, BTC, MPR, JL, DRW, and KHB are inventors on a patent related to this work. MST, LMS, SJK, BRR, and DTT's interests were reviewed and are managed by Massachusetts General Hospital and Mass General Brigham in accordance with their conflict-of-interest policies. Dr. Walt's interests were reviewed and are managed by Mass General Brigham and Harvard University in accordance with their conflict-of-interest policies. KHB's interests are managed by Dana-Farber Cancer Institute.

References

1. Siegel RL, Miller KD, Fuchs HE, Jemal A. Cancer statistics, 2022. CA Cancer J Clin **2022**;72(1):7-33 doi 10.3322/caac.21708.

2. Sawyers CL. The cancer biomarker problem. Nature **2008**;452(7187):548-52 doi 10.1038/nature06913.

3. Crosby D, Bhatia S, Brindle KM, Coussens LM, Dive C, Emberton M*, et al.* Early detection of cancer. Science **2022**;375(6586):eaay9040 doi 10.1126/science.aay9040.

4. Ignjatovic V, Geyer PE, Palaniappan KK, Chaaban JE, Omenn GS, Baker MS*, et al.* Mass Spectrometry-Based Plasma Proteomics: Considerations from Sample Collection to Achieving Translational Data. J Proteome Res **2019**;18(12):4085-97 doi 10.1021/acs.jproteome.9b00503.

5. Jamshidi A, Liu MC, Klein EA, Venn O, Hubbell E, Beausang JF*, et al.* Evaluation of cell-free DNA approaches for multi-cancer early detection. Cancer Cell **2022**;40(12):1537-49 e12 doi 10.1016/j.ccell.2022.10.022.

6. Bast RC, Jr., Lu Z, Han CY, Lu KH, Anderson KS, Drescher CW*, et al.* Biomarkers and Strategies for Early Detection of Ovarian Cancer. Cancer Epidemiol Biomarkers Prev **2020**;29(12):2504-12 doi 10.1158/1055-9965.EPI-20-1057.

7. Menon U, Gentry-Maharaj A, Burnell M, Singh N, Ryan A, Karpinskyj C*, et al.* Ovarian cancer population screening and mortality after long-term follow-up in the UK Collaborative Trial of Ovarian Cancer Screening (UKCTOCS): a randomised controlled trial. Lancet **2021**;397(10290):2182-93 doi 10.1016/S0140-6736(21)00731-5.

8. Rodic N, Sharma R, Sharma R, Zampella J, Dai L, Taylor MS*, et al.* Long interspersed
element-1 protein expression is a hallmark of many human cancers. Am J Pathol
2014;184(5):1280-6 doi 10.1016/j.ajpath.2014.01.007.
9. Ardeljan D, Wang X, Oghbaie M, Taylor MS, Husband D, Deshpande V*, et al.* LINE-1
ORF2p expression is nearly imperceptible in human cancers. Mob DNA **2020**;11:1 doi
10.1186/s13100-019-0191-2.
10. Lee E, Iskow R, Yang L, Gokcumen O, Haseley P, Luquette LJ, 3rd*, et al.* Landscape of
somatic retrotransposition in human cancers. Science **2012**;337(6097):967-71 doi
10.1126/science.1222077.
11. Helman E, Lawrence MS, Stewart C, Sougnez C, Getz G, Meyerson M. Somatic
retrotransposition in human cancer revealed by whole-genome and exome sequencing.
Genome Res **2014**;24(7):1053-63 doi 10.1101/gr.163659.113.
12. Rodriguez-Martin B, Alvarez EG, Baez-Ortega A, Zamora J, Supek F, Demeulemeester
J*, et al.* Pan-cancer analysis of whole genomes identifies driver rearrangements
promoted by LINE-1 retrotransposition. Nat Genet **2020**;52(3):306-19 doi
10.1038/s41588-019-0562-0.
13. Burns KH. Transposable elements in cancer. Nature Reviews Cancer **2017**;17(7):415-24
doi 10.1038/nrc.2017.35.
14. McKerrow W, Doudican N, Frazzette N, Evans SA, Rocha A, Kagermazova L*, et al.*
LINE-1 Retrotransposon expression in cancerous, epithelial and neuronal cells revealed
by 5' single-cell RNA-Seq. bioRxiv **2022**:2021.01.19.427347 doi
10.1101/2021.01.19.427347.
15. Pisanic TR, 2nd, Asaka S, Lin SF, Yen TT, Sun H, Bahadirli-Talbott A*, et al.* Long
Interspersed Nuclear Element 1 Retrotransposons Become Deregulated during the
Development of Ovarian Cancer Precursor Lesions. Am J Pathol **2019**;189(3):513-20 doi
10.1016/j.ajpath.2018.11.005.
16. Cohen L, Cui N, Cai Y, Garden PM, Li X, Weitz DA*, et al.* Single Molecule Protein
Detection with Attomolar Sensitivity Using Droplet Digital Enzyme-Linked
Immunosorbent Assay. ACS Nano **2020**;14(8):9491-501 doi 10.1021/acsnano.0c02378.
17. Rajurkar M, Parikh AR, Solovyov A, You E, Kulkarni AS, Chu C*, et al.* Reverse
Transcriptase Inhibition Disrupts Repeat Element Life Cycle in Colorectal Cancer.
Cancer Discov **2022** doi 10.1158/2159-8290.CD-21-1117.
18. Rajurkar M, Parikh AR, Solovyov A, You E, Kulkarni AS, Chu C*, et al.* Reverse
Transcriptase Inhibition Disrupts Repeat Element Life Cycle in Colorectal Cancer.
Cancer Discov **2022**;12(6):1462-81 doi 10.1158/2159-8290.CD-21-1117.
19. Solyom S, Ewing AD, Rahrmann EP, Doucet T, Nelson HH, Burns MB*, et al.* Extensive
somatic L1 retrotransposition in colorectal tumors. Genome Res **2012**;22(12):2328-38
doi 10.1101/gr.145235.112.
20. Ewing AD, Gacita A, Wood LD, Ma F, Xing D, Kim MS*, et al.* Widespread somatic L1
retrotransposition occurs early during gastrointestinal cancer evolution. Genome Res
2015;25(10):1536-45 doi 10.1101/gr.196238.115.
21. Scott EC, Gardner EJ, Masood A, Chuang NT, Vertino PM, Devine SE. A hot L1
retrotransposon evades somatic repression and initiates human colorectal cancer.
Genome Res **2016**;26(6):745-55 doi 10.1101/gr.201814.115.
22. Cajuso T, Sulo P, Tanskanen T, Katainen R, Taira A, Hanninen UA*, et al.*
Retrotransposon insertions can initiate colorectal cancer and are associated with poor
survival. Nat Commun **2019**;10(1):4022 doi 10.1038/s41467-019-11770-0.
23. Doucet-O'Hare TT, Rodic N, Sharma R, Darbari I, Abril G, Choi JA*, et al.* LINE-1
expression and retrotransposition in Barrett's esophagus and esophageal carcinoma.
Proc Natl Acad Sci U S A **2015**;112(35):E4894-900 doi 10.1073/pnas.1502474112.

24.   Katz-Summercorn AC, Jammula S, Frangou A, Peneva I, O'Donovan M, Tripathi M*, et*
393   *al.* Multi-omic cross-sectional cohort study of pre-malignant Barrett's esophagus reveals
394   early structural variation and retrotransposon activity. Nat Commun **2022**;13(1):1407 doi
395   10.1038/s41467-022-28237-4.
396   25.   Zhouchunyang X, Cochrane DR, Tessier-Coutier B, Leung S, Karnezis AN, Cheng AS*,*
397   *et al.* Expression of L1 retrotransposon open reading frame protein 1 (L1ORF1p) in
398   gynecologic cancers. Hum Pathol **2019** doi 10.1016/j.humpath.2019.06.001.
399   26.   Carter V, LaCava J, Taylor MS, Liang SY, Mustelin C, Ukadike KC*, et al.* High
400   Prevalence and Disease Correlation of Autoantibodies Against p40 Encoded by Long
401   Interspersed Nuclear Elements in Systemic Lupus Erythematosus. Arthritis Rheumatol
402   **2020**;72(1):89-99 doi 10.1002/art.41054.
403   27.   Wu C, Dougan TJ, Walt DR. High-Throughput, High-Multiplex Digital Protein Detection
404   with Attomolar Sensitivity. ACS Nano **2022** doi 10.1021/acsnano.1c08675.
405   28.   Jacobs IJ, Menon U, Ryan A, Gentry-Maharaj A, Burnell M, Kalsi JK*, et al.* Ovarian
406   cancer screening and mortality in the UK Collaborative Trial of Ovarian Cancer
407   Screening (UKCTOCS): a randomised controlled trial. Lancet **2016**;387(10022):945-56
408   doi 10.1016/S0140-6736(15)01224-6.

Figure Legends

Figure 1. ORF1p expression is early and pervasive in carcinomas. **a**, ORF1p immunostaining in

a cohort of 211 colorectal cancers. **b**, Representative BE case: lesional cells overexpress p53,

the L1 RNA, and ORF1p. **c**, L1 RNA and ORF1p overexpression across a cohort of 72 consensus

BE cases and 51 carcinomas. **d**, Schematic of single-molecule protein detection by Simoa; a

second generation assay is shown. Antibody/nanobody-coated magnetic beads, present in

excess relative to target, capture single target ORF1p molecules. Enzyme-labeled detection

reagent (here, a homodimeric nanobody) is added, forming an "immunosandwich", beads are

loaded into microwells that each can hold at most one bead, and ORF1p molecules are then

digitally detected using a fluorogenic substrate by counting "on" wells. First generation Simoa

instead uses Nb5-coated beads and Ab6 detector. **e**, First-generation ORF1p Simoa detects

plasma ORF1p with high specificity across major carcinomas. Pie charts indicate percentage of

samples with detectable levels; dashed red line, LOD. **, this control patient was thought to be

'healthy' at the time blood was donated to the biobank but was later found to have prostate cancer

and lymphoma.

Figure 2. Improved detection of ORF1p with second- and third-generation assays. **a**, 34H7::Nb5-5LL second-generation assay measurements across a multi-cancer cohort. **b**, Ovarian cancer patients with age- and gender-matched controls in first- and second-generation assays; patients are a subset of those in 2a; red dots: stage I disease, orange dots: stage II disease. **c**, Schematic of affinity reagents used. 34H7 and 62H2 are custom mAbs; Nb5-5LL and Nb5-9 are an engineered homodimeric and heterodimeric nanobodies, respectively. **d**, ROC curves with single marker ORF1p across all healthy and ovarian cancer patients (top, n=128-132 cancer, 447-455 healthy), and multivariate models for ovarian (bottom, n=51-53 cancer, 50 healthy). **e**, Targeted proteomics measurements of plasma ORF1p from a gastric cancer patient using two quantotypic peptides (LSFISEGEIK and NLEECIR) with internal standards. **f**, Correlation between measured ORF1p by Simoa and targeted proteomics assays; r=0.97 (Simoa vs LSFISEGEII) and r=0.99 (Simoa vs NLEECIR, t test), p<0.0001 for both. **g**, Comparison of 2^{nd} and 3^{rd} generation Simoa assays (25 µL) in 25 mostly undetectable gastroesophageal (GE) cancer and healthy control patients. **h**, Schematic of MOSAIC assays. Captured single molecule "immunosandwiches" are formed analogously to Simoa assays. DNA-conjugated streptavidin enables rolling circle amplification to be carried out, generating a strong local fluorescent signal on the bead surface, and then "on" and "off" beads are quantified by flow cytometry. **i**, 37H7::Nb5-5LL MOSAIC and Simoa assays in 10 previously-undetectable GE cancer and healthy control patients.

Figure 3. ORF1p is an early predictor of response in 19 gastroesophageal patients undergoing chemo/chemoradiotherapy. **a**, Plasma ORF1p as measured by all three second-generation Simoa assays before and during/post treatment; Responders and Non-Responders were characterized by post-therapy, pre-surgery imaging; p<0.0001, Fisher's exact test. Non-Responders also have higher pre-treatment ORF1p than Responders (p=0.02, t-test). **b**, Representative CT and PET-CT from patients in the cohort.