Instant polarized light microscopy pi (IPOLπ) for quantitative imaging of collagen architecture and dynamics in ocular tissues

Po-Yi Lee[1,2], Hannah Schilpp[2], Nathan Naylor[2],
Simon C. Watkins,[3] Bin Yang[4], Ian A Sigal[1,2*]

[1] Department of Bioengineering, Swanson School of Engineering,
[2] Department of Ophthalmology, School of Medicine,
[3] Department of Cell Biology, School of Medicine,
University of Pittsburgh, Pittsburgh, PA
[4] Department of Engineering, Rangos School of Health Sciences,
Duquesne University, Pittsburgh, PA

Short Title: IPOLπ for fast quantitative imaging of eye collagen

* Correspondence:
Ian A. Sigal, Ph.D.
Laboratory of Ocular Biomechanics
Department of Ophthalmology, University of Pittsburgh School of Medicine
203 Lothrop Street, Eye and Ear Institute, Room 930, Pittsburgh, PA 15213
Phone: (412) 864-2220; Fax: (412) 647-5880
Email: ian@OcularBiomechanics.com
www.OcularBiomechanics.com

Key Words: polarized light microscopy; collagen; biomechanics; optic nerve head

Disclosures: Nothing to disclose.

Funding: Supported in part by National Institutes of Health R01-EY023966, R01-EY031708, 1S10RR028478, P30-EY008098 and T32-EY017271 (Bethesda, MD), the Eye and Ear Foundation (Pittsburgh, PA), and Research to Prevent Blindness (unrestricted grant to UPMC ophthalmology, and Stein innovation award to Sigal, IA).

Highlights

- We introduce IPOLπ, addressing IPOL limitations for characterizing eye collagen.

- IPOLπ orientation-encoded color cycle is 180º (π radians) instead of 90º in IPOL.

- IPOLπ requires a lower exposure time than IPOL, allowing faster imaging speed.

- IPOLπ visualizes non-birefringent tissues and backgrounds from brightness.

- IPOLπ is cheaper and less sensitive to imperfectly collimated light than IPOL.

Abstract

Collagen architecture determines the biomechanical environment in the eye, and thus
characterizing collagen fiber organization and biomechanics is essential to fully understand eye
physiology and pathology. We recently introduced instant polarized light microscopy (IPOL) that
encodes optically information about fiber orientation and retardance through a color snapshot.
Although IPOL allows imaging collagen at the full acquisition speed of the camera, with excellent
spatial and angular resolutions, a limitation is that the orientation-encoding color is cyclic every
90 degrees ($\pi/2$ radians). In consequence, two orthogonal fibers have the same color and
therefore the same orientation when quantified by color-angle mapping. In this study, we
demonstrate IPOLπ, a new variation of IPOL, in which the orientation-encoding color is cyclic
every 180 degrees (π radians). Herein we present the fundamentals of IPOLπ, including a
framework based on a Mueller-matrix formalism to characterize how fiber orientation and
retardance determine the color. The improved quantitative capability of IPOLπ enables further
study of essential biomechanical properties of collagen in ocular tissues, such as fiber anisotropy
and crimp. We present a series of experimental calibrations and quantitative procedures to
visualize and quantify ocular collagen orientation and microstructure in the optic nerve head, a
region in the back of the eye. There are four important strengths of IPOLπ compared to IPOL.
First, IPOLπ can distinguish the orientations of orthogonal collagen fibers via colors, whereas
IPOL cannot. Second, IPOLπ requires a lower exposure time than IPOL, thus allowing faster
imaging speed. Third, IPOLπ allows visualizing non-birefringent tissues and backgrounds from
tissue absorption, whereas both appear dark in IPOL images. Fourth, IPOLπ is cheaper and less
sensitive to imperfectly collimated light than IPOL. Altogether, the high spatial, angular, and
temporal resolutions of IPOLπ enable a deeper insight into ocular biomechanics and eye
physiology and pathology.

1. Introduction

Collagen is a primary load-bearing component in ocular tissues. Its architecture determines the ocular biomechanical environment and susceptibility to several vision threatening conditions. (Coudrillier et al., 2012; Ethier et al., 2004) Therefore, characterizing collagen fiber organization and biomechanics is important to fully understand eye physiology and preventing vision loss.

Polarized light microscopy (PLM) has been used to study collagen architecture in soft tissues for several decades. (Canham et al., 1991; Diamant et al., 1972; Keefe et al., 1997; Tower et al., 2002) A major advantage of PLM over other more conventional histology is that it does not require labels or stains, avoiding potential artifacts and simplifying preparation. (Koike-Tani et al., 2015)

Over the past few years widefield PLM has continued to be refined and recently demonstrated robust for characterizing collagen architecture of posterior pole ocular tissues. (Brazile et al., 2018; Filas et al., 2014; Gogola et al., 2018a; Gogola et al., 2018b; Jan et al., 2018; Jan et al., 2017a; Jan and Sigal, 2018; Jan et al., 2015b; Jan et al., 2017c; Yang et al., 2018a; Yang et al., 2018b) Other imaging techniques leveraging polarized light have also been used extensively to study ocular collagen, including polarization sensitive second harmonic generation (PS-SHG) (Cisek et al., 2021; Gusachenko et al., 2012; Mansfield et al., 2008) and polarization sensitive optic coherent tomography (PS-OCT) (Baumann et al., 2014; Willemse et al., 2020; Yamanari et al., 2014). Nevertheless, widefield PLM has several strengths over raster-scanning techniques that continue to make it highly useful for soft tissues. PLM produces images with high, micrometer-scale, resolution over wide regions at high imaging speeds. PLM is substantially simpler and therefore cheaper. (Higgins, 2010; Whittaker and Przyklenk, 2009) The high angular resolution and sub-pixel information of PLM allows accurate measurement of small fiber undulations, or crimp. (Jan et al., 2015a; Kalwani et al., 2013) This is of particular interest for the study of the relationship between tissue architecture and biomechanics as it allows quantifying fiber crimp without the need to discern or trace individual fibers. Fiber tracing is very demanding on image resolution and analysis time, often leading to a reduced number of measurements. The above have made PLM a preferred technique for measuring collagen crimp.

Conventional quantitative PLM, however, requires multiple images at different polarization states to compute the collagen structure and orientation. (Shribak and Oldenbourg, 2003) Multi-image acquisition not only limits the imaging speed to evaluate the static architecture or quasi-static behavior in ocular tissues but may also introduce quantitative errors from post-processing among images, such as image registration.

Instant polarized light microscopy (IPOL) is a recently introduced technique that optically
encodes information about fiber orientation and retardance through a color snapshot. (Lee et al.,
2022; Yang et al., 2021) IPOL allows quantitative imaging of collagen at the full acquisition speed
of the camera, with excellent spatial and angular resolutions. IPOL, however, has the limitation
that the orientation-encoded colors are cyclic every 90 degrees ($\pi/2$ radians). In consequence,
two orthogonal fibers have the same color and therefore the same orientation when quantified by
color-angle mapping. Techniques to distinguish the orientations of orthogonal fibers have been
suggested, (Keikhosravi et al., 2021) but the detailed methodology to quantify orientation and
retardance have not been described.

Our goal in this work was to demonstrate IPOLπ, a new variation of IPOL, in which the
orientation-encoding color is cyclic every 180 degrees (π radians). We describe how to use IPOLπ
to conduct quantitative analysis on static and dynamics of ocular tissues, with high spatial and
temporal resolutions. We present the fundamentals of IPOLπ, including a framework based on a
Mueller-matrix formalism to characterize how fiber orientation and retardance determine the color.
The improved quantitative capability of IPOLπ enables further study of essential biomechanical
properties of collagen in ocular tissues, such as fiber anisotropy and crimp. We present a series
of experimental calibrations and quantitative procedures to visualize and quantify ocular collagen
orientation and microstructure in the optic nerve head, a region in the back of the eye.

2. Methods

This section is organized into five parts. In Section 2.1, we describe the configuration of IPOLπ imaging system. In Section 2.2, we introduce a framework based on a Muller-matrix formalism to characterize how the color changes in fiber orientation and retardance. This simulation helps builds a foundation of quantitative analysis from the proposed imaging. In Section 2.3, we demonstrate how to experimentally calibrate the relationship between color and fiber orientation and the relationship between color and retardance. In Section 2.4, we introduce how to use the interpolating functions obtained from the experimental calibration in Section 2.3 to quantify images acquired by IPOLπ and then to visualize the quantitative results. Finally, in Section 2.5, we demonstrate the applications of IPOLπ for visualizing and quantifying static collagen architecture and dynamic collagen deformation under uniaxial stretch testing of optic nerve head tissues.

2.1 IPOLπ imaging system

The optic design of IPOLπ was to retrofit a commercial inverted microscope (Olympus IX83, Olympus, Tokyo, Japan) with a white light source, circular polarizer, polarization decoder, and color camera (DP74, Olympus, Tokyo, Japan) (**Figure 1**). In the absence of a birefringent sample, the spectrum of the with light was unchanged and the image background appeared grey. With a birefringent sample, such as collagen, the spectrum of the white light was changed and thus the light appeared colorful. The frame rate of IPOLπ was limited only by that of the color camera, which in our setup was 60 frames per second.

2.2 IPOLπ simulation

The polarization simulation was used to characterize the relationship between the output RGB color from IPOLπ and the material properties of collagen. We applied a Mueller-matrix formalism to simulate how the polarization states of the broadband white-light spectrum were altered through optical elements and a sample. In the Mueller-matrix formalism, briefly, a series of 4×4 transfer matrices, i.e., Mueller matrices, was introduced to operate on an incident 1×4 Stokes vector to obtain the corresponding transmitted Stokes vector. (Collett, 2005) The Stokes vector S_{out} of the output light from IPOLπ can be described as

$$S_{out}(\varphi, \delta, \lambda) = M_{P(0°)} \cdot M_{ROT}(\lambda) \cdot M_{\text{sample}}(\varphi, \delta) \cdot M_{QWP}(\lambda) \cdot M_{P(45°)} S_{in}(\lambda) \qquad (1)$$

where M were the Mueller matrices and S were the Stokes vectors. The spectrum of the incident light S_{in} was referred to our LED white light source (IX3-LJLEDC, Olympus, Tokyo, Japan). A linear polarizer M_P orientated at 45 degrees and a quarter waveplate was equivalent to a circular

polarizer. We simulated and compared the effects on the output light of four quarter waveplates:
ideal (i.e., wavelength independent), achromatic (AQWP05M-580, Thorlabs, NJ, USA), zero-
order (WPQ05ME-546, Thorlabs, NJ, USA), and multi-order (WPMQ05M-532, Thorlabs, NJ,
USA). A phase retarder, M_{sample}, was used to represent the birefringent sample, such as collagen
fibers, where the slow axis orientation ϕ varying from 0 to 180 degrees, and the retardance δ
varying from 0 to π radians. A z-cut quartz was model as a polarization rotator, M_{ROT}, where the
rotated angles were wavelength dependent. The polarization rotator allowed diverged the
polarization directions of the visible spectrum (400-700 nm) within 180 degrees. The last one was
a linear polarizer, M_P, orientated at 0 degrees, also called an analyzer. For a given fiber orientation
and retardance, S_{out} was calculated for the wavelengths between 400nm and 700 nm, where the
first parameter of the 1×4 Stroke vector represented intensity. Each wavelength had a
corresponding RGB (Red, Green, Blue) value, an additive color model. The RGB color was
subsequently obtained by spectrally mixing the wavelength-dependent intensity of the output. The
simulated color was converted from the RGB color space to the HSV (Hue, Saturation, Value)
color space (Schanda, 2007). Note that the meanings of the terms "value" and "brightness" were
the same in this study.

2.3 RGB color calibration

With the understanding of the relationship between color and material properties of collagen
from the Section 2.2, we performed an experimental calibration and developed an algorithm to
build system-specific color-angle and color-energy conversion maps. A chicken Achilles tendon
was dissected and fixed with 10% formalin for 24 hours while under load. (Yang et al., 2018b;
Yang et al., 2021) Following fixation, the tendon was cryo-sectioned longitudinally into 20-µm-
thick sections. The tendon section fixed while under load was considered a uniform collagen
organization with consistent fiber orientation. IPOLπ images were acquired with the chicken
tendon section at several controlled angles relative to the longitudinal fiber direction, from 0 to
180 degrees, every 2 degrees.

Post-processing was necessary to extract the averaged color from the section and actual
rotation angles due to regional variations in fiber content and orientation and the limited rotational
accuracy. The individual images were registered and the rotation angles were obtained from the
transformation matrices using Fiji software. (Schindelin et al., 2012) Then, with a region of interest
manually placed on the color-uniform area in the stack, the RGB colors were extracted and
averaged from the region of interest from the corresponding frames. The calibrated background
was the mean of the colors obtained from all frames. The distance between color on experimental

calibration and calibrated background as a unit of "energy", which was related to retardance. We
used the experimental calibration to generate pseudo calibrations with corresponding orientations
and energies (**Figure 2**). The energy in low-brightness pseudo calibrations was weighted by the
brightness to avoid image noise. We interpolated a combination of pseudo and experimental
calibrations to build interpolating functions for orientation and energy inquiry. (Amidror, 2002)

2.4 Quantification and visualization

Fiber orientation and energy maps for all images were obtained by searching RGB values for all
pixels over the color-angle and color-energy interpolating functions, respectively, obtained from
Section 2.3. For visualization, the post-processed image was made from the orientation map of
the collagen fibers rendered by a HSV colormap and the brightness was weighted by the energy
map.

2.5 Imaging collagen architecture and deformation

Sample preparation for static imaging. A normal one-year-old sheep eye was obtained from a
local abattoir within four hours after death and formalin-fixed for 24 hours at 22 mmHg of
intraocular pressure. (Jan et al., 2015a) The muscles, fat, and episcleral tissues were carefully
removed. The optic nerve head region was isolated using an 11-mm-diameter trephine and
embedded in optimum cutting temperature compound (Tissue-Plus; Fisher Healthcare, TX, USA).
For static imaging,

Sample preparation for uniaxial stretch testing. An unfixed sheep eye section was prepared
as described in detail previously (Jan et al., 2022). Briefly, a normal one-year-old sheep eye was
obtained from a local abattoir within four hours after death. The muscles, fat, and episcleral tissues
were carefully removed. The optic nerve head region was isolated using an 11-mm-diameter
trephine and embedded in optimum cutting temperature compound (Tissue-Plus; Fisher
Healthcare, TX, USA). Samples were then snap frozen in liquid nitrogen-cooled isopentane and
sectioned coronally at a thickness of 16 μm. OCT was washed with multiple PBS baths. To
prevent curling or tears at the clamp points, a tissue section was sandwiched between two pieces
of silicone sheet (Medical Grade, 0.005"; BioPlexus, AZ, USA). The sheets also allowed using
PBS to maintain tissue hydration without lensing.

Static imaging for optic nerve head collagen architecture. Tissue samples were imaged using
IPOLπ and IPOL with a 10x strain-free objective (numerical aperture [NA] = 0.3). Due to the limited
field of view of the objective, mosaicking was used to image the whole section. The mosaics were
obtained with 20% overlap and stitched using Fiji. (Schindelin et al., 2012)

Dynamic imaging for sclera while under uniaxial stretching. Each section was mounted to a

custom uniaxial stretching device and then stretched on both sides equally and dynamically. The

section was imaged using IPOLπ with a 4x strain-free objective (NA = 0.13) to visualize a sclera

region. The frame rate was 60 frames per second. Angles within a scleral region were extracted

from the initial and three stretching states. Each two states had a 240-frame gap.

3. Results

3.1 IPOLπ simulation

A simulated RGB color map of IPOLπ presented the interaction with collagen fiber orientation angles from 0 to 180 degrees and retardance from 0 to π radian (**Figure 3a**). Although the colors looked like a continuous rainbow with orientations in the colormap, the relationship of retardance to hue was not a constant and the relationship to brightness was not a strictly increasing function (**Figure 3b**). Therefore, neither hue nor brightness was a suitable parameter for quantifying collagen fiber orientation and retardance, particularly in low-retardance tissues. In the RGB color space, the simulated result appeared as concentric circles, where the radius increased with an increase in retardance (**Figure 3c**). In addition, a non-birefringent sample was located at the central point of the circles. We found that the normalized radius followed the sine function of the retardance, which was equivalent with what we have called normalized "energy" in our previous studies. (Jan et al., 2015a; Jan et al., 2017b) (**Figure 3d**). The explanation of the relationship between the normalized radius and the energy is addressed in more detail in the Discussion. Types of quarter waveplates to build IPOLπ would remarkably impact how the simulated curves within 180-degree orientation in the RGB color space (**Figure 3e**).

3.2 RGB color calibration

Images of the chicken tendon section were acquired every 2 degrees from 0 to 180 degrees. The color extracted from the images is presented in the RGB color space with a representative subset of 8 images (**Figure 4a**). The color of the sample acquired by IPOLπ depended on its orientation. All colors formed an enclosing ring in the RGB color space, with the background color located at the center of the ring. A combination of pseudo and experimental calibrations was generated with the corresponding orientations (**Figure 4b**) and energies (**Figure 4c**). These scatter points were used to build two interpolating functions that produced interpolated orientations and energies, respectively, at inquiry RGBs. Note that these interpolating functions are specific to our system. Other systems will potentially vary and therefore require calibration.

3.3 Collagen architecture and deformation

IPOLπ allowed visualizing the collagen microstructure and orientation in high spatial and angular resolutions (**Figure 5**). The raw image allowed identifying birefringent (e.g., collagen and neural tissues) and non-birefringent (e.g., pigment, and background) components. Collagen and neural tissues are easy to distinguish due to the large differences in birefringence. These are, in part because of their composition, but also because the axons are primarily perpendicular to the

section and therefore have lower birefringence than the collagen fibers that are primarily in the
plane. (Axer et al., 2001; Yang et al., 2018b) Collagen fibers in the lamina beams and in the scleral
canal have clear differences in color, illustrating the strength of IPOLπ having color cycles every
180 degrees compared with conventional IPOL with color cycles every 90 degrees. This help
distinguish fibers, even in dense sclera and in complex lamina beams. The post-processed image
shows collagen orientation quantitatively. This grey background or region was converted into dark
in the energy map. This enhanced the contrast between birefringent and non-birefringent
architecture. Figure 6a shows an example IPOLπ image mosaic of a coronal section of sheep
optic nerve head. The post-processed image shows orientation and energy information (**Figure
6b**), allowing further quantitative analysis at different scales. At a large scale, IPOLπ allows the
calculation, for instance, of regional fiber anisotropy. At a small scale, IPOLπ allows identifying
collagen microscale properties such as crimp, or the natural waviness of collagen fibers,
discernible in both the PPS and LC regions (**Figure 6c**). IPOLπ can capture dynamic
deformations such as uniaxial stretch testing (**Figure 7a**). All post-processing can be conducted
after the testing (**Figure 7b**). High spatial and temporal resolutions of IPOLπ enabled the
visualization and quantification of the process of load-induced collagen fiber re-orientations
(**Figure 7c**).
233

4. Discussion

Our goal was to demonstrate IPOLπ, a new variation of IPOL in which the orientation-encoding color is cyclic every 180 degrees (π radians). We have shown that IPOLπ produces color images with high spatial, angular and temporal resolutions, suitable for visualization and quantitative analysis of the architecture and dynamics of optic nerve head tissues. Using simulations we built a foundation to convert RGB color into orientation and energy information, and have proved that the orientation-encoded color is cyclic every 180 degrees. The quantitative capability of IPOLπ enables further study on essential biomechanical properties of collagen, such as fiber anisotropy and crimp. We have illustrated the application of IPOLπ high frame rates to show fiber reorientation in uniaxial stretch tests. Altogether, the high spatial and temporal resolutions of IPOLπ enable a deeper insight into ocular structure and biomechanics, and through this on eye physiology and pathology. IPOLπ combines features of conventional PLM and IPOL and that we discuss in detail below. Interestingly, through this work we have also shown that IPOLπ has several convenient properties beyond the color cycle. Further down in the discussion we address these, explaining the strengths and potential applications.

Both IPOLπ and conventional PLM allow distinguishing fiber orientations ranging from 0 to 180 degrees since both system configurations include a circular polarizer that can differentiate the polarization directions of light ranging from 0 to 180 degrees. (Jan et al., 2015a; Kalwani et al., 2013) Images acquired using conventional PLM are monochromatic, and thus the selection of the circular polarizer can be specific to the use of the imaging wavelength. In contrast, in IPOLπ, white light is used to acquire images. Since there is no ideal circular polarization (i.e., perfectly wavelength-independent), the output polarization of each wavelength other than the specific wavelength through the circular polarizer is not ideal circular polarization. The ability of the circular polarizer to minimize the difference in the output polarization of each wavelength affects how the output color changes with collagen orientation, which further affects the capability of distinguishing collagen orientation via color. In Figure 3E, we demonstrate IPOLπ configured with a circular polarizer made by an achromatic or zero-order quarter waveplate can produce a clear color for visualization and quantification of collagen orientation.

Both IPOLπ and conventional PLM allow quantifying the retardance of the collagen structure. Conventional PLM requires images with different polarization states to derive information such as fiber orientation and retardance (Jan et al., 2015a; Mehta et al., 2013). In the four-frame algorithm

265 without extinction setting, the measured retardance δ can be calculated as (Shribak and
266 Oldenbourg, 2003)

$$\delta = \sin^{-1}\left(\frac{2\sqrt{\left(I_{90°} - I_{0°}\right)^2 + \left(I_{135°} - I_{45°}\right)^2}}{I_{0°} + I_{45°} + I_{90°} + I_{135°}}\right). \tag{2}$$

267 where four intensity measurements, $I_{0°}$, $I_{45°}$, $I_{90°}$, and $I_{135°}$, are taken with linear polarizer analyzer
268 set to 0°, 45°, 90°, and 135°, respectively. We previously defined an "energy" for visualization as
269 (Jan et al., 2015a; Jan et al., 2017b)

$$energy = \sqrt{\left(I_{90°} - I_{0°}\right)^2 + \left(I_{135°} - I_{45°}\right)^2}. \tag{3}$$

From Equations 2 and 3, we know energy is the sine function of the retardance without
normalization. In Figure 3D, we found that the normalized RGB distance (i.e., radius) between
the calibrated color and calibrated background color followed the sine function of the retardance.
Therefore, it indicates the RGB distance is equivalent with "energy".

IPOLπ has faster imaging speed and preserves higher resolution than conventional PLM due
to the need in PLM of acquiring images with different polarization states. Imaging speed limited
by the time required to change the analyzer orientation often precludes conventional PLM from
applications on dynamic tissue deformation. (Tower et al., 2002) Several techniques have been
demonstrated to improve the imaging speed, such as multiplexing on a single image sensor
(Gruev et al., 2010; Kaminsky et al., 2007) and switching polarization states quickly (Keikhosravi
et al., 2017). However, these techniques still suffer from post-processing, i.e., image registration
and denoising, which takes time and may introduce errors. The interpolation for image registration
may introduce errors in orientation and retardance calculations. For large samples, mosaicking is
required. Stitching among multiple views also interferes with image registration in a single view,
sometimes causing in a large error. Using a camera with built-in multi-analyzer grids on the
imaging sensor is a powerful solution for conventional PLM, which acquires images with four
polarization states in a snapshot (York et al., 2014). However, by having the analyzers integrated
such camera is not suitable for general applications. In IPOLπ, placing z-cut quartz beside the
analyzer allowed diverging the polarization directions of the visible spectrum, generating a colorful
light related to collagen fiber microstructure and orientation. IPOLπ only needs a color snapshot.
Post-processing is just for color inquiry at a single pixel and thus can be computed in parallel or
after imaging. Therefore, IPOLπ preserves the original spatial and temporal resolution offered by
the imaging system.

Both IPOLπ and IPOL imaging techniques are based on z-cut quartz to modulate the
polarization directions of the visible spectrum, and thus can acquire different polarizations of light
via a snapshot. We note three similarities in imaging capability. First, both allow direct
visualization of collagen architecture in color. Second, both leverage the full spatial resolution of
the microscope-camera system since both use a single snapshot to acquire birefringent
information. Third, theoretically, both acquisition speeds are limited only by the camera, thus both
are suitable for imaging tissue dynamics and biomechanics.

The configuration of IPOLπ includes a quarter waveplate but not in IPOL, resulting in different
imaging features. We highlight three important strengths of imaging capability of IPOLπ compared
to the traditional 90-degree IPOL. First, the IPOLπ image can identify the orientations of
orthogonal collagen fibers (e.g., laminar beams and collagen fibers in the scleral canal) through
colors since orientation-encoded color in IPOLπ was 180-degree cyclic in contrast to 90-degree
cyclic in conventional IPOL. Second, acquiring a IPOLπ image only required 1/100 exposure time
than acquiring a conventional IPOL image, thus allowing faster imaging speed. The conventional
IPOL includes a cross-polarizer design. The design causes a dark background and also causes
a low optical transmission rate except for birefringent regions with high retardance. In contrast, a
combination of a circular polarizer and an analyzer in IPOLπ does not block the light, and thus
the optical transmission rate is high, independent of retardance. Third, IPOLπ allows identifying
tissues with high absorption (e.g., pigment and axon) but without birefringence. These non-
birefringent tissues appear dark gray. This allows identifying the spatial relationship between
collagen and non-birefringent tissues. For example, IPOLπ can visualize the relative
displacement and deformation between collagen and axon under stretch testing. For visualization,
post-processing for IPOLπ allows for making the non-birefringent tissues invisible. In contrast, no
information about these non-birefringent tissues using conventional IPOL since they appear very
dark in IPOL.

The IPOLπ system setup is cheaper and simpler than the IPOL's. IPOL requires two pieces
of z-cut quartz, one right-handed and another left-handed. IPOLπ only requires one, which can
be of either type. In our experience the left-handed z-cut quartz costs more than triple that of the
right-handed quartz. This alone should result in reduced costs. In addition, IPOLπ can be
implemented with an imperfectly collimated light source. Polarization rotation through the z-cut
quartz is sensitive to the optical path length. In IPOL, the imaging system (e.g., dissecting
microscope) requires an extra light source collimator to make sure that a collimated light passes
through z-cut quartz in the illumination path. (Lee et al., 2020) This requirement increases the

system's complexity and cost. In IPOLπ, the z-cut quartz can be placed directly into the imaging
path. Background artifacts are minimal since the light passing through the objective lens is
collimated.

We acknowledge that when we set out to develop IPOLπ our motivation was to extend the
color cycle to 180deg, which was essential to our applications. Identifying the other strengths was
somewhat fortuitous. It is not difficult now to imagine situations where these strengths could be
the primary motivation for choosing IPOLπ over regular IPOL. An example of the serendipity of
science and the importance of research. We have also shown how to establish quantitative color-
angle and color-energy mappings using an experimental calibration. The mapping technique can
be applied to any regular color cameras without automatic white balance control and single-color
cameras with switching RGB color filters. The mapping function is system specific. All optical
elements, such as the light source, optical filters, and camera parameters, would impact the
mapping functions. It is worth noting that adjusting the brightness of the image should be based
on the change in exposure time or the use of a neutral density filter. Changing the power of the
light source may change the spectrum and thus affect the colors observed in IPOLπ and in turn
the mapping functions. Since evaluating the interpolating functions at every color pixel takes time,
we produced two 256×256×256 RGB lookup tables for orientation and energy inquiry,
respectively, which replaced time-consuming interpolation with an indexing operation. For a
20000×20000 mosaic image, sufficient to include a full section of the posterior pole of the eye at
0.5μm/pixel resolution, the computational time was reduced from 24 minutes to 6 seconds after
using the lookup table (1/240th of the time). The approach to quantifying local orientation and
energy is powerful and fast, as shown elsewhere in this paper. The color information at each pixel
allows us to characterize collagen fiber architecture without tracing individual fibers. Other
imaging techniques trace individual fibers through recognizing fiber edges (Cheng et al., 2018;
Wu et al., 2003), but these suffer in regions of high fiber density where the edges may be difficult
to discern. To avoid this problem it is common to increase the magnification, but this limits the
field of view.

It is important to consider the limitations of IPOLπ. First, IPOLπ is currently based on
transmitted light illumination, which requires the tissue samples to be cut into sections. This may
limit the potential biomechanics applications. To image thick tissues, IPOLπ would have to be
implemented in a reflected light mode, for example with structured light illumination (Yang et al.,
2018a). Since IPOLπ has a grey background, implementing it with coaxial illumination may cause
strong reflected light. Instead, thick-tissue imaging with IPOLπ might work best if based on off-

axis illumination that reflects from an object at an oblique angle. Second, the diffraction artifacts

from fiber edges in IPOLπ lead to a sharply high energy obtained after RGB color inquiry.

Therefore, to build a color-energy interpolating function for visualization and avoid artifacts, the

energy in low-brightness pseudo calibrations was weighted by the brightness. Third, tissue with

selective absorption (e.g., stain) may affect the output color, and thus induce an error in orientation

and energy mapping. For example, hematoxylin and eosin (H&E) stain is common in medical

diagnosis. (Titford, 2005) Collagen appears strong pink with H&E stain and its orientation color in

IPOLπ would be depressed by the stain. Fourth, the colorful light of IPOLπ contains information

on all birefringent tissue components, including various types of collagens and non-collagenous

components, such as elastin and neural tissue microtubules. Since the birefringence of collagen

is substantially larger than that of other birefringent elements in the eye (Inoué and Oldenbourg,

1998; Waxman et al., 2021) we assumed that the majority of the polarized light interaction was

from collagen. In addition, we acknowledge that IPOLπ cannot distinguish among various

collagen types. The ability to detect birefringence of other components could also be seen as a

strength, as it allows applying of IPOLπ to study the structure and biomechanics of neural tissues.

This could be useful, for example, to study the optic nerve head of rodents, which have a glial,

non-collagenous lamina. (Tamm et al., 2017)

In conclusion, we present IPOLπ as a label-free imaging technique for imaging collagen

tissues in the eye with high spatial, angular, and temporal resolutions. IPOLπ optically encodes

collagen orientation and retardance in color at each pixel, and thus a color snapshot allows

visualization and quantification of collagen architecture. This study provides a novel imaging

modality to collagen microstructure and biomechanics in the eye, which could help understand

the role of collagen microstructure in eye physiology, aging, and in biomechanics-related

diseases.

Acknowledgements

Disclosures: Nothing to disclose.

Funding: Supported in part by National Institutes of Health R01-EY023966, R01-EY031708,

1S10RR028478, P30-EY008098 and T32-EY017271 (Bethesda, MD), the Eye and Ear

Foundation (Pittsburgh, PA), and Research to Prevent Blindness (unrestricted grant to UPMC

ophthalmology, and Stein innovation award to Sigal, IA).

References

Amidror, I., 2002. Scattered data interpolation methods for electronic imaging systems: a survey. Journal of electronic imaging 11, 157-176.

Axer, H., Axer, M., Krings, T., Keyserlingk, D.G.v., 2001. Quantitative estimation of 3-D fiber course in gross histological sections of the human brain using polarized light. Journal of neuroscience methods 105, 121-131.

Baumann, B., Rauscher, S., Glösmann, M., Götzinger, E., Pircher, M., Fialová, S., Gröger, M., Hitzenberger, C.K., 2014. Peripapillary rat sclera investigated in vivo with polarization-sensitive optical coherence tomography. Investigative ophthalmology & visual science 55, 7686-7696.

Brazile, B.L., Hua, Y., Jan, N.-J., Wallace, J., Gogola, A., Sigal, I.A., 2018. Thin lamina cribrosa beams have different collagen microstructure than thick beams. Investigative ophthalmology & visual science 59, 4653-4661.

Canham, P.B., Finlay, H.M., Dixon, J.G., Ferguson, S.E., 1991. Layered collagen fabric of cerebral aneurysms quantitatively assessed by the universal stage and polarized light microscopy. The Anatomical Record 231, 579-592.

Cheng, F., Birder, L.A., Kullmann, F.A., Hornsby, J., Watton, P.N., Watkins, S., Thompson, M., Robertson, A.M., 2018. Layer-dependent role of collagen recruitment during loading of the rat bladder wall. Biomechanics and modeling in mechanobiology 17, 403-417.

Cisek, R., Joseph, A., Harvey, M., Tokarz, D., 2021. Polarization-sensitive second harmonic generation microscopy for investigations of diseased collagenous tissues. Frontiers in Physics 9, 726996.

Collett, E., 2005. Field guide to polarization. Spie Bellingham, WA.

Coudrillier, B., Tian, J., Alexander, S., Myers, K.M., Quigley, H.A., Nguyen, T.D., 2012. Biomechanics of the human posterior sclera: age-and glaucoma-related changes measured using inflation testing. Investigative ophthalmology & visual science 53, 1714-1728.

Diamant, J., Keller, A., Baer, E., Litt, M., Arridge, R., 1972. Collagen; ultrastructure and its relation to mechanical properties as a function of ageing. Proceedings of the Royal Society of London. Series B. Biological Sciences 180, 293-315.

Ethier, C.R., Johnson, M., Ruberti, J., 2004. Ocular biomechanics and biotransport. Annu. Rev. Biomed. Eng. 6, 249-273.

Filas, B.A., Shah, N.S., Zhang, Q., Shui, Y.-B., Lake, S.P., Beebe, D.C., 2014. Quantitative imaging of enzymatic vitreolysis-induced fiber remodeling. Investigative Ophthalmology & Visual Science 55, 8626-8637.

Gogola, A., Jan, N.-J., Brazile, B., Lam, P., Lathrop, K.L., Chan, K.C., Sigal, I.A., 2018a. Spatial patterns and age-related changes of the collagen crimp in the human cornea and sclera. Investigative ophthalmology & visual science 59, 2987-2998.

Gogola, A., Jan, N.-J., Lathrop, K.L., Sigal, I.A., 2018b. Radial and circumferential collagen fibers are a feature of the peripapillary sclera of human, monkey, pig, cow, goat, and sheep. Investigative ophthalmology & visual science 59, 4763-4774.

Gruev, V., Perkins, R., York, T., 2010. CCD polarization imaging sensor with aluminum nanowire optical filters. Optics express 18, 19087-19094.

Gusachenko, I., Tran, V., Houssen, Y.G., Allain, J.-M., Schanne-Klein, M.-C., 2012. Polarization-resolved second-harmonic generation in tendon upon mechanical stretching. Biophysical journal 102, 2220-2229.

Higgins, M.D., 2010. Imaging birefringent minerals without extinction using circularly polarized light. The Canadian Mineralogist 48, 231-235.

Inoué, S., Oldenbourg, R., 1998. Microtubule dynamics in mitotic spindle displayed by polarized light microscopy. Molecular biology of the cell 9, 1603-1607.

Jan, N.-J., Brazile, B.L., Hu, D., Grube, G., Wallace, J., Gogola, A., Sigal, I.A., 2018. Crimp around the globe; patterns of collagen crimp across the corneoscleral shell. Experimental eye research 172, 159-170.

Jan, N.-J., Gomez, C., Moed, S., Voorhees, A.P., Schuman, J.S., Bilonick, R.A., Sigal, I.A., 2017a. Microstructural crimp of the lamina cribrosa and peripapillary sclera collagen fibers. Investigative ophthalmology & visual science 58, 3378-3388.

Jan, N.-J., Grimm, J.L., Tran, H., Lathrop, K.L., Wollstein, G., Bilonick, R.A., Ishikawa, H., Kagemann, L., Schuman, J.S., Sigal, I.A., 2015a. Polarization microscopy for characterizing fiber orientation of ocular tissues. Biomedical optics express 6, 4705-4718.

Jan, N.-J., Lathrop, K., Sigal, I.A., 2017b. Collagen architecture of the posterior pole: high-resolution wide field of view visualization and analysis using polarized light microscopy. Investigative ophthalmology & visual science 58, 735-744.

Jan, N.-J., Lee, P.-Y., Wallace, J., Iasella, M., Gogola, A., Sigal, I.A., 2022. Stretch-Induced Uncrimping of Equatorial Sclera Collagen Bundles. bioRxiv.

Jan, N.-J., Sigal, I.A., 2018. Collagen fiber recruitment: a microstructural basis for the nonlinear response of the posterior pole of the eye to increases in intraocular pressure. Acta biomaterialia 72, 295-305.

Jan, N.J., Grimm, J.L., Tran, H., Lathrop, K.L., Wollstein, G., Bilonick, R.A., Ishikawa, H., Kagemann, L., Schuman, J.S., Sigal, I.A., 2015b. Polarization microscopy for characterizing fiber orientation of ocular tissues. Biomed Opt Express 6, 4705-4718.

Jan, N.J., Lathrop, K., Sigal, I.A., 2017c. Collagen Architecture of the Posterior Pole: High-Resolution Wide Field of View Visualization and Analysis Using Polarized Light Microscopy. Invest Ophthalmol Vis Sci 58, 735-744.

Kalwani, N.M., Ong, C.A., Lysaght, A.C., Haward, S.J., McKinley, G.H., Stankovic, K.M., 2013. Quantitative polarized light microscopy of unstained mammalian cochlear sections. Journal of biomedical optics 18, 026021.

Kaminsky, W., Gunn, E., Sours, R., Kahr, B., 2007. Simultaneous false‑colour imaging of birefringence, extinction and transmittance at camera speed. Journal of microscopy 228, 153-164.

Keefe, D., Tran, P., Pellegrini, C., Oldenbourg, R., 1997. Polarized light microscopy and digital image processing identify a multilaminar structure of the hamster zona pellucida. Human reproduction (Oxford, England) 12, 1250-1252.

Keikhosravi, A., Liu, Y., Drifka, C., Woo, K.M., Verma, A., Oldenbourg, R., Eliceiri, K.W., 2017. Quantification of collagen organization in histopathology samples using liquid crystal based polarization microscopy. Biomedical optics express 8, 4243-4256.

Keikhosravi, A., Shribak, M., Conklin, M.W., Liu, Y., Li, B., Loeffler, A., Levenson, R.M., Eliceiri, K.W., 2021. Real-time polarization microscopy of fibrillar collagen in histopathology. Scientific reports 11, 1-11.

Koike‑Tani, M., Tani, T., Mehta, S.B., Verma, A., Oldenbourg, R., 2015. Polarized light microscopy in reproductive and developmental biology. Molecular reproduction and development 82, 548-562.

Lee, P.-Y., Yang, B., Hua, Y., Brazile, B., Ji, F., Zhu, Z., Fryc, G.A., Sigal, I.A., 2020. Instant polarized light microscopy for real-time wide-field visualization of collagen architecture, Label-free Biomedical Imaging and Sensing (LBIS) 2020. International Society for Optics and Photonics, p. 112510Y.

Lee, P.-Y., Yang, B., Hua, Y., Waxman, S., Zhu, Z., Ji, F., Sigal, I.A., 2022. Real-time imaging of optic nerve head collagen microstructure and biomechanics using instant polarized light microscopy. Experimental Eye Research 217, 108967.

Mansfield, J.C., Winlove, C.P., Moger, J.J., Matcher, S.J., 2008. Collagen fiber arrangement in
normal and diseased cartilage studied by polarization sensitive nonlinear microscopy.
Journal of biomedical optics 13, 044020.
Mehta, S.B., Shribak, M., Oldenbourg, R., 2013. Polarized light imaging of birefringence and
diattenuation at high resolution and high sensitivity. Journal of Optics 15, 094007.
Schanda, J., 2007. Colorimetry: understanding the CIE system. John Wiley & Sons.
Schindelin, J., Arganda-Carreras, I., Frise, E., Kaynig, V., Longair, M., Pietzsch, T., Preibisch, S.,
Rueden, C., Saalfeld, S., Schmid, B., 2012. Fiji: an open-source platform for biological-
image analysis. Nature methods 9, 676-682.
Shribak, M., Oldenbourg, R., 2003. Techniques for fast and sensitive measurements of two-
dimensional birefringence distributions. Applied Optics 42, 3009-3017.
Tamm, E.R., Ethier, C.R., Dowling, J.E., Downs, C., Ellisman, M.H., Fisher, S., Fortune, B.,
Fruttiger, M., Jakobs, T., Lewis, G., 2017. Biological aspects of axonal damage in
glaucoma: a brief review. Experimental eye research 157, 5-12.
Titford, M., 2005. The long history of hematoxylin. Biotechnic & histochemistry 80, 73-78.
Tower, T.T., Neidert, M.R., Tranquillo, R.T., 2002. Fiber alignment imaging during mechanical
testing of soft tissues. Annals of biomedical engineering 30, 1221-1233.
Waxman, S., Brazile, B.L., Yang, B., Lee, P.-Y., Hua, Y., Gogola, A.L., Lam, P., Voorhees, A.P.,
Rizzo III, J.F., Jakobs, T.C., 2021. Lamina cribrosa vessel and collagen beam networks
are distinct. Experimental Eye Research, 108916.
Whittaker, P., Przyklenk, K., 2009. Fibrin architecture in clots: a quantitative polarized light
microscopy analysis. Blood Cells, Molecules, and Diseases 42, 51-56.
Willemse, J., Gräfe, M.G., Verbraak, F.D., de Boer, J.F., 2020. In vivo 3D determination of
peripapillary scleral and retinal layer architecture using polarization-sensitive optical
coherence tomography. Translational vision science & technology 9, 21-21.
Wu, J., Rajwa, B., Filmer, D.L., Hoffmann, C.M., Yuan, B., Chiang, C.-S., Sturgis, J., Robinson,
J.P., 2003. Analysis of orientations of collagen fibers by novel fiber-tracking software.
Microscopy and Microanalysis 9, 574-580.
Yamanari, M., Nagase, S., Fukuda, S., Ishii, K., Tanaka, R., Yasui, T., Oshika, T., Miura, M.,
Yasuno, Y., 2014. Scleral birefringence as measured by polarization-sensitive optical
coherence tomography and ocular biometric parameters of human eyes in vivo.
Biomedical optics express 5, 1391-1402.
Yang, B., Brazile, B., Jan, N.-J., Hua, Y., Wei, J., Sigal, I.A., 2018a. Structured polarized light
microscopy for collagen fiber structure and orientation quantification in thick ocular tissues.
Journal of biomedical optics 23, 106001.
Yang, B., Jan, N.J., Brazile, B., Voorhees, A., Lathrop, K.L., Sigal, I.A., 2018b. Polarized light
microscopy for 3‑dimensional mapping of collagen fiber architecture in ocular tissues.
Journal of biophotonics 11, e201700356.
Yang, B., Lee, P.Y., Hua, Y., Brazile, B., Waxman, S., Ji, F., Zhu, Z., Sigal, I.A., 2021. Instant
polarized light microscopy for imaging collagen microarchitecture and dynamics. Journal
of Biophotonics 14, e202000326.
York, T., Powell, S.B., Gao, S., Kahan, L., Charanya, T., Saha, D., Roberts, N.W., Cronin, T.W.,
Marshall, J., Achilefu, S., 2014. Bioinspired polarization imaging sensors: from circuits and
optics to signal processing algorithms and biomedical applications. Proceedings of the
IEEE 102, 1450-1469.

Figure 1. IPOLπ imaging configuration. A circular polarizer and a polarization decoder are retrofitted into the illumination and imaging paths, respectively. An alternative to the circular polarizer is to use a linear polarizer followed by a quarter-wave plate whose slow and fast axes are at 45°. The locations of the circular polarizer and the polarization decoder are swappable. The polarization of the white light is converted into an elliptically polarized light by passing it through the circular polarizer and the sample, where the orientation of the elliptical polarization depends on local fiber orientation. A new spectrum is then generated after the elliptically polarized light passed through the polarization decoder, and thus the output light is colorful. A color camera acquires the colorful light to produce true-color images indicating collagen fiber microstructure.

Figure 2. Generation of pseudo calibrations from the experimental calibration. An experimental
calibration created a ring with a calibrated background (i.e., central point) and a radius r_1 in the
color space. We defined the RGB distance between color on experimental calibration and
calibrated background (i.e., radius) as a unit of "energy", which was related to retardance. Pseudo
calibrated rings were generated based on changes in brightness and energy. On one hand, an
increase in brightness results in an increase in RGB values in proportion due to exposure time or
tissue absorptance. For example, a pseudo calibration ring with a radius r_2 represents a higher
exposure time or lower tissue absorption than the experimental calibrated result. Although their
radiuses are different, both have the same energy. On the other hand, an increase in energy
results in an increase in the radius of the calibrated ring in proportion. For example, a sample
generating a calibrated ring with a radius r_3 has higher energy than a sample generating a
calibrated ring with a radius r_2. A combination of pseudo and experimental calibrations was used
to build interpolating functions for orientation and energy inquiry.

Figure 3. IPOLπ simulation. (a) A simulated RGB color map of IPOLπ based on an achromatic quarter waveplate covering fiber orientation angles from 0 to 180 degrees and retardance from 0 to π radians. (b) Extracting collagen oriented at a 0 degree from (a), the relationship between retardance and hue is not a constant and the relationship between retardance and brightness is not a strictly increasing function. (c) Extracting from (a), each simulated ring in the RGB color space corresponds to a collagen fiber with a constant retardance oriented at 0 to 180 degrees. The RGB curves with increased retardance present as concentric circles with increased

radiuses, whereas the RGB color for a sample with no retardance is located at the central point of the circle independent of the

orientations. (d) The normalized radius obtained from (c) follows the sine function of the retardance, which is equivalent with the

normalized "energy". (e) Each simulated curve corresponds to IPOLπ configured with a different type of quarter waveplate, to acquire

a collagen fiber oriented at 0 to 180 degrees. An RGB curve obtained from using an achromatic quarter waveplate is similar to the

curve obtained from using an ideal quarter waveplate. A distorted RGB ring obtained from a zero-order waveplate is still colorful, and

thus allows building an interpolating function that produces interpolated orientations and energies at inquiry RGBs. Using a multi-order

waveplate to build IPOLπ lacks the ability to generate colorful images to visualize collagen microstructure.

Figure 4. The calibration curve of IPOLπ in the RGB color space. (a) RGB values were obtained from chicken tendon sections at different orientations using IPOLπ. The calibration curve was a ring, where the average of RGB values from all images was the background color (red spot). (b) A combination of pseudo and experimental calibrations was used to build an interpolating function that produced interpolated orientations at inquiry RGBs. Color represents fiber orientation. (c) A combination of pseudo and experimental calibrations was used to build an interpolating function that produced interpolated energies at inquiry RGBs. Color represents energy. Note that the energy in low-brightness pseudo calibrations was weighted by the brightness to avoid noise artifacts.

Figure 5. Example IPOLπ and the corresponding post-processing images, and an IPOL image. The raw RGB image acquired by IPOLπ had a grey background and the color had a low saturation. The dark regions and grey textured areas were pigments and axons, respectively, due to high optical absorption. Orientation and energy maps were extracted from the IPOLπ image using corresponding interpolating functions. The post-processed color image was obtained from the orientation map of the collagen fibers masked by the energy map. The post-processed image helps us understand the architecture of collagen fibers.

Figure 6. (a) A IPOLπ image acquired as a mosaic of a coronal section from the optic nerve head of a sheep eye. (b) The post-processed color image was calculated from (a). The color disc on the top left-hand side of the image represents local fiber orientation, and the brightness in the image represents energy. (c) Close-up images from (b) shows interweaving in the peripapillary sclera (red box) and collagen beam networks in the lamina cribrosa (yellow box). The high-resolution image allows identifying crimp (white arrows), an important element of the microstructure for ocular biomechanics.

Figure 7. IPOLπ can capture dynamic deformations of ocular tissues. (a) IPOLπ images of an optic nerve head section before stretch and after uniaxial stretch testing. The diagrams on the top right side indicate the stretch directions. (b) The corresponding post-processed color images obtained from the orientation maps masked by the energy maps. (c) Time-sequence images and the corresponding angle distributions from a small interest of region (white box). In the 720th frame, the high frequencies of fiber orientations are close to the stretch directions. The angle distribution plots show the expected change in preferential fiber orientation in the direction of stretch.