Thermal Stimulus Task fMRI in the Cervical Spinal Cord at 7 Tesla

Alan C. Seifert[1,2,3,*], Junqian Xu[4], Yazhuo Kong[5,6], Falk C. Eippert[5,7], Karla L. Miller[5], Irene Tracey[5], S. Johanna Vannesjo[5,8]

[1]Biomedical Engineering and Imaging Institute, Icahn School of Medicine at Mount Sinai, New York, NY USA

[2]Department of Diagnostic, Molecular, and Interventional Radiology, Icahn School of Medicine at Mount Sinai, New York, NY USA

[3]Graduate School of Biomedical Sciences, Icahn School of Medicine at Mount Sinai, New York, NY USA

[4]Departments of Radiology and Psychiatry, Baylor College of Medicine, Houston, TX USA

[5]Wellcome Centre for Integrative Neuroimaging, FMRIB, Nuffield Department of Clinical Neurosciences, University of Oxford, Oxford, UK

[6]Institute of Psychology, Chinese Academy of Sciences, Beijing, China

[7]Max Planck Institute for Human Cognitive and Brain Sciences, Leipzig, Germany

[8]Department of Physics, Norwegian University of Science and Technology (NTNU), Trondheim, Norway

*Corresponding author:
Alan C. Seifert:
E-mail: alan.seifert@mssm.edu
Address: One Gustave L. Levy Place
Box 1234
New York, NY 10029
Phone: 212-824-8456
Fax: 646-537-9589

Running Title: 7T CSC Thermal Task fMRI

Submitted to:

Submission Type: Research Article (Full Paper)

Disclosures: All authors state that they have no conflicts of interest.

Preliminary results from this manuscript have previously been presented in the following conference abstracts:

Seifert AC, Vannesjo SJ. Spatial Specificity of BOLD Signal in the Spinal Cord at 7T Using a Noxious Thermal Stimulus. Proceedings of the International Society for Magnetic Resonance in Medicine 2020:1168.

Seifert AC, Vannesjo SJ. Comparison of Physiological Noise Models for Thermal Stimulus fMRI in the Cervical Spinal Cord at 7T. Proceedings of the International Society for Magnetic Resonance in Medicine 2022:1462.

Abstract

Purpose: Although functional MRI is widely applied in the brain, fMRI of the spinal cord is more technically demanding. Proximity to the vertebral column and lungs results in strong spatial inhomogeneity and temporal fluctuations in B_0. Increasing field strength enables higher spatial resolution and improved sensitivity to BOLD signal, but amplifies the effects of B_0 inhomogeneity. In this work, we present the first stimulus task fMRI in the spinal cord at 7 T. Further, we compare the performance of single-shot and multi-shot 2D EPI protocols, as they differ in sensitivity to spatial and temporal B_0 inhomogeneity.

Methods: The cervical spinal cords of 11 healthy volunteers were scanned at 7 T using single-shot 2D EPI at 0.75 mm in-plane resolution and multi-shot 2D EPI at 0.75 and 0.6 mm in-plane resolutions. For each protocol, the BOLD response to thirteen 10-second noxious thermal stimuli applied to the right thumb was acquired in a 10-minute fMRI run. Image quality, temporal SNR, and BOLD activation (percent signal change and z-stat) at both individual- and group-level were evaluated between the protocols.

Results: Temporal SNR was highest in single-shot and multi-shot 0.75 mm protocols. In group-level analyses, activation clusters appeared in all protocols in the ipsilateral dorsal quadrant at the expected C6 neurological level. In individual-level analyses, activation clusters at the expected level were detected in some, but not all subjects and protocols. Single-shot 0.75 mm generally produced the highest mean z-statistic, while multi-shot 0.60 mm produced the best-localized activation clusters and the least geometric distortion. Larger than expected within-subject segmental variation of BOLD activation along the cord was observed.

Conclusion: Group-level sensory task fMRI of the cervical spinal cord is feasible at 7 T with single-shot or multi-shot EPI. The best choice of protocol will likely depend on the relative importance of sensitivity to activation versus spatial localization of activation for a given experiment.

Highlights

First stimulus task fMRI results in the spinal cord at 7 T

Single-shot 0.75 mm 2D EPI produced the highest mean z-statistic

Multi-shot 0.60 mm 2D EPI provided the best-localized activation and least distortion

Group-level sensory task fMRI of the spinal cord is feasible at 7 T

Key Words

Spinal cord fMRI, 7T MRI, ultra-high field, fMRI

Introduction

The spinal cord houses many scientifically and clinically important neural circuits, such as central pattern generators related to locomotion (1), sensorimotor reflex arcs, and circuits involved in the modulation of nociception (2). Although functional MRI is a well-established and widely-applied imaging modality to probe neural activity in the brain, fMRI of the spinal cord is more technically demanding.

The adult human spinal cord is approximately 15 mm in diameter at its widest point, with gray matter structures only approximately 1-4 mm wide. High spatial resolution is therefore required to image the blood oxygenation level-dependent (BOLD) signal in the spinal cord. In addition to its small size, several additional factors complicate fMRI of the spinal cord. The spinal cord is directly adjacent to the vertebral column, which is composed of alternating vertebral bones and soft tissues that differ in magnetic susceptibility. The spatially periodic B_0 inhomogeneities created in the spinal cord by the vertebral column cannot be adequately cancelled by a counter-field generated by shim currents from the scanner's spherical harmonic-based shim coils, leaving residual focal B_0 inhomogeneities near the intervertebral junctions. The spinal cord is also near the lungs, whose volume and concentration of paramagnetic O_2 varies over time due to respiration. This creates temporal fluctuations in B_0 in the spinal cord (3), producing frequency variation that can exceed 100 Hz at C7 (4). These spatially and temporally varying B_0 inhomogeneities produce geometric distortion and signal loss, as well as apparent motion between volumes in the time series. Pulsatile rostral-caudal flow of cerebrospinal fluid (CSF) over the cardiac cycle, and intermittent swallowing and motion, can also contribute nuisance variation to fMRI signal (5,6).

Increasing the static magnetic field (B_0) improves signal to noise ratio (SNR), which translates to higher achievable spatial resolution. Increased B_0 also increases sensitivity to BOLD signal in gradient-echo (GRE) EPI, leading to better sensitivity and spatial specificity (7,8). These advantages motivate the use of 7 T fMRI to investigate circuits in the spinal cord. However, several difficulties are magnified at or unique to 7 T. As field strength increases, susceptibility-mediated effects are magnified. Hence, although SNR and BOLD signal are enhanced, the geometric distortions and signal loss associated with B_0 inhomogeneity arising from the lungs and vertebral column are also intensified (9,10). 7 T scanners also lack a body RF transmit coil, and standard head RF coils do not cover the cervical spinal cord, which necessitates specialized RF hardware for the spinal cord that remains only accessible at a handful of research centers (11).

These challenges have hampered the development of spinal cord fMRI at 7 T, despite the
potential advantages of higher field strength. To date there are only three studies of spinal cord
fMRI at 7 T (12–14) all of which are resting-state fMRI studies, investigating connectivity between
gray matter regions in the absence of external stimuli or motor tasks. A significant motivation for
the development of spinal cord fMRI methods, however, is to study pain processing in healthy
and pathological states (15–20). This involves measuring the BOLD response to external painful
stimuli. There is thus a need for investigating the potential of 7 T for task-based spinal cord fMRI.

One important step towards fully realizing the potential benefits of 7 T is to establish suitable
acquisition sequences and protocols that yield high BOLD sensitivity while being sufficiently
robust against the challenges posed by static and dynamic B_0 inhomogeneities. At lower field
strengths, various sequences have been investigated for spinal cord fMRI. Single-shot echo-
planar imaging (EPI) is currently the dominant method for fMRI in the human spinal cord, as in
the brain. The initial motor task fMRI studies in the cervical spinal cord, however, used a variety
of pulse sequences, including fast low-angle shot (FLASH) sequences at 1.5 T (21) and 3 T (22),
multi-shot gradient echo (GRE) EPI at 1.5 T (23), spin-echo EPI at 1.5 T (24), and single-shot fast
spin echo at 1.5 T (25,26). Spin-echo-based pulse sequences continued to be used for fMRI in
the spinal cord at 1.5 T based on signal enhancement by extravascular water protons (SEEP)
contrast for many years (27,28). Work on BOLD fMRI in the spinal cord by multiple groups over
the next several years focused mainly on 2D single-shot (29–35) and 3D multi-shot (36) GRE EPI
at 1.5 T and 3 T.

Single-shot EPI sequences are heavily affected by geometric distortion and signal loss caused
by static B_0 inhomogeneity. Multi-shot EPI sequences have strong potential to reduce these
artifacts by shortening the echo train length and echo time. In the previous 7 T resting-state
studies, highly segmented multi-shot 3D EPI with very short echo train lengths was used
(12,13,37), yielding images largely uncorrupted by signal loss and distortion near intervertebral
junctions. However, the transition from single-shot to multi-shot EPI entails a penalty in volume
acquisition time (VAT), which can reduce sensitivity to BOLD activation by reducing the number
of observation time points for a given scan duration. Multi-shot EPI is also more sensitive to
dynamic B_0 field fluctuations, resulting in variable ghosting between volumes in the time series,
which can reduce the temporal SNR (38). In single-shot EPI, on the other hand, time-variable B_0
fields (at the respiratory time-scale) largely translate into apparent motion in the images, which
can be addressed by retrospective motion correction. Considering the relative advantages and
disadvantages of single-shot vs. multi-shot EPI sequences, it is not apparent which is more

suitable for fMRI of the spinal cord at ultra-high field.

In this work, we present the first stimulus task fMRI in the cervical spinal cord at 7 T and compare
the performance of one single-shot 2D EPI and two multi-shot 2D EPI protocols in terms of
sensitivity and specificity to stimulus-related activation.

Methods

The cervical spinal cords (C4-C7 vertebral levels, approximately corresponding to spinal cord
segments C5-C8) of eleven healthy volunteers (ages 22-39, 5 males, 6 females) were scanned
using a 7 T whole-body MRI system (Magnetom 7 T AS, Siemens Healthineers, Erlangen,
Germany) equipped with a G_{max} = 70 mT/m and maximum slew rate = 200 mT/m/ms gradient coil
and a single-channel four-element transmit, 22-channel receive array brainstem and cervical
spine radiofrequency (RF) coil (39). All imaging experiments were performed under an IRB-
approved protocol and in full compliance with all applicable human subject protection regulations.
Shim adjustment (up to 2^{nd} order) was performed during free breathing. B_1^+ adjustment was
performed by setting transmit reference voltage to 250 V, a level which avoids clipping of RF
pulses, and produces correctly-calibrated RF pulses in most subjects at approximately the level
of the C5/6 intervertebral disc. Anatomical T_1-weighted sagittal MP2RAGE (40) images at 0.7
mm isotropic resolution were acquired for every subject to aid in registration to template space.

One single-shot 2D EPI protocol at 0.75 mm in-plane resolution and two multi-shot 2D EPI
protocols at 0.75 mm and 0.6 mm in-plane resolution were used (Figure 1). These three protocols
will be abbreviated SS75, MS75, and MS60, respectively. Each protocol was optimized
independently, rather than attempting to match parameters between protocols. Research pulse
sequences were used for both the single-shot and multi-shot acquisitions. (At the time of this
study, a multi-shot 3D EPI pulse sequence analogous to what was used in references (12–14)
was not available on our scanner platform.) In all protocols, the FOV was intentionally centered
towards the back of the neck to include posterior neck muscles, which, due to their greater
distance from the trachea and esophagus, serve as spatial landmarks that are relatively robust to
swallowing to aid subsequent motion correction.

In the single-shot EPI, a single saturation slab was placed anterior to the spinal cord for outer
volume suppression, thereby reducing the required echo-train length. For in-plane acceleration
using generalized autocalibrating partially parallel acquisition (GRAPPA) (41), 52 lines of
autocalibration signal reference data were acquired using GRE (42), which has been shown to be

optimal for single-shot EPI in the spinal cord at 7 T (43). Partial Fourier (factor 6/8) acquisition

was judged to be necessary in single-shot EPI to further reduce TE to avoid excessive signal loss

due to through-slice dephasing. Image reconstruction of the single-shot EPI was performed

online with the standard reconstruction pipeline of the manufacturer, including Nyquist ghost

correction based on a three-line navigator.

The multi-shot protocols were acquired without a spatial saturation slab due to SAR limitations

given the shorter per-shot TR. Extending the volume acquisition time to accommodate spatial

saturation was judged to be an unfavorable tradeoff. Therefore, a larger FOV had to be acquired

in the multi-shot protocols, than in the single-shot protocol, to encompass signal from the entire

neck. Note that phase-encode FOVs cannot be defined to be larger than readout-encode FOV,

and the readout-encode FOV can be enlarged with no SNR or timing penalties. For this reason,

the FOV in the multi-shot protocol was set larger in the readout (right-left) dimension than in the

single-shot protocol to achieve the intended increase in the phase-encode FOV. Phase-encoding

in the left-right dimension is undesirable because the patterns of spatial distortion in the spinal

cord would be asymmetric across the left-right dimension of the spinal cord. Unlike the single-

shot sequence, the custom multi-shot sequence did not allow for GRAPPA acceleration via the

console, therefore the FOV in the phase-encode dimension was set to half of the readout FOV in

the multi-shot protocols to achieve the intended phase-encode undersampling (Fig. 1). Image

reconstruction of the multi-shot acquisitions was performed offline, with an iterative conjugate

gradient SENSE algorithm (44). The coil sensitivities were estimated based on a low-resolution

gradient-echo acquisition. For SNR-optimal coil combination, the noise correlation between coils

was calculated based on data acquired in vivo without RF excitation. For Nyquist ghost correction,

three-line phase reference data (which characterize the gradient waveforms) were acquired in a

phantom immediately following each session, using the same multi-shot protocols as performed

in vivo. In pilot studies, this approach was observed to be more effective in reducing Nyquist

ghosting than using navigators from the in vivo data. The in vivo data did include a three-line

navigator for each shot, which was used for per-shot frequency offset demodulation. The

frequency offset per shot was estimated based on a magnitude-weighted coil combination of the

phase at the center of k-space in the second navigator echo, unwrapping the phase estimates

between shots. The frequency offset demodulation was necessary to reduce ghosting caused by

temporal B_0 field fluctuations (38).

	Single-Shot	4-Shot	4-Shot
In-Plane Resolution	0.75 mm	0.75 mm	0.60 mm
Slices and Thickness	24 at 3 mm	24 at 3 mm	24 at 3 mm
TR	1500 ms	746 ms	1100 ms
VAT	1500 ms	2984 ms	4400 ms
TE	20.8 ms	18 ms	26 ms
Nominal Flip Angle	68°	45°	56°
FOV Read	84 mm	128 mm	128 mm
Matrix	112	168	216
Bandwidth	1088 Hz/px	1144 Hz/px	890 Hz/px
Echo Spacing	1.04 ms	1.06 ms	1.31 ms
Effective PE Bandwidth	23 Hz/mm	59 Hz/mm	48 Hz/mm
Lines per Shot	41	21	27
Phase Partial Fourier	6/8	Off	Off
Acceleration	GRAPPA R = 2	SENSE R = 2	SENSE R = 2
Measurements	400	202	137
Saturation Slab	1 Anterior	None	None

Figure 1: Pulse sequence parameters and slice positioning for three fMRI
protocols.

To compare expected advantages and disadvantages between the protocols, Table 1 lists
performance considerations relevant to BOLD fMRI, and for each of these gives the governing
protocol parameters, and the expected relative ranking of the three candidate protocols. The
relative importance of the different performance considerations, and the holistic ranking of
protocol performance in cervical spinal cord BOLD fMRI at 7 T, is not obvious; the goal of this
study is to determine this experimentally.

Performance Consideration	Governing Parameter(s)	Expected Ranking
BOLD Sensitivity	TE	**MS60** > SS75 > MS75
Signal Dropout	TE	**MS75** > SS75 > MS60
Ghosting	Number of Shots, TE	**SS75** > MS75 > MS60
Partial Fourier Blurring	Partial Fourier Factor	**MS60 ≈ MS75** > SS75
T2* Decay Blurring	Effective Phase Encode Bandwidth (dk/dt)	**MS75** > MS60 > SS75
Susceptibility-Induced Distortion	Effective Phase Encode Bandwidth (dk/dt)	**MS75** > MS60 > SS75

Table 1: Performance considerations relevant to BOLD fMRI, the protocol
parameters governing performance regarding each image quality or sensitivity
consideration, and the expected performance of the three candidate protocols
relative to each other.

For each of the three imaging protocols, a 10-minute fMRI acquisition was performed with
noxious thermal stimulation applied to the lateral surface of the base of the right thumb with a 30
mm x 30 mm flat thermode at a pre-calibrated intensity of 3/10 (range 42-51°C, average 46.5°C)

using an fMRI-compatible thermal stimulator (TSA-II, Medoc, Israel). Pain scale anchors were 0
= no pain and 10 = worst pain imaginable. Stimulus blocks were 10 seconds in duration, calculated
by including ramp-up and plateau, but excluding ramp-down to baseline. Ramp speed was set to
10°C/sec. Stimulus blocks were separated by pseudo-randomized inter-stimulus intervals ranging
from 25 to 45 seconds, plus initial and final stimulus-free blocks. This stimulus was expected to
produce sensory activation in the ipsilateral dorsal horn at the neurological C6 (approximately
vertebral C4-C5) level (45,46). Subjects were coached to remain still and maintain steady
respiration during thermal stimuli. Pulse-oximeter and respiratory traces were acquired (TSD200,
PPG100C, TSD201, and RSP100C, Biopac, Goleta, CA, USA).

Spinal cord toolbox v5.4.0 (47) was used for image pre-processing. Anatomical MP2RAGE
images were segmented using *sct_deepseg_sc*, vertebral levels were labeled using
sct_label_vertebrae and/or *sct_label_utils*. Registration to the PAM50 template (48) was
performed using *sct_register_to_template*. Each 4-D fMRI dataset was motion-corrected using
sct_fmri_moco with an inclusive cylindrical mask that contains the posterior neck muscle, the
spinal canal and spinal cord were manually masked in FSLeyes (49), and the fMRI dataset was
registered to the PAM50 template using *sct_register_multimodal*, initialized using the
transformation calculated between the anatomical image and PAM50. Each fMRI dataset was
straightened, spatially smoothed using *sct_smooth_spinalcord* with an anisotropic 2 mm x 2 mm
x 6 mm Gaussian kernel oriented along the spinal cord centerline, and unstraightened, to
minimize blurring of CSF signal into the cord.

First-level GLM analyses was performed in individual-subject space using FSL FEAT (50),
incorporating autocorrelation correction and a 35-term physiological noise model (PNM) (31,51).
The PNM was generated separately for each slice, incorporating slice timing correction, using
FSL PNM, and included 8 cardiac phase, 8 respiratory phase, 16 cardiac-respiratory interaction,
CSF signal derived from the spinal canal mask (canal mask minus cord mask), and two in-plane
translational motion correction terms from *sct_fmri_moco*. Group-level GLM analysis was also
performed in FSL FEAT using a mixed-effects model after warping the results of first-level
analyses into PAM50 template space.

Temporal signal to noise ratio (tSNR) was calculated voxelwise as the ratio of temporal mean
to standard deviation, then averaged within neurological level C6 in the PAM50 cord, gray matter,
and white matter masks after motion correction, but before smoothing and application of the PNM.
The extent of this neurological C6 mask was defined generously as the region where the
probabilistic mask of C6 contained within SCT (52) is greater than zero. Several additional ROI

masks were defined for further analysis (Figure 2). The PAM50 cord mask, eroded by 1 voxel,

was used to mask parameter maps for visualization. A mask for expected true-positive activation

was constructed by summing the PAM50 masks for the right dorsal horn, right fasciculus gracilis,

right fasciculus cuneatus, and right lateral corticospinal tract (PAM50 atlas mask entries 1, 3, 5,

and 35), binarizing, eroding by one voxel from the cord-CSF border only, and calculating the

intersection of the result with the aforementioned PAM50 mask for spinal (i.e., neurological) level

C6 thresholded at zero. White matter structures were included to enclose potential BOLD signal

in veins draining the expected gray matter site of activation (53), but the spinal venous plexus on

the exterior of the spinal cord, though expected to also exhibit BOLD activation (54), was excluded

due to the strong effects of CSF signal that cannot be fully accounted for by a PNM and the

resulting strong potential for false-positive activation. A control mask, for comparison purposes,

was calculated similarly using PAM50 masks for the left lateral corticospinal tract, left rubrospinal

tract, left spinothalamic and spinoreticular tracts, and left ventral horn (PAM50 mask entries 4, 8,

12, and 30). These masks were also warped to each individual subject space. Although studies

in rats have shown that some contralateral ventral horn activation does occur in response to

noxious stimuli (55,56), this was judged to be the most suitable control region in light of the

stronger resting-state connectivity between ipsilateral dorsal and ventral horns, and between

ipsilateral and contralateral dorsal horns, in humans (12,13,57).

268   Figure 2: Masks used for analysis of task fMRI activation, displayed in PAM50

269   template space. The PAM50 cord mask, eroded by 1 voxel (blue), was used to

270   mask parameter maps for visualization. Masks representing regions of interest in

271   the C6 spinal level for expected true positive activation (green) and control (red)

are also displayed.

An uncorrected threshold of z > 1.64, which corresponds to a one-tailed p-value of 0.05, was

chosen for thresholding of activation maps. For the group-level and each individual-subject result,

the mean and maximum z-statistic across activated (z > 1.64) voxels within the true-positive and

control ROI masks were tabulated. The ratio of the maximum z-statistic in the true-positive ROI

mask to the maximum z-statistic in the control ROI mask in each subject was taken as a measure

of the specificity of each protocol to true-positive activation. Mean BOLD percent signal change

(PSC) was calculated in activated native space voxels in the true-positive region for each subject,

and the BOLD percent signal change in the voxel in the true-positive region having the highest z-

statistic in each subject was also tabulated.

Results

Individual subject-level mean and single-timeframe images and tSNR maps of datasets

representing the best-case and worst-case image quality in the study are shown in Figure 3.

These best-case and worst-case datasets were chosen by first selecting the subsets of cases

with the highest and lowest tSNR, and from those subsets, choosing the dataset with the lowest

and highest artifact burden by visual inspection, respectively. In axial slices near intervertebral

junctions, where residual B_0 inhomogeneity is large, signal loss due to through-slice dephasing

and geometric distortion are evident. Geometric distortion was less severe in multi-shot

acquisitions, which had higher effective phase encoding bandwidth. The single-shot EPI had a

blurrier appearance than the multi-shot acquisitions, while multi-shot 0.60 mm produced the

sharpest appearing images. A spatial gradient in tSNR was also evident, with lower tSNR at more

inferior levels. This resulted from a combination of reduced B_1^+ efficiency, increased static B_0

inhomogeneity, and larger time-varying B_0 fluctuations due to the proximity to the lungs.

Figure 3: Individual subject-level images and temporal signal to noise ratio
(tSNR) maps spanning C4-C7 vertebral levels in datasets representing best and
worst image quality in the study. Locations of the three axial slices are indicated
with yellow lines in the sagittal frames. The first and third displayed axial slices are
superior and inferior to an intervertebral disc, and the second axial slice is located
at an intervertebral disc.

Descriptive statistics of tSNR are given in Table 2. Temporal SNR was generally highest in
single-shot 0.75 mm and multi-shot 0.75 mm, with lower variability between subjects in multi-shot
0.75 mm, while tSNR was significantly lower in multi-shot 0.6 mm, yet with lowest between-subject
variability. Temporal SNR was slightly greater in gray matter than in white matter in all three
protocols, driven predominantly by the higher signal in gray matter.

Temporal SNR in Neurological Level C6

	SS75	MS75	MS60
tSNR in Cord	14.6 ± 4.2	14.2 ± 3.3	10.0 ± 2.5
tSNR in GM	15.4 ± 4.2	15.1 ± 3.7	10.8 ± 2.6
tSNR in WM	14.2 ± 4.2	13.7 ± 3.0	9.7 ± 2.5

Table 2: Summary statistics of temporal signal to noise ratio (tSNR)
measurements in data from three protocols. Tabulated means and standard
deviations are calculated from individual subject-level means within PAM50
masks. NB: standard deviations across subjects of mean tSNR in the cord, GM,
and WM do indeed all round to 4.2; differences appear at the third decimal place.

Maps of z-statistics and BOLD percent signal changes in group-level GLM analyses are shown
in Figure 4. Z-statistic maps are thresholded at an uncorrected threshold of z = 1.64. Clusters of
activation appear in all three protocols in the right (ipsilateral) dorsal quadrant at the expected C6
neurological level, which is centered approximately at the C4-C5 intervertebral disc (see Figure 2
for superior-inferior extent). All images are displayed in radiological orientation convention. The
ipsilateral activation appears in the same cross-sectional location in single-shot 0.75 mm and
multi-shot 0.6 mm; however, this cluster is weaker and located more peripherally in multi-shot
0.75 mm. In single-shot 0.75 mm, a cluster also appears in the contralateral (left) dorsal horn.

Figure 4: Maps of z-statistics and BOLD percent signal changes in group-level
GLM analyses. Z-statistic maps are displayed using an uncorrected threshold of z
= 1.64. Maps of BOLD percent signal change are thresholded for ease of
visualization. Maps are masked to the spinal cord, encompassing all neurological
levels, and the slice locations are centered on the voxel inside the spinal cord mask
with the highest z-statistic.

Details of z-statistics and BOLD percent signal change in group-level GLM analyses are given
in Table 3. Voxels within the true-positive and control ROI masks with z > 1.64 are considered
activated. Mean z-statistics in activated true-positive voxels are relatively consistent in the range
of 1.83 to 2.04 across the three protocols, with single-shot 0.75 mm yielding the highest mean z-
statistic contrary to initial expectation (Table 1). The maximum true-positive z-statistic is highest
in single-shot 0.75 mm at 2.85, compared to 2.07 in multi-shot 0.75 mm and 2.38 in multi-shot 0.6
mm. All three protocols yielded a ratio of true-positive to control max z-statistic between 1.27 and
1.37.

The mean BOLD percent signal change in activated voxels was higher in multi-shot datasets
than in single-shot. Multi-shot 0.75 mm yielded the highest mean BOLD percent signal change,
at 0.43%, and the BOLD percent signal change in the voxel with the highest z-statistic was 0.50%.

Group-Level Activation Results

Threshold for Significance is Z > 1.64	SS75	MS75	MS60
Mean Z-Stat in Activated Voxels in True Positive Region	2.04	1.83	1.84
Max Z-Stat in True Positive Region	2.85	2.07	2.38
Max Z-Stat in Control Region	2.24	1.51	1.83
Ratio of Max Z-Stat in True Positive vs. Control	1.27	1.37	1.30
Mean BOLD PSC in Activated Voxels	0.21	0.43	0.33
BOLD PSC in Voxel w/ Highest Z-Stat	0.22	0.50	0.38

Table 3: Summary statistics of group-level GLM results.

Maps of z-statistics and BOLD percent signal changes in two representative individual-level
GLM analyses are shown in Figure 5 (best-quality dataset) and Figure 6 (worst-quality dataset).
Z-statistic maps are thresholded at an uncorrected threshold of z = 1.64.

In the best-quality dataset in Figure 5, clusters of activation appear in the ipsilateral dorsal horn
in all three protocols. The superior-inferior location of the peak activated cluster in the three
protocols varies within this single subject to a greater extent than the expected size of a spinal
segment in a single subject, although all three locations are contained within the group-level
probabilistic PAM50 atlas entry for the C6 neurological level. In single-shot 0.75 mm, the cluster

extends into the ipsilateral ventral horn as well, and a smaller cluster appears in the contralateral

deep dorsal horn. In multi-shot 0.75, a disconnected cluster of weak activation also appears in

the ipsilateral ventral horn. The cluster in multi-shot 0.6 mm is intense and well-localized,

coinciding with a cluster of high BOLD percent signal change approaching 2%. In the displayed

slices, one likely false-positive voxel is visible at the contralateral cord-CSF boundary in the multi-

shot 0.60 mm, and the contralateral cluster in single-shot 0.75 mm is neither clearly true-positive

nor false-positive, but for the purpose of this analysis, a portion of it falls within the control mask

ROI, and is treated as such in descriptive statistics.

Figure 5: Individual subject-level maps of activation z-statistic and BOLD percent
signal change, warped into template space, in the "best quality" dataset from
Figure 2. Slice locations are centered on the voxel with the highest z-statistic.

In the worst-quality dataset in Figure 6, potentially plausible true-positive activation is visible in
the ipsilateral dorsal quadrant in the multi-shot 0.6 mm protocol, though this cluster extends
partially out of the superior boundary of the neurological C6 level. No activation exceeding the
threshold of z = 1.64 is visible in either the single-shot or multi-shot 0.75 mm protocols. The
location of the peak (although still sub-threshold) z-statistic within the true-positive mask ROI is
indicated with a green crosshair in the axial plane, and yellow lines in the sagittal and coronal
planes.

Figure 6: Individual subject-level activation maps, warped into template space, in the "worst quality" dataset from Figure 2. For single-shot 0.75 mm and 4-shot 0.75 mm, the images are centered at the voxel in the true positive region with the highest z-statistic (indicated with a green crosshair), but this z-statistic is below the threshold of 1.64.

Details of z-statistics and BOLD percent signal change across all individual-level GLM analyses are given in Table 4. As in group-level analyses, voxels within the true-positive and control ROI masks (warped to each individual subject space) with z > 1.64 are considered activated. Mean z-

statistics in activated true-positive voxels are relatively consistent in the range of 1.96 to 2.18
across the three protocols. Single-shot 0.75 mm yields the highest mean z-statistic, but the
standard deviation of each subject's mean z-statistic in activated true-positive voxels is double
that of the multi-shot protocols. Single-shot 0.75 mm also yields the highest mean maximum z-
statistic at 2.68, compared to 2.27 for multi-shot 0.75 mm and 2.45 for multi-shot 0.6 mm. Single-
shot 0.75 mm and multi-shot 0.6 mm have true-positive-to-control ratios substantially greater than
unity, while the ratio in multi-shot 0.75 mm is closer to unity.

Similar to group-level analyses, the mean BOLD percent signal change in activated voxels was
higher in multi-shot protocols. Multi-shot 0.6 mm yielded the highest mean BOLD percent signal
change, at 1.78% ± 0.89%, followed by multi-shot 0.75 mm at 1.29% ± 0.76%, and single-shot
0.75 mm at 0.67% ± 0.31%. The same relationship among the protocols is evident in the BOLD
percent signal change in the voxel with the highest z-statistic in each dataset.

Individual-Level Activation Results

Threshold for Significance in Z > 1.64	SS75	MS75	MS60
Mean Z-Stat in Activated Voxels in True Positive Region	2.18 ± 0.37	1.99 ± 0.17	1.96 ± 0.19
Max Z-Stat in True Positive Region	2.68 ± 1.41	2.27 ± 0.77	2.45 ± 0.55
Max Z-Stat in Control Region	2.27 ± 1.10	2.14 ± 0.76	2.11 ± 0.78
Ratio of Max Z-Stat in True Positive vs. Control	1.18	1.06	1.16
Mean BOLD PSC in Activated Voxels	0.67 ± 0.31	1.29 ± 0.76	1.78 ± 0.89
BOLD PSC in Voxel w/ Highest Z-Stat	0.88 ± 0.84	1.51 ± 1.07	2.11 ± 1.18

Table 4: Summary statistics of individual subject GLM analyses. In individual
subjects, mean z-statistics and mean BOLD percent signal change (PSC) are
calculated within voxels having z > 1.64, as well as BOLD PSC in the voxel having
the highest z-statistic. Maximum z-statistics in true positive and control regions
are calculated without a threshold on z-statistic. Means and standard deviations
across these individual-subject mean and maximum values are then calculated
and tabulated here.

Discussion

Image quality and tSNR are strongly impaired in spinal cord fMRI by static focal B_0
inhomogeneity near the intervertebral junctions and temporally variable B_0 inhomogeneity arising
from the lungs. These two sources of B_0 inhomogeneity were clearly evident in the images and
tSNR maps shown in Figure 3. Near the intervertebral junctions, tSNR was sharply reduced and
geometric distortion and signal loss were markedly more severe than in slices further from a
junction. The gradual reduction in tSNR in lower slices was partially due to static and dynamic B_0

inhomogeneity from the lungs, but reduced B_1^+ efficiency in the lower cervical spine was also a
significant contributor (39); signal intensity in mean and single images was lower adjacent to
vertebral level C7, compared to C4-C6, even in the best-quality dataset.

In general, the single-shot 0.75 mm protocol provided the greatest sensitivity (albeit highest
between subject variability in z-statistics) to the expected activation, but also appeared to have
the lowest effective in-plane resolution and largest geometric distortion. This was in line with
expectations (Table 1), as the lower effective phase-encode bandwidth makes it more susceptible
to T_2^* blurring and geometric distortion due to static B_0 inhomogeneity, on top of blurring due to
partial Fourier acquisition. Multi-shot 0.6 mm produced the most well-localized activation clusters,
and the sharpest images, while being qualitatively less affected by geometric distortion caused
by static B_0 inhomogeneity than single-shot 0.75 mm. However, sensitivity to activation was
reduced due to its longer VAT, which reduced the number of time points in an experiment of fixed
duration, and the lower tSNR, both of which reduce the statistical power. Multi-shot 0.75 mm
generally performed poorly; it neither showed the sensitivity to activation of the single-shot 0.75
mm protocol nor the spatial specificity of the multi-shot 0.60 mm protocol, as if sharing the
drawbacks of each of the other protocols, without enjoying the benefits of either protocol to a
similar degree.

Decades of results in the brain have consistently shown that longer TEs (i.e., close to the T_2^* of
gray matter) are associated with greater sensitivity to activation. In the spinal cord at 7 T, however,
the detrimental effects of B_0 inhomogeneity severely impair data quality at longer Tes. A balance
must be struck between image quality, which favors shorter TE, and sensitivity to BOLD signal,
which favors TE close to T_2^* of gray matter, which is 29.3 ± 4.5 ms at 7 T (58). The optimal TE for
a given spinal cord fMRI pulse sequence will therefore likely be shorter than the T_2^* of gray matter.

In group-level GLM analyses, single-shot 0.75 mm and multi-shot 0.6 mm appeared to yield
similarly-sized activation clusters, and the maximum z-statistic in single-shot 0.75 mm was
substantially greater than the maximum z-statistic in multi-shot 0.6 mm (see Figure 4). In
individual-level results, however, multi-shot 0.6 mm generally yielded more sharply-defined
activation clusters in the ipsilateral dorsal quadrant (see Figure 5), and the difference between
the maximum z-statistic in single-shot 0.75 mm and the maximum z-statistic in multi-shot 0.6 mm
was much smaller than in the group level analysis. A possible explanation that unifies these
observations is that multi-shot 0.6 mm was better able to spatially resolve activation in individual
subjects, resulting in small clusters; however, when multiple subjects' data was combined, the
inherent physiological variability in the location of activation among different subjects (52), as well

as potential limitations in group-level registration of distorted spinal cord EPI images, resulted in
these sharply-defined individual-level activation clusters spatially coinciding less fully in multi-shot
0.6 mm than did the broader activation clusters in single-shot 0.75 mm. These broader activation
clusters in single-shot EPI may have arisen due to its greater statistical power, having more
temporal frames due to a shorter volume acquisition time and higher tSNR due to the lower
effective resolution.

Based on TE, BOLD percent signal change would be expected to be highest in multi-shot 0.60
mm, followed by single-shot 0.75 mm and then multi-shot 0.75 mm. In individual-level results,
multi-shot 0.60 mm did indeed yield the highest BOLD percent signal change. However, of the
remaining two protocols, multi-shot 0.75 mm yielded higher BOLD percent signal change than
single-shot 0.75 mm. One potential explanation for this is that the effect of spatial resolution
outweighed the effect of the magnitude of the underlying BOLD activation in these individual-level
measurements; the multi-shot 0.75 mm acquisition is expected to better resolve BOLD activation
than the single-shot 0.75 mm acquisition. An alternative explanation specifically concerning the
difference in the mean BOLD percent signal change in activated voxels between the multi-shot
0.75 mm and single-shot 0.75 mm acquisitions is that the single-shot 0.75 mm acquisition, which
contained 400 temporal frames, was statistically better powered to detect low-BOLD-percent-
signal-change activation than the multi-shot 0.75 mm acquisition, which contained only 202
temporal frames and may have been additionally hampered by greater physiologic noise
contributions (e.g., ghosting). In other words, a voxel with a low BOLD percent signal change is
more likely to be identified as significantly activated using a protocol with a larger number of
temporal frames and less susceptibility to ghosting. This would result in a greater number of low-
BOLD-percent-signal-change voxels being included in the set of significantly activated voxels in
the single-shot 0.75 mm acquisition, depressing the mean BOLD percent signal change
calculated for that set of voxels.

The superior-inferior location of peak activation was similar across the three protocols in group-
level analyses, and appeared at the expected location in the center of neurological C6
(determined from the probabilistic spinal levels from the PAM50 atlas), based on the anatomical
location of the thermal stimulus. The superior-inferior location of the peak activation was expected
to vary across subjects due to natural physiological variation (52), but was not expected to vary
among the three protocols within each subject. Contrary to this expectation, we did observe
substantial variation in the superior-inferior location of peak activation of up to 1 ½ vertebral levels
across the three protocols within individual subjects (see Figure 5). The cause for this variation is

unclear without conducting a reproducibility study by repeating all three protocols on the same
participants. If the true underlying stimulus-related activation occurs across a large spatial extent
in the superior-inferior dimension, then random noise as well as variations in physiological noise
may shift the center of the detected activation between acquisitions in an unsystematic way, as
observed. In that case, increasing the sensitivity to detect BOLD responses, e.g. by acquiring for
longer, should yield extended activation clusters. The true underlying activation is however
unknown, and the data in this work cannot convincingly support any conclusion. From resting-
state studies, we would expect networks of correlated activity stretching over approximately one
vertebral level (57). Task-activated networks, and patterns of nociceptive activation in particular
may, however, have a broader spatial extent to resting-state networks due to the fact that primary
afferent projections may ascend or descend in Lissauer's tract before synapsing (59–61). A large
spatial extent of the activated networks in the superior-inferior dimension may even be a
prerequisite to detect activation clusters in group-level analyses with relatively few subjects. If the
activation were narrowly localized within one subject, with a variable center along the vertebral
levels between subjects, the statistical power to detect activation at the group level would be much
diluted. Further studies would be needed to investigate the true extent of the activation, and the
reproducibility of individual-level activation clusters.

Several groups have developed methods to address distortion and through-slice dephasing in
single-shot GRE EPI at 3 T through slicewise dynamic shimming methods using either slice-
specific gradient offsets (62) or z-gradient refocusing blips (63,64). These methods, applied to
single-shot EPI, may be promising alternatives to multi-shot 2D or 3D EPI at 7 T but, at the time
of this study, were not sufficiently refined to implement routinely at 7 T. Also 3D multi-shot EPI
sequences suitable for implementation in fMRI experiments were not readily available at the time
of this study, despite encouraging results demonstrating their robustness to B_0 inhomogeneity in
resting-state fMRI studies at 7 T (12,13). Such sequences remain a highly promising option worthy
of future investigation.

Conclusions

This study provides, to our knowledge, the first stimulus task fMRI results achieved in the spinal
cord at 7 T. For a given scan time, single-shot 0.75 mm was most sensitive to the expected
activation, likely due to its shorter volume acquisition time and greater statistical power, while
multi-shot 0.60 mm provided the best-localized activation clusters and the least geometric
distortion. The best choice of protocol will likely depend on the relative importance of sensitivity
to activation versus spatial localization of activation for a given experiment.

Acknowledgements

This study was supported by National Institutes of Health (NINDS) award K01NS105160 (ACS) and Department of Defense MSRP grant W81XWH-17-1-0204 (JX). The work was also supported by the Wellcome Trust (Strategic Award 102645/Z/13/Z, Senior Research Fellowship 202788/Z/16/Z). The Wellcome Centre for Integrative Neuroimaging is supported by core funding from the Wellcome Trust (203139/Z/16/Z). Data in this study was acquired using the University of Minnesota Center for Magnetic Resonance Research (CMRR) Multiband EPI pulse sequence software package. Portions of this study in smaller numbers of subjects have previously been presented as conference abstracts (53,65).

References

1. Minassian K, Hofstoetter US, Dzeladini F, Guertin PA, Ijspeert A. The Human Central Pattern Generator for Locomotion: Does It Exist and Contribute to Walking? Neuroscientist 2017;23:649–663 doi: 10.1177/1073858417699790.

2. Bingel U, Tracey I. Imaging CNS modulation of pain in humans. Physiology (Bethesda) 2008;23:371–380 doi: 10.1152/physiol.00024.2008.

3. Raj D, Anderson AW, Gore JC. Respiratory effects in human functional magnetic resonance imaging due to bulk susceptibility changes. Phys. Med. Biol. 2001;46:3331–3340 doi: 10.1088/0031-9155/46/12/318.

4. Vannesjo SJ, Miller KL, Clare S, Tracey I. Spatiotemporal characterization of breathing-induced B0 field fluctuations in the cervical spinal cord at 7T. NeuroImage 2017;167:191–202 doi: 10.1016/j.neuroimage.2017.11.031.

5. Krüger G, Glover GH. Physiological noise in oxygenation-sensitive magnetic resonance imaging: Physiological Noise in MRI. Magn. Reson. Med. 2001;46:631–637 doi: 10.1002/mrm.1240.

6. Cohen-Adad J. Functional Magnetic Resonance Imaging of the Spinal Cord: Current Status and Future Developments. Semin. Ultrasound CT MR 2017;38:176–186 doi: 10.1053/j.sult.2016.07.007.

7. Yacoub E, Shmuel A, Pfeuffer J, et al. Imaging brain function in humans at 7 Tesla. Magn. Reson. Med. 2001;45:588–594.

8. Gati JS, Menon RS, Ugurbil K, Rutt BK. Experimental determination of the BOLD field strength dependence in vessels and tissue. Magn. Reson. Med. 1997;38:296–302 doi: 10.1002/mrm.1910380220.

9. Triantafyllou C, Hoge RD, Krueger G, et al. Comparison of physiological noise at 1.5 T, 3 T and 7 T and optimization of fMRI acquisition parameters. NeuroImage 2005;26:243–250 doi: 10.1016/j.neuroimage.2005.01.007.

10. Stroman PW, Wheeler-Kingshott C, Bacon M, et al. The current state-of-the-art of spinal cord
imaging: methods. NeuroImage 2014;84:1070–1081 doi: 10.1016/j.neuroimage.2013.04.124.

11. Barry RL, Vannesjo SJ, By S, Gore JC, Smith SA. Spinal cord MRI at 7T. NeuroImage
2018;168:437–451 doi: 10.1016/j.neuroimage.2017.07.003.

12. Barry RL, Smith SA, Dula AN, Gore JC. Resting state functional connectivity in the human
spinal cord. Elife 2014;3:e02812 doi: 10.7554/eLife.02812.

13. Barry RL, Rogers BP, Conrad BN, Smith SA, Gore JC. Reproducibility of resting state spinal
cord networks in healthy volunteers at 7 Tesla. NeuroImage 2016;133:31–40 doi:
10.1016/j.neuroimage.2016.02.058.

14. Conrad BN, Barry RL, Rogers BP, et al. Multiple sclerosis lesions affect intrinsic functional
connectivity of the spinal cord. Brain 2018 doi: 10.1093/brain/awy083.

15. Weber KA, Chen Y, Wang X, Kahnt T, Parrish TB. Functional magnetic resonance imaging of
the cervical spinal cord during thermal stimulation across consecutive runs. NeuroImage
2016;143:267–279 doi: 10.1016/j.neuroimage.2016.09.015.

16. Brooks J, Tracey I. From nociception to pain perception: imaging the spinal and supraspinal
pathways. J. Anat. 2005;207:19–33 doi: 10.1111/j.1469-7580.2005.00428.x.

17. Eippert F, Finsterbusch J, Bingel U, Buchel C. Direct evidence for spinal cord involvement in
placebo analgesia. Science 2009 doi: 10.1126/science.1067176).

18. Geuter S, Büchel C. Facilitation of pain in the human spinal cord by nocebo treatment. J.
Neurosci. 2013;33:13784–13790 doi: 10.1523/JNEUROSCI.2191-13.2013.

19. Tinnermann A, Geuter S, Sprenger C, Finsterbusch J, Buchel C. Interactions between brain
and spinal cord mediate value effects in nocebo hyperalgesia. Science 2017;358:105–108 doi:
10.1126/science.aan1221.

20. Staud R, Boissoneault J, Lai S, et al. Spinal Cord Neural Activity of Fibromyalgia Patients and
Healthy Controls During Temporal Summation of Pain: An fMRI Study. Journal of
Neurophysiology 2021:jn.00276.2021 doi: 10.1152/jn.00276.2021.

21. Yoshizawa T, Nose T, Moore GJ, Sillerud LO. Functional magnetic resonance imaging of
motor activation in the human cervical spinal cord. NeuroImage 1996;4:174–182 doi:
10.1006/nimg.1996.0068.

22. Stroman PW, Nance PW, Ryner LN. BOLD MRI of the human cervical spinal cord at 3 tesla.
Magn Reson Med 1999;42:571–576 doi: 10.1002/(sici)1522-2594(199909)42:3<571::aid-
mrm20>3.0.co;2-n.

23. Backes WH, Mess WH, Wilmink JT. Functional MR Imaging of the Cervical Spinal Cord by
Use of Median Nerve Stimulation and Fist Clenching. 2001:6.

24. Stroman PW, Ryner LN. Functional MRI of motor and sensory activation in the human spinal
cord. Magnetic Resonance Imaging 2001;19:27–32 doi: 10.1016/S0730-725X(01)00226-0.

25. Stroman PW, Krause V, Malisza KL, Frankenstein UN, Tomanek B. Functional magnetic
resonance imaging of the human cervical spinal cord with stimulation of different sensory
dermatomes. Magnetic Resonance Imaging 2002;20:1–6 doi: 10.1016/S0730-725X(02)00468-X.

26. Stroman P. Mapping of Neuronal Function in the Healthy and Injured Human Spinal Cord with
Spinal fMRI. NeuroImage 2002;17:1854–1860 doi: 10.1006/nimg.2002.1305.

27. Stroman PW, Krause V, Malisza KL, Frankenstein UN, Tomanek B. Extravascular proton-
density changes as a non-BOLD component of contrast in fMRI of the human spinal cord. Magn.
Reson. Med. 2002;48:122–127 doi: 10.1002/mrm.10178.

28. Lawrence JM, Stroman PW, Kollias SS. Functional magnetic resonance imaging of the human
spinal cord during vibration stimulation of different dermatomes. Neuroradiology 2008;50:273–
280 doi: 10.1007/s00234-007-0338-6.

29. Govers N, Béghin J, Van Goethem JWM, et al. Functional MRI of the cervical spinal cord on
1.5 T with fingertapping: to what extent is it feasible? Neuroradiology 2007;49:73–81 doi:
10.1007/s00234-006-0162-4.

30. Giulietti G, Giove F, Garreffa G, Colonnese C, Mangia S, Maraviglia B. Characterization of
the functional response in the human spinal cord: Impulse-response function and linearity.
NeuroImage 2008;42:626–634 doi: 10.1016/j.neuroimage.2008.05.006.

31. Brooks JCW, Beckmann CF, Miller KL, et al. Physiological noise modelling for spinal
functional magnetic resonance imaging studies. NeuroImage 2008;39:680–692 doi:
10.1016/j.neuroimage.2007.09.018.

32. Piché M, Cohen-Adad J, Nejad MK, et al. Characterization of cardiac-related noise in fMRI of
the cervical spinal cord. Magnetic Resonance Imaging 2009;27:300–310 doi:
10.1016/j.mri.2008.07.019.

33. Eippert F, Kong Y, Winkler AM, et al. Investigating resting-state functional connectivity in the
cervical spinal cord at 3T. NeuroImage 2016;147:589–601 doi:
10.1016/j.neuroimage.2016.12.072.

34. Weber KA, Chen Y, Wang X, Kahnt T, Parrish TB. Lateralization of cervical spinal cord activity
during an isometric upper extremity motor task with functional magnetic resonance imaging.
NeuroImage 2016;125:233–243 doi: 10.1016/j.neuroimage.2015.10.014.

35. Weber KA, Chen Y, Paliwal M, et al. Assessing the Spatial Distribution of Cervical Spinal Cord
Activity during Tactile Stimulation of the Upper Extremity in Humans with Functional Magnetic
Resonance Imaging. NeuroImage 2020:116905 doi: 10.1016/j.neuroimage.2020.116905.

36. Barry RL, Conrad BN, Maki S, et al. Multi-shot acquisitions for stimulus-evoked spinal cord
BOLD fMRI. Magn Reson Med 2020:mrm.28570 doi: 10.1002/mrm.28570.

37. Barry RL, Strother SC, Gatenby JC, Gore JC. Data-driven optimization and evaluation of 2D
EPI and 3D PRESTO for BOLD fMRI at 7 Tesla: I. Focal coverage. NeuroImage 2011;55:1034–
1043 doi: 10.1016/j.neuroimage.2010.12.086.

38. Vannesjo SJ, Clare S, Kasper L, Tracey I, Miller KL. A method for correcting breathing-induced

field fluctuations in T2*-weighted spinal cord imaging using a respiratory trace. Magn. Reson.
Med. 2019 doi: 10.1002/mrm.27664.

39. Zhang B, Seifert AC, Kim J-W, Borrello J, Xu J. 7 Tesla 22-channel wrap-around coil array for
cervical spinal cord and brainstem imaging. Magn. Reson. Med. 2017;78:1623–1634 doi:
10.1002/mrm.26538.

40. Marques JP, Kober T, Krueger G, van der Zwaag W, Van de Moortele P-F, Gruetter R.
MP2RAGE, a self bias-field corrected sequence for improved segmentation and T1-mapping at
high field. NeuroImage 2010;49:1271–1281 doi: 10.1016/j.neuroimage.2009.10.002.

41. Griswold MA, Jakob PM, Heidemann RM, et al. Generalized autocalibrating partially parallel
acquisitions (GRAPPA). Magn. Reson. Med. 2002;47:1202–1210 doi: 10.1002/mrm.10171.

42. Talagala SL, Sarlls JE, Liu S, Inati SJ. Improvement of temporal signal-to-noise ratio of
GRAPPA accelerated echo planar imaging using a FLASH based calibration scan. Magn. Reson.
Med. 2016;75:2362–2371 doi: 10.1002/mrm.25846.

43. Seifert AC, Xu J. Impact of autocalibration method on accelerated EPI of the cervical spinal
cord at 7 T. Magnetic Resonance in Med 2022;88:2583–2591 doi: 10.1002/mrm.29415.

44. Pruessmann KP, Weiger M, Börnert P, Boesiger P. Advances in sensitivity encoding with
arbitrary k-space trajectories: SENSE With Arbitrary k-Space Trajectories. Magn. Reson. Med.
2001;46:638–651 doi: 10.1002/mrm.1241.

45. Keegan JJ, Garrett FD. The segmental distribution of the cutaneous nerves in the limbs of
man. Anat. Rec. 1948;102:409–437 doi: 10.1002/ar.1091020403.

46. Lee MWL, McPhee RW, Stringer MD. An evidence-based approach to human dermatomes.
Clin. Anat. 2008;21:363–373 doi: 10.1002/ca.20636.

47. De Leener B, Lévy S, Dupont SM, et al. SCT: Spinal Cord Toolbox, an open-source software
for processing spinal cord MRI data. NeuroImage 2017;145:24–43 doi:
10.1016/j.neuroimage.2016.10.009.

48. De Leener B, Fonov VS, Collins DL, Callot V, Stikov N, Cohen-Adad J. PAM50: Unbiased
multimodal template of the brainstem and spinal cord aligned with the ICBM152 space.
NeuroImage 2017;165:170–179 doi: 10.1016/j.neuroimage.2017.10.041.

49. Jenkinson M, Beckmann CF, Behrens TEJ, Woolrich MW, Smith SM. FSL. Neuroimage
2012;62:782–790 doi: 10.1016/j.neuroimage.2011.09.015.

50. Woolrich MW, Ripley BD, Brady M, Smith SM. Temporal autocorrelation in univariate linear
modeling of FMRI data. Neuroimage 2001;14:1370–1386 doi: 10.1006/nimg.2001.0931.

51. Kong Y, Jenkinson M, Andersson J, Tracey I, Brooks JCW. Assessment of physiological noise
modelling methods for functional imaging of the spinal cord. NeuroImage 2012;60:1538–1549 doi:
10.1016/j.neuroimage.2011.11.077.

52. Cadotte DW, Cadotte A, Cohen-Adad J, et al. Characterizing the location of spinal and
vertebral levels in the human cervical spinal cord. AJNR Am J Neuroradiol 2015;36:803–810 doi:

10.3174/ajnr.A4192.

53. Seifert AC, Vannesjo SJ. Spatial Specificity of BOLD Signal in the Spinal Cord at 7T Using a
Noxious Thermal Stimulus. Proceedings of the International Society for Magnetic Resonance in
Medicine 2020:1168.

54. Cohen-Adad J, Gauthier CJ, Brooks JCW, et al. BOLD signal responses to controlled
hypercapnia in human spinal cord. NeuroImage 2010;50:1074–1084 doi:
10.1016/j.neuroimage.2009.12.122.

55. Coghill RC, Price DD, Hayes RL, Mayer DJ. Spatial distribution of nociceptive processing in
the rat spinal cord. Journal of Neurophysiology 1991;65:133–140 doi: 10.1152/jn.1991.65.1.133.

56. Coghill RC, Mayer DJ, Price DD. The roles of spatial recruitment and discharge frequency in
spinal cord coding of pain: a combined electrophysiological and imaging investigation. Pain
1993;53:295–309 doi: 10.1016/0304-3959(93)90226-F.

57. Kong Y, Eippert F, Beckmann CF, et al. Intrinsically organized resting state networks in the
human spinal cord. Proc. Natl. Acad. Sci. U.S.A. 2014;111:18067–18072 doi:
10.1073/pnas.1414293111.

58. Massire A, Taso M, Besson P, Guye M, Ranjeva J-P, Callot V. High-resolution multi-
parametric quantitative magnetic resonance imaging of the human cervical spinal cord at 7T.
NeuroImage 2016;143:58–69 doi: 10.1016/j.neuroimage.2016.08.055.

59. Coggeshall RE, Chung K, Chung JM, Langford LA. Primary afferent axons in the tract of
lissauer in the monkey. J. Comp. Neurol. 1981;196:431–442 doi: 10.1002/cne.901960307.

60. Light AR, Perl ER. Spinal termination of functionally identified primary afferent neurons with
slowly conducting myelinated fibers. J. Comp. Neurol. 1979;186:133–150 doi:
10.1002/cne.901860203.

61. Sugiura Y, Lee CL, Perl ER. Central Projections of Identified, Unmyelinated (C) Afferent Fibers
Innervating Mammalian Skin. Science 1986;234:358–361 doi: 10.1126/science.3764416.

62. Islam H, Law CSW, Weber KA, Mackey SC, Glover GH. Dynamic per slice Shimming for
Simultaneous Brain and Spinal Cord fMRI. Magn. Reson. Med. 2018 doi: 10.1002/mrm.27388.

63. Finsterbusch J, Eippert F, Büchel C. Single, slice-specific z-shim gradient pulses improve T2*-
weighted imaging of the spinal cord. NeuroImage 2012;59:2307–2315 doi:
10.1016/j.neuroimage.2011.09.038.

64. Kaptan M, Vannesjo SJ, Mildner T, et al. Automated slice-specific z-shimming for functional
magnetic resonance imaging of the human spinal cord. Human Brain Mapping 2022:hbm.26018
doi: 10.1002/hbm.26018.

65. Seifert AC, Vannesjo SJ. Comparison of Physiological Noise Models for Thermal Stimulus
fMRI in the Cervical Spinal Cord at 7T. Proceedings of the International Society for Magnetic
Resonance in Medicine 2022:1462.

685