Thermal Stimulus Task fMRI in the Cervical Spinal Cord at 7 Tesla

Alan C. Seifert[1,2,3,*], Junqian Xu[4], Yazhuo Kong[5,6], Falk Eippert[5,7], Karla L. Miller[5], Irene Tracey[5], S. Johanna Vannesjo[5,8]

[1]Biomedical Engineering and Imaging Institute, Icahn School of Medicine at Mount Sinai, New York, NY USA

[2]Department of Diagnostic, Molecular, and Interventional Radiology, Icahn School of Medicine at Mount Sinai, New York, NY USA

[3]Graduate School of Biomedical Sciences, Icahn School of Medicine at Mount Sinai, New York, NY USA

[4]Departments of Radiology and Psychiatry, Baylor College of Medicine, Houston, TX USA

[5]Wellcome Centre for Integrative Neuroimaging, FMRIB, Nuffield Department of Clinical Neurosciences, University of Oxford, Oxford, UK

[6]Institute of Psychology, Chinese Academy of Sciences, Beijing, China

[7]Max Planck Institute for Human Cognitive and Brain Sciences, Leipzig, Germany

[8]Department of Physics, Norwegian University of Science and Technology (NTNU), Trondheim, Norway

*Corresponding author:
Alan C. Seifert:
E-mail: alan.seifert@mssm.edu
Address: One Gustave L. Levy Place
            Box 1234
            New York, NY 10029
Phone: 212-824-8456
Fax: 646-537-9589

Running Title: 7T CSC Thermal Task fMRI

Submitted to: Human Brain Mapping

Submission Type: Research Article (Full Paper)

Word Count: 5631 excluding tables, captions, headings, acknowledgements, reference citations, and references (Abstract: 314; Supplemental: 323)

Preliminary results from this manuscript have previously been presented in the following conference abstracts:

Seifert AC, Vannesjo SJ. Spatial Specificity of BOLD Signal in the Spinal Cord at 7T Using a Noxious Thermal Stimulus. Proceedings of the International Society for Magnetic Resonance in Medicine 2020:1168.

Seifert AC, Vannesjo SJ. Comparison of Physiological Noise Models for Thermal Stimulus fMRI in the Cervical Spinal Cord at 7T. Proceedings of the International Society for Magnetic Resonance in Medicine 2022:1462.

Abstract

Purpose: Although functional MRI is widely applied in the brain, fMRI of the spinal cord is more technically demanding. Proximity to the vertebral column and lungs results in strong spatial inhomogeneity and temporal fluctuations in B_0. Increasing field strength enables higher spatial resolution and improved sensitivity to BOLD signal, but amplifies the effects of B_0 inhomogeneity. In this work, we present the first task fMRI in the spinal cord at 7 T. Further, we compare the performance of single-shot and multi-shot 2D EPI protocols, which differ in sensitivity to spatial and temporal B_0 inhomogeneity.

Methods: The cervical spinal cords of 11 healthy volunteers were scanned at 7 T using single-shot 2D EPI at 0.75 mm in-plane resolution and multi-shot 2D EPI at 0.75 and 0.6 mm in-plane resolutions. All protocols used 3 mm slice thickness. For each protocol, the BOLD response to thirteen 10-second noxious thermal stimuli applied to the right thumb was acquired in a 10-minute fMRI run. Image quality, temporal SNR, and BOLD activation (percent signal change and z-stat) at both individual- and group-level were evaluated between the protocols.

Results: Temporal SNR was highest in single-shot and multi-shot 0.75 mm protocols. In group-level analyses, activation clusters appeared in all protocols in the ipsilateral dorsal quadrant at the expected C6 neurological level. In individual-level analyses, activation clusters at the expected level were detected in some, but not all subjects and protocols. Single-shot 0.75 mm generally produced the highest mean z-statistic, while multi-shot 0.60 mm produced the best-localized activation clusters and the least geometric distortion. Larger than expected within-subject segmental variation of BOLD activation along the cord was observed.

Conclusion: Group-level sensory task fMRI of the cervical spinal cord is feasible at 7 T with single-shot or multi-shot EPI. The best choice of protocol will likely depend on the relative importance of sensitivity to activation versus spatial localization of activation for a given experiment.

Highlights

First stimulus task fMRI results in the spinal cord at 7 T

Single-shot 0.75 mm 2D EPI produced the highest mean z-statistic

Multi-shot 0.60 mm 2D EPI provided the best-localized activation and least distortion

Key Words

Spinal cord fMRI, 7T MRI, ultra-high field, fMRI

Required Statements

Data availability statement: The data that support the findings of this study are available from the corresponding author upon reasonable request.

Funding statement: This study was supported by National Institutes of Health (NINDS) award K01NS105160 (ACS) and Department of Defense MSRP grant W81XWH-17-1-0204 (JX). FE is supported by the Max Planck Society and the European Research Council (under the European Union's Horizon 2020 research and innovation programme; grant agreement No 758974). The work was also supported by the Wellcome Trust (Strategic Award 102645/Z/13/Z, Senior Research Fellowship 202788/Z/16/Z). The Wellcome Centre for Integrative Neuroimaging is supported by core funding from the Wellcome Trust (203139/Z/16/Z).

Conflict of interest disclosure: All authors state that they have no conflicts of interest regarding this work.

Ethics approval statement and patient consent statement: All imaging experiments were performed under an IRB-approved protocol with subject consent and in full compliance with all applicable human subject protection regulations.

Permission to reproduce material from other sources: Preliminary results from this manuscript have previously been presented in conference abstracts. No material contained within this submission are encumbered by any restrictions related to their previous publication.

Clinical trial registration: This work does not meet the definition of a clinical trial.

Introduction

The spinal cord houses many scientifically and clinically important neural circuits, such as central pattern generators related to locomotion[1], sensorimotor reflex arcs, and circuits involved in the modulation of nociception[2]. Although functional MRI is a well-established and widely-applied imaging modality to probe neural activity in the brain, fMRI of the spinal cord is more technically demanding.

The adult human spinal cord is approximately 15 mm in diameter at its widest point, with gray matter structures only approximately 1-4 mm wide. High spatial resolution is therefore required to image the blood oxygenation level-dependent (BOLD) signal in the spinal cord. In addition to its small size, several additional factors complicate fMRI of the spinal cord. The spinal cord is directly adjacent to the vertebral column, which is composed of alternating vertebral bones and soft tissues that differ in magnetic susceptibility, introducing distortions in the B_0 field. These spatially periodic B_0 inhomogeneities in the spinal cord cannot be adequately cancelled by a counter-field generated by the scanner's shim coils, leaving residual focal B_0 inhomogeneities near the intervertebral junctions. The spinal cord is also near the lungs, whose volume and concentration of paramagnetic O_2 varies over time due to respiration, also introducing B_0 distortions. The resulting temporal fluctuations in B_0 in the spinal cord[3] produce frequency variation that can exceed 100 Hz at C7[4]. These spatially and temporally varying B_0 inhomogeneities produce geometric distortion and signal loss, as well as apparent motion between volumes in the time series. Finally, pulsatile rostral-caudal flow of cerebrospinal fluid (CSF) over the cardiac cycle, and intermittent swallowing and motion, can also contribute nuisance variations to fMRI data that are more extreme than in the cerebrum[5,6].

Increasing the static magnetic field (B_0) improves signal to noise ratio (SNR), which translates to higher achievable spatial resolution. Increased B_0 also increases the BOLD signal acquired by gradient-echo (GRE) EPI, leading to better sensitivity and spatial specificity[7,8]. These advantages motivate the use of 7 T fMRI to investigate circuits in the spinal cord. However, several difficulties are magnified and become significant at 7 T or higher field strengths. As field strength increases, susceptibility-mediated effects are magnified. Hence, although SNR and BOLD signal are enhanced, the geometric distortions and signal loss associated with B_0 inhomogeneity arising from the lungs and vertebral column are also intensified[9,10]. Most 7 T scanners also lack a body RF transmit coil, and standard head RF coils do not cover the cervical spinal cord, which necessitates specialized RF hardware for the spinal cord that remains only accessible at a handful of research centers[11].

These challenges have hampered the development of spinal cord fMRI at 7 T, despite the

potential advantages of higher field strength. To date there are only three studies of spinal cord

fMRI at 7 T[12–14] all of which are resting-state fMRI studies investigating connectivity between

gray matter regions in the absence of external stimuli or motor tasks. A significant motivation for

the development of spinal cord fMRI methods, however, is to study pain processing in healthy

and pathological states[15–20]. This involves measuring the BOLD response to external painful

stimuli. There is thus a need for investigating the potential of 7 T for task-based spinal cord

fMRI.

One important step towards fully realizing the potential benefits of 7 T is to establish suitable

acquisition sequences and protocols that yield high BOLD sensitivity while being sufficiently

robust against the challenges posed by static and dynamic B_0 inhomogeneities. At lower field

strengths, various sequences have been investigated for spinal cord fMRI. Single-shot echo-

planar imaging (EPI) is currently the dominant method for fMRI in the human spinal cord, as in

the brain. The initial motor task fMRI studies in the cervical spinal cord, however, used a variety

of pulse sequences, including fast low-angle shot (FLASH) sequences at 1.5 T[21] and 3 T[22],

multi-shot gradient echo (GRE) EPI at 1.5 T[23], spin-echo EPI at 1.5 T[24], and single-shot fast

spin echo at 1.5 T[25,26]. Spin-echo-based pulse sequences continued to be used for fMRI in the

spinal cord at 1.5 T based on signal enhancement by extravascular water protons (SEEP)

contrast for many years[27,28]. Work on BOLD fMRI in the spinal cord by multiple groups over the

next several years focused mainly on 2D single-shot[29–35] and 3D multi-shot[36] GRE EPI at 1.5 T

and 3 T.

Most relevant to the present work, others have previously implemented 2D single-shot EPI in

combination with noxious thermal stimulation to study cognitive manipulation of pain[17,37] the

effects of noxious thermal stimulation on functional connectivity in the cervical spinal cord[38], and

several have developed and implemented methods for simultaneous brain and spinal cord

MRI[39–41] and utilized them in studies of, for example, the fundamental organization of resting-

state networks spanning the brain and the cervical spinal cord[42,43] and the interaction between

brain and spinal cord regions on sensitivity to pain[19,44].

Single-shot EPI sequences are heavily affected by geometric distortion and signal loss caused

by static B_0 inhomogeneity. Multi-shot EPI sequences have strong potential to reduce these

artifacts by shortening the echo train length and echo time. In the previous 7 T resting-state

studies, highly segmented multi-shot 3D EPI with very short echo train lengths was used[12,13,45],

yielding images largely uncorrupted by signal loss and distortion near intervertebral junctions.

However, the transition from single-shot to multi-shot EPI entails a penalty in volume acquisition
time (VAT), which can reduce statistical power to detect to BOLD activation by reducing the
number of observation time points for a given scan duration. Multi-shot EPI is also more
vulnerable to dynamic B_0 field fluctuations, resulting in variable ghosting between volumes in the
time series, which can reduce the temporal SNR[46]. In single-shot EPI, on the other hand, time-
variable B_0 fields (at the respiratory time-scale) largely translate into apparent motion in the
images, which can be addressed by retrospective motion correction. Considering the relative
advantages and disadvantages of single-shot vs. multi-shot EPI sequences, it is not apparent
which is more suitable for fMRI of the spinal cord at ultra-high field.

In this work, we present the first stimulus task fMRI in the cervical spinal cord at 7 T and
compare the performance of one single-shot 2D EPI and two multi-shot 2D EPI protocols in
terms of sensitivity and specificity to stimulus-related activation.

Methods

Image Acquisition

The cervical spinal cords (C4-C7 vertebral levels, approximately corresponding to spinal cord
segments C5-C8) of eleven healthy volunteers (ages 22-39, 5 males, 6 females) were scanned
using a 7 T whole-body MRI system (Magnetom 7 T AS, Siemens Healthineers, Erlangen,
Germany) equipped with a G_{max} = 70 mT/m and maximum slew rate = 200 mT/m/ms gradient
coil and a single-channel four-element transmit, 22-channel receive array brainstem and
cervical spine radiofrequency (RF) coil[47]. All imaging experiments were performed under an
IRB-approved protocol with subject consent and in full compliance with all applicable human
subject protection regulations. Shim adjustment (up to 2^{nd} order) was performed during free
breathing. B_1^+ adjustment was performed by setting transmit reference voltage to 250 V, a level
which avoids clipping of RF pulses, and produces correctly-calibrated RF pulses in most
subjects at approximately the level of the C5/6 intervertebral disc. Anatomical T_1-weighted
sagittal MP2RAGE[48] images at 0.7 mm isotropic resolution were acquired for every subject to
aid in registration to template space.

One single-shot 2D EPI protocol at 0.75 mm in-plane resolution and two multi-shot 2D EPI
protocols at 0.75 mm and 0.6 mm in-plane resolution were used (Figure 1). These three
protocols will be abbreviated SS75, MS75, and MS60, respectively. Each protocol was
optimized independently, rather than attempting to match parameters between protocols.

Research pulse sequences were used for both the single-shot and multi-shot acquisitions. (At
the time of this study, a multi-shot 3D EPI pulse sequence analogous to what was used in
references[12–14] was not available on our scanner platform.) In all protocols, the FOV was
intentionally centered towards the back of the neck to include posterior neck muscles, which,
due to their greater distance from the trachea and esophagus, serve as spatial landmarks that
are relatively robust to swallowing to aid subsequent motion correction.

In the single-shot EPI, a single saturation slab was placed anterior to the spinal cord for outer
volume suppression, thereby reducing the required echo-train length. For in-plane acceleration
using generalized autocalibrating partially parallel acquisition (GRAPPA)[49], 52 lines of
autocalibration signal reference data were acquired using GRE[50], which has been shown to be
optimal for single-shot EPI in the spinal cord at 7 T[51]. Partial Fourier (factor 6/8) acquisition was
judged to be necessary in single-shot EPI to further reduce TE to avoid excessive signal loss
due to through-slice dephasing. Image reconstruction of the single-shot EPI was performed
online with the standard reconstruction pipeline of the manufacturer, including partial Fourier
reconstruction by zero-filling and sine-window filtering, and Nyquist ghost correction based on a
three-line navigator.

The multi-shot protocols were acquired without a spatial saturation slab due to SAR limitations
given the shorter per-shot TR. Extending the volume acquisition time to accommodate spatial
saturation was judged to be an unfavorable tradeoff. Therefore, a larger FOV had to be
acquired in the multi-shot protocols, than in the single-shot protocol, to encompass signal from
the entire neck. Note that phase-encode FOVs cannot be defined to be larger than readout-
encode FOV, and the readout-encode FOV can be enlarged with no SNR or timing penalties.
For this reason, the FOV in the multi-shot protocol was set larger in the readout (right-left)
dimension than in the single-shot protocol to achieve the intended increase in the phase-encode
FOV. Phase-encoding in the left-right dimension is undesirable because the patterns of spatial
distortion in the spinal cord would be asymmetric across the left-right dimension of the spinal
cord. Unlike the single-shot sequence, the custom multi-shot sequence did not allow for
GRAPPA acceleration via the console, therefore the FOV in the phase-encode dimension was
set to half of the readout FOV in the multi-shot protocols to achieve the intended phase-encode
undersampling (Fig. 1). Image reconstruction of the multi-shot acquisitions was performed
offline, with an iterative conjugate gradient SENSE algorithm[52]. The coil sensitivities were
estimated based on a low-resolution gradient-echo acquisition. For SNR-optimal coil
combination, the noise correlation between coils was calculated based on data acquired in vivo

without RF excitation. A three-line navigator was acquired for each shot, which was used for
per-shot frequency offset demodulation to reduce ghosting caused by temporal B_0 field
fluctuations[46]. The frequency offset per shot was estimated based on a magnitude-weighted coil
combination of the phase at the center of k-space in the second navigator echo. For Nyquist
ghost correction, three-line phase reference data (which characterize the gradient waveforms)
were also acquired in a phantom immediately following each session, using the same multi-shot
protocols as performed in vivo. In pilot studies, this approach was observed to be more effective
in reducing Nyquist ghosting than using the three-line navigators from the in vivo data.

	Single-Shot	4-Shot	4-Shot
In-Plane Resolution	0.75 mm	0.75 mm	0.60 mm
Slices and Thickness	24 at 3 mm	24 at 3 mm	24 at 3 mm
TR	1500 ms	746 ms	1100 ms
VAT	1500 ms	2984 ms	4400 ms
TE	20.8 ms	18 ms	26 ms
Nominal Flip Angle	68°	45°	56°
FOV Read	84 mm	128 mm	128 mm
Matrix	112	168	216
Bandwidth	1088 Hz/px	1144 Hz/px	890 Hz/px
Echo Spacing	1.04 ms	1.06 ms	1.31 ms
Effective PE Bandwidth	23 Hz/mm	59 Hz/mm	48 Hz/mm
Lines per Shot	41	21	27
Phase Partial Fourier	6/8	Off	Off
Acceleration	GRAPPA R = 2	SENSE R = 2	SENSE R = 2
Measurements	400	202	137
Saturation Slab	1 Anterior	None	None

Figure 1: Pulse sequence parameters and slice positioning for three fMRI
protocols. Yellow boxes indicate image acquisition fields of view, and green
boxes indicate frequency and shim adjustment volumes.

To compare expected advantages and disadvantages between the protocols, Table 1 lists
performance considerations relevant to BOLD fMRI, and for each of these gives the governing
protocol parameters, and the expected relative ranking of the three candidate protocols. The
relative importance of the different performance considerations, and the holistic ranking of
protocol performance in cervical spinal cord BOLD fMRI at 7 T, is not obvious; the goal of this
study is to determine this experimentally.

Performance Consideration	Governing Parameter(s)	Expected Ranking
BOLD Sensitivity	TE	**MS60** > SS75 > MS75
Signal Dropout	TE	**MS75** > SS75 > MS60
Ghosting	Number of Shots, TE	**SS75** > MS75 > MS60
Partial Fourier Blurring	Partial Fourier Factor	**MS60** ≈ **MS75** > SS75
T2* Decay Blurring	Effective Phase Encode Bandwidth (dk/dt)	**MS75** > MS60 > SS75
Susceptibility-Induced Distortion	Effective Phase Encode Bandwidth (dk/dt)	**MS75** > MS60 > SS75

Table 1: Performance considerations relevant to BOLD fMRI, the protocol
parameters governing performance regarding each image quality or sensitivity
consideration, and the expected performance of the three candidate protocols
relative to each other.

Functional MRI Task Design

For each of the three imaging protocols, a 10-minute fMRI acquisition was performed with
noxious thermal stimulation applied to the lateral surface of the base of the right thumb with a 30
mm x 30 mm flat thermode at a pre-calibrated intensity of 3/10 (range 42-51°C, average 46.5°C)
using an fMRI-compatible thermal stimulator (TSA-II, Medoc, Israel). Pain scale anchors were 0
= no pain and 10 = worst pain imaginable. Stimulus blocks were 10 seconds in duration,
calculated by including ramp-up and plateau, but excluding ramp-down to baseline. Ramp
speed was set to 10°C/sec. Stimulus blocks were separated by pseudo-randomized inter-
stimulus intervals ranging from 25 to 45 seconds, plus initial and final stimulus-free blocks. This
stimulus was expected to produce sensory activation primarily in the ipsilateral dorsal horn at
the neurological C6 (approximately vertebral C4-C5) level[53,54]. Subjects were coached to remain
still and maintain steady respiration during thermal stimuli. Pulse-oximeter and respiratory
traces were acquired (TSD200, PPG100C, TSD201, and RSP100C, Biopac, Goleta, CA, USA).

Data Processing

Spinal cord toolbox v5.4.0[55] was used for image pre-processing. Anatomical MP2RAGE
images were segmented using *sct_deepseg_sc*, and vertebral levels were labeled using
sct_label_vertebrae and/or *sct_label_utils*. Registration to the PAM50 template[56] was performed
using *sct_register_to_template*. Each 4-D fMRI dataset was motion-corrected using
sct_fmri_moco with an inclusive cylindrical mask that contains the posterior neck muscle, the
spinal canal and spinal cord were manually masked in FSLeyes[57], and the fMRI dataset was
registered to the PAM50 template using *sct_register_multimodal*, initialized using the
transformation calculated between the anatomical image and PAM50. Each fMRI dataset was
straightened, spatially smoothed using *sct_smooth_spinalcord* with an anisotropic 2 mm x 2 mm

x 6 mm Gaussian kernel oriented along the spinal cord centerline, and unstraightened, to
minimize blurring of CSF signal into the cord.

First-level GLM analyses was performed in individual-subject space using FSL FEAT[58],
incorporating the default autocorrelation correction using FSL FILM[58] and a 35-term
physiological noise model (PNM)[31,59]. The PNM was generated separately for each slice,
incorporating slice timing correction, using FSL PNM, and included 8 cardiac phase, 8
respiratory phase, 16 cardiac-respiratory interaction, CSF signal derived from the spinal canal
mask (canal mask minus cord mask), and two in-plane translational motion correction terms
from *sct_fmri_moco*.

Group-level GLM analysis was performed in FSL FEAT using a mixed-effects model
(FLAME1). After warping the results of first-level analyses into PAM50 template space, group-
level activation maps were calculated within a search region constructed by eroding the PAM50
cord mask by one voxel. First, we performed thresholding at z > 2.7 followed by cluster-level
correction at a cluster threshold of p = 0.05. This cluster-level correction method was tested
because of its widespread use and familiarity in the field of brain fMRI, but has not yet been
rigorously validated in the spinal cord, and so these results should be interpreted with caution.
Next, because cluster-level correction did not yield any supra-threshold activation for any of the
three protocols, we applied a low, uncorrected threshold of z > 1.64, which corresponds to a
voxel p-value threshold of 0.05. This uncorrected threshold of z > 1.64 was subsequently used
for tabulation of activated voxel counts and visualization in support of the goal of comparing the
three acquisition protocols.

Data Analysis

Temporal signal to noise ratio (tSNR) was calculated voxelwise as the ratio of temporal mean
to standard deviation, then averaged within neurological level C6 in the PAM50 cord, gray
matter, and white matter masks after motion correction, but before smoothing and application of
the PNM. The extent of this neurological C6 mask was defined generously as the region where
the probabilistic mask of C6 contained within SCT[60] is greater than zero. Several additional ROI
masks were defined for further analysis (Figure 2). The PAM50 cord mask, eroded by 1 voxel,
was used as the search region in group-level analyses and to mask parameter maps for
visualization. A mask for expected true-positive activation was constructed by summing the
PAM50 masks for the right dorsal horn, right fasciculus gracilis, right fasciculus cuneatus, and
right lateral corticospinal tract (PAM50 atlas mask entries 1, 3, 5, and 35), binarizing, eroding by
one voxel from the cord-CSF border only, and calculating the intersection of the result with the

aforementioned PAM50 mask for spinal (i.e., neurological) level C6 thresholded at zero. White

matter structures were included to enclose potential BOLD signal in veins draining the expected

gray matter site of activation[61], but the spinal venous plexus on the exterior of the spinal cord,

though expected to also exhibit BOLD activation[62], was excluded due to the strong effects of

CSF signal that cannot be fully accounted for by a PNM and the resulting strong potential for

false-positive activation. A control mask, for comparison purposes, was calculated similarly

using PAM50 masks for the left lateral corticospinal tract, left rubrospinal tract, left spinothalamic

and spinoreticular tracts, and left ventral horn (PAM50 mask entries 4, 8, 12, and 30). These

masks were also warped to each individual subject space. Although studies in rats have shown

that some contralateral ventral horn activation does occur in response to noxious stimuli[63,64], this

was judged to be the most suitable control region in light of the stronger resting-state

connectivity between ipsilateral dorsal and ventral horns, and between ipsilateral and

contralateral dorsal horns, in humans[12,13,65].

Figure 2: Masks used for analysis of task fMRI activation, displayed in PAM50

template space. The PAM50 cord mask, eroded by 1 voxel (blue), was used as

the search region in group-level analyses, and to mask all parameter maps for

visualization. Masks representing regions of interest in the C6 spinal level for

expected true positive activation (green) and control (red) are also displayed.

For the group-level and each individual-subject result, the mean and maximum z-statistic

across voxels with z > 1.64 within the true-positive and control ROI masks were tabulated. The

ratio of the maximum z-statistic in the true-positive ROI mask to the maximum z-statistic in the

control ROI mask in each subject was taken as a measure of the specificity of each protocol to
true-positive activation. Mean BOLD percent signal change (PSC) was calculated in activated
native space voxels in the true-positive region for each subject, and the BOLD percent signal
change in the voxel in the true-positive region having the highest z-statistic in each subject was
also tabulated.

Results

Individual subject-level mean and single-timeframe images and tSNR maps of datasets
representing the best-case and worst-case image quality in the study are shown in Figure 3.
These best-case and worst-case datasets were chosen by first selecting the subsets of cases
with the highest and lowest tSNR, and from those subsets, choosing the dataset with the lowest
and highest artifact burden by visual inspection, respectively. In axial slices near intervertebral
junctions, where residual B_0 inhomogeneity is large, signal loss due to through-slice dephasing
and geometric distortion are evident. Geometric distortion was less severe in multi-shot
acquisitions, which had higher effective phase encoding bandwidth. The single-shot EPI had a
blurrier appearance than the multi-shot acquisitions, while multi-shot 0.60 mm produced the
sharpest appearing images. A spatial gradient in tSNR was also evident, with lower tSNR at
more inferior levels. This is consistent with a combination of reduced B_1^+ efficiency, increased
static B_0 inhomogeneity, and larger time-varying B_0 fluctuations due to the proximity to the
lungs.

Figure 3: Individual subject-level images and temporal signal to noise ratio

(tSNR) maps spanning C4-C7 vertebral levels in datasets representing best and
worst image quality in the study. Temporal SNR was calculated after motion
correction, but before spatial smoothing or application of the physiological noise
model. Locations of the three axial slices are indicated with yellow lines in the
sagittal frames. The first and third displayed axial slices are superior and inferior
to an intervertebral disc, and the second axial slice is located at an intervertebral
disc.

Summary statistics of tSNR are given in Table 2. Temporal SNR was generally highest in
single-shot 0.75 mm and multi-shot 0.75 mm, with lower variability between subjects in multi-
shot 0.75 mm, while tSNR was significantly lower in multi-shot 0.6 mm, yet with lowest between-
subject variability. Temporal SNR was slightly greater in gray matter than in white matter in all
three protocols, driven predominantly by the higher signal in gray matter.

Temporal SNR in Neurological Level C6

	SS75	MS75	MS60
tSNR in Cord	14.6 ± 4.2	14.2 ± 3.3	10.0 ± 2.5
tSNR in GM	15.4 ± 4.2	15.1 ± 3.7	10.8 ± 2.6
tSNR in WM	14.2 ± 4.2	13.7 ± 3.0	9.7 ± 2.5

Table 2: Summary statistics of temporal signal to noise ratio (tSNR)
measurements in data from three protocols. Tabulated means and standard
deviations are calculated as mean and standard deviation of the set of individual
subject-level means within PAM50 masks. NB: standard deviations across
subjects of mean tSNR in the cord, GM, and WM do indeed all round to 4.2;
differences appear at the third decimal place.

Using cluster-level correction for multiple comparisons within the eroded PAM50 cord mask,
no significantly activated voxels were identified in the spinal cord on group level analyses for
any of the three protocols. Maps of z-statistics, thresholded at an uncorrected threshold of z >
1.64, and BOLD percent signal changes in group-level GLM analyses are shown in Figure 4.
Clusters of activation under this uncorrected threshold appear in all three protocols in the
ipsilateral (right) dorsal quadrant at the expected C6 neurological level, which is centered
approximately at the C4-C5 intervertebral disc (see Figure 2 for superior-inferior extent). All
images are displayed in radiological orientation convention. The ipsilateral (right) activation
appears in the same cross-sectional location in single-shot 0.75 mm and multi-shot 0.6 mm;
however, this cluster is weaker and located more peripherally in multi-shot 0.75 mm. In single-
shot 0.75 mm, a cluster also appears in the contralateral (left) dorsal horn.

Figure 4: Maps of z-statistics and BOLD percent signal changes in group-level GLM analyses. Z-statistic maps are displayed using a low, uncorrected threshold of z > 1.64, and as such, should be interpreted with caution. Maps of BOLD percent signal change are thresholded at 0.125% for ease of visualization. Maps are masked to the spinal cord, encompassing all neurological levels, and the slice locations in all three planes are centered on the voxel inside the spinal cord mask with the highest z-statistic (indicated with a green crosshair).

Details of z-statistics and BOLD percent signal change in group-level GLM analyses are given

in Table 3. Voxels within the true-positive and control ROI masks with z > 1.64 are considered "activated" for the purpose of comparisons between protocols, although as stated previously, these tabulated voxels do not survive multiple-comparison correction using cluster-level correction. Mean z-statistics in activated true-positive voxels are relatively consistent in the range of 1.83 to 2.04 across the three protocols, with single-shot 0.75 mm yielding the highest mean z-statistic contrary to initial expectation (Table 1). The maximum true-positive z-statistic is highest in single-shot 0.75 mm at 2.85, compared to 2.07 in multi-shot 0.75 mm and 2.38 in multi-shot 0.6 mm. All three protocols yielded a ratio of true-positive to control max z-statistic between 1.27 and 1.37.

The mean BOLD percent signal change in activated voxels was higher in multi-shot datasets than in single-shot. Multi-shot 0.75 mm yielded the highest mean BOLD percent signal change, at 0.43%, and the BOLD percent signal change in the voxel with the highest z-statistic was 0.50%.

Group-Level Activation Results

Threshold for Significance is Z > 1.64	SS75	MS75	MS60
Mean Z-Stat in Activated Voxels in True Positive Region	2.04	1.83	1.84
Max Z-Stat in True Positive Region	2.85	2.07	2.38
Max Z-Stat in Control Region	2.24	1.51	1.83
Ratio of Max Z-Stat in True Positive vs. Control	1.27	1.37	1.3
% of Voxels Activated in True Positive Region	2.76%	0.71%	2.34%
% of Voxels Activated in Control Region	2.20%	0.00%	0.37%
Ratio of % of Voxels Activated in True Positive vs. Control	1.25	∞	6.36
Mean BOLD PSC in Activated Voxels	0.21	0.43	0.33
BOLD PSC in Voxel w/ Highest Z-Stat	0.22	0.5	0.38
Mean BOLD PSC in Voxels w/ Top 10 %ile Z-Stats	0.19	0.26	0.33

Table 3: Summary statistics of group-level GLM results. An uncorrected threshold of z > 1.64 was used to tabulate activated voxel counts for the purpose of comparing acquisition protocols; therefore, these results should be interpreted with caution.

To estimate the number of required subjects to achieve a supra-threshold cluster using stringent cluster-level correction methods, we performed additional group-level analyses with cluster-level correction as described in the methods section, on datasets containing replicates of the set of 11 individual subject-level datasets. Based on these analyses, we estimate that using thresholding at z > 2.7 followed by cluster-level correction at a cluster threshold of p = 0.05, the required number of subjects is estimated to be between 23 and 33 in single-shot 0.75 mm,

between 177 and 187 in multi-shot 0.75 mm, and between 34 and 44 in multi-shot 0.60 mm.

Details of z-statistics and BOLD percent signal change across all individual-level GLM
analyses are given in Table 4. As in group-level analyses, voxels within the true-positive and
control ROI masks (warped to each individual subject space) with z > 1.64 are considered
activated. Mean z-statistics in activated true-positive voxels are relatively consistent in the range
of 1.96 to 2.18 across the three protocols. Single-shot 0.75 mm yields the highest mean z-
statistic, but the standard deviation of each subject's mean z-statistic in activated true-positive
voxels is double that of the multi-shot protocols. Single-shot 0.75 mm also yields the highest
mean maximum z-statistic at 2.68, compared to 2.27 for multi-shot 0.75 mm and 2.45 for multi-
shot 0.6 mm. Single-shot 0.75 mm and multi-shot 0.6 mm have true-positive-to-control ratios
substantially greater than unity, while the ratio in multi-shot 0.75 mm is closer to unity.

Similar to group-level analyses, the mean BOLD percent signal change in activated voxels
was higher in multi-shot protocols. Multi-shot 0.6 mm yielded the highest mean BOLD percent
signal change, at 1.78% ± 0.89%, followed by multi-shot 0.75 mm at 1.29% ± 0.76%, and
single-shot 0.75 mm at 0.67% ± 0.31%. The same relationship among the protocols is evident in
the BOLD percent signal change in the voxel with the highest z-statistic in each dataset.

Individual-Level Activation Results

Threshold for Significance is Z > 1.64	SS75	MS75	MS60
Mean Z-Stat in Activated Voxels in True Positive Region	2.18 ± 0.37	1.99 ± 0.17	1.96 ± 0.19
Max Z-Stat in True Positive Region	2.68 ± 1.41	2.27 ± 0.77	2.45 ± 0.55
Max Z-Stat in Control Region	2.27 ± 1.10	2.14 ± 0.76	2.11 ± 0.78
Ratio of Max Z-Stat in True Positive vs. Control	1.18	1.06	1.16
% of Voxels Activated in True Positive Region	10.68% ± 13.94%	5.14% ± 5.55%	5.01% ± 3.57%
% of Voxels Activated in Control Region	6.91% ± 10.01%	3.53% ± 4.37%	5.07% ± 6.43%
Ratio of % of Voxels Activated in True Positive vs. Control	1.55	1.46	0.99
Mean BOLD PSC in Activated Voxels	0.67 ± 0.31	1.29 ± 0.76	1.78 ± 0.89
BOLD PSC in Voxel w/ Highest Z-Stat	0.88 ± 0.84	1.51 ± 1.07	2.11 ± 1.18
Mean BOLD PSC in Voxels w/ Top 10 %ile Z-Stats	0.59 ± 0.50	1.02 ± 0.75	1.52 ± 0.78

Table 4: Summary statistics of individual subject GLM analyses.   An
uncorrected threshold of z > 1.64 was used to tabulate activated voxel counts for
the purpose of comparing acquisition protocols; therefore, these results should
be interpreted with caution. In individual subjects, mean z-statistics and mean
BOLD percent signal change (PSC) are calculated within voxels having z > 1.64,
as well as BOLD PSC in the voxel having the highest z-statistic and in the set of
voxels having z-statistics in the top 10[th] percentile. Maximum z-statistics in true
positive and control regions are also calculated in individual subjects without a

threshold on z-statistic. Means and standard deviations across these individual-
subject mean and maximum values are then calculated and tabulated here.

Discussion

Image quality and tSNR are strongly impaired in spinal cord fMRI by static focal B_0
inhomogeneity near the intervertebral junctions and temporally variable B_0 inhomogeneity
arising from the lungs. These two sources of B_0 inhomogeneity were clearly evident in the
images and tSNR maps shown in Figure 3. Near the intervertebral junctions, tSNR was sharply
reduced and geometric distortion and signal loss were markedly more severe than in slices
further from a junction. The gradual reduction in tSNR in lower slices was partially due to static
and dynamic B_0 inhomogeneity from the lungs, but reduced B_1^+ efficiency in the lower cervical
spine was also a significant contributor[47]; signal intensity in mean and single images was lower
adjacent to vertebral level C7, compared to C4-C6, even in the best-quality dataset.

In general, the single-shot 0.75 mm protocol provided the greatest sensitivity (albeit highest
between-subject variability in z-statistics) to the expected activation, but also appeared to have
the lowest effective in-plane resolution and largest geometric distortion. This was in line with
expectations (Table 1), as the lower effective phase-encode bandwidth makes it more
susceptible to T_2^* blurring and geometric distortion due to static B_0 inhomogeneity, on top of
blurring due to partial Fourier acquisition. Multi-shot 0.6 mm produced the most well-localized
activation clusters, and the sharpest images, while being qualitatively less affected by geometric
distortion caused by static B_0 inhomogeneity than single-shot 0.75 mm. However, sensitivity to
activation was reduced due to its longer VAT, which reduced the number of time points in an
experiment of fixed duration, and the lower tSNR, both of which reduce the statistical power.
Multi-shot 0.75 mm generally performed poorly; it neither showed the sensitivity to activation of
the single-shot 0.75 mm protocol nor the spatial specificity of the multi-shot 0.60 mm protocol,
as if sharing the drawbacks of each of the other protocols, without enjoying the benefits of either
protocol to a similar degree.

Decades of results in the brain have consistently shown that longer TEs (i.e., close to the T_2^*
of gray matter) are associated with greater sensitivity to activation. In the spinal cord at 7 T,
however, the detrimental effects of B_0 inhomogeneity severely impair data quality at longer TEs.
A balance must be struck between image quality, which favors shorter TE, and sensitivity to
BOLD signal, which favors TE close to T_2^* of gray matter, which is 29.3 ± 4.5 ms at 7 T[66]. The
optimal TE for a given spinal cord fMRI pulse sequence will therefore likely be shorter than the

T_2^* of gray matter.

In group-level GLM analyses, single-shot 0.75 mm and multi-shot 0.6 mm appeared to yield

similarly-sized activation clusters, and the maximum z-statistic in single-shot 0.75 mm was

substantially greater than the maximum z-statistic in multi-shot 0.6 mm (see Figure 4). In

individual-level results, however, multi-shot 0.6 mm generally yielded more sharply-defined

activation clusters in the ipsilateral dorsal quadrant (see Figure S1), and the difference between

the maximum z-statistic in single-shot 0.75 mm and the maximum z-statistic in multi-shot 0.6

mm was much smaller than in the group level analysis. A possible explanation that unifies these

observations is that multi-shot 0.6 mm was better able to spatially resolve activation in individual

subjects, resulting in small clusters; however, when multiple subjects' data was combined, the

inherent physiological variability in the location of activation among different subjects[60], as well

as potential limitations in group-level registration of distorted spinal cord EPI images, resulted in

these sharply-defined individual-level activation clusters spatially coinciding less fully in multi-

shot 0.6 mm than did the broader activation clusters in single-shot 0.75 mm. These broader

activation clusters in single-shot EPI may have arisen due to its greater statistical power, having

more temporal frames due to a shorter volume acquisition time and higher tSNR due to the

lower effective resolution.

As stated previously, the noxious thermal stimulus was expected to produce sensory activation

in the ipsilateral dorsal horn at the neurological C6 (approximately vertebral C4-C5) level based

on the fact that the majority of first-order neurons in the lateral spinothalamic tract synapse in

the ipsilateral dorsal horn[53,54]. Indeed, virtually all previous spinal cord fMRI studies of noxious

thermal stimulation at lower field strengths or in animals have observed the strongest activation

in the ipsilateral dorsal horn. However, the lateral spinothalamic tract does send some

projections to mainly interneurons in the contralateral hemicord as well[67–70], so it is highly likely

that relatively weaker activation observed in the contralateral dorsal horn is not spurious but is,

in fact, physiologically plausible and expected, and this has been observed in prior spinal cord

fMRI work[15,18,71–74].

Based on TE, BOLD percent signal change would be expected to be highest in multi-shot 0.60

mm, followed by single-shot 0.75 mm and then multi-shot 0.75 mm. In individual-level results,

multi-shot 0.60 mm did indeed yield the highest BOLD percent signal change. However, of the

remaining two protocols, multi-shot 0.75 mm yielded higher BOLD percent signal change than

single-shot 0.75 mm. One potential explanation for this is that the effect of spatial resolution

outweighed the effect of the magnitude of the underlying BOLD activation in these individual-

level measurements; the multi-shot 0.75 mm acquisition is expected to better resolve BOLD
activation than the single-shot 0.75 mm acquisition. An alternative explanation specifically
concerning the difference in the mean BOLD percent signal change in activated voxels between
the multi-shot 0.75 mm and single-shot 0.75 mm acquisitions is that the single-shot 0.75 mm
acquisition, which contained 400 temporal frames, was statistically better powered to detect low-
BOLD-percent-signal-change activation than the multi-shot 0.75 mm acquisition, which
contained only 202 temporal frames and may have been additionally hampered by greater
physiologic noise contributions (e.g., ghosting). In other words, a voxel with a low BOLD percent
signal change is more likely to be identified as significantly activated using a protocol with a
larger number of temporal frames and less susceptibility to ghosting. This would result in a
greater number of low-BOLD-percent-signal-change voxels being included in the set of
significantly activated voxels in the single-shot 0.75 mm acquisition, decreasing the mean BOLD
percent signal change calculated for that set of voxels. However, when the mean BOLD percent
signal change is calculated for each protocol in the set of voxels having z-statistics in the top
10[th] percentile within each individual subject, BOLD percent signal change remains lower in the
single-shot 0.75 mm acquisition than in the multi-shot 0.75 mm acquisition. In the same
percentile-based set of voxels in group-level results, single-shot 0.75 mm likewise has the
lowest mean BOLD percent signal change. In light of these results, the first potential
explanation based on spatial resolution may be more likely.

The superior-inferior location of peak activation was similar across the three protocols in
group-level analyses, and appeared at the expected location in the center of neurological C6
(determined from the probabilistic spinal levels from the PAM50 atlas), based on the anatomical
location of the thermal stimulus. The superior-inferior location of the peak activation was
expected to vary across subjects due to natural physiological variation[60], but was not expected
to vary among the three protocols within each subject. Contrary to this expectation, we did
observe substantial variation in the superior-inferior location of peak activation of up to 1 ½
vertebral levels across the three protocols within individual subjects (see Figure S1). The cause
for this variation is unclear without conducting a reproducibility study by repeating all three
protocols on the same participants. If the true underlying stimulus-related activation occurs
across a large spatial extent in the superior-inferior dimension, then random noise as well as
variations in physiological noise may shift the center of the detected activation between
acquisitions in an unsystematic way, as observed. In that case, increasing the sensitivity to
detect BOLD responses, e.g. by acquiring for longer, should yield extended activation clusters.
The true underlying activation is however unknown, and the data in this work cannot

convincingly support any conclusion. From resting-state studies, we would expect networks of
correlated activity stretching over approximately one vertebral level[65]. Task-activated networks,
and patterns of nociceptive activation in particular may, however, have a broader spatial extent
to resting-state networks due to the fact that primary afferent projections may ascend or
descend in Lissauer's tract before synapsing[67,75,76]. A large spatial extent of the activated
networks in the superior-inferior dimension may even be a prerequisite to detect activation
clusters in group-level analyses with relatively few subjects. If the activation were narrowly
localized within one subject, with a variable center along the vertebral levels between subjects,
the statistical power to detect activation at the group level would be much diluted. Furthermore,
it has been shown in the cerebral cortex that significant variation in activation maps exists
across sessions, even in a single subject[77], and when evaluating activation based on
thresholded single-subject activation maps, even consistent activation can appear highly
variable when the significance of this activation is near the defined threshold for visualization[78].
Further studies would be needed to investigate the true extent of the activation, and the
reproducibility of individual-level activation clusters.

Several groups have developed methods to address distortion and through-slice dephasing in
single-shot GRE EPI at 3 T through slicewise dynamic shimming methods using either slice-
specific gradient offsets[39] or z-gradient refocusing blips[79,80]. These methods, applied to single-
shot EPI, may be promising alternatives to multi-shot 2D or 3D EPI at 7 T but, at the time of this
study, were not sufficiently refined to implement routinely at 7 T. Also 3D multi-shot EPI
sequences suitable for implementation in fMRI experiments were not readily available at the
time of this study, despite encouraging results demonstrating their robustness to B_0
inhomogeneity in resting-state fMRI studies at 7 T[12,13]. Such sequences remain a highly
promising option worthy of future investigation. This study also used different reconstruction
methods for single-shot and multi-shot EPI: on-scanner GRAPPA and offline SENSE,
respectively. An ideal solution would be to reconstruct single-shot EPI data offline using the
same SENSE method as the multi-shot EPI data; however, due to data size and transfer rate
limitations, we were not able to save raw data for single-shot fMRI for all subjects.

This study is limited by its small sample size of 11 healthy subjects. Because of this, no
significantly activated voxels were identified in the spinal cord on group level analyses using
cluster-level correction for multiple comparisons. To enable visual and quantitative comparisons
between protocols, maps and tables used a low, uncorrected threshold of z > 1.64. Future
studies in a larger number of subjects would be needed to achieve significance under

appropriately stringent thresholding methods.

It should also be noted that the BOLD percent signal change maps displayed in Figures 4, S1,
and S2 display the amplitude of the BOLD signal but do not provide information on the variability
of the BOLD signal, and so they should be interpreted with appropriate caution, and in
combination with the accompanying activation z-score maps.

568 Conclusions

This study provides, to our knowledge, the first stimulus task fMRI results achieved in the
spinal cord at 7 T. For a given scan time, single-shot 0.75 mm was most sensitive to the
expected activation, likely due to its shorter volume acquisition time and greater statistical
power, while multi-shot 0.60 mm provided the best-localized activation clusters and the least
geometric distortion. The best choice of protocol will likely depend on the relative importance of
sensitivity to activation versus spatial localization of activation for a given experiment.

575 Acknowledgements

This study was supported by National Institutes of Health (NINDS) award K01NS105160
(ACS) and Department of Defense MSRP grant W81XWH-17-1-0204 (JX). FE is supported by
the Max Planck Society and the European Research Council (under the European Union's
Horizon 2020 research and innovation programme; grant agreement No 758974). The work was
also supported by the Wellcome Trust (Strategic Award 102645/Z/13/Z, Senior Research
Fellowship 202788/Z/16/Z). The Wellcome Centre for Integrative Neuroimaging is supported by
core funding from the Wellcome Trust (203139/Z/16/Z). Data in this study was acquired using
the University of Minnesota Center for Magnetic Resonance Research (CMRR) Multiband EPI
pulse sequence software package.

The authors thank Dr. David O'Connor for helpful guidance on sample size estimation under
cluster-level correction.

Portions of this study in smaller numbers of subjects have previously been presented as
conference abstracts[61,81]. The data that support the findings of this study are available from the
corresponding author upon reasonable request.

Supplementary Materials

Individual Subject-Level Activation Maps

Maps of z-statistics and BOLD percent signal changes in two representative individual-level GLM analyses are shown in Figure S1 (best-quality dataset) and Figure S2 (worst-quality dataset). Z-statistic maps are thresholded at an uncorrected threshold of $z > 1.64$.

In the best-quality dataset in Figure S1, clusters of activation appear in the ipsilateral dorsal horn in all three protocols. The superior-inferior location of the peak activated cluster in the three protocols varies within this single subject to a greater extent than the expected size of a spinal segment in a single subject, although all three locations are contained within the group-level probabilistic PAM50 atlas entry for the C6 neurological level. In single-shot 0.75 mm, the cluster extends into the ipsilateral ventral horn as well, and a smaller cluster appears in the contralateral deep dorsal horn. In multi-shot 0.75, a disconnected cluster of weak activation also appears in the ipsilateral ventral horn. The cluster in multi-shot 0.6 mm is intense and well-localized, coinciding with a cluster of high BOLD percent signal change approaching 2%. In the displayed slices, one likely false-positive voxel is visible at the contralateral cord-CSF boundary in the multi-shot 0.60 mm, and the contralateral cluster in single-shot 0.75 mm is neither clearly true-positive nor false-positive, but for the purpose of this analysis, a portion of it falls within the control mask ROI, and is treated as such in descriptive statistics.

Figure S1: Individual subject-level maps of activation z-statistic and BOLD percent signal change, warped into template space, in the "best quality" dataset from Figure 2. Z-statistic maps are displayed using a low, uncorrected threshold of z > 1.64, and as such, should be interpreted with caution. Slice locations in all three planes are centered on the voxel with the highest z-statistic (indicated with a green crosshair).

In the worst-quality dataset in Figure S2, potentially plausible true-positive activation is visible in the ipsilateral dorsal quadrant in the multi-shot 0.6 mm protocol, though this cluster extends

partially out of the superior boundary of the neurological C6 level. No activation exceeding the threshold of z = 1.64 is visible in either the single-shot or multi-shot 0.75 mm protocols. The location of the peak (although still sub-threshold) z-statistic within the true-positive mask ROI is indicated with a green crosshair in the axial plane, and yellow lines in the sagittal and coronal planes.

Figure S2: Individual subject-level activation maps, warped into template space, in the "worst quality" dataset from Figure 2. Z-statistic maps are displayed using a low, uncorrected threshold of z > 1.64, and as such, should be interpreted with

627   caution. Images are centered at the voxel in the true positive region with the

628   highest z-statistic (indicated with a green crosshair), but in single-shot 0.75 mm

629   and 4-shot 0.75 mm, this z-statistic is below the threshold of 1.64.

References

1. Minassian K, Hofstoetter US, Dzeladini F, Guertin PA, Ijspeert A. The Human Central
Pattern Generator for Locomotion: Does It Exist and Contribute to Walking?
Neuroscientist. 2017;23(6):649-663. doi:10.1177/1073858417699790

2. Bingel U, Tracey I. Imaging CNS modulation of pain in humans. *Physiology (Bethesda)*.
2008;23(6):371-380. doi:10.1152/physiol.00024.2008

3. Raj D, Anderson AW, Gore JC. Respiratory effects in human functional magnetic resonance
imaging due to bulk susceptibility changes. *Phys Med Biol*. 2001;46(12):3331-3340.
doi:10.1088/0031-9155/46/12/318

4. Vannesjo SJ, Miller KL, Clare S, Tracey I. Spatiotemporal characterization of breathing-
induced B0 field fluctuations in the cervical spinal cord at 7T. *NeuroImage*. 2017;167:191-
202. doi:10.1016/j.neuroimage.2017.11.031

5. Krüger G, Glover GH. Physiological noise in oxygenation-sensitive magnetic resonance
imaging: Physiological Noise in MRI. *Magn Reson Med*. 2001;46(4):631-637.
doi:10.1002/mrm.1240

6. Cohen-Adad J. Functional Magnetic Resonance Imaging of the Spinal Cord: Current Status
and Future Developments. *Semin Ultrasound CT MR*. 2017;38(2):176-186.
doi:10.1053/j.sult.2016.07.007

7. Yacoub E, Shmuel A, Pfeuffer J, et al. Imaging brain function in humans at 7 Tesla. *Magn
Reson Med*. 2001;45(4):588-594.

8. Gati JS, Menon RS, Ugurbil K, Rutt BK. Experimental determination of the BOLD field
strength dependence in vessels and tissue. *Magn Reson Med*. 1997;38(2):296-302.
doi:10.1002/mrm.1910380220

9. Triantafyllou C, Hoge RD, Krueger G, et al. Comparison of physiological noise at 1.5 T, 3 T
and 7 T and optimization of fMRI acquisition parameters. *NeuroImage*. 2005;26(1):243-
250. doi:10.1016/j.neuroimage.2005.01.007

10. Stroman PW, Wheeler-Kingshott C, Bacon M, et al. The current state-of-the-art of spinal
cord imaging: methods. *NeuroImage*. 2014;84:1070-1081.
doi:10.1016/j.neuroimage.2013.04.124

11. Barry RL, Vannesjo SJ, By S, Gore JC, Smith SA. Spinal cord MRI at 7T. *NeuroImage*.
2018;168:437-451. doi:10.1016/j.neuroimage.2017.07.003

12. Barry RL, Smith SA, Dula AN, Gore JC. Resting state functional connectivity in the human
spinal cord. *Elife*. 2014;3:e02812. doi:10.7554/eLife.02812

13. Barry RL, Rogers BP, Conrad BN, Smith SA, Gore JC. Reproducibility of resting state spinal
cord networks in healthy volunteers at 7 Tesla. *NeuroImage*. 2016;133:31-40.
doi:10.1016/j.neuroimage.2016.02.058

14. Conrad BN, Barry RL, Rogers BP, et al. Multiple sclerosis lesions affect intrinsic functional

connectivity of the spinal cord. *Brain*. April 2018. doi:10.1093/brain/awy083

15. Weber KA, Chen Y, Wang X, Kahnt T, Parrish TB. Functional magnetic resonance imaging
of the cervical spinal cord during thermal stimulation across consecutive runs.
NeuroImage. 2016;143:267-279. doi:10.1016/j.neuroimage.2016.09.015

16. Brooks J, Tracey I. From nociception to pain perception: imaging the spinal and supraspinal
pathways. *J Anat*. 2005;207(1):19-33. doi:10.1111/j.1469-7580.2005.00428.x

17. Eippert F, Finsterbusch J, Bingel U, Buchel C. Direct evidence for spinal cord involvement in
placebo analgesia. *Science*. January 2009. doi:10.1126/science.1067176)

18. Geuter S, Büchel C. Facilitation of pain in the human spinal cord by nocebo treatment. *J
Neurosci*. 2013;33(34):13784-13790. doi:10.1523/JNEUROSCI.2191-13.2013

19. Tinnermann A, Geuter S, Sprenger C, Finsterbusch J, Buchel C. Interactions between brain
and spinal cord mediate value effects in nocebo hyperalgesia. *Science*.
2017;358(6359):105-108. doi:10.1126/science.aan1221

20. Staud R, Boissoneault J, Lai S, et al. Spinal Cord Neural Activity of Fibromyalgia Patients
and Healthy Controls During Temporal Summation of Pain: An fMRI Study. *Journal of
Neurophysiology*. August 2021:jn.00276.2021. doi:10.1152/jn.00276.2021

21. Yoshizawa T, Nose T, Moore GJ, Sillerud LO. Functional magnetic resonance imaging of
motor activation in the human cervical spinal cord. *NeuroImage*. 1996;4(3 Pt 1):174-182.
doi:10.1006/nimg.1996.0068

22. Stroman PW, Nance PW, Ryner LN. BOLD MRI of the human cervical spinal cord at 3 tesla.
Magn Reson Med. 1999;42(3):571-576. doi:10.1002/(sici)1522-
2594(199909)42:3<571::aid-mrm20>3.0.co;2-n

23. Backes WH, Mess WH, Wilmink JT. Functional MR Imaging of the Cervical Spinal Cord by
Use of Median Nerve Stimulation and Fist Clenching. 2001:6.

24. Stroman PW, Ryner LN. Functional MRI of motor and sensory activation in the human
spinal cord. *Magnetic Resonance Imaging*. 2001;19(1):27-32. doi:10.1016/S0730-
725X(01)00226-0

25. Stroman PW, Krause V, Malisza KL, Frankenstein UN, Tomanek B. Functional magnetic
resonance imaging of the human cervical spinal cord with stimulation of different sensory
dermatomes. *Magnetic Resonance Imaging*. 2002;20(1):1-6. doi:10.1016/S0730-
725X(02)00468-X

26. Stroman P. Mapping of Neuronal Function in the Healthy and Injured Human Spinal Cord
with Spinal fMRI. *NeuroImage*. 2002;17(4):1854-1860. doi:10.1006/nimg.2002.1305

27. Stroman PW, Krause V, Malisza KL, Frankenstein UN, Tomanek B. Extravascular proton-
density changes as a non-BOLD component of contrast in fMRI of the human spinal cord.
Magn Reson Med. 2002;48(1):122-127. doi:10.1002/mrm.10178

28. Lawrence JM, Stroman PW, Kollias SS. Functional magnetic resonance imaging of the

human spinal cord during vibration stimulation of different dermatomes. *Neuroradiology*. 2008;50(3):273-280. doi:10.1007/s00234-007-0338-6

29. Govers N, Béghin J, Van Goethem JWM, et al. Functional MRI of the cervical spinal cord on 1.5 T with fingertapping: to what extent is it feasible? *Neuroradiology*. 2007;49(1):73-81. doi:10.1007/s00234-006-0162-4

30. Giulietti G, Giove F, Garreffa G, Colonnese C, Mangia S, Maraviglia B. Characterization of the functional response in the human spinal cord: Impulse-response function and linearity. *NeuroImage*. 2008;42(2):626-634. doi:10.1016/j.neuroimage.2008.05.006

31. Brooks JCW, Beckmann CF, Miller KL, et al. Physiological noise modelling for spinal functional magnetic resonance imaging studies. *NeuroImage*. 2008;39(2):680-692. doi:10.1016/j.neuroimage.2007.09.018

32. Piché M, Cohen-Adad J, Nejad MK, et al. Characterization of cardiac-related noise in fMRI of the cervical spinal cord. *Magnetic Resonance Imaging*. 2009;27(3):300-310. doi:10.1016/j.mri.2008.07.019

33. Eippert F, Kong Y, Winkler AM, et al. Investigating resting-state functional connectivity in the cervical spinal cord at 3T. *NeuroImage*. 2016;147:589-601. doi:10.1016/j.neuroimage.2016.12.072

34. Weber KA, Chen Y, Wang X, Kahnt T, Parrish TB. Lateralization of cervical spinal cord activity during an isometric upper extremity motor task with functional magnetic resonance imaging. *NeuroImage*. 2016;125:233-243. doi:10.1016/j.neuroimage.2015.10.014

35. Weber KA, Chen Y, Paliwal M, et al. Assessing the Spatial Distribution of Cervical Spinal Cord Activity during Tactile Stimulation of the Upper Extremity in Humans with Functional Magnetic Resonance Imaging. *NeuroImage*. May 2020:116905. doi:10.1016/j.neuroimage.2020.116905

36. Barry RL, Conrad BN, Maki S, et al. Multi-shot acquisitions for stimulus-evoked spinal cord BOLD fMRI. *Magn Reson Med*. November 2020:mrm.28570. doi:10.1002/mrm.28570

37. Sprenger C, Eippert F, Finsterbusch J, Bingel U, Rose M, Büchel C. Attention modulates spinal cord responses to pain. *Curr Biol*. 2012;22(11):1019-1022. doi:10.1016/j.cub.2012.04.006

38. Weber KA, Sentis AI, Bernadel-Huey ON, et al. Thermal Stimulation Alters Cervical Spinal Cord Functional Connectivity in Humans. *Neuroscience*. 2017;369:40-50. doi:10.1016/j.neuroscience.2017.10.035

39. Islam H, Law CSW, Weber KA, Mackey SC, Glover GH. Dynamic per slice Shimming for Simultaneous Brain and Spinal Cord fMRI. *Magn Reson Med*. October 2018. doi:10.1002/mrm.27388

40. Finsterbusch J, Sprenger C, Büchel C. Combined T2*-weighted measurements of the human brain and cervical spinal cord with a dynamic shim update. *NeuroImage*. 2013;79:153-161. doi:10.1016/j.neuroimage.2013.04.021

41. Chu Y, Fricke B, Finsterbusch J. Improving T2*-weighted human cortico-spinal acquisitions
with a dedicated algorithm for region-wise shimming. *NeuroImage*. 2023;268:119868.
doi:10.1016/j.neuroimage.2023.119868

42. Vahdat S, Khatibi A, Lungu O, et al. Resting-state brain and spinal cord networks in humans
are functionally integrated. *PLoS Biol*. 2020;18(7):e3000789.
doi:10.1371/journal.pbio.3000789

43. Vahdat S, Lungu O, Cohen-Adad J, Marchand-Pauvert V, Benali H, Doyon J. Simultaneous
Brain-Cervical Cord fMRI Reveals Intrinsic Spinal Cord Plasticity during Motor Sequence
Learning. 2015;13(6):e1002186. doi:10.1371/journal.pbio.1002186

44. Sprenger C, Finsterbusch J, Büchel C. Spinal cord-midbrain functional connectivity is
related to perceived pain intensity: a combined spino-cortical FMRI study. *J Neurosci*.
2015;35(10):4248-4257. doi:10.1523/JNEUROSCI.4897-14.2015

45. Barry RL, Strother SC, Gatenby JC, Gore JC. Data-driven optimization and evaluation of 2D
EPI and 3D PRESTO for BOLD fMRI at 7 Tesla: I. Focal coverage. *NeuroImage*.
2011;55(3):1034-1043. doi:10.1016/j.neuroimage.2010.12.086

46. Vannesjo SJ, Clare S, Kasper L, Tracey I, Miller KL. A method for correcting breathing-
induced field fluctuations in T2*-weighted spinal cord imaging using a respiratory trace.
Magn Reson Med. February 2019. doi:10.1002/mrm.27664

47. Zhang B, Seifert AC, Kim JW, Borrello J, Xu J. 7 Tesla 22-channel wrap-around coil array
for cervical spinal cord and brainstem imaging. *Magn Reson Med*. 2017;78(4):1623-1634.
doi:10.1002/mrm.26538

48. Marques JP, Kober T, Krueger G, van der Zwaag W, Van de Moortele PF, Gruetter R.
MP2RAGE, a self bias-field corrected sequence for improved segmentation and T1-
mapping at high field. *NeuroImage*. 2010;49(2):1271-1281.
doi:10.1016/j.neuroimage.2009.10.002

49. Griswold MA, Jakob PM, Heidemann RM, et al. Generalized autocalibrating partially parallel
acquisitions (GRAPPA). *Magn Reson Med*. 2002;47(6):1202-1210.
doi:10.1002/mrm.10171

50. Talagala SL, Sarlls JE, Liu S, Inati SJ. Improvement of temporal signal-to-noise ratio of
GRAPPA accelerated echo planar imaging using a FLASH based calibration scan. *Magn
Reson Med*. 2016;75(6):2362-2371. doi:10.1002/mrm.25846

51. Seifert AC, Xu J. Impact of autocalibration method on accelerated EPI of the cervical spinal
cord at 7 T. *Magnetic Resonance in Med*. 2022;88(6):2583-2591. doi:10.1002/mrm.29415

52. Pruessmann KP, Weiger M, Börnert P, Boesiger P. Advances in sensitivity encoding with
arbitrary k-space trajectories: SENSE With Arbitrary k-Space Trajectories. *Magn Reson
Med*. 2001;46(4):638-651. doi:10.1002/mrm.1241

53. Keegan JJ, Garrett FD. The segmental distribution of the cutaneous nerves in the limbs of
man. *Anat Rec*. 1948;102(4):409-437. doi:10.1002/ar.1091020403

54. Lee MWL, McPhee RW, Stringer MD. An evidence-based approach to human dermatomes.
Clin Anat. 2008;21(5):363-373. doi:10.1002/ca.20636

55. De Leener B, Lévy S, Dupont SM, et al. SCT: Spinal Cord Toolbox, an open-source
software for processing spinal cord MRI data. *NeuroImage*. 2017;145(Pt A):24-43.
doi:10.1016/j.neuroimage.2016.10.009

56. De Leener B, Fonov VS, Collins DL, Callot V, Stikov N, Cohen-Adad J. PAM50: Unbiased
multimodal template of the brainstem and spinal cord aligned with the ICBM152 space.
NeuroImage. 2017;165:170-179. doi:10.1016/j.neuroimage.2017.10.041

57. Jenkinson M, Beckmann CF, Behrens TEJ, Woolrich MW, Smith SM. FSL. *Neuroimage*.
2012;62(2):782-790. doi:10.1016/j.neuroimage.2011.09.015

58. Woolrich MW, Ripley BD, Brady M, Smith SM. Temporal autocorrelation in univariate linear
modeling of FMRI data. *Neuroimage*. 2001;14(6):1370-1386. doi:10.1006/nimg.2001.0931

59. Kong Y, Jenkinson M, Andersson J, Tracey I, Brooks JCW. Assessment of physiological
noise modelling methods for functional imaging of the spinal cord. *NeuroImage*.
2012;60(2):1538-1549. doi:10.1016/j.neuroimage.2011.11.077

60. Cadotte DW, Cadotte A, Cohen-Adad J, et al. Characterizing the location of spinal and
vertebral levels in the human cervical spinal cord. *AJNR Am J Neuroradiol*.
2015;36(4):803-810. doi:10.3174/ajnr.A4192

61. Seifert AC, Vannesjo SJ. Spatial Specificity of BOLD Signal in the Spinal Cord at 7T Using a
Noxious Thermal Stimulus. *Proceedings of the International Society for Magnetic
Resonance in Medicine*. August 2020:1168.

62. Cohen-Adad J, Gauthier CJ, Brooks JCW, et al. BOLD signal responses to controlled
hypercapnia in human spinal cord. *NeuroImage*. 2010;50(3):1074-1084.
doi:10.1016/j.neuroimage.2009.12.122

63. Coghill RC, Price DD, Hayes RL, Mayer DJ. Spatial distribution of nociceptive processing in
the rat spinal cord. *Journal of Neurophysiology*. 1991;65(1):133-140.
doi:10.1152/jn.1991.65.1.133

64. Coghill RC, Mayer DJ, Price DD. The roles of spatial recruitment and discharge frequency in
spinal cord coding of pain: a combined electrophysiological and imaging investigation.
Pain. 1993;53(3):295-309. doi:10.1016/0304-3959(93)90226-F

65. Kong Y, Eippert F, Beckmann CF, et al. Intrinsically organized resting state networks in the
human spinal cord. *Proc Natl Acad Sci USA*. 2014;111(50):18067-18072.
doi:10.1073/pnas.1414293111

66. Massire A, Taso M, Besson P, Guye M, Ranjeva JP, Callot V. High-resolution multi-
parametric quantitative magnetic resonance imaging of the human cervical spinal cord at
7T. *NeuroImage*. 2016;143:58-69. doi:10.1016/j.neuroimage.2016.08.055

67. Light AR, Perl ER. Spinal termination of functionally identified primary afferent neurons with
slowly conducting myelinated fibers. *J Comp Neurol*. 1979;186(2):133-150.

doi:10.1002/cne.901860203

68. Willis WD, Westlund KN. Neuroanatomy of the pain system and of the pathways that
modulate pain. *J Clin Neurophysiol*. 1997;14(1):2-31. doi:10.1097/00004691-199701000-
00002

69. Sotgiu ML, Brambilla M, Valente M, Biella GEM. Contralateral input modulates the
excitability of dorsal horn neurons involved in noxious signal processes. Potential role in
neuronal sensitization. *Somatosensory & Motor Research*. 2004;21(3-4):211-215.
doi:10.1080/08990220400012539

70. Wang X, Zhang J, Eberhart D, et al. Excitatory Superficial Dorsal Horn Interneurons Are
Functionally Heterogeneous and Required for the Full Behavioral Expression of Pain and
Itch. *Neuron*. 2013;78(2):312-324. doi:10.1016/j.neuron.2013.03.001

71. Malisza KL, Stroman PW. Functional imaging of the rat cervical spinal cord. *J Magn Reson
Imaging*. 2002;16(5):553-558. doi:10.1002/jmri.10185

72. Cahill CM, Stroman PW. Mapping of neural activity produced by thermal pain in the healthy
human spinal cord and brain stem: a functional magnetic resonance imaging study.
Magnetic Resonance Imaging. 2011;29(3):342-352. doi:10.1016/j.mri.2010.10.007

73. Rempe T, Wolff S, Riedel C, et al. Spinal and supraspinal processing of thermal stimuli: an
fMRI study. *J Magn Reson Imaging*. 2015;41(4):1046-1055. doi:10.1002/jmri.24627

74. Yang PF, Wang F, Chen LM. Differential fMRI Activation Patterns to Noxious Heat and
Tactile Stimuli in the Primate Spinal Cord. *J Neurosci*. 2015;35(29):10493-10502.
doi:10.1523/JNEUROSCI.0583-15.2015

75. Coggeshall RE, Chung K, Chung JM, Langford LA. Primary afferent axons in the tract of
lissauer in the monkey. *J Comp Neurol*. 1981;196(3):431-442. doi:10.1002/cne.901960307

76. Sugiura Y, Lee CL, Perl ER. Central Projections of Identified, Unmyelinated (C) Afferent
Fibers Innervating Mammalian Skin. *Science*. 1986;234(4774):358-361.
doi:10.1126/science.3764416

77. McGonigle DJ, Howseman AM, Athwal BS, Friston KJ, Frackowiak RSJ, Holmes AP.
Variability in fMRI: An Examination of Intersession Differences. *NeuroImage*.
2000;11(6):708-734. doi:10.1006/nimg.2000.0562

78. Smith SM, Beckmann CF, Ramnani N, et al. Variability in fMRI: A re-examination of inter-
session differences. *Hum Brain Mapp*. 2005;24(3):248-257. doi:10.1002/hbm.20080

79. Finsterbusch J, Eippert F, Büchel C. Single, slice-specific z-shim gradient pulses improve
T2*-weighted imaging of the spinal cord. *NeuroImage*. 2012;59(3):2307-2315.
doi:10.1016/j.neuroimage.2011.09.038

80. Kaptan M, Vannesjo SJ, Mildner T, et al. Automated slice-specific z-shimming for functional
magnetic resonance imaging of the human spinal cord. *Human Brain Mapping*. August
2022:hbm.26018. doi:10.1002/hbm.26018

81. Seifert AC, Vannesjo SJ. Comparison of Physiological Noise Models for Thermal Stimulus
fMRI in the Cervical Spinal Cord at 7T. *Proceedings of the International Society for*
Magnetic Resonance in Medicine. May 2022:1462.

859