1 **Colocalization of Cancer Associated Biomarkers on Single Extracellular Vesicles for Early**
2 **Cancer Detection**
3

4 Daniel P. Salem[1*], Laura T. Bortolin[1*], Dan Gusenleitner[1,2], Jonian Grosha[1], Ibukunoluwapo O.
5 Zabroski[1], Kelly M. Biette[1,3], Sanchari Banerjee[1], Christopher R. Sedlak[1], Delaney M. Byrne[1],
6 Bilal F. Hamzeh[1], MacKenzie S. King[1], Lauren T. Cuoco[1], Timothy Santos-Heiman[1], Peter A.
7 Duff[1], Emily S. Winn-Deen[1], Toumy Guettouche[1], Dawn R. Mattoon[1], Eric K. Huang[1,4], Randy
8 Schekman[5], Anthony D. Couvillon[1#], and Joseph C. Sedlak[1,6]
9
10 [*]co-first authors
11 [#] corresponding author
12
Affiliations:
[1] Mercy BioAnalytics Inc. 880 Winter Street, Suite 2003, Waltham, Massachusetts, 02451
[2] Current address: Bayer, Cambridge, MA, 02142
[3] Current address: Recursion Pharmaceuticals, 41 400 W Salt Lake City, Utah, 84101
[4] Current address: Saga Diagnostics, Lund, Sweden
[5] HHMI Investigator, Dept. of Mol. and Cell Biology, Li Ka Shing Center, UC Berkeley, Berkeley,
CA, 94720-3370
[6] Current address: Harvard Medical School, Boston MA, 02115

Acknowledgements:
The authors would like to thank Maciej Pacula (Flagstaff Solutions, Acton, MA), David Ransohoff
(UNC School of Medicine, Chapel Hill, NC), Steven Skates (Mass General Hospital, Harvard
University, Boston, MA), Floyd Taub (Golden, CO), Daria Filonov (AlphaNano Tech, Durham,
NC), and Ed Ha (Iskuda Therapeutics, New York, NY) for their generous support during the
design, execution, and analysis of this work.

28
29
Abstract

Detection of cancer early, when it is most treatable, remains a significant challenge due to the lack
of diagnostic methods sufficiently sensitive to detect nascent tumors. Early-stage tumors are small
relative to their tissue of origin, heterogeneous, and infrequently manifest in clinical symptoms.
Detection of their presence is made more difficult by a lack of abundant tumor-specific indicators
(i.e., protein biomarkers, circulating tumor DNA, etc.) that would enable detection using a non-
invasive diagnostic assay. In addition, many benign conditions manifest in a similar manner, thus
discriminating an early-stage cancerous lesion from a benign tumor can present additional
challenges and result in unnecessary medical procedures. To overcome these obstacles, we have
developed a liquid biopsy assay that interrogates circulating extracellular vesicles (EVs) to detect
tumor-specific biomarkers colocalized on the surface of individual EVs. Extracellular vesicles
from all cell types, including early-stage tumors, are known to be abundant in blood, are
remarkably stable, and serve as a biopsy of their cell of origin. The detection of a colocalized
combination of cancer associated biomarkers that provide tumor specificity on the surface of
extracellular vesicles enables the discrimination of early- and late-stage cancer from non-
malignant conditions.

Introduction

Cancer remains a significant public health challenge globally and is the second leading cause of death in the United States after heart disease (Siegel et al., 2023). Significant advances in the availability and efficacy of therapy have improved survival for some cancers, but the economic burden associated with cancer care has risen precipitously, and is projected to approach $245 billion by 2030, with annual and cumulative costs of care increasing with cancer stage (Mariotto et al., 2020; McGarvey et al., 2022). Cancer diagnosed at an early stage is addressable through an array of treatment modalities and is typically more responsive to therapy (Siegel et al., 2023; Sung et al., 2021). Given the significant clinical and economic benefits, the development and implementation of improved diagnostic tools for the early detection of cancer is a public health imperative.

To deliver net clinical benefit at a population scale, cancer screening tests must incorporate a foundational set of features. Tests intended for the early detection of cancer must have sufficient sensitivity to identify small, early-stage tumors, and sufficient specificity to minimize over-diagnosis and false positives due to confounding co-morbidities that may be present in the target population. Tests must be designed to facilitate broad adoption, including geographic and economic accessibility, and an experience that is convenient and noninvasive for the patient. Ideally, diagnostic tests also offer insights into the molecular mechanisms of the underlying cancer that can help guide clinical management. Tests designated for screening in otherwise healthy asymptomatic individuals, whether deployed in an average-risk or high-risk patient population, have the added requirement of demonstrated mortality benefit, requiring large, randomized control trials (Crosby et al., 2022). To date, such trials have only been successfully completed for select high-prevalence cancers, resulting in routine screening recommendations only for colorectal, lung, and breast cancers (US Preventive Services Task Force, 2009; US Preventive Services Task Force et al., 2021b, 2021a). As a result, more than 70% of cancer deaths in the United States result from malignancies for which there is no recommendation for screening (Beer, 2021).

Traditional diagnostic modalities aren't well suited for early cancer detection. Imaging-based techniques require specialized equipment that may not be broadly accessible, and often necessitate a separate clinical appointment. Low-cost imaging methods such as low-dose computed tomography (LDCT) for lung cancer screening, and trans-vaginal ultrasound for ovarian cancer screening lack the specificity to accurately differentiate malignancies from benign masses (Pavlik and van Nagell, 2013; Pinsky, 2014; van Nagell et al., 2007). While higher resolution imaging techniques such as PET/CT help to address this lack of specificity, the higher cost and more limited availability make them poorly suited as front-line diagnostic tools. Blood-based biomarker tests are an attractive alternative to imaging due to the low cost and broad accessibility, but single biomarker tests including PSA for prostate cancer screening and CA125 for ovarian cancer screening have exhibited limited specificity and their use has not been associated with a demonstrated mortality benefit (Menon et al., 2021, 2009; US Preventive Services Task Force et al., 2018). More recently, liquid biopsy approaches have been developed that are designed to measure genomic material released into circulation by apoptotic tumor cells. These methods leverage next generation sequencing (NGS) to measure a variety of genomic features including the presence of somatic mutations, methylation patterns, and fragment lengths, in some cases integrating thousands of features into a classifier that yields a determination of the presence or absence of cancer (Ignatiadis et al., 2021; Lone et al., 2022). In addition to the cost and complexity associated with NGS-based testing, these methods must overcome the biological limitation associated with the low abundance of circulating tumor DNA present in the context of small, early-

stage tumors, and the limited stability of genomic material in standard blood collection tubes.
These challenges may be addressed through the testing of large volumes of blood collected in
specialized collection tubes designed to stabilize circulating DNA and RNA but introduce new
logistic considerations (Dang and Park, 2022; Desai and Lovly, 2023; Yu et al., 2022, 2021).
Finally, classifiers trained on thousands of features run the risk of overfitting to chance effects, and
even if proper statistical care is taken, the detected effects may be statistically valid but biologically
false (e.g., predicting slight differences in sample storage or handling instead of the disease, or
other confounders). While these challenges can be overcome with large sample sizes and rigorous
validation protocols, it makes such classifiers difficult to develop, validate and operationalize in
practice (Cabús et al., 2022; Han and Jiang, 2014).

Extracellular vesicles (EVs) represent a diverse class of lipid-bilayer bounded nanoparticles
ranging from 30 to 10,000 nanometers in diameter (van Niel et al., 2018; Willms et al., 2018).
They are released by all living cells and serve an array of biological functions including cell-to-
cell communication and the facilitation of cancer progression by establishing a metastatic niche
and suppressing the immune response (Bebelman et al., 2018; Jerabkova-Roda et al., 2022; Lone
et al., 2022; van Niel et al., 2018; Willms et al., 2018). EVs have been shown to contain numerous
biomarkers derived from their cell of origin including proteins, lipids, and nucleic acids
(Ghodasara et al., 2023). Tumor derived EVs are a promising analyte for early cancer detection
due to their abundance in blood and molecular similarity with their cell of origin. Unlike genomic-
based detection methods, cell death is not required for the release of EVs. (Amelio al., 2020; Yu et
al., 2022) As a result, tumor-derived EVs outnumber tumor-derived DNA copies in circulation by
orders of magnitude in early-stage disease (Avanzini et al., 2020; Ferguson and Weissleder, 2020).
This enables the utilization of smaller sample volumes (~100 µL) relative to genomic-based liquid
biopsy assays, which require several milliliters of blood per run (Liu et al., 2021). The high plasma
EV concentration of $\sim 10^{10}$ EVs per mL and estimated tumor-derived EV shedding rates per cubic
millimeter of tumor volume make EVs an abundant source of tumor-derived biomarkers to target
in cancer screening assays designed for detection of smaller, early-stage tumors (Ferguson et al.,
2022; Ferguson and Weissleder, 2020; Johnsen et al., 2019). Given their abundance, stability, and
molecular similarity to their cell of origin, EVs offer a unique diagnostic test analyte that may offer
advantages over other modalities.

We have developed a method that interrogates the billions of EVs within a plasma or serum sample
for the simultaneous detection of colocalized, cancer-associated biomarkers on the surface of
individual EVs. Following the rational design of biomarker combinations, the clinical feasibility
of this approach was applied to the detection of high-grade serous ovarian cancer (HGSOC), the
most prevalent and deadliest histotype of ovarian cancer (Koshiyama et al., 2017).

133
Figure 1. Overview of the assay method. Schematic representation of the assay workflow. (A) Extracellular vesicles
are enriched from human cell line conditioned media, plasma, or serum by size-exclusion chromatography (SEC).
Enriched EVs contain cargo and surface biomarkers from their cell-of-origin. (B) Antibodies targeting cancer-
associated biomarker antibodies are conjugated to magnetic beads (capture antibody) or dsDNA oligonucleotides
(detection antibodies). (C) SEC-enriched EVs are captured in solution with magnetic bead-antibody conjugates
targeting a specific surface biomarker. (D) Immunoaffinity captured EVs are incubated with detection antibodies
conjugated to complementary double-stranded DNA probes. The dsDNA oligonucleotides contain single-stranded
overhangs which ligate only when in proximity to a complementary probe to generate a template for PCR. The
abundance of the detection biomarkers captured on the EVs is read out using quantitative PCR. Positive signal or low
cycle threshold (low Ct) is the result of all three biomarker antibody epitopes being present simultaneously on the
surface of the same EV, the absence of one or more biomarkers results in failed capture or failed detection and low
assay signal (high Ct).

Results
Measurement of colocalized biomarkers using immunocapture and detection. A conceptual
schematic of the assay is depicted in Figure 1. EVs from cultured human cell lines, human plasma,
or serum are enriched using size exclusion chromatography (SEC, Figure 1A). For capture and
detection of biomarkers on the surface of EVs, monoclonal antibodies specific to the extracellular
domains of cancer-associated targets are conjugated to magnetic beads (capture antibody) or
complementary dsDNA oligonucleotides (detection antibody) as indicated in Figure 1B. Enriched
EVs are isolated using a antibodies to the capture biomarker (Figure 1C) followed by detection
with two antibodies to a second and third biomarker conjugated to complementary dsDNA
oligonucleotides (Figure 1D). These complementary oligonucleotides hybridize and ligate if two
DNA-conjugated antibodies are bound in proximity on the same EV (Figure 1D). The ligated DNA
contains primer sites serving as a template for qPCR amplification to quantify the number of
colocalization events in the sample. Based on this assay design, the complete set of biomarkers,
detected by the antibody conjugated to the magnetic bead as well as the antibodies conjugated to

the complementary oligonucleotides (a "biomarker combination")—must be present on the surface

of a single EV to generate a signal. In the following figures the assay readout is generally presented

as a boxplot using the PCR cycle threshold (Ct) to discriminate positive signal (low Ct) from

negative signal (high Ct).

Identification and selection of biomarkers.   Bone marrow stromal antigen-2/CD317/Tetherin

(BST2), Folate receptor alpha (FOLR1), and Mucin-1 (MUC1) are cell surface proteins previously

shown to be overexpressed in ovarian cancer (Gautam et al., 2023; Giampaolino et al., 2020, 2019;

Januchowski et al., 2017; Kufe, 2009; Trinidad et al., 2023; Varaganti et al., 2023; Yang et al.,

2022).   Protein expression of all three biomarkers was demonstrated by western blotting to be

present in the EV fractions from human ovarian cancer cell lines (Figure 2) and human plasma

(not shown) enriched by size exclusion.

174
Figure 2. Characterization of human cancer cell line extracellular vesicles. SEC-enriched EV fractions from
human ovarian cancer cell line (COV413A) conditioned media were characterized according to MISEV guidelines.
(A) Nanoparticle tracking analysis (NTA, grey bars) and protein quantification (black line) of each fraction collected
from the SEC column demonstrating the bulk of the nanoparticles elute in EV1-EV3, distinct from the bulk of soluble
protein (EV6-EV8). The first fraction (Pre-EV) is the sum of the void and wash volumes of the column which generally
contains larger particles, such as lipoproteins, which are not observed in cell line conditioned media. The purified
collection volume (PCV), fractions EV1-EV3, containing the EVs used in downstream assay steps are highlighted.
(B) Transmission electron microscopy of EVs derived from the total PCV (pooled fractions EV1-EV3) from
COV413A conditioned media at two different magnifications demonstrates intact EVs ranging from ~40-300 nm.
Scale bars are 500 nm (40,000x, left) and 200 nm (120,000x, right), respectively. (C) Particle size distribution of
pooled SEC fractions EV1-EV3 demonstrating particles consistent with the size range of small EVs ranging from ~40-
400 nm. (D) Western blot analysis of each of the initial fractions from the SEC column (Pre-EV, EV1-EV6) from
COV413A conditioned media demonstrating the presence of known EV surface biomarkers (CD9, CD63, CD81), as
well as BST2, FOLR1, and MUC1. Note that an antibody detecting the cleaved fragment of MUC1 (MUC1-C) was
used to facilitate immunoblotting due to the very high molecular weight of MUC1 (250-500 kDa). Evaluation of

additional MISEV protein markers, the presence of EV cargo proteins (flotillin, GAPDH) and absence of non-EV
biomarkers (RPS3, Tom20) can be found in supplemental figure 2. Whole cell extracts (WCE, 10 ug total protein per
lane) from COV413A were analyzed in parallel as a control for protein expression and biotinylated molecular weight
(Mol. Wt.) markers were used to determine apparent molecular mass.
194
Analyte Characterization. EVs enriched by SEC from human cell line conditioned media (Figure
2A, Supplemental Figure 2) and human plasma (Supplemental Figure 1) were characterized
according to MISEV2018 guidelines by nanoparticle tracking analysis (NTA), transmission
electron microscopy (TEM), and western blot to evaluate particle size, number, morphology, and
phenotype (Théry et al., 2018). NTA for EVs from human cell lines (Figure 2A, C) and human
plasma (Supplemental Figure 2A and 2B) demonstrates EV SEC fractions 1-3 (EV1-3) contain
particles ranging in size from 40-400 nm consistent with the known size distribution for EVs. We
selected and pooled EV1-3 for immunocapture and detection in our assay. In human plasma, EV4
was excluded due to the higher background introduced by the abundance of soluble protein in
human plasma (Supplemental Figure 2A) and absence of EV biomarkers (CD9, CD81) as
determined by western blot (Supplemental Figure 2C). The fundamental differences in NTA counts
demonstrated between human plasma (Supplemental Figure 2) and human cell line conditioned
media (Figure 2) is explained by the difference in heterogeneity between nanoparticles derived
from human blood versus those enriched from a clonal cell line in culture (Lobb et al., 2015;
Veerman et al., 2021). Enriched nanoparticles from both cell line and plasma EV fractions were
further characterized by TEM demonstrating lipid-bounded nanoparticles ranging from 40-400 nm
consistent with the NTA results (Figure 2B and Supplemental Figure 2D)(Cizmar and Yuana, 2017;
Veerman et al., 2021). Additionally, western blotting was used to confirm the presence of accepted
EV markers and absence of non-EV proteins as well as the presence of BST2, FOLR1 and MUC1
proteins within the EV fraction (Figure 2D and Supplemental Figure 2).
215
Antibody Screening. Commercially available monoclonal antibodies to FOLR1, BST2 and
MUC1 were screened for their ability to capture or detect EVs enriched from human cell lines
predicted to express high (positive cell line) or low/no (negative cell line) levels of the selected
biomarker (Supplemental Figure 3). Representative screening data for anti-FOLR1 antibodies
used for capture (Supplemental Figure 3A) and detection (Supplemental Figure 3B) are shown
using EVs derived from COV413A (FOLR1 positive) and SK-MEL-1 (FOLR1 negative) cell lines.
Of the six antibodies screened for capture using CD81 antibodies for detection, antibodies 1 and 5
were selected for additional characterization based on the signal to background difference between
the positive and negative cell line among other criteria. Antibodies 4 and 6 illustrate examples of
reagents that failed validation in the capture assay due to lack of signal in the positive cell line
(Supplemental Figure 3A). Antibodies 1 and 5 were further validated for detection using anti-
CD81 conjugated beads for capture prior to detection with anti-FOLR1 antibodies (Supplemental
Figure 3B). As with capture performance, both antibodies perform well in the detection step as
illustrated by their ability to discriminate FOLR1 positive cell line EVs (COV413A) from FOLR1
negative cell line EVs (SK-MEL-1). Screening data for antibodies against BST2 and MUC1
described in this study are shown in Supplemental Figure 3 C-F.
232

233
Figure 3: BST2, FOLR1, and MUC1 are colocalized with CD63 and CD81 in cell line EVs. EVs derived from
cell lines predicted to express high (red), medium (white) or low (blue) levels of the indicated biomarkers based on
mRNA analysis were captured with antibody-bead conjugates using anti-CD63 or anti-CD81 antibodies. Detection of
bound EVs was determined using oligo-conjugated antibodies against (A) BST2, (B) FOLR1, or (C) MUC1.
COV413A EVs were used as a positive control since this line expresses high levels of all three biomarkers studied.
For BST2, FOLR1, and MUC1, a second positive cell line, NCI-H1781, NCI-H1819, and HCC-4006, respectively,
was also included as a control. Note that despite moderate FOLR1 mRNA expression in NCI-H1819 cells, proteomics
and western blot data would suggest otherwise (data not shown) suggesting this cell line may not express FOLR1
protein. SK-MEL-1, which expresses low levels of each marker was used as a negative control. Heat maps were
included to demonstrate relative expression levels (log2 TPM) for each marker, expression data are derived from the
CCLE database at the Broad Institute. No EV controls were used to eliminate antibodies incompatible with the assay
(antibodies that demonstrated non-specific binding to the beads or other antibodies under the assay conditions).

246
Cancer biomarkers are colocalized on the surface of EVs. To demonstrate colocalization of
cancer-associated biomarkers with canonical EV biomarkers, the assay was applied to EVs derived
from human cell lines known to express high or low levels of each cancer-associated biomarker.
After capturing SEC-enriched EVs with antibodies directed against canonical EV tetraspanin
biomarkers (CD63 or CD81), bound EVs were detected using detection antibodies directed against
the extracellular domains of BST2, FOLR1, or MUC1 (Figure 3A, B and C, respectively). Only
EVs derived from cell lines positive for both the tetraspanin and the cancer-associated biomarker
result in a strong qPCR signal.

255
Assay signal operates akin to a Boolean logic "AND gate" whereby all three surface epitopes must
be confined to the same EV for PCR template generation to occur. To demonstrate this, we
characterized EVs derived from four human cancer cell lines (COV413A, OVKATE, SW900, and
G-401) exhibiting a range of BST2, FOLR1, and MUC1 expression levels (Figure 4A-C). We
tested the functionality of the assay using two-biomarker (BST2 capture/FOLR1 + FOLR1
detection and BST2 capture/MUC1 + MUC1 detection) and three-biomarker (BST2
capture/FOLR1 + MUC1 detection) combinations in EVs from cell lines that were
BST2^{high}/FOLR1^{high}/MUC1^{low} (OVKATE), BST2^{high}/FOLR1^{low}/MUC1^{high} (G-401 and SW900),
and BST2^{high}/FOLR1^{high}/MUC1^{high} (COV413A) to demonstrate that assay signal is dependent on

all of the capture and detection biomarkers being present for signal. Cell line EVs were tested at
an equivalent concentration of 1.6×10^9 particles/mL. As shown in Figure 4, signal was observed
to correlate with biomarker colocalization in expected patterns. BST2 capture/FOLR1 + FOLR1
detection and BST2 capture/MUC1 + MUC1 detection generated strong signals for the cell lines
that express both biomarkers (COV413A/OVKATE and COV413A/SW900, respectively). Weak
BST2 capture/FOLR1 + FOLR1 detection and BST2 capture/MUC1 + MUC1 detection signals
were observed for cell lines expressing low levels of at least one marker (SW900/G-401 and
OVKATE/G-401, respectively). The three-biomarker combination BST2/FOLR1/MUC1 showed
a clear separation between COV413A EVs (triple-positive) and cell lines missing at least one
biomarker, suggesting assay signal is only robust when all biomarkers are colocalized on the same
EV.

Figure 4: Colocalization of biomarkers is required for assay performance and improves signal to noise ratio.
All three biomarkers must be present on the same EV for productive readout. If only two biomarkers are present on
the EVs (one capture + one detect), no ligation occurs, and Ct is high. (A, B) Doublet (i.e. BST2 capture, FOLR1 +
FOLR1 detection or BST2 capture, MUC1 + MUC1 detection) or (C) triplet combinations (BST2 capture, FOLR1 +
MUC1 detection) of antibodies tested on cell line EVs from COV413A (BST2[high]/FOLR1[high]/MUC1[high]), OVKATE
(BST2[high]/FOLR1[high]/MUC1[low]), SW900 (BST2[high]/FOLR1[low]/MUC1[high]), and G-401 (BST2[high]/FOLR1[low]/MUC1[low])
showing that doublets of BST2 capture/FOLR1 + FOLR1 detection (A), BST2 capture/MUC1 + MUC1 detection (B),
or a triplet of BST2 capture/FOLR1 + MUC1 detection (C) only yield positive qPCR signal in cell lines with matching
high biomarker expression. Biomarker mRNA abundance (log2 TPM) is represented for each biomarker in each cell
line. (D) Signal from cell line EV-spiked (dark blue) or unspiked (light blue) healthy human plasma from a
representative set of one-biomarker (MUC1 capture, MUC1 + MUC1 detection), two-biomarker (MUC1
capture/FOLR1 + FOLR1 detection or MUC1 capture/BST2 + BST2 detection), and three-biomarker (MUC1
capture/FOLR1 + BST2 detection) combinations. Increasing the number of unique, colocalized biomarkers results in
reduced healthy background and improved discrimination of cancer cell line EV-spiked plasma. Signal difference
(ΔCt) between unspiked healthy human plasma and spiked plasma is indicated.

Detecting multiple colocalized biomarkers on EVs improves the signal-to-noise ratio. We
hypothesized that the colocalization of multiple cancer-associated biomarkers could improve the
specificity of distinguishing dilute cancer EVs from normal EVs. To test this, we spiked plasma
from healthy control subjects with EVs derived from ovarian cancer cell lines and evaluated
discrimination performance using combinations of one, two, and three unique biomarkers. Spiked

plasma was prepared by mixing purified COV413A EVs with a female plasma pool derived from
ninety healthy women at a final concentration of $1x10^8$ COV413A particles/mL. Unspiked and
spiked plasma pools were characterized in the assay using biomarker combinations containing
BST2, FOLR1, and/or MUC1 in various capture/detection assay configurations. Single-biomarker
combinations comprising capture and detection antibodies targeting MUC1 only (MUC1 capture,
MUC1 + MUC1 detection, Figure 4D; left panel) BST2-only, or FOLR1-only (data not shown),
generally exhibited poor discrimination between unspiked and spiked plasma samples (median
$\Delta Ct = 0.2$). However, discrimination performance improves in some cases for biomarker
combinations requiring the colocalization of two unique cancer-associated biomarkers in a
biomarker-dependent manner (Figure 4D; middle panels). For example, MUC1 capture/FOLR1 +
FOLR1 capture resulted in no discrimination of unspiked versus spiked plasma ($\Delta Ct = 0.6$)
whereas MUC1 capture/BST2 + BST detection resulted in improved discrimination ($\Delta Ct = 3.2$).
An assay configuration with all three biomarkers, MUC1 capture/FOLR1 + BST2 detection
resulted in a ~260,000 (2^{18}) fold reduction of signal in the unspiked plasma sample, enabling the
greatest separation of unspiked versus spiked plasma ($\Delta Ct = 6.1$, right panel).

315
Figure 5: Colocalization of BST2, FOLR1, and MUC1 on single EVs using super-resolution microscopy. EVs
derived from COV413A cells were captured on an S4 microfluidic slide (ONI) prior to staining with an antibody
cocktail containing Alexa Fluor antibody conjugates to BST2 (magenta), MUC1 (blue), and FOLR1 (yellow). Triple-
positive EVs were identified using CODI software and verified manually. Representative images are shown. The scale
bar represents 100 nm.

Biomarkers are colocalized on individual EVs from human cell lines. To confirm biomarker
colocalization on individual EVs using an orthogonal approach, we utilized super-resolution
microscopy to identify triple positive (BST2+/FOLR+/MUC1+) EVs. Figure 5 demonstrates the
colocalization and simultaneous detection of FOLR1, BST2, and MUC1 on individual,
representative COV413A cell line EVs captured on the surface of a microfluidic chip using
phosphatidylserine directed S4 affinity from Oxford Nanotechnology (ONI, Oxford, UK).

Assay performance depends on intact EVs. To demonstrate that assay signal is derived from
colocalized biomarkers on an intact EV, SEC enriched EVs from human cell line conditioned
media (COV413A) or pooled patient plasma from late-stage HGSOC were subjected to capture
with MUC1 antibodies. Prior to detection, captured EVs were incubated with buffer alone (-) or
buffer containing 0.075% Triton-X100 (+) sufficient to disrupt intact EVs (Supplemental Figure
4)(Osteikoetxea et al., 2015). Assay signal was generated using detection antibodies to BST2 and
FOLR1. Detergent treatment results in a significant increase in Ct suggesting that most of the
signal is derived from intact EVs that are detergent sensitive.
337

338
339 **Figure 6. Discrimination of High-Grade Serous Ovarian Cancer (HGSOC) from Healthy Controls and Benign**
340 **Ovarian Masses.** Performance of the MUC1 capture/BST2 + FOLR1 detection biomarker combination in a cohort of
341 healthy controls (n=24), patients with benign ovarian masses (n=26), and patients with stage I (n=9), II (n=9), III
342 (n=15), and IV (n=4) HGSOC in a proof-of-concept assay.
343

Assay signal requires biomarker colocalization on the same EV. Theoretically, complementary
DNA-conjugated antibodies located on adjacent EVs could generate a DNA template if within
sufficient proximity ("bridging"). We characterized a mixture of EVs from two different cancer
cell lines, OVKATE and SW900, each predicted to express high levels of two of the three
biomarkers (BST2^{high}/FOLR1^{high}/MUC1^{low}, and BST2^{high}/FOLR1^{low}/MUC1^{high}, respectively). To
assess bridging, we characterized the individual cell line EVs (4x10^8 particles/mL) and a 1:1
mixture (2x10^8 particles/mL each) using the triplet biomarker combination BST2 capture/FOLR1
+ MUC1 detection). As expected, EVs from COV413A that express all three biomarkers produced
high signal (18.4 Ct) relative to EVs from G-401 cells which express low levels of both FOLR1
and MUC1 (33.4 Ct). EVs from double-biomarker positive cell lines OVKATE
(BST2^{high}/FOLR1^{high}/MUC1^{low}) or SW900 (BST2^{high}/FOLR1^{low}/MUC1^{high}) generated assay
signals of 26.7 Ct and 30.5 Ct, respectively, while the 1:1 mixture of EVs from both cell lines
generated a signal of 28.2 Ct (Supplementary Figure 5). The 1:1 mixture signal was consistent
with the arithmetic mean of the individual cell line EVs, suggesting that bridging does not
contribute significantly to assay signal.
359

Assay linearity and reproducibility. The linearity of the MUC1, BST2+FOLR1 combination was
evaluated using plasma spiked with a 4-fold dilution series of EVs collected from the human
ovarian cancer cell line COV413A (see Supplemental Figure 6). Samples prepared in this manner

were expected to cover the healthy control and ovarian cancer cycle threshold range of ~20 to ~40
Ct. The dilution series was prepared 4 times and then each sample was run in duplicate (8
replicates) in a single assay plate. There was no high-dose hook effect observed, even at the highest
spike-in level of $3.2x10^8$ EVs per sample well, which is well above the combination Ct values
observed in either healthy or ovarian cancer samples. This combination was linear over the 4-log
range of 19,500 to 320 million cell line EVs per well. Based on the results from the linearity study
5-level controls were targeted at $3.2x10^8$ EVs/well, $8x10^7$ EVs/well, $2x10^7$EVs/well, $5x10^6$
EVs/well, and unspiked plasma background. Each of these full process controls was tested with
the MUC1, BST2+FOLR1 combination on 16 separate days and the between-day reproducibility
was calculated from these data. This combination had between-day coefficients of variation (CVs)
of under 2% for all 5-level controls.

Assay signal correlates with cell line gene expression and protein expression. To evaluate the
correlation between assay signal and cell line gene and protein expression, purified EVs from 32
human cancer cell lines were characterized at a constant nanoparticle concentration using single-
biomarker capture/detect combinations directed towards BST2, FOLR1, and MUC1. Assay signal
for each biomarker is summarized in Supplemental Figure 7 along with published gene expression
and protein expression data (when available). Signal strength was positively correlated with gene
and protein expression for all three biomarkers (Spearman's correlation coefficients between 0.47
and 0.87), suggesting that cell line biomarker expression is predictive of EV protein composition.

Discrimination of high grade serous ovarian cancer (HGSOC) in patient samples. To assess
feasibility for the detection of early-stage cancer, we applied the assay to a clinical cohort of 92
patients comprising 42 HGSOC cases (18 stage I/II, 24 stage III/IV), 26 females with benign
adnexal masses, and 24 female healthy controls. Plasma from each subject was characterized using
the BST2/FOLR1/MUC1 biomarker combination targeting BST2/FOLR1/MUC1-positive EVs.
The results are summarized in Figure 6. In this limited proof of concept study, the biomarker
combination of BST2/FOLR1/MUC1 achieved a 55.6% sensitivity (95% CI: 30.8%-78.5%) for
stage I/II disease. More importantly, this biomarker combination readily discriminates between
early-stage and benign disease. This performance has been further improved in subsequently
developed assay configurations (data not shown), suggesting this approach may have clinical
utility in detecting early stage HGSOC and discriminating ovarian malignancy from benign
adnexal masses.

Discussion
Cancer screening has been shown to reduce cancer-associated mortality through the detection of
localized disease that is more amenable to treatment, underscoring the clinical need for more
accurate and accessible screening tests (Crosby et al., 2022). Despite considerable effort
developing blood-based screening approaches, technical challenges have constrained test
performance. Early-stage tumors are small, limiting the abundance of the available analytes for
measurement in liquid biopsies. Cancer mis-regulates normal physiological processes, which
means the analytes produced by cancers are often also produced by normal tissues, either in the
tissue of origin or other parts of the body. Even when arising from the same tissue type, cancer
utilizes blind evolutionary opportunism and therefore the mechanisms that lead to carcinogenesis
can vary widely. Measuring colocalized cancer-associated biomarkers on intact EVs addresses

each of these challenges. Specifically, EVs are a stochastic sampling of the cell of origin bearing
cancer-associated biomarkers both on their surface and as cargo (Doyle and Wang, 2019; Li et al.,
2020). They are released continuously from tumor cells due to their anti-inflammatory and tumor-
supporting functions and are stable enough to tolerate processing and amenable to at-home
collection tests (Johnsen et al., 2019).

Here we describe a novel, non-invasive method for the early detection of cancer through the
measurement of co-localized cancer associated biomarkers on the surface of extracellular vesicles.
EVs are an emerging analyte class with features well suited to use in diagnostic testing and
represent an area of science that is undergoing rapid advances in the understanding of basic biology
as well as research and clinical applications (Ghodasara et al., 2023). There are biological and
technical limitations to the approach described here that are important considerations informing
ongoing assay development. EVs derived from hematopoietic lineage cells comprise 99.8% of all
EVs in human plasma, and thus biomarker targets that are highly expressed in blood may generate
assay background (Auber and Svenningsen, 2022; Ferguson and Weissleder, 2020; Johnsen et al.,
2019). Assay performance is dependent on the quality and availability of compatible affinity
reagents, such as antibodies. Promising cancer-associated biomarkers lacking affinity reagents
with suitable performance characteristics may restrict potentially useful biomarker combinations
from inclusion in assays. Among biomarkers with suitable affinity reagents, we observed a range
in discrimination performance, potentially resulting from variability in healthy background signal.
Variability in assay specificity may be driven by several factors including biomarker expression,
antibody affinity and specificity, and the orientation of the assay design (i.e., the choice of
biomarkers for use in immunoaffinity capture versus the detection step). The studies we conducted
evaluating test performance in clinical samples provide encouraging evidence that the assay will
have utility in identifying early-stage cancers and discriminating malignant from healthy samples
but represent proof of concept only. Further studies to assess the clinical performance of the assay
for the detection of high-grade serous ovarian cancer (HGSOC) in well powered case-control
studies utilizing clinical samples collected prospectively under standardized protocols, are in
progress.

Measuring colocalized biomarkers on single EVs using the assay described herein is a promising
approach for early-stage cancer detection. It addresses the biological abundance limitations of
genomic liquid biopsy approaches while offering improved specificity over bulk EV
measurements. The assay measures EV surface markers and thereby leverages the natural
enrichment of membrane-bound epitopes relative to cargo proteins due to the 100-fold increase in
surface area-to-volume ratio between an EV and its cell-of-origin. Moreover, it can interrogate
billions of EVs simultaneously enabling the detection of dilute (<1 in 100,000)(Ferguson and
Weissleder, 2020) cancer associated EVs reaffirming the observation that the assay can be used to
detect EVs from early-stage tumors. The qPCR-based readout enhances assay technical sensitivity
relative to conventional enzyme-amplified sandwich immunoassays, similar to other approaches
using immuno-PCR (Niemeyer et al., 2005; Stiller et al., 2021), proximity ligation assays
(Tavoosidana et al., 2011; Weibrecht et al., 2010), and proximity extension assays (Larssen et al.,
2017; Lundberg et al., 2011).

The small sample volume requirement, compatibility with standard blood collection tubes, and use
of low-cost qPCR for signal generation are features that may help to enable broad adoption,
including in lower resource settings. Advances in affinity reagent development and EV biomarker
discovery may be leveraged to improve existing assays by incorporating new reagents or designing

orthogonal combinations to boost assay performance. We anticipate this assay format is readily
extensible to a broad array of clinical indications, across cancer types and beyond cancer. For
example, neuronally derived EV biomarkers have been reported to track with cognitive decline in
Alzheimer's disease and may represent another opportunity for diagnostic utility (Eren et al.,
2022). EV-based measurement of tumor-derived biomarkers may be helpful in guiding clinical
management through use in companion diagnostic applications. For example, folate receptor
alpha, one of the ovarian cancer biomarkers included in the combination described in this study, is
the target of several ongoing efforts focused on the development of antibody-drug-conjugate
(ADC) therapeutics (El Bairi et al., 2021). A liquid biopsy assay indicative of the presence of folate
receptor alpha positive tumors may be helpful in identifying patients more likely to respond to
targeted ADC therapy. The EV-based test method described here has the potential for immediate
clinical utility in single cancer early detection, and potentially for multi-cancer early detection
either through the rational combination of single-cancer EV tests, or as part of an integrated multi-
omic approach.
470
471 Materials and Methods
472
Cell culture. All human cancer cell lines used in this study, and their respective sources, are
indicated in Supplemental Table 1. SK-MEL-1, and SK-MES-1 were used under license from the
Memorial Sloan Kettering Cancer Center. NCI-H146, NCI-H441, NCI-H520, and NCI-H1781
were developed by Dr. Gazdar and Dr. Minna and used under license from the National Cancer
Institute. COV362 and COV413A were used under license from the European Collection of
Authenticated Cell Cultures (ECACC). OVSAHO and OVISE were obtained from Sekisui
XenoTech (Kansas City, KS, USA) under license from the Japanese Collection of Research
Bioresources (JCRB, Osaka, Japan). All cells were thawed, subcultured, and maintained in
complete growth media as recommended by the supplier. For EV production, cells were passaged
into large capacity tissue culture-treated flasks or dishes until 70-80% confluent. The flasks or
dishes were then transitioned to EV-depleted FBS (A2720801, ThermoFisher Scientific, Waltham,
MA) or serum-free Cancer Cell Line Media XF (CCLM XF, C-28077, PromoCell GmbH,
Heidelberg, Germany), depending on the individual cell line as follows; growth media was
removed and the cells were rinsed twice with 1/2 volume of either complete media supplemented
with exosome-depleted (EV-) FBS or CCLM XF serum free media prior to being replaced in basal
media with 10% EV- FBS, or CCLM XF. After 48 hours, the conditioned media (CM) was
collected and cellular debris removed by centrifugation at $300 \times g$ for 5-minutes followed by two
subsequent centrifugations at $1500 \times g$ for 10 minutes each. All centrifugations were done at 21°C.
The cleared CM was transferred to 50 mL conical tubes in ~45 mL aliquots and stored at -80°C
prior to purification.
493
Concentration of conditioned media. Typically, 200 mL of CM was thawed at room-temperature
and centrifuged at $1,300 \times g$ for 10 minutes to remove any insoluble debris. The supernatant was
transferred via serological pipet to fresh 50 mL conical tubes then re-centrifuged at $2,000 \times g$ for 30
minutes. The supernatant was carefully transferred to a fresh tube and concentrated using Amicon
Ultra Centrifugal filter units with a 10,000 molecular weight cut-off (Millipore-Sigma, St. Louis,
MO). This process was repeated until all the retained material was concentrated to approximately
1 mL. All centrifugations were done at 21°C unless specified otherwise.
501

EV Isolation from concentrated conditioned cell culture media or plasma. EVs were purified
from either 1 mL of concentrated CM, acid-citrate-dextrose (ACD) plasma, or human potassium
EDTA plasma. In each instance, 1 mL of concentrated CM or plasma was applied to an Izon qEV
original 70 nm Legacy SEC column (ICO-70, Izon Science US, Medford, MA, USA) following
the vendor-recommended protocol. Briefly, qEV columns were allowed to equilibrate to room
temperature and washed in 10 mL room-temperature 1x PBS. Concentrated CM (described above)
was applied to the top of the column and the flow through material was discarded. Two milliliters
of 1X PBS were applied to the column and collected as the pre-EV fraction and analyzed as
described. EV-elution fractions (EV1, EV2…EV8) were then collected by applying 0.5 mL 1X
PBS sequentially until all desired fractions were collected. The PCV was determined based on
physical and functional characterization of each EV fraction.   Purified plasma EVs were
immediately analyzed using the assay as described. Concentrated CM EVs were quantified using
ZetaView Classic S model nanoparticle tracking analysis (NTA) instrument (Particle Metrix,
Amersee, Germany) prior to aliquoting and freezing at -80°C. The protocol for purifying plasma
and CM EVs was identical, with the exception that 0.02 uM-filtered 1X PBS, pH 7.4 was used for
the concentrated CM EV purification (so particles in the PBS will not confound the downstream
quantification) while unfiltered 1X PBS, pH 7.4 was used to purify EVs from plasma.

Western blot analysis of human cell line or plasma extracellular vesicle fractions. Whole-cell
extracts (WCE) of COV413A cells were prepared by scrape-harvesting 90%-95% confluent cells
in 1X #9803 Cell lysis buffer (Cell Signaling Technology, Danvers, MA). Lysates were incubated
at 4°C for 30 minutes prior to ultrasonic cavitation (3 x 15 seconds at 33% power) using QSonica
Q125 Ultrasonic Disruptor (QSonica, Newtown, CT, USA) fitted with a 2 mm titanium probe tip.
Protein concentrations were determined, and the lysate was further diluted to 1mg/mL in 4X LDS
Sample Buffer (NP0008, ThermoFisher Scientific). Human plasma lysate was prepared by mixing
whole human plasma 1:1 with 4X LDS Sample Buffer followed by ultrasonic cavitation to
homogenize the sample. Individual EV fractions from concentrated CM were diluted 1:1 in 4X
LDS Sample Buffer and human plasma EVs were concentrated 10:1 prior to 1:1 dilution in 4X
LDS Sample Buffer. In all cases, samples were heated to 70°C for five minutes prior to use. For
polyacrylamide gel electrophoresis (PAGE) and western blot experiments, 10-20 µL of WCE,
plasma lysate, or EV fraction was loaded alongside a biotinylated ladder and/or pre-stained protein
molecular weight markers (Catalog numbers 7727 and 59329, respectively, Cell Signaling
Technology) on NuPAGE™ Bis-Tris™ pre-cast gels (ThermoFisher Scientific). Gel percentage,
running buffer, and transfer conditions were dictated by the apparent molecular weight of the
proteins being immunoblotted. In most cases, we used 1.0 mm, 4-12% Bis-Tris mini gels with
MOPS-based running buffer (NP0001, ThermoFisher) according to the manufacturer's
instructions. After PAGE, gels were either stained with Gel-Code™ Protein Stain (24590,
ThermoFisher) or transferred to 0.2-micron PVDF membranes (1620174, Bio-Rad, Hercules, CA,
USA) using a XCell SureLock Mini-Cell (EI9051, ThermoFisher) according to the manufacturer's
recommendations. After electrophoretic transfer, all membranes were blocked at room temperature
for one hour in 1X TBST (#9997, Cell Signaling Technology) supplemented with 5% non-fat dry
milk (AB10109, American Bioanalytical, Canton MA). Primary antibodies (Supplemental Table
1) were incubated overnight (16 hours) in 1X TBST supplemented with either 5% non-fat dry milk
or bovine serum albumin (#9997, #9998, Cell Signaling Technology) at 4°C with constant rocking
according to the manufacturer's recommendation. Each blot was washed thrice for 5 minutes in
1X TBST at room temperature prior to addition of HRP-conjugated secondary antibodies and
HRP-anti-biotin antibody, both diluted in 1X TBST + 5% non-fat dry milk, for one hour at room

temperature. Each blot was then washed thrice in 1X TBST for 5 minutes and proteins were
detected using either LumiGLO/Peroxide Reagent or SignalFire™ ECL reagent (7003 and 6883,
respectively, Cell Signaling Technology) and an iBright™ imaging system (ThermoFisher). All
images are presented with no image adjustments (other than cropping to highlight the area of
interest) and are typically between 5 and 15-minute exposures, depending on the signal strength.
All antibodies used for western blot and their working dilution are listed in Supplemental Table 1.
555
Characterization of EV size distribution and concentration. Size distribution and particle
counts were determined using a ZetaView NTA instrument after enrichment of the EVs by SEC.
ZetaView software 8.05.14 SP7 (Particle Metrix) acquired and exported the data. The instrument
was calibrated as recommended by the manufacturer prior to all measurements. Following SEC
isolation and prior to NTA measurement, the isolated particles were diluted 1:250 - 1:2000, which
allowed the instrument to accurately measure the particle distribution (50-200 particles per field
of view). As recommended by the manufacturer for the quantification of EVs, the sensitivity was
set at 80, shutter at 100, min brightness 20, min area 10, and max area 10000. All 11 positions of
the instrument were analyzed, with outlier positions excluded automatically by the instrument. The
size distribution results were binned in 1nm intervals from 1 to 1000 nm.
566
Antibody Panels. All antibodies used for assay development were obtained from commercial
sources (Supplemental Table 3). Initially, around 50 different antibodies against the three
biomarker targets were screened for functionality and compatibility using cell line EVs. Antibody
clones that passed primary screening were further optimized to determine the optimal conjugation
concentration of both probes and beads in biomarker-expressing cell lines vs biomarker-negative
cell lines or buffer alone controls. The top antibodies for the MUC1, BST2, and FOLR1
combination were then validated in human plasma and further down-selected based on those
antibodies that produced the best discrimination of HGSOC from healthy women and those with
benign adnexal masses.
576
Plasma Samples. The 124 healthy donor plasma samples were collected under a WCG IRB-
approved protocol (20212722). Each donor provided 2 tubes of blood which were collected using
a 10-mL collection volume K_2EDTA tube (0265732, Thermo Fisher Scientific). After collection,
each tube of K_2EDTA plasma was inverted 6-8 times, centrifuged ($1500{\times}g$ at room temperature
for 15 minutes) at the collection site within 60 minutes collection to separate the plasma from the
other blood components. The resultant plasma was divided into 1 mL aliquots in 2 mL cryovials,
frozen and stored at -80ºC. The frozen aliquots were shipped from the collection site to Mercy
BioAnalytics on dry ice and stored at -80°C. A plasma pool from ninety healthy women was
prepared in-house using samples purchased from ProteoGenex Inc. (Inglewood, CA). Late- and
early-stage HGSOC cases were purchased from ProteoGenex, Inc. and Reprocell (Beltsville, MD).
All ProteoGenex plasma samples were sourced from Moscow, Russia and collected in either
K_2EDTA or K_3EDTA blood collection tubes. Reprocell, Inc. samples were sourced from the United
States and collected in acid citrate dextrose (ACD) blood collection tubes. Plasma samples from
women with benign ovarian masses were purchased from ProteoGenex were also sourced from
Moscow, Russia and collected in K_2EDTA blood collection tubes. Patient sample attributes are
shown in Supplemental Table 4.
593
594   **Antibody-functionalized magnetic beads.** Antibodies containing bovine serum albumin (BSA)
595   were purified using protein A and G Spintrap purification kits (GE28-9031-34 or GE28-9031-32,

Cytiva, Marlborough, MA) according to manufacturers' instructions. All antibodies used in the assay underwent a buffer exchange process to ensure that any carriers or impurities that might interfere with conjugation were removed. Up to 500 µg of antibody material was added to 2 mL of sterile 1X PBS (10010031, ThermoFisher Scientific) in 30K MW Amicon Ultra Filter Units (UFC203024, Sigma-Aldrich, St. Louis, MO, USA) and centrifuged for ten minutes at 4,375×g . This process was repeated 6 times, after which the antibodies were recovered and sterile filtered by centrifugation in 0.22 µM Spin-X tube filters (8160, Corning-Costar, Corning, NY) for 2 minutes at 12,000×g. Antibody concentrations were measured using the NanoDrop spectrophotometer. Immunoaffinity magnetic beads were prepared by coating antibodies onto Dynabeads® M-270 Epoxy beads using the Dynabeads Antibody Coupling Kit (14311D, ThermoFisher Scientific) according to the manufacturer's instructions. Briefly, the beads were suspended at 10 mg mL^{-1} in C1 buffer and mixed thoroughly by vortexing. 1 mL aliquots of the beads were transferred to 1.5 mL tubes, which were then placed on a DynaMag™-2 magnetic rack (12321D, ThermoFisher) for 1 minute to separate the beads from the solution and the supernatant was removed. C1 buffer and 20-160 µg of capture antibody, totaling a volume of 500 µL, were added and the beads were resuspended by gentle vortexing. After adding 500 µL C2 buffer, the beads were gently vortexed, and the tubes were placed on a HulaMixer™ (ThermoFisher Scientific) for conjugation at 37°C for 19 ± 0.5 h. The tubes were briefly centrifuged at 300×g for 10 seconds, placed on the magnetic rack for 1 minute to separate the beads from the solution, and the supernatant was removed. The beads were washed using four cycles of resuspension and pulse centrifugation in Dynabeads wash solutions containing 0.05% Tween 20 (H5152, Promega, Madison, WI, USA), followed by three washes in 1X PBS (10010072, ThermoFisher Scientific) with 0.1% BSA (0332, VWR, Radnor, PA, USA) and 0.1% Pluronic F-68 (P1169, Spectrum Chemical, New Brunswick, NJ, USA). The antibody-functionalized magnetic beads were resuspended in 1 mL of 1X PBS, pH 7.4 with 50% glycerol (49767, Sigma-Aldrich), 0.1% BSA, and 0.1% Pluronic F-68 and stored at -20°C.

DNA-conjugated antibody preparation Antibodies were buffer exchanged and purified, if needed, as described above. Antibodies were subsequently conjugated to DNA using a copper-free click chemistry reaction. Oligonucleotides (Supplementary Table 2) were synthesized and purchased as single-stranded DNA (ssDNA) from Integrated DNA Technologies, Inc. (Coralville, IA, USA) or Bio-Synthesis, Inc. (Lewisville, TX, USA). Before performing the conjugation, the ssDNA was annealed to form a double-stranded product. Complementary ssDNA strands (Strands 1+3 and Strand 2+4, Table 1) were mixed to a final concentration of 50 µM in 1X PBS and annealed using a standard thermal cycler (Applied Biosystems, ProFlex PCR System). The annealing protocol was as follows: 95°C for 2 minutes, 90°C for 3 minutes, and then reducing the temperature by 5°C in 3-minutes intervals until 25°C was reached. Monoclonal antibodies at 1 mg mL^{-1} in 1X PBS were incubated with DBCO-PEG5-NHS Ester linker (A102P, Click Chemistry Tools, Scottsdale, AZ, USA) at a 4:1 molar ratio of crosslinker to antibody for at least 2 h at room temperature. Following incubation, the antibody-linker solution was washed once using a 30 kDa Amicon Ultra-2 centrifugal filter (UFC203024, Sigma-Aldrich). Annealed DNA oligonucleotides with a 5′ azide modification (Integrated DNA Technologies) were added to the activated antibodies at a 3:1 molar ratio of oligonucleotide to antibody and were allowed to react overnight at room temperature. The conjugate concentrations were measured using the Qubit® protein assay kit (Q33211, ThermoFisher Scientific) according to the manufacturer's instructions, and the DNA-conjugated antibodies were stored at 4°C in 1X PBS.

Table 1: Oligonucleotide Sequences for Primers and TaqMan Probes

Name	Sequence (5' to 3')
Strand 1	/Azide/CAGTCTGACACAGCAGTCGTGACTGGCTAGACAGAGGTGT
Strand 2	/Azide/GACCTGACCTACAGTGACCATTGGCTCCTGGTCTCACTAG

Strand 3	/Phos/GAGTACACCTCTGTCTAGCCAGTCACGACTGCTGTGTCAGACTG
Strand 4	/Phos/ACTCCTAGTGAGACCAGGAGCCAATGGTCACTGTAGGTCAGGTC
Primer 1	CAGTCTGACACAGCAGTCGT
Primer 2	GACCTGACCTACAGTGACCA
TaqMan probe	/56-FAM/TGGCTAGAC/ZEN/AGAGGTGTACTCCTAGTGAGA /31ABkFQ/

Assay protocol. Following EV isolation by SEC, a concentrated buffer solution was added to the
purified EVs to obtain working concentrations of 0.1% BSA, 0.1% Pluronic F-68, 100 µg mL^{-1}
salmon sperm DNA (15632011, ThermoFisher), and 0.1% ProClin 950 (46878-U, Sigma-Aldrich).
Antibody-functionalized magnetic beads directed against Biomarker 1 were diluted 1:2 with 0.1%
BSA, 0.1% Pluronic F-68, 100 µg mL^{-1} salmon sperm DNA, and 0.1% ProClin 950, and 4.8 µL
aliquots of the diluted beads per well (corresponding to 16 µg of beads/well) were added to the
wells of a 96 deep-well KingFisher plate. 250 µL of the diluted, purified EVs was added to each
well of a deep-well KingFisher plate. The sample plate was covered with an adhesive plastic film,
was placed on an Eppendorf ThermoMixer C (Eppendorf, Enfield CT, USA) and incubated at 21°C
for 2 hours, mixing at 1100×g During this incubation, all the wash and reaction plates required for
the assay were prepared and are described in Table 2:

Table 2: Configuration and Composition of the 96-Well KingFisher Plates

Reaction plate type	#	Volume/well	Components
Antibody plate	1	100 µL	DNA-conjugated antibody, 0.1% BSA, 0.1% Pluronic F-68, 100 µg mL^{-1} salmon sperm DNA
TEPBS wash	2	1 mL	1X PBS, pH 7.4, 0.05% Tween 20, 5 mM EDTA
PBST wash	1	1 mL	1X PBS, pH 7.4, 0.05% Tween 20
Ligation plate	1	100 µL	75,000 units mL^{-1} of T7 DNA ligase (New England Biolabs, M0318), 33 mM Tris-HCl, 5 mM MgCl$_2$, 0.5 mM ATP, 0.5 mM DTT 3.75% PEG 6000, 0.5% Tween 20
PBST transfer plate	1	150 µL	1X PBS, pH 7.4, 0.05% Tween 20
qPCR plate	1	30 µL	TaqMan Fast Advanced Master Mix (4444557, Applied Biosystems), 500 nM primers, 200 nM TaqMan probe

After the 2-hour EV-bead-binding incubation was complete, a KingFisher Flex System
(ThermoFisher Scientific) was loaded with the sample plate, the antibody-probe plate, the wash
plates, the ligation reaction plate, and the PBST transfer plate and the custom KingFisher program
initiated. Briefly, the beads are magnetically captured and incubated in the antibody-probe plate
for 32 minutes and 50 seconds, then washed twice in TEPBS for three minutes and 30 seconds,
then washed in PBST for three minutes and 30 seconds and placed in the ligase reaction plate for
21 minutes and 40 seconds. Finally, the beads are transferred into a shallow plate containing PBST.
The beads are then manually transferred to a qPCR plate using a KingFisher Flex 96 PCR magnetic
head. qPCR plates were covered with a clear plastic adhesive and placed into QuantStudio Dx
Real-Time PCR instruments (ThermoFisher Scientific) run using the following amplification
conditions: 50°C for 2 minutes, 95°C for 2 minutes, 40 cycles at 95°C for 5 seconds, and 60°C for
20 seconds. See Table 2 for primer and TaqMan probe sequences. All samples were run in either
duplicate or triplicate and those showing a Ct spread >3 among the replicates were excluded from
subsequent data analysis.

Detergent disruption of EVs. To demonstrate the assay is EV-based, a detergent-sensitivity
experiment was performed. Exposure of EVs to detergent would disrupt their membranes, resulting
in loss of MH assay signal. As described in the paragraph above, EVs purified with SEC were
diluted in a buffer containing 0.1% BSA, 0.1% Pluronic F-68, 100 µg mL^{-1} salmon sperm DNA,
and 0.1% ProClin 950. Antibody-functionalized beads were used to capture the purified EVs in a
2-hour incubation under agitation. The beads with the captured EVs were then transferred, using
the KingFisher Flex system, to deep-well plates containing 0.1% BSA, 0.1% Pluronic F-68, 100
µg mL^{-1} salmon sperm DNA, 0.1% ProClin 950, and 0.075% Triton X-100 (T8787, Sigma-
Aldrich). The plates were incubated for 30 minutes, at 21°C, under agitation (1100×g) in a
ThermoMixer C instrument to allow for EV lysis. After incubation with detergent, the deep-well
plate containing the beads were washed and read out with detection antibodies to either
tetraspanins or the indicated biomarkers as described above (Assay Protocol).

Transmission Electron Microscopy of cell-line and plasma EVs.
Cancer cell-line and HGSOC cancer EVs were prepared as described above and sent to Alpha
Nano Tech (400 Park Offices Drive, STE 108, Morrisville, NC, 27709) in 1X PBS, on dry ice, for
transmission electron microscope imaging using a Gatan Orius SC1000 CCD camera with Gatan
Microscopy Suite 3.0 software. Copper-carbon formvar grids were glow discharged immediately
prior to loading with the sample. Samples were processed undiluted. Grids were floated on 10 µL
sample drop for 10 minutes, washed two times with water by floating on the drop of water for 30
seconds, and negatively stained with 2% uranyl acetate by floating on the drop of stain for 30
seconds. The grid was blot dried with Whatman paper and imaged with Jeol 1230 electron
microscope. Pixel sizes for 40,000X, 80,000X, and 120,000X are 1.28 nm, 0.7 nm, and 0.47 nm,
respectively.

Super-resolution microcopy. EV biomarker colocalization was evaluated with direct Stochastic
Optical Reconstruction Microscopy (dSTORM) using an ONI Nanoimager S and an EV Profiler
Kit (Oxford Nanoimaging Inc., Oxford, UK). Antibodies against BST2, FOLR1, and MUC1 were
conjugated to fluorophores using the Zip Alexa Fluor™ Rapid Antibody Labeling Kit from
ThermoFisher Scientific with Alexa Fluor 488 (Z11233), Alexa Fluor 647 (Z11235), or Alexa
Fluor 555 (Z11234), respectively. EVs collected from conditioned media from COV413A cells
were purified by SEC as described above, characterized by NTA, then captured on imaging slides
using S4 protein phosphatidylserine-dependent, protein-independent capture following the
provided protocol (EVProfiler Kit v 1.5). Briefly, the microfluidic imaging chip was functionalized
with S4 to capture phosphatidylserine containing vesicles then incubated with 5.6x10^8 COV413A-
derived EVs. After allowing the EVs to bind the surface, an antibody cocktail consisting of ONI
N1 solution and 1X PBS containing 0.05% Tween with MUC1-AF647 (1:200), FOLR1-AF555
(1:500), and BST2-AF488 (1:50), was used to stain S4-bound EVs. After washing, the chip was
prepared and imaged. A total of 7500 frames at 30 ms exposure time were collected with 1500
frames at 640 nm, 3000 frames at 561 nm and 3000 frames at 488 nm at 30%, 50% and 50% laser
power respectively. The CODI software package (https://alto.codi.bio/, release date June 2023)
was used for subsequent image analysis. EVs that were triple positive for MUC1, BST2, and
FOLR1 were defined as clusters identified by density-based clustering algorithm (DBSCAN) with
a radius of ≤80 nm and with ≥8 localizations of each biomarker.

Statistical Analysis. Data is presented as PCR cycle threshold (Ct). When assay data was
collapsed to a single value, the median of replicates was displayed. Cell line gene expression and

mass spectrometry data was obtained from the Cancer Cell Line Encyclopedia (CCLE). Mass
spectrometry data was not available for all cell lines. Heatmap data are ordered by signal strength.
Heatmaps containing multiple data types were row-scaled with a cap of +/- 3 standard deviations
to enable comparisons between different data modalities. When only gene expression was shown,
heatmaps were ordered to match the cell lines displayed in the figure, log2 transcripts per million
(TPM) displayed, and no scaling was used. Correlation analysis between assay signal and gene
expression or mass spectrometry data was done using Spearman's Rank Correlation. Assay
sensitivity at 95% specificity was calculated using the epiR R package (version 2.0.40) and ROC
curves were generated with the R package ROCit (version 2.1.1). All analyses were performed in
R v.4.0.5.
732

SUPPLEMENTAL FIGURES

Supplemental Figure 1: Characterization of human plasma EVs. EVs were enriched from a pool of healthy human
plasma samples by SEC. Individual fractions were collected and subject to NTA. Unconcentrated SEC fractions (EV1,
EV2, EV3) were characterized to determine particle counts (A) and overall size distribution (B) of the enriched
nanoparticles. (A) Plot showing nanoparticle counts (grey bars) and protein concentration (black line) for the EV-
containing SEC fractions from human plasma. The pooled collection volume (PCV), consisting of fractions EV1-
EV3, used in the downstream assay, are highlighted. The Pre-EV sample is a sum of the void volume and the wash
fraction containing lipoproteins and larger, non-EV particles (data not shown). (B) NTA size-distribution plot of
individual fractions EV1, EV2, and EV3 demonstrating particles within the size range of 40-400 nm. Buffer alone
(PBS) was used to determine the background. (C) In parallel, concentrated SEC fractions were analyzed by western
blot against normal human plasma (input) and COV413A whole cell extract (WCE) as a positive control. CD9 (left
panel) and CD81 (right panel) are undetectable in the input sample due to the low relative abundance in human plasma.
The PCV is indicated demonstrating CD9 and CD81 presence in fractions EV1-EV3. (D) Representative 80,000x
TEM images of EVs present in each of the three PCV fractions (EV1, EV2, and EV3) are shown, indicating intact
particles between 40 and 400 nm. Scale bars are 500 nm.

752
Supplemental Figure 2: Expression of EV and Ovarian Cancer Biomarkers in EV Fractions. Full-panel blots of
data shown in Figure 2 and additional western blot analysis of whole cell extract (WCE) from COV413A cells and
SEC purified fractions from COV413A conditioned media (EV1-EV6) using antibodies against EV surface biomarkers
CD9 (A), CD81 (B), and CD63 (C); EV cargo biomarkers Flotillin (D) and GAPDH (E) to demonstrate the EVs were
intact EVs during enrichment; and non-EV internal membrane proteins Tom20 (F) and RPS3 (G) to demonstrate the
EV fractions are the result of EV biogenesis and do not include intracellular vesicles. Total protein staining
(Coomassie) was used to validate SEC separation of the EV fractions from the soluble protein component and was
confirmed by measuring the protein concentration in each fraction (Figure 2). (I, J, K) Full panel blots of data shown
in Figure 2 demonstrating abundance of BST2 (I), FOLR1 (J), and MUC1-C (K) in COV413A EVs. Representative
molecular weight makers with indicated observed molecular weights are shown in L.
763

Supplemental Figure 3: Screening antibodies for functionality and specificity for capture and detection of EVs. Distinct monoclonal antibody clones against FOLR1 (A, B), BST2 (C, D), and MUC1 (E, F) were screened for their ability to capture (A, C, E) and detect (B, D, F) SEC-purified EVs from human cell lines known to be positive (COV413A (+)) or negative (SKMEL1 (-)) for FOLR1, BST2, and MUC1. (A, C, E) Each antibody was independently conjugated to magnetic beads at a fixed density and used to capture cell line EVs in solution. Captured EVs were detected using complementary anti-CD81 dsDNA-conjugated detection antibodies. Antibody clones demonstrating both functionality (greatest signal to noise between the positive and negative cell line) and specificity (low Ct in positive cell line and high Ct in negative cell line) were further tested and optimized (not shown). No EV controls were used to eliminate cross-reactive or "sticky" antibodies. Based on performance in immunocapture and other criteria, selected antibodies were subsequently tested for their ability to function as a probe to detect cell line EVs captured using anti-CD81 conjugated beads. (B, D, F) EVs from COV413A and SKMEL were captured with anti-CD81 bead conjugates and read out in the assay using monoclonal anti-FOLR1 (B), -BST2 (D), or -MUC1 (F) detection antibodies. Antibodies demonstrating a high degree of functionality and specificity were subject to additional testing and optimization (not shown).

781
Supplemental Figure 4: Detergent Sensitivity of Assay Performance. Buffer alone (No EV), EVs from COV413A
or pooled human plasma from late-stage HGSOC patients' plasma (OC Plasma) were assessed using MUC1
capture/FOLR1 + BST2 detection antibody combination as indicated. After capture on MUC1 antibody-bead
conjugates, bead-bound EVs were incubated in buffer alone (-) or buffer containing 0.075% Triton-X100 (+) to disrupt
the EV membranes as described in the methods. After a wash step to remove detergent, intact EVs were detected with
anti-BST2 and anti-FOLR1 antibodies. Note that the presence of detergent results in a significant reduction in assay
signal suggesting intact EVs are required for productive readout.
789
790
791

792
793 **Supplemental Figure 5: Demonstration of Single-EV Colocalization**. EVs from OVKATE (BST2^{high}, FOLR^{high},
794 MUC1^{low}), SW900 (BST2^{high}, FOLR1^{low}, MUC1^{high}), COV413A (BST2^{high}, FOLR1^{high}, MUC1^{high}), and G401
795 (BST2^{high}, FOLR1^{low}, MUC1^{low}) were assayed using the BST2 capture/FOLR1 + MUC1 detection combination (A).
796 EVs from OVKATE and SW900 were assayed independently or mixed 1:1 to determine whether oligo-conjugated
797 antibodies on two independent EVs can "bridge" to generate a PCR signal.
798
799

Capture Ab: MUC1
Detection Abs: BST2 + FOLR1

Control Tested	Days	Mean Ct Value	s.d.	%CV
Unspiked Plasma	16	33.27	0.60	**1.80%**
Plasma + 5x10^6 COV413A EVs	16	23.47	0.22	**0.92%**
Plasma + 2x10^7 COV413A EVs	16	21.51	0.30	**1.38%**
Plasma + 8x10^7 COV413A EVs	16	19.39	0.22	**1.12%**
Plasma + 3.2x10^8 COV413A EVs	16	17.38	0.33	**1.87%**

800
801 **Supplemental Figure 6: Reproducibility and Linearity of the Assay in Cell Line EV-Spiked Human Plasma.**
802 Representative reproducibility (top) and linearity (bottom) of assay performance in COV413A EV spiked human
803 plasma using BST2/FOLR1/MUC1. Mean Ct values represent the mean of duplicate assays evaluated over 16 days.
804 Points for each spike-in represent n=8 technical replicates. The last dilution point in linear range was 1.95 x 10^4 which
805 indicates a linear range over more than 15 Ct.
806

808
Supplemental Figure 7. Correlation of assay signal with biomarker expression in cell line EVs. Heat maps
comparing assay signal (top row), mRNA expression (middle row, log2 TPM), and protein abundance as determined
by mass spectroscopy (bottom row, MS/MS) for the biomarkers tested in this study, BST2 (A), FOLR1 (B), and MUC1
(C). Gene expression and MS/MS data was sourced from the Cancer Cell Line Encyclopedia (CCLE) from the Broad
Institute. The top row (Assay Signal) of each heat map is 40 minus the cycle threshold, the middle row (mRNA) is
log2 transcripts per million (TPM), and the bottom row (MS/MS) is the log2 ratio of tandem mass tag (TMT) reporters
vs control. All data was row scaled to allow comparison across data types. Spearman's rank correlation (r_s) and p-
value (p) for each data set are shown.
817

Salem and Bortolin et. al. (2023) - Colocalization of Cancer Associated Biomarkers on Single Extracellular Vesicles for Early Cancer Detection

818 **Supplemental Table 1: Human cancer cell lines used in this study.**

Cell line name	Vendor	Catalog Number	RRID
A549	ATCC	CCL-185	CVCL_0023
A704	ATCC	HTB-45	CVCL_1065
AsPC1	ATCC	CRL-1682	CVCL_0152
AU565	ATCC	CRL-2351	CVCL_1074
BT-20	ATCC	HTB-19	CVCL_0178
BxPC3	ATCC	CRL-1687	CVCL_0186
COR-L-95	Sigma-Aldrich/ECACC	96020733	CVCL_2418
COV362	Sigma-Aldrich/ECACC	07071910	CVCL_2420
COV413A	Sigma-Aldrich/ECACC	07071905	CVCL_2422
D341	ATCC	HTB-187	CVCL_0018
G-401	ATCC	CRL-1441	CVCL_0270
HCC4006	ATCC	CRL-2871	CVCL_1269
HCT-116	ATCC	CCL-247	CVCL_0291
HT-29	ATCC	HTB-38	CVCL_0320
K-562	ATCC	CCL-243	CVCL_0004
LnCaP (clone FGC)	ATCC	CRL-1740	CVCL_1379
MCF7	ATCC	HTB-22	CVCL_0031
MeWo	ATCC	HTB-65	CVCL_0445
NCI-H146	ATCC	HTB-173	CVCL_1473
NCI-H1781	ATCC	CRL-5894	CVCL_1494
NCI-H441	ATCC	HTB-174	CVCL_1561
NCI-H520	ATCC	HTB-182	CVCL_1566
OVCAR3	ATCC	HTB-161	CVCL_0465
OVISE	JCRB	JCRB1043	CVCL_3116
OVKATE	JCRB	JCRB1044	CVCL_3110
OVSAHO	JCRB	JCRB1046	CVCL_3114
PC-3	ATCC	CRL-1345	CVCL_0035
SK-MEL-1	ATCC	HTB-67	CVCL_0068
SK-MES-1	ATCC	HTB-58	CVCL_0630
SK-OV-3	ATCC	HTB-77	CVCL_0532
SUP-M2	DSMZ	ACC 509	CVCL_2209
SW900	ATCC	HTB-59	CVCL_1731
T47D	ATCC	HTB-133	CVCL_0553
T84	ATCC	CCL-248	CVCL_0555

820
821 Vendor Abbreviations: ATCC, American Type Culture Collection; ECACC, European Collection
822 of Authenticated Cell Cultures; JCRB, Japanese Collection of Research Bioresources; DSMZ,
823 Deutsche Sammlung von Mikroorganismen und Zellkulturen (German Collection of
824 Microorganisms and Cell Cultures).

825   **Supplemental Table 2: Antibodies used for western blot in this study.**

Target	Antibody Clone	Catalog #	Vendor	Dilution	RRID
CD63	E1W3T	52090	CST	1:1000	AB_2924771
CD81	D3N2D	56039	CST	1:1000	AB_2924772
CD9	D801A	13174	CST	1:1000	AB_2798139
Flotillin	D2V7J	26836	CST	1:1000	AB_2773040
BST2	D5V5Z	19277	CST	1:1000	AB_2798815
FOLR1	EPR20277	ab235140	Abcam	1:1000	AB_2924727
TOM20	D8T4N	42406	CST	1:1000	AB_2687663
RPS3	D50G7	9538	CST	1:1000	AB_10622028
CD81	E2K9V	52892	CST	1:1000	AB_2924773
GAPDH	D16H11	5174	CST	1:1000	AB_10622025
MUC 1-C	D5K9I	16564	CST	1:1000	AB_2798765
Anti-Rabbit IgG, HRP-linked	Polyclonal	7074	CST	1:1000	AB_2099233
Anti-Mouse IgG, HRP-linked	Polyclonal	7076	CST	1:1000	AB_330924
Anti-Biotin IgG, HRP-linked	Polyclonal	7075	CST	1:1000	AB_10696897

828   Vendor Abbreviations: CST, Cell Signaling Technology

829 **Supplemental Table 3: Antibodies used for capture and/or detection in this study.**
830

Name	Antibody Clone	Vendor	Catalog #	RRID
Recombinant Anti-BST2/Tetherin antibody	EPR20202-169	Abcam	ab243563	AB_2941823
Anti-BST2/Tetherin antibody [EPR20202-150] - BSA and Azide free	EPR20202-150	Abcam	ab243561	AB_2941851
Anti-human CD317	Y129 (HM1.24)	BD Biosciences	Custom order	Custom
BST2 Antibody (4F6)	4F6	Novus Biologicals	NBP2-29622	AB_2941840
BST2 monoclonal antibody (3H4)	3H4	Invitrogen	H00000684-M02	AB_1136901
Purified anti-human CD317 (BST2, Tetherin) Antibody	RS38E	Bio Legend	348402	AB_10588013
CD317 (BST2, PDCA-1) Monoclonal Antibody (26F8), Functional Grade, eBioscience™	26F8	Invitrogen Antibodies	16-3179-82	AB_1518775
Anti-BST2/Tetherin antibody [EPR20202-169] - BSA and Azide free	EPR20202-169	Abcam	ab243563	AB_2941823
Recombinant Anti-BST2/Tetherin antibody [EPR23597-202] - BSA and Azide free	EPR23597-202	Abcam	ab272175	AB_2941830
Anti-BST2/Tetherin antibody - BSA and Azide free (Capture)	EPR20202-150	Abcam	ab242967	AB_2941821
Anti-BST2/Tetherin antibody - BSA and Azide free (Detector)	EPR20202-169	Abcam	ab243001	AB_2941822
Folate Receptor alpha monoclonal antibody (548908)	548908	ThermoFisher	MA5-23917	AB_2609390
Anti-Folate Binding Protein/FBP antibody [EPR23387-276], BSA and Azide free	EPR23387-276	Abcam	ab273159	AB_2941825
Recombinant Anti-Folate Binding Protein/FBP antibody (Detector) [EPR23387-239]	EPR23387-239	Abcam	ab278006	AB_2941829
Recombinant Anti-Folate Binding Protein/FBP antibody (Capture) [EPR23387-232]	EPR23387-232	Abcam	ab278005	AB_2941828
Human FOLR1 Antibody	548908	R&D Systems	MAB5646	AB_2278620
FOLR1 Antibody (2G5C12)	2G5C12	Novus Biologicals	NBP2-61773	AB_2941842
Anti-Folate Binding Protein / FBP antibody [EPR20277] - BSA and Azide free	EPR20277	Abcam	ab235140	AB_2924727
Anti-Human folate receptor 1 Recombinant Antibody (Farletuzumab)	Farletuzumab	Creative Biolabs	TAB-113	AB_2941836
Monoclonal Mouse anti Human FOLR1 / Folate Receptor Alpha Antibody	3G12B7	LSBio	LS-C682639	AB_2941839
Anti-Folate Binding Protein / FBP antibody [LK26]	LK26	Abcam	ab3361	AB_303740
FOLR1 antibody	2B4B7	Proteintech	60307-1-Ig	AB_2881421
Human MUC-1 Antibody	604804	R&D Systems	MAB6298	AB_10640403
Anti-MUC1 antibody [EP1024Y] - Low endotoxin, Azide free	EP1024Y	Abcam	ab218998	AB_2941819
MUC1 (D9O8K) XP® Rabbit mAb	D9O8K	Cell Signaling Technology	14161	AB_2798408
MUC1 (VU4H5) Mouse mAb	VU4H5	Cell Signaling Technology	4538	AB_2148549
Mouse anti Human CD227	VU-3C6	Bio-Rad	6378-0135	AB_620067
MUC1 Monoclonal Antibody (GP1.4)	GP1.4	Invitrogen Antibodies	MA1-06503	AB_558126
MUC1 Monoclonal Antibody (GP1.4)	GP1.4	Novus Biologicals	NBP2-33174	AB_2941841
Epithelial Membrane Antigen (Concentrate)	E29	DAKO	M061301-2 (1 mL)	AB_2148557
Anti-Human MUC1 Recombinant Antibody (1B2)	1B2	Creative BioLabs	TAB-439MZ	AB_2941838
Anti-MUC1 [12D10]	12D10	Absolute Antibody	Ab02843-1.1	AB_2941833
Human MUC-1 Antibody	2336A	R&D Systems	MAB62981	AB_2941844
CD227 antibody	C595 (NCRC48)	C595 (NCRC48)	Bio-Rad	MCA1742	AB_322211
MUC1 Antibody (SM3)	SM3	Novus Biologicals	NB120-22711	AB_789083
MUC1 monoclonal antibody (115D8)	115D8	Invitrogen	MA1-35039	AB_1070963
Anti-MUC1 Antibody [115D8] - BSA and Azide free	115D8	Antibodies.com	A252591	AB_2941835
Anti-Human MUC1 (CD227) Antibody, Clone 16A	16A	Stem cell Technologies	60155	AB_2941845
Purified anti-human CD227 (MUC-1) Antibody	16A	Bio Legend	355602	AB_2561642
Anti-Human MUC1 Therapeutic Antibody (5E5)	5E5	Creative Biolabs	TAB-418MZ	AB_2941837
Anti-MUC1 antibody [HMFG1 (aka 1.10.F3)] - BSA and Azide free	HMFG1 (aka 1.10.F3)	Abcam	ab230298	AB_2941820
Anti-MUC1 HMFG1 (1.10.F3)	HMFG1	Absolute Antibody	AB02206-1.1	AB_2941832
Anti-MUC1 antibody [HMFG2]	HMFG2	Abcam	ab245693	AB_2941824
Human MUC-1 Antibody	604804	R&D Systems	MAB6298	AB_10640403

832 **Supplemental Table 4: Patient Sample Demographics (Attached)**

References

Auber M, Svenningsen P. 2022. An estimate of extracellular vesicle secretion rates of human blood cells. *J Extracell Biol* **1**. doi:10.1002/jex2.46

Avanzini S, Kurtz DM, Chabon JJ, Moding EJ, Hori SS, Gambhir SS, Alizadeh AA, Diehn M, Reiter JG. 2020. A mathematical model of ctDNA shedding predicts tumor detection size. *Sci Adv* **6**. doi:10.1126/sciadv.abc4308

Bebelman MP, Smit MJ, Pegtel DM, Baglio SR. 2018. Biogenesis and function of extracellular vesicles in cancer. *Pharmacol Ther* **188**:1–11.

Beer TM. 2021. Examining developments in multicancer early detection: highlights of new clinical data from recent conferences. *Am J Manag Care* **27**:S347–S355.

Cabús L, Lagarde J, Curado J, Lizano E, Pérez-Boza J. 2022. Current challenges and best practices for cell-free long RNA biomarker discovery. *Biomark Res* **10**:62.

Cizmar P, Yuana Y. 2017. Detection and Characterization of Extracellular Vesicles by Transmission and Cryo-Transmission Electron Microscopy. *Methods Mol Biol* **1660**:221–232.

Crosby D, Bhatia S, Brindle KM, Coussens LM, Dive C, Emberton M, Esener S, Fitzgerald RC, Gambhir SS, Kuhn P, Rebbeck TR, Balasubramanian S. 2022. Early detection of cancer. *Science* **375**:eaay9040.

Dang DK, Park BH. 2022. Circulating tumor DNA: current challenges for clinical utility. *J Clin Invest* **132**. doi:10.1172/JCI154941

Desai A, Lovly CM. 2023. Challenges in the implementation of ultrasensitive liquid biopsy approaches in precision oncology. *J Immunother Cancer* **11**:e006793.

Doyle LM, Wang MZ. 2019. Overview of Extracellular Vesicles, Their Origin, Composition, Purpose, and Methods for Exosome Isolation and Analysis. *Cells* **8**. doi:10.3390/cells8070727

El Bairi K, Al Jarroudi O, Afqir S. 2021. Revisiting antibody-drug conjugates and their predictive biomarkers in platinum-resistant ovarian cancer. *Semin Cancer Biol* **77**:42–55.

Eren E, Leoutsakos J-M, Troncoso J, Lyketsos CG, Oh ES, Kapogiannis D. 2022. Neuronal-Derived EV Biomarkers Track Cognitive Decline in Alzheimer's Disease. *Cells* **11**. doi:10.3390/cells11030436

Ferguson S, Weissleder R. 2020. Modeling EV Kinetics for Use in Early Cancer Detection. *Adv Biosyst* **4**:e1900305.

Ferguson S, Yang KS, Weissleder R. 2022. Single extracellular vesicle analysis for early cancer detection. *Trends Mol Med* **28**:681–692.

Gautam SK, Khan P, Natarajan G, Atri P, Aithal A, Ganti AK, Batra SK, Nasser MW, Jain M. 2023. Mucins as Potential Biomarkers for Early Detection of Cancer. *Cancers* **15**. doi:10.3390/cancers15061640

Ghodasara A, Raza A, Wolfram J, Salomon C, Popat A. 2023. Clinical Translation of Extracellular Vesicles. *Adv Healthc Mater* e2301010.

Giampaolino P, Della Corte L, Foreste V, Vitale SG, Chiofalo B, Cianci S, Zullo F, Bifulco G. 2019. Unraveling a difficult diagnosis: the tricks for early recognition of ovarian cancer. *Minerva Med* **110**:279–291.

Giampaolino P, Foreste V, Della Corte L, Di Filippo C, Iorio G, Bifulco G. 2020. Role of biomarkers for early detection of ovarian cancer recurrence. *Gland Surg* **9**:1102–1111.

Han H, Jiang X. 2014. Overcome support vector machine diagnosis overfitting. *Cancer Inform* **13**:145–158.

Ignatiadis M, Sledge GW, Jeffrey SS. 2021. Liquid biopsy enters the clinic - implementation issues and future challenges. *Nat Rev Clin Oncol* **18**:297–312.

Januchowski R, Sterzyńska K, Zawierucha P, Ruciński M, Świerczewska M, Partyka M,
Bednarek-Rajewska K, Brązert M, Nowicki M, Zabel M, Klejewski A. 2017. Microarray-
based detection and expression analysis of new genes associated with drug resistance in
ovarian cancer cell lines. *Oncotarget* **8**:49944–49958.

Jerabkova-Roda K, Dupas A, Osmani N, Hyenne V, Goetz JG. 2022. Circulating extracellular
vesicles and tumor cells: sticky partners in metastasis. *Trends Cancer Res* **8**:799–805.

Johnsen KB, Gudbergsson JM, Andresen TL, Simonsen JB. 2019. What is the blood concentration
of extracellular vesicles? Implications for the use of extracellular vesicles as blood-borne
biomarkers of cancer. *Biochim Biophys Acta Rev Cancer* **1871**:109–116.

Koshiyama M, Matsumura N, Konishi I. 2017. Subtypes of Ovarian Cancer and Ovarian Cancer
Screening. *Diagnostics (Basel)* **7**. doi:10.3390/diagnostics7010012

Kufe DW. 2009. Mucins in cancer: function, prognosis and therapy. *Nat Rev Cancer* **9**:874–885.

Larssen P, Wik L, Czarnewski P, Eldh M, Löf L, Ronquist KG, Dubois L, Freyhult E, Gallant CJ,
Oelrich J, Larsson A, Ronquist G, Villablanca EJ, Landegren U, Gabrielsson S, Kamali-
Moghaddam M. 2017. Tracing Cellular Origin of Human Exosomes Using Multiplex
Proximity Extension Assays. *Mol Cell Proteomics* **16**:502–511.

Li Yuchen, He X, Li Q, Lai H, Zhang H, Hu Z, Li Yan, Huang S. 2020. EV-origin: Enumerating
the tissue-cellular origin of circulating extracellular vesicles using exLR profile. *Comput*
Struct Biotechnol J **18**:2851–2859.

Liu J, Chen Y, Pei F, Zeng C, Yao Y, Liao W, Zhao Z. 2021. Extracellular Vesicles in Liquid
Biopsies: Potential for Disease Diagnosis. *Biomed Res Int* **2021**:6611244.

Lobb RJ, Becker M, Wen SW, Wong CSF, Wiegmans AP, Leimgruber A, Möller A. 2015.
Optimized exosome isolation protocol for cell culture supernatant and human plasma. *J*
Extracell Vesicles **4**:27031.

Lone SN, Nisar S, Masoodi T, Singh M, Rizwan A, Hashem S, El-Rifai W, Bedognetti D, Batra
SK, Haris M, Bhat AA, Macha MA. 2022. Liquid biopsy: a step closer to transform
diagnosis, prognosis and future of cancer treatments. *Mol Cancer* **21**:79.

Lundberg M, Eriksson A, Tran B, Assarsson E, Fredriksson S. 2011. Homogeneous antibody-based
proximity extension assays provide sensitive and specific detection of low-abundant
proteins in human blood. *Nucleic Acids Res* **39**:e102.

Mariotto AB, Enewold L, Zhao J, Zeruto CA, Yabroff KR. 2020. Medical Care Costs Associated
with Cancer Survivorship in the United States. *Cancer Epidemiol Biomarkers Prev*
29:1304–1312.

McGarvey N, Gitlin M, Fadli E, Chung KC. 2022. Increased healthcare costs by later stage cancer
diagnosis. *BMC Health Serv Res* **22**:1155.

Menon U, Gentry-Maharaj A, Burnell M, Singh N, Ryan A, Karpinskyj C, Carlino G, Taylor J,
Massingham SK, Raikou M, Kalsi JK, Woolas R, Manchanda R, Arora R, Casey L,
Dawnay A, Dobbs S, Leeson S, Mould T, Seif MW, Sharma A, Williamson K, Liu Y,
Fallowfield L, McGuire AJ, Campbell S, Skates SJ, Jacobs IJ, Parmar M. 2021. Ovarian
cancer population screening and mortality after long-term follow-up in the UK
Collaborative Trial of Ovarian Cancer Screening (UKCTOCS): a randomised controlled
trial. *Lancet* **397**:2182–2193.

Menon U, Gentry-Maharaj A, Hallett R, Ryan A, Burnell M, Sharma A, Lewis S, Davies S, Philpott
S, Lopes A, Godfrey K, Oram D, Herod J, Williamson K, Seif MW, Scott I, Mould T,
Woolas R, Murdoch J, Dobbs S, Amso NN, Leeson S, Cruickshank D, McGuire A,
Campbell S, Fallowfield L, Singh N, Dawnay A, Skates SJ, Parmar M, Jacobs I. 2009.
Sensitivity and specificity of multimodal and ultrasound screening for ovarian cancer, and

stage distribution of detected cancers: results of the prevalence screen of the UK
Collaborative Trial of Ovarian Cancer Screening (UKCTOCS). *Lancet Oncol* **10**:327–340.

Niemeyer CM, Adler M, Wacker R. 2005. Immuno-PCR: high sensitivity detection of proteins by
nucleic acid amplification. *Trends Biotechnol* **23**:208–216.

Osteikoetxea X, Sódar B, Németh A, Szabó-Taylor K, Pálóczi K, Vukman KV, Tamási V, Balogh
A, Kittel Á, Pállinger É, Buzás EI. 2015. Differential detergent sensitivity of extracellular
vesicle subpopulations. *Org Biomol Chem* **13**:9775–9782.

Pavlik EJ, van Nagell JR Jr. 2013. Early detection of ovarian tumors using ultrasound. *Womens*
Health **9**:39–55; quiz 56–7.

Pinsky PF. 2014. Assessing the benefits and harms of low-dose computed tomography screening
for lung cancer. *Lung Cancer Manag* **3**:491–498.

Siegel RL, Miller KD, Wagle NS, Jemal A. 2023. Cancer statistics, 2023. *CA Cancer J Clin* **73**:17–
48.

Stiller C, Viktorsson K, Paz Gomero E, Hååg P, Arapi V, Kaminskyy VO, Kamali C, De Petris L,
Ekman S, Lewensohn R, Karlström AE. 2021. Detection of Tumor-Associated Membrane
Receptors on Extracellular Vesicles from Non-Small Cell Lung Cancer Patients via
Immuno-PCR. *Cancers* **13**. doi:10.3390/cancers13040922

Sung H, Ferlay J, Siegel RL, Laversanne M, Soerjomataram I, Jemal A, Bray F. 2021. Global
Cancer Statistics 2020: GLOBOCAN Estimates of Incidence and Mortality Worldwide for
36 Cancers in 185 Countries. *CA Cancer J Clin* **71**:209–249.

Tavoosidana G, Ronquist G, Darmanis S, Yan J, Carlsson L, Wu D, Conze T, Ek P, Semjonow A,
Eltze E, Larsson A, Landegren UD, Kamali-Moghaddam M. 2011. Multiple recognition
assay reveals prostasomes as promising plasma biomarkers for prostate cancer. *Proc Natl*
Acad Sci U S A **108**:8809–8814.

Théry C, Witwer KW, Aikawa E, Alcaraz MJ, Anderson JD, Andriantsitohaina R, Antoniou A,
Arab T, Archer F. 2018. Minimal information for studies of extracellular vesicles 2018
(MISEV2018): a position statement of the International Society for Extracellular Vesicles
and update of the MISEV2014 guidelines. *J Extracell Vesicles* **7**:1535750.

Trinidad C, Pathak H, Cheng S, Tzeng S-C, Madan R, Sardiu M, Bantis L, Deighan C, Jewell A,
Zeng Y, Godwin A. 2023. Lineage specific extracellular vesicle-associated protein
biomarkers for the early detection of high grade serous ovarian cancer. *Res Sq*.
doi:10.21203/rs.3.rs-2814022/v1

US Preventive Services Task Force. 2009. Screening for breast cancer: U.S. Preventive Services
Task Force recommendation statement. *Ann Intern Med* **151**:716–26, W-236.

US Preventive Services Task Force, Davidson KW, Barry MJ, Mangione CM, Cabana M, Caughey
AB, Davis EM, Donahue KE, Doubeni CA, Krist AH, Kubik M, Li L, Ogedegbe G, Owens
DK, Pbert L, Silverstein M, Stevermer J, Tseng C-W, Wong JB. 2021a. Screening for
Colorectal Cancer: US Preventive Services Task Force Recommendation Statement. *JAMA*
325:1965–1977.

US Preventive Services Task Force, Grossman DC, Curry SJ, Owens DK, Barry MJ, Davidson
KW, Doubeni CA, Epling JW Jr, Kemper AR, Krist AH, Kurth AE, Landefeld CS,
Mangione CM, Phipps MG, Silverstein M, Simon MA, Tseng C-W. 2018. Screening for
Ovarian Cancer: US Preventive Services Task Force Recommendation Statement. *JAMA*
319:588–594.

US Preventive Services Task Force, Krist AH, Davidson KW, Mangione CM, Barry MJ, Cabana
M, Caughey AB, Davis EM, Donahue KE, Doubeni CA, Kubik M, Landefeld CS, Li L,
Ogedegbe G, Owens DK, Pbert L, Silverstein M, Stevermer J, Tseng C-W, Wong JB.

2021b. Screening for Lung Cancer: US Preventive Services Task Force Recommendation Statement. *JAMA* **325**:962–970.

van Nagell JR Jr, DePriest PD, Ueland FR, DeSimone CP, Cooper AL, McDonald JM, Pavlik EJ, Kryscio RJ. 2007. Ovarian cancer screening with annual transvaginal sonography: findings of 25,000 women screened. *Cancer* **109**:1887–1896.

van Niel G, D'Angelo G, Raposo G. 2018. Shedding light on the cell biology of extracellular vesicles. *Nat Rev Mol Cell Biol* **19**:213–228.

Varaganti P, Buddolla V, Lakshmi BA, Kim Y-J. 2023. Recent advances in using folate receptor 1 (FOLR1) for cancer diagnosis and treatment, with an emphasis on cancers that affect women. *Life Sci* **326**:121802.

Veerman RE, Teeuwen L, Czarnewski P, Güclüler Akpinar G, Sandberg A, Cao X, Pernemalm M, Orre LM, Gabrielsson S, Eldh M. 2021. Molecular evaluation of five different isolation methods for extracellular vesicles reveals different clinical applicability and subcellular origin. *J Extracell Vesicles* **10**:e12128.

Weibrecht I, Leuchowius K-J, Clausson C-M, Conze T, Jarvius M, Howell WM, Kamali-Moghaddam M, Söderberg O. 2010. Proximity ligation assays: a recent addition to the proteomics toolbox. *Expert Rev Proteomics* **7**:401–409.

Willms E, Cabañas C, Mäger I, Wood MJA, Vader P. 2018. Extracellular Vesicle Heterogeneity: Subpopulations, Isolation Techniques, and Diverse Functions in Cancer Progression. *Front Immunol* **9**:738.

Yang L-Q, Hu H-Y, Han Y, Tang Z-Y, Gao J, Zhou Q-Y, Liu Y-X, Chen H-S, Xu T-N, Ao L, Xu Y, Che X, Jiang Y-B, Xu C-W, Zhang X-C, Jiang Y-X, Heger M, Wang X-M, Cheng S-Q, Pan W-W. 2022. CpG-binding protein CFP1 promotes ovarian cancer cell proliferation by regulating BST2 transcription. *Cancer Gene Ther* **29**:1895–1907.

Yu D, Li Y, Wang M, Gu J, Xu W, Cai H, Fang X, Zhang X. 2022. Exosomes as a new frontier of cancer liquid biopsy. *Mol Cancer* **21**:56.

Yu W, Hurley J, Roberts D, Chakrabortty SK, Enderle D, Noerholm M, Breakefield XO, Skog JK. 2021. Exosome-based liquid biopsies in cancer: opportunities and challenges. *Ann Oncol* **32**:466–477.