Measuring single-cell susceptibility to antibiotics within monoclonal bacterial populations

Lena Le Quellec[1,2†]**, Andrey Aristov**[1†]**, Salomé Gutiérrez Ramos**[1,2]**,
Gabriel Amselem**[1,2]**, Julia Bos**[3]**, Zeynep Baharoglu**[3]**, Didier Mazel**[3]**,
Charles N. Baroud**[1,2]*

*For correspondence:
charles.baroud@pasteur.fr (CNB)

†These authors contributed
equally to this work

[1]Physical Institut Pasteur, Université Paris Cité, Microfluidics and
Bioengineering Unit,75015 Paris, France; [2]LadHyX, CNRS, Ecole
polytechnique, Institut polytechnique de Paris, 91120, Palaiseau, France;
[3]Institut Pasteur, Université Paris Cité, CNRS UMR3525, Bacterial
Genome Plasticity Unit,F-75015 Paris, France

Abstract   Given the emergence of antimicrobial drug resistance, it is critical to
understand the heterogeneity of response to an antibiotic within a population of cells.
Since the drug can exert a selection pressure that leads to the emergence of resistant
phenotypes. To date, neither bulk nor single-cell methods are able to link the
heterogeneity of single-cell susceptibility to the population-scale response to antibiotics.
Here we present a platform that measures the ability of individual *E. coli* cells to form
small colonies at different ciprofloxacin concentrations, by using anchored microfluidic
drops and an image and data analysis pipelines. The microfluidic results are
benchmarked against classical microbiology measurements of antibiotic susceptibility,
showing an agreement between the pooled microfluidic chip and replated bulk
measurements. Further, the experimental likelihood of a single cell to form a colony is
used to provide a probabilistic antibiotic susceptibility curve. In addition to the
probabilistic viewpoint, the microfluidic format enables the characterization of
morphological features over time for a large number of individual cells. This pipeline
can be used to compare the response of different bacterial strains to antibiotics with
different action mechanisms.

Introduction

Antimicrobial resistance is considered by the World Health Organization as one of the
biggest threats to public health [25]. Developing new tools and methods to better under-
stand bacterial resistance is becoming necessary. When addressing the question of bac-
terial response to antibiotics, most microbiology studies report the Minimum Inhibitory
Concentration (MIC) at which the cells stop growing, for a given initial inoculum size (usu-
ally 10^5 cells) and after a growth time of over 24 hours. However, interpreting these MIC

measurements is far from trivial and could indicate many different phenomena taking place in the cultures (high variability, inoculum effect, mechanism of action of the antibiotic) [16]. Indeed, the response of individual cells, which may then lead to colonies, can display large heterogeneity [11, 12]. For this reason the MIC is sometimes complemented with a more precise measurement, the Minimum Bactericidal Concentration (MBC). The process of determining the MBC is heavy and time-consuming and, like the MIC, only gives information about the antibiotic susceptibility on the population level. This has motivated significant effort to measure the bacterial response at a more microscopic level.

For instance, a single-cell MIC was estimated by measuring the MIC for different inoculum sizes and extrapolating the value to a single cell [5]. Classical laboratory methods, however, are difficult to scale to single-cell manipulation, both in terms of the volumes of interest and also the number of experiments that are required to obtain a significant number of replicates. In contrast, the development of microfabrication methods and microfluidics has allowed measurements to be made on individual cells in controlled environments [6]. In this area, two platforms which address questions of antibiotic response have emerged. First the so-called mother machine and its variations, where individual cells are trapped in thin channels and observed over a large number of generations [28]. These devices rely on the tracking of the initial *mother* cells by time-lapse microscopy, while removing its daughter cells as they push out of the microchannels. By fine analysis of the images under different antibiotic treatments it is possible to learn about relations between mother and daughter cells [9], or to detect the effect of rare mutations on the fitness at the single-cell level [26].

In parallel, the field of droplet microfluidics has allowed studies of a different kind. By encapsulating one or a few bacterial cells within water-in-oil droplets, the development of small colonies from individual cells [15], or the signature of their metabolism [10], can be detected with optical readouts. The addition of antibiotics in solution within the droplets can then be used to determine the bacteria's susceptibility. This basic principle has, in recent years, been developed in two main directions. Either to improve the simplicity of use [13, 19] or to improve the precision of the measurements [22, 23, 27]. These droplets approaches have the potential to be transferred to clinical studies, as reviewed in recent papers [18, 30, 14].

Although these droplet methods constitute important milestones, they suffer from several drawbacks: First, they require specific and sophisticated equipment including precise flow control systems to ensure droplet size homogeneity, as well as high-speed electronics, lasers, and data acquisition, to perform the measurements on flowing droplets. Second, the latest methods do not allow the droplets to be followed in time or to relate the final state to the initial state of the droplets. Finally, the link between the single-cell measurements and the classical biological measurements has never been explicitly tested. As such, it is difficult to relate the droplet-based measurements with the vast quantity of data obtained in traditional experiments.

Here, we present an open-access microfluidic platform that addresses some of these issues. The platform is based on *rails and anchors* that were introduced a few years ago [1]. Those droplets are formed within microfabricated wells and remain stationary for the duration of the experiment, including for the observation of biological processes within them [2, 3]. As such, the platform only requires simple microfabrication, low-precision flow control and allows to complete time-lapse measurements. The microflu-

idic setup is augmented with an original and dedicated image acquisition and analysis pipeline that extracts the relevant information from the chips in an automated manner. By providing the image and data analysis as open-source code, the platform will be easy to integrate in most academic laboratories. Moreover, the current study addresses the interpretation of the biological measurements for the first time. By doing so it links the droplet-level approach, used in most droplet-based experiments, with a single-cell analysis. This analysis is used to obtain unique measurements of the single-cell susceptibility to antibiotics.

Results

Microfluidic platform and initial observations

The microfluidic device used in this study is based on a geometry described previously [3, 4]. It consists of a triangular array of 501 individual wells, or anchors, that can each hold a single aqueous droplet (Fig. 1a). The chip format is well suited for imaging on an inverted microscope, either using wide-field illumination or confocal mode, as discussed below.

Loading the chip and distributing the bacterial solution into droplets is straightforward and takes about 5 min (see Method for the detailed steps, and Fig. 1 b, c, d). Briefly, the device is first entirely filled with oil (FC40, 3M) containing 0.5% surfactant (FluoSurf, Emulseo, France). Then, the aqueous suspension of bacteria is flowed through the device. In a last step, the oil is flowed again through the device, which leads to the formation and immobilization of droplets containing bacteria, directly on the anchors. The volume of each droplet is 2 nl and is primarily determined by the volume of the anchor [2]. This loading procedure is simple to set up and can be mastered in only a couple of trials. Indeed, the droplet formation step is robust to fluctuations in flow rate. It can be performed using syringe pumps or using hand-held syringes. Also note that the whole loading procedure is performed while the output is connected to a waste tube, and the bacterial suspension is constantly encapsulated in an inert oil layer. This makes the protocol suitable in principle to handling pathogens, since the bacterial solution remains in a closed circuit.

E. coli W3110 with a fluorescent reporter (RFP) in the ptac site (ptac::RFP) allows us to detect bacterial cells and the growth of the colonies using the red fluorescent protein (RFP) . A sample fluorescence image of the trapping area of the chip, acquired after 24 hours of incubation, is shown in Fig. 1e. The bacteria at this stage form bright fluorescent colonies that can readily be imaged using epifluorescence or confocal microscopy. Empty droplets are also present on the device, since the cells are randomly distributed. The number of empty droplets is related to the average number of bacteria per drop, which is in turn related to the initial concentration of the bacterial suspension. Starting from different initial dilutions, therefore, allows us to tune the average number of cells per droplet and the number of positive droplets within the chip in the absence of antibiotics.

Before moving on to testing the effects of antibiotics in the microfluidic device, we to benchmark the bacterial fitness in the microfluidic device compared with standard multi-well plate experiment. This is done by following the growth of the fluorescent intensity in the droplets, using time-lapse epifluorescence microscopy, while performing in parallel a standard growth-curve measurement on a fluorescence plate reader from the same batch culture. The growth curves for the two cases are shown in Fig. 1f. The curves from the microfluidic device show a large variability between individual droplets, as a result

Figure 1. (a) Microfluidic chip design. Left: a unit cell consists of a triangular arrangement of square anchors, each of size $a = 120$ μm and distance $b = 360$ μm. Right: design of the full chip, which includes a main chamber with 501 anchors, 2 input channels on the right and 1 outlet channel on the left. **(b)** Side-view (not to scale), showing the channel height $c = 30$ μm and anchor depth $d = 100$ μm. The loading protocol begins by filling the chamber with oil, then **(c)** replacing the oil with bacterial suspension, and **(d)** breaking the bacterial suspension into individual droplets, anchored in their respective wells. **(e)** Z-projection of a confocal stack of the chip after 24 h incubation (scalebar: 500 μm). **(f)** Growth curves in 501 individual droplets on the chip (top), and in each well of a 96-well plate (bottom). **(g)** Measured growth rates for bacteria in the microfluidics chip (2 replicates), and in the 96-well plate (3 replicates). Growth rates were obtained by fitting an exponential function to the growth curves during the first 10 hours of growth. P-value=0.24 obtained with Welch's t-test for indetependent samples.

of distribution of the number of initial cells and of the cell-to-cell variability [7]. In contrast, the multiwell plate experiments are insensitive to these stochastic elements and grow in a reproducible and regular manner. Fitting the individual growth curves with an exponential function for the first 10 hours of growth shows that the difference in growth rates in the microfluidic device and in the multiwell plates is statistically not significant, see Fig. 1g. Hence, the microfluidic results can be compared to standard microbiology techniques.

Imaging and analysis pipelines

Once the microfluidic device is loaded, the aim of the experiments is to identify which droplets within the array produce a population of cells after 24h and to link the final state with the initial number of cells in each droplet. These measurements are performed by first imaging the chip shortly after the loading and then after overnight incubation. Image analysis of the initial and final time points then yields a table that identifies each droplet in the array. Each droplet is then associated with measured quantities such as the initial number of cells, its final state, as well as the antibiotic concentration for a given experiment (see schematic in Fig. 2.

As such, the ability to acquire and analyze large amounts of imaging data is fundamental to obtain the antibiotic response curves. While the array format lends itself naturally to measuring droplet contents at different time points, the requirement to detect single cells at early times imposes high-resolution imaging. This runs into data handling limitations associated with the large file sizes and large number of experiments. These different constraints led us to develop automated imaging and analysis pipelines whose implementation was instrumental for obtaining the results below. The imaging steps are described below and the image and data analysis pipelines are provided as open-source packages at the following GitHub repository: https://github.com/BaroudLab/anchor-droplet-chip.

The aim of the pipeline is to generate a table with one line for each droplet, which includes a unique label, the antibiotic concentration, as well as the initial number of cells and final state for the droplet, as described in Fig. 2.

Initial state.
The initial number of cells per droplet is obtained by imaging the freshly loaded chip first in bright-field, and second by acquiring a confocal z-stack of the device in the RFP channel. The bright-field image is used to detect the positions of the droplets, by adjusting for arbitrary shift and tilt of each acquisition (see Methods). In turn, the confocal stack is reduced to a single image by using a maximum projection of the fluorescent intensity. The resulting fluorescent image allows us to count the number of individual bacteria within each droplet (Fig. 2c). The algorithm proved sufficiently robust to perform unsupervised automated analysis of the data from the chips.

Since the cells are expected to follow a Poisson distribution in the droplets [3], a quick quality control is performed at the end of the first scan, by verifying the distribution of number of cells per droplet, and checking that it indeed follows the expected shape, see Fig. 2i. This calculation also allows us to obtain a value of the Poisson parameter λ and to adjust the cell dilution if necessary in order to work in the desired range of λ.

Note that the loading and scanning each take about 10 minutes and the initial image

Figure 2. Imaging and image analysis pipeline. **(a)** The experiment begins by loading several chips with bacterial suspensions at different antibiotic concentrations. **(b)** Overlay of the bright-field image of the chip, and of the maximum projection of a fluorescent confocal z-stack. Images are acquired minutes after loading (0 h). **(c)** The bright-field image is used to identify the alignment with respect to the microscope stage and to create a unique mask around the positions of the individual anchors. **(d)** Detecting the fluorescent peaks in the maximum projection enables to count the initial number of cells per droplet. **(e)**-**(g)** Similar imaging and analysis operations are performed 24 h after the loading, to identify positive and negative droplets. **(h)** The data from the initial and final images are collected in a table that provides a unique label for each droplet, as well as the number of cells at 0 h and at 24 h. Each droplet is assigned a final status of positive or negative, depending on whether the final number of cells is larger than a predefined threshold of 15 cells. **(i)** A quality control step is performed using the single-cell detection at $t = 0$ by comparing the cell number distribution with a Poisson distribution. This leads to an estimate of the Poisson parameter λ for each chip.

₁₇₁ validation occurs within a few minutes as well. This yields a first measurement of the
₁₇₂ loading in under 30 minutes for each microfluidic chip.

Final state.

₁₇₄ The contents of each droplet are later measured at the experimental endpoint, typically
₁₇₅ after an overnight incubation ($t = 18 - 24$ h): each chip is scanned, this time using simple
₁₇₆ epi-fluorescence to optimize time and disk storage. The final image of each chip is re-
₁₇₇ aligned with the template acquired at initial times to identify each droplet, a registration
₁₇₈ step made straightforward by the fact that droplets are anchored at predefined positions.
₁₇₉ Then, cells are detected within each droplet, as shown in Fig. 2e-g. In the current exper-
₁₈₀ iments, we focus on the bottom of the microfluidic device where cells are more likely to
₁₈₁ be detected if the droplet is positive. Different quantities can be obtained from the final
₁₈₂ image as proxies for the ability of cells to grow within the drop. We count the number of
₁₈₃ cells in the final image, although mean or total fluorescence intensity can also be used.
₁₈₄ The main requirement for the measurements is to be sufficiently robust to yield a cutoff
₁₈₅ between positive and negative droplets, a classification which is also included in the data
₁₈₆ table. The final result is a csv table that contains the relevant information on the initial
₁₈₇ and final states of each droplet within each microfluidic device.

₁₈₈ Note that the protocols described above can be modified for particular situations. For
₁₈₉ instance time-lapse microscopy can be performed on the chips to obtain time-resolved
₁₉₀ measurements. Similarly, confocal imaging can be used at later times to obtain a more
₁₉₁ precise cell count or fluorescent intensity, or to identify the morphology of the cells. Al-
₁₉₂ though these cases would require small modification in the pipeline, the main bricks of
₁₉₃ the analysis discussed above can still be used in a modular fashion.

Microfluidic *vs.* microplate antibiogram

₁₉₅ The microfluidics and imaging protocols described above can be combined to obtain an
₁₉₆ antibiotic susceptibility curve, by loading several chips in parallel, using known concen-
₁₉₇ trations of antibiotics and bacteria in each chip. Performing these measurements is sim-
₁₉₈ plified by the standardized microfluidic format and analysis codes, making it possible to
₁₉₉ run six to twelve chips in parallel, each with a different concentration. Confronting the
₂₀₀ microfluidic measurements against standard microbiological techniques proves crucial
₂₀₁ to understand how to interpret the microfluidics data.

₂₀₂ In each experimental run, one chip is loaded with the same bacterial concentration
₂₀₃ as the others but without any antibiotic. This control chip allows us to estimate the value
₂₀₄ of the average number of cells per droplet at initial times, λ, for the given run:

$$\lambda \simeq -\ln(\hat{p}_-), \tag{1}$$

₂₀₅ where $\hat{p}_-$ is the fraction of empty to total droplets on the test chip. This estimate comes
₂₀₆ from the assumption that the initial number of cells per droplet follows a Poisson distri-
₂₀₇ bution [7]. A higher value of λ means that drops contain a larger average number of cells
₂₀₈ initially.

₂₀₉ A typical set of images from the experimental endpoint is shown in Fig. 3a. In these
₂₁₀ images the bright spots correspond to droplets where bacteria grew, while dark spots
₂₁₁ within the regular matrix correspond to droplets that do not contain a sufficient popula-
₂₁₂ tion of cells. These dark positions correspond either to droplets that did not contain any

Figure 3. (a) Antibiogram chips with different concentrations of ciprofloxacin, and an initial average number of cells per droplet $\lambda = 1.3$. Bright spots correspond to droplets where bacterial growth occured. Scale bar: 1 mm. Fraction of positive droplets at the end of the experiment, normalized by the total number of droplets containing at least one bacterium at the beginning of the experiment, for different concentrations of ciprofloxacin. Each color corresponds to a different run, at a different date and with a different value of λ. **(c)** Result of MBC experiments. Bacteria were grown in 96-well plates under different antibiotic concentrations, and starting with different inoculum sizes. The wells were scanned after 24 hours, and the contents of negative wells were replated on antibiotic-free petri dishes for 24 addidtional hours. The number of colonies on the petri dishes after incubation directly represents the number of surviving bacteria in each well. The color corresponds to the number of colonies on the petri dishes. **(d)** Values of the MBC (petri dishes), MIC (96-well plates) and µf·MIC on the chip, where the total number of cells per chip is used as inoculum size. P-values 6e-8 for MBC vs. MIC and 0.9 for ufMIC vs MBC obtained with Welch's t-test for independent samples.

cells initially or to droplets where the cells did not form colonies e.g. due to the antibiotic stress. This antibiogram, allows to determine a measure of the antibiotic susceptibility of the bacteria, which we denote μf·MIC. The μf·MICcorresponds to the lowest antibiotic concentration that will inhibits the growth in 95% of the droplets.

As expected, the fraction of positive droplets decreases as the concentration of antibiotics is increased. This decrease is quantified in Fig. 3b, where the number of positive droplets is shown to decrease towards zero as the concentration of the antibiotic ciprofloxacin increases, independently of the value of λ.

The interpretation of this "digital" measurement and its relation with classical microbiology measurements is not obvious. To understand its significance, measurements from the microfluidic format were confronted with measurements in a standard multiwell plate, using the same samples. Two classical microbiology measurements were performed on the samples. First, the minimum inhibitory concentration (MIC) was determined with a standard microtitre broth dilution method (see Methods). The MIC corresponds to the lowest antibiotic concentration that will inhibits visible growth. Then, the minimum bactericidal concentration (MBC) was obtained by replating the contents of the negative wells on antibiotic-free petri dishes. The number of colonies on the petri dishes after 24 hours of incubation directly represents the number of surviving bacteria in each well. The MBC thus provides a more precise measure of the antibiotic concentration that is lethal to the bacteria, compared to the MIC. We find that the number of colonies growing from negative wells is very large below a critical value of the ciprofloxacin concentration of 10 ng/mL, after which it drops dramatically to below $\approx$ 100 colonies, and eventually asymptotes to zero with increasing antibiotic concentration, see Fig. 3c.

A comparison of the measurements obtained from the three techniques (μf·MIC, MIC, and MBC) is shown in Fig. 3d, as a function of inoculum size. The inoculum size used for the μf·MIC corresponds to the total number of cells per chip. The classical MIC measurement (at $t = 24$ h) is shown to be the least sensitive of the three measurements, since it finds a critical value of the concentration that is well below the value that is necessary to kill all of the cells. In contrast, the values of the MBC and of the μf·MIC are in the same range. They both show similar trends with the inoculum size, namely a slow but detectable increase with the initial number of bacterial cells.

This close correspondence between the MBC and the μf·MIC, obtained by pooling together the total number of cells in the chip, indicates that the results in the microfluidic chip can be treated as a population-level measurement: all the bacteria from a single chip form a small population, whose ability to survive to a given concentration of antibiotics depends on its initial size. This result indicates that cells are behaving in an independent manner.

Computing single-cell antibiotic susceptibility

Beyond population level measurements, the objective is to provide insights about the antibiotic susceptibility of individual cells within a monoclonal population. This is achieved by taking a probabilistic viewpoint on the ability for a single cell to produce a colony at a given antibiotic concentration. Experimentally, we count the number of bacteria in each droplet at the beginning and at the end of an experiment for six antibiotic concentrations, see Fig. 4a. At the end of an experiment, the image analysis algorithm counts the number of bacteria at the bottom of the droplets. While this number does not reflect the exact

total number of cells in the droplet at late times, it is an acceptable proxy to differentiate positive from negative droplets .

A threshold value to distinguish positive from negative droplets is chosen at $n_{\mathrm{final}} = 15$ cells to separate best the two populations (colony or no colony). The number of positive and negative droplets for each condition and each initial number of cells is then determined. As a result, the probability of a droplet to contain a colony at the end-point can be plotted as a function of its initial number of bacteria i, and of the antibiotic concentration C_{AB}. We call this probability $p(i, C_{\mathrm{AB}})$: it represents the probability to produce a colony starting from i cells, under a concentration of antibiotics C_{AB}. The evolution of p is shown as a heat-map in Fig. 4b. We observe that $p(i, C_{\mathrm{AB}})$ decreases as the antibiotic concentration increases, but that it increases with the initial number of cells in a droplet; when more cells are in a droplet initially, more antibiotics are needed to prevent the growth of a colony after an overnight incubation.

Figure 4. Single-cell antibiotic susceptibility. **(a)** Final number of cells, counted at the bottom of each droplet at $t = 24\ h$, as a function of antibiotic concentration. Colors indicate the initial number of cells i in the droplet at $t = 0$. The horizontal dashed line is the threshold, fixed at 15 cells, chosen to define the final droplet state as positive or negative. **(b)** Survival probability, computed using Eq. (3), as a function of the initial cell number i and antibiotic concentration. **(c)** The probability to produce a colony as a function of initial number of cells, for $C_{\mathrm{AB}} = 8$ ng/mL. Bars indicate experimental measurements from counting positive cells. Dashed line shows fit according to Eq. (2), assuming independent outcomes for each cell and fitting for q. For additional fits, see Supp. Fig. S2 **(d)** The single-cell antibiotic susceptibility is plotted for all antibiotic concentrations and all experimental replicates (points). The data is well-fitted with a two-parameter Hill function (line): $h(C_{\mathrm{AB}}) = C_{\mathrm{AB}}^n / (K + C_{\mathrm{AB}}^n)$, with best fit values $n = 3.9$ and $K = 4.4$.

Counting the number of bacteria at initial times, and the fraction of positive droplets at the end of the experiment, enables to infer the susceptibility of a single cell to a con-

centration of antibiotic C_{AB}. This single-cell susceptibility, which we denote $q(C_{AB}) = 1 - p(1, C_{AB})$, is defined as the probability for a single cell to die (equivalently not form a colony) at concentration C_{AB}. If additionally all bacteria are assumed to behave independently, the probability for i cells to die is $q(C_{AB})^i$. The probability for i cells to form a colony is therefore the probability for at least one of them to form a colony, and we then have:

$$p(i, C_{AB}) = 1 - q(C_{AB})^i. \tag{2}$$

For each concentration of antibiotics, the probability $p(i, C_{AB})$ was estimated by counting the number $\mathcal{N}_i$ of droplets containing exactly i cells at the beginning of the experiment. Among these $\mathcal{N}_i$ droplets, a number $N^+(i, C_{AB})$ droplets were positive at the end of the experiment. We then have:

$$p(i, C_{AB}) = \frac{N^+(i, C_{AB})}{\mathcal{N}_i}. \tag{3}$$

The probability $p(i, C_{AB})$ was fitted to the functional form of Eq. (2) with $q(C_{AB})$ as a single fit parameter, as shown in Fig. 4c (see SI Fig. S2 for all data). The good agreement between the data and the theory confirms that the bacteria may indeed be considered as independent of each other (see in Fig. 4c), at least for the low number of cells present initially at the beginning of the experiments. Then if we write $q^\star(C_{AB})$ the best fit value of $q(C_{AB})$, this value provides the best estimate of the single-cell susceptibility to drug concentration C_{AB}, using all experimental data at hand and assuming that all cells are independent.

The single-cell susceptibility $q^\star(C_{AB})$ is expected to take a sigmoidal shape, with a value near 0 in the absence of drugs (all bacteria form colonies), and increases non-linearly with the drug concentration until reaching a plateau at $q^\star(C_{AB}) = 1$. This is confirmed by the experimental measurements, as shown in Fig. 4d. These data are arbitrarily fitted to a Hill function: $h(C_{AB}) = C_{AB}^n/(K + C_{AB}^n)$, with K and n being two fit parameters, providing a good match between experiments and the fitted function.

Identifying morphology changes under antibiotic stress

This microfluidic device and image analysis pipeline is a very powerful tool to obtain quantitative information at the single-cell level as seen in Fig. 4. But more than quantitative information, the anchored droplet format provides a unique ability to access qualitative data and to follow the evolution of the bacterial colonies within each droplet over time. The tracking of the droplet contents can be performed by time-lapse microscopy on the chips. Since the droplet position is invariant throughout the experiment, identifying the progeny of an individual cell can be achieved. Examples of a sample droplet contents can be seen in Fig. 5 and the accompanying movies (Supplementary movie S5), first in the absence of antibiotics (Fig. 5a) or under a sub-MIC concentration of ciprofloxacin (Fig.5b).

In the absence of antibiotics, almost all droplets display bacteria in their planktonic state, swimming in the droplets and showing the typical size and shape for *E. coli* in culture. A few hours into the experiments and as the populations begin to grow, cells in some droplets start to adhere to each other, eventually forming clumps. The presence of these clumps could be attributed to the presence of adhesion proteins like fimbrial adhesins that may then contribute to biofilm formation of *E. coli* [8, 21].

The morphological change of bacteria in the presence of antibiotics often begins with the elongation of the cells into long filaments: this corresponds to 83% of cases in the

Figure 5. Example of bacterial growth in three independent droplets growth **(a)** without antibiotic, and **(b)** with sub-MIC concentrations of ciprofloxacin (7 ng/mL). Scale bar: $50~\mu$m. Even at sub-MIC concentrations, the presence of ciprofloxacin leads to a filamentous bacterial morphology.

experiments at sub-MIC antibiotic concentration (7 ng/mL). The filamentation is due to the SOS response triggered by the presence of ciprofloxacin; cell-division stops but the cell metabolism continues, leading to cell volume growth. The elongated cells can then begin to produce offspring after several hours (50% of all cases) or can arrest their growth and stay in the filamentous form (33% of all cases). This is in agreement with observations on agar plates [11].

In some cases, no elongation is found. These droplets correspond either to cells that were in a dormant state (10% of all cases is expected) [17] or to cells that suffered too much damage and died due to the antibiotics (7% of all cases) as no fluorescence can be detected. As a result of these dynamics, nearly all of the droplets that form colonies contain cells having the filamentous form as well. Nevertheless, the ability to image the contents of the droplets provides a very precise indication of the state of the cells within them. This in turn is informative about the ability of the cells to overcome the antibiotic stress.

Discussion

As the emergence of antibiotic resistance is accelerating, it is crucial to understand the variability of antibiotic response on the single-cell level. This has motivated work using dilution methods to determine the MIC on petri dishes [12] or in liquid media [5], or using precision microscopy on agar [11]. In parallel, microfluidic methods have been used to provide better controlled conditions and statistics over a large number of individual cells [6, 28]. In recent years a flurry of droplet-based approaches has provided information showing the heteroresistance of a bacterial population [23], digital antibiotic susceptibility [27, 29], or even providing pathogen identification [14]. These droplet-based methods have shown that the encapsulation within droplets can be used to explore the progeny of individual cells as they respond to antibiotic stress. By relying on snapshots of moving droplets, these methods cannot relate the initial and final states of each droplet, nor can they identify the biological mechanisms that allow cells to overcome the antibiotic stress at sub-MIC concentrations.

In this context the device and analysis pipeline presented here combine the advantages of microscopy with those of droplet-based methods. Indeed the ability to identify the initial state of each droplet allows us to work at much higher cell numbers per

droplet, which translates to similar statistics as previous papers [27] while using a much lower number of droplets in total. Moreover the ability to provide time-resolved measurements on each droplet provides unique qualitative information about the adaptation of the cells, similarly to other microscpy-based methods [11]. Finally, the streamlined experimental and analysis pipeline allows us to load and image the chip in under 30 minutes, followed by an overnight incubation and a second scan requiring only a few minutes. As a result the complete campaign for obtaining an antibiogram can be performed robustly in a few hours. In contrast, the classical (petri-dish) method requires an initial overnight incubation in a 96 well plate, followed by a second overnight incubation on 50 to 80 petri dishes and manual counting of the colonies thereafter.

Beyond the performance aspects, it is important to benchmark the microfluidic measurements against standard microbiology protocols. The comparison of MIC, MBC, and µf·MIC shows that the results in the microfluidic chip can be treated as a population-level measurement, since the µf·MIC matches the MBC when accounting for the starting inoculum size. In addition to this, the detection of individual cells and the ability to perform a large number of single-cell assays in parallel allow us to develop a probabilistic treatment of the outcome within each droplet forming a colony after 24 hours. The probability for an individual cell to produce a colony at a given antibiotic concentration thus describes the heterogeneity of responses in the population.

This platform can now be used to address the single-cell response to antibiotics while screening different bacteria and molecules having different mechanisms of action. The expectation is that this screening will translate into both quantitative (shape of the single cell susceptibility curve) and qualitative (shape of the cells) differences among the conditions. The ability to encapsulate tens or hundreds of cells within the droplets, in a controlled manner, will then allow the exploration of collective behaviors and non-monotonic time-evolution of the response to antibiotics [24]. More complex experiments can also be envisaged, e.g. by recovering the contents of individual drops and performing -omics measurements on them or by varying the antibiotic concentration in time, as described previously [4]. Taken together, the different operations that can be combined into this platform constitute a major step forward in the study of antibiotic response both for scientific questions and for medical applications.

Methods

Microfluidics and microfabrication

Microfabrication and chip design

To produce microfluidic chips for the experiments, a custom mold was made using 2-layer-SU-8 photoresist lithography. The bottom layer contains the channels with 2 inputs and 1 output. The upper layer contains inverted microwells of squares shapes 120x120 um and space 240 um apart. As shown in Fig. 1 (a), the wells are organized in 13 rows of 39 and 38 wells respectively for odds and even rows. The channel height is 30 um and the well height is 130 um.

Next, PDMS and its curing agent (PDMS SYLGARD 184, Dow Corning) are mixed at a 1:10 ratio and poured into the mold. The mold is placed in a vacuum chamber for 30 minutes to eliminate air bubbles and then cured for 3 hours at 70 °C. Once the PDMS is cured, the chip is cut off from the mold and plasma bonded (CUTE Plasma, Femto

390 Science) to the coverslip.

Hydrophobic treatment

392 Prior to loading, the chips are surface treated with hydrophobic solution (NOVEC 1720
393 surface modifier/electronic grade coating 3M). To do so, two surface treatments were
394 done after the plasma bonding by filling the chip with the hydrophobic solution and cur-
395 ing it for 10 minutes at 110°C. A third surface treatment is done prior to loading.

Chip loading

As described in Fig. 1, using a syringe pump (NEMESYS), the chip is first filled through
the first input with a continuous oil phase (3M Fluorinert FC40 with non-ionic surfactant
RAN fluoSurf final concentration 1 %) while purging the air bubbles. Then, through the
second input, the continuous oil phase is replaced by the bacterial suspension. Finally,
the continuous oil phase is injected again at a very low flow rate, breaking the droplets
apart and leaving them locked in the wells.

Cell culture and preparation

Strain

405 The experiments were performed using the *E. coli* W3110 strain JEK1037 [20] labeled with
406 red fluorescent protein (lacYZ:RFP).

Antibiotic Solution

408 Ciprofloxacin (Sigma-Aldrich) was solubilized in 0.1 N HCl (Sigma-Aldrich) at 25 mg/mL.
409 The stock was then diluted with MiliQ water to 1 μg/mL. The final concentration of ciprofloxacin
410 used was between 0 and 32 ng/mL.

Cell culture

412 From the -80°C stock, the cells were streaked on LB agar plate and incubated overnight
413 at 37°C. The next day one isolated colony is inoculated in supplemented minimal media
414 (MOPS with glucose final concentration 0.4%) and IPTG is added at 0.05 mM to induce
415 the expression of the RFP. The bacterial suspension is incubated overnight at 37°C with
416 shaking.

Cell dilution

In order to get one to five cells per droplet, the optical density of the solution, measured
at 600nm, was calibrated. This was achieved as follows: the calibration was known by
using digital counting : the chip was loaded with the diluted bacterial suspension without
antibiotics, incubated at 37°C overnight, and then imaged. The empty wells were counted,
and assuming a well can only be empty if no cells is loaded, the initial loading parameter
λ was computed. λ is the Poisson parameter which corresponds to the mean number
of cells per droplet and can be directly obtained by $-\ln(N_{(-)}/N_{total})$. Where $N_{(-)}$ is the
number of negative droplets and N_{total} is the total number of droplets. λ is monitored by
the concentration of the bacterial suspension and directly linked to the optical density.

Growth characterization and antibiotic susceptibility

Growth curves

For growth characterization cells were loaded at different dilutions (500 to 500,000 cells per well) in a 96-well plate. The plate was then placed in the plate reader (Thermo Scientific Varioskan LUX) for 24 hours at 37°C with shaking. The optical density at 600 nm (OD) and the RFP fluorescence signal (excitation 488 nm and emission 520 nm) were measured every 10 minutes.

In parallel, cells were loaded to the chip with an average of one cell per droplet. This corresponds to one cell per droplet on average. The chip was then placed under the microscope and the RFP signal was measured every 30 minutes.

MIC and MBC

The minimum inhibitory concentration (MIC) and minimum bactericidal concentration (MBC) were obtained to characterize the antibiotic susceptibility at the population level.

Cells are loaded in a 96 well plate at different dilutions (500 to 500,000 cells per well) and with different ciprofloxacin concentration ranging from 2 to 36 ng/mL. The plate is then incubated for 24 hours at 37°C with shaking.

The MIC is determined as the antibiotic concentration of the first negative well, i.e. the well at which the OD is the same as the OD of an empty well.
Then, the contents of the negative wells are plated on LB agar plates and incubated for 24 hours at 37°C. The number of colonies from each of these plates is counted. The MBC is determined as the concentration where the number of colonies decrease sharply from more than a hundred cells to less than a dozen cells.

Single-cell susceptibility

To characterize the single-cell susceptibility to antibiotic, at least 10 microfluidic chips are fabricated and loaded per experiment : a control chip without antibiotic, and nine more with serial concentrations of ciprofloxacin (from 2 to 36 ng/mL). The bacterial suspension is prepared with the antibiotic and immediately loaded into the chip. The chip is imaged, using fluorescence microscopy, right after in order to determine the exact number of cells in each well.
Then, the chip is immersed in a container filled with Milli Q water to prevent evaporation and incubated overnight at 37°C. After incubation, a scan of each chip, measuring the RFP signal is performed, using fluorescence microscopy.

Microscopy and image acquisition

Microscopy images are acquired using spinning disk confocal microscope (Nikon Ti2 + Yokogawa) with a 20x 0.7 NA air objective lens (Nikon Inc.) and with 2x2 pixels binning (set directly in camera properties (Hammamatsu Orca 4)). Images of the complete chip are obtained by stitching individual images with a 5% overlap. The imaging rate is optimized by acquiring first a bright-field image of the complete chip. The RFP signal is then obtained either in confocal mode, using 3D stack.

For the 3D stack, using triggered NIDAQ Piezo Z, planes are acquired with a step of
1 μm, for a total penetration of 120 μm.

The total area of the device is about 1.4 by 0.4 mm, which translates to 40 kpix long
dataset using 350 nm pixel size. However, some chips could be tilted due to manual
bonding, which effectively increases the necessary scanning area, resulting in much big-
ger images.

Image analysis

Image registration

First the 3D fluorescence stack is converted to 2D using maximum projection (see Fig. S3
b,e). Both channels, 2D bright-field and 2D fluorescence images, are merged together
and saved as a tif stack. Then a well-labelled template image and a well-labelled mask
are made and aligned to the experimental images (see Fig. S3 c,f).

Note that the protocols described above can be modified for particular situations. For
instance time-lapse microscopy can be performed on the chips, in order to obtain time-
resolved measurements. Similarly, confocal imaging can be used at later times in order
to obtain a more precise cell count or fluorescent intensity, or to identify the morphology
of the cells in particular cases. Although these cases would require small modification
in the pipeline, the main bricks of the analysis discussed above can still be used in a
modular fashion, without major changes in the general approach.

Cell counting

After image registration, every droplet is associated with the labeled area defined by the
mask. Inside this mask, peak detection is performed to detect single fluorescent cells.
To avoid false detection, due to noise, a preprocessing is performed as follows: First, a
Gaussian filter, and subsequent peak detection using Scipy function peak_local_max with
an absolute threshold of two. The number of peak per label is then recorded into a table
for further processing.

Acknowledgements

497 This project was partially funded by ANR grant UniBAC (ANR-17-CE13-0010). The authors
498 would like to thank the support of the Microfluidics and Biomaterials platform at Institut
499 Pasteur. We also thank Erik Maikranz for useful discussions.

Supplementary information

Segmenting the whole chip images

502 Imaging very large areas with single-bacteria resolution – Challenges

503   Extracting the bacterial counts from the images at t0 and later times requires identi-
504 fying the positions of individual droplets and of the cells within them.

505   This is managed by a complete pipeline that...

Growth rate measurements

507 For the chips we used mean z projection of timelapse movies and extracted intensity
508 profiles using the labelled mask into to table. We then manually selected the curves with
509 positive growth and cropped them at 10 hours so to avoid saturation of the colonies

Figure S1. Bloc-diagram of the chip alignment procedure.

and only focus on the exponential phase of growth. We then applied an exponential fit
to individual curves using scipy fit_curve function in Python. The fit results were then
filtered to exclude poor precision using covariance matrix.

For the plate we used excel sheet produces by the plate-reader software to extract
curves and group them by concentrations. The curves were them cropped up half-maximum
intensity (0.6 a.u. in our case) and fitted the same way as the curves from the chip above.

Doubling time was then calculated from the exponential rate using the formula $t_d = $
$ln(2)/rate$ where the rate is the direct fit result, and t_d is the doubling time in hours.

Figure S2. Fitting susceptibility to susceptibility for different concentrations of ciprofloxacin.

Figure S3. Template (grayscale) + mask (blue) used for alignment and segmenting the droplets. Scalebar 500 um

References

[1] **Abbyad P**, Dangla R, Alexandrou A, Baroud CN. Rails and anchors: guiding and trapping droplet microreactors in two dimensions. Lab on a Chip. 2011; 11(5):813–821.

[2] **Amselem G**, Brun P, Gallaire F, Baroud CN. Breaking anchored droplets in a microfluidic Hele-Shaw cell. Physical Review Applied. 2015; 3(5):054006.

[3] **Amselem G**, Guermonprez C, Drogue B, Michelin S, Baroud CN. Universal microfluidic platform for bioassays in anchored droplets. Lab Chip. 2016; 16(21):4200–4211. http://xlink.rsc.org/?DOI=C6LC00968A, doi: 10.1039/C6LC00968A, publisher: Royal Society of Chemistry.

[4] **Amselem G**, Sart S, Baroud CN. Universal anchored-droplet device for cellular bioassays. In: *Methods in Cell Biology*, vol. 148 Elsevier; 2018.p. 177–199.

[5] **Artemova T**, Gerardin Y, Dudley C, Vega NM, Gore J. Isolated cell behavior drives the evolution of antibiotic resistance. Molecular Systems Biology. 2015 Jul; 11(7):822. https://www.embopress.org/doi/full/10.15252/msb.20145888, doi: 10.15252/msb.20145888, publisher: John Wiley & Sons, Ltd.

[6] **Balaban NQ**, Merrin J, Chait R, Kowalik L, Leibler S. Bacterial Persistence as a Phenotypic Switch. Science. 2004 Sep; 305(5690):1622–1625. doi: 10.1126/science.1099390.

[7] **Barizien A**, Suryateja Jammalamadaka M, Amselem G, Baroud CN. Growing from a few cells: combined effects of initial stochasticity and cell-to-cell variability. Journal of the Royal Society Interface. 2019; 16(153):20180935.

[8] **Beloin C**, Michaelis K, Lindner K, Landini P, Hacker J, Ghigo JM, Dobrindt U. The transcriptional antiterminator RfaH represses biofilm formation in Escherichia coli. Journal of bacteriology. 2006; 188(4):1316–1331.

[9] **Bergmiller T**, Andersson AMC, Tomasek K, Balleza E, Kiviet DJ, Hauschild R, Tkačik G, Guet CC. Biased partitioning of the multidrug efflux pump AcrAB-TolC underlies long-lived phenotypic heterogeneity. Science. 2017; 356(6335):311–315. doi: 10.1126/science.aaf4762, iSBN: 1095-9203 (Electronic) 0036-8075 (Linking).

[10] **Boedicker JQ**, Li L, Kline TR, Ismagilov RF. Detecting bacteria and determining their susceptibility to antibiotics by stochastic confinement in nanoliter droplets using plug-based microfluidics. Lab on a Chip. 2008; 8(8):1265–1272.

[11] **Bos J**, Zhang Q, Vyawahare S, Rogers E, Rosenberg SM, Austin RH. Emergence of antibiotic resistance from multinucleated bacterial filaments. Proceedings of the National Academy of Sciences. 2015; 112(1):178–183. http://www.pnas.org/lookup/doi/10.1073/pnas.1420702111, doi: 10.1073/pnas.1420702111, iSBN: 0027-8424 (Print).

[12] **Coates J**, Park BR, Le D, Şimşek E, Chaudhry W, Kim M. Antibiotic-induced population fluctuations and stochastic clearance of bacteria. Elife. 2018; 7:e32976.

[13] **Derzsi L**, Kaminski TS, Garstecki P. Antibiograms in five pipetting steps: precise dilution assays in sub-microliter volumes with a conventional pipette. Lab Chip. 2016; 16(5):893–901. http://xlink.rsc.org/?DOI=C5LC01151E, doi: 10.1039/C5LC01151E, publisher: Royal Society of Chemistry ISBN: 1473-0197.

[14] **Hsieh K**, Mach KE, Zhang P, Liao JC, Wang TH. Combating Antimicrobial Resistance via Single-Cell Diagnostic Technologies Powered by Droplet Microfluidics. Acc Chem Res. 2022 Jan; 55(2):123–133. https://pubs.acs.org/doi/10.1021/acs.accounts.1c00462, doi: 10.1021/acs.accounts.1c00462.

[15] **Huebner A**, Srisa-Art M, Holt D, Abell C, Hollfelder F, Demello A, Edel J. Quantitative detection of protein expression in single cells using droplet microfluidics. Chemical communications. 2007; (12):1218–1220.

[16] **Jepson AK**, Schwarz-Linek J, Ryan L, Ryadnov MG, Poon WC. What is the 'minimum inhibitory concentration'(mic) of pexiganan acting on escherichia coli?—a cautionary case study. In: *Biophysics of infection* Springer; 2016.p. 33–48.

[17] **Jõers A**, Kaldalu N, Tenson T. The Frequency of Persisters in <i>Escherichia coli</i> Reflects the Kinetics of Awakening from Dormancy. Journal of Bacteriology. 2010; 192(13):3379–3384. https://journals.asm.org/doi/abs/10.1128/JB.00056-10, doi: 10.1128/JB.00056-10.

[18] **Kaminski TS**, Scheler O, Garstecki P. Droplet microfluidics for microbiology: techniques, applications and challenges. Lab Chip. 2016; 16:2168–2187. http://pubs.rsc.org/en/Content/ArticleLanding/2016/LC/C6LC00367B, doi: 10.1039/C6LC00367B, publisher: Royal Society of Chemistry ISBN: 1473-0197.

[19] **Kao YT**, S Kaminski T, Postek W, Guzowski J, Makuch K, Ruszczak A, Stetten Fv, Zengerle R, Garstecki P. Gravity-driven microfluidic assay for digital enumeration of bacteria and for antibiotic susceptibility testing. Lab on a Chip. 2020; 20(1):54–63. https://pubs.rsc.org/en/content/articlelanding/2020/lc/c9lc00684b, doi: 10.1039/C9LC00684B, publisher: Royal Society of Chemistry.

[20] **Keymer JE LGLDAR Galajda P**. . ; .

[21] **Klemm P**. Fimbrial Adhesins of Escherichia coli. Reviews of Infectious Diseases. 1985 05; 7(3):321–340. https://doi.org/10.1093/clinids/7.3.321, doi: 10.1093/clinids/7.3.321.

[22] **Liu X**, Painter RE, Enesa K, Holmes D, Whyte G, Garlisi CG, Monsma FJ, Rehak M, Craig FF, Smith CA. High-throughput screening of antibiotic-resistant bacteria in picodroplets. Lab Chip. 2016; 16:1636–1643. http://dx.doi.org/10.1039/C6LC00180G, doi: 10.1039/C6LC00180G, publisher: Royal Society of Chemistry ISBN: 8620871129.

[23] **Lyu F**, Pan M, Patil S, Wang JH, Matin AC, Andrews JR, Tang SKY. Phenotyping antibiotic resistance with single-cell resolution for the detection of heteroresistance. Sensors and Actuators B: Chemical. 2018 Oct; 270:396–404. https://linkinghub.elsevier.com/retrieve/pii/S0925400518309493, doi: 10.1016/j.snb.2018.05.047.

[24] **Meredith HR**, Andreani V, Ma HR, Lopatkin AJ, Lee AJ, Anderson DJ, Batt G, You L. Applying ecological resistance and resilience to dissect bacterial antibiotic responses. Science Advances. 2018; 4(12):eaau1873. https://www.science.org/doi/abs/10.1126/sciadv.aau1873, doi: 10.1126/sciadv.aau1873.

[25] **O'neill J**. Antimicrobial resistance. Tackling a crisis for the health and wealth of nations. 2014; .

[26] **Robert L**, Ollion J, Robert J, Song X, Matic I, Elez M. Mutation dynamics and fitness effects followed in single cells. Science. 2018; 359(6381):1283–1286.

[27] **Scheler O**, Makuch K, Debski PR, Horka M, Ruszczak A, Pacocha N, Sozański K, Smolander OP, Postek W, Garstecki P. Droplet-based digital antibiotic susceptibility screen reveals single-cell clonal heteroresistance in an isogenic bacterial population. Sci Rep. 2020 Feb; 10(1):3282. https://www.nature.com/articles/s41598-020-60381-z, doi: 10.1038/s41598-020-60381-z, number: 1 Publisher: Nature Publishing Group.

[28] **Wang P**, Robert L, Pelletier J, Dang WL, Taddei F, Wright A, Jun S. Robust growth of Escherichia coli. Current biology. 2010; 20(12):1099–1103.

[29] **Zhang P**, Kaushik AM, Hsieh K, Li S, Lewis S, Mach KE, Liao JC, Carroll KC, Wang TH. A Cascaded Droplet Microfluidic Platform Enables High-Throughput Single Cell Antibiotic Susceptibility Testing at Scale. Small Methods. 2022; 6(1):2101254. https://onlinelibrary.wiley.com/doi/abs/10.1002/smtd.202101254, doi: 10.1002/smtd.202101254, _eprint: https://onlinelibrary.wiley.com/doi/pdf/10.1002/smtd.202101254.

[30] **Zhou W**, Le J, Chen Y, Cai Y, Hong Z, Chai Y. Recent advances in microfluidic devices for bacteria and fungus research. TrAC Trends in Analytical Chemistry. 2019; 112:175–195. https://www.sciencedirect.com/science/article/pii/S0165993618305193, doi: https://doi.org/10.1016/j.trac.2018.12.024.