A versatile high-throughput assay based on 3D ring-shaped cardiac tissues generated from human induced pluripotent stem cell derived cardiomyocytes

Magali Seguret[1], Patricia Davidson[2], Stijn Robben[2], Charlène Jouve[1], Céline Pereira[1], Quitterie Lelong[1], Lucille Deshayes[1], Cyril Cerveau[2], Maël Le Berre[2], Rita S. Rodrigues Ribeiro[2*], Jean-Sébastien Hulot[1*]

*For correspondence:
jean-sebastien.hulot@inserm.fr (FMS); ana-rita.ribeiro@4dcell.com (FS)

Present address: [§]Paris Cardiovascular Research Center, 56 Rue Leblanc, Paris, France

[1]Université de Paris Cité, PARCC, INSERM, F-75015 Paris, France; [2]4DCell, Montreuil, France

Abstract   We developed a 96-well plate assay which allows fast, reproducible and high-throughput generation of 3D cardiac rings around a deformable optically transparent hydrogel (PEG) pillar of known stiffness. Human induced pluripotent stem cell-derived cardiomyocytes, mixed with normal human adult dermal fibroblasts in an optimized 3:1 ratio, self-organized to form ring-shaped cardiac constructs. Immunostaining showed that the fibroblasts form a basal layer in contact with the glass, stabilizing the muscular fiber above. Tissues started contracting around the pillar at D1 and their fractional shortening increased until D7, reaching a plateau at 25±1%, that was maintained up to 14 days. The average stress, calculated from the compaction of the central pillar during contractions, was 1.4±0.4 mN/mm2. The cardiac constructs recapitulated expected inotropic responses to calcium and various drugs (isoproterenol, verapamil) as well as the arrhythmogenic effects of dofetilide. This versatile high-throughput assay allows multiple in situ mechanical and structural read-outs.

Introduction

Cardiac tissue engineering (CTE) aims to generate *in vitro* cell constructs that recapitulate the intricate structural and functional properties of the native myocardium (*van der Velden et al., 2022*). As the heart lacks regenerative capacities, CTE emerged in the field of regenerative medicine with the objective of producing cardiac grafts that can be implanted in patients with heart failure (*Madonna et al., 2019*; *Jabbour et al., 2021*; *Eschenhagen et al., 2022*). In parallel, the rapid development of biomaterials and microfabrication techniques, combined with progress in pluripotent stem cell biology have enabled the generation of miniature engineered heart tissues (EHTs) used as "hearts-on-chips" or cardiac organoids models. While these EHTs recapitulate a limited number of functions *in vitro*, they have provided essential platforms for cardiac disease modeling (*Filippo Buono et al.* (*2020*); *Cashman et al.* (*2016*); *Williams et al.* (*2021*); *Stillitano et al.* (*2016*); *Bliley et al.* (*2021*); *Goldfracht et al.* (*2019*); *Richards et al.* (*2020*)) and drug development (*Beauchamp et al.* (*2015*); *Polonchuk et al.* (*2017*); *Zhao et al.* (*2019*); *Mannhardt et al.* (*2020*)), reducing the need for animal studies.

Different combinations of cells, biomaterials, and scaffolds have successfully generated 3D EHTs (*Seguret et al.* (*2021*); *Zhuang et al.* (*2022*)) with various geometries including spheroids

(*Beauchamp et al.* (*2015*); *Polonchuk et al.* (*2017*); *Giacomelli et al.* (*2020*); *Richards et al.* (*2020*); *Hofbauer et al.* (*2021*); *Lewis-Israeli et al.* (*2021*)), cardiac strips (*Hansen et al.* (*2010*); *Legant et al.* (*2009*); *Turnbull et al.* (*2014*); *Mannhardt et al.* (*2016*); *Nunes et al.* (*2013*); *Zhao et al.* (*2019*)), circular bundles (*Goldfracht et al.* (*2020*); *Tiburcy et al.* (*2017*); *Li et al.* (*2020*)), myocardium-like sheets (*Shadrin et al.* (*2017*)) or ventricle-like chambers (*MacQueen et al.* (*2018*); *Li et al.* (*2018*); *Lee et al.* (*2019*)). The EHTs are optimally formed by mixing human-induced pluripotent stem cells (hiPSC)-derived cardiomyocytes, fibroblasts, and/or other cell types such as endothelial cells to more closely mimic the structure and cellular complexity of native myocardium (*Giacomelli et al.* (*2020*); *Saini et al.* (*2015*)). However, a major obstacle in the development of EHTs is the inevitable trade-off between the need for miniaturization to increase throughput (i.e. several organoids in each well of a 96- or 384-well plate) and the biological complexity of the tissues (*Cho et al.* (*2022*)). The first EHTs were generated in small batches, using pre-designed molds, and having a millimet- ric size (*Boudou et al.* (*2012*); *Turnbull et al.* (*2014*); *Legant et al.* (*2009*); *Mannhardt et al.* (*2016*)). Since then, two different avenues of development have emerged. On the one hand, to optimize throughput, more straightforward, low complexity assays such as spheroids were developed to obtain smaller tissues (*Beauchamp et al., 2015*; *Polonchuk et al., 2017*). These multicellular con- structs are valuable models to study drug responses and the human heart micro-environment but they lack the geometric intricacy of a native cardiac tissue. On the other hand, the search for higher tissue complexity and maturity led to the development of chambered organoids (*MacQueen et al.* (*2018*); *Li et al.* (*2018*); *Lee et al.* (*2019*)). These constructs require the use of complex techniques such as bioprinting or the use of bioreactors, and tissues are generated one at a time, making these approaches time-consuming and impractical on a large scale. Therefore, the next step in tissue en- gineering needs to be directed towards high-throughput physiologically-relevant assays that are simple and straightforward for the end-user. Scaling down the tissue size to increase throughput, while controlling their geometry represents a major engineering challenge. Moreover in actual de- signs, many EHTs are attached on static posts that can impact tissue formation, result in isometric contraction and can interfere with optical acquisitions when made with opaque material.

Here, we report the development of a novel assay which allows fast, reproducible and high- throughput generation of 3D cardiac rings in a 96-well plate using hiPSC-derived cardiomyocytes. The design allows the formation of multiple ring-shaped cardiac tissues in a well. The EHTs form around a central pillar made from an optically-transparent and deformable polymer that enables *in situ* monitoring of the tissue contraction with simultaneous measurements of force generation.

Results

Design and characterization of the molds

3D steel molds were designed to obtain a structure on which cells can be directed to ring-shaped cavities during sedimentation (Figure 1A). The mold fits in the well of a 96-well plate, and consists of 21 identical conical structures with a central hole (Figure 1A left). The molds obtained from the manufacturer were imaged by scanning electron microscopy and found to correspond to the intended design (Figure 1A, right). Casts of the mold were created in PEG to check the shape of the inner cavities of the mold, but the resulting structures were too soft to maintain their shape in air, and were transparent in water and therefore could not be imaged. More rigid polydimethylsiloxane (PDMS) was cast from the mold to verify the shape of the resulting structures (Figure 1B). These showed that the interior profile of the cavity was shaped in a reversed hourglass, as expected.

Generation and characterization of the gels

3D gel structures were obtained by adding a drop of PEG gel solution on top of the molds, in- verting these onto a glass coverslip, polymerizing the gel and attaching the resulting sample on a 96-well plate. To determine the stiffness of the gels, a sample of the aqueous PEG solution was polymerized. A capillary was placed perpendicularly to the edge of the gel and negative pressure was applied to the gel through the capillary. The deformation of the gel into the capillary was mea- sured and used to determine the Young's modulus of the gel (Figure 1 supplement 1). The values obtained (11.4$\pm$ 0.5 kPa) are within the physiological range for human cardiac tissue (*Gershlak*

Figure 1. Description of the mold and seeding procedure. **A.** Micromachined stainless steel mold used to shape the gel – Design (left) and scanning electron microscopy image (right) of the mold – Scale bar: 500 μm. **B.** Molded PEG gel (left - scale bar 2 mm) and zoomed view of a pillar from its PDMS replica (middle- scale bar 100 μm). Design and size of the pillar (right). **C.** Timeline of the seeding procedure. **D.** Cardiac rings in a well one day after seeding. Images stitched with ImageJ plugin - Scale bar: 1 mm. **E.** Representative compaction of a ring with time after seeding (from day 0 to day 14), in brightfield - Scale bars: 200 μm.

Figure 1—figure supplement 1. Measurement of the Young's Modulus of PEG gel

Figure 1—figure supplement 2. Measurement of the efficiency of differentiation the the iPSCs into cardiomyocytes.

Figure 1—video 1. Multiple rings beating at D14 - Brightfield imaging x4 magnification.

92 *et al.* (*2013*)).

Generation of ring-shaped cardiac tissues

The efficiency of the differentiation of iPSCs into cardiomyocytes was in average of 96.9 ± 0.8 % (Fig-
ure 1 Supplement 2).Cardiomyocytes derived from iPSCs (iPSC-CMs) with fibroblasts were seeded
on the gels. After centrifugation, the cells sedimented into the ring-shaped cavities (Figure 1D), and
compacted around the pillars to rapidly form ring-shaped cardiac constructs (Figure 1D-E). Initial
tests with cardiomyocytes resulted in cell cluster formation and malformation of the rings (Figure
2A). To stabilize the tissues, iPSC-CMs were mixed with fibroblasts and different ratios of hiPSC-
CMs to fibroblasts were tested in order to obtain the highest number of full stable cardiac rings
with time. Three ratios of hiPSC-CMs:fibroblasts were tested (Figure 2B) based on previous reports
(*Tiburcy et al.* (*2017*); *Saini et al.* (*2015*); *Giacomelli et al.* (*2020*); *Beauchamp et al.* (*2020*)). The
3:1 hiPSC-CMs:fibroblasts ratio was determined to be optimal as it allowed the generation of the
highest number of tissues stable in time, and we could not see any difference in the contractile
parameters of the tissues with different fibroblast ratios (see Figure 2 Supplement 1). Therefore,
the 3:1 hiPSC-CMs:fibroblasts ratio was used for all of the following experiments. Tissues started
beating within 24 hours after seeding (Video 1). The rings compacted in time as shown in Figure 1E.
In these optimization experiments, we obtained an average of 10.75±0.48 tissues per well which
were stable over time for at least 20 days with the 3:1 hiPSC-CMs:fibroblasts ratio.

Structure and organization of the engineered cardiac tissues

The 3D organization of the tissues was assessed by 3D immunofluorescence imaging 14 days after
seeding. The fibroblasts are concentrated at the base of the ring in contact with the glass (Figure 2C
– bottom panel), whereas the cardiomyocytes form a compact ring above the fibroblasts around
the central pillar (Figure 2C – top panel). Troponin T staining evidenced the formation of striated
elongated fibers (Figure 2D), which are typical of cardiac tissue. The obtained cardiac tissues are
75.5±1.8 μm high and their shape is toric, as shown in the 3D reconstruction (Figure 2E and video
1). To investigate the kinetics of the fibroblasts-cardiomyocytes segregation, the same immunos-
tainings were performed respectively 1 and 7 days after seeding (see Figure 2 Supplement 3). It
appears that 1 day after seeding the fibroblasts are not yet attached, although the cardiac fiber has
already started to be formed. Seven days after seeding, fibroblasts are fully spread and attached,
and the contractile ring is formed and well-aligned.

Contractility analysis

Tissues started contracting less than 24 hours after seeding, which was visible through the de-
formation of their central pillar (Supplemental video 1). Each tissue was recorded in brightfield
imaging, at a 10X magnification with a high-speed camera (Supplemental video 1). An in-house
Matlab code was developed to quantify the contraction and relaxation phases of the tissues by
monitoring the area of the central pillar over time (Figure 3A). The beat rate of the tissues first
decreased between day 1 and day 3 and increased again to reach 0.58±0.030 Hz on day 14 (Figure
3B). The contraction stress developed by the tissues increased from day 1 until day 7 (Figure 3C).
At day 7, the contractility parameters stabilized with an average stress of 1.4±0.1 mN/mm^2 (Figure
3C.). The corresponding developed strain ϵ_A was of 24.84±0.92 % (Figure 3 supplement 1). Consis-
tently, the relaxation and contraction speeds also increased during the first 7 days, then reaching
a plateau (Figure 3D and E). Since long-term culture of iPSC-CMs is gaining interest we studied the
contractile parameters of our rings 28 days after seeding in comparison to their contractile param-
eters at D14 (see Figure 3 Supplement 2). We found a slight increase for all the parameters, which
is significant for the maximum contraction speed. Nevertheless, the data is much more variable
and the number of tissues is lower (29 for D14 against 17 for D28). Therefore, we demonstrated
that long-term culture of our tissues is possible, however not yet optimized. Hence, the following
physiological and pharmacological tests have been done at D14.

Figure 2. Composition and structure of the cardiac rings. **A.**Example brightfield images of cardiomyocytes seeded with or without fibroblasts one day after seeding - Scale bar : 200 µm. **B.**Number of full cardiac rings per well in time according to the hiPSC-CMs:fibroblasts ratio they contain (4:1, 3:1 or 2:1). **C.**Confocal imaging of immunostained tissues at different heights: cardiac ring (top panel), and basal layer (bottom panel) of a tissue at 40X magnification. Vimentin, stained in green, corresponds to fibroblasts, troponin T, in red, is specific to cardiomyocytes, and DAPI is in blue. Scale bars: 100 µm. **D.**Picture of the immunostained contractile fibers at 63X magnification. Vimentin (green), Troponin T (red) and DAPI (blue) – Scale bar: 50µm. **E.**3D reconstruction of a ring. x, y and z scale bars : 100 µm.

Figure 2—figure supplement 1. Contractile parameters at D14 for the different fibroblasts ratios

Figure 2—figure supplement 2. Larger scale pictures of panel C.

Figure 2—figure supplement 3. Immunostaining of rings at D1 and D7

Figure 2—video 1. 3D reconstruction of a ring

Figure 3. Contractility analysis of the tissues. **A.**Principle of the in-house Matlab code used for contractility analysis: detection of the central pillar and monitoring of the evolution of its area in time, calculation and plot of the strain ϵ_A in time (ratio between the contraction amplitude in time and the maximum area of the central pillar) – Representative plots of the strain in time and its derivative in time, for a tissue at day 14. **B. to E.**Evolution of beating parameters through time after seeding at days 1, 3, 7, 10 and 14. The changes of all the parameters through time are significant (p<0.0001 - ANOVA for repeated measures -D1: n=57, D3: n=59 D7: n=47, D10: n=43, D14: n=36 tissues, from 3 differentiations). Beating parameters at each time point are compared to their value at day 1. **B.** Evolution of beat rate through time after seeding (***: p<0.002). **C.**Evolution of contraction stress through time after seeding (****: p<0.0001). **D.**Evolution of maximum contraction speed through time after seeding (****: p<0.0001). **E.**Evolution of maximum relaxation speed through time after seeding (****: p<0.0001).

Figure 3—figure supplement 1. Strain ϵ_A developed by the rings at D14.

Figure 3—figure supplement 2. Contractile parameters of rings at D28.

Figure 3—video 1. Contractility of a ring at day 14.

Figure 4. Study of arrhythmia in ring-shaped cardiac tissues. **A.** Poincaré plot for tissues at D14. n=24 tissues from 3 differentiations. **B.** Tissue with FluoVolt dye in its relaxed/hyperpolarized state and contracted/depolarized state. **C.** Representative plot of contraction strain (black) and fluorescence signal (green) of a tissue derived from video analysis of the FluoVolt dye.

Figure 4—video 1. FluoVolt live imaging of a ring at D15

Arrhythmia analysis

The regularity of the rings was first evaluated by plotting each period (time between two contraction peaks) as a function of the following one. We found that the Poincaré plot was fitted with a y=x line with $R^2 = 0.9633$ (Figure 4 A) and with a narrow confidence interval, showing that the vast majority of the n^{th} periods are equal to the following ones, and therefore that the rings beat regularly without spontaneous extra beats. Also, following the study of *Li et al.* (*2020*) which shows the occurrence of re-entrant waves in the case of circular geometry of tissues, we performed live imaging of a voltage-sensitive fluorescent dye in our tissues and looked for re-entrant waves. We found that, as expected, the tissue depolarization induced a visible increase in fluorescence (see Figure 4 B.). All the pixels were synchronous and we could not detect any sparks or re-entrant waves (see the video 1). We then developed a Matlab script to analyze the average FluoVolt fluorescence signal together with the contraction over time, which was also obtained from the same videos. We found that the tissues beat regularly, with a depolarization occurring just before the contraction of the ring, causing a large increase in tissue fluorescence (from 0 to 6 $\Delta F/F_0$). We also observed that the repolarization signal is noisy, as it occurs during the time when the tissue deformation is greatest and is probably affected by tissue motion.

Physiological testing on the engineered cardiac tissues

The response in contractility of the cardiac rings to an increase in the extracellular calcium concentration was assessed by sequential addition of calcium in the medium, starting from a concentration of 0.5 mM up to 3.5 mM (Figure 5A). The changes in the contraction stress and the contraction and relaxation speeds with calcium concentration were significant (p<0.0001 for each parameter, non significant for the beat rate). We observed a stabilization of the beat rate around 0.5Hz at 2mM and a positive inotropic response as the contraction amplitude increased by 600 % with a corresponding EC_{50} of 1.346 mM with a confidence interval (CI) of [1.159;1.535] mM.

Pharmacological testing on the engineered cardiac tissues

We next evaluated the dose-response to increasing concentrations of several cardiotropic drugs. The changes in the beating parameters with verapamil concentration were found to be significant (p=0.0053 for the beat rate and p<0.0001 for the other parameters). The addition of verapamil, an L-type calcium channel inhibitor, induced a negative inotropic effect until the complete stop of contraction at 10µM, as expected (Figure 5B). The IC_{50} was estimated to be 0.677 µM, CI=$[4.278;12.81].10^7$ M. The dose-response to the β-adrenergic receptor agonist, isoproterenol was tested. We observed a trend towards increase of the contraction amplitude, as well as a significant increase in the contraction and relaxation speeds (Figure 5C), without change in the beat rate. The changes in the contraction stress and the contraction and relaxation speeds with isoproterenol concentration are significant (p=0.0004 for the contraction stress; p<0.0001 for the contraction and relaxation speeds). Finally, dofetilide, which is an inhibitor of the I_{Kr} current, was also tested at different concentrations on the cardiac rings. This drug has been shown to be pro-arrhythmic and to induce arrhythmic events such as Torsades-de-Pointes in patients. The changes in the beat rate, the contraction stress and the contraction and relaxation speeds with dofetilide concentration were found to be significant (p<0.0001 for all the parameters). As shown in Figure 5D, dofetilide significantly reduced the contraction stress and speeds until the complete stop of the contraction of the cardiac tissues for a concentration of 10 nM. Arrhythmic events were noted after adding dofetilide at concentrations of 3 and 7 nM. Indeed, the Poincaré diagram for the different dofetilide concentrations (Figure 5E) shows an increase in the contraction-to-contraction intervals (equivalent to RR intervals) for some cells at 1 nM of dofetilide and an increasing number of erratic beats with increasing dofetilide concentrations.

Discussion

The combination of pluripotent stem cells technologies, advanced cardiomyocyte differentiation protocols and microfabrication methods have already resulted in the generation of engineered human cardiac muscle tissue-constructs, with different shapes and that can reproduce some cardiac functions (*Seguret et al.* (*2021*); *Cho et al.* (*2022*)). However, until now, approaching the complex geometry and structure of the human myocardium required complex techniques, which limit the number of tissues produced, thus hampering their application for high-throughput experiments. In this study, we aimed to address this challenge by developing a novel strategy to create ring-shaped miniature cardiac tissues around a central pillar that is made from an optically-transparent and deformable polymer. Our main findings are as follows : (1) this novel strategy enables the fast and reproducible formation of an important number of cardiac tissues (up to 21 per well in a 96-well format) with a limited number of hiPSC-cardiomyocytes (112,500 per well); (2) In this assay, fibroblasts are essential to generate and maintain the tissue structure, with a 3:1 hiPSC-CMs:fibroblast ratio providing with the best yield of tissue generation; (3) The self-organized ring-shaped cardiac tissues contract and deform the optically-transparent central pillar, which can be recorded and analyzed to estimate the contractile force developed by the tissue; (4) These ring-shaped tissues are compatible with live fluorescence imaging and allow the simultaneous measurement of contractile parameters and another readout provided by a fluorescent probe; (5) These EHTs display concordant responses to positive or negative inotropes, and to arrhythmogenic drugs, as required for drug testing applications.

The miniaturized cardiac constructs develop by self-organization of the hiPSC-derived cardiomy-

Figure 5. Physiological and drug testing on the cardiac tissues. Effect of the concentration in extracellular calcium (**A.**), verapamil (**B.**), isoproterenol (**C.**) or dofetilide (**D.**) on tissues contractility: beat rate, contraction stress and the maximum contraction and relaxation speeds. These parameters are expressed as a ratio between their value for each concentration and the value at basal state ([Ca^{2+}]=0.5 mM for calcium test and [Drug]=0M for drug testing). An ANOVA for repeated measures was carried out for each parameter. For each test, the value of each parameter at each concentration is compared to its value at the minimal concentration of the drug (respectively [Ca2+]=0.5mM, [Verapamil]=10^{-9} M, [Isoproterenol]=10^{-9} M and [Dofetilide]=10^{-9} M). **E.** Poincaré plot for 0 nM, 1 nM, 3 nM and 7 nM of dofetilide. For each concentration of each drug, more than 20 tissues from 3 different concentrations could be analyzed. Data is presented as mean ± SEM. *: p<0.05; **: p<0.01; ***: p<0.001; ****: p<0.0001.

ocytes and fibroblasts, which are guided by a specific design to acquire a ring shape. The resulting
EHTs are in a convenient 96-well plate format and easy to use for a user familiar with cell biology
techniques. The 96-well plate format is adaptable to many plate readers and automated platforms.
In addition, the optical clarity of the gels and the absence of gel directly under the structures en-
sures optimal imaging conditions. We showed that the tissues can be fixed, stained by immunoflu-
orescence and imaged *in situ* at high resolution. Moreover, our EHTs are also be compatible with
live cell imaging experiments with fluorescent reporters, such as the FluoVolt voltage-sensitive dye,
for which we presented a proof-of-concept. Indeed, we demonstrated that while measuring the
depolarization of the cells by imaging the fluorescence over time, we could simultaneously mon-
itor the contraction parameters *in situ* through the deformation of the pillar. This possibility to
simultaneously record two parameters represents a significant asset, as this proof-of-concept can
be extended to other fluorescent probes.

The current design of our EHTs consists of up to 21 rings in a single well, thus providing many
replicates within a single experiment, even if some of the rings fail to form due to biological variabil-
ity. This high throughput can compensate for potential intra-batch effects. Moreover, the possibil-
ity to seed several wells at the same time allows to easily study several differentiations, compen-
sating for the variability which is inherent to the differentiation of hiPSCs. The predictable circular
shape of the tissues makes them easy to recognize and track using software. Coupled with the
known organization of the 21 rings for easy automated imaging, a single well can quickly generate
data on many tissues that is quickly and automatically analyzed. The EHTs we described here are
thus easy to use, image, manipulate and obtain readouts from.

We found that the circular geometry of these self-organized cardiac tissues induces a homo-
geneous distribution of the cardiomyocytes and consequently of the contraction forces around
the pillar and in the tissue. Importantly, the presence of the central pillar not only allows resis-
tance to be felt by the tissue, but the control of its stiffness facilitates the calculation of the force
developed by the tissue. Our results showed that the tissues develop a fractional shortening (or
contraction strain) of about 25%, which is close to the contraction of a human heart (*Cheng et al.*
(*2010*)). The corresponding stress was 1.4 mN/mm^2 which is in a similar range to other previously
reported tissue-engineered cardiac muscle models with hiPSC-CMs in the literature (*Seguret et al.*
(*2021*); *Turnbull et al.* (*2014*); *Shadrin et al.* (*2017*); *Zhao et al.* (*2019*); *Ronaldson-Bouchard et al.*
(*2018*); *Goldfracht et al.* (*2020*); *Li et al.* (*2020*)). Some studies have reported higher force values
(*Ronaldson-Bouchard et al.* (*2018*)) but in tissues that have a much large size than the tissues pro-
posed here. The force generated by the adult human heart muscle, around 44 mM/mm^2 for a tissue
strip (*Hasenfuss et al.* (*1991*)), however remains significantly higher than the force developed by
the currently available platforms. Few cardiac assays can precisely monitor the force exerted by the
tissues. Indeed, cardiac cells seeded on micropillars (*Rodriguez et al.* (*2014*)) and gels containing
tracking beads (*Dou et al.* (*2021*); *Feyen et al.* (*2020*)) can provide information on local (subcellular)
force exertion. The force measurements do not result from the concerted effort of an organized
tissue composed of aligned cardiomyocytes, as proposed by ring-shaped tissues. Tissues attached
on two posts (*Turnbull et al.* (*2014*); *Mannhardt et al.* (*2016*); *Abilez et al.* (*2018*)) can also provide
some information on the force generated by a tissue but the force exertion is not radial as it is in the
native heart. Moreover, the presence of the posts can impact the contraction by affecting the force
distribution in the tissue and creating potential edge effects in the vicinity of the posts, inducing the
formation of v-necks (*Abilez et al.* (*2018*)). Lastly, some other ring-shaped cardiac constructs have
to be transferred on posts or on a force transducer to measure their contraction force (*Goldfracht*
et al. (*2020*); *Tiburcy et al.* (*2017*)). Therefore, the easy *in situ* monitoring of the force generated by
our ring-shaped tissues with an image analysis routine represents a significant advantage, as less
manipulation improves the degree of standardization. Finally, the stiffness of the central pillar can
easily be adapted to model a pathological increase or decrease of the extracellular matrix stiffness.

The EHTs we developed are composed of cardiomyocytes that rely on fibroblasts for their as-
sembly and inter-cellular adhesion. Fibroblasts have been shown to play a prominent role in the
heart as they are crucial in the constitution of the supporting extracellular matrix and contribute to
cardiomyocyte electrical coupling, conduction system insulation, vascular maintenance, and stress-

sensing (*Ivey and Tallquist* (*2016*); *Baudino et al.* (*2006*)). Other studies have highlighted the importance of a multi-cellular approach to generate EHTs (*Giacomelli et al.* (*2020*); *Saini et al.* (*2015*); *Amano et al.* (*2016*); *Caspi et al.* (*2007*)). In our system, the fibroblasts additionally play a role in providing a basement support under the cardiomyocytes, in contact with the glass support of the EHTs. Immunofluorescence imaging showed an organization of the cardiomyocytes parallel to each other, perpendicular to the direction of contraction, with an increased alignment of sarcomeres. Our EHTs mimic the composition and organization of the native myocardial tissue, even if only composed of cardiomyocytes and fibroblasts. Future experiments will determine whether other, less abundant, cell types of the native myocardium, such as endothelial or smooth muscle cells, can further improve heart tissue modeling.

The constructs showed a spontaneous beating in the day following seeding and this activity was maintained over several weeks thus allowing contractile measurements and pharmacological testing in the conditions of auto-pacing. Since *Li et al.* (*2020*) highlighted the occurrence of re-entrant waves in their circular cardiac tissues, we investigated the arrhythmogenicity of our tissues and performed live imaging of depolarization in the tissues to evaluate the presence of re-entrant waves. We were able to show that our cardiac tissues beat regularly, and we were able to correlate the depolarization with contraction over time. The next step would be to correct the fluorescence profile to account for tissue deformation in order to study tissue repolarization. This would require the development of a complex pixel-tracking code. The analysis of re-entrant waves is however hampered by a technical limitation. Indeed, it has been shown that the conduction velocity in iPSC-CMs is around 10-20 $cm.s^{-1}$, compared to 40-60 $cm.s^{-1}$ in adult cardiomyocytes *Ahmed et al.* (*2020*). Here we image our tissues at the maximum speed allowed by our camera, which is between 40 and 50 frames per second, which is faster than in *Li et al.* (*2020*). However, our constructs are smaller, with a perimeter of about 1.3 mm, and therefore, the required camera speed to detect the conduction in tissues should be at least of 120 fps if we consider a conduction speed of 15 $cm.s^{-1}$. Hence we were not able to detect conduction in our tissues, nor re-entrant waves.

This first study is nevertheless a proof of concept for live imaging in miniature ring-shaped tissues, already allowing us to correlate the depolarization peaks with contraction peaks thus reflecting the excitation-contraction coupling efficiency in the tissues. Additional efforts, especially with a faster camera, will be needed to more accurately analyze tissue conduction and detect re-entrant waves. As a proof-of-concept, we also studied the effects of well-established negative and positive inotropes and showed the ability of our platform to reproduce the anticipated effects. We observed that our tissues presented a positive inotropic response to increasing extracellular calcium concentration with a corresponding EC_{50} which is in agreement with the values reported in the literature for engineered heart tissues made of hiPSC-CM, which are between 0.4 and 1.8mM for the EC_{50} (*Feric and Radisic* (*2016*); *Schaaf et al.* (*2011*); *Streckfuss-Bömeke et al.* (*2013*); *Mannhardt et al.* (*2016*); *Turnbull et al.* (*2014*); *Goldfracht et al.* (*2020*)). The negative inotropic response to increasing concentrations of verapamil of our constructs was also described in the literature and our IC_{50} value is also in the range of reported values in other EHTs: Turnbull et al. *Turnbull et al.* (*2014*) found IC_{50}=0.61 µM and other papers range from 0.3 µM to 0.6 µM (*Mannhardt et al.* (*2016*); *Thavandiran* (*2019*)). The β-adrenergic stimulation of the tissues with increasing concentrations of isoproterenol-induced a trend towards increase in the contraction force and significantly increased contraction and relaxation speeds without change in the beat rate. This result may be surprising, as many articles report a positive chronotropic behavior for spontaneous hiPSC-CMs, while the inotropic response is less often highlighted (*de Lange et al.* (*2021*); *Chen et al.* (*2015*); *Ronaldson-Bouchard et al.* (*2018*)). Nevertheless the increase in contraction and relaxation speeds that we observed is consistent with data reported in the literature. Our hypothesis is that this increase in contraction and relaxation speeds induced by isoproterenol is translated, on average in our study, into an increase in contractile force rather than in an increase in contraction frequency. This may depend on the cell line used, and is very well illustrated in *Mannhardt et al.* (*2020*): of the 10 different cell lines tested in engineered heart tissues, all show an increase in contraction and relaxation speeds after isoproterenol administration, but this is translated either into an increase in contractile force (4 cell lines) or into a shortening of the beat (3 cell lines), and only 2 cell lines show

an increase in both parameters. Indeed, since hiPSC-CMs are immature cardiac cells, it is rare to obtain a positive force-frequency relationship without any maturation medium or mechanical or electrical training. Last, applying increasing dofetilide concentrations on our constructs showed a decrease in the contraction amplitude and contraction and relaxation speeds. This is in line with the literature as *Zhao et al.* (*2019*) depicted a significant reduction of the action potential amplitude and an increase in action potential durations of the tissues with increasing dofetilide concentration, in line with *Huo et al.* (*2019*); *Blinova et al.* (*2017*). Above a concentration of 10nM, dofetilide shows cardiotoxicity in our tissues as tissues completely stop beating. By assessing the beating regularity by plotting each period as a function of the previous one for doses of 0nM to 7nM of dofetilide, we highlighted an increasing number of erratic beats with the increase in dofetilide concentration. Other papers have also underlined the pro-arrhythmic effect of dofetilide on cardiomyocytes derived from iPSCs (*Goldfracht et al.* (*2019*); *Patel et al.* (*2019*); *Altrocchi et al.* (*2020*); *Blinova et al.* (*2017*)) and this drug has been shown to induce atrial fibrillation, torsade-de-pointes and heart failure in patients (*Abraham et al., 2015*; *Shantsila et al., 2007*). Therefore, we demonstrated that our platform recapitulated the expected cardiac responses to several inotropes and arrhythmogenic drugs.

While representing a significant improvement, our platform presents some limitations. First, the size of the rings is very small compared to native tissue (400 µm). This size limitation is necessary for the purposes of high throughput screening, but for applications requiring larger tissues, wider rings can be considered in the future. Second, all measurements were performed under spontaneous pacing as a specific technique should be developed to allow external pacing of the tissues in the 96-well format. This includes electrical field stimulation or an optogenetics system with light stimulation. The precise and regular organization of the rings in the well makes this design very amenable to the introduction of microelectrodes for pacing, which could also allow the measurement of the action potentials of individual micro-tissues. Nevertheless, our main results show that pharmacological responses, including the detection of the pro-arrhythmogenic effects of dofetilide, were correctly measured under spontaneous beating conditions. Third, the miniaturized format of our platform creates some limitations and challenges regarding some experiments, including tissue stretching. Therefore, the length-tension relationships (Franck-Starling mechanism) were not determined in our tissues.

Overall the novel platform we describe here is a versatile, easy-to-use and high-throughput tool to generate cardiac rings in a reproducible way and analyze them *in situ*. This platform will be useful in several cardiac research fields, including disease modeling and pharmacological testing.

Methods and Materials

Mold design and characterization

Micromachined stainless steel molds were obtained from high resolution CNC machining according to our CAD designs. The designs consisted of a mold that would be slightly smaller than the footprint of a well of a 96-well plate, composed of 21 structures (see Figure 1A). To check the shape of the interior cavities of the mold, a cast of the mold was created out of polydimethylsiloxane (PDMS) (Sylgard 184, Samaro), according to manufacturer's instructions.

3D hydrogel substrate preparation

A solution of 5% w/w photopolymerizable polyethylene glycol (PEG) was prepared in distilled water. Circular coverslips (16 mm, Paul Marienfeld GmbH, Germany) were silanized to ensure adequate bonding between the PEG and glass. A drop of 10-15 μL of gel was placed on the stainless steel mold, which was then placed in contact with a silanized coverslip. The gel solution was briefly exposed to light of the appropriate wavelength to polymerize the hydrogel. The resulting 3D structured hydrogel, bound to the coverslip, was unmolded and immediately placed in distilled water. The hydrogel was rinsed at least three times with distilled water and incubated in water overnight. For cell experiments, the coverslip coated in a 3D hydrogel was briefly dried and attached to a commercial black polystyrene 96-well plate with an adhesive backing (Grace Bio-labs, USA). The

resulting well was filled with distilled water. Prior cell seeding, the plate was sterilized by UV irradiation for at least 20 minutes.

Young's modulus measurements of the molds

The PEG-based hydrogel solutions were prepared and 100µL of solution was deposited into a ring-shaped mold with an inner diameter of 1cm positioned on a 16mm glass coverslip. The solution was polymerized using 365nm UV light for 1 minute at a power of 70 mW/cm^2, after which the mold was removed. The coverslip containing the gel was then transferred to the microscope (Leica camera AG, Germany). A hollow glass capillary with an inner diameter 0.75mm (World Precision Instruments, USA) was attached to a platform movable in the xy-plane which was used to position the capillary right next to the edge of the circular polymerized hydrogel. Once the capillary and the hydrogel made contact, a negative pressure was applied, using the Cobalt autonomous pressure pump (Elvesys, France), creating a small convexity in the surface of the hydrogel. Images of this convexity were taken (5X magnification) for a pressure ranging from -100mbar to -200mbar, with 10 mbar increments (Figure 1A.). Two measurements were done on two separate gels, resulting in 4 measurements per gel with 10 datapoints each. The images (and thus the size of the convexities) were analyzed using ImageJ (version 1.53c). The Young's modulus was calculated with the formula $E = \frac{p}{0.872\frac{l}{a}+0.748(\frac{l}{a})^2}$ according to literature (*Gandin et al.* (*2021*)).

Cardiomyocyte differentiation and culture

The SKiPSC-31.3 hiPSC cell line, from a healthy control subject, has been used here and was previously reported (*Galende et al.* (*2010*)). Induced pluripotent stem cells were seeded on Matrigel and cultivated in mTESR Plus medium (Stemcell Technologies). At 80 % confluency, cells were passaged with ReleSR reagent (StemCell Technologies). When hiPSCs reached a 90 % confluency, differentiation was carried out according to a protocol adapted from *Garg et al.* (*2018*). Briefly, on day 0 of differentiation, mTESR Plus was changed to RPMI-1640 (Life Technologies) supplemented with B27 without insulin (RPMI-B27 minus insulin) and 6 µM CHIR-99021 (Abcam). After 48 hours, the medium was changed to RPMI-B27 minus insulin and then to RPMI-B27 minus insulin with 5µM IWR-1 (Sigma) for 48 hours. On day 5, the medium was replaced by RPMI-B27 minus insulin, and switched to RPMI-B27 with insulin on day 7. At day 11 a glucose starvation was carried out by replacing the medium with RPMI 1640 without glucose (Life Technologies), supplemented by B27 with insulin for 3 days. Cells were then dissociated with 0.05% trypsin (Life Technologies) and re-plated in a 12-well plate at a density of 0.3 10^6 cells/cm^2 in RPMI-B27 with insulin. The following day, the medium was switched back to RPMI 1640 without glucose (Life Technologies), supplemented by B27 with insulin for 3 days again. These two rounds of glucose starvation have been shown to increase dramatically the percentage of cardiomyocytes obtained *Sharma et al.* (*2015*). After day 18, the cells were cultured in RPMI-B27 with insulin and the medium was changed every two days.

Cardiomyocyte characterization

The efficiency of the differentiation was assessed by cardiac troponin T flow cytometry at D21. On day 21 of differentiation, hiPSC-CMs were dissociated by enzymatic digestion (Miltenyi Multi Tissue dissociation kit 3) and stained with the Zombie NIR™ Fixable Viability Kit (BioLegend). Then, collected cell pellets were fixed and permeabilized using Inside Stain kit (Miltenyi Biotech, 130-090-477) at room temperature for 10 minutes. Cells were incubated with either APC anti-cardiac troponin T (CTNT) antibody (Miltenyi Biotech; 130-106-689 1:100) or APC isotype control (Miltenyi Biotech, 130-104-615 1:100) for 10 min at room temperature. Cells were analyzed using the BD Biosciences FACS LSR Fortessa X-20 instrument with at least 30,000 cells. Results were processed using FlowJo v10 (FlowJo, LLC).

Human fibroblasts culture

Commercially available normal adult human dermal fibroblasts (NHDF-Ad Lonza CC-2511, Lot 545147) were used up to passage 6 to generate the tissues. They were cultured in T75 flasks, in DMEM high glucose, supplemented with 10% FBS, 1% NEAA and 1% penicillin-streptomycin.

Tissues generation and culture

The wells of the assay were first rinsed twice with PBS and incubated at 37°C in DMEM high glucose, 2.3 mM CaCl2, 10%FBS, 0.1% penicillin-streptomycin. To generate the tissues, cardiomyocytes were dissociated by enzymatic digestion (Miltenyi Multi Tissue dissociation kit 3) on day 22 after differentiation. Fibroblasts were dissociated with TrypLE Express Enzyme (12605010 Thermofischer). Three ratios of hiPSC-CMs:fibroblasts were tested: 4:1, 3:1 and 2:1. For the following experiments the tissues are generated with a ratio of 3:1 hiPSC-CMs:fibroblasts as this ratio allowed the best preservation of tissues in time. A total of 150k cells were seeded in each well. The required numbers of fibroblasts and cardiomyocytes were then mixed and resuspended in 200µL per well of DMEM high glucose, 2.3 mM CaCl2, 10%FBS, 0.1% penicillin-streptomycin. The plate was then incubated at 37°C for 5 minutes and centrifuged 3 times in "short" mode to ensure that the cells fall into the circular molds. The plate was then kept at 37°C and the medium was changed every two days.

Tissues recording and analysis

The cardiac rings were recorded for 10 seconds in brightfield imaging with a high speed CCD camera (PL-D672MU, Pixelink) mounted on a microscope (Primovert, Zeiss), at a 10X magnification. The videos were then analyzed with a custom Matlab script. Briefly, the area of the central pillar is monitored in time and allows to recover the contraction amplitude in time Amplitude(t)= Max(Area)-Area(t) and therefore the contraction strain in time (in terms of area) ϵ_A calculated as $\epsilon_A(t) = (Max(Area) - Area(t))/Max(Area)$ as well as the derivative of this strain in time. $\epsilon_A(t)$ is linked to the linear strain ϵ with the formula $\epsilon = \frac{\Delta r}{r_{max}} = \frac{\epsilon_A}{1+\sqrt{1-\epsilon_A}}$. As we measured the Young's Modulus of the gel, we can recover the stress developed by the cardiac rings $\sigma = E.\epsilon(t)$. From these traces we also derive the contraction frequency by Fourier transform. It was considered that below a contraction of 0.5 %, the signal to noise ratio was not sufficient enough to detect the contraction of the tissue and the beating frequency was therefore set to 0 Hz. Finally, the maximum contraction and relaxation speeds of the tissue were also calculated respectively from the maxima and minima of the derivative of this strain in time.

Live imaging in tissues

Around 15 days after seeding, tissues were washed with a Tyrode solution (NaCl 14 mM, KCl 5 mM, 10 mM Hepes, Glucose 10 mM, $MgCl_2$ 1 mM, $CaCl_2$ 1.8 mM, pH 7.4) and incubated 30 minutes at 37°C, 5% Co_2 with the FluoVolt Loading solution (FluoVolt dye dilution 1:1000, Powerload solution dilution 1:100 in the Tyrode solution) as per the manufacturer's recommendations (FluoVolt kit - Thermofisher). They were then washed with Tyrode and incubated in 200 µL Tyrode per well at 37°C, 5% CO_2 to recover. Live fluorescence imaging of the Fluovolt dye was performed *in situ* with a Leica TSC SP8 confocal imaging system in resonant mode at 37°C incubation, with a 20X objective and a 256x256 pixels resolution. This allowed a frame rate of 43.96 frames per second.

Response to increasing extracellular calcium concentration and drug tests

Tissues response to different drugs and increasing calcium concentrations was assessed after day 14. For the increasing extracellular calcium concentrations, tissues were first changed to Tyrode solution 0.5 mM in calcium (NaCl 140 mM, KCl 5 mM, 10 mM Hepes, Glucose 10 mM, $MgCl_2$ 1 mM, $CaCl_2$ 0.5 mM, pH 7.4) and let to equilibrate for 30 minutes at 37°C, 5 % CO_2. Calcium concentration was then increased by sequential addition of a Tyrode solution at 10mM in calcium. After each addition, the tissues were incubated for 5 minutes before recording. For drug tests, tissues were changed to Tyrode solution at 1.8mM in calcium (NaCl 14 mM, KCl 5 mM, 10 mM Hepes, Glucose 10 mM, $MgCl_2$ 1 mM, $CaCl_2$ 1.8 mM, pH 7.4) and let to equilibrate for 30 minutes at 37°C, 5% CO2. To evaluate the dose-response of the cardiac rings to the different drugs (Verapamil- sigma v0100000, isoproterenol - sigma 1351005, dofetilide - sigma PZ0016), the contractility was first recorded for 30 seconds in basal state and the drug was then added sequentially. Between each addition of drug, the tissues were incubated for 5 minutes at 37°C before being recorded for 30 seconds.

Immunofluorescence staining and confocal microscopy

Tissues were fixed at day 15 with 4% paraformaldehyde for 15 minutes and rinsed 3 times in PBS for 5 minutes. They were then blocked and permeabilized with blocking solution(BSA 2%, Triton 0.5%) at 4°C overnight. Primary antibodies Vimentin diluted at 1:250 (MA5-11883, ThermoFischer Scientific) and Troponin T diluted at 1:500 (ab45932, Abcam) were added the next day and incubated at 4°C overnight. Tissues were rinsed 3 times in PBS for 5 minutes and secondary antibodies (Alexa Fluor secondary antibodies, LifeTechnologies, at 1:1000) and DAPI diluted at 1:1000 were added and incubated at 4°C overnight. Immunostaining pictures were taken with a Leica TCS SP8 confocal system.

Statistical analysis

All numerical results are expressed as mean $\pm$ standard error mean (SEM) of three independent experiments. Differences between experimental groups were analyzed with the appropriate statistical tests, specified each time. P values < 0.05 were considered significant for all statistical tests. Statistical analyses were performed with GraphPad Prism software.

Acknowledgments

We thank the iPS core facility of Paris Cardiovascular Research Center for its support for cell culture. We thank Camille Knops and Yunling Xu from the Flow Cytometry and the Microscopy platforms, respectively, from Université de Paris, Paris Cardiovascular Research Center, Paris, France. Schemes were created with BioRender.com (access date:10/03/2023).

References

Abilez OJ, Tzatzalos E, Yang H, Zhao MT, Jung G, Zöllner AM, Tiburcy M, Riegler J, Matsa E, Shukla P, et al. Passive stretch induces structural and functional maturation of engineered heart muscle as predicted by computational modeling. Stem cells. 2018; 36(2):265–277.

Abraham JM, Saliba WI, Vekstein C, Lawrence D, Bhargava M, Bassiouny M, Janiszewski D, Lindsay B, Militello M, Nissen SE, et al. Safety of oral dofetilide for rhythm control of atrial fibrillation and atrial flutter. Circulation: Arrhythmia and Electrophysiology. 2015; 8(4):772–776.

Ahmed RE, Anzai T, Chanthra N, Uosaki H. A brief review of current maturation methods for human induced pluripotent stem cells-derived cardiomyocytes. Frontiers in cell and developmental biology. 2020; 8:178.

Altrocchi C, de Korte T, Bernardi J, Spätjens RL, Braam SR, Heijman J, Zaza A, Volders PG. Repolarization instability and arrhythmia by I Kr block in single human-induced pluripotent stem cell-derived cardiomyocytes and 2D monolayers. EP Europace. 2020; 22(9):1431–1441.

Amano Y, Nishiguchi A, Matsusaki M, Iseoka H, Miyagawa S, Sawa Y, Seo M, Yamaguchi T, Akashi M. Development of vascularized iPSC derived 3D-cardiomyocyte tissues by filtration Layer-by-Layer technique and their application for pharmaceutical assays. Acta biomaterialia. 2016; 33:110–121.

Baudino TA, Carver W, Giles W, Borg TK. Cardiac fibroblasts: friend or foe? American Journal of Physiology-Heart and Circulatory Physiology. 2006; 291(3):H1015–H1026.

Beauchamp P, Jackson CB, Ozhathil LC, Agarkova I, Galindo CL, Sawyer DB, Suter TM, Zuppinger C. 3D co-culture of hiPSC-derived cardiomyocytes with cardiac fibroblasts improves tissue-like features of cardiac spheroids. Frontiers in molecular biosciences. 2020; 7:14.

Beauchamp P, Moritz W, Kelm JM, Ullrich ND, Agarkova I, Anson BD, Suter TM, Zuppinger C. Development and characterization of a scaffold-free 3D spheroid model of induced pluripotent stem cell-derived human cardiomyocytes. Tissue Engineering Part C: Methods. 2015; 21(8):852–861.

Bliley JM, Vermeer MC, Duffy RM, Batalov I, Kramer D, Tashman JW, Shiwarski DJ, Lee A, Teplenin AS, Volkers L, et al. Dynamic loading of human engineered heart tissue enhances contractile function and drives a desmosome-linked disease phenotype. Science Translational Medicine. 2021; 13(603):eabd1817.

Blinova K, Stohlman J, Vicente J, Chan D, Johannesen L, Hortigon-Vinagre MP, Zamora V, Smith G, Crumb WJ, Pang L, et al. Comprehensive translational assessment of human-induced pluripotent stem cell derived cardiomyocytes for evaluating drug-induced arrhythmias. Toxicological Sciences. 2017; 155(1):234–247.

509 **Boudou T**, Legant WR, Mu A, Borochin MA, Thavandiran N, Radisic M, Zandstra PW, Epstein JA, Margulies KB,
510 Chen CS. A microfabricated platform to measure and manipulate the mechanics of engineered cardiac
511 microtissues. Tissue Engineering Part A. 2012; 18(9-10):910–919.

512 **Cashman TJ**, Josowitz R, Johnson BV, Gelb BD, Costa KD. Human engineered cardiac tissues created using
513 induced pluripotent stem cells reveal functional characteristics of BRAF-mediated hypertrophic cardiomy-
514 opathy. PloS one. 2016; 11(1):e0146697.

515 **Caspi O**, Lesman A, Basevitch Y, Gepstein A, Arbel G, Habib IHM, Gepstein L, Levenberg S. Tissue engineering of
516 vascularized cardiac muscle from human embryonic stem cells. Circulation research. 2007; 100(2):263–272.

517 **Chen G**, Li S, Karakikes I, Ren L, Zi-Ying Chow M, Chopra A, Keung W, Yan B, Chan CW, Costa KD, et al. Phos-
518 pholamban as a Crucial Determinant of the Inotropic Response of Human Pluripotent Stem Cell–Derived
519 Ventricular Cardiomyocytes and Engineered 3-Dimensional Tissue Constructs. Circulation: Arrhythmia and
520 Electrophysiology. 2015; 8(1):193–202.

521 **Cheng S**, Xanthakis V, Sullivan LM, Lieb W, Massaro J, Aragam J, Benjamin EJ, Vasan RS. Correlates of echocar-
522 diographic indices of cardiac remodeling over the adult life course: longitudinal observations from the Fram-
523 ingham Heart Study. Circulation. 2010; 122(6):570–578.

524 **Cho S**, Discher DE, Leong KW, Vunjak-Novakovic G, Wu JC. Challenges and opportunities for the next generation
525 of cardiovascular tissue engineering. Nature Methods. 2022; 19(9):1064–1071.

526 **Dou W**, Wang L, Malhi M, Liu H, Zhao Q, Plakhotnik J, Xu Z, Huang Z, Simmons CA, Maynes JT, et al. A mi-
527 crodevice platform for characterizing the effect of mechanical strain magnitudes on the maturation of iPSC-
528 Cardiomyocytes. Biosensors and Bioelectronics. 2021; 175:112875.

529 **Eschenhagen T**, Ridders K, Weinberger F. How to repair a broken heart with pluripotent stem cell-derived
530 cardiomyocytes. Journal of Molecular and Cellular Cardiology. 2022; 163:106–117.

531 **Feric NT**, Radisic M. Towards adult-like human engineered cardiac tissue: Maturing human pluripotent stem
532 cell-derived cardiomyocytes in human engineered cardiac tissues. Advanced drug delivery reviews. 2016;
533 96:110.

534 **Feyen DA**, McKeithan WL, Bruyneel AA, Spiering S, Hörmann L, Ulmer B, Zhang H, Briganti F, Schweizer M, Hegyi
535 B, et al. Metabolic maturation media improve physiological function of human iPSC-derived cardiomyocytes.
536 Cell reports. 2020; 32(3):107925.

537 **Filippo Buono M**, von Boehmer L, Strang J, P Hoerstrup S, Y Emmert M, Nugraha B. Human cardiac organoids
538 for modeling genetic cardiomyopathy. Cells. 2020; 9(7):1733.

539 **Galende E**, Karakikes I, Edelmann L, Desnick RJ, Kerenyi T, Khoueiry G, Lafferty J, McGinn JT, Brodman M, Fuster
540 V, et al., Amniotic fluid cells are more efficiently reprogrammed to pluripotency than adult cells. Mary Ann
541 Liebert, Inc. 140 Huguenot Street, 3rd Floor New Rochelle, NY 10801 USA; 2010.

542 **Gandin A**, Murugesan Y, Torresan V, Ulliana L, Citron A, Contessotto P, Battilana G, Panciera T, Ventre M, Netti
543 AP, et al. Simple yet effective methods to probe hydrogel stiffness for mechanobiology. Scientific reports.
544 2021; 11(1):1–17.

545 **Garg P**, Oikonomopoulos A, Chen H, Li Y, Lam CK, Sallam K, Perez M, Lux RL, Sanguinetti MC, Wu JC. Genome
546 editing of induced pluripotent stem cells to decipher cardiac channelopathy variant. Journal of the American
547 College of Cardiology. 2018; 72(1):62–75.

548 **Gershlak JR**, Resnikoff JI, Sullivan KE, Williams C, Wang RM, Black III LD. Mesenchymal stem cells ability to
549 generate traction stress in response to substrate stiffness is modulated by the changing extracellular matrix
550 composition of the heart during development. Biochemical and biophysical research communications. 2013;
551 439(2):161–166.

552 **Giacomelli E**, Meraviglia V, Campostrini G, Cochrane A, Cao X, Van Helden RW, Garcia AK, Mircea M, Kostidis
553 S, Davis RP, et al. Human-iPSC-derived cardiac stromal cells enhance maturation in 3D cardiac microtissues
554 and reveal non-cardiomyocyte contributions to heart disease. Cell stem cell. 2020; 26(6):862–879.

555 **Goldfracht I**, Efraim Y, Shinnawi R, Kovalev E, Huber I, Gepstein A, Arbel G, Shaheen N, Tiburcy M, Zimmermann
556 WH, et al. Engineered heart tissue models from hiPSC-derived cardiomyocytes and cardiac ECM for disease
557 modeling and drug testing applications. Acta biomaterialia. 2019; 92:145–159.

558 **Goldfracht I**, Protze S, Shiti A, Setter N, Gruber A, Shaheen N, Nartiss Y, Keller G, Gepstein L. Generating
559 ring-shaped engineered heart tissues from ventricular and atrial human pluripotent stem cell-derived car-
560 diomyocytes. Nature communications. 2020; 11(1):1–15.

Hansen A, Eder A, Bönstrup M, Flato M, Mewe M, Schaaf S, Aksehirlioglu B, Schwörer A, Uebeler J, Eschenhagen T. Development of a drug screening platform based on engineered heart tissue. Circulation research. 2010; 107(1):35–44.

Hasenfuss G, Mulieri LA, Blanchard EM, Holubarsch C, Leavitt B, Ittleman F, Alpert N. Energetics of isometric force development in control and volume-overload human myocardium. Comparison with animal species. Circulation research. 1991; 68(3):836–846.

Hofbauer P, Jahnel SM, Papai N, Giesshammer M, Deyett A, Schmidt C, Penc M, Tavernini K, Grdseloff N, Meledeth C, et al. Cardioids reveal self-organizing principles of human cardiogenesis. Cell. 2021; 184(12):3299–3317.

Huo J, Wei F, Cai C, Lyn-Cook B, Pang L. Sex-related differences in drug-induced QT prolongation and torsades de pointes: a new model system with human iPSC-CMs. Toxicological Sciences. 2019; 167(2):360–374.

Ivey MJ, Tallquist MD. Defining the cardiac fibroblast. Circulation Journal. 2016; p. CJ–16.

Jabbour RJ, Owen TJ, Pandey P, Reinsch M, Wang B, King O, Couch LS, Pantou D, Pitcher DS, Chowdhury RA, et al. In vivo grafting of large engineered heart tissue patches for cardiac repair. JCI insight. 2021; 6(15).

de Lange WJ, Farrell ET, Kreitzer CR, Jacobs DR, Lang D, Glukhov AV, Ralphe JC. Human iPSC-engineered cardiac tissue platform faithfully models important cardiac physiology. American Journal of Physiology-Heart and Circulatory Physiology. 2021; 320(4):H1670–H1686.

Lee A, Hudson A, Shiwarski D, Tashman J, Hinton T, Yerneni S, Bliley J, Campbell P, Feinberg A. 3D bioprinting of collagen to rebuild components of the human heart. Science. 2019; 365(6452):482–487.

Legant WR, Pathak A, Yang MT, Deshpande VS, McMeeking RM, Chen CS. Microfabricated tissue gauges to measure and manipulate forces from 3D microtissues. Proceedings of the National Academy of Sciences. 2009; 106(25):10097–10102.

Lewis-Israeli YR, Wasserman AH, Gabalski MA, Volmert BD, Ming Y, Ball KA, Yang W, Zou J, Ni G, Pajares N, et al. Self-assembling human heart organoids for the modeling of cardiac development and congenital heart disease. Nature communications. 2021; 12(1):1–16.

Li J, Zhang L, Yu L, Minami I, Miyagawa S, Hörning M, Dong J, Qiao J, Qu X, Hua Y, et al. Circulating re-entrant waves promote maturation of hiPSC-derived cardiomyocytes in self-organized tissue ring. Communications biology. 2020; 3(1):1–12.

Li RA, Keung W, Cashman TJ, Backeris PC, Johnson BV, Bardot ES, Wong AO, Chan PK, Chan CW, Costa KD. Bioengineering an electro-mechanically functional miniature ventricular heart chamber from human pluripotent stem cells. Biomaterials. 2018; 163:116–127.

MacQueen LA, Sheehy SP, Chantre CO, Zimmerman JF, Pasqualini FS, Liu X, Goss JA, Campbell PH, Gonzalez GM, Park SJ, et al. A tissue-engineered scale model of the heart ventricle. Nature biomedical engineering. 2018; 2(12):930–941.

Madonna R, Van Laake LW, Botker HE, Davidson SM, De Caterina R, Engel FB, Eschenhagen T, Fernandez-Aviles F, Hausenloy DJ, Hulot JS, Lecour S, Leor J, Menasché P, Pesce M, Perrino C, Prunier F, Van Linthout S, Ytrehus K, Zimmermann WH, Ferdinandy P, et al. ESC Working Group on Cellular Biology of the Heart: position paper for Cardiovascular Research: tissue engineering strategies combined with cell therapies for cardiac repair in ischaemic heart disease and heart failure. Cardiovascular Research. 2019 01; 115(3):488–500. https://doi.org/10.1093/cvr/cvz010, doi: 10.1093/cvr/cvz010.

Mannhardt I, Breckwoldt K, Letuffe-Brenière D, Schaaf S, Schulz H, Neuber C, Benzin A, Werner T, Eder A, Schulze T, et al. Human engineered heart tissue: analysis of contractile force. Stem cell reports. 2016; 7(1):29–42.

Mannhardt I, Saleem U, Mosqueira D, Loos MF, Ulmer BM, Lemoine MD, Larsson C, Amèen C, de Korte T, Vlaming ML, et al. Comparison of 10 control hPSC lines for drug screening in an engineered heart tissue format. Stem cell reports. 2020; 15(4):983–998.

Nunes SS, Miklas JW, Liu J, Aschar-Sobbi R, Xiao Y, Zhang B, Jiang J, Massé S, Gagliardi M, Hsieh A, et al. Biowire: a platform for maturation of human pluripotent stem cell–derived cardiomyocytes. Nature methods. 2013; 10(8):781–787.

Patel D, Stohlman J, Dang Q, Strauss DG, Blinova K. Assessment of proarrhythmic potential of drugs in optogenetically paced induced pluripotent stem cell-derived cardiomyocytes. Toxicological Sciences. 2019; 170(1):167–179.

613 **Polonchuk L**, Chabria M, Badi L, Hoflack JC, Figtree G, Davies MJ, Gentile C. Cardiac spheroids as promising in
614   vitro models to study the human heart microenvironment. Scientific reports. 2017; 7(1):1–12.

615 **Richards DJ**, Li Y, Kerr CM, Yao J, Beeson GC, Coyle RC, Chen X, Jia J, Damon B, Wilson R, et al. Human cardiac
616   organoids for the modelling of myocardial infarction and drug cardiotoxicity. Nature Biomedical Engineering.
617   2020; 4(4):446–462.

618 **Rodriguez ML**, Graham BT, Pabon LM, Han SJ, Murry CE, Sniadecki NJ. Measuring the contractile forces of
619   human induced pluripotent stem cell-derived cardiomyocytes with arrays of microposts. Journal of biome-
620   chanical engineering. 2014; 136(5):051005.

621 **Ronaldson-Bouchard K**, Ma SP, Yeager K, Chen T, Song L, Sirabella D, Morikawa K, Teles D, Yazawa M, Vunjak-
622   Novakovic G. Advanced maturation of human cardiac tissue grown from pluripotent stem cells. Nature.
623   2018; 556(7700):239–243.

624 **Saini H**, Navaei A, Van Putten A, Nikkhah M. 3D cardiac microtissues encapsulated with the co-culture of
625   cardiomyocytes and cardiac fibroblasts. Advanced healthcare materials. 2015; 4(13):1961–1971.

626 **Schaaf S**, Shibamiya A, Mewe M, Eder A, Stör A, Hirt MN, Rau T, Zimmermann WH, Conradi L, Eschenhagen T,
627   et al. Human engineered heart tissue as a versatile tool in basic research and preclinical toxicology. PloS
628   one. 2011; 6(10):e26397.

629 **Seguret M**, Vermersch E, Jouve C, Hulot JS. Cardiac Organoids to Model and Heal Heart Failure and Cardiomy-
630   opathies. Biomedicines. 2021; 9(5):563.

631 **Shadrin IY**, Allen BW, Qian Y, Jackman CP, Carlson AL, Juhas ME, Bursac N. Cardiopatch platform enables
632   maturation and scale-up of human pluripotent stem cell-derived engineered heart tissues. Nature commu-
633   nications. 2017; 8(1):1–15.

634 **Shantsila E**, Watson T, Lip GY. Drug-induced QT-interval prolongation and proarrhythmic risk in the treatment
635   of atrial arrhythmias. Europace. 2007; 9(suppl_4):iv37–iv44.

636 **Sharma A**, Li G, Rajarajan K, Hamaguchi R, Burridge PW, Wu SM. Derivation of highly purified cardiomyocytes
637   from human induced pluripotent stem cells using small molecule-modulated differentiation and subsequent
638   glucose starvation. JoVE (Journal of Visualized Experiments). 2015; (97):e52628.

639 **Stillitano F**, Turnbull IC, Karakikes I, Nonnenmacher M, Backeris P, Hulot JS, Kranias EG, Hajjar RJ, Costa KD. Ge-
640   nomic correction of familial cardiomyopathy in human engineered cardiac tissues. European heart journal.
641   2016; 37(43):3282–3284.

642 **Streckfuss-Bömeke K**, Wolf F, Azizian A, Stauske M, Tiburcy M, Wagner S, Hübscher D, Dressel R, Chen S, Jende
643   J, et al. Comparative study of human-induced pluripotent stem cells derived from bone marrow cells, hair
644   keratinocytes, and skin fibroblasts. European heart journal. 2013; 34(33):2618–2629.

645 **Thavandiran N**. Design of Human Pluripotent Stem Cell-derived Cardiac Microtissue-based Platforms for Func-
646   tional Screening. PhD thesis, University of Toronto; 2019.

647 **Tiburcy M**, Hudson JE, Balfanz P, Schlick S, Meyer T, Chang Liao ML, Levent E, Raad F, Zeidler S, Wingender E,
648   et al. Defined engineered human myocardium with advanced maturation for applications in heart failure
649   modeling and repair. Circulation. 2017; 135(19):1832–1847.

650 **Turnbull IC**, Karakikes I, Serrao GW, Backeris P, Lee JJ, Xie C, Senyei G, Gordon RE, Li RA, Akar FG, et al. Advancing
651   functional engineered cardiac tissues toward a preclinical model of human myocardium. The FASEB Journal.
652   2014; 28(2):644–654.

653 **van der Velden J**, Asselbergs FW, Bakkers J, Batkai S, Bertrand L, Bezzina CR, Bot I, Brundel BJ, Carrier L,
654   Chamuleau S, et al. Animal models and animal-free innovations for cardiovascular research: current sta-
655   tus and routes to be explored. Consensus document of the ESC working group on myocardial function and
656   the ESC Working Group on Cellular Biology of the Heart. Cardiovascular research. 2022; 118(15):3016–3051.

657 **Williams K**, Liang T, Massé S, Khan S, Hatkar R, Keller G, Nanthakumar K, Nunes SS. A 3-D human model of
658   complex cardiac arrhythmias. Acta Biomaterialia. 2021; 132:149–161.

659 **Zhao Y**, Rafatian N, Feric NT, Cox BJ, Aschar-Sobbi R, Wang EY, Aggarwal P, Zhang B, Conant G, Ronaldson-
660   Bouchard K, et al. A platform for generation of chamber-specific cardiac tissues and disease modeling. Cell.
661   2019; 176(4):913–927.

662 **Zhuang RZ**, Lock R, Liu B, Vunjak-Novakovic G. Opportunities and challenges in cardiac tissue engineering from
663   an analysis of two decades of advances. Nature Biomedical Engineering. 2022; 6(4):327–338.

A.

1 UV-Curing the PEG-based hydrogel and removing the ring-shaped mold.

2 Positioning the hollow glass capillary to be in contact with the cured hydrogel.

3 Applying a pressure p to the hollow glass capillary, creating a convexity of the surface of the hydrogel.

4 $$E = \frac{p}{0.872\frac{l}{a} + 0.748(\frac{l}{a})^2}$$

Imaging and analyzing using Image J. Calculation of the Young's Modulus E.

B.

Figure 1—figure supplement 1. A. Procedure to measure the Young's Modulus of a PEG-gel disk by pipette aspiration. **B.** Young's modulus of 5% PEG gels (n=44).

Quantification of differentiation efficiency

Figure 1—figure supplement 2. Percentage of positive cells for Troponin T2 (TNNT2) determined by flow cytometry in the 6 different differentiations of wild-type (WT) iPSCs, used for these experiments.

Figure 2—figure supplement 1. Contractile parameters at D14 for the different fibroblasts ratios. **A.** Contraction frequency. **B.** Contraction stress. **C.** Maximum contraction speed. **D.** Maximum relaxation speed.

Figure 2—figure supplement 2. Larger scale confocal pictures of immunostained tissues at different heights basal layer(left) and cardiac ring (right) at 40X magnification: vimentin is stained in green, troponin in red, and DAPI in blue. Scale bars: 50 μm.

Figure 2—figure supplement 3. Immunostaining of rings at D1 and D7. **Top panels:** Troponin T (red), Vimentin (green) and DAPI (blue) staining of a ring at D1, basal layer (left) and cardiac ring (right). **Bottom panels:** Troponin T (red), Vimentin (green) and DAPI (blue) staining of a ring at D7, basal layer (left) and cardiac ring (right) - Scale bars: 100μm

Contraction strain ε$_A$

Figure 3—figure supplement 1. Evolution of contraction strain ϵ_A through time after seeding at days 1, 3, 7, 10 and 14. Contraction strain changes through time are significant (p<0.0001 - ANOVA for repeated measures - D1: n=57, D3: n=59 D7: n=47, D10: n=43, D14: n=36 tissues, from 3 differentiations). Contraction strain for each concentration is compared to contraction strain at day 1 (****: p<0.0001).

Figure 3—figure supplement 2. Contractile parameters of rings at D28 expressed as a fold change to D14. Mann-Whitney test - D14: n=29 and D28: n=17, from 2 differentiations