1 **Title:**

2 A high-throughput multiplex array for antigen-specific serology with automated analysis.

3 Rosenberg, A.F.[1], Killian, J.T. Jr.[1], Green, T.J.[1], Akther, J.[3], Hossain, M.E.[1], Qiu, S

4 Randall, T.D.[2], Lund, F.E.[1], and King, R.G.[1]

5
6
7 **Affiliations:**
8 [1]Department of Microbiology, University of Alabama at Birmingham (UAB), Birmingham,
9 AL 35294, USA
10
11 [3]Department of Medicine, Division of Clinical Immunology and Rheumatology, UAB,
12 Birmingham, AL 35294, USA
13
14

16 *Corresponding author: rgking@uab.edu
17

Abstract.

High-throughput customizable CBA automated script-based analysis.

The utility of high-throughput systems to evaluate antigen-specific Ab has been highlighted by the SARS-CoV-2 pandemic. Pathogen specific antibody levels are often used to assess protection following vaccination and, in the case of novel pathogens, an indication of prior exposure. Several platforms exist to visualize antigen-specific Ab, however most are not quantitative and difficult to scale for population levels studies. Additionally, the sensitivity across platforms differs making direct comparisons between studies difficult. Cytometric Bead Arrays are an attractive platform for antigen-specific Ab measurements as they allow antibodies reactive against several antigens and of several isotypes to be performed simultaneously. Additionally, cytometric arrays exhibit a high sensitivity and can be designed to provide quantitative measurements. Using commercially available particles, a biotin-Streptavidin loading strategy, and the inclusion of indirect standards, we describe here a flexible system that can be modified to include a variety of antigens. We generated two arrays, one focused on β-Coronavirus antigens and one focused on Influenza. To support the high throughput capacity of this system, we developed a suit of automated tools to process raw data into antigen-reactive IgM, IgA, and IgG. We describe quality control requirements, assay performance, and normalizations to accurately quantitate antigen specific Ig.

1. Introduction

Serum antibody (Ab) levels targeting pathogens have been used to predict protection, determine exposure history, and provide a measure of efficacy of vaccination. The ongoing COVID pandemic has highlighted the need for rapid and accurate quantification of antigen-specific Abs in human serum or fluid samples. Early in 2020, the human immune response to SARS-CoV-2 was uncharacterized and considerable effort was devoted to determining how individuals responded to infection, and whether there were correlates of Ab responses to disease outcomes. Even following the approval of multiple SARS-CoV-2 vaccines, it was unclear what serological features of the humoral immunity afforded protection from infection or re-infection. Unfortunately, the evolution of the Sars-2-CoV-2 pandemic over the last couple years have illuminated the shifting landscape of variants and the uneven protection afforded by the original vaccination formulations [1, 2]. Despite numerous published studies, some involving large-cohorts, addressing the magnitude and longevity of SARS-CoV-2 humeral immunity, there remains a need for inexpensive, high-throughput methodologies to evaluate Ab responses following or infection or vaccination [3, 4]. Additionally, it is now clear that not all Abs directed toward the Sars-2-CoV spike protein (C SP) are protective, necessitating the visualization of C SP sub-domain reactivity as well as the distribution of Abs across isotypes and toward new emerging variants. Although several platforms have been described to evaluate SARS-CoV-2-reactive Abs including [5-12], few of these methodologies are quantitative, making direct comparisons across platforms or the cohorts targeted in these studies difficult, thereby hampering our ability to define levels of Ab reactive with C SP or subdomains that correlate with protection. Such analyses are becoming increasingly important as SARS-CoV-2 variants resistant to prior vaccine responses continue to emerge [2]. Although the covid pandemic has increased the awareness of the utility of platforms to evaluate antigen specific Ab, such platforms are important for a host of other pathogens including seasonal influenza (INF) virus. To address this need, we developed a highly customizable cytometric bead array platform to quantitate specific Abs against several pathogen associated antigens of IgG, IgM, and IgA subclasses. Coupled with a

automated MATLAB analysis pipeline, this system allows for the high-throughput
quantification of Abs reactive to a wide variety of antigens.

Cytometric bead arrays have been widely employed to determine the
concentration of diverse analytes from biological samples including Abs [cite CBA review].
They offer several advantages over other traditional approaches including the ability to
multiplex ligands, enhanced sensitivity compared to traditional ELISA assays, and a
broad dynamic range[10, 13]. Because of these properties, CBAs are an attractive
platform to quantitate Abs reactive to many antigens simultaneously in a high-throughput
format.

Here we describe the generation of a highly multiplexed CBA system to evaluate
antigen specific Ab responses and describe its use targeting two pathogens. One array
consists of 12 antigens derived from B-CoV virus spike ectodomains (SP) and
nucleocapsid proteins (NP). The second incorporates 14 INF hemagglutinins (HA) in
addition to influenza NP and non-structural protein 1 (NS1). To support this platform, we
developed a post-acquisition, MATLAB-based analytic pipeline to streamline data
analysis and enable high-throughput use. In addition, we have evaluated the performance
of the array using panels of monoclonal and polyclonal human Abs providing antigen
specific and generalizable corrections allowing for the accurate quantification of serum
Ab of the IgM, IgA, and IgG subclasses toward each of the array constituents. Although
the configuration of the arrays described here are focused on B-CoV and INF derived
antigens, the platform is flexible and utilizes indirect standards which allow for the Ab
reactive to essentially any protein antigen to be quantitated.

2.   Material and methods

2.1   *Array generation*.

Cytometric array described here was constructed using spherotech 4um and 5um
Carboxy Blue Particle Array Kits (Spherotech CPAK-4067-8K and CPAK-5067-10K
respectively). Antigens were loaded by first conjugating Streptavidin (SA) to via amide
using commonly employed 1-Ethyl-3-(3-dimethylaminopropyl)carbodiimide (EDC)
chemistry. Streptavidin (SA) (Southernbiotech 7105-01) was buffered exchanged using
PD-10 columns (Cytiva 17085101) into PBS and diluted to 2mg/mL. For conjugation,

10e8 shperotech particles were isolated by centrifugation at 10,000Xg for 3 min. After
carefully removing the supernatant, the bead pellet was resuspended in 0.5 mLs of SA in
PBS. Following complete resuspension by pipetting, 0.5 mLs of 6mM EDC dissolved in
0.05M MES buffer PH 5.0 was added, and the reaction mixture was rotated at room
temperature overnight. After the conjugation reaction was complete, 0.1 mL of 1M tris PH
8.0 was added to quench the reaction. Following a 1 hr incubation, the beads were
harvested by centrifugation as described above and washed twice in 1mL of PBS.
Following the final wash, beads were resuspended in $1e10^8$ particles/mL of PBS with
0.25%NaN3 and stored at 4deg.

Following SA conjugation, quality control experiments were performed to
determine the degree and uniformity (when multiple particle sizes and/or peak identities
are used) of labeling by staining the SA-conjugated particles with fluorescently- labeled-
biotinylated hemagglutinin (PR8). Individual array constituents were mixed and diluted to
1e6 of each particle/mL. 40ul of serial dilutions of PR8 were prepared in a 96 well U
bottom plate (Corning 3797) ranging from 1ug/mL to 2ng/mL. 5ul of the bead suspension
was added, mixed by pipetting, and incubated for 15 min at RT. 200ul of PBS was then
added and the plate was centrifuged at 3000Xg for 5min. The beads were resuspended
in 80uL of PBS. Samples were then analyzed by flow cytometry.

Following SA coupling and quality control procedures described above, biotinylated
recombinant array antigens were passively absorbed onto the individual particles. For
antigens described here (table 1), a single biotin site was added enzymatically onto a
carboxy terminal AVI tag. SA conjugated particles were harvested by centrifugation as
described above and resuspended in 1mg/mL of the recombinant proteins in 1%BSA in
PBS. Antigen loading was carried out by rotating overnight at 4deg. Following absorption,
the beads were harvested by centrifugation as described, and washed twice with 1% BSA
in PBS. Finally, the antigen coated beads were resuspended at 1e8 particles/mL 1% BSA
in PBS, 0.25% NaN3 and stored at 4deg until use.

2.2 Recombinant antigens.

Production of pre-fusion recombinant Spike (S) proteins. Ectodomain S pre-fusion trimers
(SARS-CoV-2 $S_{14-1211}$) were produced by co-transfecting S-AviTag and S-6X-HisTag

constructs into FreeStyle 293-F Cells at a 1:2 ratio. Transfected cells were cultured in

FreeStyle 293 Medium for 3 days and recombinant S trimers purified from culture

supernatant by FPLC using Nickel-affinity chromatography. Purified proteins were

biotinylated by BirA enzyme in vitro.

Production of S subdomains. NTD (SARS-CoV-2 S_{14-305}) and RBD (SARS-CoV-2 S_{319-}

$_{541}$) monomers with a C-terminal dual AviTag/6X-HisTag sequence were produced by co-

transfecting single constructs into FreeStyle 293-F Cells. Following a 3-day expression,

subdomains were purified from culture supernatant by FPLC using nickel-affinity

chromatography and biotinylated by addition of BirA.

Production of CoV nucleocapsid protein (N). Recombinant N containing the full-length N

and tandem AviTag/6X-HisTag sequence were produced by co-transforming Rosetta

cells with the N expression plasmid and an inducible BirA expression plasmid. Cells were

grown in the presence of chloramphenicol, ampicillin, and streptomycin, induced with

IPTG and supplemented with biotin. Biotinylated N protein was purified by FPLC using a

nickel-affinity column and subsequent size exclusion chromatography.

Recombinant influenza HA protein production. The nucleic acid sequence corresponding

to amino acids 18-524 of influenza HA ectodomains purchased and synthetic genes

(GeneArt, Regensburg, Germany) from influenza virus strains listed in table 1 as

previously described [14], synthetic constructs were cloned into the pCXpoly(+) vector

modified with a human CD5 leader sequence and a GCN4 isoleucine zipper trimerization

domain (GeneArt) followed by either a 6XHIS tag (HA-6XHIS construct) or an AviTag

(HA-AviTag construct). The HA–6XHIS and HA–AviTag constructs for each HA were co–

transfected using 293Fectin Transfection Reagent into FreeStyle™ 293–F Cells

(ThermoFisher Scientific) at a 2:1 ratio. Transfected cells were cultured in FreeStyle 293

Expression Medium for 3 days. The supernatant was then cleared by centrifugation.

Recombinant HA molecules were purified from these culture supernatants by FPLC using

a HisTrap HP Column (GE Healthcare) and eluted with 250 mM imidazole.

Production of influenza NP and NS1 proteins.

The coding sequence of NP from PR8 was synthesized in frame with the coding sequence

for a 15–amino acid biotinylation consensus site [15] on the 3′ end (GeneArt). The

modified NP sequence was cloned in frame to the 6X His tag in the pTRC–His2c

expression vector (Invitrogen). NP protein was expressed in E. coli strain CVB101

(Avidity), purified by FPLC using a HisTrap HP Column (GE Healthcare) and eluted with

a 50–250 mM gradient of imidazole. Coding sequence for the influenza virus A/PR8/34

NS1 gene (amino acids 3–230, Accession Number: AAA43536) was chemically

synthesized (GeneArt) in frame with a 3' addition of the coding sequence for the BirA

enzymatic biotinylation consensus site136 and the 6X-His purification tag. Two mutations,

R38A and K41A, demonstrated to prevent aggregation of NS1 at high concentrations137,

were introduced into the NS1 gene by site-directed mutagenesis. The recombinant NS1

gene was cloned into the pTRC–His2c expression vector (Invitrogen) and expressed in

the BirA-enzyme containing E. coli strain CVB101 (Avidity) Biotinylated recombinant NS1

was purified by FPLC using a HisTrap HP Column (GE Healthcare), eluted with a 50–250

mM imidazole gradient, buffer exchanged and concentrated.

Antibodies and standards.

Detection of IG fluorescent goat polyclonal anti-IG F(ab')2 secondaries, Southern Biotech

(IgM cat 2022-02, IgG cat# 2062-09, IgA cat# 2052-09) Isotype standards were generated

by performing the array on mixtures of IG capture beads with 0.75x serial dilutions of

purified human Abs southern biotech (IgG cat# 0150-01, IgM Cat# 0158L-01, IgA

cat#0155L-01) in ranging from 1ug/mL to 1.3 ng/mL. Anti-SARS Abs JTK137_H22 and

JTK137_K6 were generated from the C SP binding B cells derived from a vaccinated

donor. The C SP reactive Ab CR3022 was acquired from IDT Biologika. C SP reactive ab

AHA001 was purchased from Sanyou Bio. HA-reactive Abs (P52d7_K12, P52d7_I1, and

P52d7_D20) were generated from HA-binding B cells from a vaccinated donor [16].

Polyclonal anti-C SP reactive Ab was purified from the serum of a vaccinated donor using

affinity chromatography. Briefly, 0.5 mg of purified C SP trimers generated as described

was conjugated to cyanogen bromide activated Sepharose (GE) following manufactures

protocol. 10mLs of serum was diluted into 40mLs of PBS filtered using a 0.45um syringe

filter (Corning 431220). C SP conjugated Sepharose was added and rotated overnight at

4deg. The resin was collected by centrifugation at 2000g for 10min and washed twice

with 50 mL of cold PBS. After the final wash, the resin was loaded into a disposal micro

column (Pierce 89868) and eluted with 5 fractions of 200ul of 0.1M glycine pH 2.8 into

25ul of 1M tris pH 9.0. Eluted fractions were pooled and concentrated by centrifugation

through an 30K centrifugal filter (Millipore Sigma UFC500324). Following concentration,

the samples was buffer exchanged into PBS using a 0.5mL Zeba column (Fisher scientific

PI87767) and quantitated.

2.3   Assay protocol

Serum samples were diluted into PBS (1/7150 for IgG detection, or 1/500 for IgM and IgA

detection) and arrayed in 96 well u-bottom plates. A 5µl suspension containing 5x1e5 of

each antigen coated microparticles was added to the samples. In the case of Ig standards,

anti-IgM, anti-IgA, and anti-IgG beads were added to 50ul of the serial dilutions of

standard Abs. The suspensions were mixed by pipetting and incubated for 15 mins at

room temperature. The beads were washed by the addition of 200µl of PBS and

centrifugation at 3000g for 5min at room temperature. The CBA particles were

resuspended in a secondary staining solution consisting of the appropriate anti-Ig

secondary diluted 1/400 in 1%BSA in PBS. The suspension was incubated for 15 min in

the dark at room temperature. The beads were washed by the addition of 200 µl of PBS

and pelleted by centrifugation at 3000g for 5 min at room temperature. The particles were

resuspended in 80µl of PBS and directly analyzed on a BD Cytoflex flow cytometer in

plate mode at sample rate of 100ul per minute. Sample collection was stopped following

the acquisition of 75 µL. Following acquisition, the resulting FCS files were processed

using the software described below.

2.4   Automated analysis and software

We developed custom software to automatically process FCS files to rapidly quantify the

Ab reactivities of serum samples. This software was developed in Matlab (The

Mathworks, Inc. Natick MA, USA) version R2020a on MacOS. It requires the Statistics

and Machine Learning Toolbox, the Curve Fitting Toolbox and the Signal Processing

Toolbox. Additionally, code from Matlab Central (www.mathworks.com/matlabcentral/)

public code repository was incorporated into this program [17-20]. This software is

available on GitHub [link here].

2.5 *Human serum samples.*

Human serum samples were collected as part of the UAB Cooperative Centers for

Human Immunology program from organ donors. Peripheral blood from human subjects

was drawn into K2-EDTA tubes (BD Bioscience). Plasma were isolated by density

gradient centrifugation over Lymphocyte Separation Medium (CellGro), aliquoted, and

stored in LN2. Cryopreserved samples were thawed at room temperature and diluted

1/100 in PBS in a 96 well U bottom plate to generate a 'master plate' stored at -20 deg.

Final dilutions were generated in assay plates as needed. For assay optimization and

dilution determination serum was derived from consented participants under the UAB

Institutional Review Board approved the study protocol (IRB-300005127).

3.   Results and discussion

3.1   Generation of the B-COV and INF arrays.

To simultaneously interrogate Ab reactivity to multiple viral protein antigens, we

adopted a commercially available cytometric array platform by spherotech and an indirect

conjugation methodology. Streptavidin is conjugated to carboxy modified microparticles

using commonly employed EDC chemistry to allow for the passive absorption of

biotinylated antigen onto each individual bead. Our strategy is to maximize SA

conjugation such that the binding capacity of each bead is high and uniform, followed by

the absorption of antigens at saturating concentrations. The uniformity of SA loading is

evaluated by staining bead preparations with a recombinant influenza hemagglutinin

(PR8) trimer, containing a single N-terminal biotin site, fluorescently labeled with Alexa

488. The intensity of labeling relates to the absorption of the fluorescent trimer and is

typically uniform across the array (FIG 1). The particles sizes exhibit different maximum

intensities, in part due to the difference of surface area between the 4um and 5um

particles (Fig 1c). SA loading is uniform within the constituents of each particle size,

Figure 1. Array generation and loading controls. A. Forward and side scatter histograms for differentiating the indicated size particle. **B.** Flow cytometric histogram of the array constituents before (left, unshaded) and following incubation with biotinylated / alexa488 influenza hemagglutinin PR8 (right, shaded). **C.** Mean fluorescent intensity of each array constituent (left) and corrected (right) by relative surface area when stained with the indicated concentration of PR8:biotin:488.

however, with a coefficient of variation of 5.39% and 5.34% for all 5um and 4um beads

respectively. The array exhibits a highly uniform decay of signal as a function of analyte

concentration, suggesting both that the binding capacity of each bead is uniform and that

the binding is highly dependent on analyte concentration.

To determine the binding capacity of the SA conjugated beads, we adopted two

independent approaches. First, we incubated the SA conjugated particles with an excess

of a known but variable amount of biotinylated influenza hemagglutinin trimers,

Figure 2. Dilution curves of antigen specific antibodies. **A.** Plots of the MFI of antigen-coated and anti-IgG CBA beads stained with serial dilutions of the indicated CS SP (green lines) or CA09 (blue lines) reactive mAbs. **B**. MFI of CS SP Beads stained with serial dilutions of human serum from COVID convalescent patients visualized with secondaries reactive to the indicated human Ig isotype.

───

and measured the protein absorbed from these suspensions. For the two independent

lots of SA-conjugated beads tested, an average equivalent of 220ug per 1×10^8 particles

was absorbed. Next, we incubated the SA beads with decreasing concentrations of

fluorescently labeled PR8 as described above, as assayed the mean fluorescence

intensity. The point of inflection at which concentration the beads appear to be

subsaturated equates to the same value, 200ug of PR8 per 1X1e8 beads, thus we

estimate the typical biotin binding capacity of SA beads to be 0.1 picomole per particle.

3.2 Array validation and characterization:

To generate a B-CoV focus array, we absorbed SARS-CoV-1, and SARS-CoV-2 full

length, pre-fusion stabilized trimers as well as monomeric RBD subunits of these proteins

Table 1.

Influenza			
Peak #	**Antigen**	**Antigen description**	**Protein Accession #**
5uM-1	PR8	Influenza HA (H1)	MH785011.1
5uM-2	CA09	Influenza HA (H1)	NC_026433.1
5uM-3	Hawaii2019	Influenza HA (H1)	MN976432.1
5uM-4	MICH15	Influenza HA (H1)	MK622940.1
5uM-5	Brisbane	Influenza HA (H1)	MN569031.1
5uM-6	X31	Influenza HA (H3)	DQ874876.1
5uM-7	TX50	Influenza HA (H3)	KC892952.1
5uM-8	SW13	Influenza HA (H3)	EPI543739
5uM-9	SING2016	Influenza HA (H3)	EPI2053182
4um-2	HK2014	Influenza HA (H3)	NA
4um-3	NETH1999	Influenza HA (H2)	CY122252.1
4um-4	VN2004	Influenza HA (H5)	HM006759.1
4um-5	CAN04	Influenza HA (H7)	CY015013.1
4um-6	HongKong2013	Influenza HA (H9)	NA
4um-7	INF NP	Influenza Nucleocapsid protein	NCAP_I34A1
4um-8	INF NS1	Influenza Non-structural protein	AAA43536
B-CoV			
4um-2	S NP	Sars-CoV-1 Nucleocapisid Protien	AAP13445.1
4um-3	M NP	MERS Nucleocapisid Protien	P_009047211.1
4um-4	O NP	OC43 Nucleocapisid Protien	KR055606.1
4um-5	H NP	HKU1 Nucleocapisid Protien	Q5MQC6.1
4um-6	C NP	Sars-CoV-2 Nucleocapisid Protien	QHD43423
5uM-1	S SP	Sars-CoV-1 Spike ectodomain	AAX16192.1
5uM-2	S RBD	Sars-CoV-1 Receptor binding domain	AAX16192.1
5uM-3	CS SP	Sars-CoV-2 Spike ectodomain	QHD43416.1
5uM-4	CS RBD	Sars-CoV-2 Receptor binding domain	QHD43416.1
5uM-5	CS NTD	Sars-CoV-2 N-Terminal domain	QHD43416.1
5uM-6	M SP	MERS Spike ectodomain	AKI29255.1
5uM-7	O SP	OC43 Spike ectodomain	AAT84362.1
5uM-8	H SP	HKU1 Spike ectodomain	Q5MQD0.1
5uM-9	CA09	Influenza HA (H1)	NC_026433.1

Table 1. List of antigens used to generate the Inf and CoV antigen Arrays.

The size and peak number of each array element is listed, with the biotinylated

recombinant protein loaded to generate the two arrays. The protein accession numbers

are provided.

____

and the NTD of SARS-CoV-2 onto constituents of the array. Additionally, we chose two

related B-CoV spike proteins derived from HKU1 and OC43, as well as the nucleocapsid

proteins from all of these viruses, and finally, INF HA (CA09) to serve as an unrelated

positive control. The distribution of these molecules within the array is listed in Table 1.

For the antigens for which mAbs were available, antigen loading was validated by specific

mAb staining. When the mAbs I1 and CR3022, reactive with CA09 and C SP respectively,

were subjected to the assay, the MFI derived from antigen coated beads were highly

correlated with IgG capture beads conjugated with anti-human IgG, and nearly identical

for each mAb-antigen at a range of 1ug/mL to 1ng/mL. In contrast, the staining of I1 and

CR3022 to non-cognate antigen coated beads was minimal for both mAbs (Fig 2a). These

data demonstrate the effective range and sensitivity of the assay, which is linear for nearly

three orders of magnitude and can reliably detect Abs present at 10ng/mL in this

configuration.

3.3 Assay functional range and dilution factor determination

Constituents of the B-CoV array, like most cytometric bead arrays, exhibits significant

prozone effect at high Ab concentrations (Fig 2b) . This is especially true when polyclonal

sera are analyzed. Despite the large dynamic range, it is important to choose dilutions at

which the resulting MFI can be accurately extrapolated to concentration. To determine

the Ab concentration at which the prozone effect is observed and to identify a dilution

range allowing for the determination of antigen reactive IgG, IgM, and IgA. We analyzed

ten convalescent serum samples a dilution factor between 1/500 and 1/20517. As can be

seen in fig 3, the detection of IgG yielded significantly different maximum intensities and

dilution curves then IgA and IgM. For samples that contained detectable anti-C SP IgG,

the prozone effect was apparent at dilutions up to 1/7000 after which the signal decayed

as a function of Ab concentration. IgM and IgA signals, although detectable at 1/7000 in

some samples, never reached a concentration that exhibited prozone effects. This

difference is likely due to both the relative concentrations of these Ig subclasses and

differences in the signal intensity derived from the isotype specific secondaries. Using this

experiment as a guide, we selected a 1/7100 dilution for IgG 1/500 for IgM/IgA in

subsequent experiments. Depending on the Ab levels within a cohort and the antigen of

interest, these parameters require empirical identification.

3.4 Antibody standards and quantification:

Although the data described above suggested that the CBA would provide sensitive
quantitative assessment of the presence of specific Ab, we sought to refine the assay to
accurately measure antigen specific Ab present in biological specimens. We first
attempted to utilize monoclonal Abs of defined reactivity toward C SP and HA to generate
antigen-specific standards. Although this is a traditional approach, the relationship
between Ab concentration and MFI resulting from monoclonal Abs is dependent several
factors, including the distribution of Ab isotype, which are unequally recognized by pan
IgG secondaries, the affinity of the Ab, and the valency of Ab-antigen interactions.
Because the derived MFI results from the association of the fluorescently labeled
secondary Abs with the beads is independent of the nature of the antigen, we passively
absorbed biotinylated anti-Ig secondaries onto the matched SA conjugated beads used
to generate the antigen array. To determine the accuracy of these assessments, we
compared the MFI of antigen coated and standard beads following staining with mAbs
reactive against either C SP or HA visualized with fluorescently conjugated anti-IgG
secondaries. As seen in figure 2, the signals derived from both the antigen coated and
Ig capture conjugated bead were highly similar across the dilution range, although the
ratio differed for each monoclonal. This likely reflects differences in the affinity of these
mABs, as has been described for HA reactive Abs [16]. These data suggest that
biotinylated anti-isotype capture Abs serve as an accurate surrogate for the
quantification of bead-bound antigen and, assuming the antigen density was equivalent,
as an indirect standard for the Ab quantification. To control for differences in the binding
capacity between different particle sizes, and minor deviations between the individual
array components, we generated a panel of ig-capture beads, distributing the IgM, IgA,
and IgG capture Abs on every third bead peak across all the individual constituents of
the array. Purified IgM, IgA, or IgG was serially diluted from 1ug/mL to ~1ng/mL to
correlate the MFI of the capture beads to the quantity of the standard, and these curves
were then used to extrapolate the Ab isotype bound to the antigen specific beads
providing an accurate quantification (Fig S2).

3.5 Automated Analytic Workflow

Although the array data is collected using a traditional flow cytometer and can be analyzed
by any number of flow cytometric software packages, the inclusion of multiple standard
curves and the large number of individual flow files make manual data processing
laborious. Additionally, although antigen specific Ig concentrations are derived values
determined following the export of MFI values corresponding to each bead peak and the
application of a curve derived from the standards, the quality of each sample must be
determined individually with traditional flow analysis software prior to data export
rendering this method labor intensive and error prone. To circumvent these problems and
perform automated sample quality assessments and batch processing, we developed
software pipeline that automatically processes raw data files from the cytometer to
construct sample-by-feature tables of Ab concentrations. This software, written in Matlab
(The Mathworks, Inc., Natick MA, USA), processes individual raw FCS files to identify
events associated with beads, find peaks, summarize data, and enables a batch mode
which can process many files at once including standard controls useful for back-
calculating concentration units from mean fluorescence intensity (MFI) data. The overall
workflow is shown in Figure 3. The command line program requires four arguments: a
directory containing all the FCS files, a sample manifest containing metadata for each of
the corresponding FCS files (including the flow cytometry channel mappings for the bead
channel and the secondary isotype channel), a configuration file describing the
association of bead peak order with protein for the specific assay, and a configuration file
containing parameters for peak identification. The output is a Matlab structure array
where each element contains a table of isotype reactivity for each feature in the assay as
well as other program settings for complete analytical provenance to support
reproducibility. This structure array, in turn is used for an analysis quality assessment
program to visually summarize the consistency of the bead gates for each bead size as
368
369

Figure 3. Quantitated subunit and polyvalency corrections. A. Extrapolated

concentration of anti- CS SP pAb and pIgG plotted against the known concentration. **B**.

Calculated polyvalency correction factor for anti-CS SP pAb IgG. Fold difference in MFI

between CS SP and CS RBD across the linear range of dilutions is shown. **C**. ArcSIn

transformed MFI values derived from the array plotted against known concentration of

anti-CS SP pAb and uncorrected values derived from the indicated NISBC standards.

Corrected values were generated by linear scaling of the extrapolated concentrations.

D. Extrapolated anti-CS SP antibody concentrations (before and after polyvalent

correction) of the indicated NISBC standards. Dashed lines indicated the high

confidence Ab concentration of the indicated standards. **E.** Mean MFI of the C SP
coated (black) or C RBD coated (blue) beads stained the indicated concentrations of
anti-CS RBD reactive rAbs. The resulting mean and SD of the MFIs of CS SP and CS
RBD are shown. **F.** The ratio of the MFI of CS SP to CS RBD of the indicated anti-CS
RBD rAbs at the indicated concentration.

____

well as the numbers of beads for each bead gate as shown in Figure S1. This structure
array is also used to generate standard curves for each isotype, bead size and sample
dilution (multiple dilutions can be run together and are annotated in the sample manifest,
and each requires their own set of standard curve samples).   The samples used to
generate the standard curves are annotated as such in the sample manifest. These data
are used to compute a four-parameter logistic (4PL) fit for each bead size/isotype/dilution
as shown in Figure S2. Finally, the structure array and the 4PL fits are used to back-
calculate concentration units for the MFI data, and write the entire data set as a single,
tabular text file suitable for importing into any analysis program or a relational database
for large-scale studies. This software is suitable for high-throughput use. In a typical
configuration, data derived from five 96-well plates run on a Cytoflex (Beckman Coulter,
Indianapolis IN, USA) – 480 samples including standards – can be processed in about a
minute on a current model Macbook Pro with no special memory or processor
requirements.

3.6 Deriving MFI Reactivity Data from a Sample
The computational workflow parallels the process for manually quantifying reactivity and
is shown in Figure 4. First, MFI data are extracted from an FCS file [17] and transformed
using the hyperbolic arcsine [18]. Next, a forward-scatter vs. side-scatter plot will reveal
events corresponding to the different sized beads as densities of points. These are
automatically detected by looking for the expected number of peaks along the forward-
scatter (x) axis to define gates, and then, within each, identify single peaks on the side-
scatter (y) axis. The intersection of these gates in 2-D space are used to define oval
gates (Figure 4a), and events within these gates are annotated as distinct populations of

Figure 4.

Figure 4. Software components are shown within red dashed box. Green boxes are

modules that perform a single element of processing. The light blue box is a program

that processes a single FCS file to extract transformed MFI data for all features in the

assay. The darker blue box represents a wrapper to batch process many FCS files

(represented in black dashed box) and annotates whether an FCS file is a normal

sample (blue) or a standard control sample (orange). A sample manifest is used in

batch mode to provide sample metadata. Additionally, a bead array configuration file

and a peak finding parameter file contain parameters that describe association of beads

with proteins and settings applied to each file. After the program is run in batch mode,

there are modules for bead gate quality assessment, construction of standard curves

and back-calculation to concentration units. When running for an individual file, figures

can be generated for diagnostic purposes (right).

beads. Next, forward-scatter height vs. area plots for events in each bead size gate are

evaluated using density-based clustering to identify outliers representing non-singlets that

are to be excluded (Fig 4b). Singlets from this step are then used for peak finding (for

each bead size gate) based on transformed MFI of the specified bead channel specified

in the manifest header and compared against the expected number of peaks as indicated

in the array configuration file (Fig 4c). Finally, events from each bead gate are evaluated

Figure 5. Raw FCS data processing steps.

All intensity data is hyperbolic arcsin transformed. A. Forward scatter vs. side scatter to

identify densities of events associated with multiple bead sizes. Red and blue ovals are

gates for 4um and 5um beads, respectively. B. Density-based clustering on forward

scatter height vs. area to identify outliers, or non-singlets for each bead size gate. C.

Identification of peaks on the specified bead channel for each bead size after non-

singlets filtered out. Line segments above each peak are individual bead gates. D.

Hyperbolic arcsine MFI intensity plots on the secondary isotype channels (columns of

plots) for each bead size/peak gate (rows of plots). For each row, number of beads per

gate indicated in parentheses. Antigen features for this example panel are: C = SARS-

CoV-2; S = SARS-CoV-1; M = MERS; O = OC43; H = HKU-1; CA07 HA = California 07

influenza hemagglutinin; Sp = spike protein; RBD = receptor binding domain; NP =

nuclear protein; NTD = n-terminal domain.

____

on the secondary isotype flow channel(s) specified in the manifest header to quantify the

location of the peak (in transformed MFI) for each bead feature and isotype (Fig 4d).

Details about program use and file formats can be found in the supplemental materials.

Comparison of MFIs computed with the automated software to manual processing using

FlowJo (BD Biosciences, Franklin Lakes NJ, USA) led to similar results, as shown in

Figure S1 for four representative samples.

3.7 Polyvalency and subunit corrections:

When we began to analyze serum samples and polyclonal Ab preparations, we noted

that the highest MFI values observed often eclipsed that recorded using antigen specific

monoclonal Abs, suggesting that antigen-monoclonal Abs standards would result in an

over extrapolation of the total antigen specific Ab in these samples, and that the

cytometric array is sensitive to effects resulting from the polyvalency of Abs present in

sera or other polyclonal Ab preparations. We hypothesized that multiple epitopes present

on antigens were simultaneously bound resulting in higher maximum MFI values than

exhibited by an equivalent concentration of a mAb, a phenomenon that would influence

Ab quantification. To determine if this were the case, and to quantitate this polyvalency

effect, we assayed a polyclonal C SP reactive Ab preparation derived from the serum of

a vaccinated donor. Using antigen specific capture beads to quantitate the Ab isotype in

this preparation, we determine it was composed of 90 percent IgG, 10 percent IgA, with

low but detectable levels of IgM. Because IgA was present as a significant portion C SP

reactive Ig, we further purified IgG using protein G chromatography yielding an anti-C SP

pIgG with a mixture of IgG subclasses. We then evaluated both the pAb and IgG anti-C

SP preparations in the CBA. As can be seen in figure 5a, the MFI of the IgG detection

Abs was noticeably but marginally higher in the IgG fraction resulting from the presence

of IgM and IgA Abs in the Ig preparation. We then used the array to quantitate C SP Abs

and compared these calculated values to the known concentrations. Although the

calculated values were higher than the known concentrations, they exhibited a linear

relationship up to concentrations of approximately 0.75 ug/mL, a concentration in which

the prozone effects of anti-SP Abs become apparent (fig 5c). By comparing the known

IgG concentration with the calculated values along the serial dilutions within the linear

range we could determine the increase in apparent concentration resulting from the

polyclonality of the purified anti-C SP Ab, which was an apparent 3.16 fold inflation of the

calculated concentration (fig 5b). These data demonstrate the average polyvalency effect

within the assay is constant within the range of concentrations indicated and provided a

correction factor for extrapolated Ig concentration values. To further confirm this

normalization was accurate, we assayed the NIBSC anti-C SP standards at serial

dilutions ranging from 1/1e3 to 1/8.65e4. When the resulting anti-C SP values are

corrected and plotted against the known dilutions of the pAb anti-SP they collapse to near

identity, confirming the accuracy of this correction (fig 5c). Using this approach, we

calculate the total anti-SP IgG to be 221.7ug/mL and 90.4ug/mL in the NISBC standards

2-/162 and 21/324 respectively (Fig 5D).

In addition to the polyvalency effects described above, we noted that the MFI values

were higher for the CS RBD particles than for full length SP trimers. This effect was

manifest as exceptionally high extrapolated anti-CS RBD values in some samples.

Comparing the MFI of the CS RBD and C SP beads resulting from staining with dilutions

of four RBD reactive mAbs, we consistently observed higher values for the RBD particles

(fig 5e). As these Abs recognize an identical epitope on both antigens and the bead

constituents exhibit identical biotin binding capacity, we reasoned that because the RBD

is expressed as a monomer and has variable accessibility in the full-length trimers (RBD

up-down configurations [21]) the epitope density is unequal between these configurations.

To determine if this was a consistent effect, we analyzed the fold difference in MFI

between SP and RBD beads at a range of concentrations of four anti-RBD Abs (Fig 5f).

Although the CBA signal derived from mAbs is dependent on Ab affinity in addition to

concentration, three of these Abs exhibit a concentration independent 1.5 fold increase

in   RBD   MFI   compared   to   that   of

504   C SP

Figure 6. Reproducibility of the CBA. **A.** Scatter plot of the ArcSin MFI of each bead for

all samples for two independent measurements of the CCHI U19 cohort serum. The

values of the two replicates are plotted pairwise. **B**. Extrapolated antibody concentration

for the two replicates ploted pairwise for the indicated HA antigens. **C**. Scatter plot of the

derived concentration of Ab reactive with CA09 HA derived from the COVID and INF

arrays. **D**. Mean absolute deviation for each data point of two replicates across the two

replicates for the indicated antigens and Ab isotypes.

trimers, allowing us to correct for differences in RBD epitope availability between full

length SP and monomers.

3.8 Sensitivity, accuracy, and reproducibility.

Although the array results are highly reproducible when assaying monoclonal antigen-specific Abs or isotype specific Ab standards, we sought to determine the variance of array across repeated measurements. To quantitate the reproducibility of the array derived Ab values, we analyzed serum available through the UAB Comprehensive Center for Human Immunology (CCHI) donor repository twice, independently, with COVID and INF arrays. The MFI of each antigen in the arrays was nearly identical in the between the two replicates (fig 6a). Because acquisition for most of these samples pre-dated the COVID19 pandemic, they did not contain detectable SARS-CoV-2 antigen reactivity (fig S3). However, IgG reactivity toward one or more HA antigens was observed in all samples. The calculated value of the most prominently targeted HA antigens (including both H1 and H3) were in close agreement between the replicates (fig 6b). Although deviations between replicates were noted at the highest levels detected, these fall outside of the linear range of the assay. Within the range of confidence, we observed a high level of agreement between the two measurements. Importantly, as shown in Figure 6c, the calculated values of HA CA09, which was included in the COVID array as a positive control, were nearly identical in both arrays. As these reagents were generated independently using different lots of SA beads, this demonstrates the reproducibility of both bead generation and assay measurements. To quantitate the reproducibility of the assay, we examined the mean absolute deviation of IgG and IgA reactivity toward the three most frequently targeted HA H1 and H3 antigens (fig 6d). The average deviation is less than 5% across all these antigens, and slightly lower for IgG (2.5%) than for IgA (4%). Collectively these data suggest the arrays are highly reproducible both at the level of bead generation and in their performance.

We did not observe Ab binding to C SP coated beads in serum samples collected before the spring of 2021, despite IgG reactive with O SP being readily quantifiable in all samples (serum antigen specific Ab concentrations for all antigens is shown in figure 7). Following the spring of 2021, C SP reactivity was observed in 13/19 samples whereas C NP reactivity was only observed in 4. In every case, C NP reactivity was corelated with C SP Ab, suggestive of a prior infection (fig S3). These data are in line with the infection

and vaccination rates within the general population from which these donors were derived

[22].

Although HA reactivity was observed in all samples, there were significant

differences in the HA reactivity and Ab levels observed across individuals, representative

array histograms shown in Figure S4. A bias toward either H1 or H3 HA was observed in

some individuals, while others exhibited a balanced anti-HA Ab repertoire, although the

recent vaccine formulations were apparent, with CA09, Brisbane (H1) and MI15

demonstrated the highest level of reactivity. Reactivity toward PR8 was observed to be

the lowest of H1 HA (fig 7). Reactivity toward H3 were generally lower than that of H1

while reactivity toward the H2, H5, H7 and H9 included in the array were rarely detected.

Most individuals exhibited IgG and IgM Abs toward INF NP while only 10% harbored

serum reactivity toward INF NS1, perhaps indicative of a recent INF infection. When we

compared the reactivity of these antigens, we noted a significant correlation between H1

and H3 HA, which were poorly correlated with one another (fig 8a). Although we cannot

determine if this is due to cross-reactive Abs, or that similar levels of Abs that target these

antigens, the sequence similarity, as is the case for CA09 and MI15, are associated with

the highest correlation of Ab levels (fig 8bc). Interestingly, although IgG and IgA Ab levels

are both highly corelated toward similar HA antigens, the IgG and IgA levels themselves

do not correlate (fig 8c), perhaps suggestive of different origins of HA reactive Abs of

these isotypes. Collectively these data demonstrate the assay is suitable for high-

dimensional, high-throughput Ab reactivity determination.

Figure 7. The reactivity profiles of 96 sera samples toward the full BCov and INF arrays. **A.** The concentration of IgM, IgA, and IgG reactive with the indicated antigen as measured by the Coba and INF array. **B.** Heatmap of the log10 Ab concentration (ug/mL) for the indicated antigen. Each row corresponds to an individual serum, and each column corresponds to the indicated antigen, IgM, IgA, and IgG are shown.

---

4. Conclusions

Here we describe an CBA platform designed to conduct high-throughput determination of Abs reactive with viral antigens. The assay provides a rich view of the humoral immune landscape by multiplexing the analysis of 12 B-CoV or 16 INF antigens into a single assays. By altering dilutions factors and secondary detection reagents, the levels of IgG, IgM, and IgA can be accurate determined. This platform exhibits high sensitivity, although the lower limit of detection is dependent on dilution factor, as little as 2ng/mL of antigen-specific IgG in the final dilution can be measured. The assay demonstrates high specificity, as demonstrated by multiple monoclonal Abs, the reactivity to each antigen in the array is discrete. The use of multiple indirect standards provides accurate quantification for Ab reactive to each antigen across all assayed isotypes. Additionally, the assay requires minute amounts of both antigen and serum tested, typically consuming less than 0.5ug of antigen and less than 1uL of sample per assay.

Although, the CBA array described here is focused on B-CoV and INF serology, the same strategy can be implemented to evaluate serum Ab to a wide array of antigenic targets.

Figure 8. Correlation of Ab levels for HA antigens in the array**. A.** The pairwise Pearson correlation coefficient for either IgG (lower left) or IgA (upper right) of Ab levels for each antigen in the INF Array. **B.** Scatter profiles of antibody levels toward the indicated H1 and H3 hemagglutinins for the 59 samples in the cohort.

The use of recombinant antigens, like the various SP ectodomains and subunits or HA and NP described here, allow for the rapid replacement or addition of similar antigens into the CBA with no required modification to the analytics. Although the corrections for polyvalency and subunit epitope availability are specific for C SP, these are the product of the physical nature of the antigens included in the current array, and whether such corrections are applicable to other antigens necessitates empirical determination. Polyclonal corrections may be generalizable, however, as the number of individual Abs that may simultaneously bind their cognate antigen is influenced by immunodominance of epitopes and steric hindrance, an idea supported by structural studies [23-25].

The high-dimensional assay we described greatly benefits from the use of the automated analytic platform. When combined, this system facilitates the rapid and accurate determination of antigen reactive Ab. Although the source material used as examples in this study are serum, the assay is amenable to many biological samples including nasal washes and saliva. The capacity to visualize both the magnitude and breadth of Ab responses to vaccination or infection with respiratory pathogens is valuable in determining correlates of protection derived from vaccination or pathogen exposure, and the ability to easily and rapidly expand the antigens incorporated into the array with minimal effects on the established analytics makes this approach an attractive option to fill the continuing need for large-scale serology studies.

1. Garcia-Beltran, W.F., et al., *Multiple SARS-CoV-2 variants escape neutralization by vaccine-induced humoral immunity.* Cell, 2021. **184**(9): p. 2372-2383 e9.
2. Yadav, P.D. and S. Kumar, *Global emergence of SARS-CoV-2 variants: new foresight needed for improved vaccine efficacy.* Lancet Infect Dis, 2022. **22**(3): p. 298-299.
3. Meyer, B., C. Drosten, and M.A. Muller, *Serological assays for emerging coronaviruses: challenges and pitfalls.* Virus Res, 2014. **194**: p. 175-83.
4. Pisanic, N., et al., *COVID-19 Serology at Population Scale: SARS-CoV-2-Specific Antibody Responses in Saliva.* J Clin Microbiol, 2020. **59**(1).
5. Caceres-Martell, Y., et al., *Single-reaction multi-antigen serological test for comprehensive evaluation of SARS-CoV-2 patients by flow cytometry.* Eur J Immunol, 2021. **51**(11): p. 2633-2640.
6. Cameron, A., et al., *A Multiplex Microsphere IgG Assay for SARS-CoV-2 Using ACE2-Mediated Inhibition as a Surrogate for Neutralization.* J Clin Microbiol, 2021. **59**(2).
7. Dawson, E.D., et al., *Multiplexed, microscale, microarray-based serological assay for antibodies against all human-relevant coronaviruses.* J Virol Methods, 2021. **291**: p. 114111.
8. Egia-Mendikute, L., et al., *Sensitive detection of SARS-CoV-2 seroconversion by flow cytometry reveals the presence of nucleoprotein-reactive antibodies in unexposed individuals.* Commun Biol, 2021. **4**(1): p. 486.
9. Guarino, C., et al., *Development of a quantitative COVID-19 multiplex assay and its use for serological surveillance in a low SARS-CoV-2 incidence community.* PLoS One, 2022. **17**(1): p. e0262868.
10. Jolley, M.E., et al., *Particle concentration fluorescence immunoassay (PCFIA): a new, rapid immunoassay technique with high sensitivity.* J Immunol Methods, 1984. **67**(1): p. 21-35.
11. Kenny, G., et al., *Performance and validation of an adaptable multiplex assay for detection of serologic response to SARS-CoV-2 infection or vaccination.* J Immunol Methods, 2022: p. 113345.
12. Roy, D.R., et al., *Development, Validation, and Utilization of a Luminex-Based SARS-CoV-2 Multiplex Serology Assay.* Microbiol Spectr, 2023: p. e0389822.
13. Morgan, E., et al., *Cytometric bead array: a multiplexed assay platform with applications in various areas of biology.* Clin Immunol, 2004. **110**(3): p. 252-66.
14. Allie, S.R., et al., *The establishment of resident memory B cells in the lung requires local antigen encounter.* Nat Immunol, 2019. **20**(1): p. 97-108.
15. Beckett, D., E. Kovaleva, and P.J. Schatz, *A minimal peptide substrate in biotin holoenzyme synthetase-catalyzed biotinylation.* Protein Sci, 1999. **8**(4): p. 921-9.
16. Nellore, A., et al., *A transcriptionally distinct subset of influenza-specific effector memory B cells predicts long-lived antibody responses to vaccination in humans.* Immunity, 2023.
17. fca, B.L. *readfcs.m* balkay.laszlo@med.unideb.hu, Department of Medical Imaging, University of Debrecen. Version 22/Jun/2020. Matlab File Exchange. 2020; Available from: https://www.mathworks.com/matlabcentral/fileexchange/9608-fca_readfcs.
18. Finak, G., et al., *Optimizing transformations for automated, high throughput analysis of flow cytometry data.* Bmc Bioinformatics, 2010. **11**.

19. G., C. *L4P.m, Four parameters logistic regression - There and back again (2012).* Matlab
File   Exchange.   2012;   Available   from:
https://it.mathworks.com/matlabcentral/fileexchange/38122.
20. R., H. *Flow Cytometry Data Reader and Visualization* 2021; Available from:
https://www.mathworks.com/matlabcentral/fileexchange/8430-flow-cytometry-data-
reader-and-visualization.
21. Henderson, R., et al., *Controlling the SARS-CoV-2 spike glycoprotein conformation.* Nat
Struct Mol Biol, 2020. **27**(10): p. 925-933.
22. Health, A.D.o.P. *Coronavirus Disease 2019 (COVID-19).* 2023; Available from:
https://www.alabamapublichealth.gov/covid19/.
23. Barnes, C.O., et al., *Structures of Human Antibodies Bound to SARS-CoV-2 Spike Reveal*
Common Epitopes and Recurrent Features of Antibodies. Cell, 2020. **182**(4): p. 828-842
e16.
24. Nogal, B., et al., *Mapping Polyclonal Antibody Responses in Non-human Primates*
Vaccinated with HIV Env Trimer Subunit Vaccines. Cell Rep, 2020. **30**(11): p. 3755-3765
e7.
25. Han, J., et al., *Polyclonal epitope mapping reveals temporal dynamics and diversity of*
human antibody responses to H5N1 vaccination. Cell Rep, 2021. **34**(4): p. 108682.
680