Hydrogel Capsule-Based Digital Quantitative Polymerase Chain Reaction

Zheng Lin Tan[1, 2, *], Masato Yasuura[1, 3], Yukichi Horiguchi[1, 4], Hiroki Ashiba[1, 5], Takashi Fukuda[1, 6, *]

[1] Sensing System Research Center, National Institute of Advanced Industrial Science and Technology, Central 5, 1-1-1 Higashi, Tsukuba, Ibaraki, 305-8565, Japan

[2] ORCID: 0000-0001-6447-3788

[3] ORCID: 0000-0003-2334-8088

[4] ORCID: 0000-0002-6748-9643

[5] ORCID: 0000-0002-8764-8357

[6] ORCID: 0000-0001-9599-6274

* Corresponding to Takashi Fukuda (t-fukuda@aist.go.jp) and Zheng Lin Tan (tan.zhenglin@aist.go.jp)

Abstract

Droplet digital PCR (ddPCR) is accurate in nucleic acid quantification owing to its linearity and high sensitivity. Amplification of nucleic acid in droplets, however, is limited by the stability of droplets against thermal cycling. While the use of fluorinated oil or supplementation of surfactant could improve the stability of droplets, this process has also greatly increased the cost of ddPCR and limited post-PCR analysis. Here, we report a novel method known as gel capsule-based digital PCR (gc-dPCR) which includes method to prepare hydrogel capsules encapsulating PCR reaction mix, conducting PCR reaction, and readout by either qPCR system or fluorescence microplate reader. We have compared our method to existing methods, i.e., quantitative PCR or vortex ddPCR. Our approach results in higher fluorescence intensity compared to ddPCR suggesting higher sensitivity of the system. As hydrogel capsules are more stable than droplets in fluorinated oil throughout thermal cycling, all partitions can be quantified, thus preventing loss of information from low-concentration samples. Our approach should extend to all droplet-based PCR methods. Generally, our approach has greatly improved ddPCR by increasing droplets stability and sensitivity, and reducing the cost of ddPCR, which help to remove the barrier of ddPCR in settings with limited resources.

Keywords: hydrogel; droplet digital polymerase chain reaction; nucleic acid quantification; gel capsule digital polymerase chain reaction

Competing Interests

The authors declare no competing interests.

Technologies for nucleic acid quantification are important for science, industry, public health, and medical diagnosis.

Quantitative polymerase chain reaction (qPCR) has emerged as the gold standard for its wide dynamic range and high

specificity depends on probe[1, 2]. However, cumbersome optimization process is required for qPCR to obtain reliable and

reproducible result[3]. In contrast, digital PCR (dPCR) based on limiting dilution of sample, such that each part of sample

contains only one or no target molecule, detect sample after end-point amplification, enumerate number of positive parts of

sample and yield concentration of target by statistics[4–6]. dPCR possess numerous advantages compared to qPCR, including

absolute quantification without the needs of calibration curve, enhanced accuracy for small concentration (samples with

quantification cycle, $C_q \geq 29$)[7, 8], increased resistance against reaction inhibition[9], and ability to differentiate fragmented

from intact molecules[10, 11], which make dPCR a valuable tool for clinical applications.

Droplet dPCR (ddPCR) is a variation of dPCR which rely on active emulsification method to compartmentalize samples into

small volume of monodisperse droplets in oil[12, 13]. While ddPCR is exhibit an improved accuracy compared to other PCR

approaches, requirement of equipment for active emulsification increases the cost of PCR and the barrier of adoption, which

has limited the widespread of this technology. Furthermore, concerns remained for the stability of droplets throughout multiple

thermal cycles required for nucleic acid amplification. While previous study has suggested that interaction of sodium bis(2-

ethylhexyl) sulfosuccinate or polyglycerol polyricinoleate with bovine serum albumin at the interface of droplets could

stabilize droplet during thermal cycle[14], use of proteins is not favorable, as autofluorescence from proteins might affect

sensitivity of ddPCR system. To address this issue, fluorinated oil and fluorinated surfactant were used to reduce droplet

coalescence during thermal cycling. Nevertheless, fluorinated surfactant does not completely prevent coalescence of droplets,

i.e., fluorinated oils have densities between $1.4 – 1.9$ g cm^{-3}, e.g., FluoOil series by Emulseo has density ranging from $1.61 –$

1.85 g cm^{-3} [15], which results in the floating of droplets on oil, and can cause aggregation and coalescence. Furthermore,

global supply of fluorinated surfactant and fluorinated oil is limited, and the cost is high. For instance, 10 ml 2% fluorinated

surfactant (FluoSurf by Emulseo) in fluorinated oil (Novec HFE 7500 by 3M) cost US$ 240.00 [16] (12th January 2023).

Furthermore, as the world leading manufacturer of fluorinated oil, has announced to exit production of fluorinated oil by 2025

due to environmental pollution[17], there is an urgent need to develop a novel method to replace the use of fluorinated oil for

droplet stabilization. On the other hand, irreversible interfacial stability provided by fluorinated oil has limited the downstream

application of droplets, e.g., next generation sequencing due to complexity in nucleic acid recovery. To further improve the

stability of droplet, to reduce the cost of ddPCR, and to expand potential application of ddPCR by enabling post-ddPCR

droplet manipulation, thus yield its benefit for wide range of applications, a new approach for fluorinated oil-independent

droplets stabilization is required.

In this study, we demonstrate droplet stabilization with hydrogel. Rather than stabilization of phase interface, reversible

solidification of reactor will provide higher mechanical and thermal stability to droplets. We mix the nucleic acid and PCR

reagent with sodium alginate and crosslink sodium alginate gel with group 2 metal ion. We have selected barium as crosslinker

against calcium. Whilst calcium has been widely used to crosslink alginate hydrogel, it is known to be a polymerase inhibitor.

Compression of hydrogel capsules against 2 pieces of glass slides suggested that the hydrogel capsules is stable against

mechanical forces. With this stability, we could generate and maintain droplet in soybean oil or mineral oil, which the price

is $< 10^{-3} – 10^{-4}$-fold of fluorinated oil with fluorinated surfactant, even after considering the price of alginate salt. We

demonstrate this method by quantifying influenza A virus (IAV) nucleic acid in hydrogel capsules by vortex ddPCR

(vddPCR)[18]. While we use vddPCR as an example, our method should apply to any dPCR approach in which the reactors

can be sealed, polymerization can be easily reversed, and the samples can be easily recovered.

Experimental

Preparation of barium alginate hydrogel capsule

Barium chloride (Wako Pure Chemical, Cat # 023-00195, Osaka, Japan) was dissolved in soybean oil (Wako Pure Chemical, Cat # 190-03776, Osaka, Japan) at final concentration of 20 mmol L^{-1}. While stirring, dropped 2% (w/v) sodium alginate (Wako Pure Chemical, Cat # 194-13321, Osaka, Japan) with or without red dye into soybean oil with barium chloride. The mixture was stirred for 60 s and observed with an optical microscope. To conduct qualitative analysis of compression of hydrogel capsule, hydrogel capsules were sandwiched between 2 pieces of glass slides. Then, stress was applied at the end of glass slide with a pair of clamps. The mechanical stability of hydrogel capsules was evaluated by observation of compression and leakage of content from hydrogel capsules with microscope when the diameter of hydrogel capsules increased by at least 10%.

qPCR and Bulk Readouts of ddPCR

Hydrolysis probe based on fluorescein and carboxytetramethylrhodamine was used in qPCR. The TaqMan Assays were rigorously designed, and the specificity of these assays were determined by the vendor (Thermofisher Scientific). The reaction mixtures were prepared by mixing 4× TaqMan Fast Virus 1-Setp Master Mix (Thermofisher Scientific, Ref # 4444427, Vilnus, Lithuania), 10^7 copies mL^{-1} influenza A virus standard DNA, primers and probes for matrix protein of IAV as in instruction. For qPCR and vddPCR assays, distilled deionized water (Nippon Gene, Cat # 318-90105, Toyama, Japan) was used as diluent, while 2% (w/v) sodium alginate solution in distilled deionized water (1% (w/v) final concentration) was used as diluent for gc-dPCR assays. For each assay, 20 μL reaction mix was used.

To generate droplets for ddPCR, reaction mix were vortex in 15 μL Automated Droplet Generation Oil for Probes (Bio-Rad, Cat # 1864110, USA) for 5 min, followed by addition of 15 μL Automated Droplet Generation Oil for Probes with barium chloride and vortex for another 5 min. For qPCR assays, the reaction mix were covered with 30 μL Automated Droplet Generation Oil for Probes with barium chloride to prevent measurement bias. These assays were incubated in qPCR system (Roche, LightCycler 96, Switzerland) as described in manufacturer's instruction of 4× TaqMan Fast Virus 1-Setp Master Mix, and the C_q were recorded.

After incubation, the assays were transferred into a black, flat-bottom 96-well plate, and the fluorescence intensity with excitation/emission wavelength at 488/530 nm was measured with fluorescence microplate reader (Thermo Scientific, VarioSkan Lux, Singapore). For each sample, 6 replicates were prepared with at least 2 different aliquots of reagents to ensure reproducibility.

Comparison of gc-dPCR in Fluorinated Oil and Soybean Oil

Hydrogel capsules were prepared with either fluorinated oil or soybean oil as described in previous section. Then, the hydrogel capsules were incubated in thermal cycler. After incubation, the fluorescence intensity with excitation/emission wavelength at 488/530 nm was measured with fluorescence microplate reader. The fluorescence intensity of equal volume of fluorinated oil and soybean oil were also measured to evaluate autofluorescence. After measurement, the hydrogel capsule from each condition was observed with fluorescence microscope (Olympus, BX3-RFAS, China; objective lens: Olympus, UPFLN10X2, Tokyo, Japan; fluorescence mirror unit: Evident Corporation, U-FBW, Nagano, Japan; illuminator: Thorlabs, M455F1, New Jersey, USA; complementary metal-oxide semiconductor camera: Hamamatsu Photonics, C13440-20CU, Hamamatsu, Japan). For each sample, 11 replicates were prepared with at least 2 different aliquots of reagents. Two measurements were conducted on different day with 6 replicates on the one day and 5 replicates on another day to ensure reproducibility.

Detection of Influenza A Virus RNA with gc-dPCR

To ensure that the proposed method is applicable in virus detection, nucleic acid was extracted from sample with unknown

concentration of influenza A virus (A/Puerto Rico/8/1934(H1N1)). The nucleic acid extract was diluted by 3000-fold.

Reaction mixtures were prepared as described in previous section, except influenza A virus standard DNA was replaced with

nucleic acid extract. Mixture without sample and mixture without primers were used as control. Hydrogel capsules was

generated with soybean oil as described in previous section. RNA was amplified as manufacturer's instruction for 45 cycles.

Both C_q and endpoint fluorescence intensity were measured. For each sample, 3 independent tests were conducted.

Results and Discussion

ddPCR is more sensitive than qPCR due to its nature of linearity, which allow it to accurately quantify nucleic acid,

particularly at low copy number. However, high temperature will decrease viscosity of oil and increase kinetic energy of

droplets, which results in increased probability of droplet coalescence. Condensation of droplets increases surface charge,

which attract small droplets, and further increased the probability of coalescence. While it has been demonstrated that

fluorinated oil with fluorinated surfactant could increase stability of droplets throughout PCR, some properties, e.g., high

density, low surface tension, and high evaporation rates of fluorinated oil[19] will expose droplets to air, and result in

evaporation of droplets during PCR, in addition to its high cost. Furthermore, stability of fluorinated oil at low temperature

limits demulsification of droplets, thus limited further analysis of post-PCR products. By adding sodium alginate into reaction

mix, and polymerize droplets divalent ions[20], it is possible to stabilize droplets without fluorinated oil, while increasing

mechanical stability of droplets (Figure 1). In our study, barium was used instead of other divalent ions as these ions are PCR

inhibitors[21]. Furthermore, conventional calcium alginate is sensitive to the presence of phosphate, citrate, sodium, and

magnesium which will result in gel swelling followed by leakage of content.

Furthermore, we have also demonstrated that barium alginate is stable against mechanical stress. Compression has resulted in

increased in diameter of hydrogel capsules from 6.65 μm to 7.38 μm. Despite of the deformation, no leakage of dye from

hydrogel capsule was observed (Figure 1b, c), which suggested that hydrogel capsule is stable against mechanical stress.

Moreover, a layer of shell-like structure was observed under high magnification when the hydrogel capsule was compressed,

which suggested that surface polymerization might have occurred, and formed a barium alginate gel shell bead. Mørch et al.

has investigated the distribution of alginate after cross-linking with various types of type 2 metal ions. Their study suggested

that with low concentration of type 2 metal ions[22], i.e., ≤ 20 mmol L^{-1}, a rapid gelling could occur, which results in

accumulation of alginate at the outer layer. In our study, we have provided evident that a gel shell might have formed at the

outer layer of gel microdroplets based on the difference of color between core and outer layer, which stemmed from the

difference in optical properties of gel and sol, and their response toward mechanical stress. Formation of gel shell bead with

liquid core is more favorable than complete polymerization of alginate hydrogel as lower density of gel core could facilitate

rapid reaction compared to solid core. Furthermore, supplementation of chelating agent, i.e., ethylenediaminetetraacetic acid

to barium alginate gel could dissolved the gel shell (Figure S1) of gel microdroplets, which will enable post-analysis

manipulation of droplets.

Figure 1 Mechanical stability of barium alginate hydrogel capsules. (a) barium alginate hydrogel capsules generated with soybean oil. Black arrows show the positions of gel microdroplets. Mechanical stability of gel against mechanical stress was evaluated by micrographs of hydrogel capsules encapsulating red dye (b) before and (c) during compression. Expansion of droplets due to compression was compared with black dashed line. The scale bars are 10 μm.

Recently, Abate's group has demonstrated accurate quantitation of severe acute respiratory syndrome coronavirus 2 sequences with vddPCR[18]. The process is convenient as it does not required microfluidics, and it can be conducted entirely in bulk processing, which the size distribution of droplets is not an important parameter. Furthermore, the quantitative property of vddPCR was also demonstrated by Abate's group. In this study, we have compared our proposed method, gel capsule-based dPCR (gc-dPCR) to vddPCR. To ensure that all reactions were conducted under similar conditions, fluorinated oil with barium chloride was supplemented to all assays, either to cover the reaction mix or as continuous phase to disperse droplets. Prior study has described that dispersion of reaction mix, even with vortex emulsification could reduce C_q of qPCR[18]. Our result has confirmed the effect of compartmentalization in reducing C_q of qPCR (Figure 2a). Endpoint measurement with fluorescence microplate reader has shown significant increase in fluorescence intensity in vddPCR and gc-dPCR assay compared to qPCR assay, while no statistical significant was observed between vddPCR and gc-dPCR assay (Figure 2b). Considered both the C_q and fluorescence intensity, although supplementation of hydrogel into reaction mix has significantly increased C_q of qPCR, however, it did not increase the fluorescence intensity, suggesting that the difference in C_q is not practical significant. These results suggested that both vddPCR and gc-dPCR could increase sensitivity of qPCR.

Figure 2 Gellification of droplets retain the efficiency of vddPCR. (a) qPCR readout of vddPCR and gc-dPCR suggested that both vddPCR and gc-dPCR are more sensitive than qPCR. (b) Endpoint measurement of each assay. vddPCR and gc-dPCR produced stronger fluorescence intensity compared to qPCR. The boxplots represent statistical analysis of data obtained from 6 independent tests. Statistical significance was evaluated by one-way analysis of variance followed by post-hoc Tukey honest significance difference test. * $p < 0.05$, ** $p < 0.01$, *** $p < 0.001$, **** $p < 0.0001$. ns indicates not significant.

While it is generally regarded that higher viscosity of hydrogel (> 80 mPa s) compared to water (10 mPa s) will result in retarded reaction rate, in our study, PCR efficiency remained the same for both vddPCR and gc-dPCR as diffusion length of reactant in droplets is short enough to overcome the effect of viscosity. On the other hand, larger variability of C_q and fluorescence intensity were observed for gc-dPCR, which might be due to larger size distribution of hydrogel capsules resulted from higher viscosity of sodium alginate sol compared to water. Nevertheless, higher viscosity of sodium alginate sol will result in generation of microdroplets with larger size, which will result in production of more PCR products per target molecule, thus increase the sensitivity of system. Higher sensitivity and signal intensity from microdroplets with larger size has compensated the viscosity of sodium alginate sol, which results in similar efficiency between vddPCR and gc-dPCR. Another advantage of gc-dPCR is its capability to maintain its stability without the needs of fluorinated oil or surfactant. However, it remained uncertain whether hydrogel capsules generated with soybean oil without surfactant can be used in ddPCR. To evaluate the effect of oil on gc-dPCR, endpoint measurement PCR products after thermal cycling of hydrogel capsules generated with fluorinated oil and soybean oil were conducted (Figure 3).

Figure 3 gc-dPCR in soybean oil. (a) The effect of oil on gc-dPCR. The boxplots represent statistical analysis of data obtained from 11 independent tests. Statistical significance was evaluated by Student's t-test. * $p < 0.05$, ** $p < 0.01$. (b) Micrographs of hydrogel capsules after gc-dPCR. The hydrogel capsules from each condition was observed under fluorescence filter and merged image of brightfield and fluorescence micrographs. Green in merge channel was signal originated from fluorescein in fluorescence channel. The scale bars are 100 μm.

Although the fluorescence intensity of PCR products encapsulated in hydrogel capsules generated with soybean oil was higher than those encapsulated in hydrogel capsules generated with fluorinated oil, the variance was also larger (Figure 3a). The difference of variance of fluorescence intensity might be resulted from larger droplet size dispersity for droplets generated with soybean oil compared to fluorinated oil (Figure 3b). Previous studies have experimentally shown that viscosity of continuous phase (oil) is a parameter of droplet size and its distribution[23, 24]. Higher viscosity of continuous phase will result in larger size dispersity of droplets. Consider the viscosity of soybean oil at 25°C is 50.09 mPa s, which is 12.5 – 41.7-fold higher than fluorinated oil (1.2 – 4 mPa s), thus, hydrogel capsules generated with soybean oil will have larger dispersity compared to hydrogel capsules generated with fluorinated oil. The variance can be reduced by generating droplets with active emulsification methods which produce higher monodispersity of droplets.

On the other hand, the density of soybean oil (0.917 g cm^{-3}) approximates the density of water and lower than fluorinated oil (1.4 – 1.6 g cm^{-3}). While barium alginate hydrogel capsules float on fluorinated oil, they distributed evenly in soybean oil (Figure S2), which results in efficient amplification of nucleic acid due to even heating of hydrogel capsules. Efficient amplification results in higher signal intensity, hence, the fluorescence intensity of PCR products from hydrogel capsules generated with soybean oil is higher than those generated with fluorinated oil.

In addition, thermal stability of hydrogel capsules was also demonstrated. Fluorophore remained encapsulated in capsule after PCR thermal cycling (Figure 3b), suggesting that no breakage or merger of capsule which will resulted in leakage of fluorophore has occurred. As the outer layer of hydrogel capsule is gellified, and the onset decompose temperature of barium alginate is 150°C[25], barium alginate capsule is stable within the range of temperature for thermal cycling.

Table 1 Autofluorescence of oils. Statistical analysis of fluorescence intensity of fluorinated oil and soybean oil from 3 independent tests.

	Mean	Median	Standard deviation
Fluorinated oil	0.16	0.15	0.014
Soybean oil	0.29	0.27	0.048

Furthermore, the autofluorescence of fluorinated oil and soybean oil was confirmed (Table 1). Autofluorescence of soybean oil originated from fatty acid chain is a concern in fluorescence measurement. Although the autofluorescence intensity of soybean oil is higher than fluorinated oil by 2-fold, the autofluorescence is negligible compared to the fluorescence intensity generated from PCR products (Figure 3a). As fluorescent tag binds to nucleic acid, and the phosphate backbone of nucleic acid which expose to the external environment is hydrophilic, leakage of fluorescent tag into oil is negligible. These results suggested that higher fluorescence intensity from hydrogel capsules generated with soybean oil was the result of even distribution of droplets in oil.

Table 2 Detection of RNA of influenza A virus from actual sample. Each set of data is obtained from 3 independent tests. S.D. indicates standard deviation and N.D. indicates non-detectable.

Sample	C_q			Fluorescence intensity		
	Mean	Median	S.D.	Mean	Median	S.D.
Control without target RNA	N.D.	N.D.	N.D.	0.22	0.16	0.13
Control without primers	N.D.	N.D.	N.D.	0.10	0.10	0.00
Sample containing influenza A virus	19.10	19.05	0.18	28.64	30.92	5.90

To demonstrate the applicability of our method on actual sample, gc-dPCR was performed on actual sample containing unknown concentration of influenza A virus (Table 2). Increase in fluorescence intensity and decrease in C_q suggested that influenza A virus was detected, and our method is applicable in actual sample.

Conclusions

Our approach is a major advance because it has greatly enhanced the stability of droplets, reduced the running cost of ddPCR by approximately 4000-fold through replacement of fluorinated oil with soybean oil and sodium alginate, and provided the option for post-PCR droplet manipulation. This method is compatible with various droplet generation approaches, e.g., microfluidic-based droplets generation, membrane emulsification and droplet generation in turbulent flow. While hydrogel increases the stability of microdroplets, viscosity of sol could also increase the pressure on membrane when membrane with small pore size was used in emulsification. For the case when smaller capsule size and monodisperse gel capsule is required to absolute quantification by positive capsule quantification, microfluidic device[26, 27] or low viscosity sol can be used. The stability provided by hydrogel capsules afford other valuable benefits, including the ability to alter the composition of continuous phase to ensure even distribution of droplets and heat conduction, reduce scattering at the edge of droplets, and post-quantification analysis of microdroplets of interest, e.g., isolating and sequencing of nucleic acid encapsulated in droplets. Furthermore, owing to the benefit of high mechanical stability of gel microdroplets, high-throughput fluorescence-activated cell sorter or fluorescence-activated droplet sorter can be used to analyze and isolate droplets[26, 27] of interest at throughput of > 500 kHz[28]. These features are particularly useful in clinical diagnosis and to investigate mutation in sequence. Moreover, even distribution of droplets in oil could ensure even heat transfer to all droplets, which results in even reaction. Even amplification reaction in turn increases the yield of amplicon (fluorescence intensity), which can help to reduce false

positive test results by setting a higher threshold value. Hydrogel capsules have demonstrated high mechanical and thermal

resistance without surfactant, which could greatly reduce the cost of ddPCR, and helps to accelerate employment of ddPCR

method even in settings with limited resources.

Acknowledgement

Not applicable

Author Contributions

Conceptualization: TF and ZLT; Experimental design and data curation: ZLT, TF, MY; Data analysis: ZLT; Interpretation of

results: ZLT, MY, TF, YH, HA; Funding and administration: TF, MY; Writing – Original draft: ZLT, MY, TF; Writing –

Review and editing: HA, YH, TF, MY, ZLT

Reference

1.   Holland PM, Abramson RD, Watson R, Gelfand DH (1991) Detection of specific polymerase chain reaction product
by utilizing the 5'----3' exonuclease activity of Thermus aquaticus DNA polymerase. Proc Natl Acad Sci USA 88:7276–
7280. https://doi.org/10.1073/pnas.88.16.7276

2.   Higuchi R, Dollinger G, Walsh PS, Griffith R (1992) Simultaneous Amplification and Detection of Specific DNA
Sequences. Nat Biotechnol 10:413–417. https://doi.org/10.1038/nbt0492-413

3.   Bustin SA, Benes V, Garson J, et al (2013) The need for transparency and good practices in the qPCR literature. Nat
Methods 10:1063–1067. https://doi.org/10.1038/nmeth.2697

4.   Simmonds P, Balfe P, Peutherer JF, et al (1990) Human immunodeficiency virus-infected individuals contain provirus
in small numbers of peripheral mononuclear cells and at low copy numbers. J Virol 64:864–872.
https://doi.org/10.1128/jvi.64.2.864-872.1990

5.   Ruano G, Kidd KK, Stephens JC (1990) Haplotype of multiple polymorphisms resolved by enzymatic amplification of
single DNA molecules. Proc Natl Acad Sci USA 87:6296–6300. https://doi.org/10.1073/pnas.87.16.6296

6.   Vogelstein B, Kinzler KW (1999) Digital PCR. Proc Natl Acad Sci USA 96:9236–9241.
https://doi.org/10.1073/pnas.96.16.9236

7.   Pohl G, Shih I-M (2004) Principle and applications of digital PCR. Expert Review of Molecular Diagnostics 4:41–47.
https://doi.org/10.1586/14737159.4.1.41

8.   Taylor SC, Laperriere G, Germain H (2017) Droplet Digital PCR versus qPCR for gene expression analysis with low
abundant targets: from variable nonsense to publication quality data. Sci Rep 7:2409. https://doi.org/10.1038/s41598-
017-02217-x

9.   Hindson CM, Chevillet JR, Briggs HA, et al (2013) Absolute quantification by droplet digital PCR versus analog real-
time PCR. Nat Methods 10:1003–1005. https://doi.org/10.1038/nmeth.2633

10. Han J, Lee JY, Bae Y-K (2019) Application of digital PCR for assessing DNA fragmentation in cytotoxicity response.
Biochimica et Biophysica Acta (BBA) - General Subjects 1863:1235–1242.
https://doi.org/10.1016/j.bbagen.2019.05.001

11. Pavšič J, Žel J, Milavec M (2016) Digital PCR for direct quantification of viruses without DNA extraction. Anal Bioanal
Chem 408:67–75. https://doi.org/10.1007/s00216-015-9109-0

12. Hindson BJ, Ness KD, Masquelier DA, et al (2011) High-Throughput Droplet Digital PCR System for Absolute
Quantitation of DNA Copy Number. Anal Chem 83:8604–8610. https://doi.org/10.1021/ac202028g

13.   Xu P, Zheng X, Tao Y, Du W (2016) Cross-Interface Emulsification for Generating Size-Tunable Droplets. Anal Chem
292   88:3171–3177. https://doi.org/10.1021/acs.analchem.5b04510

293   14.   Romero-Peña M, Ng EK, Ghosh S (2021) Development of thermally stable coarse water-in-oil emulsions as potential
294   DNA bioreactors. J Dispers Sci Technol 42:2075–2084. https://doi.org/10.1080/01932691.2020.1794886

295   15.   Darwin   Microfluidics   (2023)   Emulseo   Fluo-Oil   Fluorinated   Oil   for   Microfluidics.   https://darwin-
296   microfluidics.com/products/fluorinated-oil-for-microfluidics?variant=36766353850532. Accessed 6 Mar 2023

297   16.   Dolomite (2019) FluoSurf. https://www.dolomite-microfluidics.com/product/fluosurf/. Accessed 12 Jan 2023

298   17.   3M (2022) 3M to Exit PFAS Manufacturing by the End of 2025. https://news.3m.com/2022-12-20-3M-to-Exit-PFAS-
299   Manufacturing-by-the-End-of-2025. Accessed 1 Jan 2023

300   18.   Sun C, Liu L, Vasudevan HN, et al (2021) Accurate Bulk Quantitation of Droplet Digital Polymerase Chain Reaction.
301   Anal Chem 93:9974–9979. https://doi.org/10.1021/acs.analchem.1c00877

302   19.   Gopu M, Mampallil D (2022) Distributed evaporation of water-in-oil emulsion drops on solid surfaces. Physics of
303   Fluids 34:102110. https://doi.org/10.1063/5.0107471

304   20.   Haug A, Bjerrum J, Buchardt O, et al (1961) The Affinity of Some Divalent Metals for Different Types of Alginates.
305   Acta Chem Scand 15:1794–1795. https://doi.org/10.3891/acta.chem.scand.15-1794

306   21.   Kuffel A, Gray A, Daeid NN (2021) Impact of metal ions on PCR inhibition and RT-PCR efficiency. Int J Legal Med
307   135:63–72. https://doi.org/10.1007/s00414-020-02363-4

308   22.   Mørch ÝA, Donati I, Strand BL, Skjåk-Bræk G (2006) Effect of Ca^{2+} , Ba^{2+} , and Sr^{2+} on Alginate Microbeads.
309   Biomacromolecules 7:1471–1480. https://doi.org/10.1021/bm060010d

310   23.   Yao J, Lin F, Kim H, Park J (2019) The Effect of Oil Viscosity on Droplet Generation Rate and Droplet Size in a T-
311   Junction Microfluidic Droplet Generator. Micromachines 10:808. https://doi.org/10.3390/mi10120808

312   24.   Kukizaki M, Goto M (2006) Effects of Interfacial Tension and Viscosities of Oil and Water Phases on Monodispersed
313   Droplet   Formation   Using   a   Shirasu-porous-glass   (SPG)   Membrane.   Membrane   31:215–220.
314   https://doi.org/10.5360/membrane.31.215

315   25.   Liu Y, Zhang C-J, Zhao J-C, et al (2016) Bio-based barium alginate film: Preparation, flame retardancy and thermal
316   degradation behavior. Carbohydrate Polymers 139:106–114. https://doi.org/10.1016/j.carbpol.2015.12.044

317   26.   Ma C, Tan ZL, Lin Y, et al (2019) Gel microdroplet–based high-throughput screening for directed evolution of
318   xylanase-producing   Pichia   pastoris.   Journal   of   Bioscience   and   Bioengineering   128:662–668.
319   https://doi.org/10.1016/j.jbiosc.2019.05.008

320   27.   Tan ZL (2019) Development of Function-Driven Screening Bioprospecting Platform Based on Droplet-Based
321   Microfluidics. Tsinghua University

322   28.   Interuniversity Microelectronics Centre (2017) Extremely fast on-chip cell sorting. https://www.imec-
323   int.com/drupal/sites/default/files/inline-files/Extremely%20fast%20on-chip%20cell%20sorter_WEB_0.pdf. Accessed
324   30 Jan 2023