An open-source, high resolution, automated fluorescence microscope

Ando C. Zehrer[1], Ana Martin-Villalba[3], Benedict Diederich[2*] and Helge Ewers[1*]

[1]Institut für Chemie und Biochemie, Freie Universität Berlin, 14195 Berlin, Germany

[2]Leibniz-IPHT Jena, Albert-Einstein-Str. 9, 07745 Jena, Germany

[3]Department of Molecular Neurobiology, German Cancer Research Center (DFKZ), Im Neuenheimer Feld 581, 69120 Heidelberg, Germany

[*]correspondence to: benedict.diederich@leibniz-ipht.de or helge.ewers@fu-berlin.de

Abstract

Fluorescence microscopy is a fundamental tool in the life sciences, but the availability of sophisticated equipment required to yield high-quality, quantitative data is a major bottleneck in data production in many laboratories worldwide. This problem has long been recognized and the abundancy of low-cost electronics and the simplification of fabrication through 3D-printing have led to the emergence of open-source scientific hardware as a research field. Cost effective fluorescence microscopes can be assembled from cheaply mass-produced components, but lag behind commercial solutions in image quality. On the other hand, blueprints of sophisticated microscopes such as light-sheet or super-resolution systems, custom-assembled from high quality parts, are available, but require a high level of expertise from the user.

Here we combine the UC2 microscopy toolbox with high-quality components and integrated electronics and software to assemble an automated high-resolution fluorescence microscope. Using this microscope, we demonstrate high resolution fluorescence imaging for fixed and live samples. When operated inside an incubator, long-term live-cell imaging over several days was possible. Our microscope reaches single molecule sensitivity, and we performed single particle tracking and SMLM super-resolution microscopy experiments in cells. Our setup costs a fraction of its commercially available counterparts but still provides a maximum of capabilities and image

quality. We thus provide a proof of concept that high quality scientific data can be generated by
lay users with a low-budget system and open-source software. Our system can be used for routine
imaging in laboratories that do not have the means to acquire commercial systems and through its
affordability can serve as teaching material to students.

Keywords: light microscopy, low-cost, live cell imaging, *d*STORM, open-source, UC2, modular,
adaptable, SMLM, single particle tracking, ImSwitch

Introduction

Fluorescence microscopy is, due to its excellent contrast, specificity and compatibility with live
cell imaging, an essential tool in the life sciences. Digital imaging, quantitative analysis and ever
higher throughput and resolution in new techniques make fluorescence microscopy more
important than ever. Nevertheless, while the development of high-end instrumentation pushes
forward our progress in understanding cellular processes, spatial and temporal resolution are not
the only challenges impeding progress in scientific discovery. This is because necessarily, the
access to cutting-edge instrumentation is reserved to a subset of scientists as a result of its cost
and technical complexity, effectively limiting data production to laboratories with access to
substantial resources. At the same time, cost of electronic or optical components, powerful
processors and high-quality cameras are at an all-time low. Together with simple prototyping
through 3D-printing, this has led to the emergence of open-source scientific hardware as a
research field (Wenzel, 2023). The open-source microscopy community has experienced the
rapid development of different approaches aiming to make microscopy more available (Almada
et al., 2019; Alsamsam et al., 2022; Ambrose et al., 2020; Auer et al., 2018; Danial et al., 2022;
Diederich et al., 2019a; Hohlbein et al., 2022; Sharkey et al., 2016; Voigt et al., 2019; Wenzel,
2023).

The first iterations of open-source microscopes came from microscopy laboratories that tried to
provide similar quality data as commercial systems while being assembled from components for
lower overall cost, however still requiring significant expertise. Solutions now exist for single
molecule localization microscopy (SMLM) (Alsamsam et al., 2022; Auer et al., 2018; Holm et
al., 2014; Kwakwa et al., 2016; Martens et al., 2019), they include modular platforms (H. Li et

al., 2020), light sources (Berry et al., 2021; Schröder et al., 2020) and even sophisticated special

solutions such as light sheet microscopy (Voigt et al., 2019), spinning disc microscopes (Halpern

et al., 2022) and high-throughput applications (Li et al., 2019; Walzik et al., 2015).

On the other hand, another group of investigators tried to provide extremely cost-effective

microscopy solutions that are accessible to everyone. The result are solutions that can be useful in

detecting e.g. parasites in inaccessible regions of the world for simple diagnosis (Li et al., 2019),

but are incompatible with quantitative scientific imaging (Chagas et al., 2017; Cybulski et al.,

2014; Diederich et al., 2020, 2019b; Sharkey et al., 2016; Vera et al., 2016). Most of these

solutions lack excitation and emission for fluorescence imaging.

Here, we propose an instrument filling the gap between the low-cost open-source instruments for

the general public and high-quality fluorescence imaging, as required for scientific research in

molecular cell biology. We provide an automated high-resolution microscope with single

molecule sensitivity that can be fully assembled and steered by lay users. Our system is based on

the UC2 toolbox (Diederich et al., 2020, 2019a; Wang et al., 2022) using common optical

elements, affordable electronic components and 3D printed parts and is combined with a high NA

objective, motorized stage and intuitive GUI-based operation.

The result is a cost efficient and user-friendly high-resolution fluorescence microscope that is

compatible with long-term live-cell imaging, single particle tracking and single molecule

localization microscopy. All designs for the 3D parts, software and electronics are open-source

(Table 1). We here describe microscope assembly, benchmarking and a number of high-

resolution microscopy assays performed with our microscope.

Results

Assembly of a high-resolution UC2 microscope.

We aimed to generate a cost-effective, high-resolution widefield fluorescence microscope that

allows us to perform a variety of common cell biological imaging experiments. These include

standard high-resolution diffraction limited immunofluorescence imaging, long-term live-cell

imaging, single particle tracking and super-resolution microscopy. To fulfill these criteria, we

reasoned that our system should have: i) excitation and detection powerful and sensitive enough

for single molecule localization, thus requiring a laser and an appropriate detector. ii) capability

to operate inside an incubator for long-term live-cell imaging, thus requiring small size and

automation. iii) an intuitive and user-friendly GUI for maximal accessibility. Lastly, the entire

system including the physical microscope, the electronics and the software must be open-source

and be possible to assemble and operate with moderate technical know-how (Figure 1a). We

reasoned that the quality of the objective is paramount to image quality, while most other optical

parts of a light microscope can be mass-produced components. Furthermore, such a high-quality

objective allows for the swift diagnosis of problems in the assembly of the optical path due to its

high resolution and minimal chromatic and spatial aberrations. We decided to use an Olympus

1.49 NA oil objective, but any high-quality objective can be incorporated in our system, by

adapting the objective mount to the requirements of the objectives' manufacturer.

Figure 1: Schematic of the high-resolution-UC2 widefield microscope. a. Overview of the different component
categories of the microscope, from the hardware, through the electronics controlling, to the software and the
ImSwitch-based GUI. **b.** Schematic of the complete setup. Red and blue arrows (**i.** and **ii.**) are references to help
visualize the 3D structure. Top layer is the commercially bought XY-stage that was motorized for this project.
Timing belts convert motor torque (two grey gears top left) into stage motion (bigger black gear middle) but have
been omitted in this representation for clarity. Sample holder can be printed according to the sample used, circular in
this case. **c.** Commercially bought precision Z-stage and high-NA objective (Olympus 60x/1.49 NA TIRF) **d.**

Excitation layer. Laser emission (bright red light-path) is focused by a f = 100 mm lens onto the back focal plane of
the objective. Orange arrow (**iii.**) marks the filter cube which has an excitation filter (Chroma ZET635/20x EX), a
dichroic mirror (Chroma ZT640rdc) and on the bottom an emission filter (Chroma ET655lp long-pass for λ > 655
nm) to separate excitation from fluorescence. Red rectangle shows a telescope build including a diffusor (rotating
piece of cling foil) in the focus point between both lenses. Telescope build magnifies and homogenizes the laser
beam and can be used optionally**. e.** Detection layer, corresponds to the bottom layer in the setup (red star as
reference). Emitted fluorescence (dark red light-path) is depicted with a f = 100 mm lens on the detector (Alvium
1800 U-158c CMOS camera from Allied Vision). All active optical elements, e.g. mirrors and lenses are mounted
into the injection moulded cubes (50x50x50 mm, green) via custom 3D printed mounts. Cubes are connected
laterally and vertically via the puzzle shaped connection pieces. Green arrows (**iv.**) indicate the connecting puzzle
layers between the cubes.

Based on the guidelines described above and to ensure adaptability and modularity, we decided to

build our open-source automated microscope on the basis of the UC2 toolbox using its injection

molded cubes for the structural assembly. For ease of usage, we used an inverted configuration

and assembled the microscope in three layers to keep the size and design compact (Figure 1b).

The high-precision x-y sample table and z-stage (Figure 1c) were low-cost commercial solutions

that we motorized and implemented into the open-source, python-based microscope GUI as

automated elements of the microscope. The emission pathway at the bottom of the microscope

assembly is based on an Alvium 1800 U-158c CMOS camera from Allied Vision, which

compared to conventional scientific cameras is more affordable by a factor of 30 to 60 times at

the time of submission (Figure 1d), more compact and consumes less energy. One important

point here is that the camera needs to be monochromatic as the Bayer pattern on a polychromatic

camera would significantly reduce the sampling as well as the quantum efficiency of a camera

chip, which is already lower compared to that of scientific cameras. The camera was connected to

the computer via USB 3.0, allowing fast data transfer and maximal compatibility. Between the

emission pathway and the sample stage, the excitation pathway contained a low-cost 638 nm

diode laser with an optional beam expander (Figure 1e), that depending on the illumination

requirements could be inserted or removed, modulating the illumination density and homogeneity

of the excitation field (see Figure 2c). To trigger the laser, control laser power and to steer the x,

y and z motors for focus and sample translation, we used a low-cost Arduino-based board that

can be connected via USB3.0 to any computer. Electronics on the board and the USB-camera are

then directly controlled via an adapted GUI of the python-based open-source software ImSwitch

(Moreno et al. 2021), resulting in a fully automated microscope. Additionally, ImSwitch includes

the "ImScript" an automation feature, which allows to execute a sequence of written *Python*

commands with user defined parameters e.g. overnight time-series imaging experiments.

Furthermore, we took advantage of the open-source (OS) nature of ImSwitch and implemented

the recently presented OS SMLM algorithms microEye (Alsamsam et al., 2022) to realize an

online-running localization framework that gives instant qualitative feedback if the sample

preparation works as expected. In addition, the feedback of these online-running image

processing algorithms can be looped back into the hardware control to eventually compensate

possible drift (e.g. autofocus). A large variety of available Napari (Ahlers et al., 2023) plugins

can be used to directly work on the data without moving them to a different computer.

All steerable electronics could alternatively be controlled with a PS4 controller for increased

user-friendliness. The assembly can be housed with 3D printed walls or even black cardboard to

block ambient light and increase laser safety. In table 1, a list of the different microscope and

software elements and the repository addresses to the corresponding materials and methods is

displayed. A comprehensive online repository provides the bill of materials (BOM), as well as a

set of instructions on how to setup the soft- and hardware (Table 1).

The 638 nm Red Laser Module used for illumination induced a recurrent pattern in the

detection path when imaging a coverslip homogeneously covered with fluorescent molecules.

Spectrometric measurement of the excitation laser beam confirmed that our low-cost laser had

minor bands in the wavelength range of 650 - 670 nm, which we could remove using an

additional band pass excitation filter (ZET635/20X). Our finding encouraged us in our approach

to mix low-budget components with high quality optics. Focusing and positioning of the sample

is done with commercial low-budget stages that offer great long-term stability. To render all three

axes of movement computer controlled, we first motorized an x-y stage via torque transmission

form the motor to the micrometer screws of the stage via timing belts. Secondly, we incorporated

a motorized lead screw-driven (NEMA11, 50mm, Amazon, 50€) z-stage holding the objective

into the modular design of the microscope.

Imaging experiments.

We first used our microscope to image a sample of fixed mammalian cells stained against tubulin

via immunofluorescence using AF647. When we imaged these cells with moderate laser intensity

(35 W/cm^2), beam expander and 50 ms integration time, we found that cells were evenly

illuminated (Figure 2a) and microtubules appeared as diffraction-limited lines inside cells (Figure

2b). To quantify flatness of illumination, we made use of a red auto-fluorescent benchmarking

slide (Chroma). The measured fluorescence intensity was quite homogenous across the field of

view, with variations of up to 25% of the maximal fluorescence intensity, excluding the bottom

left and right corners (Figure 2c).

176
Figure 2: Widefield imaging of immuno-stained CV1 cells against tubulin (α and β-tubulin). a. Widefield
image of a tubulin stained CV1 cell. Two regions of interest are highlighted. Scale bar represents 20 µm. **b.** Zoomed
in images of the selected ROIs in **a.** with respective microtubule profile plotted across the white line in the ROI
images. ROI dimensions are 8x8 µm. Grey values are plotted as columns and fitted with gaussian functions (red
line). Microtubule diameter can be estimated to 340 to 420 nm according to the full width at half maximum of the
fits. **c.** Illumination profile of whole camera chip (1456x1088 px) at the sample plane as imaged on a far-red
fluorescent Chroma slide. Scale bar represents 35 µm. **d.** Tetra-speck beads imaged over a period of three hours and
color coded over time. Scale bar represents 1 µm. Bead positions are localized over time to measure their
displacement to their initial position over time.

These conditions are compatible with high quality fluorescence microscopy in life science

research. To test the stability of our system, we then imaged sub-diffraction fluorescent beads

repeatedly over time. When we then calculated the displacement of the beads over time, we

consistently found drift of less than 1 µm over 2 hours (Figure 2d). We concluded that our

microscope may be suitable for imaging experiments that require repeated measurements of the

same field of view such as live-cell imaging. We thus mounted live cells at room temperature on

the optical table on our microscope and stained them with SiR-actin (Spirochrome). When we

then repeatedly illuminated the cells with moderate laser power (18 W/cm^2) to excite SiR-actin

fluorescence and with white light in transmission mode to image the cells, we found that the cells

tolerated acquisition conditions for several hours without damage visible in fluorescence or

widefield image (Figure 3). As expected, drift was negligible. Importantly, we found similar

stability against drift over time, when we performed this experiment not on an optical table, but a

wetlab bench over a layer of foam material (made of Polyethylene or alternatively Polyurethane)

as a damping layer. This suggested that our microscope might be suitable for long-term imaging

in an environment that does not provide vibration damping, such as a cell culture incubator.

Given that our microscope assembly is of small size, and connected via two flexible USB cables,

we then moved on to perform live-cell imaging over longer periods in a cell culture incubator.

Figure 3: Live-cell imaging of actin in cultured cells. Shown are still images of the same CV-1 fibroblast cells in
SiR-actin fluorescence and widefield (background corrected) at different timepoints over 5 hours. Scalebar on the
right represents 30 µm.

To do so, we used a standard 37 ˚C CO_2 incubator. The longer USB cables connecting the

microscope to the exterior allowed for steering from the outside. When we plated live T98G

cancer cells on coverslips and mounted them on the microscope sample holder in warm medium

inside the incubator, they exhibited normal morphology and growth over many hours. When we

performed a long-term imaging experiment using Syto deep-red fluorescence staining of the

nucleus (Thermo Fisher), we found that cells kept dividing and moving on the coverslip for

several days (Figure 4). We concluded that our setup was compatible with long-term live-cell

imaging of cells.

Figure 4: Long-term live-cell imaging of cultured cells inside the incubator. Shown are still images of the same
population of T98G glioblastoma cells over time in the same field of view. Cells are stained with SYTO nucleic-acid
fluorescent dye and imaged in widefield (background corrected) every 20 minutes for over 32 hours inside an
incubator. Scalebar on the right represents 30 µm.

Next, we aimed to further explore the single molecule capabilities of our microscope and

performed single particle tracking (SPT) experiments on live CV-1 African green monkey cells

stably expressing GPI-GFP. When we incubated these cells with streptavidin 655 nm quantum

dots (QDs) coupled to biotin-tagged anti-GFP nanobodies (Figure 5a), we detected QDs readily

as single diffraction limited entities on cells (Figure 5b). When we then localized emission

centers and tracked the QDs over time, we could generate trajectories over many frames (Figure

5c,d) and extract parameters describing particle movement such as the diffusion coefficient (D,

Figure 5e) and the moment scaling spectrum (MSS, Figure 5f (Ewers et al., 2005)). The MSS

introduces a parameter that allows to quantitatively describe the type of motion a particle

undergoes (MSS = 0 being immobility, < 0,5 Brownian diffusion and > 0,5 super-diffusive). Our

measurements were in agreement with our previous observations (J. H. Li et al., 2020).

234
Figure 5: Single particle tracking of GPI-GFP in live CV1 cells using functionalized Quantum dots. a.
Schematic of the construct used to track single GPI- molecules with quantum dots. **b.** Exemplary raw images of
single fluorescent Quantum Dots. Image dimension is 1.6 µm x 1.6 µm. **c.** Trajectories of several single molecules
visualized on a top-light illuminated image of CV1 cells. Scale bar represents 10 µm. **d.** Selected trajectory of a
single particle over 167 time-points. Sampling rate is 25 ms, pixel size 104 nm. **e.** Histogram of the diffusion
coefficient of around 5000 tracks with 20 or more detected time points each. **f.** MSS (moment scaling spectrum) of
the same population as plotted in **e.**

Since we were readily able to detect single quantum dots, which are single emitters of few

nanometers in size, we reasoned, it might be possible to perform single molecule localization

microscopy (SMLM) experiments as well. We immuno-stained CV-1 cells against tubulin using

AF647 coupled secondary antibodies and incubated coverslips in BME/GLOX buffer for

effective pumping into the dark state. When we then illuminated our sample with high intensity

laser light (in our case ~ 1kW/cm^2), we found that the intense staining quickly faded and gave

way to the typically observed blinking in SMLM experiments. We took a timeseries of 30'000

frames and localized the detected emitters using *Thunderstorm* (Ovesný et al., 2014). Even

though we were limited to widefield illumination, we achieved a drastic improvement in

resolution in reconstructed SMLM micrographs (Figure 6a). When we measured the profile of

individual microtubules, we could easily detect the typical "railroad track" pattern with a spacing

of 38 – 43 nm, indicative of high quality SMLM imaging (Figure 6b,c). To quantify the achieved

resolution, we used Fourier ring correlation (Nieuwenhuizen et al., 2013) and found that the

overall achieved resolution was 93 nm, clearly below the optical resolution limit (Figure 6d).

256
Figure 6: dSTORM reconstruction of immuno-stained CV1 cells against tubulin (α and β-tubulin). a.
Widefield image and reconstructed super-resolution image of 30'000 widefield frames, localized and reconstructed
using *ThunderSTORM*, applied parameters described in the methods part. Reconstruction is displayed with the
average shifted histogram method, including a 25-fold magnification. Scale bar represents 5 µm. Two regions of
interest in the reconstruction are marked (red square for **b.** and green square for **c.**) and zoomed in (scale bar 1 µm).
Profiles within these ROIs are measured by averaging the profile of the reconstructed image over 250 nm along the
microtubule (white dashed rectangles). White histogram bars are the intensity values of the pixels in the
reconstruction. Two peaks arise from the circular shape of microtubules. These were analyzed with a double
gaussian fit (red curves are the cumulated single black fits). Distances of (38 ± 2) nm (red box) and (43 ± 2) nm
(green box) could be extrapolated. **d.** Fourier ring correlation of the complete reconstructed image. Threshold is left
to the preset 1/7. Overall resolution is estimated at 93 nm.

DISCUSSION.

Here we developed a cost effective, high-quality automated microscope that can be assembled by

laypeople with minimal background knowledge in optics, electronics and informatics. The

assembly guidelines including CAD files of the 3D printed parts, list of used components and

software are openly accessible on the Github repositories (see table 1), thus offering maximal

availability.

Our microscope is based on a simple modular assembly system called UC2 (You.See.Too. ,

(Diederich et al., 2019a)). This framework is based on modular cubes that can hold different

optical components in place and allow for simplified alignment and assembly. This flexibility

furthermore facilitates the replacement of any lens, filter, laser camera or other components,

allowing the user to implement already available parts into their setup, or to further trim down the

setup by removing unused parts. We made use of an improved version where the cubes are made

through injection molding, which provides a significantly higher manufacturing precision than

the same cubes fabricated using 3D-printing, resulting in greater mechanical stability and

reproducibility, while being available at low cost through mass production (Wang et al. 2022).

The hereby provided increased structural integrity makes the setup portable allows it to be

reproducibly built and dismantled within hours. This facilitates its use in remote areas. Improved

versions or updated parts could easily be shared digitally and printed locally, further promoting

the advantage of digital manufacturing in the laboratory. The here created assembly can easily be

replicated multiple times to increase imaging throughput still for a fraction of the cost of a single

high-end microscope.

We present data in standard immunofluorescence microscopy, live-cell microscopy and long-

term live-cell microscopy assays performed inside an incubator. Furthermore, we show single

particle tracking data of semiconductor quantum dots on live cells and single molecule

localization based super-resolution microscopy. In all cases, the quality of data we generated is

comparable to those generated by systems orders of magnitude more expensive.

Our system thus allows for sophisticated imaging assays to be performed at a professional level

with greatly reduced cost. At the same time, electronic control of components and remote

steering of the microscope can be implemented without background knowledge in electronics.

This makes the generation of high-quality image data accessible especially for life science

research groups with limited monetary means also in lower income countries. The main

limitation of our system in this direction is the requirement of a high-NA objective, the by far

most expensive component of the setup. Since for scientific imaging, high quality data and thus

high-NA and corrected objectives are essential, this component cannot be replaced. However, the

quality of low-cost objectives has recently been improving, even with objectives allowing for

effective single molecule detection. The reduced NA diminishes the localization precision and the

signal to noise ratio, thus limiting the possible applications. Development of brighter dyes might

also be an alternative to improve the imaging quality enough to account for low budget lasers and
filters.

Due to its modular assembly, extensions may be easily implemented to our microscope.
Depending on the specific research aim, multiple excitation and emission wavelengths may be
desirable, but would require a redesign of the frame to incorporate motorized filter switching and
more light sources. Alternatively, multiband filters in combination with fiber-coupled and thus
combinable laser sources are also possible. In the future, the possibilities of this system could be
significantly expanded by adding more modules specific for additional capabilities and we
specifically invite the community to do so. Especially the combination with a low-budget
spinning disc (Halpern et al., 2022), TIRF illumination or light-sheet imaging (Diederich et al.,
2020) are possible future applications.

Further steps towards user-friendliness include the replacement of the laser light-source by a
high-power LED to reduce the requirement of laser safety. Another step would be to render the
electronics assembly fully plugable, thus eliminating the soldering steps that are required in the
current assembly.

Besides the application in basic research, this cheap and easy to assemble setup would be useful
for teaching the next generation of scientists that more than ever have to be able to synthesize
capabilities spanning several disciplines. We endorse making scientific instruments available for
more scientists, students, as well as the general public. Our system builds on efforts in open
science regarding software (ImSwitch,(Moreno et al., 2021) and microscope hardware (UC2
(Diederich et al., 2019a)) while benefitting from a growing field of open-source microscopy and
laboratory hardware in general (Almada et al., 2019; Auer et al., 2018; Hohlbein et al., 2022;
Martens et al., 2019; Wenzel, 2023). Taken together with efforts in making fluorescent dyes
available (Lavis, 2021), fundamental obstacles to the free diffusion of capabilities and know-how
in science are being removed. We hope that our live-cell high-resolution incubator microscope
(U.C.STORM) will catalyze these efforts.

Methods

Component fabrication

In the components list, a detailed table describes all the different pieces used for the microscope, including the price of each component. The CAD designs as well as further information about UC2 or setup building can be found on the Github repository (see table 1). The XY-stage was a manual commercial model that was modified and motorized for the application. The Z stage was also commercially bought, but already manufactured with a motor. Both stages were chosen because of their low price and convenient use. Custom-made stage solutions to print and build can be found on the UC2 repository (Table 1). In the slicing software Prusa slicer, the original Prusa i3 MK3S & MK3S+ printer was set. The parts were printed with the 0.15 mm SPEED or 0.15 mm QUALITY settings. The two materials used in this project were the Prusament PLA (Polylactic Acid) and Prusament ABS (Acrylnitril-Butadien-Styrol) filaments (both from Prusa), which can both be selected in the slicing software ABS was used wherever heavy loads may deform the setup in the event of temperature gradients (e.g. objective mount). Materials were printed with PLA infill (20 - 40%) ($T_{print}= 215$ ˚C; $T_{bed}= 60$ ˚C). Materials printed with ABS infill (20 - 40%) ($T_{print}= 255$ ˚C; $T_{bed}= 100$ ˚C).

UC2 setup assembly.

In the UC2 system, optical components were mounted in standardized cubes. To make alignment possible over an optical excitation or emission path over several cubes distance, the components must experience a large enough degree of freedom within the cubes so they can be moved manually in angle and exact position, but low enough to keep components immobile over prolonged time periods. The challenge presented by tilted optics can on the one side be addressed during the design process; broader mounts/adapters are less prone to tilt due to higher contact surface with the cube. Once the optics were orthogonal to the light path, their position in relation to the optical axis was set manually with highest possible accuracy, using appropriate mounts. The fine adjustment was done with tiltable mirrors, as using translational mounts for every optical element would be complex to produce or costly. A helpful tool for adjustment was to position an aperture in the excitation pathway. The maximum of the laser spot was set to be in the center of the aperture hole and was adjusted using the mirror. The emission path was adapted in the same way, as imaging a fluorescent surface allows to position the maximal laser intensity spot on the center of the detector field of view (also with a mirror).

The setup entails two imaging channels: the laser induced fluorescence (excitation path) and the
LED top-light illumination (650 nm Star LED, Cree, 3W, brightfield). The top-light light source
is mechanically detangled from the stages to remain unaltered by sample motion. As the filter
cube in the microscope is immobile, the top-light illumination requires either a high-power white
LED or an LED that emits light in a wavelength range that is not blocked by the dichromatic
mirror or the emission filter (in our case 650-660 nm).

Setup parts availability and price:

The different 3D printed parts, electronics and optics used in the assembly, the corresponding
CAD files as well as an estimation of the price can be found in the corresponding Github
repositories, as listed in table 1. General information about the UC2 system can be found on the
UC2 repository.

Repository	Link:
General openUC2	https://github.com/openUC2/
Specific to this setup: UC.STORM	https://github.com/openUC2/UC2-STORM-and-Fluorescence
ImSwitch (UC2 specific)	https://github.com/openUC2/ImSwitch/.
UC2-Rest	https://github.com/openUC2/UC2-REST
UC2-ESP32	https://github.com/youseetoo/uc2-esp32
UC2-ESP32 Firmware Flashintool (Online)	https://youseetoo.github.io/

Table 1: Github repositories for hardware, electronics and software.

Image acquisition and control software:

For the microscopy control, acquisition and reconstruction software, we chose ImSwitch because
its modular design adapts very well to the inherently modular nature of the cube-based optics

platform. A novel firmware "UC2-ESP32" was developed, which is fully integrated into the

python-based ImSwitch structure. Using a unified REST API-like interface, lasers, motors,

LEDs, LED arrays and readout sensors can be controlled serially via WiFi and USB. To ensure

wide distribution of this device adapter, we used a readily available hardware module that can

control multiple stepper motors and easily adapt other hardware components such as the LEDs

and lasers. An online-based flash tool loads the latest firmware onto the Wemos ESP32 D1 R32

in combination with the CNC Shield to create a fully functional microscopy control unit. Detailed

documentation can be found in the online documentation (https://youseetoo.github.io/). A Python

library "UC2-REST" adapts the USB serial and integrates it into the ImSwitch software. In this

project, the strengths of open-source were fully unleashed by extending the control and

acquisition algorithm ImSwitch with an online SMLM reconstruction algorithm by (Alsamsam et

al., 2022). The user can select different parameters from the MicroEye framework to locate

blinking events directly in the Napari viewer (Sofroniew et al., 2022) to directly track possible

reconstruction or sample artefacts. The UC2-specific fork amongst other repositories can be

found in Table 1.

Imaging modalities.

Illumination: The illumination is provided by the 638nm Red Laser Module 500mW Round

Dot Focusable TTL 3050 (acquired on lasernets.com). When implementing the telescope and

beam homogenizer, the illumination density reaches up to 90 W/cm^2 but could be tuned to any

required value below that. In the focal point between both lenses a rotating piece of cling film

diffuses the beam. The variance in the laser density over the field of view excluding the corners

was measured with a fluorescent slide and is reproducibly lower than 25%.

Additionally, for long term experiments, lacking fluorescence intensity can to an extent be

compensated with longer detector exposure times. To perform SMLM i.e. dSTORM in our case,

the telescope is removed to have maximal laser density. The total available power at the sample

plane is 83 mW thus a laser density of around 520 W/cm^2. But since the laser spot is smaller

than the FOV, sufficient laser density can be reached when only using a subset of the FOV

(approximately a third of the FOV) can be used with sufficient laser power (over 970 W/cm^2)

and a variation in laser intensity of about 25% of the laser intensity maximum.

<u>Objectives and pixel sizes</u>: For the widefield imaging, the room temperature live cell imaging, the
single particle tracking as well as the *d*STORM experiment, a 60x Olympus 1.49NA oil
immersion objective (UPLAPO60XOHR) is used. The setup optics have accordingly been chosen
for maximal resolution thus matching the Nyquist sampling rate on the camera. The pixel-size is
determined with a calibrated grid and a value of 104.1 nm is extrapolated. For the incubator
measurements, a 20x Olympus 0.6 NA air (MXPLFLN20X) is used. The optics in the setup are
not adjusted to this objective, resulting in a pixel-size of 319.4 nm.

Coverslip and cell seeding:

<u>Coverslip preparation:</u> Metal holders were used in combination with circular coverslips (25 mm)
which were first cleaned with a 2% Hellmanex solution for 10 minutes then cleaned in ethanol
for 10 minutes and were finally placed in the plasma cleaner for 15 minutes after having dried.
The 8-well sample-holders (Ibidi) were washed with 1 M KOH for 10 min then rinsed with H_2O
and dried.

<u>Cell culture medium:</u> Cell culture medium was on a phenol red free DMEM base with 1%
GlutaMAX and 10% FBS.

<u>Cell seeding:</u> Cells were detached from the cell culture dish with a 2 mM EDTA solution. 30'000
cells were seeded on a 25 mm circular coverslip for *d*STORM experiments and general wide-
field imaging. For the live cell experiments, 55'000 to 80'000 cells were seeded onto 25 mm
coverslips, alternatively 15'000 - 25'000 cells per well when using an 8 well chamber.

Microtubule staining:

CV-1 cells (Wild-type Cercopithecus aethiops kidney fibroblasts) were first washed with 37 ˚C
warm PEM buffer. The cells were fixed with warm PEM with 4% PFA, 0.05% GA and 0.1%
TX 100 for 20 minutes. After fixation, sample was quenched in 8 minutes with NH_4Cl/PBS
(50 mM) then washed 3 times with PEM buffer. The cells were then permeabilized with 0.3 %
TX100 in PEM for 5 minutes and washed with PEM buffer. The sample was then blocked for 30
minutes in imageIT followed by 1 hour in a 4 % HS, 1 % BSA, 0.1 % TX100 PEM based
solution.

The sample was incubated overnight (12□hours) at 4 ˚C with a primary mouse anti alpha- and

beta- tubulin IgG antibodies (Sigma-Aldrich (REF: T5168) and Sigma-Aldrich (REF: T5293))

diluted by a factor of 1 to 350 in a 1% BSA and 0.1 % TX100 PEM based solution. Sample was

thoroughly washed 3 times for 5 minutes with a 1 % BSA and 0.1 % TX100 PEM-based solution.

The cells were then incubated with the secondary donkey anti-mouse antibody conjugated with

the AF647 fluorophore, diluted by a factor of 1 to 250 in a 1% BSA and 0.1 % TX100 PEM

based solution at room temperature for 2 hours. The 3 washing steps were then repeated, as

previously mentioned for the primary antibody.

The sample was the post-fixated with a 4% PFA diluted in PEM solution, for 15 minutes. The

cells were then quenched with a 50 mM NH_4Cl solution diluted in PBS for 8 minutes. Finally, the

cells was washed 3 times.

Single particle tracking:

Anti-GFP nanobodies (LaG-16 anti-GFP, own production according to (Fabricius et al., 2018))

were added to a biotin (α-Biotin-ω-(succinimidyl propionate)-24(ethylene glycol)) solution at

double molar excess. After 1 hour of incubation on a shaker (300 rpm), the nanobody-biotin

construct was purified by sequential filtration through three 7 kDa MWCO desalting columns.

The biotin conjugated anti-GFP nanobodies were mixed in a 1 to 1 ratio with the streptavidin

coated quantum dots and incubated for 10 minutes at room temperature. The solution was then

diluted into 1 ml of live cell medium. The cells were then incubated at room temperature with

this solution for 15 minutes. Afterwards the supernatant was removed and replaced with regular

live cell-medium (Thermo Fischer).

SiR actin staining:

The CV-1 cells were washed in 37°C warm Medium (DMEM without phenol red + FBS +

glutamax). 1 µl of the original SiR Actin (Spirochrome, 100 nM) solution was diluted in 1 ml of

37°C warm Medium (DMEM without phenol red + 10 % FBS + 1 **%** glutamax). To this

solution, 1 µl of verapamil was added. The sample medium was removed and exchanged with this

solution. After 1 hour of incubation in the incubator, the sample was imaged in a live cell

imaging solution (Invitrogen) for the following hours at room temperature.

Nucleic staining.

T98G cells were washed in 37°□C prewarmed medium (DMEM without phenol red + 10 % FBS

+ 1 % glutamax). For staining, 1□µl of SYTO™ Red Fluorescent Nucleic Acid Stain

(Invitrogen) was diluted in 1□ml of complete medium and incubated with the sample for 30

minutes at 37°□C in the incubator, then imaged over prolonged time.

(*d*)**STORM imaging.**

<u>Imaging buffer:</u> Imaging buffer was made of 150 mM Tris-HCl, 1.5 % β-

Mercaptoethanol, 0.5 % (v/w) glucose, 0.25 mg/ml glucose oxidase and 20 µg/ml catalase

(pH 8.8).

<u>Image acquisition:</u> The 30'000 frames of the (*d*)STORM data set have been acquired with

ImSwitch on the full camera chip (1456 x 1088 px). The exposure time was 20 ms, the gain was

set to 20 (maximal value), the black-level and offset to 0 and no time delay was set between two

frames. Laser power was set to 1024 (maximal value). Setup was positioned on a regular table

with a layer of foam material to dump vibrations. The room was left during the measurement.

The images were acquired within around 15 minutes and saved directly on the computer disc in a

.hdf5 file format for time efficiency and data handling advantages.

<u>Analysis and reconstruction:</u> The (d)STORM data-set was imported as a whole into ImageJ as a

hdf5 file. Acquiring the full CMOS chip 30'000 times generated enormous datasets (95.6 Gb). To

load the file, a Fiji plugin to load the stack as virtual stack was used. (N5, (Saalfeld et al., 2022)).

The stack of images was then cropped to retain the region of interest, which has been positioned

to match the region where sufficient laser density induces blinking of single molecules during the

acquisition. The analysis was done using ThunderSTORM (Ovesný et al., 2014), a publicly

available SMLM reconstruction plugin on ImageJ. The camera parameters mentioned in the

image acquisition section were used for the analysis. Image filtering was done with a Difference-

of-Gaussian filter $\sigma_1 = 1.1$ px; $\sigma_2 = 1.7$ px. The approximate localization of molecules was done

using the local maximum method with default settings (std(Wave.F1) as peak intensity threshold

and 8-neighborhood connectivity). Sub-pixel localization of molecules was done with the PSF:

Integrated Gaussian (Fitting radius 3px) and maximum likelihood fitting method (initial sigma

1.7 px). Reconstruction of localized the raw-data was visualized using the average shifted

histograms method (Magnification: 25x; lateral shifts: 3).

Author contributions:

HE and AMV acquired funding and conceptualized the study. BD developed the UC2 platform

and adapted the electronics and software. ACZ and BD build and improved the microscope and

its application. ACZ conducted the experiments and data analysis. ACZ, HE and BD wrote the

paper.

Acknowledgements:

This work was funded by the Deutsche Forschungsgemeinschaft (DFG, German Research

Foundation) as part of TRR 186 (Project Number 278001972) and by Freie Universität Berlin.

The authors would like to thank all members of the Ewers laboratory for helpful discussions. We

thank Haoran Wang for his technical support and helpful insights. We also thank the free state of

Thuringia to support his project.

References

513
514
Ahlers J., Althviz Moré D., Amsalem O., Anderson A., Bokota G., Boone P., Bragantini J.,
Buckley G., Burt A., Bussonnier M., Can Solak A., Caporal C., Doncila Pop D., Evans
K., Freeman J., Gaifas L., Gohlke C., Gunalan K., Har-Gil H., … Yamauchi K. (2023).
napari: a multi-dimensional image viewer for Python (v0.4.18). Zenodo.
https://doi.org/10.5281/ZENODO.8115575
Almada P, Pereira PM, Culley S, Caillol G, Boroni-Rueda F, Dix CL, Charras G, Baum B, Laine
RF, Leterrier C, Henriques R. 2019. Automating multimodal microscopy with NanoJ-
Fluidics. *Nat Commun* **10**:1223. doi:10.1038/s41467-019-09231-9
Alsamsam MN, Kopūstas A, Jurevičiūtė M, Tutkus M. 2022. The miEye: Bench-top super-
resolution microscope with cost-effective equipment. *HardwareX* **12**.
doi:10.1016/j.ohx.2022.e00368
Ambrose B, Baxter JM, Cully J, Willmott M, Steele EM, Bateman BC, Martin-Fernandez ML,
Cadby A, Shewring J, Aaldering M, Craggs TD. 2020. The smfBox is an open-source
platform for single-molecule FRET. *Nat Commun* **11**:5641. doi:10.1038/s41467-020-
19468-4
Auer A, Schlichthaerle T, Woehrstein JB, Schueder F, Strauss MT, Grabmayr H, Jungmann R.
2018. Nanometer-scale Multiplexed Super-Resolution Imaging with an Economic 3D-
DNA-PAINT Microscope. *Chemphyschem* **19**:3024–3034. doi:10.1002/cphc.201800630
Berry K, Taormina M, Maltzer Z, Turner K, Gorham M, Nguyen T, Serafin R, Nicovich PR.
2021. Characterization of a fiber-coupled EvenField illumination system for fluorescence
microscopy. *Opt Express, OE* **29**:24349–24362. doi:10.1364/OE.430440

Chagas AM, Prieto-Godino LL, Arrenberg AB, Baden T. 2017. The €100 lab: A 3D-printable open-source platform for fluorescence microscopy, optogenetics, and accurate temperature control during behaviour of zebrafish, Drosophila, and Caenorhabditis elegans. *PLOS Biology* **15**:e2002702. doi:10.1371/journal.pbio.2002702

Cybulski JS, Clements J, Prakash M. 2014. Foldscope: Origami-Based Paper Microscope. *PLOS ONE* **9**:e98781. doi:10.1371/journal.pone.0098781

Danial JSH, Lam JYL, Wu Y, Woolley M, Dimou E, Cheetham MR, Emin D, Klenerman D. 2022. Constructing a cost-efficient, high-throughput and high-quality single-molecule localization microscope for super-resolution imaging. *Nat Protoc* **17**:2570–2619. doi:10.1038/s41596-022-00730-6

Diederich B, Diederich B, Richter R, Carlstedt S, Uwurukundo X, Uwurukundo X, Wang H, Mosig A, Heintzmann R, Heintzmann R. 2019a. UC2 – A 3D-printed General-Purpose Optical Toolbox for Microscopic ImagingImaging and Applied Optics 2019 (COSI, IS, MATH, PcAOP) (2019), Paper ITh3B.5. Presented at the Imaging Systems and Applications. Optica Publishing Group. p. ITh3B.5. doi:10.1364/ISA.2019.ITh3B.5

Diederich B, Lachmann R, Carlstedt S, Marsikova B, Wang H, Uwurukundo X, Mosig AS, Heintzmann R. 2020. A versatile and customizable low-cost 3D-printed open standard for microscopic imaging. *Nat Commun* **11**:5979. doi:10.1038/s41467-020-19447-9

Diederich B, Then P, Jügler A, Förster R, Heintzmann R. 2019b. cellSTORM—Cost-effective super-resolution on a cellphone using dSTORM. *PLOS ONE* **14**:e0209827. doi:10.1371/journal.pone.0209827

Ewers H, Smith AE, Sbalzarini IF, Lilie H, Koumoutsakos P, Helenius A. 2005. Single-particle tracking of murine polyoma virus-like particles on live cells and artificial membranes. *Proc Natl Acad Sci USA* **102**:15110–15115. doi:10.1073/pnas.0504407102

Fabricius V, Lefebre J, Geertsema H, Marino SF, Ewers H. 2018. Rapid and efficient C-terminal labeling of nanobodies for DNA-PAINT. *J Phys D Appl Phys* **51**:474005. doi:10.1088/1361-6463/aae0e2

Halpern AR, Lee MY, Howard MD, Woodworth MA, Nicovich PR, Vaughan JC. 2022. Versatile, do-it-yourself, low-cost spinning disk confocal microscope. *Biomed Opt Express, BOE* **13**:1102–1120. doi:10.1364/BOE.442087

Hohlbein J, Diederich B, Marsikova B, Reynaud EG, Holden S, Jahr W, Haase R, Prakash K. 2022. Open microscopy in the life sciences: quo vadis? *Nat Methods* **19**:1020–1025. doi:10.1038/s41592-022-01602-3

Holm T, Klein T, Löschberger A, Klamp T, Wiebusch G, van de Linde S, Sauer M. 2014. A Blueprint for Cost-Efficient Localization Microscopy. *ChemPhysChem* **15**:651–654. doi:10.1002/cphc.201300739

Kwakwa K, Savell A, Davies T, Munro I, Parrinello S, Purbhoo MA, Dunsby C, Neil MAA, French PMW. 2016. easySTORM: a robust, lower-cost approach to localisation and TIRF microscopy. *Journal of Biophotonics* **9**:948–957. doi:10.1002/jbio.201500324

Lavis LD. 2021. What if we just give everything away? *eLife* **10**:e74981. doi:10.7554/eLife.74981

Li H, Krishnamurthy D, Li E, Vyas P, Akireddy N, Chai C, Prakash M. 2020. Squid: Simplifying Quantitative Imaging Platform Development and Deployment. doi:10.1101/2020.12.28.424613

Li H, Soto-Montoya H, Voisin M, Valenzuela LF, Prakash M. 2019. Octopi: Open configurable high-throughput imaging platform for infectious disease diagnosis in the field. doi:10.1101/684423

Li JH, Santos-Otte P, Au B, Rentsch J, Block S, Ewers H. 2020. Directed manipulation of
membrane proteins by fluorescent magnetic nanoparticles. *Nat Commun* **11**:4259.
doi:10.1038/s41467-020-18087-3
Martens KJA, van Beljouw SPB, van der Els S, Vink JNA, Baas S, Vogelaar GA, Brouns SJJ,
van Baarlen P, Kleerebezem M, Hohlbein J. 2019. Visualisation of dCas9 target search in
vivo using an open-microscopy framework. *Nat Commun* **10**:3552. doi:10.1038/s41467-
019-11514-0
Moreno XC, Al-Kadhimi S, Alvelid J, Bodén A, Testa I. 2021. ImSwitch: Generalizing
microscope control in Python. *Journal of Open Source Software* **6**:3394.
doi:10.21105/joss.03394
Nieuwenhuizen RPJ, Lidke KA, Bates M, Puig DL, Grünwald D, Stallinga S, Rieger B. 2013.
Measuring image resolution in optical nanoscopy. *Nat Methods*. doi:10.1038/nmeth.2448
Ovesný M, Křížek P, Borkovec J, Švindrych Z, Hagen GM. 2014. ThunderSTORM: a
comprehensive ImageJ plug-in for PALM and STORM data analysis and super-resolution
imaging. *Bioinformatics* **30**:2389–2390. doi:10.1093/bioinformatics/btu202
Saalfeld S, Pisarev I, Hanslovsky P, Champion A, Rueden C, Bogovic J, Kittisopikul M,
jakirkham. 2022. saalfeldlab/n5: n5-2.5.1. doi:10.5281/zenodo.6578232
Schröder D, Deschamps J, Dasgupta A, Matti U, Ries J. 2020. Cost-efficient open source laser
engine for microscopy. *Biomed Opt Express, BOE* **11**:609–623. doi:10.1364/BOE.380815
Sharkey JP, Foo DCW, Kabla A, Baumberg JJ, Bowman RW. 2016. A one-piece 3D printed
flexure translation stage for open-source microscopy. *Review of Scientific Instruments*
87:025104. doi:10.1063/1.4941068
Sofroniew N, Lambert T, Evans K, Nunez-Iglesias J, Bokota G, Winston P, Peña-Castellanos G,
Yamauchi K, Bussonnier M, Doncila Pop D, Can Solak A, Liu Z, Wadhwa P, Burt A,
Buckley G, Sweet A, Migas L, Hilsenstein V, Gaifas L, Bragantini J, Rodríguez-Guerra J,
Muñoz H, Freeman J, Boone P, Lowe A, Gohlke C, Royer L, PIERRÉ A, Har-Gil H,
McGovern A. 2022. napari: a multi-dimensional image viewer for Python.
doi:10.5281/zenodo.7276432
Vera RH, Schwan E, Fatsis-Kavalopoulos N, Kreuger J. 2016. A Modular and Affordable Time-
Lapse Imaging and Incubation System Based on 3D-Printed Parts, a Smartphone, and
Off-The-Shelf Electronics. *PLOS ONE* **11**:e0167583. doi:10.1371/journal.pone.0167583
Voigt FF, Kirschenbaum D, Platonova E, Pagès S, Campbell RAA, Kastli R, Schaettin M, Egolf
L, van der Bourg A, Bethge P, Haenraets K, Frézel N, Topilko T, Perin P, Hillier D,
Hildebrand S, Schueth A, Roebroeck A, Roska B, Stoeckli ET, Pizzala R, Renier N,
Zeilhofer HU, Karayannis T, Ziegler U, Batti L, Holtmaat A, Lüscher C, Aguzzi A,
Helmchen F. 2019. The mesoSPIM initiative: open-source light-sheet microscopes for
imaging cleared tissue. *Nat Methods* **16**:1105–1108. doi:10.1038/s41592-019-0554-0
Walzik MP, Vollmar V, Lachnit T, Dietz H, Haug S, Bachmann H, Fath M, Aschenbrenner D,
Abolpour Mofrad S, Friedrich O, Gilbert DF. 2015. A portable low-cost long-term live-
cell imaging platform for biomedical research and education. *Biosensors and
Bioelectronics* **64**:639–649. doi:10.1016/j.bios.2014.09.061
Wang H, Lachmann R, Marsikova B, Heintzmann R, Diederich B. 2022. UCsim2: two-
dimensionally structured illumination microscopy using UC2. *Philosophical Transactions
of the Royal Society A: Mathematical, Physical and Engineering Sciences* **380**:20200148.
doi:10.1098/rsta.2020.0148
Wenzel T. 2023. Open hardware: From DIY trend to global transformation in access to
laboratory equipment. *PLOS Biology* **21**:e3001931. doi:10.1371/journal.pbio.3001931
630

a.

Cubes, 3D printed parts & optics

Stages | Laser | LED

Electronics

UC2-Rest

ImSwitch GUI

b.

i.

ii.

iii.

iv.

iv.

iv.

iv.

z-stage

c.

iv.

d.

e.

ii.

z-stage

iii.

i.

	0 min	75 min	150 min	300 min
Fluorescence				
Bright-field				
Merged				

	0 h	8 h	16 h	24 h	32 h
SYTO far-red					
Brightfield					
Merge					