Spatial redundancy transformer for self-supervised fluorescence image denoising

Xinyang Li[1,2,3†], Xiaowan Hu[2†], Xingye Chen[1,3†], Jiaqi Fan[2,4], Zhifeng Zhao[1,3], Jiamin Wu[1,3,5,6*], Haoqian Wang[2,7*] & Qionghai Dai[1,3,5,6*]

[1]*Department of Automation, Tsinghua University, Beijing 100084, China.*

[2]*Tsinghua Shenzhen International Graduate School, Tsinghua University, Shenzhen, 518055, China.*

[3]*Institute for Brain and Cognitive Sciences, Tsinghua University, Beijing 100084, China.*

[4]*Department of Electronic Engineering, Tsinghua University, Beijing 100084, China.*

[5]*Beijing Key Laboratory of Multi-dimension & Multi-scale Computational Photography (MMCP), Tsinghua University, Beijing 100084, China.*

[6]*IDG/McGovern Institute for Brain Research, Tsinghua University, Beijing, China.*

[7]*The Shenzhen Institute of Future Media Technology, Shenzhen 518071, China*

[†]*These authors contributed equally to this work.*

*Correspondence: wujiamin@tsinghua.edu.cn; wanghaoqian@tsinghua.edu.cn; qhdai@tsinghua.edu.cn

Abstract

Fluorescence imaging with high signal-to-noise ratios has become the foundation of accurate visualization and analysis of biological phenomena. However, the inevitable photon shot noise poses a formidable challenge on imaging sensitivity. In this paper, we provide a spatial redundancy denoising transformer (SRDTrans) to remove noise from fluorescence time-lapse images in a self-supervised manner. First, a sampling strategy based on spatial redundancy is proposed to extract adjacent orthogonal training pairs, which eliminates the dependence on high imaging speed. Secondly, to break the performance bottleneck of convolutional neural networks (CNNs), we designed a lightweight spatiotemporal transformer architecture to capture long-range dependencies and high-resolution features at a low computational cost. SRDTrans can overcome the inherent spectral bias of CNNs and restore high-frequency information without producing over-smoothed structures and distorted fluorescence traces. Finally, we demonstrate the state-of-the-art denoising performance of SRDTrans on single-molecule localization microscopy and two-photon volumetric calcium imaging. SRDTrans does not contain any assumptions about the imaging process and the sample, thus can be easily extended to a wide range of imaging modalities and biological applications.

Introduction

The rapid development of intravital imaging techniques enables researchers to observe biological structures and activities at micrometer and even nanometer scales[1-4]. As an imaging method with great prevalence, fluorescence microscopy has contributed to the discovery of a

series of new physiological and pathological mechanisms due to its high spatiotemporal

resolution and molecular specificity[5-8]. The fundamental goal of fluorescence microscopy is to

obtain clean and sharp images containing sufficient information about the sample, which can

guarantee the accuracy of downstream analysis and support convincing conclusions. However,

limited by multiple biophysical and biochemical factors (*e.g.*, labeling concentration,

fluorophore brightness, phototoxicity, photobleaching, *etc.*), fluorescence imaging is conducted

in photon-limited conditions and the inherent photon shot noise degrades the image signal-to-

noise ratio (SNR) severely, especially in low-illumination and high-speed observations[9-11].

Various methods have been proposed to remove noise from fluorescence images.

Conventional denoising algorithms based on numerical filtering and mathematical

optimization suffer from unsatisfactory performance and limited applicability[12-14]. In the past

few years, deep learning has shown remarkable performance in image denoising[15-17]. After

iterative training on a dataset with ground truth (GT), deep neural networks can learn the

mapping between noisy images and their clean counterparts. Such a supervised manner

depends heavily on paired GT images[18-22]. When observing the activity of living organisms,

obtaining pixel-wise registered clean images is a great challenge because the sample often

undergoes fast dynamics. To alleviate this contradiction, some self-supervised methods have

been proposed for more applicable and practical denoising in fluorescence imaging[9, 23-29].

Among them, the first kind of methods rely on the similarity between adjacent frames[9,23-25]. But

when the sample changes very quickly or the imaging speed is too slow, the time-lapse data

cannot provide enough temporal redundancy. This is a common problem in volumetric imaging

since the volume rate decreases proportionally to the number of imaging planes. The

dissimilarity between adjacent frames will result in inferior performance and distorted

structures and fluorescence kinetics. The other kind of methods learn to denoise only from

spatially adjacent pixels in 2D frames[26-29]. However, without utilizing endogenous temporal

correlations, these methods perform poorly on time-lapse imaging. Therefore, to achieve better

denoising performance, the ability to simultaneously extract global spatial information and

long-range temporal correlations is essential, which is lacking in CNNs because of the locality

of convolutional kernels[30,31]. Moreover, the inherent spectral bias makes CNNs tend to fit

low-frequency features preferentially while ignoring high-frequency features, inevitably

producing over-smoothed denoising results[32, 33].

Here, we present the spatial redundancy denoising transformer (SRDTrans) to address

these dilemmas. On the one hand, a spatial redundancy sampling strategy is proposed to extract

three-dimensional (3D) training pairs from the original time-lapse data in two orthogonal

directions. This scheme has no dependence on the similarity between two adjacent frames, so

SRDTrans is applicable to very fast activities and extremely low imaging speed, which is

complementary to our previously proposed DeepCAD that leverages temporal redundancy[9,26].

On the other hand, we designed a lightweight spatiotemporal transformer network to fully

exploit long-range correlations. The optimized feature interaction mechanism allows our model

to obtain high-resolution features with a small number of parameters. Compared with classical

CNNs, the proposed SRDTrans has stronger abilities for global perception and high-frequency

maintenance, enabling the revelation of fine-grained spatiotemporal patterns that were

previously indiscernible. We demonstrate the superior denoising performance of SRDTrans on

two representative applications. The first one is single-molecule localization microscopy

(SMLM) with adjacent frames being random subsets of fluorophores[34,35]. The other one is two-

photon calcium imaging of large 3D neuronal populations with a volumetric speed as low as

0.3 Hz. Extensive qualitative and quantitative results indicate that SRDTrans can serve as a

fundamental denoising tool for fluorescence imaging to observe various cellular and subcellular

phenomena.

Results

Principle of SRDTrans. The self-supervised framework of SRDTrans is schematized in Fig.

1a. For spatial redundancy sampling, spatially adjacent sub-stacks are sampled by orthogonal

masks from the original low-SNR image stack. Each target is adjacent to the input stack

horizontally or vertically to exploit spatial correlations fully and isotropically. A simplified

implementation of sampling is depicted in Fig. 1b. Since the noise of adjacent pixels is

independent while the signals are closely correlated, the sub-stack filled by the central

pixels can be used as the training input and the other two spatially adjacent sub-stacks are used

as corresponding targets to optimize the network parameters (Supplementary Note 1).

Compared with other sampling methods[26-28], our sampling strategy is more effective and

comprehensive in preserving both spatial and temporal information (Supplementary Fig. 1). In

the inference stage, low-SNR stacks will be fed into pre-trained SRDTrans models without

spatial down-sampling. To overcome the locality of convolutional kernels, we constructed a

transformer network to capture endogenous non-local spatial features and long-range temporal

dependencies using the self-attention mechanism (Fig. 1c). The restoration of each pixel can

simultaneously integrate the temporal information of all frames and the spatial information of

all pixels, even if they are far from each other. Besides, the network does not contain any spatial

down-sampling module, allowing more high-frequency components to flow through the

network and avoiding the loss of spatial resolution. Furthermore, as the amount of data in

fluorescence time-lapse imaging is often very large, sometimes at the scale of petabyte, the

transformer network was designed to be as lightweight as possible to relieve the computational

burden of large-scale data processing. Compared with other transformer networks[36-39], our

architecture can achieve the best performance with more than one order of magnitude fewer

parameters (Supplementary Tab. 1 and Supplementary Tab. 2). The lightweight architecture of

SRDTrans also makes it easy to train a good model even with a small amount of training data

(*e.g.*, 500 frames, 490×490 pixels each frame), relieving the pressure of capturing large-scale

datasets. (Supplementary Fig. 2).

To demonstrate the predominance of our transformer network over CNNs, we generated

simulated calcium imaging data (Method and Supplementary Fig. 3) and used them to train a

3D-UNet[16], as well as the transformer of SRDTrans, using the same spatial redundancy

sampling strategy. We term the former as spatial redundancy denoising CNN (SRDCNN). The

visualized feature maps of a typical deep layer intuitively show the superiority of SRDTrans

in revealing fine-grained patterns (Fig. 1d). As features flow through the network, the limited

receptive field of convolutional kernels makes CNN-based methods only focus on rough

features while our transformer architecture still has a strong perception of sophisticated

structures. We also compare the denoised images of the two architectures (Fig. 1e). The result

of SRDCNN is obviously over-smoothed, especially in regions with sharp edges, which is a

manifestation of spectral bias that the low-frequency information is overfitted while the high-

frequency information is hardly preserved (Supplementary Fig. 4). This deficiency is largely

alleviated by SRDTrans, and more subcellular structures such as dendritic fibers can be restored

accurately. The intensity profile deconstructed from the SRDTrans denoised image is more

consistent with GT (Fig. 1f). Moreover, lacking the ability to capture long-range temporal

correlations also drags down the denoising performance of SRDCNN on time-lapse imaging

data. However, for SRDTrans, the output SNR continuously grows as the input temporal scale

(T) increases (Fig. 1g). A more comprehensive evaluation of the influence of input temporal

scale suggests that SRDTrans can make full use of the information offered by temporally distant

pixels (Supplementary Fig. 5). To show the practicality of SRDTrans at extremely low imaging

speed, we compared the performance of methods combining different networks and sampling

schemes on simulated calcium imaging data sampled at 0.3 Hz (Supplementary Video 1).

When the similarity between adjacent frames is low, using spatial redundancy sampling is more

reasonable. Succinctly, the synergy between spatial redundancy sampling and dedicated

transformer architecture endows SRDTrans with the ability to restore high-resolution structures

and fast dynamics.

SRDTrans improves the performance of SMLM. Given N detected fluorescence photons,

the lower bound of the precision of SMLM scales to $1/\sqrt{N}$[34], which is the mathematical

formula of the shot-noise limit or the well-known standard quantum limit[40,41]. This indicates

that the fundamental physical limit of localization precision is shot noise. To investigate the

benefits that our denoising method can bring to SMLM, we first applied SRDTrans to simulated

stochastic optical reconstruction microscopy (STORM) data with GT for quantitative

evaluation[42,43]. The noise-free single-molecule emitting images were synthesized by

TestSTORM[44] and corresponding noisy images with different SNRs were then generated by

applying different levels of Mixed Poisson-Gaussian noise (Method and Supplementary Fig.

6). For the image stack of each SNR, we trained a specific model for it and then processed it

using the model to obtain the denoised image stack. Quantitative comparisons of the visualized

images (Fig. 2a and Supplementary Video 2) and the extracted intensity profiles (Fig. 2b) show

that the results of SRDTrans are highly consistent with GT. Over a wide range of imaging SNR,

including some extremely low-SNR conditions, SRDTrans can substantially improve the

image quality evaluated by SNR at a pixel-intensity level and structural similarity (SSIM) at a

perception level (Fig. 2c). Compared with other self-supervised methods[23-28], SRDTrans can

better preserve the distribution and intensity of emitters owing to its strong ability in exploiting

high-resolution features and long-range dependencies (Supplementary Fig. 7).

Next, we evaluate the improvement of single-molecule localization performance with the

enhancement of image SNR. To reconstruct super-resolved images, we applied

ThunderSTORM[45], one of the most frequently used localization software with an excellent

balance between accuracy and execution runtime[46,47]. Contaminated by noise, the image

reconstructed from the original noisy data contains many misidentified molecules (Fig. 2d).

By contrast, the reconstructed image from SRDTrans denoised images can reveal clear and

continuous cytoskeletal filaments that are not previously recognizable because of suppressed

localization error and improved resolution (Fig. 2e and Supplementary Fig. 8). For better

quantitative analysis, we matched detected fluorescent molecules with GT using the Hungarian

algorithm[46]. From raw images, few fluorescent molecules can be detected and most of them

are wrongly localized. After SRDTrans denoising, almost all molecules can be detected in high

agreement with GT (Fig. 2f and Supplementary Fig. 9). Using Jaccard index and root mean

squared error (RMSE) as the metrics to respectively quantify the proportion of correctly

detected molecules and the localization accuracy of those detected molecules, we found that

Jaccard index was improved by ~6-fold (85.7 % versus 14.1 %) and RMSE was reduced by

~3.4-fold (24.86 nm versus 85.94 nm) after denoising (Fig. 2g). From a more comprehensive

perspective, we further adopted the metric termed efficiency that combines Jaccard index and

RMSE[47]. The results show that SRDTrans improved the efficiency of single-molecule

localization from -21.54 to 71.33 (Fig. 2h), fully demonstrating the benefits of SRDTrans on

SMLM.

We further applied SRDTrans to experimental SMLM data of microtubules to validate its

ability in revealing subcellular structures. Raw frames were captured with low excitation power

and short exposure time to reduce phototoxicity and emitter density. The experimentally

obtained single-molecule emission images and SRDTrans denoised images are shown

in Fig. 3a. Disturbed by the noise, the reconstruction algorithm can hardly localize the

fluorescent molecules in the raw frames. The reconstructed super-resolution image contains

many erroneous spots and cannot reveal any meaningful structures (Fig. 3b). Whereas

SRDTrans can effectively suppress the noise and remove localization artifacts in the

reconstructed image, making the distribution and extension directions of microtubules visible.

We computed the Fourier ring correlation (FRC)[48,49] curve to quantify the resolution

from the SMLM images (Fig. 3c). The image resolution is defined as the inverse of the

spatial frequency at the intersection of the FRC curve and the threshold line (~0.143).

Benefitting from the removal of noise, the resolution of SRDTrans denoised data was

improved from 52.4 nm to 36.0 nm and the localization uncertainty was reduced from 8.0

nm to 5.0 nm (Fig. 3d). In addition to the data acquired by our instrument, we also used

SRDTrans to denoise publicly available SMLM data contributed by other laboratories[50] (Fig.

3e). The reconstructed super-resolution images indicate that SRDTrans can eliminate the

artifacts and bring more complete organelle structures (Fig. 3f,g). We applied Gaussian fitting

to the intensity profile perpendicular to the microtube filaments and measured the full width at

half maximum (FWHM) to quantify the image resolution (Fig. 3h). The SRDTrans denoised

data shows improved resolution as the averaged FWHM dropped from 187.89 nm to 60.96 nm

(Fig. 3i). As SMLM is heading towards live-cell imaging and long-term observation[51,52] our

denoising method promises to be a beneficial tool to reduce the laser power by resolving

fluorescent molecules from very low-SNR frames. For 3D SMLM, since the axial positions of

molecules are estimated through PSF engineering[34], SRDTrans is expected to offer great help

by resolving single-molecule emission patterns from low-SNR images.

SRDTrans enhances two-photon volumetric calcium imaging in the mouse cortex. In

multiphoton imaging, the volumetric imaging speed decreases linearly as the number of

scanning planes increases. Thus, the achievable sampling rate for observing neuronal

populations with large axial ranges is often quite low, making those denoising methods that rely

on the similarity between temporally adjacent frames infeasible[9,23-25]. However, SRDTrans

provides an opportunity to restore those highly degraded fluorescence signals in large-scale

volumetric calcium imaging by utilizing the similarity between spatially adjacent pixels. To

evaluate the denoising performance of SRDTrans on calcium imaging with different sampling

rates, we generated realistic calcium imaging data with synchronized GT using NAOMi[53]. We

started from denoising high-sampling-rate (30 Hz) data and found that SRDTrans can

effectively remove noise and recover previously indiscernible structures such as soma, neurites,

and vascular shadows (Fig. 4a, Supplementary Fig. 10, and Supplementary Video 3). The

enhancement is not only manifested in the visual effect, but more importantly, in the accurate

restoration of pixel intensities (Fig. 4b). Visualization in the frequency domain (calculated by

discrete Fourier transform) shows that SRDTrans can restore most of the frequency

components (Fig. 4c), especially those high-frequency components lost by DeepCAD[9,25] and

DeepInterpolation[24], thereby leading to the best performance that has ever been achieved in

denoising calcium imaging data (Supplementary Fig. 11). Such a remarkable denoising

capability can be maintained over a wide range of input SNRs (from -2.08 dB to 17.68 dB),

and the average SNR improvement is about 22 dB (Fig. 4d).

Then, we investigate the denoising performance of SRDTrans on calcium imaging data

sampled at 0.3 Hz, which is 100 times lower than the imaging speed demonstrated above.

Bilateral assessments both in the space domain and the frequency domain reveal that SRDTrans

can accurately retrieve the fluorescence signals from the original highly degraded images

without structural blurring and frequency deficiency (Fig. 4e and Supplementary Fig. 11).

When the sampling rate is much lower than the fluorescence dynamics, the large discrepancy

between the signals in two adjacent frames cannot provide the temporal correlation required by

DeepCAD and DeepInterpolation, so they are not accurate enough to be used in conditions of

low imaging speed or fast activity (Supplementary Fig. 12). To figure out how SRDTrans works

at different imaging speeds, we performed an ablation study on different sampling strategies

and network architectures (Fig. 4f and Supplementary Tab. 3). The results indicate that spatial

redundancy sampling performs better at low imaging speeds, whereas temporal redundancy

sampling performs better at high imaging speeds. Almost equally for all imaging speeds, our

transformer architecture offers an additional improvement (~2.05 dB on average) over

conventional CNNs. In general, the synergistic combination of the spatial redundancy sampling

and the transformer architecture in SRDTrans provides better performance than DeepCAD at

all imaging speeds (Supplementary Fig. 13). In the time domain, the superior ability of

SRDTrans can reveal high-fidelity calcium transitions without distorting fluorescence kinetics

(Fig. 4g and Supplementary Fig. 14). Moreover, we also simulated fast-moving objects to

imitate migrating cells that are widely existed in living organisms. The quantitative evaluation

shows that SRDTrans can preserve the structure of densely distributed objects even if they are

moving at a speed of up to 9 pixel/frame (Supplementary Fig. 15 and Supplementary Video 4),

alleviating the shortage of denoising methods for fast-moving cells and organelles.

Finally, we went a step further in denoising calcium imaging data by applying SRDTrans

to volumetric recordings, which is not achievable for other self-supervised denoising

methods[9,24-25] because of their heavy reliance on high sampling rates. We used transgenic mice

expressing the GCaMP6f genetically encoded calcium indicator[54] and imaged a large brain

volume of 500×500×200 μm in the mouse cortex using a two-photon microscope. We scanned

100 planes with a frame rate of 30 Hz, and thus the overall volume rate was 0.3 Hz. 3D

visualization of the neural volume shows that the spatial profiles and firing states of the neurons

can be revealed after denoising, which otherwise would be swamped by severe shot noise (Fig.

5a and Supplementary Video 5). For better comparison, we present the snapshots of a certain

imaging plane at two different moments. With the enhancement of SRDTrans, the structure and

distribution of the neurons become clearly observable (Fig. 5b). We also extracted the

fluorescence traces along the time axis and found that a large number of calcium transients can

be restored after denoising (Fig. 5c). The dramatically improved SNR would propel the

decoding of underlying neural activity from fluorescence signals. Since neural circuits in the

mammalian brain are spatially coordinated and temporally orchestrated, deciphering the

function of large neuronal ensembles requires large-scale volumetric imaging with high SNR.

The superior denoising performance of SRDTrans provides an opportunity to implement high-

sensitivity volumetric calcium imaging for investigating functionally concerted neurons and

recognizing circuit motifs, especially those distributed across multiple cortical layers.

Discussion

In this work, we proposed SRDTrans, a self-supervised denoising method for fluorescence

microscopy that achieves excellent performance on various time-lapse imaging data, even for

samples with very fast activities or microscopes with extremely low imaging speed. The spatial

redundancy sampling strategy eliminates the reliance on temporal redundancy, making

SRDTrans robust to image sequences with very low correlations between adjacent frames.

Benefiting from the spatiotemporal transformer architecture with global receptive fields,

SRDTrans can alleviate the inherent spectral bias in CNNs and restore more high-frequency

components. The strong capability of long-range perception also endows SRDTrans with better

performance than all previous methods. For demonstration, we focused on applications with

low temporal redundancy, such as SMLM of cytoskeletal filaments and two-photon calcium

imaging of large neural volumes, which verifies the effectiveness and superiority of SRDTrans

in visualizing and analyzing complex biological structures and functions. Additionally, our

method does not rely on any assumptions about the contrast mechanism, noise model, sample

dynamics, and imaging speed. Thus, it can be readily extended to other biological samples and

imaging modalities, such as membrane voltage imaging, single-protein detection, light sheet

microscopy, light field microscopy, and super-resolution microscopy[55-59].

As the development of fluorescent indicators heading towards faster kinetics to monitor

biological dynamics at the millisecond scale[60-62], the imaging speed required to record these

fast activities is continuously growing. Obtaining adequate sampling rates is becoming more

and more challenging for denoising methods relying on temporal redundancy. Our rationale is

to fill the niche by seeking to utilize spatial redundancy as an alternative to enable self-

supervised denoising in more imaging applications. Although the perfect case for spatial

redundancy sampling is that the spatial sampling rate is two times higher than the Nyquist

sampling of the diffraction limit to ensure that two adjacent pixels have nearly identical optical

signals, the endogenous similarity between two spatially downsampled sub-stacks is sufficient

to guide the training of the network. However, this does not mean that the proposed spatial

redundancy sampling strategy can fully replace temporal redundancy sampling, since the

ablation study (Fig. 3f) shows that, if equipped with the same network architecture, the temporal

redundancy sampling can achieve better performance in high-speed imaging. The superiority

of SRDTrans over DeepCAD at high imaging speeds actually derives from the transformer

architecture. In general, spatial redundancy and temporal redundancy are two complementary

sampling strategies to enable self-supervised training of denoising networks for fluorescence

time-lapse imaging. Using which sampling strategy depends on which kind of redundancy is

more sufficient in the data. It is noteworthy that there are still many cases where neither

redundancy is sufficient to support current sampling strategies. Developing specific or

more general self-supervised denoising methods is of persistent value for fluorescence

imaging.

Methods

The spatial redundancy sampling scheme. In SRDTrans, we employed a spatial redundancy sampling strategy to produce training pairs. The detailed implementation for generating training pairs in SRDTrans is shown in Fig. 1b and Supplementary Fig. 1d. For each image of an input training stack with H×W×T pixels, we spatially divided it into many adjacent small patches with 2×2 pixels. Next, we randomly selected three adjacent pixels from each patch. The central pixel was used to compose the input sub-stack and the two spatially adjacent pixels were used to compose the two target sub-stacks. After traversing all patches, we can finally obtain three downsampled sub-stacks with the size of H/2×W/2×T pixels. Since the fluorescence signals of spatially adjacent pixels are closely correlated, the input sub-stack and each target sub-stack can be considered as two independent samples of the same underlying pattern. Thus, the input sub-stack and the two target sub-stack can form two training pairs, which can be used for the self-supervised training of denoising networks (Supplementary Note 1).

Network architecture and loss function. The transformer architecture of SRDTrans is composed of three parts: a temporal encoder module, a spatiotemporal transformer block (STB), and a temporal decoder module (Fig. 1c). The temporal encoder module is equipped with two temporal encoders. Each temporal encoder can compress the temporal scale of the input sub-stack by a factor of r (r is 4 in this work) using a convolutional layer with 3×3 kernels. Different from those UNet-type architectures with many up-sampling and down-sampling operations, SRDTrans do not reduce the size of feature maps in these encoders (Supplementary

Fig. 16). Thus, for an input sub-stack with a size of H/2×W/2×T pixels, the output size of the

temporal encoder module is H/2×W/2×T/r^2. The output from the temporal encoder module will

be fed into the STB to extract global information both in space and time. In the STB, the input

is divided into patches with the size of p×p×T/r^2 (p is 7 in this work). These patches are first

flattened into one-dimensional vectors and input into the position embedding layer, where

spatial concatenation and linear transformation are performed. Two multi-head self-attention

(MSA) layers are then stacked to calculate the global attention between these vectors. Local

features flow fully in MSA layers and densely interact with long-range global features. Finally,

the output of the STB is remapped by the temporal decoder module, and its temporal scale can

be rescaled to T. This decompression operation is implemented by two cascaded temporal

decoders using convolutional layers with 3×3 kernels.

We used a linear combination of L1-norm loss and L2-norm loss as the loss function to

optimize the parameters of SRDTrans. We defined the input sub-stack filled with central

pixels as S_c, the target sub-stack filled with its vertically adjacent pixels as S_v, and the

target sub-stack filled with its horizontally adjacent pixels as S_h. The total loss consists

of two pairs of training losses, which is defined as:

$$
\begin{aligned}
L_{ver} &= \left\| \mathrm{F}_{SRDTrans}\left(S_i\right) - S_v \right\|_2^2 + \left| \mathrm{F}_{SRDTrans}\left(S_i\right) - S_v \right|_1, \\
L_{hor} &= \left\| \mathrm{F}_{SRDTrans}\left(S_i\right) - S_h \right\|_2^2 + \left| \mathrm{F}_{SRDTrans}\left(S_i\right) - S_h \right|_1, \\
L_{total} &= L_{ver} + L_{hor}.
\end{aligned}
$$

where L_{ver} and L_{hor} denote the loss of the vertically and horizontally adjacent sub-stacks,

respectively.

Training and inference. To achieve optimal performance, specific models were trained for

stacks with different SNRs. One or more training stacks were divided into a specified number

of 3D (xy-t) training pairs (6000 by default). The batch size for all experiments was set to the

number of GPUs (NVIDIA GeForce RTX 3090) being used. The patch size was set to

128×128×128 pixels. All extracted training pairs were geometrically transformed by random

flipping or rotation for 8-fold data augmentation. The compression factor of each temporal

encoder was set to 4. In the STB, we set the internal patch size to 7, the number of heads in the

multi-headed self-attention block to 8, and the embedded feature channels to 128. For model

optimization, we used the Adam optimizer and the exponential decay rate for the first moment

was 0.9, the exponential decay rate for the second moment was 0.999, and the learning rate was

0.00001. PyTorch was used to construct the network and implement all operations. In the

interference stage, the model of the last training epoch was selected for final processing. The

denoised result of each image stack was saved as a separate TIF file.

Data simulation. Quantitative evaluations were performed on simulated data because noise-

free images (ground truth) are available. To synthesize noise-free two-photon calcium imaging

data, we used the Neural Anatomy and Optical Microscopy (NAOMi)[53], which can generate

realistic calcium imaging data with high-fidelity tissue characteristics and fluorescence kinetics.

Then we applied different levels of Mixed Poisson Gaussian (MPG) noise to generate calcium

imaging data of different imaging SNRs[9,25]. We first synthesized calcium imaging data sampled

at 30 Hz and 1 Hz. The data of other sampling rates were obtained by extracting frames at

different intervals. The image size for all simulated calcium imaging data was 490×490 pixels

and the pixel size was 1.02 μm.

To generate realistic SMLM data, we first acquired reconstructed super-resolution images

from the ShareLoc.XYZ platform (https://shareloc.xyz/)[50]. These images were experimentally

obtained on a Nikon N-STORM microscope and contained densely distributed microtubules

immuno-labeled with Alexa-647[63]. The tracks of all microtubules in a selected ROI (region of

interest) were extracted semi-automatically using the JFilament plugin of ImageJ[64]. We then

generated synthetic single-molecule emitting image stacks (ground-truth images without noise)

using the TestSTORM[44] simulator by loading the microtubule patterns from JFilament. All

fluorescent molecules were linked on the microtubule pattern with a radius of 12.5 nm. For

imaging parameters, the numerical aperture was 1.4 and the frame rate was 200 Hz (5-ms

exposure time). The image size was 328×328 pixels and the pixel size was 30 nm. Noisy stacks

were generated by applying MPG noise post hoc with a customized Matlab script[9,25].

We synthesized moving objects with different moving speeds using the MNIST

dataset[65,66]. Each frame was defined as an image of 512×512 pixels with a black

background and many bright moving digits. Each digit was an image patch (28×28 pixels)

randomly extracted from the MNIST dataset, moved in a random direction, and

appeared or disappeared only once. The total number of digits in the field of view was

500. We first generated the data with a moving speed of 0.5 pixel/frame. Those data

with higher moving speeds were then generated by extracting frames at different

intervals. The total frame number for all moving speeds was 5000. The final experiment

was implemented on twenty datasets with moving speeds from 0.5 to 10 pixel/frame. The

sampling interval of the moving speed was 0.5 pixel/frame.

SMLM imaging. The imaging samples (including the buffer solution and the sample holder)

for the SMLM experiments were purchased from Standard Imaging (Beijing) Biotechnology

Co., Ltd (https://www.standardimaging.cn/standardsample?lang=en). The SMLM

experiments were performed on a commercial microscope (Nikon N-STORM) equipped with

laser sources of 405 nm and 640 nm, which were used for activation and excitation, respectively.

An sCMOS camera (Hamamatsu Flash 4.0) was placed on the image plane to capture the

emission signals. To mimic living-cell imaging, we used low excitation power to reduce

phototoxicity and short exposure time to obtain images with low emitter density. The detailed

settings are summarized in Supplementary Tab. 5.

SMLM sample preparation. The Biologics Standards-Cercopithecus-1 (BSC-1) cell line

purchased from Pricella Life Technology Co. LTD was used for SMLM imaging. BSC-1 cells

were cultured in DMEM (Invitrogen, #11965-118) supplemented with 10% fetal bovine serum

(FBS) (Gibco, #16010-159). To prevent bacterial contamination, 100 μg/ml penicillin and

streptomycin (Invitrogen, #15140122) were added into the DMEM medium. Cells were grown

under standard cell culture conditions (5% CO_2, humidified atmosphere at 37°C). BSC-1 cells

were plated on #1.5 glass-bottom dishes over 48 h before sample preparation. For cell passage,

cells were washed with pre-warmed PBS (Life Technologies, #14190500BT) 3 times and

digested with 25% trypsin (Gibco, #25200-056) for 30 s. BSC-1 l cell lines were tested for

potential mycoplasma contamination (MycoAlert, Lonza), and all tests showed negative results.

For immunofluorescence staining, cells were grown on 35 mm, #1.5 glass coverslips. We pre-

treated glass-bottom dishes with fibronectin (Invitrogen, #33016015) for 1 h at 37°C to increase

cell adhesion. On the day of sample preparation, the cell density should be about 50%-70%.

Cells were fixed with 37°C pre-warmed fixation buffer for 10 min, containing 4%

paraformaldehyde (EMS) and 0.1% glutaraldehyde in PBS. Then the sample was washed three

times with PBS. For quenching the background fluorescence, we incubated the cells with 2 ml

0.1% NaBH4 solution in PBS for 7 min, optionally shaking on the shaker (<1 Hz). The sample

was washed three times with 2 ml PBS and then incubated for 30 min in PBS containing 5%

BSA (Jackson, #001-000-162) and 0.5% Triton X-100 (Fisher Scientific) at 37°C. All

antibodies were diluted in the 5% BSA+ 0.5% triton solution described above. Next, we

incubated the sample for 40 min with the appropriate dilution of primary antibodies: beta-

tubulin (E7, DSHB) at 25°C. After primary antibodies incubation, the cells were washed three

times with 2 ml PBS for 5 min. Secondary antibodies were incubated for 60 min with the

appropriate dilutions of secondary antibodies (Jackson Immuno Research, conjugated with

Cy5 in our laboratory) at 25°C. After being washed 3 times with PBS, cells were fixed with

post-fixation buffer for 10 min. The sample was stored at 4°C in PBS and protected from light.

Before imaging, we used an imaging buffer (Standard Imaging (Beijing) Biotechnology Co.,

Ltd, #ST031) to replace PBS.

SMLM reconstruction. The super-resolution SMLM images were reconstructed by the Fiji

plugin ThunderSTORM[45]. The detailed configuration is set as: the image filter was the

wavelet filter (B-spline) with the order of 3 and the scale of 2; the algorithm for determining

the approximate position of molecules was the local maximum algorithm; the sub-pixel

localization is performed by the integrated Gaussian PSF model with a fitting radius of 3 pixels;

a fitting method of "weighted least squares", and an initial sigma of 1.6 pixels. Both

visualization images are generated by averaged shifted histogram with a magnification of 5.

Mouse preparation and calcium imaging. All experiments involving mice were performed

in accordance with the institutional guidelines for animal welfare and have been approved by

the Animal Care and Use Committee of Tsinghua University. Transgenic mice hybridized

between Rasgrf2-2A-dCre mice and Ai148 (TIT2L-GC6f-ICL-tTA2)-D mice expressing Cre-

dependent GCaMP6f genetically encoded calcium indicator (GECI) were used for calcium

imaging of neural circuits. Craniotomy surgeries were conducted to remove the skull and a

coverslip was implanted on the craniotomy region for chronic imaging. Two-photon volumetric

imaging of the mouse cortex was performed on head-fixed mice without anesthesia using a

standard two-photon microscope. The neural volume being recorded was located at the primary

visual cortex with a depth of about 150~350 μm below the dura, and was scanned for 100

planes with an axial step of 2 μm. The whole imaging session lasted 30 minutes with a volume

rate of 0.3 Hz.

Three-dimensional visualization of neural activity. For volumetric calcium imaging, we

used Imaris 9.0 (Oxford Instruments) to visualize the calcium activity of the neuronal

population before and after denoising. Specifically, we imported the 4D (xyz-t) data into Imaris,

applied pseudo-color to the images, and then performed 3D rendering using the maximum

intensity projection mode. The contrast and brightness were adjusted to make structures in the

volume as clear as possible. All values for gamma correction were set to one.

Method comparison. We compared the performance of SRDTrans with six baseline

self-supervised methods, including Noise2Noise[23], Noise2Void[26], Noise2Self[27],

Neighbor2Neighbor[28], DeepInterpolation[24], and DeepCAD[9,25]. These methods were all

implemented by open-source codes released by relevant papers. The denoising model

of each method was trained and tested on the same datasets. For the methods designed

for 2D images, we split the time-lapse (xy-t) image stack into a series of 2D frames to

match the input dimension. Training and inference were performed frame-by-frame. We

followed the default training settings about network architectures and hyperparameters

for all methods. Specifically, the model of DeepInterpolation was fine-tuned based on

a public pre-trained model (pre-trained with 225,000 two-photon images of the Ai93

reporter line). Other methods were trained from scratch. The detailed settings of each

method are listed in Supplementary Tab. 4.

Evaluation metrics. We used several metrics to evaluate the performance of different

denoising methods. For an image (or an image stack) S_x and its ground truth S_y, the

metrics are defined as follows:

Signal-to-noise ratio (SNR) measures the pixel-level deviation between two images

using the logarithmic decibel scale, which is formulated as

$$SNR = 10 \log_{10} \frac{\left\| S_y \right\|_2^2}{\left\| S_x - S_y \right\|_2^2}.$$

468   *Structural similarity (SSIM)* measures the similarity between two images on a

469   perceptual level, including luminance, contrast, and structure. The definition is

470
$$SSIM = \frac{(2\mu_x\mu_y + c_1)(2\sigma_{xy} + c_2)}{(\mu_x^2 + \mu_y^2 + c_1)(\sigma_x^2 + \sigma_y^2 + c_2)},$$

where $\{\mu_x, \mu_y\}$ and $\{\sigma_x, \sigma_y\}$ are the means and variances of S_x and S_y, respectively. σ_{xy} is

the covariance of S_x and S_y. The two constants c_1 and c_2 are defined as $c_1=(k_1L)^2$ and

$c_2=(k_2L)^2$ with k_1=0.01, k_2=0.03, and L=65535.

Jaccard index measures the proportion of correctly detected molecules in SMLM.

Those correctly localized fluorescent molecules are true positives (TP). Those

incorrectly localized molecules are false positives (FP) and those undetected molecules

are false negatives (FN). The Jaccard index is formulated as:

478
$$Jaccard = 100\frac{TP}{TP+FP+FN}\% .$$

479   *Root mean squared error (RMSE)* quantifies the mean difference between the

480   localized positions (P_x) and ground-truth positions (P_y) of all detected fluorescent

481   molecules:

482
$$RMSE = \sqrt{\frac{1}{TP}\sum_{TP}\left\|P_y - P_x\right\|_2^2},$$

483   *Efficiency (E)* is a comprehensive metric combining the Jaccard index and RMSE

484   to measure the performance of single-molecule localization[43]. It can simultaneously

485   reflect the ability to detect molecules from images (measured by Jaccard) and the ability

486   to precisely locate molecules (measured by RMSE), which is defined as:

487
$$Efficiency = 100 - \sqrt{(100 - JAC)^2 + \alpha^2 RMSE^2},$$

488   where α=1 nm^{-1} controls the trade-off between Jaccard and RMSE.

489   *Pearson correlation coefficient* measures the similarity between a variable (images

490   and fluorescence traces) and its ground truth, which is formulated as

491
$$R = \frac{E\left[(S_x - \mu_x)(S_y - \mu_y)\right]}{\sigma_x \sigma_y},$$

492   where E represents the arithmetic mean. $\{\mu_x, \mu_y\}$ and $\{\sigma_x, \sigma_y\}$ are the means and

493   variances of S_x and S_y, respectively.

494   *Logarithmic frequency distance (LFD)* quantifies the spectral difference between

495   two images in the frequency domain. For images with a size of H×W pixels, LFD is

496   formulated as:

497
$$d(u,v) = \left\| F_{S_x}(u,v) - F_{S_y}(u,v) \right\|^2,$$
$$LFD = \log_{10}\left[\frac{1}{HW} (\sum_{h=0}^{H-1}\sum_{w=0}^{W-1}|d(u,v)|) + 1 \right].$$

F_{Sx} and F_{Sx} are the discrete Fourier transform of S_x and S_y, respectively.

Statistics and reproducibility. To ensure the reproducibility of the findings, we report

the sample size and statistics in the legend and text of each experiment. All boxplots

are drawn in the standard Tukey boxplot format. The upper and lower quartiles are

represented by boxes, and the lines in the boxes indicate the median. The lower whisker

represents the minimum observed value, equal to the lower quartile minus 1.5× the

interquartile range. The upper whisker the maximum observed value, equal to the upper

quartile plus 1.5× the interquartile range. Results obtained through experimental or

observational studies or statistical analysis of data sets can be reproduced with high

reliability when the study is repeated. In the figure shown are representative frames,

and all experiments obtained similar results on more than 1,000 frames.

Data availability

We have no restrictions on data availability. Both the simulated and experimental data of single-

molecule localization microscopy and two-photon calcium imaging in this work are available

at https://github.com/cabooster/SRDTrans/tree/main/datasets.

Code availability

The source code and pre-trained models are publicly available at

https://github.com/cabooster/SRDTrans.

Acknowledgements

This research was supported by the National Natural Science Foundation of China (62088102,

62222508, 62071272) and National Key Research and Development Program of China

(Project No. 2022YFB36066), in part by the Shenzhen Science and Technology Project under

Grant (CJGJZD20200617102601004, JCYJ20220818101001004). This work was also

supported by the China Nation Postdoctoral Program for Innovative Talents (BX2021159) and

Shuimu Tsinghua Scholar Program. We thanked Dr. Huiwen Hao from Standard Imaging

(Beijing) Biotechnology Co., Ltd for providing information about SMLM sample preparation.

Author Contributions

Q.D., H.W. and J.W. supervised this research. Q.D., H.W., J.W. and X.L. conceived and

initiated this project. X.L., X.H. and X.C. designed detailed implementations and

performed imaging experiments. X.H. and X.L. developed the Python code, performed

simulations, and processed relevant imaging data. J.F. prepared samples and provided

models animals. Z.Z. gave critical support on the two-photon imaging system and

imaging procedures. X.H., X.L., and X.C. analyzed the data, prepared figures and

videos. X.L., X.H., X.C., J.F., Z.Z. and J.W. participated in discussions about the results

and gave valuable advice. All authors participated in the drafting of the manuscript.

Competing interests

The authors declare no competing interests.

Materials & Correspondence

Correspondence and requests for materials should be addressed to Jiamin Wu, Haoqian Wang,

or Qionghai Dai.

References

1. Royer, L. A. et al. Adaptive light-sheet microscopy for long-term, high-resolution imaging in living organisms. *Nat. Biotechnol.* **34**, 1267-1278 (2016).

2. Fan, J. et al. Video-rate imaging of biological dynamics at centimetre scale and micrometre resolution. *Nat. Photon.* **13**, 809-816 (2019).

3. Keller, P.J. & Ahrens, M.B. Visualizing whole-brain activity and development at the single-cell level using light-sheet microscopy. *Neuron* **85**, 462-483 (2015).

4. Balzarotti, F. et al. Nanometer resolution imaging and tracking of fluorescent molecules with minimal photon fluxes. *Science* **355**, 606-612 (2017).

5. Baruch, K. et al. PD-1 immune checkpoint blockade reduces pathology and improves memory in mouse models of Alzheimer's disease. *Nature medicine* **22**, 135-137 (2016).

6. Wu, J. et al. Iterative tomography with digital adaptive optics permits hour-long intravital observation of 3D subcellular dynamics at millisecond scale. *Cell* **184**, 3318–3332 (2021).

7. Verweij, F. J. et al. The power of imaging to understand extracellular vesicle biology in vivo. *Nat. Methods* **18**, 1013–1026 (2021).

8. Mulay, S.V. et al. Enhanced fluorescence turn‐on imaging of hypochlorous acid in living immune and cancer cells. *Chemistry－A European Journal* **22**, 9642-9648 (2016).

9. Li, X. et al. Real-time denoising enables high-sensitivity fluorescence time-lapse imaging beyond the shot-noise limit. *Nat. Biotechnol.* **1-11** (2022).

10. Casacio, C.A. et al. Quantum-enhanced nonlinear microscopy. *Nature* **594**, 201-206 (2021).

11. Brida, G., Genovese, M. & Ruo Berchera, I. Experimental realization of sub-shot-noise quantum imaging. *Nat.Photon* **4**, 227–230 (2010).

12. Meiniel, W., Olivo-Marin, J. C. & Angelini, E. D. Denoising of microscopy images: a review of the state-of-the-art, and a new sparsity-based method. *IEEE Trans. Image Process.* **27**, 3842–3856 (2018).

13. Dabov, K., Foi, A., Katkovnik, V. & Egiazarian, K. Image denoising by sparse 3-D transform-domain collaborative filtering. *IEEE Trans. Image Process.* **16**, 2080-2095 (2007).

14. Gu, S., Zhang, L., Zuo, W. & Feng, X. Weighted nuclear norm minimization with application to image denoising. In *Proc. of the IEEE/CVF conference on computer vision and pattern recognition* 2862-2869 (2014).

15. Zhang K. et al.. Beyond a gaussian denoiser: Residual learning of deep CNN for image denoising. *IEEE Trans. Image Process.* **26**, 3142-3155 (2017).

16. Mao, X., Shen, C. & Yang, Y.-B. Image restoration using very deep convolutional encoder-decoder networks with symmetric skip connections. In *Advances in Neural Information Processing Systems* **29** (2016).

17. Tai, Y., Yang, J., Liu, X. & Xu, C. Memnet: A persistent memory network for

image restoration. in *Proc.* In *Proc. of the IEEE/CVF conference on computer*
vision and pattern recognition 4539-4547 (2017).

18.　Weigert, M. et al. Content-aware image restoration: pushing the limits of
fluorescence microscopy. *Nat. Methods* **15**, 1090–1097 (2018).

19.　Belthangady, C. & Royer, L. A. Applications, promises, and pitfalls of deep
learning for fluorescence image reconstruction. *Nat. Methods* **16**, 1215–1225
(2019).

20.　Chen, J. et al. Three-dimensional residual channel attention networks denoise
and sharpen fluorescence microscopy image volumes. *Nat. Methods* **18**, 678–
687 (2021)

21.　Chaudhary, S., Moon, S. & Lu, H. Fast, efficient, and accurate neuro-imaging
denoising via supervised deep learning. *Nat Commun.* **13**, 5165 (2022)

22.　Wang, Z., Xie, Y. & Ji, S. Global voxel transformer networks for augmented
microscopy. *Nat. Mach. Intell.* **3**, 161–171 (2021).

23.　Lehtinen, J. et al. Noise2Noise: learning image restoration without clean data.
In *Proc. of the 35th International Conference on Machine Learning* (eds Dy, J.
& Krause, A.) 2965–2974 (PMLR, 2018).

24.　Lecoq, Jérôme, et al. Removing independent noise in systems neuroscience data
using DeepInterpolation. *Nat. Methods* **18**, 1401-1408 (2021).

25.　Li, X. et al. Reinforcing neuron extraction and spike inference in calcium
imaging using deep self-supervised denoising. *Nat. Methods* **18**, 1395-1400
(2021).

26.　Krull, A., Buchholz, T.-O. & Jug, F. Noise2Void—learning denoising from
single noisy images. In *Proc. of the IEEE/CVF conference on computer vision*
and pattern recognition 2129-2137 (2019).

27.　Batson, J. & Royer, L. Noise2Self: blind denoising by self-supervision. In *Proc.*
of the 36th International Conference on Machine Learning 524–533 (PMLR,
2019).

28.　Huang, T. et al. Neighbor2neighbor: Self-supervised denoising from single
noisy images. In *Proc. of the IEEE/CVF conference on computer vision and*
pattern recognition 14781-14790 (2021).

29.　Lequyer, J. et al. A fast blind zero-shot denoiser. *Nat. Mach. Intell.* **4**, 953–963
(2022).

30.　Luo, W. et al. Understanding the effective receptive field in deep convolutional
neural networks. In *Advances in Neural Information Processing Systems* **29**
(2016).

31.　Araujo, André, Wade Norris, and Jack Sim. Computing receptive fields of
convolutional neural networks." *Distill* **4**.11 (2019).

32.　Rahaman N. et al. On the spectral bias of neural networks. In *International*
Conference on Machine Learning 5301-5310 (PMLR, 2019).

33.　Khayatkhoei M, Elgammal A. Spatial frequency bias in convolutional
generative adversarial networks. In *Proc. of the AAAI Conference on Artificial*

Intelligence, Vol. **36** 7152-7159 (2022).

34.   Lelek, M. et al. Single-molecule localization microscopy. *Nat. Rev. Methods*
Primers **1**, 39 (2021).

35.   Speiser, A. et al. Deep learning enables fast and dense single-molecule
localization with high accuracy. *Nat. Methods* **18**, 1082–1090 (2021).

36.   Liu Z. et al. Video swin transformer. In *Proc. of the IEEE/CVF conference on*
computer vision and pattern recognition 3202-3211 (2022).

37.   Zhou H. et al. nnformer: Interleaved transformer for volumetric segmentation.
arXiv preprint arXiv:2109.03201 (2021).

38.   Hatamizadeh A. et al. Unetr: Transformers for 3d medical image segmentation.
In *Proc. of the IEEE/CVF winter conference on applications of computer vision*
(2022).

39.   Hatamizadeh A. et al. Swin unetr: Swin transformers for semantic segmentation
of brain tumors in mri images. In *7th International Workshop of Medical Image*
Computing and Computer-Assisted Intervention (2021).

40.   Taylor, M. A. & Bowen, W. P. Quantum metrology and its application in biology.
Phys. Rep. **615**, 1–59 (2016).

41.   T. Nagata. et al. Beating the standard quantum limit with four-entangled
photons. *Science* **316**, 726–729 (2007).

42.   Rust, M., Bates, M. & Zhuang, X. Sub-diffraction-limit imaging by stochastic
optical reconstruction microscopy (STORM). *Nat. Methods* **3**, 793–796 (2006).

43.   Elias Nehme, Lucien E. Weiss, Tomer Michaeli, and Yoav Shechtman. Deep-
STORM: super-resolution single-molecule microscopy by deep learning.
Optica **5**, 458-464 (2018).

44.   Sinkó J. et al. TestSTORM: Simulator for optimizing sample labeling and image
acquisition in localization based super-resolution microscopy. *Biomedical*
optics express **5**, 778-787 (2014).

45.   Ovesný, M. et al. ThunderSTORM: a comprehensive ImageJ plug-in for PALM
and STORM data analysis and super-resolution imaging. *Bioinformatics* **30**,
2389-2390 (2014).

46.   Sage, D. et al. Quantitative evaluation of software packages for singlemolecule
localization microscopy. *Nat. Methods* **12**, 717–724 (2015).

47.   Sage, D. et al. Super-resolution fight club: assessment of 2D and 3D single-
molecule localization microscopy software. *Nat. Methods* **16**, 387–395 (2019).

48.   Nieuwenhuizen, R., Lidke, K., Bates, M. *et al.* Measuring image resolution in
optical nanoscopy. *Nat. Methods* **10**, 557–562 (2013).

49.   Descloux, A., Grußmayer, K.S. & Radenovic, A. Parameter-free image
resolution estimation based on decorrelation analysis. *Nat. Methods* **16**, 918–
924 (2019).

50.   Ouyang, W. et al. ShareLoc — an open platform for sharing localization
microscopy data. *Nat. Methods* **19**, 1331–1333 (2022).

51.   Jones, S. et al. Fast, three-dimensional super-resolution imaging of live

cells. *Nat. Methods* **8**, 499–505 (2011).

52.   Godin A G, Lounis B, Cognet L. Super-resolution microscopy approaches for
live cell imaging. *Biophys. J.* **107**.8, 1777-1784 (2014).

53.   Song, A., Gauthier, J.L., Pillow, J.W., Tank, D.W. & Charles, A.S. Neural
anatomy and optical microscopy (NAOMi) simulation for evaluating calcium
imaging methods. *J. Neurosci. Methods* **358**, 109173 (2021).

54.   Chen, TW. et al. Ultrasensitive fluorescent proteins for imaging neuronal
activity. *Nature* **499**, 295–300 (2013).

55.   Huang, F., Schwartz, S. L., Byars, J. M. & Lidke, K. A. Simultaneous multiple-
emitter fitting for single molecule super-resolution imaging. *Biomed. Opt.*
Express **2**, 1377–1393 (2011).

56.   Platisa, J. et al. High-speed low-light in vivo two-photon voltage imaging of
large neuronal populations. *Nat. Methods* (2023).

57.   Zhao, W. et al. Sparse deconvolution improves the resolution of live-cell super-
resolution fluorescence microscopcy. *Nat. Biotechnol.* **40**, 606–617 (2022).

58.   Dahmardeh, M. et al. Self-supervised machine learning pushes the sensitivity
limit in label-free detection of single proteins below 10 kDa. *Nat. Methods* **20**,
442–447 (2023).

59.   Li, X. et al. Unsupervised content-preserving transformation for optical
microscopy. *Light. Sci. Appl.* **10**, 44 (2021).

60.   Zhang, Y. et al. Fast and sensitive GCaMP calcium indicators for imaging neural
populations. *Nature* **615**, 884–891 (2023).

61.   Knöpfel, T., Song, C. Optical voltage imaging in neurons: moving from
technology development to practical tool. *Nat. Rev. Neurosci.* **20**, 719–727
(2019).

62.   Liu Z. et al. Sustained deep-tissue voltage recording using a fast indicator
evolved for two-photon microscopy. *Cell* **185**, 3408-3425(2022).

63.   Jimenez A, Friedl K, Leterrier C. About samples, giving examples: optimized
single molecule localization microscopy. *Methods* **174** , 100-114 (2020).

64.   Smith M B. et al. Segmentation and tracking of cytoskeletal filaments using
open active contours. *Cytoskeleton* **67**, 693-705 (2010).

65.   LeCun, Y., Bottou, L., Bengio, Y. & Haffner, P. Gradient-based learning applied
to document recognition. In *Proc. IEEE* **86**, 2278-2324 (1998).

66.   He, Z. et al. Tracking by animation: Unsupervised learning of multi-object
attentive trackers. In *Proc. of the IEEE/CVF conference on computer vision and*
pattern recognition 1318-1327 (2019).

Fig. 1 | Principle of SRDTrans and performance evaluation. a, Self-supervised training strategy of SRDTrans. In a single instance of spatial redundancy sampling, the original low-SNR stack of H×W×T pixels is sampled by orthogonal masks, producing three downsampled sub-stacks (input, target 1, and target 2) of H/2×W/2×T pixels. The "input" sub-stack is fed into the transformer network, and the corresponding output is compared with the "target" sub-stacks to calculate the loss function for parameter optimization. **b**, Simplified schematic of spatial redundancy sampling (H=4, W=4, T=1). A 4×4 patch is split into four 2×2 blocks and three adjacent pixels are randomly selected in each block. The central pixel (labeled as "2") is horizontally or vertically adjacent to the other two pixels (labeled as "1" and "3"). Pixels with the same label are recombined into new 2×2 patches. **c**, The architecture of the lightweight spatiotemporal transformer. It consists of a temporal encoder module, a spatiotemporal transformer block (STB), and a temporal decoder module. Each temporal encoder compresses the temporal scale (t) of the input by a factor

of r (adjustable and set as 4 in this work) using convolution. In the STB module, the input is divided into small patches, and different feature maps of the same spatial position are stitched together in the patch flattening layer. The position embedding layer records the spatial position of each patch so that it can be mapped back after the global interaction in the multi-head self-attention layer. The self-attention mechanism can calculate the spatiotemporal correlation between all local patches. Then the output patches are remapped to their original spatial positions. The output of the STB module will be uncompressed to the original temporal scale by the following temporal decoder module. Multiple skip connections are configured to stabilize the training of the network. **d**, Visualizing the feature responses in deep layers of SRDCNN and SRDTrans. SRDCNN represents the method that replaces the transformer network in SRDTrans with a 3D-UNet. Scale bar, 60 μm. **e**, Comparing the denoising performance of SRDCNN and SRDTrans on simulated calcium imaging data (30 Hz). Magnified views of boxed regions are shown at the top left of the images. Purple arrowheads point to a dendritic fiber. Scale bar, 40 μm for the whole field of view (FOV) and 10 μm for magnified views. **f**, Pixel intensity along the red dashed line in **e**. **g**, Evaluating the ability of SRDCNN and SRDTrans to capture long-range dependencies. Models were trained and validated on simulated calcium imaging data (30 Hz) of different input temporal scales (T).

Fig. 2 | Validation of SRDTrans on simulated SMLM data. a, Single-molecule emission images before
and after denoising. Left, raw data. Middle, SRDTrans denoised data. Right, GT. Magnified views of boxed
regions show the emission pattern of a bunch of fluorescent molecules. Scale bars, 2 μm for the whole FOV
and 0.5 μm for magnified views. The SNR value of the raw data and denoised are noted. **b**, Intensity profiles
along the white dashed line in **a**. **c**, Quantitative evaluation of the denoising performance with SNR and
SSIM. Left, image SNR before and after denoising. Right, image SSIM before and after denoising. Each
data point is the statistical result of 24000 frames. **d**, Reconstructed super-resolution images of microtubules
before and after denoising. Left, the image reconstructed from raw data. Middle, the image reconstructed

from SRDTrans denoised data. Right, GT. Scale bar, 5 μm. **e**, Merged image of the yellow boxed region in
d. Magenta, the image reconstructed from raw data. Green, the image reconstructed from SRDTrans
denoised data. Red, GT. The overlapping positions of red and green appear yellow. Scale bar, 1 μm. **f**,
Consistency analysis of the localized fluorescent molecules in raw images (top) and SRDTrans denoised
images (bottom) referring to GT. A magnified view of the boxed region is shown at the bottom left of each
panel. **g**, Boxplots showing the localization precision quantified with the Jaccard index (left, higher is better)
and RMSE (right, lower is better) before and after SRDTrans denoising (N=5000). **h**, Evaluating the
performance of single-molecule localization before (blue) and after (yellow) denoising with a more
comprehensive metric termed efficiency.
746

Fig. 3 | SRDTrans enhances experimental SMLM imaging. a, Experimentally obtained single-molecule
emission images. Left, raw data. Right, SRDTrans denoised data. The magnified views of two boxed
regions are shown below each image. Scale bars, 2 μm for the whole FOV and 0.5 μm for magnified views.
b, Reconstructed super-resolution images. The microtubules in fixed BSC-1 cells were labeled with Cy5.

Scale bars, 2 μm for the whole FOV and 0.5 μm for magnified views. **c,** Fourier-ring correlation (FRC) curves of the raw reconstructed image (blue) and the SRDTrans enhanced reconstructed image (yellow) in **b**. The estimated resolution (52.4 nm for raw image and 36.0 nm for SRDTrans denoised image) is the inverse of the spatial frequency where the FRC curve drops below the cutoff threshold (~0.143). **d,** Boxplots showing the localizing uncertainty before and after denoising (N=1048575 for raw data, N=395908 for SRDTrans). The uncertainty was calculated by the ThunderSTORM plugin. **e,** Single-molecule emission images from the open-source platform ShareLoc. Left, raw data. Right, SRDTrans denoised data. Scale bars, 2 μm. **f,** Reconstructed super-resolution images of microtubules (immuno-labeled with Alexa 647). Scale bars, 2 μm. **g**, Magnified views of boxed regions in **f**. Scale bars, 0.5 μm. **h**, Intensity profiles perpendicular to the microtubule filaments as indicated in **f** (blue lines: raw data, yellow lines: SRDTrans denoised data, dashed line: the Gaussian fitting result). The corresponding FWHM value is quantified as 2.335σ, where σ denotes the standard deviation of Gaussian fitting results. **i**, Boxplots showing the FWHM of randomly selected filaments from **e** (blue: raw data, yellow: SRDTrans denoised data, N=80).

Fig. 4 | Evaluating the performance of SRDTrans on simulated calcium imaging data. a, SRDTrans denoised calcium imaging data sampled at 30 Hz. Magnified views show the neural activity of yellow boxed regions in a short period (~2 s). Left, the original low-SNR data. Middle, SRDTrans denoised data. Right, GT. Scale bars, 60 μm for the whole FOV and 30 μm for magnified views. The magenta arrowhead indicates a dendritic fiber and the yellow arrowhead indicates two neighboring somas. **b**, Pixel intensity along the yellow dashed line in **a**. Top left, raw data. Middle left, SRDTrans denoised data. Bottom left, GT. Right, plotting the three intensity profiles in one coordinate. The similarity with GT is quantified by Pearson correlation coefficients (R). **c**, Frequency spectrum calculated by discrete Fourier transform before and after denoising. Magnified views of the boxed regions show the frequency components within the optical transfer function. The similarity in the frequency domain is quantified by logarithmic frequency distance (LFD). **d**, The performance of SRDTrans at different SNR levels (averaged on 6000 frames). **e**. Comparing the performance of DeepCAD and SRDTrans on calcium imaging data sampled at 0.3 Hz. Magnified views show the neural activity of yellow boxed regions in a 20-s time window. Scale bars, 100 μm for the whole FOV and 40 μm for magnified views. The yellow and purple arrowheads point to a firing neuron and a resting neuron, respectively. **f**, Ablation experiments to investigate the effects of different sampling strategies and network architectures. A wide range of imaging speeds (from 0.1 Hz to 30 Hz) are surveyed and summarized in a logarithmic coordinate. SRDTrans (yellow) uses spatial redundancy sampling and a lightweight spatiotemporal transformer. DeepCAD (purple) combines temporal redundancy sampling and a CNN (3D-UNet). SRDCNN (green) is the method combining spatial redundancy sampling and a CNN (3D-UNet). The input patch size is 128×128×128 pixels for all methods. **g**, Fluorescence traces extracted from 50 randomly selected neuronal pixels. The similarity with GT is quantified by Pearson correlation coefficients (R). Top, traces extracted from raw data. Middle, traces extracted from DeepCAD denoised data. Middle bottom, traces extracted from SRDTrans denoised data. Bottom, GT.

Fig. 5 | High-sensitivity calcium imaging of large neural volumes. a, 3D visualization of the neural activity of a 510×510×200 μm volume (100 planes, 0.3 Hz volume rate) in the mouse cortex. Left, the original low-SNR volume. Right, the same volume denoised with SRDTrans. Magnified views of yellow boxed regions are shown under each snapshot. Scale bars, 100 μm for the whole FOV and 10 μm for magnified views. **b**, Raw frames and SRDTrans denoised counterparts of a single imaging plane at two different moments. The x-z and y-z cross-sections of the volume indicated by the dashed lines are shown beside each x-y plane. Magnified views of yellow boxed regions are shown at the bottom right of the images. Scale bars, 70 μm for the whole FOV and 10 μm for magnified views. **c**, Fluorescence traces extracted from all pixels on the yellow dashed line in **b**, Left, traces extracted from raw data. Right, SRDTrans denoised traces. Yellow arrowheads point to some representative calcium transients.

a Spatial redundancy sampling

Images — Orthogonal masks — Training pairs

Mask 1 Mask 2 Mask 3

Target 1 — Learning — Network — Output — Learning — Target 2

Input

b

4×4 Patch — Mask ID — Orthogonal mask

Mask 1 Mask 2 Mask 3

2×2 Patches

c

STB

Temporal encoder Temporal decoder

Spatiotemporal transformer block (STB)

Patch flattening

Position embedding

Multi-head self-attention

d

Raw data — SRDCNN — SRDTrans

(norm.)

e

Raw data — SRDCNN — SRDTrans — GT

f

Distance (μm)

Normalized intensity (AU)

g

T (slices)

Output SNR (dB)

Raw — SRDCNN — SRDTrans — GT

a

Raw data (SNR = -0.05 dB)   SRDTrans (SNR = 26.53 dB)   GT   Raw data   SRDTrans   GT

b

Raw data — SRDTrans — GT (dashed)

Normalized intensity (AU) vs Distance (µm)

c

Output SNR (dB) vs Input SNR (dB)

Output SSIM vs Input SSIM

d

Raw data   SRDTrans   GT

e

Merge

Raw data   SRDTrans   GT

f

Raw data   SRDTrans

• GT molecules   ○ Detected molecules

g

Raw data   SRDTrans

Jaccard (%)

RMSE (nm)

h

RMSE (nm) vs Jaccard (%)

Efficiency=-20, Efficiency=0, Efficiency=20, Efficiency=40, Efficiency=60, Efficiency=80

▲ Raw data   ● SRDTrans

a Raw data / SRDTrans

b Time = 243.33 s / Time = 273.33 s

c Raw data / SRDTrans

ΔF/F (norm.) 0 — 1.0