Restoration of metabolic functional metrics from label-free, two-photon cervical tissue images using multiscale deep-learning-based denoising algorithms

Nilay Vora[1,+] , Christopher M. Polleys[1,+], Filippos Sakellariou[2], Georgios Georgalis[3], Hong-Thao Thieu[4ξ], Elizabeth M. Genega[5#], Narges Jahanseir[5], Abani Patra[3,6], Eric Miller[7,8], Irene Georgakoudi[1]*

[1] Department of Biomedical Engineering, Tufts University, Medford, MA 02155, USA.

[2] Anatolia College, Thessaloniki, Greece.

[3] Data Intensive Studies Center, Tufts University, Medford, MA 02155, USA.

[4] Department of Obstetrics and Gynecology, Tufts University School of Medicine, Tufts Medical Center, Boston, MA 02111, USA.

[5] Department of Pathology and Laboratory Medicine, Tufts University School of Medicine, Tufts Medical Center, Boston, MA 02111, USA.

[6] Department of Mathematics, Tufts University, Medford, MA 02155, USA.

[7] Department of Electrical and Computer Engineering, Tufts University, Medford, MA 02155, USA.

[8] Tufts Institute for Artificial Intelligence, Tufts University, Medford, MA 02155, USA.

[+] These two authors contributed equally to this work.

[ξ] Current affiliation: Department of Obstetrics and Gynecology, Newton-Wellesley Hospital, Newton, MA 02462, USA.

[#] Current affiliation: Department of Pathology & Laboratory Medicine, Emory University Hospital, Atlanta, GA 30322, USA.

* Corresponding Author

Contact Information

Nilay Vora: nilay.vora@tufts.edu

Christopher M. Polleys: christopher.polleys@tufts.edu

Filippos Sakellariou: 20193005@student.anatolia.edu.gr

Georgios Georgalis: georgios.georgalis@tufts.edu

Hong-Thao Thieu: Htthieu@mgb.org

Elizabeth M. Genega: egenega@emory.edu

Narges Jahanseir: njahanseir@tuftsmedicalcenter.org

Abani Patra: abani.patra@tufts.edu

32   Eric Miller: eric.miller@tufts.edu

33   Irene Georgakoudi: irene.georgakoudi@tufts.edu

Abstract

Label-free, two-photon imaging captures morphological and functional metabolic tissue changes and enables enhanced understanding of numerous diseases. However, this modality suffers from low signal arising from limitations imposed by the maximum permissible dose of illumination and the need for rapid image acquisition to avoid motion artifacts. Recently, deep learning methods have been developed to facilitate the extraction of quantitative information from such images. Here, we employ deep neural architectures in the synthesis of a multiscale denoising algorithm optimized for restoring metrics of metabolic activity from low-SNR, two-photon images. Two-photon excited fluorescence (TPEF) images of reduced nicotinamide adenine dinucleotide (phosphate) (NAD(P)H) and flavoproteins (FAD) from freshly excised human cervical tissues are used. We assess the impact of the specific denoising model, loss function, data transformation, and training dataset on established metrics of image restoration when comparing denoised single frame images with corresponding six frame averages, considered as the ground truth. We further assess the restoration accuracy of six metrics of metabolic function from the denoised images relative to ground truth images. Using a novel algorithm based on deep denoising in the wavelet transform domain, we demonstrate optimal recovery of metabolic function metrics. Our results highlight the promise of denoising algorithms to recover diagnostically useful information from low SNR label-free two-photon images and their potential importance in the clinical translation of such imaging.

Introduction

Metabolism refers to the set of chemical reactions that occur within a cell to produce energy and to build the necessary macromolecules to sustain life[1]. The energetic and macromolecular demands of a cell often change with aging and the onset of several diseases, including cancer, diabetes, neurodegenerative disorders, and cardiovascular diseases[2]. Therefore, it is clear that understanding the nature of such metabolic changes at the cellular level to characterize heterogeneity and dynamic interactions among different cell populations is critical for the development of improved diagnostic and treatment methods[3]. However, established methods to assess metabolic function in the clinic and the laboratory either lack resolution[4] or are destructive[5].

One approach that is capable of probing tissue metabolic state with high three-dimensional resolution in a non-destructive manner is two-photon excited fluorescence (TPEF) microscopy[6]. TPEF is a non-linear imaging technique that benefits from intrinsic optical sectioning and the ability to penetrate hundreds of micrometers into bulk tissue[7]. TPEF is also uniquely suited to capture images from endogenous fluorophores such as NAD(P)H and FAD[8]. NADH and FAD are coenzymes that facilitate energy generation and biomolecular synthesis via a number of pathways[9]. Several of these pathways, including the tricarboxylic acid cycle, glutaminolysis, fatty acid oxidation, and oxidative phosphorylation occur in the mitochondria[10]. NADPH plays an important role in anti-oxidant pathways and has similar fluorescence characteristics to those of NADH[11]. Thus, the term NAD(P)H is used throughout this paper to refer to the fluorescence of both NADH and NADPH. A large fraction of the flavin-associated cellular fluorescence is attributed to FAD bound to lipoamide dehydrogenase (LipDH), even though

contributions from free FAD and FAD bound to Complex II (electron transfer

flavoprotein) may also be significant. Here, we use the term FAD to refer to all flavin-

associated fluorescence detected from cells.

Despite the lack of specificity in the origins of the fluorescence signals, the ratio of

FAD/NAD(P)H or its normalized definition of FAD/(NAD(P)H+FAD) have been shown to

correlate to the oxido-reductive state of the cells in many studies[12–15]. Mitochondria are

also characterized by the ability to fuse and fission to enhance energy production and

delivery in response to stress or to facilitate removal of damaged mitochondria[16]. Such

differences in mitochondrial organization have also been quantified based on analysis of

NAD(P)H TPEF images[17,18]. NAD(P)H fluoresces more efficiently when bound to

enzymes typically in the mitochondria; therefore variations in NAD(P)H TPEF intensity

fluctuations can be exploited for label-free quantitative assessments of mitochondrial

organization (clustering) in cells, tissues, and living humans[17,19]. Changes in

mitochondrial organization have in turn been attributed to metabolic function changes[20–

22]. The heterogeneity of parameters such as the redox ratio and mitochondrial clustering

within a tissue have also been identified as important indicators of metabolic state[23–25].

A number of studies have already highlighted the diagnostic potential of such

assessments in living humans and there is growing interest in performing such

measurements at the bedside or via endoscopes to expand the range of diagnostic

applications to several organs beyond the skin[26–29]. Fast image acquisition in these

settings is critical; however, endoscope designs typically include relatively low numerical

aperture (NA) (0.5-0.7) objectives and are not as efficient in the generation and

collection of TPEF[30]. As a result, low resolution, noise, and other degradations may

mask the diagnostically useful functional features. Thus, approaches to enhance label-

free, TPEF images could play a transformative role in the successful translation of this

technique to improve tissue metabolic function assessments in the context of diagnosis

or treatment.

Traditionally, both standard image processing methods as well as inverse techniques

have been used to enhancing the interpretability of TPEF data[31]. These methods are

most appropriate when one can easily model the physics associated with the sensing

modality and the stochasticity of the data is captured in a computationally convenient

distribution. Neither is the case for TPEF sensing where the interaction of light with

tissue leads to a highly complex forward model and the data are a mix of Poisson

statistics and additive Gaussian noise[32]. Motivated by these challenges as well as the

recent success of machine learning methods for addressing a range of image analysis

and interpretation problems, we consider the use of deep-learning methods for

enhancing TPEF images to improve the extraction of metabolically-relevant information.

Deep-learning-based methods have already been shown to enhance quality and

resolution of a wide range of images, including label-free two-photon images[33–37].

Convolutional neural network-based content-aware image restoration (CARE), residual

channel attention networks (RCAN), and super-resolution generative adversarial

networks (SRGAN) have been developed for this purpose[33–35]. While these models

have been applied to fluorescence microscopy data, their use has been limited to

exogenously labeled samples which have enhanced contrast compared to label-free

images. However, recently, *Shen et al.* (2022) demonstrated the application of a

generative adversarial networks (GAN) for the restoration of label-free multimodal

nonlinear images[36]. We note that in these and related studies, standard metrics, such

as peak SNR (PSNR) and structural similarity index measure (SSIM) are used widely as

indicators of the quality of image restoration, even though they may not always match

the human visual system's assessment of image quality (MOS)[33,34,36,37].

Here, we report on the ability of deep-learning based denoising approaches to restore

functional metabolic metrics extracted from label-free TPEF images. Specifically, we

consider recovery of average and depth dependent variations in the redox ratio

(FAD/(NAD(P)H+FAD)) and mitochondrial clustering extracted from analysis of TPEF

images acquired from freshly excised human cervical epithelia, including healthy and

precancerous lesions. In addition, we assess whether PSNR and SSIM improvements

are correlated with the restoration of the functional metabolic metrics. We consider

CARE (a U-net), GANs (SRGAN), and RCAN networks, and assess five loss functions,

including mean average error (MAE), mean square error (MSE), SSIM, frequency focal

loss (FFL), coefficient of variation (R2), and three combinations of these loss functions

(see Supplementary Discussion S1 and Supplementary Fig. S2 online). We also

examine whether training on FAD or NAD(P)H images impacts the successful

restoration of metabolic function metrics from the corresponding denoised images.

We find that a novel combination of a one level wavelet transformation and CARE

models trained to denoise each of the four wavelet domain sub-bands yields denoised

images that enable optimized recovery of all metabolic function metrics. Interestingly,

we observe that the architecture most successful in recovering metabolic metrics is not

optimal in terms of more standard metrics such as PSNR and SSIM used to measure

performance. Thus, our results indicate that deep-learning based denoising algorithms

may require distinct multiscale training and testing approaches for the recovery of

functional metrics needed for improved diagnosis and for understanding the drivers of

disease and development of novel therapeutics, instead of traditional morphological

image quality metrics.

Results

<u>Identification of the optimal deep-learning model architecture for denoising label-free,</u>

<u>optical TPEF images to enable recovery of metabolic function metrics</u>

Human cervical tissue biopsies were collected from 54 patients and imaged immediately

upon excision, as described in *Methods: Optical Instrumentation and Image Acquisition*

(Figure 1). Several regions of interest (ROIs) were imaged from each biopsy. Multiple

optical sections (OS) were imaged from each ROI at distinct depths. At each OS, we

acquired TPEF images at a combination of two excitation wavelengths (775 and 860

nm) and three or four emission bands. Images collected at 775 nm excitation 435-485

nm emission were attributed primarily to NAD(P)H, while images at 860 nm excitation

500-550 nm emission were considered to contain signal primarily from FAD and FAD

bound to lipoamide dehydrogenase. Six frames were acquired at each wavelength

setting. To reduce the contribution of noise, these six frames were averaged together.

Metrics extracted from these averaged images were previously observed to enable

highly sensitive and specific detection of cervical pre-cancer[25]. The averaged image

was therefore considered the ground truth used for training and testing the denoising

success of single frames. Single frames, the corresponding denoised, and ground truth

images were analyzed using established procedures to extract the redox ratio *(*RR),

defined as FAD/(NAD(P)H+FAD) in this study, and mitochondrial clustering (β) (Figure

1). All models (Figure **1**) were trained and evaluated with identically generated image

stacks. Various combinations of model architectures, loss functions, data

transformations, and training data combinations, as outlined in Table 1, were evaluated

on 3229 total OSs representing healthy/benign cervical tissues as well as precancerous

(low-grade and high grade) squamous intra-epithelial lesions (LSIL and HSIL,

respectively).

Figure 1: Summary of deep learning pipeline. Human cervical tissue biopsies are
collected and subsequently imaged within 4 hours post-excision. Collected biopsies are
plated on glass bottom dishes and imaged using a Leica SP8 commercial microscope.

At a minimum, three ROIs are imaged per sample. At each ROI, multiple optical
sections are imaged at distinct depths through the epithelium. Depth-resolved, two-
photon OSs are collected using 2 excitation wavelengths and several bandpass-filtered
detectors. Six images are captured for each excitation/emission wavelength and every
OS at a given depth, z. These six images are averaged together to generate the ground
truth image set. A random image from the six per depth z is selected as the input (RAW)
image. The paired image stacks are provided to the neural network for training and
denoising. Four-leading denoising networks are used in this study to denoise input
images: a previously described CARE model, an RCAN model, an SRGAN model, and
a WU-net[33–35,37]. Denoised images and Input images are compared against 6x
Averaged images to determine 3D PSNR and SSIM along with metabolic metrics. Scale
bar = 25 μm.

PSNR and SSIM improvements are standard metrics of image visual quality and have

been used in other studies focused on denoising biomedical images as an indicator of

model success[33,34,36,37]. We aimed to assess whether images restored by models that

yield optimized PSNR and SSIM values result in accurate recovery of metabolic metrics

(Figure 2). For evaluation of model architecture, loss function, and signal type, only

results from models trained on NAD(P)H data from tissues of known benign status were

included.

Table 1: Summary of all parameters explored during training and optimization of the
final model (highlighted in bold). Results shown below are focused on the optimized
model, but all combinations were trained and evaluated.

Model Architecture	**CARE**	RCAN		SRGAN		
Loss Functions	MAE / L1	MSE / L2		SSIM + L2	SSIM + FFL	**SSIM + R2**
Signal Pre-processing Method	**Wavelet Transform**	None				
Training Data Format	Healthy Data Only	**Healthy and Diseased (Mix)**				
Training Data Type	NAD(P)H Data	**FAD Data**				

Leading denoising model architectures were selected for evaluation based on a

comprehensive literature search. CARE, RCAN, and SRGAN (Figure 1) models were

trained as described in *Methods: Deep Learning Model Description* and *Deep Learning*

Performance Benchmark. A representative OS from a HSIL biopsy is shown in Figure

3**a**. Results shown were generated by models trained using an SSIM + Mean Squared

Error (MSE or L_2) loss function. A summary of all parameters used to generate the

figures and tables are listed in Table 2.

Figure 2: β and RR metrics are extracted from each OS *(See Materials and Methods:*
Morphological and Functional Metrics for greater detail). Depth-dependent trends
across the multiple cell layers of the cervical squamous epithelium are assessed for
input images (RAW 1X), denoised images (Restored), and six-frame averaged, ground
truth images (GT 6X). Measurements of mean values and corresponding variability
across all depths are shown as a dashed line and shaded region in the mitochondrial
clustering, β, and RR (FAD/(NAD(P)H+FAD)) panels for the GT 6X image. The
distribution of RR values for each OS is used to extract the interquartile range (IQR),
representing the range of values within the 25% and 75% of the RR distributions and
providing an assessment of intra-field RR heterogeneity. IQR variability is a metric of
inter-field (depth-dependent) RR heterogeneity.

Prior to denoising, standard image quality metrics were calculated for input (RAW 1X)

images by comparing the RAW 1X images to ground truth (GT 6X) images. PSNR and

SSIM values were calculated using the GT 6X image as a reference and RAW 1X or

denoised images as the distorted image[38]. Across all images, FAD image PSNR was

greater than NAD(P)H image PSNR (Table 3), even though FAD images featured lower

cytoplasmic signal compared to NAD(P)H images (Figure 3**a**). During PSNR calculation,

the reduced signal intensity led to smaller differences between RAW 1X and GT 6X

images and yielded a greater observed PSNR value. This observation was also

consistent with results from other studies[36]. SSIM values were consistent between

NAD(P)H and FAD images (Table 3). Corruption of the GT 6X images for both channels

by noise was expected to have similar effects on structural similarity and calculated

SSIM values.

Table 2: Summary of parameters used to generate Figures 3-6 and Table 3-6.
Parameters are bolded when all combinations from Table 1 are used.

	Model Architecture	Loss Functions	Pre-Processing Method	Training Data Format	Training Data Type
Figure 3 / Table 3	**All**	SSIM + L2 Loss	None	Healthy Only	NAD(P)H
Figure 4 / Table 4	CARE	SSIM + R2	**All**	Healthy Only	NAD(P)H
Figure 5 / Table 5	CARE	SSIM +R2	Wavelet Transform	**All**	**All**
Figure 6 / Table 6	CARE	SSIM + R2 MSE MAE	**All**	**All**	**All**

237   We used 777 and 109 RAW 1X NAD(P)H OSs for training and validation of the models,

238   respectively. Each 512 x 512 OS was patched into four-256 x 256 image patches (OSP)

239   prior to training and validation (3108 and 436 OSPs, respectively). All three models

Figure 3: **(a)** A 290 x 290 µm^2 field of view from a low-grade squamous intraepithelial lesion (LSIL) cervical tissue biopsy. NAD(P)H and FAD images for the same region are shown along with the corresponding denoised image from each of the three trained models (CARE, RCAN, and SRGAN). Scale bar = 50 µm. **(b)** A 44.2 x 44.2 µm^2 field of view (white square in **a**) of three cells. NAD(P)H and FAD images are shown for all models and the input and ground truth images. Scale bar = 10 µm. **(c)** Bar plots of the coefficient of determination of all downstream metrics for images denoised by all models and RAW 1X vs. the GT 6X image. Fisher r to z transformation was used to measure significance. *p<0.05 and **p<0.01.

were trained before being evaluated on an independent set of 2343 OSs (9372 OSPs).

Metrics of image quality and metabolic function were calculated as described in

Methods: Deep Learning Metrics and *Methods: Morphological and Functional Metrics*

sections.

CARE-generated image stacks demonstrated higher PSNR for FAD images and higher

SSIM for both NAD(P)H and FAD images compared to restored-image stacks

generated by RCAN and SRGAN. Across all test set images, standard metrics of image

quality (Table 3) and visual inspection (Figure 3**b**) suggested RCAN- and CARE-

denoised images had similar image quality. Across the entire test set, we observed

SRGAN failed to restore cellular features within the GT 6X images (Figure 3**b**) and

underperformed even relative to RAW 1X images in standard image quality metrics

(Table 3). Perceptual loss was believed to impact content restoration in the SRGAN

architecture[34]. Inputs for perceptual loss calculations have been shown to impact

significantly SRGAN performance and were likely the cause of SRGAN's poor recovery

of image quality[34].

To assess restoration of metabolic activity, depth-dependent optical RR and

mitochondrial clustering (β) values were calculated for the restored images, input (RAW

1X) images, and ground truth (GT 6X) images (Figure 2). Pearson correlation coefficient

values were calculated between the metabolic function metrics from the GT 6X and

either the RAW 1X or restored images. Statistical significance was derived from Fisher-

r-to-z transformation for all metrics of interest. Interestingly, analysis of the RAW 1X

images led to very high correlations with metrics of RR intra- and inter-field variability

compared to GT 6X images. We hypothesized that similar sources of noise in both FAD

and NAD(P)H images led to this outcome since RR metrics were calculated using a

ratio of FAD and NAD(P)H intensity measurements. It was for this reason that in this

initial comparison, we trained models on NAD(P)H images and applied the same

weights to NAD(P)H and FAD images. RCAN-generated images demonstrated

statistically significant recovery of β variability ($\sigma^2(\beta)$) (Figure 3**c**). However, recovery of

mean RR variability by this model was poor (Figure 3**c**). CARE-denoised images,

overall, featured higher (albeit not statistically significant) correlations with RR metrics

compared to all other models (Figure 3**c**). Thus, the U-net architecture of CARE was

utilized for all further optimization steps.

Table 3: Summary of standard metrics of image quality for RAW 1X images and
denoised images generated from various model architectures. Values are reported for
mean performance (± standard deviation) across all test set ROIs.

Model Name	NAD(P)H Images		FAD Images	
	PSNR (dB) ↑	SSIM ↑	PSNR (dB) ↑	SSIM ↑
RAW 1X	19.2 ± 2.8	0.48 ± 0.09	23.1 ± 5.5	0.49 ± 0.13
CARE	22.7 ± 2.9	**0.63 ± 0.08**	**26.8 ± 3.1**	**0.60 ± 0.07**
RCAN	**23.1 ± 1.7**	0.62 ± 0.08	24.3 ± 2.0	0.51 ± 0.12
SRGAN	19.6 ± 1.1	0.31 ± 0.08	20.2 ± 1.5	0.25 ± 0.07

A Multiscale Image Transformation Enhances Quantification of Mitochondrial Clustering:

Although denoising improved the restoration of $\sigma^2(\beta)$, the mean β ($\bar{\beta}$) values of the

denoised images were not well correlated with the values from the GT 6X images. We

considered discrete wavelet transformation (DWT) to enhance high spatial frequency

restoration necessary for β metric calculations. A single level DWT, transformed each

image into four sub-band images: a coarser scale approximation and three *detail*

images; one horizontal, one vertical, and one diagonal[39]. To generate the three subband

images, a basis function, called a mother wavelet, was convolved along both

dimensions of the original image while an associated scaling function was used to

generate the coarser approximation. During standard wavelet-based denoising,

thresholds are used to remove noise from wavelet-transformed detail images, before

implementing an inverse-transform to recover the restored image[40,41]. The DWT has

been shown to be advantageous compared to traditional low-pass filtering as the pixel-

by-pixel convolution with the mother wavelet preserves correlations of high frequency

features. In this study, we used deep learning models trained on each of the

transformed images to adaptively learn the best threshold for denoising of low

frequencies (approximation) and high frequencies (details) rather than relying on

arbitrary thresholding for denoising (see Supplementary Discussion S2 online)[42]. As

with any DWT-denoising model, the selection of the correct mother wavelet played a

significant role in model performance. For all models, mother wavelets from the

biorthogonal, coiflets, and Daubechies families were evaluated. These mother wavelets

families were selected due to their frequent use in denoising tasks[43]. Multiple models

were trained and evaluated, with biorthogonal 1.1 yielding the highest recovery of

Figure 4: **(a)** A 290 x 290 µm^2 field of view from a LSIL cervical tissue biopsy. NAD(P)H and FAD images for the same region are shown along with the corresponding denoised image from each signal pre-processing method utilized (Single frame image and Wavelet transformation). Scale bar = 50 µm. **(b)** A 44.2 x 44.2 µm^2 field of view (white square in **a**) of three cells. NAD(P)H and FAD images are shown for all models based on the corresponding signal pre-processing method used during training and the input and ground truth images. Scale bar = 10 µm. **(c)** Bar plots of the coefficient of determination of all downstream metrics for images denoised by all models trained based on the corresponding signal pre-processing method used during training and RAW 1X vs. the GT 6X image. Fisher r to z transformation was used to measure significance. *$p<0.05$ and **$p<0.01$.

metabolic metrics (data not shown). As such, biorthogonal 1.1 was used for all

subsequent model optimization.

Application of DWT before training four CARE models and inverse DWT (iDWT) after

evaluation yielded images with improved FAD and NAD(P)H PSNR with slight

decreases in SSIM (Figure 4**a**). Across the entire test set, NAD(P)H PSNR improved

using WU-net while FAD PSNR and SSIM both decreased compared to CARE (Table

4). All loss functions were evaluated for WU-net, with SSIM + R2 loss (results shown in

Figure 4) and SSIM + FFL loss (see Supplementary Table S3 online) yielding the best

overall performance. WU-net denoised NAD(P)H images extracted similar cellular

structures as the CARE derived images but featured lower background signal and small

fluctuations in cytoplasmic signal leading to the observed higher PSNR values (Figure

5**b**). WU-net led to statistically significant improvements in the correlation of extracted

$\sigma^2(\beta)$ with GT 6X images relative to analysis of the RAW 1X images. Extracted $\bar{\beta}$

values were also better correlated to GT 6X images, albeit improvements were not

significant.

Comparing WU-net to an identical CARE model, we observed that WU-net achieved

improved performance on β metrics while maintaining recovery of RR metrics (Figure

4**c**). The overall improved β restoration suggested that WU-net was better able to

capture true signals from noise in the high spatial frequencies found in NAD(P)H

images. WU-net further preserved the relationship between NAD(P)H and FAD channel

images, enabling high correlations for RR metrics. Due to the observed performance of

WU-net for β metric recovery, we explored further optimization of WU-net which could

be achieved by varying training datasets.

Table 4: Summary of standard metrics of image quality for RAW 1X images and
denoised images generated after signal pre-processing. Values are reported for mean
performance (± standard deviation) across all test set ROIs.

Data Transform	NAD(P)H Images		FAD Images	
	PSNR (dB) ↑	SSIM ↑	PSNR (dB) ↑	SSIM ↑
RAW 1X	19.2 ± 2.8	0.48 ± 0.09	23.1 ± 5.5	0.49 ± 0.13
CARE	22.8 ± 3.0	**0.63 ± 0.08**	**27.2 ± 3.7**	**0.62 ± 0.07**
Wavelet	**23.6 ± 2.3**	**0.63 ± 0.08**	26.1 ± 2.3	0.57 ± 0.09

Selection of Training Data:

Initial model development focused on a limited training set of cervical tissues of known

benign status (Healthy). Benign tissue samples comprised of cell layers with consistent

changes in differentiation as a function of depth among image stacks. Training on such

images was expected to enable the model to learn characteristics of noise without

having to account for feature heterogeneity found in pre-cancerous cervical tissue

samples. We further sought to assess whether training on a data set that was expanded

to include image stacks from tissues with both benign and pre-cancerous lesions (Mix)

Figure 5: **(a)** A 290 x 290 µm² field of view from a LSIL cervical tissue biopsy. NAD(P)H
and FAD images for the same region are shown along with the corresponding denoised
image from each data type used as training data for the WU-net model (NAD(P)H
Healthy Only-, NAD(P)H Mixed Diagnosis-, and FAD Mixed Diagnosis- Wavelet
transformed images). All models were equally constructed with only the data type and
diagnosis type varied. Scale bar = 50 µm. **(b)** A 44.2 x 44.2 µm² field of view (white
square in **a**) of three cells. NAD(P)H and FAD images are shown for all data types used
during training and the input and ground truth images. Scale bar = 10 µm. **(c)** Bar plots
of the coefficient of determination of all downstream metrics for images denoised by
models trained using varying data types and diagnosis types and RAW 1X versus the
GT 6X image. Fisher r to z transformation was used to measure significance. *$p<0.05$
and **$p<0.01$

impacted performance. In this new training set, 1657 and 554 RAW 1X NAD(P)H OSs

(6628 and 2216 OSPs) were used for training and validation of the models, respectively.

An independent test set of 1018 OSs (4072 OSPs) was used to evaluate model

performance after training.

An additional consideration we explored was the impact of the source of image contrast,

i.e., NAD(P)H or FAD, used for training. NAD(P)H images featured greater structural

information compared to FAD images, and they were utilized in our analysis for

extraction of mitochondrial clustering-focused metabolic function metrics (Figures 3-5).

Thus, training was focused on NAD(P)H images, and optimized model weights from

NAD(P)H image training were used to denoise FAD images for extraction of RR-based

metrics. However, since similar noise characteristics were assumed to be present in

both RAW 1X NAD(P)H and FAD images, we sought to confirm that training on

NAD(P)H images was optimal. Thus, we used FAD images to train WU-net models

using the same hyperparameters and settings as the ones used when NAD(P)H images

were used. Post-training, NAD(P)H images were denoised using the weights of the FAD

image trained model to extract RR and mitochondrial clustering-based metrics.

The use of training sets with mixed diagnosis images resulted in minimal differences in

the denoised images when compared to training just on healthy sample images (Figure

5**a**). PSNR and SSIM values for images were observed to be nearly identical because

of these insignificant differences (Table 5). Both models led to denoised images with

consistent cell boundary and intracellular structures given the same RAW 1X images

(Figure 5**b**) and had similar levels of restoration of downstream metrics, with the mixed

diagnosis data set leading to slightly improved correlations in most cases (Figure 5**c**).

The increase in correlation could be attributed to the large training set available for a

mixture of diagnoses compared to only training on healthy data.

An identical model was trained using the FAD image data from the mixed diagnosis

dataset. While the denoised images from the FAD-trained model looked like those from

the corresponding NAD(P)H-trained model (Figure 5**a** and 5**b**), standard metrics of

image quality were slightly lower. Images denoised by the FAD-trained model

demonstrated higher background signal compared to images denoised by NAD(P)H-

trained models (Figure 5**b**). However, despite FAD images lacking much of the

structural and morphological information of their NAD(P)H counterparts, their use in

training led to further improvements in β metric recovery and mean RR restoration from

the RAW 1X images (Figure 5**c**). We hypothesize high frequency information in the FAD

images originated primarily from noise in comparison to NAD(P)H images. As a result of

the high frequency information containing primarily noise, the model improved in its'

learning of noise characteristics in the images, enabling improved denoising and

recovery of metrics of metabolic activity (Figure 5**c**).

Table 5: Summary of standard metrics of image quality for RAW 1X images and denoised images generated after training models on various data types. Values are reported for mean performance (± standard deviation) across all test set ROIs.

Training Data	NAD(P)H Images		FAD Images	
	PSNR (dB) ↑	SSIM ↑	PSNR (dB) ↑	SSIM ↑
RAW 1X	19.2 ± 2.8	0.48 ± 0.09	23.1 ± 5.5	0.49 ± 0.13
Healthy Only	**23.6 ± 2.3**	**0.63 ± 0.08**	26.1 ± 2.3	**0.57 ± 0.09**
Mixed NAD(P)H	23.4 ± 2.5	**0.63 ± 0.08**	**26.3 ± 3.1**	**0.57 ± 0.08**
Mixed FAD	23.5 ± 2.6	0.62 ± 0.09	24.8 ± 3.7	0.52 ± 0.08

Summary of Final Model Performance:

Across all models, image quality improved after denoising based on PSNR and SSIM (Table 6). Based on standard image quality metrics of all models discussed in this study, it could be assumed that models trained using NAD(P)H images and the CARE architecture with standard loss functions of MAE and MSE would perform best at the restoration of downstream metrics (Figure 6**a**). CARE models trained with MAE and MSE loss functions both demonstrated statistically significant improvement in denoised FAD and NAD(P)H image PSNR and SSIM ($p < 0.001$). Comparatively, Wavelet-transformed- FAD images denoised using WU-net with SSIM + R2 loss had poorer standard metric performance (Table 6). Images restored with this model did not achieve statistically significant improvement of FAD image PSNR and SSIM (Figure 6**a**). As PSNR and SSIM are commonly used as indicators of model performance, it was expected that improvements in these metrics would correspond to better recovery of downstream metabolic metrics. However, the WU-net model trained on mixed diagnosis, FAD images led to denoised images whose extracted metabolic metrics were

Table 6: Summary of standard metrics of image quality (PSNR and SSIM) for RAW 1X images, standard implementation of CARE, and the best performing model from this study. Values are reported for mean performance (± standard deviation).

Final Model	NAD(P)H Images		FAD Images	
	PSNR (dB) ↑	SSIM ↑	PSNR (dB) ↑	SSIM ↑
RAW 1X	19.2 ± 2.8	0.48 ± 0.09	23.1 ± 5.5	0.49 ± 0.13
Healthy NAD(P)H CARE MAE	23.6 ± 2.8	0.63 ± 0.08	**26.9 ± 2.7**	**0.59 ± 0.08**
Healthy NAD(P)H CARE MSE	**23.7 ± 2.6**	**0.64 ± 0.08**	26.8 ± 2.7	**0.59 ± 0.08**
Mixed FAD CARE SSIM + R2 Wavelet Transform	23.5 ± 2.6	0.62 ± 0.09	24.8 ± 3.7	0.52 ± 0.08

consistently correlated with the metrics extracted from GT 6X images (Figure 6**b**). The

final correlations of the models shown in Figure 6 are reported in Table 7.

Discussion

Tissue morphological and functional metrics extracted from label-free, 2PM images

could provide significant clinical utility for disease diagnosis[25]. Neural networks will likely

play a critical role in enabling accurate extraction of such metrics from images that are

likely to be acquired in an in vivo imaging setting. Previous studies by multiple groups

have demonstrated deep learning-based denoising models can be used to improve the

PSNR and SSIM of fluorescence images acquired using 2PM[33,35,36]. Here, we

demonstrated PSNR and SSIM, while relevant in the assessment of image quality, were

not representative of functional metric recovery needed for clinical utility (Figure 6).

Different algorithms have been reported for denoising of fluorescence images, however,

only *Shen et al. (2022)* have reported a network used for denoising of label-free

Figure 6: (**a**) Box and whisker plots of PSNR and SSIM of 40 test set ROIs. Denoised images demonstrated an improvement in standard metrics of image quality. (**b**) Bar plots of the coefficient of determination of all downstream metrics for images denoised by models trained using various data types, loss functions, and diagnosis types versus the GT 6X image. A one-way ANOVA with Tukey Kramer post-hoc test was used to measure significance of PSNR and SSIM. Fisher r to z transformation was used to measure significance of improvement in metabolic metric correlations *p<0.05, **p<0.01, and ***p<0.001.

autofluorescence images[36]. In this study, a modified enhanced SRGAN model was used

to denoise ex-vivo, multi-modal label-free images of human ovarian cancer tissue

sections[36]. The trained GAN demonstrated an ~4.5 dB improvement in PSNR and 79%

improvement in SSIM after denoising[36]. In comparison, we demonstrated ~4.3 dB and

~2.7 dB improvement in PSNR and ~30% and 6% improvement in SSIM for NAD(P)H

and FAD images, respectively (Figure 6**a**). While improvement in image PSNR and

SSIM were lower, RAW 1X and Denoised images in this study have higher PSNR and

SSIM for all images suggesting differences in enhancement are due to limits in image

improvement and not a lack of network performance (Table 6).

Table 7: Correlation values of models in Figure 6. Fisher r to z transformation was used
to measure significance. **p<0.01

Final Model	Downstream Metrics					
	Mean RR ↑	Mean β ↑	Mean RR IQR ↑	σ^2(Mean RR) ↑	σ^2(Mean β) ↑	σ^2(Mean RR IQR) ↑
RAW 1X	0.71	0.40	**0.87**	0.84	0.33	0.57
Healthy NAD(P)H CARE MAE	0.82	0.57	0.85	**0.89**	0.78**	**0.70**
Healthy NAD(P)H CARE MSE	0.87	0.53	0.84	0.84	0.81**	0.66
Mixed FAD CARE SSIM + R2 Wavelet Transform	**0.96****	**0.68**	**0.87**	0.84	**0.90****	**0.70**

We further observed that GAN models did not perform well on our dataset. GANs aim to

emulate characteristics of high SNR images in low SNR images through an adversarial

training process[34]. To improve image quality, GANs learn the manifold of high SNR data

which is assumed to be composed of images that have similar image quality metrics[44].

Thus, it is important for image quality to be consistent across all high SNR images.

High-SNR images from a single depth in a thin OS, like those used to train the

enhanced SRGAN model in *Shen et al.* (2022), have similar image quality for all ground

truth images leading to improved GAN performance[36]. In our study, bulk tissues were

imaged at multiple depths leading to inconsistent image quality in our ground truth

images as SNR is known to change as a function of depth. As such, we hypothesized

the GAN model implemented in this study failed to learn the manifold of high SNR

images and improve our images whereas the enhanced SRGAN model implemented by

Shen et al. (2022) succeeded.

While multiple studies demonstrate models capable of improving PSNR and SSIM,

assessment of morphofunctional metrics of metabolic activity after denoising has not

been examined previously[33–37]. Here, we calculate restored image PSNR and SSIM

along with metabolic metric recovery and observe that higher PSNR and SSIM values

did not ensure the greatest restoration of RR and β metrics (Figure 6). While PSNR and

SSIM values between models are observed to be within <5% of each other (Table 3-5),

many studies indicate maximum improvement of PSNR and SSIM values as indicators

of model performance[33–37]. In this study, we observe that models with optimal PSNR

and SSIM values did not yield the greatest recovery of metabolic metrics. Altogether,

PSNR and SSIM are not well suited for assessment of model performance on label-free

2PM images, necessitating further validation using metrics of metabolic activity.

Application of denoising algorithms on label-free 2PM datasets to date have been

limited by the lack of available large clinical datasets[36,45]. Deep learning models have

shown promise with small datasets (*Shen et al.* (2022) used only 24 paired images) in

image restoration; however, larger datasets are needed for consistent reconstruction of

high-SNR images[36,44]. Here we presented a denoising network trained on a larger

training set of 1657 OSs (6628 OSPs) and evaluated on an independent test set of

1018 OSs (4072 OSPs).

Using CARE, we observed improvements in image quality based on standard metrics

(Table 3). However, the pre-packaged, standard models showed poor recovery of β

metrics. Custom-loss functions improved metabolic metric recovery by penalizing

models for both failing to generate similar images and reducing pixel correlation (see

Supplementary Table S2 online). More interestingly, we observed the use of DWT to

separate the frequency information in an image before training independent models

(WU-net) produces images that had high metabolic metric correlations with GT 6X

metrics (Figure 4**c**). By training on independent frequency-band images, the models

were forced to learn the noise characteristics of different frequency bands without

convolving the bands[42].

A key advantage of WU-net, in comparison to identically trained (non-wavelet) U-nets,

was the denoising of higher frequencies where noise was expected to be dominant.

Denoising of high frequency noise led to enhanced recovery of β metrics as WU-net

was more consistent in reducing noise in these frequencies (See Supplementary

Discussion S3 and Supplementary Fig. S3 online). WU-net led to a statistically

significant decrease in high frequencies compared to a comparable CARE model (See

Supplementary Fig. S3 online). Further, the incorporation of SSIM + R2 as a loss

function promoted the models to restore similar frequencies from the GT 6X image in

the denoised image while minimizing the loss of correlation between pixels.

Further, we observed that models trained on FAD images outperform their NAD(P)H

counterparts (Figure 5**c**). To explain this phenomenon, we examined the correlation of

optical RR metrics between RAW 1X images and GT 6X images. RR metrics from RAW

1X images correlated well with RR metrics from GT 6X images, suggesting that the

noise characteristics in FAD and NAD(P)H images are similar. However, as the FAD

images contain less signal compared to their NAD(P)H paired images, high-frequency

contributions are mostly noise in the RAW 1X FAD images. Thus, training on FAD

images likely improved the model's learning of noise characteristics. This led to

improvement in downstream metric recovery and translation of model weights to

NAD(P)H images.

WU-net with a custom loss (SSIM + R2) function and training on FAD data

demonstrates improved restoration of most metrics of metabolic activity from label-free,

2PM images (Table 7); however, further improvements in the restoration of $\bar{\beta}$ are

desired. One potential method of improving $\bar{\beta}$ restoration would be to design a loss

function that minimizes the differences in the power spectral density maps of paired

images that are used for β calculation. A challenge of such a method would be the

computational time required for generating these maps[22,24,25]. Future studies may

examine simpler predictors of mitochondrial clustering using a modified GAN network,

where the discriminator network will estimate β from the input images and optimize the

generator to achieve accurate β metric recovery.

In this study, we specifically focused on restoration of morphological and functional

metrics from label-free, 2PM images of human cervical tissue, relying on a single

denoising algorithm. Future studies will examine the application of the trained denoising

model and model architecture on datasets acquired from different microscope systems,

objective lenses, and tissue types. Validation of the model on these datasets would

support the broad use of WU-net for denoising label-free 2PM images. Further,

successful implementation of pre-trained models on other datasets would reduce the

need for large clinical datasets[36,45]. As the model advances, improvements in ground

truth data collection are needed. Ground truth data used in this study contain noise and

therefore are not truly representative of mitochondrial signal. Alternative techniques for

image acquisition such as slower line scan speed could be utilized to improve ground

truth image quality.

In summary, we demonstrated that maximizing standard metrics of image quality

(PSNR and SSIM) did not necessarily lead to improved recovery of functional tissue

metrics, especially ones associated with mitochondrial organization (Table 7). Using

WU-net with a custom loss function, we demonstrated improved recovery of functional

metrics of metabolic activity, even though PSNR and SSIM metrics were not optimal.

Results from this study support the application of deep learning algorithms for the

restoration of RR and β metrics from low-SNR 2PM images. As more data becomes

available both from varying microscope systems, objective lenses, and tissue types, a

more robust algorithm could be generated for rapid image collection and classification,

eventually improving patient health during *in vivo* image collection.

Materials and Methods

Sample Acquisition

All activities pertaining to cervical tissue biopsy handling were done in accordance with

approved Tufts Health Sciences IRB protocol #10283. Patients over the age of 18 with a

recent LSIL or HSIL pap smear diagnosis undergoing colposcopy or loop electrosurgical

excision procedure (LEEP) were recruited to the study. Informed consent was acquired

from all study subjects before participation. During the routine procedure, a second

biopsy from a colposcopically abnormal region of the cervix was taken and placed in a

custom-built tissue carrier containing keratinocyte serum-free media (KSFM; Lonza).

Biopsies were transported via personal vehicle to the Tufts Advanced Microscopy

Imaging Center (TAMIC) for imaging. All imaging was conducted within 4 hours post-

biopsy. Immediately after imaging, biopsies were fixed in 10% neutral buffered formalin.

Biopsies were returned within 5 business days to the Tufts Medical Center Department

of Pathology for standard histopathological diagnosis.

Patients over the age of 18 undergoing hysterectomies for benign gynecological

disease were also recruited to the study as healthy controls. The only difference

between healthy and precancerous biopsy acquisition was in the actual biopsy excision.

Healthy biopsies were sampled from the resected cervix by a pathologist after

macroscopic inspection to rule out abnormalities.

Deep Learning Dataset Details

A total of 151 ROIs (image stacks) were collected from 54 patients. The training and

validation set was comprised of 100 ROIs featuring 5-50 OSs per ROI. 75% of the ROIs

were randomly selected for training and the remaining 25% were set aside as a

validation set (1657 training OSs and 554 validation OSs). The test set featured 51

ROIs (with 10-50 OSs per ROI) and was excluded from all training (1018 OSs). For *k-*

fold validation, training and validation sets were shuffled for up to five seeds to ensure

robustness of denoising on a constant test set (see Supplementary Fig. S6 online). The

dataset features images from tissues with three diagnoses: Benign, LSIL, and HSIL.

The test set was composed of 25 Benign ROIs (49.02%), 14 LSIL ROIs (27.45%), and

12 HSIL ROIs (23.53%). The training and validation sets were composed of 55 Benign

ROIs (54.45%), 25 LSIL ROIs (24.75%), and 21 HSIL ROIs (20.79%). Based on

training/validation splitting seed, these values could range from 52-57.3% Benign, 25.3-

26.7% LSIL, and 18.7-22.7% HSIL in the training set and 48-64% Benign, 20-24% LSIL,

and 16-28% HSIL in the validation set. An alternative training scheme was initially

attempted. In this scheme, only benign ROIs were used in training with 112 ROIs of

mixed diagnosis being used in the test set and 39 benign ROIs being used for training.

The training set was later modified as it became evident that training on a mixture of

diagnoses resulted in superior restoration of downstream metrics (Figure 5).

Optical Instrumentation and Image Acquisition

Images were collected using a commercially available Leica SP8 inverted microscope

system equipped with an Insight fs laser. Tissue biopsies were placed epithelial side

down onto a glass bottom dish and light was delivered using an epi-illumination

scheme. Tissue biopsies were excited with 755 nm and 860 nm light. Two hybrid

photodetectors (HyDs) were set up to collect 460 ± 25 and 525 ± 25 nm light. Two

photomultiplier tubes (PMTs) were set up to collect 430 ± 12 and 624 ± 20 nm light.

Light was delivered and collected using a 40X/1.1 NA water-immersion objective lens

(290 x 290-µm field-of-view). Images were collected through the full thickness of the

epithelium using a depth-sampling rate of 4-μm. Six individual frames were collected at

each depth. On average, 3 – 5 ROIs were sampled from each biopsy.

Morphological and Functional Metrics

Images were calibrated and processed as described in detail previously to extract

images that represented NAD(P)H and FAD TPEF intensity fluctuations[23–25,46]. At each

optical depth, NAD(P)H and FAD images were used to define a corresponding redox

ratio for each pixel of the field, as:

$$Optical\ Redox\ Ratio\ (RR) = \frac{FAD}{FAD + NAD(P)H} \qquad (1)$$

From the RR distributions for each OS, we calculated the mean RR and the interquartile

range (IQR) as metrics of the overall oxidation-reduction tissue state and the

corresponding heterogeneity, respectively. The mean and sample variance (variability)

of the mean OS RR and the OS RR IQR for all images in an epithelial stack were

calculated to assess the depth-dependence of these metrics.

NAD(P)H images were analyzed as described previously [17,18,21,22] using a Fourier based

approach to extract a value for the parameter β, as a metric of the level of mitochondrial

fragmentation and networking, which also depends highly on the metabolic activity of

the tissue. Briefly, an inverse power law was fit to the power spectral density (PSD) of

the 2D Fourier transform of the cytoplasmic NAD(P)H intensity fluctuation images, as:

$$R(k) = Ak^{-\beta} \qquad (2)$$

where R is the fit to the PSD, k is the magnitude of the spatial frequency, β is the power

law exponent, and A is a constant. The mean and sample variance of β were assessed

as a function of depth for each image stack.

Deep Learning Model Description

The basic structure of the CARE network has been described extensively (See

Supplementary Fig. S4**a** online)[33]. The network was implemented through Keras and

TensorFlow[47,48]. A copy of the CSBDeep repository

(https://github.com/CSBDeep/CSBDeep) was locally imported into an anaconda

environment[49]. The network was configured to take a 256 x 256 x 1 input image and

generate a 256 x 256 x 1 denoised image. A 40-gigabyte Nvidia Tesla A100 GPU card

was used for all training and evaluation. Typically, a 1 x 512 x 512 x z-depths image

stack was split into 4 x 256 x 256 x z-depths image patches before training. A starting

learning rate of 1×10^{-5} was used with an Adam optimizer[50]. Training was allowed to

continue for 300 epochs with a scheduler reducing the learning rate when network

performance stagnated for more than 20 epochs. Loss functions were varied to find the

optimal function to improve downstream analysis performance. Loss functions used

include SSIM Loss, R2 Loss, Focal Frequency Loss (FFL), MAE (L_1) Loss, MSE (L_2)

Loss, and combined losses such as a combined SSIM + L_2, SSIM + FFL, SSIM + R2

Loss[51]. Six down-sampling and up-sampling layers were generated with the first layer

expanding the single-channel images to thirty-two channels. Residual connections were

used to preserve encoded information from each down sampled layer and pass it

forward to the decoder layers (see Supplementary Figure S4 online).

For the WU-net architecture, four CARE networks, one per subband, were built as

described above. A DWT was used to decompose a 1 x 256 x 256 OSP into 1 x 128 x

128 x 4 frequency band images. The four frequency bands would then be individually

input to each CARE network for denoising. After denoising, an IDWT was used to

reconstruct the 1 x 256 x 256 OSP (for greater detail see Supplementary Fig. S5

online).

Training time typically varied from one to two hours, with an evaluation time of

approximately twenty-four seconds per image stack. For all trained CARE networks, 3D

SSIM, PSNR, Mean β, β Variability, Mean RR, RR Variability, RR IQR, RR IQR

Variability were analyzed. All final metrics were assessed using a single frame input,

denoised, and ground truth (6 frame averages) images with built-in and custom

MATLAB (MathWorks; Natick, MA) functions.

Statistics

For Figures 3-5**c** and Table 7, Fisher r-to-z transformation was used to convert

Pearson's correlation coefficients (r) to z_r values[52]. This transform was calculated using

Equation 3:

$$z_r = \frac{1}{2} \ln \left(\frac{1+r}{1-r} \right) \tag{3}$$

654   The z_r value, unlike r, belongs to a normal distribution, allowing for the calculation of a Z-

655   statistic to determine confidence intervals. To calculate the test Z-statistic for

656   comparison of z_r values to determine significance, Equation 4 was used:

$$Z_{test} = \frac{z_{r1} - z_{r2}}{\sqrt{\left(\frac{1}{n_1 - 3} + \frac{1}{n_2 - 3}\right)}} \qquad (4)$$

657   where n_1 and n_2 are the sample size of r_1 and r_2, respectively[53]. The Z_{test} value was then

658   compared to the critical Z-values to determine significance and p-values using a two-

659   tailed distribution.

Data Availability

661   The raw datasets used for model generation in the current study along with the trained

662   model weights are available from the corresponding author on reasonable request.

663   Codes for network training and prediction are publicly available at

664   https://gitlab.tufts.edu/georgakoudi-lab/Denoising2PImages

Acknowledgments

We would like to thank the National Institute of Biomedical Imaging and Bioengineering

(R01 EB030061), the National Institute of Health, Office of the Director (S10

OD021624), and the National Cancer Institute for funding this work (R03 CA235053).

The authors acknowledge the Tufts University High Performance Compute Cluster

(https://it.tufts.edu/high-performance-computing) which was utilized for the research

reported in this paper. The support of the Data Intensive Center (DISC) is

acknowledged. We would also like to thank Jasmine Kwan for her help in preparing

figures for this paper.

Author Contribution

I.G. conceived the initial goals of the study. Under the guidance of I.G., C.P. collected

all tissue biopsies used in this study and subsequently collected the image datasets.

H.T. screened and recruited patients for this study. E.G. and N.J. rendered biopsy

diagnoses after imaging was completed. N.V. wrote the initial implementations for

denoising and with F.S. further advanced the algorithms and developed a repository

used to denoise the collected images. C.P. aided in modifying existing downstream

analysis scripts for application on denoised images and completed statistical analysis

on downstream metrics. G.G. provided guidance in code implementation and data

augmentation. E.M., A.P., G.G., I.G., led discussions on interpretation of results and

methods for further optimization of generated models. I.G. supervised the study and

with N.V. and C.P. prepared the manuscript text. All authors have reviewed and

approved the manuscript.

Conflict of interest

The authors declare no competing interests.

References

1. Kaelin, W. G. & Thompson, C. B. Clues from cell metabolism. *Nature* **465**, 562–564 (2010).

2. DeBerardinis, R. J. & Thompson, C. B. Cellular Metabolism and Disease: What Do Metabolic Outliers Teach Us? *Cell* **148**, 1132–1144 (2012).

3. Kim, J. & DeBerardinis, R. J. Mechanisms and Implications of Metabolic Heterogeneity in Cancer. *Cell Metab.* **30**, 434–446 (2019).

4. Gialleonardo, V. Di, Wilson, D. M. & Keshari, K. R. The Potential of Metabolic Imaging. *Semin. Nucl. Med.* **46**, 28–39 (2016).

5. Lu, X. *et al.* LC–MS-based metabonomics analysis. *J. Chromatogr. B* **866**, 64–76 (2008).

6. So, P. T. C., Dong, C. Y., Masters, B. R. & Berland, K. M. Two-Photon Excitation Fluorescence Microscopy. *Annu. Rev. Biomed. Eng.* **2**, 399–429 (2000).

7.   Helmchen, F. & Denk, W. Deep tissue two-photon microscopy. *Nat. Methods* **2**, 932–940 (2005).

8.   Georgakoudi, I. & Quinn, K. P. Optical Imaging Using Endogenous Contrast to Assess Metabolic State. *Annu. Rev. Biomed. Eng.* **14**, 351–367 (2012).

9.   Heikal, A. A. Intracellular coenzymes as natural biomarkers for metabolic activities and mitochondrial anomalies. *Biomark. Med.* **4**, 241–263 (2010).

10. Spinelli, J. B. & Haigis, M. C. The multifaceted contributions of mitochondria to cellular metabolism. *Nat. Cell Biol.* **20**, 745–754 (2018).

11. Blacker, T. S., Duchen, M. R. & Bain, A. J. NAD(P)H binding configurations revealed by time-resolved fluorescence and two-photon absorption. *Biophys. J.* **122**, 1240–1253 (2023).

12. Chance, B., Schoener, B., Oshino, R., Itshak, F. & Nakase, Y. Oxidation-reduction ratio studies of mitochondria in freeze-trapped samples. NADH and flavoprotein fluorescence signals. *J. Biol. Chem.* **254**, 4764–4771 (1979).

13. Mayevsky, A. & Chance, B. Oxidation–reduction states of NADH in vivo: From animals to clinical use. *Mitochondrion* **7**, 330–339 (2007).

14. Kolenc, O. I. & Quinn, K. P. Evaluating Cell Metabolism Through Autofluorescence Imaging of NAD(P)H and FAD. *Antioxid. Redox Signal.* **30**, 875–889 (2019).

15. Georgakoudi, I. & Quinn, K. P. Label-Free Optical Metabolic Imaging in Cells and Tissues. *Annu. Rev. Biomed. Eng.* **25**, 413–443 (2023).

16. Tilokani, L., Nagashima, S., Paupe, V. & Prudent, J. Mitochondrial dynamics: overview of molecular mechanisms. *Essays Biochem.* **62**, 341–360 (2018).

17. Levitt, J. M. *et al.* Diagnostic cellular organization features extracted from autofluorescence images. *Opt. Lett.* **32**, 3305 (2007).

18. Xylas, J., Quinn, K. P., Hunter, M. & Georgakoudi, I. Improved Fourier-based characterization of intracellular fractal features. *Opt. Express* **20**, 23442 (2012).

19. Blinova, K. *et al.* Mitochondrial NADH Fluorescence Is Enhanced by Complex I Binding. *Biochemistry* **47**, 9636–9645 (2008).

20. Wai, T. & Langer, T. Mitochondrial Dynamics and Metabolic Regulation. *Trends Endocrinol. Metab.* **27**, 105–117 (2016).

21. Xylas, J. *et al.* Noninvasive assessment of mitochondrial organization in three-dimensional tissues reveals changes associated with cancer development. *Int. J. Cancer* **136**, 322–332 (2015).

22. Pouli, D. *et al.* Imaging mitochondrial dynamics in human skin reveals depth-dependent hypoxia and malignant potential for diagnosis. *Sci. Transl. Med.* **8**, (2016).

23. Quinn, K. P. *et al.* Quantitative metabolic imaging using endogenous fluorescence to detect stem cell differentiation. *Sci. Rep.* **3**, 3432 (2013).

24. Varone, A. *et al.* Endogenous Two-Photon Fluorescence Imaging Elucidates Metabolic Changes Related to Enhanced Glycolysis and Glutamine Consumption in Precancerous Epithelial Tissues. *Cancer Res.* **74**, 3067–3075 (2014).

25. Pouli, D. *et al.* Label-free, High-Resolution Optical Metabolic Imaging of Human Cervical Precancers Reveals Potential for Intraepithelial Neoplasia Diagnosis. *Cell Reports Med.* **1**, 100017 (2020).

26. Balu, M. *et al.* In Vivo Multiphoton Microscopy of Basal Cell Carcinoma. *JAMA Dermatology* **151**, 1068 (2015).

27. You, S. *et al.* Label-Free Deep Profiling of the Tumor Microenvironment. *Cancer Res.* **81**, 2534–2544 (2021).

28. Pshenay-Severin, E. *et al.* Multimodal nonlinear endomicroscopic imaging probe using a double-core double-clad fiber and focus-combining micro-optical concept. *Light Sci. Appl.* **10**, 207 (2021).

29. Shiu, J. *et al.* Multimodal analyses of vitiligo skin identifies tissue characteristics of stable disease. *JCI Insight* **7**, (2022).

30. Kučikas, V., Werner, M. P., Schmitz-Rode, T., Louradour, F. & van Zandvoort, M. A. M. J. Two-Photon Endoscopy: State of the Art and Perspectives. *Mol. Imaging Biol.* (2021). doi:10.1007/s11307-021-01665-2

31. Fan, L., Zhang, F., Fan, H. & Zhang, C. Brief review of image denoising techniques. *Vis. Comput. Ind. Biomed. Art* **2**, 7 (2019).

32. Monakhova, K., Richter, S. R., Waller, L. & Koltun, V. Dancing under the stars: video denoising in starlight. *Proc. IEEE Comput. Soc. Conf. Comput. Vis. Pattern Recognit.* **2022**-**June**, 16220–16230 (2022).

33. Weigert, M. *et al.* Content-aware image restoration: pushing the limits of fluorescence microscopy. *Nat. Methods* **15**, 1090–1097 (2018).

34. Ledig, C. *et al.* Photo-Realistic Single Image Super-Resolution Using a Generative Adversarial Network. *2017 IEEE Conf. Comput. Vis. Pattern Recognit.* **2017**-**Jan**, 105–114 (2016).

35. Chen, J. *et al.* Three-dimensional residual channel attention networks denoise and sharpen fluorescence microscopy image volumes. *Nat. Methods* **18**, 678–687 (2021).

36. Shen, B. *et al.* Deep learning autofluorescence-harmonic microscopy. *Light Sci. Appl.* **11**, 76 (2022).

37. Zhang, Y. *et al.* Image Super-Resolution Using Very Deep Residual Channel Attention Networks. *Lect. Notes Comput. Sci. (including Subser. Lect. Notes Artif. Intell. Lect. Notes Bioinformatics)* **11211 LNCS**, 294–310 (2018).

38. Hore, A. & Ziou, D. Image Quality Metrics: PSNR vs. SSIM. in *2010 20th*
International Conference on Pattern Recognition 2366–2369 (IEEE, 2010).
doi:10.1109/ICPR.2010.579

39. Ergen, B. Signal and Image Denoising Using Wavelet Transform. in *Advances in*
Wavelet Theory and Their Applications in Engineering, Physics and Technology
(ed. Baleanu, D.) (InTech, 2012). doi:10.5772/36434

40. Hongqiao, L. & Shengqian, W. A New Image Denoising Method Using Wavelet
Transform. in *2009 International Forum on Information Technology and*
Applications **1**, 111–114 (IEEE, 2009).

41. Luisier, F., Blu, T. & Unser, M. A New SURE Approach to Image Denoising:
Interscale Orthonormal Wavelet Thresholding. *IEEE Trans. Image Process.* **16**,
593–606 (2007).

42. Aytekin, C., Alenius, S., Paliy, D. & Gren, J. A Sub-band Approach to Deep
Denoising Wavelet Networks and a Frequency-adaptive Loss for Perceptual
Quality. *2021 IEEE 23rd Int. Work. Multimed. Signal Process.* 1–6 (2021).
doi:10.1109/MMSP53017.2021.9733576

43. Han, B. Wavelet Filter Banks. in 67–151 (2017). doi:10.1007/978-3-319-68530-
4_2

44. Goodfellow, I., Bengio, Y. & Courville, A. *Deep Learning*. (MIT Press, 2016).

45. Melanthota, S. K. *et al.* Deep learning-based image processing in optical
microscopy. *Biophys. Rev.* **14**, 463–481 (2022).

46. Quinn, K. P. *et al.* Characterization of metabolic changes associated with the
functional development of 3D engineered tissues by non-invasive, dynamic
measurement of individual cell redox ratios. *Biomaterials* **33**, 5341–5348 (2012).

47. Developers, T. TensorFlow. (2023). doi:10.5281/ZENODO.7916447

48. Chollet, F. & others. Keras. (2015).

49. Anaconda Software Distribution. *Anaconda Documentation* (2021).

50. Kingma, D. P. & Ba, J. Adam: A Method for Stochastic Optimization. *3rd Int. Conf.*
Learn. Represent. ICLR 2015 - Conf. Track Proc. 1–15 (2014).

51. Jiang, L., Dai, B., Wu, W. & Loy, C. C. Focal Frequency Loss for Image
Reconstruction and Synthesis. *Proc. IEEE Int. Conf. Comput. Vis.* 13899–13909
(2020). doi:10.1109/ICCV48922.2021.01366

52. Fisher, R. A. Frequency Distribution of the Values of the Correlation Coefficient in
Samples from an Indefinitely Large Population. *Biometrika* **10**, 507 (1915).

53. Fisher, R. A. On the 'Probable Error' of a Coefficient of Correlation as Deduce
From a Small Sample. *Metron* **1**, 205–235 (1921).

813