Baited reconstruction with 2D template matching for high-resolution structure determination *in vitro* and *in vivo* without template bias

Bronwyn A. Lucas[a,b,c,*], Benjamin A. Himes[a,d] and Nikolaus Grigorieff[a,d,*]

[a] RNA Therapeutics Institute, University of Massachusetts Chan Medical School, Worcester, MA

[b] Department of Molecular and Cell Biology, University of California Berkeley, Berkeley, CA

[c] Center for Computational Biology, University of California Berkeley, Berkeley, CA

[d] Howard Hughes Medical Institute

*Correspondence: bronwynlucas@berkeley.edu, niko@grigorieff.org

Abstract

Cryogenic electron microscopy (cryo-EM) has revolutionized structural biology, rapidly

increasing the number of available molecular structures. Because of this, as well as advances in

structure prediction, the focus of structural biology has begun to shift to studying

macromolecular structures in their native cellular environment. A dominant feature of cryo-EM

images is shot noise, making the identification of small particles of interest difficult. This is

further compounded by structural noise if these particles are imaged against a background of

other molecules, such as inside a cell. 2D template matching (2DTM) can be used to localize

complexes with high precision, even in the presence of cellular background. Once localized,

these particles may be averaged together in 3D reconstructions; however, regions included in the

template may suffer from template bias, leading to inflated resolution estimates and making the

interpretation of high-resolution features unreliable. We evaluate conditions that minimize

template bias and show that molecular features *not* present in the template can be reconstructed

at high resolution from targets found by 2DTM, extending prior work at low-resolution.

Moreover, we present a quantitative metric for template bias to aid the interpretation of 3D

reconstructions calculated with particles localized using high-resolution templates and fine

angular sampling.

Introduction

Over the last decade, single-particle electron cryo-microscopy (cryo-EM) has emerged as a high-resolution technique to study molecules and their assemblies in a near-native state (Guaita et al., 2022). In the most favorable cases, close to 1 Å resolution can be achieved, rivaling results obtained by protein crystallography (Nakane et al., 2020; Yip et al., 2020). The resolution obtained from a single-particle dataset depends on the quality of the images, the accuracy of particle alignment and imaging parameters, the structural integrity of the sample, and the number of particles contributing to a reconstruction. For a high-quality dataset, between 20,000 and 70,000 asymmetric units of well-aligned and homogeneous particles have to be averaged to reach sub 2 Å resolution (Nakane et al., 2020; Yip et al., 2020). Methods development in a related cryo-EM technique has also enabled imaging particles *in situ* at high resolution using tomography and subtomogram averaging. These *in situ* subtomogram averages now approach 3 Å resolution (Tegunov et al., 2021), a resolution obtained routinely for single-particle reconstructions *in vitro*. Data collection for tomography requires more time compared to the single-particle technique due to the need for acquiring a tilt series, and processing tends to be computationally more expensive due to the additional degrees of freedom, compared to 2D images used in the single-particle technique. An additional complication of averaging images of molecules for *in situ* structure determination is the selection of valid targets, which have to be identified against a background of other molecules inside the cell or tissue being imaged. This is in contrast to a typical single-particle dataset, in which the particles have undergone a purification step that enriches the particle of interest, and imaged in solvent which makes particle selection more reliable.

2D template matching (2DTM) is an approach that can be used to identify target molecules and complexes in cryo-EM images of cells and cell sections, using single images of nominally untilted specimens (Lucas et al., 2022, 2021; Rickgauer et al., 2020, 2017). This approach can be used in combination with 3D template matching (3DTM) to identify targets in tomograms collected from the same areas imaged for 2DTM. 2DTM is fundamentally limited by the background generated by overlapping molecules in untiled views of the sample, imposing a size limit on what can be detected (Rickgauer et al., 2017). A combined approach of 2D and 3DTM

could benefit from the strengths of both approaches, with better target detection (lower false

negative rate) of 3DTM in the tomograms, and the improved overall precision of 2DTM in the

untiled views (Lucas et al., 2021). In our previous studies, we have demonstrated that the targets

detected using 2DTM can be used to calculate 3D reconstructions showing novel details not

present in the template (Lucas et al., 2022, 2021; Rickgauer et al., 2017). 3D reconstruction is

straightforward because for every detected target, 2DTM also determines their x,y location in the

image, three Euler angles and image defocus, i.e., all the parameters needed to calculate a single-

particle reconstruction. Using this approach, we revealed non-modeled density for the viral

polymerase (VP1) bound to a rotavirus capsid (Rickgauer et al., 2017), for the small ribosomal

subunit (SSU) and tRNAs (Lucas et al., 2022, 2021), as well as structural differences between *M.*

pneumoniae and *B. subtilis* large ribosomal subunits (LSUs) (Lucas et al., 2021). Interpretation

of reconstructions obtained from 2DTM targets can be hindered by template bias (Lucas et al.,

2022, 2021), i.e., the reproduction of modeled features included in the template, that are

reproduced in the reconstructions, even though they do not correspond to structural features in

the detected particles. This could result from inclusion of pure noise particles and/or local

overfitting of particle parameters in the presence of signal. Our previous studies showed that

template bias does not prevent the discovery of new structural features at low resolution that

were not represented by the template, but it has yet to be determined if this is true for high-

resolution features, which are more susceptible to noise overfitting (Stewart and Grigorieff,

2004). 2DTM could in principle be used to study the structure of targets at high-resolution, that

would otherwise be too small to identify on their own, as long as they bind or are otherwise

rigidly attached to a larger target that can be located by 2DTM. However, due to limitations in

the number and heterogeneity of particles in previous studies, it was unclear whether this

approach could indeed recover reliable high-resolution information.

In the present study, we explore this possibility further to assess the resolution that can be

obtained in unmodeled regions omitted from the template. We analyze a published single particle

dataset of β-galactosidase using 2DTM, and 60S LSUs detected in images of *S. cerevisiae*

lamellae. In both cases, we show high resolution in areas of the reconstruction that were omitted

in the template, demonstrating the utility of 2DTM for structure discovery. We present a new

metric to quantify template bias in a template-based 3D reconstruction, making reconstruction
from 2DTM targets a more broadly useful tool.
97
98 **Results**
99

Reconstruction of the beta-galactosidase ligand binding pocket
To show the potential of 2DTM to reveal new structural details at high resolution, we analyzed a
published single-particle cryo-EM dataset of *E. coli* β-galactosidase (Bgal) bound to phenylethyl
β-D-thiogalactopyranoside (PETG) (Saur et al., 2020). The dataset was used previously to
calculate a 2.2 Å single-particle reconstruction (EMDB-10574) that displays density for a
number of specifically bound water molecules in the structure, including in the PETG binding
pocket. The authors also built an atomic model into the high-resolution map (PDB: 6TTE). For
our template, however, we used an atomic model of ligand-free Bgal determined by X-ray
crystallography at 1.7 Å (PDB: 1DP0) (Juers et al., 2000). Using a model that was built into a
map that is independent from the data analyzed by 2DTM aids our demonstration of 2DTM as a
tool that can make use of atomic models experimentally unrelated to the data being analyzed.
111
To demonstrate high resolution in areas omitted in the template, we removed atoms in the
vicinity of all D2 symmetry-related ligand binding pockets, within a 10 Å radius centered on the
side chain amide nitrogen atom of asparagine 102 (in PDB: 1DP0). The truncated atomic model
was used to generate a template with *cis*TEM's simulator (Himes and Grigorieff, 2021) (see
Methods). We searched 558 micrographs downloaded from the EMPIAR database (EMPIAR-
10644) and obtained 59,259 targets with 2DTM SNRs above a threshold of 7.3 (**Figure 1A**), the
standard threshold calculated to limit the average number of false positives (false positive rate) to
one per micrograph, based on the given search parameters and a Gaussian noise model
(Rickgauer et al., 2017) (see Eq. (2) below). To reduce the particles to a number closer to the
final dataset used to calculate the 2.2 Å cryo-EM map in (49,895), and to enrich for the particles
most similar to the template (similar to selecting the best classes in Saur et al., 2020), we limited
our targets to those with 2DTM SNRs above 9.0 and obtained a final dataset of 55,627 particles.
124

The identified 55,627 targets were extracted together with their template-matched x,y positions,

Euler angles and CTFFIND4-derived defocus values using prepare_stack_matchtemplate (Lucas

et al., 2021), and the particle stack and alignment parameters were imported into *cis*TEM as a

refinement package for further single particle processing. The Fourier Shell Correlation (FSC)

(Harauz and Heel, 1986) for the initial reconstruction calculated from the template-matched

alignment parameters indicated a resolution of 2.4 Å (FSC = 0.143) (Rosenthal and Henderson,

2003). We performed further refinement against the template while keeping the refinement

resolution limit of 3.0 Å – one cycle of defocus and beam tilt refinement, followed by a

refinement of alignment parameters and another cycle of defocus and beam tilt parameters. The

final reconstruction (**Figure 1A-C**) displayed a resolution according to the FSC of 2.2 Å (**Figure**

1—figure supplement 1F). As mentioned above and previously discussed (Lucas et al., 2021),

resolution estimates based on the FSC can be affected by template bias. Therefore, the present

estimate has to be considered unreliable, and has to be supported by additional evidence, such as

inspection of features visible in the density map.

Peaks corresponding to detected targets are clearly visible (**Figure 1—figure supplement 1B**).

The average 2DTM SNR for this dataset was 11.6, and a maximum of 16.3, which is in the range

of what is expected for a 465-kDa target (Rickgauer et al., 2020, 2017). The refined

reconstruction shows clear density for PETG and water molecules in the ligand binding pocket

that were omitted in the template (**Figure 1C**). Comparison of this reconstruction with the

published map (**Figure 1G**) suggests that they are virtually identical and that there is little or no

evidence of template bias in the 2DTM reconstruction. An assessment of the local resolution

using Phenix (Liebschner et al., 2019) further indicates a resolution of about 1.8 Å in the binding

pocket, consistent with the clear density for water.

Baited reconstruction visualizes ribosomes at near atomic resolution in FIB-milled lamellae

To investigate whether 2DTM can be used to generate reliable high-resolution reconstructions

from images derived from cellular samples, we used a previously published dataset of 37 images

of 4 FIB-milled lamellae generated from *S. cerevisiae* cells treated with the translation inhibitor

cycloheximide (CHX) to enrich the ribosome population in a single state (**Figure 2A**) (Lucas and

Grigorieff, 2023). The lamella samples were not tilted during data collection and therefore

exhibited a small tilt with respect to the electron beam of about 8° due to the milling angle

during sample preparation. We identified 12210 large ribosomal subunits with 2DTM in the

cytoplasm using a threshold of 7.85 which corresponds to an expectation of one false positive per

image across the dataset, or ~0.3% of the particles (**Figure 2—figure supplement 1A-C**). Local

positional and orientational refinement was performed using the *cis*TEM program

refine_template (Lucas et al., 2021) and the original template as a reference. The refined 2DTM

coordinates were used to calculate an initial reconstruction with a nominal resolution of 3.15 Å

(FSC = 0.143) (Rosenthal and Henderson, 2003). One cycle of beam tilt refinement against the

reconstruction improved the resolution to 3.1 Å (**Figure 2A,B**). Unlike for the *in vitro* Bgal

reconstruction, further refinement of the other alignment parameters using the reconstruction as a

reference caused the resolution to decrease to 8 Å. The reconstruction has reduced signal at high

spatial frequencies relative to the template as indicated by the half-map FSC (**Figure 2—figure**

supplement 1D). This, combined with the higher background and lack of low-resolution contrast

of the ribosomes in the cellular lamella relative to a purified sample, may reduce the alignment

accuracy relative to the high-resolution 2DTM template. This highlights the importance of high

spatial frequencies for alignment of particles in images with strong background, in contrast to

images of purified samples that show strong low resolution features, which are important for

reliable particle alignment (Stewart and Grigorieff, 2004).

As previously reported (Lucas et al., 2022, 2021), we found density consistent with the small

ribosomal subunit (SSU) and tRNAs, that did not derive from the template. In the present case,

local resolution estimation shows that parts of the SSU are resolved at < 4 Å resolution (**Figure**

2B). The SSU is conformationally variable and shows considerable positional heterogeneity

relative to the LSU. Therefore, this value is likely an underestimate of the potential attainable

resolution in reconstructions from 2DTM targets in cells. Although the map represents an

average of all states identified, we observed clear density for tRNAs in the A/A and P/P state

with apparent density for the polypeptide on the A site tRNA (**Figure 2—figure supplement 2**)

and no clear density for E-site tRNAs. This allowed us to conclude that CHX stalls ribosomes in

the classical PRE translocation state *in vivo*, likely by preventing transition of the P/P tRNA to

the P/E state consistent with an *in vitro* structure of the translating *Neurospora crassa* ribosome

(Shen et al., 2021) and inference from ribosome profiling data (Lareau et al., 2014; Wu et al.,

2019). The relatively low resolution of the tRNAs likely reflects the mixed pool of tRNA

depending on the codons on which the ribosome stalled as well as a mixture of states.

Visualization of drug-target interactions in cells

Drug-target interactions can be visualized at high-resolution *in vitro* with cryo-EM and X-ray

crystallography. However, it is unclear whether this recapitulates the binding site *in vivo,*

possibly missing weak interactions that are disrupted during purification. Visualizing drug-target

interactions in cells is therefore an important goal. We observed additional density near the

ribosomal E site not present in the template that is consistent with the position of CHX in a

previously published crystal structure (Loubresse et al., 2014) (**Figure 2C**). The density was

sufficiently well resolved to dock CHX and provide *in vivo* confirmation for the position and

orientation of CHX binding in the E site.

We noted several key differences between the model built from the *in vitro* CHX-bound structure

and the *in situ* CHX-bound structure. Firstly, we did not observe density for eIF5A but did

observe density consistent with binding of spermidine (**Figure 2D**), as has been observed

previously for the *in vitro* CHX-bound *Neurospora crassa* ribosome (PDB: 7R81) (Shen et al.,

2021). This demonstrates that spermidine can bind to ribosomes within cells, however, whether

spermidine binds as part of the translation cycle or whether stalling of translation with CHX

allowed for spermidine to bind is unclear. Baited reconstruction with 2DTM could be used to

further probe the function of polyamides to regulate translation *in vivo*.

Baited reconstruction using the large ribosomal subunit as a template model allowed us to

visualize the binding of small molecules such as drugs and polyamides to the ribosome within

cells. This demonstrates the power of baited reconstruction to reveal biologically relevant

features that would only be evident at high resolution.

Omit templates reveal high resolution features without template bias

The local resolution of parts of the LSU were measured at ~3 Å, however, this region overlapped

with the template and therefore the resolution measure using standard tools may be unreliable.

To assess the resolution in this region we repeated this experiment with a template that lacked

the ribosomal protein L7A. Since this protein was not present in the template, any density in this

region cannot be due to template bias. We found that the local resolution of L7A was

indistinguishable from the surrounding density and showed varying local resolution from 3.2 to

4.5 Å (**Figure 3A-B**). The density was sufficiently well resolved to observe side chains in

regions that were lacking from the template (**Figure 3C**). This suggests that baited reconstruction

with 2DTM coordinates can be used to generate high-resolution reconstructions from cellular

samples, free from template bias, and demonstrates an approach to verify local resolution

estimates.

To examine the recovery of high-resolution information with single residue precision, we

generated another truncated template by removing every 20^{th} residue from each chain. This

resulted in a total reduction of 51 kDa or ~3% of the template mass. We then localized 12090

targets using the same 2DTM protocol as for the full-length template. The small difference in

template mass minimally affected target detection, only 120 targets (<1%) were missed, and

there were minimal deviations in the locations and orientations for the remaining targets. The 3D

reconstruction generated from the detected targets showed clear density corresponding to

nucleotides (**Figure 3D,E**) and various amino acids (**Figure 3F-K**) that were missing from the

template and therefore cannot derive from template bias. This demonstrates that omitting small

randomly scattered regions from a 2DTM template can be used to assess template bias

throughout the reconstruction.

Quantifying template bias

The calculation of reconstructions from targets identified by 2DTM, which relies on *a priori*

structural models, bears the danger of generating results that reproduce features of the template

even when these features are absent from the targets to be detected. In the field of cryo-EM, this

is often referred to as the "Einstein from noise" problem (Henderson, 2013). The risk of template

bias increases with dataset size (number of images), as well as the ratio of false positives vs true

positives. Template bias in reconstructions generated from 2DTM targets is generally avoided

because the scoring function (SNR threshold) is set to reject most false positives. To quantify

template bias in reconstructions at various 2DTM SNR thresholds, we generated a series of

reconstructions at different thresholds using targets identified with a full-length LSU template

("full" template) and the template lacking 3% of the residues ("omit" template) covering

different areas of the model, while retaining most detections relative to the full template as

described above (**Figure 4A**). We wrote a new *cis*TEM program measure_template_bias (see

Methods) that calculates the difference between map densities ρ_{full} and ρ_{omit} in these

reconstructions, in the omitted regions:

$$\Omega = \frac{\rho_{full} - \rho_{omit}}{\rho_{full}}$$

(1)

As expected, for high 2DTM SNR thresholds (few or no false positives), the template bias Ω was

only a few percent, while for lower thresholds, it approached 100% (**Figure 4A**). This was

consistent with increased density in the reconstructions using the full template relative to the

omit template (**Figure 4C**). The observed lower limit of Ω of ~8% (**Figure 4A**) is likely due to

some overfitting of noise when template-matching true particles, rather than inclusion of false

positives. This overfitting may manifest itself in small alignment errors of the targets against the

matching template, and a bias of these errors towards compensating for any mismatch between

target and template, such as omitted regions in the template. Further work to quantify Ω at

different spatial frequencies will be informative to assess the contribution of local overfitting to

template bias.

If we assume that the template bias is proportional to the rate of false positive detection, r_f, we

can plot the expected false positive rate, $r_{f,model}$, against the observed template bias Ω (**Figure

4B**). The expected false positive rate is given by the complementary error function (Rickgauer et

al., 2017) as

$$r_{f,model} = \frac{1}{2} erfc\left(\frac{SNR_t}{\sqrt{2}}\right)$$

(2)

where SNR_t is the 2DTM SNR threshold applied to the template search results. The plot shows

that the template bias is not proportional to the expected false positive rate (**Figure 4B**). This is

likely due to the variable background found in images of lamellae, which means that the spectral

whitening that is applied to the images before the search (Rickgauer et al., 2017) does not whiten

all areas of the images evenly. This results in local deviations of the background (noise)

distribution from the Gaussian noise model implied in Eq. (2), leading to higher-than-expected

false positive ratios at low SNR thresholds.

If we estimate the number of true targets at 13456 (the number of targets identified by both

templates at a threshold of 7.85) and recalculate the number of false positives as the overall

number of detected targets in excess of this number, the template bias is approximately

proportional to the false positive rate (red line in **Figure 4B**). Further work is required to develop

an improved noise model that predicts the correct number of false positives in images of variable

contrast, such as images of cellular lamellae. Furthermore, it is important to note that the 2DTM

SNR threshold used here to exclude most of the false positives also leads to a rejection of true

positives. The number of these false negatives depends on the 2DTM SNR generated by the

targets, which is proportional to their molecular mass (**Figure 4—figure supplement 1**). For our

data and 150 nm thick lamellae, this means that targets below about 300 kDa will not be detected.

Improvements in cryo-EM instrumentation, sample preparation, image processing and 2DTM

methods will lower this limit (Russo et al., 2022).

Discussion

We show here that baited reconstruction with 2DTM can reveal high resolution detail in regions

not modeled in the template. Using a previously published single particle dataset we observe

interactions between specific sidechains with water and a ligand. Using particles localized in

FIB-milled yeast lamellae, we observe specific binding of the drug CHX and polyamides to the

ribosome in cells. We show that baited reconstruction can be used to recover high-resolution

features in cells without template bias in regions omitted from the template, and quantify

template bias in regions overlapping the template. The use of 2D images to generate high

resolution reconstructions makes this process significantly faster and less computationally

expensive relative to tomography. Baited reconstruction is analogous to a "pull down" assay in

molecular biology, wherein a "bait" molecule is used to capture and identify novel interacting

"prey". This strategy is distinct from prior structure determination strategies because it makes

use of a high-resolution template, traditionally avoided to prevent introducing template bias

artifacts (Henderson, 2013). Baited reconstruction leverages the advantages of precise targeting

with a high-resolution template, while avoiding the template bias by focusing on regions omitted

from, or external to the template. Baited reconstruction can therefore leverage the wealth of

existing structural data, as well as molecular models generated by the newly available structure

prediction tools (Baek et al., 2021; Evans et al., 2022; Jumper et al., 2021), to approach

biological and pharmacological questions *in vitro* and *in vivo*.

Implications for drug discovery

One of the most direct applications of this approach is to drug discovery. During the drug

development pipeline, potentially thousands of variants of a lead compound are tested relative to

a single protein target. Determining the structures of each in complex with its protein partner

using the traditional single-particle cryo-EM workflow can be time-consuming and laborious,

and often requires image processing expertise. The strategy presented here could be used to

streamline this process substantially.

The ribosome is a major target of antibiotic and anticancer drugs. We have demonstrated that

baited reconstruction with 2DTM can reveal drug-ribosome interactions directly in cells. The

reconstructions are at comparable resolution relative to the state-of-the-art from tomography,

while using a more streamlined data collection and processing pipeline that could be easily

automated. This approach could therefore be used to more efficiently characterize the

mechanism of action of antibiotic drugs directly in cells. Since 2DTM does not require

purification, the interactions with other cellular complexes can also be investigated.

2DTM accelerates high-resolution in situ structure determination

Baited reconstruction is substantially faster and a more streamlined pipeline for *in situ* structure

determination compared to cryo-ET and subtomogram averaging. Current pipelines for *in situ*

structure determination using cryo-ET and subtomogram averaging are time-consuming and

require expert knowledge to curate an effective pipeline. We expect focused classification to

identify sub-populations to further improve the resolution of *in situ* reconstructions from 2DTM

targets. To help classify particles against a cellular background without introducing alignment

errors (see Results), alignment parameters (Euler angles, x,y shifts) can remain fixed. While

tomograms are required to provide the cellular 3D context of molecules, our work shows that it

is not always necessary to use tomography to generate high-resolution reconstructions of

macromolecular complexes in cells. 2DTM could reduce the manual effort and time for structure

determination in cells when compared to subtomogram averaging, depending on the time it takes

to annotate, refine and classify the subtomograms.

Our approach also differs from the recently published isSPA method (Cheng et al., 2023, 2021).

isSPA follows the traditional single-particle workflow, applied to particles in their native

environment. Particles are selected using a template that is limited to an intermediate resolution

of 8 Å, resulting in an initial particle stack that contains many false positives. Selection of the top

scores, followed by standard single-particle classification and alignment protocols then yields

reconstructions of the detected targets. This approach is particularly successful in situations

where there is a high concentration of the particle of interest, such as Rubisco inside the

carboxysome (Cheng et al., 2021), capsid proteins in viral capsids (Cheng et al., 2021), and

phycobilisome and photosystem II in the thylakoid membranes inside *Pennisetum purpureum*

cells (Cheng et al., 2023). In contrast, using 2DTM, we select targets with high fidelity, avoid

false positives, and determine the molecule pose to high accuracy without the need for an

intermediate reconstruction from the detected targets to act as reference for further refinement.

By using the full resolution of the signal present in the images, 2DTM is also more sensitive than

isSPA, detecting particles of 300 kDa in 150 nm lamellae (Figure 4-figure supplement 1). These

differences mean that 2DTM can be used with fewer, and potentially smaller particles to achieve

high-resolution structures compared to isSPA and other techniques following the canonical

single-particle averaging workflow. As demonstrated here, the detection criterion used in 2DTM

largely avoids overfitting artifacts in reconstructions by eliminating images that are not

statistically distinguishable from noise. This makes 2DTM particularly useful for *in situ* structure

determination, which is often limited by the low abundance of the target complexes inside the

cell. By reducing the number of particles needed to achieve high-resolution reconstructions in

cells, baited reconstruction with 2DTM will make it possible to determine the structures of less

abundant complexes in cells.

Application to in vitro single particle analysis

Our results of a single particle dataset of purified Bgal demonstrates another use case for 2DTM.

In the original analysis of this dataset using the traditional single-particle workflow, 136013

particles were initially selected using template-based particle picking (Gautomatch 0.56,

http://www.mrc-lmb.cam.ac.uk/kzhang/) (Saur et al., 2020). 2D classification, ab-initio

reconstruction, and further 3D classification eventually yielded a 2.2 Å reconstruction showing

the bound ligand (PETG). The same result was achieved with a simple run of 2DTM, without

requiring manual intervention or expert knowledge in the image processing workflow. In a

separate 2DTM search using the first 277 images of the dataset and a crystal structure of GroEL

(PDB: 1GRL) as a template— a particle of comparable size to Bgal— we detected only 53

targets above the default SNR threshold (excluding two images that had sharp black lines across

them), and none above a threshold of 9.0. This further demonstrates the high level of

discrimination of 2DTM between true and false positives, as shown earlier (Rickgauer et al.,

2017). Besides the streamlined workflow, 2DTM can therefore also be used in the presence of

impurities to reliably select the particles of interest. Using multiple templates, particles could be

classified to arrive at quantitative estimates of particles occupying defined conformational states.

The reduced need for sample purity and dataset size to perform such analyses may further

accelerate the 2DTM workflow, compared to the traditional single particle workflow, provided

appropriate templates are available.

Furthermore, validation of map and model quality is a major challenge in cryo-EM. Current

methods use low-pass filtered templates to avoid template bias at high-spatial frequencies. We

here present a quantitative estimate of local and global template bias in sequence space. This will

allow the full resolution of the template to be used to localize particles more specifically and

avoid false positives. This may assist in identification of particle classes in the localization stage

and can streamline the reconstruction process. Estimating template bias with baited

reconstruction can provide a quantitative metric of map and model quality that may find broad

utility in single particle and in situ workflows.

Application to subtomogram averaging

Recently, higher resolution template matching and finer angular sampling have also been

explored for the analysis of cryo-ET 3D reconstructions (Chaillet et al., 2023; Cruz-León et al.,

2023). This approach has clear advantages because it reduced false positives due to low-
resolution overlap (Chaillet et al., 2023; Cruz-León et al., 2023) and provides more specific
localization of targets in a crowded cellular environment. However, if the identified targets are
subsequently used for subtomogram averaging, the reconstructions may exhibit template bias.
Both baited reconstruction and the quality metrics we describe above could be applied to
subtomogram averaging pipelines.
407
Future applications
We have shown that it is possible to recover single residue detail, and even the location of water
molecules in the most favorable cases, using baited reconstruction with cryo-EM. This approach
is analogous to the use of OMIT maps in X-ray crystallography to avoid model bias (Bhat and
Cohen, 1984; Hodel et al., 1992) and the M-free used to estimate reference bias in subtomogram
averaging (Yu and Frangakis, 2014). Our approach differs by sampling random residues
throughout the sequence and consequently provides higher precision in the estimation of
template bias at high resolution. The observation that reconstructions with negligible template
bias can be determined using particles identified with high-resolution template matching depends
fundamentally on the noise model and threshold used to identify true positives and exclude false
positives. We observe that the number of false positives does not perfectly match predictions
based on a white Gaussian noise model, suggesting that the background is not perfectly Gaussian
everywhere, for example due to local features with strong low-resolution contrast. When the
noise model is uncertain or inaccurate, thresholding alone may not be sufficient to remove false
positives. It is therefore important to validate features in reconstructions from targets found by
template matching if they overlap with the template. In addition, overfitting could be assessed
using the Omega metric described here, to quantify template bias in regions important for the
study.
426
By further analogy to X-ray crystallography, the strategy we presented here could be extended
by tiling through the template model, omitting overlapping features and combining the densities
in each omitted region to form a continuous 3D map in which the density for each residue was
omitted from the template, comparable in principle to a composite OMIT map (Terwilliger et al.,
2008). While currently computationally expensive and therefore not feasible in most cases, this

strategy could be regarded as a "gold standard", yielding reconstructions that are devoid of

template bias while retaining the benefits of precise localization and identification of the targets.

If only some map regions are validated, as was done in the examples presented here, it is likely

that the rest of the 3D map is also reliable, based on the assumption that false positives were

excluded from the reconstruction. However, this reasoning may not strictly hold when there is

partial and variable mismatch between the targets and the template, for example due to

conformational heterogeneity in the detected target population. In such a situation, template bias

may not be uniform across the reconstruction, and template bias has to be assessed more

rigorously.

Materials and Methods

Yeast culture and FIB-milling

Saccharomyces cerevisiae strains BY4741 (ATCC) colonies were grown to mid log phase in

YPD, diluted to 10,000 cells/mL and treated with 10 µg/mL cycloheximide (Sigma) for 10 min

at 30 °C with shaking as described in (Lucas and Grigorieff, 2023). 3 µL was applied to a 2/1 or

2/2 Quantifoil 200 mesh SiO_2 Cu grid, allowed to rest for 15 s, back-side blotted for 8 s at 27 °C,

95% humidity followed by plunge freezing in liquid ethane at –184 °C using a Leica EM GP2

plunger. Frozen grids were stored in liquid nitrogen until FIB-milled. FIB-milling was performed

as described in (Lucas and Grigorieff, 2023).

Cryo-EM data collection and image processing

Bgal micrograph movie data were downloaded from the EMPIAR database (EMPIAR-10644)

and processed with the *cis*TEM image processing package (Grant et al., 2018) using Unblur

(Grant and Grigorieff, 2015) to align and average the exposure-weighted movie frames, and

CTFFIND4 (Rohou and Grigorieff, 2015) to determine image defocus values. Four of the 562

micrographs were discarded based on lack of clear CTF Thon rings or ice crystal contamination.

The remaining 558 images were processed using *cis*TEM's template matching implementation

(Lucas et al., 2021), yielding 59,259 targets with 2DTM SNRs above a threshold of 7.3.

Cryo-EM images of the yeast cytoplasm were previously published using imaging and

processing pipelines as described in (Lucas and Grigorieff, 2023), except that an additional 7

images were included that were previously excluded because they contained organelle regions.

Simulating 3D templates

The atomic coordinates from the indicated PDBs were used to generate a 3D volume using the

*cis*TEM (Grant et al., 2018) program simulate (Himes and Grigorieff, 2021). For the Bgal

template, we used a pixel size of 0.672 Å, which is slightly smaller than published for this

dataset (0.68 Å). The smaller pixel size was obtained by fitting the 1.7 Å X-ray structure (PDB:

1DP0) into the published 2.2 Å cryo-EM map of PETG-bound Bgal, and adjusting the pixel size

of the map to achieve optimal density overlap between model and map in UCSF Chimera

(Pettersen et al., 2004). Details on template generation are summarized in Table 1.

2D Template Matching

2DTM was performed using the program *match_template* (Lucas et al., 2021) implemented in

the *cis*TEM graphical user interface (Grant et al., 2018). For the Bgal searches, an in-plane

angular step of 1.5º and an out-of-plane angular step of 2.5º, and D2 symmetry was used (no

defocus search). This yielded a threshold of 7.30 calculated from a total number of $\sim 6.88 \times 10^{12}$

search locations, identifying targets with an average of one false positive per image.

For the LSU, an in-plane angular step of 1.5º and an out-of-plane angular step of 2.5º, C1

symmetry and defocus search of ± 1200 Å with a 200 Å step was used. This yielded a threshold

of 7.85 calculated from a total number of $\sim 4.88 \times 10^{14}$ search locations, identifying targets with

an average of one false positive per image.

Generating 3D reconstructions

The *cis*TEM program prepare_stack_matchtemplate (Lucas et al., 2021) was used to generate

particle stacks from the refined coordinates from the 2DTM searches followed by reconstruction

using the *cis*TEM program reconstruct3d as described in the text. Local resolution estimation

was performed using the local resolution tool in Phenix (Liebschner et al., 2019) using a box size

of 7 Å (Bgal) or 12 Å (ribosome). To visualize regions of the ribosome reconstruction outside of

the LSU template we used the UCSF ChimeraX (Pettersen et al., 2021) volume tools to segment

the map using a radius of 3 Å from the template atoms. UCSF Chimera (Bgal) (Pettersen et al.,

2004) or ChimeraX (ribosome) (Pettersen et al., 2021) were used for visualization.

Quantifying template bias

We wrote a program, measure_template_bias, which is part of the *cis*TEM software (Grant et al.,

2018, source code available at github.com/timothygrant80/cisTEM, executibles available at

cistem.org), to assess the degree of template bias present in a reconstruction, calculated from

detected 2DTM targets. The program requires two templates on input, one template representing

the full structure of the targets to be found (full template), and one containing omitted elements

of the structure that serve as test regions to assess template bias (omit template). The program

also requires the two reconstructions that were calculated form targets detected by these two

templates (full reconstruction and omit reconstruction). The two templates and the two

reconstructions have to be identically density-scaled, respectively. Using the two templates,

measure_template_bias calculates a difference map that leaves only non-zero densities in areas

omitted in the omit template. The difference map is then used as a mask to identify the test

regions used to assess template bias. The densities in the test regions are summed for the two

input reconstructions, yielding ρ_{full} and ρ_{omit}, respectively. The average degree of template

bias (Ω) is then defined as the difference between ρ_{full} and ρ_{omit}, relative to ρ_{full} (Eq. (1)). Ω

can assume values between 0 and 1 (100%), with 0 representing the least degree of template bias,

and 1 representing the highest degree of template bias. If the degree of template bias has to be

evaluated more locally, measure_template_bias also accepts a difference map, instead of the two

templates, that will be used to identify the areas to be used for measuring template bias.

Acknowledgements

The authors thank the members of the Grigorieff lab for helpful discussions. We are also grateful

for the use of and support from the cryo-EM facility at UMass Chan Medical School.

Competing interests

The authors are listed as inventors on a closely related patent application named "Methods and

Systems for Imaging Interactions Between Particles and Fragments", filed on behalf of the

University of Massachusetts.

Funding

BAL and NG gratefully acknowledge funding from the Chan Zuckerberg Initiative, grant #
2021-234617 (5022).

Author contributions

Bronwyn A Lucas, Conceptualization, Data curation, Formal analysis, Funding acquisition,
Validation, Investigation, Visualization, Methodology, Writing - original draft, Project
administration, Writing – review and editing; Benjamin A Himes, Conceptualization, Formal
analysis, Writing – review and editing; Nikolaus Grigorieff, Conceptualization, Software, Formal
analysis, Supervision, Funding acquisition, Methodology, Writing - original draft, Project
administration, Writing – review and editing

References

Baek M, DiMaio F, Anishchenko I, Dauparas J, Ovchinnikov S, Lee GR, Wang J, Cong Q, Kinch LN, Schaeffer RD, Millán C, Park H, Adams C, Glassman CR, DeGiovanni A, Pereira JH, Rodrigues AV, Dijk AAV, Ebrecht AC, Opperman DJ, Sagmeister T, Buhlheller C, Pavkov-Keller T, Rathinaswamy MK, Dalwadi U, Yip CK, Burke JE, Garcia KC, Grishin NV, Adams PD, Read RJ, Baker D. 2021. Accurate prediction of protein structures and interactions using a three-track neural network. *Science* **373**:871–876.

Bhat TN, Cohen GH. 1984. OMITMAP: An electron density map suitable for the examination of errors in a macromolecular model. *J Appl Crystallogr* **17**:244–248. doi:10.1107/S0021889884011456

Chaillet ML, van der Schot G, Gubins I, Roet S, Veltkamp RC, Förster F. 2023. Extensive Angular Sampling Enables the Sensitive Localization of Macromolecules in Electron Tomograms. *Int J Mol Sci* **24**:13375. doi:10.3390/ijms241713375

Cheng J, Li B, Si L, Zhang X. 2021. Determining structures in a native environment using single-particle cryoelectron microscopy images. *The Innovation* **2**:100166. doi:10.1016/j.xinn.2021.100166

Cheng J, Liu T, You X, Zhang F, Sui S-F, Wan X, Zhang X. 2023. Determining protein structures in cellular lamella at pseudo-atomic resolution by GisSPA. *Nat Commun* **14**:1282. doi:10.1038/s41467-023-36175-y

Cruz-León S, Majtner T, Hoffmann PC, Kreysing JP, Tuijtel MW, Schaefer SL, Geißler K, Beck M, Turoňová B, Hummer G. 2023. High-confidence 3D template matching for cryo-electron tomography. doi:10.1101/2023.09.05.556310

Evans R, O'Neill M, Pritzel A, Antropova N, Senior A, Green T, Žídek A, Bates R, Blackwell S, Yim J, Ronneberger O, Bodenstein S, Zielinski M, Bridgland A, Potapenko A, Cowie A, Tunyasuvunakool K, Jain R, Clancy E, Kohli P, Jumper J, Hassabis D. 2022. Protein complex prediction with AlphaFold-Multimer. *bioRxiv* 2021.10.04.463034. doi:10.1101/2021.10.04.463034

Grant T, Grigorieff N. 2015. Measuring the optimal exposure for single particle cryo-EM using a 2.6 Å reconstruction of rotavirus VP6. *eLife* **4**:e06980. doi:10.7554/eLife.06980

Grant T, Rohou A, Grigorieff N. 2018. CisTEM, user-friendly software for single-particle image processing. *eLife* **7**:e35383. doi:10.7554/eLife.35383

Guaita M, Watters SC, Loerch S. 2022. Recent advances and current trends in cryo-electron microscopy. *Curr Opin Struct Biol* **77**:102484. doi:10.1016/j.sbi.2022.102484

Harauz G, Heel M van. 1986. Exact filters for general geometry three dimensional reconstruction. *Optik* **73**:146–156.

Henderson R. 2013. Avoiding the pitfalls of single particle cryo-electron microscopy: Einstein from noise. *Proc Natl Acad Sci U S A* **110**:18037–18041. doi:10.1073/PNAS.1314449110

Himes B, Grigorieff N. 2021. Cryo-TEM simulations of amorphous radiation-sensitive samples using multislice wave propagation. *IUCrJ* **8**:943–953. doi:10.1107/S2052252521008538

Hodel A, Kim S-H, Brünger AT. 1992. Model bias in macromolecular crystal structures. *Acta Crystallogr A* **48**:851–858. doi:10.1107/S0108767392006044

Juers DH, Jacobson RH, Wigley D, Zhang X-J, Huber RE, Tronrud DE, Matthews BW. 2000. High resolution refinement of β-galactosidase in a new crystal form reveals multiple

metal-binding sites and provides a structural basis for α-complementation. *Protein Sci* **9**:1685–1699. doi:10.1110/PS.9.9.1685

Jumper J, Evans R, Pritzel A, Green T, Figurnov M, Ronneberger O, Tunyasuvunakool K, Bates R, Žídek A, Potapenko A, Bridgland A, Meyer C, Kohl SAA, Ballard AJ, Cowie A, Romera-Paredes B, Nikolov S, Jain R, Adler J, Back T, Petersen S, Reiman D, Clancy E, Zielinski M, Steinegger M, Pacholska M, Berghammer T, Bodenstein S, Silver D, Vinyals O, Senior AW, Kavukcuoglu K, Kohli P, Hassabis D. 2021. Highly accurate protein structure prediction with AlphaFold. *Nature* **596**:583–589. doi:10.1038/s41586-021-03819-2

Lareau LF, Hite DH, Hogan GJ, Brown PO. 2014. Distinct stages of the translation elongation cycle revealed by sequencing ribosome-protected mRNA fragments. *eLife* **3**:e01257. doi:10.7554/eLife.01257

Liebschner D, Afonine PV, Baker ML, Bunkoczi G, Chen VB, Croll TI, Hintze B, Hung LW, Jain S, McCoy AJ, Moriarty NW, Oeffner RD, Poon BK, Prisant MG, Read RJ, Richardson JS, Richardson DC, Sammito MD, Sobolev OV, Stockwell DH, Terwilliger TC, Urzhumtsev AG, Videau LL, Williams CJ, Adams PD. 2019. Macromolecular structure determination using X-rays, neutrons and electrons: recent developments in Phenix. *Urnissn2059-7983* **75**:861–877. doi:10.1107/S2059798319011471

Loubresse NGD, Prokhorova I, Holtkamp W, Rodnina MV, Yusupova G, Yusupov M. 2014. Structural basis for the inhibition of the eukaryotic ribosome. *Nat 2014 5137519* **513**:517–522. doi:10.1038/nature13737

Lucas BA, Grigorieff N. 2023. Quantification of gallium cryo-FIB milling damage in biological lamellae. *Proc Natl Acad Sci* **120**:e2301852120. doi:10.1073/pnas.2301852120

Lucas BA, Himes BA, Xue L, Grant T, Mahamid J, Grigorieff N. 2021. Locating macromolecular assemblies in cells by 2D template matching with cisTEM. *eLife* **10**. doi:10.7554/elife.68946

Lucas BA, Zhang K, Loerch S, Grigorieff N. 2022. In situ single particle classification reveals distinct 60S maturation intermediates in cells. *eLife* **11**. doi:10.7554/ELIFE.79272

Nakane T, Kotecha A, Sente A, McMullan G, Masiulis S, Brown PMGE, Grigoras IT, Malinauskaite L, Malinauskas T, Miehling J, Uchański T, Yu L, Karia D, Pechnikova EV, Jong E de, Keizer J, Bischoff M, McCormack J, Tiemeijer P, Hardwick SW, Chirgadze DY, Murshudov G, Aricescu AR, Scheres SHW. 2020. Single-particle cryo-EM at atomic resolution. *Nat 2020 5877832* **587**:152–156. doi:10.1038/s41586-020-2829-0

Pettersen EF, Goddard TD, Huang CC, Couch GS, Greenblatt DM, Meng EC, Ferrin TE. 2004. UCSF Chimera—A visualization system for exploratory research and analysis. *J Comput Chem* **25**:1605–1612. doi:10.1002/JCC.20084

Pettersen EF, Goddard TD, Huang CC, Meng EC, Couch GS, Croll TI, Morris JH, Ferrin TE. 2021. UCSF ChimeraX: Structure visualization for researchers, educators, and developers. *Protein Sci Publ Protein Soc* **30**:70. doi:10.1002/PRO.3943

Rickgauer JP, Choi H, Lippincott-Schwartz J, Denk W. 2020. Label-free single-instance protein detection in vitrified cells. *bioRxiv* 2020.04.22.053868. doi:10.1101/2020.04.22.053868

Rickgauer JP, Grigorieff N, Denk W. 2017. Single-protein detection in crowded molecular environments in cryo-EM images. *eLife* **6**:e25648. doi:10.7554/elife.25648

Rohou A, Grigorieff N. 2015. CTFFIND4: Fast and accurate defocus estimation from electron micrographs. *J Struct Biol* **192**:216–221. doi:10.1016/J.JSB.2015.08.008

Rosenthal PB, Henderson R. 2003. Optimal Determination of Particle Orientation, Absolute Hand, and Contrast Loss in Single-particle Electron Cryomicroscopy. *J Mol Biol* **333**:721–745. doi:10.1016/J.JMB.2003.07.013

Russo C, Dickerson JL, Naydenova K. 2022. Cryomicroscopy in situ: what is the smallest molecule that can be directly identified without labels in a cell? Faraday Discuss. **240**:277-302. doi:10.1039/d2fd00076h

Saur M, Hartshorn MJ, Dong J, Reeks J, Bunkoczi G, Jhoti H, Williams PA. 2020. Fragment-based drug discovery using cryo-EM. *Drug Discov Today* **25**:485–490. doi:10.1016/J.DRUDIS.2019.12.006

Shen L, Su Z, Yang K, Wu C, Becker T, Bell-Pedersen D, Zhang J, Sachs MS. 2021. Structure of the translating Neurospora ribosome arrested by cycloheximide. *Proc Natl Acad Sci U S A* **118**:e2111862118. doi:10.1073/PNAS.2111862118

Stewart A, Grigorieff N. 2004. Noise bias in the refinement of structures derived from single particles. *Ultramicroscopy* **102**:67–84. doi:10.1016/j.ultramic.2004.08.008

Tegunov D, Xue L, Dienemann C, Cramer P, Mahamid J. 2021. Multi-particle cryo-EM refinement with M visualizes ribosome-antibiotic complex at 3.5 Å in cells. *Nat Methods* **18**:186–193. doi:10.1038/s41592-020-01054-7

Terwilliger TC, Grosse-Kunstleve RW, Afonine PV, Moriarty NW, Adams PD, Read RJ, Zwart PH, Hung L-W. 2008. Iterative-build OMIT maps: map improvement by iterative model building and refinement without model bias. *Acta Crystallogr D Biol Crystallogr* **64**:515–524. doi:10.1107/S0907444908004319

Wu ACC-C, Zinshteyn B, Wehner KA, Correspondence RG, Wu CC-C, Green R. 2019. High-Resolution Ribosome Profiling Defines Discrete Ribosome Elongation States and Translational Regulation during Cellular Stress Article High-Resolution Ribosome Profiling Defines Discrete Ribosome Elongation States and Translational Regulation during Cellular Stress. *Mol Cell* **73**:959–970. doi:10.1016/j.molcel.2018.12.009

Yip KM, Fischer N, Paknia E, Chari A, Stark H. 2020. Atomic-resolution protein structure determination by cryo-EM. *Nat 2020 5877832* **587**:157–161. doi:10.1038/s41586-020-2833-4

Yu Z, Frangakis AS. 2014. M-free: Scoring the reference bias in sub-tomogram averaging and template matching. *J Struct Biol* **187**:10–19. doi:10.1016/J.JSB.2014.05.007

Figure Legends

Figure 1: Baited reconstruction for visualization of Beta galactosidase ligand binding pocket at high resolution. (A) Reconstruction of Bgal from 2DTM coordinates using images from a previously published dataset (Saur et al., 2020) using a Bgal crystal structure (PDB: 1DP0) (Juers et al., 2000) as a template, with a 10 Å sphere around the PETG ligand omitted. **(B)** A 2D slice through the reconstruction in (A) including the region deleted from the density shows no obvious discontinuity in the density. **(C)** A view of the density in (A) indicated with a red box, with regions within 1.8 Å of the template model highlighted in red. Gray indicates density of Bgal outside of the template, purple indicates density consistent with the position of PETG and blue indicates additional density that likely represent water molecules. **(D)** A stick diagram showing the locations of the atoms in the template used for template matching. **(E)** Published density from (Saur et al., 2020) aligned and scaled as in (A). **(F)** As in (B), showing a region of the published density in (E). **(G)** As in (C), showing the same region of the published density in (E). **(H)** As in (D), showing all atoms annotated in the crystal structure, including those omitted before generating the 2DTM template.

Figure 2: Visualizing drugs and small molecules bound to the ribosome *in vivo*. (A) A reconstruction of the ribosome from 2DTM coordinates identified in the cytoplasm of FIB-milled *Sacchromyces cerevisiae* cell sections showing clear density for both the 60S (part of the template) and the 40S (outside of the template). **(B)** A slice of the reconstruction in (A), indicating the local resolution using the indicated color coding. The arrow indicates the P-site tRNA. **(C)** Regions of the density >3 Å from the template model are indicated in pink. The crystal structure PDB: 4U3U was aligned with the template model and the position of cycloheximide (CHX) was not altered. **(D)** As in C, showing density corresponding to a spermidine (PDB: 7R81) and unaccounted for density outside of the template (black arrow), which may also represent a polyamide.

Figure 3: Baited reconstruction reveals high-resolution features *in vivo* without template bias. (A) Slice of a reconstruction using 2DTM coordinates identified with a template lacking the protein L7A. Color coding indicates the local resolution as indicated in the key. **(B)** As in (A),

pink indicates the 2DTM template used to identify the targets used in the reconstruction. **(C)** The

model PDB: 6Q8Y is shown in the density. Red corresponds to the protein L7A, which was

omitted from the template used to identify targets for the reconstruction. Blue corresponds to

model features that were present in the template. **(D)** Single nucleotide omit template and **(E)**

reconstruction showing emergence of density outside of the template, including a phosphate

bulge, black arrow. Single amino acid omit templates lacking Phe **(F)**, Arg **(H)** or Ser **(J)** and

density **(G, I, K)**, respectively showing emergence of features consistent with each amino acid.

Figure 4: Baited reconstruction provides quantitative metric for template bias. (A)

Observed template bias (Ω) calculated using the *cis*TEM program measure_template_bias as a

function of the 2DTM SNR threshold used to select targets from images of yeast lamellae. Blue

arrows indicate the reconstructions shown in **Figure 4C**. **(B)** Plot showing a comparison of the

predicted false positive rate and the observed Ω. The plotted straight line indicates the best fit

linear function $y = 0.96x - 0.05$. **(C)** Images showing the same region of maps resulting from

reconstruction using targets identified with the indicated template at the indicated 2DTM SNR

threshold. Red indicates the location of the omitted residue in the omit template.

Figure 1—figure supplement 1: (A) Example micrograph from (Saur et al., 2020). Scale bar

indicates 200 Å. **(B)** Maximum intensity projection showing the maximum 2DTM SNR at each

pixel after a 2DTM search of the image in (A). **(C)** 2D projections indicating the best matching

2D templates of Bgal in the image in (A). **(D)** Estimated local resolution of the reconstruction in

Figure 1A, as indicated by the color key. **(E)** Region of the reconstruction in (D) shown in

Figure 1B. (F) FSC evaluated between the half-maps as a function of spatial frequency.

Figure 2—figure supplement 1: (A) Area of an example micrograph showing the yeast

cytoplasm from (Lucas et al., 2023). Scale bar indicates 200 Å. **(B)** Maximum intensity

projection showing the highest 2DTM SNR at each pixel after a 2DTM search of the image in (A)

with an LSU template. **(C)** 2D projections indicating the best matching 2D templates in the

image in (A). **(D)** Plot showing the FSC of the two half maps of the reconstruction shown in

Figure 2. **(E)** Density showing the region of the map in **Figure 2D** color coded by the local

density.

Figure 2—figure supplement 2: Cycloheximide stalls the ribosome in a non-rotated PRE-translocation state *in vivo*. (A) Local resolution filtered map shown in Figure 2, only showing features outside of the template shows density consistent with A/A and P/P tRNAs. **(B)** The same region of the map rotated to show weak density in the polypeptide exit tunnel consistent with the nascent polypeptide attached to the A-site tRNA, consistent with the non-rotated PRE-translocation state.

Figure 4—figure supplement 1: Plot showing the average relative number of LSUs detected in 7 images of ~150 nm thick lamellae using templates generated from subsections of the LSU of the indicated molecular mass. The x-intercept indicates the smallest particle that was detectable on average, which in this case is 300 kDa.

735   **Table 1:** Preparation and simulation of the 3D templates used in this study

Template name	PDB	PDB modified?	Resolution of PDB map (Å)	PDB B-factor scaling	Additional B-factor applied (Å^2)	Pixel size (Å)	Box size (pixels)
Bgal	1DP0	10 Å sphere around Asp 102 deleted HETATOMs excluded	1.7	0.0	50.0	0.672	512
LSU	6Q8Y	Only atomic coordinates corresponding to the LSU included HETATOMs excluded	3.1	0	30	1.06	384
LSU (ΔL7A)	6Q8Y	Only atomic coordinates corresponding to the LSU included Atomic coordinates corresponding to L7A excluded HETATOMs excluded	3.1	0	30	1.06	384

Figure 1

Figure 1-figure supplement 1

Figure 2

Figure 2-figure supplement 1

Figure 2-figure supplement 2

Figure 3

A

B
2DTM template lacking L7A

C
L7A
Model in 2DTM template
Model not in 2DTM template

D Template

E Reconstruction from 2DTM coordinates

F Template

G Reconstruction from 2DTM coordinates

H

I

J

K

Figure 4

A

B

Gaussian Prediction

Consistent #TP

C

Reconstruction (Full template)

Reconstruction (Omit template)

Threshold: 7.04
Particles: 63994, 61433
Observed Ω: 88.2%

Threshold: 7.51
Particles: 20254, 19713
Observed Ω: 36.4%

Threshold: 9
Particles: 11410, 11081
Observed Ω: 8.04%

Figure 4-figure supplement 1