ATUM-Tomo: A multi-scale approach to cellular ultrastructure by combined volume scanning electron microscopy and electron tomography

Georg Kislinger[1,2], Gunar Fabig[3], Antonia Wehn[4,5], Lucia Rodriguez[1,2], Hanyi Jiang[1,2,6], Cornelia Niemann[1,2], Andrey S. Klymchenko[7], Nikolaus Plesnila[4,8], Thomas Misgeld[1,2,8], Thomas Müller-Reichert[3], Igor Khalin[4,9,#] and Martina Schifferer[2,8,#,*]

[1]Institute of Neuronal Cell Biology, Technical University Munich, Munich, Germany

[2]German Center for Neurodegenerative Diseases (DZNE), Munich, Germany

[3]Experimental Center, Faculty of Medicine Carl Gustav Carus, Technische Universität Dresden, Dresden, Germany

[4]Institute for Stroke and Dementia Research (ISD), LMU University Hospital, LMU Munich, Germany

[5]Department of Neurosurgery, University of Munich Medical Center, Munich, Germany

[6]Department of Psychiatry and Psychotherapy, University Medicine Greifswald, Greifswald, Germany

[7]Laboratoire de Biophotonique et Pharmacologie, CNRS UMR 7213, Université de Strasbourg, Illkirch, France

[8]Munich Cluster of Systems Neurology (SyNergy), Munich, Germany

[9]Normandie University, UNICAEN, INSERM UMR-S U1237, Physiopathology and Imaging of Neurological Disorders (PhIND), GIP Cyceron, Institute Blood and Brain @ Caen-Normandie (BB@C), 14 074 Bd Henri Becquerel, 14000 Caen, France

[#]These authors contributed equally to this work.

*For correspondence: martina.schifferer@dzne.de

Abstract

Like other volume electron microscopy approaches, Automated Tape Collecting Ultramicrotomy (ATUM) enables imaging of serial sections deposited on thick plastic tapes by scanning electron microscopy (SEM). ATUM is unique in enabling hierarchical imaging and thus efficient screening for target structures, as needed for correlative light and electron microscopy. However, SEM of sections on tape can only access the section surface, thereby limiting the axial resolution to the typical size of cellular vesicles with an order of magnitude lower than the acquired xy resolution. In contrast, serial-section electron tomography (ET), a transmission electron microscopy-based approach, yields isotropic voxels at full EM resolution, but requires deposition of sections on electron-stable thin and fragile films, thus making screening of large section libraries difficult and prone to section loss. To combine the strength of both approaches, we developed 'ATUM-Tomo', a hybrid method, where sections are first reversibly attached to plastic tape via a dissolvable coating, and after screening detached and transferred to the ET-compatible thin films. ATUM-SEM of serial semi-thick sections and

consecutive ET of selected re-mounted sections combines SEM's fast target recognition and coarse rendering capability with high-resolution volume visualizations, respectively, thus enabling multi-scale interrogation of cellular ultrastructure. As a proof-of-principle, we applied correlative ATUM-Tomo to study ultrastructural features of blood brain barrier (BBB) leakiness around microthrombi in a mouse model of traumatic brain injury. Microthrombi and associated sites of BBB leakiness were identified by confocal imaging of injected fluorescent and electron-dense nanoparticles, then relocalized by ATUM-SEM, and finally interrogated by correlative ATUM-Tomo. This workflow is a seamless zoom-in on structural BBB pathology from the micro- to the nanometer scale. Overall, our new ATUM-Tomo approach will substantially advance ultrastructural analysis of biological phenomena that require cell- and tissue-level contextualization of the finest subcellular textures.

Introduction

Volume electron microscopy (EM) provides high-resolution data of target structures that can be used to render complex biological morphologies in three dimensions (Peddie *et al.*, 2022). While volume EM thus provides high spatial resolution, searching for a rare nanometer-sized structure of interest in sections of several hundreds of square micrometers is tedious. To bridge these scales, multimodal approaches like correlative light and electron microscopy (CLEM) (Scher and Avinoam, 2021) or micro-computed tomography combined with volume EM (Karreman *et al.*, 2017) have been applied. Among the available volume EM approaches, array tomography techniques (Wacker and Schroeder, 2013), such as automated tape collecting ultramicrotomy (ATUM) (Kasthuri *et al.*, 2015), have proven particularly powerful for correlating specific biological structures (Schifferer *et al.*, 2021; Snaidero *et al.*, 2020). Array tomography techniques require section collection onto a solid substrate, such as coated glass, silicon (Micheva and Smith, 2007) or plastic tape, and subsequent serial-section scanning electron microscopy (SEM). In ATUM-SEM, a reel-to-reel system transfers the sections from the diamond knife onto tape (Kubota *et al.*, 2018). The tape strips are assembled on silicon wafers (Baena *et al.*, 2019; Kislinger *et al.*, 2023), thus generating tissue libraries that can be repetitively imaged by SEM at different resolution regimes. This separates the correlation task into two phases: (1) low-resolution screening for the region previously imaged by fluorescence microscopy and (2) imaging of re-located structures by high-resolution volume EM. High-resolution imaging by array tomography is, however, currently restricted by physical sectioning, resulting in typical voxels of 3×3×30-100 nm, where SEM determines the lateral resolution of ≈3 nm. As a consequence, array tomography can reveal fine cellular processes down to ≈40 nm, but potentially misses smaller subcellular structures, such as vesicles or contact sites that can be hidden within a section volume. For higher-resolution imaging, typically transmission electron microscopy (TEM) is applied, with electron tomography (ET) being the method of choice for high-resolution imaging of sub-cellular volumes (West *et al.*, 2011; Mastronarde, 2005; He and He, 2014; Kiewisz *et al.*, 2022; Redemann *et al.*, 2017). In ET, semi-thick sections (200-400 nm) are tilted in the electron beam and Fourier transformed for image reconstruction (Baumeister, Grimm and Walz, 1999; Young and Villa, 2023; Frank, 2005), but even

with serial semi-thick sections (Soto *et al.*, 1994), the overall volume that can be reconstructed and
visualized in three dimensions is limited.

So far, ATUM could not be combined with ET because the former necessitates attaching sections on a
solid support, while the latter as a projection method, needs sections to be deposited on electron-
transparent films. While section deposition is usually irreversible in ATUM, we reasoned that plastic
tape coating could enable section removal after SEM and transfer onto ET-compatible films. Thus,
selected semi-thick sections could be re-inspected by ET – a novel multimodal approach, which we
call 'ATUM-Tomo'. We exemplify the ATUM-Tomo approach by assessing structural determinants of
blood-brain barrier (BBB) leakiness, a key adaptation of the neurovascular unit that controls substance
exchange between blood and central nervous system parenchyma and can be locally disturbed in
pathologies, including after ischemia, traumatic brain injury or multiple sclerosis (Zhang, Jin and
Faber, 2019; Shi *et al.*, 2020). We had previously characterized BBB leakage around impact-induced
microclots using correlative confocal microscopy and ATUM of fluorescent lipid nanodroplets (LNDs)
and gold nanoparticles ((Khalin *et al.*, 2022). At the BBB, the actual barrier function is determined by
distinctive ultrastructural features, mostly within the endothelial layer, including tight junctions,
pinocytosis and vesicular transport (O'Brown, Megason and Gu, 2019; Andreone *et al.*, 2017;
Nahirney, Reeson and Brown, 2016; Hirano, Kawanami and Llena, 1994). As these structures lie
below the 30 nm z-resolution limit of ATUM, they eluded our previous correlative imaging. Here we
demonstrate that triple correlation using confocal microscopy, volume SEM in combination with ET
provides an integrated view of the cellular environment as well as the ultrastructural features of the
compromised BBB. Our application shows that ATUM-Tomo in combination with CLEM is a powerful
multimodal approach, enabling a bridging of scales from the mm to the nm range.

Results

Correlative confocal imaging and ATUM-SEM reveals sites of BBB disruption

Controlled cortical impact in mice is a model for traumatic brain injury that involves BBB damage. To
mark sites of vessel injury, we injected LNDs and colloidal gold particles of similar size (30 nm) and
surface chemistry (PEGylation) ensuring similar biodistribution. LNDs are highly fluorescent, while gold
is electron dense – and both accumulate in microthrombi formed in the impact penumbra within 30-60
min post-injection (Khalin *et al.*, 2022), enabling the tracking of microthrombi and BBB leakage sites.

These particles were injected systemically 60 min after impact, followed by DyLight 649 lectin as a

vessel lumen marker just before perfusion with fixative containing glutaraldehyde, which all

fluorophores withstand (Fig. 1A). We sliced the fixed brains in 80 µm thick coronal vibratome sections

(Fig. 1A) to allow confocal imaging and preserve some longer stretches of blood vessels, while

restricting the axial depth of the search volume for correlation. After post-fixation in glutaraldehyde,

sections were temporarily mounted on a glass slide with a cover slip and recorded ≈2×2 mm confocal

overview of all cortical layers bearing the lesion and peri-lesional areas (Fig. 1B). For correlative EM,

we identified areas of vascular occlusions indicated by LNDs accumulations and sites of extravasation

marked by perivascular LND fluorescence and locally acquired higher magnification confocal image

stacks (Fig. 1C). Tissue preservation was ensured by imaging for a maximum of one hour, followed by

gentle removal of the cover slip by floating-off in the buffer.

In order to keep the section flat during EM processing and to minimize the search volume, we
dissected the smallest possible tissue block around the region of interest based on tissue fiducials as
a guide. The tissue was processed for EM by applying a standard reduced-osmium-
thiocarbohydrazide-osmium (rOTO) *en bloc* staining protocol (Hua, Laserstein and Helmstaedter,
2015) in a way that facilitates the axial and lateral relocation of the area of interest (Kislinger *et al.*,
2020a). Production of serial thin (100 nm) sections and collection of these sections on either carbon-
coated Kapton (ccKapton) (Kasthuri *et al.*, 2015) or carbon nanotube (Kubota *et al.*, 2018) tape
generated a library of ribbons. A typical 1000 section library was collected, which could then be
imaged hierarchically and repetitively. For an overview to facilitate correlation, we imaged the entire
area of every 10[th] section (i.e., 1000 nm axial sampling) with 200×200 nm lateral sampling. This
allowed us to efficiently screen such a serial-section library for regions of interest, even if the confocal
imaging and ultramicrotomy planes were not perfectly aligned.

Regions with microthrombi and extravasation were identified by correlation and selected for high-
resolution (4×4×100 nm) data acquisition by volume SEM. Microthrombi were characterized by a high
LND signal, while sites of extravasation showed a perivascular halo. The microclot illustrated here
consisted of luminal erythrocyte stalls (Fig. 1D, G) in combination with platelets and immune cells (Fig.
1D, H). An accumulation of platelets, interspersed with filamentous, electron-dense fibrin was found in
the center and at the ends, "plugging" the occlusion (Fig. 1D, F, H). Correlating to the fluorescence
pattern, we detected gold particles and brighter but less discrete particles of the same size,
presumably representing LNDs, between luminal cells (Fig. 1D). Healthy endothelium was
characterized by an intact glycocalyx and hence lectin staining in the absence of LND fluorescence. In
contrast, at sites of extravasation indicated by an LND halo, we found a neutrophil undergoing
diapedesis (Fig. 1E) and abluminal gold particles in its proximity (Fig. 1F). Thus, we can assume that
some microthrombi could serve not only as entry points for nanoparticles, but also for cells from blood.
Within the microthrombus periphery, regions of the endothelium were swollen with intact tight
junctions, while in the core, regions with almost no discrete endothelial layer prevailed (Fig. 1H).
Colloidal gold was found at pathologically thinned endothelium, basement membrane and in pericytes
(Fig. 1F-H).

Development of reversible section attachment on coated tape for ATUM-Tomo

ATUM-SEM revealed the cellular composition of vascular occlusions and ultrastructural features of

BBB leakiness. While transcellular or paracellular pathways are supposed to be involved, so far, early

transport mechanisms have not been fully understood. The existence of morphologically normal tight

junctions visualized by ATUM-SEM suggests that vesicular transport might be involved. However,

resolving gold-particle loaded vesicles at 20-40 nm diameter or tight junctions challenges the axial

resolution of ATUM-SEM (≈30 nm). Therefore, we sought to combine ATUM-SEM and ET, as the

latter would provide the required resolution (Michel, 2012; Feng *et al.*, 2002; Wagner *et al.*, 2012), but

the former would allow efficient screening across an expanded field of view (Peddie *et al.*, 2022). For

this purpose, we devised a new method to reversibly attach semi-thick ATUM sections to plastic tape

for SEM and then transfer them to TEM grids for ET.

We first tested if conventionally collected sections of mouse cortical tissue could be removed from

ccKapton or carbon nanotube tape. However, such sections were stuck to the tape despite treatment

with organic solvents and any other chemical or physical treatments that we tested (Fig. 2A). We then

hypothesized that adding a thin, dissolvable coating between tape and section could allow gentle

detachment (Fig. 2B, C). Polyvinyl formal (Formvar) was a good candidate, as thin films of this highly

solvable polymer are commonly used to support sections for TEM. Moreover, it has been shown that

sections can be transferred from Formvar-coated glass coverslips onto TEM grids by hydrofluoric acid

(Paez-Segala *et al.*, 2015). Therefore, we tested if an ultrathin tissue section could be removed from

Formvar-coated tape by organic solvents. Short ccKapton tape strips were Formvar-coated using a

standard film casting device (c**f**cKapton), onto which 100-300 nm tissue sections were collected from

the diamond knife water bath. Indeed, such sections could be recovered from the tape by incubation in

chloroform or 1,2-dichlorethan. To coat longer stretches of tape, we designed a reel-to-reel coating

device with a spraying unit (Figure 2 - figure supplement 1). We mounted such longer cfcKapton tape

with tissue sections onto wafers, as required for ATUM-SEM, and subjected them to SEM imaging. We

found that Formvar coating induced charging, depending on the thickness of the coat (Fig. 2D). In

addition, imaging at high resolution (<10 nm) prevented subsequent section detachment (Figure 2 -

figure supplement 2).

In order to avoid these problems, we considered other materials that are soluble in organic solvents,

and settled on the simple and readily available solution of permanent markers. Marker pens consist of

genuine pigmented alcohol-based ink and humectants while a special resin ensures that it is

waterproof. We labeled ccKapton with a thin film of permanent marker with wide tip (cpcKapton) and

found that sections attached to such tape could be detached by intense rinsing with acetone. The

resulting floating sections could be transferred to a water bath in a drop of acetone using a single-use

Pasteur pipette. For longer stretches of tape as needed for serial sections, we rebuilt the coating unit

by installation of a permanent marker instead of the spraying unit (Fig. 2E).

Next, we assessed the SEM imaging characteristics of tissue sections mounted on cpcKapton

attached to silicon wafers. Coatings with permanent marker of red color, but not black color partially

detached from the plastic support and lead to strong charging during SEM imaging (Fig. 2D). Tissue

sections attached to cpcKapton, using pens of either color, showed major tissue section folds (Fig.

2D). In order to remove these folds, we plasma discharged black cpcKapton tape extensively (see
Material and Methods) until the film interference color changed from black to magenta (mild
discharging) to gold (extensive discharging) (Fig. 2F). Such plasma-discharged cpcKapton enabled
reversible section attachment at certain SEM imaging conditions (10 nm lateral resolution at 3 µs dwell
time) with ultrastructural quality comparable to ccKapton (Fig. 2G).

We then assessed the tissue ultrastructure in the TEM and the possibility to relocate regions
previously imaged by SEM. For this purpose, 80 nm sections on cpcKapton were wafer-mounted,
imaged by SEM and transferred to TEM grids. Despite extensive acetone rinsing the sections stayed
intact with well-preserved ultrastructure. The *en bloc* embedding provided strong TEM image contrast
without further poststaining. If the section area exceeds the slot size, we deposited the region of
interest at the grid center. We could efficiently relocate the previously SEM-imaged region in the TEM
by using tissue-intrinsic fiducials at low resolution (Fig. 2H). However, we detected dark deposits
(Figure 2 - figure supplement 3A) on some regions of the section, which probably originate from tissue
contact with the coating and accordingly are only superficial. Because ET mainly visualizes the inner
tissue areas of the semi-thick sections, we considered these deposits as unproblematic, and
cpcKapton hence as a potentially suitable support material for combining ATUM-SEM and ET to
bridge across scales (Fig. 3A, B). We further applied this approach to assess BBB integrity after
traumatic brain injury.

ATUM-CLEM and ATUM-Tomo of nanoparticles at the leaky BBB

To this end, confocal imaging and EM processing of a traumatic brain injury lesion were performed as
described for correlative ATUM-SEM. We collected serial 400 nm semi-thick sections of the lesion
area covering the entire volume of the vibratome section on cpcKapton. Typically, this resulted in 250
sections that can be mounted onto one or two wafers, dramatically speeding up the correlation
process compared to ATUM-SEM (1000 sections on four wafers). We imaged the entire section areas
at 100 nm lateral resolution to screen for rough anatomical landmarks. Two vessels with high LND
fluorescence signal (Fig. 3C) were correlated and selected for imaging at higher resolution (20×20
nm). Based on this ultrastructural volume and in order to unambiguously identify the localization of
nanoparticles, we acquired single images of these target regions at 10 nm lateral resolution. Alignment
and three-dimensional rendering of the volume EM data (Fig. 3D) provided an overview of the
underlying neurovascular unit despite the coarser axial resolution compared to ATUM-SEM. Based on
these reconstructions and inspection of the ultrastructural volume, we selected sections with gold
particles potentially crossing the BBB. These tissue sections were detached by acetone treatment and
collected onto Formvar-coated slot grids, as needed for subsequent ET. We found that colloidal gold
fiducials (20 nm diameter) needed for image reconstruction could be attached to the surface of these
sections despite previous SEM imaging. Tomographic reconstructions of semi-thick sections removed
from cpcKapton showed surface contamination only at the side that was facing the permanent marker
coat (Figure 2 – figure supplement 3B). The superficial contamination caused reconstruction artifacts
in the back projection down to down to a depth of up to 50 nm into the section. Sections that were
collected at a nominal cutting thickness of 400 nm showed an actual thickness of 260-300 nm after the
tomographic reconstruction. This collapse is due to a mass loss after interaction of the electron beam

with the specimen (Luther, Lawrence and Crowther, 1988; McEwen and Marko, 1998; O'Toole,
Morphew and McIntosh, 2020). Moreover, we detected variable lateral shrinkage of about 20% and
minor distortions between the SEM and TEM images of ultrathin sections, probably due to beam
damage and acetone treatment (Figure 2 - figure supplement 4). The sections could be recovered
even four months after collection and nine months after SEM imaging. Besides Epon/Araldite resin
(LX112), we tested the detachment and re-imaging procedure with tissue sections embedded in
standard Epon and Durcupan resins (Figure 2 - figure supplement 5). For all three types of resins we
were able to image by SEM, detach sections and collect them onto grids for re-imaging by TEM
tomography.

Despite the above-mentioned minor collapse of the section, the features of the targeted microthrombi
and the surrounding neurovascular unit familiar from conventional thin-sectioning ATUM-SEM could
be readily discerned: The unit's cellular layers with endothelial cells, basement membrane and
pericytic processes could be visualized, even at pathological sites. Moreover, the injected colloidal
gold particles and LNDs were accumulated in the vessel lumen, with the identification of LNDs —
compared to SEM images—actually facilitated in the ET by the increased resolution (Fig. 3E;
Supplemental movies 1-3). We observed that the endothelial layer showed local edema and
detachment from the basement membrane, while tight junctions were still intact (Fig. 3E). Gold
particles were detected in platelet vesicles, vacuoles of endothelial cell swellings, as well as in areas
of endothelial detachment from the basement membrane (Fig. 3E). We also recovered and
reconstructed tomograms from nine serial semi-thick sections (Fig. 4). Comparing the SEM
reconstruction with the one from the corresponding serial ETs (Fig. 4A) clearly showed that SEM
imaging visualizes only the section surface resulting in a coarse three-dimensional rendering, while ET
provided information of the entire section thickness (Fig. 4B-D; Supplemental movies 4-6). Taken
together, this correlative ATUM-Tomo approach enabled the targeted visualization of extravasating
nanoparticles and their ultrastructural organelle carriers within reconstructed CNS neuropil as
contextualized by confocal and volume SEM imaging. With our valuable approach we bridged micro-
to nanoscales for the examination of healthy and diseased tissue samples by electron microscopy.

Discussion

We and others have shown previously that non-destructive array tomography-type volume EM
methods like ATUM-SEM are particularly powerful for correlative workflows, because tissue libraries
with re-imaging capabilities are generated (Schifferer *et al.*, 2021). In contrast, destructive block-face
volume EM methods only allow a single shot acquisition and are limited in field of view. The ATUM
imaging process can be separated into screening at low resolution for the sites of interest and
subsequent high-resolution volume acquisition. ATUM-SEM allows the rendering of cellular and
organellar morphologies at the ultrastructural level. However, the visualization of the detailed
subcellular environment of nanoscale structures is limited to the maximal lateral resolution achievable
by SEM and the axial resolution set by physical section thickness (30 nm). To increase the axial
resolution of target structures, we previously combined ATUM with Focused Ion Beam (FIB)-SEM.
ATUM-FIB takes advantage of thick section (2-10 µm) imaging for screening combined with FIB-SEM

acquisition of the target region at isotropic voxels (Kislinger *et al.*, 2020a; Kislinger *et al.*, 2020b). It
enables targeted volume imaging at 3×3×3 nm resolution. The only room temperature volume EM
methods that can provide higher resolution are serial-section TEM or ET (Peddie *et al.*, 2022;
Schneider, Hegermann and Wrede, 2021). While serial-section ET has proven suitable to resolve and
reconstruct volumes of abundant (Soto *et al.*, 1994; Rachel *et al.*; Horstmann *et al.*, 2012) or relatively
small biological targets such as spindles (Kiewisz *et al.*, 2022; Redemann *et al.*, 2017), the
visualization of rare or correlated structures is tedious and time consuming. So far, an automated way
of serial TEM can only be done by serial ultrathin sections collected onto grid tape, a plastic tape with
holes covered with a thin support film at defined distances (Graham *et al.*, 2019). Besides high costs
for the tape, the ultramicrotome has to be automatized to make sure that the sections are positioned
onto the slot and not the grid bar. Moreover, a standard TEM has to be re-designed for a reel-to-reel
sample loading system (Maniates-Selvin *et al.*, 2020; Yin *et al.*, 2020), which is not compatible with ET
and obviates its application in a standard EM facility.

Our novel ATUM-Tomo approach provides a simple and inexpensive way of combining large volume
SEM with targeted TEM at isotropic, high-resolution voxels. As TEM is a projection method, it requires
the removal of the solid substrate, which is needed for SEM imaging. Here we developed a simple
tape coating technique that allows consecutive section inspection by the two EM modalities. Formvar
coating resulted in reversible section collection but heterogeneity in film thickness led to charging
artifacts, background signal and sites of irreversible section attachment. We found that marker pen
coating provided a simple, reproducible and homogenous coating. Using pen-coated tape, we were
able to image hierarchically up to a lateral resolution of 10 nm by SEM, while retaining the recovery
option for transfer onto grids. Detachment of semi-thick sections by acetone did not result in major
section distortion. The detachment of ultrathin sections proved more difficult due to section folding and
fragility during acetone rinsing. Moreover, surface contamination might potentially cover sites of
interest which has more pronounced consequences for the information content of ultrathin than semi-
thick sections. Surprisingly, previous SEM imaging does not affect the ultrastructural quality of the ET
images. Section folds usually occur at sites within the tissue that have not been imaged by SEM
before. This might be caused by reinforced section attachment to the tape caused by prolonged
exposure to the electron beam. This effect is even beneficial as it guides the positioning of tissue
sections onto the TEM grid slot if it exceeds its area. The SEM volume reconstructed from serial semi-
thick sections shows a coarser morphology compared to serial ultrathin sections but still provides the
information needed for the selection of sections to be subjected to ET. There was minor tissue loss
between consecutive tomograms. This was probably caused by a combination of standard material
loss due to sectioning (Hayworth *et al.*, 2015; Kislinger *et al.*, 2020b) and the incomplete tomographic
reconstruction (missing cone) (Barth *et al.*, 1988; Ding *et al.*, 2019) (Figure 2 - figure supplement 3,
Fig. 4B). While we observed tissue loss because of surface contamination and sectioning, we were
able to recover and reconstruct a volume of up to nine consecutive semi-thick sections by ET. Thus,
we demonstrate that *en bloc* stained tissue can be sequentially imaged by volume SEM and
subsequent ET providing high ultrastructural details.

We inspected microthrombi in traumatic brain injury by both correlative ATUM and ATUM-Tomo. In
both cases, CLEM was facilitated by preservation of lectin and LND fluorescence despite
glutaraldehyde fixation, a prerequisite for optimal ultrastructural preservation. While LNDs enable the
detection of microthrombi and sites of BBB leakage by confocal microscopy, their size and contrast
show variability in SEM and TEM micrographs. Here, detection of abluminal gold particles served as a
clear indicator of extravasation by virtue of high electron density and discrete size of the gold particles.
Nanoparticles were found among luminal cells and close to sites of immune cell diapedesis. By virtue
of the novel ATUM-Tomo approach, we captured nanoparticles at sites of endothelial swellings. In line
with our results, earlier studies already described endothelial edema preceding FITC-albumin early
after middle cerebral artery occlusion (Krueger *et al.*, 2015; Krueger *et al.*, 2019), another model
system for studying BBB leakiness. At later stages, these authors observed loss of endothelial cell
integrity with discontinuous plasma membranes and detachment from the basement membrane.
However, the DAB staining of FITC-albumin applied in these studies masks the endothelial
ultrastructure and thus provides compromised EM information (Krueger *et al.*, 2019). In contrast, our
CLEM approach provides high ultrastructural quality by chemical fixation with glutaraldehyde.
Endothelial swellings were identified at regions without lectin staining, only showing LND fluorescence.
Lectin serves as a marker for intact glycocalyx and its loss indicates early pathological events leading
to activation of transcellular transport. It was shown that the glycocalyx is critical for BBB integrity by
suppressing caveolin1-dependent endothelial transcytosis following ischemic stroke (Zhu *et al.*, 2022).
This is in line with the detection of intact tight junctions and gold particles in vesicles at sites of BBB
leakiness in the current and previous studies (Krueger *et al.*, 2015). In contrast, the paracellular
pathway seems not to be the first loophole and gains importance starting 48 h after the injury
(Knowland *et al.*, 2014).

In summary, we developed a simple and low-cost method for multimodal EM imaging that combines
the strengths of both volume SEM and ET. Our novel ATUM-Tomo approach enables the consecutive
inspection of selected areas of interest by correlative serial SEM and TEM, optionally in combination
with CLEM. ATUM-Tomo and also correlative ATUM-Tomo can bridge several scales, which is
particularly important for ultrastructural analysis of focal pathologies or rare cellular and subcellular
events. Both imaging modalities are non-destructive, thus allowing re-imaging and hierarchical
imaging at the SEM and TEM levels, which is especially important for precious samples including
human biopsies and complex CLEM experiments. While we demonstrate a neuropathology-related
application, further biological targets that require high-resolution isotropic voxels and spatial orientation
within a larger ultrastructural context can potentially be studied by ATUM-Tomo. This includes the
detection of gap junctions for connectomics or for the study of long-range projections (Holler *et al.*,
2021) and the subcellular location of virus particles (Wu *et al.*, 2022; Roingeard, 2008; Pelchen-
Matthews and Marsh, 2007). Thus, ATUM-Tomo opens up new avenues for multimodal volume EM
imaging of diverse biological research areas.

Material and Methods

Controlled cortical impact model of traumatic brain injury

All animal experiments were approved by the Ethical Review Board of the Government of Upper Bavaria. The data were collected in accordance with the ARRIVE guidelines (Kilkenny *et al.*, 2010) recommendations (Guillen, 2012). Animal husbandry, health screens, and hygiene management checks were performed in accordance with Federation of European Laboratory Animal Science Associations (FELASA) guidelines and recommendations. 12 C57Bl/6N male mice, 23–25 g, underwent controlled cortical impact as described above and our in previous studies (Wehn *et al.*, 2021).

This model induces a highly reproducible lesion that shares many characteristics of human traumatic brain injury. In short, after induction of anesthesia with buprenorphine (100 mg×kg−1 Bw) and isoflurane (4%, 30s), animals were sedated with 1.5–2.5% isoflurane in oxygen/air under continuous monitoring of body temperature and heart rate. After right parietal craniotomy, the impact is directly applied to the intact dura with a pressure-controlled custom made controlled cortical impact device (L. Kopacz, University of Mainz, Germany; 6 m s−1, 0.5 mm penetration depth, 150 ms contact time). After re-fixation of the skull plate and surgical closure, animals are ventilated with 100% oxygen until they re-gained consciousness, then were kept in an incubator at 34°C and 60% humidity in order to prevent hypothermia. Perfusion, confocal imaging and correlative ATUM-SEM were performed as previously described (Khalin *et al.*, 2022).

Lipid nanodroplets

Dye-loaded lipid nanoemulsions droplets (LNDs) were produced by spontaneous nanoemulsification by adapting previously described protocols (Bouchaala *et al.*, 2016). Briefly, the 1% solution of R18/TPB (prepared as described earlier (Reisch *et al.*, 2014)) in LabrafacWL® was mixed with Kolliphor ELP®, and the mixture was homogenized under magnetic stirring at 37°C for 10 min. The amounts of LabrafacWL® and Kolliphor ELP® were 40 and 60 mg, respectively. Finally, LNDs were obtained by addition of 230 mg Milli-Q water. Hydrodynamic diameter was measured by dynamic light scattering (DLS) using a Zetasizer Nano ZSP (Malvern Instruments S.A., Worcestershire, UK), based on volume statistics. The size of LNDs was 35.66±1.66 nm, PDI index 0.113±0.002.

Injection

For injection of lectin and NPs we used an established femoral arterial catheterization protocol (Khalin *et al.*, 2020). 1 mg/ml DyLight 649 Labeled Lycopersicon Esculentum (Tomato) Lectin (Vector Laboratories, Burlingame, CA, US) for labeling of blood vessels was injected in a quantity of 0.1 ml per

368 mouse. Commercially available 30 nm colloidal gold nanoparticles with density 1.00 g/cm^3, methyl
369 terminated, PEG 2000 coated (Sigma-Aldrich, St. Louis, Missouri, USA) were injected through femoral
370 catheter into the mouse in the same dose as LNDs (4 µL/g animal) (Khalin *et al.*, 2022).

Tissue collection

Animals were transcardially perfused with 4% PFA and 2.5% glutaraldehyde in PBS pH 7.4 in deep
anesthesia. The brain was collected and the lesion area of fixed brains was subsequently vibratome-
sectioned in rostral-caudal direction into 80-100 µm sections using a vibratome as previously
described (Ghosh *et al.*, 2015). In order to ensure high ultrastructural preservation, sections were
postfixed for 16-24h in the same fixative mentioned above, washed in buffer and carefully transferred
using brushes. Instead of standard mounting which could harm the tissue, sections were mounted
onto a glass slide using mounting medium (Aqueous Mounting Medium (ab128982), Abcam,
Cambridge, UK) diluted with PBS at a ratio of 1:5.   A cover slip (0.17 µm, Menzel Gläser,
Braunschweig, Germany) was then placed onto it for imaging. After performing confocal microscopy,
the mounted sections were submerged in PBS and the cover slip carefully removed. Sections could
then be transferred back into 24-well plates containing fresh buffer for further processing.

Confocal imaging for correlation

Imaging was performed using confocal microscopy (ZEISS LSM 900, Carl Zeiss Microscopy GmbH,
Jena Germany). Confocal imaging revealed that lectin and LND fluorescence was possible with
minimal autofluorescence despite perfusion fixation with glutaraldehyde. A 10x objective (EC Plan-
Neofluar 10x/0.30 M27) was used with an image matrix of 1024×1024 pixel, a pixel scaling of 0.2×0.2
µm and a color depth of 8 bit. The scanned region of approximately 2×2 mm covered the whole
cortical thickness of the lesion and peri-lesional area in order to generate overview maps for later
correlation. Whole brain images were collected in z-stacks as tile scans with a slice-distance of 5 µm
and a total range of 50 µm. Specific regions of interest including the microthrombi and sites of
extravasation were collected in single plane using 25x magnification (objective: EC Plan-Neofluar
25x/0.8 Imm Korr DIC M27) with an image matrix of 1024×1024 pixel, a pixel scaling of 0.2×0.2 µm
and a depth of 8 bit. After imaging for maximal 1 hour, the glass slide was immersed into a petri dish
filled with 1x PBS. As soon as the tissue section was floating from the glass sandwich, it was
transferred to fresh buffer for storage.

Embedding of vibratome sections for electron microscopy

The sections were stored in PBS at 4°C for up to one week until the start of the post-embedding. For
correlation, the confocal imaging plane was ideally parallel to the sectioning plane during
ultramicrotomy. The searchable volume in the axial dimension is predetermined by and restricted to
the vibratome section thickness. Usually, heavy metal impregnation leads to bending of the section. In
order to avoid this sample preparation artifact and minimize the lateral searchable area, we dissected
a region of approximately 1.5×2 mm capturing the lesion and the perilesional area under the binocular
using fine scalpels. The fluorescence overview tile scan was used as a template to relocate the
previously imaged region. The overall section shape, including big blood vessels and bleedings that

represent non-fluorescent regions in the confocal template were used for this coarse lateral re-
location.

We applied a standard reduced osmium thiocarbohydrazide osmium (rOTO) *en bloc* staining protocol
(Kislinger *et al.*, 2020a) (Kislinger *et al.*, 2023) including post-fixation in 2% osmium tetroxide (EMS),
1.5% potassium ferrocyanide (Sigma) in 0.1 M sodium cacodylate (Science Services) buffer (pH 7.4).
Staining was enhanced by reaction with 1% thiocarbohydrazide (Sigma) for 45 min at 40°C. The tissue
was washed in water and incubated in 2% aqueous osmium tetroxide, washed and further contrasted
by overnight incubation in 1% aqueous uranyl acetate at 4°C and 2h at 50°C. Samples were
dehydrated in a graded series of ethanol and infiltrated with LX112 (LADD) resin. For the resin test
(Figure 2 - figure supplement 5), we followed the same protocol while infiltrating samples in Durcupan
(11.4 g component A, 10 g component B, 0.3 g component C and 0.1 g component D (Sigma-Aldrich,
USA)) and Epon (21.4 g glycidether 100, 14.4 g DDSA, 11.3 g MNA and 0.84 g DMP-30 (Serva,
Heidelberg, Germany)), respectively.

Plastic tape plasma discharging

Plasma discharging increases hydrophilicity and thereby prevents wrinkle formation during serial
section collection for ATUM and ATUM-Tomo approaches. In addition, ATUM-Tomo requires several
rounds of plasma treatment after marker pen coating as the marker introduces hydrophobicity.

For this end, we built a plasma treatment instrument for tape reels based on a standard instrument
(easiGlow, PELCO) with a custom-made vacuum chamber according to the plans proposed by Jeff
Lichtman (Schalek et al., 2012) and Mark Terasaki (Baena et al., 2019), with slight modifications
regarding the motor placement and the type of motor (Figure 2 - figure supplement 6, see also
(Kislinger *et al.*, 2023)). We used a dynamic O-ring shaft seal (Kurt J. Lesker Company) and a gear
motor with a 150:1 reduction and a suitable controller (Pololu Corporation, Las Vegas, USA) and fixed
it directly to the enclosure. In this setting, the motor achieves slow tape speeds (at 12 rpm) while
maintaining necessary pulling forces. A more detailed description of the construction and the use as
well as a parts list are deposited here: https://github.com/georgkislinger/GKislinger_elife_ATUM-Tomo.

We plasma discharged (negative polarity, 0.35 mBar, 15 mA) as many times as needed to achieve a
golden instead of a black appearance (Fig. 2F; Figure 2 - figure supplement 6). In addition,
hydrophobicity was tested by estimating the water contact angle by placing a drop of water onto the
tape. It is ready to use if it spreads well with a contact angle below 45° (see also (Kislinger *et al.*,
2023)). This tape can be used for up to one month with one round of plasma discharging before each
collection process. However, additional rounds of plasma discharging may increase charging artifacts
and impede section recovery as the conductive carbon and pen coatings become thinner.

Plastic tape marker coating

Initially, we coated ~10 cm long carbon-coated Kapton (ccKapton) (kindly provided by Jeff Lichtman
and Richard Schalek) or carbon nanotube (Science Services) tape strips with 1% (w/v) Formvar in
chloroform (Sigma) using a standard film casting device for grids (Science Services). In order to
generate a homogenous film on longer ccKapton tape pieces, we designed a spray-coating device.

This custom-built coating machine was equipped with a holder for the spraying unit (Lubrimat L60,
Steidle) and two tape reels driven by a motor (Figure 2 - figure supplement 1). However, this method
proved to be unreliable, as the chloroform caused deterioration of the spraying unit over time which, in
turn, the poststain (i.e., lead) to an uneven and contaminated coating. Heterogeneous coating resulted
in charging during or detachment problems after SEM imaging.

In order to enable section removal, we marked ccKapton tape on its coated side with a permanent
marker (edding 3000, black, edding International GmbH, Germany). We used the before-mentioned
coating machine and replaced the spraying unit with a permanent parker. Thereby, the marker was
constantly pressed onto the tape as it was pulled under the marker tip to ensure uniform tape coating
(Fig. 2E).

Section collection onto coated tape

For the volume analysis, the block was trimmed at a depth of 100 µm and a 45° angle on four sides
using a trimming machine (EM TRIM, Leica). Serial sections were taken with a 35° ultra-maxi diamond
knife (Diatome) at a nominal cutting thickness of 80-100 nm for ATUM-SEM and the TEM tests and
400 nm for ATUM-Tomo at the ATUMtome (Powertome, RMC). For ATUM-SEM, sections were
collected onto freshly plasma-treated carbon nanotube tape (Science Services); for ATUM-Tomo,
sections were collected on either cpcKapton (carbon- and permanent marker coated) or cfcKapton
(carbon- and Formvar-coated) tape. Covering the whole sample thickness of 80 µm and considering
its unevenness after contrasting, either 1000 sections at 100 nm thickness (ATUM-SEM) or 400
sections at 400 nm thickness (ATUM-Tomo) were taken. Plastic tape stripes were assembled onto
adhesive carbon tape (Science Services) attached to 4-inch silicon wafers (Siegert Wafer) and
grounded by adhesive carbon tape strips (Science Services). For each sample, 4-6 (ATUM-SEM) or 1-
2 (ATUM-Tomo) wafers each bearing 200-300 sections were generated.

Serial-section acquisition by SEM

ATUM-SEM was performed on a Crossbeam Gemini 340 SEM (Zeiss) with a four-quadrant
backscatter detector at 8 kV. Hierarchical imaging captures large areas at low resolution that can, in
turn, be screened for structures of interest that are subsequently reimaged at higher resolution and so
forth. In ATLAS5 Array Tomography (Fibics), wafer overview images were generated (1000 nm/pixel).
We acquired whole section low resolution images (0.1×0.1×1 µm) of every 10[th] section (thus
simulating the thickness of one confocal plane covering 1 µm thickness) to generate overviews that
can be correlated with the confocal imaging data. We used blood vessel patterns, bleedings and the
different tissue layers as landmarks. Target regions were identified and acquired at medium resolution
(20×20×200 nm). Within this volume, clots were selected for high-resolution imaging (4-20×4-20×200
nm).

For ATUM-Tomo, electron micrographs were acquired in a similar hierarchical scheme on an Apreo
S2 SEM (Thermo Fisher Scientific) with the T1 detector. Serial-section acquisition was performed
using the Maps2 (Thermo Fisher Scientific) software. We acquired images of whole sections at low
resolution (100×100×400 nm) to facilitate re-location. Anatomical landmarks and vasculature patterns

of single SEM overview images and a summed (in Fiji) stack of sections were compared to the
confocal data sets in Fiji. Within this volume, we selected a region with clots and transition zones and
extravasation for high-resolution imaging at 20×20×400 nm. We acquired subsets of these regions on
each or every second section at 10×10×400 nm or 10×10×800 nm resolution, respectively.

Correlation and volume image analysis

There were two levels of correlation: first, we dissected the tissue according to the field of view of
confocal imaging and second, we identified occlusions of interest in the low-resolution SEM data for
high-resolution imaging. At each step, anatomical landmarks (bleedings with accumulation of
erythrocytes, the section outline etc.) and vasculature patterns of single and summed sections were
compared to the confocal data sets using either Fiji (Schindelin *et al.*, 2012) or the ec-CLEM ICY
plugin (Paul-Gilloteaux *et al.*, 2017). We used both a single SEM image and the sum (in Fiji) of a stack
of SEM images for relocation of the occlusions of interest. Serial-section data were aligned by a
sequence of automatic and manual processing steps in Fiji TrakEM2 (Kislinger *et al.*, 2023). The
VAST software was used for manual segmentation of blood vessels (Berger, Seung and Lichtman,
2018) that were rendered in Blender (Brito, 2018) for better visualization.

Section retrieval from coated tape onto TEM grids

Based on SEM-imaging results and the corresponding three-dimensional SEM data, we selected
sections of interest for subsequent TEM imaging. We cut the carbon adhesive tape around the
selected section on the wafer using a scalpel and peeled off the section-containing piece of cpcKapton
from the adhesive tape (Fig. 2B, C). Next, we submerged the piece of tape containing one single
section of interest in a glass petri dish containing acetone. We then created a steady acetone flow with
a plastic transfer pipette until the coating between tape and section was dissolved, resulting in a
section floating in the acetone bath. In our experience, the required rinsing time increases with
previous beam exposure of the section in the SEM. Once floating in the acetone bath, we collected the
single section in a plastic transfer pipette and released it into a water-filled boat (as used for sectioning
with a diamond knife, kindly provided by Diatome). Through the surface tension of the water, the
section stretched and floated on the water surface. However, through the detachment from the tape in
the acetone bath, it could happen that the sections were physically flipped on the water bath relative to
the sectioning direction.

TEM of ultrathin sections

512 Ultrathin sections were removed from either cfcKapton or cpcKapton tape using chloroform or
513 acetone. Retrieved sections were collected onto Formvar-coated copper grids (Science Services).
514 Sections were imaged using a JEM-1400+ (JEOL) transmission electron microscope (TEM) equipped
515 with a XF416 (TVIPS) camera operated by the EM-Menu software (TVIPS).

Electron Tomography of semi-thin sections

For electron tomography (ET), single serial semi-thick (400 nm) sections were released as described
above and collected on slot grids covered with a Formvar layer of golden interference color. Colloidal

gold particles (diameter 20 nm, BBI) serving as fiducials for tomographic reconstruction were then
attached to the sections by submerging the grids into a droplet of the pure colloidal gold solution for 1
min. Samples were blotted to remove residual solution and air-dried. Semi-thick sections were imaged
using a Tecnai F30 transmission electron microscope (Thermo Fisher Scientific) operated at 300 kV
and equipped with a 4k×4k CMOS camera (OneView, Gatan) as previously described (Kiewisz *et al.*,
2022). Using a dual-axis specimen holder (Type 2040, Fishione, USA), tilt series were recorded from -
60° to +60° with 1° increments at a magnification of 4700x and a pixel size of 2.572 nm applying the
SerialEM software package (Mastronarde, 2003; Mastronarde, 2005). For dual-tilt electron
tomography, the grid was rotated for 90° in the XY-plane and the second tilt series was acquired using
identical microscope settings (Mastronarde, 1997). Both tomographic image stacks of the same
positions were reconstructed, combined and flattened using IMOD (Kremer, Mastronarde and
McIntosh, 1996; Mastronarde and Held, 2017). Serial semi-thick sections were stitched and combined
by using the ZIB Amira (Zuse Institute Berlin, Germany) software package (Stalling, Westerhoff and
Hege, 2005; Lindow *et al.*, 2021). After converting the image data into TIFF format, the image
registration was manually refined in Fiji TrakEM2., followed by a final elastic alignment. Segmentation
of cellular features was performed manually with the program VAST (Berger, Seung and Lichtman,
2018) and subsequently rendered in Blender (Brito, 2018). All movies were generated using Amira.

Acknowledgements

This work was supported by the DFG under Germany's Excellence Strategy within the framework of
the Munich Cluster for Systems Neurology (SyNergy; EXC 2145 – ID 390857198), TRR 274/1 2020
(projects Z01 and B03 – ID 408885537) and FOR Immunostroke (Mi 694/9-1 A03 – ID 428663564).
Research in the Müller-Reichert lab was funded by the DFG (grant MU 1423/8-2 and 8-3 to TMR). All
animal experiments were supported by DFG grant 457586042. We thank Felix Beyer and Gero Finck
(TUM, SyNergy NeuroCore), Dr. Tobias Fürstenhaupt (EM facility at MPI-CBG, Dresden) for excellent
technical assistance and Théo Cerciat (TUM) for segmentation.

References

Andreone, B. J., Chow, B. W., Tata, A., Lacoste, B., Ben-Zvi, A., Bullock, K., Deik, A. A., Ginty, D. D.,
Clish, C. B. and Gu, C. (2017) 'Blood-Brain Barrier Permeability Is Regulated by Lipid Transport-
Dependent Suppression of Caveolae-Mediated Transcytosis', *Neuron,* 94(3), pp. 581-594.e5.

Baena, V., Schalek, R. L., Lichtman, J. W. and Terasaki, M. (2019) 'Serial-section electron microscopy
using automated tape-collecting ultramicrotome (ATUM)', *Methods Cell Biol,* 152, pp. 41-67.

Barth, M., Bryan, R. K., Hegerl, R. and Baumeister, W. (1988) 'Estimation of missing cone data in
three-dimensional electron microscopy', *Scanning Microsc Suppl,* 2, pp. 277-84.

Baumeister, W., Grimm, R. and Walz, J. (1999) 'Electron tomography of molecules and cells', *Trends*
in Cell Biology, 9(2), pp. 81-85.

Berger, D. R., Seung, H. S. and Lichtman, J. W. (2018) 'VAST (Volume Annotation and Segmentation
Tool): Efficient Manual and Semi-Automatic Labeling of Large 3D Image Stacks', *Front Neural Circuits,*
12, pp. 88.

Bouchaala, R., Mercier, L., Andreiuk, B., Mély, Y., Vandamme, T., Anton, N., Goetz, J. G. and
Klymchenko, A. S. (2016) 'Integrity of lipid nanocarriers in bloodstream and tumor quantified by near-
infrared ratiometric FRET imaging in living mice', *J Control Release,* 236, pp. 57-67.

Brito, A. (2018) *Blender Quick Start Guide: 3D Modeling, Animation, and Render with Eevee in*
Blender 2.8. Packt Publishing.

Ding, G., Liu, Y., Zhang, R. and Xin, H. (2019) 'A joint deep learning model to recover information and reduce artifacts in missing-wedge sinograms for electron tomography and beyond', *Scientific Reports,* 9, pp. 1-13.

Feng, D., Nagy, J. A., Dvorak, H. F. and Dvorak, A. M. (2002) 'Ultrastructural studies define soluble macromolecular, particulate, and cellular transendothelial cell pathways in venules, lymphatic vessels, and tumor-associated microvessels in man and animals', *Microscopy Research and Technique,* 57(5), pp. 289-326.

Frank, J. 'Electron tomography : methods for three-dimensional visualization of structures in the cell'.

Ghosh, M., Balbi, M., Hellal, F., Dichgans, M., Lindauer, U. and Plesnila, N. (2015) 'Pericytes are involved in the pathogenesis of cerebral autosomal dominant arteriopathy with subcortical infarcts and leukoencephalopathy', *Ann Neurol,* 78(6), pp. 887-900.

Graham, B. J., Hildebrand, D. G. C., Kuan, A. T., Maniates-Selvin, J. T., Thomas, L. A., Shanny, B. L. and Lee, W.-C. A. (2019) 'High-throughput transmission electron microscopy with automated serial sectioning', *bioRxiv*, pp. 657346.

Guillen, J. (2012) 'FELASA guidelines and recommendations', *J Am Assoc Lab Anim Sci,* 51(3), pp. 311-21.

Hayworth, K. J., Xu, C. S., Lu, Z., Knott, G. W., Fetter, R. D., Tapia, J. C., Lichtman, J. W. and Hess, H. F. (2015) 'Ultrastructurally smooth thick partitioning and volume stitching for large-scale connectomics', *Nature Methods,* 12(4), pp. 319-322.

He, W. and He, Y. (2014) 'Electron tomography for organelles, cells, and tissues', *Methods Mol Biol,* 1117, pp. 445-83.

Hirano, A., Kawanami, T. and Llena, J. F. (1994) 'Electron microscopy of the blood-brain barrier in disease', *Microscopy Research and Technique,* 27(6), pp. 543-556.

Holler, S., Köstinger, G., Martin, K. A. C., Schuhknecht, G. F. P. and Stratford, K. J. (2021) 'Structure and function of a neocortical synapse', *Nature,* 591(7848), pp. 111-116.

Horstmann, H., Körber, C., Sätzler, K., Aydin, D. and Kuner, T. (2012) 'Serial Section Scanning Electron Microscopy (S3EM) on Silicon Wafers for Ultra-Structural Volume Imaging of Cells and Tissues', *PLOS ONE,* 7(4), pp. e35172.

Hua, Y., Laserstein, P. and Helmstaedter, M. (2015) 'Large-volume en-bloc staining for electron microscopy-based connectomics', *Nature Communications,* 6(1), pp. 7923.

Karreman, M. A., Ruthensteiner, B., Mercier, L., Schieber, N. L., Solecki, G., Winkler, F., Goetz, J. G. and Schwab, Y. (2017) 'Chapter 13 - Find your way with X-Ray: Using microCT to correlate in vivo imaging with 3D electron microscopy', in Müller-Reichert, T. and Verkade, P. (eds.) *Methods in Cell Biology*: Academic Press, pp. 277-301.

Kasthuri, N., Hayworth, Kenneth J., Berger, Daniel R., Schalek, Richard L., Conchello, José A., Knowles-Barley, S., Lee, D., Vázquez-Reina, A., Kaynig, V., Jones, Thouis R., Roberts, M., Morgan, Josh L., Tapia, Juan C., Seung, H. S., Roncal, William G., Vogelstein, Joshua T., Burns, R., Sussman, Daniel L., Priebe, Carey E., Pfister, H. and Lichtman, Jeff W. (2015) 'Saturated Reconstruction of a Volume of Neocortex', *Cell,* 162(3), pp. 648-661.

Khalin, I., Adarsh, N., Schifferer, M., Wehn, A., Groschup, B., Misgeld, T., Klymchenko, A. and Plesnila, N. (2022) 'Size-Selective Transfer of Lipid Nanoparticle-Based Drug Carriers Across the Blood Brain Barrier Via Vascular Occlusions Following Traumatic Brain Injury', *Small,* n/a(n/a), pp. 2200302.

Khalin, I., Heimburger, D., Melnychuk, N., Collot, M., Groschup, B., Hellal, F., Reisch, A., Plesnila, N. and Klymchenko, A. S. (2020) 'Ultrabright Fluorescent Polymeric Nanoparticles with a Stealth Pluronic Shell for Live Tracking in the Mouse Brain', *ACS Nano,* 14(8), pp. 9755-9770.

Kiewisz, R., Fabig, G., Conway, W., Baum, D., Needleman, D. and Müller-Reichert, T. (2022) 'Three-dimensional structure of kinetochore-fibers in human mitotic spindles', *eLife,* 11, pp. e75459.

Kilkenny, C., Browne, W. J., Cuthill, I. C., Emerson, M. and Altman, D. G. (2010) 'Improving Bioscience Research Reporting: The ARRIVE Guidelines for Reporting Animal Research', *PLOS Biology,* 8(6), pp. e1000412.

Kislinger, G., Gnägi, H., Kerschensteiner, M., Simons, M., Misgeld, T. and Schifferer, M. (2020a) 'ATUM-FIB microscopy for targeting and multiscale imaging of rare events in mouse cortex', *STAR Protocols,* 1(3), pp. 100232.

Kislinger, G., Gnägi, H., Kerschensteiner, M., Simons, M., Misgeld, T. and Schifferer, M. (2020b) 'Multiscale ATUM-FIB Microscopy Enables Targeted Ultrastructural Analysis at Isotropic Resolution', *iScience,* 23(7), pp. 101290.

Kislinger, G., Niemann, C., Rodriguez, L., Jiang, H., Fard, M. K., Snaidero, N., Schumacher, A.-M., Kerschensteiner, M., Misgeld, T. and Schifferer, M. (2023) 'Neurons on tape: Automated Tape Collecting Ultramicrotomy-mediated volume EM for targeting neuropathology', *Methods in Cell Biology*: Academic Press.

Knowland, D., Arac, A., Sekiguchi, Kohei J., Hsu, M., Lutz, Sarah E., Perrino, J., Steinberg, Gary K.,
Barres, Ben A., Nimmerjahn, A. and Agalliu, D. (2014) 'Stepwise Recruitment of Transcellular and
Paracellular Pathways Underlies Blood-Brain Barrier Breakdown in Stroke', *Neuron,* 82(3), pp. 603-
617.
Kremer, J. R., Mastronarde, D. N. and McIntosh, J. R. (1996) 'Computer visualization of three-
dimensional image data using IMOD', *J Struct Biol,* 116(1), pp. 71-6.
Krueger, M., Bechmann, I., Immig, K., Reichenbach, A., Härtig, W. and Michalski, D. (2015) 'Blood-
brain barrier breakdown involves four distinct stages of vascular damage in various models of
experimental focal cerebral ischemia', *J Cereb Blood Flow Metab,* 35(2), pp. 292-303.
Krueger, M., Mages, B., Hobusch, C. and Michalski, D. (2019) 'Endothelial edema precedes blood-
brain barrier breakdown in early time points after experimental focal cerebral ischemia', *Acta
Neuropathologica Communications,* 7(1), pp. 17.
Kubota, Y., Sohn, J., Hatada, S., Schurr, M., Straehle, J., Gour, A., Neujahr, R., Miki, T., Mikula, S.
and Kawaguchi, Y. (2018) 'A carbon nanotube tape for serial-section electron microscopy of brain
ultrastructure', *Nature Communications,* 9(1), pp. 437.
Kiewisz R., Fabig G., Conway W., Baum D., Needleman D. and Müller-Reichert T. (2022) 'Three-
dimensional structure of kinetochore-fibers in human mitotic spindles.' *Elife*, 11:e75459.
Lindow, N., Brünig, F. N., Dercksen, V. J., Fabig, G., Kiewisz, R., Redemann, S., Müller-Reichert, T.,
Prohaska, S. and Baum, D. (2021) 'Semi-automatic stitching of filamentous structures in image stacks
from serial-section electron tomography', *J Microsc,* 284(1), pp. 25-44.
Luther, P. K., Lawrence, M. C. and Crowther, R. A. (1988) 'A method for monitoring the collapse of
plastic sections as a function of electron dose', *Ultramicroscopy,* 24(1), pp. 7-18.
Maniates-Selvin, J. T., Hildebrand, D. G. C., Graham, B. J., Kuan, A. T., Thomas, L. A., Nguyen, T.,
Buhmann, J., Azevedo, A. W., Shanny, B. L., Funke, J., Tuthill, J. C. and Lee, W.-C. A. (2020)
'Reconstruction of motor control circuits in adult Drosophila using automated transmission
electron microscopy', *bioRxiv,* pp. 2020.01.10.902478.
Mastronarde, D. N. (1997) 'Dual-axis tomography: an approach with alignment methods that preserve
resolution', *J Struct Biol,* 120(3), pp. 343-52.
Mastronarde, D. N. (2003) 'SerialEM: A Program for Automated Tilt Series Acquisition on Tecnai
Microscopes Using Prediction of Specimen Position', *Microscopy and Microanalysis,* 9(S02), pp. 1182-
1183.
Mastronarde, D. N. (2005) 'Automated electron microscope tomography using robust prediction of
specimen movements', *J Struct Biol,* 152(1), pp. 36-51.
Mastronarde, D. N. and Held, S. R. (2017) 'Automated tilt series alignment and tomographic
reconstruction in IMOD', *J Struct Biol,* 197(2), pp. 102-113.
McEwen, B. F. and Marko, M. (1998) 'Chapter 5 Three-Dimensional Transmission Electron
Microscopy and Its Application to Mitosis Research', in Rieder, C.L. (ed.) *Methods in Cell Biology*:
Academic Press, pp. 81-111.
Michel, C. C. (2012) 'Electron Tomography of Vesicles', *Microcirculation,* 19(6), pp. 473-476.
Micheva, K. D. and Smith, S. J. (2007) 'Array Tomography: A New Tool for Imaging the Molecular
Architecture and Ultrastructure of Neural Circuits', *Neuron,* 55(1), pp. 25-36.
Nahirney, P. C., Reeson, P. and Brown, C. E. (2016) 'Ultrastructural analysis of blood-brain barrier
breakdown in the peri-infarct zone in young adult and aged mice', *Journal of cerebral blood flow and
metabolism : official journal of the International Society of Cerebral Blood Flow and Metabolism,* 36(2),
pp. 413-425.
O'Brown, N. M., Megason, S. G. and Gu, C. (2019) 'Suppression of transcytosis regulates zebrafish
blood-brain barrier function', *eLife,* 8, pp. e47326.
O'Toole, E., Morphew, M. and McIntosh, J. R. (2020) 'Electron tomography reveals aspects of spindle
structure important for mechanical stability at metaphase', *Mol Biol Cell,* 31(3), pp. 184-195.
Paez-Segala, M. G., Sun, M. G., Shtengel, G., Viswanathan, S., Baird, M. A., Macklin, J. J., Patel, R.,
Allen, J. R., Howe, E. S., Piszczek, G., Hess, H. F., Davidson, M. W., Wang, Y. and Looger, L. L.
(2015) 'Fixation-resistant photoactivatable fluorescent proteins for CLEM', *Nature methods,* 12(3), pp.
215-218.
Paul-Gilloteaux, P., Heiligenstein, X., Belle, M., Domart, M.-C., Larijani, B., Collinson, L., Raposo, G.
and Salamero, J. (2017) 'eC-CLEM: flexible multidimensional registration software for correlative
microscopies', *Nature Methods,* 14(2), pp. 102-103.
Peddie, C. J., Genoud, C., Kreshuk, A., Meechan, K., Micheva, K. D., Narayan, K., Pape, C., Parton,
R. G., Schieber, N. L., Schwab, Y., Titze, B., Verkade, P., Weigel, A. and Collinson, L. M. (2022)
'Volume electron microscopy', *Nature Reviews Methods Primers,* 2(1), pp. 51.
Pelchen-Matthews, A. and Marsh, M. (2007) 'Electron microscopy analysis of viral morphogenesis',
Methods Cell Biol, 79, pp. 515-42.

Rachel, R., Heinz, V., Flechsler, J., Witzgall, R. and Heimerl, T. 'Serial section (S)TEM tomography of
prokaryotic cells', *European Microscopy Congress 2016: Proceedings*, pp. 101-102.

Redemann, S., Baumgart, J., Lindow, N., Shelley, M., Nazockdast, E., Kratz, A., Prohaska, S.,
Brugués, J., Fürthauer, S. and Müller-Reichert, T. (2017) 'C. elegans chromosomes connect to
centrosomes by anchoring into the spindle network', *Nat Commun*, 8, pp. 15288.

Reisch, A., Didier, P., Richert, L., Oncul, S., Arntz, Y., Mély, Y. and Klymchenko, A. S. (2014)
'Collective fluorescence switching of counterion-assembled dyes in polymer nanoparticles', *Nature*
Communications, 5(1), pp. 4089.

Roingeard, P. (2008) 'Viral detection by electron microscopy: past, present and future', *Biol Cell,*
100(8), pp. 491-501.

Scher, N. and Avinoam, O. (2021) '50 Shades of CLEM: How to choose the right approach for you',
Methods Cell Biol, 162, pp. 1-11.

Schifferer, M., Snaidero, N., Djannatian, M., Kerschensteiner, M. and Misgeld, T. (2021) 'Niwaki
Instead of Random Forests: Targeted Serial Sectioning Scanning Electron Microscopy With
Reimaging Capabilities for Exploring Central Nervous System Cell Biology and Pathology', *Front*
Neuroanat, 15, pp. 732506.

Schindelin, J., Arganda-Carreras, I., Frise, E., Kaynig, V., Longair, M., Pietzsch, T., Preibisch, S.,
Rueden, C., Saalfeld, S., Schmid, B., Tinevez, J.-Y., White, D. J., Hartenstein, V., Eliceiri, K.,
Tomancak, P. and Cardona, A. (2012) 'Fiji: an open-source platform for biological-image analysis',
Nature Methods, 9(7), pp. 676-682.

Schneider, J., Hegermann, J. and Wrede, C. (2021) 'Volume electron microscopy: analyzing the lung',
Histochemistry and cell biology, 155.

Shi, Y., Thrippleton, M. J., Blair, G. W., Dickie, D. A., Marshall, I., Hamilton, I., Doubal, F. N., Chappell,
F. and Wardlaw, J. M. (2020) 'Small vessel disease is associated with altered cerebrovascular
pulsatility but not resting cerebral blood flow', *J Cereb Blood Flow Metab,* 40(1), pp. 85-99.

Snaidero, N., Schifferer, M., Mezydlo, A., Zalc, B., Kerschensteiner, M. and Misgeld, T. (2020) 'Myelin
replacement triggered by single-cell demyelination in mouse cortex', *Nature Communications,* 11(1),
pp. 4901.

Soto, G. E., Young, S. J., Martone, M. E., Deerinck, T. J., Lamont, S., Carragher, B. O., Hama, K. and
Ellisman, M. H. (1994) 'Serial section electron tomography: a method for three-dimensional
reconstruction of large structures', *Neuroimage,* 1(3), pp. 230-43.

Stalling, D., Westerhoff, M. and Hege, H.-C. 2005. Amira: A highly interactive system for visual data
analysis. *In:* Hansen, C. and Johnson, C. (eds.) *The Visualization Handbook.* Elsevier.

Wacker, I. and Schroeder, R. R. (2013) 'Array tomography', *Journal of Microscopy,* 252(2), pp. 93-99.

Wagner, R., Modla, S., Hossler, F. and Czymmek, K. (2012) 'Three-Dimensional Analysis and
Computer Modeling of the Capillary Endothelial Vesicular System with Electron Tomography',
Microcirculation, 19(6), pp. 477-484.

Wehn, A. C., Khalin, I., Duering, M., Hellal, F., Culmsee, C., Vandenabeele, P., Plesnila, N. and
Terpolilli, N. A. (2021) 'RIPK1 or RIPK3 deletion prevents progressive neuronal cell death and
improves memory function after traumatic brain injury', *Acta Neuropathologica Communications,* 9(1),
pp. 138.

West, M., Zurek, N., Hoenger, A. and Voeltz, G. K. (2011) 'A 3D analysis of yeast ER structure reveals
how ER domains are organized by membrane curvature', *J Cell Biol,* 193(2), pp. 333-46.

Wu, H., Fujioka, Y., Sakaguchi, S., Suzuki, Y. and Nakano, T. (2022) 'Three-dimensional
reconstruction by electron tomography for the application to ultrastructural analysis of SARS-CoV-2
particles', *Medical Molecular Morphology,* 55(1), pp. 60-67.

Yin, W., Brittain, D., Borseth, J., Scott, M. E., Williams, D., Perkins, J., Own, C. S., Murfitt, M., Torres,
R. M., Kapner, D., Mahalingam, G., Bleckert, A., Castelli, D., Reid, D., Lee, W. A., Graham, B. J.,
Takeno, M., Bumbarger, D. J., Farrell, C., Reid, R. C. and da Costa, N. M. (2020) 'A petascale
automated imaging pipeline for mapping neuronal circuits with high-throughput transmission electron
microscopy', *Nat Commun,* 11(1), pp. 4949.

Young, L. N. and Villa, E. (2023) 'Bringing Structure to Cell Biology with Cryo-Electron Tomography',
Annu Rev Biophys, 52, pp. 573-595.

Zhang, H., Jin, B. and Faber, J. E. (2019) 'Mouse models of Alzheimer's disease cause rarefaction of
pial collaterals and increased severity of ischemic stroke', *Angiogenesis,* 22(2), pp. 263-279.

Zhu, J., Li, Z., Ji, Z., Wu, Y., He, Y., Liu, K., Chang, Y., Peng, Y., Lin, Z., Wang, S., Wang, D., Huang,
K. and Pan, S. (2022) 'Glycocalyx is critical for blood-brain barrier integrity by suppressing caveolin1-
dependent endothelial transcytosis following ischemic stroke', *Brain Pathology,* 32(1), pp. e13006.

Figures

Figure 1. Correlative confocal microscopy and ATUM-SEM reveals hallmarks of vascular occlusions and extravasation. (**A**) Schematic drawing of the ATUM-SEM CLEM workflow. Controlled cortical impact (middle) is followed by systemic injection of 30 nm diameter LNDs (magenta) and 30 nm diameter colloidal gold nanoparticles (black). DyLight 649 lectin (green) was injected 5 min before perfusion (1). After fixation, coronal vibratome sections are generated (2) and positioned onto glass slides with only laying the cover slip on top of the section for confocal imaging (3). The vibratome section is recovered by immersion into the petri dish (4). The previously imaged region of interest is dissected (5) and processed for EM including embedding into resin and contrast enhancement (6).

Serial ultramicrotomy and tape collection (7) is followed by wafer mounting and SEM imaging (8). (**B**)

Top: Sum projection of a confocal tile scan (lectin, green; LNDs, magenta; left) and corresponding

binocular image (right) of the dissected vibratome section. The lesion area is indicated by a dashed

line. Scale bars 200 μm. Middle: Lesion area identified in the overview tile scan (box in top image) is

relocated in the sum projection confocal image and in the serial section low resolution SEM (right).

Scale bars 50 μm. The region of interest (ROI, box) is chosen in the confocal image and re-located in

the sum projection SEM image. Bottom: Region of interest confocal image (left) and three single SEM

micrographs overlaid with the correlated confocal images (right). Scale bars 5 μm. (**C**) Scheme of the

ATUM-SEM strategy for CLEM. A blood vessel of interest (magenta) in an 80 μm thick vibratome

section is relocated by screening serial ultrathin sections at low resolution. The target region is

reimaged at high resolution (up to 5×5×50 nm). (**D**) The correlated region in (B) was segmented and

reconstructed from the ultrastructural data. Endothelium (ec, dark green), LNDs and gold particles (Au,

magenta), monocyte (mc, orange), neutrophil (np, yellow), platelet (pt, blue), fibrin (fi, grey),

erythrocytes (ery, red), pericyte (pc, bright green) are shown. Inset: SEM image of vessel lumen filled

with equally sized, putative LNDs (magenta arrowhead) and gold particles (black arrowhead) and

bigger aggregates thereof. Scale bar 100 nm. (**E-G**) Segmented SEM image (top) and three-

dimensional rendering thereof (bottom), showing (**E**) the extravasation site from of a neutrophil

(yellow), (**F**) extraluminal gold particles (arrows) next to vessel lumen clotted with platelets (blue) and

fibrin (grey) and (**G**) stalled erythrocytes (red) interspersed with colloidal gold particles. Scale bars 1

μm. (**H**) Endothelial morphologies from left to right: normal endothelium (dark green) with tight junction

(white arrows) and gold particles (magenta arrows); thinned endothelium covered by a pericyte (bright

green), swollen endothelium with mitochondria and tight junction. Immune cells (orange), erythrocytes

(red), platelets (blue). Scale bars 1 μm.

Figure 2. with 6 supplements. Principle of reversible section attachment on coated tape. (**A**) Schematic showing that sections on ccKapton mounted on a silicon wafer cannot be recovered for TEM. (**B**) Schematics of section removal from cpcKapton or cfcKapton. From left to right: The adhesive tape around the selected section (yellow border) is excised by a razor blade (cutting line: yellow dashed lines) and detached from the wafer. The section is detached from the tape by acetone rinsing using a pipette. The section is collected from the acetone bath onto a TEM slot grid. (**C**) Detachment workflow of a particular selected section (yellow border). From left to right: SEM overview image of a section (scale bar 200 nm); the imaged section mounted on the wafer; the section is removed together with the underlying adhesive tape (cutting line: yellow dashed lines); the section is

floating on a water bath of a diamond knife after acetone treatment; slot grid with the section. (**D**) SEM
images of mouse cortex specimens collected onto different Kapton tapes. From left to right: ccKapton;
cfcKapton showing a charging artifact (arrow); cpcKapton red with major charging; cpcKapton black
without plasma treatment showing section folds; cpcKapton black after intense plasma treatment.
Scale bars 50 μm. (**E**) Photograph of the coating unit with a permanent pen and halfway labeled
ccKapton. (**F**) Photograph of cpcKapton before discharging (top), after mild (middle) and extensive
(bottom) plasma discharging. (**G**) Ultrastructural quality of a cortical tissue section on ccKapton (left)
and cpcKapton after plasma treatment (right) imaged at 10×10 nm resolution. Scale bars 2 μm. (**H**)
Recovery of an ultrathin section. From left to right: SEM overview image of cortex with corpus
callosum (scale bar 100 μm). SEM medium resolution image thereof (scale bar 20 μm). SEM high-
resolution image (10×10 nm) of a blood vessel cross section (scale bar 2 μm). TEM image of the same
section showing the selected blood vessel (scale bar 2 μm). High magnification image of a tight
junction (arrowheads), scale bar 200 nm. High-magnification images of endothelial vesicles (scale bar
50 nm). Black boxes indicate location of the image to the right.

The online version of this article includes the following figure supplements for figure 3:

Figure supplement 1. Formvar coating for reversible section collection.

Figure supplement 2. Optimizing SEM imaging conditions for section detachment.

Figure supplement 3. Contamination of section surface after detachment.

Figure supplement 4. Estimation of distortion between SEM and TEM images.

Figure supplement 5. Resin test for section recovery.

Figure supplement 6. Tape plasma discharging unit construction and use.

Fig. 3. Correlative ATUM-Tomo of vascular occlusions. (**A**) Schematic of the ATUM-Tomo workflow.
Semi-thick sections are serially imaged by SEM. Targeted sections (yellow) bearing the site of interest
(blood vessel, magenta) are selected for ET. (**B**) Schematic highlighting how ATUM-Tomo and CLEM-
ATUM-Tomo bridge scales in resolution and field of view. TEM tomography enables high-resolution
imaging (e.g. vesicular structures (diameter ~30 nm), ATUM can visualize the cellular composition and
morphology (e.g. a neurovascular unit) (diameter ~30 nm) and confocal provides a large field of view
(e.g. highlighting a lesion site in cortex, 3×3 mm). (**C**) Confocal microscopy images of the traumatic
brain injury lesion. Lectin (green), LNDs (magenta). Numbered boxes indicate positions selected for
SEM imaging. Scale bars 200 μm, 20 μm (**D**) Reconstruction of selected vessels as indicated in (C).
Scale bar 10 μm. (**E**) Regions from the vessels 1 (a, b) and 2 (c) in (D) selected for ET. SEM image
(first column), corresponding ET low- (second column) and high-magnification images with
segmentation (third column) and three-dimensional reconstruction thereof (fourth column). (a)
Example of particles in endothelial swellings (red arrow gold particle, magenta arrow LND). (b) Intact
tight junctions (white arrow). (c) Endothelial disruption and gold particle localization in a platelet.
Endothelium (green), pericyte (yellow), platelet (blue), basement membrane (cyan), LNDs (magenta),
gold nanoparticles (red), tight junctions (white).

Fig. 4. Correlative ATUM-Tomo with serial-section ET. (**A**) Schematic of the ATUM-Tomo workflow.
Semi-thick sections are serially imaged by SEM. Targeted sections (yellow) bearing the site of interest
(blood vessel, magenta) are selected for serial section ET. Tomograms from consecutive ET volumes
can be aligned. (B-D) The boxed region in Fig. 3 D (number 2) has been subjected to ET including
nine consecutive sections. (endothelium in green, pericyte in yellow, platelet in blue, basement

membrane in cyan, LNDs in magenta, gold nanoparticles in red. (**B**) Non-smoothed, raw

reconstructions of nine consecutive SEM images. (**C**) Corresponding region of nine detached sections

subjected to ET and reconstructed. (**D**) Side view (left) and longitudinal cut view (right) of the

smoothed serial-section ET reconstruction. Scale bars 1 µm.

 Supplementary Figures

 Figure 2 – Figure supplement 1. Formvar coating for reversible section collection. Ultrathin (80 nm)
 (**A**) and semi-thin (300 nm) (**B**) sections imaged by SEM (BSD detector, 8 kV (increase in
 magnification from left to right)). Irregular Formvar coating can be detected on the left and middle
 image of ultrathin sections (**A**), while the semi-thin sections in (**B**) do not show any Formvar-related
 background. High-resolution images of both ultra- and semi-thin sections (right images in A and B) are
 free of background. Scale bars 100 μm (left), 50 μm (middle), 5 μm (right). (**C**) Photos of the tape
 coating unit attached to a spraying unit (Lubrimat L60) for dispersion of Formvar onto the tape. The top
 image shows the red spraying unit with Formvar reservoir (Falcon tube) as well as the reel-to-reel
 system for movement of the ccKapton tape. The bottom photo shows that the spray nozzle dispersing
 the Formvar is surrounded by a Plexiglas cylinder that the tape can transverse.

 Figure 2 – Figure supplement 2. Optimizing SEM imaging conditions for section detachment. (**A**)
 Ultrathin (100 nm thick) and (**B**) Semi-thin (400 nm thick) sections imaged at 10 nm (left) and 5 nm
 (right) lateral resolution at 5 kV, 4 kV beam deceleration. Only sections imaged at 10 nm lateral
 resolution could be detached from cpcKapton after imaging, the ones imaged at 5 nm resolution were
 stuck to the tape. Scale bars 2 μm.

Figure 2 – Figure supplement 3. Contamination of section surface after detachment. (**A**) Ultrathin sections imaged by SEM (left) and TEM (right). The TEM image shows contaminations (arrows). Scale bars 2 µm (left), 5 µm (right). (**B**) Schematic illustration of a section collected on either cpcKapton (permanent marker coat, blue) tape (top row) or on a slot grid (bottom row). The side of the section facing the coating is contaminated (light blue) up to a thickness of 50 nm while the rest of the section thickness (grey) does not show any dirt originating from the coating. Single TEM images of a tomographic reconstruction at different depths, indicated in the scheme (z_1-z_4) are shown. High-magnification images (bottom row) of the boxed regions (top row) reveal details of back projection artifacts due to the contamination. The left (z1, top image) and the middle left image (z2) show no artifacts; the middle right image (z3, 50 nm distance to the support) shows some grey spots (arrow), while the right image (z4, bottom section) appears 'dirty'. Scale bars 1 µm (top row), 500 nm (bottom row).

Figure 2 – Figure supplement 4. Estimation of distortion between SEM and TEM images. Ultrathin
section SEM and TEM images were compared after rigid (top) and affine (bottom) alignment to assess
image distortion. We picked a clean example (**A**) and one region with some contamination (**B**). For
both examples and alignment algorithms we calculated the Pearson coefficient between the SEM and
TEM image. The images show the spatial distribution of colocalization (left, ImageJ 'Colocalization
Finder') and mislocalization (middle, ImageJ 'Colocalization Highlighter'). The right images are
calculated from the transformation of rigid to affine alignment (ImageJ 'Image Registration with SIFT').
Matching is indicated in green, diverging positions are shown in red (right column). Scale bars 2 µm
(A), 10 µm (B).

Figure 2 – Figure supplement 5. Resin test for section recovery. Section recovery was possible for
all tested resin formulations, Durcupan (**A**), Epon (**B**) and LX112 (**C**). The first two columns show SEM
images at increasing magnifications. The third and fourth column show single tomographic slices after
reconstruction at increasing magnifications. The white boxes indicate the position of the zoomed in
image to the right, respectively. Scale bars 10 µm (SEM) and 2 µm (ET).

A

B

C

D

Figure 2 – Figure supplement 6. Tape plasma discharging unit construction and use. (**A**) Schematics
with dimensions of the vacuum chamber and with the dynamic O-ring shaft seals. (**B**) Photo of the
spool holder design: Plastic burls (red arrow) that fit into the spool indentation and prevent spool slip
off as well as tightening of the turning rod (blue arrow) are highlighted. (**C**) Photos of the discharging
unit: Front (top row), side (bottom left) and top (bottom right) views showing the easiGlow plasma
discharging unit with its custom-built chamber and the gear motor mounted directly onto the reel
holder. (**D**) Photos demonstrating the insertion of the reel into the vacuum chamber. Tape reel
orientation (left), reel fixation using a finger (middle) and mounting onto the spool holder (right) are
shown.

Supplemental movie 1. Rotation of the 3D model as shown in Figure 3Ea.

Supplemental movie 2. Rotation of the 3D model as shown in Figure 3Eb.

Supplemental movie 3. Rotation of the 3D model as shown in Figure 3Ec.

Supplemental movie 4. Rotation of the 3D model as shown in Figure 4B.

Supplemental movie 5. Rotation of the 3D model as shown in Figure 4C.

Supplemental movie 6. Rotation of the 3D model as shown in Figure 4D.