Feasibility of high-resolution perfusion imaging using Arterial Spin Labelling MRI at 3 Tesla

Sriranga Kashyap[1], **Ícaro Agenor Ferreira Oliveira**[1], **and Kâmil Uludağ**[1,2,3,4]

[1]Krembil Brain Institute, University Health Network, Toronto, ON, Canada
[2]Department of Medical Biophysics, University of Toronto, Toronto, ON, Canada
[3]Center for Neuroscience Imaging Research, Institute for Basic Science, Suwon, Republic of Korea
[4]Department of Biomedical Engineering, Sungkyunkwan University, Suwon, Republic of Korea

ABSTRACT

Cerebral blood flow (CBF) is a critical physiological parameter of brain health and it can be non-invasively measured with arterial spin labelling (ASL) MRI. In this study, we evaluated and optimised whole-brain, high-resolution ASL as an alternative to the low-resolution ASL employed in the routine assessment of CBF in both healthy participants and patients. Two high-resolution protocols (i.e., pCASL and FAIR-Q2TIPS (PASL) with 2 mm isotropic voxels) were compared to a default clinical pCASL protocol ($3.4 \times 3.4 \times 4$ mm^3), all of whom had an acquistion time of ≈ 5 min. We assessed the impact of high-resolution acquisition on reducing partial voluming and improving sensitivity to the perfusion signal, and evaluated the effectiveness of z-deblurring on the ASL data. We compared the quality of whole-brain ASL acquired using three available head coils with differing number of receive channels (i.e., 20, 32 and 64 ch). In conclusion, this study demonstrates the feasibility of high-spatial resolution whole brain ASL within clinical scanning duration, offering higher spatial fidelity and resolving power than those obtained with current standard clinical ASL protocols.

Keywords : Arterial Spin Labelling | Brain Perfusion | High spatial resolution | 3 Tesla | RF-coil | Neuroimaging

Introduction

Arterial spin labelling (ASL) is a non-invasive neuroimaging technique that uses magnetically labeled arterial blood water as an endogenous tracer to measure Cerebral Blood Flow (CBF)[1,2]. ASL provides a safe and repeatable method for assessing brain state and function without any risk of toxicity or allergic reactions from exogenous contrast agents. ASL can also be utilised to assess quantitative CBF in units of mL/100g/min at an individual voxel level[1,3].

In recent years, technological advances in MRI scanner hardware and software and new cutting-edge analysis methods have positively impacted the range of ASL applications and resulted in notable growth in publications[4–6]. Another factor for its increasing popularity in clinical research is the community effort to standardise acquisition methods, data structures, and analyses[7–9]. However, widely adopted standards (e.g., described in the ASL 'white paper'[7]) prescribe spatial resolutions of 3-4 mm in-plane and 4-8 mm slice thickness for ASL scans that are maximally 5-6 minutes long (typical length of clinical / standard-of-care MRI protocols). While these protocols may suffice for macroscopic effects (such as pattern of large regions of hypo-perfusion), it is insufficient to detect subtle abnormalities that may represent early-stage of neurological diseases, such as stroke, or small lesions[10]. Therefore, ASL at higher spatial resolution ($<$ 3 mm nominal, isotropic) is highly desirable.

Another reason for going to high spatial resolutions is to reduce partial volume (PV) effects, which occur when the voxel signal contains fractional contributions from more than one tissue type, e.g., grey matter (GM), white matter (WM) and cerebrospinal fluid (CSF). This can introduce inaccuracies in perfusion quantification of the tissue of interest resulting in either or both under-estimation and over-estimation, depending on the PV fractions in the voxels. For instance, Asllani et al.[11] show that a voxel mixture of 80:20% grey:white matter (this ratio would be inclusive after threshold, in most cases), would result in a 24% perfusion under-estimation. Another example is the study by Donahue et al.[12], investigating the impact of a higher-resolution ASL protocol compared to low-resolution Positron Emission Tomography (PET) scans, and they demonstrate that uncorrected CBF PET images might underestimate the grey matter (GM) CBF by 20%. In fact, in 2006, Donahue et al.[12] actually envision the future of ASL imaging at 3 T to be higher resolutions of 2.5mm in-plane or higher. 17 years since, 3 T ASL imaging is still routinely carried out with voxel sizes $>$ 3 mm, and the voxels are almost never isotropic, which can lead to underestimation of lesions and even misdiagnosis in the direction of the lowest spatial resolution. While there haven been methods and algorithms developed that can provide a means to post-hoc correct for PV effects[13–15], they can usually not recreate lost information and, therefore, the most straight-forward and preferred approach is to just acquire the data at higher spatial resolutions.

This is notwithstanding high-resolution ASL studies carried out at field strengths higher (4.7 T, 7 T) than that typically used in the clinic (1.5 T, 3 T). For example, Mora Álvarez et al.[10] demonstrated the feasibility of a high-resolution Continuous ASL (CASL) at 4.7 T within a clinical time frame of 6 minutes. The study also observed reduced PV averaging at $1.5{\times}1.5{\times}3mm^3$ resolution. Another interesting example

is the study published by[16] where they employed Turbo-FLASH (Fast Low Angle SHot) ASL, both pseudo-continuous ASL (pCASL) and pulsed ASL (PASL) at 7 T showing the feasibility of achieving an in-plane resolution of $0.85 \times 1.7mm^2$. At 7 T, recent functional MRI (fMRI) studies also showed the feasibility of using perfusion-weighted contrast with ASL at sub-millimetre spatial resolutions of 0.9 mm isotropic[17] and 0.7 mm isotropic[18,19] using a 3D-EPI[20] readout with a FAIR[21] QUIPSS II[22,23] labelling scheme.

While there is evidence of the transformative potential that ultra-high field scanners can have for clinical applications, they are limited in availability compared to the ubiquity of 3 T scanners. Therefore, a translation of high-resolution ASL to widely available 3 T clinical platforms is urgently needed to catalyse research as well as further advance the clinical standards-of-care. This requires systematic optimisation attuned to easily accessible workflows, currently not explored in existing ASL literature.

The current study addresses these aforementioned challenges and gaps in literature by first, developing, testing, and evaluating high-resolution ASL protocols at 3 T in clinically feasible times and compare them to a vendor default protocol that are typically used in routine clinical scanning. To this end, we developed optimised 2 mm isotropic pCASL and PASL protocols that balances the trade-off between signal-to-noise (SNR) and acquisition time (TA) for clinical feasibility (TA $\approx$ 5 minutes). Furthermore, we also evaluated the impact on 3 T perfusion imaging of the the choice of standard head coils. Therefore, we systematically evaluated our protocols and the clinical default with all three commercially available head coils (20 channel head & neck, 32 channel head only, 64 channel head & neck) to ascertain the optimal hardware for high-resolution acquisitions. Additionally, we quantify and demonstrate the reduction in partial voluming enabled by the high-resolution acquisitions. Finally, the lengthening of the readout with 3D-GRASE is recognised to result in through-plane (z-axis) blurring resulting in loss of spatial resolution[24,25]. We also assess the impact of advanced post-processing methods such as z-deblurring to improve spatial fidelity of the acquired data.

Methods

Participants

Eight healthy volunteers (4 females and 4 males, mean age $= 29 \pm 4$ years) participated in the study and provided written informed consent prior to scanning. All participants were screened healthy individuals, non-smokers, not taking any medications, and with no history of neurological or neurovascular conditions. All procedures in this study conformed to the standards set by the Declaration of Helsinki and was approved by the Research Ethics Board of University Health Network according to the guidelines of Health Canada.

Data acquisition

Data were acquired on a Siemens Magnetom Prisma 3 T MRI scanner (Siemens Healthineers, Erlangen, Germany) at the Slaight Family Centre for Advanced MRI (Toronto Western Hospital, Toronto ON, Canada) having a maximum gradient strength of 80 mT/m, slew rate of 200 T/m/s and running on the XA30A IDEA software platform. We used three commercial MRI coils available at the centre, namely, a 20ch Head and Neck, a 32ch Head only, and a 64ch Head and Neck for receive, and the transmit was carried out by the body coil. Participants were positioned by taking the the eye centres as reference for the magnet isocentering to minimise B_0 offsets for the labelling in the neck. All data in participant were acquired in the same scan session. The participants were brought out of the scanner, coils were exchanged, and the participants were repositioned to the magnet's isocentre. The sequential order of coils was pseudo-randomised between participants to avoid any systematic biases.

Anatomical imaging

Structural scans were acquired with the 32ch Head coil. Whole-brain anatomical data were acquired using a 3D Multi-Echo Magnetisation-Prepared Rapid Gradient Echo (3D-MEMPRAGE) sequence[26] that uses volumetric EPI navigators combined with selective data reacquisition[27] to produce (prospectively) motion-corrected T_1w images[28] that were used in the study. 3D-MEMPRAGE data were acquired at 0.8 mm isotropic resolution, TI = 1000 ms, TEs_{1-4} = 1.81, 3.6, 5.39, 7.18 ms, TR = 2500 ms, $\alpha = 8°$, 208 sagittal slices, matrix = 320×320, GRAPPA = 2, Ref. lines = 32, partial-Fourier$_{slice}$ = 6/8, echo spacing = 11.2 ms, bandwidth = 740 Hz/pixel, turbo factor = 168, total acquisition time $\approx$ 8 min). The four echoes were combined (using root mean squares, RMS) into a high fidelity T_1-weighted image following the scanner's on-line reconstruction. Quantitative T_1 mapping was carried out using a 3D Magnetisation-Prepared 2 Rapid Gradient Echoes (3D-MP2RAGE) sequence[29]. The MP2RAGE T_1 maps were only used to facilitate perfusion quantification and thus were acquired at a 1.2 mm isotropic resolution, TIs_{1-2} = 700, 2500 ms, $\alpha_{1-2} = 4°, 5°$, TE = 4.04 ms, TR = 3200 ms, 144 axial slices, matrix = 192×192, GRAPPA = 2, Ref. lines = 32, partial-Fourier$_{phase}$ = 6/8, echo spacing = 9.08 ms, bandwidth = 150 Hz/px, turbo factor = 144, total acquisition time $\approx$ 4 min). T_1 maps were calculated in-line using the Siemens MapIt package (Siemens Healthineers, Erlangen, Germany).

Perfusion imaging

All ASL protocols were developed using the Siemens Advanced 3D-ASL work-in-progress (WIP) sequence (courtesy of Siemens Healthineers, Erlangen, Germany) available for the XA30A baseline platform. The ASL data were acquired with a segmented 3D-GRASE readout for improved SNR[30–32]. Three ASL protocols were acquired per coil in each participant: 1) the clinical default protocol ($3.4 \times 3.4 \times 4$ mm, 'Clinical' in Table 1); 2) a high-resolution (or hires) pCASL protocol (2 mm isotropic, 'Hires' in Table 1); and 3) a hires PASL protocol employing a FAIR Q2TIPS[33] labelling scheme (2 mm isotropic, 'PASL Hires' in Table 1). For clinical and hires ASL variants, two steady-state magnetisation (M_0) calibration images were acquired without any labelling, but with matched readout and TR increased to 20 s, one of the M_0 had the opposite phase-encoding for distortion-correction. The new hires protocols developed in this study were acquired in approximately the same total time as the spatially anisotropic clinical ASL scan ($\approx 5min$).

Parameters	Clinical	Hires	PASL Hires
Labelling	pCASL	pCASL	FAIR Q2TIPS
TR / TE	4720 / 20.40 ms	4000 / 16.80 ms	4000 / 16.80 ms
Tag Control pairs	12	12	12
BS	Grey-White Strong	Grey-White	Grey-White
Bolus (TI_1) / PLD (TI_2)	1800 / 1800 ms	1600 / 1800 ms	700 / 1800 ms
FOV	220×220	192×192	192×192
Matrix	64×64	96×96	96×96
Slice Oversampling	20%	22.5%	22.5%
Acceleration	GRAPPA 2	2D-CAIPI 3	2D-CAIPI 3
Slices	30	66	66
Partial Fourier	off	6/8 yz	6/8 yz
Bandwidth	2442 Hz/Px	1930 Hz/Px	1930 Hz/Px
Echo spacing	0.49 ms	0.68 ms	0.68 ms
Turbo factor	12	20	20
Segments	3	3	3
EPI factor	31	23	23
TA	05:09 min	05:11 min	05:11 min

Table 1. Table of sequence parameters for the three ASL protocols in the present study. Although the study focuses on the two pCASL protocols, the PASL protocol is included here for completeness.
Abbreviations: TR=Repetition time, TE=Echo time, FA=Flip angle, BS=Background Suppression (Grey-White = 2, Grey-White Strong = 4 non-selective pulses), PLD=Post-labelling delay, TI=Inversion Time (for PASL), FOV=Field-of-view, TA=Acquisition time

Data Processing

Anatomical imaging

The RMS-combined, motion-corrected T_1-weighted 3D-MEMPRAGE was processed using Freesurfer v 7.3.2[34–36](https://surfer.nmr.mgh.harvard.edu/) using a brain mask that was generated using *mri_synthstrip*[37] and was provided as an additional input to the *recon-all* pipeline.

Perfusion imaging

The first volume of the ASL timeseries was discarded as separate M_0 scans had been acquired for quantification. The pre-processing steps were carried out using FSL[38] included motion and distortion-correction, where all control and label volumes were independently realigned to the first volume of the ASL scan. The separately acquired M_0 scans were rigidly registered to the first volume of the ASL scan and then distortion correction was done using FSL's *topup*[39] with the two M_0 images. The perfusion timeseries were calculated using sinc-subtraction as implemented in FSL's *perfusion_subtract*. The M_0 images, perfusion-weighted data, and the MP2RAGE T_1 maps (coregistered to M_0) were used as input to *oxasl*[40] (https://github.com/physimals/oxasl) for voxelwise perfusion quantification. The M_0s were co-registered to the anatomical image using Freesurfer's *bbregister*[41] to obtain CBF maps in both native and structural space. No adaptive spatial smoothing[42] or partial-volume correction[43] was applied. Next, all anatomical scans were carefully registered to the 1 mm isotropic MNI Non-linear 2009c asymmetric template space[44,45] using ANTs SyN algorithm[46,47] (https://github.com/ANTsX/ANTs). Native space maps from *oxasl* were resampled in a single step to the MNI space using *antsApplyTransforms*. The stability of the perfusion signal over time (temporal SNR, tSNR) was calculated dividing the temporal mean by the temporal standard deviation of the perfusion-weighted data (also referred to as, perfusion tSNR). The SNR (consequently, tSNR) of a voxel is expected to scale proportionally with its volume and thus, makes it challenging to compare datasets of highly different spatial resolutions. Therefore, to better appreciate the tSNR relative to a dataset's spatial resolution, the perfusion tSNR map from the hires scan was scaled by the ratio of the voxel volumes of clinical to hires datasets (46.24 / 8 mm^3 = 5.78).

Partial volume analysis

In order to visualise the impact of the higher spatial resolution acquisition, participant-wise T_1-weighted images were resampled to the nominal spatial resolution of the clinical protocol ($3.4 \times 3.4 \times 4.0$ mm) or the hires protocol (2.0 mm isotropic). The resampled T_1-weighted images were segmented using FSL's *fsl_anat*[48] to obtain PV estimates. The cortical grey matter segmentation from Freesurfer was morphologically dilated by 1 voxel and resampled to the two resolutions and this resampled, dilated cortical mask was used as the ROI for the PV analyses. To this end, we used a histogram-based analysis to first sort the voxels into different PV fraction bins. Then, to compare the two different acquisition resolutions, the number of voxels in each histogram bin was scaled by their voxel volumes of 46.24 mm^3 and 8 mm^3, respectively, for the clinical and hires protocols giving us the volume of PV voxels in each bin. This normalisation enabled direct comparison of the PV. A difference between the hires and clinical

¹⁵³ histograms (after rescaling) were computed for all values above a PV fraction threshold of 0.5 for each of
¹⁵⁴ the three tissue classes, namely, GM, WM and CSF.

Deblurring analysis

¹⁵⁶ In an additional analysis, ASL data acquired from the 32ch coil were pre-processed using the *oxasl_deblur*
¹⁵⁷ (https://github.com/physimals/oxasl_deblur). We evaluated two different methods for
¹⁵⁸ deblurring the data, namely, fast Fourier Tranform division (FFT) and Lucy-Richardson deconvolution
¹⁵⁹ (Lucy) as implemented in *oxasl_deblur* with three different kernel options (direct estimation, lorentzian,
¹⁶⁰ lorentzian with a weiner filter). Smoothness of the deblurred data was estimated using AFNI's[49,50]
¹⁶¹ *3dFWHMx*[51] function.

Results

Comparison of clinical and hires ASL data

Figure 1 shows the group average relative CBF (rCBF) maps from clinical and hires pCASL protocols presented in three orthogonal views (middle panel) for the three head-coils used to acquire the data (drawing in left panel). The panel on the right shows the distribution of the rCBF values in GM across the participants' data as a violin plot with the rCBF values represented on the y-axis for the two protocols. The figure annotation represents the mean $\pm$ standard deviation of the distribution. A comparison of the group average perfusion-weighting and quantitative CBF (in units of mL/100 g/min) for the two protocols and three head-coils is shown in Supplementary Figure S1. Summary statistics for all the different perfusion metrics calculated from the data are tabulated in Supplementary Tables S2-S6. While the mean rCBF values from the hires data are very similar to that from the clinical protocol (e.g., 32ch: 1072.07 ± 561.64 vs 1016.20 ± 435.08 a. u.), the difference between the datasets is in their inter-quartile ranges (IQR). The hires data has $\approx 25\%$ greater IQR (Table S3, 32ch: 759.0) compared to the clinical data (Table S3, 32ch: 566.97).

Analysis of the partial voluming

One of the primary advantages of acquiring higher spatial resolution data is to reduce partial voluming of the signal of interest. In Figure 2, the difference in voxel volumes (hires−clinical) are plotted at each partial volume fraction bin ranging from 0.5-1.0 (50% to 'pure' single tissue composition) for three tissue classes, that is, GM (a), WM (b) and CSF (c) using a dilated GM ROI. Data from each participant is shown as a coloured dot, with the mean across participants plotted as a black dashed line. In Figure 2a, we observe that on average, above a PV fraction of 0.6 (60% GM), there is a net positive change in the volume of GM and remains positive for all higher PV fractions. In other words, even within the dilated GM ROI, there is a larger volume (total ≈ 6713 mm^3) of 'pure' GM in the hires compared to the clinical data, therefore, partial voluming is reduced. This finding is corroborated by the spatial maps of PV as illustrated in the right panels with the PV map of hires and clinical shown on top and bottom rows, respectively. A similar pattern is observed for Figure 2b and 2c that quantifies the PV in WM and CSF respectively. That is, within the dilated GM ROI used to extract these results, there is a significantly larger volume of 'pure' WM (total ≈ 16938 mm^3) and 'pure' CSF (total ≈ 15422 mm^3). As GM is bound on either side with WM and CSF, we can infer that greater the number of 'pure' non-GM voxels, lower is the amount of voxels which are PV with GM, and this finding is corroborated by the spatial maps of PV.

Figure 1. Mosaic of orthogonal views of the relative CBF (arbitrary units) for data acquired using the three respective head coils (drawing on the left). In the middle column, the maps obtained from the clinical and hires pCASL acquisitions are displayed in the top and bottom rows, respectively. (Right) Violin plots of the rCBF distribution across all participants' data for the two acquisitions. The annotation represents the mean ± standard deviation of the distribution.

Figure 2. Histogram difference plot for all bins ≥ 0.5 threshold of PV fraction for (a) GM, (b) WM, (c) CSF tissue classes of the hires and clinical acquisitions. Value from each participant is represented as a colour-coded circle and the group average is plotted as a black dotted line. A single-participant PV estimate map is shown in the right panel for the clinical and hires spatial scales, spatially illustrating the findings of the histogram-analysis. Please note, sub-07 is excluded from this analysis as *fsl_anat* could not be completed.

Deblurring analysis of 3D-GRASE ASL

Table 2 shows that the effective spatial resolution of both the clinical and hires datasets is different to is indicated in the protocol, also referred to as the nominal spatial resolution (in this study, 3.4×3.4×4.0 mm and 2.0 mm isotropic respectively). Systematic evaluation of five parameter combinations in *oxasl_deblur* (Table S2) shows that all five combinations result in an improvement in the image fidelity, or in other words, a reduction in the Full Width at Half Maximum (FWHM). We found that using the FFT method with direct kernel estimation to yield the smallest effective FWHM (clinical: 6.38 ± 0.33 mm vs hires: 2.38 ± 0.13 mm, Table S2). Table 2 shows the FWHM estimated from AFNI's *3dFWHMx* for x-,y- and z-axes as well as the effective FWHM (ACF). We observe that irrespective of the acquisition resolution, the smoothness is maximal along the z-axis (Clinical: 8.73 ± 0.65 mm, Hires: 4.01 ± 0.31 mm) and this is the axis along which *oxasl_deblur* is most effective, reducing the smoothness estimate to 5.22 ± 0.49 mm and 1.41 ± 0.12 mm for Clinical and Hires data respectively. The change in the smoothness estimate after deblurring method is 2.7 times greater for the hires dataset than the clinical data. Figure **??** shows the group-average rCBF maps for the clinical and hires datasets before ('orig') and after ('deblurred') deblurring and the distribution of rCBF values across all participants' data is also shown as a violin plot.

Table 2. Table of FWHM (in mm) estimated using AFNI's *3dFWHMx* for the clinical and hires 3D-GRASE datasets deblurred using the FFT method and direct kernel estimation as implemented in *oxasl_deblur*. Comparison of FWHM for different deblurring methods can be found in Table S2 (Supplementary Material). Numerical values presented are mean $\pm$ std. dev across participants.

FWHM	Clinical	Clinical Deblurred	Hires	Hires Deblurred
x	5.51 ± 0.27	5.49 ± 0.76	3.08 ± 0.18	2.65 ± 0.21
y	6.24 ± 0.20	5.91 ± 0.55	2.85 ± 0.17	2.42 ± 0.22
z	8.73 ± 0.65	5.22 ± 0.49	4.01 ± 0.31	1.41 ± 0.12
ACF	10.08 ± 0.67	7.24 ± 0.58	4.83 ± 0.36	2.69 ± 0.12

Figure 3. (Top) Orthogonal views of the grou-average relative CBF (rCBF) maps obtained using the 32 channel head coil, before and after deblurring using the FFT method and direct kernel estimation as implemented in *oxasl_deblur*. (Bottom) Violin plots of the rCBF distribution across all participants' data for the two acquisitions before and after deblurring. The annotation represents the mean $\pm$ standard deviation of the distribution.

Impact of head-coil choice for imaging perfusion

Figure 1 demonstrates that robust rCBF maps can be acquired independently of the coil choice. However, the spatial distribution of the rCBF maps from the hires protocol show a preference for the 32 and 64 ch. Figure 4 top and middle rows illustrate the impact of perfusion tSNR across the three coils. In the case of the clinical protocol, the increasing coil count has an $\approx$ 2-2.5% gain in perfusion tSNR, whereas, the hires protocol has an $\approx$ 34-42% gain in perfusion tSNR with increasing coil count (Table S5). The perfusion tSNR maps of the hires data, rescaled by the ratio of voxel volume (Figure 4, bottom row), illustrates the improvement of tSNR with 32 and 64 coils over the 20 ch. Additionally, Table S3 shows that the inter-quartile range (IQR) of the rCBF values increases with increasing coil count (20/32/64 ch: for clinical, 533.11/566.97/574.38 a. u., and for hires 693.79/759/788.37 a. u.) for both protocols. The IQR of rCBF between the three coils behaves similarly with the hires PASL protocol (20/32/64 ch: 978.52/1103.48/1121.65 a. u.). Therefore, it is the SNR benefits afforded by higher coil count rather than quality or type of labelling used that is responsible of the improvement in the higher IQR of rCBF values.

Figure 4. Orthogonal views of the group average perfusion tSNR maps where the pCASL Clinical and Hires maps are presented in the top and middle rows respectively. The bottom row shows the hires tSNR data (middle row) but with rescaled values.

Discussion

In this study, we demonstrate that it is feasible to measure perfusion robustly and repeatably using ASL at a high-spatial resolution of 2 mm isotropic within clinically-feasible times of $\approx$ 5 minutes. The hires protocols showcased in the present study were developed using vendor-provided sequences only, in order to allow wide-spread usage without the necessity to developing custom MR sequences or image reconstruction, despite significant progress being made on the development of ASL methods[8]). We show that the increased spatial resolution does result in a reduction of partial voluming compared to the default clinical protocol. We show that through-plane blurring is a problem for 3D-GRASE ASL independent of the protocol being used. We find z-deblurring to be more effective on the hires than the clinical data. Finally, the choice of head coil for imaging perfusion with ASL at 3 T does play an important role with the 32 and 64 ch being particularly well-suited. Consistent with the results of deblurring, the hires datasets benefit most from perfusion tSNR improvments with higher coil counts.

Impact of spatial resolution on ASL imaging

We show that increasing the spatial resolution of ASL 5.78$\times$ the clinical resolution does not have a detrimental effect on the measuring perfusion (Figure 1) and takes the same duration as a clinical scan ($\approx$ 5 min). The mean rCBF values from the high-resolution data were found to be similar to the clinical data (e.g., 32-channel: 1072.07 $\pm$ 561.64 vs. 1016.20 $\pm$ 435.08 a. u.) (Table S3). Importantly, however, the hires rCBF data exhibited approximately 25% greater IQR (32-channel: 759.0 vs. 566.97 a. u.) than the clinical data. Since the IQR is a measure of spread around the mean, this measure is indicative of the dynamic range of perfusion in the data. Being capable of resolving a wider range of perfusion values is critical to detect subtle abnormalities and early-detection of neurological diseases[52], therefore, emphasising the importance of high-spatial resolution imaging[10].

In addition to the improvement in the dynamic range, acquiring data at a higher spatial resolution reduces PV effects. The cortical GM is bound on either side by WM and CSF and PV occurs when a GM voxel contains fractional distributions from these adjacent tissue classes that influence cortical perfusion measures. Figure 2 shows that hires ASL data consistently yields a greater volume of 'pure' tissue voxels than the clinical data (GM: $\approx$ 6713 mm^3, WM: $\approx$ 16938 mm^3, CSF: $\approx$ 15422 mm^3). It is important to note that these PV fractions were derived from a dilated, cortical GM ROI, that is, the ROI does not consist of the large ventricles or the majority of WM in the brain. The increased number of 'pure' WM and CSF voxels indicate that the hires data can enable a more effective separation of non-GM signal contributors to the perfusion signal of interest.

We only discuss the rCBF maps here as they are the most clinically-relevant maps, and all ASL sequences on the Siemens Prisma automatically generate this map and is available in the hospitals' picture archiving and communication system (PACS) soon after. Other maps including that of the perfusion-weighting and quantitative CBF are shown in Figures S1 (clinical vs hires) and S2 (pCASL vs PASL). We found that the hires PASL results are in good agreement with the hires pCASL (Figure S2). Consistent with previous

₂₅₆ work[53,54], the pCASL labelling scheme exhibits about 22-26% higher perfusion tSNR than FAIR Q2TIPS
₂₅₇ for the hires acquisitions in our study.

Impact of deblurring on 3D-GRASE ASL data

₂₅₉ Due to high spatial resolution required, the total echo train length (TE * TF) can exceed 300 ms ($>>$
₂₆₀ $\pi \times T_2^*$ of tissue) resulting in increased blurring[55–57], that occurs maximally in the slice direction (through-
₂₆₁ plane or z-axis). Thus, requiring post-processing correction or making compromises that would render
₂₆₂ whole-brain acquisitions infeasible. We find that application of z-deblurring has a demonstrable effect
₂₆₃ on the improvement of the spatial fidelity (or reducing the estimated FWHM) of the ASL data as shown
₂₆₄ in Table 2. Interesting to note that the FWHM along z for the deblurred clinical data (5.22 ± 0.49 mm)
₂₆₅ is still larger than the non-deblurred hires data (4.01 ± 0.31 mm). This has an important implication in
₂₆₆ clinical setting where advanced image post-processing is often unavailable. Importantly, the hires ASL
₂₆₇ protocols enable researchers and clinicians to resolve perfusion changes with a higher spatial fidelity
₂₆₈ (without requiring advanced image processing) than the post-hoc deblurred clinical datasets. Furthermore,
₂₆₉ post-processing deblurring methods have their limitations and they cannot synthesise resolution from
₂₇₀ information lost in acquisition. While lengthening the echo-train is an important concern, our findings
₂₇₁ (Table 2) indicate that deblurring methods are more effective for high-resolution ASL imaging.

Impact of coil choice on ASL imaging

₂₇₃ We demonstrate that robust relative CBF maps can be acquired independently of the coil choice (Figure 1),
₂₇₄ however, higher coil counts(32 and 64 ch), offer substantial gains in perfusion tSNR compared to the 20 ch
₂₇₅ coil (Figure 4. We find that increasing coil count has an $\approx$ 2-2.5% gain in perfusion tSNR for the clinical
₂₇₆ protocol compared to $\approx$ 34-42% gain for the hires protocol (Table S5). One reason for this difference
₂₇₇ could be that data acquired with clinical protocol in Figure 1 is relatively insensitive to the choice of
₂₇₈ coil due to its low spatial resolution (i.e., low thermal noise) and acceleration (i.e., no g-factor penalty)
₂₇₉ requirements. On the other hand, the hires protocols accelerate higher and have increased thermal noise
₂₈₀ relative the clinical protocol owing to the smaller voxel sizes, and therefore, benefits from the increased
₂₈₁ number of coils (Figure 4).
₂₈₂ Interestingly, Figure 4 shows that the reducing the voxel size (i.e., higher spatial resolution) actually
₂₈₃ results in a gain in perfusion SNR (clinical vs hires (scaled)) which may seem counter-intuitive from the
₂₈₄ standpoint of conventional fMRI where the SNR of the BOLD signal decreases with increasing resolution.
₂₈₅ However, this is due to the different signal origins of the BOLD and perfusion contrasts. By reducing PV
₂₈₆ with veins and macrovasculature, we are reducing the signal contributors of the BOLD signal, whereas,
₂₈₇ these same signal components are sources of noise in perfusion imaging, as they have very low perfusion
₂₈₈ signals. In addition, reducing WM contribution of voxels dominated by GM improves the fidelity of GM
₂₈₉ perfusion values and reduced influence of physical noise stemming from WM. Therefore, reducing PV
₂₉₀ increases our sensitivity to the cortical microvasculature signal. In other words, higher spatial resolution
₂₉₁ decreases image SNR in both BOLD and perfusion methods due to reduction in porton density (i.e., voxel

292 volume), but also reduces noise sources in perfusion imaging stemming from CSF, veins, and WM.

Limitations

For the purposes of the present study, we opted for modest acceleration schemes (Table 1) as the protocols
were to be compared on all three available head coils and the 20 ch coil would be the lowest common
denominator. The availability of the 3D-GRASE readout with 2D-CAIPIRINHA undersampling enabled
us to achieve higher isotropic spatial resolution for perfusion imaging[32,58,59]. For a future non-comparison
type of study, this protocol optimisation can be pushed further, to take advantage of the higher coil count
and achieve higher acceleration. A systematic exploration of different CAIPI acceleration schemes or
trajectories, impact of reduced g-factor noise amplification on image quality is, unfortunately, beyond the
scope of the present work.

Concluding remarks

Taking together, this study demonstrates the feasibility and benefits of imaging perfusion using high-
resolution isotropic ASL for research and clinical applications at 3 T. We have shown that increasing the
spatial resolution does not compromise accuracy, quality of the perfusion maps, and allows for a wider
dynamic range of perfusion values. We show the high-resolution data can more effectively separate out the
non-GM signal contributors (reduce PV effects) which improves sensitivity to cortical microvasculature.
Additionally, post-processing methods such as z-deblurring are important considerations for whole-brain
perfusion imaging using 3D-GRASE ASL to enhance the spatial fidelity of the data. High-resolution
acquisitions take advantage of the higher coil counts and offer substantial gains in perfusion tSNR with
the 32 and 64 ch coils. Echoing what Donahue and colleagues envisioned in 2006[12], we strongly believe
that high-resolution ASL can be adopted into clinical research and establish a new standard for perfusion
imaging using ASL at 3 T.

Conflict of Interest Statement

315 The authors declare that the research was conducted in the absence of any commercial or financial
316 relationships that could be construed as a potential conflict of interest.

Data Availability Statement

318 The data will be made available upon publication.

Author Contributions

320 SK: Conceputalisation, Methodology, Investigation, Data curation, Formal analysis, Visualization, Writing
321 – original draft, Writing – review & editing; IAFO: Methodology, Investigation, Validation, Visualization,
322 Writing – original draft, Writing – review & editing; KU: Conceputalisation, Methodology, Project

administration, Funding acquisition, Resources, Supervision, Writing – original draft, Writing – review & editing.

Funding

The study was supported by the Institute for Basic Science, Suwon, Republic of Korea (IBS-R015-D1) to Kâmil Uludağ.

Acknowledgments

We thank Asma Naheed, MRT for the scanning support at the Slaight Family Centre for Advanced MRI, Toronto Western Hospital, University Health Network, Toronto, Canada. We like to thank Dr. Thomas F. Kirk for the invaluable discussions about the partial voluming problem in MRI data. We like to thank Dr. Dimo Ivanov for his insights into ASL acquisition. We thank Dr. Gerald Moran and Dr. Josef Pfeuffer for their collaboration and providing the Advanced 3D ASL Work-in-progress (WIP) package for the Siemens XA30 platform.

References

1. Williams, D. S., Detre, J. A., Leigh, J. S. & Koretsky, A. P. Magnetic resonance imaging of perfusion using spin inversion of arterial water. *Proc Natl Acad Sci U S A* **89**, 212–216, DOI: 10.1073/pnas.89.1.212 (1992).

2. Detre, J. A., Leigh, J. S., Williams, D. S. & Koretsky, A. P. Perfusion imaging. *Magn Reson. Med* **23**, 37–45, DOI: 10.1002/mrm.1910230106 (1992).

3. Buxton, R. B. *et al.* A general kinetic model for quantitative perfusion imaging with arterial spin labeling. *Magn Reson. Med* **40**, 383–396, DOI: 10.1002/mrm.1910400308 (1998).

4. Lindner, T. *et al.* Current state and guidance on arterial spin labeling perfusion mri in clinical neuroimaging. *Magn. Reson. Medicine* **89**, 2024–2047, DOI: https://doi.org/10.1002/mrm.29572 (2023). https://onlinelibrary.wiley.com/doi/pdf/10.1002/mrm.29572.

5. Detre, J. A., Rao, H., Wang, D. J. J., Chen, Y. F. & Wang, Z. Applications of arterial spin labeled mri in the brain. *J Magn Reson. Imaging* **35**, 1026–1037, DOI: 10.1002/jmri.23581 (2012).

6. Iutaka, T. *et al.* Arterial spin labeling: Techniques, clinical applications, and interpretation. *Radio-Graphics* **43**, e220088, DOI: 10.1148/rg.220088 (2023). PMID: 36367822, https://doi.org/10.1148/rg.220088.

7. Alsop, D. C. *et al.* Recommended implementation of arterial spin-labeled perfusion MRI for clinical applications: A consensus of the ISMRM perfusion study group and the european consortium for ASL in dementia. **73**, 102–116, DOI: 10.1002/mrm.25197.

8. Hernandez-Garcia, L. *et al.* Recent technical developments in ASL: A review of the state of the art. *Magn. Reson. Medicine* **88**, 2021–2042, DOI: 10.1002/mrm.29381 (2022).

9. Clement, P., Castellaro, M., Okell, T. & et al. Asl-bids, the brain imaging data structure extension for arterial spin labeling. *Sci. Data* **9**, 543, DOI: 10.1038/s41597-022-01615-9 (2022).

10. Mora Álvarez, M. G., Stobbe, R. W. & Beaulieu, C. High resolution continuous arterial spin labeling of human cerebral perfusion using a separate neck tagging RF coil. *PLoS ONE* **14**, e0215998, DOI: 10.1371/journal.pone.0215998 (2019).

11. Asllani, I., Borogovac, A. & Brown, T. R. Regression algorithm correcting for partial volume effects in arterial spin labeling mri. *Magn. Reson. Medicine* **60**, 1362–1371, DOI: 10.1002/mrm.21670 (2008).

12. Donahue, M. J., Lu, H., Jones, C. K., Pekar, J. J. & van Zijl, P. C. An account of the discrepancy between mri and pet cerebral blood flow measures. a high-field mri investigation. *NMR Biomed.* **19**, 1043–1054, DOI: 10.1002/nbm.1075 (2006).

13. Chappell, M. A. *et al.* Partial volume correction in arterial spin labeling perfusion mri: A method to disentangle anatomy from physiology or an analysis step too far? *Neuroimage* **238**, 118236, DOI: 10.1016/j.neuroimage.2021.118236 (2021).

14. Kirk, T. F., Coalson, T. S., Craig, M. S. & Chappell, M. A. Toblerone: Surface-based partial volume estimation. *IEEE Transactions on Med. Imaging* **39**, 1501–1510, DOI: 10.1109/TMI.2019.2951080 (2020).

15. Kirk, T. *et al.* Partial volume effect correction of arterial spin labelling data using surface segmentations. In *Proceedings of the Annual Meeting of the International Society for Magnetic Resonance in Medicine (ISMRM)*, 3292 (Virtual, 2020).

16. Zuo, Z. *et al.* Turbo-FLASH based arterial spin labeled perfusion MRI at 7 t. *PLOS ONE* **8**, e66612, DOI: 10.1371/journal.pone.0066612 (2013).

17. Kashyap, S. *et al.* Sub-millimetre resolution laminar fMRI using arterial spin labelling in humans at 7 T. *PLOS ONE* **16**, e0250504, DOI: 10.1371/journal.pone.0250504 (2021).

18. Ivanov, D. *et al.* Whole-brain sub-millimeter isotropic resolution cerebral blood flow map in humans. In *Proceedings of the Annual Meeting of the International Society for Magnetic Resonance in Medicine (ISMRM)*, 2301 (Paris, France, 2018).

19. Kashyap, S. *et al.* The impact of optimal RF coil combination on whole-brain sub-millimetre resolution perfusion imaging at 7T. In *Proceedings of the Annual Meeting of the International Society for Magnetic Resonance in Medicine (ISMRM)*, 1145 (London, UK, 2022).

20. Poser, B. A., Koopmans, P. J., Witzel, T., Wald, L. L. & Barth, M. Three-dimensional echo-planar imaging at 7 tesla. *NeuroImage* **51**, 261–266, DOI: 10.1016/j.neuroimage.2010.01.108 (2010). 20139009.

21. Kim, S.-G. Quantification of relative cerebral blood flow change by flow-sensitive alternating inversion recovery (FAIR) technique: application to functional mapping. *Magn. Reson. Medicine* **34**, 293–301, DOI: 10.1002/mrm.1910340303 (1995).

22. Wong, E. C., Buxton, R. B. & Frank, L. R. Implementation of quantitative perfusion imaging techniques for functional brain mapping using pulsed arterial spin labeling. *NMR Biomed.* **10**, 237–249, DOI: 10.1002/(sici)1099-1492(199706/08)10:4/5<237::aid-nbm475>3.0.co;2-x (1997).

23. Wong, E. C., Buxton, R. B. & Frank, L. R. Quantitative imaging of perfusion using a single subtraction (QUIPSS and QUIPSS II). *Magn. Reson. Medicine* **39**, 702–708, DOI: 10.1002/mrm.1910390506 (1998).

24. Paschoal, A. M., Leoni, R. F., Pastorello, B. F. & van Osch, M. J. P. Three-dimensional gradient and spin-echo readout for time-encoded pseudo-continuous arterial spin labeling: Influence of segmentation factor and flow compensation. *Magn. Reson. Medicine* **86**, 1454–1462, DOI: 10.1002/mrm.28807 (2021).

25. Tan, H., Hoge, W., Hamilton, C., Günther, M. & Kraft, R. 3d grase propeller: Improved image acquisition technique for arterial spin labeling perfusion imaging. *Magn. Reson. Medicine* **66**, 168–173, DOI: 10.1002/mrm.22768 (2011).

26. van der Kouwe, A. J. W., Benner, T., Salat, D. H. & Fischl, B. Brain morphometry with multiecho mprage. *Neuroimage* **40**, 559–569, DOI: 10.1016/j.neuroimage.2007.12.025 (2008).

27. Tisdall, M. D. *et al.* Volumetric navigators for prospective motion correction and selective reacquisition in neuroanatomical mri. *Magn Reson. Med* **68**, 389–399, DOI: 10.1002/mrm.23228 (2012).

28. Tisdall, M. D. *et al.* Prospective motion correction with volumetric navigators (vnavs) reduces the bias and variance in brain morphometry induced by subject motion. *Neuroimage* **127**, 11–22, DOI: 10.1016/j.neuroimage.2015.11.054 (2016).

29. Marques, J. P. *et al.* Mp2rage, a self bias-field corrected sequence for improved segmentation and t1-mapping at high field. *Neuroimage* **49**, 1271–1281, DOI: 10.1016/j.neuroimage.2009.10.002 (2010).

30. Feinberg, D. A., Ramanna, S. & Guenther, M. Evaluation of new asl 3d grase sequences using parallel imaging, segmented and interleaved k-space at 3T with 12- and 32-channel coils. In *Proceedings of the 17Th Annual Meeting of ISMRM; Honolulu, Hawaii, USA* (Honolulu, Hawaii, USA, 2009).

31. Fernández-Seara, M. A. *et al.* Minimizing acquisition time of arterial spin labeling at 3T. *Magn Reson. Med* **59**, 1467–1471, DOI: 10.1002/mrm.21633 (2008).

32. Vidorreta, M. *et al.* Evaluation of segmented 3d acquisition schemes for whole-brain high-resolution arterial spin labeling at 3 t. *NMR Biomed* **27**, 1387–1396, DOI: 10.1002/nbm.3201 (2014).

33. Luh, W.-M., Wong, E. C., Bandettini, P. A. & Hyde, J. S. QUIPSS II with thin-slice TI1 periodic saturation: a method for improving accuracy of quantitative perfusion imaging using pulsed arterial spin labeling. *Magn. Reson. Medicine* **41**, 1246–1254, DOI: 10.1002/(sici)1522-2594(199906)41:6<1246::aid-mrm22>3.0.co;2-n (1999).

34. Fischl, B., Sereno, M. I. & Dale, A. M. Cortical surface-based analysis. ii: Inflation, flattening, and a surface-based coordinate system. *Neuroimage* **9**, 195–207, DOI: 10.1006/nimg.1998.0396 (1999).

35. Dale, A. M., Fischl, B. & Sereno, M. I. Cortical surface-based analysis. i. segmentation and surface reconstruction. *Neuroimage* **9**, 179–194, DOI: 10.1006/nimg.1998.0395 (1999).

36. Fischl, B. *et al.* Whole brain segmentation: automated labeling of neuroanatomical structures in the human brain. *Neuron* **33**, 341–355, DOI: 10.1016/s0896-6273(02)00569-x (2002).

37. Hoopes, A., Mora, J. S., Dalca, A. V., Fischl, B. & Hoffmann, M. Synthstrip: skull-stripping for any brain image. *NeuroImage* **260**, 119474, DOI: 10.1016/j.neuroimage.2022.119474 (2022).

38. Smith, S. M. *et al.* Advances in functional and structural mr image analysis and implementation as fsl. *Neuroimage* **23**, S208–S219, DOI: 10.1016/j.neuroimage.2004.07.051 (2004).

39. Andersson, J. L., Skare, S. & Ashburner, J. How to correct susceptibility distortions in spin-echo echo-planar images: application to diffusion tensor imaging. *Neuroimage* **20**, 870–888, DOI: 10.1016/S1053-8119(03)00336-7 (2003).

40. Chappell, M. A., Groves, A. R., Whitcher, B. & Woolrich, M. W. Variational bayesian inference for a nonlinear forward model. *Trans. Sig. Proc.* **57**, 223–236, DOI: 10.1109/TSP.2008.2005752 (2009).

41. Greve, D. N. & Fischl, B. Accurate and robust brain image alignment using boundary-based registration. *Neuroimage* **48**, 63–72, DOI: 10.1016/j.neuroimage.2009.06.060 (2009).

42. Groves, A. R., Chappell, M. A. & Woolrich, M. W. Combined spatial and non-spatial prior for inference on mri time-series. *Neuroimage* **45**, 795–809, DOI: 10.1016/j.neuroimage.2008.12.027 (2009).

43. Chappell, M. A. *et al.* Partial volume correction of multiple inversion time arterial spin labeling mri data. *Magn Reson. Med* **65**, 1173–1183, DOI: 10.1002/mrm.22641 (2011).

44. Fonov, V., Evans, A., McKinstry, R., Almli, C. & Collins, D. Unbiased nonlinear average age-appropriate brain templates from birth to adulthood. *NeuroImage* **47**, S102, DOI: 10.1016/S1053-8119(09)70884-5 (2009).

45. Fonov, V. *et al.* Unbiased average age-appropriate atlases for pediatric studies. *Neuroimage* **54**, 313–327, DOI: 10.1016/j.neuroimage.2010.07.033 (2011).

46. Avants, B. B., Tustison, N. J., Song, G. & Gee, J. C. Ants: advanced open-source normalization tools for neuroanatomy. *Penn Image Comput. Sci. Lab.* 1–35 (2009).

47. Avants, B. B. *et al.* A reproducible evaluation of ants similarity metric performance in brain image registration. *Neuroimage* **54**, 2033–2044, DOI: 10.1016/j.neuroimage.2010.09.025 (2011).

48. Zhang, Y., Brady, M. & Smith, S. Segmentation of brain mr images through a hidden markov random field model and the expectation-maximization algorithm. *IEEE Trans Med Imaging* **20**, 45–57, DOI: 10.1109/42.906424 (2001).

49. Cox, R. W. AFNI: Software for analysis and visualization of functional magnetic resonance neuroimages. *Comput. Biomed Res* **29**, 162–173, DOI: 10.1006/cbmr.1996.0014 (1996).

50. Cox, R. W. & Hyde, J. S. Software tools for analysis and visualization of fmri data. *NMR Biomed* **10**, 171–178, DOI: 10.1002/(sici)1099-1492(199706/08)10:4/5<171::aid-nbm453>3.0.co;2-l (1997).

51. Cox, R. W., Chen, G., Glen, D. R., Reynolds, R. C. & Taylor, P. A. FMRI Clustering in AFNI: False-Positive Rates Redux. *Brain Connect.* **7**, 152–171, DOI: 10.1089/brain.2016.0475 (2017).

52. Clement, P. *et al.* Variability of physiological brain perfusion in healthy subjects - a systematic review of modifiers. considerations for multi-center asl studies. *J. Cereb. Blood Flow & Metab.* **38**, 1418–1437, DOI: 10.1177/0271678X17702156 (2018). Epub 2017 Apr 10.

53. Wu, W.-C., Fernandez-Seara, M., Detre, J. A., Wehrli, F. W. & Wang, J. A theoretical and experimental investigation of the tagging efficiency of pseudocontinuous arterial spin labeling. *Magn. Reson. Medicine* **58**, 1020–1027, DOI: 10.1002/mrm.21385 (2007).

54. Chen, Y., Wang, D. J. & Detre, J. A. Test-retest reliability of arterial spin labeling with common labeling strategies. *J. Magn. Reson. Imaging* **33**, 940–949, DOI: 10.1002/jmri.22550 (2011).

55. Zhao, L., Chang, C.-Y. E. & Alsop, D. C. Controlling t2 blurring in 3d rare arterial spin labeling acquisition through optimal combination of variable flip angles and k-space filtering. *Magn. Reson. Medicine* **80**, 1391–1401, DOI: 10.1002/mrm.27118 (2018).

56. Liang, X., Connelly, A., Tournier, J.-D. & Calamante, F. A variable flip angle-based method for reducing blurring in 3d grase asl. *Phys. Medicine Biol.* **59**, 5559–5573, DOI: 10.1088/0031-9155/59/18/5559 (2014).

57. Qin, Q. Point spread functions of the t2 decay in k-space trajectories with long echo train. *Magn. Reson. Imaging* **30**, 1134–1142, DOI: 10.1016/j.mri.2012.04.017 (2012).

58. Ivanov, D., Pfeuffer, J., Gardumi, A., Uludağ, K. & Poser, B. A. a. 2d caipirinha improves accelerated 3d grase asl. In *Proceedings of the Annual Meeting of the International Society for Magnetic Resonance in Medicine (ISMRM)*, 3630 (Honolulu, USA, 2017).

59. Boland, M. *et al.* Accelerated 3d-grase imaging improves quantitative multiple post labeling delay arterial spin labeling. *Magn. Reson. Medicine* **80**, 2475–2484, DOI: 10.1002/mrm.27226 (2018). https://onlinelibrary.wiley.com/doi/pdf/10.1002/mrm.27226.

 # Supplementary Material

 ## Impact of deblur options on the 3D-GRASE ASL data

	FFT			Lucy	
	Direct	**Lorentz**	**Lorwien**	**Direct**	**Lorentz**
Clinical	6.38 ± 0.33	6.57 ± 0.3	6.65 ± 0.33	6.55 ± 0.3	6.55 ± 0.35
Hires	2.38 ± 0.13	2.53 ± 0.12	2.43 ± 0.14	2.54 ± 0.09	2.62 ± 0.09

Table S1. Table of FWHM (in mm) estimated using AFNI's *3dFWHMx* for different deblurring methods available in *oxasl_deblur*. Numerical values presented are mean ± std. dev across participants. This evaluation was carried out on data from the 32 ch coil only.

Figure S1. Mosaic of orthogonal views of the group average perfusion-weighted (MRI signal units), relative CBF (arbitrary units) and CBF (mL/100g/min) maps for data acquired using the three respective head coils.. In each panel, the maps obtained from the clinical and hires pCASL acquisitions are displayed in the top and bottom rows, respectively.

 Tables of Summary Statistics : Clinical vs Hires pCASL

CBF (mL/100 g/min)
pCASL Clinical

Statistic	20 ch	32 ch	64 ch
Mean	72.57	68.43	67.36
Median	70.81	65.42	64.42
Standard Deviation	27.85	27.96	27.61
Inter-quartile Range	36.85	37.70	37.12
1st Quartile	53.55	48.65	47.79
3rd Quartile	90.40	86.35	84.91
Outlier Percentage	0.79%	0.80%	0.88%

CBF (mL/100 g/min)
pCASL Hires

Statistic	20 ch	32 ch	64 ch
Mean	57.90	58.30	59.56
Median	56.55	56.66	57.75
Standard Deviation	26.15	25.07	25.77
Inter-quartile Range	35.58	33.59	34.85
1st Quartile	39.42	40.83	41.44
3rd Quartile	75.00	74.42	76.28
Outlier Percentage	0.69%	0.78%	0.74%

Table S2. Summary Statistics of quantitative CBF (in mL/100 g/min) for Clinical and Hires pCASL data over all subjects in the study

Relative CBF
pCASL Clinical

Statistic	20 ch	32 ch	64 ch
Mean	960.79	1016.20	1001.84
Median	933.14	963.99	950.55
Standard Deviation	385.91	435.08	433.61
Inter-quartile Range	533.11	566.97	574.38
1st Quartile	678.91	704.31	687.94
3rd Quartile	1212.03	1271.28	1262.31
Outlier Percentage	0.81%	1.66%	1.46%

Relative CBF
pCASL Hires

Statistic	20 ch	32 ch	64 ch
Mean	944.67	1072.07	1097.49
Median	885.92	1003.30	1023.64
Standard Deviation	502.92	561.64	578.83
Inter-quartile Range	693.79	759.00	788.37
1st Quartile	568.31	656.43	666.43
3rd Quartile	1262.09	1415.43	1454.80
Outlier Percentage	0.99%	1.28%	1.18%

Table S3. Summary Statistics of relative CBF for Clinical and Hires pCASL data over all subjects in the study

Perfusion weighting (a. u.) pCASL Clinical			
Statistic	20 ch	32 ch	64 ch
Mean	575.50	608.55	600.29
Median	557.94	576.44	568.68
Standard Deviation	226.10	256.24	256.66
Inter-quartile Range	312.50	334.30	338.79
1st Quartile	409.56	424.01	414.27
3rd Quartile	722.06	758.30	753.05
Outlier Percentage	0.84%	1.69%	1.51%

Perfusion weighting (a. u.) pCASL Hires			
Statistic	20 ch	32 ch	64 ch
Mean	534.89	605.91	620.15
Median	500.93	565.98	577.39
Standard Deviation	280.25	313.30	322.90
Inter-quartile Range	386.34	423.67	440.03
1st Quartile	324.63	373.16	378.73
3rd Quartile	710.97	796.82	818.77
Outlier Percentage	1.03%	1.31%	1.21%

Table S4. Summary Statistics of Perfusion-weighting (in arbitrary units) for Clinical and Hires pCASL data over all subjects in the study

Perfusion tSNR pCASL Clinical			
Statistic	20 ch	32 ch	64 ch
Mean	5.66	5.77	5.92
Median	5.43	5.51	5.64
Standard Deviation	2.35	2.45	2.53
Inter-quartile Range	3.06	3.17	3.23
1st Quartile	3.99	4.03	4.14
3rd Quartile	7.06	7.20	7.37
Outlier Percentage	1.50%	1.67%	1.83%

Perfusion tSNR pCASL Hires			
Statistic	20 ch	32 ch	64 ch
Mean	1.59	2.14	2.27
Median	1.45	1.85	1.97
Standard Deviation	0.92	1.39	1.44
Inter-quartile Range	1.16	1.63	1.74
1st Quartile	0.93	1.16	1.24
3rd Quartile	2.09	2.79	2.98
Outlier Percentage	2.29%	3.55%	3.20%

Table S5. Summary Statistics of perfusion tSNR for Clinical and Hires pCASL data over all subjects in the study

Perfusion tSNR pCASL Clinical			
Statistic	20 ch	32 ch	64 ch
Mean	5.66	5.77	5.92
Median	5.43	5.51	5.64
Standard Deviation	2.35	2.45	2.53
Inter-quartile Range	3.06	3.17	3.23
1st Quartile	3.99	4.03	4.14
3rd Quartile	7.06	7.20	7.37
Outlier Percentage	1.50%	1.67%	1.83%

Perfusion tSNR (volume-scaled) pCASL Hires			
Statistic	20 ch	32 ch	64 ch
Mean	9.19	12.35	13.11
Median	8.40	10.66	11.40
Standard Deviation	5.30	8.03	8.34
Inter-quartile Range	6.70	9.42	10.03
1st Quartile	5.38	6.70	7.18
3rd Quartile	12.08	16.12	17.21
Outlier Percentage	2.29%	3.55%	3.20%

Table S6. Summary Statistics of perfusion tSNR after scaling the pCASL hires data in Table S7 by the ratio of voxel-volumes.

 Comparison of Hires pCASL and PASL protocols

Figure S2. Mosaic of orthogonal views of the group average perfusion-weighted (MRI signal units), relative CBF (arbitrary units) and CBF (mL/100g/min) maps for data acquired using the three respective head coils. In each panel, the maps obtained from the hires pCASL and PASL acquisitions are displayed in the top and bottom rows, respectively.

Figure S3. Orthogonal views of the group average perfusion tSNR maps for the hires acquisitions with pCASL and PASL labelling schemes.

CBF (mL/100 g/min) pCASL Hires			
Statistic	20 ch	32 ch	64 ch
Mean	57.90	58.30	59.56
Median	56.55	56.66	57.75
Standard Deviation	26.15	25.07	25.77
Inter-quartile Range	35.58	33.59	34.85
1st Quartile	39.42	40.83	41.44
3rd Quartile	75.00	74.42	76.28
Outlier Percentage	0.69%	0.78%	0.74%

CBF (mL/100 g/min) PASL Hires			
Statistic	20 ch	32 ch	64 ch
Mean	42.06	42.12	42.37
Median	39.65	39.82	40.20
Standard Deviation	25.94	25.71	25.80
Inter-quartile Range	37.10	36.93	37.09
1st Quartile	21.86	22.11	22.30
3rd Quartile	58.96	59.04	59.39
Outlier Percentage	0.77%	0.72%	0.71%

Table S7. Summary Statistics of quantitative CBF (in mL/100 g/min) for pCASL Hires and PASL Hires data over all subjects in the study

Relative CBF pCASL Hires			
Statistic	20 ch	32 ch	64 ch
Mean	944.67	1072.07	1097.49
Median	885.92	1003.30	1023.64
Standard Deviation	502.92	561.64	578.83
Inter-quartile Range	693.79	759.00	788.37
1st Quartile	568.31	656.43	666.43
3rd Quartile	1262.09	1415.43	1454.80
Outlier Percentage	0.99%	1.28%	1.18%

Relative CBF PASL Hires			
Statistic	20 ch	32 ch	64 ch
Mean	988.63	1114.79	1126.50
Median	863.30	970.60	984.09
Standard Deviation	701.69	793.27	801.86
Inter-quartile Range	978.52	1103.48	1121.65
1st Quartile	435.20	489.73	492.45
3rd Quartile	1413.71	1593.21	1614.10
Outlier Percentage	1.32%	1.38%	1.30%

Table S8. Summary Statistics of relative CBF for pCASL Hires and PASL Hires data over all subjects in the study

Perfusion weighting (a. u.) pCASL Hires			
Statistic	20 ch	32 ch	64 ch
Mean	534.89	605.91	620.15
Median	500.93	565.98	577.39
Standard Deviation	280.25	313.30	322.90
Inter-quartile Range	386.34	423.67	440.03
1st Quartile	324.63	373.16	378.73
3rd Quartile	710.97	796.82	818.77
Outlier Percentage	1.03%	1.31%	1.21%

Perfusion weighting (a. u.) PASL Hires			
Statistic	20 ch	32 ch	64 ch
Mean	420.22	471.74	476.20
Median	369.42	412.66	417.72
Standard Deviation	286.37	324.00	327.51
Inter-quartile Range	396.12	447.66	454.97
1st Quartile	196.43	218.29	219.09
3rd Quartile	592.56	665.95	674.06
Outlier Percentage	1.38%	1.44%	1.36%

Table S9. Summary Statistics of Perfusion-weighting (in arbitrary units) for pCASL Hires and PASL Hires data over all subjects in the study

Perfusion tSNR pCASL Hires			
Statistic	20 ch	32 ch	64 ch
Mean	1.59	2.14	2.27
Median	1.45	1.85	1.97
Standard Deviation	0.92	1.39	1.44
Inter-quartile Range	1.16	1.63	1.74
1st Quartile	0.93	1.16	1.24
3rd Quartile	2.09	2.79	2.98
Outlier Percentage	2.29%	3.55%	3.20%

Perfusion tSNR PASL Hires			
Statistic	20 ch	32 ch	64 ch
Mean	1.26	1.74	1.84
Median	1.09	1.41	1.51
Standard Deviation	0.90	1.38	1.42
Inter-quartile Range	1.15	1.62	1.73
1st Quartile	0.58	0.74	0.79
3rd Quartile	1.73	2.36	2.52
Outlier Percentage	2.47%	3.81%	3.30%

Table S10. Summary Statistics of perfusion tSNR for pCASL Hires and PASL Hires data over all subjects in the study

Perfusion tSNR (volume-scaled)
pCASL Hires

Statistic	20 ch	32 ch	64 ch
Mean	9.19	12.35	13.11
Median	8.40	10.66	11.40
Standard Deviation	5.30	8.03	8.34
Inter-quartile Range	6.70	9.42	10.03
1st Quartile	5.38	6.70	7.18
3rd Quartile	12.08	16.12	17.21
Outlier Percentage	2.29%	3.55%	3.20%

Perfusion tSNR (volume-scaled)
PASL Hires

Statistic	20 ch	32 ch	64 ch
Mean	7.28	10.06	10.63
Median	6.28	8.12	8.72
Standard Deviation	5.18	8.00	8.24
Inter-quartile Range	6.63	9.39	10.02
1st Quartile	3.37	4.25	4.54
3rd Quartile	10.00	13.64	14.56
Outlier Percentage	2.47%	3.81%	3.30%

Table S11. Summary Statistics of perfusion tSNR after scaling the pCASL hires data in Table S7 by the ratio of voxel-volumes.

 Single participant maps for all three ASL protocols

Figure S4. Mosaic of orthogonal views of a single participant's perfusion-weighted (MRI signal units), relative CBF (arbitrary units) and CBF (mL/100g/min) maps for data acquired using the three respective head coils. In each panel, the maps obtained from the clinical pCASL, hires pCASL and hires PASL acquisitions are displayed in the top, middle and bottom rows, respectively.