Volumetric trans-scale imaging of massive quantity of heterogeneous cell populations in centimeter-wide tissue and embryo

Authors

Taro Ichimura,[1]* Taishi Kakizuka,[1,2] Yoshitsugu Taniguchi,[3] Satoshi Ejima,[3] Yuki Sato,[4]

Keiko Itano,[2] Kaoru Seiriki,[5] Hitoshi Hashimoto,[1,5] Ko Sugawara,[6] Hiroya Itoga,[6] Shuichi

Onami,[1,6] and Takeharu Nagai[1,2,7]*

Affiliations

[1]Transdimensional Life Imaging Division, Institute for Open and Transdisciplinary Research

Initiatives, Osaka University, Yamadaoka 2-1, Suita, Osaka 565-0871, Japan.

[2]Department of Biomolecular Science and Engineering, SANKEN, Osaka University,

Mihogaoka 8-1, Ibaraki, Osaka 567-0047, Japan.

[3]SIGMAKOKI CO., LTD., 1-19-9 Midori, Sumida-ku, Tokyo 130-0021, Japan.

[4]Department of Anatomy and Cell Biology, Graduate School of Medical Sciences, Kyushu

University, Maidashi 3-1-1, Higashi-ku, Fukuoka 812-8582, Japan.

[5]Laboratory of Molecular Neuropharmacology, Graduate School of Pharmaceutical Sciences,

Osaka University, Yamadaoka 1-6, Suita, Osaka 565-0871, Japan.

[6]Laboratory for Developmental Dynamics, RIKEN Center for Biosystems Dynamics

Research, Minatojima-minamimachi 2-2-3, Chuo-ku, Kobe, Hyogo 650-0047, Japan.

[7]Research Institute for Electronic Science, Hokkaido University, Kita 20, Nishi 10, Kita-ku,

Sapporo, Hokkaido 001-0020, Japan.

*Email address: ichimura@otri.osaka-u.ac.jp, ng1@sanken.osaka-u.ac.jp

Abstract

We established a volumetric trans-scale imaging system with an ultra-large field-of-view

(FOV) that enables simultaneous observation of millions of cellular dynamics in centimeter-

wide three-dimensional (3D) tissues and embryos. Using a custom-made giant lens system

with a magnification of 2× and a numerical aperture (NA) of 0.25, and a CMOS camera with

more than 100 megapixels, we built a trans-scale scope AMATERAS-2, and realized

fluorescence imaging with a transverse spatial resolution of approximately 1.1 μm across an

FOV of approximately 1.5×1.0 cm^2. The 3D resolving capability was realized through a

combination of optical and computational sectioning techniques tailored for our low-power

imaging system. We applied the imaging technique to 1.2 cm-wide section of mouse brain,

and successfully observed various regions of the brain with sub-cellular resolution in a single

FOV. We also performed time-lapse imaging of a 1-cm-wide vascular network during quail

embryo development for over 24 hours, visualizing the movement of over 4.0×10^5 vascular

endothelial cells and quantitatively analyzing their dynamics. Our results demonstrate the

potential of this technique in accelerating production of comprehensive reference maps of all

cells in organisms and tissues, which contributes to understanding developmental processes,

brain functions, and pathogenesis of disease, as well as high-throughput quality check of

tissues used for transplantation medicine.

MAIN TEXT

Introduction

Recently, life sciences have focused on comprehending the working principles of multicellular

systems, spanning from basic biology to medical applications (1–5). Multicellular systems are

generally complex, comprising heterogeneous cells rather than homogeneous assemblies. To

understand the mechanism by which numerous cells cooperate to express the function of the

entire system, it is ideal to observe all individual components within the system simultaneously

during the time span of a whole phenomenon. In particular, in scenarios where a small fraction

of cells in a large multicellular system may influence the fate of the population, or when distant

cells and tissues operate synchronously, deducing the entire system from observations of sub-

populations becomes challenging. Therefore, the trans-scale measurement of collective

dynamics of all individual cells that constitute the system of interest is desired.

In response to this demand, researchers have recently reported the development of imaging

methods with large field-of-view (FOV) (6–11). Whereas most of these methods are focused

on observing brain activity in neuroscience using two-photon excited fluorescence (7–10), our efforts have been dedicated to developing techniques as versatile tools for studying multicellular systems science and developmental biology. In our previous work, we proposed a fluorescence imaging system for the visible wavelength region, capable of spatially resolving individual cells within a centimeter FOV, which we named AMATERAS-1 (A Multi-scale/modal Analytical Tool for Every Rare Activity in Singularity), enabling simultaneous observation of dynamics in the range of 10^5–10^6 cells (11). Additionally, we successfully detected less than 1% of cells with unique roles in multicellular systems and revealed how these rare cells trigger a drastic transition from unicellular to multicellular behavior in *Dictyostelium discoideum* (12). However, the imaging system had a numerical aperture (NA) of 0.12, which is insufficient for realizing volumetric observation owing to its broad depth of field (DOF). Overcoming this limitation and enabling volumetric imaging for diverse three-dimensional (3D) tissues and small organisms has been a significant challenge.

In this study, we developed AMATERAS-2, a volumetric optical imaging system with approximately 15×10 mm^2 FOV, equivalent to that of AMATERAS-1 (11). The key component for the enhancement is the giant 2× lens system with an NA of 0.25, which offers improved 3D spatial resolution and higher sensitivity. By incorporating novel methodologies of optical sectioning and computational sectioning, we successfully added volumetric imaging capabilities. The effectiveness of these advancements was demonstrated in imaging 1.5-mm-thick brain section blocks and conducting a 25-h time-lapse observation of vascular endothelial cells during quail embryogenesis, allowing us to trace the spatiotemporal dynamics of approximately 4.0×10^5 cells in three dimensions.

RESULT

Optical configuration and performance

AMATERAS-1 utilized a telecentric macro lens with 2× magnification, thus allowing dynamic observation in the centimeter-FOV with subcellular resolution (11). Despite its efficacy in biological studies, its low NA (0.12) limits both its spatial resolution in the *z*-direction, and system brightness. To overcome these constraints and expand the possibilities for observing cell

dynamics in tissues and embryos, we developed a giant lens system with a higher NA, consisting of a pair of objective and tube lenses. The objective lens is composed of 12 optical elements in 7 groups, while the tube lens is composed of 9 lenses in 6 groups (Fig. 1A). Both lenses are infinity-corrected with NAs of 0.25 for the objective and 0.125 for the tube lens, and have a field number of 44 mm (diameter) suitable for large-area image sensors. They are effectively aberration-corrected to the diffraction limit or lower in the visible wavelength range of 436–656 nm. The objective lens and tube lens have sizes of 144 mm × 310 mm and 170 mm × 345 mm, respectively (Fig. 1B). The objective and tube lenses exhibit zero vignetting, enabling observation of peripheral areas of the FOV with high brightness.

As the NA of AMATERAS-2 is 0.25, more than double that of AMATERAS-1 (0.12), our system in principle achieves over twice the spatial resolution in the transverse (xy) direction, more than four times the resolution in the longitudinal (z) direction, and over four times the brightness of AMATERAS-1.

Using this new lens system, we constructed a wide-field epi-illumination fluorescence imaging system, termed AMATERAS-2w, with a magnification of 2×, an NA of 0.25, and a field number of 44 mm (Fig. 1C, Materials and Methods). We utilized either a 120-megapixel CMOS camera or a 250-megapixel CMOS camera, depending on the research purpose. Both the cameras have an approximately 35 mm diagonal, thus providing an observation FOV of 17.8 mm and 17.4 mm diagonal at 2× magnification, respectively. The pixel sizes for these cameras are 2.2 and 1.5 μm, with sampling intervals of 1.1 and 0.75 μm at 2× magnification, respectively.

We installed a high-brightness LED light sources for fluorescence excitation, directing it into the objective lens through a custom-made long-pass dichroic mirror designed for GFP imaging with a cut-off wavelength of 497 nm. To ensure uniform illumination across the entire FOV, we employed a lens-let array pair. Additionally, we incorporated a 2-inch band-pass fluorescence filter between the tube lens and the camera.

We experimentally obtained the optical point spread function (PSF) using green fluorescent beads (center wavelength: 515 nm) with a 0.2-μm diameter, employing both the 120-megapixel

CMOS camera (Fig. 2A) and the 250-megapixel CMOS camera (Fig. 2B). The left panels in Figs. 2A and B display the PSFs in the xy and xz planes, respectively. The single-point spatial resolution was evaluated by the full-width of half-maximum (FWHM) of the line profiles determined through Gaussian-function fit. The transverse resolutions (xy-direction) obtained with the 120- and 250-megapixel cameras were 1.24 ± 0.12 ($N = 100$) and 1.12 ± 0.072 ($N = 100$) (Figs. 2A-B, right top), both of which are slightly downgraded from the theoretical FWHM, 1.05 μm, given by $0.51 \lambda_{em}/NA$ ($\lambda_{em} = 515$ nm, $NA = 0.25$) (13,14). This can be attributed to the coarse sampling interval (1.1 μm and 0.75 μm, respectively). In particular, in the 120-megapixel case, the apparent single-point resolution varies depending on the relative position of the bead within the pixel (Figs. S1A-B). The coarse sampling also introduces uncertainty in the estimated position of the bead's center coordinate (Fig. S1E). Although this may pose some challenges in quantitatively evaluating fine shapes in cells, it is not a practical issue in imaging at cell resolution. By contrast, the 250-megapixel camera provides a nearly constant resolution (Figs. S1C-D). Thus, the 250-megapixel camera is suitable when high-resolution images are required, whereas the 120-megapixel camera is suitable for other cases where increasing the number of photons per pixel or suppressing data size is necessary.

In the depth direction (z-direction), we evaluated the DOF by fitting Gaussian curves to the line profiles (Figs. 2A-B, right bottom). The DOF was measured to be 15.6 ± 1.3 ($N = 100$) and 14.8 ± 1.1 ($N = 100$) for the 120- and 250-megapixel cameras, respectively. This difference in DOF is attributed to the distinct pixel sizes of the cameras (15). The DOF value for the 250-megapixel camera was found close to the theoretical FWHM, 14.3 μm, given by $1.78 \lambda_{em}/NA^2$ ($\lambda_{em} = 515$ nm, $NA = 0.25$) (14).

The spatial resolution in the transverse direction has been significantly improved, approximately two times compared with that of AMATERAS-1. Figures 2C and D display fluorescence images of cardiomyocytes derived from human induced pluripotent stem cells (hiPSCs) stained with Rhodamine phalloidin (C) across the entire FOV and (D) within a central local region indicated by a magenta square in (C). These images were captured using the 250-megapixel camera, and the FOV size was 14.7×9.4 mm^2. For comparison, Fig. 2E shows an

image of the same area observed with AMATERAS-1. Evidently, AMATERAS-2w delivers

more detailed images with finer spatial structures. This is further confirmed by the line profiles

depicted in Fig. 2F, where the filamentous actin stained with Rhodamine exhibits a sharper

spatial resolution with AMATERAS-2w. This result indicated the capability of observing

subcellular structures, such as myofibrils, in cell sheets with a large area, such as artificial

myocardial sheets, which would enable us to simultaneously investigate microscale structures

and macroscale multicellular dynamics.

Spinning pinhole-array disk provides optical sectioning capability in wide-FOV

This imaging system utilizes wide-field illumination and detection, which currently lacks the

capability to selectively acquire images of specific z-planes for volumetric imaging. To extend

the system's applicability, we developed two methods to enable the volumetric imaging, namely,

optical sectioning and computational sectioning.

For optical sectioning, various options of optical techniques exist for achieving 3D fluorescence

imaging with a large FOV in the visible wavelength region. These include scanning confocal

microscopy (6,13), light-sheet microscopy (4,16), light-field microscopy (17), two-photon

excitation microscopy (7-10,18), and spatiotemporal focusing (19). Among these options, we

selected the confocal imaging method for this study. This is because the light-sheet microscopy

struggles with uniform illumination over the centimeter scale FOV of AMATERAS (16,20),

and the light-field microscopy involves a trade-off between spatial resolution and the ability to

resolve three dimensions, which is incompatible with the AMATERAS's concept (21). In

addition, two-photon excitation microscopy, including that using spatiotemporal focusing, is

crucial for deep tissue imaging, but still requires further innovations in pulsed laser power and

scanning methods to be effectively applied to the wide FOV of AMATERAS.

We considered the confocal method to be the most compatible with our imaging system.

Considering the vast FOV, we adopted the multipoint-scanning confocal system. The

commonly used method involving a combination of a microlens array and a pinhole array disk

(22) could not be employed in AMATERAS owing to the large NA of the tube lens (NA =

0.125) compared with that of standard microscopes. Matching the NA requires shortening the

distance between the two disks or enlarging the microlenses. However, shortening the inter-disk distance leads to a loss of space for inserting a dichroic mirror, and enlarging the microlenses reduces the number of foci. While confocal reflection (non-fluorescence) microscopy does not require a beam splitter, allowing for close microlens-pinhole spacing (23), the confocal fluorescence microscopy requires a beam splitter due to the essential separation of fluorescence and excitation light. Therefore, designing microlens-pinhole arrays suitable for AMATERAS-2 presents a significant challenge. This challenge can be circumvented by using an optical configuration where excitation light and fluorescence pass through synchronized pinhole arrays on separate paths, as initially proposed in the early development of multipoint confocal microscopy (24). However, this solution would require duplicating the optics of the giant lens, which is currently unfeasible. Moreover, even if possible, the added complexity of the system would render this approach impractical.

To address this challenge, we opted to use pinhole arrays without microlenses, which is rather a more traditional configuration of multipoint confocal microscopy (13,25,26). Although this method is less light-efficient compared with the one using microlenses, it aligns well with our imaging system's high NA and low magnification. Here, we designed the pinhole array disk for fluorescence imaging in green (*e.g.*, GFP, YFP) and red (*e.g.* RFP, mCherry) wavelengths. Commercial confocal systems using pinhole-array disks typically feature large pinhole sizes (*e.g.*, 50 μm) and wide spacing between pinholes (*e.g.*, 200 μm), designed for general microscopes. For AMATERAS-2, however, with its 2× magnification and NA of 0.25, the pinhole size must be much smaller. The diameters of the Airy disk for green and red wavelengths are about 5 μm and 6 μm, respectively, at the focal plane of the tube lens ($NA =$ 0.125). As a prototype, we set the pinhole size to 6 μm (Fig. 3A), matching the Airy disk diameter for red wavelength. A more detailed discussion on the impact of pinhole diameter dependence on 3D resolution is available in the Supplementary File (Note 1, Fig. S2). The pinhole spacing was set as 24 μm, striking a balance between throughput and crosstalk (Fig. 3A). To implement this, we custom-made a pinhole-array disk and incorporated it into a rotary machine (CrestOptics, Rome, Italy).

We positioned the pinhole array disk at the image plane of the tube lens and constructed a relay

lens system to transfer the plane to the camera. This particular configuration, termed

AMATERAS-2c, is illustrated in Fig. 3B (Materials and Methods). To split the light paths of

fluorescence excitation light, we placed a short-pass dichroic mirror beneath the disk. For

fluorescence excitation, we used a high-brightness LED with a center wavelength of 470 nm,

the same as in the wide-field imaging system (Fig. 1C). The relay lens consisted of a telecentric

macro lens with 1× magnification, and a 2-inch band-pass fluorescent filter was positioned at

the entrance of the macro lens. If necessary, the relay lens can be replaced by a 2× magnification

lens to switch the total magnification from 2× to 4×, based on resolution requirements.

Fluorescence images of 0.5 μm fluorescent beads were observed with and without the pinhole-

array disk, where fluorescent beads were three-dimensionally dispersed in an agarose gel slab

(Fig. 3C). The presence of the pinhole-array disk dramatically suppressed background light

from outside the focal plane. This verified that optical sectioning can be achieved by the

pinhole-array disk. The variations in the 3D PSF with and without the pinhole-array disk are

detailed in the Supplementary File (Fig. S3). Typical 3D PSF was obtained with the

fluorescence beads with the relay lens of (D) 1× and (E) 2× magnifications (Figs. 3D and E).

As NA of the 1× relay lens (NA = 0.079) is lower than that of the tube lens (NA = 0.125), the

transverse spatial resolution (Fig. 3D, right top) is degraded compared to the one obtained by

the wide-field configuration without the relay lens (Figs. 2A and B, Note 1). By contrast, the

2× relay lens (NA = 0.12) almost matches with the tube lens, and the reduction in the transverse

spatial resolution is relatively small (Fig. 3E). As for the longitudinal resolution, the FWHM of

the PSF was approximately 16 μm with a 1× relay lens and approximately 14 μm with a 2×

relay lens, both values significantly exceeding the ideal value about 9.5 μm (Note 1). These

discrepancies are attributed to the relatively loose focusing of the excitation light and the non-

negligible size of the pinhole.

Computational sectioning provides pseudo-depth-resolved imaging capability

In addition to optical sectioning, we developed a computational sectioning method to separate

images of the focal plane and out-of-focal planes. During measurement, a *z*-stack is acquired in

the wide-field imaging configuration (Fig. 1C) by scanning the focal position in the z-direction and capturing images at each step. The raw image in a 3D volume results from the superposition of in-focus and out-of-focus plane images. We estimated the contribution from out-of-focus planes as the baseline by iteratively applying low-pass filtering under the assumption that the image's spatial frequency in the focal plane is higher than that in out-of-focus planes. The cut-off frequency of the low-pass filter was optimized using the principles of independent component analysis (Materials and Methods, Note 2, Fig. S4-S5).

To demonstrate computational sectioning, we observed myocardial organoids derived from hiPSCs (27). This 3D dome structure of a cavity chamber is extensively studied in developmental biology and regenerative medicine as a model for human myocardial tissue development (28,29). We fabricated a large area of organoids across the entire FOV, chemically fixed and immunostained them for cardiac troponin T (cTnT) with Alexa488, and stained the nuclei with Hoechst33342 (Materials and Methods). The z-stack was obtained within a z-range involving the organoids (approximately 250 µm) with 4-µm intervals. Figures 4A and B display two-color overlaid images of the island-formed organoids across the entire FOV, without and with computational sectioning, respectively (both are maximum-intensity projection (MIP) images of the z-stacks). Figures 4C and D present images of six layers in the z-direction of the yellow square region of Figs. 4A and B, before and after computational sectioning, respectively. After the sectioning, the distribution of cTnT and nuclei in each z-layer is clearly visualized (Fig. 4D): The island areas are composed of multilayered cells, and the inter-island spaces are covered with a single layer of cells, as shown in Fig. 4D, $z = 20$ µm. A 3D isosurface representation (Fig. 4E) shows that the hollow oval structure is formed by a thin layer of cTnT-positive cells (cardiomyocytes) whereas the underlying cell layer is composed of cTnT-negative cells, consistent with previous literature (27). This demonstration validates the technique for wide-field imaging. Video files of Figs. 4C–E are available in the Supplementary files (Movies S1 and S2).

In principle, the signal component can be extracted as long as the signal intensity (in-focus component) is significantly higher than the statistical noise of the background intensity (out-of-

255 focus component). This method is effective in cellular imaging when fluorescent molecules are

256 localized within the cell, like in the nucleus, or when they exhibit a filament-like structure,

257 thereby enabling cell recognition and dynamic tracking. It relies on the spatial frequency of the

258 fluorescence image in the focal plane being uniform and clearly higher than that outside the

259 focal plane.

Observing a 1.2-cm wide and 1.5-mm thick volume of mouse brain section

To demonstrate the potential for brain imaging, we performed imaging of a 1.5-mm thick mouse

brain section in the coronal plane (Fig. 5A). The brain was chemically cleared using CUBIC

(30,31), and the cell nuclei were stained with SYTOX-Green (Materials and Methods). The 1×

relay lens was primarily employed (total magnification = 2×) to cover the coronal plane in the

FOV. Each layer's exposure time was 1 s, and a total of 378 layers were acquired to cover the

entire 1.5-mm-thick volume in steps of 4 μm in the z-direction. Optical and computational

sectioning were both applied. Figure 5B shows a fluorescence image at a single z-layer ($z = 700$

μm), revealing the 12 mm × 8 mm wide section with single cell resolution. Figure 5C displays

enlarged views of the xy-plane (coronal plane) in the dotted square region in Fig. 5B, along with

xz (transverse plane) and yz (sagittal plane) cross sections, essentially demonstrating successful

z-direction sectioning. Figure 5D shows a 3D representation of the same volume of raw data

obtained by optical sectioning alone, prior to computational sectioning. However, owing to

intense background light overlapping, individual cells are challenging to distinguish,

particularly in regions of high cell density. By utilizing computational sectioning to eliminate

background light, clear images were obtained even in areas with strong background light.

Supplementary files contain videos of the z-scan for the entire FOV and local regions (Movies

S3 and S4).

Figure 5C visualizes the characteristic 3D structures of several regions, including hippocampal

dentate gyrus, medial habenula, and choroid plexus, which are known to be associated with

memory formation (32), depression (33), and regulation of cerebrospinal fluid (34), respectively.

Close-up views of the hippocampal region (indicated with the dashed square in Fig. 5B) in the

xy and xz planes are shown in Figs. 5E and F, respectively. A comparison of the images at 2×

and 4× magnification in Figs. 5E and F reveals that the higher magnification (4×) allows for

improved spatial resolution and separation of individual cells in both the *xy* and *xz* planes.

For cell detection in the 3D imaging data of the mouse brain section, we used a recently

developed interactive platform for cell detection and tracking, called ELEPHANT (Efficient

LEarning using sParse Human Annotations for Nuclear Tracking) (35). This platform

seamlessly integrates manual annotation with deep learning and allows for result proofreading

(Materials and Methods). Using this advanced image analysis technique, cell nuclei were

detected within the 3D volume. Figure 5G illustrates the cell detection results on specific *xy*

and *xz* planes within the cortex. As the cell density in the cortex is relatively low, the ellipsoid-

shaped nuclei image does not overlap with neighboring cells, resulting in the successful

detection of nearly all cells (as clearly shown in the *z*-scan video, Movie S5). The precision and

recall of the cell detection were estimated to be 99.4 % and 97.6 %, respectively (Materials and

Methods). Based on these results, the cell density in the cortex was calculated to be 1.18×10^5

cells/mm^3, which closely matches previous reports (36). This technique was also applied to two

other brain regions, the medial habenula and choroid plexus (Figs 5H-I). In these regions, cell

density is higher than in the cortex and the signal-to-noise ratio is lower due to strong

background intensity, making it difficult for traditional methods to detect cells accurately.

However, with the assistance of ELEPHANT, we successfully detected the majority of cells.

Dynamics of over a 4.0×10^5 cells were captured over 24-hours development of a quail

embryo

We employed the imaging methods described for a time-lapse observation of cell migration

during quail development. The non-confocal optical configuration (AMATERAS-2w, Fig. 1C)

and utilizing computational sectioning alone were sufficient for this 250-μm thick sample. We

used a tie1:H2B-eYFP transgenic quail embryo (37) in which enhanced yellow fluorescent

protein (eYFP) visualizes the nuclei of vascular endothelial cells (Materials and Methods). The

embryo was cultured *ex ovo* on a 35 mm glass-bottom dish (Fig. 6A). Time-lapse fluorescence

imaging began at 36 h (HH10) after the start of egg incubation and captured eYFP signals in

the developing embryo over 24 h. For 3D imaging, we acquired *z*-stacks every 7.5 min, with

each z-stack containing 21 layers spaced 12 µm apart. This resulted in a total of 4200 image

layers and 200 time-points, with a data size of approximately 500 GB. During the time-lapse

observation, we stabilized the relative distance between the lens and the sample using a self-

developed autofocus system (Materials and Methods).

Representative images of cell distribution based on the nuclear eYFP signals at four time-points

(t = 0, 8, 16, and 24 h) are shown in Figs. 6B and C. These images were reconstructed using the

HSB color model to represent the 3D distribution after the baseline was removed by

computational sectioning (Fig. S6). The brightness indicates the intensity of the MIP image in

the z-direction, whereas the hue represents the z-position with the maximum value in the MIP

process (Materials and Methods). As observed in Figs. 6B and C, the cell nuclei distribution

undergoes significant changes over time, thus resulting in the formation of organs such as the

ventral aortae, heart, and dorsal aortae (t = 24 h, Fig. 6C). The heart region appears blurred

owing to its oscillating shape caused by beating. Supplementary files include videos of the

HSB-color images in the entire FOV and magnified local regions (Movies S6–S8). Figure 6D

shows enlarged views of the ventral aortae (indicated by the left dashed square in Fig. 6C) at

the four time-points. Additionally, a 3D isosurface of the z-stack of the dorsal aorta region (the

right dashed square in Fig. 6C) was calculated (Fig. 6E, and Movie S9), clearly showing the

developmental process of the two tube structures of the dorsal aortae (38). Note that the upper

vessel wall is partially missing in the second half of the movie because the fluorescent image

was disturbed by the blood flow that began around t = 15 h.

At the current spatiotemporal resolution, we successfully traced the movement of individual

cells. Segmentation of cell nuclei allowed us to detect approximately 4.5×10^6 cells across all

200 frames, and by linking cells in close proximity over consecutive frames, we tracked about

4.0×10^5 cells. Figures 6F and G show cellular movement along with a feature parameter related

to cell movement at t = 8 h and 16 h. Both left panels are magnified views of the square regions

in the right panels, with trajectories drawn on MIP images. By analyzing the cell segmentation

and tracking results, various feature parameters related to cell nucleus morphology and

dynamics were computed, including nuclear size, brightness, aspect ratio, velocity, acceleration,

mean square displacement, and meandering index (39,40). Here, we used the meandering index,

a measure representing the straightness of cellular movement, to color the trajectories in Figs.

6F and G with the rainbow color table. We found that the distribution of high meandering index

at t = 8 h and 16 h was significantly changed. Further detailed analysis and biological

interpretation employing the cell-tracking results will be discussed in a future paper.

344 Discussion

In this study, we presented AMATERAS-2w, a large FOV fluorescence imaging system, along

with its variant AMATERAS-2c. The system is built around a newly designed giant lens with

an NA of 0.25, magnification of 2×, and field number of 44 mm. This innovative lens

configuration allows for imaging with an impressive spatial resolution of approximately 1.1 μm

within a centimeter-scale FOV (15 × 10 mm^2). Notably, this FOV-to-spatial resolution ratio

surpasses those of previously reported large-FOV imaging methods. The key metric "optical

invariant", derived from the product of NA and FOV, has a value of 2.75 (Materials and

Methods, 41), thus indicating excellent performance of the lens system. Furthermore, the

"space-bandwidth product", which signifies the FOV to spatial resolution ratio achieved under

real measurement conditions (considering image sensor size and wavelength), reaches 4.3×10^8

(Materials and Methods, 42). Both of these metrics significantly outperform those reported in

earlier studies on large FOV imaging methods (6–10,41). Although these metrics are essential

indicators of our system's performance, they do not solely determine its superiority over other

large-FOV imaging systems. Variations in biological targets and additional optical performance

measures (such as frame rate and imaging depth) must also be considered. Our primary focus

in designing this system was to expand the FOV, *i.e.*, maximize the number of observable cells

at any instant, even at the cost of intracellular spatial resolution. Thus, in this aspect, our system

excels as an instrument for such purposes.

The spatial resolution (FWHM) of AMATERAS-2w and -2c was measured to be approximately

1.1–1.8 μm in the transverse direction (*xy*) and 13–16 μm in the longitudinal direction (*z*)

through PSF measurements (Figs. 2A, 2B, 3D, and 3E). These values represent a significant

enhancement compared with the previous AMATERAS version (2.3 and 64 μm, respectively)

(11). The improved transverse resolution enables clearer visualization of intracellular

filamentous structures (Fig. 2F). Moreover, the refinement in longitudinal resolution was more

than four times greater, thus demonstrating that the expected effect was achieved with more

than double NA. Depending on the object of observation, it may be debatable whether this

anisotropic spatial resolution elongated along the optical axis qualifies "cellular resolution".

Although this may be insufficient to resolve dense cell populations, it proves adequate for cells

with only stained nuclei or those sparsely distributed, as observed in Figs. 5F-I, and 6E. Using

modern cell detection methods compatible with 3D big data, such as ELEPHANT employed in

this study (35), we can achieve accurate cell detection, counting and tracking to some extent,

even without clear spatial separation in three dimensions. It goes without saying that the higher

spatial resolution in raw image data would enhance the performance of subsequent image

analysis. The next step in optical development involves further enhancing the NA of the lens

system, including the relay lens, to further improve volumetric resolution and expand the range

of observable biological objects.

We successfully achieved volumetric imaging of a 1.5-mm-thick mouse brain coronal section

using the multipoint-scanning confocal imaging method, along with computational sectioning

and tissue-clearing (Fig. 5). This thickness corresponds to about one-tenth of the whole brain.

The exposure time per image was 1 s, and data acquisition took less than 10 min. The ultimate

goal is to apply this approach to enable whole-brain imaging. Unlike previous methods such as

light-sheet imaging (30) and block-face serial microscopy tomography (43), which require half

a day or more for whole-brain data acquisition owing to tiling images from a narrow FOV, our

method can significantly reduce the acquisition time when extended to the whole brain. This

feature is effective not only for the brain imaging but also for imaging of other organs and whole

body of an organism. We also plan to apply it to live imaging of highly transparent tissues such

as zebrafish (44) and organoids (45). Nevertheless, as the thickness increases, challenges such

as increased background light and larger spherical aberration arise, thus resulting in weaker

focal plane images. To address these issues and realize whole-brain imaging, we propose the

incorporation of a mechanism for compensating spherical aberrations in the future. Additionally,

we aim to improve the pinhole array pattern design and light source intensity to enhance

transmittance and increase fluorescence intensity from samples, broadening the scope of

applications.

We have shown that AMATERAS-2 can perform time-lapse observation of the centimeter-sized

vascular networks in the quail embryo. It is a promising tool for understanding the multicellular

behavior, especially the mechanisms of tissue formation during development (1-3). In regard to

the quail embryo's case, the data obtained here are the first demonstration of the dynamics of

the extensive organization of the vascular network throughout the quail embryo at the single-

cell level. Conventional methods of research employing traditional microscopy have selected a

limited area for observation and thus had a difficulty in determining the distribution and

synchronicity of cellular dynamics across the entire system. By analyzing such a huge four-

dimensional data (*xyz* and *t*), we can gain insight into when and where each organ is formed, in

what order, and in what interrelationships. In addition, employing this method allows us to

break away from the conventional hypothesis-driven research manner and obtain new findings

in a data-driven manner. This advantage can be realized not only for quail embryo but also for

a variety of other species undergoing widespread and dynamic cellular events, making it a

powerful tool in the study of multicellular organisms, especially in the study of tissue formation.

To further advance multicellular systems research, information on individual cell states is

required in addition to cellular dynamics, therefore multiplex measurement is an essential

technological challenge. Although most of the observations presented in this paper were made

with monochromatic fluorescence imaging, multiplex imaging will lead to a more profound

understanding if information on the distribution of cellular states can also be obtained by

employing multiple fluorescent probes. This requires technological upgrades in optical filters

and light sources for multicolor imaging. In addition, spatial transcriptomics has been rapidly

developing in recent years and has become one of the most crucial tools for multicellular system

study (5). In the near future, dynamics imaging and spatial transcriptomics for the same sample

will be a very important technological integration to advance an integrative understanding of

self-organization in tissue formation. Rather than simply using AMATERAS sequentially with

existing spatial transcriptomics, it could be employed as an optical detection system for spatial

transcriptomics to achieve extremely high throughput in the same wide FOV of cellular

dynamics imaging.

Finally, let us discuss a practical challenge we encountered with our instrument, primarily

related to the handling of extensive image data. In our studies, the raw data for 3D imaging of

the mouse brain section and time-lapse 3D imaging of the quail embryo amounted to

approximately 100 and 500 GB, respectively. However, after multiple image processing and

analysis steps, a single experimental measurement resulted in several terabyte of image data.

Managing such vast data proves challenging for standard computers in terms of both software

and hardware capabilities. To address software limitations, we opted not to employ commercial

software for data analysis. Instead, we utilized our originally developed programs,

implementing certain optimizations to reduce analysis time, minimize write/read cycles, and

prevent memory overflows. As for hardware, we established a data server with petabyte storage,

thereby enabling efficient data sharing among collaborators from diverse research institutions

physically distant from one another. The existence of this infrastructure considerably facilitated

the smooth progress of collaborative research within this study. Considering the increasing

significance of handling image big data, we expect methodologies for image computation to

become even more critical in the future. Ideally, a comprehensive system should be developed,

encompassing not only data storage and sharing but also an integrated analysis solution in the

cloud. We would like to actively promote the development of such a system, essentially

encompassing advancements in imaging techniques and data-handling strategies.

MATERIALS AND METHODS

Wide-field imaging system (AMATERAS-2w) configuration

We aimed for AMATERAS to achieve a larger FOV-to-resolution ratio than conventional microscopes, which required a giant imaging lens system. We realized this design concept by collaborating with SIGMAKOKI CO., LTD. Inc. (Tokyo, Japan) to manufacture the giant objective and tube lenses as shown in Fig. 1. The characteristics of these lens are described in the main text. Designing such large lenses presents significant challenges in selecting optical glass materials compared to regular-sized lenses. Although there are over 200 types of optical glass available, our options are limited because the thickness or diameter of several types of glass may be unprocessable, or their transparency may degrade due to their size. Therefore, we iteratively explored the optimal design, considering the feasibility of material procurement and processing in conjunction with the optical design.

By combining these lenses, we constructed an imaging system with a 2× magnification, NA of 0.25, and a 44 mm field number. To facilitate both water-immersion and dry observation, we incorporated a glass plate whose thickness can be adjusted at the lens tip. The objective lens has a focal length of 120 mm, whereas its working distance is 14 mm owing to the presence of the attachment at the tip. The pupil positions of the objective lens and the tube lens are set externally to the lens bodies. This arrangement allows for the placement of a spatial filter or other devices at the pupil position, facilitating future enhancements in functionality.

For image acquisition, a 120-megapixel CMOS camera (VCC-120CXP1M, CIS, Tokyo, Japan) and a 250-megapixel CMOS camera (VCC-250CXP1M, CIS, Tokyo, Japan) were used, out of which we selected one depending on research purpose. Both the cameras have a chip size of 35 mm (diagonal). The pixel sizes of these two cameras are 2.2 μm and 1.5 μm, respectively. The sampling intervals are 1.1 μm and 0.75 μm at 2× magnification. Image data captured by the CMOS cameras are loaded into a workstation via a CoaXpress frame grabber board (APX-3664G3, Aval Data, Tokyo, Japan). After the imaging experiments, they were transferred to a network server with large storage capacity for long-term storage and data sharing among project members.

By use of the lens system and CMOS cameras, a wide-field epi-illumination fluorescence imaging system was constructed. The imaging system chassis was designed to accommodate both inverted and upright microscope geometries. For this study, all experiments were conducted using the inverted microscope arrangement. Epi-illumination was achieved using a high-brightness LED sources (SOLIS-470C and SOLIS-385C, Thorlabs, Newton, NJ), with an excitation filter (#87-800, Edmund Optics, Barrington, NJ) placed immediately after it. The illumination light was homogenized using a pair of lenslet arrays and projected onto the sample surface. Additionally, a similar LED light source (SOLIS-525C, Thorlabs, Newton, NJ) with an illumination homogenizer was also mounted as transmitted illumination light for bright-field observation. To split the light path, we used a custom-made long-pass dichroic mirror measuring 158 mm × 120 mm × 10 mm (BK7), which was designed for fluorescence imaging of GFP with a cut-off wavelength of 497 nm. A fluorescent filter (#86-992, Edmund Optics, Barrington, NJ) was positioned in front of the camera, which was firmly attached to the imaging lens using an F-mount.

The sample stage comprises a three-axis translational movable stage and a two-axis tilt stage. Only the z-axis stage is motorized, utilizing an electric actuator (SOM-C13E, SIGMAKOKI CO., LTD., Tokyo, Japan). For time-lapse observation under controlled conditions, we mounted a stage-top incubator (SV-141A, BLAST, Kawasaki, Japan) on the five-axis stage, providing a stable environment at 37 °C with CO_2 control. To ensure precise measurements, we enclosed the entire sample space within a dark box, shielding it from external light and temperature fluctuations.

Multipoint confocal system (AMATERAS-2c) configuration

In the multipoint confocal imaging system, we utilized a custom-made pinhole array disk with specific dimensions (pinhole size: 6 µm, spacing: 24 µm). This disk was mounted on a high-speed rotating machine (CrestOptics, Rome, Italy) and placed precisely on the image plane of the imaging lens. Alignment throughout the FOV was ensured using a translation stage (XR25P/M, Thorlabs, Newton, NJ) and tilt stage (AIS-1016B, SIGMAKOKI CO., LTD., Tokyo, Japan) to adjust height and tilt. Beneath the disk, we positioned a short-pass dichroic mirror

(DMSP490L, Edmund Optics, Barrington, NJ) to allow excitation light to be incident from

below. The excitation light source included a high-brightness LED (SOLIS-470C, Thorlabs,

Newton, NJ), an excitation filter (#87-800, Edmund Optics, Barrington, NJ), and a pair of

lenslet arrays (#63-231, Edmund Optics, Barrington, NJ), together with a 75-mm lens (f = 200

mm, LA1353-A, Thorlabs, Newton, NJ) to ensure uniform illumination onto the disk. The

fluorescent image was reflected by the dichroic mirror, transmitted through a fluorescent filter

(#86-992, Edmund Optics), and projected onto a 120-megapixel camera using either a 1× or 2×

relay lens (LSTL10H-F and LSTL20H-F, Myutron, Tokyo, Japan). The relay lens and camera

were attached with an F-mount and secured on a three-axis translational stage (PT3/M, Thorlabs,

Newton, NJ). To adjust the focus, the stage for the optical axis direction was motorized using

an electric actuator (SOM-C25E, SIGMAKOKI CO., LTD., Tokyo, Japan).

Measurement of 3D point spread functions

To evaluate the optical performance of the imaging systems, the 3D point spread function was

measured using green fluorescent beads. Fluorescent beads with a 0.2 μm diameter (FluoSphere

F8811, Invitrogen, ThermoFisher Scientific, Waltham, MA) were employed in the evaluation

of AMATERAS-2w (Figs. 2A-B) and those with a 0.5 μm diameter (FluoSphere F8813,

Invitrogen, ThermoFisher Scientific, Waltham, MA) were employed in the evaluation of

AMATERAS-2c (Figs. 3D-E). The beads were dispersed three-dimensionally in agarose gel on

a glass-bottom dish, and *z*-stacks were obtained by moving the sample in the *z*-direction in 1 or

2 μm steps.

Computational sectioning

In wide-field fluorescence imaging of a 3D volume, the superimposition of a fluorescent image

from the focal plane and background light from outside the focal plane is a fundamental issue.

This problem is not limited to wide-field imaging but also arises in multipoint confocal imaging

of thick or high-density samples. To overcome this challenge, we employed computational

sectioning, an image processing technique aimed at removing the background light component.

Various algorithms have been proposed (46,47), and Leica microscopes' standard software

includes this functionality. In this study, we developed an original algorithm as a simple and

robust method for data analysis.

The critical step in our approach involves estimating the background light component at each

layer of the *z*-stack data. To achieve this, we leverage the assumption that the background light

exhibits a low spatial frequency, whereas the focal plane image demonstrates a high spatial

frequency. Consequently, we employ an iterative low-pass filtering technique to estimate the

background light component. Specifically, as shown in Figs. S6A and B, a two-dimensional

(2D) low-pass filter is applied to the raw data ($f_0(x,y)$) to obtain a smooth surface ($L_0(x,y)$). To

obtain $f_1(x,y)$, $f_0(x,y)$ and $L_0(x,y)$ are compared and the smaller one is chosen to $f_1(x,y)$ at every

position (x,y), as expressed by Eq. (1).

538
$$f_j(x, y) = \min \left(f_{j-1}(x, y), \ L_{j-1}(x, y) \right) \tag{1}$$

539   where min(a,b) returns the smaller value of a and b. Next, the low-pass filter is applied to $f_1(x,y)$

540   to obtain $L_1(x,y)$; this process is repeated to make $L_j(x,y)$ closer to the baseline of the original

541   image. The iteration stops when the standard deviation of the difference of $f_j(x,y)$ and $L_j(x,y)$

542   reaches a preset value (ε).

543
$$\sqrt{Var \left(f_j(x,y) - \ L_j(x,y) \right)} < \varepsilon \tag{2}$$

Among various methods available as 2D low-pass filters, we sequentially applied one-

dimensional (1D) infinite impulse response (IIR) filter in two directions (x,y). This method is

faster than other methods such as a 2D Fourier transform (48). This speed difference is

especially significant when the number of pixels in an image is massive, as in the present case.

We adopted the Butterworth type of the IIR low-pass filter.

When the focal plane is empty and there is a bright object in the background, the baseline

estimation tends to be skewed by the intensity fluctuations of the background. To mitigate this,

we applied a 3×3 moving average filter to the entire image as a pre-processing step before

initiating the baseline estimation algorithm. A critical parameter in our algorithm is the cutoff

frequency of the low-pass IIR filter. If the cutoff frequency is set too high, the focal plane

component would be included in the background; if it is set to low, background light would remain in the focal plane. Thus, it is crucial to find an optimal middle ground for the cutoff frequency. During this optimization, we did not account for cell size or optical system performance. Indeed, we employed a user-friendly blind separation method based on independent component analysis (ICA) to enhance usability (49). Similar to ICA, we assumed that the fluorescence image in the focal plane deviates from the Gaussian (normal) distribution and that the superposition of images from multiple planes, including both in-focus and out-of-focus planes, results in a distribution closer to the Gaussian distribution. To quantify the non-Gaussian nature of the distribution, we considered several measures, including kurtosis, skewness, negentropy, and mutual information (49). Among these measures, we found skewness to be the most robust metric for our image dataset and incorporated it into our algorithm. The cutoff frequency was adjusted to maximize the skewness of the estimated in-focus image. The optimization of the cut-off frequency was performed on a subset of the data before applying it to the entire dataset. An example of the selection of non-Gaussianity measure and the optimization of the cut-off frequency is found in the Supplementary File (Fig. S4-S5, Note 2). The typical cut-off frequency was set at 0.06 in normalized frequency unit of 1/pixel, and the value was applied across all data.

Culture of hiPSCs and differentiation into cardiac organoid with cavity chamber structure

A hiPSC line 201B7 was routinely maintained as previously described (50) on iMatrix-511 (Matrixome, 89292) coated culture dish in StemFit medium (Ajinomoto, AK02N), and 10 μM Y-27632 (Fujifilm Wako, 030-24021) was added for the first 24 h after passage. Four days before inducing differentiation, hiPSCs were dissociated with TrypLE Select Enzyme (Thermo Fisher Scientific, A1285901) and 1.9×10^5 cells were passed into iMatrix-511 coated 12 well plate in StemFit medium supplemented with 10 μM Y-27632. After 24 h, the medium was changed to StemFit without Y-27632 and was exchanged daily for 3 days. Differentiation was initiated with the CDM3 medium (51) composed of RPMI-1640 with HEPES (Fujifilm Wako, 189-02145), 0.5 mg/ml human recombinant albumin (Sigma Aldrich, A9731) and 0.2 mg/ml L-ascorbic acid 2-phosphate (Sigma Aldrich, A8960) supplemented with 3 μM CHIR99021

(Fujifilm Wako, 038-23101) for 48 h. On day 2 of differentiation, the medium was changed to

CDM3 supplemented with 5 µM IWP-2 (Fujifilm Wako, 034-24301). From day 4 to day 8, the

cells were cultured in CDM3 and the medium was changed every other day. From day 8, the

medium was changed to RPMI-1640 with HEPES supplemented with 2% B27 (Thermo Fisher

Scientific, 17504044) and was changed every other day. The cells were fixed in 4%

paraformaldehyde (PFA; Thermo Fisher Scientific, 43368) on day 15 for immunocytochemical

analysis.

Immunocytochemical staining of cardiac organoid

Cells were fixed with 4% PFA for 15 min at room temperature (RT ~23 °C), permeabilized with

0.2% Triton X-100 (Fujifilm Wako, 807423) for 20 min at RT and incubated with blocking

solution composed of D-PBS with 5% BSA and 10 % FBS for 30 min at RT. Subsequently, the

cells were incubated with primary antibody (1:200 rabbit IgG anti-cTnT, Bioss, bs-10648R) at

4 °C overnight. After washing three times, cells were incubated with secondary antibody (1:500

alexa488 conjugated goat IgG anti-rabbit IgG, Abcam, ab150077) for 2 h at RT. The cells were

washed three times in D-PBS with 1 µg/ml Hoechst33342 (Dojin, H342). Subsequently, the

cells were incubated with D-PBS and used for fluorescent observation.

Color representation of the 3D image

Figures 6B–D show the 3D positions of the cells using the HSB (also known as HSV) color

model. After applying the method described earlier to remove background light from the z-stack

image, we conducted a maximum value intensity projection along the z-direction. The z-

position of the maximum value at each xy-position was then recorded (Fig. S6C). The color

image was reconstructed by associating the brightness and hue of the HSB table with the value

of the maximum intensity and the maximum value position, respectively (Fig. S6D). For the

hue table, we assigned red to blue to z-layer numbers ranging from 0 ($z = 0$ µm) to 20 ($z = 240$

µm).

Preparation of mouse brain section

All animal care and handling procedures were conducted with approval from the Animal Care

and Use Committee of Osaka University (approval numbers R02-8-7). Our utmost efforts were

made to minimize the number of animals used. Experiments involved adult male C57BL/6J

mice (SLC, Shizuoka, Japan) aged between 2 and 3 mo. To ensure proper anesthesia, mice were

deeply anesthetized through intraperitoneal injection of an anesthetic cocktail containing 0.3

mg/kg medetomidine (Nippon Zenyaku Kogyo, Fukushima, Japan), 4 mg/kg midazolam

(Sandoz Pharma, Basel, Switzerland), and 5 mg/kg butorphanol (Meiji Seika Pharma, Tokyo,

Japan). Subsequently, anesthetized mice underwent transcardial perfusion with saline, followed

by 4% paraformaldehyde (Nacalai Tesque, Kyoto, Japan) dissolved in phosphate-buffered

saline (PBS). Brain tissues were then excised and immersed in a 4% paraformaldehyde solution

until further use. For tissue preparation, brain tissue blocks were sliced into 1.5-mm thick

coronal sections using a vibrating microtome (LinearSlicer Pro7N, Dosaka EM, Kyoto, Japan).

Subsequent staining and tissue clearing involved immersion of sections in CUBIC-L (Tokyo

Chemical Industry, Tokyo, Japan) for delipidation, SYTOX Green (Thermo Fisher Scientific,

Waltham, MA, USA) solution (1:5,000) diluted in 20% (vol/vol) DMSO in PBS for nuclear

staining, and CUBIC-R+(M) (Tokyo Chemical Industry, Tokyo, Japan) for refractive index

matching, following previously established methods (31,52). After clearing, tissue sections

were placed on glass bottom dishes and embedded in 2% (wt/vol) agarose gel prepared with

CUBIC-R+(M) for subsequent imaging analysis.

Preparation of quail embryo

Transgenic quail line, tie1:H2B-eYFP (37) was bred in quail breeding facility at Kyushu

University. Fertilized eggs were incubated at 38 °C. The staging of quail embryos was based on

the Hamburger and Hamilton stages of chicken embryos (53). The animal study was approved

by the Institutional Animal Care and Use Committee of Kyushu University (Authorization

number: A20-019). *Ex ovo* culture was performed as previously described (54). Black filter

paper was used instead of white filter paper to avoid fluorescence background.

Autofocus in time-lapse imaging

To address *z*-directional drift in the time-lapse observation of quail, we developed an original

autofocus method. During the intervals between fluorescence z-images, we acquired bright-

field *z*-stacks and analyzed the images to identify the most in-focus position. To achieve this,

we used a black ink marker located on the substrate edge as the autofocus target instead of the

variable sample itself. For evaluating the degree of in-focus, we applied a total variation (TV)

filter to the marker image and used the kurtosis of the filtered image as the focus score. The z-

position with the maximum focus score was identified as the in-focus location, thus signifying

the sharpest image of the marker edge. Although other filter types and statistical moments are

available options for the focus score (55), our preliminary experiments demonstrated the

suitability of the kurtosis of TV for the marker image. In the actual experiment, we acquired 15

images in 15 μm steps and plotted the focus scores against z-positions. Through quadratic

function fitting of the focus scores at three z-positions (including the z-position with the

maximum value and adjacent positions), we estimated the best in-focus position with an

accuracy of 1 μm.

Deep-learning based cell detection for the mouse brain data

Cell detection was conducted using ELEPHANT, a unified platform that facilitates manual

annotation, deep learning and proofreading of results within a single user-friendly GUI (35,56).

ELEPHANT serves as an extension to Mastodon (57), an open-source framework for large-

scale tracking deployed in Fiji, a widely used image-analysis software (58). ELEPHANT

supports incremental deep learning with sparse annotation, incorporating algorithms for

detecting cells in 3D and tracking them in time-lapse 3D image datasets. In this paper, we

utilized the cell detection function exclusively for analyzing 3D imaging data of mouse brain

section (Fig. 5G-I). The incremental learning approach allows cell detection models to be

trained progressively on a dataset that begins with sparse annotations and is continuously

enhanced through human proofreading. For visualization, image data were displayed using

BigDataViewer (59) on Mastodon, accessed via BigDataServer (60). This setup permits the

visualization of large-scale image data on client computers while maintaining the data on the

server. Deep learning capabilities were provided by the ELEPHANT server, enabling remote

GPU access. This setup allows non-experts to perform image analysis on big data using deep

learning, effectively overcoming typical challenges such as the need for large amounts of high-

quality training data, the absence of an interactive user interface, and limited access to

computing resources, including substantial storage and high-end GPUs (61,62). The server

computer used in this study is equipped with an Intel(R) Xeon(R) Gold 6132 CPU @ 2.60GHz,

runs Ubuntu 20.04, and includes 384 GB DDR4 2,666 MT/s RAM, 2x NVIDIA Tesla V100-

PCIE-32GB GPUs, and a network-attached Lustre parallel file system with over 2 PB of storage.

The client computer features an Apple M1 Pro CPU, runs Sonoma 14.3, and has 16 GB of

LPDDR5-6400 RAM and a 500 GB SSD.

In the analyses presented in this paper (Fig. 5G-I), cell detection models were trained for three

regions, including the cortex. In each region, the detection model was refined by repeating the

sequence (annotation, training, prediction and proofreading) five to seven times. By annotating

approximately 100 cells in total, we were able to establish a model capable of detecting the vast

majority of the cells. The training of the detection models was conducted on volumes of 256 ×

256 x 16 voxels, which were prepared by preprocessing with a random flip in each dimension.

During the label generation step, the center ratio was set to 0.4 and the background threshold

was set to zero, meaning that only explicitly annotated voxels were used for training. In the

prediction step, volumes cropped to dimensions of $700 \times 700 \times 100$ around the target area were

used as input. In the postprocessing step, a threshold for nucleus center probabilities was set to

0.5, and r_{min}, r_{max} and d_{sup} were set to 1 μm, 5 μm, and 3 μm respectively (see details about the

parameters in 35).

To quantitatively evaluate the performance of cell detection, we calculated precision and recall,

defined as TP/(TP+FP) and TP/(TP+FN), respectively, where TP stands for true-positive, FP

for false-positive, and FN for false-negative. In the actual 3D data of the cortical region, the

values of TP, FP, and FN were manually counted by comparing the detection results of cells (N

~ 640) within a defined volume of $310 \times 170 \times 100$, one by one, with the actual images. As a

result, we obtained a precision of 99.4 % and a recall of 97.6 %.

Cell segmentation and tracking for the quail embryo data

For the analysis of quail embryo data, cell detection, segmentation and tracking were performed

using a custom-made program coded in Python. The observed 3D volume has a transverse area

(xy) of more than 1 cm^2 and a height (z) of 240 μm, which is relatively thin in the z direction

compared to the xy plane. The overlap of multiple vascular endothelial cells in the z direction

occurs with a very low probability. Therefore, to save computational costs and time, 2D cell

detection was conducted on *z*-projected MIP images instead of full 3D cell detection. To avoid

miss-detection of overlapping cells, we divided the *z*-stack data into three blocks along the *z*-

axis and created MIP images for each block, on which 2D cell detection was performed on the

MIP images. The resulting cell lists were compared to identify identical cells detected across

multiple blocks. By recognizing double-detected cell pairs as identical, we updated the cell list

accordingly. This process was applied to all double-detected cells, thus achieving effective cell

detection in the 3D volume and enabling cell tracking throughout the observation period.

Evaluation of optical system with optical invariant and space-bandwidth product

We employed two indices to evaluate the lens system's scale range and compare it with previous

research. The first index, the optical invariant, measures the lens system's performance and is

obtained by multiplying the FOV radius and NA. The second index, the space-bandwidth

product, considers the image sensor and wavelength while quantifying the ratio of actual

resolution to FOV. These indices are calculated using the following simplified formulas,

respectively (41).

711
$$I = \frac{NA \times FN}{2M} \qquad (3)$$

712
$$SBP = 8\left(\frac{R_{FOV}}{d_{xy}}\right)^2 \qquad (4)$$

713 where *I*, *FN*, *M*, *SBP*, *R$_{FOV}$*, and *dxy* denote optical invariant, field number, magnification,

714 space-bandwidth product, FOV radius, and spatial resolution (FWHM).

References

1. Y. Sasai, Cytosystems dynamics in self-organization of tissue architecture. *Nature*. **493**, 318–326 (2013).

2. K. McDole, L. Guignard, F. Amat, A. Berger, G. Malandain, L. A. Royer, S. C. Turaga, K. Branson, P. J. Keller, In Toto Imaging and Reconstruction of Post-Implantation Mouse Development at the Single-Cell Level. *Cell.* **175**, 859-876.e33 (2018).

3. M. H. Dominguez, A. L. Krup, J. M. Muncie, B. G. Bruneau, Graded mesoderm assembly governs cell fate and morphogenesis of the early mammalian heart. *Cell.* **186**, 479-496.e23 (2023).

4. H. R. Ueda, H. U. Dodt, P. Osten, M. N. Economo, J. Chandrashekar, P. J. Keller, Whole-Brain Profiling of Cells and Circuits in Mammals by Tissue Clearing and Light-Sheet Microscopy. *Neuron*. **106**, 369–387 (2020).

5. A. Rao, D. Barkley, G. S. França, I. Yanai, Exploring tissue architecture using spatial transcriptomics. *Nature*. **596**, 211–220 (2021).

6. G. McConnell, J. Trägårdh, R. Amor, J. Dempster, E. Reid, W. B. Amos, A novel optical microscope for imaging large embryos and tissue volumes with sub-cellular resolution throughout. *Elife*. **5**, e18659 (2016).

7. N. J. Sofroniew, D. Flickinger, J. King, K. Svoboda, A large field of view two-photon mesoscope with subcellular resolution for in vivo imaging. *Elife*. **5,** e14472 (2016).

8. J. Fan, J. Suo, J. Wu, H. Xie, Y. Shen, F. Chen, G. Wang, L. Cao, G. Jin, Q. He, T. Li, G. Luan, L. Kong, Z. Zheng, Q. Dai, Video-rate imaging of biological dynamics at centimetre scale and micrometre resolution. *Nat. Photonics*. **13**, 809–816 (2019).

9. K. Ota, Y. Oisi, T. Suzuki, M. Ikeda, Y. Ito, T. Ito, H. Uwamori, K. Kobayashi, M. Kobayashi, M. Odagawa, C. Matsubara, Y. Kuroiwa, M. Horikoshi, J. Matsushita, H. Hioki, M. Ohkura, J. Nakai, M. Oizumi, A. Miyawaki, T. Aonishi, T. Ode, M. Murayama, Fast, cell-resolution, contiguous-wide two-photon imaging to reveal

functional network architectures across multi-modal cortical areas. *Neuron*. **109**, 1810-1824.e9 (2021).

10. C. H. Yu, J. N. Stirman, Y. Yu, R. Hira, S. L. Smith, Diesel2p mesoscope with dual independent scan engines for flexible capture of dynamics in distributed neural circuitry. *Nat. Commun.* **12**, 6639 (2021).

11. T. Ichimura, T. Kakizuka, K. Horikawa, K. Seiriki, A. Kasai, H. Hashimoto, K. Fujita, T. M. Watanabe, T. Nagai, Exploring rare cellular activity in more than one million cells by a transscale scope. *Sci. Rep.* **11**, 16539 (2021).

12. T. Kakizuka, Y. Hara, Y. Ohta, A. Mukai, A. Ichiraku, Y. Arai, T. Ichimura, T. Nagai, K. Horikawa, *bioRxiv*, doi:10.1101/2020.06.29.176891.

13. G. S. Kino, T. R. Corle, *Confocal Scanning Optical Microscopy and Related Imaging Systems 1st Edition* (Academic Press, Cambridge, MA, 1996).

14. T. Wilson, Resolution and optical sectioning in the confocal microscope. *J. Microsc.* **244**, 113–121 (2011).

15. H. Gross, F. Blechinger, and B. Achtner. *Handbook of Optical Systems, Volume 4: Survey of Optical Instruments* (Wiley-VCH, Weinheim, 2008).

16. E. Battistella, J. Schniete, K. Wesencraft, J. F. Quintana, G. McConnell, Light-sheet mesoscopy with the Mesolens provides fast sub-cellular resolution imaging throughout large tissue volumes. *iScience*. **25**, 104797 (2022).

17. R. Prevedel, Y. G. Yoon, M. Hoffmann, N. Pak, G. Wetzstein, S. Kato, T. Schrödel, R. Raskar, M. Zimmer, E. S. Boyden, A. Vaziri, Simultaneous whole-animal 3D imaging of neuronal activity using light-field microscopy. *Nat. Methods*. **11,** 727–730 (2014).

18. J. Demas, J. Manley, F. Tejera, K. Barber, H. Kim, F. M. Traub, B. Chen, A. Vaziri, High-speed, cortex-wide volumetric recording of neuroactivity at cellular resolution using light beads microscopy. *Nat. Methods*. **18,** 1103-1111 (2021).

19. E. Papagiakoumou, E. Ronzitti, V. Emiliani, Scanless two-photon excitation with temporal focusing. *Nat. Methods* **17**, 571–581 (2020).

20. F. F. Voigt, D. Kirschenbaum, E. Platonova, S. Pagès, R. A. A. Campbell, R. Kastli, M.

Schaettin, L. Egolf, A. van der Bourg, P. Bethge, K. Haenraets, N. Frézel, T. Topilko, P.

Perin, D. Hillier, S. Hildebrand, A. Schueth, A. Roebroeck, B. Roska, E. T. Stoeckli, R.

Pizzala, N. Renier, H. U. Zeilhofer, T. Karayannis, U. Ziegler, L. Batti, A. Holtmaat, C.

Lüscher, A. Aguzzi, F. Helmchen, The mesoSPIM initiative: open-source light-sheet

microscopes for imaging cleared tissue. *Nat. Methods*. **16**, 1105–1108 (2019).

21. T. Nöbauer, Y. Zhang, H. Kim, A. Vaziri, Mesoscale volumetric light-field (MesoLF)

imaging of neuroactivity across cortical areas at 18 Hz. *Nat. Methods*. **20**, 600–609

(2023).

22. A. Ichihara, T. Tanaami, K. Isozaki, Y. Sugiyama, Y. Kosugi, K. Mikuriya, M. Abe, I.

Uemura, High-speed confocal fluorescence microscopy using a Nipkow scanner with

microlenses for 3D-imaging of single fluorescent molecule in real time. *Bioimaging* **4**,

52–62 (1996).

23. M. Ishihara, A confocal surface measurement system having improved measurement

accuracy for rough surfaces and measurement speed. *J. Robot. Mechatronics* **15**, 331–340

(2003).

24. M. Petráň, M. Hadravský, A. Boyde, The tandem scanning reflected light microscope.

Scanning **7**, 97–108 (1985).

25. G. Q. Xiao, T. R. Corle, G. S. Kino, Real-time confocal scanning optical microscope.

Appl. Phys. Lett. **53**, 716–718 (1988).

26. A. R. Halpern, M. Y. Lee, M. D. Howard, M. A. Woodworth, P. R. Nicovich, J. C.

Vaughan, Versatile, do-it-yourself, low-cost spinning disk confocal microscope. *Biomed.*

Opt. Express. **13**, 1102 (2022).

27. Z. Ma, J. Wang, P. Loskill, N. Huebsch, S. Koo, F. L. Svedlund, N. C. Marks, E. W. Hua,

C. P. Grigoropoulos, B. R. Conklin, K. E. Healy, Self-organizing human cardiac

microchambers mediated by geometric confinement. *Nat. Commun.* **61**. 6, 1–10 (2015).

28. P. Hofbauer, S. M. Jahnel, N. Papai, M. Giesshammer, A. Deyett, C. Schmidt, M. Penc,

K. Tavernini, N. Grdseloff, C. Meledeth, L. C. Ginistrelli, C. Ctortecka, Š. Šalic, M.

Novatchkova, S. Mendjan, Cardioids reveal self-organizing principles of human cardiogenesis. *Cell*. **184**, 3299-3317.e22 (2021).

29. Y. R. Lewis-Israeli, A. H. Wasserman, M. A. Gabalski, B. D. Volmert, Y. Ming, K. A. Ball, W. Yang, J. Zou, G. Ni, N. Pajares, X. Chatzistavrou, W. Li, C. Zhou, A. Aguirre, Self-assembling human heart organoids for the modeling of cardiac development and congenital heart disease. *Nat. Commun.* **121**. 12, 1–16 (2021).

30. E. A. Susaki, K. Tainaka, D. Perrin, F. Kishino, T. Tawara, T. M. Watanabe, C. Yokoyama, H. Onoe, M. Eguchi, S. Yamaguchi, T. Abe, H. Kiyonari, Y. Shimizu, A. Miyawaki, H. Yokota, H. R. Ueda, Whole-brain imaging with single-cell resolution using chemical cocktails and computational analysis. *Cell*. **157**, 726–739 (2014).

31. K. Tainaka, T. C. Murakami, E. A. Susaki, C. Shimizu, R. Saito, K. Takahashi, A. Hayashi-Takagi, H. Sekiya, Y. Arima, S. Nojima, M. Ikemura, T. Ushiku, Y. Shimizu, M. Murakami, K. F. Tanaka, M. Iino, H. Kasai, T. Sasaoka, K. Kobayashi, K. Miyazono, E. Morii, T. Isa, M. Fukayama, A. Kakita, H. R. Ueda, Chemical Landscape for Tissue Clearing Based on Hydrophilic Reagents. *Cell Rep*. **24**, 2196-2210.e9 (2018).

32. T. Hainmueller, M. Bartos, Dentate gyrus circuits for encoding, retrieval and discrimination of episodic memories. *Nat. Rev. Neurosci*. **21**, 153–168 (2020).

33. M. Agetsuma, H. Aizawa, T. Aoki, R. Nakayama, M. Takahoko, M. Goto, T. Sassa, R. Amo, T. Shiraki, K. Kawakami, T. Hosoya, S. I. Higashijima, H. Okamoto, The habenula is crucial for experience-dependent modification of fear responses in zebrafish. *Nat. Neurosci*. **13**, 1354–1356 (2010).

34. N. Dani, R. H. Herbst, C. McCabe, G. S. Green, K. Kaiser, J. P. Head, J. Cui, F. B. Shipley, A. Jang, D. Dionne, L. Nguyen, C. Rodman, S. J. Riesenfeld, J. Prochazka, M. Prochazkova, R. Sedlacek, F. Zhang, V. Bryja, O. Rozenblatt-Rosen, N. Habib, A. Regev, M. K. Lehtinen, A cellular and spatial map of the choroid plexus across brain ventricles and ages. *Cell*. **184**, 3056-3074.e21 (2021).

35. K. Sugawara, Ç. Çevrim, M. Averof, Tracking cell lineages in 3D by incremental deep learning. *Elife*. **11**, 1–19 (2022).

36. S. Herculano-Houzel, R. Lent, Isotropic fractionator: A simple, rapid method for the
quantification of total cell and neuron numbers in the brain. *J. Neurosci.* **25**, 2518–2521
(2005).

37. Y. Sato, G. Poynter, D. Huss, M. B. Filla, A. Czirok, B. J. Rongish, C. D. Little, S. E.
Fraser, R. Lansford, Dynamic analysis of vascular morphogenesis using transgenic quail
embryos. *PLoS One*. **5**, e12674 (2010).

38. Y. Sato, Dorsal aorta formation: Separate origins, lateral-to-medial migration, and
remodeling. *Develop. Growth Differ.* **55**, 113–129 (2013).

39. E. Meijering, O. Dzyubachyk, I. Smal, Methods for cell and particle tracking. *Elsevier
Inc*. **504**, 183–200 (2012).

40. J. B. Beltman, A. F. M. Marée, R. J. De Boer, Analysing immune cell migration. *Nat.
Rev. Immunol*. **9**, 789–798 (2009).

41. J. R. Bumstead, J. J. Park, I. A. Rosen, A. W. Kraft, P. W. Wright, M. D. Reisman, D. C.
Côté, J. P. Culver, J. R. Bumstead, J. J. Park, I. A. Rosen, A. W. Kraft, P. W. Wright, D.
Reisman, D. C. Côté, J. P. Culver, M. D. Reisman, D. C. Côté, J. P. Culver, Designing a
large field-of-view two- photon microscope using optical invariant analysis microscope
using optical invariant analysis. *Neurophotonics*. **5**, 025001 (2018).

42. A. W. Lohmann, R. G. Dorsch, D. Mendlovic, C. Ferreira, Z. Zalevsky, Space–
bandwidth product of optical signals and systems. *J. Opt. Soc. Am. A*. **13**, 470 (1996).

43. K. Seiriki, A. Kasai, T. Nakazawa, M. Niu, Y. Naka, M. Tanuma, H. Igarashi, K.
Yamaura, A. Hayata-Takano, Y. Ago, H. Hashimoto, Whole-brain block-face serial
microscopy tomography at subcellular resolution using FAST. *Nat. Protoc.* **14**, 1509–
1529 (2019).

44. S. Takanezawa, T. Saitou, T. Imamura, Wide field light-sheet microscopy with lens-
axicon controlled two-photon Bessel beam illumination. *Nat. Commun.* **12**, 2979 (2021).

45. L. Hof, T. Moreth, M. Koch, T. Liebisch, M. Kurtz, J. Tarnick, S. M. Lissek, M. M. A.
Verstegen, L. J. W. van der Laan, M. Huch, F. Matthäus, E. H. K. Stelzer, F. Pampaloni,

Long-term live imaging and multiscale analysis identify heterogeneity and core principles of epithelial organoid morphogenesis. *BMC Biol*. **19**, 1–22 (2021).

46. H. -C. Lee and G. Yang, Computational removal of background fluorescence for biological fluorescence microscopy. *IEEE 11th International Symposium on Biomedical Imaging (ISBI),* 205–208, doi: 10.1109/ISBI.2014.6867845 (2014).

47. K. Walter, F. Ziesche, Apparatus and Method, Particularly for microscopes and endoscopes, using baseline estimation and half-quadratic minimization for the deblurring of images, WO 2019/185174 A1. European Patent Office (2019).

48. P. Getreuer, A Survey of Gaussian Convolution Algorithms. *Image Process. Line*. **3**, 286–310 (2013).

49. A. Hyvaerinen, J. Karhunen, E. Oja, Independent Component Analysis (Adaptive and Cognitive Dynamic Systems: Signal Processing, Learning, Communications and Control) (Wiley-Interscience, New York, 2021).

50. M. Nakagawa, Y. Taniguchi, S. Senda, N. Takizawa, T. Ichisaka, K. Asano, A. Morizane, D. Doi, J. Takahashi, M. Nishizawa, Y. Yoshida, T. Toyoda, K. Osafune, K. Sekiguchi, S. Yamanaka, A novel efficient feeder-Free culture system for the derivation of human induced pluripotent stem cells. *Sci. Rep*. **4**, 1–7 (2014).

51. P. W. Burridge, E. Matsa, P. Shukla, Z. C. Lin, J. M. Churko, A. D. Ebert, F. Lan, S. Diecke, B. Huber, N. M. Mordwinkin, J. R. Plews, O. J. Abilez, B. Cui, J. D. Gold, J. C. Wu, Chemically defined generation of human cardiomyocytes. *Nat. Methods*. **11**, 855–860 (2014).

52. E. A. Susaki, C. Shimizu, A. Kuno, K. Tainaka, X. Li, K. Nishi, K. Morishima, H. Ono, K. L. Ode, Y. Saeki, K. Miyamichi, K. Isa, C. Yokoyama, H. Kitaura, M. Ikemura, T. Ushiku, Y. Shimizu, T. Saito, T. C. Saido, M. Fukayama, H. Onoe, K. Touhara, T. Isa, A. Kakita, M. Shibayama, H. R. Ueda, Versatile whole-organ/body staining and imaging based on electrolyte-gel properties of biological tissues. *Nat. Commun.* **11**, 1982 (2020).

53. V. Hamburger, H. L. Hamilton, A series of normal stages in the development of the chick embryo. *J. Morphol.* **88**, 49–92 (1951).

54. Y. Sato, R. Lansford, Transgenesis and imaging in birds, and available transgenic
reporter lines. *Develop. Growth Differ.* **55**, 406–421 (2013).

55. J. M. Mateos-Pérez, R. Redondo, R. Nava, J. C. Valdiviezo, G. Cristóbal, B. Escalante-
Ramírez, M. J. Ruiz-Serrano, J. Pascau, M. Desco, Comparative evaluation of autofocus
algorithms for a real-time system for automatic detection of Mycobacterium tuberculosis.
Cytom. Part A. **81 A**, 213–221 (2012).

56. https://github.com/elephant-track

57. https://github.com/mastodon-sc/mastodon

58. J. Schindelin, I. Arganda-Carreras, E. Frise, V. Kaynig, M. Longair, T. Pietzsch, S.
Preibisch, C. Rueden, S. Saalfeld, B. Schmid, J. Y. Tinevez, D. J. White, V. Hartenstein,
K. Eliceiri, P. Tomancak, A. Cardona, Fiji: An open-source platform for biological-image
analysis. *Nat. Methods.* **9**, 676–682 (2012).

59. T. Pietzsch, S. Saalfeld, S. Preibisch, P. Tomancak, BigDataViewer: visualization and
processing for large image data sets. *Nat. Methods.* **12**, 481–483 (2015).

60. https://github.com/bigdataviewer/bigdataviewer-server

61. E. Moen, D. Bannon, T. Kudo, W. Graf, M. Covert, D. Van Valen, Deep learning for
cellular image analysis. *Nat. Methods.* **16**, 1233–1246 (2019).

62. C. Wen, T. Miura, V. Voleti, K. Yamaguchi, M. Tsutsumi, K. Yamamoto, K. Otomo, Y.
Fujie, T. Teramoto, T. Ishihara, K. Aoki, T. Nemoto, E. M. C. Hillman, K. D. Kimura,
3Deecelltracker, a Deep Learning-Based Pipeline for Segmenting and Tracking Cells in
3D Time Lapse Images. *Elife*. **10**, 1–37 (2021).

ACKNOWLEDGMENTS

We would like to thank Prof. K. Fujita of Department of Applied Physics, Osaka University, Japan for his valuable comments on optics design. We are also grateful to Prof. S. Miyagawa, and Prof. Y. Sawa of Graduate School of Medicine, Osaka University, for their support on handling human iPS cells and valuable discussion.

Funding:

Grant-in-Aid for Scientific Research on Innovative Areas "Singularity Biology (No. 8007)" 21H00431 (YS), 18H05416 (HH), 18H05412 (SO), and 18H05410, 18H05408 (TN)

Grant-in-Aid for Transformative Research Areas (A) "Seeing through Scattering Media (No. 20A207)" 21H05590 and 23H041350 (TI)

the Research Program of "Five-star Alliance" in "NJRC Mater. & Dev." (TN)

Precursory Research for Embryonic Science and Technology (PRESTO) JPMJPR18G2 (TI)

JSPS KAKENHI JP23H00395 and JP24K22022 (HH)

AMED Brain/MINDS JP21dm0207117 (HH) and BINDS JP23ama121054 and JP23ama121052 (HH)

The Uehara Memorial Foundation (TN)

Takeda Science Foundation (TN, HH)

Core Research for Evolutionary Science and Technology (CREST) JPMJCR15N3 (TN), JPMJCR1926 (KS, SO)

RIKEN Cluster for Science, Technology and Innovation Hub (SO)

JST NBDC Grant Number JPMJND2201 (SO)

Author contributions:

Conceptualization: TI, TN

Optics design: TI, HH, TN, YT, SE

Lens design and fabrication: YT

Optical system construction: TI, YT, SE

Development of imaging system software: TI

Sample preparation: TK, YS, KS, HH

Data acquisition: TK, YS, TI

Data analysis and visualization: KI, TI, TK, KS

Setup of deep-learning platform: KS, HI, SO

Construction of data share server: HI, SO

Writing—original draft: TI

Writing—review & editing: TI, TN

Competing interests:

T.I, H.H and T.N have patent applications (2021-104163(JP), 2023-077341). K.S. is

employed part-time by LPIXEL Inc. All other authors declare no additional conflict of

interests.

Data and materials availability:

All data needed to evaluate the conclusions in the paper are present in the paper and/or

Supplementary Materials.

FIGURES

Fig. 1. AMATERAS-2, a centimeter-FOV cell-imaging system (A) Schematics of the

design of the objective and tube lenses for AMATERAS-2. (B) Appearance of the objective

and tube lenses with the width and length. (C) Diagram of the wide-field fluorescence

imaging system, AMATERAS-2w, using the two lenses in (A-B). See Materials and Methods

for detail.

Fig. 2. Optical performance of AMATERAS-2w (A),(B) Experimentally obtained PSF of

green fluorescent beads with a 0.2-μm diameter in the *xy* and *xz* planes, captured using CMOS

cameras with (A) 120 megapixels (pixel size = 2.2 μm) and (B) 250 megapixels (pixel size =

1.5 μm). The representative line profiles in the *x* and *z* directions on the beads are shown

(black lines with circular markers) along with Gaussian fit curves (red lines) and FWHMs,

and the mean and standard deviation (SD) calculated for 100 beads are described beside each

profile. (C) Fluorescence image of a cardiomyocyte sheet in the entire FOV. (D) Digitally

magnified view of the area indicated by a magenta square at the center of (C); scale bar, 100

μm. (E) Same area as (D) but observed by AMATERAS-1 for comparison. (F) Line profiles

of the lines in (D) and (E) for comparison of spatial resolution.

Fig. 3. AMATERAS-2c, a multipoint confocal imaging system for 3D volume imaging.

(A) Design of our pinhole-array disk. (B) Schematic of the optical configuration using the pinhole-array disk. (C) Effect of background light rejection by the presence of the pinhole-array disk in a $292 \times 202 \ \mu m^2$ area at the FOV center corresponding to $1/50^2$ of the entire FOV. (D),(E) Experimentally obtained PSF of green fluorescent beads with a diameter of 0.5 μm in the *xy* and *xz* planes, captured using relay lens with (D) 1× magnification (NA = 0.079) and (E) 2× magnification (NA = 0.12). The representative line profiles in the *x* and *z* directions on the beads are also shown (black lines with circular markers) along with Gaussian fit curves (red lines) and FWHMs, and the mean and standard deviation (SD) calculated for 100 beads are described beside each profile.

Fig. 4. Volumetric imaging with computational sectioning. (A),(B) Two-color images of

the cavity chamber structure of a myocardial organoid (A) before and (B) after computational

sectioning. Both (A) and (B) are MIP images of the z-stack data, captured using the 250-

megapixel CMOS camera; here, magenta and cyan represent the distribution of

immunostained cardiac troponin T and Hoechst-labeled nuclei, respectively. (C),(D) Six z-

layers with a difference of 40 µm in the yellow square region indicated in (A),(B); scale bars,

200 µm. (E) Three-dimensional isosurface representation in the dashed square region

indicated in (D).

Fig. 5. Volumetric imaging of a mouse brain section. (A) Schematic of a mouse brain showing the placement of the brain sample in the inverted configuration. (B) Schematic of a single *z*-layer in the coronal plane with size of the effective area covering the entire brain section. (C) Fluorescence images of three orthogonal cross sections of the dotted square indicated in (B); here, the yellow, cyan, and magenta lines represent the *z*-, *x*-, and *y*-positions of the cross sections, respectively. (D) Cross-sectional images of raw data in the same region as (C), where the computational sectioning is not applied. (E),(F) Magnified images of the hippocampal region (dashed square in (B)) in the (E) *xy*-plane and (F) *xz*-plane. For comparison, images obtained with 2× (left) and 4× (right) magnification systems are shown together. Note that the 3D regions observed at different magnifications do not perfectly match owing to the difficulty in optical alignment. (G-I) Cell detection by ELEPHANT in three brain regions, the cerebral cortex (G), medial habenula (H), and choroid plexus (I), where

1004   pairs of intersecting *xy*- and *xz*-planes are arranged vertically. The detected cells within each

1005   plane are marked with green and cyan ovals, and the straight line shared by the two planes is

1006   represented by a yellow dashed line, with certain cells along this line highlighted by magenta

1007   ovals. The *xy*-planes overlay the oval markers within a range of ± 25 μm above and below the

1008   focal plane, while the *xz*-plane shows only oval markers on the plane.

1009

Fig. 6. Dynamics of vascular endothelial cells in quail embryo captured by time-lapse observation. (A) Schematic showing how the specimen is mounted (left) and a photograph of a specimen placed in a glass-bottom dish. (B) Representative images obtained at $t = 0$ h, 8 h, and 16 h; scale bars, 2 mm. (C) Enlarged view of the image corresponding at $t = 24$ h; the pseudo-colors represent the z-position of the cells and fluorescence intensity by their hue and brightness, respectively. (D) Magnified views of the anterior side of ventral aortae (left dashed square in (C)) at the four time-points with intervals of 8 h; scale bars, 200 μm. (E) Three-dimensional isosurface representation of the dorsal aorta region (right dashed square in (C)) viewed from the left side. (F),(G) Cell position trajectories and spatial mapping of features related to the cell dynamics in the two time-regions of (F) $t = 8$–10 h and (G) $t = 16$–18 h. Left panels show enlarged views in the white square regions (292×202 μm^2) indicated in the right panels. In both the panels, the rainbow-colored trajectories of the cell movement

1023   are overlaid on the grayscale MIP images at the first frame in the time regions (F: $t = 8$ h, G: t

1024 $= 16$ h). The scale bars are 50 μm. The rainbow color table represents the meandering index

1025   of the cell movement in the two time-regions.

1026