Full title

Reference materials for SARS-CoV-2 molecular diagnostic quality control: validation of encapsulated synthetic RNAs for room temperature storage and shipping

Short title

Room temperature storage and shipping of RNA reference materials for covid-19 detection quality control

Authors

Marthe Colotte[1*] Aurélie Luis[1], , Delphine Coudy[1], Sophie Tuffet[1], Isabelle Robene[3], Babbitha Fenelon[3], Emmanuel Jouen[3], Nicolas Leveque[4], Luc Deroche[4], Sophie Alain[5], Dorian Plumelle[6], Camille Tumiotto[7], Laurent Busson[7], Marie-Edith Lafon[7], Jacques Bonnet[1,2*]

Affiliations

[1] Imagene Company, Gradignan France;

[2] Université de Bordeaux, BRIC, U 312 INSERM Bordeaux, France;

[3] CIRAD Pôle de protection des plantes, Saint-Pierre, La Réunion, France,

[4] Laboratoire de virologie et mycobactériologie, CHU de Poitiers, Poitiers, France;

[5] CHU Limoges, Limoges, France;

[6] Laboratoire d'Analyses Médicales, Laboratoire Plumelle, Salon De Provence, France;

[7] CHU de Bordeaux, laboratoire de virologie, UMR 5234, Université de Bordeaux, Bordeaux, France.

Corresponding authors

23   colotte@imagene.eu; jacques-emile.bonnet@u-bordeaux.fr;

Abstract

The Coronavirus pandemic unveiled the unprecedented need for diagnostic tests to rapidly

detect the presence of pathogens in the population. Real-time RT-PCR and other nucleic

acid amplification techniques are accurate and sensitive molecular techniques that

necessitate positive controls. To meet this need, Twist Bioscience has developed and

released synthetic RNA controls. However, RNA is an inherently unstable molecule needing

cold storage, costly shipping, and resource-intensive logistics. Imagene provides a solution

to this problem by encapsulating dehydrated RNA inside metallic capsules filled with

anhydrous argon, allowing room temperature and eco-friendly storage and shipping. Here,

RNA controls produced by Twist were encapsulated (RNAshells) and distributed to several

laboratories that used them for COVID-19 detection tests by amplification. One RT-LAMP

procedure, four different RT-PCR devices and 6 different PCR kits were used. The

amplification targets were genes E, N; RdRp, Sarbeco-E and Orf1a/b. RNA retrieval was

satisfactory, and the detection was reproducible. RNA stability was checked by accelerated

aging. The results for a 10-year equivalent storage time at 25 °C were not significantly

different from those for unaged samples. This room temperature RNA stability allows the

preparation and distribution of large strategic batches which can be stored for a long time

and used for standardization processes between detection sites. Moreover, it makes it also

possible to use these controls for single use and in the field where large differences in

temperature can occur. Consequently, this type of encapsulated RNA controls, processed

at room temperature, can be used as reference materials for the SARS-Cov-2 virus as well

as for other pathogens detection.

Introduction

The Coronavirus pandemic has led to an unprecedented need for diagnostic tests to rapidly

detect the presence of this pathogen in the population. Real-time RT-PCR and RT-LAMP

amplification techniques proved to be accurate and sensitive molecular tools for virological

diagnosis [1,2,3,4,5].

A key aspect of any ongoing pathogen detection effort is using quality control measures for

a wide range of applications from diagnostic assay development to day-to-day testing. This

includes verifying and validating diagnostic tests of both Next-Generation Sequencing

(NGS) and reverse transcription-amplification (RT-PCR or RT-LAMP) assays [6,7,8,9].

Controls based on viral nucleic acids extracted from infected patients or live viruses

propagated in cell culture have safety and security concerns. Synthetic RNA controls, like

those generated through gene synthesis by Twist Bioscience, mitigate the biohazard risk

and broaden access across diverse strains [10]. Additionally, as DNA synthesis is a readily

available technology, new RNA controls can be rapidly developed in response to the

emergence of new strains or pathogens.

However, RNA is an inherently unstable molecule. In solution or dry i), RNA rapidly

degrades through hydrolysis or oxidation (for a review, see [11]. This instability presents

storage and cost challenges for laboratories that require large stocks of control RNA (for a

review see [2]). Imagene's technology consists of confining dehydrated RNA, in the

presence of proprietary stabilizing solutions, inside metallic capsules filled with anhydrous

argon (RNAshells®), keeping away the samples from deleterious factors. Thus, it offers a

solution to the stability problem by allowing extensive and eco-friendly storage and

shipping a room temperature [2].

Here, Imagene encapsulated synthetic RNA controls (RNAshells) produced by Twist

Bioscience. These controls were then distributed to five laboratories for fitness-for-purpose

analysis. The objective was to verify, first, that the detection of the virus was reliable and

second that the RNA stability was insured for a minimum of a 10-year storage period at

room temperature.   For that, we used four different RT-PCR devices with six different PCR

kits and one LAMP procedure on RNA from three SARS-Cov-2 variants and five gene

targets.

Material and methods

Workflow

The RNA samples synthetized by Twist Bioscience were sent on dry ice to Imagene. One

part of these samples was aliquoted and sent on dry ice to Bordeaux CHU. The other part

was aliquoted and encapsulated. A set of these aliquots was heated to simulate 10 years of

room temperature storage and sent to Bordeaux CHU. The other aliquots were sent to

Bordeaux, Limoges, Poitiers CHUs, the CIRAD Réunion and a private clinical laboratory

(Plumelle, Salon de Provence).

The experimental strategy we followed is summarized in Fig 1.

Fig 1. **Experimental strategy for the preparation, distribution and validation of**

RNAshell® encapsulated RNA controls for SARS-CoV-2 virus detection.

89   RNA Twist RNA controls

90 Synthetic SARS-CoV-2 RNA control molecules for three SARS-CoV-2 variants were graciously

91 supplied by Twist Bioscience. Details are given Table 1.

92   **Table 1. Synthetic SARS-CoV-2 RNA controls used in this study**

Control	GenBank ID	GISAID name	Name used
Control 1	MT007544.1	Australia/VIC01/2020	RNA 1
Control 2	MN908947.3	Wuhan-Hu-1	RNA 2
Control 7	EPI_ISL_418227	France/HF2393/2020	RNA 7

93   The controls are ssRNA *in vitro* transcribed from synthetic DNA. Each one consists of

94   six non-overlapping 5 kb fragments, providing greater than 99% coverage of the viral

95   genome.

96 RNA encapsulation and accelerated ageing

RNA controls were sent to Imagene over dry ice as 100 µL solutions at a concentration of

approximately 10^6 copies per microliter [10]. They were distributed in 10 µL aliquots and

stored at -80℃. Aliquots were thawed, diluted with an Imagene proprietary stabilization

solution and distributed into glass inserts fitted in capsules to obtain either 10^4 copies (low)

or 10^5 copies (high) per capsule.

RNA samples were vacuum-dried and encapsulated under anhydrous and anoxic

argon/helium according to the process developed by Imagene [12,13] to give RNAshells®. A

set of 3 capsules of each variant RNA and each copy number (a total of 18 capsules) were

heated at 90 °C for 16 h to simulate 10 years of storage at 25 °C according to our previously

published Arrhenius equation for encapsulated RNA [11]. RNAshells® were then shipped at

room temperature to the participating laboratories.

RNAshell® opening and RNA rehydration

The RNAshells® were opened with a portable ShellOpener[TM] supplied by Imagene following

the given instructions. After opening, a given volume of RNAse-free ultrapure water was

added in the capsule and the RNA was left 10-20 minutes on the bench to rehydrate. At

that point, the RNA was ready for analysis (RT-qPCR or RT-LAMP). The different volumes

used to rehydrate and analyze the RNA by RT-qPCR in the different laboratories are given in

Table 2.

Table 2. Summary of the material and methods for the different RT-qPCR analysis.

Site	Bordeaux CHU	Bordeaux CHU	Limoges CHU	Poitiers CHU	Poitiers CHU	Private clinical Laboratory
Instrument	Light Cycler 480 (Roche)	Light Cycler 480 (Roche)	CFX96 (BioRad)	CFX96 (BioRad)	CFX96 (BioRad)	SLAN®-96P Real-Time PCR System (Sansure)
Kit	GeneFinder Covid-19 Plus, Osang Healthcare	SARS CoV-2 R Gene, Argene Biomérieux [14,15]	SuperScript™ III One-Step RT-PCR (Thermofisher)	TaqPath COVID-19 Multiplex Diagnostic	Allplex™ 2019-nCoV Assay (Seegene)	Kit Novel coronavirus 2019-ncov (Sansure)

				Solution (Thermofisher)		
	(distributed by Elitech)[2] [14]					
Nb cycles	45	45	45	40	45	45
Cycles	20 min @ 50°C 5 min @ 95°C 45 cycles : - 15 s @ 95°C - 50 s @ 58°C	5 min @ 50°C 15 min @ 95°C 45 cycles : - 10 s @ 95°C - 40 s @ 60°C - 25 s @ 72°C	ND	2 min @ 25 °C 10 min @ 53°C 2 min @ 95°C 40 cycles : - 3 s @ 95°C - 30 s @ 60°C	20 min @ 50 °C 15 min @ 95°C 45 cycles : - 15 s @ 94°C - 30 s @ 58°C	30 min @ 50°C 1 min @ 95°C 45 cycles : - 15 s @ 95°C - 30 s @ 60°C
Positivity	Ct < 43	ND	fluo signal > 6 000 for IP2 > 13 000 for IP4	Ct < 37	Ct < 40	ND
Target genes	E, N, RdRp	E, N, RdRp	IP2-RdRp, IP4-RdRp Sarbeco-E	S, orf1, N	E, N, RdRp	ORF1ab, N
RNA Controls	RNAshells #1, 2, 7 10^4 copies 10^5 copies 18 aged RNAshells (3 of each) Solution #1, 2, 7 10 µL @ 10^6	RNAshells #1, 2, 7 10^4 copies 10^5 copies 18 aged RNAshells (3 of each) Solution #1, 2, 7 10 µL @ 10^6	RNAshells #1, 2, 7 10^4 copies 10^5 copies	RNAshells #1, 2, 7 10^4 copies 10^5 copies	RNAshells #1, 2, 7 10^4 copies 10^5 copies	RNAshells #2 6x 10^4 copies 6x 10^5 copies

	copies / μL	copies / μL				
Rehydration volume	10 μL	10 μL	20 μL	25 μL	25 μL	ND
RNA volume in reaction	5 μL	5 μL	5 μL	10 μL	8 μL	ND
Theoretical number of copies / reaction	5,000 or 50,000	5,000 or 50,000	2,500 or 25,000	4,000 or 40,000	3,200 or 32,000	ND

116   ND: non disclosed

Analysis of RNA samples in solution

118   The non-diluted and non-encapsulated aliquots of RNA 1, 2 and 7, at approximately 10^6

119   copies per μL, were prepared and sent by Imagene on dry ice to Bordeaux CHU. The

120   samples were thawed and diluted to match the concentrations of the rehydrated capsules

121   (here, 10^3 and 10^4 copies per μL). They were analyzed in parallel by RT-qPCR (see Table 2).

RT-qPCR analysis

124 The different available RT-qPCR materials and methods used in the study are summed up in

125 Table 2.

RT-LAMP analysis

Here, the RT-LAMP system was first tested on a dilution series of control RNA 2 Low or

High, rehydrated with 50 µL of RNAse-free pure water. For Control 2 Low: a single analysis

was done for 1000 copies/reaction, 2 replicates were done for 500 and 250 copies/reaction

and 3 replicates for 100, 50 and 25 copies/reaction. For Control 2 High: 3 replicates were

done for 1000 copies/reaction and 4 replicates for 200, 40 and 8 copies/reaction.

To test the encapsulated RNA 2 as a positive control in the field, 60 independent RT-LAMP

runs were performed for the presence of SARS-CoV-2. Each run included 14 samples, a

negative control and a positive control. The positive control consisted of RNA 2 rehydrated

with water and further diluted in *Amies* medium to have a final concentration of 1000

copies/reaction.

RT-LAMP was performed according to the protocol RunCov developed by the CIRAD-UMR

PVBMT (3P) and validated by the National Reference Center [16]. The reactions were

performed in a 25 µL total reaction volume containing 15 µL of ISO-DR004-RT300

Isothermal Mastermix (OptiGene Ltd, Horsham, United Kingdom), 2.475 µL of each primer

pair (N and S), 0.05 µL (0.5 units) of AMV transcriptase (Promega, Charbonnières-les-Bains,

France) and 5 µL of template RNA. LAMP reactions were run on a Genie II instrument

(OptiGene) at 65°C for 25 min followed by an annealing step during which the temperature

was decreased from 92°C to 78°C at a speed of 0.05°C/s to determine the annealing

temperature (Ta) of the produced amplicon.

Statistical analysis

All statistical analyses including ANOVA, Fisher Test and previous verification of data normality (histogram and normal: Q-Q plot), were performed using the R statistical software, version 4.1.1 (2021-08-10; R Development Core Team, Vienna, Austria) with the different packages: stats, multcomp, emmeans and ggplot2).

Results and discussion

Objectives and strategy

The aim of this work was to validate whether encapsulated RNA controls stored and shipped at room temperature were fit for purpose as reference materials for internal quality control of SARS-CoV-2 detection.

Bordeaux CHU received non-diluted, non-encapsulated, frozen samples in solution, to compare the controls before and after encapsulation (i.e. to check the potential effect of the process and stabilization solution). They also received the aged RNAshells® to validate that 10 years of storage at room temperature did not alter the performance of the encapsulated controls. Two other laboratories (Poitiers CHU and Limoges CHU) received RNAshells® to verify that the controls gave the same performances when run with other RT-qPCR techniques. The CIRAD-UMR PVBMT (3P) laboratory tested the controls in multiple experiments while developing a RT-LAMP test for field use. Finally, RNAshells® were sent to a private clinical laboratory to check that the capsules can be routinely used by any operator.

Controls in solution

First, we checked the performance of the controls on non-encapsulated samples. The

results of these experiments are presented Fig 2.

Fig 2. RT-qPCR for controls in solution. The controls were thawed and diluted at

reception. Measures were done at Bordeaux CHU on 3 variants RNA (RNA 1, 2, 7), 2

copy numbers: 5,000 (low, L) and 50,000 (high, H) copies/reaction, 3 target genes

(polymerase gene, N gene and E gene).

All the controls gave positive results. For a given low or high concentration and a given

target gene, the Ct were comparable between the different variants (p = 0.647). The

highest Ct differences came from the gene targets (p=7.79 x 10^{-4}), probably due to

differences in PCR efficiency.

The chosen copy numbers gave Ct values in a suitable range (around 30 and 27 respectively

for the low and high controls), clearly different from the negative control (above 40-45).

The Ct were also not too low. This is advantageous because a positive control with too

many more copies than in patient samples could result in carry-over contamination in the

nearby wells of the PCR plate leading to false positive samples.

RNA controls encapsulation and accelerated aging

Then, we needed to verify, that encapsulation does not induce alterations in the controls. Another important point is to test whether RNA controls could be stored at room temperature for an extended period of time in RNAshell®, so a set of capsules were submitted to an accelerated aging by heating to simulate a storage period of 10 years at 25°C according to our previously published Arrhenius study [11]. The results are shown Fig 3.

Fig 3. RT-qPCR on RNAshell®-encapsulated and aged controls. RT-qPCR measures were done by Bordeaux CHU on 3 variants RNA (RNA 1, 2, 7), 2 copy numbers: 5,000 (low, L) and 50,000 (high, H) copies/reaction, 3 target genes (polymerase gene, N gene and E gene) and after 2 storage times: time 0 post-encapsulation and after heating for a period equivalent to 10 years at room temperature (0 and 10).

All analyzed RNA, whatever the variant, target gene or storage time were successfully

amplified.

To analyze the extent and sources of variations on these experiments we used variant 1.

The results are shown Fig 4.

Fig 4. Group analysis for Ct on variant 1 controls. Ct (3 replicates) are shown

according to: a) the state: encapsulated (encaps), encapsulated and aged (encaps

10) or in solution (liquid) and b) the number of copies (5,000 (low) and 50,000

(high) of each target gene. Groups of Ct values from the same state (dry vs liquid)

and copy number are circled.

The encapsulated controls and the controls in solution show significant Ct differences

(p=0.001999), due partly to the encapsulation step, including the addition of a stabilization

solution, but more probably to the fact that the controls in solution did not come from the

same diluted samples. However the Ct values of encapsulated RNA are maintained in the

same suitable range, meaning that the encapsulation process did not lead to RNA loss,

degradation or PCR inhibition to an extent that would make RNAshell®-encapsulated RNA

unfit for purpose.

On the other side, no significant difference was observed between 0- and 10-years

RNAshells® (Fig 3, p=0.7606) meaning that the simulated long-term storage of RNA did not

lead to significant degradation. One can therefore consider that the encapsulated RNA

controls are stable over at least 10 years at room temperature and are suitable reference

materials.

219   Compatibility with other RT-qPCR kits and instruments

220   To broaden the validation, capsules were distributed to laboratories that performed tests

221   with three other RT-amplification techniques, including the method used by the National

222   Reference Center for SARS-CoV 2 (on RdRp IP2 and IP4 regions). The results are shown

223   Fig 5.

Fig5. RT-qPCR results from Limoges and Poitiers CHU. For Limoges CHU, the RT-qPCR measures were done on 3 variants RNA (RNA 1, 2 and 7), 2 copy numbers: 2,500 (Low, L) and 25,000 (High, H) copies/reaction), and 3 target genes (RdRp-IP2, RdRp-IP4 and E genes). For Poitiers CHU, RNAshells® were rehydrated with 25 µL of

water. Eight μL were used for the Seegene kit and 10 μL for the Thermofisher kit. Ct

values were obtained by RT-qPCR measures done on 3 variants RNA (RNA 1, 2 and

7), 3 target genes (E, N and RdRp genes), with 2 copy numbers: 10,000 (Low, L) and

100,000 copies/capsule (high, H) giving 3,200 and 32,000 copies per reaction for the

Seegene kit and 4,000 and 40,000 copies per reaction for the Thermofisher kit. For

each technique, the control capsules were run in 3 distinct experiments (Poitiers

CHU) and the mean and SE of the 3 Ct were calculated.

Although the experiments were run over 3 days (at Poitiers CHU), the sampling from a

given capsule gave constant results. In all the experiments, techniques and targets, the

results obtained in the different laboratories were reproducible among all the capsules,

showing that the encapsulated control RNA are suitable for their use as internal quality

controls. The Ct for variants were strictly equivalent. The highest variability came from

the analysis kits and target genes as demonstrated above with the controls in solution.

Compatibility with a RT-LAMP procedure

SARS-Cov-2 genome was successfully amplified using the RT-LAMP assay from both RNA 2

low and RNA 2 high capsules with a high sensitivity (100% positive samples at 25 copies per

reaction for RNA2 low, 40 copies for RNA2 high and 50% positive samples at 8 copies for

RNA 2 high). Annealing temperature peaks were obtained for both S and N regions for 90%

of the positive samples with annealing temperatures ranging between 88.4-88.7 °C and

83.4-83.8 °C, for N and S regions, respectively. For 10% positive samples only one annealing

temperature peak (S or N) was obtained. This last result was principally obtained for low

copy numbers. This was already observed in other RT-LAMP experiments involving other

types of samples and is probably due to differences in primers binding probabilities to their

respective targets.

In the field, 1000 copies per reaction were taken so that the signal came out early,

validating the run before or at the same time as positive samples came out.

The results are given Table 3.

Table 3. RT-LAMP results from CIRAD.

RNAshell RNA 2	Copies / reaction	Amplification time (min)	Annealing temperature	
			N gene	S gene
Low	1000	10.00	88.5	83.50
Low	500	11.75	88.5	83.7
Low	500	11.00	88.5	83.7
Low	250	12.75	88.5	83.6
Low	250	13.50	88.6	83.7
Low	200	14.50	88.6	83.8
Low	200	14.50	88.4	83.8
Low	200	16.75	88.7	83.7
Low	100	13.00	88.6	83.8
Low	100	12.50	88.8	83.7
Low	100	13.00	88.5	83.7
Low	50	13.50	88.7	83.8
Low	50	17.00	88.7	83.8
Low	50	14.00	88.7	83.7
Low	25	16.75	88.3	83.8
Low	25	14.00	88.6	83.8
Low	25	13.50	88.4	83.6
High	1000	10.30	88.7	83.7
High	1000	10.15	88.7	83.7
High	1000	10.15	88.6	83.7
High	200	12.00	88.6	83.7
High	200	11.30	88.6	83.6
High	200	11.30	88.6	83.6
High	200	10.45	88.7	83.7
High	40	12.30	88.7	83.7
High	40	13.45	-	83.7
High	40	13.15	-	83.7
High	40	13.45	88.6	88.5
High	8	19.45	-	83.4

High	8	23.30	88.7	-
High	8	-	-	-
High	8	-	-	-

RNA 2 samples High and Low were rehydrated with 50 µL of water and diluted to

have between 1000 and 8 copies per reaction and amplified using the Runcov RT-

LAMP protocol of the CIRAD La Réunion. The amplification time and annealing

temperatures are given.

” Real life” use of the capsules as internal quality controls

For the monitoring of SARS-CoV 2 detection in the field by RT-LAMP, the RNA 2 variant

(1000 copies) was used as a positive control for 60 RT-LAMP tests run on samples taken

and analyzed in 4 different sites (A, B, C, D). This control was successfully amplified in each

run (Fig 6).

Fig 6. LAMP tests The RNA 2 variant (1000 copies) was used as a positive control

270   for 60 (real life) tests run on samples analyzed at 4 different sites (A, B, C, D). For

271   each test, both N and S regions were amplified.

For each test, both N and S regions were amplified, generating annealing peaks within the

expected annealing temperature ranges (not shown).

This field study shows that the capsules can be used as positive controls at the point of care

or on site with simplified logistics due to their stability at room temperature.

Validation of RNA controls for private clinical laboratories

To test the robustness of the controls, a series of experiments we conducted in a private

clinical laboratory. The laboratory received the capsules and the instructions for use and

included the rehydrated RNA in their routine diagnostic test. The results are shown Fig 7.

Fig 7. Private clinical laboratory-. Results on 2 target genes and 2 levels on 6 capsules

283   for each.

The results showed a good reproducibility among the capsules except for 2 capsules with a

low RNA copy number where the operator probably had rehydrated the RNA with a higher

volume of water, decreasing the number of copies in the test.

So, the capsules can easily be used as reference materials for internal quality control of the

monitoring of the diagnostics procedures, without the need for extensive training of the

operators.

291 Conclusion

This work shows that e RNAs synthetized by Twist Bioscience and encapsulated in

RNAshells® by Imagene have been validated as positive controls for SARS-CoV-2 detection.

Both 10,000 and 100,000 copy numbers per capsule are suitable for their use in RT-qPCR

and RT-LAMP analysis. The signal is reproducible among the variants, the targets and the

analysis methods. Moreover, the RNAshell®-encapsulated RNA showed an excellent

stability, with an unprecedented shelf life of more than 10 years at room temperature. This

stability is a great advantage because, first, it facilitates the use of these controls in the

field, in particular when large variations of temperature can be expected, and second,

importantly, it allows the preparation and the distribution of large strategic batches which

can be kept for a long time and used for standardization between processes or detection

sites.

Bibliography

1. Pascarella G, Strumia A, Piliego C, Bruno F, Del Buono R, Costa F, et al. COVID-19 diagnosis and management: a comprehensive review. J Intern Med. 2020; 288(2): 192-206. doi: 10.1111/joim.13091

2. Thomas E, Delabat S, Carattini YL, Andrews DM. SARS-CoV-2 and Variant Diagnostic Testing Approaches in the United States. Viruses. 2021; 13(12). doi: 10.3390/v13122492

3. Yoo HM, Kim IH, Kim S. Nucleic Acid Testing of SARS-CoV-2. Int J Mol Sci. 2021; 22(11). doi: 10.3390/ijms22116150

4. Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DK, et al. Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. Euro Surveill. 2020; 25(3). doi: 10.2807/1560-7917.Es.2020.25.3.2000045

5. Lescure FX, Bouadma L, Nguyen D, Parisey M, Wicky PH, Behillil S, et al. Clinical and virological data of the first cases of COVID-19 in Europe: a case series. Lancet Infect Dis. 2020; 20(6): 697-706. doi: 10.1016/s1473-3099(20)30200-0

6. John G, Sahajpal NS, Mondal AK, Ananth S, Williams C, Chaubey A, et al. Next-Generation Sequencing (NGS) in COVID-19: A Tool for SARS-CoV-2 Diagnosis, Monitoring New Strains and Phylodynamic Modeling in Molecular Epidemiology. Curr Issues Mol Biol. 2021; 43(2): 845-867. doi: 10.3390/cimb43020061

7. Muzzey D, Evans EA, Lieber C. Understanding the Basics of NGS: From Mechanism to Variant Calling. Curr Genet Med Rep. 2015; 3(4): 158-165. doi: 10.1007/s40142-015-0076-8

8. Rahbari R, Moradi N, Abdi M. rRT-PCR for SARS-CoV-2: Analytical considerations. Clin Chim Acta. 2021; 5161-7. doi: 10.1016/j.cca.2021.01.011

9. Yang S, Rothman RE. PCR-based diagnostics for infectious diseases: uses, limitations, and future applications in acute-care settings. Lancet Infect Dis. 2004; 4(6): 337-348. doi: 10.1016/s1473-3099(04)01044-8

10. twist. https://www.twistbioscience.com/productchild/tools-addressing-sars-cov-2-virus.

11. Fabre AL, Colotte M, Luis A, Tuffet S, Bonnet J. An efficient method for long-term room temperature storage of RNA. Eur J Hum Genet. 2014; 22(3): 379-385. doi: 10.1038/ejhg.2013.145

12. Bonnet J, Colotte M, Coudy D, Couallier V, Portier J, Morin B, et al. Chain and conformation stability of solid-state DNA: implications for room temperature storage. Nucleic Acids Res. 2010; 38(5): 1531-1546. doi: 10.1093/nar/gkp1060

13. Imagene. http://www.imagene.eu/.

14. Ong DSY, Claas ECJ, Breijer S, Vaessen N. Comparison of the GeneFinder(TM) COVID-19 Plus RealAmp Kit on the sample-to-result Platform ELITe InGenius to the national reference method: An added value of N gene target detection? J Clin Virol. 2020; 132104632. doi: 10.1016/j.jcv.2020.104632

343   15. Lefeuvre C, Pivert A, Przyrowski E, Bouthry E, Darviot E, Mahieu R, et al. Comparison of
344   performance between three SARS-CoV-2 molecular assays (Aptima™, Laboratory
345   Developed Test-Fusion, and R-GENE®) with special attention to turnaround time, a
346   key point in laboratory management. J Med Virol. 2022; 94(7): 3399-3403. doi:
347   10.1002/jmv.27675

348   16. https://covid-19.sante.gouv.fr/tests. covid-19.sante.gouv.fr.

Acknowlegdements

351   We thank Twist bioscience for the gift of the RNA controls and suggestions for the writing

352   of the paper.

Poitiers CHU (Seegene Kit)

Poitiers CHU (Thermofisher Kit)

Limoges CHU

fig 6. CIRAD